

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: WALTON COUNTY SO
Time of Inspection: 12:10

Date of Inspection: 01/30/2024

Serial Number: 80-007364
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202206B Exp: 06/14/2024	0.20g/210L Test (g/210L) Lot#:202206C Exp: 06/15/2024	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402370891 Exp: 02/23/2025
0.000	0.048	0.077	0.198	0.082
0.000	0.048	0.078	0.199	0.081
0.000	0.048	0.078	0.199	0.081

Number of Simulators Used: 5

Remarks:

A F / M A: Range Exceeded USER ERROR REPEAT CORRECTED - *Too much MA Solution used corrected*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

D/S William P May #F383

WILLIAM P MAY

Signature and Printed Name

01/30/2024
Date