

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HIGHLANDS COUNTY SO
Time of Inspection: 02:44

Date of Inspection: 09/25/2024

Serial Number: 80-007507
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 14323080A3 Exp: 09/05/2025
0.000	0.046	0.077	0.193	0.080
0.000	0.047	0.077	0.194	0.080
0.000	0.047	0.077	0.196	0.080


Number of Simulators Used: 5

Remarks:

Int Det: I repeated the Interferent test due to not putting Acetone into Simulator. I put in 5mil in waited 1/2 hr and repeated test. No Issues.

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

GREGORY S GLASCO

09/25/2024
Date

Gutschow, Taylor

From: Glasco, Gregory, S <gglasco@highlandssheriff.org>
Sent: Friday, September 27, 2024 6:21 PM
To: Gutschow, Taylor
Subject: Re: some issues

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

3 we have a 3 mil pipette. I may have said 5 by mistake but our pipette only holds 3.

From: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Sent: Friday, September 27, 2024 3:10 PM
To: Glasco, Gregory, S <gglasco@highlandssheriff.org>
Cc: Abeln, Shon W <sabeln@highlandssheriff.org>
Subject: RE: some issues

Did you put 3 mL of acetone in the Interferent simulator or 8 mL? If you put in 8 mL, that would invalidate the inspection(s) because the Form 39 procedure specifies how much is to be included.

I'll get the Form 40s added to our website once I receive them.

From: Glasco, Gregory, S <gglasco@highlandssheriff.org>
Sent: Wednesday, September 25, 2024 4:54 AM
To: Gutschow, Taylor <TaylorGutschow@fdle.state.fl.us>
Cc: Abeln, Shon W <sabeln@highlandssheriff.org>
Subject: some issues

CAUTION: This email originated outside of FDLE. Please use caution when opening attachments, clicking links, or responding to this email.

I did a very rookie mistake, I started to put back into service 007505 after getting it back from it's annual. I stopped it at the interferent section and aborted the inspection but said it was in compliance. Printed that #40 out and wrote the explanation on it then repeated the entire inspection successfully. What happen was that I did not realize that I failed to put acetone in the interferent simulator and did not catch it. So I figured there must be an error with the instrument I just received and started the second instruments inspection 0007507. I got to the interferent and nothing happened the first time around (no tone) so I realized what I had done and then put 8mil in the interferent simulator and waited a half house then repeated that section since it was still asking me to repeat.

Now since I then knew what had happened I proceeded without issue and completed the inspection successfully on both instruments,

So 007507 had a repeated interferent test and 007505 was completely started over. I wrote on the instruments 40 form what the issue was. Both form 40s have copies in an envelope for you in 007505 that

is on it's way to you soon. Well I took it over to my purchasing department on Wed the 25th labeled for it to be shipped out. I had a very bad night that night....lol

Under Florida Law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing. Our vision is a safer Highlands County where citizens and law enforcement are joined together, the rights of all persons are respected, and community values are reflected.

Under Florida Law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing. Our vision is a safer Highlands County where citizens and law enforcement are joined together, the rights of all persons are respected, and community values are reflected.

Under Florida Law, e-mail addresses are public records. If you do not want your e-mail address released in response to a public-records request, do not send electronic mail to this entity. Instead, contact this office by phone or in writing. Our vision is a safer Highlands County where citizens and law enforcement are joined together, the rights of all persons are respected, and community values are reflected.