

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 13:13

Date of Inspection: 07/11/2024

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Number of Simulators Used: _____

Remarks:

AI not conducted. Bypassed to bring out of Disabled Mode.
ML
7/11/24

Not determined ML
7/11/24

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

07/11/2024
Date

Flow Calibration Adjustment(s)

Performed by TDG

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
07/11/2024
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 7.141

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.832

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.855

Dependent Data Scale Factor = 100000 L/min





Independent Data Scale Factor = 256

Rounded Slope = 704

Rounded Intercept = -724774

Correlation = 0.99786

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<div>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2024 Software: 8100.27</div> <div>SN 80-000229</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:49</td></tr><tr><td>Control Test</td><td>0.048</td><td>13:50</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:50</td></tr><tr><td>Control Test</td><td>0.048</td><td>13:51</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:52</td></tr><tr><td>Control Test</td><td>0.048</td><td>13:52</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:53</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0480</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <div>Operator's Signature </div>	Test	g/210L	Time	Air Blank	0.000	13:49	Control Test	0.048	13:50	Air Blank	0.000	13:50	Control Test	0.048	13:51	Air Blank	0.000	13:52	Control Test	0.048	13:52	Air Blank	0.000	13:53	Control Test Stats			Average	0.0480		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<div>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2024 Software: 8100.27</div> <div>SN 80-000229</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:38</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:39</td></tr><tr><td>Control Test</td><td>0.079</td><td>13:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:40</td></tr><tr><td>Control Test</td><td>0.078</td><td>13:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:41</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0783</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7370</td><td></td></tr></tbody></table> <div>Operator's Signature </div>	Test	g/210L	Time	Air Blank	0.000	13:38	Control Test	0.078	13:38	Air Blank	0.000	13:39	Control Test	0.079	13:39	Air Blank	0.000	13:40	Control Test	0.078	13:41	Air Blank	0.000	13:41	Control Test Stats			Average	0.0783		Std Dev	0.0006		Rel Std Dev(%)	0.7370		<div>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2024 Software: 8100.27</div> <div>SN 80-000229</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>14:03</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:04</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:05</td></tr><tr><td>Control Test</td><td>0.197</td><td>14:05</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:06</td></tr><tr><td>Control Test</td><td>0.196</td><td>14:07</td></tr><tr><td>Air Blank</td><td>0.000</td><td>14:07</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1967</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.2936</td><td></td></tr></tbody></table> <div>Operator's Signature </div>	Test	g/210L	Time	Air Blank	0.000	14:03	Control Test	0.197	14:04	Air Blank	0.000	14:05	Control Test	0.197	14:05	Air Blank	0.000	14:06	Control Test	0.196	14:07	Air Blank	0.000	14:07	Control Test Stats			Average	0.1967		Std Dev	0.0006		Rel Std Dev(%)	0.2936		<div>FDLE Intoxilyzer - Alcohol Analyzer Model 8000 07/11/2024 Software: 8100.27</div> <div>SN 80-000229</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>13:43</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:43</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr><tr><td>Control Test</td><td>0.080</td><td>13:44</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr><tr><td>Control Test</td><td>0.081</td><td>13:45</td></tr><tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></tbody></table> <div>Operator's Signature </div>	Test	g/210L	Time	Air Blank	0.000	13:43	Control Test	0.080	13:43	Air Blank	0.000	13:44	Control Test	0.080	13:44	Air Blank	0.000	13:45	Control Test	0.081	13:45	Air Blank	0.000	13:45	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:49																																																																																																																																																	
Control Test	0.048	13:50																																																																																																																																																	
Air Blank	0.000	13:50																																																																																																																																																	
Control Test	0.048	13:51																																																																																																																																																	
Air Blank	0.000	13:52																																																																																																																																																	
Control Test	0.048	13:52																																																																																																																																																	
Air Blank	0.000	13:53																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0480																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:38																																																																																																																																																	
Control Test	0.078	13:38																																																																																																																																																	
Air Blank	0.000	13:39																																																																																																																																																	
Control Test	0.079	13:39																																																																																																																																																	
Air Blank	0.000	13:40																																																																																																																																																	
Control Test	0.078	13:41																																																																																																																																																	
Air Blank	0.000	13:41																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0783																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7370																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:03																																																																																																																																																	
Control Test	0.197	14:04																																																																																																																																																	
Air Blank	0.000	14:05																																																																																																																																																	
Control Test	0.197	14:05																																																																																																																																																	
Air Blank	0.000	14:06																																																																																																																																																	
Control Test	0.196	14:07																																																																																																																																																	
Air Blank	0.000	14:07																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1967																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2936																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	13:43																																																																																																																																																	
Control Test	0.080	13:43																																																																																																																																																	
Air Blank	0.000	13:44																																																																																																																																																	
Control Test	0.080	13:44																																																																																																																																																	
Air Blank	0.000	13:45																																																																																																																																																	
Control Test	0.081	13:45																																																																																																																																																	
Air Blank	0.000	13:45																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 12:42

Date of Inspection: 07/12/2024

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.047	0.078	0.196	0.080
0.000	0.047	0.078	0.196	0.080
0.000	0.047	0.078	0.196	0.081
0.000	0.047	0.078	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.047	0.078	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.047	0.078	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.047	0.078	0.196	0.080


Standard Deviations	0.0004	0.0000	0.0000	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0001 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.



I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

07/12/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000229, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000229</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>07/12/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:42</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.
*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).
The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.
This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

07/12/2024

Date
TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Stabilities MG

13:13 02/12/2024

065

#1

FDLE
Intoxilyzer - Alcohol
Model 8000
02/12/2024
Software: 8100.27

IN 80-000229

Test	g/2	Time
Air Blank	0.01	14:16
Control Test	0.08	14:16
Air Blank	0.01	14:17
Control Test	0.01	14:17
Air Blank	0.01	14:18
Control Test	0.08	14:18
Air Blank	0.01	14:18
Control Test Stats		
Average	0.08	
Std Dev	0.01	
Rel Std Dev(%)	1.28	

Operator's Sig

065

#2

FDLE
Intoxilyzer - Alcohol
Model 8000
02/12/2024
Software: 8100.27

IN 80-000229

Test	g/2	Time
Air Blank	0.01	14:19
Control Test	0.08	14:20
Air Blank	0.01	14:20
Control Test	0.08	14:20
Air Blank	0.01	14:21
Control Test	0.08	14:21
Air Blank	0.01	14:22
Control Test Stats		
Average	0.08	
Std Dev	0.01	
Rel Std Dev(%)	0.01	

Operator's Sig

Stability checks printed from the internal printer are not fully legible. Reseated printer paper between printouts 1 and 2 to no effect. Will print stabilities using external printer.

MG
02/12/2024

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000229

02/12/2024

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:30
Control Test	0.049	14:31
Air Blank	0.000	14:31
Control Test	0.048	14:32
Air Blank	0.000	14:33
Control Test	0.048	14:33
Air Blank	0.000	14:34
Control Test Stats		
Average	0.0483	
Std Dev	0.0006	
Rel Std Dev(%)	1.1945	



Operator's Signature

FDLE
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-000229
02/12/2024
Software: 8100.27

ARS

Test	g/210L	Time
Air Blank	0.000	14:37
Control Test	0.078	14:38
Air Blank	0.000	14:39
Control Test	0.078	14:39
Air Blank	0.000	14:40
Control Test	0.078	14:40
Air Blank	0.000	14:41
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

mg

Operator's Signature

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000229

02/12/2024

Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:44
Control Test	0.195	14:44
Air Blank	0.000	14:45
Control Test	0.196	14:46
Air Blank	0.000	14:46
Control Test	0.196	14:47
Air Blank	0.000	14:48
Control Test Stats		
Average	0.1957	
Std Dev	0.0006	
Rel Std Dev(%)	0.2951	

ML

Operator's Signature

FDLE

Intoxilyzer - Alcohol Analyzer

Model 8000

SN 80-000229

02/12/2024

Software: 8100.27

065

Test	g/210L	Time
Air Blank	0.000	14:23
Control Test	0.080	14:23
Air Blank	0.000	14:24
Control Test	0.080	14:24
Air Blank	0.000	14:25
Control Test	0.080	14:25
Air Blank	0.000	14:26
Control Test Stats		
Average	0.0800	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

MG

Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Time of Inspection: 12:48

Date of Inspection: 02/13/2024

Serial Number: 80-000229

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.078	0.195	0.079
0.000		0.078	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079


Standard Deviations	0.0000	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/13/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FDLE

Serial Number: 80-000229

Time of Inspection: 12:48

Date of Inspection: 02/13/2024

Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#: 202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#: 202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 01923080A3 Exp: 02/05/2025
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.077	0.196	0.079
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.078	0.196	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.080
0.000	0.048	0.078	0.195	0.079
0.000	0.048	0.078	0.195	0.079

Standard Deviations	0.0000	0.0004	0.0005	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply () with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

02/13/2024

Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-000229, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-000229</u>	UNCERTAINTY* \pm
Owning Agency:	<u>FDLE</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>02/13/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:48</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021
Issuing Authority: Alcohol Testing Program

02/13/2024 Date
Taylor D Gutschow
TAYLOR D GUTSCHOW,
Department Inspector

Service • Integrity • Respect • Quality

Return Material Authorization

Ship to:



CMI, Inc.



Enforcement Electronics

Shipment to repair facility authorized by: Shayla Platt on 2/13/2024

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-000229

Bill To Address:

Florida Department of Law Enforcement

Attn: Shayla Platt

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

When the instrument prints to the internal printer, a white line obscures a portion of the results.

While the instrument printed stability checks to the external printer without issue, printed Form

41s were missing a data point. See attached paperwork for above examples. Copies reprinted

from COBRA contained the missing data point.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Shayla Platt

Phone #: 850-728-4200

Email: ShaylaPlatt@fdle.state.fl.us

ATP Contact Name: Taylor Gutschow

ATP Email: TaylorGutschow@fdle.state.fl.us