



INSTRUMENT PROCESSING SHEET

Agency Davie PDS/N 80-001056Florida Department of
Law EnforcementDate In 09/06/2024 DI Completion Date 09/19/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>09/17/2024</u>		Quality Checks By <u>TDG</u> Date <u>09/17/2024</u>		Flow Calibration By _____ Date _____																																																													
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>213</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.167</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																														
Simulator	Serial #	Lot #/Exp																																																															
0.050	MP6286	202303K 03/29/2025																																																															
0.080	MP6287	202303L 03/29/2025																																																															
0.200	MP6288	202304C 04/05/2025																																																															
0.080 DGS	N/A	01923080A3 02/05/2025																																																															
				Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____																																																													
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																																																													
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1012</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																														
0.000		N/A	N/A																																																														
0.040																																																																	
0.100																																																																	
0.200																																																																	
0.300																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000	MP6284																																																																
Interferent	MP6285																																																																
0.050	MP6286																																																																
0.080	MP6287																																																																
0.200	MP6288																																																																
Notes/Suggested Service: <u>Added plastic caps for shipping to the agency. (TDG)</u> _____ _____ _____ _____ _____ _____				Attachments <table border="1" style="width:100%"><tr><td><input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____</td></tr></table> <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use		<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																										
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																																
Destinee <u>Armstrong</u> Digitally signed by Destinee Armstrong Date: 2024.09.24 12:38:07 -04'00'				Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.10.01 08:17:21 -04'00'																																																													
Tech Review / Date _____				Admin Review / Date _____																																																													

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
<div>DAUIE PD Intoxilyzer - Alcohol Analyzer Model 8000 09/17/2024 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:38</td></tr><tr><td>Control Test</td><td>0.051</td><td>12:39</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:39</td></tr><tr><td>Control Test</td><td>0.051</td><td>12:40</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:41</td></tr><tr><td>Control Test</td><td>0.051</td><td>12:41</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:42</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0510</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:38	Control Test	0.051	12:39	Air Blank	0.000	12:39	Control Test	0.051	12:40	Air Blank	0.000	12:41	Control Test	0.051	12:41	Air Blank	0.000	12:42	Control Test Stats			Average	0.0510		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<div>DAUIE PD Intoxilyzer - Alcohol Analyzer Model 8000 09/17/2024 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:45</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:46</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:46</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:47</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:48</td></tr><tr><td>Control Test</td><td>0.081</td><td>12:48</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:49</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0803</td><td></td></tr><tr><td>Std Dev</td><td>0.0006</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.7187</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:45	Control Test	0.080	12:46	Air Blank	0.000	12:46	Control Test	0.080	12:47	Air Blank	0.000	12:48	Control Test	0.081	12:48	Air Blank	0.000	12:49	Control Test Stats			Average	0.0803		Std Dev	0.0006		Rel Std Dev(%)	0.7187		<div>DAUIE PD Intoxilyzer - Alcohol Analyzer Model 8000 09/17/2024 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:53</td></tr><tr><td>Control Test</td><td>0.200</td><td>12:53</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:54</td></tr><tr><td>Control Test</td><td>0.199</td><td>12:54</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:55</td></tr><tr><td>Control Test</td><td>0.198</td><td>12:56</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:56</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.1990</td><td></td></tr><tr><td>Std Dev</td><td>0.0010</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.5025</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:53	Control Test	0.200	12:53	Air Blank	0.000	12:54	Control Test	0.199	12:54	Air Blank	0.000	12:55	Control Test	0.198	12:56	Air Blank	0.000	12:56	Control Test Stats			Average	0.1990		Std Dev	0.0010		Rel Std Dev(%)	0.5025		<div>DAUIE PD Intoxilyzer - Alcohol Analyzer Model 8000 09/17/2024 Software: 8100.27</div> <table><thead><tr><th>Test</th><th>g/210L</th><th>Time</th></tr></thead><tbody><tr><td>Air Blank</td><td>0.000</td><td>12:29</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:29</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:30</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:30</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr><tr><td>Control Test</td><td>0.080</td><td>12:31</td></tr><tr><td>Air Blank</td><td>0.000</td><td>12:31</td></tr><tr><td>Control Test Stats</td><td></td><td></td></tr><tr><td>Average</td><td>0.0800</td><td></td></tr><tr><td>Std Dev</td><td>0.0000</td><td></td></tr><tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr></tbody></table> <div>Operator's Signature</div>	Test	g/210L	Time	Air Blank	0.000	12:29	Control Test	0.080	12:29	Air Blank	0.000	12:30	Control Test	0.080	12:30	Air Blank	0.000	12:31	Control Test	0.080	12:31	Air Blank	0.000	12:31	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:38																																																																																																																																																	
Control Test	0.051	12:39																																																																																																																																																	
Air Blank	0.000	12:39																																																																																																																																																	
Control Test	0.051	12:40																																																																																																																																																	
Air Blank	0.000	12:41																																																																																																																																																	
Control Test	0.051	12:41																																																																																																																																																	
Air Blank	0.000	12:42																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0510																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:45																																																																																																																																																	
Control Test	0.080	12:46																																																																																																																																																	
Air Blank	0.000	12:46																																																																																																																																																	
Control Test	0.080	12:47																																																																																																																																																	
Air Blank	0.000	12:48																																																																																																																																																	
Control Test	0.081	12:48																																																																																																																																																	
Air Blank	0.000	12:49																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0803																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7187																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:53																																																																																																																																																	
Control Test	0.200	12:53																																																																																																																																																	
Air Blank	0.000	12:54																																																																																																																																																	
Control Test	0.199	12:54																																																																																																																																																	
Air Blank	0.000	12:55																																																																																																																																																	
Control Test	0.198	12:56																																																																																																																																																	
Air Blank	0.000	12:56																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1990																																																																																																																																																		
Std Dev	0.0010																																																																																																																																																		
Rel Std Dev(%)	0.5025																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	12:29																																																																																																																																																	
Control Test	0.080	12:29																																																																																																																																																	
Air Blank	0.000	12:30																																																																																																																																																	
Control Test	0.080	12:30																																																																																																																																																	
Air Blank	0.000	12:31																																																																																																																																																	
Control Test	0.080	12:31																																																																																																																																																	
Air Blank	0.000	12:31																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0800																																																																																																																																																		
Std Dev	0.0000																																																																																																																																																		
Rel Std Dev(%)	0.0000																																																																																																																																																		

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: DAVIE PD
Time of Inspection: 14:17

Date of Inspection: 09/19/2024

Serial Number: 80-001056
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.052	0.080	0.199	0.079
0.000	0.051	0.081	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.079
0.000	0.052	0.080	0.200	0.079
0.000	0.051	0.080	0.200	0.078
0.000	0.051	0.080	0.200	0.079
0.000	0.051	0.081	0.201	0.078
0.000	0.052	0.081	0.200	0.078
0.000	0.051	0.081	0.200	0.078

Standard Deviations	0.0004	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.


I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

09/19/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001056, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001056</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>DAVIE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>09/19/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>14:17</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

09/19/2024

Date



TAYLOR D GUTSCHOW,
Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality