

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: Orange City PD
Time of Inspection:

Date of Inspection: N/A

Serial Number: 80-001159
Software:

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK			Date and/or Time Adjusted		
Minimum Sample Volume Check: OK			Barometric Pressure Sensor Check: OK		
Alcohol Free Subject Test: 0.000			Mouth Alcohol Test: Slope Not Met		
Interferent Detect Test: Interferent Detect			Diagnostic Check (Post-Inspection): OK		

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: _____

Remarks:

As of 12/27/24, instrument is out for repair so unable to perform 2024 Department Inspection.

N/A - Compliance not determined 12/27/24 DA

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



 Signature and Printed Name

12/27/2024
Date



INSTRUMENT PROCESSING SHEET

Agency Orange City PDS/N 80-001159Florida Department of
Law EnforcementDate In 11/13/2024 DI Completion Date n/a☐ Ship ☐ P/U ☐ H/D ☐ CMI ☒ EE

Intake By <u>ALL</u> Date <u>11/13/2024</u>	Quality Checks By _____ Date _____	Flow Calibration By _____ Date _____																																																													
<div><input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE</div> <div>Visual Inspection: <div><input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle</div><div><input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf</div><div><input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube</div><div><input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight</div></div> <div>Other Equipment/ Accessories: <input checked="" type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable</div> <div>Notes: _____ _____ _____ _____ _____ _____ _____ _____</div>	<div><input type="checkbox"/> Breath Tube Screen <input type="checkbox"/> Replace External O-Rings <input type="checkbox"/> Instrument Set Up Verified <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Flow Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) <input type="checkbox"/> Barometric Pressure Check Gauge ID # _____ <input type="checkbox"/> Stability Checks</div> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr style="background-color: #d3d3d3;"><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050			0.080			0.200			0.080 DGS	N/A		<div>Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)</div> <div style="background-color: #d3d3d3; padding: 2px;">Maintenance By _____ Date _____</div> <div><input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____</div>																																														
Simulator	Serial #	Lot #/Exp																																																													
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Calibration Adjustment By _____	Department Inspection By _____																																																														
<div>Barometric Pressure Gauge _____ ID # _____</div> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr style="background-color: #d3d3d3;"><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> <div><input type="checkbox"/> Post Calibration Adjustment Stability Checks</div> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr style="background-color: #d3d3d3;"><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> <div>Notes/Suggested Service: <u>Instrument will not enter READY mode due to DSP fail. DA 11/18/2024</u> _____ _____ _____ _____ _____</div>	Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<div>Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____</div> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr style="background-color: #d3d3d3;"><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table> <div style="background-color: #d3d3d3; padding: 2px;">Attachments</div> <table border="1" style="width:100%; border-collapse: collapse;"><tbody><tr><td style="width:50%; vertical-align: top;"><input type="checkbox"/> Form 41 <input type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment</td><td style="width:50%; vertical-align: top;"><input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____</td></tr></tbody></table> <div style="background-color: #d3d3d3; padding: 2px;"><input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC</div> <div style="background-color: #d3d3d3; padding: 2px;"><input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use</div> <div style="background-color: #d3d3d3; padding: 2px;"><input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use</div> <div style="background-color: #d3d3d3; padding: 2px; display: flex; justify-content: space-between;"><div>Tech Review / Date _____</div><div>Admin Review / Date _____</div></div>	Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200		<input type="checkbox"/> Form 41 <input type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____
Simulator	Serial #	Lot #	Expiration																																																												
0.000		N/A	N/A																																																												
0.040																																																															
0.100																																																															
0.200																																																															
0.300																																																															
0.080 DGS	N/A																																																														
Simulator	Serial #	Lot #	Expiration																																																												
0.050																																																															
0.080																																																															
0.200																																																															
0.080 DGS	N/A																																																														
Simulator	Serial Number																																																														
0.000																																																															
Interferent																																																															
0.050																																																															
0.080																																																															
0.200																																																															
<input type="checkbox"/> Form 41 <input type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____																																																														

Return Material Authorization

Ship to: ☐ CMI, Inc.
☒ Enforcement Electronics

Shipment to repair facility authorized by: Brian Kibbe on 11/18/2024

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-001159

Bill To Address:

Orange City PD

ATTN: Brian Kibbe

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Department of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Rd

Tallahassee, FL 32304

Reason for Return:

Instrument will not enter READY MODE due to DSP fail.

Please choose one of the following options:

- ☐ 1. I _____, authorize all repairs.
- ☐ 2. I _____, authorize repairs up to \$_____.
- ☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Brian Kibbe

Phone #: 386-561-8223

Email: bkibbe@orangecityfl.gov

ATP Contact Name: Destinee Armstrong ATP Email: destinee.armstrong@fdle.state.fl.us