



INSTRUMENT PROCESSING SHEET

Agency Gulf Breeze PDS/N 80-001306Florida Department of
Law EnforcementDate In 05/08/2024 DI Completion Date 5/9/2024☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Intake	By <u>ALL</u>	Date <u>05/08/2024</u>	Quality Checks	By <u>DA</u>	Date <u>5/8/24</u>	Flow Calibration	By _____	Date _____																															
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☒ Power cord ☐ Printer Cable ☒ Static Bag ☐ 12V DC Cable Notes: AI reports no interferent detect during acetone test in agency inspections - removed from service. I bypassed AI to operate instrument DA 5/8/24 RFI detected within the stability check of the 0.05g/210L. 0.05g/210L stability check repeated DA 5/8/24			☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>236</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP105</u> 32 mm <u>0.156</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.246</u> (.228 - .278) 103 mm <u>0.503</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>28662</u> ☒ Stability Checks			☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																																	
			<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6291</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6292</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6293</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5 04/05/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K 03/29/2025	0.080	MP6292	202303L 03/29/2025	0.200	MP6293	202304C 04/05/2025	0.080 DGS	N/A	06723080A5 04/05/2025	<table border="1"><thead><tr><th colspan="2">Maintenance By _____ Date _____</th></tr><tr><td>☐ Battery Replacement</td><td></td></tr><tr><td>☐ Dry Gas Regulator Replacement</td><td></td></tr><tr><td>☐ Breath Tube Replacement</td><td></td></tr><tr><td>☐ Other _____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr><tr><td>_____</td><td></td></tr></thead></table>			Maintenance By _____ Date _____		☐ Battery Replacement		☐ Dry Gas Regulator Replacement		☐ Breath Tube Replacement		☐ Other _____		_____		_____		_____	
Simulator	Serial #	Lot #/Exp																																					
0.050	MP6291	202303K 03/29/2025																																					
0.080	MP6292	202303L 03/29/2025																																					
0.200	MP6293	202304C 04/05/2025																																					
0.080 DGS	N/A	06723080A5 04/05/2025																																					
Maintenance By _____ Date _____																																							
☐ Battery Replacement																																							
☐ Dry Gas Regulator Replacement																																							
☐ Breath Tube Replacement																																							
☐ Other _____																																							

Calibration Adjustment By _____			Department Inspection By <u>SP</u>																																				
Barometric Pressure Gauge _____ ID # _____			Barometric Pressure ID# <u>28663</u>																																				
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1011</u> Instrument <u>1009</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u>								
Simulator	Serial #	Lot #	Expiration																																				
0.000		N/A	N/A																																				
0.040																																							
0.100																																							
0.200																																							
0.300																																							
0.080 DGS	N/A																																						
☐ Post Calibration Adjustment Stability Checks			<table border="1"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP5086</td></tr><tr><td>Interferent</td><td>MP5087</td></tr><tr><td>0.050</td><td>MP5088</td></tr><tr><td>0.080</td><td>MP5089</td></tr><tr><td>0.200</td><td>MP5090</td></tr></tbody></table>						Simulator	Serial Number	0.000	MP5086	Interferent	MP5087	0.050	MP5088	0.080	MP5089	0.200	MP5090																			
Simulator	Serial Number																																						
0.000	MP5086																																						
Interferent	MP5087																																						
0.050	MP5088																																						
0.080	MP5089																																						
0.200	MP5090																																						
<table border="1"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1"><thead><tr><th colspan="2">Attachments</th></tr><tr><td>☒ Form 41</td><td>☐ Post-Stability Checks</td></tr><tr><td>☒ Stability Checks</td><td>☐ Flow Calibration</td></tr><tr><td>☒ Calibration Certificate</td><td>☒ Form 40</td></tr><tr><td>☐ Calibration Adjustment</td><td>☒ Other <u>INT Stabilities</u></td></tr></thead></table>						Attachments		☒ Form 41	☐ Post-Stability Checks	☒ Stability Checks	☐ Flow Calibration	☒ Calibration Certificate	☒ Form 40	☐ Calibration Adjustment	☒ Other <u>INT Stabilities</u>	
Simulator	Serial #	Lot #	Expiration																																				
0.050																																							
0.080																																							
0.200																																							
0.080 DGS	N/A																																						
Attachments																																							
☒ Form 41	☐ Post-Stability Checks																																						
☒ Stability Checks	☐ Flow Calibration																																						
☒ Calibration Certificate	☒ Form 40																																						
☐ Calibration Adjustment	☒ Other <u>INT Stabilities</u>																																						
Notes/Suggested Service: Agency reported issues with interferant detection. Conducted stabilities all flagged by instrument. Sent agency new bottle of acetone solution. SP.			☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use																																				
			Taylor Gutschow Digitally signed by Taylor Gutschow Date: 2024.05.09 15:55:36 -0400																																				
			Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.05.13 11:57:02 -0400																																				
			Tech Review / Date _____ Admin Review / Date _____																																				

Stability Checks 80-001306 5/8/24 DA

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
05/ 8/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	RFI*	13:09
Air Blank	0.000	13:09

*RFI Detect


Operator's Signature

RFI detected so I will repeat the 0.05g/210L
stability check. DA 5/8/24

0.05g/210L repeat DA 5/8/24

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
05/08/2024
Software: 8100.27

Test	g/210L	Time

Air Blank	0.000	13:10
Control Test	0.051	13:11
Air Blank	0.000	13:11
Control Test	0.050	13:12
Air Blank	0.000	13:13
Control Test	0.049	13:13
Air Blank	0.000	13:14
Control Test Stats		
Av	0.0500	
St	0.0010	
Re	Dev(%)	2.0000



Operator's Signature

GUL EEZE PD
Int yzer - Alcohol Analyzer
Mod 000 SN 80-001306
05/ 024
Soft ve: 8100.27

Test		g/210L	Time
Air	nk	0.000	13:18
Cor.	Test	0.079	13:19
Air	nk	0.000	13:20
Cor.	Test	0.079	13:20
Air	nk	0.000	13:21
Cor.	Test	0.079	13:21
Air	nk	0.000	13:22
Cor.	Test Stats		
A		0.0790	
St		0.0000	
R	Dev (%)	0.0000	


Operator's Signature

Stability Checks 80-001306 5/8/24 DA

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
05/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:24
Control Test	0.199	13:25
Air Blank	0.000	13:26
Control Test	0.198	13:26
Air Blank	0.000	13:27
Control Test	0.198	13:28
Air Blank	0.000	13:28
Control Test Stats		
Average	0.1983	
Std Dev	0.0006	
Rel Std Dev(%)	0.2911	



Operator's Signature

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
05/08/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	13:29
Control Test	0.078	13:30
Air Blank	0.000	13:30
Control Test	0.078	13:30
Air Blank	0.000	13:31
Control Test	0.078	13:31
Air Blank	0.000	13:32
Control Test Stats		
Average	0.0780	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

DGS



Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 12:57

Date of Inspection: 05/08/2024

Serial Number: 80-001306
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK		No
Alcohol Free Subject Test: 0.000		No
Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No
Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08g/210L Test (g/210L) Lot#: _____ Exp: _____	0.20g/210L Test (g/210L) Lot#: _____ Exp: _____	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: _____ Exp: _____

Number of Simulators Used: _____


Remarks:

BYPASS AI TO OPERATE INSTRUMENT COMPLIANCE NOT DETERMINED

Did not determine compliance DA 5/8/24

The above instrument complies () does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



DESTINEE N ARMSTRONG

Signature and Printed Name

05/08/2024
Date

GULF BREEZE PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001306
05/09/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	09:40
Control Test	INT*	09:41
Air Blank	0.000	09:41
Control Test	INT*	09:42
Air Blank	0.000	09:42
Control Test	INT*	09:43
Air Blank	0.000	09:44
Control Test	INT*	09:44
Air Blank	0.000	09:45
Control Test	INT*	09:46
Air Blank	0.000	09:46
Control Test	INT*	09:47
Air Blank	0.000	09:47
Control Test	INT*	09:48
Air Blank	0.000	09:49
Control Test	INT*	09:49
Air Blank	0.000	09:50
Control Test	INT*	09:51
Air Blank	0.000	09:51
Control Test	INT*	09:52
Air Blank	0.000	09:52
Control Test	INT*	09:53
Air Blank	0.000	09:54
Control Test	INT*	09:54
Air Blank	0.000	09:55
Control Test	INT*	09:55
Air Blank	0.000	09:56
Control Test	INT*	09:57
Air Blank	0.000	09:57
Control Test	INT*	09:58
Air Blank	0.000	09:58
Control Test	INT*	09:59
Air Blank	0.000	10:00
Control Test	INT*	10:00
Air Blank	0.000	10:01
Control Test	INT*	10:02
Air Blank	0.000	10:02
Control Test	INT*	10:03
Air Blank	0.000	10:03
Control Test	INT*	10:04
Air Blank	0.000	10:05
Control Test Stats		
Average	0.0000	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

*Interferent Detect

Int. Detect
Stabilities
SP 5/9/24

SP
Operator's Signature

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: GULF BREEZE PD
Time of Inspection: 12:16

Date of Inspection: 05/09/2024

Serial Number: 80-001306
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.050	0.079	0.197	0.077
0.000	0.050	0.079	0.197	0.077
0.000	0.049	0.080	0.198	0.077
0.000	0.050	0.079	0.197	0.077
0.000	0.050	0.079	0.198	0.078
0.000	0.050	0.080	0.197	0.077
0.000	0.050	0.079	0.197	0.078
0.000	0.050	0.080	0.198	0.078
0.000	0.050	0.080	0.197	0.078
0.000	0.050	0.080	0.197	0.077


Standard Deviations	0.0003	0.0005	0.0004	0.0005
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



SHAYLA D PLATT

Signature and Printed Name

05/09/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-001306, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-001306</u>	UNCERTAINTY* $\pm$	
Owning Agency:	<u>GULF BREEZE PD</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>05/09/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>12:16</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt

Digitally signed by Shayla Platt
Date: 2024.05.09 13:36:30 -04'00'

05/09/2024

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality

Page 1 of 1