



Agency	S/N
--------	-----

Date In _____ DI Completion Date _____ ☐ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

Calibration Adjustment	By	Department Inspection	By
------------------------	----	-----------------------	----

Barometric Pressure Gauge _____ ID # _____

❑ Post Calibration Adjustment Stability Checks

[illegible]Department Inspection ByBarometric Pressure ID# _____

Gauge _____ Instrument _____

Mouth Alcohol Solution Lot #Acetone Stock Solution Lot #

Attachments

- ☐ Instrument Complies with Chapter 11D-8, FAC
- ☐ Instrument Does Not Comply with Chapter 11D-8, FAC

- ☐ Return to/Place into Evidentiary Use
- ☐ Remain Out of Evidentiary Use

- ☐ Conduct an Agency Inspection Before Evidentiary Use

Tech Review / Date		Admin Review / Date	
--------------------	--	---------------------	--

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001443
02/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:27
Control Test	0.051	11:27
Air Blank	0.000	11:28
Control Test	0.050	11:29
Air Blank	0.000	11:29
Control Test	0.050	11:30
Air Blank	0.000	11:30
Control Test Stats		
Average	0.0503	
Std Dev	0.0006	
Rel Std Dev(%)	1.1471	

Operator's Signature *AZ*

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001443
02/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:32
Control Test	0.081	11:33
Air Blank	0.000	11:33
Control Test	0.081	11:34
Air Blank	0.000	11:34
Control Test	0.081	11:35
Air Blank	0.000	11:35
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Wet

Operator's Signature *AZ*

CLAY COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-001443
02/20/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	11:37
Control Test	0.200	11:38
Air Blank	0.000	11:38
Control Test	0.200	11:39
Air Blank	0.000	11:39
Control Test	0.199	11:40
Air Blank	0.000	11:41
Control Test Stats		
Average	0.1997	
Std Dev	0.0006	
Rel Std Dev(%)	0.2892	

Operator's Signature *AZ*

Test	g/210L	Time
Air Blank	0.000	11:44
Control Test	0.081	11:44
Air Blank	0.000	11:45
Control Test	0.081	11:45
Air Blank	0.000	11:45
Control Test	0.081	11:46
Air Blank	0.000	11:46
Control Test Stats		
Average	0.0810	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

D65

Operator's Signature *AZ*