



INSTRUMENT PROCESSING SHEET

Agency Florida Highway Patrol, Troop KS/N 80-003405Florida Department of
Law EnforcementDate In 12/09/2024DI Completion Date 12/10/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By <u>TDG</u> Date <u>12/09/2024</u>		Quality Checks By <u>TDG</u> Date <u>12/09/2024</u>		Flow Calibration By _____ Date _____																													
☒ Annual ☐ Registration ☐ Return from CMI / EE Visual Inspection: ☒ Case ☒ Handle ☒ Keyboard ☒ Dry Gas Shelf ☒ Feet ☒ Breath Tube ☒ Ports ☒ Screws Tight Other Equipment/ Accessories: ☐ Power cord ☐ Printer Cable ☐ Static Bag ☐ 12V DC Cable Notes: <u>Dropped off by Armando</u> <u>Rovirosa. Agency keyboard not</u> <u>working. Used FDLE keyboard.</u>		☒ Breath Tube Screen ☒ Replace External O-Rings ☒ Instrument Set Up Verified ☒ R-Value <u>170</u> ☒ Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.152</u> (.139 - .169) 36 mm <u>0.171</u> (.156 - .190) 53 mm <u>0.238</u> (.228 - .278) 103 mm <u>0.507</u> (.447 - .547) ☒ Barometric Pressure Check Gauge ID # <u>26932</u> ☒ Stability Checks		☐ Flow Column # _____ ☐ 5L/min - 17mm ☐ 15L/min - 53mm ☐ 30L/min - 103mm ☐ R-Value _____ ☐ Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)																													
		<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5099</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP5100</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP5101</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG429602 10/22/2026</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP5099	202303K 03/29/2025	0.080	MP5100	202303L 03/29/2025	0.200	MP5101	202304C 04/05/2025	0.080 DGS	N/A	AG429602 10/22/2026	Maintenance By _____ Date _____ ☐ Battery Replacement ☐ Dry Gas Regulator Replacement ☐ Breath Tube Replacement ☐ Other _____														
Simulator	Serial #	Lot #/Exp																															
0.050	MP5099	202303K 03/29/2025																															
0.080	MP5100	202303L 03/29/2025																															
0.200	MP5101	202304C 04/05/2025																															
0.080 DGS	N/A	AG429602 10/22/2026																															
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																													
Barometric Pressure Gauge _____ ID # _____				Barometric Pressure ID# <u>26932</u>																													
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Gauge <u>1020</u> Instrument <u>1016</u> Mouth Alcohol Solution Lot # <u>2024-A</u> Acetone Stock Solution Lot # <u>2023-B</u>	
Simulator	Serial #	Lot #	Expiration																														
0.000		N/A	N/A																														
0.040																																	
0.100																																	
0.200																																	
0.300																																	
0.080 DGS	N/A																																
☐ Post Calibration Adjustment Stability Checks				<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288																
Simulator	Serial Number																																
0.000	MP6284																																
Interferent	MP6285																																
0.050	MP6286																																
0.080	MP6287																																
0.200	MP6288																																
<table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Attachments ☒ Form 41 ☒ Stability Checks ☒ Calibration Certificate ☐ Calibration Adjustment ☐ Post-Stability Checks ☐ Flow Calibration ☐ Form 40 ☐ Other _____									
Simulator	Serial #	Lot #	Expiration																														
0.050																																	
0.080																																	
0.200																																	
0.080 DGS	N/A																																
Notes/Suggested Service: _____ _____ _____ _____ _____ _____				☒ Instrument Complies with Chapter 11D-8, FAC ☐ Instrument Does Not Comply with Chapter 11D-8, FAC ☒ Return to/Place into Evidentiary Use ☐ Remain Out of Evidentiary Use ☒ Conduct an Agency Inspection Before Evidentiary Use Digitally signed by Shayla Platt Date: 2024.12.12 13:43:00 -05'00'																													
				Destinee Armstrong Tech Review / Date _____ Shayla Platt Admin Review / Date _____																													

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																																
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																																
✓	✓	✓	✓																																																																																																																																																
FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 12/09/2024 Software: 8100.27 SN 80-003405 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:15</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:16</td></tr> <tr><td>Control Test</td><td>0.050</td><td>14:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:17</td></tr> <tr><td>Control Test</td><td>0.051</td><td>14:18</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:18</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:15	Control Test	0.050	14:15	Air Blank	0.000	14:16	Control Test	0.050	14:16	Air Blank	0.000	14:17	Control Test	0.051	14:18	Air Blank	0.000	14:18	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 12/09/2024 Software: 8100.27 SN 80-003405 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:08</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:09</td></tr> <tr><td>Control Test</td><td>0.079</td><td>14:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:10</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:11</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:08	Control Test	0.079	14:08	Air Blank	0.000	14:09	Control Test	0.079	14:09	Air Blank	0.000	14:10	Control Test	0.080	14:11	Air Blank	0.000	14:11	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 12/09/2024 Software: 8100.27 SN 80-003405 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:24</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:25</td></tr> <tr><td>Control Test</td><td>0.200</td><td>14:26</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:26</td></tr> <tr><td>Control Test</td><td>0.199</td><td>14:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:28</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1993</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2896</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:24	Control Test	0.199	14:25	Air Blank	0.000	14:25	Control Test	0.200	14:26	Air Blank	0.000	14:26	Control Test	0.199	14:27	Air Blank	0.000	14:28	Control Test Stats			Average	0.1993		Std Dev	0.0006		Rel Std Dev(%)	0.2896		13:21 12/09/2024 FL HIGHWAY PATROL Intoxilyzer - Alcohol Analyzer Model 8000 12/09/2024 Software: 8100.27 SN 80-003405 <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>14:01</td></tr> <tr><td>Control Test</td><td>0.080</td><td>14:02</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:02</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:03</td></tr> <tr><td>Control Test</td><td>0.081</td><td>14:03</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>14:04</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0807</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7157</td><td></td></tr> </tbody> </table>	Test	g/210L	Time	Air Blank	0.000	14:01	Control Test	0.080	14:02	Air Blank	0.000	14:02	Control Test	0.081	14:03	Air Blank	0.000	14:03	Control Test	0.081	14:03	Air Blank	0.000	14:04	Control Test Stats			Average	0.0807		Std Dev	0.0006		Rel Std Dev(%)	0.7157	
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:15																																																																																																																																																	
Control Test	0.050	14:15																																																																																																																																																	
Air Blank	0.000	14:16																																																																																																																																																	
Control Test	0.050	14:16																																																																																																																																																	
Air Blank	0.000	14:17																																																																																																																																																	
Control Test	0.051	14:18																																																																																																																																																	
Air Blank	0.000	14:18																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0503																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	1.1471																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:08																																																																																																																																																	
Control Test	0.079	14:08																																																																																																																																																	
Air Blank	0.000	14:09																																																																																																																																																	
Control Test	0.079	14:09																																																																																																																																																	
Air Blank	0.000	14:10																																																																																																																																																	
Control Test	0.080	14:11																																																																																																																																																	
Air Blank	0.000	14:11																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0793																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7277																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:24																																																																																																																																																	
Control Test	0.199	14:25																																																																																																																																																	
Air Blank	0.000	14:25																																																																																																																																																	
Control Test	0.200	14:26																																																																																																																																																	
Air Blank	0.000	14:26																																																																																																																																																	
Control Test	0.199	14:27																																																																																																																																																	
Air Blank	0.000	14:28																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.1993																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.2896																																																																																																																																																		
Test	g/210L	Time																																																																																																																																																	
Air Blank	0.000	14:01																																																																																																																																																	
Control Test	0.080	14:02																																																																																																																																																	
Air Blank	0.000	14:02																																																																																																																																																	
Control Test	0.081	14:03																																																																																																																																																	
Air Blank	0.000	14:03																																																																																																																																																	
Control Test	0.081	14:03																																																																																																																																																	
Air Blank	0.000	14:04																																																																																																																																																	
Control Test Stats																																																																																																																																																			
Average	0.0807																																																																																																																																																		
Std Dev	0.0006																																																																																																																																																		
Rel Std Dev(%)	0.7157																																																																																																																																																		
Operator's Signature	Operator's Signature	Operator's Signature	Operator's Signature																																																																																																																																																

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FL HIGHWAY PATROL
Time of Inspection: 11:43

Date of Inspection: 12/10/2024

Serial Number: 80-003405
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG429602 Exp: 10/22/2026
0.000	0.049	0.079	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.079	0.199	0.080
0.000	0.050	0.080	0.199	0.080
0.000	0.050	0.079	0.199	0.079
0.000	0.050	0.079	0.198	0.080
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.079	0.199	0.080


Standard Deviations	0.0003	0.0004	0.0005	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

12/10/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-003405, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-003405</u>	UNCERTAINTY* $\pm$
Owning Agency:	<u>FL HIGHWAY PATROL</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>12/10/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>11:43</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

12/10/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality