



Agency Osceola County SO

S/N 80-004587

Date In 06/25/2024 DI Completion Date 6/29/2024

☒ Ship ☐ P/U ☐ H/D ☐ CMI ☐ EE

SP

6/29/24

Intake By <u>ALL</u> Date <u>06/25/2024</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input checked="" type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____ _____ _____ _____ 	Quality Checks By <u>ALL</u> Date <u>06/25/2024</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>233</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP-103</u> 32 mm <u>.152</u> (.139 - .169) 36 mm <u>.160</u> (.156 - .190) 53 mm <u>.218</u> (.228 - .278) 103 mm <u>.476</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28662</u> <input checked="" type="checkbox"/> Stability Checks	Flow Calibration By <u>6/29/24</u> Date <u>SP- 7/1/24</u> Flow Column # <u>ATP102</u> <input checked="" type="checkbox"/> 5L/min – 17mm <input checked="" type="checkbox"/> 15L/min – 53mm <input checked="" type="checkbox"/> 30L/min – 103mm <input checked="" type="checkbox"/> R-Value <u>239</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP103</u> 32 mm <u>.148</u> (.139 - .169) 36 mm <u>.164</u> (.156 - .190) 53 mm <u>.238</u> (.228 - .278) 103 mm <u>.503</u> (.447 - .547)															
	<table border="1"> <thead> <tr> <th>Simulator</th> <th>Serial #</th> <th>Lot #/Exp</th> </tr> </thead> <tbody> <tr> <td>0.050</td> <td>MP6291</td> <td>202303K 03/29/2025</td> </tr> <tr> <td>0.080</td> <td>MP6292</td> <td>202303L 03/29/2025</td> </tr> <tr> <td>0.200</td> <td>MP6293</td> <td>202304C 04/05/2025</td> </tr> <tr> <td>0.080 DGS</td> <td>N/A</td> <td>AG310901 04/19/2025</td> </tr> </tbody> </table>	Simulator	Serial #	Lot #/Exp	0.050	MP6291	202303K 03/29/2025	0.080	MP6292	202303L 03/29/2025	0.200	MP6293	202304C 04/05/2025	0.080 DGS	N/A	AG310901 04/19/2025	Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____ _____
Simulator	Serial #	Lot #/Exp															
0.050	MP6291	202303K 03/29/2025															
0.080	MP6292	202303L 03/29/2025															
0.200	MP6293	202304C 04/05/2025															
0.080 DGS	N/A	AG310901 04/19/2025															

Calibration Adjustment				By _____	
Barometric Pressure Gauge _____				ID # _____	
Simulator	Serial #	Lot #	Expiration		
0.000		N/A	N/A		
0.040					
0.100					
0.200					
0.300					
0.080 DGS	N/A				
<input type="checkbox"/> Post Calibration Adjustment Stability Checks					
Simulator	Serial #	Lot #	Expiration		
0.050					
0.080					
0.200					
0.080 DGS	N/A				
Notes/Suggested Service: _____ Tech review: corrected date and initial placement. SP 7/1/24 _____ _____ _____ _____ _____ _____					

Department Inspection		By <u>SP</u>	
Barometric Pressure ID# <u>28662</u>			
Gauge <u>1016</u>	Instrument <u>1015</u>		
Mouth Alcohol Solution Lot # <u>2024-A</u>			
Acetone Stock Solution Lot # <u>2023-B</u>			
Simulator	Serial Number		
0.000	MP5086		
Interferent	MP5087		
0.050	MP5088		
0.080	MP5089		
0.200	MP5090		
Attachments			
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment		<input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input type="checkbox"/> Other _____	
<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC			
<input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use			
<input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use			
Taylor Gutschow <small>Digitally signed by Taylor Gutschow Date: 2024.07.01 11:19:22 -04'00'</small>		Phil Nicodemo <small>Digitally signed by Phil Nicodemo Date: 2024.07.12 10:30:10 -04'00'</small>	
Tech Review / Date		Admin Review / Date	

Stability checks 80-004587 06/25/24

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
06/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:56
Control Test	0.051	14:57
Air Blank	0.000	14:57
Control Test	0.050	14:58
Air Blank	0.000	14:59
Control Test	0.051	14:59
Air Blank	0.000	15:00
Control Test Stats		
Average	0.0507	
Std Dev	0.0006	
Rel Std Dev(%)	1.1395	

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
06/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:03
Control Test	0.082	15:03
Air Blank	0.000	15:04
Control Test	0.081	15:05
Air Blank	0.000	15:05
Control Test	0.080	15:06
Air Blank	0.000	15:07
Control Test Stats		
Average	0.0810	
Std Dev	0.0010	
Rel Std Dev(%)	1.2346	

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
06/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:08
Control Test	0.202	15:09
Air Blank	0.000	15:10
Control Test	0.200	15:10
Air Blank	0.000	15:11
Control Test	0.199	15:12
Air Blank	0.000	15:12
Control Test Stats		
Average	0.2003	
Std Dev	0.0015	
Rel Std Dev(%)	0.7625	

Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
06/25/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	15:15
Control Test	0.079	15:15
Air Blank	0.000	15:15
Control Test	0.079	15:16
Air Blank	0.000	15:16
Control Test	0.079	15:16
Air Blank	0.000	15:17
Control Test Stats		
Average	0.0790	
Std Dev	0.0000	
Rel Std Dev(%)	0.0000	

Operator's Signature

FLOW
CAL
ADJUST

SP

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
06/29/2024
Software: 8100.27

Flow Rate Calibration*****
1: Rate (Liters/min) = 5
SQRT(Diff)) = 7.000
2: Rate (Liters/min) = 15
SQRT(Diff)) = 11.355
3: Rate (Liters/min) = 30
SQRT(Diff)) = 20.711
Dependent Data Scale Factor = 100000 L/min
Independent Data Scale Factor = 256
Rounded Slope = 699
Rounded Intercept = -662227
Correlation = 0.99572

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 10:41

Date of Inspection: 06/29/2024

Serial Number: 80-004587
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:AG310901 Exp: 04/19/2025
0.000	0.050	0.081	0.202	0.078
0.000	0.049	0.080	0.202	0.078
0.000	0.049	0.081	0.202	0.077
0.000	0.050	0.081	0.202	0.077
0.000	0.049	0.081	0.202	0.077
0.000	0.049	0.080	0.202	0.077
0.000	0.049	0.081	0.202	0.077
0.000	0.050	0.081	0.203	0.077
0.000	0.050	0.081	0.202	0.078
0.000	0.049	0.081	0.202	0.077

Standard Deviations	0.0005	0.0004	0.0003	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0004 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Shayla Platt
Signature and Printed Name

SHAYLA D PLATT

06/29/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
2331 Phillips Road.
Suite B1032
Tallahassee, FL 32308

This is to certify the calibration of Intoxilyzer 8000 serial number 80-004587, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-004587</u>	UNCERTAINTY* \pm	
Owning Agency:	<u>OSCEOLA COUNTY SO</u>	0.050 g/ 210 L	0.004
Calibration Date:	<u>06/29/2024</u>	0.080 g/ 210 L	0.004
Calibration Time:	<u>10:41</u>	0.200 g/ 210 L	0.007
		0.080 g/ 210 L Dry Gas Control	0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards.

This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

Shayla Platt
Digitally signed by Shayla Platt
Date: 2024.06.29 12:12:55 -04'00'

06/29/2024

Date

SHAYLA D PLATT,
Department Inspector

FDLE/ATP Form 69 March 2022
Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality



INSTRUMENT PROCESSING SHEET

Agency Osceola County SOS/N 80-004587Florida Department of
Law EnforcementDate In 01-10-2024 DI Completion Date N/A☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake	By <u>ALL</u>	Date <u>01-10-2024</u>	Quality Checks	By <u>BS</u>	Date <u>1/23/2024</u>	Flow Calibration	By _____	Date _____																																																									
<div><input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE</div> <div>Visual Inspection: <div><input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle</div><div><input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf</div><div><input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube</div><div><input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight</div></div> <div>Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input checked="" type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable</div> <div>Notes: _____ _____ _____ _____ _____ _____ _____</div>			<div><input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>221</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP102</u> 32 mm <u>0.148</u> (.139 - .169) 36 mm <u>0.164</u> (.156 - .190) 53 mm <u>0.234</u> (.228 - .278) 103 mm <u>0.488</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>28421</u> <input checked="" type="checkbox"/> Stability Checks</div> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP5088</td><td>202303K 3/29/2025</td></tr><tr><td>0.080</td><td>MP5089</td><td>202303L 3/29/2025</td></tr><tr><td>0.200</td><td>MP5090</td><td>202304C 4/5/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG310901 4/19/2025</td></tr></tbody></table>			Simulator	Serial #	Lot #/Exp	0.050	MP5088	202303K 3/29/2025	0.080	MP5089	202303L 3/29/2025	0.200	MP5090	202304C 4/5/2025	0.080 DGS	N/A	AG310901 4/19/2025	<div>Flow Column # _____ <input type="checkbox"/> 5L/min – 17mm <input type="checkbox"/> 15L/min – 53mm <input type="checkbox"/> 30L/min – 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547)</div> <div>Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____</div>																																												
						Simulator	Serial #	Lot #/Exp																																																									
						0.050	MP5088	202303K 3/29/2025																																																									
						0.080	MP5089	202303L 3/29/2025																																																									
0.200	MP5090	202304C 4/5/2025																																																															
0.080 DGS	N/A	AG310901 4/19/2025																																																															
Calibration Adjustment			Department Inspection																																																														
Barometric Pressure Gauge <u>1019/1018</u> ID # <u>30793/30793</u>			Barometric Pressure ID# _____ Gauge _____ Instrument _____ Mouth Alcohol Solution Lot # _____ Acetone Stock Solution Lot # _____																																																														
<table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5091</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5082</td><td>22460</td><td>12/28/2024</td></tr><tr><td>0.100</td><td>MP5083</td><td>23390</td><td>10/17/2025</td></tr><tr><td>0.200</td><td>MP5084</td><td>23340</td><td>9/18/2025</td></tr><tr><td>0.300</td><td>MP5085</td><td>23070</td><td>3/6/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>06723080A5</td><td>4/5/2025</td></tr></tbody></table> <div><input type="checkbox"/> Post Calibration Adjustment Stability Checks</div> <table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>			Simulator	Serial #	Lot #	Expiration	0.000	MP5091	N/A	N/A	0.040	MP5082	22460	12/28/2024	0.100	MP5083	23390	10/17/2025	0.200	MP5084	23340	9/18/2025	0.300	MP5085	23070	3/6/2025	0.080 DGS	N/A	06723080A5	4/5/2025	Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			<table border="1" style="width:100%; border-collapse: collapse;"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td></td></tr><tr><td>Interferent</td><td></td></tr><tr><td>0.050</td><td></td></tr><tr><td>0.080</td><td></td></tr><tr><td>0.200</td><td></td></tr></tbody></table> <div>Attachments</div> <div><input type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Calibration Certificate <input checked="" type="checkbox"/> Calibration Adjustment x2</div> <div><input type="checkbox"/> Post-Stability Checks <input type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u></div>			Simulator	Serial Number	0.000		Interferent		0.050		0.080		0.200	
Simulator	Serial #	Lot #	Expiration																																																														
0.000	MP5091	N/A	N/A																																																														
0.040	MP5082	22460	12/28/2024																																																														
0.100	MP5083	23390	10/17/2025																																																														
0.200	MP5084	23340	9/18/2025																																																														
0.300	MP5085	23070	3/6/2025																																																														
0.080 DGS	N/A	06723080A5	4/5/2025																																																														
Simulator	Serial #	Lot #	Expiration																																																														
0.050																																																																	
0.080																																																																	
0.200																																																																	
0.080 DGS	N/A																																																																
Simulator	Serial Number																																																																
0.000																																																																	
Interferent																																																																	
0.050																																																																	
0.080																																																																	
0.200																																																																	
<div>Notes/Suggested Service: <u>Stability checks had ambient, purge, and interferent messages. Need to perform an optical bench calibration adjustment.</u> <u>Performed calibration adjustment 1/25/2024, faults detected at 0.100. Stopped adjustment, confirmed MP5083 had no leaks, and performed a new calibration adjustment with same serial #s, lot #s, and expiration dates. Still detecting fault, calibration adjustment failed.</u> <u>Sending instrument to repair. (BS 1/25/2024)</u></div>			<div><input type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC</div> <div><input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use</div> <div><input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use</div>																																																														
						Tech Review / Date _____ Admin Review / Date _____																																																											

Stability Checks

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
01/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	0.000	14:30
Control Test	INT*	14:31
Air Blank	AMB**	14:32
Air Blank	AMB**	14:32

*Interferent Detect
**Ambient Fail

Benjamin S. Sibley
Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
01/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	14:34
Air Blank	AMB*	14:34

*Ambient Fail

Benjamin S. Sibley
Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
01/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	14:37
Air Blank	PUR**	14:38

*Ambient Fail
**Purge Fail

Benjamin S. Sibley
Operator's Signature

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-004587
01/23/2024
Software: 8100.27

Test	g/210L	Time
Air Blank	AMB*	14:36
Air Blank	AMB*	14:36

*Ambient Fail

DGS

Benjamin S. Sibley
Operator's Signature

OSCEOLA COUNTY, SO
Intoxilyzer - Alcohol Analyzer
Model 8000
SN 80-004587
12:05:36
01/25/2024
Auto Calibration
Max Power Res Value = 86
Auto Range Res Value = 63

```

<<<< CHANNEL 2 >>>>
Sample      (% Abs
Sample #1 = 1.5510  (0.0000
Sample #2 = 1.5580  (0.0200
Sample #3 = 1.5470  (0.0200
Sample #4 = 1.5360  (0.0500
Aug % Abs = 1.5450  (0.0313)
STD DEV = 0.0141  (0.0163)
REL STD DEV = 0.913  (51.986)

```

```

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/L %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12825, Sum Io = 13622
<<<< CHANNEL 1 >>>>
Sample      & Abs      (% Abs Re
Sample #1 = 0.0620 (-0.0600)
Sample #2 = 0.3250 (-0.1730)
Sample #3 = 0.1750 (-0.0390)
Sample #4 = -0.1820 (-0.0800)
Avg % Abs = 0.1063 (-0.0973)
STD DEV = 0.2506 (0.0687)
REL STD DEV = 245.160 (70.542)

```

```

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/L ****
Samples Taken = 4, Discarded = 1
Sum Io = 12786, Sum Io = 13618
<<<< CHANNEL 1 >>>>

Sample      % Abs      (% Abs Ref)
Sample #1 = 0.4160      (0.4120)
Sample #2 = 1.6510      (-0.0760)
Sample #3 = 1.3060      (-0.0690)
Sample #4 = 0.8910      (-0.0060)
Avg % Abs = 1.0593      (-0.0503)
STD DEV = 0.2475      (0.0386)
REL STD DEV = 23.364      (76.955)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 0.2110      (-0.0130)
Sample #2 = 0.1890      (0.0090)
Sample #3 = 0.1520      (0.0310)
Sample #4 = 0.1880      (0.0170)
Avg. % Abs = 0.1763      (0.0190)
STD Dev = 0.0211      (0.0111)
REL STD DEV = 11.954      (58.608)

```

```

Sol. Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l 222%
Samples Taken = 4, Discarded = 1
3um Io = 12761, 9um Io = 13619
<<<< CHANNEL 1 >>>>
Sample      % Abs Re
Sample #1 = 0.7170 (-0.0620)
Sample #2 = 1.0680 (-0.3450)
Sample #3 = 0.9470 (-0.3600)
Sample #4 = 0.4680 (0.0380)
Avg % Abs = 0.8277 (-0.2227)
STD DEV = 0.3173 (0.2259)
REL STD DEV = 36.337 (101.431)

```

Sample	% Abs	(% Abs Ref)
Sample #1	= 0.7170	(-0.0620)
Sample #2	= 1.0680	(-0.3450)
Sample #3	= 0.9470	(-0.3600)
Sample #4	= 0.4680	(-0.0380)
Avg % Abs	= 0.8277	(-0.2227)
STD DEV	= 0.3173	(0.2259)
REL STD DEV	= 38.337	(101.431)

```
<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.5510      (0.0070)
Sample #2 = 1.5580      (0.0200)
Sample #3 = 1.5470      (0.0240)
Sample #4 = 1.5300      (0.0500)
Avg % Abs = 1.5450      (0.0313)
STD DEV = 0.0141      (0.0163)
REL STD DEV = 0.913      (51.986)
```

```
Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l 222%
Samples Taken = 4, discarded = 1
Sum io = 12672, Sum lo = 13613
<<<< CHANNEL I >>>>
Sample      % Abs      (% Abs Ref
Sample #1 = 0.8400      (-0.3490)
Sample #2 = 1.4610      (-1.6740)
Sample #3 = 1.6550      (-1.9380)
Sample #4 = 2.0160      (-2.2390)
Avg % Abs = 1.7140      (-1.9503)
STD DEV = 0.2807 (0.2827)
REL STD DEV = 16.378 (14.495)
```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.5660  (0.0110)
Sample #2 = 3.5360  (0.0410)
Sample #3 = 3.5260  (0.0410)
Sample #4 = 3.5570  (0.0270)
Avg % Abs = 3.5397 (0.0363)
STD DEV = 0.0158 (0.0091)
REL STD DEV = 0.447 (22.247)

```

```

Sol Value = 0.100 g/210L ***
Fit Value = 0.4762 mg/l 222%
Samples Taken = 4, Discarded = 1
Sum Io = 12924, 9um Io = 13613
<<<< CHANNEL I >>>>
Sample      (% Abs R
Sample #1 = 1.6290 (-0.2510)
Sample #2 = 2.1090 (-0.0860)
Sample #3 = 1.7550 (0.4340)
Sample #4 = 2.0800 (-0.5050)
Avg % Abs = 1.9813 (-0.0530)
STD DEV = 0.1965 (0.4705)
REL STD DEV = 9.920 (887.693)

```

```

Sample      % Abs      (% Abs Ref)
Sample #1 = 1.6290      (-0.2500)
Sample #2 = 2.1090      (-0.0680)
Sample #3 = 1.7550      (-0.4340)
Sample #4 = 2.0800      (-0.5150)
Avg % Abs = 1.9813      (-0.0520)
STD DEV = 0.1965      (-0.4705)
REL STD DEV = 9.920      (887.693)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs
Sample #1 = 3.5500 (% Abs Ref)
Sample #2 = 3.5430 (0.0000)
Sample #3 = 3.5340 (0.0190)
Sample #4 = 3.5340 (0.0430)
Sample #5 = 3.5390 (0.0310)
Avg % Abs = 3.5387 (0.0310)
STD DEV = 0.0045 (0.0120)
REL STD DEV = 0.127 (38.710)

```

```

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/L %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12888, Sum Io = 13613
<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.2650 (-0.1980)
Sample #2 = 1.8230 (-0.0170)
Sample #3 = 1.9980 (-0.1200)
Sample #4 = 2.2500 (-0.0890)
Avg % Abs = 2.0237 (-0.0307)
STD DEV = 0.2147 (-0.0538)
REL STD DEV = 10.607 (175.432)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.5830   (-0.0210)
Sample #2 = 3.5440   (0.0350)
Sample #3 = 3.5440   (0.0350)
Sample #4 = 3.4960   (0.0540)
Avg % Abs = 3.5280   (0.0407)
STD DEV = 0.0277   (0.0116)
REL STD DEV = 0.786   (28.501)

```

**** AUTO CAL FAIL
Benjamin S. Soble

Optical Bench Calibration Adjustment 2

OSCEOLA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
14:23:06
10/25/2024
Auto Calibration
Max Power Res Value = 86
Auto Range Res Value = 63

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5710 (-0.0080)
Sample #2 = 1.5540 (0.0100)
Sample #3 = 1.5650 (0.0060)
Sample #4 = 1.5340 (0.0340)
Avg % Abs = 1.5510 (0.0167)
STD DEV = 0.0157 (0.0151)
REL STD DEV = 1.013 (90.863)

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12746, Sum Io = 13618
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = -0.4080 (-0.2580)
Sample #2 = -0.5320 (0.1360)
Sample #3 = 0.5440 (-1.0870)
Sample #4 = -0.0530 (-0.9920)
Avg % Abs = -0.0137 (-0.6477)
STD DEV = 0.5391 (0.6003)
REL STD DEV = 3944.473 (105.044)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 0.1750 (-0.0120)
Sample #2 = 0.1780 (-0.0150)
Sample #3 = 0.1580 (0.0100)
Sample #4 = 0.1560 (-0.0020)
Avg % Abs = 0.1640 (-0.0023)
STD DEV = 0.0122 (0.0125)
REL STD DEV = 7.418 (535.857)

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12884, Sum Io = 13616
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.7810 (-0.0240)
Sample #2 = 0.4120 (-0.0870)
Sample #3 = 1.0240 (-0.0180)
Sample #4 = 1.5480 (0.2290)
Avg % Abs = 0.9947 (0.0413)
STD DEV = 0.5686 (0.1661)
REL STD DEV = 57.162 (401.965)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5510 (-0.0090)
Sample #2 = 1.5560 (-0.0100)
Sample #3 = 1.5490 (0.0150)
Sample #4 = 1.5560 (0.0130)
Avg % Abs = 1.5537 (0.0027)
STD DEV = 0.0040 (0.0117)
REL STD DEV = 0.260 (437.857)

Sol Value = 0.040 g/210L ***
Fit value = 0.1915 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12818, Sum Io = 13615
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 0.8820 (0.0770)
Sample #2 = 1.1360 (0.1010)
Sample #3 = 1.2340 (-0.5850)
Sample #4 = 0.6880 (0.0990)
Avg % Abs = 1.0193 (-0.1283)
STD DEV = 0.2911 (0.3955)
REL STD DEV = 28.558 (308.171)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 1.5360 (-0.0090)
Sample #2 = 1.5610 (-0.0020)
Sample #3 = 1.5750 (0.0010)
Sample #4 = 1.5250 (0.0170)
Avg % Abs = 1.5537 (0.0053)
STD DEV = 0.0258 (0.0102)
REL STD DEV = 1.660 (191.519)

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12837, Sum Io = 13616
Channel 1
Sample % Abs (% Abs Ref)
Sample #1 = 1.6970 (0.0340)
Sample #2 = 1.9420 (0.2580)
Sample #3 = 1.9450 (0.1860)
Sample #4 = 2.5620 (0.2110)
Avg % Abs = 2.1497 (0.2317)
STD DEV = 0.3571 (0.0588)
REL STD DEV = 16.612 (25.377)

Channel 2
Sample % Abs (% Abs Ref)
Sample #1 = 3.5490 (0.0140)
Sample #2 = 3.5330 (0.0430)
Sample #3 = 3.5330 (0.0300)
Sample #4 = 3.5570 (0.0360)
Avg % Abs = 3.5410 (0.0363)
STD DEV = 0.0139 (0.0065)
REL STD DEV = 0.391 (17.908)

**** AUTO CAL FAIL
Benjamin Sully

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Daniel Lyons on 2/6/2024

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-004587

Bill To Address:

Daniel Lyons

Osceola County Sheriff's Office

Ship to Address:

FDLE Off-Site Mail Facility

c/o Florida Dept of Law Enforcement

Alcohol Testing Program

813 B Lake Bradford Road

Tallahassee, FL 32304

Reason for Return:

Stability checks have interferent detect, ambient fail, purge fail. Attempted optical bench calibration adjustment twice, but individual solution calibrations kept repeating and could not be adjusted.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$_____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Daniel Lyons

Phone #: 407-799-1711

Email: Daniel.Lyons@osceolasheriff.org

ATP Contact Name: Benjamin Siddoway

ATP Email: BenjaminSiddoway@fdle.state.fl.us