

Flow Calibration Adjustment(s)

Performed by TDG

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005076
11/08/2024
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff) = 6.926

2: Rate (Liters/min) = 15

SQRT(Diff) = 11.957

3: Rate (Liters/min) = 30

SQRT(Diff) = 20.758

Dependent Data Scale Factor = 10000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 701

Rounded Intercept = -706193

Correlation = 0.99916

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083 <input checked="" type="checkbox"/> ≤0.003 of Wet <input checked="" type="checkbox"/>																																																																																																																																				
<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:44</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:44</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:45</td></tr> <tr><td>Control Test</td><td>0.049</td><td>13:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:46</td></tr> <tr><td>Control Test</td><td>0.050</td><td>13:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:47</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0493</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1703</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	13:44	Control Test	0.049	13:44	Air Blank	0.000	13:45	Control Test	0.049	13:46	Air Blank	0.000	13:46	Control Test	0.050	13:47	Air Blank	0.000	13:47	Control Test Stats			Average	0.0493		Std Dev	0.0006		Rel Std Dev(%)	1.1703		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:22</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>13:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:25</td></tr> <tr><td>Control Test</td><td>0.079</td><td>13:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	13:22	Control Test	0.079	13:23	Air Blank	0.000	13:23	Control Test	0.080	13:24	Air Blank	0.000	13:25	Control Test	0.079	13:25	Air Blank	0.000	13:26	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:28</td></tr> <tr><td>Control Test</td><td>0.201</td><td>13:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:30</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:30</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:31</td></tr> <tr><td>Control Test</td><td>0.200</td><td>13:32</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:32</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.2003</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.2882</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	13:28	Control Test	0.201	13:29	Air Blank	0.000	13:30	Control Test	0.200	13:30	Air Blank	0.000	13:31	Control Test	0.200	13:32	Air Blank	0.000	13:32	Control Test Stats			Average	0.2003		Std Dev	0.0006		Rel Std Dev(%)	0.2882		<p><i>ML</i></p> <p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 11/08/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>13:35</td></tr> <tr><td>Control Test</td><td>0.078</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:36</td></tr> <tr><td>Control Test</td><td>0.076</td><td>13:36</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:37</td></tr> <tr><td>Control Test</td><td>0.077</td><td>13:37</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>13:38</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0770</td><td></td></tr> <tr><td>Std Dev</td><td>0.0010</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.2987</td><td></td></tr> </table> <p>Operator's Signature <i>ML</i></p>	Air Blank	0.000	13:35	Control Test	0.078	13:36	Air Blank	0.000	13:36	Control Test	0.076	13:36	Air Blank	0.000	13:37	Control Test	0.077	13:37	Air Blank	0.000	13:38	Control Test Stats			Average	0.0770		Std Dev	0.0010		Rel Std Dev(%)	1.2987	
Air Blank	0.000	13:44																																																																																																																																					
Control Test	0.049	13:44																																																																																																																																					
Air Blank	0.000	13:45																																																																																																																																					
Control Test	0.049	13:46																																																																																																																																					
Air Blank	0.000	13:46																																																																																																																																					
Control Test	0.050	13:47																																																																																																																																					
Air Blank	0.000	13:47																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0493																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1703																																																																																																																																						
Air Blank	0.000	13:22																																																																																																																																					
Control Test	0.079	13:23																																																																																																																																					
Air Blank	0.000	13:23																																																																																																																																					
Control Test	0.080	13:24																																																																																																																																					
Air Blank	0.000	13:25																																																																																																																																					
Control Test	0.079	13:25																																																																																																																																					
Air Blank	0.000	13:26																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0793																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.7277																																																																																																																																						
Air Blank	0.000	13:28																																																																																																																																					
Control Test	0.201	13:29																																																																																																																																					
Air Blank	0.000	13:30																																																																																																																																					
Control Test	0.200	13:30																																																																																																																																					
Air Blank	0.000	13:31																																																																																																																																					
Control Test	0.200	13:32																																																																																																																																					
Air Blank	0.000	13:32																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.2003																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	0.2882																																																																																																																																						
Air Blank	0.000	13:35																																																																																																																																					
Control Test	0.078	13:36																																																																																																																																					
Air Blank	0.000	13:36																																																																																																																																					
Control Test	0.076	13:36																																																																																																																																					
Air Blank	0.000	13:37																																																																																																																																					
Control Test	0.077	13:37																																																																																																																																					
Air Blank	0.000	13:38																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0770																																																																																																																																						
Std Dev	0.0010																																																																																																																																						
Rel Std Dev(%)	1.2987																																																																																																																																						


```

>>>>> CHANNEL 2 >>>>>
Sample      % Abs      (% Abs)
Sample #1 = 1.4990      (-0.0)
Sample #2 = 1.4860      (0.00)
Sample #3 = 1.4570      (0.02)
Sample #4 = 1.4710      (0.01)
Aug % Abs = 1.4713      (0.0153)
STD DEV = 0.0145      (0.0075)
REL STD DEV = 0.986      (48.949)
-----
SN 80-005076
09:28:01
1/12/2024

```

Auto Calibration
Max Power Res Value = 77
Auto Range Res Value = 58

```

Sol Value = 0.000 g/210L ***
Fit value = 0.0000 mg/L 4.88%
Samples Taken = 4, Discarded = 1
Sum 10 = 12526, Num 10 = 12999

<<<< CHANNEL 1 >>>>

Sample      % Ads      (% Ads Ref)
Sample #1 = 0.0670      (0.0040)
Sample #2 = 0.0910      (0.0120)
Sample #3 = 0.0480      (0.0550)
Sample #4 = 0.0670      (0.0600)

Avg % ADS = 0.0687      (0.0423)
STD DEV = 0.0215      (0.0264)

REL STD DEV = 31.38%      (62.33%)

```

```

<<<< CHANNEL 2 >>>>      [% Abs Ref]
Sample      % Abs
Sample #1 = 0.1020      (-0.0020)
Sample #2 = 0.1300      (0.0000)
Sample #3 = 0.1100      (0.0050)
Sample #4 = 0.1160      (0.0160)
Avg % Abs = 0.1167      (0.0070)
STD DEV = 0.0103      (0.0082)
REL STD DEV = 8.649      (116.934)

```

Sol Value = 0.040 g/210L ***
Fit value = 0.1905 mg/l %22%
Samples Taken = 4, Discarded = 1
Sum Io = 12514, Sum Io = 12984

```

<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 0.8280    (-0.0160)
Sample #2 = 0.8104    (0.0030)
Sample #3 = 0.8110    (0.0120)
Sample #4 = 0.7970    (0.0260)
Avg % Abs = 0.8040    (0.0137)
STD DEV = 0.0070    (0.0115)
REL STD DEV = 0.871    (84.807)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 1.4990    (-0.0210)
Sample #2 = 1.4860    (0.0080)
Sample #3 = 1.4570    (0.0230)
Sample #4 = 1.4710    (0.0150)
Avg % Abs = 1.4713    (0.0153)
STD DEV = 0.0145    (0.0075)
REL STD DEV = 0.986    (48.949)

```

Sol Value = 0.100 g/210L ***
Fit value = 0.4762 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12509 g/m Io = 12983

```

Sample      % Abs      (% Abs Ref)
Sample #1 = 1.9120   (-0.0130)
Sample #2 = 1.8940   (0.0100)
Sample #3 = 1.8760   (0.0180)
Sample #4 = 1.8750   (0.0190)
Avg % Abs = 1.8783 (0.0157)
STD DEV = 0.0049
REL STD DEV = 0.263 (31.486)

```

```
<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.4930  (-0.0020)
Sample #2 = 3.4710  (0.0250)
Sample #3 = 3.4650  (0.0310)
Sample #4 = 3.4690  (0.0300)
Avg % Abs = 3.4683 (0.0287)
STD DEV = 0.0031 (0.0032)
REL STD DEV = 0.088 (0.1214)
```

```
sol value = 0.200 g/210L ***
tit value = 0.9524 mg/l %%%
samples Taken = 4, Discarded = 1
sum lo = 1.2506 gum lo = 12980
```

```

<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 3.6360   (-0.0060)
Sample #2 = 3.6260   (0.0120)
Sample #3 = 3.6370   (0.0090)
Sample #4 = 3.6070   (0.0230)
ug % Abs = 3.6233   (0.0147)
STD DEV = 0.0152   (0.0074)
EL STD DEV = 0.419   (0.258)

```

```

<<<< CHANNEL 2 >>>>
Sample      % Abs      [ % Abs
Sample #1 = 6.6540  [-0.0
Sample #2 = 6.5940  [0.14
Sample #3 = 6.5980  [-0.14
Sample #4 = 6.6150  [0.14
Avg % Abs = 6.590  [0.0427
STD DEV = 0.0156  [0.0015
REL STD DEV = 0.164  [3.580

```

Sol Value = 0.300 g/210L ***
Fit value = 1.4286 mg/l %%%
Samples Taken = 4, Discarded =
Sum Io = 12502 Sum Io = 1297

```

<<<< CHANNEL 1 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 5.3620    (-0.0230)
Sample #2 = 5.3630    (-0.0200)
Sample #3 = 5.3680    (-0.0070)
Sample #4 = 5.3410    (-0.0280)
Avg % Abs = 5.3573    (-0.0110)
STD DEV = 0.0144    (0.0154)
REL STD DEV = 0.268    (0.39.95%)

```

```
<<<< CHANNEL 2 >>>>
Sample      % Abs      (% Abs Ref)
Sample #1 = 9.7110    (0.0120)
Sample #2 = 9.7030    (0.0510)
Sample #3 = 9.6730    (0.0650)
Sample #4 = 9.6950    (0.0710)
Avg % Abs = 9.6870    (0.0623)
STD DEV = 0.0151    (0.0103)
REL STD DEV = 0.156    (0.1645)
```

```
***** AUTO CAL DATA *****
<<<<< CHANNEL 1 >>>>>
Sol Val = 0.0000 mg/l or 0.000
% ABS = 0.069
Std Dev = 0.02 Rel Std Dev =
Sol Val = 0.1905 mg/l or 0.040
% ABS = 0.804
Std Dev = 0.01 Rel Std Dev =
Sol Val = 0.4762 mg/l or 0.100
% ABS = 1.878
```

Std Dev = 0.00 Rel Std Dev = 0.26
Sol Val = 0.9524 mg/l or 0.200 g/21.0L
% Abs = 3.623
Std Dev = 0.02 Rel Std Dev = 0.42
Sol Val = 1.4288 mg/l or 0.300 g/21.0L
% Abs = 5.357
Std Dev = 0.01 Rel Std Dev = 0.27
Zero Order Coef = -194.44
First Order Coef = 2615.61
Second Order Coef = 16.58
Standard Deviation = 17.833447

```
<<<< CHANNEL 2 >>>>
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.119
Std Dev = 0.01 Rel Std Dev = 8.65
Sol Val = 0.1905 mg/l or 0.190 g/210L
% Abs = 1.471
Std Dev = 0.01 Rel Std Dev = 0.99
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.468
Std Dev = 0.00 Rel Std Dev = 0.09
Sol Val = 0.9324 mg/l or 0.200 g/210L
% Abs = 6.599
Std Dev = 0.01 Rel Std Dev = 0.08
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.687
Std Dev = 0.02 Rel Std Dev = 0.16
Zero Order Coef = -177.37
First Order Coef = 1400.83
Second Order Coef = 9.65
Standard Deviation = 26.8417
```

Solution Stats Quadratic Fit Chan 1			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	0.000	0.0003	
0.040	0.040	-0.0003	
0.100	0.100	-0.0003	
0.200	0.200	0.0005	
0.300	0.300	-0.0002	

Solution Stats Quadratic Fit Chan 2			
Act	Fit	Residual	
g/210L	g/210L	g/210L	
0.000	-0.000	0.0002	
0.040	0.040	0.0000	
0.100	0.101	-0.0007	
0.200	0.199	0.0008	
0.300	0.300	-0.0003	

```

Sol Value = 0.000 g/210L ***
Fit Value = 0.3910 mg/l 42%
Samples Taken = 4, Discarded = 1
***** CHANNEL 1 *****
Sample #1 = 3240.00
Sample #2 = 3277.00
Sample #3 = 3243.00
Sample #4 = 3224.00
Average Result = 3248.0000
STD DEV = 26.6514
REL STD DEV = 0.827

```

```

***** CHANNEL 2 *****
Sample #1 = 3315.00
Sample #2 = 3329.00
Sample #3 = 3325.00
Sample #4 = 3319.00
Average Result = 3324.333
STD DEV = 5.0332
REL STD DEV = 0.151
*****
Dry Gas H2O Adjust Results *****
Barometric Pressure = 1015
3 um H2O Adjust (mg/1*10,000) = 561
9 um H2O Adjust (mg/1*10,000) = 485
***** AUTO CAL PASS *****

```

Optical Calibration Adjustment

Bv: TDG

Post-Cal Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																				
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																				
✓	✓	✓	✓																																																																																																																																				
≤0.003 of Wet			≤0.003 of Wet																																																																																																																																				
✓			✓																																																																																																																																				
<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 11/12/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:45</td></tr> <tr><td>Control Test</td><td>0.051</td><td>10:46</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:46</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:47</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:47</td></tr> <tr><td>Control Test</td><td>0.050</td><td>10:48</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:49</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0503</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>1.1471</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	10:45	Control Test	0.051	10:46	Air Blank	0.000	10:46	Control Test	0.050	10:47	Air Blank	0.000	10:47	Control Test	0.050	10:48	Air Blank	0.000	10:49	Control Test Stats			Average	0.0503		Std Dev	0.0006		Rel Std Dev(%)	1.1471		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 11/12/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:20</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:21</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:21</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:22</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:23</td></tr> <tr><td>Control Test</td><td>0.080</td><td>10:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:24</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0800</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	10:20	Control Test	0.080	10:21	Air Blank	0.000	10:21	Control Test	0.080	10:22	Air Blank	0.000	10:23	Control Test	0.080	10:23	Air Blank	0.000	10:24	Control Test Stats			Average	0.0800		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 11/12/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:28</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:29</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:30</td></tr> <tr><td>Control Test</td><td>0.198</td><td>10:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:31</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1980</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	10:28	Control Test	0.198	10:28	Air Blank	0.000	10:29	Control Test	0.198	10:29	Air Blank	0.000	10:30	Control Test	0.198	10:31	Air Blank	0.000	10:31	Control Test Stats			Average	0.1980		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 11/12/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <table border="1"> <tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:33</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:34</td></tr> <tr><td>Control Test</td><td>0.081</td><td>10:35</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>10:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0810</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </table> <p>Operator's Signature _____</p>	Air Blank	0.000	10:33	Control Test	0.081	10:33	Air Blank	0.000	10:33	Control Test	0.081	10:34	Air Blank	0.000	10:34	Control Test	0.081	10:35	Air Blank	0.000	10:35	Control Test Stats			Average	0.0810		Std Dev	0.0000		Rel Std Dev(%)	0.0000	
Air Blank	0.000	10:45																																																																																																																																					
Control Test	0.051	10:46																																																																																																																																					
Air Blank	0.000	10:46																																																																																																																																					
Control Test	0.050	10:47																																																																																																																																					
Air Blank	0.000	10:47																																																																																																																																					
Control Test	0.050	10:48																																																																																																																																					
Air Blank	0.000	10:49																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0503																																																																																																																																						
Std Dev	0.0006																																																																																																																																						
Rel Std Dev(%)	1.1471																																																																																																																																						
Air Blank	0.000	10:20																																																																																																																																					
Control Test	0.080	10:21																																																																																																																																					
Air Blank	0.000	10:21																																																																																																																																					
Control Test	0.080	10:22																																																																																																																																					
Air Blank	0.000	10:23																																																																																																																																					
Control Test	0.080	10:23																																																																																																																																					
Air Blank	0.000	10:24																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0800																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:28																																																																																																																																					
Control Test	0.198	10:28																																																																																																																																					
Air Blank	0.000	10:29																																																																																																																																					
Control Test	0.198	10:29																																																																																																																																					
Air Blank	0.000	10:30																																																																																																																																					
Control Test	0.198	10:31																																																																																																																																					
Air Blank	0.000	10:31																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.1980																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						
Air Blank	0.000	10:33																																																																																																																																					
Control Test	0.081	10:33																																																																																																																																					
Air Blank	0.000	10:33																																																																																																																																					
Control Test	0.081	10:34																																																																																																																																					
Air Blank	0.000	10:34																																																																																																																																					
Control Test	0.081	10:35																																																																																																																																					
Air Blank	0.000	10:35																																																																																																																																					
Control Test Stats																																																																																																																																							
Average	0.0810																																																																																																																																						
Std Dev	0.0000																																																																																																																																						
Rel Std Dev(%)	0.0000																																																																																																																																						

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 11:08

Date of Inspection: 11/12/2024

Serial Number: 80-005076
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000		No	Mouth Alcohol Test: Slope Not Met		No
Interferent Detect Test: Interferent Detect		No	Diagnostic Check (Post-Inspection): OK		No

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: Exp:	0.08g/210L Test (g/210L) Lot#: Exp:	0.20g/210L Test (g/210L) Lot#: Exp:	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: Exp:

Standard Deviations				
---------------------	--	--	--	--

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: _____ Number of Simulators Used: 5

Remarks:

A F / M A: No Sample Provided. DID NOT GIVE SAMPLE. RETEST.. Non-compliance: .

See notes on Instrument Processing Sheet.

ML
11/12/2024

Not determined

ML
11/12/2024

The above instrument complies () does not comply (☒) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Taylor D Gutschow

TAYLOR D GUTSCHOW

Signature and Printed Name

11/12/2024
Date

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 12:55

Date of Inspection: 11/12/2024

Serial Number: 80-005076
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.080	0.198	0.081
0.000	0.050	0.080	0.197	0.081
0.000	0.050	0.080	0.198	0.081
0.000	0.050	0.080	0.198	0.080
0.000	0.050	0.080	0.198	0.081
0.000	0.050	0.080	0.198	0.081
0.000	0.050	0.080	0.198	0.081
0.000	0.051	0.081	0.198	0.081
0.000	0.050	0.080	0.198	0.081
0.000	0.050	0.081	0.198	0.081


Standard Deviations	0.0003	0.0004	0.0003	0.0003
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

11/12/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005076, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005076</u>	UNCERTAINTY* \pm
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>11/12/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>12:55</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.



11/12/2024

Date

TAYLOR D GUTSCHOW,

Department Inspector

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality







INSTRUMENT PROCESSING SHEET

Agency Sarasota CSOS/N 80-005076Florida Department of
Law EnforcementDate In 02/27/2024 DI Completion Date 03/04/2024☐ Ship ☒ P/U ☐ H/D ☐ CMI ☐ EE

Intake By TDG _____ Date <u>02/27/2024</u> <input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: <u>Dropped off. Missing two leftmost screws on regulator.</u>	Quality Checks By TDG _____ Date <u>02/28/2024</u> <input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>246</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP106</u> 32 mm <u>0.167</u> (.139 - .169) 36 mm <u>0.175</u> (.156 - .190) 53 mm <u>0.257</u> (.228 - .278) 103 mm <u>0.519</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP4864</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>	Simulator	Serial #	Lot #/Exp	0.050	MP4864	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Flow Calibration By _____ Date _____ Flow Column # _____ <input type="checkbox"/> 5L/min - 17mm <input type="checkbox"/> 15L/min - 53mm <input type="checkbox"/> 30L/min - 103mm <input type="checkbox"/> R-Value _____ <input type="checkbox"/> Post Calibration Verification (L/s) Flow Column # _____ 32 mm _____ (.139 - .169) 36 mm _____ (.156 - .190) 53 mm _____ (.228 - .278) 103 mm _____ (.447 - .547) Maintenance By TDG _____ Date <u>2/27/2024</u> <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input checked="" type="checkbox"/> Other <u>Replaced missing screws on regulator.</u>																																												
Simulator	Serial #	Lot #/Exp																																																											
0.050	MP4864	202303K 03/29/2025																																																											
0.080	MP6287	202303L 03/29/2025																																																											
0.200	MP6288	202304C 04/05/2025																																																											
0.080 DGS	N/A	01923080A3 02/05/2025																																																											
Calibration Adjustment By TDG _____ Barometric Pressure Gauge <u>1022</u> ID # <u>28199</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td>MP5097</td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td>MP5098</td><td>23400</td><td>10/24/2025</td></tr><tr><td>0.100</td><td>MP5099</td><td>23390</td><td>10/17/2025</td></tr><tr><td>0.200</td><td>MP5100</td><td>23340</td><td>09/18/2025</td></tr><tr><td>0.300</td><td>MP5101</td><td>23070</td><td>03/06/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>AG222203</td><td>08/10/2024</td></tr></tbody></table> <input checked="" type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td>MP4864</td><td>202303K</td><td>03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L</td><td>03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C</td><td>04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3</td><td>02/05/2025</td></tr></tbody></table> Notes/Suggested Service: <u>Failed the baro pressure check during Quality Checks. The instrument read 1001, and the gauge read 1020. (TDG)</u>	Simulator	Serial #	Lot #	Expiration	0.000	MP5097	N/A	N/A	0.040	MP5098	23400	10/24/2025	0.100	MP5099	23390	10/17/2025	0.200	MP5100	23340	09/18/2025	0.300	MP5101	23070	03/06/2025	0.080 DGS	N/A	AG222203	08/10/2024	Simulator	Serial #	Lot #	Expiration	0.050	MP4864	202303K	03/29/2025	0.080	MP6287	202303L	03/29/2025	0.200	MP6288	202304C	04/05/2025	0.080 DGS	N/A	01923080A3	02/05/2025	Department Inspection By TDG _____ Barometric Pressure ID# <u>26932</u> Gauge <u>1018</u> Instrument <u>1020</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2022-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments <input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Stability Checks <input type="checkbox"/> Flow Calibration <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Calibration Adjustment <input type="checkbox"/> Other _____ <input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input checked="" type="checkbox"/> Return to/Place into Evidentiary Use <input type="checkbox"/> Remain Out of Evidentiary Use <input checked="" type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use <div style="display: flex; justify-content: space-between;"><div>Shayla Platt Digitally signed by Shayla Platt Date: 2024.03.07 13:22:00 Tech Review / Date _____</div><div>Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.03.14 14:59:09 -04'00' Admin Review / Date _____</div></div>	Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288
Simulator	Serial #	Lot #	Expiration																																																										
0.000	MP5097	N/A	N/A																																																										
0.040	MP5098	23400	10/24/2025																																																										
0.100	MP5099	23390	10/17/2025																																																										
0.200	MP5100	23340	09/18/2025																																																										
0.300	MP5101	23070	03/06/2025																																																										
0.080 DGS	N/A	AG222203	08/10/2024																																																										
Simulator	Serial #	Lot #	Expiration																																																										
0.050	MP4864	202303K	03/29/2025																																																										
0.080	MP6287	202303L	03/29/2025																																																										
0.200	MP6288	202304C	04/05/2025																																																										
0.080 DGS	N/A	01923080A3	02/05/2025																																																										
Simulator	Serial Number																																																												
0.000	MP6284																																																												
Interferent	MP6285																																																												
0.050	MP6286																																																												
0.080	MP6287																																																												
0.200	MP6288																																																												

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083 ✓ ≤0.003 of Wet ✓
<p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005076 02/28/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:21</p> <p>Control Test 0.050 11:21</p> <p>Air Blank 0.000 11:22</p> <p>Control Test 0.049 11:22</p> <p>Air Blank 0.000 11:23</p> <p>Control Test 0.050 11:24</p> <p>Control Test Stats</p> <p>Average 0.0497</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 1.1625</p> <p>Operator's Signature </p>	<p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005076 02/28/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:29</p> <p>Control Test 0.080 11:30</p> <p>Air Blank 0.000 11:30</p> <p>Control Test 0.079 11:31</p> <p>Air Blank 0.000 11:32</p> <p>Control Test 0.080 11:32</p> <p>Air Blank 0.000 11:33</p> <p>Control Test Stats</p> <p>Average 0.0797</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.7247</p> <p>Operator's Signature </p>	<p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005076 02/28/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:36</p> <p>Control Test 0.199 11:36</p> <p>Air Blank 0.000 11:37</p> <p>Control Test 0.199 11:38</p> <p>Air Blank 0.000 11:39</p> <p>Control Test 0.200 11:39</p> <p>Air Blank 0.000 11:39</p> <p>Control Test Stats</p> <p>Average 0.1993</p> <p>Std Dev 0.0006</p> <p>Rel Std Dev(%) 0.2896</p> <p>Operator's Signature </p>	<p>065</p> <p>SARASOTA COUNTY SO Intoxilizer - Alcohol Analyzer Model 8000 SN 80-005076 02/28/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 11:42</p> <p>Control Test 0.082 11:42</p> <p>Air Blank 0.000 11:43</p> <p>Control Test 0.083 11:43</p> <p>Air Blank 0.000 11:43</p> <p>Control Test 0.081 11:44</p> <p>Air Blank 0.000 11:44</p> <p>Control Test Stats</p> <p>Average 0.0820</p> <p>Std Dev 0.0010</p> <p>Rel Std Dev(%) 1.2195</p> <p>Operator's Signature </p>

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.4960 (0.0000)
Sample #2 = 1.4910 (0.0150)
Sample #3 = 1.4750 (0.0260)
Sample #4 = 1.5030 (0.0290)
Avg % Abs = 1.4897 (0.0233)
STD DEV = 0.0140 (0.0074)
REL STD DEV = 0.943 (31.590)

SARASOTA COUNTY SO
Intoxilyzer - Alcohol Analyzer
Model 8000
SN 80-005076
13:34:46
02/29/2024

Auto Calibration
Max Power Res Value = 76
Auto Range Res Value = 57

Sol Value = 0.000 g/210L ***
Fit Value = 0.0000 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12494, Sum Io = 12971
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 1.9000 (0.0080)
Sample #2 = 1.9150 (0.0170)
Sample #3 = 1.8950 (0.0570)
Sample #4 = 1.8670 (0.0650)
Avg % Abs = 1.8803 (0.0463)
STD DEV = 0.0303 (0.0257)
REL STD DEV = 1.611 (55.503)

***** CHANNEL 2 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.5240 (0.0160)
Sample #2 = 3.5030 (0.0130)
Sample #3 = 3.4640 (0.0470)
Sample #4 = 3.4460 (0.0550)
Avg % Abs = 3.4710 (0.0383)
STD DEV = 0.0291 (0.0223)
REL STD DEV = 0.839 (58.176)

Sol Value = 0.200 g/210L ***
Fit Value = 0.9524 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12451, Sum Io = 12946
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 3.6760 (0.0280)
Sample #2 = 3.6300 (0.0080)
Sample #3 = 3.6560 (0.0030)
Sample #4 = 3.6210 (0.0330)
Avg % Abs = 3.6357 (0.0147)
STD DEV = 0.0182 (0.0161)
REL STD DEV = 0.500 (109.587)

Sol Value = 0.040 g/210L ***
Fit Value = 0.1905 mg/l %%%
Samples Taken = 4, Discarded = 1
Sum Io = 12477, Sum Io = 12962
***** CHANNEL 1 *****
Sample % Abs (% Abs Ref)
Sample #1 = 0.7940 (0.0090)
Sample #2 = 0.8340 (0.0050)
Sample #3 = 0.8020 (0.0250)
Sample #4 = 0.8090 (0.0320)
Avg % Abs = 0.8150 (0.0160)
STD DEV = 0.0168 (0.0219)
REL STD DEV = 2.064 (137.073)

***** AUTO CAL DATA *****
***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.094
Std Dev = 0.02 Rel Std Dev = 16.55
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.815
Std Dev = 0.02 Rel Std Dev = 2.06
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.880
Std Dev = 0.03 Rel Std Dev = 1.61
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.635
Std Dev = 0.02 Rel Std Dev = 0.50
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.400
Std Dev = 0.02 Rel Std Dev = 0.42
Zero Order Coef = -262.06
First Order Coef = 2669.18
Second Order Coef = 4.79
Standard Deviation = 14.14138





***** CHANNEL 2 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.124
Std Dev = 0.01 Rel Std Dev = 10.11
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 1.490
Std Dev = 0.01 Rel Std Dev = 0.94
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 3.471
Std Dev = 0.03 Rel Std Dev = 0.84
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 6.678
Std Dev = 0.02 Rel Std Dev = 0.32
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 9.782
Std Dev = 0.01 Rel Std Dev = 0.09
Zero Order Coef = -180.34
First Order Coef = 1394.06
Second Order Coef = 8.69
Standard Deviation = 7.828565

***** CHANNEL 1 *****
Sol Val = 0.0000 mg/l or 0.000 g/210L
% Abs = 0.094
Std Dev = 0.02 Rel Std Dev = 16.55
Sol Val = 0.1905 mg/l or 0.040 g/210L
% Abs = 0.815
Std Dev = 0.02 Rel Std Dev = 2.06
Sol Val = 0.4762 mg/l or 0.100 g/210L
% Abs = 1.880
Std Dev = 0.03 Rel Std Dev = 1.61
Sol Val = 0.9524 mg/l or 0.200 g/210L
% Abs = 3.635
Std Dev = 0.02 Rel Std Dev = 0.50
Sol Val = 1.4286 mg/l or 0.300 g/210L
% Abs = 5.400
Std Dev = 0.02 Rel Std Dev = 0.42
Zero Order Coef = -262.06
First Order Coef = 2669.18
Second Order Coef = 4.79
Standard Deviation = 14.14138

Optical Calibration
Adjustment

By: TDG

Post-Cal Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083
<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 02/29/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:09 Control Test 0.051 15:09 Air Blank 0.000 15:10 Control Test 0.050 15:11 Air Blank 0.000 15:11 Control Test 0.051 15:12 Air Blank 0.000 15:12 Control Test Stats Average 0.0507 Std Dev 0.0006 Rel Std Dev(%) 1.1395</p> <p>Operator's Signature </p>	<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 02/29/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:15 Control Test 0.080 15:16 Air Blank 0.000 15:17 Control Test 0.080 15:17 Air Blank 0.000 15:18 Control Test 0.081 15:19 Air Blank 0.000 15:19 Control Test Stats Average 0.0803 Std Dev 0.0006 Rel Std Dev(%) 0.7187</p> <p>Operator's Signature </p>	<p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 02/29/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:23 Control Test 0.198 15:24 Air Blank 0.000 15:25 Control Test 0.197 15:25 Air Blank 0.000 15:26 Control Test 0.197 15:26 Air Blank 0.000 15:27 Control Test Stats Average 0.1973 Std Dev 0.0006 Rel Std Dev(%) 0.2926</p> <p>Operator's Signature </p>	<p>AGS</p> <p>SARASOTA COUNTY SO Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005076 02/29/2024 Software: 8100.27</p> <p>Test g/210L Time</p> <p>Air Blank 0.000 15:05 Control Test 0.080 15:06 Air Blank 0.000 15:06 Control Test 0.080 15:07 Air Blank 0.000 15:07 Control Test 0.080 15:07 Air Blank 0.000 15:08 Control Test Stats Average 0.0800 Std Dev 0.0000 Rel Std Dev(%) 0.0000</p> <p>Operator's Signature </p>

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: SARASOTA COUNTY SO
Time of Inspection: 13:00

Date of Inspection: 03/04/2024

Serial Number: 80-005076
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.081	0.197	0.079
0.000	0.050	0.080	0.197	0.080
0.000	0.050	0.081	0.197	0.079
0.000	0.050	0.081	0.198	0.079
0.000	0.050	0.080	0.198	0.079
0.000	0.050	0.081	0.198	0.079
0.000	0.050	0.080	0.197	0.079
0.000	0.050	0.081	0.198	0.080
0.000	0.050	0.081	0.198	0.079


Standard Deviations	0.0000	0.0005	0.0005	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



Signature and Printed Name

TAYLOR D GUTSCHOW

03/04/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005076, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005076</u>	UNCERTAINTY* \pm
Owning Agency:	<u>SARASOTA COUNTY SO</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>03/04/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:00</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence (k=3).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.


TAYLOR D GUTSCHOW,
Department Inspector

03/04/2024

Date

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Service • Integrity • Respect • Quality