



INSTRUMENT PROCESSING SHEET

Agency Fort Myers PDS/N 80-005468Florida Department of
Law EnforcementDate In 10/18/2024DI Completion Date 10/23/2024☐ Ship ☐ P/U ☐ H/D ☒ CMI ☐ EE

Intake By <u>TDG</u> Date <u>10/21/2024</u>		Quality Checks By <u>TDG</u> Date (See Notes)		Flow Calibration By <u>TDG</u> Date <u>10/23/2024</u>																																																															
<input checked="" type="checkbox"/> Annual <input type="checkbox"/> Registration <input type="checkbox"/> Return from CMI / EE Visual Inspection: <input checked="" type="checkbox"/> Case <input checked="" type="checkbox"/> Handle <input checked="" type="checkbox"/> Keyboard <input checked="" type="checkbox"/> Dry Gas Shelf <input checked="" type="checkbox"/> Feet <input checked="" type="checkbox"/> Breath Tube <input checked="" type="checkbox"/> Ports <input checked="" type="checkbox"/> Screws Tight Other Equipment/ Accessories: <input type="checkbox"/> Power cord <input type="checkbox"/> Printer Cable <input type="checkbox"/> Static Bag <input type="checkbox"/> 12V DC Cable Notes: _____ _____ _____ _____ _____ _____ _____		<input checked="" type="checkbox"/> Breath Tube Screen <input checked="" type="checkbox"/> Replace External O-Rings <input checked="" type="checkbox"/> Instrument Set Up Verified <input checked="" type="checkbox"/> R-Value <u>169</u> <input checked="" type="checkbox"/> Flow Verification (L/s) Flow Column # <u>ATP104</u> 32 mm <u>0.132*</u> (.139 - .169) 36 mm <u>0.144*</u> (.156 - .190) 53 mm <u>0.222*</u> (.228 - .278) 103 mm <u>0.500</u> (.447 - .547) <input checked="" type="checkbox"/> Barometric Pressure Check Gauge ID # <u>26932</u> <input checked="" type="checkbox"/> Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #/Exp</th></tr></thead><tbody><tr><td>0.050</td><td>MP6286</td><td>202303K 03/29/2025</td></tr><tr><td>0.080</td><td>MP6287</td><td>202303L 03/29/2025</td></tr><tr><td>0.200</td><td>MP6288</td><td>202304C 04/05/2025</td></tr><tr><td>0.080 DGS</td><td>N/A</td><td>01923080A3 02/05/2025</td></tr></tbody></table>		Simulator	Serial #	Lot #/Exp	0.050	MP6286	202303K 03/29/2025	0.080	MP6287	202303L 03/29/2025	0.200	MP6288	202304C 04/05/2025	0.080 DGS	N/A	01923080A3 02/05/2025	Flow Column # <u>ATP101 (x3)</u> <input checked="" type="checkbox"/> 5L/min - 17mm <input checked="" type="checkbox"/> 15L/min - 53mm <input checked="" type="checkbox"/> 30L/min - 103mm <input checked="" type="checkbox"/> R-Value <u>169 / 169 / 169</u> <input checked="" type="checkbox"/> Post Calibration Verification (L/s) Flow Column # <u>ATP104 (x3)</u> 32 mm <u>0.140 / 0.136* / 0.140</u> (.139 - .169) 36 mm <u>0.152* / 0.152* / 0.152*</u> (.156 - .190) 53 mm <u>0.226* / 0.222* / 0.226*</u> (.228 - .278) 103 mm <u>0.500 / 0.492 / 0.500</u> (.447 - .547) Maintenance By _____ Date _____ <input type="checkbox"/> Battery Replacement <input type="checkbox"/> Dry Gas Regulator Replacement <input type="checkbox"/> Breath Tube Replacement <input type="checkbox"/> Other _____ _____ _____ _____																																																
Simulator	Serial #	Lot #/Exp																																																																	
0.050	MP6286	202303K 03/29/2025																																																																	
0.080	MP6287	202303L 03/29/2025																																																																	
0.200	MP6288	202304C 04/05/2025																																																																	
0.080 DGS	N/A	01923080A3 02/05/2025																																																																	
Calibration Adjustment By _____				Department Inspection By <u>TDG</u>																																																															
Barometric Pressure Gauge _____ ID # _____ <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.000</td><td></td><td>N/A</td><td>N/A</td></tr><tr><td>0.040</td><td></td><td></td><td></td></tr><tr><td>0.100</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.300</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table> <input type="checkbox"/> Post Calibration Adjustment Stability Checks <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial #</th><th>Lot #</th><th>Expiration</th></tr></thead><tbody><tr><td>0.050</td><td></td><td></td><td></td></tr><tr><td>0.080</td><td></td><td></td><td></td></tr><tr><td>0.200</td><td></td><td></td><td></td></tr><tr><td>0.080 DGS</td><td>N/A</td><td></td><td></td></tr></tbody></table>				Simulator	Serial #	Lot #	Expiration	0.000		N/A	N/A	0.040				0.100				0.200				0.300				0.080 DGS	N/A			Simulator	Serial #	Lot #	Expiration	0.050				0.080				0.200				0.080 DGS	N/A			Barometric Pressure ID# <u>26932</u> Gauge <u>1017</u> Instrument <u>1015</u> Mouth Alcohol Solution Lot # <u>2023-A</u> Acetone Stock Solution Lot # <u>2023-B</u> <table border="1" style="width:100%"><thead><tr><th>Simulator</th><th>Serial Number</th></tr></thead><tbody><tr><td>0.000</td><td>MP6284</td></tr><tr><td>Interferent</td><td>MP6285</td></tr><tr><td>0.050</td><td>MP6286</td></tr><tr><td>0.080</td><td>MP6287</td></tr><tr><td>0.200</td><td>MP6288</td></tr></tbody></table> Attachments <table border="1" style="width:100%"><tr><td><input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment</td><td><input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u></td></tr></table>		Simulator	Serial Number	0.000	MP6284	Interferent	MP6285	0.050	MP6286	0.080	MP6287	0.200	MP6288	<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u>
Simulator	Serial #	Lot #	Expiration																																																																
0.000		N/A	N/A																																																																
0.040																																																																			
0.100																																																																			
0.200																																																																			
0.300																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial #	Lot #	Expiration																																																																
0.050																																																																			
0.080																																																																			
0.200																																																																			
0.080 DGS	N/A																																																																		
Simulator	Serial Number																																																																		
0.000	MP6284																																																																		
Interferent	MP6285																																																																		
0.050	MP6286																																																																		
0.080	MP6287																																																																		
0.200	MP6288																																																																		
<input checked="" type="checkbox"/> Form 41 <input checked="" type="checkbox"/> Stability Checks <input checked="" type="checkbox"/> Calibration Certificate <input type="checkbox"/> Calibration Adjustment	<input type="checkbox"/> Post-Stability Checks <input checked="" type="checkbox"/> Flow Calibration <input type="checkbox"/> Form 40 <input checked="" type="checkbox"/> Other <u>Form 51</u>																																																																		
Notes/Suggested Service: <u>Checked breath tube screen and replaced o-rings on 10/22. Completed the rest of the Quality Checks on 10/23. (TDG)</u> _____ <u>*Outside nominal range. (TDG)</u> _____ <u>Could not get post-flow cal verifications to pass. (TDG)</u> _____				<input checked="" type="checkbox"/> Instrument Complies with Chapter 11D-8, FAC <input type="checkbox"/> Instrument Does Not Comply with Chapter 11D-8, FAC <input type="checkbox"/> Return to/Place into Evidentiary Use <input checked="" type="checkbox"/> Remain Out of Evidentiary Use <input type="checkbox"/> Conduct an Agency Inspection Before Evidentiary Use Destinee Armstrong Digitally signed by Destinee Armstrong Date: 2024.11.14 09:30:53 -05'00' Phil Nicodemo Digitally signed by Phil Nicodemo Date: 2024.11.18 10:45:44 -05'00' Tech Review / Date _____ Admin Review / Date _____																																																															

Flow Calibration Adjustment(s)

Performed by TDG

#1

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005468
10/23/2024
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 7.348

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.957

3: Rate (Liters/min) = 30

SQRT(Diff)) = 21.422

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 683

Rounded Intercept = -706077

Correlation = 0.99667

#3

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005468
10/23/2024
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 7.348

2: Rate (Liters/min) = 15

SQRT(Diff)) = 11.957

3: Rate (Liters/min) = 30

SQRT(Diff)) = 21.258

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 692

Rounded Intercept = -727232

Correlation = 0.99701

#2

FORT MYERS PD
Intoxilyzer - Alcohol Analyzer
Model 8000 SN 80-005468
10/23/2024
Software: 8100.27

Flow Rate Calibration*****

1: Rate (Liters/min) = 5

SQRT(Diff)) = 7.547

2: Rate (Liters/min) = 15

SQRT(Diff)) = 12.039

3: Rate (Liters/min) = 30

SQRT(Diff)) = 21.445

Dependent Data Scale Factor = 100000 L/min

Independent Data Scale Factor = 256

Rounded Slope = 690

Rounded Intercept = -750770

Correlation = 0.99627

Return Material Authorization

Ship to: ☒ CMI, Inc.

☐ Enforcement Electronics

Shipment to repair facility authorized by: Daniel Aguilera on 11/08/2024

Items Returned: Instrument ☒ Supplies ☐ Other ☐ Describe: _____

Instrument Model: Intoxilyzer 8000 Serial Number: 80-005468

Bill To Address:

Fort Myers Police Department

Attn: Daniel Aguilera

Ship to Address:

Florida Department of Law Enforcement

Fort Myers Regional Operations Center

Attn: Taylor Gutschow

4700 Terminal Drive, Suite 1

Fort Myers, FL 33907

Reason for Return:

Instrument needs flow calibration adjustment. Conducted several in-house cal adjusts but could not get the post-cal verification to pass.

Please choose one of the following options:

☐ 1. I _____, authorize all repairs.

☐ 2. I _____, authorize repairs up to \$ _____.

☒ 3. I require an estimate **BEFORE** any repairs will be authorized and/ or conducted.

Please contact: Name: Daniel Aguilera

Phone #: 239-321-8238 Email: daguilera@fmpolice.com

ATP Contact Name: Taylor Gutschow ATP Email: TaylorGutschow@fdle.state.fl.us

Stability Checks

0.05g/210L	0.08g/210L	0.20g/210L	DGS 0.08g/210L																																																																																																																																										
0.047 to 0.053	0.077 to 0.083	0.194 to 0.206	0.077 to 0.083																																																																																																																																										
<p>#1</p> <p>FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005468 10/23/2024 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:14</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:15</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:15</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:16</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:17</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	11:14	Control Test	0.049	11:15	Air Blank	0.000	11:15	Control Test	0.049	11:16	Air Blank	0.000	11:17	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005468 10/23/2024 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:07</td></tr> <tr><td>Control Test</td><td>0.080</td><td>11:08</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:08</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:09</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:10</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:10</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:11</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0793</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7277</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	11:07	Control Test	0.080	11:08	Air Blank	0.000	11:08	Control Test	0.079	11:09	Air Blank	0.000	11:10	Control Test	0.079	11:10	Air Blank	0.000	11:11	Control Test Stats			Average	0.0793		Std Dev	0.0006		Rel Std Dev(%)	0.7277		<p>065</p> <p>FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005468 10/23/2024 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:22</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:23</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:23</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:24</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:25</td></tr> <tr><td>Control Test</td><td>0.196</td><td>11:25</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:26</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.1960</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	11:22	Control Test	0.196	11:23	Air Blank	0.000	11:23	Control Test	0.196	11:24	Air Blank	0.000	11:25	Control Test	0.196	11:25	Air Blank	0.000	11:26	Control Test Stats			Average	0.1960		Std Dev	0.0000		Rel Std Dev(%)	0.0000		<p>065</p> <p>FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005468 10/23/2024 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:27</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:27</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:28</td></tr> <tr><td>Control Test</td><td>0.078</td><td>11:28</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test</td><td>0.079</td><td>11:29</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:29</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0787</td><td></td></tr> <tr><td>Std Dev</td><td>0.0006</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.7339</td><td></td></tr> </tbody> </table> <p>Operator's Signature <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	11:27	Control Test	0.079	11:27	Air Blank	0.000	11:28	Control Test	0.078	11:28	Air Blank	0.000	11:29	Control Test	0.079	11:29	Air Blank	0.000	11:29	Control Test Stats			Average	0.0787		Std Dev	0.0006		Rel Std Dev(%)	0.7339	
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	11:14																																																																																																																																											
Control Test	0.049	11:15																																																																																																																																											
Air Blank	0.000	11:15																																																																																																																																											
Control Test	0.049	11:16																																																																																																																																											
Air Blank	0.000	11:17																																																																																																																																											
Control Test Stats																																																																																																																																													
Average	0.0490																																																																																																																																												
Std Dev	0.0000																																																																																																																																												
Rel Std Dev(%)	0.0000																																																																																																																																												
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	11:07																																																																																																																																											
Control Test	0.080	11:08																																																																																																																																											
Air Blank	0.000	11:08																																																																																																																																											
Control Test	0.079	11:09																																																																																																																																											
Air Blank	0.000	11:10																																																																																																																																											
Control Test	0.079	11:10																																																																																																																																											
Air Blank	0.000	11:11																																																																																																																																											
Control Test Stats																																																																																																																																													
Average	0.0793																																																																																																																																												
Std Dev	0.0006																																																																																																																																												
Rel Std Dev(%)	0.7277																																																																																																																																												
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	11:22																																																																																																																																											
Control Test	0.196	11:23																																																																																																																																											
Air Blank	0.000	11:23																																																																																																																																											
Control Test	0.196	11:24																																																																																																																																											
Air Blank	0.000	11:25																																																																																																																																											
Control Test	0.196	11:25																																																																																																																																											
Air Blank	0.000	11:26																																																																																																																																											
Control Test Stats																																																																																																																																													
Average	0.1960																																																																																																																																												
Std Dev	0.0000																																																																																																																																												
Rel Std Dev(%)	0.0000																																																																																																																																												
Test	g/210L	Time																																																																																																																																											
Air Blank	0.000	11:27																																																																																																																																											
Control Test	0.079	11:27																																																																																																																																											
Air Blank	0.000	11:28																																																																																																																																											
Control Test	0.078	11:28																																																																																																																																											
Air Blank	0.000	11:29																																																																																																																																											
Control Test	0.079	11:29																																																																																																																																											
Air Blank	0.000	11:29																																																																																																																																											
Control Test Stats																																																																																																																																													
Average	0.0787																																																																																																																																												
Std Dev	0.0006																																																																																																																																												
Rel Std Dev(%)	0.7339																																																																																																																																												

Inadvertently ran 3 stabilities instead of 3. Repeated. *ML* 10/23/24

Stability Checks

0.05g/210L 0.047 to 0.053	0.08g/210L 0.077 to 0.083	0.20g/210L 0.194 to 0.206	DGS 0.08g/210L 0.077 to 0.083	DGS 0.08g/210L ≤0.003 of Wet																																				
<p>#2</p> <p>FORT MYERS PD Intoxilyzer - Alcohol Analyzer Model 8000 SN 80-005468 10/23/2024 Software: 8100.27</p> <table border="1"> <thead> <tr> <th>Test</th> <th>g/210L</th> <th>Time</th> </tr> </thead> <tbody> <tr><td>Air Blank</td><td>0.000</td><td>11:31</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:31</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:32</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:33</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:33</td></tr> <tr><td>Control Test</td><td>0.049</td><td>11:34</td></tr> <tr><td>Air Blank</td><td>0.000</td><td>11:35</td></tr> <tr><td>Control Test Stats</td><td></td><td></td></tr> <tr><td>Average</td><td>0.0490</td><td></td></tr> <tr><td>Std Dev</td><td>0.0000</td><td></td></tr> <tr><td>Rel Std Dev(%)</td><td>0.0000</td><td></td></tr> </tbody> </table> <p>Operator's Signature: <i>ML</i></p>	Test	g/210L	Time	Air Blank	0.000	11:31	Control Test	0.049	11:31	Air Blank	0.000	11:32	Control Test	0.049	11:33	Air Blank	0.000	11:33	Control Test	0.049	11:34	Air Blank	0.000	11:35	Control Test Stats			Average	0.0490		Std Dev	0.0000		Rel Std Dev(%)	0.0000					
Test	g/210L	Time																																						
Air Blank	0.000	11:31																																						
Control Test	0.049	11:31																																						
Air Blank	0.000	11:32																																						
Control Test	0.049	11:33																																						
Air Blank	0.000	11:33																																						
Control Test	0.049	11:34																																						
Air Blank	0.000	11:35																																						
Control Test Stats																																								
Average	0.0490																																							
Std Dev	0.0000																																							
Rel Std Dev(%)	0.0000																																							

Florida Department of Law Enforcement Alcohol Testing Program

DEPARTMENT INSPECTION REPORT - INTOXILYZER 8000

Agency: FORT MYERS PD
Time of Inspection: 13:28

Date of Inspection: 10/23/2024

Serial Number: 80-005468
Software: 8100.27

Check or Test	YES	NO	Check or Test	YES	NO
Diagnostic Check (Pre-Inspection): OK	Yes		Date and/or Time Adjusted		No
Minimum Sample Volume Check: OK	Yes		Barometric Pressure Sensor Check: OK	Yes	
Alcohol Free Subject Test: 0.000	Yes		Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes		Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303K Exp: 03/29/2025	0.08g/210L Test (g/210L) Lot#:202303L Exp: 03/29/2025	0.20g/210L Test (g/210L) Lot#:202304C Exp: 04/05/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:01923080A3 Exp: 02/05/2025
0.000	0.049	0.079	0.196	0.079
0.000	0.050	0.079	0.196	0.078
0.000	0.049	0.078	0.197	0.078
0.000	0.049	0.079	0.197	0.079
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.079
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.197	0.078
0.000	0.049	0.079	0.196	0.078
0.000	0.049	0.079	0.196	0.078

Standard Deviations	0.0003	0.0003	0.0004	0.0004
---------------------	--------	--------	--------	--------

Average Standard Deviation of 0.05, 0.08 and 0.20 g/210L Tests: 0.0003 Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



TAYLOR D GUTSCHOW

Signature and Printed Name

10/23/2024
Date



Calibration Certificate

Florida Department of Law Enforcement
Alcohol Testing Program
4700 Terminal Drive, Suite 1
Ft. Myers, FL 33907

This is to certify the calibration of Intoxilyzer 8000 serial number 80-005468, manufactured by CMI, Inc. was calibrated in accordance with FDLE/ATP Form 36 - Department Inspection Procedures - Intoxilyzer 8000.

Serial Number:	<u>80-005468</u>	UNCERTAINTY* \pm
Owning Agency:	<u>FORT MYERS PD</u>	0.050 g/ 210 L 0.004
Calibration Date:	<u>10/23/2024</u>	0.080 g/ 210 L 0.004
Calibration Time:	<u>13:28</u>	0.200 g/ 210 L 0.007
		0.080 g/ 210 L Dry Gas Control 0.005

All results are reported in g/ 210 L.

Bias is limited by calibration acceptance criteria. All calibration results must be within ± 0.005 or 5%, whichever is greater, of the target alcohol concentration.

*Uncertainty is based on fleet-wide data and is expressed to a 99.73% level of confidence ($k=3$).

The instrument results before and after any adjustment are found in the associated pre and post stability checks.

TRACEABILITY INFORMATION

This instrument was calibrated using solutions prepared by Alcohol Countermeasure Systems, Inc. (ACS). ACS prepared and certified these CRMs in accordance with ISO 17034 and ISO/ IEC 17025 Standards.

Simulator temperatures are traceable to NIST. Simulator temperatures are checked with NIST traceable digital thermometers calibrated by Precision Metrology in accordance with ISO/ IEC 17025 standards.

Dry gas control measurements are traceable to NIST through the use of CRMs supplied by an accredited CRM supplier. The supplier of dry gas standard controls prepared and certified the CRMs in accordance with ISO Guide 34 and ISO/ IEC 17025 standards. This document shall not be reproduced except in full, without written approval of the Florida Department of Law Enforcement Alcohol Testing Program.

FDLE/ATP Form 69 December 2021

Issuing Authority: Alcohol Testing Program

Date

10/23/2024

TAYLOR D GUTSCHOW,

Department Inspector

Service • Integrity • Respect • Quality