

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT – INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 11:27

Date of Inspection: 03/12/2025

Serial Number: 80-000892

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08g/210L Test (g/210L) Lot#: 202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 402783528 Exp: 06/27/2026
0.000	0.048	0.077	0.198	0.080
0.000	0.048	0.078	0.198	0.080
0.000	0.048	0.078	0.198	0.080

Number of Simulators Used: 4

Remarks:

INPUT WRONG LOT # AND EXP DATE

for .05 solution.

The correct solution was used (.05) but the wrong lot # and Exp. date was inputted.

correct Lot # is 202303J

correct Exp Date: 03/28/25 Use Basford, N362

The above instrument complies (X) does not comply () with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Lt. Lise Basford N362

NICOLE M BASFORD

Signature and Printed Name

03/12/2025
Date