

Florida Department of Law Enforcement

Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC TAMPA

Serial Number: 80-000901

Time of Inspection: 18:59

Date of Inspection: 01/06/2025

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#: 202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 24923080A3 Exp: 10/05/2025
0.000	0.047	0.000 / 0.080	0.197	0.079
0.000	0.048	0.000 / 0.079	0.197	0.078
0.000	0.049	0.000 / 0.079	0.198	0.079


Number of Simulators Used: 5

Remarks:

08: Control Outside Tolerance. *SIMULATOR WAS NOT ATTACHED.*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


Signature and Printed Name

DAMON J PULASKI

01/06/2025
Date