

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: TITUSVILLE PD

Time of Inspection: 01:08

Date of Inspection: 01/24/2025

Serial Number: 80-000996

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	


Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#:202410F Exp: 10/30/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:231592 Exp: 11/05/2025
0.000	0.049	0.080	0.200	0.081
0.000	0.050	0.080	0.201	0.081
0.000	0.049	0.080	0.201	0.081

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.


167
MATTHEW S RASMUSSEN

 Signature and Printed Name

01/24/2025
 Date