

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: PONCE INLET PD
Time of Inspection: 17:08

Date of Inspection: 01/26/2025

Serial Number: 80-001158
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#:202410F Exp: 10/30/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:402558895 Exp: 10/03/2025
0.000	0.047	0.077	0.197	0.080
0.000	0.047	0.078	0.198	0.080
0.000	0.047	0.078	0.198	0.081

Number of Simulators Used: 4

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 PI B34

SCOTT B GASTON
Signature and Printed Name

01/26/2025
Date