

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: NAS PENSACOLA PD  
Time of Inspection: 20:06

Date of Inspection: 02/01/2025

Serial Number: 80-001205  
Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202303J<br>Exp: 03/28/2025 | 0.08g/210L Test (g/210L)<br>Lot#:202307A<br>Exp: 07/11/2025 | 0.20g/210L Test (g/210L)<br>Lot#:202307C<br>Exp: 07/12/2025 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:223083<br>Exp: 04/05/2025 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------------------|
| 0.000                      | 0.047                                                       | 0.076                                                       | 0.189 / 0.192                                               | 0.080                                                                   |
| 0.000                      | 0.047                                                       | 0.076                                                       | 0.191 / 0.193                                               | 0.081                                                                   |
| 0.000                      | 0.048                                                       | 0.077                                                       | 0.192 / 0.193                                               | 0.081                                                                   |

Number of Simulators Used: 5

**Remarks:**

20: Control Outside Tolerance.

*RETESTED 0.20g/210L.*

The above instrument complies ( ☒ ) does not comply ( ☐ ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



KEITH H KELLEY

Signature and Printed Name

02/01/2025  
Date