

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: OSCEOLA COUNTY SO
Time of Inspection: 17:27

Date of Inspection: 01/20/2025

Serial Number: 80-003935
Software: 8100.27

| Check or Test | YES | NO |
|---|-----|----|
| Date and/or Time Adjusted | | No |
| Diagnostic Check (Pre-Inspection): OK | Yes | |
| Alcohol Free Subject Test: 0.000 | Yes | |
| Mouth Alcohol Test: Slope Not Met | Yes | |
| Interferent Detect Test: Interferent Detect | Yes | |
| Diagnostic Check (Post-Inspection): OK | Yes | |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025 | 0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025 | 0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025 | 0.08 g/210L Dry Gas Std Test (g/210L) Lot#:403207507 Exp: 11/26/2027 |
|----------------------------|---|---|---|--|
| 0.000 | 0.048 | 0.079 | 0.197 | 0.082 |
| 0.000 | 0.048 | 0.080 | 0.197 | 0.082 |
| 0.000 | 0.049 | 0.079 | 0.197 | 0.082 |

Number of Simulators Used: 5

Remarks:

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

DANIEL E LYONS

Signature and Printed Name

01/20/2025
Date