

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: LAKELAND PD

Time of Inspection: 13:01

Date of Inspection: 01/29/2025

Serial Number: 80-003945

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#: 202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#: 202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#: 202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#: 14323080A3 Exp: 09/05/2025
0.000	0.048	0.078	0.194	0.081
0.000	0.050	0.079	0.196	0.081
0.000	0.050	0.080	0.196	0.081

Number of Simulators Used: 5

Remarks:

A F / M A: RFI Detect. *RADIO ENTERED ROOM. RADIO REMOVED, MET STANDARD. 01-29-25*

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

CAMILO J ALMEIDA

Signature and Printed Name

01/29/2025

Date