

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 21:34

Date of Inspection: 03/31/2025

Serial Number: 80-005249

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202410E Exp: 10/22/2026	0.20g/210L Test (g/210L) Lot#:202410F Exp: 10/30/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.048	0.080	0.199	0.080
0.000	0.049	0.081	0.200	0.079
0.000	0.049	0.081	0.200	0.080

Number of Simulators Used: 5

Remarks

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

 PATRICK M BAKER

Signature and Printed Name

03/31/2025
Date

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: HERNANDO COUNTY SO

Time of Inspection: 01:19

Date of Inspection: 02/28/2025

Serial Number: 80-005249

Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202402E Exp: 02/27/2026	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202410F Exp: 10/30/2026	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:28423080A3 Exp: 11/05/2025
0.000	0.048	0.078	0.200	0.078
0.000	0.049	0.079	0.201	0.079
0.000	0.049	0.079	0.201	0.078

Number of Simulators Used: 5

Remarks

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

PATRICK M BAKER

Signature and Printed Name

02/28/2025
Date