

Florida Department of Law Enforcement Alcohol Testing Program

AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: COLLIER COUNTY SO
Time of Inspection: 10:57

Date of Inspection: 02/05/2025

Serial Number: 80-007077
Software: 8100.27

Check or Test	YES	NO
Date and/or Time Adjusted		No
Diagnostic Check (Pre-Inspection): OK	Yes	
Alcohol Free Subject Test: 0.000	Yes	
Mouth Alcohol Test: Slope Not Met	Yes	
Interferent Detect Test: Interferent Detect	Yes	
Diagnostic Check (Post-Inspection): OK	Yes	

Alcohol Free Test (g/210L)	0.05g/210L Test (g/210L) Lot#:202303J Exp: 03/28/2025	0.08g/210L Test (g/210L) Lot#:202307A Exp: 07/11/2025	0.20g/210L Test (g/210L) Lot#:202307C Exp: 07/12/2025	0.08 g/210L Dry Gas Std Test (g/210L) Lot#:20524080A2 Exp: 09/05/2026
0.000	0.044 / 0.047	0.079	0.198	0.079
0.000	0.044 / 0.048	0.080	0.199	0.078
0.000	0.045 / 0.047	0.080	0.198	0.079


Number of Simulators Used: 5

Remarks:

05 Control Outside Tolerance. ORING ON SIMULATOR RETURN REPLACED and tested w/in range on the second attempt. LG #3223

The above instrument complies (☒) does not comply (☐) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.



YVETTE R GONZALES

Signature and Printed Name

02/05/2025
Date