

# Florida Department of Law Enforcement Alcohol Testing Program

## AGENCY INSPECTION REPORT - INTOXILYZER 8000

Agency: FFWCC

Time of Inspection: 11:48

Date of Inspection: 03/12/2025

Serial Number: 80-007498

Software: 8100.27

| Check or Test                               | YES | NO |
|---------------------------------------------|-----|----|
| Date and/or Time Adjusted                   |     | No |
| Diagnostic Check (Pre-Inspection): OK       | Yes |    |
| Alcohol Free Subject Test: 0.000            | Yes |    |
| Mouth Alcohol Test: Slope Not Met           | Yes |    |
| Interferent Detect Test: Interferent Detect | Yes |    |
| Diagnostic Check (Post-Inspection): OK      | Yes |    |

| Alcohol Free Test (g/210L) | 0.05g/210L Test (g/210L)<br>Lot#:202303J<br>Exp: 03/28/2025 | 0.08g/210L Test (g/210L)<br>Lot#:202307A<br>Exp: 07/11/2025 | 0.20g/210L Test (g/210L)<br>Lot#:202307C<br>Exp: 07/12/2025 | 0.08 g/210L Dry Gas Std Test (g/210L)<br>Lot#:402648278<br>Exp: 01/27/2026 |
|----------------------------|-------------------------------------------------------------|-------------------------------------------------------------|-------------------------------------------------------------|----------------------------------------------------------------------------|
| 0.000                      | 0.049                                                       | 0.079                                                       | 0.197                                                       | 0.077                                                                      |
| 0.000                      | 0.049                                                       | 0.079                                                       | 0.197                                                       | 0.077                                                                      |
| 0.000                      | 0.049                                                       | 0.079                                                       | 0.197                                                       | 0.077                                                                      |

Number of Simulators Used: 4

Remarks:

NAIT IS IN COMPLIANCE. ACCIDENTALLY HIT NO..

\* Instrument is in compliance as seen above.  
When the instrument asked if it was in compliance  
I accidentally selected NO. The instrument is in  
compliance.

The above instrument complies ( ) does not comply ( X ) with Chapter 11D-8, FAC.

I certify that I hold a valid Florida Department of Law Enforcement Agency Inspector Permit and that I performed this inspection in accordance with the provisions of Chapter 11D-8, FAC.

Lt. Nicole M. Basford N342 NICOLE M BASFORD  
Signature and Printed Name

03/12/2025  
Date