



Florida Department of
Law Enforcement

James T. "Tim" Moore
Commissioner

Criminal Justice
Professionalism Program
Alcohol Testing Program

P.O. Box 1489
Tallahassee, Florida 32302
(850) 410-8304
(850) 488-1760 Fax
<http://www.fdle.state.fl.us>

March 18, 2002

Mr. Russ Arrend
Institute of Police Technology and Management
University of North Florida
12000 Alumni Drive
Jacksonville, Florida 32224

SUBJECT: Technical Advisory Committee on DUI Enforcement and Prosecution
Subcommittee on Breath Test Instrumentation

Dear Mr. Arrend:

Enclosed is the information concerning the recommendation of the Technical Advisory Committee on DUI Enforcement and Prosecution, Subcommittee on Breath Test Instrumentation, for a new breath test instrument for approval for use in the state of Florida.

A meeting was held on January 29 and 30, 2002, to review the results of field testing and to recommend a breath test instrument to undergo the approval process. This meeting was a follow up to a meeting held in July 2001, where subcommittee members had an opportunity to receive basic information, view the instruments, and ask questions of instrument representatives.

As a result of the above process, the subcommittee has unanimously recommended the CMI Intoxilyzer 8000 for approval by the Florida Department of Law Enforcement.

If you have any questions regarding the above information, please feel free to contact me at (850) 410-7814.

Sincerely,

James T. Moore
Commissioner

Laura D. Barfield
Chairman
Subcommittee on Breath Test Instrumentation

LDB

Enclosures

IPTM Technical Advisory Committee
Report of the Subcommittee on Breath Test Instrumentation
Maitland, Florida

Subcommittee Members Present

Laura Barfield	Florida Department of Law Enforcement (Chairman)
Rafael Madrigal	Florida Department of Law Enforcement
Jay Zager	Broward County Sheriff's Office
Roger Doherty	Florida Department of Transportation
Don Hartery	Office of the State Attorney-12 th Circuit
Bob Sondel	Institute of Police Technology and Management
Lisa Hunter-Sellers	Florida Highway Patrol (for Ron Castleberry)
Jimmy Branch	Jacksonville Sheriff's Office
Ira Karmelin	Office of the State Attorney-15 th Circuit

Tuesday, January 29, 2002

Time: 1:00 p.m. until 5:00 p.m.

Topic of Discussion: Recommendation of a breath test instrument for Approval
Evaluation for use in the state of Florida.

Presentation of Field Testing Data: Roger Skipper presented field testing data obtained from five different evidentiary breath test instruments including: the Datamaster CDM, the Drager 7110 Mark IIIC, the Intoxilyzer 8000, the Drager 7410 EPAS, and the Intoximeter Alcosensor IV-XL @ POA. A copy of the PowerPoint presentation and the Field Test Report is included with this report. Performance issues included portability, computer networking of data, use in marine environments, interferent detection, and practical usage at two DUI checkpoints in Brevard and Orange counties.

There was a general discussion on software and remote accessibility before the presentation. All instruments had this capability. This discussion was started based on questions from the subcommittee members.

The presentation included question and answer periods for each performance issue addressed.

Following the presentation, all members of the subcommittee participated in a discussion with extensive question and answer sessions. Questions included: fuel cell recovery time and if the entire test (including control tests) could be performed within 15 minutes as required by the rule. Fuel cell recovery time at higher alcohol concentrations was a problem during the field testing. The effect of humidity on some instruments during the DUI Checkpoint

field tests and the instrument locking up was of particular concern. Dual technology capability (fuel cell and infrared) was a concern and deemed undesirable by consensus.

Recommendation of the Subcommittee: A motion was made to recommend the CMI Intoxilyzer 8000 as the most reliable, practical, and best performer for use in Florida. The motion was seconded and the subcommittee voted unanimously to recommend the CMI Intoxilyzer 8000 for approval evaluation as prescribed in Chapter 11D-8, FAC.

Wednesday, January 30, 2002

Time: 9:00 a.m. until 11:00 a.m.

Topic of Discussion: Anticipating approval of the CMI Intoxilyzer 8000, the subcommittee discussed rule requirements and implementation procedures for a new instrument. This was a roundtable session to share general ideas and concerns on how to effectively and efficiently obtain scientifically reliable and legally admissible breath test results.

Recommendations from the Subcommittee:

- ✓ Continue to perform monthly agency inspections of the new instrument using dry gas only at three different levels (0.05, 0.08, and 0.20 g/210L).
- ✓ Continue to inspect the Intoxilyzer 5000 in the current manner, using the same procedures and forms for department inspections, agency inspections and evidentiary breath tests.
- ✓ Discontinue approval of alcohol reference solution. There is sufficient data that has been collected to show that the solutions are of the correct concentration and do not need an approval analyses before distribution. The solutions should remain to be certified by the manufacturer and only certified solutions should be distributed after the Alcohol Testing Program reviews the data regarding the certification of a lot. Complete a "Form 32" showing that the data was reviewed, the solutions are certified and that they are approved for use in Florida.
- ✓ Use the certified concentration from the manufacturer of dry gas. Review data from the manufacturer and complete a "Form 32" stating that the data has been reviewed by FDLE/ATP, the dry gas lot has been certified, and the lot is approved for use in Florida.
- ✓ Performing hands-on inspections of the new instrument by the Regional Alcohol Breath Test Inspector (FDLE). A question arose that was not answered during the discussion: Should this "hands-on" FDLE inspection be the same procedural steps that are taken by the agency inspector during an agency inspection?

- ✓ Address concerns regarding the remote access feature of the new instrument and the Regional Alcohol Breath Test Inspector's ability to remotely access a breath test instrument to check its operation and calibration.

Florida Department of Law Enforcement

Alcohol Testing Program

Florida Department of Law Enforcement

Field Testing of Evidential Breath Test Instruments

Instruments Tested

- **CMI Intoxilyzer 8000**
- **Drager Alcotest 7110 Mark III C**
- **Drager Alcotest 7410 EPAS**
- **Intoximeter Alcosensor IV-XL**
- **National Patent Datamaster .cdm**

Order of listing does not imply a ranking of any kind.

Technology

- **7110 - Infrared and fuel cell**
- **8000 - Infrared only**
- **7410 - Fuel cell only**
- **IV-XL - Fuel cell only**

Alco Sensor IV-XL @POA

- **System weight: approx 10 lbs.**
- **Operating Temp: 0 - 104 degrees F**
- **Range: .000 - .400 g/210 L**
- **Network compliant**
- **Operates on 120 VAC, 12 VDC, or internal batteries**

Drager Alcotest 7110 MK III C

- **System weight: 16.5 lbs**
- **Operating Temp: 32-104 degrees F**
- **Range: .000 - .450 g/210L**
- **Network compliant**
- **Operates on 120 VAC or 12 VDC**

CMI Intoxilyzer 8000

- **System weight: 17 lbs**
- **Operating Temp: 32 - 104 degrees F**
- **Range: .000 - .600 g/210L**
- **Network compliant**
- **Operates on 120 VAC or 12 VDC, or external battery**

Dräger 7410 EPAS

- **System weight: 14 lbs**
- **Operating Temp: 23 - 104 degrees F**
- **Range: .000 - .400 g/210L**
- **Network compliant**
- **Operates on 120 VAC, 12 VDC, or internal batteries**

National Patent Datamaster .cdm

- **System weight: 30 lbs.**
- **Operating Temp: 65 - 78 degrees F**
- **Range: .000 - .600 g/210L**
- **Network compliant**
- **Operates on 120 VDC**

Instrument Field Test

- **Portability**
- **Officer Safety/Ease of Use**
- **Interferent Rejection**

Portability

- Each instrument was tested for portability
- Each instrument was tested for ease of use and officer safety.
- Each instrument was tested for ergonomics

Portability

- **7110: Acceptable**
- **8000: Acceptable**
- **7410: Acceptable**
- **Alco IV-XL: Acceptable**
- **.cdm: Not acceptable**

.cdm Portability Issues

- **Weight: Heavy and bulky.**
- **Ergonomics: No carry handles**
- **Power: Requires 120 VAC. Must use inverter or generator in field use.**
- **Temp: Limited use range.**
- **Overall, instrument designed for use in a facility designed to overcome ambient conditions and power issues.**

.cdm Portability

- **No further testing was conducted on the .cdm.**

Officer Safety/Ease of Use

- Each instrument was tested on number of steps required by officer to complete a test sequence.
- Number of steps can directly relate to divided attention between subject and instrument.

Manual Actions Required to Complete One Test

- 7110: 7
- 7410: 15
- 8000: 7
- IV-XL: 20

Measured from pushing the first button to retrieving the printed result

Officer Safety/Ease of Use

- **7410 and Alco VI-XL requires the dry gas bottle to be physically connected and disconnected during each control test.**

Interferent Evaluation

- Each instrument tested for reaction to acetone, methanol, n-propanol, isopropanol, n-butanol, xylene, and an industrial “cocktail”
- “Cocktail” consisted of a mixture of methanol, methylene chloride, toluene, and acetone.
- “Cocktail” used to simulate an individual exposed in an industrial environment

.03 Acetone

- **7110: .000 but not flagged**
- **8000: .000 and flagged**
- **ALCO IV: .000 But not flagged**
- **7410: .00 but not flagged**

.08 ETHOL with .03 acetone

- **.08 ETHOL was ran first to establish baseline.**
- **Next, .03 ml of acetone stock was added to see if detected and compensated for.**

Test Results

- **7110 did not add or flag the acetone**
- **8000 did not add and flagged**
- **Alco VI did not add or flag the acetone**
- **7410 did not add or flag the acetone**

.10 N-Propanol (LD-50)

- **7110: Interferent**
- **8000: .025 average and not flagged**
- **Alco IV: VOID**
- **7410: .04 average and not flagged**

.04 N-propanol

- **7110: IR .024 average EC .031 average and not flagged**
- **8000: Interferent**
- **Alco VI: .031 average and not flagged**
- **7410: .03 average and not flagged**

.24 Isopropanol (LD-50)

- **7110: Interferent**
- **8000: Interferent**
- **Alco IV: .110 average and not flagged**
- **7410: .11 average and not flagged**

.10 Isopropanol

- **7110: Interferent**
- **8000: Interferent**
- **Alco IV: .047 average and not flagged**
- **7410: .03 average and not flagged**

.04 Methanol (LD-50)

- **7110: Interferent**
- **8000: Interferent**
- **Alco IV: VOID**
- **7410: .03 average but not flagged**

.02 Methanol

- **7110: Interferent**
- **8000: .000**
- **Alco IV: VOID**
- **7410: .01 average but not flagged**

Xylene

- 7110: mouth alcohol
- 8000: flagged and locked
- Alco IV: would not take sample
- 7410: would not take sample

Cocktail

- 7110: Interferent
- 8000: .000/.000
- Alco IV: .000/VOID
- 7410: .06 average and did not flag

N-Butanol

- 7110: mouth alcohol
- 8000: mouth alcohol
- Alco IV: did not read sample
- 7410: did not read sample

RADAR/LIDAR

- **RADAR and LIDAR did not appear to affect the instruments.**

Field Use Comments

- **Used in Brevard and Orange County Roadblocks**
- **Volunteer subject tests**
- **Officer use**
- **Ambient condition effects**

Florida Department of Law Enforcement

- **This concludes this presentation.**
- **Questions?**



Florida Department of
Law Enforcement

Alcohol Testing Program


700 South Babcock Street
Suite 401
Melbourne, FL 32901
(407) 984-4880
Fax (407) 984-4889

James T. "Tim" Moore
Commissioner

March 13, 2002

MEMORANDUM

TO: Laura Barfield, Chairman
Technical Advisory Committee on DUI Enforcement and Prosecution
Subcommittee on Breath Test Instrumentation

FROM: Roger Skipper 
Regional Alcohol Breath Test Inspector
FDLE Alcohol Testing Program

SUBJECT: Field Testing of Evidential Breath Test Instruments

During the period of September 2001 through January 2002 field tests on five evidential breath test instruments were conducted for the Technical Advisory Committee on DUI Enforcement and Prosecution. The five instruments were the *Intoxilyzer 8000* (CMI, Inc.), *Alcotest 7110 MK III C* (Drager Safety, Inc.), *Alcotest 7410 EPAS* (Drager Safety, Inc.), *Alco Sensor IV-XL P.O.A.* (Intoximeter, Inc.), and *DataMaster cdm* (National Patent, Inc.).

The criteria used for the basis of the field testing are contained in the Instrument Field Testing Procedures, which is attached.

The instruments were tested in the workshop located at the Melbourne Field Office and at two D.U.I. checkpoints, one in Brevard County and one in Orange County.

Attached you will find the results of the field testing process. This document will summarize each item of the criteria used as the basis for testing, and documents the information presented to the subcommittee on January 30, 2002.

RGS/rgs

3 Atchmts: Testing Procedure; Testing Results; Test Data Sheet

Committed to
Service - Integrity - Respect - Quality

**INSTRUMENT FIELD TESTING PROCEDURES
FLORIDA DEPARTMENT OF LAW ENFORCEMENT
ALCOHOL TESTING PROGRAM**

The following criteria will be used to determine the suitability of an instrument for use in the State of Florida, and to recommend the instrument for formal approval evaluation. This criteria is based on the requirements of existing Florida law, the needs of local law enforcement agencies, and available technology.

1. Portability. The instrument must be portable and capable of operating on land and on water. The selected instrument must possess the following traits:

a. Weight. Light in weight enough to be easily transferred from vehicle to vehicle, across vessel docks, and from fixed facilities to mobile facilities. Light in weight enough to be easily removed from vehicle trunks and other storage areas.

b. Power. Uses power from a variety of sources such as 110 VAC, 12 VDC, and internal/external battery.

c. Ruggedness. Not easily damaged from bumps and shocks normally encountered in a mobile environment. Moisture proof or moisture resistant.

2. Accuracy and Precision. Must conform to current Federal standards and appear on the most recent U. S. Department of Transportation Conforming Products List of Evidential Breath Measurement Devices. Accuracy must be better than 0.005 g/210L or 5%, whichever is greater, and precision must be better than 0.0042 or less.

3. Ease of Use. Efficient, practical and convenient use in a mobile environment without compromising officer/operator safety.

a. Display should be easily read in a variety of lighting environments.

b. The operation prompts should be clear and easily understood.

c. Recycle time between samples and tests should be as short as possible. The sampling time for the current instrument for an evidential test is approximately 4.5 minutes, calculated from pushing the *Start Test* button to the print card being ejected.

4. Communications and Software. The ability to store, transmit, and receive test data to/from a remote computer by modem, Internet, or both. The ability to record, store and print demographic and historical data. The ability for certain portions of the software, such as demographics; to be edited by authorized persons (highly desirable but not required).

Attachment 1.1

5. Control Test. The ability to conduct control tests with either a wet bath simulator or dry gas. The ability to insert control tests in the subject test protocol. The ability for authorized persons to choose the number of control tests for calibration check purposes (highly desirable but not required).
6. Strip or barcode compliant. The ability, through software and hardware, to read magnetic strip or single/double/triple level bar code data from driver licenses and identification/user cards.
7. Environmental Operating Conditions. The ability to operate in a broad ambient temperature and humidity conditions. The greatest possible temperature and humidity operating ranges are preferable.
8. Interfering Substances. The ability to detect and flag substances other than ethanol, and prohibit subsequent sampling after such detection. Must not give a perceived alcohol (ethanol) result when subjected to interfering substances.
9. Radio Frequency Interference Detection. The ability to detect and cancel tests when radio interference is present, OR, must not be affected by radio interference.

Attachment 1.2

Committed to
Service - Integrity - Respect - Quality

RESULTS OF INSTRUMENT FIELD TESTING

ABSTRACT: Five evidential breath test instruments were evaluated from September 27, 2001 through January 28, 2002. One, the National Patent Systems Datamaster .cdm, was determined to be "not acceptable" after the portability tests due to its dependence on 120 VAC current and its excessive weight (26 pounds). It was not subjected to any further testing. Four other instruments, Drager Alcotest 7110 MKIIIC, CMI Intoxilyzer 8000, Intoximeter Alco Sensor IV POA, and Drager Alcotest 7410 EPAS, were tested using the criteria stated in the Instrument Field Testing Procedures.

1. Portability. All instruments, with the exception of the Datamaster .cdm, are portable. The remaining four are less than 17 pounds, use standard 120 VAC current, 12 VDC current, and/or battery power. All are enclosed in pelican type cases or have hardened cases.
2. Accuracy and Precision: All meet or exceed the standard of accuracy and precision as defined in the procedures. The Drager 7110, because of its dual technology sampling system, allows a 0.008 g/210L difference between the infrared analysis and fuel cell analysis before invalidating the test, but reports only the lowest reading of the two.
3. Ease of Use. The Drager 7110 and the Intoxilyzer 8000 requires seven individual steps to complete a breath test sequence as defined in Rule 11D-8, F.A.C. The Drager 7410 requires fifteen steps, and the Intoximeter POA requires 20 steps. The steps are counted from the input of data to the removal of the printed results. The Drager 7410 and Intoximeter POA require additional steps due to having to connect and disconnect mouthpieces and control test gas cylinders, not required for the first two.

The display on the Intoxilyzer 8000 is clear and easy to read from several angles and in low light. The display on the Drager 7110 is clear when viewed directly. It tends to "washout" when viewed at an angle. The display on the handheld components of the Drager 7410 and Intoximeter POA are clear when viewed from several angles, but the display on the computer components of these two instruments "washout" when viewed at an angle, and when viewed in low light/glare conditions.

The Drager 7410 and Intoximeter POA use a small stylus to assist in inputting data. The Intoximeter has a small keyboard also. The Drager 7410 depends entirely on the stylus for data input. The stylus was easily misplaced, rendering the instrument unusable for data input until the stylus was recovered.

The recycle time between test sequences varied between the instruments, from a maximum of 16 minutes on the Drager 7410 to seven minutes on the Intoxilyzer 8000. The process for this test was the introduction of a 0.30 g/210L ethanol sample, then a 0.20 g/210L ethanol sample to cause a "no 0.020" agreement, and then a third sample consisting of another 0.30 g/210L ethanol sample.

Attachment 2.1

4. Communications and Software: All four instruments satisfied the communications and software criteria.
5. Control Test. All four instruments support both wet bath and dry gas control tests. However, the Drager 7410 and Intoximeter POA required the simulator or dry gas cylinder to be attached and detached for each control test. This is an automatic function in the Drager 7110 and the Intoxilyzer 8000.
6. Strip or Barcode Compliant. All four instruments accept strip or barcode readers.
7. Environmental Operating Conditions. All four instruments operate in a temperature range of 30 to 104 degrees Fahrenheit and humidity levels from 95 to 100%. On December 7, 2001, the instruments were tested at a DUI Checkpoint in Palm Bay, Brevard County, Florida. Ambient temperature during the period was an average of 65 degrees Fahrenheit. Relative humidity was approximately 90% and condensing, with a light mist. The Drager 7110 and Intoximeter POA ceased to function after being exposed to these conditions after approximately two hours. Both had to be taken inside and allowed to dry out. After drying, both worked properly. The Intoxilyzer 8000 and Drager 7410 functioned properly throughout the evening. No instrument seemed to be affected by exhaust fumes generated by vehicles passing through the checkpoint.
8. Interfering Substances. All four instruments reacted to interfering substances in slightly different ways. See Attachment #1 for details.
9. Radio Interference Detection. All instruments were radiated with kA band police radar, LIDAR (laser speed gun), and an 800mhz police radio. All performed satisfactorily.

General Comments. The instruments were tested at two DUI checkpoints, one in Brevard County (see para. 7) and one in Orange County in the City of Apoka on December 14, 2001. During these two checkpoints, police officers and civilian technicians were allowed to use the instruments in non-evidential testing and their comments were solicited. The majority of the officers preferred the Intoxilyzer 8000, with the Drager 7110 next, the Drager 7410 third, and the Intoximeter POA fourth. When asked why they preferred the Intoxilyzer 8000 to the other three, the most common response was short recycle time between tests on the instrument and ease of use.

Attachment 2.2

Committed to
Service - Integrity - Respect - Quality

TEST DATA SHEET

TEST	7110	8000	ALCO IV	7410**
.03 ACET	.000/.000 NF .000/.000 NF .000/.000 NF	.000/.000 .000/.000 .000/.000	.000/.000 NF .000/.000 NF .000/.000 NF	.00/.00 NF .00/.00 NF .00/.00 NF
.08 ETHOL	IR:.075/.078 EC:.076/.078	.080/.077	.075/.075	.07/.07
.08 ETHOL W/ .03 ACET	IR:.079/.078 NF EC:.078/.078 NF	.085/.084	.073/.075 NF	.07/.07 NF
.10 N-PROP	INF INF INF	.024/.025 NF .024/.025 NF .026/.026 NF	VOID VOID VOID	.04/.03 NF .04/.04 NF .04/.04 NF
.04 N-PROP	IR:.024/.024 NF EC:.031/.031 NF IR:.024/.024 NF EC:.032/.032 NF IR:.024/.024 NF EC:.030/.030 NF	INF INF INF	.032/.031 NF .032/.032 NF* .031/.031 NF*	.03/.03 NF .03/.03 NF .03/.03 NF
.24 ISOP	INF INF INF	INF INF INF	.112/.111 NF .112/.109 NF* .108/.110 NF*	.11/.11 NF .11/.11 NF .11/.11 NF
.10 ISOP	INF INF INF	INF INF INF	.048/.048 NF .047/.047 NF* .047/.046 NF*	.04/.04 NF .04/.04 NF .04/.04 NF
.04 METH	INF INF INF	INF INF INF	VOID VOID VOID	.04/.04 NF .04/.03 NF .03/.03 NF
.02 METH	INF INF INF	.000/.000 .000/.000 .000/.000	VOID VOID VOID	.02/.02 NF .02/.01 NF .01/.02 NF
XYLENE	MOUTH ALCO	INF	DNRS	DNRS

Attachment 3.1

Committed to
Service - Integrity - Respect - Quality

TUFF	INF	.000/.000	.000/VOID	.07/.05 NF
N-BUTANOL	MOUTH ALCO	MOUTH ALCO	DNRS	DNRS
RFI	NO EFFECT	NO EFFECT	NO EFFECT	NO EFFECT

LEGEND:

* = took from 10 to 23 minutes to reset for next series of tests

** = instrument not set to read third didget

NF = not flagged as interfering substance

DNRS = did not read or recognize sample as breath. remained at sample request step.

TUFF = BIX® "Tuff-Job" brand industrial stripper mixture of methanol (67-56-1), toluene (108-88-3), methylene chloride (75-09-2), and acetone (67-64-1)

ACET = acetone

ETHOL = ethanol

ISOP = isopropanol

N-PROP = n-propanol

METH = methanol

Simulators were prepared with distilled water and proper amount of prepared stock solution for value. The n-butanol, xylene, and Tuff tests were conducted by placing a droplet of the substance in a dry, cold simulator.

Instruments were operated using the manufacturers' operator and/or technician's manuals.

Attachment 3.2

Skipper, Roger

From: Barfield, Laura
Sent: Saturday, November 17, 2001 5:19 PM
To: Skipper, Roger
Subject: Dilutions for Various Solutions

Methanol Stock (10g/100mL): 126 mL methanol in 1000 mL volumetric flask
Isopropanol Stock (10g/100mL): 127 mL isopropanol in 1000 mL volumetric flask
n-Propanol Stock (10g/100mL): 124 mL n-propanol in 1000 mL volumetric flask
n-Butanol (10g/100mL): 124 mL n-butanol in 1000 mL volumetric flask

Concentration(g/100mL)	Volumetric Flask	Volume Stock to add
0.10	500 mL	5 mL
0.08	500 mL	4 mL
0.06	500 mL	3 mL
0.04	500 mL	2 mL
0.02	500 mL	1 mL

The LD50 for methanol is 0.05 g/100mL

The LD50 for isopropanol is 0.25 g/100mL.

I will need to research n-propanol and n-butanol further. I can only find the LD50 for rats!

Let me know if you have any questions on this...Laura

Laura D. Barfield
Florida Department of Law Enforcement
Alcohol Testing Program
Program Manager
P.O. Box 1489
Tallahassee, Florida 32302
(850) 410-8304

8000

WALT - PRINTER COVER - SCREW NEEDS TO BE REPLACED W/ SOMETHING ELSE, BUTTON, SNAP OR THE LIKE. ALLOWS BREATH TEST W/O PAPER IN PRINTER, CAN TESTS BE RECALLED.

- Air Blow 20 SECS.
- Remove "Subj. Signature" line on print out
- "FDHE Implied Consent, Program" - Prints this now
- "Please Blow Times Out at 2 mins"
- "Please Blow Until Tare Stops"
- Recommend display showing "Please blow until told to stop"
- No Blow = $\overline{026}$ (Deficient Sample)
Recommend (LSB) for low Sample Volume
- Paper test mechanism allows paper to fall back into slot, requiring lid to be inserted.
- When using 105L BOTTLE, GAS tube + ELECTRICAL connex is too short
- Change "Markt" to Accuracy check
- Can test be recalled from unit or is ext computer needed

11/21 DOWNLOAD NEW PROGRAM

- NO 2 MIN WAIT (ADD)

SOFTWARE change

IN ACABABAB(CA)CA

~~ACABABAB(CA)~~ NO 2min WAIT BETWEEN samples, butt, goes into wait and stays there (at least 10 minutes)

in ABABA(BA) NO 2min WAIT

DATA ENTRY Deactivated, will not turn on

DOS not print Mouth Alcohol in TICKET.

Audible RFI ALGRT

010202 when :02 meth introduced then BT, would not accept sample & stayed at "Please Blow". when simulator disconnected & Breath introduced then took sample.

CMI / B. Taylor - called

ACABAWABA(020)CA

PAS IV

UACT - 4 yrs use of Alco-IV in .02 enforcement -
NOT RELIABLE FROM MONTH TO MONTH ON INSPECTIONS
HAS TO BE RE-CALIBRATED OFTEN - DIFFICULT TO USE
WITH CONTROL TEST - OFFICER SAFETY ISSUE -
SMALL ~~1 1/2~~ BOARD
STYLOS EASY TO LOSE
MUST ENFORCE 20 MIN WAIT - WILL GIVE ^{READING} ALCOHOL ONLY
PROBLEM WITH PRINTER UNIT - POWER CORD WAS NOT CONNECTED
ON ONE UNIT. MOUTH P.E.E. INTENSIVE
POWER/BATTERY CHARGING ISSUE

12-05

TEST Sequence times .105 STANDARD 4 mins = NO CONTROL TESTS
Attempt to introduce .03 meth sample. TOOK FIRST
sample & INDICATED INVALID ON ~~the~~ PBT SCREEN. When
Attempted to print results, got "Invalid test number" "Possible
Memory corruption" GRAB MESSAGE

12-07 CHECKPOINT = Relative Humidity 80% - instrument lockup
up (keyboard) AFTER 1 1/2 hr exposure
- AFTER GETTING DRY OUT, WORKS AGAIN

1-7-02 ~~path~~ LAG TIME in clearing FC between ISOP tests
of @ least 10 minutes

REQUIRES MORE FORCEFUL BREATH samples with longer
duration than others, could be a problem ^{Recovery high #} of NO-GOS

Difficult to properly seat HAND HELD in case,
case flexes + IR contact is not correct resulting
in data transfer errors

MODE TONES not loud enough drowned out
by ^{ambient} ~~background~~ noise.

ON HIGH CONCENTRATION TEST FOR TOTAL TIME OF TEST,
UNIT LOCKED up AFTER SERIES + DID NOT RESET FOR
30 minutes. Battery removed + unit RESET.

Stylus GAGE too MISPLACED

7110

WALT - HIGH BREATH READING TAKE TOO LONG TO PURGE AND
IF 3rd SAMPLE REQUIRE MAY BE LONGER THAN 15 MIN FOR
TEST. BCSO ABOUT 25% OF BREATH TESTS ARE OVER .20

~~1/20/01~~

Unit sent back to fact on 1-2-02 because
of purge loop problem.

RFI monitor

Function

RFI

ENTER

#1 show

Key radio by extension (rose), watch for
value change.

7410

WANT - POA DELIBERANT - HARD TO ENTER DATA

POWER ISSUE, WHEN POA BATT. IS DISCHARGED - PROGRAM
FOR BALANCE TEST IS DELETED

Small stylus is ~~HARD TO USE~~ EASY TO LAY ASIDE
+ MISPLACE.

DEMO EQUIPMENT OUT TO FLORIDA DEPT. OF LAW ENFORCEMENT

Customer #1459 1/23/02					
Item	Part #	Qty	Serial #	Date Out	Invoice #
Demo Mag Strips Rdr	DL13024	1 ✓	A002314	6/15/01	29466
Demo Cable	DL13026	1 ✓			
Mag Cards	13002	2 ✓			
Demo PDA	DL13007	1 ✓	J650EAU2BB0A014104		
Regulator - 7410+ w	13059	1 ✓			
Demo Power Supply	DL13067	1 ✓			
Demo EPAS	DL13900	1 ✓	ARNB-0368	6/15/01	29466
Demo EPAS Printer	DL13902	1 ✓	ARHF-0057		
Demo EPAS Case	DL13901	1 ✓	DDRDE-1-0001		
Demo 7110	DL7110	1 ✓	ARMM-0448	7/19/01	30040
Demo Keyboard	DL1013	1 ✓			
Demo Simulator	DLSIM	1 ✓	DDNNS3-0046		
Demo Sim. Temp. Probe	DL12025	1 ✓	DDLAP1-0098		
Demo Power Cord	DL34901	1 ✓			
Demo Power Cord	DL12041	1 ✓			
Demo Cable	DL12108	1 ✓			
Demo Hose Assy	DL12043	1 ✓			
Demo Hose, Gum Rubber	DL12035	1 ✓			
Dem Regulator-2 stage	DL12065	1 ✓			

PACKED. *[Signature]*

Attn: Roger Skipper

RMA # 32349

(Return Merchandise Authorization #)

Thank you.

[Signature]

2/11/02

1 From This portion can be removed for Recipient's records.

Date 11-12-01 FedEx Tracking Number 826225609440

Sender's Name Toby Hall Phone 800 839-0690

Company MPD INC

Address 316 E 9TH ST

City OWENSBORO State KY ZIP 42303

2 Your Internal Billing Reference

To Recipient's Name Roger Skipper Phone 321-984-4880

Company FDLE/Melbourne Office

Address 700 South Babcock Street, #401

To 700 S Babcock St, Suite 401, Owensboro, KY 42301

City Melbourne State FL ZIP 32901



0175361373

402

4a Express Package Service

☒ FedEx Priority Overnight ☐ FedEx Standard Overnight ☐ FedEx First Overnight

☐ FedEx 2Day* ☐ FedEx Express Saver*

4b Express Freight Service

☐ FedEx 1Day Freight* ☐ FedEx 2Day Freight ☐ FedEx 3Day Freight

☐ FedEx Envelope/Letter* ☐ FedEx Pak* ☒ Dingy/Pkg

6 Special Handling

☐ SATURDAY Delivery ☐ SUNDAY Delivery ☐ HOLD Saturday at FedEx Location

Does this shipment contain dangerous goods?

☒ No ☐ Yes

☐ Fragile ☐ Fragile (Fragile Marking Required) ☐ Dry Ice

☐ Payment Bill to: ☐ Recipient ☐ Third Party ☐ Credit Card ☐ Cash/Check

Total Packages 1

Total Weight 30

Total Charges

8 Release Signature

By signing you and once us to deliver this shipment without obtaining a signature and agree to indemnify and hold FedEx harmless from any resulting claims.

Customer Order Value: \$100.00 Please visit us at www.fedex.com

0175361373

1000

FedEx USA Airbill

Express

826225609439

0215

1 From This portion can be removed for Recipient's records.

Date 11-12-01

FedEx Tracking Number

826225609439

Sender's Name Toby Hall

Phone

800 835-0690

Company MPD INC

Address 316 E 9TH ST

City OWENSBORO

State KY

ZIP 42303

2 Your Internal Billing Reference

3 To Recipient's Name Roger Skipper

Phone

321 984-4880

Company FDLE/Melbourne Office

Address 700 South Babcock Street, #401

To: "TOOL" of FedEx location, point FedEx address.

We cannot deliver to P.O. boxes or P.O. ZIP codes.

City Melbourne

State

ZIP 32901



8262 2560 9439

0175361373

402

4a Express Package Service

☒ FedEx Priority Overnight

☐ Next business day

☐ FedEx 2Day

☐ Second business day

☐ FedEx Express Saver

☐ Third business day

4b Express Freight Service

☐ FedEx 1Day Freight

☐ Next business day

☐ FedEx 2Day Freight

☐ Second business day

☐ FedEx 3Day Freight

☐ Third business day

* Call for Confirmation

5 Packaging

☐ FedEx Envelope/Letter

☐ FedEx Pak

* Declared value limit \$200

☒ Other Pkg.

Weight and dimensions pkg.

6 Special Handling

☐ SUNDAY Delivery

☐ SUNDAY Delivery

☐ HOLD Saturday

☐ HOLD Sunday

☐ Signature Required

☐ Signature Required

☐ Signature Required

☐ Signature Required

Does this shipment contain dangerous goods?

☒ No

☐ Yes

☐ Yes

☐ Yes

One box must be checked.

☐ Shipper's Declaration

☐ Shipper's Declaration

☐ Shipper's Declaration

Dangerous Goods cannot be shipped in FedEx packaging.

☐ Cargo Aircraft Only

☐ Cargo Aircraft Only

☐ Cargo Aircraft Only

7 Payment Bill to:

☒ Sender

☐ Recipient

☐ Third Party

☐ Credit Card

☐ Cash/Check

☐ Cash/Check

☐ Cash/Check

☐ Cash/Check

Enter FedEx Acct. No. or Credit Card No. below.

☐ Obtain Receipt

☐ Obtain Receipt

☐ Obtain Receipt

Sender's Section

Sender's Section

Sender's Section

Sender's Section

Total Packages

Total Packages

Total Packages

Total Packages

Total Weight

Total Weight

Total Weight

Total Weight

Your liability is limited to \$100 unless you declare a higher value. See the FedEx Service Guide for details.

Your liability is limited to \$100 unless you declare a higher value. See the FedEx Service Guide for details.

Your liability is limited to \$100 unless you declare a higher value. See the FedEx Service Guide for details.

Your liability is limited to \$100 unless you declare a higher value. See the FedEx Service Guide for details.

8 Release Signature

Sign to authorize delivery without obtaining signature.

Sign to authorize delivery without obtaining signature.

Sign to authorize delivery without obtaining signature.

By signing this Airbill, you agree to the terms and conditions of the FedEx Service Guide, including the limitation of liability, and agree to indemnify and hold FedEx harmless from any claims, damages, or losses, including attorney's fees, arising from the use of FedEx services. Questions? Visit our Web site at www.fedex.com or call 1-800-Go-FedEx (646-3333).

402

SEARS

ORDNH 05

SEARS

SEARS
MELBOURNE FLORIDA, FL 02245

SEARS

RETAIN FOR COMPARISON WITH MONTHLY
STATEMENT OR FOR RETURN OR EXCHANGE

SEARS

SALESCHECK #
022450094427TRAN# PG/STORE REG# ASSOC#
4427 30 02245 009 927

SALE

34 83688	SURGE STRI MDS	24.99T
34 83688	SURGE STRI MDS	24.99T
	SUBTOTAL	49.98
	TAX EXEMPT	.00

SEARS

CARD TYPE: VISA

ACCT #: W5258

AUTH CODE: 099692-E

C.O. NUMBER: 112001

11/20/01 VISA TOTAL 49.98

CARDHOLDER ACKNOWLEDGES RECEIPT
OF GOODS AND/OR SERVICES IN THE
AMOUNT OF

\$49.98

AND AGREES TO PERFORM THE
OBLIGATIONS SET FORTH IN THE
CARDHOLDER'S AGREEMENT WITH THE
ISSUER IDENTIFIED HEREON.

SEARS

SEARS

SEARS

SEARS

PURCHASED BY

SEA

SATISFACTION GUARANTEED
OR YOUR MONEY BACK
PLEASE RETAIN THIS RECEIPT TO
RECEIVE A REFUND IN THE FORM OF YOUR
ORIGINAL PAYMENT

SEARS



Scott Specialty Gases, Inc.

CERTIFICATE OF ANALYSIS EBS™ - Ethanol Breath Standard

Customer Name
CMI
316 East Ninth Street
Owensboro, KY 42303

Assay Laboratory
Scott Specialty Gases
6141 Easton Road, Bldg. 1
Plumsteadville, Pa. 18949

Purchase Order# : 362923
Scott Project #:01-58825
Lot No.: 107402I
Expiration Date:
August 17, 2003

REFERENCE STANDARD	CYLINDER #	CONCENTRATION ppm
N.I.S.T. Traceable Standards*	ALM-060154	103.6
	ALM-060204	118.1
	ALM-059962	222.2
	ALM-037472	251.0

CONTAINER TYPE:V

LOT CERTIFIED VALUE CONCENTRATION

COMPONENTS	ppm (BrAC)	ACCURACY
Ethanol	273.5 (0.105)	+/- 2%
Nitrogen	Balance	

*Certification: Traceable to N.I.S.T. RGM Ethanol Standards

Analyst:

Floyd Barof

Date: 8/17/2001

三、四、五、六

8290 5921 9349

5720

Form I.D. No.

FROM 1

Sender's FedEx
Account Number

Date 1/20/01

Sender's FedEx
Account Number

Sender's Name

4

Phone

Company

Address 700 S BABCOCK ST STE 400

MELBOURNE

FILE 33901

2. Your Internal Billing Reference

To Recipient's Name
Futometer Inc (800)

ATTN: Dan Kline
8110 Lackland Rd
Company
address

address 0110 LFKCMAKd

ST LOUIS MO 63114

Peel and Stick FedEx USA Airbill

See back for application instructions.

Questions? Visit our Web site at fedex.com

or call 1-800-Go-FedEx® (800)463-3339.

By using this Airtel you agree to the service conditions on the back of this Airtel and in our current Service Guide, including terms that limit our liability.


உயிர்ப்பாதுகாத்துக் கொடுக்க வேண்டிய நிலைமை ஏற்பட்டால் உடனடியாக உதவி செய்ய வேண்டும் என்று கேட்டுக் கொள்கிறோம்.

0188339073

and agree to indemnify and hold us harmless from any resulting claim.

子

PULL AND RETAIN THIS COPY BEFORE AFFIXING TO THE PACKAGE



**FedEx
Tracking
Number**

1 From <i>Passes print and postcard</i>	Date	2/5/01	Sender's FedEx Account Number
---	------	--------	-------------------------------

Sender's Name

Company FDZ

Address 700 S
City Melbourne

2 Your Interval Billing Reference

3. To Recipient's Name ATTN: 1864

Company CMI, Inc.
Address 316 E. 10th St. N. Fargo, ND 58102

Owensboro

Questions? Visit our Web site at fedex.com or call 1-800-Go-FedEx® (800)463-3339.

By using this Airbill you agree to the service conditions on the back of this Airbill and the current Service Guide, including terms that limit our liability.

Shyred 2008

0020

☐ **2a Express Package Service**
☐ **2b FedEx Priority Overnight**
Next business morning

☐ FedEx 2Day

Second business day
FedEx Express rate not available. Minimum
4b Express Freight Service

☐ FedEx 1Day Freight®
Next business day

☐ 5 Packaging

☐ FedEx Envelope®

6 Special Handling

☐ **SATURDAY Delivery RESTRICTIONS:**
Available only for FedEx Priority Overnight and FedEx 2Day
except ZIP codes

Does this shipment contain dangerous goods?
☐ No ☐ Yes
 (If Yes, see attached Shipper's Declaration for Dangerous Goods label. They label cannot be shipped without it.)

☐ **Student**
 Section No. in Section
 I will be called
 12807
 Section No.
 Credit Card No.
 Total Packages
 Total W

8 Release Signature Sign to receive

By signing you authorize us to deliver

Rev. Ramo 17/00-Part 155315-21994-200

RETAIN THIS COPY FOR YOUR RECORDS.

+

Packages up to 150 lbs.
 Payment may be later in some areas.
FedEx First Overnight
 Earliest next business morning

EVERY FedEx Extra Hour

Packages over 150 lbs.
Free pickup and delivery on orders over \$2000.

FedEx 3Day Freight
Two business day

www.oxfordjournals.org

HOLD Saturday
at FedEx Location
INSTRUCTIONS
Available only for FedEx Priority

☐ **Cargo Aircraft Only**

1

404

1

From: BECKY TAYLOR (270)885-6348
MPD INC
316 EAST 9TH STREET
OWENSBORO, KY, 42303

REVENUE BARCODE



FedEx.


To: Roger Skipper (321)984-4880
FDLE/ Melbourne office
700 South Babcock Street 401

12 JDC & JWS
Card

SHIP DATE: 12DEC01
WEIGHT: 2 LBS

Melbourne, FL, 32901

Ref:



DELIVERY ADDRESS BARCODE (EDS-EDS)

FedEx PRIORITY OVERNIGHT

TRK # 7902 4405 4837 6281


MCO

XHMLBA

32901-FL-US

THU
AA

Deliver by:
13DEC01



Intoximeters

8110 LACKLAND ROAD
ST. LOUIS, MO 63114
PHONE 314-429-4000
FAX 314-429-4170

PAGE NO: 1

INVOICE NO: 082501

INVOICE DATE: 11/27/2001

I N V O I C E

CUSTOMER NO: FLDLE1
CUSTOMER PH: 321/984-4880
BILL TO:

YOUR ORDER NO: LOANER
OUR ORDER NO: CO-000000050857-0000
SHIP TO:

FLORIDA DEPT OF LAW ENFORCEMEN
MELBOURNE FIELD OFFICE
700 S BABCOCK 401
MELBOURNE FL
32901 USA

FLORIDA DEPT OF LAW ENFORCEMEN
MELBOURNE FIELD OFFICE
700 S BABCOCK 401
MELBOURNE FL
32901 USA

TERMS: CASH #1: DISC:
CASH #2: DISC:
CASH #3: DISC:
CASH #4: DISC:
NET DUE DATE: 122701

SHIPPED: 11/21/2001
SHIP VIA: FEDEX
F.O.B.: ORIGIN
SHIPMENT NO: 052188 REF: ML
PRO NO: CSR: AMA

QTY ORDERED	QTY SHIPPED	QTY B.O.	ITEM NUMBER	UNIT PRICE US DOLLARS	EXTD PRICE US DOLLARS
1	1	0	99-0065-00	0.00	0.00
MISC SHIPMENT					

POA system that came back do to printer problems, we are now sending back for evaluation.

Unless otherwise stipulated by contract, a finance charge of 1.5% per month (or the highest rate allowed by law, whichever is less) will be assessed on all due balances. A restocking fee may be applied for returned merchandise.

FEDERAL ID#: 43-0906533

Thank you

SALES TOTAL:	0.00
SALES TAX:	0.00
FREIGHT:	0.00
LESS:	0.00
OTHER CHARGES:	0.00

INVOICE TOTAL:	0.00
US DOLLARS	=====

TERMS AND CONDITIONS

Effective May, 1996

1. Seller shall in no event be liable to Buyer or to any person who shall purchase from Buyer or use any products supplied pursuant to this invoice for damages of any kind, including, but not limited to, direct, indirect, special or consequential damages or loss of production or loss of profits or claims of third parties against Seller, resulting from any cause whatsoever including, but not limited to any delay, act, error, or omission of Seller, and Seller's sole liability to Buyer for any cause of actions, whether based upon negligence, contract, or strict liability, shall be to replace any products covered by this invoice which do not conform to the specifications set forth herein at the time of delivery by Seller.
2. All sales and quotations made by Seller are subject to each of the within terms and conditions. These terms and conditions shall replace all terms and conditions of Buyer's order and of any proposal or quotation to Buyer not agreed to by Buyer and Seller prior to the date of this invoice. In the absence of Buyer's written acceptance, the first to occur of: (i) an acceptance of any goods covered by Buyer's order, or (ii) Buyer's failure to object in writing to the terms of this invoice within ten (10) days following its receipt, shall constitute Buyer's acceptance of these terms and conditions. No waiver, alteration or modification of these provisions shall be valid unless made in writing and signed by a duly authorized representative of Seller.
3. All unpaid items will be subject to a late payment fee computed at the rate of one and one-half percent (1.5%) per month (an effective rate of eighteen percent (18%) per annum) on the declining balance unpaid for more than thirty (30) days after the date of this invoice is received. Buyer shall pay Seller all costs of collection on past due accounts including but not limited to reasonable attorneys' fees, whether or not litigation is commenced in aid thereof.
4. SELLER WARRANTS THAT THE GOODS ARE AS DESCRIBED IN THIS AGREEMENT. SELLER MAKES NO OTHER WARRANTY OF ANY KIND, EXPRESS OR IMPLIED, AND ALL IMPLIED WARRANTIES OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE HEREBY DISCLAIMED BY SELLER AND EXCLUDED FROM THIS WARRANTY. No agent, employee or representative of the Seller has any authority to bind the Seller to any affirmation, representation or warranty concerning the Goods sold under this agreement, and unless any affirmation, representation or warranty made by an agent, employee or representative is specifically included within this written agreement, it has not formed a part of the basis of this bargain and shall not in any way be enforceable by the Buyer.
5. Quotations are subject to prior sales, change without notice, strikes, accidents and/or other causes beyond Seller's control. All sales and quotations are F.O.B. point of shipment, unless otherwise stipulated. Freight is prepaid and charged to Buyer, unless otherwise stipulated. Seller reserves the right to cancel Buyer's order in the event the selling price, or prices, shall be established by the Government with respect to the items covered by Buyer's order which shall be lower than the price, or prices, indicated herein. All claims of whatever nature, including but not limited to claims for shortage, deductions, and defective Goods must be made in writing to Seller within five (5) days of Buyer's receipt of Goods. Failure to comply with the aforesaid procedure shall constitute Buyer's waiver of any such claim. Buyer must make all claims for damage or loss in transit to the transportation company. Any return of Goods to Seller by Buyer must be authorized by Seller in writing and is subject to Seller's 25% restocking fee. Buyer is responsible for all freight charges, both outbound and incoming. No obsolete items may be returned. Seller will not be liable for Buyer's claims of loss, damage, cost of repairs, or incidental or consequential damages of any kind whatsoever, whether such claims are based upon warranty, contract, or negligence, or whether such claims arise in connection with the sale, use or repair of the Goods.
6. The amount of sales, excise or other taxes, if any, applicable to the Goods covered by this invoice, shall be added to the purchase price and shall be paid by Buyer unless Buyer timely provides Seller with an exemption certificate accepted by the taxing authorities.
7. This agreement shall be deemed for all purposes to have been made in Missouri and shall be governed by and construed in accordance with the laws of Missouri without regard to its principles of conflicts under law. Any cause of action involving the goods sold hereunder shall be brought only in state and federal courts sitting in the State of Missouri and both parties consent to the exclusive jurisdiction of such courts in any such action between them. Any action for breach of contract must be commenced within two (2) years after the cause of action has accrued.
8. Seller's delivery of the Goods to carrier shall constitute delivery to Buyer. All risk of loss or damage in transit shall be borne by Buyer.
9. Shipping dates are approximate and are not guaranteed. Seller shall not be liable for failure to deliver or perform or for delays in delivery or performance occasioned by causes beyond its control, including without limitation, strikes, lockouts, fires, accidents, interruptions in the supply of materials, breakdowns, delays in carriers or suppliers, and governmental actions and regulations.
10. If Buyer fails to fulfill the terms of payment on any order, Seller, in addition to its rights and remedies, but not in limitation thereof, may, defer all further shipments until such payments are made, or may, at its option, cancel the order. All decisions with respect to the extension and/or continuation of credit terms to Buyer shall be in the sole discretion of Seller. Seller reserves the right to require prepayment of orders.
11. Unless otherwise specified herein, Seller reserves the right to make deliveries in installments. Delay in delivery of all installments shall not relieve Buyer of 1) its obligation to pay for all installments received prior to such delay and 2) its obligation to accept remaining deliveries. Seller reserves the right, even after partial shipment on account of any order, to require satisfactory security from Buyer for performance of Buyer's obligation. Buyer's refusal to furnish such security will entitle Seller to suspend shipments until such security is furnished, and, will entitle Seller, at its option, to cancel order. In such event, Buyer shall nevertheless be required to pay for such shipments as may have been received by Buyer prior to Seller's cancellation of the order.
12. Seller's failure strictly to enforce any terms or conditions of this agreement or to exercise any right arising hereunder shall not constitute a waiver of Seller's right strictly to enforce such term and condition or exercise such right thereafter. Each right or remedy granted to Seller hereunder shall be deemed cumulative and may be exercised from time to time. Any waiver of Buyer's default hereunder shall be in writing and shall not operate as a waiver of any other default or of the same default thereafter.

Skipper, Roger

From: Skipper, Roger
Sent: Wednesday, January 02, 2002 4:48 PM
To: 'Taylor, Becky'
Subject: RE:

OK

also, the second cal checks seem to fall off each and every time, especially with a high simulator (.30 and .40). I noticed there was a strong smell of ethanol around the instrument during the purge. I think it may be because the outlet port is recessed within the case and the purged air is diffused inside and just outside the case and not venting properly. the last test I just ran with a .40 sim and a .105 gas standard resulted in a first control check of .104 but the second was .087. There is a similar difference in the breath result also. Maybe extending the breath pump exhaust port to the outside of the case where a tube could be connected to route the exhaust down away from the case would help. This could explain the drop off I seen in the field also.

Just a thought. If ya'll want to talk about it email me right back or call in the morning. I'll be in the office for about 20 more minutes.

Roger

-----Original Message-----

From: Taylor, Becky [SMTP:bjtaylor@alcoholtest.com]
Sent: Wednesday, January 02, 2002 4:28 PM
To: Skipper, Roger
Subject:

I tried to Call but they said the phones had been forwarded. Bill ask if you could dilute the Methanol solution in half and try it again.

thanks alot.
Becky

PACKING SLIP

Draeger Safety, Inc., Breathalyzer
185 Suttle Street
Suite 105
Durango, CO 81303-7911 USA
Telephone: 970/385-5555

Sales Order 31222
Order Date 12/04/01

Ship To:

FLORIDA DEPT. LAW ENFORCEMENT
ATTN: ROGER SKIPPER
700 SOUTH BARECOCK ST. STE 201
MELBOURNE FL 32901
USA

Customer	Ship Via	F.O.B.	Terms	Purchase Order Number	Salesperson	Reference No.
001459	FDX-2DAY	DESTINATION	Need Copy of PO	EVAL/UPGRADE	N/A	
Qty. Ordered	Qty. Shipped	Item Number	Unit of Measure	Required Date		
1	0	DL13007	EACH	12/04/01		
		DEMO PDA, MODIFIED FOR EPAS-SERIAL#J650EAU2BBOAO14104				

Dräger

Alcotest® 7110 MK III C

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest 7110 MK III C has been tested for accuracy and found to be in compliance with the National Highway Traffic Safety Administration Standard for evidential breath testing devices. The Alcotest MK III C is compliant as a "mobile" and "nonmobile" EBT with 49 FR 48854, 49 FR 48864 and 58 FR 48705. The manufacturer recommends accuracy verification of this instrument within 12 months of the calibration date below, or sooner, according to your State Specifications.

Certification date:

01/14/2002

SERIAL NUMBER

ARMH-D448

DRAEGER SAFETY, INC.
Breathalyzer Division
Durango, CO

EW

3 TESTS
THRU BT
METH

$\frac{104}{104} \cdot \frac{.04}{.03} \leftarrow \frac{.03}{.03}$

also took 10 minutes to
clean PC for next
survey

	.04			.10			.02		
7110	IF .024/.024	.024/.023	24/022	INF	INF	INF	INF	INF	INF
8000	EC .031 .031	.032/.031	630/630	INF	INF	INF	.000	.000	.000
POA	INF	INF	INF	INF	INF	INF	.000	.000	.000
EPAS	.032 .032	.031	*	.048	.047	.047	VOIDED	VOIDED	VOIDED
	.031 .032	.031	*	.048	.047	.047	.02	.02	.01
	.03 .03	.03	not	.04	.04	.04	.02	.02	.01
	.03 .03	.03	flagged	.04	.04	.04	.02	.01	.02

REMAN AS
ETROL

7110
8000
POA
GPAS

ACC TONS + .08
.078
.079

COCCPUUP

RADAR/LADAR RFI TESTS

3508
27

238

NF = NOT FLAGGED INF = INTERFERENT DNRS = DID NOT RECOGNIZE sample
 * = WAIT BETWEEN SAMPLE SETS VARIED 10-23 MINUTES

	7110	8000	ALCO IV	EPAS
.10 ISOP ^{N-PROP}	INF	024/025 NF	VOID	04/03 NF
	INF	024/025 NF	VOID	04/04 NF
	INF	026/025 NF	VOID	04/04 NF
.24 ISOP	INF	INF	112/111 NF*	11/11 NF
	INF	INF	112/109 NF*	11/11 NF
	INF	INF	108/110 NF*	11/11 NF
.04 METH	INF	DNRS	VOID *	04/04 NF
	INF	DNRS	VOID *	04/03 NF
	INF ^{NOT 1}	DNRS	VOID *	03/03 NF
.04 NPROP	IR 024/024 EZ 031/031 NF JR 024/023 EZ 032/031 NF IR 024/022 EZ 030/030 NF	INF	032/031 NF*	03/03 NF
		INF	032/032 NF*	03/03 NF
		INF	031/031 NF*	03/03 NF
.10 ISOP	INF	INF	048/048 NF*	04/04 NF
	INF	INF	047/047 NF*	04/04 NF
	INF	INF	047/046 NF*	04/04 NF
.02 METH	INF	000/000	VOID	02/02 NF
	INF	000/000	VOID	02/01 NF
	INF	000/000	VOID	01/02 NF

NE = NO EFFECT
NCT = NO CONTROL TEST

7110 RADAR NE	8000 NE	ALCO IV NE	TPAS NE
03/02/03 NO.02 TIME 16m	7m	9m NCT	11 NCT
03 ACET 000/000		000/000	00/00
000/000		000/000	00/00
000/000		000/000	00/00
08+03 ACET 79/78 NF 78/77 NF 75/78		73/75 NF	07/07 NF
08 BASE 12 76/78		75/75	07/07
XYL MOUTH ALCOHOL		DNRS	DNRS
TUFF INF		000/VOID	07/05 DNF
N-BUTANOL N-BUTYL MOUTH ALCOHOL		DNRS	DNRS

OPERATOR REQUIRED STOPS

* DATA INPUT IS ONE STOP

	7110	8000	POA	EPAS
1	PUSH START	PUSH START	SELECT BEGIN TEST	SELECT END TEST
2	INPUT DATA	INPUT DATA	INPUT DATA	INPUT DATA
3	connect mp	connect mmpieces	INSERT mmpieces	CONNECT CONTROL
4	Present BT	PRESNT BT	PUSH SET	Deliver sample
5	PRESnt BT	PRESnt BT	CONNECT CONTROL	Remove control
6	Remove MP	Remove mP	PRESNT UNIT	CONNECT mmpieces
7	collect ticket	collect ticket	DELIVERY SAMPLE	connect mP
8			PUSH SET	PRESNT UNIT
9			PUSH SET	Present unit
10			DISCONNECT CONTROL	Remove mmpieces
11			PRESNT UNIT (BT)	Remove mP
12				connect
13			PUSH SET	connect control
14			PRESNT UNIT (BT)	Remove control
15			PUSH SET	^{unit} place on printer
16			CONNECT CONTROL	collect tickets
17			DELIVER SAMPLE	^{unit} Remove from printer
18			DISCONNECT CONTROL	collect ticket
19			PUSH SET PUSH SET	
20			DISCONNECT CONTROL	
			POT IN CRADLE	
			SELECT PRINT	
			INSERT mmpieces	
			Tear off ticket	

HIGH

NO .02

.30 - .20 - .30

OR .20 .30 .20

with Data input + 2 min wait enabled if on option

7110
16 min

8000
7 minutes
NO Recovery
NEEDED FOR
next series
2m

POA
NO CONTROL TESTS
9 minutes
But long 10-14 min
Recovery for next
series
2m

PAS
NO CONTROL
11 minutes TESTS
3 minute Recovery
FOR next series
2m

.000 .000 .000
.000 .000 .000

103 ACETONE - water
.000 .000 .000
.000 .000 .000

.000 .00 .000
.00 .00 .00

7110
8000
POA
EPAS
XVI
NO RESPONSE
at all
NO RESPONSE
at all
NO RESPONSE
at all

METH TOL / METH ETH CLOR
PETS
TUFF
INF
000 / 000
000 / VOID
.07 not
.05 FLAG

TIME
N-Butyl
Romo as
mouth Accoro
Mouth
would not recognize
sample
would not
recognize sample
as being placed
same

me
1089

08
08+.03

75/78 76/78
79/78 78/77
NO FLAG

.08/.027
.085/.084

.73
.75
.75
.75

.07
.07
.07
.07
one
set

CUSTOMER RECEIPT NP
FEDEX GROUND CALL TAG NUMBER
9973291342534 CDM

RETURN

TO: [illegible]
FROM: [illegible]
DATE: [illegible]
TIME: [illegible]
BY: [illegible]
SIGNATURE: [illegible]
FEDERAL EXPRESS
10000 WILSON BLVD
MEMPHIS, TN 38117
(901) 874-1000
FAX (901) 874-1001
WWW.FEDEX.COM

FEDEX GROUND
SERVICE
1/66

**IF UNABLE TO PICK-UP
MERCHANDISE, INDICATE REASON**

- (A) Incorrect Address
— (B) Not In 3 Attempts
— (F) Business Permanently Closed
— (H) Out Of Service Area

Customer Signature Required:

- (C) Not Ready in 3 Attempts
— (D) Changed Mind – Will Keep
— (E) Duplicate Call Tag
— (G) Picked-Up By Another Carrier
— (I) Consignee Unable To Find Package

Signature _____ Date _____

Printed Name _____

1. Except. Code _____	1. Time _____
2. Except. Code _____	2. Time _____
3. Except. Code _____	3. Time _____
1. Date ____/____/____	1. Van # _____
2. Date ____/____/____	2. Van # _____
3. Date ____/____/____	3. Van # _____