

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONCALVES,GUSTAVO,R
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: GONCALVES,GUSTAVO
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: LANGDALE

DATE/TIME OF TEST: 04/03/2016 01:36:04

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.076	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.205	01:31
Breath Volume	1.816	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:33
Subject Sample # 2	0.201	01:34
Breath Volume	1.910	01:34
Air Blank	0.000	01:35
Control Test	0.078	01:35
Air Blank	0.000	01:35
Diagnostics Check	OK	01:36

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SANCHEZ-JIMENEZ

DATE/TIME OF TEST: 04/04/2016 01:33:34

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.179	01:28
Breath Volume	3.875	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.180	01:31
Breath Volume	4.117	01:31
Air Blank	0.000	01:32
Control Test	0.076	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GREGORY,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 04/04/2016 06:49:06

Diagnostics Check	OK	06:42
Air Blank	0.000	06:42
Control Test	0.076	06:43
Air Blank	0.000	06:43
Subject Sample # 1	0.055	06:44
Breath Volume	2.242	06:44
Air Blank	0.000	06:45
Air Blank	0.000	06:47
Subject Sample # 2	0.053	06:47
Breath Volume	3.355	06:47
Air Blank	0.000	06:48
Control Test	0.075	06:48
Air Blank	0.000	06:49
Diagnostics Check	OK	06:49

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 04/04/2016 22:53:21

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.076	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.187	22:48
Breath Volume	3.570	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.183	22:51
Breath Volume	3.652	22:51
Air Blank	0.000	22:52
Control Test	0.076	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN, JOSEPH, G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: THOMPSON, SCOTT
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: HOVDE

DATE/TIME OF TEST: 04/04/2016 23:11:23

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:05
Subject Sample # 1	0.039	23:06
Breath Volume	4.175	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:08
Subject Sample # 2	0.031	23:09
Breath Volume	1.128	23:09
Air Blank	0.000	23:10
Control Test	0.077	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMOND,COREY,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,COREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 04/08/2016 00:34:33

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.077	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.150	00:30
Breath Volume	1.617	00:30
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.135	00:32
Breath Volume	1.804	00:32
Air Blank	0.000	00:33
Control Test	0.075	00:33
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN, JOSEPH, G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RONGISH, SHERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BUFFEY

DATE/TIME OF TEST: 04/09/2016 01:37:25

Diagnostics Check	OK	01:30
Air Blank	0.000	01:31
Control Test	0.076	01:31
Air Blank	0.000	01:32
Subject Sample # 1	0.157	01:32
Breath Volume	1.593	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.151	01:35
Breath Volume	1.500	01:35
Air Blank	0.000	01:36
Control Test	0.075	01:36
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLEURANT, MICHELSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLEURANT, MICHELSON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ARROYO GARCIA

DATE/TIME OF TEST: 04/11/2016 01:26:21

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.077	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.173	01:21
Breath Volume		1.519	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.170	01:24
Breath Volume		1.207	01:24
Air Blank		0.000	01:25
Control Test		0.076	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRATO,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BARRINGER,BRIAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09014080A1
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: PARROTT

DATE/TIME OF TEST: 04/12/2016 15:40:59

Diagnostics Check	OK	15:34
Air Blank	0.000	15:35
Control Test	0.076	15:35
Air Blank	0.000	15:35
Subject Sample # 1	0.297	15:36
Breath Volume	4.253	15:36
Air Blank	0.000	15:37
Air Blank	0.000	15:38
Subject Sample # 2	0.289	15:39
Breath Volume	1.914	15:39
Air Blank	0.000	15:40
Control Test	0.076	15:40
Air Blank	0.000	15:40
Diagnostics Check	OK	15:40

Data Download Date/Time: 04/13/2016 08:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: STROTMAN

DATE/TIME OF TEST: 04/13/2016 23:37:53

Diagnostics Check	OK	23:31
Air Blank	0.000	23:31
Control Test	0.077	23:32
Air Blank	0.000	23:32
Subject Sample # 1	0.222	23:33
Breath Volume	1.195	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.230	23:36
Breath Volume	2.042	23:36
Air Blank	0.000	23:36
Control Test	0.076	23:37
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMOND,COREY,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,COREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MILSTID

DATE/TIME OF TEST: 04/15/2016 00:10:10

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.077	00:04
Air Blank	0.000	00:05
Subject Sample # 1	0.067	00:05
Breath Volume	2.402	00:05
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.062	00:08
Breath Volume	2.023	00:08
Air Blank	0.000	00:09
Control Test	0.077	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONCALVES,GUSTAVO,R
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: GONCALVES,GUSTAVO
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 04/17/2016 01:53:42

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.077	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.163	01:49
Breath Volume	2.093	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.165	01:52
Breath Volume	2.046	01:52
Air Blank	0.000	01:52
Control Test	0.077	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN, JOSEPH, G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: AQUILAR, ALEXANDER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:47

SUBJECT'S LAST NAME: AYALA

DATE/TIME OF TEST: 04/17/2016 18:26:46

Diagnostics Check	OK	18:20
Air Blank	0.000	18:20
Control Test	0.077	18:20
Air Blank	0.000	18:21
Subject Sample # 1	0.000	18:22
Breath Volume	2.093	18:22
Air Blank	0.000	18:22
Air Blank	0.000	18:24
Subject Sample # 2	0.000	18:25
Breath Volume	1.277	18:25
Air Blank	0.000	18:25
Control Test	0.077	18:26
Air Blank	0.000	18:26
Diagnostics Check	OK	18:26

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN, JOSEPH, G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RONGISH, SHERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MENDOZA SAN MARTIN

DATE/TIME OF TEST: 04/18/2016 00:33:23

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.077	00:27
Air Blank		0.000	00:28
Subject Sample	# 1	0.108	00:28
Breath Volume		3.015	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.101	00:31
Breath Volume		2.230	00:31
Air Blank		0.000	00:32
Control Test		0.076	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN, JOSEPH, G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN, EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: NIESZ

DATE/TIME OF TEST: 04/18/2016 02:32:46

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.180	02:28
Breath Volume	1.535	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.177	02:31
Breath Volume	1.355	02:31
Air Blank	0.000	02:31
Control Test	0.076	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONIECZKI,D,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KONIECZKI,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: NOVINGER

DATE/TIME OF TEST: 04/19/2016 11:26:36

Diagnostics Check	OK	11:18
Air Blank	0.000	11:19
Control Test	0.078	11:19
Air Blank	0.000	11:20
Subject Sample # 1	0.000	11:22
Breath Volume	1.667	11:22
Air Blank	0.000	11:22
Air Blank	0.000	11:24
Subject Sample # 2	0.000	11:25
Breath Volume	1.187	11:25
Air Blank	0.000	11:25
Control Test	0.076	11:25
Air Blank	0.000	11:26
Diagnostics Check	OK	11:26

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRATO, JOSHUA, M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WESTERLUND, TRENT
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 04/20/2016 13:48:15

Diagnostics Check	OK	13:41
Air Blank	0.000	13:41
Control Test	0.077	13:42
Air Blank	0.000	13:42
Subject Sample # 1	0.016	13:43
Breath Volume	1.613	13:43
Air Blank	0.000	13:44
Air Blank	0.000	13:45
Subject Sample # 2	0.015	13:46
Breath Volume	1.656	13:46
Air Blank	0.000	13:47
Control Test	0.077	13:47
Air Blank	0.000	13:48
Diagnostics Check	OK	13:48

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROBEL, RAYMOND, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: STROBEL, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: WILBUR

DATE/TIME OF TEST: 04/20/2016 21:06:33

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:03	
Control Test	0.077	21:04	
Air Blank	0.000	21:04	
Subject Sample	# 1	21:05	Subject Test Refused
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Control Test	0.077	21:05	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GUDDAL

DATE/TIME OF TEST: 04/23/2016 00:47:19

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.078	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.000	00:43
Breath Volume	1.730	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.000	00:45
Breath Volume	1.707	00:45
Air Blank	0.000	00:46
Control Test	0.077	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUBOIS, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/23/2016 03:33:12

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.077	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.131	03:28
Breath Volume		2.148	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.123	03:31
Breath Volume		1.316	03:31
Air Blank		0.000	03:32
Control Test		0.076	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SABEAN,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: HARLOW

DATE/TIME OF TEST: 04/23/2016 08:59:43

Diagnostics Check	OK	08:51
Air Blank	0.000	08:51
Control Test	0.078	08:51
Air Blank	0.000	08:52
Subject Sample # 1	0.177	08:55
Breath Volume	1.625	08:55
Air Blank	0.000	08:56
Air Blank	0.000	08:57
Subject Sample # 2	0.177	08:58
Breath Volume	1.671	08:58
Air Blank	0.000	08:58
Control Test	0.076	08:59
Air Blank	0.000	08:59
Diagnostics Check	OK	08:59

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DECREMER

DATE/TIME OF TEST: 04/24/2016 19:27:02

Diagnostics Check	OK	19:20
Air Blank	0.000	19:20
Control Test	0.077	19:20
Air Blank	0.000	19:21
Subject Sample # 1	0.147	19:22
Breath Volume	1.558	19:22
Air Blank	0.000	19:22
Air Blank	0.000	19:24
Subject Sample # 2	0.136	19:25
Breath Volume	1.519	19:25
Air Blank	0.000	19:26
Control Test	0.077	19:26
Air Blank	0.000	19:26
Diagnostics Check	OK	19:27

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: BERNOKEITS

DATE/TIME OF TEST: 04/26/2016 15:48:53

Diagnostics Check	OK	15:40
Air Blank	0.000	15:41
Control Test	0.078	15:41
Air Blank	0.000	15:41
Subject Sample # 1	0.219	15:43
Breath Volume	1.714	15:43
Air Blank	0.000	15:44
Air Blank	0.000	15:45
Subject Sample # 2	0.223	15:47
Breath Volume	1.562	15:47
Air Blank	0.000	15:47
Control Test	0.075	15:48
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILERA,DANIEL,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NOEL,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 04/27/2016 00:32:45

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.077	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.162	00:28
Breath Volume	2.019	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.157	00:31
Breath Volume	1.472	00:31
Air Blank	0.000	00:31
Control Test	0.077	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: KRASINSKI

DATE/TIME OF TEST: 04/27/2016 23:39:44

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.078	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.122	23:35
Breath Volume	1.617	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.117	23:38
Breath Volume	1.753	23:38
Air Blank	0.000	23:38
Control Test	0.078	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 05/17/2016 10:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SEREQUE

DATE/TIME OF TEST: 04/02/2016 04:06:07

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.078	03:58	
Air Blank	0.000	03:58	
Subject Sample	# 1 0.161	04:01	
Breath Volume	1.968	04:01	
Air Blank	0.000	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 2 0.152	04:04	
Breath Volume	1.578	04:04	
Air Blank	0.000	04:05	
Control Test	0.073	04:05	Control Outside Tolerance
Air Blank	0.000	04:06	

Data Download Date/Time: 04/12/2016 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: SEREQUE

DATE/TIME OF TEST: 04/02/2016 04:17:49

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.076	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.154	04:12
Breath Volume	1.515	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.150	04:16
Breath Volume	1.421	04:16
Air Blank	0.000	04:16
Control Test	0.076	04:17
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/12/2016 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: CARSWELL

DATE/TIME OF TEST: 04/05/2016 23:14:15

Diagnostics Check	OK	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:07
Subject Sample # 1	0.200	23:09
Breath Volume	2.070	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:11
Subject Sample # 2	0.186	23:12
Breath Volume	2.066	23:12
Air Blank	0.000	23:13
Control Test	0.075	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 04/12/2016 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 04/06/2016 22:49:47

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.078	22:47	
Air Blank	0.000	22:48	
Subject Sample	# 1	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:48	
Control Test	0.078	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 04/12/2016 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 04/09/2016 21:05:14

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.078	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.209	21:00
Breath Volume	2.500	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.199	21:03
Breath Volume	2.320	21:03
Air Blank	0.000	21:04
Control Test	0.076	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 04/12/2016 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: NGO

DATE/TIME OF TEST: 04/10/2016 04:33:56

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.155	04:29
Breath Volume	1.585	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample # 2	0.165	04:32
Breath Volume	2.000	04:32
Air Blank	0.000	04:33
Control Test	0.075	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 04/12/2016 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 04/15/2016 03:03:20

Diagnostics Check	OK	03:00	
Air Blank	0.000	03:00	
Control Test	0.077	03:01	
Air Blank	0.000	03:01	
Subject Sample	# 1	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.078	03:02	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 05/03/2016 08:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 04/16/2016 04:26:41

Diagnostics Check	OK	04:26	
Air Blank		04:26	Sequence Aborted
Air Blank	0.000	04:26	

Data Download Date/Time: 05/03/2016 08:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 04/16/2016 04:40:18

Diagnostics Check	OK	04:33
Air Blank	0.000	04:34
Control Test	0.076	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.144	04:35
Breath Volume	2.250	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.140	04:38
Breath Volume	1.601	04:38
Air Blank	0.000	04:39
Control Test	0.076	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 05/03/2016 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: ACKER

DATE/TIME OF TEST: 04/16/2016 21:45:29

Diagnostics Check	OK	21:42	
Air Blank	0.000	21:42	
Control Test	0.078	21:43	
Air Blank	0.000	21:43	
Subject Sample	# 1	21:44	Subject Test Refused
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Control Test	0.076	21:44	
Air Blank	0.000	21:45	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 05/03/2016 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,OFFICER,B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KARAYIANES,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: GALLAGHER

DATE/TIME OF TEST: 04/20/2016 02:21:08

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.150	02:16
Breath Volume	2.273	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.148	02:19
Breath Volume	1.191	02:19
Air Blank	0.000	02:20
Control Test	0.076	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 05/03/2016 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 04/23/2016 02:36:47

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.077	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.188	02:32
Breath Volume	1.523	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.191	02:35
Breath Volume	1.667	02:35
Air Blank	0.000	02:35
Control Test	0.076	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 05/03/2016 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORICE,DOMINIC,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SORICE,DOMINIC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:12

SUBJECT'S LAST NAME: MCENTIRE

DATE/TIME OF TEST: 04/24/2016 13:40:58

Diagnostics Check	OK	13:33
Air Blank	0.000	13:33
Control Test	0.078	13:34
Air Blank	0.000	13:34
Subject Sample # 1	0.000	13:35
Breath Volume	2.277	13:35
Air Blank	0.000	13:35
Air Blank	0.000	13:38
Subject Sample # 2	0.000	13:39
Breath Volume	2.234	13:39
Air Blank	0.000	13:40
Control Test	0.077	13:40
Air Blank	0.000	13:40
Diagnostics Check	OK	13:40

Data Download Date/Time: 05/03/2016 08:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: STRANGE

DATE/TIME OF TEST: 04/02/2016 21:59:43

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:48	
Control Test		0.080	21:48	
Air Blank		0.000	21:49	
Subject Sample	# 1	0.228	21:52	Slope Not Level
Breath Volume		1.195	21:52	
Air Blank		0.000	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 2	0.218	21:58	Volume Not Met
Breath Volume		0.207	21:58	
Air Blank		0.000	21:58	
Control Test		0.078	21:59	
Air Blank		0.000	21:59	
Diagnostics Check		OK	21:59	

Data Download Date/Time: 05/18/2016 15:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WHITE,C
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: MIRANDA ALVAREZ

DATE/TIME OF TEST: 04/07/2016 00:50:45

Diagnostics Check	OK	00:43
Air Blank	0.000	00:44
Control Test	0.080	00:44
Air Blank	0.000	00:45
Subject Sample # 1	0.132	00:45
Breath Volume	2.261	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.142	00:49
Breath Volume	2.363	00:49
Air Blank	0.000	00:49
Control Test	0.079	00:50
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 05/18/2016 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PENA-POTRERO

DATE/TIME OF TEST: 04/30/2016 02:15:50

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.168	02:11
Breath Volume	2.238	02:11
Air Blank	0.000	02:12
Air Blank	0.000	02:13
Subject Sample # 2	0.160	02:14
Breath Volume	1.488	02:14
Air Blank	0.000	02:14
Control Test	0.079	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 05/18/2016 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINKMAN,KELLY,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRINKMAN,KELLY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RAPA

DATE/TIME OF TEST: 04/30/2016 02:32:25

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.081	02:31	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 05/18/2016 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: RIEMER

DATE/TIME OF TEST: 04/30/2016 20:14:27

Diagnostics Check	OK	20:11	
Air Blank	0.000	20:11	
Control Test	0.081	20:11	
Air Blank	0.000	20:12	
Subject Sample	# 1	20:13	Subject Test Refused
Breath Volume	0.000	20:13	
Air Blank	0.000	20:13	
Control Test	0.081	20:13	
Air Blank	0.000	20:14	
Diagnostics Check	OK	20:14	

Data Download Date/Time: 05/18/2016 15:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COPPOLA,TYSON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 04/01/2016 03:56:45

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.078	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.161	03:45	Volume Not Met
Breath Volume		0.269	03:45	
Air Blank		0.000	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.161	03:49	
Breath Volume		1.839	03:49	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 3	0.150	03:55	
Breath Volume		1.972	03:55	
Air Blank		0.000	03:55	
Control Test		0.075	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 04/23/2016 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVE,JASON,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: BINZ,ANTHONY
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: SHOVER

DATE/TIME OF TEST: 04/05/2016 20:17:03

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.078	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:16	
Control Test	0.078	20:16	
Air Blank	0.000	20:16	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 04/23/2016 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVE,JASON,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: BINZ,ANTHONY
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/10/2016 22:15:56

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.079	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.182	22:10
Breath Volume	1.835	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.191	22:14
Breath Volume	3.046	22:14
Air Blank	0.000	22:15
Control Test	0.076	22:15
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 04/23/2016 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: KIRKLAND,RICHARD
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: DICKEY

DATE/TIME OF TEST: 04/13/2016 05:03:15

Diagnostics Check	OK	05:00	
Air Blank	0.000	05:00	
Control Test	0.079	05:01	
Air Blank	0.000	05:01	
Subject Sample	# 1	05:01	Subject Test Refused
Breath Volume	0.000	05:01	
Air Blank	0.000	05:02	
Control Test	0.079	05:02	
Air Blank	0.000	05:03	
Diagnostics Check	OK	05:03	

Data Download Date/Time: 04/23/2016 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELAHANTY,MICHAEL,T
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HANSEN,AARON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: PLEBANEK

DATE/TIME OF TEST: 04/15/2016 21:47:26

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.079	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.099	21:42
Breath Volume	3.710	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:45
Subject Sample # 2	0.097	21:45
Breath Volume	3.371	21:45
Air Blank	0.000	21:46
Control Test	0.078	21:46
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 04/23/2016 15:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: KIRKLAND,RICHARD'
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: GERENA

DATE/TIME OF TEST: 04/16/2016 06:17:02

Diagnostics Check	OK	06:06
Air Blank	0.000	06:06
Control Test	0.078	06:07
Air Blank	0.000	06:07
Subject Sample # 1	0.172	06:10
Breath Volume	1.453	06:10
Air Blank	0.000	06:11
Air Blank	0.000	06:13
Subject Sample # 2	0.185	06:15
Breath Volume	1.746	06:15
Air Blank	0.000	06:16
Control Test	0.076	06:16
Air Blank	0.000	06:16
Diagnostics Check	OK	06:17

Data Download Date/Time: 04/23/2016 15:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEVERANCE,MARK,D
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: WATSON,BRADLEY
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: ZAKREWSKY

DATE/TIME OF TEST: 04/28/2016 16:59:12

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:56	
Control Test	0.079	16:56	
Air Blank	0.000	16:57	
Subject Sample	# 1	16:57	Subject Test Refused
Breath Volume	0.000	16:57	
Air Blank	0.000	16:58	
Control Test	0.078	16:58	
Air Blank	0.000	16:59	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 05/14/2016 11:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,MARCELINO,R
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: RODRIGUEZ,E
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MARTINEZ CANIZARES

DATE/TIME OF TEST: 04/07/2016 02:41:53

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.210	02:37
Breath Volume	2.648	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.214	02:40
Breath Volume	2.695	02:40
Air Blank	0.000	02:40
Control Test	0.078	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 04/19/2016 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMAGO,MEGAN,L
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ROMERO,JOSE
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: CASTILLO CASTANEDA

DATE/TIME OF TEST: 04/01/2016 21:57:56

Diagnostics Check	OK	21:54	
Air Blank	0.000	21:55	
Control Test	0.078	21:55	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:56	Subject Test Refused
Breath Volume	0.000	21:56	
Air Blank	0.000	21:57	
Control Test	0.078	21:57	
Air Blank	0.000	21:57	
Diagnostics Check	OK	21:57	

Data Download Date/Time: 04/19/2016 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: JANVIER,FRITZ
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/03/2016 19:44:52

Diagnostics Check	OK	19:37
Air Blank	0.000	19:37
Control Test	0.078	19:38
Air Blank	0.000	19:38
Subject Sample # 1	0.207	19:40
Breath Volume	1.753	19:40
Air Blank	0.000	19:40
Air Blank	0.000	19:42
Subject Sample # 2	0.218	19:43
Breath Volume	2.925	19:43
Air Blank	0.000	19:43
Control Test	0.077	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 04/19/2016 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: JANVIER,FRITZ
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 03/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: PICHARDO CALDERON

DATE/TIME OF TEST: 04/17/2016 20:07:14

Diagnostics Check		OK	19:59
Air Blank		0.000	20:00
Control Test		0.078	20:00
Air Blank		0.000	20:00
Subject Sample	# 1	0.206	20:02
Breath Volume		2.292	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.211	20:05
Breath Volume		2.332	20:05
Air Blank		0.000	20:06
Control Test		0.078	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 04/19/2016 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNOZ,A,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MUNOZ,A
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:16

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/29/2016 09:01:37

Diagnostics Check	OK	08:54
Air Blank	0.000	08:55
Control Test	0.077	08:55
Air Blank	0.000	08:56
Subject Sample # 1	0.204	08:56
Breath Volume	4.082	08:56
Air Blank	0.000	08:57
Air Blank	0.000	08:59
Subject Sample # 2	0.196	08:59
Breath Volume	3.699	08:59
Air Blank	0.000	09:00
Control Test	0.075	09:01
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 05/18/2016 12:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SMERLING

DATE/TIME OF TEST: 04/18/2016 21:43:37

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.078	21:43	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 04/29/2016 14:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1497755
DRY GAS EXPIRY DATE: 05/02/2016
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: CIFUENTES

DATE/TIME OF TEST: 04/03/2016 00:13:20

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.157	00:08
Breath Volume	1.328	00:08
Air Blank	0.000	00:08
Air Blank	0.000	00:10
Subject Sample # 2	0.157	00:11
Breath Volume	1.523	00:11
Air Blank	0.000	00:12
Control Test	0.080	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 04/21/2016 09:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: LOPEZ,CRISTIAN
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/23/2016 04:57:48

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:56	
Control Test	0.160	04:57	Control Outside Tolerance
Air Blank	0.000	04:57	

Data Download Date/Time: 05/07/2016 10:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,CRISTIAN,
BREATH TEST OPERATOR AGENCY: Medley PD
ARRESTING OFFICER: LOPEZ,CRISTIAN
ARRESTING OFFICER AGENCY: Medley PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/23/2016 05:12:02

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.184	05:07
Breath Volume	2.179	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.173	05:10
Breath Volume	4.359	05:10
Air Blank	0.000	05:11
Control Test	0.079	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:12

Data Download Date/Time: 05/07/2016 10:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: THORSEN

DATE/TIME OF TEST: 04/01/2016 00:54:53

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.078	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.155	00:50
Breath Volume	1.445	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:52
Subject Sample # 2	0.160	00:53
Breath Volume	1.460	00:53
Air Blank	0.000	00:53
Control Test	0.077	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: ANDERTON

DATE/TIME OF TEST: 04/01/2016 01:17:24

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.078	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Slope Not Met
Breath Volume	1.320	01:15	
Air Blank	0.000	01:16	
Control Test	0.079	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ANDERTON

DATE/TIME OF TEST: 04/01/2016 01:47:33

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.326	01:42
Breath Volume		2.039	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.336	01:45
Breath Volume		1.917	01:45
Air Blank		0.000	01:46
Control Test		0.078	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/01/2016 04:26:49

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.078	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.128	04:22
Breath Volume	1.394	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:24
Subject Sample # 2	0.131	04:25
Breath Volume	1.546	04:25
Air Blank	0.000	04:25
Control Test	0.078	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: BIEGLER JR

DATE/TIME OF TEST: 04/02/2016 01:33:30

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.079	01:27
Air Blank	0.000	01:28
Subject Sample # 1	0.187	01:28
Breath Volume	2.316	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.188	01:31
Breath Volume	2.867	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 04/02/2016 01:52:08

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Subject Sample # 1	0.143	01:47
Breath Volume	1.558	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.137	01:50
Breath Volume	1.503	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 04/02/2016 02:38:52

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.124	02:33
Breath Volume	1.429	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.130	02:37
Breath Volume	1.230	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 04/02/2016 04:01:54

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.103	03:57
Breath Volume	1.531	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.107	04:00
Breath Volume	2.515	04:00
Air Blank	0.000	04:00
Control Test	0.078	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/02/2016 06:41:29

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.078	06:36
Air Blank	0.000	06:36
Subject Sample # 1	0.144	06:37
Breath Volume	1.671	06:37
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.146	06:39
Breath Volume	2.382	06:39
Air Blank	0.000	06:40
Control Test	0.079	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: GONSALVES

DATE/TIME OF TEST: 04/03/2016 01:18:09

Diagnostics Check	OK	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Subject Sample # 1	0.000	01:13
Breath Volume	3.820	01:13
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.000	01:16
Breath Volume	3.773	01:16
Air Blank	0.000	01:17
Control Test	0.080	01:17
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/03/2016 02:45:17

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.200	02:40
Breath Volume	1.324	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.203	02:43
Breath Volume	2.425	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VALDES MIRANDA

DATE/TIME OF TEST: 04/03/2016 03:17:39

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.078	03:11
Air Blank	0.000	03:12
Subject Sample # 1	0.112	03:13
Breath Volume	1.546	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.118	03:16
Breath Volume	1.859	03:16
Air Blank	0.000	03:16
Control Test	0.078	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MACFADDEN

DATE/TIME OF TEST: 04/03/2016 04:45:10

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.078	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.116	04:39
Breath Volume	1.156	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.117	04:43
Breath Volume	1.156	04:43
Air Blank	0.000	04:44
Control Test	0.077	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: GURI,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/03/2016 05:12:06

Diagnostics Check	OK	05:03	
Air Blank	0.000	05:03	
Control Test	0.078	05:04	
Air Blank	0.000	05:04	
Subject Sample	# 1	05:07	No Sample Provided
Breath Volume	0.000	05:07	
Air Blank	0.000	05:08	
Air Blank	0.000	05:10	
Subject Sample	# 2	05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:11	
Control Test	0.078	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: LECOUNT III

DATE/TIME OF TEST: 04/03/2016 06:44:40

Diagnostics Check		OK	06:38
Air Blank		0.000	06:38
Control Test		0.078	06:39
Air Blank		0.000	06:39
Subject Sample	# 1	0.108	06:40
Breath Volume		2.066	06:40
Air Blank		0.000	06:40
Air Blank		0.000	06:42
Subject Sample	# 2	0.113	06:43
Breath Volume		2.179	06:43
Air Blank		0.000	06:43
Control Test		0.078	06:44
Air Blank		0.000	06:44
Diagnostics Check		OK	06:44

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS, V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: GONZALES

DATE/TIME OF TEST: 04/04/2016 00:50:10

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.193	00:45
Breath Volume		1.160	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:47
Subject Sample	# 2	0.197	00:48
Breath Volume		1.617	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: AMADOR FUXENCH

DATE/TIME OF TEST: 04/04/2016 02:12:14

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.078	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:10	Subject Test Refused
Breath Volume	0.000	02:10	
Air Blank	0.000	02:11	
Control Test	0.079	02:11	
Air Blank	0.000	02:12	
Diagnostics Check	OK	02:12	

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: NERI-LOPEZ

DATE/TIME OF TEST: 04/04/2016 04:41:30

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.088	04:36
Breath Volume	1.320	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.093	04:39
Breath Volume	2.281	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BREDESEN

DATE/TIME OF TEST: 04/05/2016 01:05:05

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.079	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.062	01:00
Breath Volume		4.632	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:02
Subject Sample	# 2	0.064	01:03
Breath Volume		4.527	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: DALAMA GONZALEZ

DATE/TIME OF TEST: 04/05/2016 03:11:11

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:05
Subject Sample # 1	0.267	03:06
Breath Volume	4.164	03:06
Air Blank	0.000	03:07
Air Blank	0.000	03:08
Subject Sample # 2	0.266	03:09
Breath Volume	2.765	03:09
Air Blank	0.000	03:10
Control Test	0.079	03:10
Air Blank	0.000	03:11
Diagnostics Check	OK	03:11

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ANDINO

DATE/TIME OF TEST: 04/06/2016 02:38:14

Diagnostics Check	OK	02:31
Air Blank	0.000	02:32
Control Test	0.078	02:32
Air Blank	0.000	02:33
Subject Sample # 1	0.182	02:33
Breath Volume	2.953	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.182	02:36
Breath Volume	2.929	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: MILLEDGE

DATE/TIME OF TEST: 04/06/2016 03:58:16

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.009	03:53
Breath Volume	2.289	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.009	03:56
Breath Volume	2.003	03:56
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 04/06/2016 19:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 04/07/2016 01:25:07

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.088	01:20
Breath Volume	1.730	01:20
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.086	01:23
Breath Volume	1.566	01:23
Air Blank	0.000	01:24
Control Test	0.078	01:24
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 05/11/2016 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PORTEE

DATE/TIME OF TEST: 04/08/2016 01:07:02

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank		0.000	00:56	
Subject Sample	# 1	0.132	00:59	Volume Not Met
Breath Volume		0.371	00:59	
Air Blank		0.000	00:59	
Air Blank		0.000	01:01	
Subject Sample	# 2	0.133	01:02	
Breath Volume		1.351	01:02	
Air Blank		0.000	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 3	0.139	01:05	
Breath Volume		1.843	01:05	
Air Blank		0.000	01:06	
Control Test		0.078	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 05/11/2016 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: JANVIER

DATE/TIME OF TEST: 04/09/2016 00:41:48

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.119	00:37
Breath Volume	1.675	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.127	00:40
Breath Volume	2.164	00:40
Air Blank	0.000	00:40
Control Test	0.077	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 05/11/2016 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PREP

DATE/TIME OF TEST: 04/09/2016 02:39:27

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.143	02:34
Breath Volume	1.492	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.146	02:37
Breath Volume	1.144	02:37
Air Blank	0.000	02:38
Control Test	0.078	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 05/11/2016 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 04/09/2016 03:56:04

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.077	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.186	03:50
Breath Volume		2.289	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.187	03:54
Breath Volume		1.984	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:56

Data Download Date/Time: 05/11/2016 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILCHER,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: FAUSTO

DATE/TIME OF TEST: 04/09/2016 04:33:22

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.078	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.035	04:28
Breath Volume	2.378	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.032	04:31
Breath Volume	2.519	04:31
Air Blank	0.000	04:32
Control Test	0.078	04:32
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 05/11/2016 11:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PELL,H
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: ELLINGTON

DATE/TIME OF TEST: 04/09/2016 05:10:40

Diagnostics Check	OK	05:07	
Air Blank	0.000	05:07	
Control Test	0.078	05:08	
Air Blank	0.000	05:08	
Subject Sample	# 1	05:09	Subject Test Refused
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Control Test	0.079	05:10	
Air Blank	0.000	05:10	
Diagnostics Check	OK	05:10	

Data Download Date/Time: 05/11/2016 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: MEUSLING

DATE/TIME OF TEST: 04/09/2016 06:32:59

Diagnostics Check		OK	06:25
Air Blank		0.000	06:26
Control Test		0.079	06:26
Air Blank		0.000	06:26
Subject Sample	# 1	0.130	06:28
Breath Volume		3.152	06:28
Air Blank		0.000	06:28
Air Blank		0.000	06:30
Subject Sample	# 2	0.133	06:31
Breath Volume		2.281	06:31
Air Blank		0.000	06:32
Control Test		0.078	06:32
Air Blank		0.000	06:32
Diagnostics Check		OK	06:32

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: ALVAREZ-ZARZGOZA

DATE/TIME OF TEST: 04/09/2016 21:30:33

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.079	21:24
Air Blank	0.000	21:25
Subject Sample # 1	0.118	21:25
Breath Volume	1.503	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.126	21:28
Breath Volume	2.238	21:28
Air Blank	0.000	21:29
Control Test	0.078	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 04/10/2016 03:13:38

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.079	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1	0.076	03:04	Volume Not Met
Breath Volume		1.089	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 2	0.080	03:07	
Breath Volume		1.781	03:07	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 3	0.081	03:12	
Breath Volume		2.207	03:12	
Air Blank		0.000	03:12	
Control Test		0.079	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: COLON ANZUETA

DATE/TIME OF TEST: 04/10/2016 05:16:26

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.117	05:12
Breath Volume		1.347	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.120	05:14
Breath Volume		1.207	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SALAZAR-MOLINA

DATE/TIME OF TEST: 04/10/2016 21:22:52

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.078	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.283	21:18
Breath Volume	1.281	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:20
Subject Sample # 2	0.273	21:21
Breath Volume	1.710	21:21
Air Blank	0.000	21:21
Control Test	0.078	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: DILL

DATE/TIME OF TEST: 04/10/2016 22:47:35

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.163	22:42
Breath Volume	1.242	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.175	22:45
Breath Volume	1.265	22:45
Air Blank	0.000	22:46
Control Test	0.078	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SYEDIN

DATE/TIME OF TEST: 04/11/2016 03:35:23

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.206	03:30
Breath Volume		2.214	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.210	03:33
Breath Volume		3.726	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 04/12/2016 01:07:16

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.079	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.000	01:03
Breath Volume	3.246	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.000	01:05
Breath Volume	2.695	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/12/2016 02:25:05

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Subject Sample # 1	0.181	02:20
Breath Volume	2.351	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.183	02:23
Breath Volume	2.792	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MICHELINI

DATE/TIME OF TEST: 04/14/2016 01:45:41

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:31	
Control Test		0.078	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.043	01:35	Volume Not Met
Breath Volume		0.203	01:35	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2	0.048	01:40	
Breath Volume		1.363	01:40	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 3	0.046	01:44	
Breath Volume		1.351	01:44	
Air Blank		0.000	01:44	
Control Test		0.079	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PEREZ JR

DATE/TIME OF TEST: 04/14/2016 03:52:51

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.117	03:47
Breath Volume	1.316	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.117	03:51
Breath Volume	2.003	03:51
Air Blank	0.000	03:51
Control Test	0.079	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SEQUEIRA,N
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: EDDINGS

DATE/TIME OF TEST: 04/15/2016 01:06:48

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.225	01:02
Breath Volume		2.675	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.234	01:05
Breath Volume		2.855	01:05
Air Blank		0.000	01:05
Control Test		0.082	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 04/15/2016 02:01:20

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.079	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:54	No Sample Provided
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Air Blank	0.000	01:56	
Subject Sample	# 2	01:59	No Sample Provided
Breath Volume	0.000	01:59	
Air Blank	0.000	02:00	
Control Test	0.078	02:00	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 04/15/2016 21:05:02

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.183	21:00
Breath Volume	1.781	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:02
Subject Sample # 2	0.191	21:03
Breath Volume	3.777	21:03
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:05

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HERNANDEZ-PEREZ

DATE/TIME OF TEST: 04/15/2016 23:22:52

Diagnostics Check	OK	23:16
Air Blank	0.000	23:16
Control Test	0.078	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.101	23:17
Breath Volume	1.296	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.106	23:21
Breath Volume	2.191	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: RAYYAN

DATE/TIME OF TEST: 04/16/2016 01:21:49

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.079	01:16
Air Blank	0.000	01:16
Subject Sample # 1	0.033	01:17
Breath Volume	4.156	01:17
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.031	01:20
Breath Volume	2.847	01:20
Air Blank	0.000	01:20
Control Test	0.078	01:21
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 04/16/2016 02:10:07

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.078	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.142	02:05
Breath Volume	1.855	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.133	02:08
Breath Volume	1.296	02:08
Air Blank	0.000	02:09
Control Test	0.077	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ROSALES-GOMEZ

DATE/TIME OF TEST: 04/16/2016 03:29:54

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank		03:29	RFI Detect
Air Blank	0.000	03:29	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ROSALES-GOMEZ

DATE/TIME OF TEST: 04/16/2016 03:38:22

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.043	03:33
Breath Volume	1.757	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.042	03:36
Breath Volume	1.429	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: GUZMA-PEDROZA

DATE/TIME OF TEST: 04/16/2016 23:07:04

Diagnostics Check	OK	22:59
Air Blank	0.000	22:59
Control Test	0.079	23:00
Air Blank	0.000	23:00
Subject Sample # 1	0.311	23:02
Breath Volume	2.437	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.306	23:05
Breath Volume	1.425	23:05
Air Blank	0.000	23:06
Control Test	0.077	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:07

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ, R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 04/17/2016 00:39:52

Diagnostics Check		OK	00:33
Air Blank		0.000	00:33
Control Test		0.078	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.241	00:35
Breath Volume		2.445	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:37
Subject Sample	# 2	0.231	00:38
Breath Volume		2.515	00:38
Air Blank		0.000	00:38
Control Test		0.078	00:39
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 04/17/2016 03:04:09

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.000	02:59
Breath Volume		1.226	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.000	03:02
Breath Volume		1.859	03:02
Air Blank		0.000	03:03
Control Test		0.078	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY, D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 04/17/2016 04:46:28

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.079	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.030	04:41
Breath Volume	1.617	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.031	04:44
Breath Volume	1.816	04:44
Air Blank	0.000	04:45
Control Test	0.078	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANGELICCHI, J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: VIXAMAR

DATE/TIME OF TEST: 04/17/2016 05:10:00

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.078	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.121	05:05
Breath Volume	2.003	05:05
Air Blank	0.000	05:06
Air Blank	0.000	05:07
Subject Sample # 2	0.121	05:08
Breath Volume	2.117	05:08
Air Blank	0.000	05:09
Control Test	0.078	05:09
Air Blank	0.000	05:09
Diagnostics Check	OK	05:10

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 04/18/2016 00:42:29

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.151	00:37
Breath Volume	1.542	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.163	00:40
Breath Volume	1.152	00:40
Air Blank	0.000	00:41
Control Test	0.077	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: BLONDER

DATE/TIME OF TEST: 04/18/2016 01:48:51

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.163	01:44
Breath Volume	1.425	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.162	01:47
Breath Volume	1.226	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BLOMBERG

DATE/TIME OF TEST: 04/18/2016 02:09:24

Diagnostics Check	OK	02:03
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.175	02:04
Breath Volume	1.328	02:04
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.177	02:07
Breath Volume	1.113	02:07
Air Blank	0.000	02:08
Control Test	0.078	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: VAN SPARRENTAK

DATE/TIME OF TEST: 04/19/2016 02:00:33

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.149	01:55
Breath Volume	2.519	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.155	01:58
Breath Volume	2.613	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: YOUNGBLOOD

DATE/TIME OF TEST: 04/20/2016 01:03:21

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.079	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1	0.144	00:55	Volume Not Met
Breath Volume		0.972	00:55	
Air Blank		0.000	00:56	
Air Blank		0.000	00:57	
Subject Sample	# 2	0.148	00:58	
Breath Volume		1.984	00:58	
Air Blank		0.000	00:59	
Air Blank		0.000	01:01	
Subject Sample	# 3	0.145	01:01	
Breath Volume		1.773	01:01	
Air Blank		0.000	01:02	
Control Test		0.078	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/21/2016 00:16:04

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:08	
Control Test	0.079	00:08	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:11	No Sample Provided
Breath Volume	0.000	00:11	
Air Blank	0.000	00:12	
Air Blank	0.000	00:14	
Subject Sample	# 2	00:14	Subject Test Refused
Breath Volume	0.000	00:14	
Air Blank	0.000	00:15	
Control Test	0.078	00:15	
Air Blank	0.000	00:15	
Diagnostics Check	OK	00:16	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MINAYA,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MATZZA TAGLE

DATE/TIME OF TEST: 04/21/2016 01:09:36

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.201	01:03
Breath Volume	1.187	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.214	01:07
Breath Volume	2.195	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SANGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/21/2016 23:53:43

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.078	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.224	23:48
Breath Volume	1.113	23:48
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.223	23:51
Breath Volume	1.664	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:53
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: EK

DATE/TIME OF TEST: 04/22/2016 03:42:52

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.079	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1 0.228	03:32	
Breath Volume	2.523	03:32	
Air Blank	0.000	03:33	
Air Blank	0.000	03:34	
Subject Sample	# 2 0.218	03:38	Volume Not Met
Breath Volume	1.062	03:38	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 3 0.215	03:41	
Breath Volume	1.160	03:41	
Air Blank	0.000	03:41	
Control Test	0.076	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BROERS

DATE/TIME OF TEST: 04/23/2016 00:40:32

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:39	Subject Test Refused
Breath Volume	0.000	00:39	
Air Blank	0.000	00:39	
Control Test	0.079	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: BROERS

DATE/TIME OF TEST: 04/23/2016 00:48:02

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.177	00:43
Breath Volume	2.214	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.176	00:46
Breath Volume	2.937	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:48

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 04/23/2016 01:52:28

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.143	01:47
Breath Volume	1.609	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.145	01:50
Breath Volume	1.289	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MUJICA ABELLA

DATE/TIME OF TEST: 04/23/2016 03:47:00

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.187	03:42
Breath Volume		2.078	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.180	03:45
Breath Volume		2.566	03:45
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: GASPAR ESTEBAN

DATE/TIME OF TEST: 04/23/2016 22:44:54

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.186	22:39
Breath Volume	1.324	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.184	22:43
Breath Volume	2.703	22:43
Air Blank	0.000	22:43
Control Test	0.078	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MANGANARO,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VOGEL

DATE/TIME OF TEST: 04/24/2016 00:38:40

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.153	00:34
Breath Volume	2.800	00:34
Air Blank	0.000	00:34
Air Blank	0.000	00:36
Subject Sample # 2	0.157	00:37
Breath Volume	2.921	00:37
Air Blank	0.000	00:37
Control Test	0.079	00:38
Air Blank	0.000	00:38
Diagnostics Check	OK	00:38

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SEQUEIRA,NC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: SALGADO-MONTES

DATE/TIME OF TEST: 04/24/2016 01:17:51

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.115	01:13
Breath Volume	2.023	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.115	01:16
Breath Volume	2.199	01:16
Air Blank	0.000	01:16
Control Test	0.077	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: TONEY

DATE/TIME OF TEST: 04/24/2016 02:19:10

Diagnostics Check	OK	02:16	
Air Blank	0.000	02:16	
Control Test	0.078	02:17	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GILCHRIST

DATE/TIME OF TEST: 04/24/2016 02:47:09

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.119	02:42
Breath Volume	1.816	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.117	02:45
Breath Volume	1.789	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: FREYRE

DATE/TIME OF TEST: 04/24/2016 04:14:34

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.078	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.126	04:10
Breath Volume		2.371	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.123	04:13
Breath Volume		2.121	04:13
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BRITO-BOLMEY

DATE/TIME OF TEST: 04/24/2016 23:54:48

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:45	
Control Test		0.078	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.065	23:49	Volume Not Met
Breath Volume		0.531	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:51	
Subject Sample	# 2		23:53	Subject Test Refused
Breath Volume		0.457	23:53	
Air Blank		0.000	23:53	
Control Test		0.078	23:54	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 04/26/2016 04:51:48

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.096	04:47
Breath Volume	1.808	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.099	04:50
Breath Volume	1.730	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: KROPP

DATE/TIME OF TEST: 04/26/2016 23:53:09

Diagnostics Check	OK	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.148	23:48
Breath Volume	1.648	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:50
Subject Sample # 2	0.152	23:51
Breath Volume	1.562	23:51
Air Blank	0.000	23:52
Control Test	0.079	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 04/27/2016 02:52:08

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.078	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.128	02:47
Breath Volume	3.835	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.129	02:50
Breath Volume	3.796	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 04/28/2016 02:45:45

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.111	02:41
Breath Volume	1.277	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.114	02:44
Breath Volume	1.304	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/29/2016 04:26:17

Diagnostics Check	OK	04:19
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.155	04:21
Breath Volume	2.199	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.160	04:24
Breath Volume	1.949	04:24
Air Blank	0.000	04:25
Control Test	0.079	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/30/2016 01:50:13

Diagnostics Check	OK	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:44
Air Blank	0.000	01:44
Subject Sample # 1	0.114	01:45
Breath Volume	1.570	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:47
Subject Sample # 2	0.124	01:48
Breath Volume	1.777	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 05/11/2016 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: AVENDANO

DATE/TIME OF TEST: 04/30/2016 02:10:07

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:04
Air Blank	0.000	02:05
Subject Sample # 1	0.000	02:05
Breath Volume	2.195	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.000	02:08
Breath Volume	2.507	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 05/11/2016 11:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: ARELLANO-CERVANTES

DATE/TIME OF TEST: 04/30/2016 03:19:01

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.016	03:14
Breath Volume	1.558	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.015	03:17
Breath Volume	1.339	03:17
Air Blank	0.000	03:18
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:19

Data Download Date/Time: 05/11/2016 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 04/30/2016 06:24:18

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:12	
Control Test		0.079	06:12	
Air Blank		0.000	06:13	
Subject Sample	# 1	0.097	06:16	Volume Not Met
Breath Volume		0.742	06:16	
Air Blank		0.000	06:16	
Air Blank		0.000	06:18	
Subject Sample	# 2	0.117	06:19	
Breath Volume		1.210	06:19	
Air Blank		0.000	06:19	
Air Blank		0.000	06:21	
Subject Sample	# 3	0.130	06:22	
Breath Volume		1.136	06:22	
Air Blank		0.000	06:23	
Control Test		0.078	06:23	
Air Blank		0.000	06:24	
Diagnostics Check		OK	06:24	

Data Download Date/Time: 05/11/2016 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 04/01/2016 03:42:48

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.081	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.140	03:38
Breath Volume	3.363	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.140	03:41
Breath Volume	3.359	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: BURLESON

DATE/TIME OF TEST: 04/01/2016 23:12:50

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.082	23:06
Air Blank	0.000	23:07
Subject Sample # 1	0.153	23:07
Breath Volume	1.222	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.166	23:11
Breath Volume	2.265	23:11
Air Blank	0.000	23:11
Control Test	0.081	23:12
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/02/2016 01:58:28

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.237	01:50
Breath Volume		2.886	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.200	01:53
Breath Volume		1.175	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 3	0.233	01:56
Breath Volume		2.835	01:56
Air Blank		0.000	01:57
Control Test		0.080	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/02/2016 02:20:14

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.080	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.152	02:15
Breath Volume	2.851	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.154	02:18
Breath Volume	2.789	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/02/2016 04:40:49

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:29	
Control Test		0.081	04:29	
Air Blank		0.000	04:30	
Subject Sample	# 1	0.111	04:33	Slope Not Level
Breath Volume		1.191	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2	0.127	04:39	Slope Not Level
Breath Volume		1.820	04:39	
Air Blank		0.000	04:39	
Control Test		0.079	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/02/2016 04:53:18

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.131	04:48
Breath Volume	2.496	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:50
Subject Sample # 2	0.129	04:51
Breath Volume	1.882	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GURI,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 04/02/2016 06:13:57

Diagnostics Check	OK	06:07
Air Blank	0.000	06:07
Control Test	0.081	06:07
Air Blank	0.000	06:08
Subject Sample # 1	0.000	06:09
Breath Volume	3.027	06:09
Air Blank	0.000	06:09
Air Blank	0.000	06:11
Subject Sample # 2	0.000	06:12
Breath Volume	3.593	06:12
Air Blank	0.000	06:13
Control Test	0.080	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:13

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:41

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/02/2016 20:08:15

Diagnostics Check	OK	20:01
Air Blank	0.000	20:02
Control Test	0.082	20:02
Air Blank	0.000	20:03
Subject Sample # 1	0.000	20:04
Breath Volume	2.101	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.000	20:06
Breath Volume	2.921	20:06
Air Blank	0.000	20:07
Control Test	0.082	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 04/03/2016 03:40:58

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.081	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.119	03:35
Breath Volume	1.367	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.118	03:39
Breath Volume	1.773	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: VELA MOTA

DATE/TIME OF TEST: 04/03/2016 04:50:18

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.154	04:45
Breath Volume	3.519	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.155	04:48
Breath Volume	2.375	04:48
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Diagnostics Check	OK	04:50

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: MARTINEZ ZAMBRANA

DATE/TIME OF TEST: 04/03/2016 05:24:57

Diagnostics Check	OK	05:18
Air Blank	0.000	05:19
Control Test	0.080	05:19
Air Blank	0.000	05:19
Subject Sample # 1	0.167	05:20
Breath Volume	3.421	05:20
Air Blank	0.000	05:21
Air Blank	0.000	05:22
Subject Sample # 2	0.170	05:23
Breath Volume	3.593	05:23
Air Blank	0.000	05:24
Control Test	0.080	05:24
Air Blank	0.000	05:24
Diagnostics Check	OK	05:24

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARRO,TAYLOR,C
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: AGENT

DATE/TIME OF TEST: 04/03/2016 22:44:32

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.082	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.243	22:39
Breath Volume	2.316	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.246	22:42
Breath Volume	2.628	22:42
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: SNELL

DATE/TIME OF TEST: 04/04/2016 01:44:49

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.082	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.173	01:40
Breath Volume	3.351	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.171	01:43
Breath Volume	1.519	01:43
Air Blank	0.000	01:43
Control Test	0.081	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SUAREZ PENA

DATE/TIME OF TEST: 04/04/2016 03:55:09

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.081	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.169	03:50
Breath Volume	1.539	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.172	03:53
Breath Volume	1.988	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 04/06/2016 19:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: KALBERER

DATE/TIME OF TEST: 04/05/2016 03:16:07

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.132	03:11
Breath Volume	3.730	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.135	03:14
Breath Volume	4.058	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 04/06/2016 19:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEVORE,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/05/2016 15:56:44

Diagnostics Check	OK	15:50
Air Blank	0.000	15:51
Control Test	0.081	15:51
Air Blank	0.000	15:51
Subject Sample # 1	0.000	15:52
Breath Volume	2.617	15:52
Air Blank	0.000	15:52
Air Blank	0.000	15:54
Subject Sample # 2	0.000	15:55
Breath Volume	1.164	15:55
Air Blank	0.000	15:55
Control Test	0.081	15:56
Air Blank	0.000	15:56
Diagnostics Check	OK	15:56

Data Download Date/Time: 04/06/2016 19:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 04/06/2016 01:55:55

Diagnostics Check	OK	01:47
Air Blank	0.000	01:48
Control Test	0.082	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.210	01:51
Breath Volume	1.472	01:51
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.224	01:54
Breath Volume	1.140	01:54
Air Blank	0.000	01:55
Control Test	0.081	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 04/06/2016 19:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALE,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PALTE

DATE/TIME OF TEST: 04/06/2016 03:46:27

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:45	
Control Test	0.081	03:45	
Air Blank		03:45	RFI Detect
Air Blank	0.000	03:46	

Data Download Date/Time: 04/06/2016 19:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALE,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PALTE

DATE/TIME OF TEST: 04/06/2016 03:57:11

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.149	03:52
Breath Volume	2.144	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.142	03:55
Breath Volume	1.382	03:55
Air Blank	0.000	03:56
Control Test	0.080	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/06/2016 19:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/06/2016 05:06:17

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.081	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.164	05:01
Breath Volume	1.625	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.168	05:04
Breath Volume	2.078	05:04
Air Blank	0.000	05:05
Control Test	0.080	05:05
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 04/06/2016 19:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ELIZABETH,B
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/07/2016 00:46:56

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:42	
Control Test	0.082	00:43	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	Sequence Aborted
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: DANIEL-CHAVEZ

DATE/TIME OF TEST: 04/08/2016 01:35:21

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.082	01:33	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.000	01:33	
Air Blank	0.000	01:34	
Control Test	0.082	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,ALIEK,O
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SIERRA,ALIEK
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/08/2016 01:55:19

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.082	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.133	01:50
Breath Volume		2.183	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.128	01:53
Breath Volume		1.468	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN-CHARLES,SANDRA,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JEAN-CHARLES,SANDRA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 12:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/09/2016 01:56:49

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.141	01:52
Breath Volume	3.355	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.142	01:55
Breath Volume	3.707	01:55
Air Blank	0.000	01:55
Control Test	0.081	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: VILLEDA GUERRERO

DATE/TIME OF TEST: 04/09/2016 06:25:51

Diagnostics Check	OK	06:20	
Air Blank	0.000	06:20	
Control Test	0.081	06:20	
Air Blank	0.000	06:21	
Subject Sample	# 1 0.000	06:24	Volume Not Met
Breath Volume	0.000	06:24	
Air Blank		06:25	Ambient Fail
Air Blank		06:25	Purge Fail

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: VILLEDA GUERRERO

DATE/TIME OF TEST: 04/09/2016 06:29:16

Diagnostics Check	OK	06:27	
Air Blank	0.000	06:28	
Control Test	0.068	06:28	Control Outside Tolerance
Air Blank	0.000	06:29	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: VILLEDA GUERRERO

DATE/TIME OF TEST: 04/09/2016 06:34:00

Diagnostics Check	OK	06:30	
Air Blank	0.000	06:31	
Control Test	0.080	06:31	
Air Blank	0.000	06:32	
Subject Sample	# 1	06:32	Subject Test Refused
Breath Volume	0.019	06:32	
Air Blank	0.000	06:33	
Control Test	0.081	06:33	
Air Blank	0.000	06:33	
Diagnostics Check	OK	06:34	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON, RAYMOND, E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ADDISON, RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 14:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/09/2016 16:44:02

Diagnostics Check	OK	16:39	
Air Blank	0.000	16:39	
Control Test	0.081	16:39	
Air Blank	0.000	16:40	
Subject Sample	# 1	16:42	Subject Test Refused
Breath Volume	0.000	16:42	
Air Blank	0.000	16:43	
Control Test	0.082	16:43	
Air Blank	0.000	16:43	
Diagnostics Check	OK	16:44	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SEGUI,YOSVANY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 04/10/2016 05:07:25

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.081	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:06	
Control Test	0.081	05:06	
Air Blank	0.000	05:07	
Diagnostics Check	OK	05:07	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOZANO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RODRIGUEZ,LUIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: PUENTE CANAY

DATE/TIME OF TEST: 04/10/2016 20:21:38

Diagnostics Check	OK	20:14
Air Blank	0.000	20:15
Control Test	0.082	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.218	20:17
Breath Volume	3.410	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.211	20:19
Breath Volume	2.609	20:19
Air Blank	0.000	20:20
Control Test	0.079	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: LASSO SILVA,NICOLAS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MOLINARES

DATE/TIME OF TEST: 04/14/2016 02:30:46

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:28	
Control Test	0.081	02:28	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:29	Subject Test Refused
Breath Volume	0.000	02:29	
Air Blank	0.000	02:29	
Control Test	0.081	02:30	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADDISON, RAYMOND, E
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA, RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BIARRIETA

DATE/TIME OF TEST: 04/16/2016 02:48:39

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.054	02:44
Breath Volume	3.234	02:44
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.055	02:47
Breath Volume	4.000	02:47
Air Blank	0.000	02:47
Control Test	0.082	02:48
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/16/2016 06:06:14

Diagnostics Check	OK	06:01	
Air Blank	0.000	06:02	
Control Test	0.082	06:02	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:04	Subject Test Refused
Breath Volume	0.000	06:04	
Air Blank	0.000	06:05	
Control Test	0.083	06:05	
Air Blank	0.000	06:06	
Diagnostics Check	OK	06:06	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,DANIEL,C
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VASQUEZ,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:56

SUBJECT'S LAST NAME: ROLAND

DATE/TIME OF TEST: 04/16/2016 06:23:57

Diagnostics Check	OK	06:16
Air Blank	0.000	06:17
Control Test	0.082	06:17
Air Blank	0.000	06:17
Subject Sample # 1	0.099	06:19
Breath Volume	1.609	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:21
Subject Sample # 2	0.096	06:22
Breath Volume	1.871	06:22
Air Blank	0.000	06:23
Control Test	0.078	06:23
Air Blank	0.000	06:23
Diagnostics Check	OK	06:23

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN-CHARLES,SANDRA,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JEAN-CHARLES,SANDRA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 04/17/2016 00:44:38

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:41	
Control Test	0.082	00:41	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.081	00:44	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEAN-CHARLES,SANDRA,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JEAN-CHARLES,SANDRA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 04/17/2016 01:17:24

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.081	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.081	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 04/21/2016 10:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: GONSALVES

DATE/TIME OF TEST: 04/03/2016 00:42:27

Diagnostics Check	OK	00:41	
Air Blank		00:41	RFI Detect
Air Blank	0.000	00:42	

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MACFADDEN

DATE/TIME OF TEST: 04/03/2016 03:49:44

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:47	RFI Detect
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.080	03:49	
Air Blank		03:49	Sequence Aborted
Air Blank	0.000	03:49	

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: MACFADDEN

DATE/TIME OF TEST: 04/03/2016 04:03:53

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.080	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.140	03:56	Volume Not Met
Breath Volume		0.015	03:56	
Air Blank		0.000	03:57	
Air Blank		0.000	03:59	
Subject Sample	# 2	0.110	04:02	Volume Not Met
Breath Volume		0.433	04:02	
Air Blank		0.000	04:02	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CALZADA CRUZ

DATE/TIME OF TEST: 04/05/2016 03:47:04

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.000	03:42
Breath Volume	1.980	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.000	03:45
Breath Volume	2.050	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:47

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOKENDORF,BRIAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DOKENDORF,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 04/08/2016 03:54:23

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.080	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.034	03:50
Breath Volume	3.062	03:50
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.034	03:52
Breath Volume	2.730	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CRANSTON

DATE/TIME OF TEST: 04/10/2016 05:37:29

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:26	
Control Test	0.080	05:26	
Air Blank	0.000	05:27	
Subject Sample	# 1	05:30	No Sample Provided
Breath Volume	0.000	05:30	
Air Blank	0.000	05:30	
Air Blank	0.000	05:32	
Subject Sample	# 2	05:36	No Sample Provided
Breath Volume	0.000	05:36	
Air Blank	0.000	05:36	
Control Test	0.079	05:36	
Air Blank	0.000	05:37	
Diagnostics Check	OK	05:37	

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: EHRHART,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: KOVACS

DATE/TIME OF TEST: 04/16/2016 05:06:05

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:57
Subject Sample # 1	0.110	04:59
Breath Volume	1.304	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:02
Subject Sample # 2	0.109	05:04
Breath Volume	1.515	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:05
Diagnostics Check	OK	05:06

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 04/17/2016 02:27:25

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.080	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:26	RFI Detect
Breath Volume	0.000	02:26	
Air Blank		02:26	RFI Detect
Air Blank	0.000	02:27	

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: TURPIN

DATE/TIME OF TEST: 04/23/2016 01:58:55

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.080	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.092	01:54
Breath Volume	1.945	01:54
Air Blank	0.000	01:54
Air Blank	0.000	01:56
Subject Sample # 2	0.089	01:57
Breath Volume	2.070	01:57
Air Blank	0.000	01:57
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 04/28/2016 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DONLEY

DATE/TIME OF TEST: 04/30/2016 01:42:59

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.080	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:41	Interferent Detect
Breath Volume	1.421	01:41	
Air Blank	0.000	01:42	
Control Test	0.080	01:42	Interferent Detect
Air Blank		01:42	RFI Detect

Data Download Date/Time: 05/22/2016 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003412

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,GEORGE,M
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RODRIGUEZ,JOSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: CONEY

DATE/TIME OF TEST: 04/09/2016 08:38:31

Diagnostics Check	OK	08:31
Air Blank	0.000	08:31
Control Test	0.080	08:32
Air Blank	0.000	08:32
Subject Sample # 1	0.136	08:33
Breath Volume	1.667	08:33
Air Blank	0.000	08:34
Air Blank	0.000	08:36
Subject Sample # 2	0.135	08:36
Breath Volume	2.707	08:36
Air Blank	0.000	08:37
Control Test	0.078	08:37
Air Blank	0.000	08:38
Diagnostics Check	OK	08:38

Data Download Date/Time: 05/03/2016 15:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: PARKER-MARLEY

DATE/TIME OF TEST: 04/01/2016 20:19:45

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.078	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.208	20:15
Breath Volume	1.519	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:17
Subject Sample # 2	0.219	20:18
Breath Volume	1.152	20:18
Air Blank	0.000	20:18
Control Test	0.077	20:19
Air Blank	0.000	20:19
Diagnostics Check	OK	20:19

Data Download Date/Time: 04/10/2016 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PURANJIT

DATE/TIME OF TEST: 04/02/2016 00:03:04

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:00	
Control Test	0.079	00:01	
Air Blank	0.000	00:01	
Subject Sample # 1	0.237	00:02	
Breath Volume	1.277	00:02	
Air Blank		00:02	RFI Detect
Air Blank		00:02	RFI Detect
Air Blank	0.000	00:03	

Data Download Date/Time: 04/10/2016 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: PURANJIT

DATE/TIME OF TEST: 04/02/2016 00:11:24

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.078	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.248	00:06
Breath Volume		1.160	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.238	00:09
Breath Volume		1.632	00:09
Air Blank		0.000	00:10
Control Test		0.077	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 04/10/2016 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: PASTAL

DATE/TIME OF TEST: 04/17/2016 02:20:42

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.079	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:19	Subject Test Refused
Breath Volume	0.000	02:19	
Air Blank	0.000	02:19	
Control Test	0.080	02:20	
Air Blank	0.000	02:20	
Diagnostics Check	OK	02:20	

Data Download Date/Time: 05/05/2016 17:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 04/22/2016 03:46:29

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.175	03:42
Breath Volume	1.546	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.168	03:45
Breath Volume	1.105	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 05/05/2016 17:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CANCEL

DATE/TIME OF TEST: 04/23/2016 02:32:32

Diagnostics Check	OK	02:26
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Subject Sample # 1	0.136	02:28
Breath Volume	2.441	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.143	02:30
Breath Volume	2.597	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 05/05/2016 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: BALAGUER-GARCIA

DATE/TIME OF TEST: 04/24/2016 00:59:27

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.079	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.202	00:55
Breath Volume	3.113	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.207	00:57
Breath Volume	2.820	00:57
Air Blank	0.000	00:58
Control Test	0.078	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/05/2016 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DIAZ-COLON

DATE/TIME OF TEST: 04/25/2016 02:46:45

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.169	02:42
Breath Volume	2.074	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.174	02:45
Breath Volume	1.503	02:45
Air Blank	0.000	02:45
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 05/05/2016 17:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLOW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 04/03/2016 02:40:58

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.078	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.137	02:36
Breath Volume	3.371	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.138	02:39
Breath Volume	2.691	02:39
Air Blank	0.000	02:40
Control Test	0.076	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/10/2016 13:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRIS,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLOW,CHRIS
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: WESTPHAL

DATE/TIME OF TEST: 04/29/2016 21:09:40

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.078	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.218	21:02
Breath Volume	2.757	21:02
Air Blank	0.000	21:03
Air Blank	0.000	21:05
Subject Sample # 2	0.206	21:07
Breath Volume	2.839	21:07
Air Blank	0.000	21:08
Control Test	0.077	21:09
Air Blank	0.000	21:09
Diagnostics Check	OK	21:09

Data Download Date/Time: 05/05/2016 17:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

INSTRUMENT REGISTERED TO: HURLBURT FIELD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWEESE,BRETT,M
BREATH TEST OPERATOR AGENCY: Hurlburt Field PD
ARRESTING OFFICER: DEWEESE,BRETT
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DEWEESE

DATE/TIME OF TEST: 04/11/2016 01:52:15

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.000	01:48
Breath Volume	1.238	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.000	01:50
Breath Volume	2.343	01:50
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 05/31/2016 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 04/01/2016 21:41:38

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:40	
Control Test	0.079	21:41	
Air Blank	0.000	21:41	
Diagnostics Check	OK	21:41	

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 04/02/2016 03:31:38

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:23	
Control Test	0.079	03:23	
Air Blank	0.000	03:24	
Subject Sample	# 1	03:27	No Sample Provided
Breath Volume	0.000	03:27	
Air Blank	0.000	03:27	
Air Blank	0.000	03:29	
Subject Sample	# 2	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.080	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,HENRY,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COOPER,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/03/2016 04:01:34

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.000	03:57
Breath Volume	3.921	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.000	04:00
Breath Volume	2.878	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MAY EHUAN

DATE/TIME OF TEST: 04/05/2016 04:47:21

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.081	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.197	04:42
Breath Volume	3.257	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.199	04:45
Breath Volume	3.011	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: MURCHISON

DATE/TIME OF TEST: 04/09/2016 00:33:18

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:31	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 04/09/2016 04:26:46

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.067	04:21
Breath Volume	2.957	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.060	04:25
Breath Volume	1.722	04:25
Air Blank	0.000	04:25
Control Test	0.078	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRUILL,KOURTNEY,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,LANCE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/10/2016 06:24:48

Diagnostics Check		OK	06:12	
Air Blank		0.000	06:13	
Control Test		0.080	06:13	
Air Blank		0.000	06:14	
Subject Sample	# 1	0.160	06:17	Volume Not Met
Breath Volume		0.777	06:17	
Air Blank		0.000	06:17	
Air Blank		0.000	06:19	
Subject Sample	# 2	0.166	06:20	
Breath Volume		2.667	06:20	
Air Blank		0.000	06:21	
Air Blank		0.000	06:22	
Subject Sample	# 3	0.171	06:23	
Breath Volume		2.554	06:23	
Air Blank		0.000	06:23	
Control Test		0.079	06:24	
Air Blank		0.000	06:24	
Diagnostics Check		OK	06:24	

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,JAMIESON,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROSS,JAMIESON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 04/11/2016 05:50:11

Diagnostics Check	OK	05:42
Air Blank	0.000	05:43
Control Test	0.080	05:43
Air Blank	0.000	05:43
Subject Sample # 1	0.165	05:45
Breath Volume	1.875	05:45
Air Blank	0.000	05:45
Air Blank	0.000	05:47
Subject Sample # 2	0.168	05:48
Breath Volume	1.343	05:48
Air Blank	0.000	05:49
Control Test	0.079	05:49
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVINGSTON, RONALD, C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ADKINS, JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: FERRIS

DATE/TIME OF TEST: 04/11/2016 20:46:51

Diagnostics Check	OK	20:40
Air Blank	0.000	20:41
Control Test	0.080	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.143	20:42
Breath Volume	1.242	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:44
Subject Sample # 2	0.143	20:45
Breath Volume	1.847	20:45
Air Blank	0.000	20:45
Control Test	0.079	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCFALL,CHARLES,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MCFALL,CHARLES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 04/11/2016 21:42:20

Diagnostics Check	OK	21:38	
Air Blank	0.000	21:39	
Control Test	0.079	21:39	
Air Blank	0.000	21:40	
Subject Sample	# 1	21:40	Subject Test Refused
Breath Volume	0.000	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREJEAN,JESSICA,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PREJEAN,JESSICA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:54

SUBJECT'S LAST NAME: KONTOS

DATE/TIME OF TEST: 04/13/2016 12:28:39

Diagnostics Check	OK	12:14	
Air Blank	0.000	12:14	
Control Test	0.079	12:14	
Air Blank	0.000	12:15	
Subject Sample	# 1 0.303	12:15	
Breath Volume	1.281	12:15	
Air Blank	0.000	12:16	
Air Blank	0.000	12:18	
Subject Sample	# 2 0.255	12:21	Volume Not Met
Breath Volume	0.480	12:21	
Air Blank	0.000	12:22	
Air Blank	0.000	12:23	
Subject Sample	# 3 0.273	12:27	Volume Not Met
Breath Volume	0.503	12:27	
Air Blank	0.000	12:27	
Control Test	0.074	12:28	Control Outside Tolerance
Air Blank	0.000	12:28	

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CUTLER,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PETERSEN,CUTLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: OAKES

DATE/TIME OF TEST: 04/13/2016 23:12:29

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.189	23:07
Breath Volume	2.617	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.196	23:10
Breath Volume	3.390	23:10
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Diagnostics Check	OK	23:12

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 04/16/2016 01:50:25

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.079	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 04/30/2016 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: OGUIN

DATE/TIME OF TEST: 04/16/2016 17:50:56

Diagnostics Check	OK	17:44
Air Blank	0.000	17:45
Control Test	0.079	17:45
Air Blank	0.000	17:46
Subject Sample # 1	0.000	17:46
Breath Volume	2.410	17:46
Air Blank	0.000	17:47
Air Blank	0.000	17:49
Subject Sample # 2	0.000	17:49
Breath Volume	2.464	17:49
Air Blank	0.000	17:50
Control Test	0.079	17:50
Air Blank	0.000	17:50
Diagnostics Check	OK	17:50

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 04/17/2016 00:46:58

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.079	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.080	00:46	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE,SCOTT,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPENCE,SCOTT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: POOLE

DATE/TIME OF TEST: 04/17/2016 05:16:02

Diagnostics Check	OK	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:09
Subject Sample # 1	0.165	05:11
Breath Volume	2.835	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:13
Subject Sample # 2	0.149	05:14
Breath Volume	2.535	05:14
Air Blank	0.000	05:15
Control Test	0.079	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:16

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAGGAN,WESLEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 04/17/2016 05:46:18

Diagnostics Check		OK	05:31	
Air Blank		0.000	05:31	
Control Test		0.080	05:32	
Air Blank		0.000	05:32	
Subject Sample	# 1	0.186	05:35	Volume Not Met
Breath Volume		0.125	05:35	
Air Blank		0.000	05:36	
Air Blank		0.000	05:38	
Subject Sample	# 2	0.181	05:39	No .020 agreement
Breath Volume		1.757	05:39	
Air Blank		0.000	05:40	
Air Blank		0.000	05:42	
Subject Sample	# 3	0.205	05:44	No .020 agreement
Breath Volume		1.878	05:44	
Air Blank		0.000	05:45	
Control Test		0.076	05:45	
Air Blank		0.000	05:46	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAGGAN,WESLEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 04/17/2016 06:04:15

Diagnostics Check		OK	05:50	
Air Blank		0.000	05:51	
Control Test		0.078	05:51	
Air Blank		0.000	05:52	
Subject Sample	# 1	0.209	05:53	
Breath Volume		1.515	05:53	
Air Blank		0.000	05:53	
Air Blank		0.000	05:55	
Subject Sample	# 2	0.203	05:58	Volume Not Met
Breath Volume		0.128	05:58	
Air Blank		0.000	05:59	
Air Blank		0.000	06:01	
Subject Sample	# 3	0.209	06:02	
Breath Volume		1.707	06:02	
Air Blank		0.000	06:03	
Control Test		0.075	06:03	
Air Blank		0.000	06:04	
Diagnostics Check		OK	06:04	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CUTLER,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 04/17/2016 17:22:29

Diagnostics Check	OK	17:16
Air Blank	0.000	17:16
Control Test	0.079	17:16
Air Blank	0.000	17:17
Subject Sample # 1	0.106	17:18
Breath Volume	2.843	17:18
Air Blank	0.000	17:18
Air Blank	0.000	17:20
Subject Sample # 2	0.104	17:20
Breath Volume	2.308	17:20
Air Blank	0.000	17:21
Control Test	0.078	17:21
Air Blank	0.000	17:22
Diagnostics Check	OK	17:22

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: JENNINGS,BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 04/17/2016 22:22:17

Diagnostics Check	OK	22:15
Air Blank	0.000	22:16
Control Test	0.080	22:16
Air Blank	0.000	22:17
Subject Sample # 1	0.086	22:17
Breath Volume	2.019	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.084	22:20
Breath Volume	1.890	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CUTLER,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PETERSEN,CUTLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOYET

DATE/TIME OF TEST: 04/18/2016 02:34:20

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:31	
Control Test	0.079	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:32	Subject Test Refused
Breath Volume	0.000	02:32	
Air Blank	0.000	02:33	
Control Test	0.079	02:33	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NAYLOR,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 04/18/2016 20:42:24

Diagnostics Check	OK	20:39	
Air Blank	0.000	20:39	
Control Test	0.080	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:41	Subject Test Refused
Breath Volume	0.000	20:41	
Air Blank	0.000	20:41	
Control Test	0.079	20:41	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL, COLLEEN, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES, DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/23/2016 00:03:59

Diagnostics Check	OK	00:00	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:01	
Subject Sample	# 1	00:02	Subject Test Refused
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.080	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 04/23/2016 00:29:23

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.080	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.156	00:24
Breath Volume	1.195	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.165	00:27
Breath Volume	2.226	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CUTLER,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PETERSEN,CUTLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DIZDAREVIC

DATE/TIME OF TEST: 04/23/2016 01:35:17

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.079	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.259	01:30
Breath Volume		4.410	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.264	01:33
Breath Volume		4.199	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CUTLER,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PETERSEN,CUTLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: NEVAREZ-RODRIGUEZ

DATE/TIME OF TEST: 04/23/2016 03:59:10

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.000	03:54
Breath Volume	3.281	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.000	03:57
Breath Volume	3.726	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: GODFREY

DATE/TIME OF TEST: 04/23/2016 23:20:50

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:16	No Sample Provided
Breath Volume	0.000	23:16	
Air Blank	0.000	23:17	
Air Blank	0.000	23:19	
Subject Sample	# 2	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.079	23:20	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: FANELLA

DATE/TIME OF TEST: 04/24/2016 02:10:49

Diagnostics Check	OK	02:02
Air Blank	0.000	02:03
Control Test	0.080	02:03
Air Blank	0.000	02:04
Subject Sample # 1	0.174	02:05
Breath Volume	1.710	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.173	02:09
Breath Volume	1.621	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAND,GLENN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 04/24/2016 23:34:45

Diagnostics Check	OK	23:30	
Air Blank	0.000	23:31	
Control Test	0.079	23:31	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.079	23:34	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSEN,CUTLER,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PETERSEN,CUTLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ABADIE

DATE/TIME OF TEST: 04/27/2016 22:35:05

Diagnostics Check	OK	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Subject Sample # 1	0.220	22:30
Breath Volume	2.015	22:30
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.200	22:33
Breath Volume	1.800	22:33
Air Blank	0.000	22:34
Control Test	0.079	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:35

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS,JAMIESON,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROSS,JAMIESON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GLERUM

DATE/TIME OF TEST: 04/28/2016 02:37:12

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NABULSI,MOOTAZ
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/28/2016 23:14:26

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.079	23:12	
Air Blank	0.000	23:12	
Subject Sample	# 1	23:13	Subject Test Refused
Breath Volume	0.000	23:13	
Air Blank	0.000	23:13	
Control Test	0.078	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAGGAN,WESLEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: HERRIN

DATE/TIME OF TEST: 04/29/2016 05:41:28

Diagnostics Check	OK	05:38	
Air Blank	0.000	05:39	
Control Test	0.079	05:39	
Air Blank	0.000	05:39	
Subject Sample	# 1	05:40	Subject Test Refused
Breath Volume	0.000	05:40	
Air Blank	0.000	05:40	
Control Test	0.079	05:40	
Air Blank	0.000	05:41	
Diagnostics Check	OK	05:41	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: JENNINGS,BRANDON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 04/30/2016 01:09:00

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,CASEY,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DENNIS,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 03/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MOREY

DATE/TIME OF TEST: 04/30/2016 03:26:33

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.169	03:22
Breath Volume		4.027	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.167	03:25
Breath Volume		4.078	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 04/30/2016 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEROUCHE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: AUSSERER

DATE/TIME OF TEST: 04/08/2016 20:14:43

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.081	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.255	20:09
Breath Volume		2.527	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:11
Subject Sample	# 2	0.254	20:13
Breath Volume		1.257	20:13
Air Blank		0.000	20:13
Control Test		0.081	20:14
Air Blank		0.000	20:14
Diagnostics Check		OK	20:14

Data Download Date/Time: 06/01/2016 16:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEROUCHE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: IRELAND

DATE/TIME OF TEST: 04/08/2016 23:47:14

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.253	23:42
Breath Volume	2.074	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.243	23:45
Breath Volume	1.644	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 06/01/2016 16:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ZIPPERER

DATE/TIME OF TEST: 04/01/2016 01:00:29

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.078	00:54
Air Blank	0.000	00:54
Subject Sample # 1	0.129	00:55
Breath Volume	1.789	00:55
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.133	00:58
Breath Volume	1.433	00:58
Air Blank	0.000	00:59
Control Test	0.076	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: ARRIETA MERCADO

DATE/TIME OF TEST: 04/02/2016 02:33:18

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.182	02:27
Breath Volume	1.984	02:27
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.177	02:31
Breath Volume	1.128	02:31
Air Blank	0.000	02:32
Control Test	0.075	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: LOPEZ SIERRA

DATE/TIME OF TEST: 04/03/2016 03:24:42

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.079	03:13	
Air Blank		0.000	03:13	
Subject Sample	# 1	0.207	03:16	Volume Not Met
Breath Volume		0.066	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 2	0.244	03:19	No .020 agreement
Breath Volume		1.867	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 3	0.219	03:23	No .020 agreement
Breath Volume		2.015	03:23	
Air Blank		0.000	03:23	
Control Test		0.075	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: RODRIGUEZ QUINTERO

DATE/TIME OF TEST: 04/09/2016 23:43:39

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.124	23:38
Breath Volume	1.421	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.136	23:42
Breath Volume	2.535	23:42
Air Blank	0.000	23:42
Control Test	0.078	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 04/14/2016 18:12:12

Diagnostics Check	OK	18:03
Air Blank	0.000	18:03
Control Test	0.078	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.259	18:05
Breath Volume	1.152	18:05
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.257	18:10
Breath Volume	1.535	18:10
Air Blank	0.000	18:11
Control Test	0.077	18:11
Air Blank	0.000	18:12
Diagnostics Check	OK	18:12

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTAYL, MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 04/16/2016 01:37:56

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.079	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.116	01:33
Breath Volume		2.152	01:33
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.121	01:36
Breath Volume		2.394	01:36
Air Blank		0.000	01:37
Control Test		0.077	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SOLANA

DATE/TIME OF TEST: 04/16/2016 23:43:06

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.078	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.072	23:38
Breath Volume	4.562	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.068	23:41
Breath Volume	4.234	23:41
Air Blank	0.000	23:42
Control Test	0.078	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, TYLER, J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER, TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: GOUX

DATE/TIME OF TEST: 04/24/2016 01:43:16

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.125	01:38
Breath Volume		2.207	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.134	01:41
Breath Volume		2.472	01:41
Air Blank		0.000	01:42
Control Test		0.077	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 04/28/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: LATORRE

DATE/TIME OF TEST: 04/01/2016 02:59:32

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.078	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.117	02:55
Breath Volume	2.558	02:55
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.114	02:58
Breath Volume	2.335	02:58
Air Blank	0.000	02:58
Control Test	0.078	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 04/06/2016 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,BRIAN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/16/2016 02:48:06

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.077	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.184	02:43
Breath Volume	2.820	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.175	02:46
Breath Volume	2.804	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/23/2016 01:15:15

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.078	01:09
Air Blank	0.000	01:10
Subject Sample # 1	0.197	01:10
Breath Volume	2.316	01:10
Air Blank	0.000	01:11
Air Blank	0.000	01:13
Subject Sample # 2	0.209	01:13
Breath Volume	2.738	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MERCADOROSARIO

DATE/TIME OF TEST: 04/23/2016 04:56:08

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.078	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.152	04:51
Breath Volume	1.953	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:53
Subject Sample # 2	0.156	04:54
Breath Volume	2.230	04:54
Air Blank	0.000	04:55
Control Test	0.078	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: FIGUEORAGONZALEZ

DATE/TIME OF TEST: 04/24/2016 03:55:29

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.078	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.018	03:51
Breath Volume	2.445	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.016	03:54
Breath Volume	1.796	03:54
Air Blank	0.000	03:54
Control Test	0.078	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: JACINTO

DATE/TIME OF TEST: 04/24/2016 18:34:14

Diagnostics Check	OK	18:28
Air Blank	0.000	18:28
Control Test	0.078	18:29
Air Blank	0.000	18:29
Subject Sample # 1	0.329	18:29
Breath Volume	3.425	18:29
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.324	18:32
Breath Volume	1.453	18:32
Air Blank	0.000	18:33
Control Test	0.078	18:33
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: LUKAS

DATE/TIME OF TEST: 04/29/2016 00:09:28

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.078	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	RFI Detect
Breath Volume	0.187	00:08	
Air Blank		00:08	RFI Detect
Air Blank	0.000	00:09	

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: LUKAS

DATE/TIME OF TEST: 04/29/2016 00:20:22

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.078	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.140	00:13
Breath Volume	1.628	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:16
Subject Sample # 2	0.135	00:18
Breath Volume	1.425	00:18
Air Blank	0.000	00:19
Control Test	0.078	00:19
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/29/2016 03:20:00

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.078	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1 0.186	03:13	
Breath Volume	2.488	03:13	
Air Blank	0.000	03:14	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:18	Sequence Aborted
Breath Volume	0.132	03:18	
Air Blank	0.000	03:20	

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/29/2016 03:28:26

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.079	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1 0.190	03:23	
Breath Volume	2.265	03:23	
Air Blank	0.000	03:23	
Air Blank	0.000	03:25	
Subject Sample	# 2 0.176	03:26	
Breath Volume	1.687	03:26	
Air Blank	0.000	03:27	
Control Test	0.095	03:27	Interferent Detect
Air Blank		03:28	Purge Fail

Data Download Date/Time: 05/11/2016 10:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 04/06/2016 19:12:21

Diagnostics Check	OK	19:06
Air Blank	0.000	19:06
Control Test	0.083	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.000	19:08
Breath Volume	2.628	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.000	19:10
Breath Volume	2.460	19:10
Air Blank	0.000	19:11
Control Test	0.084	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 04/12/2016 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERRY,TAMMY,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERALL,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 04/07/2016 00:24:10

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.083	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.141	00:19
Breath Volume	1.500	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.141	00:22
Breath Volume	1.531	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 04/12/2016 10:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,BEN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: BANDALY

DATE/TIME OF TEST: 04/08/2016 23:56:16

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.084	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.000	23:51
Breath Volume	1.902	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.000	23:54
Breath Volume	1.902	23:54
Air Blank	0.000	23:55
Control Test	0.083	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 04/12/2016 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GARDNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: DZIEGIELEWSKI

DATE/TIME OF TEST: 04/10/2016 02:14:37

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.083	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 1	0.118	02:04	Volume Not Met
Breath Volume		0.785	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 2	0.125	02:08	
Breath Volume		1.187	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 3	0.121	02:13	
Breath Volume		1.136	02:13	
Air Blank		0.000	02:13	
Control Test		0.081	02:14	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 04/12/2016 10:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEETZ,JUSTIN,W
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MELANSON,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: KWEDOR

DATE/TIME OF TEST: 04/13/2016 17:43:08

Diagnostics Check	OK	17:40	
Air Blank	0.000	17:40	
Control Test	0.081	17:40	
Air Blank	0.000	17:41	
Subject Sample	# 1	17:41	Subject Test Refused
Breath Volume	0.000	17:41	
Air Blank	0.000	17:42	
Control Test	0.082	17:42	
Air Blank	0.000	17:43	
Diagnostics Check	OK	17:43	

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MARRERO

DATE/TIME OF TEST: 04/16/2016 04:19:09

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.082	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.217	04:13
Breath Volume	1.558	04:13
Air Blank	0.000	04:13
Air Blank	0.000	04:15
Subject Sample # 2	0.200	04:17
Breath Volume	1.695	04:17
Air Blank	0.000	04:18
Control Test	0.082	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: STOCKTON,C
ARRESTING OFFICER AGENCY: Brooksville PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 04/16/2016 21:57:22

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:46	
Control Test		0.082	21:46	
Air Blank		0.000	21:46	
Subject Sample	# 1	0.265	21:50	Volume Not Met
Breath Volume		0.679	21:50	
Air Blank		0.000	21:50	
Air Blank		0.000	21:52	
Subject Sample	# 2	0.272	21:55	Volume Not Met
Breath Volume		0.890	21:55	
Air Blank		0.000	21:56	
Control Test		0.079	21:56	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MASCITELLI,V
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: WAY

DATE/TIME OF TEST: 04/16/2016 23:39:04

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.082	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.179	23:34
Breath Volume	2.082	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.177	23:37
Breath Volume	1.664	23:37
Air Blank	0.000	23:38
Control Test	0.080	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:39

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 04/17/2016 20:18:04

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.082	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.143	20:13
Breath Volume	3.058	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:15
Subject Sample # 2	0.143	20:16
Breath Volume	2.425	20:16
Air Blank	0.000	20:17
Control Test	0.081	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:18

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLITO,BRAD,R
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FRYE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: RYDEEN

DATE/TIME OF TEST: 04/18/2016 02:58:37

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.082	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:57	Subject Test Refused
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Control Test	0.082	02:58	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BELFRY,NICK
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: QUIONES

DATE/TIME OF TEST: 04/19/2016 05:33:30

Diagnostics Check	OK	05:26
Air Blank	0.000	05:26
Control Test	0.082	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.204	05:27
Breath Volume	3.425	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.185	05:31
Breath Volume	3.187	05:31
Air Blank	0.000	05:32
Control Test	0.081	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA, MICHAEL, P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL, B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 04/30/2016 02:04:32

Diagnostics Check	OK	01:57
Air Blank	0.000	01:58
Control Test	0.083	01:58
Air Blank	0.000	01:59
Subject Sample # 1	0.140	01:59
Breath Volume	3.300	01:59
Air Blank	0.000	02:00
Air Blank	0.000	02:02
Subject Sample # 2	0.139	02:02
Breath Volume	4.085	02:02
Air Blank	0.000	02:03
Control Test	0.082	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEJESUS,K,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BRIDGES,D
ARRESTING OFFICER AGENCY: Brooksville PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: STROTHER

DATE/TIME OF TEST: 04/30/2016 07:32:57

Diagnostics Check	OK	07:23
Air Blank	0.000	07:23
Control Test	0.083	07:24
Air Blank	0.000	07:24
Subject Sample # 1	0.136	07:27
Breath Volume	2.152	07:27
Air Blank	0.000	07:27
Air Blank	0.000	07:29
Subject Sample # 2	0.134	07:31
Breath Volume	1.214	07:31
Air Blank	0.000	07:32
Control Test	0.082	07:32
Air Blank	0.000	07:32
Diagnostics Check	OK	07:32

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRITZ,STEPHEN,A
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: GABRIELE

DATE/TIME OF TEST: 04/30/2016 09:56:59

Diagnostics Check	OK	09:51
Air Blank	0.000	09:51
Control Test	0.083	09:51
Air Blank	0.000	09:52
Subject Sample # 1	0.000	09:52
Breath Volume	2.148	09:52
Air Blank	0.000	09:53
Air Blank	0.000	09:55
Subject Sample # 2	0.000	09:55
Breath Volume	2.527	09:55
Air Blank	0.000	09:56
Control Test	0.082	09:56
Air Blank	0.000	09:56
Diagnostics Check	OK	09:56

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA, MICHAEL, P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL, B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: GARCIA-ALMANTE

DATE/TIME OF TEST: 04/30/2016 20:36:35

Diagnostics Check	OK	20:33	
Air Blank	0.000	20:34	
Control Test	0.083	20:34	
Air Blank	0.000	20:34	
Subject Sample	# 1	20:35	Subject Test Refused
Breath Volume	0.000	20:35	
Air Blank	0.000	20:35	
Control Test	0.082	20:35	
Air Blank	0.000	20:36	
Diagnostics Check	OK	20:36	

Data Download Date/Time: 05/20/2016 08:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: HILL-TIGLAO

DATE/TIME OF TEST: 04/01/2016 00:27:50

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:22
Subject Sample # 1	0.000	00:23
Breath Volume	1.285	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.000	00:26
Breath Volume	1.718	00:26
Air Blank	0.000	00:26
Control Test	0.080	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ROYALS

DATE/TIME OF TEST: 04/02/2016 01:01:40

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1	01:00	Subject Test Refused
Breath Volume	0.000	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:01	
Air Blank	0.000	01:01	
Diagnostics Check	OK	01:01	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 04/02/2016 01:55:09

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.081	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.000	01:50
Breath Volume	4.171	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.000	01:53
Breath Volume	3.894	01:53
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/02/2016 03:59:55

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.082	03:55
Breath Volume	1.382	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.077	03:58
Breath Volume	1.359	03:58
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: SEIDEL

DATE/TIME OF TEST: 04/02/2016 04:51:16

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:45
Subject Sample # 1	0.147	04:46
Breath Volume	1.808	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.165	04:49
Breath Volume	1.222	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/02/2016 21:47:59

Diagnostics Check	OK	21:45	
Air Blank	0.000	21:45	
Control Test	0.080	21:45	
Air Blank	0.000	21:46	
Subject Sample	# 1	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:47	
Control Test	0.080	21:47	
Air Blank	0.000	21:47	
Diagnostics Check	OK	21:47	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 04/03/2016 02:55:51

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:53	
Control Test	0.081	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:54	Subject Test Refused
Breath Volume	0.000	02:54	
Air Blank	0.000	02:54	
Control Test	0.080	02:55	
Air Blank	0.000	02:55	
Diagnostics Check	OK	02:55	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,AMANDA,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WARD,AMANDA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 04/04/2016 01:07:54

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:05	
Control Test	0.082	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:06	Subject Test Refused
Breath Volume	0.000	01:06	
Air Blank	0.000	01:06	
Control Test	0.080	01:07	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SKOGLUND

DATE/TIME OF TEST: 04/05/2016 04:14:56

Diagnostics Check	OK	04:12	
Air Blank	0.000	04:12	
Control Test	0.081	04:12	
Air Blank	0.000	04:13	
Subject Sample	# 1	04:13	Subject Test Refused
Breath Volume	0.000	04:13	
Air Blank	0.000	04:14	
Control Test	0.080	04:14	
Air Blank	0.000	04:14	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MIELKE

DATE/TIME OF TEST: 04/06/2016 02:06:52

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:02	
Subject Sample # 1	0.177	02:02	
Breath Volume	1.523	02:02	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample # 2	0.176	02:06	
Breath Volume	1.507	02:06	
Air Blank		02:06	RFI Detect
Air Blank	0.000	02:06	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,CHESTER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BECKFORD

DATE/TIME OF TEST: 04/06/2016 04:05:00

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.081	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.000	03:59
Breath Volume	1.230	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.000	04:03
Breath Volume	1.871	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:05

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 04/06/2016 23:02:27

Diagnostics Check	OK	22:56
Air Blank	0.000	22:56
Control Test	0.081	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.000	22:58
Breath Volume	2.664	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.000	23:01
Breath Volume	3.527	23:01
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: NAVES

DATE/TIME OF TEST: 04/08/2016 02:28:59

Diagnostics Check	OK	02:22
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.123	02:24
Breath Volume	1.730	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.125	02:27
Breath Volume	1.734	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/08/2016 04:37:40

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:35	
Control Test	0.080	04:35	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:36	Subject Test Refused
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Control Test	0.081	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RIDEOUT

DATE/TIME OF TEST: 04/09/2016 01:07:22

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.096	01:02
Breath Volume	1.667	01:02
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.094	01:05
Breath Volume	1.140	01:05
Air Blank	0.000	01:06
Control Test	0.079	01:06
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: LITTELL

DATE/TIME OF TEST: 04/09/2016 03:25:44

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.080	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.080	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 04/09/2016 04:34:20

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.133	04:29
Breath Volume	1.871	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.136	04:32
Breath Volume	1.394	04:32
Air Blank	0.000	04:33
Control Test	0.079	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/09/2016 05:19:34

Diagnostics Check	OK	05:16	
Air Blank	0.000	05:17	
Control Test	0.079	05:17	
Air Blank	0.000	05:17	
Subject Sample	# 1	05:18	Subject Test Refused
Breath Volume	0.000	05:18	
Air Blank	0.000	05:18	
Control Test	0.079	05:18	
Air Blank	0.000	05:19	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 04/09/2016 23:37:19

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.082	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	2.277	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	1.382	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: DOBER

DATE/TIME OF TEST: 04/10/2016 00:25:15

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.082	00:23	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:24	
Control Test	0.082	00:24	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: RIGSBEE

DATE/TIME OF TEST: 04/10/2016 04:38:35

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.081	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.209	04:33
Breath Volume	2.917	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.204	04:36
Breath Volume	2.406	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,R,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NOVAK,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/11/2016 01:57:44

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.081	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.000	01:53
Breath Volume	1.554	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.000	01:56
Breath Volume	1.703	01:56
Air Blank	0.000	01:56
Control Test	0.081	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: HECKER

DATE/TIME OF TEST: 04/12/2016 22:25:50

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.081	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.098	22:21
Breath Volume	1.402	22:21
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.106	22:24
Breath Volume	1.660	22:24
Air Blank	0.000	22:24
Control Test	0.080	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MARTIN,V
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARLAN,PATRICK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 11:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: SOUDER

DATE/TIME OF TEST: 04/13/2016 13:10:33

Diagnostics Check	OK	13:04
Air Blank	0.000	13:04
Control Test	0.081	13:05
Air Blank	0.000	13:05
Subject Sample # 1	0.200	13:06
Breath Volume	2.136	13:06
Air Blank	0.000	13:06
Air Blank	0.000	13:08
Subject Sample # 2	0.199	13:08
Breath Volume	1.621	13:08
Air Blank	0.000	13:09
Control Test	0.080	13:09
Air Blank	0.000	13:10
Diagnostics Check	OK	13:10

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ALVA

DATE/TIME OF TEST: 04/15/2016 04:22:58

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.082	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.079	04:18
Breath Volume	2.371	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:20
Subject Sample # 2	0.081	04:21
Breath Volume	2.199	04:21
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIBLEY, TRAVIS, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SIBLEY, TRAVIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: GEHRINGER

DATE/TIME OF TEST: 04/15/2016 18:27:50

Diagnostics Check	OK	18:20
Air Blank	0.000	18:21
Control Test	0.081	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.096	18:22
Breath Volume	1.781	18:22
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.094	18:26
Breath Volume	1.757	18:26
Air Blank	0.000	18:26
Control Test	0.080	18:27
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM, TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: GASTON

DATE/TIME OF TEST: 04/16/2016 01:28:33

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.080	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 04/23/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARR,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: HAGER

DATE/TIME OF TEST: 04/16/2016 02:32:10

Diagnostics Check	OK	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:26
Air Blank	0.000	02:26
Subject Sample # 1	0.009	02:27
Breath Volume	1.398	02:27
Air Blank	0.000	02:27
Air Blank	0.000	02:29
Subject Sample # 2	0.009	02:30
Breath Volume	2.062	02:30
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 04/16/2016 04:08:32

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.081	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.137	04:04
Breath Volume	1.585	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.137	04:07
Breath Volume	1.230	04:07
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANIAGUA,SGT. M,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PANIAGUA,SGT. M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: FERUS

DATE/TIME OF TEST: 04/16/2016 05:11:44

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.207	05:07
Breath Volume	2.753	05:07
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.219	05:10
Breath Volume	2.066	05:10
Air Blank	0.000	05:10
Control Test	0.080	05:11
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: WESCOM

DATE/TIME OF TEST: 04/16/2016 21:01:35

Diagnostics Check	OK	20:55
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:56
Subject Sample # 1	0.000	20:57
Breath Volume	2.191	20:57
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.000	21:00
Breath Volume	1.632	21:00
Air Blank	0.000	21:00
Control Test	0.081	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIBSON,JOHN,A
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: GIBSON,JOHN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: FEOLA

DATE/TIME OF TEST: 04/17/2016 05:13:01

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:10	
Control Test	0.081	05:10	
Air Blank	0.000	05:10	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Control Test	0.081	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SAPINO,MARK
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: HOUGLAND

DATE/TIME OF TEST: 04/17/2016 20:31:14

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:28	
Control Test	0.082	20:29	
Air Blank	0.000	20:29	
Subject Sample	# 1	20:29	Subject Test Refused
Breath Volume	0.000	20:29	
Air Blank	0.000	20:30	
Control Test	0.081	20:30	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/18/2016 23:28:45

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.081	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:27	Subject Test Refused
Breath Volume	0.000	23:27	
Air Blank	0.000	23:27	
Control Test	0.080	23:28	
Air Blank	0.000	23:28	
Diagnostics Check	OK	23:28	

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: KERN

DATE/TIME OF TEST: 04/21/2016 00:29:59

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.081	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.108	00:25
Breath Volume	2.335	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:27
Subject Sample # 2	0.104	00:28
Breath Volume	2.019	00:28
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 04/21/2016 03:29:45

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.186	03:20
Breath Volume	1.335	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.211	03:25
Breath Volume	1.898	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 3	0.227	03:28
Breath Volume	1.445	03:28
Air Blank	0.000	03:28
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: DUNBAR

DATE/TIME OF TEST: 04/21/2016 05:14:36

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.081	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.249	05:10
Breath Volume		2.273	05:10
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.245	05:13
Breath Volume		2.152	05:13
Air Blank		0.000	05:13
Control Test		0.080	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARD,ALEXANDRA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:11

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 04/21/2016 16:34:36

Diagnostics Check	OK	16:31	
Air Blank	0.000	16:32	
Control Test	0.080	16:32	
Air Blank	0.000	16:32	
Subject Sample	# 1	16:33	Subject Test Refused
Breath Volume	0.000	16:33	
Air Blank	0.000	16:33	
Control Test	0.080	16:34	
Air Blank	0.000	16:34	
Diagnostics Check	OK	16:34	

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 04/22/2016 03:56:56

Diagnostics Check	OK	03:54	
Air Blank	0.000	03:54	
Control Test	0.080	03:55	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:56	Sequence Aborted
Breath Volume	0.000	03:56	
Air Blank	0.000	03:56	

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 04/22/2016 04:11:45

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.071	04:07
Breath Volume	1.371	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.070	04:10
Breath Volume	1.308	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: METZLER

DATE/TIME OF TEST: 04/23/2016 03:46:56

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:44	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:46	
Control Test	0.081	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 04/23/2016 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLANCO,KEDWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POLANCO,KEDWIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 04/24/2016 00:21:03

Diagnostics Check	OK	00:17	
Air Blank	0.000	00:18	
Control Test	0.081	00:18	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:20	
Control Test	0.080	00:20	
Air Blank	0.000	00:20	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STILLWELL,KRISTEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: STILLWELL,KRISTEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 04/25/2016 19:37:12

Diagnostics Check	OK	19:30
Air Blank	0.000	19:30
Control Test	0.081	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.181	19:32
Breath Volume	1.871	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:35
Subject Sample # 2	0.180	19:35
Breath Volume	1.320	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMAR,GIOVANNI,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMAR,GIOVANNI
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 04/26/2016 00:34:33

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:32	
Control Test	0.080	00:32	
Air Blank	0.000	00:33	
Subject Sample	# 1	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:33	
Control Test	0.080	00:33	
Air Blank	0.000	00:34	
Diagnostics Check	OK	00:34	

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 17:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: ABREU

DATE/TIME OF TEST: 04/27/2016 18:07:28

Diagnostics Check	OK	18:01
Air Blank	0.000	18:01
Control Test	0.081	18:02
Air Blank	0.000	18:02
Subject Sample # 1	0.000	18:03
Breath Volume	2.457	18:03
Air Blank	0.000	18:03
Air Blank	0.000	18:05
Subject Sample # 2	0.000	18:06
Breath Volume	1.953	18:06
Air Blank	0.000	18:06
Control Test	0.080	18:06
Air Blank	0.000	18:07
Diagnostics Check	OK	18:07

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GROWE

DATE/TIME OF TEST: 04/28/2016 03:07:48

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.080	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.163	03:03
Breath Volume	1.121	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.165	03:06
Breath Volume	1.238	03:06
Air Blank	0.000	03:06
Control Test	0.081	03:07
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: MCDEVITT

DATE/TIME OF TEST: 04/28/2016 18:25:25

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:22	
Control Test	0.080	18:23	
Air Blank	0.000	18:23	
Subject Sample	# 1	18:24	Subject Test Refused
Breath Volume	0.000	18:24	
Air Blank	0.000	18:24	
Control Test	0.080	18:24	
Air Blank	0.000	18:25	
Diagnostics Check	OK	18:25	

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: RECTOR

DATE/TIME OF TEST: 04/29/2016 02:37:05

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.081	02:35	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FERGUSON,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 04/29/2016 04:16:42

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.080	04:16	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GONZALES JR

DATE/TIME OF TEST: 04/30/2016 01:04:00

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.081	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.065	00:59
Breath Volume	2.722	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.069	01:02
Breath Volume	2.812	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:04

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/30/2016 02:05:50

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.080	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 04/30/2016 04:18:32

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:16	
Control Test	0.081	04:16	
Air Blank	0.000	04:16	
Subject Sample	# 1	04:17	Subject Test Refused
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.081	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 05/25/2016 22:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31113080A1
DRY GAS EXPIRY DATE: 12/01/2015
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: MITRI

DATE/TIME OF TEST: 04/30/2016 19:55:15

Diagnostics Check	OK	19:51	
Air Blank	0.000	19:52	
Control Test	0.078	19:52	
Air Blank	0.000	19:53	
Subject Sample	# 1	19:53	Subject Test Refused
Breath Volume	0.000	19:53	
Air Blank	0.000	19:54	
Control Test	0.079	19:54	
Air Blank	0.000	19:55	
Diagnostics Check	OK	19:55	

Data Download Date/Time: 05/05/2016 01:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SHOWLEY

DATE/TIME OF TEST: 04/29/2016 02:45:28

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.174	02:40
Breath Volume	1.511	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.172	02:43
Breath Volume	1.441	02:43
Air Blank	0.000	02:44
Control Test	0.077	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 05/18/2016 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: MERCADO DIAZ

DATE/TIME OF TEST: 04/30/2016 21:26:50

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.079	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.038	21:22
Breath Volume	2.832	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.038	21:25
Breath Volume	2.816	21:25
Air Blank	0.000	21:25
Control Test	0.078	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 05/18/2016 20:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: MOBLEY

DATE/TIME OF TEST: 04/01/2016 23:54:17

Diagnostics Check	OK	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:48
Subject Sample # 1	0.096	23:49
Breath Volume	2.257	23:49
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.099	23:52
Breath Volume	2.445	23:52
Air Blank	0.000	23:53
Control Test	0.079	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 04/17/2016 01:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES, MICHAEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES, MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/03/2016 21:11:21

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.080	21:05
Air Blank	0.000	21:05
Subject Sample # 1	0.044	21:06
Breath Volume	1.699	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.043	21:09
Breath Volume	2.050	21:09
Air Blank	0.000	21:10
Control Test	0.078	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 04/17/2016 01:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MENDAK,MICHAEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: LAFLEUR

DATE/TIME OF TEST: 04/08/2016 08:13:15

Diagnostics Check	OK	08:10	
Air Blank	0.000	08:10	
Control Test	0.080	08:11	
Air Blank	0.000	08:11	
Subject Sample	# 1	08:11	Subject Test Refused
Breath Volume	0.000	08:11	
Air Blank	0.000	08:12	
Control Test	0.079	08:12	
Air Blank	0.000	08:13	
Diagnostics Check	OK	08:13	

Data Download Date/Time: 04/17/2016 01:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ALEMAN

DATE/TIME OF TEST: 04/14/2016 05:13:15

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.081	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.000	05:08
Breath Volume	1.656	05:08
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.000	05:11
Breath Volume	1.664	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 04/17/2016 01:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 04/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: PERAZA REINOSO

DATE/TIME OF TEST: 04/30/2016 21:14:25

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Subject Sample # 1	0.173	21:09
Breath Volume	1.593	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.175	21:12
Breath Volume	1.109	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 05/24/2016 09:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ, MICHEAL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 04/16/2016 00:16:49

Diagnostics Check	OK	00:10
Air Blank	0.000	00:10
Control Test	0.078	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.000	00:12
Breath Volume	1.128	00:12
Air Blank	0.000	00:12
Air Blank	0.000	00:14
Subject Sample # 2	0.000	00:15
Breath Volume	1.234	00:15
Air Blank	0.000	00:15
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: RUSSO

DATE/TIME OF TEST: 04/16/2016 00:32:59

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:30	
Control Test	0.080	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VAN GILS, PATRICK
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LOPEZ MONTERO

DATE/TIME OF TEST: 04/16/2016 03:59:16

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.078	03:57	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.078	03:58	
Air Blank	0.000	03:59	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TORRES,DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VAUGHEN

DATE/TIME OF TEST: 04/16/2016 04:35:46

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.163	04:30
Breath Volume	1.656	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:32
Subject Sample # 2	0.165	04:34
Breath Volume	1.617	04:34
Air Blank	0.000	04:34
Control Test	0.078	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO, GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/17/2016 00:51:53

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.079	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.181	00:42
Breath Volume	1.292	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:44
Subject Sample # 2	0.202	00:46
Breath Volume	2.433	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 3	0.188	00:50
Breath Volume	1.195	00:50
Air Blank	0.000	00:50
Control Test	0.077	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH, FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: MARTINS RODRIGUES

DATE/TIME OF TEST: 04/17/2016 02:44:11

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:39
Breath Volume	1.875	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.000	02:42
Breath Volume	1.542	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: RICARDO VILLALOBOS

DATE/TIME OF TEST: 04/23/2016 23:47:40

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.194	23:43
Breath Volume	1.777	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.196	23:46
Breath Volume	1.351	23:46
Air Blank	0.000	23:46
Control Test	0.077	23:47
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ, MICHEAL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: RODRIGUEZ PEREZ

DATE/TIME OF TEST: 04/24/2016 02:07:53

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.079	02:05	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.080	02:07	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 05/06/2016 09:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/01/2016 21:37:18

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.078	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.000	21:33
Breath Volume	1.792	21:33
Air Blank	0.000	21:33
Air Blank	0.000	21:35
Subject Sample # 2	0.000	21:35
Breath Volume	2.187	21:35
Air Blank	0.000	21:36
Control Test	0.078	21:36
Air Blank	0.000	21:37
Diagnostics Check	OK	21:37

Data Download Date/Time: 04/06/2016 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 04/02/2016 02:58:31

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.078	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.154	02:54
Breath Volume	2.312	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.145	02:56
Breath Volume	1.121	02:56
Air Blank	0.000	02:57
Control Test	0.076	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 04/06/2016 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 04/02/2016 04:31:45

Diagnostics Check	OK	04:25
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:26
Subject Sample # 1	0.194	04:27
Breath Volume	2.699	04:27
Air Blank	0.000	04:27
Air Blank	0.000	04:29
Subject Sample # 2	0.189	04:30
Breath Volume	2.011	04:30
Air Blank	0.000	04:30
Control Test	0.078	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 04/06/2016 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: BARRETO MARRERO

DATE/TIME OF TEST: 04/02/2016 06:40:24

Diagnostics Check	OK	06:31
Air Blank	0.000	06:32
Control Test	0.078	06:32
Air Blank	0.000	06:33
Subject Sample # 1	0.170	06:35
Breath Volume	1.660	06:35
Air Blank	0.000	06:36
Air Blank	0.000	06:38
Subject Sample # 2	0.171	06:38
Breath Volume	2.207	06:38
Air Blank	0.000	06:39
Control Test	0.077	06:39
Air Blank	0.000	06:40
Diagnostics Check	OK	06:40

Data Download Date/Time: 04/06/2016 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: PERDOMO BACA

DATE/TIME OF TEST: 04/06/2016 05:43:10

Diagnostics Check		OK	05:34
Air Blank		0.000	05:35
Control Test		0.079	05:35
Air Blank		0.000	05:35
Subject Sample	# 1	0.116	05:38
Breath Volume		2.195	05:38
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.122	05:41
Breath Volume		1.632	05:41
Air Blank		0.000	05:42
Control Test		0.077	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 04/06/2016 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DEMATEO

DATE/TIME OF TEST: 04/07/2016 00:42:06

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.078	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.160	00:37
Breath Volume	2.273	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.158	00:40
Breath Volume	1.675	00:40
Air Blank	0.000	00:41
Control Test	0.078	00:41
Air Blank	0.000	00:42
Diagnostics Check	OK	00:42

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: UNTORIA GONZALEZ

DATE/TIME OF TEST: 04/07/2016 01:11:13

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.078	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.128	01:05
Breath Volume	1.378	01:05
Air Blank	0.000	01:06
Air Blank	0.000	01:07
Subject Sample # 2	0.128	01:09
Breath Volume	1.414	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/07/2016 04:00:31

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.077	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.134	03:55
Breath Volume	1.890	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.141	03:58
Breath Volume	2.261	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: LINARES

DATE/TIME OF TEST: 04/07/2016 06:29:48

Diagnostics Check	OK	06:23
Air Blank	0.000	06:24
Control Test	0.078	06:24
Air Blank	0.000	06:24
Subject Sample # 1	0.174	06:25
Breath Volume	2.539	06:25
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.175	06:28
Breath Volume	1.984	06:28
Air Blank	0.000	06:28
Control Test	0.078	06:29
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: NATIVIDAD-ALVAREZ

DATE/TIME OF TEST: 04/13/2016 03:29:03

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.078	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.189	03:24
Breath Volume	2.246	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.182	03:27
Breath Volume	1.992	03:27
Air Blank	0.000	03:28
Control Test	0.077	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: FELIZOLA II

DATE/TIME OF TEST: 04/14/2016 01:14:34

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.077	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.078	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TABILO

DATE/TIME OF TEST: 04/14/2016 03:36:09

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.000	03:31
Breath Volume	1.511	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.000	03:34
Breath Volume	1.847	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,DAEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ORDAZ VAZQUEZ

DATE/TIME OF TEST: 04/15/2016 01:36:38

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.077	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.189	01:31
Breath Volume	1.175	01:31
Air Blank	0.000	01:32
Air Blank	0.000	01:34
Subject Sample # 2	0.198	01:34
Breath Volume	2.085	01:34
Air Blank	0.000	01:35
Control Test	0.077	01:35
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/15/2016 02:35:55

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:33	
Control Test	0.077	02:33	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:34	Subject Test Refused
Breath Volume	0.000	02:34	
Air Blank	0.000	02:34	
Control Test	0.077	02:35	
Air Blank	0.000	02:35	
Diagnostics Check	OK	02:35	

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: HECHAVARRIA

DATE/TIME OF TEST: 04/17/2016 03:10:38

Diagnostics Check	OK	03:04
Air Blank	0.000	03:05
Control Test	0.079	03:05
Air Blank	0.000	03:06
Subject Sample # 1	0.000	03:06
Breath Volume	2.765	03:06
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.000	03:09
Breath Volume	3.011	03:09
Air Blank	0.000	03:09
Control Test	0.077	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ZORBA

DATE/TIME OF TEST: 04/17/2016 03:19:40

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.077	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.135	03:15
Breath Volume	2.285	03:15
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.137	03:18
Breath Volume	1.632	03:18
Air Blank	0.000	03:18
Control Test	0.077	03:19
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/20/2016 03:44:07

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.196	03:39
Breath Volume	1.992	03:39
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.177	03:42
Breath Volume	1.679	03:42
Air Blank	0.000	03:43
Control Test	0.078	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: AMBROSIO JR

DATE/TIME OF TEST: 04/22/2016 04:07:26

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:04	
Control Test	0.078	04:05	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:06	Subject Test Refused
Breath Volume	0.000	04:06	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:07	
Diagnostics Check	OK	04:07	

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DE POOL

DATE/TIME OF TEST: 04/23/2016 01:03:04

Diagnostics Check	OK	00:56
Air Blank	0.000	00:57
Control Test	0.077	00:57
Air Blank	0.000	00:58
Subject Sample # 1	0.190	00:58
Breath Volume	1.675	00:58
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.196	01:01
Breath Volume	1.933	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:03

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 04/23/2016 02:42:03

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.077	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.000	02:37
Breath Volume	2.667	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.000	02:40
Breath Volume	2.937	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:42

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 04/24/2016 00:39:08

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.078	00:33
Air Blank	0.000	00:34
Subject Sample # 1	0.126	00:34
Breath Volume	4.347	00:34
Air Blank	0.000	00:35
Air Blank	0.000	00:37
Subject Sample # 2	0.120	00:37
Breath Volume	3.949	00:37
Air Blank	0.000	00:38
Control Test	0.077	00:38
Air Blank	0.000	00:39
Diagnostics Check	OK	00:39

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: VALLADARES

DATE/TIME OF TEST: 04/24/2016 02:25:21

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.078	02:23	
Air Blank	0.000	02:23	
Subject Sample	# 1	02:23	Subject Test Refused
Breath Volume	0.000	02:23	
Air Blank	0.000	02:24	
Control Test	0.078	02:24	
Air Blank	0.000	02:25	
Diagnostics Check	OK	02:25	

Data Download Date/Time: 05/06/2016 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RODRIGUEZ-DIAZ

DATE/TIME OF TEST: 04/28/2016 01:18:54

Diagnostics Check	OK	01:10
Air Blank	0.000	01:11
Control Test	0.077	01:11
Air Blank	0.000	01:12
Subject Sample # 1	0.179	01:13
Breath Volume	2.546	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.173	01:17
Breath Volume	1.886	01:17
Air Blank	0.000	01:17
Control Test	0.077	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 05/06/2016 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/28/2016 23:43:41

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.078	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.000	23:39
Breath Volume	2.109	23:39
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.000	23:42
Breath Volume	3.324	23:42
Air Blank	0.000	23:42
Control Test	0.080	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 05/06/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LOSADA-PEREZ

DATE/TIME OF TEST: 04/29/2016 03:22:42

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.077	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.223	03:15
Breath Volume		1.929	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.198	03:18
Breath Volume		1.769	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 3	0.209	03:21
Breath Volume		3.390	03:21
Air Blank		0.000	03:21
Control Test		0.077	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 05/06/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: ORTEGA ROJAS

DATE/TIME OF TEST: 04/29/2016 04:34:48

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.077	04:28
Air Blank	0.000	04:29
Subject Sample # 1	0.088	04:30
Breath Volume	2.722	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.091	04:33
Breath Volume	3.375	04:33
Air Blank	0.000	04:33
Control Test	0.078	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 05/06/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 04/30/2016 03:31:59

Diagnostics Check	OK	03:29	
Air Blank	0.000	03:29	
Control Test	0.078	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:31	
Control Test	0.077	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 05/06/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CHANG,BRADLEY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LEON MONTALVO

DATE/TIME OF TEST: 04/30/2016 04:03:18

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.077	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.077	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 05/06/2016 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/01/2016 04:03:18

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.145	03:58
Breath Volume	3.050	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.147	04:01
Breath Volume	2.785	04:01
Air Blank	0.000	04:02
Control Test	0.078	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 04/06/2016 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/02/2016 04:18:14

Diagnostics Check	OK	04:11
Air Blank	0.000	04:11
Control Test	0.078	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.141	04:13
Breath Volume	2.593	04:13
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.138	04:16
Breath Volume	2.289	04:16
Air Blank	0.000	04:17
Control Test	0.078	04:17
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 04/06/2016 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: ROSENTHAL

DATE/TIME OF TEST: 04/06/2016 03:36:14

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:33	
Control Test	0.079	03:34	
Air Blank	0.000	03:34	
Subject Sample	# 1	03:34	Subject Test Refused
Breath Volume	0.000	03:34	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:36	
Diagnostics Check	OK	03:36	

Data Download Date/Time: 04/06/2016 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: MENDOZA-PERALTA

DATE/TIME OF TEST: 04/08/2016 05:36:49

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.079	05:30
Air Blank	0.000	05:31
Subject Sample # 1	0.176	05:31
Breath Volume	1.496	05:31
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.183	05:35
Breath Volume	1.867	05:35
Air Blank	0.000	05:35
Control Test	0.078	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GODOY

DATE/TIME OF TEST: 04/09/2016 02:24:37

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:17
Air Blank	0.000	02:17
Subject Sample # 1	0.125	02:18
Breath Volume	1.945	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:20
Subject Sample # 2	0.127	02:22
Breath Volume	2.183	02:22
Air Blank	0.000	02:23
Control Test	0.079	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: SALGUERA

DATE/TIME OF TEST: 04/14/2016 07:17:51

Diagnostics Check	OK	07:11
Air Blank	0.000	07:11
Control Test	0.079	07:11
Air Blank	0.000	07:12
Subject Sample # 1	0.203	07:13
Breath Volume	2.734	07:13
Air Blank	0.000	07:13
Air Blank	0.000	07:15
Subject Sample # 2	0.198	07:16
Breath Volume	2.289	07:16
Air Blank	0.000	07:16
Control Test	0.077	07:17
Air Blank	0.000	07:17
Diagnostics Check	OK	07:17

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BABUN CHAVARRIA

DATE/TIME OF TEST: 04/15/2016 02:57:00

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.092	02:51
Breath Volume	1.285	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.091	02:55
Breath Volume	1.308	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:57

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/21/2016 01:01:15

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.014	00:57
Breath Volume	2.714	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.011	00:59
Breath Volume	1.347	00:59
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: GUILFORD-LANG

DATE/TIME OF TEST: 04/22/2016 03:14:29

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.117	03:09
Breath Volume		3.156	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.118	03:12
Breath Volume		2.875	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: OHAYON

DATE/TIME OF TEST: 04/23/2016 04:00:59

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.092	03:56
Breath Volume	1.789	03:56
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.101	03:59
Breath Volume	1.859	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: QUINCOSES

DATE/TIME OF TEST: 04/24/2016 00:08:07

Diagnostics Check	OK	23:53	
Air Blank	0.000	23:54	
Control Test	0.079	23:54	
Air Blank	0.000	23:55	
Subject Sample	# 3	23:58	No Sample Provided
Breath Volume	0.000	23:58	
Air Blank	0.000	23:58	
Air Blank	0.000	00:00	
Subject Sample	# 1	0.118	00:02
Breath Volume	1.265	00:02	
Air Blank	0.000	00:03	
Air Blank	0.000	00:05	
Subject Sample	# 2	0.118	00:06
Breath Volume	1.843	00:06	
Air Blank	0.000	00:07	
Control Test	0.077	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MARTINEZ JUNCO

DATE/TIME OF TEST: 04/27/2016 04:58:05

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.184	04:53
Breath Volume	2.734	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.180	04:56
Breath Volume	2.277	04:56
Air Blank	0.000	04:57
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:58

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 04/28/2016 03:22:21

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.115	03:16
Breath Volume	1.167	03:16
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.121	03:20
Breath Volume	1.757	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CARDENAS

DATE/TIME OF TEST: 04/29/2016 00:54:59

Diagnostics Check	OK	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.082	00:50
Breath Volume	2.156	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.078	00:53
Breath Volume	1.406	00:53
Air Blank	0.000	00:54
Control Test	0.078	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GUGLIOTTA

DATE/TIME OF TEST: 04/30/2016 00:31:33

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:25
Subject Sample # 1	0.184	00:26
Breath Volume	1.195	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:28
Subject Sample # 2	0.185	00:29
Breath Volume	2.003	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:30
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 04/30/2016 04:27:21

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.078	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.200	04:22
Breath Volume		2.746	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.198	04:25
Breath Volume		2.371	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 05/05/2016 13:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER, MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RAMOS DELGADO

DATE/TIME OF TEST: 04/06/2016 00:49:57

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.000	00:39	Volume Not Met
Breath Volume		0.183	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2	0.000	00:44	
Breath Volume		2.308	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:47	
Subject Sample	# 3	0.000	00:48	
Breath Volume		1.757	00:48	
Air Blank		0.000	00:49	
Control Test		0.078	00:49	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 04/26/2016 12:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER, MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CATTELL

DATE/TIME OF TEST: 04/09/2016 02:19:08

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.078	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.148	02:14
Breath Volume		1.652	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.153	02:17
Breath Volume		1.421	02:17
Air Blank		0.000	02:18
Control Test		0.077	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 04/26/2016 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER, MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 04/09/2016 03:04:21

Diagnostics Check	OK	02:57
Air Blank	0.000	02:57
Control Test	0.078	02:58
Air Blank	0.000	02:58
Subject Sample # 1	0.127	02:59
Breath Volume	2.503	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.126	03:02
Breath Volume	2.207	03:02
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 04/26/2016 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BING

DATE/TIME OF TEST: 04/11/2016 22:34:05

Diagnostics Check	OK	22:27
Air Blank	0.000	22:28
Control Test	0.080	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.166	22:29
Breath Volume	2.109	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.170	22:32
Breath Volume	1.464	22:32
Air Blank	0.000	22:33
Control Test	0.079	22:33
Air Blank	0.000	22:33
Diagnostics Check	OK	22:34

Data Download Date/Time: 04/26/2016 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: KRAFFERT

DATE/TIME OF TEST: 04/12/2016 22:01:37

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:56	
Control Test	0.078	21:56	
Air Blank	0.000	21:56	
Subject Sample	# 1 0.066	21:57	
Breath Volume	1.597	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample	# 2	22:00	RFI Detect
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.078	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 04/26/2016 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SUTTLE,CORY,A

BREATH TEST OPERATOR AGENCY: Lakeland PD

ARRESTING OFFICER: SUTTLE,CORY

ARRESTING OFFICER AGENCY: Lakeland PD

LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22315080A2

DRY GAS EXPIRY DATE: 09/05/2017

VIOLATION CODE: DUI

ARREST TIME: 21:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: KRAFFERT

DATE/TIME OF TEST: 04/12/2016 22:10:10

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.080	22:03
Air Blank	0.000	22:04
Subject Sample # 1	0.073	22:05
Breath Volume	1.171	22:05
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.064	22:08
Breath Volume	1.710	22:08
Air Blank	0.000	22:09
Control Test	0.078	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 04/26/2016 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MELANIE,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SULLIVAN,MELANIE
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 04/13/2016 20:28:29

Diagnostics Check	OK	20:22
Air Blank	0.000	20:22
Control Test	0.079	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.000	20:24
Breath Volume	1.839	20:24
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.000	20:27
Breath Volume	1.199	20:27
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 04/26/2016 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: AGUILERA BLANCHARD

DATE/TIME OF TEST: 04/16/2016 19:32:41

Diagnostics Check	OK	19:25
Air Blank	0.000	19:25
Control Test	0.079	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.164	19:27
Breath Volume	1.699	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.164	19:31
Breath Volume	1.554	19:31
Air Blank	0.000	19:31
Control Test	0.078	19:32
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 04/26/2016 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LOZADA ROBLES

DATE/TIME OF TEST: 04/10/2016 00:17:05

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.083	00:14	
Air Blank	0.000	00:14	
Subject Sample # 1	0.000	00:15	
Breath Volume	2.507	00:15	
Air Blank		00:15	RFI Detect
Air Blank	0.000	00:16	
Control Test	0.083	00:16	
Air Blank	0.000	00:16	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 04/10/2016 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LOZADA ROBLES

DATE/TIME OF TEST: 04/10/2016 00:18:51

Diagnostics Check	OK	00:18	
Air Blank		00:18	RFI Detect
Air Blank		00:18	RFI Detect

Data Download Date/Time: 04/10/2016 12:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LOZADA ROBLES

DATE/TIME OF TEST: 04/10/2016 00:26:29

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.083	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.000	00:22
Breath Volume	2.011	00:22
Air Blank	0.000	00:22
Air Blank	0.000	00:24
Subject Sample # 2	0.000	00:25
Breath Volume	1.933	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Diagnostics Check	OK	00:26

Data Download Date/Time: 04/10/2016 12:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUF AHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: ONEAL

DATE/TIME OF TEST: 04/08/2016 22:44:42

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:39
Air Blank	0.000	22:39
Subject Sample # 1	0.173	22:40
Breath Volume	1.324	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.165	22:43
Breath Volume	1.117	22:43
Air Blank	0.000	22:43
Control Test	0.078	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 04/28/2016 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/22/2016 03:27:22

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.032	03:22
Breath Volume	2.644	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.029	03:25
Breath Volume	1.101	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 04/28/2016 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON, THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 04/01/2016 02:14:42

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.146	02:09
Breath Volume	2.750	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.139	02:12
Breath Volume	3.093	02:12
Air Blank	0.000	02:13
Control Test	0.081	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 04/15/2016 14:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON, THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MIRALDA

DATE/TIME OF TEST: 04/06/2016 01:13:09

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.082	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.167	01:08
Breath Volume	3.605	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.171	01:11
Breath Volume	2.941	01:11
Air Blank	0.000	01:12
Control Test	0.083	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 04/15/2016 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON, THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: VES

DATE/TIME OF TEST: 04/07/2016 03:44:15

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:39	
Control Test	0.083	03:39	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:42	Subject Test Refused
Breath Volume	0.000	03:42	
Air Blank	0.000	03:43	
Control Test	0.083	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 04/15/2016 14:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON, THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: RODRIGUEZ-GOMEZ

DATE/TIME OF TEST: 04/17/2016 23:14:07

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.082	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.225	23:09
Breath Volume	2.468	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.228	23:12
Breath Volume	2.421	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 05/20/2016 13:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON, THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON, THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/18/2016 22:02:50

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.082	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.234	21:58
Breath Volume	2.976	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:00
Subject Sample # 2	0.220	22:01
Breath Volume	2.511	22:01
Air Blank	0.000	22:01
Control Test	0.081	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 05/20/2016 13:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CARNENAS

DATE/TIME OF TEST: 04/09/2016 02:24:11

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.082	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.127	02:19
Breath Volume	2.539	02:19
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.124	02:22
Breath Volume	2.386	02:22
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 04/15/2016 11:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SNELGROVE,TROY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PETRIRENA

DATE/TIME OF TEST: 04/10/2016 01:23:31

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.081	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.133	01:18
Breath Volume	2.535	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.134	01:21
Breath Volume	2.531	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 04/15/2016 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: FELIPE

DATE/TIME OF TEST: 04/10/2016 02:44:40

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.172	02:40
Breath Volume	2.339	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.177	02:43
Breath Volume	1.820	02:43
Air Blank	0.000	02:43
Control Test	0.079	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 04/15/2016 11:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGIGHAN,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MCGIGHAN,WILLIAM
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/30/2016 01:59:17

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.077	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.167	01:54
Breath Volume	1.984	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.158	01:57
Breath Volume	2.480	01:57
Air Blank	0.000	01:58
Control Test	0.075	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 05/18/2016 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006169

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: RICE,MICHAEL
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/30/2016 02:17:10

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.077	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.176	02:12
Breath Volume	2.417	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.184	02:15
Breath Volume	1.609	02:15
Air Blank	0.000	02:16
Control Test	0.075	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 05/18/2016 14:27