Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: COLT

DATE/TIME OF TEST: 04/06/2016 04:19:22

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.182	04:14
Breath Volume		2.062	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:16
Subject Sample	# 2	0.185	04:17
Breath Volume		3.167	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 04/27/2016 01:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PLUSSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: BUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: TURNBULL

DATE/TIME OF TEST: 04/06/2016 22:20:19

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.080	22:18	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:19	
Control Test		0.080	22:19	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 04/27/2016 01:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: PUSHCAR

DATE/TIME OF TEST: 04/08/2016 22:54:58

Diagnostics Check		OK	22:47
Air Blank		0.000	22:48
Control Test		0.080	22:48
Air Blank		0.000	22:48
Subject Sample	# 1	0.199	22:50
Breath Volume		1.730	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.208	22:53
Breath Volume		1.273	22:53
Air Blank		0.000	22:54
Control Test		0.075	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 04/27/2016 01:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/10/2016 02:31:05

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.143	02:24
Breath Volume		1.519	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.132	02:29
Breath Volume		1.808	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 04/27/2016 01:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BUXTON,JAMIE
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: DOUGHERTY

DATE/TIME OF TEST: 04/10/2016 23:08:31

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:06	
Control Test		0.081	23:06	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:07	
Control Test		0.081	23:07	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:08	

Data Download Date/Time: 04/27/2016 01:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: OERMAN

DATE/TIME OF TEST: 04/13/2016 02:31:24

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Control Test		0.079	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 04/27/2016 01:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY,SPENCER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,AUSTIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: GODSEY

DATE/TIME OF TEST: 04/05/2016 22:02:19

Diagnostics Check		OK	21:59	
Air Blank		0.000	21:59	
Control Test		0.080	21:59	
Air Blank		0.000	22:00	
Subject Sample	# 1		22:00	Subject Test Refused
Breath Volume		0.000	22:00	
Air Blank		0.000	22:01	
Control Test		0.080	22:01	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 04/26/2016 08:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN,EVAN,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CALHOUN,EVAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 04/07/2016 02:09:36

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.080	02:09	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: RUPP

DATE/TIME OF TEST: 04/08/2016 21:29:55

Diagnostics Check		OK	21:26	
Air Blank		0.000	21:26	
Control Test		0.080	21:27	
Air Blank		0.000	21:27	
Subject Sample	# 1		21:28	Subject Test Refused
Breath Volume		0.000	21:28	
Air Blank		0.000	21:29	
Control Test		0.080	21:29	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 04/26/2016 08:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETTEE,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CLARK,THOMAS
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:26

SUBJECT'S LAST NAME: MARTINEZ CAMPOS

DATE/TIME OF TEST: 04/10/2016 11:06:09

Diagnostics Check		OK	10:59
Air Blank		0.000	10:59
Control Test		0.081	10:59
Air Blank		0.000	11:00
Subject Sample	# 1	0.061	11:01
Breath Volume		1.410	11:01
Air Blank		0.000	11:01
Air Blank		0.000	11:03
Subject Sample	# 2	0.060	11:04
Breath Volume		1.574	11:04
Air Blank		0.000	11:05
Control Test		0.079	11:05
Air Blank		0.000	11:06
Diagnostics Check		OK	11:06

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BURGHARDT

DATE/TIME OF TEST: 04/10/2016 23:37:30

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.081	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.059	23:33
Breath Volume		1.445	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.059	23:35
Breath Volume		3.832	23:35
Air Blank		0.000	23:36
Control Test		0.078	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: GARRISON,ADAM
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 04/11/2016 19:02:51

Diagnostics Check		OK	18:59	
Air Blank		0.000	18:59	
Control Test		0.080	19:00	
Air Blank		0.000	19:00	
Subject Sample	# 1		19:01	Subject Test Refused
Breath Volume		0.000	19:01	
Air Blank		0.000	19:01	
Control Test		0.079	19:02	
Air Blank		0.000	19:02	
Diagnostics Check		OK	19:02	

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CHAUVIN

DATE/TIME OF TEST: 04/15/2016 03:26:24

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.192	03:21
Breath Volume		3.859	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.190	03:24
Breath Volume		3.730	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGEE,JOHN,W
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MCGEE,JOHN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/15/2016 08:46:10

Diagnostics Check		OK	08:41	
Air Blank		0.000	08:42	
Control Test		0.080	08:42	
Air Blank		0.000	08:42	
Subject Sample	# 1		08:44	Subject Test Refused
Breath Volume		0.000	08:44	
Air Blank		0.000	08:45	
Control Test		0.080	08:45	
Air Blank		0.000	08:46	
Diagnostics Check		OK	08:46	

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 04/15/2016 22:39:36

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.145	22:35
Breath Volume		2.156	22:35
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.136	22:38
Breath Volume		1.921	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Diagnostics Check		OK	22:39

Data Download Date/Time: 04/26/2016 08:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: BARAHONA

DATE/TIME OF TEST: 04/17/2016 01:45:21

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.143	01:40
Breath Volume		2.085	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.141	01:43
Breath Volume		1.812	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/17/2016 03:59:43

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.296	03:55
Breath Volume		3.039	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.295	03:58
Breath Volume		3.140	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 04/17/2016 18:55:50

Diagnostics Check		OK	18:48
Air Blank		0.000	18:49
Control Test		0.080	18:49
Air Blank		0.000	18:49
Subject Sample	# 1	0.217	18:51
Breath Volume		1.652	18:51
Air Blank		0.000	18:51
Air Blank		0.000	18:53
Subject Sample	# 2	0.232	18:54
Breath Volume		1.644	18:54
Air Blank		0.000	18:54
Control Test		0.079	18:55
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 04/26/2016 08:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: PULLANO

DATE/TIME OF TEST: 04/19/2016 03:19:20

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.079	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:11	No Sample Provided
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 2	0.196	03:14	
Breath Volume		1.773	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 3	0.200	03:17	
Breath Volume		1.976	03:17	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY,SPENCER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,AUSTIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: BLANKENSHIP

DATE/TIME OF TEST: 04/19/2016 04:06:46

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:04	
Control Test		0.080	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Control Test		0.080	04:06	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 04/26/2016 08:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GABAY,JOHN,K
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GABAY,JOHN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: SNOPEK

DATE/TIME OF TEST: 04/19/2016 23:46:07

Diagnostics Check		OK	23:43	
Air Blank		0.000	23:43	
Control Test		0.081	23:43	
Air Blank		0.000	23:44	
Subject Sample	# 1		23:44	Subject Test Refused
Breath Volume		0.000	23:44	
Air Blank		0.000	23:45	
Control Test		0.080	23:45	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURRY,SPENCER,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: REILLY,TOM
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: KRABBE

DATE/TIME OF TEST: 04/20/2016 00:21:10

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.000	00:16
Breath Volume		2.187	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.000	00:19
Breath Volume		1.105	00:19
Air Blank		0.000	00:20
Control Test		0.080	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 04/26/2016 08:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 04/22/2016 03:35:02

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.138	03:30
Breath Volume		2.445	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.144	03:33
Breath Volume		1.437	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:35

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: SCHARF

DATE/TIME OF TEST: 04/23/2016 22:28:51

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:26	
Control Test		0.080	22:26	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:27	Subject Test Refused
Breath Volume		0.000	22:27	
Air Blank		0.000	22:27	
Control Test		0.080	22:28	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 04/26/2016 08:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,KEVIN,M
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: O'CONNELL,KEVIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 04/30/2016 14:26:52

Diagnostics Check		OK	14:24	
Air Blank		0.000	14:25	
Control Test		0.079	14:25	
Air Blank		0.000	14:26	
Subject Sample	# 1		14:26	RFI Detect
Breath Volume		0.000	14:26	
Air Blank			14:26	RFI Detect
Air Blank			14:26	RFI Detect

Data Download Date/Time: 05/04/2016 08:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,KEVIN,M
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: O'CONNELL,KEVIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 04/30/2016 14:33:19

Diagnostics Check		OK	14:29	
Air Blank		0.000	14:29	
Control Test		0.080	14:29	
Air Blank		0.000	14:30	
Subject Sample	# 1	0.046	14:32	
Breath Volume		1.914	14:32	
Air Blank			14:32	RFI Detect
Air Blank			14:32	RFI Detect
Air Blank		0.000	14:33	

Data Download Date/Time: 05/04/2016 08:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,KEVIN,M
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: O'CONNELL,KEVIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 04/30/2016 14:43:57

Diagnostics Check	OK	14:42	
Air Blank	0.000	14:42	
Control Test	0.078	14:43	
Air Blank		14:43	RFI Detect
Air Blank	0.000	14:43	

Data Download Date/Time: 05/04/2016 08:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,KEVIN,M
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: O'CONNELL,KEVIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 04/30/2016 14:50:49

Diagnostics Check		OK	14:45	
Air Blank		0.000	14:45	
Control Test		0.080	14:46	
Air Blank		0.000	14:46	
Subject Sample	# 1	0.042	14:47	
Breath Volume		1.945	14:47	
Air Blank		0.000	14:47	
Air Blank		0.000	14:49	
Subject Sample	# 2	0.040	14:50	
Breath Volume		2.121	14:50	
Air Blank			14:50	RFI Detect
Air Blank		0.000	14:50	

Data Download Date/Time: 05/04/2016 08:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JESSICA,Y
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: GUERRA,GABRIEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: LORA

DATE/TIME OF TEST: 04/02/2016 00:30:58

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:28	
Control Test		0.081	00:28	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 04/29/2016 11:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JESSICA,Y
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: FUENTES,ADALBERTO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:13

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 04/09/2016 05:40:57

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:38	
Control Test		0.080	05:38	
Air Blank		0.000	05:39	
Subject Sample	# 1		05:39	Subject Test Refused
Breath Volume		0.000	05:39	
Air Blank		0.000	05:40	
Control Test		0.080	05:40	
Air Blank		0.000	05:40	
Diagnostics Check		OK	05:40	

Data Download Date/Time: 04/29/2016 11:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HAMMANN,A
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 04/11/2016 00:34:49

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:32	
Control Test		0.080	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 04/29/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FELICIANO

DATE/TIME OF TEST: 04/12/2016 01:16:50

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.218	01:12
Breath Volume		2.308	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:14
Subject Sample	# 2	0.207	01:15
Breath Volume		2.710	01:15
Air Blank		0.000	01:15
Control Test		0.080	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 04/29/2016 11:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMANN,ANDRE,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: FUENTES,A
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ALBURY

DATE/TIME OF TEST: 04/12/2016 20:31:40

Diagnostics Check		OK	20:23
Air Blank		0.000	20:23
Control Test		0.079	20:24
Air Blank		0.000	20:24
Subject Sample	# 1	0.219	20:26
Breath Volume		1.578	20:26
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.217	20:30
Breath Volume		1.714	20:30
Air Blank		0.000	20:30
Control Test		0.079	20:31
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 04/29/2016 11:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JANVIER,FRITZ,G
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: ARTIME,JOEL
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 04/22/2016 22:21:37

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.077	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.197	22:16
Breath Volume		1.679	22:16
Air Blank		0.000	22:17
Air Blank		0.000	22:19
Subject Sample	# 2	0.193	22:20
Breath Volume		1.792	22:20
Air Blank		0.000	22:20
Control Test		0.076	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 05/18/2016 11:46

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MURGUIDO,ALEX
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 03/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/14/2016 16:40:49

Diagnostics Check		OK	16:34
Air Blank		0.000	16:34
Control Test		0.082	16:35
Air Blank		0.000	16:35
Subject Sample	# 1	0.148	16:36
Breath Volume		1.937	16:36
Air Blank		0.000	16:36
Air Blank		0.000	16:38
Subject Sample	# 2	0.145	16:39
Breath Volume		2.218	16:39
Air Blank		0.000	16:39
Control Test		0.082	16:40
Air Blank		0.000	16:40
Diagnostics Check		OK	16:40

Data Download Date/Time: 04/28/2016 00:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: REYES,GEORGE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FELISOLA PEREZ

DATE/TIME OF TEST: 04/28/2016 21:27:37

Diagnostics Check		OK	21:18	
Air Blank		0.000	21:19	
Control Test		0.081	21:19	
Air Blank		0.000	21:20	
Subject Sample	# 1		21:23	No Sample Provided
Breath Volume		0.000	21:23	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 2		21:26	Subject Test Refused
Breath Volume		0.000	21:26	
Air Blank		0.000	21:26	
Control Test		0.081	21:27	
Air Blank		0.000	21:27	
Diagnostics Check		OK	21:27	

Data Download Date/Time: 06/04/2016 10:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 04/29/2016 20:02:31

Diagnostics Check		OK	19:56
Air Blank		0.000	19:56
Control Test		0.078	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.186	19:57
Breath Volume		1.367	19:57
Air Blank		0.000	19:58
Air Blank		0.000	20:00
Subject Sample	# 2	0.197	20:00
Breath Volume		1.738	20:00
Air Blank		0.000	20:01
Control Test		0.077	20:01
Air Blank		0.000	20:02
Diagnostics Check		OK	20:02

Data Download Date/Time: 06/04/2016 10:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,ERIC,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FERNANDEZ,JAVIER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: AURELIEN

DATE/TIME OF TEST: 04/02/2016 21:58:56

Diagnostics Check		OK	21:56	
Air Blank		0.000	21:56	
Control Test		0.081	21:56	
Air Blank		0.000	21:57	
Subject Sample	# 1		21:57	Subject Test Refused
Breath Volume		0.000	21:57	
Air Blank		0.000	21:58	
Control Test		0.081	21:58	
Air Blank		0.000	21:58	
Diagnostics Check		OK	21:58	

Data Download Date/Time: 04/27/2016 13:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MOLINA,DANNY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RIBEIRO

DATE/TIME OF TEST: 04/05/2016 22:43:04

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.081	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.182	22:35
Breath Volume		3.847	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:37
Subject Sample	# 2	0.152	22:38
Breath Volume		1.726	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:40
Subject Sample	# 3	0.154	22:41
Breath Volume		1.656	22:41
Air Blank		0.000	22:42
Control Test		0.081	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:43

Data Download Date/Time: 04/27/2016 13:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER,STEVE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LACONCHA

DATE/TIME OF TEST: 04/07/2016 04:27:15

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.168	04:22
Breath Volume		3.070	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.169	04:25
Breath Volume		3.929	04:25
Air Blank		0.000	04:26
Control Test		0.081	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 04/27/2016 13:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER,STEVE
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HOBSON JR

DATE/TIME OF TEST: 04/08/2016 03:17:36

Diagnostics Check		OK	03:10
Air Blank		0.000	03:10
Control Test		0.081	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.155	03:13
Breath Volume		1.554	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.159	03:16
Breath Volume		3.000	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 04/27/2016 13:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RODRIGUEZ,AGUSTIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: MONTENEGRO

DATE/TIME OF TEST: 04/08/2016 04:55:16

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.082	04:53	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:53	Subject Test Refused
Breath Volume		0.000	04:53	
Air Blank		0.000	04:54	
Control Test		0.081	04:54	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 04/27/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,GIRANNY,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MEDINA,FULGENCIO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/14/2016 20:23:59

Diagnostics Check		OK	20:08	
Air Blank		0.000	20:08	
Control Test		0.079	20:08	
Air Blank		0.000	20:09	
Subject Sample	# 1	0.278	20:11	
Breath Volume		2.101	20:11	
Air Blank		0.000	20:11	
Air Blank		0.000	20:13	
Subject Sample	# 2	0.253	20:16	Volume Not Met
Breath Volume		0.031	20:16	
Air Blank		0.000	20:17	
Air Blank		0.000	20:19	
Subject Sample	# 3	0.210	20:22	Volume Not Met
Breath Volume		0.378	20:22	
Air Blank		0.000	20:23	
Control Test		0.079	20:23	
Air Blank		0.000	20:23	
Diagnostics Check		OK	20:23	

Data Download Date/Time: 04/27/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,GIRANNY,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: MEDINA,FULGENCIO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/14/2016 20:48:53

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:44	
Control Test		0.081	20:45	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:47	Subject Test Refused
Breath Volume		0.609	20:47	
Air Blank		0.000	20:47	
Control Test		0.079	20:48	
Air Blank		0.000	20:48	
Diagnostics Check		OK	20:48	

Data Download Date/Time: 04/27/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CALLAN,TONY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/15/2016 01:32:22

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.081	01:28	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:31	
Control Test		0.081	01:31	
Air Blank		0.000	01:32	
Diagnostics Check		OK	01:32	

Data Download Date/Time: 04/27/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,JOSE,D
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GARCIA,MICHAEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 08:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:15

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/17/2016 08:55:04

Diagnostics Check		OK	08:52	
Air Blank		0.000	08:52	
Control Test		0.081	08:52	
Air Blank		0.000	08:53	
Subject Sample	# 1		08:53	Subject Test Refused
Breath Volume		0.000	08:53	
Air Blank		0.000	08:54	
Control Test		0.081	08:54	
Air Blank		0.000	08:54	
Diagnostics Check		OK	08:55	

Data Download Date/Time: 04/27/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONFESSORE,STEVE,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GARCIA,ALFREDO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: SOLOMAN

DATE/TIME OF TEST: 04/22/2016 04:04:43

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:02	
Control Test		0.080	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:03	Subject Test Refused
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Control Test		0.081	04:04	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 04/27/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: CALLAN,TONY
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: CAMPO GIRALDO

DATE/TIME OF TEST: 04/30/2016 01:23:19

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.081	01:17
Air Blank		0.000	01:17
Subject Sample	# 1	0.143	01:18
Breath Volume		1.492	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.146	01:21
Breath Volume		1.960	01:21
Air Blank		0.000	01:22
Control Test		0.081	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 05/26/2016 10:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUFFNER,HAL,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUFFNER,HAL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: MILU

DATE/TIME OF TEST: 04/21/2016 15:49:51

Diagnostics Check		OK	15:44	
Air Blank		0.000	15:44	
Control Test		0.079	15:44	
Air Blank		0.000	15:45	
Subject Sample	# 1	0.339	15:45	
Breath Volume		2.089	15:45	
Air Blank		0.000	15:46	
Air Blank		0.000	15:48	
Subject Sample	# 2	0.329	15:48	
Breath Volume		1.910	15:48	
Air Blank		0.000	15:49	
			15:49	RFI Detect
Air Blank		0.000	15:49	

Data Download Date/Time: 05/24/2016 09:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 04/24/2016 01:35:03

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.152	01:30
Breath Volume		3.523	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.146	01:33
Breath Volume		1.914	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:35

Data Download Date/Time: 05/24/2016 09:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: PEYNADO

DATE/TIME OF TEST: 04/27/2016 01:56:00

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.124	01:51
Breath Volume		1.632	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.118	01:54
Breath Volume		1.644	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:56

Data Download Date/Time: 05/24/2016 09:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: GIBREAL

DATE/TIME OF TEST: 04/05/2016 19:22:50

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.078	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.056	19:18
Breath Volume		2.246	19:18
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.045	19:21
Breath Volume		1.199	19:21
Air Blank		0.000	19:21
Control Test		0.075	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/07/2016 10:35:40

Diagnostics Check		OK	10:20	
Air Blank		0.000	10:21	
Control Test		0.077	10:21	
Air Blank		0.000	10:22	
Subject Sample	# 1	0.182	10:25	Volume Not Met
Breath Volume		0.406	10:25	
Air Blank		0.000	10:26	
Air Blank		0.000	10:27	
Subject Sample	# 2	0.193	10:28	
Breath Volume		1.359	10:28	
Air Blank		0.000	10:29	
Air Blank		0.000	10:31	
Subject Sample	# 3		10:33	Subject Test Refused
Breath Volume		0.449	10:33	
Air Blank		0.000	10:34	
Control Test		0.073	10:35	Control Outside Tolerance
Air Blank		0.000	10:35	

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 04/07/2016 10:39:36

Diagnostics Check		OK	10:36	
Air Blank		0.000	10:37	
Control Test		0.075	10:37	
Air Blank		0.000	10:38	
Subject Sample	# 1		10:38	Subject Test Refused
Breath Volume		0.000	10:38	
Air Blank		0.000	10:38	
Control Test		0.076	10:38	
Air Blank		0.000	10:39	
Diagnostics Check		OK	10:39	

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 04/10/2016 07:35:10

Diagnostics Check		OK	07:31	
Air Blank		0.000	07:31	
Control Test		0.078	07:31	
Air Blank		0.000	07:32	
Subject Sample	# 1		07:33	Subject Test Refused
Breath Volume		0.000	07:33	
Air Blank		0.000	07:34	
Control Test		0.079	07:34	
Air Blank		0.000	07:35	
Diagnostics Check		OK	07:35	

Data Download Date/Time: 04/28/2016 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: CABRERA-VELOZ

DATE/TIME OF TEST: 04/13/2016 17:56:13

Diagnostics Check		OK	17:49
Air Blank		0.000	17:49
Control Test		0.078	17:49
Air Blank		0.000	17:50
Subject Sample	# 1	0.065	17:51
Breath Volume		1.628	17:51
Air Blank		0.000	17:52
Air Blank		0.000	17:53
Subject Sample	# 2	0.065	17:54
Breath Volume		1.515	17:54
Air Blank		0.000	17:55
Control Test		0.077	17:55
Air Blank		0.000	17:56
Diagnostics Check		OK	17:56

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH,ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: QUIJADA

DATE/TIME OF TEST: 04/14/2016 10:19:03

Diagnostics Check		OK	10:16	
Air Blank		0.000	10:16	
Control Test		0.079	10:16	
Air Blank		0.000	10:17	
Subject Sample	# 1		10:17	Subject Test Refused
Breath Volume		0.000	10:17	
Air Blank		0.000	10:18	
Control Test		0.078	10:18	
Air Blank		0.000	10:18	
Diagnostics Check		OK	10:19	

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/17/2016 02:16:43

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.078	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1	0.191	02:05	
Breath Volume		1.867	02:05	
Air Blank		0.000	02:06	
Air Blank		0.000	02:08	
Subject Sample	# 2	0.204	02:11	Volume Not Met
Breath Volume		0.093	02:11	
Air Blank		0.000	02:12	
Air Blank		0.000	02:13	
Subject Sample	# 3	0.214	02:14	
Breath Volume		1.582	02:14	
Air Blank		0.000	02:15	
Control Test		0.074	02:16	Control Outside Tolerance
Air Blank		0.000	02:16	

Data Download Date/Time: 04/28/2016 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/17/2016 02:25:25

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.076	02:19	
Air Blank		0.000	02:19	
Subject Sample	# 1	0.234	02:21	
Breath Volume		3.441	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2	0.222	02:23	
Breath Volume		1.675	02:23	
Air Blank		0.000	02:24	
Control Test		0.074	02:24	Control Outside Tolerance
Air Blank		0.000	02:25	

Data Download Date/Time: 04/28/2016 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 04/17/2016 02:35:37

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.076	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.212	02:30
Breath Volume		2.894	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.213	02:34
Breath Volume		3.285	02:34
Air Blank		0.000	02:34
Control Test		0.077	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 04/28/2016 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,RICARDO,J
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: MORENO,RICARDO
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 04/19/2016 03:18:48

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.078	03:13
Air Blank		0.000	03:13
Subject Sample	# 1	0.140	03:14
Breath Volume		3.105	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.139	03:17
Breath Volume		1.347	03:17
Air Blank		0.000	03:17
Control Test		0.077	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAESTREH,HANS,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: CELESTIN

DATE/TIME OF TEST: 04/19/2016 20:07:28

Diagnostics Check		OK	20:03	
Air Blank		0.000	20:03	
Control Test		0.078	20:04	
Air Blank		0.000	20:04	
Subject Sample	# 1	0.114	20:05	
Breath Volume		1.929	20:05	
Air Blank			20:05	Improper Sample
Air Blank		0.000	20:06	
Control Test		0.079	20:06	
Air Blank		0.000	20:07	
Diagnostics Check		OK	20:07	

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAESTRE,HANS,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: CELESTIN

DATE/TIME OF TEST: 04/19/2016 20:16:44

Diagnostics Check		OK	20:09
Air Blank		0.000	20:10
Control Test		0.076	20:10
Air Blank		0.000	20:10
Subject Sample	# 1	0.111	20:12
Breath Volume		1.980	20:12
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.111	20:15
Breath Volume		1.820	20:15
Air Blank		0.000	20:15
Control Test		0.076	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTELO,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BURNS,JEFFREY
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 04/21/2016 00:57:49

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.075	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.076	00:57	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 04/28/2016 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANKOFF,BRIAN,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: HANKOFF,BRIAN
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 04/23/2016 02:12:54

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.078	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.158	02:07
Breath Volume		3.222	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.153	02:11
Breath Volume		2.226	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 04/28/2016 12:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 04/01/2016 00:22:31

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.078	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.177	00:17
Breath Volume		2.929	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.178	00:20
Breath Volume		2.589	00:20
Air Blank		0.000	00:21
Control Test		0.078	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 04/20/2016 10:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RODRIGUEZ FERNANDEZ

DATE/TIME OF TEST: 04/01/2016 00:52:58

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.190	00:48
Breath Volume		2.796	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:50
Subject Sample	# 2	0.196	00:51
Breath Volume		2.101	00:51
Air Blank		0.000	00:52
Control Test		0.076	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: MILANESE

DATE/TIME OF TEST: 04/01/2016 04:42:50

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:28	
Control Test		0.078	04:29	
Air Blank		0.000	04:29	
Subject Sample	# 1		04:32	No Sample Provided
Breath Volume		0.000	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:35	
Subject Sample	# 2	0.161	04:38	
Breath Volume		1.765	04:38	
Air Blank		0.000	04:39	
Air Blank		0.000	04:40	
Subject Sample	# 3	0.160	04:41	
Breath Volume		1.546	04:41	
Air Blank		0.000	04:41	
Control Test		0.078	04:42	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 04/20/2016 10:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: PEREZ MONZON

DATE/TIME OF TEST: 04/02/2016 07:19:44

Diagnostics Check		OK	07:13
Air Blank		0.000	07:13
Control Test		0.079	07:13
Air Blank		0.000	07:14
Subject Sample	# 1	0.108	07:15
Breath Volume		4.000	07:15
Air Blank		0.000	07:15
Air Blank		0.000	07:17
Subject Sample	# 2	0.107	07:18
Breath Volume		3.019	07:18
Air Blank		0.000	07:18
Control Test		0.077	07:19
Air Blank		0.000	07:19
Diagnostics Check		OK	07:19

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: NOGUEIRA

DATE/TIME OF TEST: 04/03/2016 02:20:03

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.216	02:15
Breath Volume		1.628	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.200	02:18
Breath Volume		2.285	02:18
Air Blank		0.000	02:19
Control Test		0.076	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:20

Data Download Date/Time: 04/20/2016 10:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: AGUERO-VEITIA

DATE/TIME OF TEST: 04/03/2016 22:17:06

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:02	
Control Test		0.078	22:02	
Air Blank		0.000	22:02	
Subject Sample	# 1	0.236	22:05	
Breath Volume		2.265	22:05	
Air Blank		0.000	22:06	
Air Blank		0.000	22:07	
Subject Sample	# 2	0.206	22:10	Volume Not Met
Breath Volume		0.585	22:10	
Air Blank		0.000	22:11	
Air Blank		0.000	22:13	
Subject Sample	# 3	0.233	22:15	
Breath Volume		2.242	22:15	
Air Blank		0.000	22:16	
Control Test		0.073	22:16	Control Outside Tolerance
Air Blank		0.000	22:17	

Data Download Date/Time: 04/20/2016 10:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: AGUERO-VEITIA

DATE/TIME OF TEST: 04/03/2016 22:52:56

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.077	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.217	22:47
Breath Volume		3.285	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.229	22:51
Breath Volume		2.902	22:51
Air Blank		0.000	22:52
Control Test		0.077	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: RIVERA-CASTILLO

DATE/TIME OF TEST: 04/04/2016 02:43:06

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.078	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:42	
Control Test		0.077	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: UNTORIA-GONZALEZ

DATE/TIME OF TEST: 04/07/2016 00:27:20

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.079	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1	0.135	00:19	Volume Not Met
Breath Volume		0.101	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2	0.130	00:25	Volume Not Met
Breath Volume		0.035	00:25	
Air Blank		0.000	00:26	
Control Test		0.076	00:26	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 04/20/2016 10:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: UNTORIA GONZALEZ

DATE/TIME OF TEST: 04/07/2016 00:44:31

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.075	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1	0.110	00:35	Volume Not Met
Breath Volume		0.296	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 2	0.136	00:39	
Breath Volume		1.421	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 3	0.146	00:42	
Breath Volume		1.507	00:42	
Air Blank		0.000	00:43	
Control Test		0.074	00:43	Control Outside Tolerance
Air Blank		0.000	00:44	

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: UNTORIA GONZALEZ

DATE/TIME OF TEST: 04/07/2016 00:56:09

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:46	
Control Test		0.075	00:47	
Air Blank		0.000	00:47	
Subject Sample	# 1	0.128	00:50	
Breath Volume		1.289	00:50	
Air Blank		0.000	00:51	
Air Blank		0.000	00:53	
Subject Sample	# 2	0.128	00:54	
Breath Volume		1.269	00:54	
Air Blank		0.000	00:55	
Control Test		0.074	00:55	Control Outside Tolerance
Air Blank		0.000	00:56	

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: DIANDERAS

DATE/TIME OF TEST: 04/07/2016 01:18:10

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.075	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.076	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:37

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 04/07/2016 12:00:57

Diagnostics Check		OK	11:58	
Air Blank		0.000	11:58	
Control Test		0.078	11:58	
Air Blank		0.000	11:59	
Subject Sample	# 1		11:59	Subject Test Refused
Breath Volume		0.000	11:59	
Air Blank		0.000	12:00	
Control Test		0.077	12:00	
Air Blank		0.000	12:00	
Diagnostics Check		OK	12:00	

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/08/2016 03:04:17

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:59	
Control Test		0.078	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:02	Subject Test Refused
Breath Volume		0.101	03:02	
Air Blank		0.000	03:03	
Control Test		0.078	03:03	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 04/20/2016 10:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DESIUS,DUCLAIRE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: VALDES-DAUSSA

DATE/TIME OF TEST: 04/08/2016 03:22:51

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:12	
Control Test		0.077	03:12	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.171	03:16	Volume Not Met
Breath Volume		0.582	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 2	0.164	03:21	Volume Not Met
Breath Volume		1.035	03:21	
Air Blank			03:22	Purge Fail
Air Blank		0.000	03:22	

Data Download Date/Time: 04/20/2016 10:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DESIUS,DUCLAIRE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: VALDES-DAUSSA

DATE/TIME OF TEST: 04/08/2016 03:45:04

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.071	03:44	Control Outside Tolerance
Air Blank	0.000	03:45	

Data Download Date/Time: 04/20/2016 10:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DESIUS,DUCLAIRE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: VALDES DAUSSA

DATE/TIME OF TEST: 04/08/2016 03:49:48

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:47	
Control Test		0.075	03:47	
Air Blank		0.000	03:48	
Subject Sample	# 1	0.152	03:48	
Breath Volume		1.562	03:48	
Air Blank			03:49	Purge Fail
Air Blank		0.000	03:49	

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DESIUS,DUCLAIRE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: VALDES DAUSSA

DATE/TIME OF TEST: 04/08/2016 04:13:22

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:10	
Control Test		0.076	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.077	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FAVORS

DATE/TIME OF TEST: 04/11/2016 01:15:29

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:13	
Control Test		0.079	01:13	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.078	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUERRA,FRANCISCO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GUERRA,FRANCISCO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MASTRAPA

DATE/TIME OF TEST: 04/11/2016 03:09:12

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.079	03:07	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:08	
Control Test		0.078	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/14/2016 02:02:24

Diagnostics Check		OK	01:59	
Air Blank		0.000	01:59	
Control Test		0.079	02:00	
Air Blank		0.000	02:00	
Subject Sample	# 1		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:01	
Control Test		0.077	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 04/20/2016 10:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: AGREDA

DATE/TIME OF TEST: 04/16/2016 01:05:15

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.189	00:58
Breath Volume		2.035	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:00
Subject Sample	# 2	0.178	01:03
Breath Volume		2.058	01:03
Air Blank		0.000	01:04
Control Test		0.077	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 04/16/2016 01:51:56

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:49	
Control Test		0.077	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:50	Subject Test Refused
Breath Volume		0.000	01:50	
Air Blank		0.000	01:51	
Control Test		0.077	01:51	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOWNSEND,JESSICA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: ARANCIBIA

DATE/TIME OF TEST: 04/16/2016 02:38:35

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.078	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.078	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 04/20/2016 10:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 04/16/2016 06:14:41

Diagnostics Check		OK	06:07
Air Blank		0.000	06:08
Control Test		0.078	06:08
Air Blank		0.000	06:09
Subject Sample	# 1	0.112	06:09
Breath Volume		1.402	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.116	06:13
Breath Volume		1.132	06:13
Air Blank		0.000	06:13
Control Test		0.078	06:14
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: RUBIO

DATE/TIME OF TEST: 04/17/2016 01:44:15

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.079	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.092	01:39
Breath Volume		1.113	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.093	01:42
Breath Volume		1.714	01:42
Air Blank		0.000	01:43
Control Test		0.077	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/17/2016 05:24:42

Diagnostics Check		OK	05:15	
Air Blank		0.000	05:15	
Control Test		0.079	05:16	
Air Blank		0.000	05:16	
Subject Sample	# 1	0.164	05:19	Volume Not Met
Breath Volume		0.144	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:22	
Subject Sample	# 2		05:23	Subject Test Refused
Breath Volume		0.136	05:23	
Air Blank		0.000	05:23	
Control Test		0.075	05:24	
Air Blank		0.000	05:24	
Diagnostics Check		OK	05:24	

Data Download Date/Time: 04/20/2016 10:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/20/2016 23:54:17

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:51	
Control Test		0.078	23:52	
Air Blank		0.000	23:52	
Subject Sample	# 1		23:52	Subject Test Refused
Breath Volume		0.000	23:52	
Air Blank		0.000	23:53	
Control Test		0.078	23:53	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 04/21/2016 01:50:38

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.000	01:46
Breath Volume		2.015	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.000	01:49
Breath Volume		1.503	01:49
Air Blank		0.000	01:49
Control Test		0.077	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 04/21/2016 03:22:50

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.078	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.152	03:18
Breath Volume		2.078	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.138	03:21
Breath Volume		1.289	03:21
Air Blank		0.000	03:21
Control Test		0.077	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 05/13/2016 09:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 04/22/2016 04:04:27

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.079	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.127	03:59
Breath Volume		1.390	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.127	04:02
Breath Volume		1.203	04:02
Air Blank		0.000	04:03
Control Test		0.076	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUESTA,DANIEL,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUESTA,DANIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: PERAZA

DATE/TIME OF TEST: 04/23/2016 08:09:59

Diagnostics Check		OK	07:58	
Air Blank		0.000	07:59	
Control Test		0.078	07:59	
Air Blank		0.000	07:59	
Subject Sample	# 1		08:03	No Sample Provided
Breath Volume		0.000	08:03	
Air Blank		0.000	08:03	
Air Blank		0.000	08:05	
Subject Sample	# 2		08:08	No Sample Provided
Breath Volume		0.000	08:08	
Air Blank		0.000	08:09	
Control Test		0.078	08:09	
Air Blank		0.000	08:09	
Diagnostics Check		OK	08:09	

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUESTA,DANIEL,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUESTA,DANIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: PERAZA

DATE/TIME OF TEST: 04/23/2016 08:22:04

Diagnostics Check		OK	08:14	
Air Blank		0.000	08:14	
Control Test		0.078	08:15	
Air Blank		0.000	08:15	
Subject Sample	# 1		08:18	No Sample Provided
Breath Volume		0.000	08:18	
Air Blank		0.000	08:19	
Air Blank		0.000	08:21	
Subject Sample	# 2		08:21	RFI Detect
Breath Volume		0.000	08:21	
Air Blank			08:21	RFI Detect
Air Blank			08:22	RFI Detect

Data Download Date/Time: 05/13/2016 09:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/23/2016 21:47:45

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:45	
Control Test		0.080	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:46	
Control Test		0.078	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREOZZI,ADAM,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/26/2016 01:28:51

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.176	01:23
Breath Volume		1.894	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:25
Subject Sample	# 2	0.162	01:27
Breath Volume		2.363	01:27
Air Blank		0.000	01:27
Control Test		0.076	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: PADRON ALARCON

DATE/TIME OF TEST: 04/27/2016 03:25:22

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.079	03:19
Air Blank		0.000	03:19
Subject Sample	# 1	0.140	03:20
Breath Volume		1.531	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.140	03:23
Breath Volume		1.464	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: GIRAUD

DATE/TIME OF TEST: 04/28/2016 01:50:15

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.193	01:45
Breath Volume		1.472	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.190	01:48
Breath Volume		2.078	01:48
Air Blank		0.000	01:49
Control Test		0.076	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 05/13/2016 09:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RODRIGUEZ MOREJON

DATE/TIME OF TEST: 04/28/2016 02:03:13

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.077	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.182	01:58
Breath Volume		1.644	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.185	02:01
Breath Volume		1.347	02:01
Air Blank		0.000	02:02
Control Test		0.076	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: NUNEZ,T
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 04/29/2016 06:41:13

Diagnostics Check		OK	06:38	
Air Blank		0.000	06:38	
Control Test		0.079	06:38	
Air Blank		0.000	06:39	
Subject Sample	# 1		06:39	Subject Test Refused
Breath Volume		0.000	06:39	
Air Blank		0.000	06:40	
Control Test		0.077	06:40	
Air Blank		0.000	06:41	
Diagnostics Check		OK	06:41	

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: BASTOS PETIT

DATE/TIME OF TEST: 04/29/2016 07:49:26

Diagnostics Check	OK	07:48	
Air Blank		07:48	RFI Detect
Air Blank	0.000	07:49	

Data Download Date/Time: 05/13/2016 09:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:26

SUBJECT'S LAST NAME: BASTOS PETIT

DATE/TIME OF TEST: 04/29/2016 07:59:38

Diagnostics Check		OK	07:51	
Air Blank		0.000	07:51	
Control Test		0.079	07:51	
Air Blank		0.000	07:52	
Subject Sample	# 1		07:55	No Sample Provided
Breath Volume		0.000	07:55	
Air Blank		0.000	07:56	
Air Blank		0.000	07:58	
Subject Sample	# 2		07:58	Subject Test Refused
Breath Volume		0.000	07:58	
Air Blank		0.000	07:58	
Control Test		0.078	07:59	
Air Blank		0.000	07:59	
Diagnostics Check		OK	07:59	

Data Download Date/Time: 05/13/2016 09:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 04/30/2016 02:31:14

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.190	02:25
Breath Volume		1.738	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.193	02:29
Breath Volume		1.316	02:29
Air Blank		0.000	02:30
Control Test		0.077	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/30/2016 06:28:07

Diagnostics Check		OK	06:21
Air Blank		0.000	06:21
Control Test		0.078	06:22
Air Blank		0.000	06:22
Subject Sample	# 1	0.100	06:23
Breath Volume		2.085	06:23
Air Blank		0.000	06:24
Air Blank		0.000	06:25
Subject Sample	# 2	0.101	06:26
Breath Volume		2.214	06:26
Air Blank		0.000	06:27
Control Test		0.078	06:27
Air Blank		0.000	06:28
Diagnostics Check		OK	06:28

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: BUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: FAURA LEYVA

DATE/TIME OF TEST: 04/30/2016 21:41:18

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:38	
Air Blank		0.000	21:39	
Subject Sample	# 1		21:39	Subject Test Refused
Breath Volume		0.000	21:39	
Air Blank		0.000	21:40	
Control Test		0.077	21:40	
Air Blank		0.000	21:41	
Diagnostics Check		OK	21:41	

Data Download Date/Time: 05/13/2016 09:54

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 04/05/2016 01:37:20

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.080	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.108	01:32
Breath Volume		3.425	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.110	01:35
Breath Volume		4.132	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 04/21/2016 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BEDOYA ARIAS

DATE/TIME OF TEST: 04/05/2016 20:49:31

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.080	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.097	20:45
Breath Volume		2.234	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.096	20:47
Breath Volume		1.214	20:47
Air Blank		0.000	20:48
Control Test		0.079	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 04/21/2016 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: COLLADO GARCIA

DATE/TIME OF TEST: 04/08/2016 23:08:35

Diagnostics Check		OK	23:05	
Air Blank		0.000	23:06	
Control Test		0.080	23:06	
Air Blank		0.000	23:06	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:07	
Control Test		0.079	23:07	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:08	

Data Download Date/Time: 04/21/2016 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAYARD,CLIFFORD,P
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CAYARD,CLIFFORD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/10/2016 01:22:39

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:20	
Control Test		0.079	01:20	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:21	
Control Test		0.079	01:22	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 04/21/2016 11:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLON,RICHARD,G
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BELLON,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/11/2016 04:28:10

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.197	04:23
Breath Volume		3.101	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.198	04:26
Breath Volume		3.148	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 04/21/2016 11:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,MAURICE,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HENRY,MAURICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/12/2016 03:00:06

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.080	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 04/21/2016 11:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 04/15/2016 20:24:56

Diagnostics Check	OK	20:24	
Air Blank		20:24	RFI Detect
Air Blank		20:24	RFI Detect

Data Download Date/Time: 04/21/2016 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 04/15/2016 20:39:46

Diagnostics Check		OK	20:36	
Air Blank		0.000	20:37	
Control Test		0.079	20:37	
Air Blank		0.000	20:38	
Subject Sample	# 1		20:38	Subject Test Refused
Breath Volume		0.000	20:38	
Air Blank		0.000	20:38	
Control Test		0.079	20:39	
Air Blank		0.000	20:39	
Diagnostics Check		OK	20:39	

Data Download Date/Time: 04/21/2016 11:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: ASLANIAN

DATE/TIME OF TEST: 04/15/2016 22:07:57

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.079	22:03	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:06	Subject Test Refused
Breath Volume		0.000	22:06	
Air Blank		0.000	22:07	
Control Test		0.080	22:07	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:07	

Data Download Date/Time: 04/21/2016 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: BILLINGS

DATE/TIME OF TEST: 04/18/2016 17:47:30

Diagnostics Check		OK	17:41
Air Blank		0.000	17:41
Control Test		0.080	17:42
Air Blank		0.000	17:42
Subject Sample	# 1	0.000	17:43
Breath Volume		2.820	17:43
Air Blank		0.000	17:43
Air Blank		0.000	17:45
Subject Sample	# 2	0.000	17:46
Breath Volume		3.359	17:46
Air Blank		0.000	17:46
Control Test		0.080	17:46
Air Blank		0.000	17:47
Diagnostics Check		OK	17:47

Data Download Date/Time: 04/21/2016 11:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: HENRICHMERTINS

DATE/TIME OF TEST: 04/22/2016 22:11:40

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:09	
Control Test		0.079	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:10	Subject Test Refused
Breath Volume		0.000	22:10	
Air Blank		0.000	22:10	
Control Test		0.080	22:11	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 05/12/2016 10:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 04/24/2016 18:55:59

Diagnostics Check		OK	18:41	
Air Blank		0.000	18:41	
Control Test		0.080	18:41	
Air Blank		0.000	18:42	
Subject Sample	# 1	0.200	18:45	Volume Not Met
Breath Volume		0.828	18:45	
Air Blank		0.000	18:46	
Air Blank		0.000	18:47	
Subject Sample	# 2	0.207	18:51	
Breath Volume		1.640	18:51	
Air Blank		0.000	18:52	
Air Blank		0.000	18:53	
Subject Sample	# 3	0.200	18:54	
Breath Volume		1.378	18:54	
Air Blank		0.000	18:55	
Control Test		0.075	18:55	
Air Blank		0.000	18:55	
Diagnostics Check		OK	18:55	

Data Download Date/Time: 05/12/2016 10:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: PUENTE CANAY

DATE/TIME OF TEST: 04/24/2016 21:11:11

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.079	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.203	21:06
Breath Volume		2.882	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.201	21:09
Breath Volume		3.136	21:09
Air Blank		0.000	21:10
Control Test		0.080	21:10
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 05/12/2016 10:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 04/25/2016 18:57:32

Diagnostics Check		OK	18:51	
Air Blank		0.000	18:51	
Control Test		0.077	18:52	
Air Blank		0.000	18:52	
Subject Sample	# 1	0.189	18:52	
Breath Volume		2.117	18:52	
Air Blank		0.000	18:53	
Air Blank		0.000	18:55	
Subject Sample	# 2	0.175	18:55	
Breath Volume		1.847	18:55	
Air Blank		0.000	18:56	
Control Test		0.070	18:56	Control Outside Tolerance
Air Blank		0.000	18:57	

Data Download Date/Time: 05/12/2016 10:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 04/25/2016 19:19:48

Diagnostics Check	OK	19:19	
Air Blank		19:19	RFI Detect
Air Blank		19:19	RFI Detect

Data Download Date/Time: 05/12/2016 10:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: CENTENO

DATE/TIME OF TEST: 04/25/2016 19:27:31

Diagnostics Check		OK	19:21
Air Blank		0.000	19:21
Control Test		0.079	19:22
Air Blank		0.000	19:22
Subject Sample	# 1	0.188	19:23
Breath Volume		3.140	19:23
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.177	19:25
Breath Volume		3.238	19:25
Air Blank		0.000	19:26
Control Test		0.079	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 05/12/2016 10:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,I
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 04/30/2016 16:03:33

Diagnostics Check		OK	15:57
Air Blank		0.000	15:58
Control Test		0.080	15:58
Air Blank		0.000	15:58
Subject Sample	# 1	0.000	15:59
Breath Volume		3.519	15:59
Air Blank		0.000	15:59
Air Blank		0.000	16:01
Subject Sample	# 2	0.000	16:02
Breath Volume		3.484	16:02
Air Blank		0.000	16:02
Control Test		0.079	16:02
Air Blank		0.000	16:03
Diagnostics Check		OK	16:03

Data Download Date/Time: 05/12/2016 10:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: PEREIRA GOLAN

DATE/TIME OF TEST: 04/30/2016 21:28:40

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.079	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.205	21:23
Breath Volume		2.808	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.193	21:27
Breath Volume		3.019	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:28
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 05/12/2016 10:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,MAURICE,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HENRY,MAURICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MENA

DATE/TIME OF TEST: 04/05/2016 00:44:39

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.079	00:41	
Air Blank		0.000	00:42	
Subject Sample	# 1	0.000	00:43	
Breath Volume		2.695	00:43	
Air Blank			00:43	RFI Detect
Air Blank		0.000	00:43	
Control Test		0.081	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 04/21/2016 11:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,MAURICE,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HENRY,MAURICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: MENA

DATE/TIME OF TEST: 04/05/2016 00:54:47

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.000	00:49
Breath Volume		1.988	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.000	00:53
Breath Volume		1.492	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 04/21/2016 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,LUCY,K
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RODRIGUEZ,LUCY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: DELAGUARDIA

DATE/TIME OF TEST: 04/07/2016 22:29:07

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.079	22:22
Air Blank		0.000	22:23
Subject Sample	# 1	0.245	22:23
Breath Volume		1.597	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.246	22:27
Breath Volume		3.179	22:27
Air Blank		0.000	22:28
Control Test		0.079	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 04/21/2016 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: HELVIE

DATE/TIME OF TEST: 04/09/2016 03:16:40

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.182	03:10
Breath Volume		1.425	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.174	03:14
Breath Volume		1.628	03:14
Air Blank		0.000	03:15
Control Test		0.081	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 04/21/2016 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/10/2016 11:11:37

Diagnostics Check		OK	11:08	
Air Blank		0.000	11:08	
Control Test		0.079	11:09	
Air Blank		0.000	11:09	
Subject Sample	# 1		11:10	Interferent Detect
Breath Volume		2.605	11:10	
Air Blank		0.000	11:10	
Control Test		0.080	11:11	
Air Blank		0.000	11:11	
Diagnostics Check		OK	11:11	

Data Download Date/Time: 04/21/2016 11:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/10/2016 11:35:43

Diagnostics Check		OK	11:32	
Air Blank		0.000	11:33	
Control Test		0.079	11:33	
Air Blank		0.000	11:33	
Subject Sample	# 1		11:34	Interferent Detect
Breath Volume		3.910	11:34	
Air Blank		0.000	11:34	
Control Test		0.079	11:35	
Air Blank		0.000	11:35	
Diagnostics Check		OK	11:35	

Data Download Date/Time: 04/21/2016 11:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 04/10/2016 11:43:28

Diagnostics Check		OK	11:40	
Air Blank		0.000	11:40	
Control Test		0.079	11:41	
Air Blank		0.000	11:41	
Subject Sample	# 1		11:42	Interferent Detect
Breath Volume		3.921	11:42	
Air Blank		0.000	11:42	
Control Test		0.080	11:42	
Air Blank		0.000	11:43	
Diagnostics Check		OK	11:43	

Data Download Date/Time: 04/21/2016 11:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRONSON,SAMONYA,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: MAJOR

DATE/TIME OF TEST: 04/10/2016 18:45:44

Diagnostics Check		OK	18:39
Air Blank		0.000	18:40
Control Test		0.079	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.210	18:41
Breath Volume		2.140	18:41
Air Blank		0.000	18:41
Air Blank		0.000	18:43
Subject Sample	# 2	0.204	18:44
Breath Volume		2.519	18:44
Air Blank		0.000	18:44
Control Test		0.079	18:45
Air Blank		0.000	18:45
Diagnostics Check		OK	18:45

Data Download Date/Time: 04/21/2016 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWART,RYAN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: COWART,RYAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: JUSTO ALONSO

DATE/TIME OF TEST: 04/11/2016 01:40:33

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.080	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 1	0.183	01:38	
Breath Volume		1.761	01:38	
Air Blank			01:39	RFI Detect
Air Blank		0.000	01:39	
Control Test		0.081	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 04/21/2016 11:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWART,RYAN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: COWART,RYAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: JUSTO ALONSO

DATE/TIME OF TEST: 04/11/2016 01:47:36

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.200	01:43
Breath Volume		1.566	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.200	01:46
Breath Volume		1.554	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 04/21/2016 11:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVE,YAARA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: REEVE,YAARA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 04/18/2016 10:22:08

Diagnostics Check		OK	10:19	
Air Blank		0.000	10:19	
Control Test		0.080	10:19	
Air Blank		0.000	10:20	
Subject Sample	# 1		10:20	Subject Test Refused
Breath Volume		0.000	10:20	
Air Blank		0.000	10:21	
Control Test		0.079	10:21	
Air Blank		0.000	10:22	
Diagnostics Check		OK	10:22	

Data Download Date/Time: 04/21/2016 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,M
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 04/06/2016 23:02:12

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.196	22:57
Breath Volume		1.761	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.200	23:00
Breath Volume		1.570	23:00
Air Blank		0.000	23:01
Control Test		0.077	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 05/02/2016 08:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: MAYA,JOSE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: IRIGARAY

DATE/TIME OF TEST: 04/10/2016 00:08:43

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.079	00:06	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:07	
Control Test		0.079	00:08	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 05/02/2016 08:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 04/14/2016 17:12:14

Diagnostics Check		OK	17:06
Air Blank		0.000	17:06
Control Test		0.080	17:06
Air Blank		0.000	17:07
Subject Sample	# 1	0.000	17:07
Breath Volume		2.207	17:07
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 2	0.000	17:10
Breath Volume		1.937	17:10
Air Blank		0.000	17:11
Control Test		0.079	17:11
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 05/02/2016 08:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 04/17/2016 21:45:08

Diagnostics Check		OK	21:38	
Air Blank		0.000	21:38	
Control Test		0.079	21:39	
Air Blank		0.000	21:39	
Subject Sample	# 1	0.240	21:40	
Breath Volume		1.632	21:40	
Air Blank		0.000	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 2	0.229	21:43	
Breath Volume		1.898	21:43	
Air Blank		0.000	21:44	
Control Test		0.074	21:44	Control Outside Tolerance
Air Blank		0.000	21:45	

Data Download Date/Time: 05/02/2016 08:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: SOLIS

DATE/TIME OF TEST: 04/17/2016 21:54:36

Diagnostics Check		OK	21:46
Air Blank		0.000	21:47
Control Test		0.079	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.226	21:49
Breath Volume		1.703	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.226	21:53
Breath Volume		1.367	21:53
Air Blank		0.000	21:53
Control Test		0.077	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 05/02/2016 08:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 04/24/2016 00:36:01

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:34	Subject Test Refused
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Control Test		0.079	00:35	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 05/02/2016 08:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: CHATTAHOOCHEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POSTON,WILLIAM,A
BREATH TEST OPERATOR AGENCY: Chattahoochee PD
ARRESTING OFFICER: POSTON,WILLIAM
ARRESTING OFFICER AGENCY: Chattahoochee PD
LAST AGENCY INSPECTION: 03/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 04/17/2016 18:24:22

Diagnostics Check		OK	18:17
Air Blank		0.000	18:18
Control Test		0.077	18:18
Air Blank		0.000	18:19
Subject Sample	# 1	0.237	18:19
Breath Volume		3.320	18:19
Air Blank		0.000	18:20
Air Blank		0.000	18:22
Subject Sample	# 2	0.229	18:22
Breath Volume		3.320	18:22
Air Blank		0.000	18:23
Control Test		0.077	18:23
Air Blank		0.000	18:24
Diagnostics Check		OK	18:24

Data Download Date/Time: 04/27/2016 22:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALTAIS,ANDREW,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: PIFER,J
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PRESSLER

DATE/TIME OF TEST: 04/23/2016 19:37:18

Diagnostics Check		OK	19:30
Air Blank		0.000	19:31
Control Test		0.079	19:31
Air Blank		0.000	19:32
Subject Sample	# 1	0.129	19:32
Breath Volume		4.027	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.127	19:35
Breath Volume		4.082	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 05/16/2016 12:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MATTSON,CHRIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 04/23/2016 19:21:30

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:18	
Control Test		0.081	19:18	
Air Blank		0.000	19:19	
Subject Sample	# 1		19:20	Subject Test Refused
Breath Volume		0.000	19:20	
Air Blank			19:21	RFI Detect
Air Blank		0.000	19:21	

Data Download Date/Time: 05/12/2016 20:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MATTSON,CHRIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 04/28/2016 20:41:46

Diagnostics Check	OK	20:40	
Air Blank		20:41	Improper Sample
Air Blank	0.000	20:41	

Data Download Date/Time: 05/12/2016 20:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MATTSON,CHRIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 04/28/2016 20:49:27

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.080	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.107	20:45
Breath Volume		2.949	20:45
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.101	20:48
Breath Volume		2.613	20:48
Air Blank		0.000	20:48
Control Test		0.079	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 05/12/2016 20:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BAUMGARTNER,COURTNEY
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JACOBSON

DATE/TIME OF TEST: 04/30/2016 22:24:08

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:21	
Control Test		0.079	22:22	
Air Blank		0.000	22:22	
Subject Sample	# 1		22:22	Subject Test Refused
Breath Volume		0.000	22:22	
Air Blank		0.000	22:23	
Control Test		0.079	22:23	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 05/12/2016 20:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: OLLOQUI

DATE/TIME OF TEST: 04/10/2016 21:28:00

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:25	
Control Test		0.081	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1		21:26	Subject Test Refused
Breath Volume		0.000	21:26	
Air Blank		0.000	21:27	
Control Test		0.081	21:27	
Air Blank		0.000	21:27	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 05/13/2016 09:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERRER,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: FERRER,RAFAEL
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WEDDERBURN

DATE/TIME OF TEST: 04/10/2016 22:58:32

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.234	22:53
Breath Volume		1.507	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.242	22:56
Breath Volume		2.492	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 05/13/2016 09:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HAYLES

DATE/TIME OF TEST: 04/25/2016 01:37:45

Diagnostics Check	OK	01:37	
Air Blank		01:37	RFI Detect
Air Blank		01:37	RFI Detect

Data Download Date/Time: 05/13/2016 09:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,CRUZ,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HAYLES

DATE/TIME OF TEST: 04/25/2016 01:42:33

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:40	
Control Test		0.081	01:40	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:41	Subject Test Refused
Breath Volume		0.000	01:41	
Air Blank		0.000	01:41	
Control Test		0.080	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 05/13/2016 09:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000909

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: DELANEY

DATE/TIME OF TEST: 04/30/2016 18:55:19

Diagnostics Check		OK	18:47
Air Blank		0.000	18:48
Control Test		0.082	18:48
Air Blank		0.000	18:49
Subject Sample	# 1	0.159	18:50
Breath Volume		2.394	18:50
Air Blank		0.000	18:51
Air Blank		0.000	18:52
Subject Sample	# 2	0.153	18:53
Breath Volume		2.238	18:53
Air Blank		0.000	18:54
Control Test		0.079	18:54
Air Blank		0.000	18:55
Diagnostics Check		OK	18:55

Data Download Date/Time: 05/13/2016 09:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIKEMA,ELIZABETH,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: HUBATCH,NICHOLAS
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05514080A1
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 04/04/2016 01:40:14

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.081	01:39	
Air Blank		01:39	RFI Detect
Air Blank	0.000	01:40	

Data Download Date/Time: 05/26/2016 22:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIKEMA,ELIZABETH,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: HUBATCH,NICHOLAS
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 04/04/2016 01:50:12

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.250	01:45
Breath Volume		2.066	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.231	01:48
Breath Volume		1.789	01:48
Air Blank		0.000	01:49
Control Test		0.081	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 05/26/2016 22:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUIKEMA,ELIZABETH,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: HERON,LAUREN
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 04/08/2016 01:45:28

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.220	01:40
Breath Volume		1.835	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.228	01:43
Breath Volume		1.992	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 05/26/2016 22:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GONZALEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WHELLER

DATE/TIME OF TEST: 04/02/2016 04:29:41

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.165	04:24
Breath Volume		1.761	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.177	04:27
Breath Volume		1.574	04:27
Air Blank		0.000	04:28
Control Test		0.078	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 04/13/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/02/2016 18:38:16

Diagnostics Check		OK	18:32
Air Blank		0.000	18:32
Control Test		0.080	18:32
Air Blank		0.000	18:33
Subject Sample	# 1	0.066	18:33
Breath Volume		3.894	18:33
Air Blank		0.000	18:34
Air Blank		0.000	18:36
Subject Sample	# 2	0.064	18:36
Breath Volume		4.234	18:36
Air Blank		0.000	18:37
Control Test		0.079	18:37
Air Blank		0.000	18:38
Diagnostics Check		OK	18:38

Data Download Date/Time: 04/13/2016 13:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,CARLOS,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GONZALEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 04/03/2016 01:31:02

Diagnostics Check		OK	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.145	01:26
Breath Volume		1.957	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:28
Subject Sample	# 2	0.148	01:29
Breath Volume		1.859	01:29
Air Blank		0.000	01:30
Control Test		0.078	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:31

Data Download Date/Time: 04/13/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: GIANCASPRO

DATE/TIME OF TEST: 04/04/2016 18:22:47

Diagnostics Check		OK	18:16
Air Blank		0.000	18:17
Control Test		0.079	18:17
Air Blank		0.000	18:17
Subject Sample	# 1	0.007	18:18
Breath Volume		1.437	18:18
Air Blank		0.000	18:18
Air Blank		0.000	18:20
Subject Sample	# 2	0.000	18:21
Breath Volume		2.449	18:21
Air Blank		0.000	18:21
Control Test		0.078	18:22
Air Blank		0.000	18:22
Diagnostics Check		OK	18:22

Data Download Date/Time: 04/13/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: POLINUS

DATE/TIME OF TEST: 04/05/2016 00:45:52

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.080	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1	0.131	00:35	
Breath Volume		1.394	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 2	0.125	00:41	Volume Not Met
Breath Volume		0.832	00:41	
Air Blank		0.000	00:41	
Air Blank		0.000	00:43	
Subject Sample	# 3	0.127	00:44	
Breath Volume		1.535	00:44	
Air Blank		0.000	00:44	
Control Test		0.076	00:45	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 04/13/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ELLIS,BILLY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: GRANTHAM

DATE/TIME OF TEST: 04/05/2016 08:21:55

Diagnostics Check		OK	08:18	
Air Blank		0.000	08:19	
Control Test		0.079	08:19	
Air Blank		0.000	08:20	
Subject Sample	# 1		08:20	Subject Test Refused
Breath Volume		0.000	08:20	
Air Blank		0.000	08:21	
Control Test		0.079	08:21	
Air Blank		0.000	08:21	
Diagnostics Check		OK	08:21	

Data Download Date/Time: 04/13/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:55

SUBJECT'S LAST NAME: GROSSO

DATE/TIME OF TEST: 04/06/2016 18:23:49

Diagnostics Check		OK	18:16
Air Blank		0.000	18:17
Control Test		0.078	18:17
Air Blank		0.000	18:18
Subject Sample	# 1	0.112	18:19
Breath Volume		1.226	18:19
Air Blank		0.000	18:20
Air Blank		0.000	18:21
Subject Sample	# 2	0.123	18:22
Breath Volume		1.246	18:22
Air Blank		0.000	18:22
Control Test		0.077	18:23
Air Blank		0.000	18:23
Diagnostics Check		OK	18:23

Data Download Date/Time: 04/13/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MAURICE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR,MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/08/2016 02:13:20

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.000	02:09
Breath Volume		2.363	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.000	02:11
Breath Volume		4.078	02:11
Air Blank		0.000	02:12
Control Test		0.080	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 04/13/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOSTALY,STEVENS,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DOSTALY,STEVENS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 04/14/2016 02:01:23

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.096	01:56
Breath Volume		2.429	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.091	01:59
Breath Volume		2.519	01:59
Air Blank		0.000	02:00
Control Test		0.078	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:31

SUBJECT'S LAST NAME: MAILAND

DATE/TIME OF TEST: 04/14/2016 09:57:30

Diagnostics Check		OK	09:51
Air Blank		0.000	09:51
Control Test		0.079	09:51
Air Blank		0.000	09:52
Subject Sample	# 1	0.093	09:52
Breath Volume		4.667	09:52
Air Blank		0.000	09:53
Air Blank		0.000	09:55
Subject Sample	# 2	0.094	09:55
Breath Volume		5.179	09:55
Air Blank		0.000	09:56
Control Test		0.078	09:56
Air Blank		0.000	09:57
Diagnostics Check		OK	09:57

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SANTIBANEZ,D
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 04/16/2016 02:39:45

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.264	02:33
Breath Volume		2.441	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.257	02:38
Breath Volume		1.289	02:38
Air Blank		0.000	02:38
Control Test		0.078	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER,MICHAEL,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: OBERTING

DATE/TIME OF TEST: 04/16/2016 21:15:06

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.077	21:07
Air Blank		0.000	21:07
Subject Sample	# 1	0.135	21:09
Breath Volume		1.222	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.133	21:13
Breath Volume		1.207	21:13
Air Blank		0.000	21:14
Control Test		0.078	21:14
Air Blank		0.000	21:14
Diagnostics Check		OK	21:15

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGER,WAYNE,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CRAGIN,STEPHEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 04/17/2016 22:27:30

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.078	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.000	22:22
Breath Volume		2.468	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.000	22:26
Breath Volume		2.328	22:26
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 05/17/2016 13:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: VEGA,ROBERT
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: EIKE

DATE/TIME OF TEST: 04/17/2016 22:41:19

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.078	22:35
Air Blank		0.000	22:36
Subject Sample	# 1	0.251	22:36
Breath Volume		1.144	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.251	22:39
Breath Volume		1.347	22:39
Air Blank		0.000	22:40
Control Test		0.078	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN,ANTHONY,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROSADO,M
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 04/20/2016 02:19:20

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.216	02:14
Breath Volume		3.054	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.216	02:17
Breath Volume		2.804	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 05/17/2016 13:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGER,WAYNE,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HERON,LAUREN
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: MCGLYNN

DATE/TIME OF TEST: 04/22/2016 00:31:24

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.078	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.090	00:26
Breath Volume		1.457	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.088	00:29
Breath Volume		1.433	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: ENDRES

DATE/TIME OF TEST: 04/22/2016 21:04:39

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.107	20:59
Breath Volume		2.382	20:59
Air Blank		0.000	21:00
Air Blank		0.000	21:02
Subject Sample	# 2	0.107	21:03
Breath Volume		2.457	21:03
Air Blank		0.000	21:03
Control Test		0.080	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CATES,CURTIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SARVER

DATE/TIME OF TEST: 04/27/2016 03:29:29

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.079	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.209	03:24
Breath Volume		3.175	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.200	03:27
Breath Volume		2.164	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,NICHOLAS,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,NICHOLAS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 04/29/2016 06:50:53

Diagnostics Check		OK	06:44
Air Blank		0.000	06:44
Control Test		0.079	06:44
Air Blank		0.000	06:45
Subject Sample	# 1	0.096	06:46
Breath Volume		2.210	06:46
Air Blank		0.000	06:47
Air Blank		0.000	06:48
Subject Sample	# 2	0.098	06:49
Breath Volume		1.699	06:49
Air Blank		0.000	06:49
Control Test		0.078	06:50
Air Blank		0.000	06:50
Diagnostics Check		OK	06:50

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMART,DALLAS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FILS

DATE/TIME OF TEST: 04/30/2016 03:43:21

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.159	03:38
Breath Volume		1.800	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.158	03:41
Breath Volume		1.812	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 05/17/2016 13:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEABODY,MICHAEL,F
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TITUS,ROBERT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: CHRISTENSEN

DATE/TIME OF TEST: 04/16/2016 00:31:13

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.218	00:26
Breath Volume		2.296	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.222	00:29
Breath Volume		2.675	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 05/10/2016 16:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEABODY,MICHAEL,F
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOEL,MEGAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 04/16/2016 02:38:31

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.078	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:37	RFI Detect
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.078	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 05/10/2016 16:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEABODY,MICHAEL,F
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOEL,MEGAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 04/16/2016 02:48:40

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.000	02:44
Breath Volume		2.925	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.000	02:47
Breath Volume		2.957	02:47
Air Blank		0.000	02:47
Control Test		0.078	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 05/10/2016 16:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROLIGA,ALEXANDRA,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COLLINS,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: DELIZ

DATE/TIME OF TEST: 04/29/2016 01:10:38

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.078	01:07	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.079	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 05/11/2016 12:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEROSA,ANNAMARIE,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEROSA,ANNAMARIE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: VAQUEZZACARIAS

DATE/TIME OF TEST: 04/29/2016 01:21:02

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.189	01:16
Breath Volume		2.554	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.194	01:19
Breath Volume		2.257	01:19
Air Blank		0.000	01:20
Control Test		0.078	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:21

Data Download Date/Time: 05/11/2016 12:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EGAN,MELISSA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 20:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: RISSLER

DATE/TIME OF TEST: 04/01/2016 21:14:14

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.081	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.180	21:08
Breath Volume		2.480	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.185	21:12
Breath Volume		2.679	21:12
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SWETAVAGE,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: FITZGERALD

DATE/TIME OF TEST: 04/02/2016 01:34:33

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:28
Air Blank		0.000	01:28
Subject Sample	# 1	0.123	01:30
Breath Volume		2.289	01:30
Air Blank		0.000	01:30
Air Blank		0.000	01:32
Subject Sample	# 2	0.123	01:33
Breath Volume		1.968	01:33
Air Blank		0.000	01:33
Control Test		0.081	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LUGO,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 04/02/2016 02:57:31

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.081	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.128	02:53
Breath Volume		1.828	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.135	02:56
Breath Volume		2.496	02:56
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: EGAN,MELISSA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: RISSLER

DATE/TIME OF TEST: 04/02/2016 04:38:16

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.082	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.110	04:33
Breath Volume		2.191	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:35
Subject Sample	# 2	0.109	04:36
Breath Volume		1.980	04:36
Air Blank		0.000	04:37
Control Test		0.082	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMERON,CELESTE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: RISSLER

DATE/TIME OF TEST: 04/02/2016 09:37:39

Diagnostics Check		OK	09:31
Air Blank		0.000	09:32
Control Test		0.081	09:32
Air Blank		0.000	09:32
Subject Sample	# 1	0.055	09:33
Breath Volume		2.468	09:33
Air Blank		0.000	09:33
Air Blank		0.000	09:35
Subject Sample	# 2	0.053	09:36
Breath Volume		1.828	09:36
Air Blank		0.000	09:36
Control Test		0.082	09:37
Air Blank		0.000	09:37
Diagnostics Check		OK	09:37

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMERON,CELESTE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: RISSLER

DATE/TIME OF TEST: 04/02/2016 12:27:43

Diagnostics Check		OK	12:21
Air Blank		0.000	12:22
Control Test		0.082	12:22
Air Blank		0.000	12:23
Subject Sample	# 1	0.030	12:23
Breath Volume		2.441	12:23
Air Blank		0.000	12:24
Air Blank		0.000	12:25
Subject Sample	# 2	0.029	12:26
Breath Volume		2.339	12:26
Air Blank		0.000	12:26
Control Test		0.081	12:27
Air Blank		0.000	12:27
Diagnostics Check		OK	12:27

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTON,JOHN,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE,NICK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: DINGERS

DATE/TIME OF TEST: 04/05/2016 21:51:50

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.082	21:44
Air Blank		0.000	21:45
Subject Sample	# 1	0.117	21:47
Breath Volume		2.175	21:47
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.116	21:50
Breath Volume		2.343	21:50
Air Blank		0.000	21:50
Control Test		0.082	21:51
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LUGO,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: TROXEL

DATE/TIME OF TEST: 04/09/2016 04:32:47

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Subject Sample	# 1	0.152	04:28
Breath Volume		3.261	04:28
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.152	04:31
Breath Volume		3.285	04:31
Air Blank		0.000	04:31
Control Test		0.082	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NOVA,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: MONROY-MORAN

DATE/TIME OF TEST: 04/09/2016 18:30:24

Diagnostics Check		OK	18:22
Air Blank		0.000	18:23
Control Test		0.082	18:23
Air Blank		0.000	18:24
Subject Sample	# 1	0.240	18:25
Breath Volume		2.097	18:25
Air Blank		0.000	18:26
Air Blank		0.000	18:28
Subject Sample	# 2	0.256	18:28
Breath Volume		2.894	18:28
Air Blank		0.000	18:29
Control Test		0.081	18:29
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DOPPELBAUER,CPL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: BLACKSTOCK

DATE/TIME OF TEST: 04/09/2016 21:32:19

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.082	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.000	21:27
Breath Volume		1.105	21:27
Air Blank		0.000	21:27
Air Blank		0.000	21:30
Subject Sample	# 2	0.000	21:30
Breath Volume		1.519	21:30
Air Blank		0.000	21:31
Control Test		0.082	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAMERIS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/10/2016 00:47:05

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:34	
Control Test		0.082	00:35	
Air Blank		0.000	00:35	
Subject Sample	# 1	0.120	00:38	Volume Not Met
Breath Volume		0.472	00:38	
Air Blank		0.000	00:39	
Air Blank		0.000	00:41	
Subject Sample	# 2	0.119	00:43	
Breath Volume		1.585	00:43	
Air Blank		0.000	00:44	
Air Blank			00:45	RFI Detect
Air Blank		0.000	00:46	
Control Test		0.082	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:47	

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MONAGHAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: LOURIDAS

DATE/TIME OF TEST: 04/10/2016 03:14:23

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.082	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.203	03:08
Breath Volume		1.953	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.213	03:12
Breath Volume		2.640	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RENZ,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: TILDEN

DATE/TIME OF TEST: 04/10/2016 21:36:59

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.082	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.263	21:32
Breath Volume		2.277	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.271	21:35
Breath Volume		2.703	21:35
Air Blank		0.000	21:36
Control Test		0.081	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MURRAY,JOHN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 04/13/2016 02:19:12

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.082	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.141	02:14
Breath Volume		2.187	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.131	02:17
Breath Volume		1.679	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SILL,1
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 04/15/2016 21:10:56

Diagnostics Check		OK	21:04
Air Blank		0.000	21:05
Control Test		0.081	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.170	21:06
Breath Volume		2.789	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.170	21:09
Breath Volume		2.691	21:09
Air Blank		0.000	21:10
Control Test		0.081	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FRETIAS,ANTHONY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: CHILES

DATE/TIME OF TEST: 04/16/2016 03:44:30

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.081	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.132	03:33	Volume Not Met
Breath Volume		0.117	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2	0.139	03:37	
Breath Volume		1.222	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 3		03:43	No Sample Provided
Breath Volume		0.000	03:43	
Air Blank		0.000	03:43	
Control Test		0.081	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CARLEY,RICHARD
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: PAWLEY

DATE/TIME OF TEST: 04/16/2016 19:03:54

Diagnostics Check		OK	18:55
Air Blank		0.000	18:56
Control Test		0.082	18:56
Air Blank		0.000	18:57
Subject Sample	# 1	0.268	18:58
Breath Volume		2.031	18:58
Air Blank		0.000	18:59
Air Blank		0.000	19:01
Subject Sample	# 2	0.268	19:02
Breath Volume		2.078	19:02
Air Blank		0.000	19:02
Control Test		0.081	19:03
Air Blank		0.000	19:03
Diagnostics Check		OK	19:03

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LUGO,.
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 04/17/2016 04:56:42

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.082	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.118	04:50
Breath Volume		1.238	04:50
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 2	0.119	04:55
Breath Volume		1.480	04:55
Air Blank		0.000	04:55
Control Test		0.082	04:56
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 05/27/2016 11:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANGELINI,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: ANDERSSEN

DATE/TIME OF TEST: 04/18/2016 18:47:40

Diagnostics Check		OK	18:40
Air Blank		0.000	18:41
Control Test		0.082	18:41
Air Blank		0.000	18:42
Subject Sample	# 1	0.214	18:43
Breath Volume		2.781	18:43
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.216	18:46
Breath Volume		2.605	18:46
Air Blank		0.000	18:46
Control Test		0.081	18:47
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 05/27/2016 11:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COOPER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: BUCHACHER

DATE/TIME OF TEST: 04/18/2016 20:08:19

Diagnostics Check		OK	20:02
Air Blank		0.000	20:02
Control Test		0.082	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.139	20:03
Breath Volume		2.007	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.144	20:06
Breath Volume		2.074	20:06
Air Blank		0.000	20:07
Control Test		0.082	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GALLEGO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 04/18/2016 23:13:29

Diagnostics Check		OK	23:06
Air Blank		0.000	23:07
Control Test		0.082	23:07
Air Blank		0.000	23:08
Subject Sample	# 1	0.198	23:08
Breath Volume		2.867	23:08
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 2	0.196	23:12
Breath Volume		2.890	23:12
Air Blank		0.000	23:12
Control Test		0.082	23:12
Air Blank		0.000	23:13
Diagnostics Check		OK	23:13

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TAMERIS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 04/19/2016 23:50:08

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.082	23:44
Air Blank		0.000	23:45
Subject Sample	# 1	0.196	23:45
Breath Volume		3.410	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:47
Subject Sample	# 2	0.190	23:48
Breath Volume		3.308	23:48
Air Blank		0.000	23:49
Control Test		0.082	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREITAS,ADAM
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GLUCKSMAN

DATE/TIME OF TEST: 04/22/2016 02:24:18

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.082	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.280	02:16
Breath Volume		3.000	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:18
Subject Sample	# 2	0.305	02:19
Breath Volume		2.277	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 3	0.298	02:22
Breath Volume		3.910	02:22
Air Blank		0.000	02:23
Control Test		0.082	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRAVEL,CPL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HEITTER

DATE/TIME OF TEST: 04/23/2016 03:38:41

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.081	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.055	03:33
Breath Volume		2.273	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.051	03:37
Breath Volume		1.914	03:37
Air Blank		0.000	03:37
Control Test		0.082	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LOTTHAMMER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: KLEBS JR

DATE/TIME OF TEST: 04/23/2016 18:54:32

Diagnostics Check		OK	18:47
Air Blank		0.000	18:47
Control Test		0.082	18:47
Air Blank		0.000	18:48
Subject Sample	# 1	0.211	18:49
Breath Volume		2.164	18:49
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.203	18:53
Breath Volume		2.312	18:53
Air Blank		0.000	18:53
Control Test		0.082	18:53
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COOPER,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 04/24/2016 03:19:46

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.082	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.179	03:15
Breath Volume		1.976	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.173	03:18
Breath Volume		1.957	03:18
Air Blank		0.000	03:18
Control Test		0.081	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRAVEL,CPL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 04/24/2016 22:17:20

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.081	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.000	22:13
Breath Volume		1.234	22:13
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.000	22:15
Breath Volume		1.171	22:15
Air Blank		0.000	22:16
Control Test		0.081	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 05/27/2016 11:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: DAVIS,TIMMY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 04/02/2016 20:45:21

Diagnostics Check		OK	20:38
Air Blank		0.000	20:39
Control Test		0.083	20:39
Air Blank		0.000	20:39
Subject Sample	# 1	0.316	20:40
Breath Volume		2.050	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.324	20:43
Breath Volume		2.253	20:43
Air Blank		0.000	20:44
Control Test		0.078	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 04/21/2016 16:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAKER,RUEL,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MILLER,RYAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: BUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 04/10/2016 18:13:05

Diagnostics Check		OK	18:06
Air Blank		0.000	18:06
Control Test		0.084	18:07
Air Blank		0.000	18:07
Subject Sample	# 1	0.185	18:08
Breath Volume		3.507	18:08
Air Blank		0.000	18:09
Air Blank		0.000	18:10
Subject Sample	# 2	0.188	18:11
Breath Volume		3.335	18:11
Air Blank		0.000	18:12
Control Test		0.083	18:12
Air Blank		0.000	18:13
Diagnostics Check		OK	18:13

Data Download Date/Time: 04/21/2016 16:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,H
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAMM,JAMES
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 04/16/2016 02:17:11

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.082	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.228	02:12
Breath Volume		3.351	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.232	02:15
Breath Volume		3.335	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 04/21/2016 16:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TEMPLIN,RICHARD,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TEMPLIN,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: FRANKS

DATE/TIME OF TEST: 04/17/2016 00:11:08

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.082	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.195	00:06
Breath Volume		1.519	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.193	00:09
Breath Volume		2.003	00:09
Air Blank		0.000	00:10
Control Test		0.082	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 04/21/2016 16:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAM,RICHARD,A
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: TAYLOR,UPTON
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 03/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: CHAVOUS

DATE/TIME OF TEST: 04/10/2016 01:46:44

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.080	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.223	01:42
Breath Volume		2.792	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.215	01:45
Breath Volume		2.347	01:45
Air Blank		0.000	01:45
Control Test		0.079	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 04/21/2016 16:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,N,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: WILLIAMS,N
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 04/16/2016 23:48:40

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.077	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:47	
Control Test		0.077	23:48	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 04/22/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,N,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: LAROSA,R
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HENDRY

DATE/TIME OF TEST: 04/22/2016 03:31:32

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:23	
Control Test		0.076	03:23	
Air Blank		0.000	03:23	
Subject Sample	# 1	0.234	03:25	
Breath Volume		1.941	03:25	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 2	0.218	03:29	
Breath Volume		1.546	03:29	
Air Blank		0.000	03:30	
Control Test		0.072	03:30	Control Outside Tolerance
Air Blank		0.000	03:31	

Data Download Date/Time: 04/22/2016 15:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,N,O
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: WILLIAMS,N
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: SYS CHECK
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/22/2016 04:54:42

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.075	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 1	0.000	04:50	
Breath Volume		3.523	04:50	
Air Blank		0.000	04:50	
Air Blank		0.000	04:52	
Subject Sample	# 2	0.000	04:53	
Breath Volume		3.675	04:53	
Air Blank		0.000	04:53	
Control Test		0.074	04:54	Control Outside Tolerance
Air Blank		0.000	04:54	

Data Download Date/Time: 04/22/2016 15:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILIAMS,N,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: WILLIAMS,N
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: SYS CHECK
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/29/2016 23:17:12

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.076	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.000	23:12
Breath Volume		2.656	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.000	23:15
Breath Volume		1.542	23:15
Air Blank		0.000	23:16
Control Test		0.077	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 09/29/2016 14:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,N,
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: LAROSA,R
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1737087
DRY GAS EXPIRY DATE: 09/12/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 04/29/2016 23:29:51

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.077	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.073	23:25
Breath Volume		1.351	23:25
Air Blank		0.000	23:26
Air Blank		0.000	23:27
Subject Sample	# 2	0.071	23:28
Breath Volume		1.800	23:28
Air Blank		0.000	23:28
Control Test		0.077	23:29
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 09/29/2016 14:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLANES,GABRIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CAMPER,ALAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 03/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: KICK

DATE/TIME OF TEST: 04/08/2016 06:09:34

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:07	
Control Test		0.079	06:07	
Air Blank		0.000	06:08	
Subject Sample	# 1		06:08	Subject Test Refused
Breath Volume		0.000	06:08	
Air Blank		0.000	06:08	
Control Test		0.078	06:08	
Air Blank		0.000	06:09	
Diagnostics Check		OK	06:09	

Data Download Date/Time: 04/13/2016 10:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLEIGHER,JOHN,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WILCOX,LOGAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/24/2016 19:13:35

Diagnostics Check		OK	19:02	
Air Blank		0.000	19:03	
Control Test		0.077	19:03	
Air Blank		0.000	19:04	
Subject Sample	# 1	0.221	19:07	Volume Not Met
Breath Volume		0.898	19:07	
Air Blank		0.000	19:07	
Air Blank		0.000	19:09	
Subject Sample	# 2	0.214	19:10	
Breath Volume		1.183	19:10	
Air Blank		0.000	19:10	
Air Blank		0.000	19:12	
Subject Sample	# 3	0.208	19:12	
Breath Volume		2.496	19:12	
Air Blank			19:13	RFI Detect
Air Blank		0.000	19:13	

Data Download Date/Time: 05/24/2016 10:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLEIGHER,JOHN,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WILCOX,LOGAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/24/2016 19:28:49

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:18	
Control Test		0.077	19:19	
Air Blank		0.000	19:19	
Subject Sample	# 1	0.216	19:21	
Breath Volume		2.613	19:21	
Air Blank		0.000	19:21	
Air Blank		0.000	19:23	
Subject Sample	# 2	0.190	19:24	
Breath Volume		3.003	19:24	
Air Blank		0.000	19:24	
Air Blank		0.000	19:26	
Subject Sample	# 3	0.203	19:27	
Breath Volume		3.140	19:27	
Air Blank		0.000	19:27	
Control Test		0.073	19:28	Control Outside Tolerance
Air Blank		0.000	19:28	

Data Download Date/Time: 05/24/2016 10:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLEIGHER,JOHN,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: WILCOX,LOGAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/24/2016 19:45:57

Diagnostics Check		OK	19:39
Air Blank		0.000	19:39
Control Test		0.076	19:40
Air Blank		0.000	19:40
Subject Sample	# 1	0.219	19:41
Breath Volume		2.195	19:41
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.229	19:44
Breath Volume		3.445	19:44
Air Blank		0.000	19:45
Control Test		0.076	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:45

Data Download Date/Time: 05/24/2016 10:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOATWRIGHT,MATTHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BOATWRIGHT,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SIPLIN

DATE/TIME OF TEST: 04/26/2016 22:37:34

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:24	
Control Test		0.078	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1	0.205	22:28	Volume Not Met
Breath Volume		0.167	22:28	
Air Blank		0.000	22:28	
Air Blank		0.000	22:30	
Subject Sample	# 2	0.250	22:31	
Breath Volume		1.296	22:31	
Air Blank		0.000	22:32	
Air Blank		0.000	22:34	
Subject Sample	# 3	0.248	22:35	
Breath Volume		1.320	22:35	
Air Blank		0.000	22:36	
Control Test		0.076	22:36	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 05/24/2016 10:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,D,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 04/03/2016 04:25:38

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.081	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.151	04:21
Breath Volume		1.933	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.149	04:24
Breath Volume		1.816	04:24
Air Blank		0.000	04:24
Control Test		0.080	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 04/07/2016 11:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KISER,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 04/08/2016 00:58:51

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.080	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.170	00:54
Breath Volume		1.687	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.158	00:57
Breath Volume		1.125	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: AMISON,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HYDE

DATE/TIME OF TEST: 04/08/2016 01:10:44

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.080	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.039	01:06
Breath Volume		2.667	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.037	01:09
Breath Volume		1.976	01:09
Air Blank		0.000	01:09
Control Test		0.080	01:10
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MOUEN

DATE/TIME OF TEST: 04/09/2016 04:22:42

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.081	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.033	04:17
Breath Volume		1.996	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.032	04:21
Breath Volume		1.125	04:21
Air Blank		0.000	04:21
Control Test		0.081	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,STEPHEN,E
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SIMMONS,STEPHEN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 04/09/2016 21:43:44

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.081	21:37
Air Blank		0.000	21:38
Subject Sample	# 1	0.287	21:39
Breath Volume		1.847	21:39
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.287	21:41
Breath Volume		1.242	21:41
Air Blank		0.000	21:42
Control Test		0.076	21:43
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 04/13/2016 03:32:27

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.171	03:28
Breath Volume		3.125	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.175	03:30
Breath Volume		3.023	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLACK,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: MIMS,CHARLES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: PAULETTE

DATE/TIME OF TEST: 04/16/2016 01:35:46

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 04/18/2016 01:46:31

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.085	01:42
Breath Volume		1.390	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.094	01:44
Breath Volume		1.601	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SMERLING

DATE/TIME OF TEST: 04/18/2016 22:54:52

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.081	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.181	22:50
Breath Volume		1.675	22:50
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.181	22:53
Breath Volume		2.167	22:53
Air Blank		0.000	22:53
Control Test		0.080	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 05/18/2016 12:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CASEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HARRELL,CASEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: MEAD

DATE/TIME OF TEST: 04/21/2016 06:11:37

Diagnostics Check		OK	06:05
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.000	06:06
Breath Volume		1.625	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:09
Subject Sample	# 2	0.000	06:10
Breath Volume		3.136	06:10
Air Blank		0.000	06:10
Control Test		0.080	06:11
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OHARA,K
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 04/22/2016 04:48:07

Diagnostics Check	OK	04:47	
Air Blank		04:47	RFI Detect
Air Blank	0.000	04:48	

Data Download Date/Time: 05/18/2016 12:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OHARA,K
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 04/22/2016 04:59:12

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.081	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.198	04:54
Breath Volume		1.570	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:56
Subject Sample	# 2	0.198	04:57
Breath Volume		2.425	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KISER,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 04/23/2016 01:05:32

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:03	
Control Test		0.080	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:04	
Control Test		0.080	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,D,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 04/24/2016 06:09:26

Diagnostics Check		OK	06:03
Air Blank		0.000	06:03
Control Test		0.080	06:03
Air Blank		0.000	06:04
Subject Sample	# 1	0.122	06:05
Breath Volume		2.574	06:05
Air Blank		0.000	06:05
Air Blank		0.000	06:07
Subject Sample	# 2	0.116	06:07
Breath Volume		1.613	06:07
Air Blank		0.000	06:08
Control Test		0.080	06:08
Air Blank		0.000	06:09
Diagnostics Check		OK	06:09

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,JAMES,H
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PITTMAN,JAMES
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 04/27/2016 21:43:41

Diagnostics Check		OK	21:37
Air Blank		0.000	21:37
Control Test		0.079	21:38
Air Blank		0.000	21:38
Subject Sample	# 1	0.121	21:38
Breath Volume		2.593	21:38
Air Blank		0.000	21:39
Air Blank		0.000	21:41
Subject Sample	# 2	0.122	21:42
Breath Volume		2.656	21:42
Air Blank		0.000	21:42
Control Test		0.079	21:43
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMIDT,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: LAGASSE

DATE/TIME OF TEST: 04/30/2016 02:56:55

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.216	02:52
Breath Volume		2.468	02:52
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.213	02:55
Breath Volume		2.605	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 05/18/2016 12:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ABRAHAMS

DATE/TIME OF TEST: 04/01/2016 02:22:10

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.185	02:17
Breath Volume		1.605	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.184	02:20
Breath Volume		3.640	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 04/07/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CASEY,B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HARRELL,CASEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: GORAM

DATE/TIME OF TEST: 04/01/2016 06:42:57

Diagnostics Check		OK	06:40	
Air Blank		0.000	06:40	
Control Test		0.079	06:40	
Air Blank		0.000	06:41	
Subject Sample	# 1		06:41	Subject Test Refused
Breath Volume		0.000	06:41	
Air Blank		0.000	06:42	
Control Test		0.080	06:42	
Air Blank		0.000	06:42	
Diagnostics Check		OK	06:42	

Data Download Date/Time: 04/07/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,JAMES,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: HARRIS,JAMES
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 04/02/2016 04:04:52

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.162	04:00
Breath Volume		2.023	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.158	04:03
Breath Volume		1.777	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:04
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 04/07/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SAINT LOUIS,JUNIOR
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: POPE

DATE/TIME OF TEST: 04/05/2016 01:25:44

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.253	01:20
Breath Volume		3.984	01:20
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.239	01:23
Breath Volume		3.562	01:23
Air Blank		0.000	01:24
Control Test		0.076	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 04/07/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: PARRAMORE

DATE/TIME OF TEST: 04/07/2016 02:50:59

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.077	02:46
Breath Volume		1.546	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.075	02:49
Breath Volume		1.320	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 04/07/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PLATA

DATE/TIME OF TEST: 04/08/2016 03:34:36

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.176	03:30
Breath Volume		3.335	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.167	03:33
Breath Volume		2.136	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 04/09/2016 03:04:42

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.081	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.140	03:00
Breath Volume		2.964	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.134	03:03
Breath Volume		3.003	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 04/10/2016 02:50:06

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.211	02:45
Breath Volume		1.488	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.201	02:48
Breath Volume		2.246	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 05/18/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: AUERBACH

DATE/TIME OF TEST: 04/10/2016 05:05:49

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.079	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.197	05:01
Breath Volume		3.835	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 2	0.201	05:04
Breath Volume		3.742	05:04
Air Blank		0.000	05:04
Control Test		0.079	05:05
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: COOK,JUDITH
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: ABUTHAMRAH

DATE/TIME OF TEST: 04/10/2016 06:00:26

Diagnostics Check		OK	05:47
Air Blank		0.000	05:47
Control Test		0.079	05:48
Air Blank		0.000	05:48
Subject Sample	# 1	0.244	05:50
Breath Volume		1.886	05:50
Air Blank		0.000	05:51
Air Blank		0.000	05:52
Subject Sample	# 2	0.222	05:55
Breath Volume		2.011	05:55
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 3	0.237	05:58
Breath Volume		1.691	05:58
Air Blank		0.000	05:59
Control Test		0.077	05:59
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 04/12/2016 03:21:42

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.079	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.205	03:16
Breath Volume		1.714	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.195	03:19
Breath Volume		2.070	03:19
Air Blank		0.000	03:20
Control Test		0.078	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,EMILY,P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SEARS,MATT
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/13/2016 01:13:10

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.080	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1	0.102	01:05	Volume Not Met
Breath Volume		0.960	01:05	
Air Blank		0.000	01:05	
Air Blank		0.000	01:07	
Subject Sample	# 2	0.120	01:08	
Breath Volume		2.347	01:08	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 3	0.128	01:11	
Breath Volume		2.921	01:11	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 04/14/2016 02:59:34

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.045	02:55
Breath Volume		1.625	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.045	02:58
Breath Volume		1.480	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 04/16/2016 00:45:13

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.023	00:40
Breath Volume		2.550	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.022	00:43
Breath Volume		1.570	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WRIGHT,S
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: KOUWESHOVEN

DATE/TIME OF TEST: 04/16/2016 02:25:38

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.079	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.224	02:21
Breath Volume		1.980	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.223	02:24
Breath Volume		2.062	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BULLOCK,JOSHUA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 04/17/2016 03:01:59

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.244	02:54
Breath Volume		2.257	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.265	02:57
Breath Volume		1.464	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 3	0.270	03:00
Breath Volume		1.527	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRADLIN,BRETT,J
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SPRADLIN,BRETT
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/21/2016 17:08:59

Diagnostics Check		OK	17:02
Air Blank		0.000	17:03
Control Test		0.079	17:03
Air Blank		0.000	17:03
Subject Sample	# 1	0.163	17:04
Breath Volume		2.179	17:04
Air Blank		0.000	17:05
Air Blank		0.000	17:06
Subject Sample	# 2	0.170	17:07
Breath Volume		2.601	17:07
Air Blank		0.000	17:08
Control Test		0.078	17:08
Air Blank		0.000	17:08
Diagnostics Check		OK	17:08

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 04/22/2016 22:35:25

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.079	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.185	22:30
Breath Volume		1.207	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.190	22:33
Breath Volume		1.593	22:33
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CHITTENDEN

DATE/TIME OF TEST: 04/23/2016 01:51:24

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.079	01:46
Breath Volume		1.179	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.077	01:49
Breath Volume		3.277	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/23/2016 02:54:30

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.110	02:50
Breath Volume		1.976	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.113	02:53
Breath Volume		2.753	02:53
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: MARTINA,CHERRY
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/23/2016 03:25:37

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.217	03:20
Breath Volume		3.171	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.218	03:23
Breath Volume		2.800	03:23
Air Blank		0.000	03:24
Control Test		0.077	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HOOD

DATE/TIME OF TEST: 04/24/2016 02:07:19

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.080	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.087	02:02
Breath Volume		3.207	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.085	02:05
Breath Volume		3.320	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,C
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BINGER

DATE/TIME OF TEST: 04/24/2016 04:02:53

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.140	03:58
Breath Volume		2.015	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:00
Subject Sample	# 2	0.141	04:01
Breath Volume		2.035	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/28/2016 00:32:30

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.105	00:27
Breath Volume		3.750	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.102	00:30
Breath Volume		3.410	00:30
Air Blank		0.000	00:31
Control Test		0.078	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 05/18/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: WESTOVER

DATE/TIME OF TEST: 04/28/2016 01:34:11

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.078	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.172	01:28
Breath Volume		1.332	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.167	01:32
Breath Volume		1.101	01:32
Air Blank		0.000	01:33
Control Test		0.078	01:33
Air Blank		0.000	01:34
Diagnostics Check		OK	01:34

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEGRAW,JACK,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: MARTINA,CHERRY
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/29/2016 02:01:47

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.079	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.178	01:57
Breath Volume		2.175	01:57
Air Blank		0.000	01:58
Air Blank		0.000	01:59
Subject Sample	# 2	0.182	02:00
Breath Volume		1.839	02:00
Air Blank		0.000	02:00
Control Test		0.078	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 05/18/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BAKER,MARK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 04/30/2016 18:47:35

Diagnostics Check		OK	18:41
Air Blank		0.000	18:41
Control Test		0.079	18:41
Air Blank		0.000	18:42
Subject Sample	# 1	0.144	18:42
Breath Volume		3.187	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.142	18:46
Breath Volume		3.300	18:46
Air Blank		0.000	18:46
Control Test		0.079	18:46
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 05/18/2016 12:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,VICKI,N
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: HANKS,KYLE
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 04/01/2016 22:47:17

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.078	22:38
Air Blank		0.000	22:39
Subject Sample	# 1	0.235	22:39
Breath Volume		3.601	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.202	22:42
Breath Volume		2.011	22:42
Air Blank		0.000	22:43
Air Blank		0.000	22:45
Subject Sample	# 3	0.209	22:45
Breath Volume		3.859	22:45
Air Blank		0.000	22:46
Control Test		0.078	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 04/05/2016 08:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: PELLMAN

DATE/TIME OF TEST: 04/04/2016 04:02:55

Diagnostics Check		OK	03:59	
Air Blank		0.000	04:00	
Control Test		0.079	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:01	Subject Test Refused
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Control Test		0.078	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 04/05/2016 08:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 04/05/2016 03:29:39

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.077	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.228	03:25
Breath Volume		2.320	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.227	03:28
Breath Volume		1.687	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 04/05/2016 08:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELL,DENNIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: REVELL,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:21

SUBJECT'S LAST NAME: PILGRIM

DATE/TIME OF TEST: 04/06/2016 11:21:34

Diagnostics Check		OK	11:19	
Air Blank		0.000	11:19	
Control Test		0.080	11:20	
Air Blank		0.000	11:20	
Subject Sample	# 1		11:21	Sequence Aborted
Breath Volume		0.000	11:21	
Air Blank		0.000	11:21	

Data Download Date/Time: 05/19/2016 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REVELL,DENNIS,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: REVELL,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:21

SUBJECT'S LAST NAME: PILGRIM

DATE/TIME OF TEST: 04/06/2016 11:47:55

Diagnostics Check		OK	11:41
Air Blank		0.000	11:41
Control Test		0.079	11:42
Air Blank		0.000	11:42
Subject Sample	# 1	0.197	11:43
Breath Volume		3.054	11:43
Air Blank		0.000	11:44
Air Blank		0.000	11:45
Subject Sample	# 2	0.200	11:46
Breath Volume		3.718	11:46
Air Blank		0.000	11:47
Control Test		0.079	11:47
Air Blank		0.000	11:47
Diagnostics Check		OK	11:47

Data Download Date/Time: 05/19/2016 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: PIGOTT

DATE/TIME OF TEST: 04/09/2016 01:16:20

Diagnostics Check		OK	01:13	
Air Blank		0.000	01:13	
Control Test		0.079	01:14	
Air Blank		0.000	01:14	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:15	
Control Test		0.079	01:15	
Air Blank		0.000	01:16	
Diagnostics Check		OK	01:16	

Data Download Date/Time: 05/19/2016 10:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 04/10/2016 02:46:12

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.000	02:42
Breath Volume		2.964	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.000	02:44
Breath Volume		2.542	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 05/19/2016 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS SR,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS SR,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: BARRY

DATE/TIME OF TEST: 04/10/2016 20:13:10

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:11	
Control Test		0.080	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:12	Sequence Aborted
Breath Volume		0.000	20:12	
Air Blank		0.000	20:13	

Data Download Date/Time: 05/19/2016 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SKATULER

DATE/TIME OF TEST: 04/13/2016 01:22:12

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.221	01:17
Breath Volume		3.199	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.238	01:20
Breath Volume		3.531	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 05/19/2016 10:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLEBROOKS,ALAN,W
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: MATHEUS,LEESHENDER
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: DALESSANDRO

DATE/TIME OF TEST: 04/16/2016 00:49:15

Diagnostics Check		OK	00:46	
Air Blank		0.000	00:46	
Control Test		0.079	00:46	
Air Blank		0.000	00:47	
Subject Sample	# 1		00:47	Subject Test Refused
Breath Volume		0.000	00:47	
Air Blank		0.000	00:48	
Control Test		0.079	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 05/19/2016 10:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIEN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BRIEN,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: KETZENBERG

DATE/TIME OF TEST: 04/18/2016 19:28:04

Diagnostics Check		OK	19:22	
Air Blank		0.000	19:22	
Control Test		0.079	19:23	
Air Blank		0.000	19:23	
Subject Sample	# 1	0.117	19:24	
Breath Volume		1.886	19:24	
Air Blank		0.000	19:25	
Air Blank			19:26	RFI Detect
Air Blank		0.000	19:27	
Control Test		0.081	19:27	
Air Blank		0.000	19:27	
Diagnostics Check		OK	19:28	

Data Download Date/Time: 05/19/2016 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIEN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BRIEN,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: KETZENBERG

DATE/TIME OF TEST: 04/18/2016 19:37:48

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.080	19:32
Air Blank		0.000	19:32
Subject Sample	# 1	0.120	19:33
Breath Volume		2.738	19:33
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.113	19:36
Breath Volume		2.531	19:36
Air Blank		0.000	19:36
Control Test		0.080	19:37
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 05/19/2016 10:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DAILEY

DATE/TIME OF TEST: 04/19/2016 01:42:22

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.079	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.173	01:34
Breath Volume		1.238	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.195	01:37
Breath Volume		1.375	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 3	0.191	01:40
Breath Volume		2.445	01:40
Air Blank		0.000	01:41
Control Test		0.077	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 05/19/2016 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 04/22/2016 00:54:30

Diagnostics Check		OK	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.185	00:50
Breath Volume		2.019	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.180	00:52
Breath Volume		2.312	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 05/19/2016 10:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,ASHLEY,N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,ASHLEY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 04/24/2016 00:42:14

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.078	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:41	
Control Test		0.078	00:41	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 05/19/2016 10:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN,MITCHELL,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN,MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 04/03/2016 02:28:32

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.188	02:23
Breath Volume		2.277	02:23
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.195	02:26
Breath Volume		2.929	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/04/2016 00:19:45

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.174	00:15
Breath Volume		2.386	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.181	00:18
Breath Volume		1.996	00:18
Air Blank		0.000	00:18
Control Test		0.078	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 04/04/2016 04:25:57

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.177	04:20
Breath Volume		1.246	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.182	04:24
Breath Volume		1.691	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 04/04/2016 23:12:27

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.079	23:09	
Air Blank		0.000	23:10	
Subject Sample	# 1		23:11	RFI Detect
Breath Volume		0.000	23:11	
Air Blank		0.000	23:11	
Control Test		0.080	23:11	
Air Blank		0.000	23:12	
Diagnostics Check		OK	23:12	

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 04/04/2016 23:20:29

Diagnostics Check		OK	23:14
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.163	23:15
Breath Volume		1.542	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.158	23:18
Breath Volume		1.578	23:18
Air Blank		0.000	23:19
Control Test		0.079	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN,MITCHELL,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN,MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/07/2016 04:21:39

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.079	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1	0.209	04:20	Volume Not Met
Breath Volume		0.085	04:20	
Air Blank			04:21	Purge Fail
Air Blank		0.000	04:21	

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN,MITCHELL,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN,MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 04/07/2016 04:38:45

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:25	
Control Test		0.077	04:25	
Air Blank		0.000	04:26	
Subject Sample	# 1	0.219	04:29	Volume Not Met
Breath Volume		0.105	04:29	
Air Blank		0.000	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 2	0.213	04:33	
Breath Volume		1.421	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 3	0.217	04:37	
Breath Volume		1.492	04:37	
Air Blank		0.000	04:37	
Control Test		0.078	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BRYD

DATE/TIME OF TEST: 04/08/2016 01:14:22

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.079	01:08
Air Blank		0.000	01:08
Subject Sample	# 1	0.105	01:09
Breath Volume		1.769	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:11
Subject Sample	# 2	0.105	01:12
Breath Volume		1.847	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 04/10/2016 04:09:01

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.196	04:04
Breath Volume		1.554	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.200	04:07
Breath Volume		2.144	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:09

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: JOYNER

DATE/TIME OF TEST: 04/12/2016 22:28:41

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:23	
Control Test		0.079	22:23	
Air Blank		0.000	22:24	
Subject Sample	# 1		22:27	Subject Test Refused
Breath Volume		0.000	22:27	
Air Blank		0.000	22:27	
Control Test		0.080	22:28	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 04/14/2016 09:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPANGLER II,GARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SPANGLER II,GARY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/15/2016 05:06:34

Diagnostics Check		OK	04:59	
Air Blank		0.000	04:59	
Control Test		0.079	05:00	
Air Blank		0.000	05:00	
Subject Sample	# 1	0.183	05:02	
Breath Volume		1.273	05:02	
Air Blank		0.000	05:03	
Air Blank		0.000	05:04	
Subject Sample	# 2	0.163	05:05	
Breath Volume		1.316	05:05	
Air Blank			05:06	RFI Detect
Air Blank		0.000	05:06	

Data Download Date/Time: 05/12/2016 12:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN,MITCHELL,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN,MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 04/17/2016 03:46:05

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.109	03:41
Breath Volume		1.210	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.122	03:44
Breath Volume		1.160	03:44
Air Blank		0.000	03:45
Control Test		0.078	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 05/12/2016 12:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/22/2016 04:07:12

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.078	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.179	04:02
Breath Volume		2.195	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.192	04:05
Breath Volume		1.203	04:05
Air Blank		0.000	04:06
Control Test		0.078	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 05/12/2016 12:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PENDELTON

DATE/TIME OF TEST: 04/23/2016 02:37:48

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.079	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.079	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 05/12/2016 12:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: BARTHELL

DATE/TIME OF TEST: 04/23/2016 04:41:51

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.128	04:37
Breath Volume		2.085	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.123	04:40
Breath Volume		1.253	04:40
Air Blank		0.000	04:40
Control Test		0.078	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 05/12/2016 12:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 04/24/2016 03:50:48

Diagnostics Check		OK	03:46	
Air Blank		0.000	03:46	
Control Test		0.079	03:47	
Air Blank		0.000	03:47	
Subject Sample	# 1	0.178	03:49	
Breath Volume		2.226	03:49	
Air Blank			03:50	Purge Fail
Air Blank		0.000	03:50	

Data Download Date/Time: 05/12/2016 12:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 04/24/2016 03:58:30

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.078	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.172	03:53
Breath Volume		1.878	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.177	03:56
Breath Volume		1.550	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 05/12/2016 12:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CHARLES,E
BREATH TEST OPERATOR AGENCY: Gretna PD
ARRESTING OFFICER: BROWN,CHARLES
ARRESTING OFFICER AGENCY: Gretna PD
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 04/28/2016 00:20:46

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.078	00:19	
Air Blank	0.000	00:20	
		00:20	RFI Detect
Air Blank	0.000	00:20	

Data Download Date/Time: 05/12/2016 12:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,CHARLES,E
BREATH TEST OPERATOR AGENCY: Gretna PD
ARRESTING OFFICER: BROWN,CHARLES
ARRESTING OFFICER AGENCY: Gretna PD
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 04/28/2016 00:31:46

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.078	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.159	00:27
Breath Volume		2.195	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.164	00:30
Breath Volume		1.488	00:30
Air Blank		0.000	00:30
Control Test		0.077	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 05/12/2016 12:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ALEJO

DATE/TIME OF TEST: 04/28/2016 03:56:06

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.079	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.162	03:51
Breath Volume		1.679	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.156	03:54
Breath Volume		3.046	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 05/12/2016 12:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHRITENOUR,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LEDESMA,LEOPOLDO
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 04/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:17

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 04/29/2016 09:44:06

Diagnostics Check		OK	09:37
Air Blank		0.000	09:37
Control Test		0.079	09:38
Air Blank		0.000	09:38
Subject Sample	# 1	0.000	09:39
Breath Volume		2.925	09:39
Air Blank		0.000	09:40
Air Blank		0.000	09:42
Subject Sample	# 2	0.000	09:42
Breath Volume		3.042	09:42
Air Blank		0.000	09:43
Control Test		0.080	09:43
Air Blank		0.000	09:44
Diagnostics Check		OK	09:44

Data Download Date/Time: 05/12/2016 12:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: AMAYA-AMAYA

DATE/TIME OF TEST: 04/02/2016 03:20:34

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.207	03:15
Breath Volume		1.808	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.211	03:18
Breath Volume		2.523	03:18
Air Blank		0.000	03:19
Control Test		0.077	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 04/10/2016 13:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 04/05/2016 05:54:05

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.081	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.000	05:49
Breath Volume		2.046	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.000	05:52
Breath Volume		2.312	05:52
Air Blank		0.000	05:53
Control Test		0.079	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 04/10/2016 13:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEHL,EDWARD,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DIEHL,EDWARD
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 14:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:43

SUBJECT'S LAST NAME: OLMO

DATE/TIME OF TEST: 04/06/2016 16:13:54

Diagnostics Check		OK	16:06
Air Blank		0.000	16:07
Control Test		0.081	16:07
Air Blank		0.000	16:08
Subject Sample	# 1	0.039	16:09
Breath Volume		1.476	16:09
Air Blank		0.000	16:10
Air Blank		0.000	16:11
Subject Sample	# 2	0.038	16:12
Breath Volume		1.882	16:12
Air Blank		0.000	16:12
Control Test		0.081	16:13
Air Blank		0.000	16:13
Diagnostics Check		OK	16:13

Data Download Date/Time: 04/10/2016 13:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTHER,PAUL,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BRIGHT

DATE/TIME OF TEST: 04/08/2016 02:28:22

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.081	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.137	02:22
Breath Volume		2.046	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.126	02:26
Breath Volume		1.371	02:26
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 04/10/2016 13:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,AYLER,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CRUZ,AYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CASTRO HERNANDEZ

DATE/TIME OF TEST: 04/10/2016 02:23:02

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.081	02:12	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Air Blank		0.000	02:18	
Subject Sample	# 2		02:21	No Sample Provided
Breath Volume		0.000	02:21	
Air Blank		0.000	02:22	
Control Test		0.081	02:22	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 04/10/2016 13:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,J,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RAMIREZ,JC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KISSAM III

DATE/TIME OF TEST: 04/11/2016 03:47:05

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.000	03:42
Breath Volume		4.050	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.000	03:45
Breath Volume		4.406	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:47

Data Download Date/Time: 05/05/2016 17:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: BEREZNOFF

DATE/TIME OF TEST: 04/11/2016 20:17:37

Diagnostics Check		OK	20:11
Air Blank		0.000	20:11
Control Test		0.081	20:12
Air Blank		0.000	20:12
Subject Sample	# 1	0.216	20:12
Breath Volume		2.000	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.207	20:16
Breath Volume		2.117	20:16
Air Blank		0.000	20:16
Control Test		0.079	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 05/05/2016 17:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,AYLER,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CRUZ,AYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: LARA LOAIZA

DATE/TIME OF TEST: 04/13/2016 22:27:10

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.081	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.293	22:22
Breath Volume		1.660	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.292	22:25
Breath Volume		1.121	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 05/05/2016 17:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CRESPO

DATE/TIME OF TEST: 04/16/2016 23:56:21

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.082	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:52
Breath Volume		3.449	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		3.296	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 05/05/2016 17:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES-RAMOS,EDGARDO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: TORRES-RAMOS,EDGARDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 10:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: MATOS SOTO

DATE/TIME OF TEST: 04/17/2016 12:03:04

Diagnostics Check		OK	11:50	
Air Blank		0.000	11:51	
Control Test		0.082	11:51	
Air Blank		0.000	11:51	
Subject Sample	# 1	0.133	11:55	Volume Not Met
Breath Volume		0.539	11:55	
Air Blank		0.000	11:56	
Air Blank		0.000	11:57	
Subject Sample	# 2	0.152	11:58	
Breath Volume		1.875	11:58	
Air Blank		0.000	11:59	
Air Blank		0.000	12:00	
Subject Sample	# 3	0.147	12:01	
Breath Volume		1.750	12:01	
Air Blank		0.000	12:02	
Control Test		0.079	12:02	
Air Blank		0.000	12:02	
Diagnostics Check		OK	12:03	

Data Download Date/Time: 05/05/2016 17:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLLIER,THOMAS,L
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: MOLESTINA

DATE/TIME OF TEST: 04/18/2016 06:04:16

Diagnostics Check		OK	05:58
Air Blank		0.000	05:58
Control Test		0.082	05:59
Air Blank		0.000	05:59
Subject Sample	# 1	0.000	06:00
Breath Volume		1.882	06:00
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 2	0.000	06:02
Breath Volume		1.929	06:02
Air Blank		0.000	06:03
Control Test		0.080	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 05/05/2016 17:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTRERAS,BRYANT,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CONTRERAS,BRYANT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: PESA

DATE/TIME OF TEST: 04/19/2016 17:51:22

Diagnostics Check		OK	17:45	
Air Blank		0.000	17:45	
Control Test		0.081	17:46	
Air Blank		0.000	17:46	
Subject Sample	# 1	0.161	17:47	
Breath Volume		3.843	17:47	
Air Blank		0.000	17:48	
Air Blank		0.000	17:49	
Subject Sample	# 2	0.159	17:50	
Breath Volume		3.289	17:50	
Air Blank			17:50	RFI Detect
Air Blank		0.000	17:51	

Data Download Date/Time: 05/05/2016 17:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVILA,FABIO,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CRAWFORD,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 04/24/2016 01:38:03

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.082	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.160	01:33
Breath Volume		2.609	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:35
Subject Sample	# 2	0.162	01:36
Breath Volume		1.128	01:36
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:38

Data Download Date/Time: 05/05/2016 17:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEHL,EDWARD,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: STINSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SANCHEZ-MORALES

DATE/TIME OF TEST: 04/24/2016 01:52:56

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.081	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.057	01:48
Breath Volume		2.156	01:48
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.058	01:51
Breath Volume		2.402	01:51
Air Blank		0.000	01:52
Control Test		0.081	01:52
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 05/05/2016 17:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,ANGEL,M
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DIAZ,ANGEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: CANDELARIO

DATE/TIME OF TEST: 04/24/2016 09:07:24

Diagnostics Check		OK	09:00
Air Blank		0.000	09:01
Control Test		0.079	09:01
Air Blank		0.000	09:01
Subject Sample	# 1	0.191	09:02
Breath Volume		1.515	09:02
Air Blank		0.000	09:03
Air Blank		0.000	09:04
Subject Sample	# 2	0.191	09:05
Breath Volume		1.527	09:05
Air Blank		0.000	09:06
Control Test		0.079	09:06
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 05/05/2016 17:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOZNICK,ROBERT,J
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: KLOZNICK,ROBERT
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: PIERZO

DATE/TIME OF TEST: 04/29/2016 05:14:38

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.183	05:10
Breath Volume		1.144	05:10
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.178	05:12
Breath Volume		1.136	05:12
Air Blank		0.000	05:13
Control Test		0.077	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 05/05/2016 17:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ROBBIE,E
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: ESH,JOSHUA
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BERGET

DATE/TIME OF TEST: 04/29/2016 19:41:13

Diagnostics Check	OK	19:39	
Air Blank	0.000	19:40	
Control Test	0.026	19:40	Control Outside Tolerance
Air Blank	0.000	19:41	

Data Download Date/Time: 05/05/2016 17:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ROBBIE,E
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: ESH,JOSHUA
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BERGET

DATE/TIME OF TEST: 04/29/2016 19:44:45

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:43	
Control Test	0.000	19:44	Control Outside Tolerance
Air Blank	0.000	19:44	

Data Download Date/Time: 05/05/2016 17:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: MARADIAGA HERNANDEZ

DATE/TIME OF TEST: 04/01/2016 23:05:29

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.077	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.156	23:00
Breath Volume		1.156	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.156	23:03
Breath Volume		1.308	23:03
Air Blank		0.000	23:04
Control Test		0.076	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 04/27/2016 18:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 04/02/2016 01:20:10

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.077	01:12	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:16	No Sample Provided
Breath Volume		0.000	01:16	
Air Blank		0.000	01:16	
Air Blank		0.000	01:18	
Subject Sample	# 2		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:19	
Control Test		0.077	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 04/27/2016 18:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,ASHLEY,A
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CRUTCHER,T
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: STARBUCK

DATE/TIME OF TEST: 04/14/2016 18:15:43

Diagnostics Check		OK	18:09
Air Blank		0.000	18:09
Control Test		0.077	18:10
Air Blank		0.000	18:10
Subject Sample	# 1	0.311	18:11
Breath Volume		2.730	18:11
Air Blank		0.000	18:11
Air Blank		0.000	18:13
Subject Sample	# 2	0.317	18:14
Breath Volume		2.566	18:14
Air Blank		0.000	18:14
Control Test		0.077	18:15
Air Blank		0.000	18:15
Diagnostics Check		OK	18:15

Data Download Date/Time: 04/27/2016 18:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELEZ,H,
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: VELEZ,H
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 04/17/2016 20:56:04

Diagnostics Check		OK	20:53	
Air Blank		0.000	20:53	
Control Test		0.077	20:53	
Air Blank		0.000	20:54	
Subject Sample	# 1		20:54	Subject Test Refused
Breath Volume		0.000	20:54	
Air Blank		0.000	20:55	
Control Test		0.078	20:55	
Air Blank		0.000	20:55	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 04/27/2016 18:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VELEZ,H,
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: VELEZ,H
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 03/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: HOUMES KINCAID

DATE/TIME OF TEST: 04/24/2016 18:49:30

Diagnostics Check		OK	18:43
Air Blank		0.000	18:43
Control Test		0.078	18:44
Air Blank		0.000	18:44
Subject Sample	# 1	0.082	18:45
Breath Volume		1.941	18:45
Air Blank		0.000	18:45
Air Blank		0.000	18:47
Subject Sample	# 2	0.083	18:47
Breath Volume		1.546	18:47
Air Blank		0.000	18:48
Control Test		0.078	18:48
Air Blank		0.000	18:49
Diagnostics Check		OK	18:49

Data Download Date/Time: 04/27/2016 18:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLERIA,TROY,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: TELLERIA,TROY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 03/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 04/04/2016 21:06:35

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.081	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.317	21:01
Breath Volume		3.171	21:01
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.317	21:04
Breath Volume		3.250	21:04
Air Blank		0.000	21:05
Control Test		0.081	21:05
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 04/10/2016 13:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,C
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: RAITH

DATE/TIME OF TEST: 04/12/2016 03:13:14

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.082	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.152	03:08
Breath Volume		1.816	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.151	03:11
Breath Volume		1.527	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 05/05/2016 16:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLERIA,TROY,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: TELLERIA,TROY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 04/13/2016 04:14:41

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.082	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2		04:13	No Sample Provided
Breath Volume		0.000	04:13	
Air Blank		0.000	04:13	
Control Test		0.081	04:14	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 05/05/2016 16:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINET,MARLON,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 04/21/2016 04:23:36

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.081	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.160	04:19
Breath Volume		4.054	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.161	04:22
Breath Volume		4.207	04:22
Air Blank		0.000	04:22
Control Test		0.081	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 05/05/2016 16:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ROBBIE,E
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: ESH,JOSHUA
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: BERGET

DATE/TIME OF TEST: 04/29/2016 19:58:59

Diagnostics Check		OK	19:49
Air Blank		0.000	19:50
Control Test		0.081	19:50
Air Blank		0.000	19:50
Subject Sample	# 1	0.202	19:52
Breath Volume		1.402	19:52
Air Blank		0.000	19:53
Air Blank		0.000	19:54
Subject Sample	# 2	0.183	19:57
Breath Volume		2.125	19:57
Air Blank		0.000	19:58
Control Test		0.081	19:58
Air Blank		0.000	19:58
Diagnostics Check		OK	19:58

Data Download Date/Time: 05/05/2016 16:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.