Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMAN,JOHN,S
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRAMAN,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 05/02/2016 16:36:02

Diagnostics Check		OK	16:29
Air Blank		0.000	16:29
Control Test		0.081	16:29
Air Blank		0.000	16:30
Subject Sample	# 1	0.128	16:30
Breath Volume		1.433	16:30
Air Blank		0.000	16:31
Air Blank		0.000	16:33
Subject Sample	# 2	0.140	16:34
Breath Volume		1.710	16:34
Air Blank		0.000	16:35
Control Test		0.082	16:35
Air Blank		0.000	16:35
Diagnostics Check		OK	16:36

Data Download Date/Time: 05/06/2016 14:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MARSH

DATE/TIME OF TEST: 05/04/2016 04:40:06

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.081	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.000	04:35
Breath Volume		3.394	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.000	04:38
Breath Volume		2.378	04:38
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 05/06/2016 14:31

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,RONALD,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WINGARD,JEFF
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HARR

DATE/TIME OF TEST: 05/27/2016 22:52:17

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.080	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.230	22:45
Breath Volume		2.917	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.216	22:50
Breath Volume		1.976	22:50
Air Blank		0.000	22:51
Control Test		0.080	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 06/20/2016 17:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACKEL,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: YACKEL,BRAD
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SCANO

DATE/TIME OF TEST: 05/27/2016 23:45:42

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.213	23:40
Breath Volume		1.542	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.210	23:44
Breath Volume		2.121	23:44
Air Blank		0.000	23:44
Control Test		0.080	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 06/20/2016 17:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,RONALD,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WINGARD,JEFF
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 05/28/2016 23:32:19

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:26
Subject Sample	# 1	0.071	23:27
Breath Volume		2.160	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.070	23:30
Breath Volume		2.082	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 06/20/2016 17:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,RONALD,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TATUM,THOMAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: STIGLER

DATE/TIME OF TEST: 05/30/2016 01:23:11

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.195	01:18
Breath Volume		1.589	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.198	01:21
Breath Volume		1.585	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 06/20/2016 17:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACKEL,BRAD,K
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GRAY,DONALD
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 05/30/2016 01:36:25

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.079	01:26	
Air Blank		0.000	01:26	
Subject Sample	# 1	0.205	01:30	Volume Not Met
Breath Volume		0.839	01:30	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2		01:34	Subject Test Refused
Breath Volume		0.117	01:34	
Air Blank		0.000	01:35	
Control Test		0.079	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 06/20/2016 17:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDANIEL,RONALD,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ISAACS

DATE/TIME OF TEST: 05/30/2016 02:48:44

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.190	02:43
Breath Volume		1.910	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.185	02:47
Breath Volume		1.664	02:47
Air Blank		0.000	02:47
Control Test		0.078	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 06/20/2016 17:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRSHIKOVA,ALINA,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: DATSKO,NATHAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/01/2016 01:48:31

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.082	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.143	01:43
Breath Volume		2.804	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.142	01:47
Breath Volume		2.710	01:47
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 05/31/2016 09:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELOIT,BRIAN,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: FLOYD,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CRAFTON

DATE/TIME OF TEST: 05/02/2016 00:45:18

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.232	00:40
Breath Volume		2.632	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.226	00:43
Breath Volume		2.820	00:43
Air Blank		0.000	00:44
Control Test		0.081	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 05/31/2016 09:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,ERIN,M
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: ABEL,PAUL
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 09:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/05/2016 11:43:47

Diagnostics Check		OK	11:37
Air Blank		0.000	11:38
Control Test		0.081	11:38
Air Blank		0.000	11:39
Subject Sample	# 1	0.058	11:39
Breath Volume		2.148	11:39
Air Blank		0.000	11:40
Air Blank		0.000	11:41
Subject Sample	# 2	0.055	11:42
Breath Volume		1.800	11:42
Air Blank		0.000	11:42
Control Test		0.080	11:43
Air Blank		0.000	11:43
Diagnostics Check		OK	11:43

Data Download Date/Time: 05/31/2016 09:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUENTES,RAFAEL,
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WILSON,MARK
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/06/2016 04:07:56

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:51	
Control Test		0.081	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1	0.231	03:55	
Breath Volume		2.839	03:55	
Air Blank		0.000	03:55	
Air Blank		0.000	03:57	
Subject Sample	# 2		04:00	No Sample Provided
Breath Volume		0.000	04:00	
Air Blank		0.000	04:01	
Air Blank		0.000	04:03	
Subject Sample	# 3	0.000	04:06	Volume Not Met
Breath Volume		0.234	04:06	
Air Blank		0.000	04:06	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 05/31/2016 09:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMONDS,FRANKIE,D
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: ELLIOTT,D
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: BEAU CHAMP

DATE/TIME OF TEST: 05/16/2016 19:53:41

Diagnostics Check		OK	19:47
Air Blank		0.000	19:48
Control Test		0.081	19:48
Air Blank		0.000	19:48
Subject Sample	# 1	0.224	19:49
Breath Volume		1.796	19:49
Air Blank		0.000	19:50
Air Blank		0.000	19:51
Subject Sample	# 2	0.212	19:52
Breath Volume		1.628	19:52
Air Blank		0.000	19:52
Control Test		0.080	19:53
Air Blank		0.000	19:53
Diagnostics Check		OK	19:53

Data Download Date/Time: 05/31/2016 09:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELOIT,BRIAN,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CUBBEDGE,KYLE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:21

SUBJECT'S LAST NAME: WALLS

DATE/TIME OF TEST: 05/23/2016 16:50:54

Diagnostics Check		OK	16:44
Air Blank		0.000	16:45
Control Test		0.081	16:45
Air Blank		0.000	16:45
Subject Sample	# 1	0.300	16:46
Breath Volume		2.976	16:46
Air Blank		0.000	16:47
Air Blank		0.000	16:48
Subject Sample	# 2	0.282	16:49
Breath Volume		2.027	16:49
Air Blank		0.000	16:50
Control Test		0.081	16:50
Air Blank		0.000	16:50
Diagnostics Check		OK	16:50

Data Download Date/Time: 05/31/2016 09:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: IRVINE

DATE/TIME OF TEST: 05/24/2016 01:47:53

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.081	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.116	01:43
Breath Volume		1.582	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.114	01:46
Breath Volume		1.964	01:46
Air Blank		0.000	01:46
Control Test		0.081	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 05/31/2016 09:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HINES,JESSICA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 10:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: GUZZO

DATE/TIME OF TEST: 05/25/2016 11:47:49

Diagnostics Check		OK	11:41
Air Blank		0.000	11:41
Control Test		0.081	11:42
Air Blank		0.000	11:42
Subject Sample	# 1	0.231	11:42
Breath Volume		1.507	11:42
Air Blank		0.000	11:43
Air Blank		0.000	11:45
Subject Sample	# 2	0.225	11:46
Breath Volume		1.429	11:46
Air Blank		0.000	11:46
Control Test		0.081	11:47
Air Blank		0.000	11:47
Diagnostics Check		OK	11:47

Data Download Date/Time: 05/31/2016 09:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASH,JUSTIN,A
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CASH,JUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/26/2016 22:16:05

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:04	
Control Test		0.081	22:04	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.236	22:08	Volume Not Met
Breath Volume		1.035	22:08	
Air Blank		0.000	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.228	22:11	
Breath Volume		2.082	22:11	
Air Blank		0.000	22:12	
Air Blank		0.000	22:14	
Subject Sample	# 3	0.228	22:14	
Breath Volume		1.800	22:14	
Air Blank		0.000	22:15	
Control Test		0.079	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 05/31/2016 09:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKOSKIE,ADAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BARKOSKIE,ADAM
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 05/27/2016 05:24:30

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.080	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.000	05:20
Breath Volume		3.246	05:20
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.000	05:23
Breath Volume		1.878	05:23
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 05/31/2016 09:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,KEVIN,T
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CASSIDY,CODY
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 15:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MERRIMAN

DATE/TIME OF TEST: 05/17/2016 16:16:52

Diagnostics Check		OK	16:10
Air Blank		0.000	16:10
Control Test		0.080	16:11
Air Blank		0.000	16:11
Subject Sample	# 1	0.253	16:12
Breath Volume		1.546	16:12
Air Blank		0.000	16:13
Air Blank		0.000	16:14
Subject Sample	# 2	0.253	16:15
Breath Volume		1.625	16:15
Air Blank		0.000	16:15
Control Test		0.079	16:16
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 05/29/2016 20:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: PARSONS,TABITHA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: KRADEL

DATE/TIME OF TEST: 05/19/2016 01:03:48

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.129	00:57
Breath Volume		1.406	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.125	01:02
Breath Volume		1.574	01:02
Air Blank		0.000	01:02
Control Test		0.081	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/29/2016 20:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/29/2016 23:08:42

Diagnostics Check		OK	23:02
Air Blank		0.000	23:02
Control Test		0.080	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.235	23:04
Breath Volume		2.472	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.238	23:07
Breath Volume		2.957	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 06/24/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,STERLING,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: REARDON

DATE/TIME OF TEST: 05/30/2016 02:22:33

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.080	02:17
Air Blank		0.000	02:17
Subject Sample	# 1	0.054	02:18
Breath Volume		2.074	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.055	02:21
Breath Volume		2.031	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 06/24/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BICE,BRIAN
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: DE JESUS RODRIGUEZ

DATE/TIME OF TEST: 05/06/2016 12:04:21

Diagnostics Check		OK	12:01	
Air Blank		0.000	12:02	
Control Test		0.082	12:02	
Air Blank		0.000	12:02	
Subject Sample	# 1		12:02	Subject Test Refused
Breath Volume		0.000	12:02	
Air Blank		0.000	12:03	
Control Test		0.082	12:03	
Air Blank		0.000	12:04	
Diagnostics Check		OK	12:04	

Data Download Date/Time: 05/29/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORYANCE,JASON,R
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER,WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 05/13/2016 16:40:03

Diagnostics Check		OK	16:37	
Air Blank		0.000	16:37	
Control Test		0.081	16:37	
Air Blank		0.000	16:38	
Subject Sample	# 1		16:38	RFI Detect
Breath Volume		0.000	16:38	
Air Blank		0.000	16:39	
Control Test		0.081	16:39	
Air Blank			16:39	RFI Detect
Air Blank		0.000	16:40	

Data Download Date/Time: 05/29/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORYANCE,JASON,R
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER,WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 15:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 05/13/2016 16:48:39

Diagnostics Check		OK	16:41
Air Blank		0.000	16:42
Control Test		0.081	16:42
Air Blank		0.000	16:43
Subject Sample	# 1	0.273	16:44
Breath Volume		1.429	16:44
Air Blank		0.000	16:44
Air Blank		0.000	16:46
Subject Sample	# 2	0.262	16:47
Breath Volume		1.585	16:47
Air Blank		0.000	16:47
Control Test		0.080	16:48
Air Blank		0.000	16:48
Diagnostics Check		OK	16:48

Data Download Date/Time: 05/29/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,C,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CASH,R
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: KAISER

DATE/TIME OF TEST: 05/14/2016 22:30:54

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:28	
Control Test		0.081	22:28	
Air Blank		0.000	22:29	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:30	
Control Test		0.081	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 05/29/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOUGERE,MARCEL,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GOLDEN,JOHN
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/21/2016 04:09:51

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:06	
Control Test		0.081	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:08	Subject Test Refused
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Control Test		0.082	04:09	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 05/29/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,C,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WEST,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: NEPAMUCENO

DATE/TIME OF TEST: 05/28/2016 02:09:57

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.081	02:04
Air Blank		0.000	02:05
Subject Sample	# 1	0.187	02:05
Breath Volume		1.785	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.186	02:08
Breath Volume		2.246	02:08
Air Blank		0.000	02:09
Control Test		0.080	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 05/29/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAIN,SHEILA
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: YEROU

DATE/TIME OF TEST: 05/04/2016 06:17:19

Diagnostics Check		OK	06:10
Air Blank		0.000	06:11
Control Test		0.078	06:11
Air Blank		0.000	06:12
Subject Sample	# 1	0.182	06:12
Breath Volume		1.582	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.181	06:15
Breath Volume		2.105	06:15
Air Blank		0.000	06:16
Control Test		0.077	06:16
Air Blank		0.000	06:17
Diagnostics Check		OK	06:17

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HORNE,PHILLIP
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 05/05/2016 21:55:37

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.079	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.270	21:51
Breath Volume		2.031	21:51
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.278	21:54
Breath Volume		2.832	21:54
Air Blank		0.000	21:54
Control Test		0.078	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/06/2016 03:22:51

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.193	03:18
Breath Volume		2.058	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.202	03:21
Breath Volume		2.472	03:21
Air Blank		0.000	03:21
Control Test		0.078	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKINGER,LANI,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MITCHELL,BRAD
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 07:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: MELBEUF

DATE/TIME OF TEST: 05/06/2016 09:23:21

Diagnostics Check		OK	09:10	
Air Blank		0.000	09:11	
Control Test		0.079	09:11	
Air Blank		0.000	09:12	
Subject Sample	# 1	0.211	09:15	Volume Not Met
Breath Volume		1.070	09:15	
Air Blank		0.000	09:15	
Air Blank		0.000	09:17	
Subject Sample	# 2	0.197	09:18	
Breath Volume		1.808	09:18	
Air Blank		0.000	09:19	
Air Blank		0.000	09:21	
Subject Sample	# 3	0.195	09:21	
Breath Volume		1.382	09:21	
Air Blank		0.000	09:22	
Control Test		0.078	09:22	
Air Blank		0.000	09:23	
Diagnostics Check		OK	09:23	

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HERNANDEZ,JONATHAN
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 05/07/2016 00:46:31

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.075	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.207	00:41
Breath Volume		1.187	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.208	00:45
Breath Volume		2.019	00:45
Air Blank		0.000	00:45
Control Test		0.076	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 05/24/2016 16:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: TREAT

DATE/TIME OF TEST: 05/07/2016 04:58:12

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.079	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.195	04:53
Breath Volume		1.375	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.200	04:56
Breath Volume		1.425	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MANNING,MARK
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: GONZALEZ CRUZ

DATE/TIME OF TEST: 05/07/2016 05:36:10

Diagnostics Check		OK	05:27	
Air Blank		0.000	05:28	
Control Test		0.078	05:28	
Air Blank		0.000	05:29	
Subject Sample	# 1	0.158	05:32	Volume Not Met
Breath Volume		0.156	05:32	
Air Blank		0.000	05:32	
Air Blank		0.000	05:34	
Subject Sample	# 2		05:34	Subject Test Refused
Breath Volume		0.000	05:34	
Air Blank		0.000	05:35	
Control Test		0.077	05:35	
Air Blank		0.000	05:36	
Diagnostics Check		OK	05:36	

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 05/07/2016 23:46:52

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.078	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.198	23:42
Breath Volume		1.687	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:44
Subject Sample	# 2	0.182	23:45
Breath Volume		1.460	23:45
Air Blank		0.000	23:45
Control Test		0.077	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEE,TIMOTHY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: OVALLE

DATE/TIME OF TEST: 05/08/2016 02:43:52

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.077	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.227	02:38
Breath Volume		1.417	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.230	02:42
Breath Volume		2.222	02:42
Air Blank		0.000	02:42
Control Test		0.077	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUZQUINOS,ANTHONY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 05/08/2016 05:19:30

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.079	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.184	05:14
Breath Volume		1.156	05:14
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.178	05:18
Breath Volume		1.785	05:18
Air Blank		0.000	05:18
Control Test		0.077	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MARISTANY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DURIZIL

DATE/TIME OF TEST: 05/08/2016 22:34:02

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.078	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.122	22:29
Breath Volume		2.367	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.134	22:32
Breath Volume		2.542	22:32
Air Blank		0.000	22:33
Control Test		0.078	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:34

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: ACOSTA RAMOS

DATE/TIME OF TEST: 05/09/2016 00:24:26

Diagnostics Check		OK	00:17
Air Blank		0.000	00:17
Control Test		0.078	00:18
Air Blank		0.000	00:18
Subject Sample	# 1	0.199	00:20
Breath Volume		2.488	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.188	00:22
Breath Volume		1.542	00:22
Air Blank		0.000	00:23
Control Test		0.077	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKINGER,LANI,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAAYMAN,FRANCO
ARRESTING OFFICER AGENCY: Lighthouse Point PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:32

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/09/2016 08:05:21

Diagnostics Check		OK	07:58
Air Blank		0.000	07:59
Control Test		0.078	07:59
Air Blank		0.000	08:00
Subject Sample	# 1	0.145	08:00
Breath Volume		1.273	08:00
Air Blank		0.000	08:01
Air Blank		0.000	08:02
Subject Sample	# 2	0.135	08:03
Breath Volume		1.636	08:03
Air Blank		0.000	08:04
Control Test		0.078	08:04
Air Blank		0.000	08:05
Diagnostics Check		OK	08:05

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HORNE,PHILLIP
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: NARDI

DATE/TIME OF TEST: 05/12/2016 23:56:44

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.252	23:51
Breath Volume		2.277	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.240	23:55
Breath Volume		2.300	23:55
Air Blank		0.000	23:55
Control Test		0.077	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: THORPE

DATE/TIME OF TEST: 05/13/2016 03:45:21

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.077	03:34
Air Blank		0.000	03:35
Subject Sample	# 1	0.193	03:37
Breath Volume		1.851	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.154	03:40
Breath Volume		2.171	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 3	0.177	03:43
Breath Volume		2.562	03:43
Air Blank		0.000	03:44
Control Test		0.076	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: TROCCOLI,DAVID
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: HASAN

DATE/TIME OF TEST: 05/13/2016 05:13:17

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.079	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.119	05:09
Breath Volume		1.160	05:09
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.110	05:11
Breath Volume		1.285	05:11
Air Blank		0.000	05:12
Control Test		0.078	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 05/13/2016 21:49:19

Diagnostics Check		OK	21:42
Air Blank		0.000	21:42
Control Test		0.078	21:43
Air Blank		0.000	21:43
Subject Sample	# 1	0.256	21:44
Breath Volume		2.910	21:44
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.254	21:47
Breath Volume		2.476	21:47
Air Blank		0.000	21:48
Control Test		0.077	21:48
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DEBAH

DATE/TIME OF TEST: 05/14/2016 01:44:15

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.216	01:39
Breath Volume		1.714	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.206	01:42
Breath Volume		1.984	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LLOYD,WAYNE
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: VEYNOVICH

DATE/TIME OF TEST: 05/14/2016 03:46:04

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.076	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.157	03:41
Breath Volume		2.453	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.161	03:44
Breath Volume		2.597	03:44
Air Blank		0.000	03:45
Control Test		0.077	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:46

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DICINTIO,PATRICK
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: JEFFERS

DATE/TIME OF TEST: 05/14/2016 05:51:47

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.079	05:46
Air Blank		0.000	05:47
Subject Sample	# 1	0.180	05:47
Breath Volume		2.476	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:49
Subject Sample	# 2	0.179	05:50
Breath Volume		2.117	05:50
Air Blank		0.000	05:50
Control Test		0.077	05:51
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMPSON,BRETT
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: BOCANEGRA

DATE/TIME OF TEST: 05/14/2016 07:03:34

Diagnostics Check		OK	06:55
Air Blank		0.000	06:56
Control Test		0.079	06:56
Air Blank		0.000	06:57
Subject Sample	# 1	0.000	06:59
Breath Volume		2.230	06:59
Air Blank		0.000	06:59
Air Blank		0.000	07:01
Subject Sample	# 2	0.000	07:02
Breath Volume		1.687	07:02
Air Blank		0.000	07:02
Control Test		0.078	07:02
Air Blank		0.000	07:03
Diagnostics Check		OK	07:03

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: CULBRETH

DATE/TIME OF TEST: 05/17/2016 05:57:04

Diagnostics Check		OK	05:50
Air Blank		0.000	05:50
Control Test		0.078	05:51
Air Blank		0.000	05:51
Subject Sample	# 1	0.190	05:52
Breath Volume		2.992	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.185	05:55
Breath Volume		2.566	05:55
Air Blank		0.000	05:56
Control Test		0.078	05:56
Air Blank		0.000	05:56
Diagnostics Check		OK	05:57

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RIVERA,RAUL
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 05/17/2016 19:08:09

Diagnostics Check		OK	19:01
Air Blank		0.000	19:02
Control Test		0.079	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.175	19:03
Breath Volume		1.476	19:03
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.164	19:06
Breath Volume		2.156	19:06
Air Blank		0.000	19:07
Control Test		0.078	19:07
Air Blank		0.000	19:08
Diagnostics Check		OK	19:08

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ESCOBAR,DIEGO
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: ALLISON

DATE/TIME OF TEST: 05/17/2016 22:49:32

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:38	
Control Test		0.078	22:38	
Air Blank		0.000	22:38	
Subject Sample	# 1	0.000	22:42	Volume Not Met
Breath Volume		0.074	22:42	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 2	0.279	22:44	
Breath Volume		2.367	22:44	
Air Blank		0.000	22:45	
Air Blank		0.000	22:47	
Subject Sample	# 3	0.274	22:48	
Breath Volume		1.410	22:48	
Air Blank		0.000	22:48	
Control Test		0.077	22:48	
Air Blank		0.000	22:49	
Diagnostics Check		OK	22:49	

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 05/21/2016 01:15:40

Diagnostics Check		OK	01:09
Air Blank		0.000	01:09
Control Test		0.079	01:10
Air Blank		0.000	01:10
Subject Sample	# 1	0.113	01:11
Breath Volume		2.261	01:11
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.110	01:14
Breath Volume		1.917	01:14
Air Blank		0.000	01:14
Control Test		0.079	01:15
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BREA-DALMASI

DATE/TIME OF TEST: 05/21/2016 01:48:47

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.079	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.129	01:44
Breath Volume		1.921	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.126	01:47
Breath Volume		1.742	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECKINGER,LANI,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERNANDEZ,NICHOLAS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 05/24/2016 07:38:25

Diagnostics Check		OK	07:31
Air Blank		0.000	07:31
Control Test		0.079	07:32
Air Blank		0.000	07:32
Subject Sample	# 1	0.000	07:33
Breath Volume		1.437	07:33
Air Blank		0.000	07:33
Air Blank		0.000	07:35
Subject Sample	# 2	0.000	07:37
Breath Volume		1.328	07:37
Air Blank		0.000	07:37
Control Test		0.079	07:37
Air Blank		0.000	07:38
Diagnostics Check		OK	07:38

Data Download Date/Time: 05/24/2016 16:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001149

INSTRUMENT REGISTERED TO: LAKE HELEN P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSHWORTH,JOSEPH,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: RUSHWORTH,JOSEPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SALTER

DATE/TIME OF TEST: 05/06/2016 02:26:17

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.294	02:21
Breath Volume		3.117	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.294	02:24
Breath Volume		2.777	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 05/31/2016 17:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TINO AMBROCIO

DATE/TIME OF TEST: 05/06/2016 02:18:36

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.079	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.207	02:13
Breath Volume		1.535	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.206	02:16
Breath Volume		2.371	02:16
Air Blank		0.000	02:17
Control Test		0.076	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 05/25/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: TRACE

DATE/TIME OF TEST: 05/09/2016 19:47:13

Diagnostics Check		OK	19:44	
Air Blank		0.000	19:44	
Control Test		0.079	19:45	
Air Blank		0.000	19:45	
Subject Sample	# 1		19:45	Subject Test Refused
Breath Volume		0.000	19:45	
Air Blank		0.000	19:46	
Control Test		0.079	19:46	
Air Blank		0.000	19:47	
Diagnostics Check		OK	19:47	

Data Download Date/Time: 05/25/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: LACOURSE

DATE/TIME OF TEST: 05/21/2016 22:07:07

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.079	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.072	22:01
Breath Volume		2.167	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.070	22:05
Breath Volume		2.242	22:05
Air Blank		0.000	22:06
Control Test		0.079	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 05/25/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FINKLEA,WILLIE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 05/24/2016 10:14:53

Diagnostics Check		OK	10:08	
Air Blank		0.000	10:09	
Control Test		0.079	10:09	
Air Blank		0.000	10:10	
Subject Sample	# 1	0.331	10:13	Volume Not Met
Breath Volume		0.031	10:13	
Air Blank			10:14	Purge Fail
Air Blank			10:14	Purge Fail

Data Download Date/Time: 05/25/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FINKLEA,WILLIE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 05/24/2016 10:22:20

Diagnostics Check		OK	10:17	
Air Blank		0.000	10:18	
Control Test		0.076	10:18	
Air Blank		0.000	10:18	
Subject Sample	# 1		10:20	Subject Test Refused
Breath Volume		0.128	10:20	
Air Blank			10:21	Purge Fail
Air Blank			10:22	Purge Fail

Data Download Date/Time: 05/25/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FINKLEA,WILLIE
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: MEYERS

DATE/TIME OF TEST: 05/24/2016 10:26:45

Diagnostics Check		OK	10:23	
Air Blank		0.000	10:24	
Control Test		0.076	10:24	
Air Blank		0.000	10:25	
Subject Sample	# 1		10:25	Subject Test Refused
Breath Volume		0.000	10:25	
Air Blank		0.000	10:25	
Control Test		0.077	10:26	
Air Blank		0.000	10:26	
Diagnostics Check		OK	10:26	

Data Download Date/Time: 05/25/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 04/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: PEISHER

DATE/TIME OF TEST: 05/05/2016 04:54:13

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.080	04:52	
Air Blank		0.000	04:52	
Subject Sample	# 1		04:52	Subject Test Refused
Breath Volume		0.000	04:52	
Air Blank		0.000	04:53	
Control Test		0.079	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 05/15/2016 10:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BASTEDO

DATE/TIME OF TEST: 05/08/2016 03:36:49

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:25	
Control Test		0.081	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:29	No Sample Provided
Breath Volume		0.000	03:29	
Air Blank		0.000	03:30	
Air Blank		0.000	03:32	
Subject Sample	# 2		03:35	No Sample Provided
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.080	03:36	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 05/15/2016 10:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WHITMAN,TIMOTHY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12014080A1
DRY GAS EXPIRY DATE: 06/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RHOADES

DATE/TIME OF TEST: 05/24/2016 02:55:44

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.096	02:50
Breath Volume		1.230	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.098	02:54
Breath Volume		2.011	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/27/2016 06:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IHNKEN,MATTHE,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: YACKEL,N
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: DURKEE

DATE/TIME OF TEST: 05/06/2016 23:48:34

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.088	23:44
Breath Volume		2.488	23:44
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.088	23:47
Breath Volume		2.335	23:47
Air Blank		0.000	23:47
Control Test		0.080	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 06/20/2016 19:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECATES,C,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HOFFMAN,J
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 05/07/2016 20:53:22

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.081	20:47
Air Blank		0.000	20:48
Subject Sample	# 1	0.178	20:48
Breath Volume		1.925	20:48
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.188	20:51
Breath Volume		2.503	20:51
Air Blank		0.000	20:52
Control Test		0.081	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 06/20/2016 19:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TURNEY,JOEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DEADY

DATE/TIME OF TEST: 05/08/2016 00:49:16

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.081	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.233	00:44
Breath Volume		1.128	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.220	00:47
Breath Volume		1.242	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 06/20/2016 19:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKNIGHT,OMAR,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: IGO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: SALVANO

DATE/TIME OF TEST: 05/17/2016 22:05:14

Diagnostics Check		OK	22:01	
Air Blank		0.000	22:01	
Control Test		0.081	22:02	
Air Blank		0.000	22:02	
Subject Sample	# 1		22:03	Subject Test Refused
Breath Volume		0.000	22:03	
Air Blank		0.000	22:04	
Control Test		0.081	22:04	
Air Blank		0.000	22:05	
Diagnostics Check		OK	22:05	

Data Download Date/Time: 06/20/2016 19:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WHITTET,JAMES
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: BUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: RIVER-PEREZ

DATE/TIME OF TEST: 05/28/2016 16:41:27

Diagnostics Check		OK	16:34
Air Blank		0.000	16:35
Control Test		0.081	16:35
Air Blank		0.000	16:35
Subject Sample	# 1	0.106	16:36
Breath Volume		1.238	16:36
Air Blank		0.000	16:37
Air Blank		0.000	16:39
Subject Sample	# 2	0.106	16:39
Breath Volume		1.304	16:39
Air Blank		0.000	16:40
Control Test		0.080	16:40
Air Blank		0.000	16:41
Diagnostics Check		OK	16:41

Data Download Date/Time: 06/20/2016 19:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TATUM,THOMAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CACCHIONE

DATE/TIME OF TEST: 05/28/2016 23:07:22

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.081	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.176	23:02
Breath Volume		2.457	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.172	23:05
Breath Volume		2.246	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MCDONOUGH,CLAY
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: BUI
ARREST TIME: 18:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: NOTZ

DATE/TIME OF TEST: 05/29/2016 19:02:00

Diagnostics Check		OK	18:55
Air Blank		0.000	18:55
Control Test		0.081	18:56
Air Blank		0.000	18:56
Subject Sample	# 1	0.102	18:57
Breath Volume		3.378	18:57
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.104	19:00
Breath Volume		3.808	19:00
Air Blank		0.000	19:01
Control Test		0.081	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:02

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRYAN,STEVEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: WOSKA

DATE/TIME OF TEST: 05/30/2016 03:17:09

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:12	No Sample Provided
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:16	
Control Test		0.081	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENELL,BENITA,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SANDERS,LAUREN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: OAKMAN

DATE/TIME OF TEST: 05/02/2016 03:42:36

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 1		03:41	Subject Test Refused
Breath Volume		0.000	03:41	
Air Blank		0.000	03:41	
Control Test		0.080	03:42	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SANDERS,LAUREN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 05/05/2016 00:54:17

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.154	00:49
Breath Volume		1.847	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.144	00:52
Breath Volume		2.082	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: BRAUN,CALEB
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 05/06/2016 17:55:23

Diagnostics Check		OK	17:46
Air Blank		0.000	17:46
Control Test		0.081	17:46
Air Blank		0.000	17:47
Subject Sample	# 1	0.354	17:49
Breath Volume		1.582	17:49
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.353	17:53
Breath Volume		1.117	17:53
Air Blank		0.000	17:54
Control Test		0.079	17:54
Air Blank		0.000	17:55
Diagnostics Check		OK	17:55

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: DELONG,JOSH
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SCHOEN

DATE/TIME OF TEST: 05/08/2016 23:24:58

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.080	23:16
Air Blank		0.000	23:16
Subject Sample	# 1	0.243	23:18
Breath Volume		1.949	23:18
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.245	23:23
Breath Volume		2.332	23:23
Air Blank		0.000	23:24
Control Test		0.081	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HOWARD,JAKE
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: NUTTER

DATE/TIME OF TEST: 05/13/2016 07:09:59

Diagnostics Check		OK	06:58	
Air Blank		0.000	06:59	
Control Test		0.080	06:59	
Air Blank		0.000	06:59	
Subject Sample	# 1		07:03	No Sample Provided
Breath Volume		0.000	07:03	
Air Blank		0.000	07:03	
Air Blank		0.000	07:05	
Subject Sample	# 2		07:08	No Sample Provided
Breath Volume		0.000	07:08	
Air Blank		0.000	07:09	
Control Test		0.081	07:09	
Air Blank		0.000	07:09	
Diagnostics Check		OK	07:09	

Data Download Date/Time: 05/28/2016 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HOWARD,JAKE
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: NUTTER

DATE/TIME OF TEST: 05/13/2016 07:22:33

Diagnostics Check		OK	07:13
Air Blank		0.000	07:13
Control Test		0.080	07:13
Air Blank		0.000	07:14
Subject Sample	# 1	0.178	07:16
Breath Volume		1.312	07:16
Air Blank		0.000	07:16
Air Blank		0.000	07:18
Subject Sample	# 2	0.191	07:21
Breath Volume		2.562	07:21
Air Blank		0.000	07:21
Control Test		0.080	07:21
Air Blank		0.000	07:22
Diagnostics Check		OK	07:22

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENELL,BENITA,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MOODY,SIERRA
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 05/15/2016 23:38:38

Diagnostics Check		OK	23:34	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:37	Subject Test Refused
Breath Volume		0.000	23:37	
Air Blank		0.000	23:37	
Control Test		0.081	23:38	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HYATT,JUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BREMER

DATE/TIME OF TEST: 05/23/2016 00:27:39

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:25	
Control Test		0.081	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1		00:26	Subject Test Refused
Breath Volume		0.000	00:26	
Air Blank		0.000	00:26	
Control Test		0.080	00:27	
Air Blank		0.000	00:27	
Diagnostics Check		OK	00:27	

Data Download Date/Time: 05/28/2016 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SANDERS,LAUREN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HOVATTER-MARTINO

DATE/TIME OF TEST: 05/24/2016 00:13:05

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.079	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:05	No Sample Provided
Breath Volume		0.000	00:05	
Air Blank		0.000	00:06	
Air Blank		0.000	00:08	
Subject Sample	# 2	0.166	00:11	Volume Not Met
Breath Volume		0.058	00:11	
Air Blank		0.000	00:12	
Control Test		0.082	00:12	
Air Blank		0.000	00:13	
Diagnostics Check		OK	00:13	

Data Download Date/Time: 05/28/2016 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SANDERS,LAUREN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HOVATTER-MARTINO

DATE/TIME OF TEST: 05/24/2016 00:20:34

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:16	
Control Test		0.079	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 1		00:19	Subject Test Refused
Breath Volume		0.058	00:19	
Air Blank		0.000	00:19	
Control Test		0.080	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: WARMINGTON,WILLIAM
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:24

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 05/25/2016 19:56:21

Diagnostics Check		OK	19:50
Air Blank		0.000	19:50
Control Test		0.079	19:50
Air Blank		0.000	19:51
Subject Sample	# 1	0.143	19:51
Breath Volume		1.871	19:51
Air Blank		0.000	19:52
Air Blank		0.000	19:54
Subject Sample	# 2	0.155	19:54
Breath Volume		1.542	19:54
Air Blank		0.000	19:55
Control Test		0.079	19:55
Air Blank		0.000	19:56
Diagnostics Check		OK	19:56

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,BRENT,M
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: JONES,BRENT
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: RENTAS COLON

DATE/TIME OF TEST: 05/04/2016 01:22:33

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.185	01:18
Breath Volume		2.683	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.187	01:20
Breath Volume		2.785	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 05/17/2016 09:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JASON,A
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: NESBITT,STEPHEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 05/16/2016 02:41:27

Diagnostics Check		OK	02:35
Air Blank		0.000	02:35
Control Test		0.077	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.164	02:37
Breath Volume		1.226	02:37
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.166	02:39
Breath Volume		1.722	02:39
Air Blank		0.000	02:40
Control Test		0.078	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 05/17/2016 09:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,BRENT,M
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: LEE,RYAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 04/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HARTLE

DATE/TIME OF TEST: 05/16/2016 23:04:35

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:02	
Control Test		0.079	23:02	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:03	Subject Test Refused
Breath Volume		0.000	23:03	
Air Blank		0.000	23:03	
Control Test		0.079	23:03	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:04	

Data Download Date/Time: 05/17/2016 09:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERDIKES,PANDELIS,C
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: PERDIKES,PANDELIS
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: OESTER

DATE/TIME OF TEST: 05/30/2016 00:59:44

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.080	00:59	
Air Blank		00:59	RFI Detect
Air Blank	0.000	00:59	

Data Download Date/Time: 06/17/2016 02:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERDIKES,PANDELIS,C
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: PERDIKES,PANDELIS
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: OESTER

DATE/TIME OF TEST: 05/30/2016 01:10:41

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.081	01:08	
Air Blank		0.000	01:09	
Subject Sample	# 1		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.079	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 06/17/2016 02:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CRAWFORD,ANDREW
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 05/01/2016 05:40:26

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:38	
Control Test		0.079	05:38	
Air Blank		0.000	05:38	
Subject Sample	# 1		05:39	Subject Test Refused
Breath Volume		0.000	05:39	
Air Blank		0.000	05:39	
Control Test		0.079	05:39	
Air Blank		0.000	05:40	
Diagnostics Check		OK	05:40	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSENTHAL,DAVID,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GANNON,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 05/01/2016 07:51:58

Diagnostics Check		OK	07:49	
Air Blank		0.000	07:49	
Control Test		0.079	07:49	
Air Blank		0.000	07:50	
Subject Sample	# 1		07:50	Subject Test Refused
Breath Volume		0.000	07:50	
Air Blank		0.000	07:51	
Control Test		0.079	07:51	
Air Blank		0.000	07:51	
Diagnostics Check		OK	07:51	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GRAHAM,DALE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 05/03/2016 04:31:14

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.078	04:16	
Air Blank		0.000	04:16	
Subject Sample	# 1	0.184	04:18	
Breath Volume		1.343	04:18	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2	0.151	04:23	Volume Not Met
Breath Volume		0.644	04:23	
Air Blank		0.000	04:24	
Air Blank		0.000	04:26	
Subject Sample	# 3	0.179	04:29	Volume Not Met
Breath Volume		0.570	04:29	
Air Blank		0.000	04:30	
Control Test		0.077	04:30	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 05/23/2016 10:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAHAM,KEITH,T
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: JESIONEK,ROBERT
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 05/06/2016 01:54:46

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:52	
Control Test		0.080	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:53	Subject Test Refused
Breath Volume		0.000	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Diagnostics Check		OK	01:54	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSOON,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CRAWFORD,ANDREW
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: ISOLA

DATE/TIME OF TEST: 05/06/2016 05:14:53

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.176	05:10
Breath Volume		2.683	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:12
Subject Sample	# 2	0.166	05:13
Breath Volume		2.046	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:14
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRIGGER,ROHALIO,E
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: COFIELD

DATE/TIME OF TEST: 05/06/2016 10:21:28

Diagnostics Check		OK	10:13
Air Blank		0.000	10:14
Control Test		0.079	10:14
Air Blank		0.000	10:15
Subject Sample	# 1	0.085	10:17
Breath Volume		1.750	10:17
Air Blank		0.000	10:17
Air Blank		0.000	10:19
Subject Sample	# 2	0.086	10:19
Breath Volume		1.132	10:19
Air Blank		0.000	10:20
Control Test		0.079	10:20
Air Blank		0.000	10:21
Diagnostics Check		OK	10:21

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENZ,JONATHAN',R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: DRISCOLL

DATE/TIME OF TEST: 05/07/2016 23:19:48

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.079	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.193	23:15
Breath Volume		4.277	23:15
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.191	23:18
Breath Volume		4.316	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: FRENZ,JONATHAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: LO PRIMO

DATE/TIME OF TEST: 05/09/2016 22:42:52

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:40	
Control Test		0.079	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:41	
Control Test		0.080	22:42	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRIGGER,ROHALIO,E
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CRAWFORD,ANDREW
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: BALTAZAR

DATE/TIME OF TEST: 05/12/2016 06:03:54

Diagnostics Check		OK	06:00	
Air Blank		0.000	06:01	
Control Test		0.080	06:01	
Air Blank		0.000	06:02	
Subject Sample	# 1		06:02	Subject Test Refused
Breath Volume		0.000	06:02	
Air Blank		0.000	06:03	
Control Test		0.079	06:03	
Air Blank		0.000	06:03	
Diagnostics Check		OK	06:03	

Data Download Date/Time: 05/23/2016 10:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: JESIONEK,ROBERT
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DE ARAUJO

DATE/TIME OF TEST: 05/13/2016 02:41:48

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:39	
Control Test		0.080	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:40	Subject Test Refused
Breath Volume		0.000	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:41	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JESIONEK,ROBERT,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: JESIONEK,ROBERT
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 05/14/2016 02:07:33

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ;6360100162001980264

DATE/TIME OF TEST: 05/14/2016 04:16:51

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:14	
Control Test		0.080	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:15	Subject Test Refused
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.079	04:16	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURKE,ZACHARY,D
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: JESIONEK,ROBERT
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CABRA

DATE/TIME OF TEST: 05/20/2016 02:56:22

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.079	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:55	
Control Test		0.079	02:55	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSOON,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CRAWFORD,ANDREW
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: RIZZO

DATE/TIME OF TEST: 05/20/2016 05:18:56

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.079	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.164	05:14
Breath Volume		2.136	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.160	05:17
Breath Volume		2.339	05:17
Air Blank		0.000	05:18
Control Test		0.080	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 05/23/2016 10:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSOON,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 04/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 05/22/2016 23:22:50

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:20	
Control Test		0.079	23:20	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:21	
Control Test		0.079	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 05/23/2016 10:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNEY,DANIEL,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FREDERICKS

DATE/TIME OF TEST: 05/24/2016 00:49:51

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.079	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.052	00:43
Breath Volume		1.148	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.051	00:48
Breath Volume		1.277	00:48
Air Blank		0.000	00:48
Control Test		0.080	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 06/09/2016 09:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: JESIONEK,R
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SHOKRY

DATE/TIME OF TEST: 05/26/2016 03:17:22

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:15	
Control Test		0.080	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:16	Subject Test Refused
Breath Volume		0.000	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 06/09/2016 09:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CRAWFORD,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SUAREZ GOMEZ

DATE/TIME OF TEST: 05/26/2016 03:47:17

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.183	03:42
Breath Volume		2.277	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.174	03:45
Breath Volume		1.632	03:45
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/09/2016 09:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: JESIONEK,ROBERT
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CADENA

DATE/TIME OF TEST: 05/29/2016 03:08:51

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:06	
Control Test		0.079	03:06	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:07	Subject Test Refused
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Control Test		0.079	03:08	
Air Blank		0.000	03:08	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 06/09/2016 09:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRIGGER,ROHALIO,E
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BETHEL,RICKEY
ARRESTING OFFICER AGENCY: FL Atlantic UPD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/30/2016 05:57:43

Diagnostics Check		OK	05:53	
Air Blank		0.000	05:53	
Control Test		0.079	05:54	
Air Blank		0.000	05:54	
Subject Sample	# 1		05:56	Subject Test Refused
Breath Volume		0.000	05:56	
Air Blank		0.000	05:56	
Control Test		0.080	05:57	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 06/09/2016 09:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHARD,ROSS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: JESIONEK,ROBERT
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: ELLISON JR

DATE/TIME OF TEST: 05/31/2016 20:48:08

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.079	20:41
Air Blank		0.000	20:41
Subject Sample	# 1	0.150	20:42
Breath Volume		1.636	20:42
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.146	20:46
Breath Volume		1.402	20:46
Air Blank		0.000	20:47
Control Test		0.079	20:47
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 06/09/2016 09:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: MICELI

DATE/TIME OF TEST: 05/01/2016 19:25:06

Diagnostics Check		OK	19:18
Air Blank		0.000	19:18
Control Test		0.081	19:19
Air Blank		0.000	19:19
Subject Sample	# 1	0.000	19:20
Breath Volume		1.117	19:20
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 2	0.000	19:23
Breath Volume		1.109	19:23
Air Blank		0.000	19:24
Control Test		0.081	19:24
Air Blank		0.000	19:25
Diagnostics Check		OK	19:25

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: VILLAFRANCO

DATE/TIME OF TEST: 05/02/2016 01:03:45

Diagnostics Check		OK	00:56
Air Blank		0.000	00:57
Control Test		0.081	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.152	00:58
Breath Volume		1.738	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.147	01:02
Breath Volume		2.660	01:02
Air Blank		0.000	01:02
Control Test		0.081	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAGALO,PHILIP,J
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LAGALO,PHILIP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: SCHERER

DATE/TIME OF TEST: 05/02/2016 15:17:24

Diagnostics Check		OK	15:10
Air Blank		0.000	15:11
Control Test		0.080	15:11
Air Blank		0.000	15:12
Subject Sample	# 1	0.186	15:12
Breath Volume		1.625	15:12
Air Blank		0.000	15:13
Air Blank		0.000	15:15
Subject Sample	# 2	0.193	15:15
Breath Volume		1.554	15:15
Air Blank		0.000	15:16
Control Test		0.079	15:16
Air Blank		0.000	15:17
Diagnostics Check		OK	15:17

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: TERCERO VELASQUEZ

DATE/TIME OF TEST: 05/02/2016 20:27:17

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.080	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.166	20:22
Breath Volume		2.097	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.160	20:25
Breath Volume		1.738	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,COREY,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEE,COREY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 05/03/2016 20:40:06

Diagnostics Check		OK	20:33
Air Blank		0.000	20:33
Control Test		0.080	20:34
Air Blank		0.000	20:34
Subject Sample	# 1	0.218	20:35
Breath Volume		1.574	20:35
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.208	20:38
Breath Volume		1.812	20:38
Air Blank		0.000	20:39
Control Test		0.079	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDIGER,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: RUEDIGER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: DEDMAN JR

DATE/TIME OF TEST: 05/03/2016 21:52:18

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.080	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.205	21:47
Breath Volume		2.152	21:47
Air Blank		0.000	21:48
Air Blank		0.000	21:49
Subject Sample	# 2	0.188	21:50
Breath Volume		2.304	21:50
Air Blank		0.000	21:51
Control Test		0.078	21:51
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: TARLTON

DATE/TIME OF TEST: 05/04/2016 01:22:32

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.106	01:17
Breath Volume		1.585	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.103	01:21
Breath Volume		1.593	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,DAVID,E
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: JACOBSON,DAVID
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: UBILLA

DATE/TIME OF TEST: 05/05/2016 15:10:18

Diagnostics Check		OK	15:01
Air Blank		0.000	15:01
Control Test		0.081	15:01
Air Blank		0.000	15:02
Subject Sample	# 1	0.087	15:03
Breath Volume		1.660	15:03
Air Blank		0.000	15:04
Air Blank		0.000	15:06
Subject Sample	# 2	0.094	15:08
Breath Volume		1.804	15:08
Air Blank		0.000	15:09
Control Test		0.080	15:09
Air Blank		0.000	15:10
Diagnostics Check		OK	15:10

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: DARGHOUTH

DATE/TIME OF TEST: 05/05/2016 23:54:12

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.090	23:49
Breath Volume		2.082	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.086	23:52
Breath Volume		1.398	23:52
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:24

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 05/09/2016 18:52:37

Diagnostics Check		OK	18:46
Air Blank		0.000	18:46
Control Test		0.081	18:46
Air Blank		0.000	18:47
Subject Sample	# 1	0.255	18:48
Breath Volume		2.222	18:48
Air Blank		0.000	18:48
Air Blank		0.000	18:50
Subject Sample	# 2	0.263	18:51
Breath Volume		2.914	18:51
Air Blank		0.000	18:51
Control Test		0.080	18:52
Air Blank		0.000	18:52
Diagnostics Check		OK	18:52

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GAGNON

DATE/TIME OF TEST: 05/10/2016 01:35:37

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.081	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.165	01:31
Breath Volume		3.562	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.157	01:34
Breath Volume		3.410	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 05/18/2016 14:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 05/11/2016 01:28:38

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.000	01:24
Breath Volume		2.070	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.000	01:27
Breath Volume		2.316	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:28
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: FEDER

DATE/TIME OF TEST: 05/13/2016 00:19:23

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.105	00:14
Breath Volume		1.386	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.103	00:17
Breath Volume		1.257	00:17
Air Blank		0.000	00:18
Control Test		0.080	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 05/13/2016 00:42:37

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.190	00:38
Breath Volume		1.441	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.197	00:41
Breath Volume		2.328	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 05/18/2016 14:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: DOMINGO PABLO

DATE/TIME OF TEST: 05/13/2016 03:46:54

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.210	03:42
Breath Volume		2.046	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.217	03:45
Breath Volume		1.992	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: MELECIO

DATE/TIME OF TEST: 05/14/2016 03:02:57

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.080	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:54	No Sample Provided
Breath Volume		0.000	02:54	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 2	0.154	02:58	
Breath Volume		1.683	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 3	0.164	03:01	
Breath Volume		2.972	03:01	
Air Blank		0.000	03:02	
Control Test		0.078	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 05/18/2016 14:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: CRICHLOW

DATE/TIME OF TEST: 05/15/2016 02:20:43

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.196	02:13
Breath Volume		1.449	02:13
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.174	02:16
Breath Volume		1.699	02:16
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 3	0.188	02:19
Breath Volume		3.179	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 05/15/2016 03:34:49

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.094	03:30
Breath Volume		2.656	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.093	03:33
Breath Volume		2.578	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: SON

DATE/TIME OF TEST: 05/15/2016 04:35:38

Diagnostics Check		OK	04:29
Air Blank		0.000	04:29
Control Test		0.080	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.183	04:31
Breath Volume		3.035	04:31
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.185	04:34
Breath Volume		3.246	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GITTO,JUSTIN,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GITTO,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: HENAO

DATE/TIME OF TEST: 05/16/2016 01:40:40

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.080	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.263	01:36
Breath Volume		2.589	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.258	01:39
Breath Volume		1.906	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LILLJA

DATE/TIME OF TEST: 05/17/2016 23:59:33

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.079	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.323	23:55
Breath Volume		3.000	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.331	23:57
Breath Volume		3.535	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 05/18/2016 14:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 05/19/2016 01:04:46

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.000	01:00
Breath Volume		2.121	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.000	01:03
Breath Volume		1.660	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: MELICE

DATE/TIME OF TEST: 05/20/2016 22:30:13

Diagnostics Check		OK	22:23
Air Blank		0.000	22:24
Control Test		0.080	22:24
Air Blank		0.000	22:25
Subject Sample	# 1	0.258	22:25
Breath Volume		1.484	22:25
Air Blank		0.000	22:26
Air Blank		0.000	22:28
Subject Sample	# 2	0.276	22:28
Breath Volume		1.203	22:28
Air Blank		0.000	22:29
Control Test		0.080	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FREUND

DATE/TIME OF TEST: 05/21/2016 02:05:27

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.080	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.089	02:01
Breath Volume		2.230	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.087	02:03
Breath Volume		2.519	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DETRUIT

DATE/TIME OF TEST: 05/21/2016 03:21:12

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.133	03:16
Breath Volume		2.796	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.140	03:19
Breath Volume		2.843	03:19
Air Blank		0.000	03:20
Control Test		0.081	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: BUCKLES

DATE/TIME OF TEST: 05/22/2016 16:48:18

Diagnostics Check		OK	16:42
Air Blank		0.000	16:42
Control Test		0.080	16:42
Air Blank		0.000	16:43
Subject Sample	# 1	0.141	16:43
Breath Volume		2.269	16:43
Air Blank		0.000	16:44
Air Blank		0.000	16:46
Subject Sample	# 2	0.137	16:46
Breath Volume		2.390	16:46
Air Blank		0.000	16:47
Control Test		0.079	16:47
Air Blank		0.000	16:48
Diagnostics Check		OK	16:48

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/23/2016 19:16:13

Diagnostics Check		OK	19:09
Air Blank		0.000	19:10
Control Test		0.080	19:10
Air Blank		0.000	19:11
Subject Sample	# 1	0.316	19:11
Breath Volume		2.800	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.311	19:14
Breath Volume		2.957	19:14
Air Blank		0.000	19:15
Control Test		0.079	19:15
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,MARK,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: STEELE,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/26/2016 02:55:43

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.081	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.189	02:48
Breath Volume		2.160	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.165	02:51
Breath Volume		1.265	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 3	0.183	02:54
Breath Volume		3.011	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOLBRECHT,JON,H
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: FOLBRECHT,JON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: SCHEFFER

DATE/TIME OF TEST: 05/26/2016 17:17:15

Diagnostics Check		OK	17:10
Air Blank		0.000	17:11
Control Test		0.080	17:11
Air Blank		0.000	17:12
Subject Sample	# 1	0.301	17:12
Breath Volume		2.179	17:12
Air Blank		0.000	17:13
Air Blank		0.000	17:15
Subject Sample	# 2	0.304	17:15
Breath Volume		2.296	17:15
Air Blank		0.000	17:16
Control Test		0.081	17:16
Air Blank		0.000	17:17
Diagnostics Check		OK	17:17

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:36

SUBJECT'S LAST NAME: WENZEL

DATE/TIME OF TEST: 05/27/2016 16:08:16

Diagnostics Check		OK	16:02
Air Blank		0.000	16:02
Control Test		0.080	16:03
Air Blank		0.000	16:03
Subject Sample	# 1	0.000	16:04
Breath Volume		2.250	16:04
Air Blank		0.000	16:04
Air Blank		0.000	16:06
Subject Sample	# 2	0.000	16:06
Breath Volume		1.750	16:06
Air Blank		0.000	16:07
Control Test		0.079	16:07
Air Blank		0.000	16:08
Diagnostics Check		OK	16:08

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YUN,DAVID,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: YUN,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/29/2016 00:13:14

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.218	00:08
Breath Volume		1.554	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.230	00:11
Breath Volume		1.433	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGIGHAN,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MCGIGHAN,WILLIAM
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:56

SUBJECT'S LAST NAME: CONNOLLY

DATE/TIME OF TEST: 05/29/2016 07:26:01

Diagnostics Check		OK	07:19	
Air Blank		0.000	07:20	
Control Test		0.080	07:20	
Air Blank		0.000	07:21	
Subject Sample	# 1	0.183	07:21	
Breath Volume		3.781	07:21	
Air Blank		0.000	07:22	
Air Blank		0.000	07:24	
Subject Sample	# 2		07:24	Subject Test Refused
Breath Volume		0.000	07:24	
Air Blank		0.000	07:25	
Control Test		0.079	07:25	
Air Blank		0.000	07:25	
Diagnostics Check		OK	07:26	

Data Download Date/Time: 06/28/2016 00:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:54

SUBJECT'S LAST NAME: LALIBERTE

DATE/TIME OF TEST: 05/29/2016 14:28:53

Diagnostics Check		OK	14:22
Air Blank		0.000	14:23
Control Test		0.080	14:23
Air Blank		0.000	14:24
Subject Sample	# 1	0.000	14:24
Breath Volume		1.644	14:24
Air Blank		0.000	14:24
Air Blank		0.000	14:26
Subject Sample	# 2	0.000	14:27
Breath Volume		1.234	14:27
Air Blank		0.000	14:27
Control Test		0.080	14:28
Air Blank		0.000	14:28
Diagnostics Check		OK	14:28

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANARO,RONNIE,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PANARO,RONNIE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 05/29/2016 18:17:36

Diagnostics Check		OK	18:09
Air Blank		0.000	18:10
Control Test		0.080	18:10
Air Blank		0.000	18:11
Subject Sample	# 1	0.274	18:12
Breath Volume		2.691	18:12
Air Blank		0.000	18:13
Air Blank		0.000	18:15
Subject Sample	# 2	0.273	18:16
Breath Volume		1.937	18:16
Air Blank		0.000	18:16
Control Test		0.078	18:17
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTERSON,KYLE,L
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PATTERSON,KYLE
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: BUI
ARREST TIME: 13:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:10

SUBJECT'S LAST NAME: BARFIELD

DATE/TIME OF TEST: 05/30/2016 16:41:18

Diagnostics Check		OK	16:35
Air Blank		0.000	16:35
Control Test		0.080	16:35
Air Blank		0.000	16:36
Subject Sample	# 1	0.117	16:36
Breath Volume		2.558	16:36
Air Blank		0.000	16:37
Air Blank		0.000	16:39
Subject Sample	# 2	0.117	16:39
Breath Volume		2.777	16:39
Air Blank		0.000	16:40
Control Test		0.080	16:40
Air Blank		0.000	16:41
Diagnostics Check		OK	16:41

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 05/30/2016 21:25:20

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.080	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.238	21:20
Breath Volume		2.238	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.236	21:23
Breath Volume		1.847	21:23
Air Blank		0.000	21:24
Control Test		0.079	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 06/28/2016 00:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1518994
DRY GAS EXPIRY DATE: 06/26/2016
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: VERDI

DATE/TIME OF TEST: 05/03/2016 21:38:38

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:35	
Control Test		0.078	21:36	
Air Blank		0.000	21:36	
Subject Sample	# 1		21:37	Subject Test Refused
Breath Volume		0.000	21:37	
Air Blank		0.000	21:37	
Control Test		0.078	21:38	
Air Blank		0.000	21:38	
Diagnostics Check		OK	21:38	

Data Download Date/Time: 05/26/2016 15:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1518994
DRY GAS EXPIRY DATE: 06/26/2016
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:36

SUBJECT'S LAST NAME: FLORES FLORES

DATE/TIME OF TEST: 05/08/2016 21:11:33

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.079	21:04
Air Blank		0.000	21:04
Subject Sample	# 1	0.115	21:06
Breath Volume		1.835	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:08
Subject Sample	# 2	0.117	21:10
Breath Volume		1.253	21:10
Air Blank		0.000	21:10
Control Test		0.079	21:10
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 05/26/2016 15:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURCO,CHARLES,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: GRIGORIAN,ARAMIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1518994
DRY GAS EXPIRY DATE: 06/26/2016
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SWISS

DATE/TIME OF TEST: 05/16/2016 22:58:00

Diagnostics Check		OK	22:51
Air Blank		0.000	22:52
Control Test		0.079	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.000	22:53
Breath Volume		3.480	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.000	22:56
Breath Volume		2.675	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:58

Data Download Date/Time: 05/26/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURCO,CHARLES,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: PICCIANO,HOLLY
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1518994
DRY GAS EXPIRY DATE: 06/26/2016
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 05/17/2016 00:38:08

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:23	
Control Test		0.079	00:24	
Air Blank		0.000	00:24	
Subject Sample	# 1	0.210	00:27	Volume Not Met
Breath Volume		0.179	00:27	
Air Blank		0.000	00:28	
Air Blank		0.000	00:30	
Subject Sample	# 2	0.240	00:31	
Breath Volume		1.773	00:31	
Air Blank		0.000	00:31	
Air Blank		0.000	00:33	
Subject Sample	# 3	0.229	00:36	
Breath Volume		1.687	00:36	
Air Blank		0.000	00:37	
Control Test		0.078	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 05/26/2016 15:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANCE,CHARLES,R
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: JOHNSON,ROBERT
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: LECHUGA

DATE/TIME OF TEST: 05/01/2016 00:43:49

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.080	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.175	00:39
Breath Volume		3.187	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.166	00:42
Breath Volume		2.125	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 06/01/2016 14:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATANI,J,V
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CATANI,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 05/03/2016 06:00:04

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.077	05:53
Air Blank		0.000	05:54
Subject Sample	# 1	0.132	05:55
Breath Volume		1.285	05:55
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.125	05:58
Breath Volume		1.730	05:58
Air Blank		0.000	05:59
Control Test		0.077	05:59
Air Blank		0.000	05:59
Diagnostics Check		OK	06:00

Data Download Date/Time: 05/11/2016 08:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ANTHONY,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: ESQUIVEL

DATE/TIME OF TEST: 05/07/2016 04:25:50

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.078	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.195	04:21
Breath Volume		2.472	04:21
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.201	04:24
Breath Volume		1.828	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 05/11/2016 08:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALDERMAN,JOSEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/16/2016 00:33:13

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.081	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.150	00:28
Breath Volume		2.406	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.147	00:31
Breath Volume		2.160	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 06/01/2016 10:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,JOHN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SIEGLER,JOSH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ROOK

DATE/TIME OF TEST: 05/28/2016 02:24:53

Diagnostics Check		OK	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.260	02:20
Breath Volume		1.859	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.247	02:23
Breath Volume		1.968	02:23
Air Blank		0.000	02:23
Control Test		0.079	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 06/01/2016 10:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOUTH,DAMIAN,W
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: SOUTH,DAMIAN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: CANO-ROSALES

DATE/TIME OF TEST: 05/30/2016 02:28:15

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.196	02:22
Breath Volume		1.710	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.204	02:26
Breath Volume		2.492	02:26
Air Blank		0.000	02:27
Control Test		0.079	02:27
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 06/01/2016 10:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERGUSON,ROBERT,K
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: TEE,D
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/27/2016 23:02:25

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.077	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.189	22:57
Breath Volume		2.292	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.206	23:00
Breath Volume		2.671	23:00
Air Blank		0.000	23:01
Control Test		0.076	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 05/29/2016 23:58

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HER,KENG,
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: STEPHEN

DATE/TIME OF TEST: 05/07/2016 00:47:32

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.062	00:43
Breath Volume		3.031	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.063	00:46
Breath Volume		2.582	00:46
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 05/11/2016 11:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDISON,DAMON,
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 04/25/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 036
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:21

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 05/07/2016 14:47:25

Diagnostics Check		OK	14:41
Air Blank		0.000	14:42
Control Test		0.080	14:42
Air Blank		0.000	14:42
Subject Sample	# 1	0.000	14:43
Breath Volume		2.996	14:43
Air Blank		0.000	14:43
Air Blank		0.000	14:45
Subject Sample	# 2	0.000	14:46
Breath Volume		3.773	14:46
Air Blank		0.000	14:46
Control Test		0.080	14:46
Air Blank		0.000	14:47
Diagnostics Check		OK	14:47

Data Download Date/Time: 05/11/2016 11:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRY,DASMINE,K
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: PAUL,JAMES
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: MOUSE

DATE/TIME OF TEST: 05/11/2016 11:50:40

Diagnostics Check		OK	11:41
Air Blank		0.000	11:41
Control Test		0.079	11:41
Air Blank		0.000	11:42
Subject Sample	# 1	0.095	11:43
Breath Volume		1.222	11:43
Air Blank		0.000	11:43
Air Blank		0.000	11:45
Subject Sample	# 2	0.011	11:46
Breath Volume		1.460	11:46
Air Blank		0.000	11:46
Air Blank		0.000	11:48
Subject Sample	# 3	0.000	11:49
Breath Volume		1.609	11:49
Air Blank		0.000	11:49
Control Test		0.079	11:50
Air Blank		0.000	11:50
Diagnostics Check		OK	11:50

Data Download Date/Time: 06/10/2016 10:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEARNED,DANIEL,K
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BENANTE

DATE/TIME OF TEST: 05/29/2016 01:44:51

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:42	
Control Test		0.079	01:43	
Air Blank		0.000	01:43	
Subject Sample	# 1	0.179	01:44	
Breath Volume		3.507	01:44	
Air Blank			01:44	RFI Detect
Air Blank			01:44	RFI Detect
Air Blank		0.000	01:44	

Data Download Date/Time: 06/10/2016 10:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENATE,MATTHEW,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 09/01/2014
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: BENANTE

DATE/TIME OF TEST: 05/29/2016 02:12:34

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.167	02:07
Breath Volume		2.687	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.171	02:11
Breath Volume		2.226	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:11
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 06/10/2016 10:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: LEAL,AUSTIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: LINDER

DATE/TIME OF TEST: 05/06/2016 05:20:12

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.077	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.184	05:14
Breath Volume		1.921	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.185	05:18
Breath Volume		1.933	05:18
Air Blank		0.000	05:19
Control Test		0.076	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 05/17/2016 02:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMS,JOHN,R
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: WALL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: SIMS

DATE/TIME OF TEST: 05/31/2016 23:55:13

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.076	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.261	23:50
Breath Volume		2.281	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.255	23:53
Breath Volume		1.757	23:53
Air Blank		0.000	23:54
Control Test		0.076	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 06/29/2016 03:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/21/2016 00:23:43

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.131	00:19
Breath Volume		1.492	00:19
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.133	00:22
Breath Volume		2.687	00:22
Air Blank		0.000	00:22
Control Test		0.079	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEROGATIS,JOSEPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: SEVILLE

DATE/TIME OF TEST: 05/21/2016 01:18:06

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.000	01:13
Breath Volume		2.242	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.000	01:16
Breath Volume		2.312	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: HAQ

DATE/TIME OF TEST: 05/21/2016 03:51:26

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.124	03:46
Breath Volume		2.597	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.129	03:49
Breath Volume		3.367	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: RICCIARDI

DATE/TIME OF TEST: 05/21/2016 23:19:42

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.080	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.000	23:15
Breath Volume		2.414	23:15
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.000	23:18
Breath Volume		2.460	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MITCHELL,CEDRIC
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 05/22/2016 04:11:55

Diagnostics Check		OK	04:04
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.184	04:06
Breath Volume		1.667	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.187	04:10
Breath Volume		1.746	04:10
Air Blank		0.000	04:11
Control Test		0.082	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 05/24/2016 19:51:07

Diagnostics Check		OK	19:43
Air Blank		0.000	19:44
Control Test		0.080	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.000	19:46
Breath Volume		2.554	19:46
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.000	19:49
Breath Volume		1.917	19:49
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: PINKERTON

DATE/TIME OF TEST: 05/26/2016 04:21:07

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:07	
Control Test		0.081	04:08	
Air Blank		0.000	04:08	
Subject Sample	# 1	0.159	04:11	Volume Not Met
Breath Volume		0.179	04:11	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.175	04:15	
Breath Volume		1.667	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 3	0.180	04:19	
Breath Volume		1.789	04:19	
Air Blank		0.000	04:20	
Control Test		0.079	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLANCO,MARCOS
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: HINO

DATE/TIME OF TEST: 05/26/2016 18:44:39

Diagnostics Check		OK	18:36
Air Blank		0.000	18:37
Control Test		0.080	18:37
Air Blank		0.000	18:37
Subject Sample	# 1	0.203	18:39
Breath Volume		1.921	18:39
Air Blank		0.000	18:40
Air Blank		0.000	18:42
Subject Sample	# 2	0.195	18:43
Breath Volume		1.679	18:43
Air Blank		0.000	18:43
Control Test		0.079	18:44
Air Blank		0.000	18:44
Diagnostics Check		OK	18:44

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SOUCY

DATE/TIME OF TEST: 05/27/2016 22:11:50

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.121	22:06
Breath Volume		2.183	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.118	22:10
Breath Volume		3.121	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:11
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALEGRIA,RAMIRO
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ROSALES-PINOTH

DATE/TIME OF TEST: 05/28/2016 01:11:41

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:58	
Control Test		0.079	00:58	
Air Blank		0.000	00:58	
Subject Sample	# 1		01:01	No Sample Provided
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 2	0.165	01:05	
Breath Volume		1.953	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 3	0.163	01:10	
Breath Volume		2.304	01:10	
Air Blank		0.000	01:10	
Control Test		0.079	01:11	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CUA

DATE/TIME OF TEST: 05/29/2016 00:44:42

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.080	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.192	00:39
Breath Volume		2.117	00:39
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.185	00:43
Breath Volume		1.359	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 05/29/2016 04:04:07

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.159	03:59
Breath Volume		1.816	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.160	04:02
Breath Volume		1.777	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/30/2016 03:45:58

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.207	03:40
Breath Volume		1.382	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.202	03:44
Breath Volume		1.363	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MONROY,SHARISKA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: ANDRES-DOMINGO

DATE/TIME OF TEST: 05/30/2016 21:07:49

Diagnostics Check		OK	21:00
Air Blank		0.000	21:00
Control Test		0.081	21:00
Air Blank		0.000	21:01
Subject Sample	# 1	0.341	21:02
Breath Volume		1.929	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.339	21:06
Breath Volume		2.265	21:06
Air Blank		0.000	21:06
Control Test		0.079	21:07
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WATSON,DAVE
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: SEARS JR

DATE/TIME OF TEST: 05/31/2016 04:38:35

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.080	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.122	04:34
Breath Volume		1.742	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.130	04:37
Breath Volume		2.816	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: KEHRT

DATE/TIME OF TEST: 05/07/2016 20:11:56

Diagnostics Check		OK	20:08	
Air Blank		0.000	20:08	
Control Test		0.079	20:09	
Air Blank		0.000	20:09	
Subject Sample	# 1		20:10	Subject Test Refused
Breath Volume		0.000	20:10	
Air Blank		0.000	20:11	
Control Test		0.079	20:11	
Air Blank		0.000	20:11	
Diagnostics Check		OK	20:11	

Data Download Date/Time: 05/25/2016 09:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 05/09/2016 02:01:00

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.080	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.080	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 05/25/2016 09:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: FERDINAND

DATE/TIME OF TEST: 05/17/2016 04:29:55

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:19	
Control Test		0.079	04:19	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:22	No Sample Provided
Breath Volume		0.000	04:22	
Air Blank		0.000	04:23	
Air Blank		0.000	04:25	
Subject Sample	# 2		04:28	No Sample Provided
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.080	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 05/25/2016 09:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SECORD,RYAN,T
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 15:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:39

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/25/2016 16:13:25

Diagnostics Check		OK	16:10	
Air Blank		0.000	16:10	
Control Test		0.080	16:11	
Air Blank		0.000	16:11	
Subject Sample	# 1		16:12	Subject Test Refused
Breath Volume		0.000	16:12	
Air Blank		0.000	16:12	
Control Test		0.080	16:12	
Air Blank		0.000	16:13	
Diagnostics Check		OK	16:13	

Data Download Date/Time: 07/01/2016 11:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ANNESE,MORRIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PRESTON

DATE/TIME OF TEST: 05/29/2016 01:51:46

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.251	01:47
Breath Volume		2.562	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.249	01:50
Breath Volume		2.304	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 07/01/2016 11:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ANNESE,MORRIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 05/29/2016 22:47:59

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:45	
Control Test		0.078	22:45	
Air Blank		0.000	22:46	
Subject Sample	# 1		22:46	Subject Test Refused
Breath Volume		0.000	22:46	
Air Blank		0.000	22:47	
Control Test		0.079	22:47	
Air Blank		0.000	22:47	
Diagnostics Check		OK	22:47	

Data Download Date/Time: 07/01/2016 11:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHURCH,JOHN,M
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHURCH,JOHN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 05/02/2016 04:07:24

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.000	04:03
Breath Volume		2.816	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.000	04:06
Breath Volume		2.882	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 05/10/2016 08:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUDAH,PATTY,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: HEWETT,MIKE
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DOWDLE

DATE/TIME OF TEST: 05/05/2016 03:01:46

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1	0.161	02:50	Volume Not Met
Breath Volume		0.261	02:50	
Air Blank		0.000	02:51	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.170	02:54	
Breath Volume		1.585	02:54	
Air Blank		0.000	02:55	
Air Blank		0.000	02:57	
Subject Sample	# 3	0.132	03:00	Volume Not Met
Breath Volume		0.371	03:00	
Air Blank		0.000	03:00	
Control Test		0.079	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 05/10/2016 08:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: ROBERSON

DATE/TIME OF TEST: 05/05/2016 23:58:30

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:47	
Control Test		0.080	23:47	
Air Blank		0.000	23:48	
Subject Sample	# 1		23:51	No Sample Provided
Breath Volume		0.000	23:51	
Air Blank		0.000	23:52	
Air Blank		0.000	23:54	
Subject Sample	# 2		23:57	No Sample Provided
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Control Test		0.080	23:57	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 05/10/2016 08:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOUD,JOHN,D
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LARA,SILVIA
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07014080A2
DRY GAS EXPIRY DATE: 05/01/2016
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 05/26/2016 23:51:20

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.233	23:46
Breath Volume		3.535	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.242	23:49
Breath Volume		3.109	23:49
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 06/07/2016 12:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ANGELO,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: RIVERA,ANGELO
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 05/28/2016 23:41:21

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.114	23:36
Breath Volume		1.273	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:38
Subject Sample	# 2	0.102	23:39
Breath Volume		1.429	23:39
Air Blank		0.000	23:40
Control Test		0.079	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 06/07/2016 12:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REIMUNDY,BRYAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: REIMUNDY,BRYAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: CUSHMAN

DATE/TIME OF TEST: 05/01/2016 21:05:11

Diagnostics Check		OK	20:58
Air Blank		0.000	20:59
Control Test		0.079	20:59
Air Blank		0.000	21:00
Subject Sample	# 1	0.144	21:00
Breath Volume		2.105	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.136	21:03
Breath Volume		2.074	21:03
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:05
Diagnostics Check		OK	21:05

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 05/03/2016 06:36:08

Diagnostics Check		OK	06:33	
Air Blank		0.000	06:33	
Control Test		0.081	06:34	
Air Blank		0.000	06:34	
Subject Sample	# 1		06:34	Subject Test Refused
Breath Volume		0.000	06:34	
Air Blank		0.000	06:35	
Control Test		0.080	06:35	
Air Blank		0.000	06:36	
Diagnostics Check		OK	06:36	

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MCDOWELL

DATE/TIME OF TEST: 05/04/2016 03:43:05

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.080	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.160	03:38
Breath Volume		2.914	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.160	03:41
Breath Volume		2.519	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 05/31/2016 15:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERS,WILLIAM,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CHAMBERS,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: WAXMAN

DATE/TIME OF TEST: 05/04/2016 18:07:09

Diagnostics Check		OK	18:00
Air Blank		0.000	18:00
Control Test		0.081	18:00
Air Blank		0.000	18:01
Subject Sample	# 1	0.090	18:02
Breath Volume		1.371	18:02
Air Blank		0.000	18:02
Air Blank		0.000	18:04
Subject Sample	# 2	0.087	18:05
Breath Volume		1.605	18:05
Air Blank		0.000	18:06
Control Test		0.078	18:06
Air Blank		0.000	18:07
Diagnostics Check		OK	18:07

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SHORT

DATE/TIME OF TEST: 05/05/2016 03:57:45

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:55	
Control Test		0.080	03:55	
Air Blank		0.000	03:56	
Subject Sample	# 1		03:56	Subject Test Refused
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Control Test		0.079	03:57	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMPTER,SCOTT,A
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: TEAL,WILLIAM
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 05/08/2016 03:31:32

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.080	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:30	No Sample Provided
Breath Volume		0.000	03:30	
Air Blank		0.000	03:30	
Air Blank		0.000	03:31	

Data Download Date/Time: 05/31/2016 15:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYSINGER,CHRISTIE,
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BEARY,TIMOTHY
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 05/08/2016 19:50:35

Diagnostics Check		OK	19:41
Air Blank		0.000	19:41
Control Test		0.080	19:41
Air Blank		0.000	19:42
Subject Sample	# 1	0.230	19:45
Breath Volume		1.746	19:45
Air Blank		0.000	19:46
Air Blank		0.000	19:47
Subject Sample	# 2	0.233	19:48
Breath Volume		2.042	19:48
Air Blank		0.000	19:49
Control Test		0.077	19:49
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,RICK,A
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: HARMON,RICK
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: MAGPIONG

DATE/TIME OF TEST: 05/12/2016 17:36:23

Diagnostics Check		OK	17:26	
Air Blank		0.000	17:26	
Control Test		0.080	17:27	
Air Blank		0.000	17:27	
Subject Sample	# 1	0.169	17:30	Volume Not Met
Breath Volume		0.031	17:30	
Air Blank		0.000	17:31	
Air Blank		0.000	17:33	
Subject Sample	# 2		17:34	Subject Test Refused
Breath Volume		0.085	17:34	
Air Blank		0.000	17:35	
Control Test		0.077	17:35	
Air Blank		0.000	17:36	
Diagnostics Check		OK	17:36	

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARPER,BILLY,J
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: OCONNORS,C
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: PICINICH

DATE/TIME OF TEST: 05/21/2016 19:17:05

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.080	19:11
Air Blank		0.000	19:11
Subject Sample	# 1	0.231	19:12
Breath Volume		2.859	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:14
Subject Sample	# 2	0.229	19:15
Breath Volume		2.667	19:15
Air Blank		0.000	19:16
Control Test		0.079	19:16
Air Blank		0.000	19:16
Diagnostics Check		OK	19:17

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALINSKI,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: JAMES,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 05/30/2016 23:08:47

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.000	23:04
Breath Volume		1.281	23:04
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.000	23:07
Breath Volume		1.816	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:08
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALINSKI'TIMOTHY,A,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: HICKMAN,J
ARRESTING OFFICER AGENCY: Groveland PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: DYAL

DATE/TIME OF TEST: 05/31/2016 02:11:34

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.191	02:05
Breath Volume		1.589	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.200	02:09
Breath Volume		3.042	02:09
Air Blank		0.000	02:10
Control Test		0.079	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLLERY,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: DOLLERY,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MAHON

DATE/TIME OF TEST: 05/31/2016 04:36:07

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.080	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:34	Subject Test Refused
Breath Volume		0.000	04:34	
Air Blank		0.000	04:35	
Control Test		0.079	04:35	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLLERY,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,JAKE
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: RIDDELL

DATE/TIME OF TEST: 05/31/2016 05:04:31

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.168	04:59
Breath Volume		1.347	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.180	05:02
Breath Volume		1.355	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 05/31/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADEMACHER,RICHARD,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: MURPHEY,DAVID
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/04/2016 21:17:39

Diagnostics Check		OK	21:10
Air Blank		0.000	21:11
Control Test		0.080	21:11
Air Blank		0.000	21:11
Subject Sample	# 1	0.240	21:12
Breath Volume		2.078	21:12
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.245	21:16
Breath Volume		1.410	21:16
Air Blank		0.000	21:16
Control Test		0.079	21:17
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 05/20/2016 00:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADEMACHER,RICHARD,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: RADEMACHER,RICHARD
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: HARTSELL

DATE/TIME OF TEST: 05/08/2016 02:39:24

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.152	02:34
Breath Volume		1.550	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.169	02:37
Breath Volume		2.074	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 05/20/2016 00:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHEY SR,DAVID,G
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: LANK,G
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 05/14/2016 02:51:52

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.079	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.000	02:47
Breath Volume		2.316	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.000	02:50
Breath Volume		1.679	02:50
Air Blank		0.000	02:50
Control Test		0.079	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 05/20/2016 00:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RADEMACHER,RICHARD,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: WILMOT,MEEGAN
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 04/26/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MOHAN

DATE/TIME OF TEST: 05/16/2016 01:56:24

Diagnostics Check		OK	01:48
Air Blank		0.000	01:49
Control Test		0.079	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.187	01:51
Breath Volume		1.660	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.190	01:54
Breath Volume		2.113	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 05/20/2016 00:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANK,WILLIAM,
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: FRANK,WILLIAM
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 05/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MCKENNA

DATE/TIME OF TEST: 05/21/2016 04:19:55

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.232	04:14
Breath Volume		2.292	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.228	04:18
Breath Volume		1.656	04:18
Air Blank		0.000	04:19
Control Test		0.079	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 06/07/2016 05:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: MENCKE,JEFFREY
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 05/20/2016 22:57:48

Diagnostics Check		OK	22:51
Air Blank		0.000	22:52
Control Test		0.076	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.169	22:53
Breath Volume		1.718	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:55
Subject Sample	# 2	0.167	22:56
Breath Volume		2.089	22:56
Air Blank		0.000	22:56
Control Test		0.075	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 06/12/2016 08:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAIGHT,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STRAIGHT,KAYLA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: YOVANOVICH

DATE/TIME OF TEST: 05/23/2016 00:20:32

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.082	00:15
Air Blank		0.000	00:15
Subject Sample	# 1	0.000	00:16
Breath Volume		1.957	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.000	00:19
Breath Volume		2.812	00:19
Air Blank		0.000	00:19
Control Test		0.082	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 05/31/2016 14:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAIGHT,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STRAIGHT,KAYLA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: BENEZETTE

DATE/TIME OF TEST: 05/26/2016 21:51:21

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.082	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.000	21:47
Breath Volume		1.675	21:47
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.000	21:49
Breath Volume		2.234	21:49
Air Blank		0.000	21:50
Control Test		0.081	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 05/31/2016 14:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SEXTON,JACKIE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 05/28/2016 00:12:44

Diagnostics Check		OK	00:03
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.247	00:04
Breath Volume		1.406	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.223	00:07
Breath Volume		1.210	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 3	0.236	00:11
Breath Volume		2.464	00:11
Air Blank		0.000	00:11
Control Test		0.081	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 05/31/2016 14:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAIGHT,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STRAIGHT,KAYLA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/30/2016 19:40:03

Diagnostics Check		OK	19:32
Air Blank		0.000	19:32
Control Test		0.082	19:32
Air Blank		0.000	19:33
Subject Sample	# 1	0.000	19:35
Breath Volume		2.914	19:35
Air Blank		0.000	19:36
Air Blank		0.000	19:38
Subject Sample	# 2	0.000	19:38
Breath Volume		3.191	19:38
Air Blank		0.000	19:39
Control Test		0.082	19:39
Air Blank		0.000	19:39
Diagnostics Check		OK	19:40

Data Download Date/Time: 05/31/2016 14:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAIGHT,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STRAIGHT,KAYLA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/30/2016 19:51:27

Diagnostics Check		OK	19:45
Air Blank		0.000	19:45
Control Test		0.081	19:46
Air Blank		0.000	19:46
Subject Sample	# 1	0.000	19:47
Breath Volume		1.718	19:47
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.000	19:50
Breath Volume		1.531	19:50
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 05/31/2016 14:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAIGHT,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STRAIGHT,KAYLA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/31/2016 23:16:08

Diagnostics Check		OK	23:09
Air Blank		0.000	23:10
Control Test		0.082	23:10
Air Blank		0.000	23:11
Subject Sample	# 1	0.000	23:11
Breath Volume		2.394	23:11
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.000	23:14
Breath Volume		2.601	23:14
Air Blank		0.000	23:15
Control Test		0.082	23:15
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: SCARSELLA,JERAMY
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 05/04/2016 20:35:49

Diagnostics Check		OK	20:27
Air Blank		0.000	20:27
Control Test		0.080	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.011	20:30
Breath Volume		2.011	20:30
Air Blank		0.000	20:31
Air Blank		0.000	20:33
Subject Sample	# 2	0.008	20:34
Breath Volume		1.695	20:34
Air Blank		0.000	20:34
Control Test		0.081	20:35
Air Blank		0.000	20:35
Diagnostics Check		OK	20:35

Data Download Date/Time: 05/31/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,KEVIN,R
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: CHARLA,MICHAEL
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MCDONNELL

DATE/TIME OF TEST: 05/13/2016 18:09:01

Diagnostics Check		OK	17:57	
Air Blank		0.000	17:57	
Control Test		0.080	17:57	
Air Blank		0.000	17:58	
Subject Sample	# 1	0.231	17:58	
Breath Volume		1.792	17:58	
Air Blank		0.000	17:59	
Air Blank		0.000	18:01	
Subject Sample	# 2	0.206	18:04	Volume Not Met
Breath Volume		0.304	18:04	
Air Blank		0.000	18:05	
Air Blank		0.000	18:07	
Subject Sample	# 3	0.224	18:07	
Breath Volume		1.160	18:07	
Air Blank		0.000	18:08	
Control Test		0.080	18:08	
Air Blank		0.000	18:08	
Diagnostics Check		OK	18:09	

Data Download Date/Time: 05/31/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MELODY,FRANCIS
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MENOYO-MONTES

DATE/TIME OF TEST: 05/18/2016 22:01:55

Diagnostics Check		OK	21:59	
Air Blank		0.000	21:59	
Control Test		0.079	21:59	
Air Blank		0.000	22:00	
Subject Sample	# 1		22:00	Subject Test Refused
Breath Volume		0.000	22:00	
Air Blank		0.000	22:01	
Control Test		0.079	22:01	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 05/31/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVES,JOSEPH,A
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: SCARSELLA,JERAMY
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 05/23/2016 04:15:03

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.000	04:10
Breath Volume		1.402	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.000	04:13
Breath Volume		1.371	04:13
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:15

Data Download Date/Time: 05/31/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: MELODY,FRANCIS
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: COBB IV

DATE/TIME OF TEST: 05/25/2016 03:40:51

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:39	Subject Test Refused
Breath Volume		0.000	03:39	
Air Blank		0.000	03:39	
Control Test		0.080	03:40	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 05/31/2016 15:02

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LOVERAS,DAISY
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: REWI

DATE/TIME OF TEST: 05/25/2016 04:47:36

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.079	04:37	
Air Blank		0.000	04:37	
Subject Sample	# 1	0.156	04:40	Volume Not Met
Breath Volume		0.406	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2		04:46	No Sample Provided
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.078	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 06/23/2016 15:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUMPKINS,HENRY
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 05/27/2016 02:50:27

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.000	02:46
Breath Volume		2.187	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.000	02:49
Breath Volume		2.785	02:49
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 06/23/2016 15:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 05/27/2016 05:01:29

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.079	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.212	04:55
Breath Volume		3.003	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.207	04:59
Breath Volume		2.949	04:59
Air Blank		0.000	05:00
Control Test		0.079	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 06/23/2016 15:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BONILLA,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: DESAUTELS

DATE/TIME OF TEST: 05/27/2016 17:52:54

Diagnostics Check		OK	17:46
Air Blank		0.000	17:47
Control Test		0.079	17:47
Air Blank		0.000	17:48
Subject Sample	# 1	0.000	17:48
Breath Volume		2.160	17:48
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.000	17:51
Breath Volume		1.714	17:51
Air Blank		0.000	17:52
Control Test		0.079	17:52
Air Blank		0.000	17:52
Diagnostics Check		OK	17:52

Data Download Date/Time: 06/23/2016 15:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: W426165784180

DATE/TIME OF TEST: 05/27/2016 21:38:43

Diagnostics Check		OK	21:32
Air Blank		0.000	21:32
Control Test		0.079	21:33
Air Blank		0.000	21:33
Subject Sample	# 1	0.336	21:34
Breath Volume		2.812	21:34
Air Blank		0.000	21:34
Air Blank		0.000	21:36
Subject Sample	# 2	0.327	21:37
Breath Volume		2.707	21:37
Air Blank		0.000	21:37
Control Test		0.078	21:38
Air Blank		0.000	21:38
Diagnostics Check		OK	21:38

Data Download Date/Time: 06/23/2016 15:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RUEDA-GARCIA

DATE/TIME OF TEST: 05/28/2016 00:30:58

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.180	00:26
Breath Volume		2.546	00:26
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.177	00:29
Breath Volume		2.371	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 05/28/2016 22:05:04

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.078	21:59
Air Blank		0.000	22:00
Subject Sample	# 1	0.273	22:00
Breath Volume		1.410	22:00
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.267	22:03
Breath Volume		1.718	22:03
Air Blank		0.000	22:04
Control Test		0.078	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:05

Data Download Date/Time: 06/23/2016 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CAROTTI,MOISES
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 05/29/2016 01:20:59

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.077	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.239	01:16
Breath Volume		1.980	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.235	01:19
Breath Volume		1.593	01:19
Air Blank		0.000	01:20
Control Test		0.076	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/29/2016 02:31:43

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.106	02:26
Breath Volume		1.628	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.101	02:30
Breath Volume		1.656	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 06/23/2016 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 05/30/2016 01:46:00

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.224	01:39
Breath Volume		1.781	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.226	01:44
Breath Volume		1.445	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:46

Data Download Date/Time: 06/23/2016 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: VALLEJO

DATE/TIME OF TEST: 05/30/2016 18:53:35

Diagnostics Check		OK	18:41	
Air Blank		0.000	18:42	
Control Test		0.080	18:42	
Air Blank		0.000	18:42	
Subject Sample	# 1	0.230	18:43	
Breath Volume		3.945	18:43	
Air Blank		0.000	18:44	
Air Blank		0.000	18:46	
Subject Sample	# 2	0.156	18:49	Volume Not Met
Breath Volume		0.621	18:49	
Air Blank		0.000	18:49	
Air Blank		0.000	18:51	
Subject Sample	# 3	0.219	18:52	
Breath Volume		4.539	18:52	
Air Blank		0.000	18:52	
Control Test		0.078	18:53	
Air Blank		0.000	18:53	
Diagnostics Check		OK	18:53	

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOUGHERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: TATIS SILVAGNOLI

DATE/TIME OF TEST: 05/06/2016 03:10:07

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:05
Subject Sample	# 1	0.154	03:05
Breath Volume		1.910	03:05
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.157	03:08
Breath Volume		3.171	03:08
Air Blank		0.000	03:09
Control Test		0.078	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 05/10/2016 11:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECTON,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 05/06/2016 23:48:11

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.000	23:43
Breath Volume		1.199	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.000	23:46
Breath Volume		1.367	23:46
Air Blank		0.000	23:47
Control Test		0.078	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 05/10/2016 11:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NIEVES,CHARLIE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: JUARES

DATE/TIME OF TEST: 05/07/2016 04:18:20

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.117	04:13
Breath Volume		2.714	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.117	04:16
Breath Volume		2.816	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 05/10/2016 11:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 05/07/2016 04:41:35

Diagnostics Check		OK	04:35
Air Blank		0.000	04:35
Control Test		0.078	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.166	04:37
Breath Volume		2.554	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.160	04:39
Breath Volume		1.992	04:39
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 05/10/2016 11:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: JIMENEZMOYA

DATE/TIME OF TEST: 05/07/2016 05:22:28

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.079	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.167	05:17
Breath Volume		3.500	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:20
Subject Sample	# 2	0.169	05:20
Breath Volume		3.351	05:20
Air Blank		0.000	05:21
Control Test		0.077	05:21
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 05/10/2016 11:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 05/11/2016 00:18:36

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.197	00:14
Breath Volume		1.910	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.195	00:16
Breath Volume		2.214	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 05/19/2016 16:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/13/2016 04:11:44

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.079	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.138	04:07
Breath Volume		1.199	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.144	04:10
Breath Volume		1.484	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 05/19/2016 16:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELEON,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 05/16/2016 20:58:04

Diagnostics Check		OK	20:51
Air Blank		0.000	20:51
Control Test		0.079	20:51
Air Blank		0.000	20:52
Subject Sample	# 1	0.266	20:53
Breath Volume		2.296	20:53
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.251	20:56
Breath Volume		1.343	20:56
Air Blank		0.000	20:57
Control Test		0.078	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:58

Data Download Date/Time: 05/19/2016 16:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: LYONS

DATE/TIME OF TEST: 05/18/2016 02:06:33

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.116	02:02
Breath Volume		3.261	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.099	02:04
Breath Volume		1.144	02:04
Air Blank		0.000	02:05
Control Test		0.077	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 05/19/2016 16:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BUTT

DATE/TIME OF TEST: 05/22/2016 03:45:42

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.057	03:36	
Breath Volume		2.808	03:36	
Air Blank		0.000	03:36	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.048	03:41	Volume Not Met
Breath Volume		0.808	03:41	
Air Blank		0.000	03:42	
Air Blank		0.000	03:44	
Subject Sample	# 3		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.078	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 05/25/2016 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE,SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: VELAZQUEZ LABOY

DATE/TIME OF TEST: 05/24/2016 01:07:12

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.134	01:02
Breath Volume		1.882	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.137	01:05
Breath Volume		1.640	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 05/25/2016 16:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MULLER,VERONICA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BENDER

DATE/TIME OF TEST: 05/25/2016 02:34:22

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.200	02:29
Breath Volume		2.777	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.197	02:32
Breath Volume		2.312	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 05/25/2016 16:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: VIVANCO

DATE/TIME OF TEST: 05/26/2016 02:35:36

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.145	02:30
Breath Volume		1.339	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.142	02:33
Breath Volume		1.820	02:33
Air Blank		0.000	02:34
Control Test		0.080	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 06/03/2016 09:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: BASTERRECHEA

DATE/TIME OF TEST: 05/28/2016 00:53:12

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.000	00:48
Breath Volume		3.707	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.000	00:51
Breath Volume		3.507	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 06/03/2016 09:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBLEDO,MIGUEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: MACEDORAMIREZ

DATE/TIME OF TEST: 05/29/2016 22:43:17

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.078	22:37
Air Blank		0.000	22:37
Subject Sample	# 1	0.173	22:38
Breath Volume		1.132	22:38
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.168	22:41
Breath Volume		1.539	22:41
Air Blank		0.000	22:42
Control Test		0.079	22:42
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 06/03/2016 09:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EWABADANIELS,ELIZABETH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 05/30/2016 13:25:14

Diagnostics Check		OK	13:18
Air Blank		0.000	13:19
Control Test		0.079	13:19
Air Blank		0.000	13:20
Subject Sample	# 1	0.358	13:20
Breath Volume		1.597	13:20
Air Blank		0.000	13:21
Air Blank		0.000	13:23
Subject Sample	# 2	0.355	13:23
Breath Volume		2.234	13:23
Air Blank		0.000	13:24
Control Test		0.077	13:24
Air Blank		0.000	13:25
Diagnostics Check		OK	13:25

Data Download Date/Time: 06/03/2016 09:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: CHASIN

DATE/TIME OF TEST: 05/30/2016 16:48:11

Diagnostics Check		OK	16:41
Air Blank		0.000	16:42
Control Test		0.078	16:42
Air Blank		0.000	16:43
Subject Sample	# 1	0.040	16:43
Breath Volume		3.835	16:43
Air Blank		0.000	16:44
Air Blank		0.000	16:46
Subject Sample	# 2	0.039	16:46
Breath Volume		4.191	16:46
Air Blank		0.000	16:47
Control Test		0.077	16:47
Air Blank		0.000	16:48
Diagnostics Check		OK	16:48

Data Download Date/Time: 06/03/2016 09:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OSORNO,EDGARD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/31/2016 20:39:08

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.078	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.000	20:34
Breath Volume		2.476	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:37
Subject Sample	# 2	0.000	20:37
Breath Volume		3.683	20:37
Air Blank		0.000	20:38
Control Test		0.077	20:38
Air Blank		0.000	20:39
Diagnostics Check		OK	20:39

Data Download Date/Time: 06/03/2016 09:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MORALES ACEVEDO

DATE/TIME OF TEST: 05/06/2016 02:51:54

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.080	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1	0.093	02:43	Volume Not Met
Breath Volume		0.246	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 2	0.097	02:47	
Breath Volume		1.871	02:47	
Air Blank		0.000	02:47	
Air Blank		0.000	02:49	
Subject Sample	# 3	0.102	02:50	
Breath Volume		1.199	02:50	
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUTTO,ROSS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CINTRON ROSARIO

DATE/TIME OF TEST: 05/06/2016 04:48:11

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.080	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.149	04:43
Breath Volume		3.187	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.157	04:46
Breath Volume		3.269	04:46
Air Blank		0.000	04:47
Control Test		0.079	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JAMES,SETH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: GAINER

DATE/TIME OF TEST: 05/06/2016 23:26:02

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.191	23:21
Breath Volume		2.214	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:23
Subject Sample	# 2	0.197	23:24
Breath Volume		1.976	23:24
Air Blank		0.000	23:25
Control Test		0.079	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:26

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATTS,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: NAIDOO

DATE/TIME OF TEST: 05/07/2016 04:13:11

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.081	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1	0.128	04:05	Volume Not Met
Breath Volume		0.148	04:05	
Air Blank		0.000	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 2	0.141	04:08	
Breath Volume		1.593	04:08	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 3	0.133	04:11	
Breath Volume		1.273	04:11	
Air Blank		0.000	04:12	
Control Test		0.079	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,ELVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: SANCHEZ CONTRERAS

DATE/TIME OF TEST: 05/07/2016 04:44:26

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.080	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.129	04:39
Breath Volume		2.406	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.134	04:42
Breath Volume		2.582	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECTON,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CUTLIP

DATE/TIME OF TEST: 05/08/2016 00:07:52

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.080	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.000	00:03
Breath Volume		3.109	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.000	00:06
Breath Volume		2.980	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 05/10/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BENTIL,GERALD
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MCCRIMON

DATE/TIME OF TEST: 05/08/2016 03:03:01

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.196	02:57
Breath Volume		2.031	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.191	03:01
Breath Volume		2.359	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:03

Data Download Date/Time: 05/10/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROTHELL,HUNTER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: STYLES

DATE/TIME OF TEST: 05/08/2016 04:44:42

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.080	04:39
Air Blank		0.000	04:39
Subject Sample	# 1	0.013	04:40
Breath Volume		1.457	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.012	04:43
Breath Volume		1.128	04:43
Air Blank		0.000	04:43
Control Test		0.080	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBBIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: MADERA

DATE/TIME OF TEST: 05/08/2016 07:07:50

Diagnostics Check		OK	07:01
Air Blank		0.000	07:01
Control Test		0.080	07:01
Air Blank		0.000	07:02
Subject Sample	# 1	0.162	07:03
Breath Volume		1.707	07:03
Air Blank		0.000	07:03
Air Blank		0.000	07:05
Subject Sample	# 2	0.167	07:06
Breath Volume		1.640	07:06
Air Blank		0.000	07:06
Control Test		0.080	07:07
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PINO,ALJANDRO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 04/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: FANORD

DATE/TIME OF TEST: 05/09/2016 22:14:25

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:12	
Control Test		0.079	22:12	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.000	22:13	
Air Blank		0.000	22:13	
Control Test		0.079	22:13	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 05/10/2016 11:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BEAUCHMAN

DATE/TIME OF TEST: 05/14/2016 02:07:49

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.080	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.149	02:03
Breath Volume		2.839	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:05
Subject Sample	# 2	0.145	02:06
Breath Volume		2.660	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 05/19/2016 16:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WALCZAK,KEVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 05/14/2016 04:48:48

Diagnostics Check		OK	04:37	
Air Blank		0.000	04:37	
Control Test		0.081	04:38	
Air Blank		0.000	04:38	
Subject Sample	# 1		04:41	No Sample Provided
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Air Blank		0.000	04:44	
Subject Sample	# 2		04:47	No Sample Provided
Breath Volume		0.000	04:47	
Air Blank		0.000	04:47	
Control Test		0.081	04:48	
Air Blank		0.000	04:48	
Diagnostics Check		OK	04:48	

Data Download Date/Time: 05/19/2016 16:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANTOS,EDWIN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 05/14/2016 06:12:53

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.081	06:07
Air Blank		0.000	06:08
Subject Sample	# 1	0.182	06:08
Breath Volume		1.175	06:08
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.174	06:11
Breath Volume		1.378	06:11
Air Blank		0.000	06:11
Control Test		0.080	06:12
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 05/19/2016 16:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:49

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 05/14/2016 09:18:58

Diagnostics Check		OK	09:12
Air Blank		0.000	09:13
Control Test		0.079	09:13
Air Blank		0.000	09:14
Subject Sample	# 1	0.061	09:14
Breath Volume		2.132	09:14
Air Blank		0.000	09:15
Air Blank		0.000	09:16
Subject Sample	# 2	0.061	09:17
Breath Volume		1.804	09:17
Air Blank		0.000	09:18
Control Test		0.080	09:18
Air Blank		0.000	09:18
Diagnostics Check		OK	09:18

Data Download Date/Time: 05/19/2016 16:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILSON,RUDY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 05/14/2016 22:15:02

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.080	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.239	22:10
Breath Volume		3.117	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.242	22:13
Breath Volume		3.371	22:13
Air Blank		0.000	22:14
Control Test		0.080	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:15

Data Download Date/Time: 05/19/2016 16:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THORNE,MICAH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: TRAN

DATE/TIME OF TEST: 05/15/2016 23:37:36

Diagnostics Check		OK	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.195	23:33
Breath Volume		1.269	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.192	23:36
Breath Volume		2.050	23:36
Air Blank		0.000	23:36
Control Test		0.078	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 05/19/2016 16:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 05/20/2016 04:09:41

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.144	04:05
Breath Volume		3.339	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.129	04:08
Breath Volume		1.769	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 05/25/2016 16:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: ORANHERNANDEZ

DATE/TIME OF TEST: 05/21/2016 02:29:06

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:18	
Control Test		0.080	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1	0.108	02:22	Volume Not Met
Breath Volume		0.160	02:22	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 2	0.098	02:27	Volume Not Met
Breath Volume		0.214	02:27	
Air Blank		0.000	02:28	
Control Test		0.080	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 05/25/2016 16:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO,GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SOCKOL

DATE/TIME OF TEST: 05/22/2016 02:04:30

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.079	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.167	02:00
Breath Volume		1.820	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.165	02:03
Breath Volume		1.277	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 05/25/2016 16:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:02

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 05/22/2016 20:32:55

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.080	20:27
Air Blank		0.000	20:28
Subject Sample	# 1	0.289	20:28
Breath Volume		1.527	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.279	20:31
Breath Volume		1.410	20:31
Air Blank		0.000	20:32
Control Test		0.080	20:32
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 05/25/2016 16:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MALDONADO,NITZA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DIEDERICH

DATE/TIME OF TEST: 05/24/2016 02:52:02

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.000	02:47
Breath Volume		1.503	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.000	02:50
Breath Volume		1.464	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:52

Data Download Date/Time: 05/25/2016 16:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CHERY

DATE/TIME OF TEST: 05/27/2016 03:18:22

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.131	03:14
Breath Volume		1.285	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.126	03:16
Breath Volume		1.218	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EDDY,NATHANAEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: MCGINLEY

DATE/TIME OF TEST: 05/27/2016 23:39:05

Diagnostics Check		OK	23:33
Air Blank		0.000	23:33
Control Test		0.079	23:33
Air Blank		0.000	23:34
Subject Sample	# 1	0.155	23:34
Breath Volume		2.273	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.154	23:37
Breath Volume		2.378	23:37
Air Blank		0.000	23:38
Control Test		0.079	23:38
Air Blank		0.000	23:39
Diagnostics Check		OK	23:39

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,LUIS
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 05/28/2016 02:51:35

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.081	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.000	02:47
Breath Volume		2.582	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.000	02:50
Breath Volume		2.484	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPEER,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/28/2016 03:32:58

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.064	03:28
Breath Volume		2.761	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:30
Subject Sample	# 2	0.061	03:31
Breath Volume		3.332	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WOEHR,PAMELA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: NEGRON

DATE/TIME OF TEST: 05/29/2016 02:01:16

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Subject Sample	# 1	0.229	01:56
Breath Volume		2.367	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.221	01:59
Breath Volume		1.667	01:59
Air Blank		0.000	02:00
Control Test		0.078	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 06/03/2016 09:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: QUILES,JEAN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 05/29/2016 03:50:25

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.033	03:46
Breath Volume		1.769	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.034	03:49
Breath Volume		2.718	03:49
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EWABADANIEL,ELIZABETH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 05/30/2016 13:09:37

Diagnostics Check	OK	13:08	
Air Blank	0.000	13:08	
Control Test	0.073	13:09	Control Outside Tolerance
Air Blank	0.000	13:09	

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: EWABADANIEL,ELIZABETH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:36

SUBJECT'S LAST NAME: FOUNTAIN

DATE/TIME OF TEST: 05/30/2016 13:12:41

Diagnostics Check	OK	13:11	
Air Blank	0.000	13:11	
Control Test	0.073	13:12	Control Outside Tolerance
Air Blank	0.000	13:12	

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO,GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 05/31/2016 05:42:03

Diagnostics Check		OK	05:33
Air Blank		0.000	05:33
Control Test		0.080	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.251	05:37
Breath Volume		1.687	05:37
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.251	05:40
Breath Volume		2.023	05:40
Air Blank		0.000	05:41
Control Test		0.078	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:42

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SEEGOBIN

DATE/TIME OF TEST: 05/11/2016 03:59:23

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:48	
Control Test		0.080	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 1	0.150	03:52	Volume Not Met
Breath Volume		0.125	03:52	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2	0.152	03:57	Volume Not Met
Breath Volume		0.636	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 05/19/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SEEGOBIN

DATE/TIME OF TEST: 05/11/2016 04:11:54

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.078	04:01	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:04	No Sample Provided
Breath Volume		0.000	04:04	
Air Blank		0.000	04:05	
Air Blank		0.000	04:07	
Subject Sample	# 2		04:10	No Sample Provided
Breath Volume		0.000	04:10	
Air Blank		0.000	04:10	
Control Test		0.080	04:11	
Air Blank		0.000	04:11	
Diagnostics Check		OK	04:11	

Data Download Date/Time: 05/19/2016 16:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KILGORE,PETER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:59

SUBJECT'S LAST NAME: HICKS II

DATE/TIME OF TEST: 05/11/2016 14:32:28

Diagnostics Check		OK	14:26
Air Blank		0.000	14:26
Control Test		0.080	14:27
Air Blank		0.000	14:27
Subject Sample	# 1	0.266	14:27
Breath Volume		2.953	14:27
Air Blank		0.000	14:28
Air Blank		0.000	14:30
Subject Sample	# 2	0.263	14:30
Breath Volume		2.453	14:30
Air Blank		0.000	14:31
Control Test		0.078	14:31
Air Blank		0.000	14:32
Diagnostics Check		OK	14:32

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MAENDER

DATE/TIME OF TEST: 05/12/2016 03:51:11

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.170	03:46
Breath Volume		1.484	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.164	03:49
Breath Volume		1.156	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAL,RAJIV
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 05/13/2016 03:06:01

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.209	03:01
Breath Volume		2.562	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.205	03:04
Breath Volume		2.207	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:06

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FRYE,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: SPICER

DATE/TIME OF TEST: 05/13/2016 16:58:52

Diagnostics Check		OK	16:52
Air Blank		0.000	16:53
Control Test		0.078	16:53
Air Blank		0.000	16:53
Subject Sample	# 1	0.280	16:54
Breath Volume		3.593	16:54
Air Blank		0.000	16:55
Air Blank		0.000	16:56
Subject Sample	# 2	0.279	16:57
Breath Volume		3.792	16:57
Air Blank		0.000	16:57
Control Test		0.078	16:58
Air Blank		0.000	16:58
Diagnostics Check		OK	16:58

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: VILLAVERDE

DATE/TIME OF TEST: 05/14/2016 00:56:31

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.000	00:52
Breath Volume		1.855	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.000	00:55
Breath Volume		2.445	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: PINTER

DATE/TIME OF TEST: 05/14/2016 03:28:28

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.156	03:23
Breath Volume		2.558	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.169	03:26
Breath Volume		3.628	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE,PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: PAULLING

DATE/TIME OF TEST: 05/14/2016 04:38:36

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.079	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.134	04:33
Breath Volume		1.140	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.143	04:37
Breath Volume		2.742	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ALTAY

DATE/TIME OF TEST: 05/15/2016 03:15:47

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.159	03:11
Breath Volume		2.398	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.165	03:14
Breath Volume		2.375	03:14
Air Blank		0.000	03:14
Control Test		0.078	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA,EDDIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SPANN JR

DATE/TIME OF TEST: 05/15/2016 03:43:40

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.237	03:38
Breath Volume		1.132	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.246	03:42
Breath Volume		2.316	03:42
Air Blank		0.000	03:42
Control Test		0.077	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WALCZAK,KEVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WOODALL

DATE/TIME OF TEST: 05/15/2016 04:19:02

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.191	04:14
Breath Volume		3.128	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.187	04:17
Breath Volume		3.191	04:17
Air Blank		0.000	04:18
Control Test		0.078	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:19

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: AMOS

DATE/TIME OF TEST: 05/15/2016 05:09:28

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.107	05:05
Breath Volume		2.417	05:05
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.113	05:07
Breath Volume		3.078	05:07
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REARDON,SARAH
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 05/16/2016 01:51:59

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.079	01:41	
Air Blank		0.000	01:41	
Subject Sample	# 1		01:45	No Sample Provided
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 2		01:50	No Sample Provided
Breath Volume		0.000	01:50	
Air Blank		0.000	01:51	
Control Test		0.080	01:51	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 05/19/2016 16:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 05/16/2016 03:57:06

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.218	03:52
Breath Volume		1.972	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.217	03:55
Breath Volume		1.335	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:48

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 05/17/2016 10:24:52

Diagnostics Check		OK	10:11	
Air Blank		0.000	10:12	
Control Test		0.080	10:12	
Air Blank		0.000	10:13	
Subject Sample	# 1	0.146	10:16	Volume Not Met
Breath Volume		0.703	10:16	
Air Blank		0.000	10:17	
Air Blank		0.000	10:18	
Subject Sample	# 2	0.188	10:19	No .020 agreement
Breath Volume		2.101	10:19	
Air Blank		0.000	10:19	
Air Blank		0.000	10:21	
Subject Sample	# 3	0.159	10:23	No .020 agreement
Breath Volume		1.390	10:23	
Air Blank		0.000	10:23	
Control Test		0.078	10:24	
Air Blank		0.000	10:24	
Diagnostics Check		OK	10:24	

Data Download Date/Time: 05/19/2016 16:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:48

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 05/17/2016 10:33:50

Diagnostics Check		OK	10:26
Air Blank		0.000	10:27
Control Test		0.079	10:27
Air Blank		0.000	10:28
Subject Sample	# 1	0.169	10:29
Breath Volume		1.148	10:29
Air Blank		0.000	10:29
Air Blank		0.000	10:31
Subject Sample	# 2	0.168	10:32
Breath Volume		1.496	10:32
Air Blank		0.000	10:32
Control Test		0.078	10:33
Air Blank		0.000	10:33
Diagnostics Check		OK	10:33

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 05/18/2016 03:55:54

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.164	03:51
Breath Volume		1.859	03:51
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.160	03:54
Breath Volume		1.945	03:54
Air Blank		0.000	03:54
Control Test		0.078	03:55
Air Blank		0.000	03:55
Diagnostics Check		OK	03:55

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZITTO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 05/19/2016 02:38:35

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.079	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.113	02:33
Breath Volume		2.996	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:35
Subject Sample	# 2	0.109	02:37
Breath Volume		1.781	02:37
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICK
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: AMANDUS

DATE/TIME OF TEST: 05/19/2016 03:22:32

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.077	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.157	03:18
Breath Volume		2.816	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.157	03:21
Breath Volume		1.636	03:21
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: BRATONOZIC

DATE/TIME OF TEST: 05/19/2016 04:09:28

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.160	04:05
Breath Volume		1.898	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.168	04:07
Breath Volume		2.507	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 05/19/2016 16:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE,PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 05/19/2016 21:09:21

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:55	
Control Test		0.079	20:55	
Air Blank		0.000	20:56	
Subject Sample	# 1	0.162	20:59	No .020 agreement
Breath Volume		1.531	20:59	
Air Blank		0.000	20:59	
Air Blank		0.000	21:01	
Subject Sample	# 2	0.197	21:02	No .020 agreement
Breath Volume		1.312	21:02	
Air Blank		0.000	21:03	
Air Blank		0.000	21:04	
Subject Sample	# 3		21:07	No Sample Provided
Breath Volume		0.000	21:07	
Air Blank		0.000	21:08	
Control Test		0.077	21:08	
Air Blank		0.000	21:09	
Diagnostics Check		OK	21:09	

Data Download Date/Time: 05/25/2016 16:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE,PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 05/19/2016 21:22:32

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.078	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:15	No Sample Provided
Breath Volume		0.000	21:15	
Air Blank		0.000	21:16	
Air Blank		0.000	21:18	
Subject Sample	# 2		21:21	No Sample Provided
Breath Volume		0.000	21:21	
Air Blank		0.000	21:21	
Control Test		0.078	21:21	
Air Blank		0.000	21:22	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 05/25/2016 16:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: TRIGALLY

DATE/TIME OF TEST: 05/20/2016 03:35:31

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.079	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1	0.077	03:25	
Breath Volume		3.148	03:25	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 2		03:31	No Sample Provided
Breath Volume		0.000	03:31	
Air Blank		0.000	03:31	
Air Blank		0.000	03:33	
Subject Sample	# 3		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.079	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 05/25/2016 16:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORBERG,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PROCHET

DATE/TIME OF TEST: 05/20/2016 04:32:33

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.079	04:21	
Air Blank		0.000	04:22	
Subject Sample	# 1	0.213	04:25	Volume Not Met
Breath Volume		0.273	04:25	
Air Blank		0.000	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 2	0.215	04:30	Volume Not Met
Breath Volume		0.324	04:30	
Air Blank		0.000	04:31	
Control Test		0.078	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAIGHT,MONTAGUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/20/2016 08:26:03

Diagnostics Check		OK	08:19
Air Blank		0.000	08:20
Control Test		0.079	08:20
Air Blank		0.000	08:21
Subject Sample	# 1	0.000	08:21
Breath Volume		1.195	08:21
Air Blank		0.000	08:21
Air Blank		0.000	08:23
Subject Sample	# 2	0.000	08:24
Breath Volume		1.550	08:24
Air Blank		0.000	08:25
Control Test		0.079	08:25
Air Blank		0.000	08:25
Diagnostics Check		OK	08:26

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORBERG,JASON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: MORALES BURGOS

DATE/TIME OF TEST: 05/21/2016 02:48:32

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.208	02:44
Breath Volume		1.746	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.211	02:46
Breath Volume		1.343	02:46
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONGSHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/21/2016 03:45:43

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.142	03:41
Breath Volume		1.527	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.152	03:44
Breath Volume		1.179	03:44
Air Blank		0.000	03:44
Control Test		0.079	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMONE,JOHN
ARRESTING OFFICER AGENCY: Eatonville PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 05/21/2016 04:58:28

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.079	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.128	04:53
Breath Volume		1.281	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.122	04:56
Breath Volume		1.113	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROTHELL,HUNTER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/22/2016 02:29:51

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.192	02:25
Breath Volume		2.718	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.179	02:28
Breath Volume		1.320	02:28
Air Blank		0.000	02:28
Control Test		0.078	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BINET,MARLON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SAVARY CUELLO

DATE/TIME OF TEST: 05/22/2016 04:12:24

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.078	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.133	04:07
Breath Volume		2.789	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.130	04:10
Breath Volume		1.589	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LORENZ,JOHN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: ESTREMERA SANTOS

DATE/TIME OF TEST: 05/23/2016 00:22:53

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.080	00:12	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:15	No Sample Provided
Breath Volume		0.000	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2		00:21	No Sample Provided
Breath Volume		0.000	00:21	
Air Blank		0.000	00:21	
Control Test		0.078	00:22	
Air Blank		0.000	00:22	
Diagnostics Check		OK	00:22	

Data Download Date/Time: 05/25/2016 16:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,ELVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: DATZ

DATE/TIME OF TEST: 05/24/2016 00:22:46

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.141	00:18
Breath Volume		1.167	00:18
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.150	00:21
Breath Volume		1.605	00:21
Air Blank		0.000	00:21
Control Test		0.079	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REMBIS-LICARI,JAZMIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: NARANJO

DATE/TIME OF TEST: 05/24/2016 05:17:25

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.084	05:12
Breath Volume		1.175	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.080	05:15
Breath Volume		2.121	05:15
Air Blank		0.000	05:16
Control Test		0.079	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOWSON,AMANDA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: FEJFAR

DATE/TIME OF TEST: 05/25/2016 04:15:10

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.232	04:10
Breath Volume		1.398	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.242	04:13
Breath Volume		1.503	04:13
Air Blank		0.000	04:14
Control Test		0.078	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 05/25/2016 16:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/26/2016 03:31:17

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.124	03:26
Breath Volume		1.335	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.130	03:29
Breath Volume		2.445	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 06/03/2016 09:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WALCZAK,KEVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: GAHUNGU

DATE/TIME OF TEST: 05/27/2016 01:26:14

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.207	01:21
Breath Volume		1.742	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:23
Subject Sample	# 2	0.215	01:24
Breath Volume		1.363	01:24
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 06/03/2016 09:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,ARNOLDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WEDDLE

DATE/TIME OF TEST: 05/27/2016 22:03:47

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:52	
Control Test		0.079	21:52	
Air Blank		0.000	21:53	
Subject Sample	# 1	0.091	21:56	Volume Not Met
Breath Volume		0.132	21:56	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2	0.083	22:02	Volume Not Met
Breath Volume		0.605	22:02	
Air Blank		0.000	22:02	
Control Test		0.078	22:03	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 06/03/2016 09:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,JUAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: LUYANDA

DATE/TIME OF TEST: 05/27/2016 22:57:52

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:42	
Control Test		0.079	22:42	
Air Blank		0.000	22:42	
Subject Sample	# 1	0.268	22:46	Volume Not Met
Breath Volume		0.085	22:46	
Air Blank		0.000	22:46	
Air Blank		0.000	22:48	
Subject Sample	# 2	0.250	22:50	
Breath Volume		1.566	22:50	
Air Blank		0.000	22:51	
Air Blank		0.000	22:53	
Subject Sample	# 3	0.239	22:56	
Breath Volume		3.683	22:56	
Air Blank		0.000	22:56	
Control Test		0.073	22:57	Control Outside Tolerance
Air Blank		0.000	22:57	

Data Download Date/Time: 06/03/2016 09:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,JUAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: LUYANDA

DATE/TIME OF TEST: 05/27/2016 23:08:47

Diagnostics Check		OK	22:59	
Air Blank		0.000	22:59	
Control Test		0.075	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 1	0.244	23:02	
Breath Volume		3.039	23:02	
Air Blank		0.000	23:03	
Air Blank		0.000	23:04	
Subject Sample	# 2	0.239	23:06	
Breath Volume		1.593	23:06	
Air Blank		0.000	23:07	
Control Test		0.074	23:08	Control Outside Tolerance
Air Blank		0.000	23:08	

Data Download Date/Time: 06/03/2016 09:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IMPARATO,FRANK
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 05/28/2016 00:17:36

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.076	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.177	00:13
Breath Volume		1.175	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.168	00:16
Breath Volume		1.207	00:16
Air Blank		0.000	00:16
Control Test		0.075	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 06/03/2016 09:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LACENTRA,DONALD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FLOYD

DATE/TIME OF TEST: 05/28/2016 02:32:08

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.078	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.189	02:27
Breath Volume		2.789	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.191	02:30
Breath Volume		2.882	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 06/03/2016 09:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSS,DAVID
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ENRIQUEZ JR

DATE/TIME OF TEST: 05/28/2016 03:49:27

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.078	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:45	No Sample Provided
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2		03:48	Subject Test Refused
Breath Volume		0.000	03:48	
Air Blank		0.000	03:48	
Control Test		0.078	03:48	
Air Blank		0.000	03:49	
Diagnostics Check		OK	03:49	

Data Download Date/Time: 06/03/2016 09:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZITO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GUILLEN PEREZ

DATE/TIME OF TEST: 05/29/2016 04:14:09

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.079	04:02	
Air Blank		0.000	04:03	
Subject Sample	# 1	0.059	04:06	Volume Not Met
Breath Volume		0.183	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 2	0.073	04:09	
Breath Volume		2.921	04:09	
Air Blank		0.000	04:10	
Air Blank		0.000	04:12	
Subject Sample	# 3	0.069	04:12	
Breath Volume		1.382	04:12	
Air Blank		0.000	04:13	
Control Test		0.078	04:13	
Air Blank		0.000	04:14	
Diagnostics Check		OK	04:14	

Data Download Date/Time: 06/03/2016 09:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HALL,N
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 05/07/2016 01:13:17

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.079	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.194	01:08
Breath Volume		2.621	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.196	01:11
Breath Volume		2.125	01:11
Air Blank		0.000	01:12
Control Test		0.077	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 05/17/2016 20:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: HALL,NICHOLAS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CASIMIRO-CASTILLO

DATE/TIME OF TEST: 05/07/2016 23:07:38

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.079	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.229	23:03
Breath Volume		2.265	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.224	23:06
Breath Volume		2.656	23:06
Air Blank		0.000	23:06
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 05/17/2016 20:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: JONES,CHRISTOPHER
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GEIB

DATE/TIME OF TEST: 05/26/2016 00:14:16

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:11	
Control Test		0.079	00:12	
Air Blank		0.000	00:12	
Subject Sample	# 1		00:12	Subject Test Refused
Breath Volume		0.000	00:12	
Air Blank		0.000	00:13	
Control Test		0.079	00:13	
Air Blank		0.000	00:14	
Diagnostics Check		OK	00:14	

Data Download Date/Time: 06/23/2016 23:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: GLEASON,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 05/08/2016 03:33:56

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.178	03:29
Breath Volume		2.136	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.172	03:32
Breath Volume		2.011	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 05/18/2016 04:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: BRADSHAW,CHRIS
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 05/09/2016 04:44:24

Diagnostics Check		OK	04:37
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.145	04:39
Breath Volume		2.785	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.142	04:42
Breath Volume		2.503	04:42
Air Blank		0.000	04:43
Control Test		0.078	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 05/18/2016 04:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SIMMONS,BRIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BACH

DATE/TIME OF TEST: 05/12/2016 19:36:55

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.079	19:31
Air Blank		0.000	19:32
Subject Sample	# 1	0.000	19:32
Breath Volume		1.375	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.000	19:35
Breath Volume		3.601	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 05/18/2016 04:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: ANDERSON,HUNTER
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 05/17/2016 00:31:20

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.079	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.121	00:26
Breath Volume		1.925	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.119	00:29
Breath Volume		1.488	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 05/18/2016 04:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: HOLBROOK,NICHOLAS
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 05/19/2016 01:21:57

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.078	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.092	01:17
Breath Volume		2.507	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.090	01:20
Breath Volume		3.253	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 06/28/2016 12:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,COREY,M
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: GARNER,TYRELL
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SCHNOTALA

DATE/TIME OF TEST: 05/25/2016 00:42:42

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.079	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.221	00:38
Breath Volume		2.339	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.230	00:41
Breath Volume		3.066	00:41
Air Blank		0.000	00:41
Control Test		0.078	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 06/28/2016 12:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: ANDERSON,HUNTER
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 05/26/2016 01:51:44

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:49	
Control Test		0.079	01:49	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:50	Subject Test Refused
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	
Control Test		0.079	01:51	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 06/28/2016 12:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LYNCH,ROBERT
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: LAPPICO

DATE/TIME OF TEST: 05/26/2016 03:42:03

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.078	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.188	03:33
Breath Volume		2.542	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.164	03:36
Breath Volume		1.750	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 3	0.159	03:40
Breath Volume		1.324	03:40
Air Blank		0.000	03:41
Control Test		0.077	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:42

Data Download Date/Time: 06/28/2016 12:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSE,DAVID,A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: JOHNSON,SHAWN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: FRIDDLE

DATE/TIME OF TEST: 05/30/2016 16:55:46

Diagnostics Check		OK	16:47
Air Blank		0.000	16:47
Control Test		0.078	16:48
Air Blank		0.000	16:48
Subject Sample	# 1	0.389	16:50
Breath Volume		1.628	16:50
Air Blank		0.000	16:50
Air Blank		0.000	16:52
Subject Sample	# 2	0.406	16:54
Breath Volume		1.933	16:54
Air Blank		0.000	16:54
Control Test		0.077	16:55
Air Blank		0.000	16:55
Diagnostics Check		OK	16:55

Data Download Date/Time: 06/28/2016 12:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OBYRNE,JAMES,W
BREATH TEST OPERATOR AGENCY: Waldo PD
ARRESTING OFFICER: OBYRNE,JAMES
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/02/2016 22:41:28

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.080	22:35
Air Blank		0.000	22:35
Subject Sample	# 1	0.000	22:36
Breath Volume		1.601	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.000	22:40
Breath Volume		2.601	22:40
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 05/31/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:32

SUBJECT'S LAST NAME: STAPLES

DATE/TIME OF TEST: 05/03/2016 12:01:54

Diagnostics Check		OK	11:55
Air Blank		0.000	11:56
Control Test		0.080	11:56
Air Blank		0.000	11:56
Subject Sample	# 1	0.000	11:57
Breath Volume		2.757	11:57
Air Blank		0.000	11:58
Air Blank		0.000	12:00
Subject Sample	# 2	0.000	12:00
Breath Volume		2.539	12:00
Air Blank		0.000	12:00
Control Test		0.080	12:01
Air Blank		0.000	12:01
Diagnostics Check		OK	12:01

Data Download Date/Time: 05/31/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HUGHES,JENNA
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RACIK

DATE/TIME OF TEST: 05/07/2016 05:25:20

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.079	05:20
Air Blank		0.000	05:20
Subject Sample	# 1	0.122	05:20
Breath Volume		3.281	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.124	05:23
Breath Volume		3.183	05:23
Air Blank		0.000	05:24
Control Test		0.081	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 05/31/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: THOMAS,ADREAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/12/2016 05:08:27

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:56	
Control Test		0.080	04:56	
Air Blank		0.000	04:57	
Subject Sample	# 1	0.129	05:00	Volume Not Met
Breath Volume		0.351	05:00	
Air Blank		0.000	05:00	
Air Blank		0.000	05:02	
Subject Sample	# 2	0.131	05:03	
Breath Volume		1.316	05:03	
Air Blank		0.000	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 3	0.120	05:06	
Breath Volume		1.375	05:06	
Air Blank		0.000	05:07	
Control Test		0.079	05:07	
Air Blank		0.000	05:08	
Diagnostics Check		OK	05:08	

Data Download Date/Time: 05/31/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,RICARDO,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: NOLES,GABRIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: LOWIE

DATE/TIME OF TEST: 05/15/2016 21:10:43

Diagnostics Check		OK	21:03
Air Blank		0.000	21:03
Control Test		0.079	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.253	21:06
Breath Volume		1.968	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.265	21:08
Breath Volume		2.027	21:08
Air Blank		0.000	21:09
Control Test		0.078	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 05/31/2016 11:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELF,BRETT,R
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: NELSON,ROBERT
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: BEVIS

DATE/TIME OF TEST: 05/16/2016 20:36:10

Diagnostics Check		OK	20:29
Air Blank		0.000	20:29
Control Test		0.080	20:30
Air Blank		0.000	20:30
Subject Sample	# 1	0.000	20:31
Breath Volume		2.007	20:31
Air Blank		0.000	20:31
Air Blank		0.000	20:33
Subject Sample	# 2	0.000	20:34
Breath Volume		2.160	20:34
Air Blank		0.000	20:35
Control Test		0.079	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 05/31/2016 11:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOVSGARD,BRITTNEY,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CLARK,MAEGAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/25/2016 13:11:26

Diagnostics Check		OK	13:08	
Air Blank		0.000	13:08	
Control Test		0.080	13:09	
Air Blank		0.000	13:09	
Subject Sample	# 1		13:10	Subject Test Refused
Breath Volume		0.000	13:10	
Air Blank		0.000	13:10	
Control Test		0.080	13:10	
Air Blank		0.000	13:11	
Diagnostics Check		OK	13:11	

Data Download Date/Time: 05/31/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKOVSGARD,BRITTNEY,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CLARK,MAEGAN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 12:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:34

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 05/25/2016 13:16:21

Diagnostics Check		OK	13:12	
Air Blank		0.000	13:13	
Control Test		0.080	13:13	
Air Blank		0.000	13:14	
Subject Sample	# 1		13:14	Subject Test Refused
Breath Volume		0.000	13:14	
Air Blank		0.000	13:15	
Control Test		0.081	13:15	
Air Blank		0.000	13:16	
Diagnostics Check		OK	13:16	

Data Download Date/Time: 05/31/2016 11:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: ARBELO

DATE/TIME OF TEST: 05/27/2016 19:01:44

Diagnostics Check		OK	18:55
Air Blank		0.000	18:55
Control Test		0.080	18:55
Air Blank		0.000	18:56
Subject Sample	# 1	0.367	18:57
Breath Volume		1.714	18:57
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.367	19:00
Breath Volume		1.148	19:00
Air Blank		0.000	19:00
Control Test		0.080	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 05/31/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,MATTHEW,G
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MILLER,MATTHEW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/29/2016 06:04:30

Diagnostics Check		OK	05:58
Air Blank		0.000	05:58
Control Test		0.079	05:59
Air Blank		0.000	05:59
Subject Sample	# 1	0.258	06:00
Breath Volume		1.437	06:00
Air Blank		0.000	06:00
Air Blank		0.000	06:02
Subject Sample	# 2	0.258	06:02
Breath Volume		1.921	06:02
Air Blank		0.000	06:03
Control Test		0.080	06:03
Air Blank		0.000	06:04
Diagnostics Check		OK	06:04

Data Download Date/Time: 05/31/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 05/01/2016 01:18:10

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.181	01:13
Breath Volume		2.054	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.199	01:16
Breath Volume		3.945	01:16
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 05/22/2016 11:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FINNEGAN

DATE/TIME OF TEST: 05/01/2016 22:51:32

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.078	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.212	22:47
Breath Volume		2.382	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.204	22:50
Breath Volume		2.683	22:50
Air Blank		0.000	22:50
Control Test		0.078	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 05/22/2016 11:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: BONER

DATE/TIME OF TEST: 05/02/2016 01:40:56

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.078	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.126	01:36
Breath Volume		1.644	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.127	01:39
Breath Volume		1.886	01:39
Air Blank		0.000	01:40
Control Test		0.078	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 05/22/2016 11:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 04/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PISCIOTTANO

DATE/TIME OF TEST: 05/07/2016 01:13:19

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Subject Sample	# 1	0.069	01:05	Volume Not Met
Breath Volume		0.046	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.076	01:08	
Breath Volume		1.800	01:08	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 3	0.077	01:11	
Breath Volume		1.750	01:11	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 05/22/2016 11:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: OPIO

DATE/TIME OF TEST: 05/30/2016 21:15:59

Diagnostics Check		OK	21:09
Air Blank		0.000	21:09
Control Test		0.078	21:10
Air Blank		0.000	21:10
Subject Sample	# 1	0.106	21:11
Breath Volume		1.746	21:11
Air Blank		0.000	21:12
Air Blank		0.000	21:14
Subject Sample	# 2	0.111	21:14
Breath Volume		1.722	21:14
Air Blank		0.000	21:15
Control Test		0.077	21:15
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: COMBS

DATE/TIME OF TEST: 05/31/2016 01:02:37

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.078	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.097	00:58
Breath Volume		2.003	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.095	01:01
Breath Volume		1.746	01:01
Air Blank		0.000	01:01
Control Test		0.078	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PACKARD

DATE/TIME OF TEST: 05/31/2016 22:54:53

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:52	
Control Test		0.078	22:52	
Air Blank		0.000	22:53	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.000	22:53	
Air Blank		0.000	22:53	
Control Test		0.077	22:54	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TOMICH,KRISTOFFOR
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: MORENO SERRANO

DATE/TIME OF TEST: 05/15/2016 06:46:44

Diagnostics Check		OK	06:41	
Air Blank		0.000	06:41	
Control Test		0.079	06:41	
Air Blank		0.000	06:42	
Subject Sample	# 1		06:45	Subject Test Refused
Breath Volume		0.000	06:45	
Air Blank		0.000	06:45	
Control Test		0.079	06:46	
Air Blank		0.000	06:46	
Diagnostics Check		OK	06:46	

Data Download Date/Time: 05/28/2016 22:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: NODZON,DAWN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: BROHARD

DATE/TIME OF TEST: 05/28/2016 03:47:38

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.078	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.103	03:43
Breath Volume		2.855	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.103	03:46
Breath Volume		1.996	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 05/28/2016 22:20

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LAQUERRE,TIMOTHY
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/21/2016 00:56:30

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.079	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1	0.364	00:49	
Breath Volume		1.753	00:49	
Air Blank		0.000	00:49	
Air Blank		0.000	00:51	
Subject Sample	# 2	0.338	00:51	
Breath Volume		1.792	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 3	0.358	00:54	
Breath Volume		3.437	00:54	
Air Blank			00:55	Purge Fail
Air Blank			00:56	Purge Fail

Data Download Date/Time: 05/28/2016 22:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LAQUERRE,TIMOTHY
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 05/21/2016 01:10:17

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.077	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.343	01:05
Breath Volume		2.925	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.350	01:08
Breath Volume		2.816	01:08
Air Blank		0.000	01:09
Control Test		0.077	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 05/28/2016 22:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SCHNEIDER,CLARK
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 05/21/2016 01:32:05

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.237	01:27
Breath Volume		3.675	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:29
Subject Sample	# 2	0.237	01:30
Breath Volume		1.699	01:30
Air Blank		0.000	01:31
Control Test		0.078	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 05/28/2016 22:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LYONNAIS,AMANDA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 04/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SCARBOROUGH

DATE/TIME OF TEST: 05/21/2016 02:12:04

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.144	02:07
Breath Volume		3.195	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.139	02:10
Breath Volume		3.195	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 05/28/2016 22:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KOZLOWSKI

DATE/TIME OF TEST: 05/08/2016 03:43:47

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.079	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Control Test		0.079	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 05/25/2016 21:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIVECOAT,CURTIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: BOZARTH

DATE/TIME OF TEST: 05/10/2016 00:00:26

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:57	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1		23:59	Subject Test Refused
Breath Volume		0.000	23:59	
Air Blank		0.000	23:59	
Control Test		0.080	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 05/25/2016 21:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MACKAIL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 05/14/2016 04:58:22

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.079	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.147	04:53
Breath Volume		1.183	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.150	04:56
Breath Volume		1.234	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 05/25/2016 21:11

COBRA ©1997-2014 CMI, Inc.