Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILERA,DANIEL,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: DALTON

DATE/TIME OF TEST: 05/01/2016 01:35:24

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.076	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.235	01:27
Breath Volume		2.558	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:29
Subject Sample	# 2	0.205	01:30
Breath Volume		1.394	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 3	0.198	01:33
Breath Volume		1.117	01:33
Air Blank		0.000	01:34
Control Test		0.075	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 05/17/2016 10:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,RAYMOND,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: BENNINGHAUS

DATE/TIME OF TEST: 05/03/2016 00:37:04

Diagnostics Check		OK	00:25	
Air Blank		0.000	00:26	
Control Test		0.077	00:26	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:30	No Sample Provided
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Air Blank		0.000	00:32	
Subject Sample	# 2		00:35	Subject Test Refused
Breath Volume		0.000	00:35	
Air Blank		0.000	00:36	
Control Test		0.076	00:36	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 05/17/2016 10:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,RAYMOND,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HOPKINS

DATE/TIME OF TEST: 05/03/2016 23:32:51

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.077	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.203	23:28
Breath Volume		1.476	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:30
Subject Sample	# 2	0.203	23:31
Breath Volume		1.945	23:31
Air Blank		0.000	23:31
Control Test		0.077	23:32
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 05/17/2016 10:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,RAYMOND,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ZIELINSKI

DATE/TIME OF TEST: 05/04/2016 02:47:38

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.077	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.220	02:43
Breath Volume		1.406	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.214	02:45
Breath Volume		1.968	02:45
Air Blank		0.000	02:46
Control Test		0.076	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 05/17/2016 10:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,RAYMOND,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 05/06/2016 23:21:53

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.078	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:14	No Sample Provided
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Air Blank		0.000	23:17	
Subject Sample	# 2		23:20	No Sample Provided
Breath Volume		0.000	23:20	
Air Blank		0.000	23:20	
Control Test		0.078	23:21	
Air Blank		0.000	23:21	
Diagnostics Check		OK	23:21	

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIAQUINTO,KRISTIN,
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: MACKERETH,TYLER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/10/2016 00:54:29

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.077	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.105	00:48
Breath Volume		1.500	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.095	00:52
Breath Volume		1.257	00:52
Air Blank		0.000	00:53
Control Test		0.075	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,FREDERICK,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: THABIT,S
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 05/10/2016 03:50:09

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.077	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.102	03:45
Breath Volume		3.050	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.106	03:48
Breath Volume		2.949	03:48
Air Blank		0.000	03:49
Control Test		0.077	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,R,T
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: FRIZONE

DATE/TIME OF TEST: 05/10/2016 22:19:03

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.078	22:13
Air Blank		0.000	22:13
Subject Sample	# 1	0.222	22:14
Breath Volume		1.351	22:14
Air Blank		0.000	22:15
Air Blank		0.000	22:16
Subject Sample	# 2	0.220	22:17
Breath Volume		1.273	22:17
Air Blank		0.000	22:18
Control Test		0.078	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:19

Data Download Date/Time: 05/17/2016 10:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,FREDERICK,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLO,A
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WALLING

DATE/TIME OF TEST: 05/14/2016 01:12:11

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.156	01:07
Breath Volume		2.992	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.151	01:10
Breath Volume		3.039	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 05/14/2016 01:19:14

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.078	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Slope Not Met
Breath Volume		0.750	01:17	
Air Blank		0.000	01:18	
Control Test		0.078	01:18	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,FREDERICK,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LUBERACKI,N
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/14/2016 02:56:30

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.077	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.131	02:51
Breath Volume		3.675	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.143	02:54
Breath Volume		2.402	02:54
Air Blank		0.000	02:55
Control Test		0.077	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,FREDERICK,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GRAY,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 05/14/2016 03:15:09

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.198	03:10
Breath Volume		1.652	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:12
Subject Sample	# 2	0.206	03:13
Breath Volume		1.593	03:13
Air Blank		0.000	03:14
Control Test		0.077	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 05/17/2016 10:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONCALVES,GUSTAVO,R
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN,LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/14/2016 23:06:22

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.077	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.143	23:00
Breath Volume		1.148	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.139	23:04
Breath Volume		1.226	23:04
Air Blank		0.000	23:05
Control Test		0.076	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILERA,DANIEL,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: THOMAS,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 05/15/2016 02:10:13

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.077	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.147	02:03
Breath Volume		1.667	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.161	02:08
Breath Volume		1.734	02:08
Air Blank		0.000	02:09
Control Test		0.076	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 05/17/2016 10:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILERA,DANIEL,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 04/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CORRALES

DATE/TIME OF TEST: 05/15/2016 02:53:24

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:39	
Control Test		0.076	02:40	
Air Blank		0.000	02:40	
Subject Sample	# 1		02:43	No Sample Provided
Breath Volume		0.000	02:43	
Air Blank		0.000	02:44	
Air Blank		0.000	02:46	
Subject Sample	# 2	0.180	02:48	
Breath Volume		2.722	02:48	
Air Blank		0.000	02:49	
Air Blank		0.000	02:51	
Subject Sample	# 3	0.171	02:51	
Breath Volume		2.214	02:51	
Air Blank		0.000	02:52	
Control Test		0.076	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 05/17/2016 10:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUILERA,DANIEL,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: HILES

DATE/TIME OF TEST: 05/18/2016 23:42:33

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.078	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.106	23:37
Breath Volume		2.300	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.103	23:40
Breath Volume		1.890	23:40
Air Blank		0.000	23:41
Control Test		0.077	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KONIECZKI,D,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KONIECZKI,D
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:47

SUBJECT'S LAST NAME: BROOKSHAW

DATE/TIME OF TEST: 05/19/2016 12:17:51

Diagnostics Check		OK	12:11
Air Blank		0.000	12:11
Control Test		0.077	12:12
Air Blank		0.000	12:12
Subject Sample	# 1	0.000	12:13
Breath Volume		2.183	12:13
Air Blank		0.000	12:13
Air Blank		0.000	12:15
Subject Sample	# 2	0.000	12:16
Breath Volume		1.671	12:16
Air Blank		0.000	12:16
Control Test		0.076	12:17
Air Blank		0.000	12:17
Diagnostics Check		OK	12:17

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GARBER,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: CHIRINOS LA CRUZ

DATE/TIME OF TEST: 05/21/2016 09:41:08

Diagnostics Check		OK	09:33
Air Blank		0.000	09:33
Control Test		0.078	09:34
Air Blank		0.000	09:34
Subject Sample	# 1	0.153	09:36
Breath Volume		3.378	09:36
Air Blank		0.000	09:37
Air Blank		0.000	09:39
Subject Sample	# 2	0.146	09:39
Breath Volume		2.347	09:39
Air Blank		0.000	09:40
Control Test		0.076	09:40
Air Blank		0.000	09:41
Diagnostics Check		OK	09:41

Data Download Date/Time: 06/08/2016 09:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,LISA,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: TAUTKUVIENE

DATE/TIME OF TEST: 05/21/2016 16:44:50

Diagnostics Check		OK	16:36
Air Blank		0.000	16:37
Control Test		0.078	16:37
Air Blank		0.000	16:37
Subject Sample	# 1	0.185	16:39
Breath Volume		1.304	16:39
Air Blank		0.000	16:40
Air Blank		0.000	16:42
Subject Sample	# 2	0.205	16:43
Breath Volume		1.199	16:43
Air Blank		0.000	16:43
Control Test		0.077	16:44
Air Blank		0.000	16:44
Diagnostics Check		OK	16:44

Data Download Date/Time: 06/08/2016 09:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,RAYMOND,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WITTROCK

DATE/TIME OF TEST: 05/22/2016 01:52:08

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.077	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.221	01:47
Breath Volume		2.238	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.221	01:50
Breath Volume		2.855	01:50
Air Blank		0.000	01:51
Control Test		0.078	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HARDING,JAMES
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: PRAYMAN

DATE/TIME OF TEST: 05/23/2016 02:48:40

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.345	02:43
Breath Volume		1.722	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.338	02:46
Breath Volume		1.386	02:46
Air Blank		0.000	02:47
Control Test		0.075	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: REYES,GUSTAVO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LOPEZ-GABRIEL

DATE/TIME OF TEST: 05/27/2016 01:30:40

Diagnostics Check	OK	01:29	
Air Blank		01:30	Sequence Aborted
Air Blank	0.000	01:30	

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: REYES,GUSTAVO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: LOPEZ-GABRIEL

DATE/TIME OF TEST: 05/27/2016 01:39:08

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.078	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.063	01:34
Breath Volume		1.203	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:36
Subject Sample	# 2	0.060	01:37
Breath Volume		2.167	01:37
Air Blank		0.000	01:38
Control Test		0.077	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MAURICE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR,MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GUTTORMSON

DATE/TIME OF TEST: 05/28/2016 02:58:07

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.078	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.070	02:53
Breath Volume		1.964	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.070	02:56
Breath Volume		1.617	02:56
Air Blank		0.000	02:57
Control Test		0.077	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WIGHT

DATE/TIME OF TEST: 05/28/2016 03:26:46

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.192	03:22
Breath Volume		1.523	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.189	03:25
Breath Volume		1.750	03:25
Air Blank		0.000	03:25
Control Test		0.078	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: KOSCO

DATE/TIME OF TEST: 05/28/2016 18:37:47

Diagnostics Check		OK	18:31
Air Blank		0.000	18:31
Control Test		0.077	18:32
Air Blank		0.000	18:32
Subject Sample	# 1	0.200	18:33
Breath Volume		3.878	18:33
Air Blank		0.000	18:33
Air Blank		0.000	18:35
Subject Sample	# 2	0.203	18:36
Breath Volume		3.843	18:36
Air Blank		0.000	18:36
Control Test		0.077	18:37
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD,LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD,LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: PERIEIRA DE OLIVEIRA

DATE/TIME OF TEST: 05/29/2016 00:15:26

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.078	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.150	00:10
Breath Volume		1.949	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.152	00:13
Breath Volume		2.644	00:13
Air Blank		0.000	00:14
Control Test		0.077	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,FREDERICK,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MOREJON,O
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 05/29/2016 02:13:30

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.077	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.132	02:08
Breath Volume		1.218	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.132	02:11
Breath Volume		1.816	02:11
Air Blank		0.000	02:12
Control Test		0.077	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER,MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/29/2016 02:40:23

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.156	02:35
Breath Volume		2.109	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.153	02:38
Breath Volume		2.484	02:38
Air Blank		0.000	02:39
Control Test		0.076	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,JEAN,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NOEL,JEAN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: VANHOOK

DATE/TIME OF TEST: 05/29/2016 02:51:22

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.077	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.107	02:46
Breath Volume		3.804	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.103	02:49
Breath Volume		2.953	02:49
Air Blank		0.000	02:50
Control Test		0.076	02:50
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,MAURICE,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR,MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 05/29/2016 23:42:01

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.077	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.207	23:37
Breath Volume		2.675	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:39
Subject Sample	# 2	0.201	23:40
Breath Volume		2.125	23:40
Air Blank		0.000	23:41
Control Test		0.077	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:42

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VENTURI,KRISTYN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: POIRIER,DARECK
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HASENFUSS

DATE/TIME OF TEST: 05/30/2016 05:20:23

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.078	05:14
Air Blank		0.000	05:15
Subject Sample	# 1	0.155	05:15
Breath Volume		1.226	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.157	05:18
Breath Volume		1.386	05:18
Air Blank		0.000	05:19
Control Test		0.077	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 06/08/2016 09:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,JOHN,H
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: COLEMAN,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SANFILIPPO

DATE/TIME OF TEST: 05/02/2016 23:48:37

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.078	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:47	
Control Test		0.076	23:48	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 05/03/2016 08:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: BROWNLOW

DATE/TIME OF TEST: 05/05/2016 21:48:32

Diagnostics Check		OK	21:45	
Air Blank		0.000	21:46	
Control Test		0.078	21:46	
Air Blank		0.000	21:46	
Subject Sample	# 1		21:47	Subject Test Refused
Breath Volume		0.000	21:47	
Air Blank		0.000	21:47	
Control Test		0.077	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 05/06/2016 01:08:38

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.078	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.080	01:04
Breath Volume		3.660	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.079	01:07
Breath Volume		2.367	01:07
Air Blank		0.000	01:07
Control Test		0.077	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 05/06/2016 03:05:14

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.078	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.155	03:00
Breath Volume		3.894	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.148	03:03
Breath Volume		3.632	03:03
Air Blank		0.000	03:04
Control Test		0.077	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KOLBICKA,LAWRENCE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: CARTLIDGE

DATE/TIME OF TEST: 05/11/2016 15:44:15

Diagnostics Check		OK	15:38
Air Blank		0.000	15:38
Control Test		0.077	15:39
Air Blank		0.000	15:39
Subject Sample	# 1	0.000	15:40
Breath Volume		2.074	15:40
Air Blank		0.000	15:40
Air Blank		0.000	15:42
Subject Sample	# 2	0.000	15:42
Breath Volume		2.195	15:42
Air Blank		0.000	15:43
Control Test		0.078	15:43
Air Blank		0.000	15:44
Diagnostics Check		OK	15:44

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: LUMSDEN

DATE/TIME OF TEST: 05/12/2016 00:59:25

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.078	00:54
Air Blank		0.000	00:54
Subject Sample	# 1	0.088	00:54
Breath Volume		1.156	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.093	00:57
Breath Volume		2.421	00:57
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/12/2016 22:55:51

Diagnostics Check		OK	22:49
Air Blank		0.000	22:50
Control Test		0.077	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.171	22:51
Breath Volume		2.414	22:51
Air Blank		0.000	22:52
Air Blank		0.000	22:53
Subject Sample	# 2	0.170	22:54
Breath Volume		2.074	22:54
Air Blank		0.000	22:54
Control Test		0.077	22:55
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HUDDLESTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:22

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 05/18/2016 12:48:37

Diagnostics Check		OK	12:42
Air Blank		0.000	12:42
Control Test		0.077	12:43
Air Blank		0.000	12:43
Subject Sample	# 1	0.000	12:44
Breath Volume		1.558	12:44
Air Blank		0.000	12:44
Air Blank		0.000	12:46
Subject Sample	# 2	0.000	12:47
Breath Volume		1.343	12:47
Air Blank		0.000	12:47
Control Test		0.077	12:48
Air Blank		0.000	12:48
Diagnostics Check		OK	12:48

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLEY,CPL,S
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BUTTERWORTH,OFC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:09

SUBJECT'S LAST NAME: VADALA

DATE/TIME OF TEST: 05/19/2016 11:47:50

Diagnostics Check		OK	11:40
Air Blank		0.000	11:41
Control Test		0.078	11:41
Air Blank		0.000	11:42
Subject Sample	# 1	0.000	11:43
Breath Volume		2.644	11:43
Air Blank		0.000	11:43
Air Blank		0.000	11:45
Subject Sample	# 2	0.000	11:46
Breath Volume		1.167	11:46
Air Blank		0.000	11:46
Control Test		0.077	11:47
Air Blank		0.000	11:47
Diagnostics Check		OK	11:47

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 05/20/2016 04:24:12

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.077	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.183	04:18
Breath Volume		3.125	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.167	04:22
Breath Volume		2.121	04:22
Air Blank		0.000	04:23
Control Test		0.076	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SELLERS

DATE/TIME OF TEST: 05/22/2016 02:50:10

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:47	
Control Test		0.077	02:48	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:49	
Control Test		0.076	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASSETT,STEVEN,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: HASSETT,STEVEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: KITTINGER

DATE/TIME OF TEST: 05/22/2016 19:40:02

Diagnostics Check		OK	19:32
Air Blank		0.000	19:33
Control Test		0.077	19:33
Air Blank		0.000	19:34
Subject Sample	# 1	0.196	19:35
Breath Volume		1.292	19:35
Air Blank		0.000	19:36
Air Blank		0.000	19:37
Subject Sample	# 2	0.200	19:38
Breath Volume		1.472	19:38
Air Blank		0.000	19:39
Control Test		0.076	19:39
Air Blank		0.000	19:39
Diagnostics Check		OK	19:40

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: PATERNOSTER

DATE/TIME OF TEST: 05/23/2016 21:11:03

Diagnostics Check		OK	21:04
Air Blank		0.000	21:05
Control Test		0.078	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.000	21:06
Breath Volume		1.320	21:06
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.000	21:09
Breath Volume		1.394	21:09
Air Blank		0.000	21:10
Control Test		0.077	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:11

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: GERDES

DATE/TIME OF TEST: 05/25/2016 03:50:30

Diagnostics Check		OK	03:47	
Air Blank		0.000	03:48	
Control Test		0.078	03:48	
Air Blank		0.000	03:48	
Subject Sample	# 1		03:49	Subject Test Refused
Breath Volume		0.000	03:49	
Air Blank		0.000	03:49	
Control Test		0.076	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 05/28/2016 19:39:44

Diagnostics Check		OK	19:33
Air Blank		0.000	19:33
Control Test		0.077	19:34
Air Blank		0.000	19:34
Subject Sample	# 1	0.155	19:35
Breath Volume		2.429	19:35
Air Blank		0.000	19:36
Air Blank		0.000	19:37
Subject Sample	# 2	0.154	19:38
Breath Volume		2.386	19:38
Air Blank		0.000	19:38
Control Test		0.075	19:39
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:12

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 05/29/2016 15:38:03

Diagnostics Check		OK	15:32
Air Blank		0.000	15:32
Control Test		0.078	15:32
Air Blank		0.000	15:33
Subject Sample	# 1	0.000	15:33
Breath Volume		1.640	15:33
Air Blank		0.000	15:34
Air Blank		0.000	15:36
Subject Sample	# 2	0.000	15:36
Breath Volume		1.363	15:36
Air Blank		0.000	15:37
Control Test		0.077	15:37
Air Blank		0.000	15:37
Diagnostics Check		OK	15:38

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 05/29/2016 20:06:12

Diagnostics Check		OK	20:03	
Air Blank		0.000	20:03	
Control Test		0.077	20:04	
Air Blank		0.000	20:04	
Subject Sample	# 1		20:04	Subject Test Refused
Breath Volume		0.000	20:04	
Air Blank		0.000	20:05	
Control Test		0.076	20:05	
Air Blank		0.000	20:06	
Diagnostics Check		OK	20:06	

Data Download Date/Time: 06/14/2016 17:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,JOHN,H
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: WILLIAMS,CAMERON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/30/2016 00:20:15

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.077	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1	0.190	00:09	
Breath Volume		1.175	00:09	
Air Blank		0.000	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.181	00:15	Volume Not Met
Breath Volume		1.062	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 3	0.178	00:18	
Breath Volume		1.136	00:18	
Air Blank		0.000	00:19	
Control Test		0.071	00:19	Control Outside Tolerance
Air Blank		0.000	00:20	

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,JOHN,H
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: WILLIAMS,CAMERON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/30/2016 00:24:58

Diagnostics Check	OK	00:23	
Air Blank	0.000	00:24	
Control Test	0.074	00:24	Control Outside Tolerance
Air Blank	0.000	00:24	

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,JOHN,H
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: WILLIAMS,CAMERON
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 05/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 05/30/2016 00:41:17

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:27	
Control Test		0.076	00:28	
Air Blank		0.000	00:28	
Subject Sample	# 1	0.166	00:31	Slope Not Level
Breath Volume		1.421	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2	0.175	00:35	
Breath Volume		1.335	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 3	0.173	00:39	
Breath Volume		1.187	00:39	
Air Blank		0.000	00:40	
Control Test		0.073	00:40	Control Outside Tolerance
Air Blank		0.000	00:41	

Data Download Date/Time: 06/14/2016 17:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCOTT,BRADLEY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 05/09/2016 11:05:11

Diagnostics Check		OK	10:58
Air Blank		0.000	10:59
Control Test		0.080	10:59
Air Blank		0.000	10:59
Subject Sample	# 1	0.022	11:00
Breath Volume		1.183	11:00
Air Blank		0.000	11:00
Air Blank		0.000	11:02
Subject Sample	# 2	0.019	11:03
Breath Volume		1.425	11:03
Air Blank		0.000	11:04
Control Test		0.080	11:04
Air Blank		0.000	11:05
Diagnostics Check		OK	11:05

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RENTAS,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BIENER

DATE/TIME OF TEST: 05/09/2016 22:06:19

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.079	21:59
Air Blank		0.000	22:00
Subject Sample	# 1	0.169	22:01
Breath Volume		1.593	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:03
Subject Sample	# 2	0.167	22:04
Breath Volume		2.609	22:04
Air Blank		0.000	22:05
Control Test		0.078	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 05/10/2016 03:37:15

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.000	03:33
Breath Volume		2.335	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.000	03:35
Breath Volume		2.156	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARDGROVE,ROB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: MAYERS

DATE/TIME OF TEST: 05/14/2016 03:56:45

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.000	03:52
Breath Volume		1.707	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.000	03:55
Breath Volume		1.796	03:55
Air Blank		0.000	03:55
Control Test		0.078	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DRAGON,JAMES
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 05/14/2016 18:53:48

Diagnostics Check		OK	18:47
Air Blank		0.000	18:47
Control Test		0.079	18:48
Air Blank		0.000	18:48
Subject Sample	# 1	0.000	18:49
Breath Volume		2.941	18:49
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.000	18:52
Breath Volume		2.214	18:52
Air Blank		0.000	18:52
Control Test		0.079	18:53
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 05/27/2016 11:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,JOHN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/15/2016 21:25:13

Diagnostics Check		OK	21:17
Air Blank		0.000	21:18
Control Test		0.079	21:18
Air Blank		0.000	21:18
Subject Sample	# 1	0.267	21:19
Breath Volume		1.976	21:19
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.251	21:23
Breath Volume		1.101	21:23
Air Blank		0.000	21:24
Control Test		0.079	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SUTTER

DATE/TIME OF TEST: 05/16/2016 03:58:01

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.118	03:53
Breath Volume		1.609	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.120	03:56
Breath Volume		1.750	03:56
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:58

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: QUEGAN,THOMAS
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: PELLEGRINO

DATE/TIME OF TEST: 05/19/2016 21:29:34

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.267	21:24
Breath Volume		2.296	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.268	21:28
Breath Volume		1.277	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: NELSON,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: HOLMSTROM

DATE/TIME OF TEST: 05/21/2016 21:29:55

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.192	21:24
Breath Volume		1.675	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:27
Subject Sample	# 2	0.190	21:28
Breath Volume		1.667	21:28
Air Blank		0.000	21:29
Control Test		0.080	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PRICE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/22/2016 01:07:04

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.216	01:02
Breath Volume		2.128	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.223	01:05
Breath Volume		1.921	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:07

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSEPH,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMPBELL,JOSEPH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/22/2016 08:20:04

Diagnostics Check		OK	08:13
Air Blank		0.000	08:13
Control Test		0.079	08:14
Air Blank		0.000	08:14
Subject Sample	# 1	0.108	08:15
Breath Volume		1.765	08:15
Air Blank		0.000	08:15
Air Blank		0.000	08:17
Subject Sample	# 2	0.106	08:18
Breath Volume		1.652	08:18
Air Blank		0.000	08:19
Control Test		0.078	08:19
Air Blank		0.000	08:19
Diagnostics Check		OK	08:20

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSEPH,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMPBELL,JOSEPH
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:30

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/22/2016 11:04:59

Diagnostics Check		OK	10:57
Air Blank		0.000	10:58
Control Test		0.079	10:58
Air Blank		0.000	10:58
Subject Sample	# 1	0.062	10:59
Breath Volume		1.527	10:59
Air Blank		0.000	11:00
Air Blank		0.000	11:02
Subject Sample	# 2	0.065	11:03
Breath Volume		1.730	11:03
Air Blank		0.000	11:04
Control Test		0.079	11:04
Air Blank		0.000	11:04
Diagnostics Check		OK	11:04

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARIAS,MONICA,S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FARIAS,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:02

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 05/22/2016 13:28:40

Diagnostics Check		OK	13:22
Air Blank		0.000	13:23
Control Test		0.079	13:23
Air Blank		0.000	13:23
Subject Sample	# 1	0.026	13:24
Breath Volume		1.500	13:24
Air Blank		0.000	13:24
Air Blank		0.000	13:26
Subject Sample	# 2	0.028	13:27
Breath Volume		1.628	13:27
Air Blank		0.000	13:27
Control Test		0.079	13:28
Air Blank		0.000	13:28
Diagnostics Check		OK	13:28

Data Download Date/Time: 05/27/2016 11:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALEY,DAWN,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: IMBRUNO,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:58

SUBJECT'S LAST NAME: STERSHIC

DATE/TIME OF TEST: 05/23/2016 15:26:09

Diagnostics Check		OK	15:20
Air Blank		0.000	15:20
Control Test		0.078	15:20
Air Blank		0.000	15:21
Subject Sample	# 1	0.000	15:21
Breath Volume		1.742	15:21
Air Blank		0.000	15:22
Air Blank		0.000	15:24
Subject Sample	# 2	0.000	15:24
Breath Volume		1.207	15:24
Air Blank		0.000	15:25
Control Test		0.078	15:25
Air Blank		0.000	15:26
Diagnostics Check		OK	15:26

Data Download Date/Time: 05/27/2016 11:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SILL,MATT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: BARTHA

DATE/TIME OF TEST: 05/23/2016 22:58:54

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.079	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.000	22:54
Breath Volume		2.843	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.000	22:57
Breath Volume		1.484	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 05/27/2016 11:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WEISSPFENNIG,WILLIAM
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: RANCK

DATE/TIME OF TEST: 05/27/2016 17:16:41

Diagnostics Check		OK	17:10
Air Blank		0.000	17:11
Control Test		0.078	17:11
Air Blank		0.000	17:11
Subject Sample	# 1	0.269	17:12
Breath Volume		2.562	17:12
Air Blank		0.000	17:12
Air Blank		0.000	17:14
Subject Sample	# 2	0.259	17:15
Breath Volume		3.996	17:15
Air Blank		0.000	17:15
Control Test		0.077	17:16
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 06/29/2016 14:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RIED,NICHOLAS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: SCHMEISKE

DATE/TIME OF TEST: 05/28/2016 00:28:29

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.215	00:23
Breath Volume		1.304	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.210	00:27
Breath Volume		1.152	00:27
Air Blank		0.000	00:27
Control Test		0.078	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 06/29/2016 14:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MINISCI

DATE/TIME OF TEST: 05/28/2016 04:09:00

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.145	04:04
Breath Volume		1.300	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.144	04:07
Breath Volume		1.386	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:09

Data Download Date/Time: 06/29/2016 14:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MORAGHAN,JOHN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PIQUETTE

DATE/TIME OF TEST: 05/30/2016 02:27:42

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.079	02:25	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:26	Subject Test Refused
Breath Volume		0.000	02:26	
Air Blank		0.000	02:26	
Control Test		0.079	02:27	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 06/29/2016 14:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VAUGHAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 05/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: MADONNA

DATE/TIME OF TEST: 05/30/2016 21:22:46

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.078	21:17
Air Blank		0.000	21:17
Subject Sample	# 1	0.098	21:18
Breath Volume		2.062	21:18
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.096	21:21
Breath Volume		1.781	21:21
Air Blank		0.000	21:21
Control Test		0.079	21:22
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 06/29/2016 14:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BOETGER,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 05/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/31/2016 03:04:04

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.147	02:59
Breath Volume		3.683	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.142	03:02
Breath Volume		4.121	03:02
Air Blank		0.000	03:03
Control Test		0.078	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:04

Data Download Date/Time: 06/29/2016 14:40

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: WHITE,C
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: TRETSKY

DATE/TIME OF TEST: 05/06/2016 18:41:49

Diagnostics Check		OK	18:30	
Air Blank		0.000	18:30	
Control Test		0.080	18:31	
Air Blank		0.000	18:31	
Subject Sample	# 1	0.000	18:34	Volume Not Met
Breath Volume		0.597	18:34	
Air Blank		0.000	18:35	
Air Blank		0.000	18:37	
Subject Sample	# 2		18:40	No Sample Provided
Breath Volume		0.000	18:40	
Air Blank		0.000	18:40	
Control Test		0.080	18:41	
Air Blank		0.000	18:41	
Diagnostics Check		OK	18:41	

Data Download Date/Time: 05/18/2016 15:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THEODORE,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLER,THEODORE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: NORTH

DATE/TIME OF TEST: 05/09/2016 05:15:24

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:05	
Control Test		0.080	05:05	
Air Blank		0.000	05:06	
Subject Sample	# 1	0.202	05:09	Volume Not Met
Breath Volume		0.429	05:09	
Air Blank		0.000	05:10	
Air Blank		0.000	05:11	
Subject Sample	# 2		05:13	Subject Test Refused
Breath Volume		0.015	05:13	
Air Blank		0.000	05:14	
Control Test		0.076	05:14	
Air Blank		0.000	05:15	
Diagnostics Check		OK	05:15	

Data Download Date/Time: 05/18/2016 15:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,THEODORE,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MILLER,THEODORE
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: NORTH

DATE/TIME OF TEST: 05/09/2016 05:23:10

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.078	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.260	05:18
Breath Volume		1.320	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:20
Subject Sample	# 2	0.252	05:21
Breath Volume		1.472	05:21
Air Blank		0.000	05:22
Control Test		0.078	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 05/18/2016 15:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINKMAN,KELLY,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRINKMAN,KELLY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/13/2016 20:40:06

Diagnostics Check		OK	20:27	
Air Blank		0.000	20:28	
Control Test		0.080	20:28	
Air Blank		0.000	20:28	
Subject Sample	# 1	0.279	20:29	
Breath Volume		1.308	20:29	
Air Blank		0.000	20:29	
Air Blank		0.000	20:31	
Subject Sample	# 2	0.226	20:34	Volume Not Met
Breath Volume		0.820	20:34	
Air Blank		0.000	20:35	
Air Blank		0.000	20:37	
Subject Sample	# 3		20:38	Subject Test Refused
Breath Volume		0.000	20:38	
Air Blank		0.000	20:39	
Control Test		0.076	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 05/18/2016 15:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BOLLMAN

DATE/TIME OF TEST: 05/14/2016 00:30:02

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:24
Subject Sample	# 1	0.089	00:25
Breath Volume		2.160	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.092	00:28
Breath Volume		2.207	00:28
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:30

Data Download Date/Time: 05/18/2016 15:22

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHEO,FRANCISCO,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MICHEO,FRANCISCO
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 05/17/2016 22:18:35

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.080	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.117	22:13
Breath Volume		2.617	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:15
Subject Sample	# 2	0.115	22:16
Breath Volume		1.953	22:16
Air Blank		0.000	22:17
Control Test		0.079	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 05/18/2016 15:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001741

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: NAVA

DATE/TIME OF TEST: 05/06/2016 03:19:54

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.080	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.193	03:15
Breath Volume		2.355	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.200	03:18
Breath Volume		2.839	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 05/20/2016 12:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/20/2016 22:35:58

Diagnostics Check	OK	22:34	
Air Blank	0.000	22:34	
Control Test	0.080	22:35	
Air Blank		22:35	RFI Detect
Air Blank	0.000	22:35	

Data Download Date/Time: 05/25/2016 09:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 05/20/2016 22:49:53

Diagnostics Check		OK	22:37
Air Blank		0.000	22:37
Control Test		0.080	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.200	22:41
Breath Volume		1.335	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:43
Subject Sample	# 2	0.225	22:44
Breath Volume		1.117	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 3	0.225	22:48
Breath Volume		2.117	22:48
Air Blank		0.000	22:48
Control Test		0.079	22:49
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 05/25/2016 09:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001742

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: PABLO-CARRILLO

DATE/TIME OF TEST: 05/22/2016 21:42:56

Diagnostics Check		OK	21:36
Air Blank		0.000	21:36
Control Test		0.081	21:37
Air Blank		0.000	21:37
Subject Sample	# 1	0.253	21:38
Breath Volume		2.894	21:38
Air Blank		0.000	21:38
Air Blank		0.000	21:40
Subject Sample	# 2	0.271	21:41
Breath Volume		2.492	21:41
Air Blank		0.000	21:41
Control Test		0.080	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 05/25/2016 09:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: SCHILL,PATRICK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: ABRAMS JR

DATE/TIME OF TEST: 05/02/2016 02:21:59

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.078	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 05/14/2016 11:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: PAGE,BRIAN
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: AURICHIO

DATE/TIME OF TEST: 05/07/2016 15:43:56

Diagnostics Check		OK	15:37
Air Blank		0.000	15:37
Control Test		0.078	15:38
Air Blank		0.000	15:38
Subject Sample	# 1	0.040	15:39
Breath Volume		3.171	15:39
Air Blank		0.000	15:39
Air Blank		0.000	15:41
Subject Sample	# 2	0.041	15:42
Breath Volume		2.816	15:42
Air Blank		0.000	15:43
Control Test		0.079	15:43
Air Blank		0.000	15:43
Diagnostics Check		OK	15:43

Data Download Date/Time: 05/14/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: SCHILL,PATRICK
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: EDMONSON

DATE/TIME OF TEST: 05/14/2016 20:17:55

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.079	20:11
Air Blank		0.000	20:11
Subject Sample	# 1	0.199	20:13
Breath Volume		2.316	20:13
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.195	20:16
Breath Volume		2.039	20:16
Air Blank		0.000	20:16
Control Test		0.077	20:17
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 07/06/2016 10:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,BRIAN,R
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: HANSEN,AARON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 05/19/2016 16:09:43

Diagnostics Check		OK	16:03
Air Blank		0.000	16:04
Control Test		0.078	16:04
Air Blank		0.000	16:04
Subject Sample	# 1	0.000	16:05
Breath Volume		2.847	16:05
Air Blank		0.000	16:05
Air Blank		0.000	16:07
Subject Sample	# 2	0.000	16:08
Breath Volume		1.886	16:08
Air Blank		0.000	16:08
Control Test		0.077	16:09
Air Blank		0.000	16:09
Diagnostics Check		OK	16:09

Data Download Date/Time: 07/06/2016 10:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYMAN,THOMAS,C
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COPPOLA,TYSON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 05/21/2016 03:25:38

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.185	03:21
Breath Volume		2.371	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.172	03:23
Breath Volume		2.316	03:23
Air Blank		0.000	03:24
Control Test		0.075	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 07/06/2016 10:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYMAN,THOMAS,C
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COPPOLA,TYSON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 05/21/2016 23:33:26

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.252	23:28
Breath Volume		4.285	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.237	23:31
Breath Volume		4.511	23:31
Air Blank		0.000	23:32
Control Test		0.077	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 07/06/2016 10:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WYMAN,THOMAS,C
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: COPPOLA,TYSON
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: HERALD

DATE/TIME OF TEST: 05/22/2016 19:22:46

Diagnostics Check		OK	19:14
Air Blank		0.000	19:15
Control Test		0.078	19:15
Air Blank		0.000	19:16
Subject Sample	# 1	0.205	19:17
Breath Volume		1.101	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.194	19:20
Breath Volume		2.148	19:20
Air Blank		0.000	19:21
Control Test		0.075	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 07/06/2016 10:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: KIRKLAND,RICHARD
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: PISULA

DATE/TIME OF TEST: 05/23/2016 22:33:47

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.078	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.318	22:29
Breath Volume		2.511	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.322	22:32
Breath Volume		2.035	22:32
Air Blank		0.000	22:32
Control Test		0.078	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 07/06/2016 10:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002231

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKLAND,RICHARD,E
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD
ARRESTING OFFICER: JENKINS,ANDRE
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 05/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12415080A5
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 05/25/2016 02:10:35

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.078	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Control Test		0.078	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 07/06/2016 10:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,MARCUS,R
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: HERNANDEZ,DERMIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: BUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 05/28/2016 13:26:52

Diagnostics Check		OK	13:23	
Air Blank		0.000	13:24	
Control Test		0.082	13:24	
Air Blank		0.000	13:24	
Subject Sample	# 1		13:25	Subject Test Refused
Breath Volume		0.000	13:25	
Air Blank		0.000	13:25	
Control Test		0.081	13:26	
Air Blank		0.000	13:26	
Diagnostics Check		OK	13:26	

Data Download Date/Time: 06/14/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBLE,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: CARPER,SEAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: BUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 05/29/2016 22:06:32

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:03	
Control Test		0.081	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:06	Sequence Aborted
Breath Volume		0.000	22:06	
Air Blank		0.000	22:06	

Data Download Date/Time: 06/14/2016 11:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBLE,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: CARPER,SEAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: BUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: KRAMER

DATE/TIME OF TEST: 05/29/2016 22:13:42

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:09	
Control Test		0.081	22:09	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:12	Subject Test Refused
Breath Volume		0.000	22:12	
Air Blank		0.000	22:12	
Control Test		0.081	22:13	
Air Blank		0.000	22:13	
Diagnostics Check		OK	22:13	

Data Download Date/Time: 06/14/2016 11:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,E,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MESA,A
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VENEREO

DATE/TIME OF TEST: 05/19/2016 01:43:32

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.078	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1		01:42	Subject Test Refused
Breath Volume		0.000	01:42	
Air Blank		0.000	01:42	
Control Test		0.077	01:42	
Air Blank		0.000	01:43	
Diagnostics Check		OK	01:43	

Data Download Date/Time: 06/22/2016 13:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTERO,JOSE,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: HOLLAND,E
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 05/28/2016 23:48:38

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.077	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:47	
Control Test		0.077	23:48	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 06/22/2016 13:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTERO,JOSE,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: FABIEN,J
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: BLANCO LORENZO

DATE/TIME OF TEST: 05/29/2016 16:41:46

Diagnostics Check		OK	16:34
Air Blank		0.000	16:35
Control Test		0.077	16:35
Air Blank		0.000	16:36
Subject Sample	# 1	0.288	16:37
Breath Volume		2.296	16:37
Air Blank		0.000	16:37
Air Blank		0.000	16:39
Subject Sample	# 2	0.279	16:40
Breath Volume		2.578	16:40
Air Blank		0.000	16:40
Control Test		0.076	16:41
Air Blank		0.000	16:41
Diagnostics Check		OK	16:41

Data Download Date/Time: 06/22/2016 13:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARREA,S,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: GOMEZ,ANDREW
ARRESTING OFFICER AGENCY: Opa Locka PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/26/2016 10:30:13

Diagnostics Check		OK	10:23
Air Blank		0.000	10:24
Control Test		0.077	10:24
Air Blank		0.000	10:25
Subject Sample	# 1	0.198	10:25
Breath Volume		1.675	10:25
Air Blank		0.000	10:26
Air Blank		0.000	10:28
Subject Sample	# 2	0.193	10:28
Breath Volume		2.160	10:28
Air Blank		0.000	10:29
Control Test		0.076	10:29
Air Blank		0.000	10:30
Diagnostics Check		OK	10:30

Data Download Date/Time: 06/22/2016 12:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARREA,S,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: DUQUE,A
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: PINO

DATE/TIME OF TEST: 05/27/2016 14:18:25

Diagnostics Check		OK	14:15	
Air Blank		0.000	14:15	
Control Test		0.078	14:16	
Air Blank		0.000	14:16	
Subject Sample	# 1		14:17	Subject Test Refused
Breath Volume		0.000	14:17	
Air Blank		0.000	14:17	
Control Test		0.079	14:17	
Air Blank		0.000	14:18	
Diagnostics Check		OK	14:18	

Data Download Date/Time: 06/22/2016 12:55

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,E,N
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: LINARES,C
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 04/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 05/07/2016 06:01:00

Diagnostics Check		OK	05:53
Air Blank		0.000	05:54
Control Test		0.078	05:54
Air Blank		0.000	05:55
Subject Sample	# 1	0.012	05:56
Breath Volume		1.132	05:56
Air Blank		0.000	05:57
Air Blank		0.000	05:58
Subject Sample	# 2	0.010	05:59
Breath Volume		1.148	05:59
Air Blank		0.000	06:00
Control Test		0.077	06:00
Air Blank		0.000	06:00
Diagnostics Check		OK	06:01

Data Download Date/Time: 05/18/2016 12:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATERS,COREY,J
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: BOHNE,NICK
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG502603
DRY GAS EXPIRY DATE: 01/26/2017
VIOLATION CODE: BUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 05/07/2016 18:28:36

Diagnostics Check		OK	18:22
Air Blank		0.000	18:23
Control Test		0.077	18:23
Air Blank		0.000	18:23
Subject Sample	# 1	0.126	18:24
Breath Volume		3.332	18:24
Air Blank		0.000	18:24
Air Blank		0.000	18:26
Subject Sample	# 2	0.127	18:27
Breath Volume		3.726	18:27
Air Blank		0.000	18:27
Control Test		0.077	18:27
Air Blank		0.000	18:28
Diagnostics Check		OK	18:28

Data Download Date/Time: 06/27/2016 18:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRYAN,P
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: SAPP,MIKE
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG502603
DRY GAS EXPIRY DATE: 01/26/2017
VIOLATION CODE: BUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 05/07/2016 19:07:31

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.078	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.194	19:02
Breath Volume		2.195	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.198	19:05
Breath Volume		2.218	19:05
Air Blank		0.000	19:06
Control Test		0.077	19:06
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 06/27/2016 18:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003196

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,BRYAN,P
BREATH TEST OPERATOR AGENCY: FFWCC (Ocala)
ARRESTING OFFICER: MCGUIRE,JON
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG502603
DRY GAS EXPIRY DATE: 01/26/2017
VIOLATION CODE: BUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:48

SUBJECT'S LAST NAME: THEMANN

DATE/TIME OF TEST: 05/07/2016 20:10:28

Diagnostics Check		OK	20:04
Air Blank		0.000	20:04
Control Test		0.077	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.123	20:06
Breath Volume		3.546	20:06
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.129	20:08
Breath Volume		3.601	20:08
Air Blank		0.000	20:09
Control Test		0.077	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 06/27/2016 18:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAKER,RUEL,W
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HUGHES,BRYANT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 05/16/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: THURMAN

DATE/TIME OF TEST: 05/29/2016 20:15:25

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:13	
Control Test		0.080	20:13	
Air Blank		0.000	20:13	
Subject Sample	# 1		20:14	Subject Test Refused
Breath Volume		0.000	20:14	
Air Blank		0.000	20:14	
Control Test		0.079	20:14	
Air Blank		0.000	20:15	
Diagnostics Check		OK	20:15	

Data Download Date/Time: 06/21/2016 16:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: SWEETWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTANO,DANIEL,A
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: MARTINEZ,ANA
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 05/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 05/25/2016 17:52:57

Diagnostics Check		OK	17:46
Air Blank		0.000	17:47
Control Test		0.081	17:47
Air Blank		0.000	17:48
Subject Sample	# 1	0.250	17:48
Breath Volume		2.976	17:48
Air Blank		0.000	17:49
Air Blank		0.000	17:51
Subject Sample	# 2	0.246	17:51
Breath Volume		3.078	17:51
Air Blank		0.000	17:52
Control Test		0.080	17:52
Air Blank		0.000	17:52
Diagnostics Check		OK	17:52

Data Download Date/Time: 07/12/2016 11:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: CULBERT

DATE/TIME OF TEST: 05/01/2016 03:53:02

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.079	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.167	03:48
Breath Volume		1.238	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.170	03:51
Breath Volume		1.941	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:53

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 05/01/2016 04:10:31

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.128	04:06
Breath Volume		1.917	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.124	04:08
Breath Volume		2.144	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: WINGFIELD

DATE/TIME OF TEST: 05/01/2016 06:12:08

Diagnostics Check		OK	06:05
Air Blank		0.000	06:05
Control Test		0.079	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.159	06:07
Breath Volume		1.812	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:09
Subject Sample	# 2	0.162	06:10
Breath Volume		2.523	06:10
Air Blank		0.000	06:11
Control Test		0.079	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VILLA MUNOZ

DATE/TIME OF TEST: 05/02/2016 04:03:57

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.174	03:59
Breath Volume		2.453	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.160	04:02
Breath Volume		1.121	04:02
Air Blank		0.000	04:03
Control Test		0.077	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LUCHT

DATE/TIME OF TEST: 05/03/2016 04:30:23

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.128	04:25
Breath Volume		2.750	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.129	04:28
Breath Volume		2.726	04:28
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 05/04/2016 02:09:14

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.078	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.173	02:01	Volume Not Met
Breath Volume		0.015	02:01	
Air Blank		0.000	02:02	
Air Blank		0.000	02:03	
Subject Sample	# 2	0.204	02:04	
Breath Volume		2.058	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 3	0.203	02:07	
Breath Volume		1.894	02:07	
Air Blank		0.000	02:08	
Control Test		0.077	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 05/05/2016 23:35:45

Diagnostics Check		OK	23:28
Air Blank		0.000	23:28
Control Test		0.079	23:29
Air Blank		0.000	23:29
Subject Sample	# 1	0.082	23:30
Breath Volume		1.128	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.085	23:34
Breath Volume		1.683	23:34
Air Blank		0.000	23:34
Control Test		0.078	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 05/06/2016 02:00:38

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:49	
Control Test		0.078	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:53	No Sample Provided
Breath Volume		0.000	01:53	
Air Blank		0.000	01:53	
Air Blank		0.000	01:55	
Subject Sample	# 2	0.097	01:56	
Breath Volume		2.101	01:56	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 3	0.094	01:59	
Breath Volume		2.773	01:59	
Air Blank		0.000	01:59	
Control Test		0.079	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: COZZAGLIO

DATE/TIME OF TEST: 05/06/2016 20:56:06

Diagnostics Check		OK	20:49
Air Blank		0.000	20:50
Control Test		0.081	20:50
Air Blank		0.000	20:51
Subject Sample	# 1	0.174	20:51
Breath Volume		1.390	20:51
Air Blank		0.000	20:52
Air Blank		0.000	20:54
Subject Sample	# 2	0.182	20:54
Breath Volume		1.437	20:54
Air Blank		0.000	20:55
Control Test		0.079	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FUSKO

DATE/TIME OF TEST: 05/07/2016 03:44:17

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.080	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1	0.098	03:34	Volume Not Met
Breath Volume		0.238	03:34	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2	0.141	03:39	
Breath Volume		1.609	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 3		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Control Test		0.078	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: PEREZ DE LA CRUZ

DATE/TIME OF TEST: 05/08/2016 00:42:35

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.181	00:38
Breath Volume		2.375	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.177	00:40
Breath Volume		1.949	00:40
Air Blank		0.000	00:41
Control Test		0.078	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MARASCO

DATE/TIME OF TEST: 05/08/2016 03:30:28

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.061	03:25
Breath Volume		2.941	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.060	03:28
Breath Volume		2.261	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRUZ,E
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: PAROLINI

DATE/TIME OF TEST: 05/08/2016 19:42:34

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.079	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.282	19:37
Breath Volume		3.007	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.289	19:40
Breath Volume		2.910	19:40
Air Blank		0.000	19:41
Control Test		0.078	19:41
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DOW

DATE/TIME OF TEST: 05/09/2016 01:47:39

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.186	01:42
Breath Volume		3.144	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.188	01:46
Breath Volume		3.039	01:46
Air Blank		0.000	01:46
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GREENE JR

DATE/TIME OF TEST: 05/09/2016 03:48:53

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.078	03:44
Breath Volume		2.250	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.078	03:47
Breath Volume		2.394	03:47
Air Blank		0.000	03:47
Control Test		0.078	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LEBOVITZ,A
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GORMAN

DATE/TIME OF TEST: 05/10/2016 00:52:13

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.000	00:47
Breath Volume		3.152	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.000	00:50
Breath Volume		2.984	00:50
Air Blank		0.000	00:51
Control Test		0.078	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: ANTONIO-FRANCO

DATE/TIME OF TEST: 05/11/2016 01:15:44

Diagnostics Check		OK	01:04	
Air Blank		0.000	01:04	
Control Test		0.079	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1	0.243	01:08	Volume Not Met
Breath Volume		0.031	01:08	
Air Blank		0.000	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 2	0.206	01:14	Volume Not Met
Breath Volume		0.066	01:14	
Air Blank		0.000	01:14	
Control Test		0.077	01:15	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: ANTONIO-FRANCO

DATE/TIME OF TEST: 05/11/2016 01:31:40

Diagnostics Check		OK	01:20
Air Blank		0.000	01:21
Control Test		0.076	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.255	01:24
Breath Volume		1.359	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.245	01:29
Breath Volume		1.675	01:29
Air Blank		0.000	01:30
Control Test		0.077	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 05/11/2016 11:42

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 05/11/2016 06:22:34

Diagnostics Check		OK	06:15
Air Blank		0.000	06:16
Control Test		0.078	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.057	06:17
Breath Volume		1.964	06:17
Air Blank		0.000	06:18
Air Blank		0.000	06:20
Subject Sample	# 2	0.051	06:21
Breath Volume		2.097	06:21
Air Blank		0.000	06:21
Control Test		0.079	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 05/11/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 05/18/2016 23:23:49

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.058	23:19
Breath Volume		1.101	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.058	23:22
Breath Volume		1.230	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 06/07/2016 13:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: ALANIS LUCIO

DATE/TIME OF TEST: 05/19/2016 04:32:11

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.046	04:26
Breath Volume		1.519	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.048	04:30
Breath Volume		2.171	04:30
Air Blank		0.000	04:31
Control Test		0.078	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 06/07/2016 13:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: AZADI

DATE/TIME OF TEST: 05/20/2016 00:47:23

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.078	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.213	00:42
Breath Volume		3.128	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.218	00:45
Breath Volume		3.046	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 06/07/2016 13:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/20/2016 02:56:03

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.154	02:51
Breath Volume		3.535	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:53
Subject Sample	# 2	0.150	02:54
Breath Volume		2.953	02:54
Air Blank		0.000	02:55
Control Test		0.078	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/07/2016 13:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 05/21/2016 01:50:33

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.000	01:46
Breath Volume		1.335	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.000	01:49
Breath Volume		1.851	01:49
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 06/07/2016 13:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 05/21/2016 02:23:44

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:19	
Control Test		0.079	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 06/07/2016 13:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: HENRY JR

DATE/TIME OF TEST: 05/21/2016 04:27:50

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:25	
Control Test		0.079	04:25	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:26	Subject Test Refused
Breath Volume		0.000	04:26	
Air Blank		0.000	04:26	
Control Test		0.078	04:27	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: GANDARILLA

DATE/TIME OF TEST: 05/22/2016 01:14:17

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.154	01:09
Breath Volume		2.320	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.159	01:12
Breath Volume		2.281	01:12
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: ALSHAREEF

DATE/TIME OF TEST: 05/22/2016 03:25:57

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.207	03:21
Breath Volume		2.265	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.211	03:24
Breath Volume		1.238	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: KOVACIC

DATE/TIME OF TEST: 05/22/2016 03:43:12

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.146	03:38
Breath Volume		1.410	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.153	03:41
Breath Volume		1.593	03:41
Air Blank		0.000	03:42
Control Test		0.079	03:42
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: RENFROE

DATE/TIME OF TEST: 05/22/2016 05:48:44

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:36	
Control Test		0.079	05:37	
Air Blank		0.000	05:37	
Subject Sample	# 1	0.087	05:40	Volume Not Met
Breath Volume		0.101	05:40	
Air Blank		0.000	05:41	
Air Blank		0.000	05:43	
Subject Sample	# 2	0.087	05:44	
Breath Volume		1.910	05:44	
Air Blank		0.000	05:44	
Air Blank		0.000	05:46	
Subject Sample	# 3	0.087	05:47	
Breath Volume		1.640	05:47	
Air Blank		0.000	05:47	
Control Test		0.077	05:48	
Air Blank		0.000	05:48	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 05/23/2016 02:54:57

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.265	02:50
Breath Volume		2.593	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.269	02:53
Breath Volume		2.652	02:53
Air Blank		0.000	02:54
Control Test		0.078	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/23/2016 05:20:46

Diagnostics Check		OK	05:13
Air Blank		0.000	05:14
Control Test		0.079	05:14
Air Blank		0.000	05:15
Subject Sample	# 1	0.202	05:15
Breath Volume		1.175	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.215	05:19
Breath Volume		3.390	05:19
Air Blank		0.000	05:19
Control Test		0.078	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 05/25/2016 01:50:12

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.079	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1	0.205	01:42	Volume Not Met
Breath Volume		0.265	01:42	
Air Blank		0.000	01:43	
Air Blank		0.000	01:45	
Subject Sample	# 2	0.237	01:48	Volume Not Met
Breath Volume		0.546	01:48	
Air Blank		0.000	01:49	
Control Test		0.077	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: MARKS

DATE/TIME OF TEST: 05/25/2016 02:08:16

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.077	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.211	01:55	
Breath Volume		1.167	01:55	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.178	02:00	Volume Not Met
Breath Volume		0.128	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:03	
Subject Sample	# 3	0.213	02:06	Volume Not Met
Breath Volume		0.277	02:06	
Air Blank		0.000	02:07	
Control Test		0.075	02:07	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:08	

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: ESTELLE

DATE/TIME OF TEST: 05/26/2016 01:35:09

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.079	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.199	01:30
Breath Volume		2.093	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.195	01:33
Breath Volume		2.000	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FLORES,L
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: SHELTON

DATE/TIME OF TEST: 05/26/2016 23:01:56

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.079	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.177	22:56
Breath Volume		1.839	22:56
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.185	23:00
Breath Volume		2.738	23:00
Air Blank		0.000	23:01
Control Test		0.078	23:01
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 05/27/2016 02:59:18

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.077	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.191	02:54
Breath Volume		1.324	02:54
Air Blank		0.000	02:55
Air Blank		0.000	02:56
Subject Sample	# 2	0.206	02:57
Breath Volume		1.398	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/27/2016 03:34:03

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.088	03:29
Breath Volume		3.761	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.090	03:32
Breath Volume		3.550	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:34

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIERRA,P
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: OSORIO OTERO

DATE/TIME OF TEST: 05/28/2016 01:50:18

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.078	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.130	01:45
Breath Volume		1.671	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.124	01:48
Breath Volume		2.042	01:48
Air Blank		0.000	01:49
Control Test		0.078	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: SQUIRES

DATE/TIME OF TEST: 05/28/2016 22:56:26

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.078	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.097	22:52
Breath Volume		1.730	22:52
Air Blank		0.000	22:52
Air Blank		0.000	22:54
Subject Sample	# 2	0.098	22:54
Breath Volume		1.429	22:54
Air Blank		0.000	22:55
Control Test		0.078	22:55
Air Blank		0.000	22:56
Diagnostics Check		OK	22:56

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 05/29/2016 03:30:23

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.000	03:26
Breath Volume		1.750	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.000	03:29
Breath Volume		2.324	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MEYER,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 05/29/2016 04:43:40

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.106	04:39
Breath Volume		2.007	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.107	04:42
Breath Volume		2.171	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: BONVISO

DATE/TIME OF TEST: 05/29/2016 23:14:41

Diagnostics Check		OK	23:07
Air Blank		0.000	23:08
Control Test		0.079	23:08
Air Blank		0.000	23:09
Subject Sample	# 1	0.164	23:10
Breath Volume		1.710	23:10
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.168	23:13
Breath Volume		1.960	23:13
Air Blank		0.000	23:13
Control Test		0.078	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 06/07/2016 13:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: PAVARINI

DATE/TIME OF TEST: 05/30/2016 03:01:28

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.082	02:57
Breath Volume		1.230	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.089	02:59
Breath Volume		2.554	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 05/31/2016 01:49:56

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.133	01:45
Breath Volume		1.480	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.117	01:48
Breath Volume		2.136	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 06/07/2016 13:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: PITTS II

DATE/TIME OF TEST: 05/11/2016 23:56:21

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.081	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.115	23:51
Breath Volume		2.195	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.110	23:54
Breath Volume		1.406	23:54
Air Blank		0.000	23:55
Control Test		0.081	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HERCULES II

DATE/TIME OF TEST: 05/12/2016 04:07:17

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.081	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.090	04:02
Breath Volume		1.105	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.096	04:05
Breath Volume		1.683	04:05
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: ODPARLIK III

DATE/TIME OF TEST: 05/13/2016 03:20:26

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.120	03:15
Breath Volume		4.328	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.121	03:18
Breath Volume		3.972	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: JORGENSEN

DATE/TIME OF TEST: 05/13/2016 03:38:19

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.201	03:33
Breath Volume		2.031	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.200	03:36
Breath Volume		1.949	03:36
Air Blank		0.000	03:37
Control Test		0.079	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: STEINLE JR

DATE/TIME OF TEST: 05/13/2016 05:08:32

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.080	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.280	05:03
Breath Volume		1.289	05:03
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.296	05:06
Breath Volume		1.937	05:06
Air Blank		0.000	05:07
Control Test		0.079	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: RIVERA OLIVO

DATE/TIME OF TEST: 05/14/2016 01:20:17

Diagnostics Check		OK	01:14
Air Blank		0.000	01:14
Control Test		0.081	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.183	01:15
Breath Volume		3.988	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.183	01:18
Breath Volume		3.804	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: PAPAS

DATE/TIME OF TEST: 05/14/2016 04:24:22

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.082	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.162	04:19
Breath Volume		1.585	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:21
Subject Sample	# 2	0.157	04:22
Breath Volume		1.410	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RODRIGUEZ,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 05/14/2016 22:55:18

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:50	
Control Test		0.081	22:50	
Air Blank		0.000	22:51	
Subject Sample	# 1		22:53	Subject Test Refused
Breath Volume		0.402	22:53	
Air Blank		0.000	22:54	
Control Test		0.081	22:54	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: VARGUEZ-COB

DATE/TIME OF TEST: 05/14/2016 23:43:28

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.179	23:38
Breath Volume		2.179	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.183	23:41
Breath Volume		1.886	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ODOM

DATE/TIME OF TEST: 05/15/2016 22:53:46

Diagnostics Check		OK	22:47
Air Blank		0.000	22:47
Control Test		0.077	22:47
Air Blank		0.000	22:48
Subject Sample	# 1	0.180	22:49
Breath Volume		3.984	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.169	22:52
Breath Volume		3.144	22:52
Air Blank		0.000	22:52
Control Test		0.076	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/16/2016 01:02:56

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.045	00:58
Breath Volume		3.320	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.039	01:01
Breath Volume		1.437	01:01
Air Blank		0.000	01:02
Control Test		0.076	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: PROSSER

DATE/TIME OF TEST: 05/16/2016 04:50:14

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:44
Subject Sample	# 1	0.113	04:45
Breath Volume		1.656	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.109	04:48
Breath Volume		1.257	04:48
Air Blank		0.000	04:49
Control Test		0.078	04:49
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: VASQUEZ JR

DATE/TIME OF TEST: 05/17/2016 02:54:38

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.081	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.000	02:50
Breath Volume		3.093	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.000	02:53
Breath Volume		3.460	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: HOWARD JR

DATE/TIME OF TEST: 05/17/2016 22:29:24

Diagnostics Check		OK	22:23
Air Blank		0.000	22:23
Control Test		0.080	22:24
Air Blank		0.000	22:24
Subject Sample	# 1	0.000	22:25
Breath Volume		1.953	22:25
Air Blank		0.000	22:25
Air Blank		0.000	22:27
Subject Sample	# 2	0.000	22:28
Breath Volume		2.148	22:28
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAYLOR,SEAN
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ORSINI

DATE/TIME OF TEST: 05/18/2016 18:36:44

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.081	18:31
Air Blank		0.000	18:31
Subject Sample	# 1	0.134	18:32
Breath Volume		3.546	18:32
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.129	18:35
Breath Volume		2.488	18:35
Air Blank		0.000	18:35
Control Test		0.081	18:36
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: KEELAN

DATE/TIME OF TEST: 05/18/2016 23:43:41

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.081	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.199	23:39
Breath Volume		1.835	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.203	23:42
Breath Volume		1.875	23:42
Air Blank		0.000	23:42
Control Test		0.080	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: LEMPEREUR

DATE/TIME OF TEST: 05/19/2016 04:16:36

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.139	04:11
Breath Volume		2.187	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.149	04:15
Breath Volume		2.625	04:15
Air Blank		0.000	04:15
Control Test		0.081	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: SALAZAR JR

DATE/TIME OF TEST: 05/19/2016 21:04:55

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.081	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.009	21:00
Breath Volume		1.843	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:03
Subject Sample	# 2	0.009	21:03
Breath Volume		1.289	21:03
Air Blank		0.000	21:04
Control Test		0.081	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: LIZARDI

DATE/TIME OF TEST: 05/20/2016 02:34:30

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.081	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.198	02:29
Breath Volume		2.164	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.200	02:32
Breath Volume		2.707	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: DEWALD

DATE/TIME OF TEST: 05/20/2016 23:04:21

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.081	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.255	22:59
Breath Volume		1.523	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.242	23:02
Breath Volume		1.636	23:02
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: STANLEY JR

DATE/TIME OF TEST: 05/21/2016 01:09:32

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.080	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.000	01:05
Breath Volume		2.085	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.000	01:08
Breath Volume		2.953	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RODRIGUEZ DIAZ

DATE/TIME OF TEST: 05/21/2016 02:24:08

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.141	02:18
Breath Volume		1.531	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.129	02:22
Breath Volume		1.636	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOBSON,D
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SARA

DATE/TIME OF TEST: 05/21/2016 03:47:53

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.082	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.128	03:43
Breath Volume		1.828	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.127	03:46
Breath Volume		2.148	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHANNON,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: HARLEE

DATE/TIME OF TEST: 05/21/2016 06:11:16

Diagnostics Check		OK	06:04
Air Blank		0.000	06:05
Control Test		0.081	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.152	06:06
Breath Volume		1.175	06:06
Air Blank		0.000	06:07
Air Blank		0.000	06:09
Subject Sample	# 2	0.149	06:09
Breath Volume		1.468	06:09
Air Blank		0.000	06:10
Control Test		0.080	06:10
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEPPARD,RAYMOND
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: ESPINOZA SANTOS

DATE/TIME OF TEST: 05/21/2016 17:09:59

Diagnostics Check		OK	17:03
Air Blank		0.000	17:04
Control Test		0.082	17:04
Air Blank		0.000	17:05
Subject Sample	# 1	0.263	17:05
Breath Volume		1.570	17:05
Air Blank		0.000	17:06
Air Blank		0.000	17:07
Subject Sample	# 2	0.261	17:08
Breath Volume		2.035	17:08
Air Blank		0.000	17:09
Control Test		0.079	17:09
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/22/2016 00:54:45

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.082	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.151	00:49
Breath Volume		1.238	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.161	00:53
Breath Volume		1.996	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 05/22/2016 01:59:01

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.197	01:51
Breath Volume		4.578	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:53
Subject Sample	# 2	0.161	01:54
Breath Volume		1.796	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:56
Subject Sample	# 3	0.164	01:57
Breath Volume		1.511	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:59

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 05/22/2016 03:50:53

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.191	03:46
Breath Volume		3.378	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.186	03:49
Breath Volume		3.699	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: ONOFRE

DATE/TIME OF TEST: 05/22/2016 06:14:06

Diagnostics Check		OK	06:07
Air Blank		0.000	06:07
Control Test		0.081	06:08
Air Blank		0.000	06:08
Subject Sample	# 1	0.058	06:09
Breath Volume		2.460	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:11
Subject Sample	# 2	0.057	06:12
Breath Volume		2.281	06:12
Air Blank		0.000	06:13
Control Test		0.080	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: WEAH

DATE/TIME OF TEST: 05/22/2016 20:41:59

Diagnostics Check		OK	20:31
Air Blank		0.000	20:32
Control Test		0.081	20:32
Air Blank		0.000	20:33
Subject Sample	# 1	0.255	20:35
Breath Volume		1.250	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:37
Subject Sample	# 2	0.253	20:40
Breath Volume		1.847	20:40
Air Blank		0.000	20:41
Control Test		0.078	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: REHO

DATE/TIME OF TEST: 05/23/2016 00:29:50

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.081	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Control Test		0.082	00:29	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: REHO

DATE/TIME OF TEST: 05/23/2016 00:46:59

Diagnostics Check		OK	00:36
Air Blank		0.000	00:37
Control Test		0.081	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.263	00:38
Breath Volume		1.457	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.288	00:42
Breath Volume		2.792	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 3	0.293	00:45
Breath Volume		3.539	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: CARPENTER

DATE/TIME OF TEST: 05/24/2016 21:02:47

Diagnostics Check		OK	20:56
Air Blank		0.000	20:56
Control Test		0.081	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.215	20:58
Breath Volume		3.550	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.204	21:01
Breath Volume		2.375	21:01
Air Blank		0.000	21:01
Control Test		0.079	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 05/26/2016 03:58:31

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.082	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.000	03:54
Breath Volume		2.835	03:54
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.000	03:57
Breath Volume		3.222	03:57
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/26/2016 05:32:27

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.081	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.200	05:27
Breath Volume		1.757	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.213	05:30
Breath Volume		1.199	05:30
Air Blank		0.000	05:31
Control Test		0.081	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 05/27/2016 03:10:11

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.081	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1	0.079	03:00	Volume Not Met
Breath Volume		0.347	03:00	
Air Blank		0.000	03:01	
Air Blank		0.000	03:03	
Subject Sample	# 2	0.096	03:04	
Breath Volume		1.558	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 3	0.099	03:08	
Breath Volume		1.765	03:08	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: REISCH,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: REAY

DATE/TIME OF TEST: 05/27/2016 15:50:18

Diagnostics Check		OK	15:41
Air Blank		0.000	15:41
Control Test		0.081	15:42
Air Blank		0.000	15:42
Subject Sample	# 1	0.000	15:45
Breath Volume		1.929	15:45
Air Blank		0.000	15:46
Air Blank		0.000	15:48
Subject Sample	# 2	0.000	15:48
Breath Volume		1.378	15:48
Air Blank		0.000	15:49
Control Test		0.081	15:49
Air Blank		0.000	15:50
Diagnostics Check		OK	15:50

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 05/28/2016 00:20:45

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.081	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.098	00:16
Breath Volume		2.417	00:16
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.099	00:19
Breath Volume		3.105	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MAYO-LOPEZ

DATE/TIME OF TEST: 05/28/2016 04:42:45

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.208	04:38
Breath Volume		3.132	04:38
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.215	04:41
Breath Volume		3.476	04:41
Air Blank		0.000	04:41
Control Test		0.081	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: UNO,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: BODLEY

DATE/TIME OF TEST: 05/28/2016 18:38:57

Diagnostics Check		OK	18:31
Air Blank		0.000	18:31
Control Test		0.082	18:32
Air Blank		0.000	18:32
Subject Sample	# 1	0.181	18:34
Breath Volume		2.097	18:34
Air Blank		0.000	18:35
Air Blank		0.000	18:36
Subject Sample	# 2	0.173	18:37
Breath Volume		1.820	18:37
Air Blank		0.000	18:38
Control Test		0.080	18:38
Air Blank		0.000	18:38
Diagnostics Check		OK	18:38

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HOLLIDAY

DATE/TIME OF TEST: 05/28/2016 21:57:11

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.193	21:52
Breath Volume		2.093	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.193	21:55
Breath Volume		1.910	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: KOPSAK

DATE/TIME OF TEST: 05/29/2016 02:04:38

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.081	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.288	02:00
Breath Volume		2.914	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.282	02:02
Breath Volume		2.332	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 06/07/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: LAUTERMILCH

DATE/TIME OF TEST: 05/29/2016 05:20:25

Diagnostics Check		OK	05:14
Air Blank		0.000	05:14
Control Test		0.080	05:15
Air Blank		0.000	05:15
Subject Sample	# 1	0.105	05:15
Breath Volume		1.371	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.100	05:18
Breath Volume		1.882	05:18
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: MELENDEZ-COLON

DATE/TIME OF TEST: 05/29/2016 21:18:07

Diagnostics Check		OK	21:06	
Air Blank		0.000	21:06	
Control Test		0.081	21:07	
Air Blank		0.000	21:07	
Subject Sample	# 1		21:10	No Sample Provided
Breath Volume		0.000	21:10	
Air Blank		0.000	21:11	
Air Blank		0.000	21:13	
Subject Sample	# 2	0.000	21:16	Volume Not Met
Breath Volume		0.000	21:16	
Air Blank		0.000	21:17	
Control Test		0.080	21:17	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: LOCKE

DATE/TIME OF TEST: 05/31/2016 00:21:58

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.083	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.000	00:17
Breath Volume		2.984	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.000	00:20
Breath Volume		3.250	00:20
Air Blank		0.000	00:21
Control Test		0.082	00:21
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/31/2016 04:38:30

Diagnostics Check		OK	04:30	
Air Blank		0.000	04:30	
Control Test		0.082	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:34	No Sample Provided
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:37	
Control Test		0.081	04:37	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 06/07/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: TATE GREENE

DATE/TIME OF TEST: 05/31/2016 05:13:14

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:01	
Control Test		0.081	05:01	
Air Blank		0.000	05:02	
Subject Sample	# 1	0.134	05:05	Volume Not Met
Breath Volume		0.250	05:05	
Air Blank		0.000	05:06	
Air Blank		0.000	05:08	
Subject Sample	# 2	0.132	05:08	
Breath Volume		1.207	05:08	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 3	0.128	05:11	
Breath Volume		1.199	05:11	
Air Blank		0.000	05:12	
Control Test		0.080	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 06/07/2016 11:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: PAULLEY

DATE/TIME OF TEST: 05/18/2016 21:48:43

Diagnostics Check		OK	21:43	
Air Blank		0.000	21:43	
Control Test		0.079	21:43	
Air Blank		0.000	21:44	
Subject Sample	# 1	0.089	21:46	
Breath Volume		1.226	21:46	
Air Blank			21:46	RFI Detect
Air Blank		0.000	21:47	
Control Test		0.080	21:47	
Air Blank			21:47	Ambient Fail
Air Blank			21:48	Purge Fail

Data Download Date/Time: 05/22/2016 22:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 05/22/2016 03:00:33

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.080	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1	0.107	02:53	Slope Not Level
Breath Volume		1.945	02:53	
Air Blank		0.000	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 2	0.095	02:58	Volume Not Met
Breath Volume		1.039	02:58	
Air Blank		0.000	02:59	
Control Test		0.080	02:59	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 05/22/2016 22:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 05/22/2016 03:12:12

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.079	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.103	03:06
Breath Volume		1.574	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.103	03:10
Breath Volume		1.773	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 05/22/2016 22:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BENAVIDES AQUILILLA

DATE/TIME OF TEST: 05/08/2016 01:07:25

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.176	01:02
Breath Volume		2.414	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.176	01:05
Breath Volume		2.539	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 06/29/2016 09:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: VARGAS BOU

DATE/TIME OF TEST: 05/21/2016 03:58:59

Diagnostics Check	OK	03:58	
Air Blank		03:58	RFI Detect
Air Blank	0.000	03:58	

Data Download Date/Time: 06/29/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: VARGAS BOU

DATE/TIME OF TEST: 05/21/2016 04:05:43

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.078	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.007	04:01
Breath Volume		1.609	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:04
Breath Volume		1.421	04:04
Air Blank		0.000	04:04
Control Test		0.079	04:05
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 06/29/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.0.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: RANK

DATE/TIME OF TEST: 05/21/2016 23:32:00

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.078	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.000	23:27
Breath Volume		5.246	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.000	23:30
Breath Volume		1.468	23:30
Air Blank		0.000	23:31
Control Test		0.078	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:32

Data Download Date/Time: 06/29/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CLOUGH

DATE/TIME OF TEST: 05/25/2016 04:37:37

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.000	04:33
Breath Volume		1.582	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.000	04:36
Breath Volume		1.820	04:36
Air Blank		0.000	04:36
Control Test		0.077	04:37
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 06/29/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003937

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: TAURONI

DATE/TIME OF TEST: 05/26/2016 01:11:41

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.080	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.217	01:06
Breath Volume		2.019	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.210	01:10
Breath Volume		2.296	01:10
Air Blank		0.000	01:10
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 06/29/2016 09:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLOW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MOLESTINA

DATE/TIME OF TEST: 05/15/2016 00:00:27

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.078	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.000	23:56
Breath Volume		2.171	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		2.097	23:59
Air Blank		0.000	23:59
Control Test		0.076	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 06/23/2016 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004771

INSTRUMENT REGISTERED TO: HURLBURT FIELD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,AUSTIN,J
BREATH TEST OPERATOR AGENCY: Hurlburt Field PD
ARRESTING OFFICER: ULLRICH,JASON
ARRESTING OFFICER AGENCY: Hurlburt Field PD
LAST AGENCY INSPECTION: 04/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 05/07/2016 01:48:24

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.028	01:44
Breath Volume		2.941	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.022	01:46
Breath Volume		2.472	01:46
Air Blank		0.000	01:47
Control Test		0.076	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 05/31/2016 17:32

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMSON,TYLER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: CUSHING

DATE/TIME OF TEST: 05/01/2016 03:00:36

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.079	02:58	
Air Blank		0.000	02:59	
Subject Sample	# 1		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	02:59	
Control Test		0.080	03:00	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EMRICHI,IAN,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EMRICH,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: QUINTANA

DATE/TIME OF TEST: 05/01/2016 06:29:10

Diagnostics Check		OK	06:22
Air Blank		0.000	06:23
Control Test		0.080	06:23
Air Blank		0.000	06:23
Subject Sample	# 1	0.113	06:24
Breath Volume		2.242	06:24
Air Blank		0.000	06:25
Air Blank		0.000	06:27
Subject Sample	# 2	0.117	06:27
Breath Volume		4.175	06:27
Air Blank		0.000	06:28
Control Test		0.080	06:28
Air Blank		0.000	06:29
Diagnostics Check		OK	06:29

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,DUSTAN,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RICE,DUSTAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 05/02/2016 07:13:03

Diagnostics Check		OK	07:04	
Air Blank		0.000	07:04	
Control Test		0.080	07:04	
Air Blank		0.000	07:05	
Subject Sample	# 1		07:08	No Sample Provided
Breath Volume		0.000	07:08	
Air Blank		0.000	07:09	
Air Blank		0.000	07:11	
Subject Sample	# 2		07:11	Subject Test Refused
Breath Volume		0.000	07:11	
Air Blank		0.000	07:12	
Control Test		0.079	07:12	
Air Blank		0.000	07:12	
Diagnostics Check		OK	07:13	

Data Download Date/Time: 05/28/2016 09:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: HALBRUEGGE

DATE/TIME OF TEST: 05/07/2016 02:03:09

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.079	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:58	No Sample Provided
Breath Volume		0.000	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:02	
Control Test		0.080	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JERARD,ELISA,K
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CULBERSON,DARYL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: BUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 05/07/2016 22:12:16

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:09	
Control Test		0.080	22:10	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:10	Subject Test Refused
Breath Volume		0.000	22:10	
Air Blank		0.000	22:11	
Control Test		0.080	22:11	
Air Blank		0.000	22:12	
Diagnostics Check		OK	22:12	

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NAYLOR,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LETOURNEAU

DATE/TIME OF TEST: 05/08/2016 04:53:08

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:48	
Control Test		0.080	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:51	Subject Test Refused
Breath Volume		0.000	04:51	
Air Blank		0.000	04:52	
Control Test		0.080	04:52	
Air Blank		0.000	04:53	
Diagnostics Check		OK	04:53	

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAND,GLENN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 05/08/2016 05:18:13

Diagnostics Check		OK	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.246	05:13
Breath Volume		2.996	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.233	05:16
Breath Volume		1.792	05:16
Air Blank		0.000	05:17
Control Test		0.078	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: NAYLOR,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: PLANTZ

DATE/TIME OF TEST: 05/13/2016 05:39:45

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:37	
Control Test		0.079	05:37	
Air Blank		0.000	05:37	
Subject Sample	# 1		05:38	Subject Test Refused
Breath Volume		0.000	05:38	
Air Blank		0.000	05:38	
Control Test		0.080	05:39	
Air Blank		0.000	05:39	
Diagnostics Check		OK	05:39	

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 05/14/2016 01:55:46

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.079	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:54	
Control Test		0.079	01:55	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 05/28/2016 09:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 05/14/2016 04:10:38

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.079	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.080	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 05/28/2016 09:13

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCFALL,CHARLES,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: GORDON,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: IVEY

DATE/TIME OF TEST: 05/14/2016 06:06:39

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.079	06:00
Air Blank		0.000	06:01
Subject Sample	# 1	0.173	06:02
Breath Volume		2.601	06:02
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.169	06:05
Breath Volume		2.492	06:05
Air Blank		0.000	06:05
Control Test		0.079	06:06
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,MARK,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KEMP,MARK
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 05/15/2016 20:28:10

Diagnostics Check		OK	20:19
Air Blank		0.000	20:20
Control Test		0.080	20:20
Air Blank		0.000	20:20
Subject Sample	# 1	0.229	20:22
Breath Volume		1.886	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.226	20:26
Breath Volume		1.281	20:26
Air Blank		0.000	20:27
Control Test		0.077	20:27
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDERSON,KELLY,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDRICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BEETS

DATE/TIME OF TEST: 05/20/2016 00:16:05

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:13	
Control Test		0.079	00:13	
Air Blank		0.000	00:14	
Subject Sample	# 1		00:14	Subject Test Refused
Breath Volume		0.000	00:14	
Air Blank		0.000	00:15	
Control Test		0.079	00:15	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 05/20/2016 23:07:47

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:53	
Control Test		0.079	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 1	0.153	22:55	
Breath Volume		3.527	22:55	
Air Blank		0.000	22:55	
Air Blank		0.000	22:57	
Subject Sample	# 2		23:00	No Sample Provided
Breath Volume		0.000	23:00	
Air Blank		0.000	23:01	
Air Blank		0.000	23:03	
Subject Sample	# 3		23:06	No Sample Provided
Breath Volume		0.000	23:06	
Air Blank		0.000	23:06	
Control Test		0.078	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 05/20/2016 23:39:46

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:25	
Control Test		0.080	23:25	
Air Blank		0.000	23:26	
Subject Sample	# 1	0.136	23:27	
Breath Volume		2.890	23:27	
Air Blank		0.000	23:27	
Air Blank		0.000	23:29	
Subject Sample	# 2		23:32	No Sample Provided
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 3		23:38	No Sample Provided
Breath Volume		0.000	23:38	
Air Blank		0.000	23:38	
Control Test		0.079	23:39	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:39	

Data Download Date/Time: 05/28/2016 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 05/21/2016 00:01:34

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.134	23:56
Breath Volume		3.730	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.140	00:00
Breath Volume		4.109	00:00
Air Blank		0.000	00:00
Control Test		0.078	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 05/21/2016 01:54:35

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.079	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.115	01:49
Breath Volume		4.074	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.111	01:53
Breath Volume		3.429	01:53
Air Blank		0.000	01:53
Control Test		0.079	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: FARNAM

DATE/TIME OF TEST: 05/21/2016 05:15:56

Diagnostics Check		OK	05:09
Air Blank		0.000	05:09
Control Test		0.079	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.242	05:11
Breath Volume		2.265	05:11
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.244	05:14
Breath Volume		1.335	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CLEVELAND

DATE/TIME OF TEST: 05/22/2016 01:26:46

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.015	01:22
Breath Volume		3.695	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.014	01:25
Breath Volume		3.562	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BROWN,RAPHEAL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 05/22/2016 05:36:36

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.080	05:30
Air Blank		0.000	05:31
Subject Sample	# 1	0.176	05:31
Breath Volume		1.699	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.173	05:35
Breath Volume		1.507	05:35
Air Blank		0.000	05:35
Control Test		0.079	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: POGOZELSKI,CHRIS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 05/22/2016 06:11:57

Diagnostics Check		OK	06:05
Air Blank		0.000	06:05
Control Test		0.079	06:05
Air Blank		0.000	06:06
Subject Sample	# 1	0.000	06:07
Breath Volume		1.593	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:10
Subject Sample	# 2	0.000	06:10
Breath Volume		1.941	06:10
Air Blank		0.000	06:11
Control Test		0.079	06:11
Air Blank		0.000	06:11
Diagnostics Check		OK	06:11

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,DUSTAN,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: FREDRICK,TAVISH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: ZOLLNER

DATE/TIME OF TEST: 05/23/2016 21:11:23

Diagnostics Check		OK	21:08	
Air Blank		0.000	21:08	
Control Test		0.079	21:09	
Air Blank		0.000	21:09	
Subject Sample	# 1		21:10	Subject Test Refused
Breath Volume		0.000	21:10	
Air Blank		0.000	21:10	
Control Test		0.080	21:10	
Air Blank		0.000	21:11	
Diagnostics Check		OK	21:11	

Data Download Date/Time: 05/28/2016 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLICK,LOREN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SLICK,LOREN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 05/25/2016 23:24:15

Diagnostics Check		OK	23:17
Air Blank		0.000	23:18
Control Test		0.080	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.173	23:19
Breath Volume		4.679	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.179	23:22
Breath Volume		4.855	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CLAUSNER,ALEZANDRA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ROCHER

DATE/TIME OF TEST: 05/26/2016 04:43:51

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.080	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.167	04:39
Breath Volume		1.457	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.166	04:42
Breath Volume		1.949	04:42
Air Blank		0.000	04:42
Control Test		0.078	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANKINS,MEGAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SCHRADER,TODD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 05/27/2016 16:04:24

Diagnostics Check		OK	15:50	
Air Blank		0.000	15:50	
Control Test		0.080	15:50	
Air Blank		0.000	15:51	
Subject Sample	# 1		15:54	No Sample Provided
Breath Volume		0.000	15:54	
Air Blank		0.000	15:54	
Air Blank		0.000	15:56	
Subject Sample	# 2	0.240	15:59	
Breath Volume		2.589	15:59	
Air Blank		0.000	15:59	
Air Blank		0.000	16:01	
Subject Sample	# 3	0.245	16:02	
Breath Volume		2.343	16:02	
Air Blank		0.000	16:03	
Control Test		0.080	16:03	
Air Blank		0.000	16:04	
Diagnostics Check		OK	16:04	

Data Download Date/Time: 05/28/2016 09:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 05/28/2016 06:36:21

Diagnostics Check		OK	06:29
Air Blank		0.000	06:30
Control Test		0.079	06:30
Air Blank		0.000	06:31
Subject Sample	# 1	0.176	06:31
Breath Volume		2.707	06:31
Air Blank		0.000	06:32
Air Blank		0.000	06:34
Subject Sample	# 2	0.187	06:34
Breath Volume		2.906	06:34
Air Blank		0.000	06:35
Control Test		0.079	06:35
Air Blank		0.000	06:36
Diagnostics Check		OK	06:36

Data Download Date/Time: 05/28/2016 09:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANKINS,MEGAN,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:13

SUBJECT'S LAST NAME: FLUCKER

DATE/TIME OF TEST: 05/29/2016 08:44:50

Diagnostics Check		OK	08:33	
Air Blank		0.000	08:33	
Control Test		0.080	08:34	
Air Blank		0.000	08:34	
Subject Sample	# 1		08:37	No Sample Provided
Breath Volume		0.000	08:37	
Air Blank		0.000	08:38	
Air Blank		0.000	08:40	
Subject Sample	# 2	0.000	08:43	Volume Not Met
Breath Volume		0.355	08:43	
Air Blank		0.000	08:43	
Control Test		0.080	08:44	
Air Blank		0.000	08:44	
Diagnostics Check		OK	08:44	

Data Download Date/Time: 06/16/2016 11:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUNTZ,MARK,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KUNTZ,MARK
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 05/30/2016 21:15:00

Diagnostics Check		OK	21:06	
Air Blank		0.000	21:06	
Control Test		0.079	21:07	
Air Blank		0.000	21:07	
Subject Sample	# 1		21:10	No Sample Provided
Breath Volume		0.000	21:10	
Air Blank		0.000	21:11	
Air Blank		0.000	21:13	
Subject Sample	# 2		21:13	Subject Test Refused
Breath Volume		0.000	21:13	
Air Blank		0.000	21:14	
Control Test		0.079	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:15	

Data Download Date/Time: 06/16/2016 11:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR,ROBERT,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BROWN,RAPHEAL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MAUS

DATE/TIME OF TEST: 05/31/2016 01:34:01

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.247	01:29
Breath Volume		3.539	01:29
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.254	01:32
Breath Volume		2.382	01:32
Air Blank		0.000	01:33
Control Test		0.078	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:34

Data Download Date/Time: 06/16/2016 11:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 05/31/2016 02:04:09

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.079	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.212	01:59
Breath Volume		1.886	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:01
Subject Sample	# 2	0.216	02:02
Breath Volume		1.671	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 06/16/2016 11:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: BURGESS,JOSHUA
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: GULLEDGE SUNS

DATE/TIME OF TEST: 05/01/2016 21:08:05

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.081	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.188	21:03
Breath Volume		1.113	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:05
Subject Sample	# 2	0.190	21:06
Breath Volume		1.335	21:06
Air Blank		0.000	21:07
Control Test		0.079	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 06/01/2016 16:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS,EVAN,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: NEWCOMB,WILLIAM
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 05/03/2016 20:55:05

Diagnostics Check		OK	20:52	
Air Blank		0.000	20:52	
Control Test		0.082	20:53	
Air Blank		0.000	20:53	
Subject Sample	# 1		20:53	Subject Test Refused
Breath Volume		0.000	20:53	
Air Blank		0.000	20:54	
Control Test		0.082	20:54	
Air Blank		0.000	20:55	
Diagnostics Check		OK	20:55	

Data Download Date/Time: 06/01/2016 16:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: ELLISON

DATE/TIME OF TEST: 05/06/2016 03:57:06

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.081	03:55	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:56	
Control Test		0.080	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 06/01/2016 16:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MINER

DATE/TIME OF TEST: 05/14/2016 23:15:35

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.081	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1	0.178	23:10	Volume Not Met
Breath Volume		0.472	23:10	
Air Blank		0.000	23:11	
Air Blank		0.000	23:13	
Subject Sample	# 2		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:14	
Control Test		0.078	23:14	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 06/01/2016 16:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: NEYMAN

DATE/TIME OF TEST: 05/16/2016 01:41:36

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.080	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.136	01:36
Breath Volume		1.316	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:38
Subject Sample	# 2	0.126	01:39
Breath Volume		1.429	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 06/01/2016 16:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: NEWCOMB,COREY
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 05/21/2016 04:27:14

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:14	
Control Test		0.081	04:15	
Air Blank		0.000	04:15	
Subject Sample	# 1	0.175	04:18	Slope Not Level
Breath Volume		1.269	04:18	
Air Blank		0.000	04:19	
Air Blank		0.000	04:21	
Subject Sample	# 2	0.188	04:21	
Breath Volume		1.578	04:21	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 3	0.181	04:25	
Breath Volume		1.722	04:25	
Air Blank		0.000	04:26	
Control Test		0.077	04:26	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 06/01/2016 16:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MACKENDREE

DATE/TIME OF TEST: 05/24/2016 01:52:02

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.045	01:47
Breath Volume		2.484	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.043	01:50
Breath Volume		2.937	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:52

Data Download Date/Time: 06/01/2016 16:11

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEALS,JOSEPH,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SEALS,JOSEPH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 04/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 05/27/2016 15:30:47

Diagnostics Check		OK	15:24
Air Blank		0.000	15:25
Control Test		0.081	15:25
Air Blank		0.000	15:25
Subject Sample	# 1	0.088	15:26
Breath Volume		3.625	15:26
Air Blank		0.000	15:26
Air Blank		0.000	15:28
Subject Sample	# 2	0.086	15:29
Breath Volume		3.082	15:29
Air Blank		0.000	15:29
Control Test		0.081	15:30
Air Blank		0.000	15:30
Diagnostics Check		OK	15:30

Data Download Date/Time: 06/01/2016 16:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: VERNSEY

DATE/TIME OF TEST: 05/17/2016 01:37:35

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.077	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.186	01:31
Breath Volume		1.402	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.178	01:36
Breath Volume		1.574	01:36
Air Blank		0.000	01:36
Control Test		0.076	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 05/26/2016 09:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: WISMER

DATE/TIME OF TEST: 05/26/2016 02:04:25

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.078	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.153	01:59
Breath Volume		2.164	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.159	02:02
Breath Volume		1.390	02:02
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 05/26/2016 09:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BUNDE

DATE/TIME OF TEST: 05/31/2016 01:14:09

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.182	01:09
Breath Volume		2.093	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.180	01:12
Breath Volume		1.742	01:12
Air Blank		0.000	01:13
Control Test		0.077	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 06/28/2016 08:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: SHILONSANTIZ

DATE/TIME OF TEST: 05/21/2016 23:02:41

Diagnostics Check		OK	22:54
Air Blank		0.000	22:55
Control Test		0.077	22:55
Air Blank		0.000	22:56
Subject Sample	# 1	0.168	22:57
Breath Volume		1.117	22:57
Air Blank		0.000	22:57
Air Blank		0.000	22:59
Subject Sample	# 2	0.162	23:01
Breath Volume		1.140	23:01
Air Blank		0.000	23:01
Control Test		0.078	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 06/07/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 05/22/2016 02:38:42

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.077	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.106	02:34
Breath Volume		2.230	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.100	02:37
Breath Volume		1.886	02:37
Air Blank		0.000	02:37
Control Test		0.077	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 06/07/2016 11:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: PENENORIESTRADA

DATE/TIME OF TEST: 05/22/2016 23:01:05

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.078	22:59	
Air Blank		0.000	23:00	
Subject Sample	# 1		23:00	RFI Detect
Breath Volume		0.000	23:00	
Air Blank			23:00	RFI Detect
Air Blank		0.000	23:01	

Data Download Date/Time: 06/07/2016 11:36

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: PENENORIESTRADA

DATE/TIME OF TEST: 05/22/2016 23:09:06

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.078	23:03
Air Blank		0.000	23:03
Subject Sample	# 1	0.209	23:04
Breath Volume		4.218	23:04
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.200	23:07
Breath Volume		3.039	23:07
Air Blank		0.000	23:08
Control Test		0.078	23:08
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 06/07/2016 11:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 05/02/2016 02:33:45

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.083	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.182	02:28
Breath Volume		2.429	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.190	02:32
Breath Volume		2.898	02:32
Air Blank		0.000	02:32
Control Test		0.082	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 05/20/2016 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 05/04/2016 20:22:21

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.083	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.069	20:17
Breath Volume		1.476	20:17
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.069	20:20
Breath Volume		1.828	20:20
Air Blank		0.000	20:21
Control Test		0.082	20:21
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 05/20/2016 08:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MEDLEY,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 05/06/2016 03:48:42

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:46	
Control Test		0.083	03:46	
Air Blank		0.000	03:47	
Subject Sample	# 1		03:47	Subject Test Refused
Breath Volume		0.000	03:47	
Air Blank		0.000	03:47	
Control Test		0.083	03:48	
Air Blank		0.000	03:48	
Diagnostics Check		OK	03:48	

Data Download Date/Time: 05/20/2016 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 05/06/2016 23:07:44

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.083	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.098	23:02
Breath Volume		2.691	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.106	23:06
Breath Volume		3.183	23:06
Air Blank		0.000	23:06
Control Test		0.082	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 05/20/2016 08:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MURPHY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 05/10/2016 23:58:40

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.083	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.142	23:54
Breath Volume		3.722	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.128	23:57
Breath Volume		2.519	23:57
Air Blank		0.000	23:57
Control Test		0.083	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 05/20/2016 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: NICHOLS,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CHEWNING

DATE/TIME OF TEST: 05/15/2016 02:43:47

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.083	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Control Test		0.083	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 05/20/2016 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HIBBARD

DATE/TIME OF TEST: 05/16/2016 01:47:18

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.084	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.197	01:42
Breath Volume		2.230	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.207	01:45
Breath Volume		2.667	01:45
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 05/20/2016 08:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LAVALLE,D
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: RICE

DATE/TIME OF TEST: 05/21/2016 01:29:10

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:18	
Control Test		0.083	01:18	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.118	01:22	Volume Not Met
Breath Volume		0.472	01:22	
Air Blank		0.000	01:22	
Air Blank		0.000	01:24	
Subject Sample	# 2	0.131	01:27	Volume Not Met
Breath Volume		0.085	01:27	
Air Blank		0.000	01:28	
Control Test		0.081	01:28	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: AULICINO,M
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/21/2016 02:13:27

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.082	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.133	02:08
Breath Volume		1.125	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.124	02:11
Breath Volume		2.437	02:11
Air Blank		0.000	02:12
Control Test		0.082	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: JOHNSON,S
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: BAITY

DATE/TIME OF TEST: 05/21/2016 23:55:30

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.083	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.000	23:51
Breath Volume		2.687	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.000	23:54
Breath Volume		3.132	23:54
Air Blank		0.000	23:54
Control Test		0.083	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 05/22/2016 22:35:32

Diagnostics Check		OK	22:28
Air Blank		0.000	22:29
Control Test		0.083	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.268	22:30
Breath Volume		2.515	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.263	22:33
Breath Volume		2.687	22:33
Air Blank		0.000	22:34
Control Test		0.082	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: GREENLEAF

DATE/TIME OF TEST: 05/26/2016 00:38:16

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:35	
Control Test		0.084	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:36	Subject Test Refused
Breath Volume		0.000	00:36	
Air Blank		0.000	00:37	
Control Test		0.084	00:37	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:38	

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVENPORT,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ZAHN

DATE/TIME OF TEST: 05/28/2016 01:57:28

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.083	01:51
Air Blank		0.000	01:52
Subject Sample	# 1	0.250	01:52
Breath Volume		2.230	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.237	01:55
Breath Volume		1.949	01:55
Air Blank		0.000	01:56
Control Test		0.082	01:56
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 06/17/2016 08:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LITTLE,J
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: BUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: GUARNERI

DATE/TIME OF TEST: 05/29/2016 20:58:49

Diagnostics Check		OK	20:55	
Air Blank		0.000	20:56	
Control Test		0.083	20:56	
Air Blank		0.000	20:57	
Subject Sample	# 1		20:57	Subject Test Refused
Breath Volume		0.000	20:57	
Air Blank		0.000	20:57	
Control Test		0.083	20:58	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BERGWERFF,R
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: BUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: MEIER

DATE/TIME OF TEST: 05/29/2016 22:09:44

Diagnostics Check		OK	22:02
Air Blank		0.000	22:02
Control Test		0.082	22:03
Air Blank		0.000	22:03
Subject Sample	# 1	0.109	22:05
Breath Volume		2.605	22:05
Air Blank		0.000	22:05
Air Blank		0.000	22:07
Subject Sample	# 2	0.112	22:08
Breath Volume		1.808	22:08
Air Blank		0.000	22:08
Control Test		0.081	22:09
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HENRY,J
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: BUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 05/30/2016 18:13:50

Diagnostics Check		OK	18:11	
Air Blank		0.000	18:11	
Control Test		0.083	18:11	
Air Blank		0.000	18:12	
Subject Sample	# 1		18:12	Subject Test Refused
Breath Volume		0.000	18:12	
Air Blank		0.000	18:12	
Control Test		0.083	18:13	
Air Blank		0.000	18:13	
Diagnostics Check		OK	18:13	

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: PULASKI,D
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 05/31/2016 04:17:31

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.083	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.145	04:12
Breath Volume		3.058	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.148	04:15
Breath Volume		2.796	04:15
Air Blank		0.000	04:16
Control Test		0.083	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 06/17/2016 08:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERY,T,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 05/31/2016 23:40:47

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.082	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.165	23:34
Breath Volume		1.218	23:34
Air Blank		0.000	23:35
Air Blank		0.000	23:37
Subject Sample	# 2	0.154	23:39
Breath Volume		1.488	23:39
Air Blank		0.000	23:39
Control Test		0.081	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 06/17/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LEAVITT

DATE/TIME OF TEST: 05/01/2016 03:50:54

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:45
Subject Sample	# 1	0.201	03:45
Breath Volume		2.171	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.201	03:49
Breath Volume		1.878	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 05/25/2016 22:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LORING,BLAINE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:42

SUBJECT'S LAST NAME: SZILAGYI

DATE/TIME OF TEST: 05/01/2016 17:08:41

Diagnostics Check		OK	17:02
Air Blank		0.000	17:03
Control Test		0.080	17:03
Air Blank		0.000	17:03
Subject Sample	# 1	0.000	17:04
Breath Volume		1.507	17:04
Air Blank		0.000	17:04
Air Blank		0.000	17:06
Subject Sample	# 2	0.000	17:07
Breath Volume		1.601	17:07
Air Blank		0.000	17:07
Control Test		0.080	17:08
Air Blank		0.000	17:08
Diagnostics Check		OK	17:08

Data Download Date/Time: 05/25/2016 22:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 05/01/2016 20:15:09

Diagnostics Check		OK	20:06
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.133	20:08
Breath Volume		1.996	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.134	20:13
Breath Volume		2.539	20:13
Air Blank		0.000	20:14
Control Test		0.080	20:14
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GODINEZ

DATE/TIME OF TEST: 05/04/2016 03:50:33

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.082	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.196	03:43
Breath Volume		1.191	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.219	03:46
Breath Volume		1.644	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 3	0.224	03:48
Breath Volume		2.023	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 05/25/2016 22:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: COWLES

DATE/TIME OF TEST: 05/04/2016 20:55:25

Diagnostics Check		OK	20:49
Air Blank		0.000	20:49
Control Test		0.081	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.000	20:51
Breath Volume		1.636	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.000	20:54
Breath Volume		1.261	20:54
Air Blank		0.000	20:54
Control Test		0.081	20:54
Air Blank		0.000	20:55
Diagnostics Check		OK	20:55

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 05/05/2016 02:01:07

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.081	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DRIVER

DATE/TIME OF TEST: 05/05/2016 02:56:30

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.163	02:51
Breath Volume		3.433	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.165	02:54
Breath Volume		3.214	02:54
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: REITER

DATE/TIME OF TEST: 05/05/2016 04:36:41

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.196	04:32
Breath Volume		1.687	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.206	04:35
Breath Volume		1.683	04:35
Air Blank		0.000	04:35
Control Test		0.080	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: BLOESCH

DATE/TIME OF TEST: 05/05/2016 21:09:36

Diagnostics Check		OK	21:06	
Air Blank		0.000	21:07	
Control Test		0.081	21:07	
Air Blank		0.000	21:07	
Subject Sample	# 1		21:08	Subject Test Refused
Breath Volume		0.000	21:08	
Air Blank		0.000	21:08	
Control Test		0.081	21:09	
Air Blank		0.000	21:09	
Diagnostics Check		OK	21:09	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: OSTROWSKY

DATE/TIME OF TEST: 05/06/2016 01:38:40

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.082	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.080	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GIBBONS

DATE/TIME OF TEST: 05/06/2016 04:31:08

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.082	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.245	04:25
Breath Volume		1.789	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.238	04:29
Breath Volume		1.804	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BLOOM

DATE/TIME OF TEST: 05/07/2016 03:12:31

Diagnostics Check	OK	03:11	
Air Blank		03:12	Ambient Fail
Air Blank	0.000	03:12	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BLOOM

DATE/TIME OF TEST: 05/07/2016 03:28:03

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.080	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1	0.188	03:18	
Breath Volume		1.667	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 2	0.166	03:23	Volume Not Met
Breath Volume		0.082	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:26	
			03:26	Improper Sample
Air Blank		0.000	03:27	
Control Test		0.077	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WELSHANS,DEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 05/07/2016 06:26:55

Diagnostics Check		OK	06:20
Air Blank		0.000	06:20
Control Test		0.079	06:21
Air Blank		0.000	06:21
Subject Sample	# 1	0.155	06:22
Breath Volume		1.378	06:22
Air Blank		0.000	06:22
Air Blank		0.000	06:24
Subject Sample	# 2	0.154	06:25
Breath Volume		3.074	06:25
Air Blank		0.000	06:26
Control Test		0.080	06:26
Air Blank		0.000	06:26
Diagnostics Check		OK	06:26

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DROSIS

DATE/TIME OF TEST: 05/07/2016 22:01:13

Diagnostics Check		OK	21:54
Air Blank		0.000	21:54
Control Test		0.082	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.134	21:56
Breath Volume		2.125	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:58
Subject Sample	# 2	0.136	21:59
Breath Volume		2.832	21:59
Air Blank		0.000	22:00
Control Test		0.080	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: KOHLER

DATE/TIME OF TEST: 05/08/2016 01:02:01

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.229	00:57
Breath Volume		2.000	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.232	01:00
Breath Volume		1.769	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 05/25/2016 22:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: RUBACH

DATE/TIME OF TEST: 05/08/2016 02:30:18

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:20	No Sample Provided
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2	0.146	02:24	
Breath Volume		2.355	02:24	
Air Blank		0.000	02:25	
Air Blank		0.000	02:27	
Subject Sample	# 3	0.155	02:28	
Breath Volume		2.351	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 05/25/2016 22:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NOCERA,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MC EWEN

DATE/TIME OF TEST: 05/08/2016 03:30:21

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.156	03:25
Breath Volume		1.285	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.160	03:28
Breath Volume		1.113	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 05/25/2016 22:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 05/08/2016 04:52:35

Diagnostics Check		OK	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.111	04:47
Breath Volume		1.636	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.116	04:51
Breath Volume		1.796	04:51
Air Blank		0.000	04:51
Control Test		0.079	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHR,K,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 05/11/2016 02:07:41

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.081	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:06	Subject Test Refused
Breath Volume		0.000	02:06	
Air Blank		0.000	02:06	
Control Test		0.081	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FISCHETTI

DATE/TIME OF TEST: 05/13/2016 03:50:26

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.000	03:46
Breath Volume		1.925	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.000	03:49
Breath Volume		1.609	03:49
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: YOCUM

DATE/TIME OF TEST: 05/14/2016 00:57:31

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.081	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.099	00:52
Breath Volume		2.078	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.097	00:55
Breath Volume		2.125	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CAMPOS CRUZ

DATE/TIME OF TEST: 05/14/2016 03:39:49

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:37	
Control Test		0.081	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:38	
Control Test		0.079	03:39	
Air Blank		0.000	03:39	
Diagnostics Check		OK	03:39	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 05/14/2016 04:19:57

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.081	04:09	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:12	No Sample Provided
Breath Volume		0.000	04:12	
Air Blank		0.000	04:13	
Air Blank		0.000	04:15	
Subject Sample	# 2	0.132	04:18	Volume Not Met
Breath Volume		0.316	04:18	
Air Blank		0.000	04:19	
Control Test		0.081	04:19	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 05/14/2016 04:36:12

Diagnostics Check		OK	04:21	
Air Blank		0.000	04:22	
Control Test		0.079	04:22	
Air Blank		0.000	04:22	
Subject Sample	# 1	0.134	04:26	Volume Not Met
Breath Volume		0.023	04:26	
Air Blank		0.000	04:26	
Air Blank		0.000	04:28	
Subject Sample	# 2	0.159	04:29	
Breath Volume		1.550	04:29	
Air Blank		0.000	04:30	
Air Blank		0.000	04:32	
Subject Sample	# 3	0.162	04:34	
Breath Volume		1.566	04:34	
Air Blank		0.000	04:35	
Control Test		0.078	04:35	
Air Blank		0.000	04:36	
Diagnostics Check		OK	04:36	

Data Download Date/Time: 05/25/2016 22:19

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PATTI

DATE/TIME OF TEST: 05/15/2016 02:34:07

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.081	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:33	
Control Test		0.080	02:33	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 05/15/2016 05:14:19

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.018	05:09
Breath Volume		3.941	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.016	05:12
Breath Volume		3.914	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LOVEGROVE

DATE/TIME OF TEST: 05/18/2016 23:26:42

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.111	23:22
Breath Volume		2.679	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.113	23:25
Breath Volume		2.089	23:25
Air Blank		0.000	23:25
Control Test		0.080	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMAR,G
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: MARASON

DATE/TIME OF TEST: 05/20/2016 02:51:24

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:49	
Control Test		0.081	02:49	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:50	Subject Test Refused
Breath Volume		0.000	02:50	
Air Blank		0.000	02:50	
Control Test		0.079	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HINTHORN

DATE/TIME OF TEST: 05/20/2016 04:16:50

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.204	04:11
Breath Volume		1.683	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.203	04:15
Breath Volume		1.242	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NOCERA,D
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/21/2016 03:23:48

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.080	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 05/21/2016 03:53:30

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.081	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.080	03:48
Breath Volume		1.203	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.093	03:51
Breath Volume		2.613	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 05/21/2016 04:31:23

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.080	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.190	04:26
Breath Volume		1.640	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.177	04:29
Breath Volume		1.183	04:29
Air Blank		0.000	04:30
Control Test		0.076	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RONCA

DATE/TIME OF TEST: 05/21/2016 04:57:37

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.169	04:52
Breath Volume		1.890	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.160	04:56
Breath Volume		2.117	04:56
Air Blank		0.000	04:56
Control Test		0.079	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEARY,F
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 05/21/2016 20:07:33

Diagnostics Check		OK	20:00
Air Blank		0.000	20:01
Control Test		0.081	20:01
Air Blank		0.000	20:02
Subject Sample	# 1	0.098	20:02
Breath Volume		1.980	20:02
Air Blank		0.000	20:03
Air Blank		0.000	20:05
Subject Sample	# 2	0.100	20:06
Breath Volume		1.976	20:06
Air Blank		0.000	20:06
Control Test		0.080	20:06
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: REDMON,R
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 05/21/2016 20:38:47

Diagnostics Check		OK	20:35	
Air Blank		0.000	20:36	
Control Test		0.080	20:36	
Air Blank		0.000	20:37	
Subject Sample	# 1		20:37	Subject Test Refused
Breath Volume		0.000	20:37	
Air Blank		0.000	20:37	
Control Test		0.080	20:38	
Air Blank		0.000	20:38	
Diagnostics Check		OK	20:38	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARR,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 05/22/2016 03:24:08

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:21	
Control Test		0.081	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:23	
Control Test		0.081	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVILA,JERRY,W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DAVILA,JERRY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 05/23/2016 00:39:08

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:36	
Control Test		0.080	00:36	
Air Blank		0.000	00:37	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,ELIZABETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: KILPATRICK

DATE/TIME OF TEST: 05/25/2016 01:52:01

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.000	01:47
Breath Volume		1.437	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.000	01:50
Breath Volume		1.531	01:50
Air Blank		0.000	01:51
Control Test		0.081	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:52

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 04/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 05/25/2016 05:16:20

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.081	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.075	05:11
Breath Volume		1.457	05:11
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.080	05:14
Breath Volume		1.386	05:14
Air Blank		0.000	05:15
Control Test		0.080	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 05/25/2016 22:19

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 05/26/2016 04:49:26

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.081	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.151	04:43
Breath Volume		1.304	04:43
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.141	04:47
Breath Volume		1.960	04:47
Air Blank		0.000	04:48
Control Test		0.079	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 06/03/2016 00:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GARBER

DATE/TIME OF TEST: 05/27/2016 00:44:58

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.081	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1	0.177	00:35	Volume Not Met
Breath Volume		0.167	00:35	
Air Blank		0.000	00:36	
Air Blank		0.000	00:38	
Subject Sample	# 2	0.177	00:39	
Breath Volume		2.367	00:39	
Air Blank		0.000	00:39	
Air Blank		0.000	00:41	
Subject Sample	# 3	0.164	00:43	
Breath Volume		2.062	00:43	
Air Blank		0.000	00:44	
Control Test		0.079	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 06/03/2016 00:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: GOLDBERG

DATE/TIME OF TEST: 05/27/2016 03:34:37

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.081	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.000	03:30
Breath Volume		2.937	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.000	03:33
Breath Volume		2.585	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 06/03/2016 00:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: GAYNOR

DATE/TIME OF TEST: 05/27/2016 04:25:18

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:22	
Control Test		0.080	04:23	
Air Blank		0.000	04:23	
Subject Sample	# 1		04:23	Subject Test Refused
Breath Volume		0.000	04:23	
Air Blank		0.000	04:24	
Control Test		0.079	04:24	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 06/03/2016 00:03

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEDY,N
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 05/27/2016 23:06:31

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:04	
Control Test		0.081	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:05	
Control Test		0.080	23:05	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NESTOR,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: BERGERON

DATE/TIME OF TEST: 05/28/2016 00:15:48

Diagnostics Check		OK	00:12	
Air Blank		0.000	00:13	
Control Test		0.081	00:13	
Air Blank		0.000	00:14	
Subject Sample	# 1		00:14	Subject Test Refused
Breath Volume		0.000	00:14	
Air Blank		0.000	00:14	
Control Test		0.080	00:15	
Air Blank		0.000	00:15	
Diagnostics Check		OK	00:15	

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BOONE,A
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROSSIN

DATE/TIME OF TEST: 05/28/2016 01:46:53

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:34	
Subject Sample	# 1	0.344	01:37	Volume Not Met
Breath Volume		0.292	01:37	
Air Blank		0.000	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.333	01:41	No .020 agreement
Breath Volume		1.140	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 3	0.309	01:45	No .020 agreement
Breath Volume		1.117	01:45	
Air Blank		0.000	01:45	
Control Test		0.077	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: RADCLIFF

DATE/TIME OF TEST: 05/28/2016 05:10:37

Diagnostics Check		OK	05:07	
Air Blank		0.000	05:08	
Control Test		0.080	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:09	Subject Test Refused
Breath Volume		0.000	05:09	
Air Blank		0.000	05:09	
Control Test		0.080	05:10	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZARASVAND,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/28/2016 18:35:40

Diagnostics Check	OK	18:35	
Air Blank		18:35	RFI Detect
Air Blank	0.000	18:35	

Data Download Date/Time: 06/03/2016 00:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ZARASVAND,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 05/28/2016 18:43:52

Diagnostics Check		OK	18:37
Air Blank		0.000	18:38
Control Test		0.081	18:38
Air Blank		0.000	18:38
Subject Sample	# 1	0.113	18:39
Breath Volume		1.578	18:39
Air Blank		0.000	18:39
Air Blank		0.000	18:41
Subject Sample	# 2	0.109	18:42
Breath Volume		1.109	18:42
Air Blank		0.000	18:42
Control Test		0.080	18:43
Air Blank		0.000	18:43
Diagnostics Check		OK	18:43

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ESTELL

DATE/TIME OF TEST: 05/29/2016 02:28:06

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.081	02:26	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:26	Subject Test Refused
Breath Volume		0.000	02:26	
Air Blank		0.000	02:27	
Control Test		0.080	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLASINI,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: LOWTHER

DATE/TIME OF TEST: 05/29/2016 04:07:33

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.081	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:03	No Sample Provided
Breath Volume		0.000	04:03	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.080	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: CROUSE

DATE/TIME OF TEST: 05/29/2016 05:25:33

Diagnostics Check		OK	05:22	
Air Blank		0.000	05:23	
Control Test		0.081	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:24	Subject Test Refused
Breath Volume		0.000	05:24	
Air Blank		0.000	05:24	
Control Test		0.080	05:24	
Air Blank		0.000	05:25	
Diagnostics Check		OK	05:25	

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHR,K,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 05/29/2016 21:55:34

Diagnostics Check		OK	21:48
Air Blank		0.000	21:49
Control Test		0.081	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.095	21:50
Breath Volume		1.667	21:50
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.099	21:53
Breath Volume		1.605	21:53
Air Blank		0.000	21:54
Control Test		0.080	21:54
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 05/30/2016 21:11:55

Diagnostics Check		OK	21:04
Air Blank		0.000	21:05
Control Test		0.081	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.308	21:07
Breath Volume		2.539	21:07
Air Blank		0.000	21:07
Air Blank		0.000	21:09
Subject Sample	# 2	0.304	21:10
Breath Volume		1.992	21:10
Air Blank		0.000	21:11
Control Test		0.079	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:11

Data Download Date/Time: 06/03/2016 00:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: RICKETSON,J
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:35

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 05/16/2016 16:14:58

Diagnostics Check		OK	16:08
Air Blank		0.000	16:08
Control Test		0.080	16:09
Air Blank		0.000	16:09
Subject Sample	# 1	0.000	16:10
Breath Volume		5.171	16:10
Air Blank		0.000	16:11
Air Blank		0.000	16:13
Subject Sample	# 2	0.000	16:13
Breath Volume		3.285	16:13
Air Blank		0.000	16:14
Control Test		0.078	16:14
Air Blank		0.000	16:14
Diagnostics Check		OK	16:14

Data Download Date/Time: 05/27/2016 12:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WALSER

DATE/TIME OF TEST: 05/03/2016 23:39:37

Diagnostics Check	OK	23:38	
Air Blank		23:39	Ambient Fail
Air Blank	0.000	23:39	

Data Download Date/Time: 05/18/2016 20:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WALSER

DATE/TIME OF TEST: 05/03/2016 23:46:00

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.078	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 1	0.138	23:42	
Breath Volume		1.285	23:42	
Air Blank		0.000	23:43	
Air Blank			23:45	Ambient Fail
Air Blank			23:46	Ambient Fail

Data Download Date/Time: 05/18/2016 20:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WALSER

DATE/TIME OF TEST: 05/03/2016 23:48:07

Diagnostics Check	OK	23:47	
Air Blank		23:47	Ambient Fail
Air Blank	0.000	23:48	

Data Download Date/Time: 05/18/2016 20:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: WALSER

DATE/TIME OF TEST: 05/03/2016 23:55:58

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.077	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.137	23:51
Breath Volume		1.894	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:53
Subject Sample	# 2	0.132	23:54
Breath Volume		1.320	23:54
Air Blank		0.000	23:55
Control Test		0.076	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 05/18/2016 20:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 04/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ALCOCER ROMERO

DATE/TIME OF TEST: 05/06/2016 03:53:04

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.162	03:48
Breath Volume		1.878	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.159	03:51
Breath Volume		1.773	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:53

Data Download Date/Time: 05/18/2016 20:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MYRTIL,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HYPPOLITE

DATE/TIME OF TEST: 05/06/2016 01:22:58

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.081	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.171	01:18
Breath Volume		1.148	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.176	01:21
Breath Volume		1.457	01:21
Air Blank		0.000	01:22
Control Test		0.078	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 05/24/2016 09:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GUTIERREZ HENRIQUEZ

DATE/TIME OF TEST: 05/29/2016 00:33:49

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:31	
Control Test		0.080	00:31	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:32	Subject Test Refused
Breath Volume		0.000	00:32	
Air Blank		0.000	00:32	
Control Test		0.080	00:33	
Air Blank		0.000	00:33	
Diagnostics Check		OK	00:33	

Data Download Date/Time: 06/16/2016 09:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,FEDERICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 05/30/2016 03:26:37

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.079	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.179	03:22
Breath Volume		1.820	03:22
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.183	03:25
Breath Volume		1.437	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/16/2016 09:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CELLA,DOMENICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 05/15/2016 01:10:31

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.215	01:06
Breath Volume		2.585	01:06
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.211	01:08
Breath Volume		2.875	01:08
Air Blank		0.000	01:09
Control Test		0.079	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 06/22/2016 09:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHEAL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LLANES

DATE/TIME OF TEST: 05/15/2016 01:45:23

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:42	
Control Test		0.077	01:43	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:44	
Control Test		0.078	01:44	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 06/22/2016 09:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: GARCIA,LUIS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 05/22/2016 22:42:59

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.080	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.166	22:37
Breath Volume		1.480	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:39
Subject Sample	# 2	0.161	22:41
Breath Volume		1.421	22:41
Air Blank		0.000	22:42
Control Test		0.078	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 06/22/2016 09:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: DARLING,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 05/22/2016 23:40:34

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:37	
Control Test		0.079	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.079	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 06/22/2016 09:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 05/23/2016 03:42:29

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.079	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:41	Subject Test Refused
Breath Volume		0.000	03:41	
Air Blank		0.000	03:41	
Control Test		0.079	03:41	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 06/22/2016 09:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNSCHY,PATRICK,R
BREATH TEST OPERATOR AGENCY: Bay Harbor Island PD
ARRESTING OFFICER: MUNSCHY,PATRICK
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: YARBOROUGH

DATE/TIME OF TEST: 05/28/2016 07:19:16

Diagnostics Check		OK	07:16	
Air Blank		0.000	07:16	
Control Test		0.079	07:16	
Air Blank		0.000	07:17	
Subject Sample	# 1		07:17	Subject Test Refused
Breath Volume		0.000	07:17	
Air Blank		0.000	07:18	
Control Test		0.079	07:18	
Air Blank		0.000	07:19	
Diagnostics Check		OK	07:19	

Data Download Date/Time: 06/22/2016 09:06

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JACQUES

DATE/TIME OF TEST: 05/29/2016 04:29:33

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.155	04:25
Breath Volume		2.203	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.142	04:27
Breath Volume		1.695	04:27
Air Blank		0.000	04:28
Control Test		0.077	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 06/22/2016 09:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 05/06/2016 02:37:16

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.103	02:32
Breath Volume		3.171	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.099	02:35
Breath Volume		3.875	02:35
Air Blank		0.000	02:36
Control Test		0.078	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 05/06/2016 10:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ALVAREZ-RIVERO

DATE/TIME OF TEST: 05/06/2016 03:37:06

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.186	03:31
Breath Volume		1.460	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.176	03:35
Breath Volume		1.234	03:35
Air Blank		0.000	03:36
Control Test		0.077	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 05/06/2016 10:05

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HYPPOLITE

DATE/TIME OF TEST: 05/11/2016 04:17:02

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.078	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.256	04:12
Breath Volume		1.800	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.249	04:15
Breath Volume		1.605	04:15
Air Blank		0.000	04:16
Control Test		0.076	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:17

Data Download Date/Time: 06/03/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 05/15/2016 04:02:12

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.077	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.214	03:57	
Breath Volume		3.105	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2	0.212	04:00	
Breath Volume		3.183	04:00	
Air Blank		0.000	04:01	
Control Test		0.000	04:01	Control Outside Tolerance
Air Blank		0.000	04:02	

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: HIDALGO

DATE/TIME OF TEST: 05/15/2016 04:10:22

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.077	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.199	04:05
Breath Volume		2.460	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.199	04:08
Breath Volume		2.726	04:08
Air Blank		0.000	04:09
Control Test		0.075	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 05/18/2016 01:33:00

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.077	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.262	01:28
Breath Volume		2.597	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.256	01:31
Breath Volume		1.375	01:31
Air Blank		0.000	01:32
Control Test		0.075	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:33

Data Download Date/Time: 06/03/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: VILLA-SALAZAR

DATE/TIME OF TEST: 05/18/2016 01:46:47

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:44	
Control Test		0.076	01:44	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:45	Subject Test Refused
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Control Test		0.077	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/19/2016 04:41:33

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.077	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1	0.155	04:31	Volume Not Met
Breath Volume		0.296	04:31	
Air Blank		0.000	04:32	
Air Blank		0.000	04:33	
Subject Sample	# 2	0.183	04:35	
Breath Volume		1.226	04:35	
Air Blank		0.000	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 3	0.188	04:39	
Breath Volume		1.492	04:39	
Air Blank		0.000	04:40	
Control Test		0.076	04:40	
Air Blank		0.000	04:41	
Diagnostics Check		OK	04:41	

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CHARLEBOIS

DATE/TIME OF TEST: 05/20/2016 01:58:27

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.077	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.139	01:53
Breath Volume		1.269	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.143	01:56
Breath Volume		1.140	01:56
Air Blank		0.000	01:57
Control Test		0.076	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 06/03/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FIGUEROA-PERDIGON

DATE/TIME OF TEST: 05/25/2016 02:31:42

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.078	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.126	02:27
Breath Volume		2.328	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.126	02:30
Breath Volume		3.843	02:30
Air Blank		0.000	02:30
Control Test		0.076	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 05/26/2016 02:18:23

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.078	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.166	02:13
Breath Volume		2.449	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.157	02:16
Breath Volume		3.566	02:16
Air Blank		0.000	02:17
Control Test		0.077	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 06/03/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 05/26/2016 22:54:39

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.078	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.000	22:50
Breath Volume		3.835	22:50
Air Blank		0.000	22:50
Air Blank		0.000	22:52
Subject Sample	# 2	0.000	22:53
Breath Volume		3.058	22:53
Air Blank		0.000	22:53
Control Test		0.077	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 06/03/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LIMONGI

DATE/TIME OF TEST: 05/27/2016 02:07:17

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.077	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.144	02:02
Breath Volume		3.246	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.149	02:05
Breath Volume		2.882	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 06/03/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 05/29/2016 04:47:37

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.077	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.176	04:42
Breath Volume		1.695	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.173	04:45
Breath Volume		2.230	04:45
Air Blank		0.000	04:46
Control Test		0.077	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 06/03/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: MUGICA III

DATE/TIME OF TEST: 05/30/2016 00:01:20

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.078	23:56
Air Blank		0.000	23:56
Subject Sample	# 1	0.000	23:57
Breath Volume		1.109	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.000	23:59
Breath Volume		1.507	23:59
Air Blank		0.000	00:00
Control Test		0.076	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CELLA,DOMENICO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CELLA,DOMENICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ANDINO

DATE/TIME OF TEST: 05/30/2016 00:48:10

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.078	00:46	
Air Blank		0.000	00:46	
Subject Sample	# 1		00:46	Subject Test Refused
Breath Volume		0.000	00:46	
Air Blank		0.000	00:47	
Control Test		0.077	00:47	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CANALES

DATE/TIME OF TEST: 05/30/2016 03:37:40

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.078	03:35	
Air Blank		0.000	03:36	
Subject Sample	# 1		03:36	Subject Test Refused
Breath Volume		0.000	03:36	
Air Blank		0.000	03:36	
Control Test		0.077	03:37	
Air Blank		0.000	03:37	
Diagnostics Check		OK	03:37	

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 05/30/2016 05:33:32

Diagnostics Check		OK	05:26
Air Blank		0.000	05:27
Control Test		0.077	05:27
Air Blank		0.000	05:28
Subject Sample	# 1	0.150	05:28
Breath Volume		2.152	05:28
Air Blank		0.000	05:29
Air Blank		0.000	05:31
Subject Sample	# 2	0.149	05:31
Breath Volume		2.035	05:31
Air Blank		0.000	05:32
Control Test		0.076	05:32
Air Blank		0.000	05:33
Diagnostics Check		OK	05:33

Data Download Date/Time: 06/03/2016 10:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 05/04/2016 01:05:22

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.079	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.204	01:00
Breath Volume		3.062	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.192	01:03
Breath Volume		1.976	01:03
Air Blank		0.000	01:04
Control Test		0.078	01:04
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 05/05/2016 13:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MONTALVAN

DATE/TIME OF TEST: 05/04/2016 04:08:43

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.078	04:06	
Air Blank		0.000	04:06	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Control Test		0.079	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 05/05/2016 13:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CAICEDO

DATE/TIME OF TEST: 05/05/2016 04:28:08

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.080	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.175	04:23
Breath Volume		1.828	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.180	04:26
Breath Volume		2.738	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 05/05/2016 13:10

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: COTES

DATE/TIME OF TEST: 05/06/2016 02:13:29

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.172	02:08
Breath Volume		1.875	02:08
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.170	02:11
Breath Volume		1.191	02:11
Air Blank		0.000	02:12
Control Test		0.078	02:12
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MEDRANO ORTEZ

DATE/TIME OF TEST: 05/14/2016 05:18:37

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.081	05:04	
Air Blank		0.000	05:05	
Subject Sample	# 1	0.156	05:07	
Breath Volume		1.332	05:07	
Air Blank		0.000	05:08	
Air Blank		0.000	05:09	
Subject Sample	# 2	0.000	05:13	Volume Not Met
Breath Volume		0.031	05:13	
Air Blank		0.000	05:13	
Air Blank		0.000	05:15	
Subject Sample	# 3	0.139	05:17	
Breath Volume		1.628	05:17	
Air Blank		0.000	05:17	
Control Test		0.078	05:17	
Air Blank		0.000	05:18	
Diagnostics Check		OK	05:18	

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: RICO ARANGO

DATE/TIME OF TEST: 05/15/2016 02:18:41

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.078	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.077	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: REINA

DATE/TIME OF TEST: 05/18/2016 02:05:09

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.078	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.100	02:00
Breath Volume		1.296	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.101	02:03
Breath Volume		1.304	02:03
Air Blank		0.000	02:04
Control Test		0.077	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: O'BRIEN

DATE/TIME OF TEST: 05/20/2016 02:30:53

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:28	
Control Test		0.078	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.078	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: NABHEN

DATE/TIME OF TEST: 05/21/2016 00:57:27

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.078	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.118	00:53
Breath Volume		2.179	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.116	00:55
Breath Volume		1.468	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HERMANTIN

DATE/TIME OF TEST: 05/26/2016 01:32:37

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.221	01:27
Breath Volume		2.457	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.233	01:31
Breath Volume		2.773	01:31
Air Blank		0.000	01:31
Control Test		0.077	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SUAREZ GRUEIRO

DATE/TIME OF TEST: 05/27/2016 04:45:59

Diagnostics Check		OK	04:34	
Air Blank		0.000	04:34	
Control Test		0.078	04:35	
Air Blank		0.000	04:35	
Subject Sample	# 1	0.133	04:38	Volume Not Met
Breath Volume		0.816	04:38	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 2	0.115	04:44	Volume Not Met
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Control Test		0.077	04:45	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SUAREZ GRUEIRO

DATE/TIME OF TEST: 05/27/2016 04:57:34

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.077	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1	0.089	04:52	Volume Not Met
Breath Volume		0.023	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:55	
Subject Sample	# 2		04:56	Subject Test Refused
Breath Volume		0.167	04:56	
Air Blank		0.000	04:56	
Control Test		0.076	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SUAREZ GRUEIRO

DATE/TIME OF TEST: 05/27/2016 05:06:14

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:02	
Control Test		0.077	05:02	
Air Blank		0.000	05:02	
Subject Sample	# 1		05:04	Subject Test Refused
Breath Volume		0.058	05:04	
Air Blank		0.000	05:05	
Control Test		0.077	05:05	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: SADLIER

DATE/TIME OF TEST: 05/28/2016 02:39:18

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.078	02:37	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:38	
Control Test		0.078	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 06/02/2016 09:23

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 05/29/2016 04:40:58

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.078	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1	0.154	04:35	
Breath Volume		1.281	04:35	
Air Blank		0.000	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 2		04:39	Subject Test Refused
Breath Volume		0.000	04:39	
Air Blank		0.000	04:40	
Control Test		0.076	04:40	
Air Blank		0.000	04:40	
Diagnostics Check		OK	04:40	

Data Download Date/Time: 06/02/2016 09:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HAMIC

DATE/TIME OF TEST: 05/31/2016 04:12:59

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.194	04:08
Breath Volume		1.550	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.183	04:11
Breath Volume		2.058	04:11
Air Blank		0.000	04:12
Control Test		0.076	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 06/13/2016 10:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JIMMY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,GREG
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: REICHENBACH

DATE/TIME OF TEST: 05/06/2016 03:15:04

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.159	03:10
Breath Volume		4.644	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.150	03:13
Breath Volume		4.804	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:15

Data Download Date/Time: 05/26/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: WEAVER,R
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: ULMER

DATE/TIME OF TEST: 05/07/2016 00:06:11

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.141	00:01
Breath Volume		3.128	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.131	00:04
Breath Volume		1.394	00:04
Air Blank		0.000	00:05
Control Test		0.081	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

Data Download Date/Time: 05/26/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HECKARD JR

DATE/TIME OF TEST: 05/13/2016 23:55:06

Diagnostics Check		OK	23:48
Air Blank		0.000	23:49
Control Test		0.080	23:49
Air Blank		0.000	23:50
Subject Sample	# 1	0.235	23:50
Breath Volume		1.398	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:52
Subject Sample	# 2	0.251	23:53
Breath Volume		2.476	23:53
Air Blank		0.000	23:54
Control Test		0.080	23:54
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 05/26/2016 09:08

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 04/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:08

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 05/18/2016 18:35:31

Diagnostics Check		OK	18:28
Air Blank		0.000	18:29
Control Test		0.079	18:29
Air Blank		0.000	18:30
Subject Sample	# 1	0.171	18:30
Breath Volume		1.718	18:30
Air Blank		0.000	18:31
Air Blank		0.000	18:33
Subject Sample	# 2	0.177	18:33
Breath Volume		4.382	18:33
Air Blank		0.000	18:34
Control Test		0.078	18:34
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 05/26/2016 09:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: ORTEGA JR

DATE/TIME OF TEST: 05/27/2016 21:22:31

Diagnostics Check		OK	21:14
Air Blank		0.000	21:15
Control Test		0.080	21:15
Air Blank		0.000	21:16
Subject Sample	# 1	0.148	21:16
Breath Volume		1.242	21:16
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.142	21:20
Breath Volume		1.226	21:20
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 06/28/2016 08:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NAUMAN

DATE/TIME OF TEST: 05/30/2016 04:47:26

Diagnostics Check		OK	04:40
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.068	04:42
Breath Volume		1.460	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:44
Subject Sample	# 2	0.073	04:45
Breath Volume		1.777	04:45
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 06/28/2016 08:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 05/30/2016 18:49:35

Diagnostics Check		OK	18:32	
Air Blank		0.000	18:33	
Control Test		0.080	18:33	
Air Blank		0.000	18:34	
Subject Sample	# 1	0.168	18:37	Volume Not Met
Breath Volume		0.121	18:37	
Air Blank		0.000	18:38	
Air Blank		0.000	18:39	
Subject Sample	# 2	0.163	18:42	
Breath Volume		1.488	18:42	
Air Blank		0.000	18:42	
Air Blank		0.000	18:44	
Subject Sample	# 3	0.165	18:47	Slope Not Level
Breath Volume		1.359	18:47	
Air Blank		0.000	18:48	
Control Test		0.079	18:48	
Air Blank		0.000	18:49	
Diagnostics Check		OK	18:49	

Data Download Date/Time: 06/28/2016 08:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 05/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:12

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 05/30/2016 19:05:56

Diagnostics Check		OK	18:51	
Air Blank		0.000	18:52	
Control Test		0.079	18:52	
Air Blank		0.000	18:52	
Subject Sample	# 1	0.134	18:56	Volume Not Met
Breath Volume		0.531	18:56	
Air Blank		0.000	18:56	
Air Blank		0.000	18:58	
Subject Sample	# 2	0.170	18:59	
Breath Volume		1.703	18:59	
Air Blank		0.000	19:00	
Air Blank		0.000	19:02	
Subject Sample	# 3	0.177	19:04	
Breath Volume		1.394	19:04	
Air Blank		0.000	19:05	
Control Test		0.078	19:05	
Air Blank		0.000	19:05	
Diagnostics Check		OK	19:05	

Data Download Date/Time: 06/28/2016 08:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006026

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YODER,SCOTT,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YODER,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:46

SUBJECT'S LAST NAME: REYES REYES

DATE/TIME OF TEST: 05/12/2016 16:15:56

Diagnostics Check		OK	16:09
Air Blank		0.000	16:09
Control Test		0.080	16:10
Air Blank		0.000	16:10
Subject Sample	# 1	0.219	16:11
Breath Volume		3.695	16:11
Air Blank		0.000	16:12
Air Blank		0.000	16:13
Subject Sample	# 2	0.219	16:14
Breath Volume		2.902	16:14
Air Blank		0.000	16:15
Control Test		0.078	16:15
Air Blank		0.000	16:15
Diagnostics Check		OK	16:15

Data Download Date/Time: 05/20/2016 12:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006026

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON,THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FLORES ESCOBAR

DATE/TIME OF TEST: 05/14/2016 00:55:41

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.237	00:50
Breath Volume		1.445	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:52
Subject Sample	# 2	0.237	00:54
Breath Volume		2.640	00:54
Air Blank		0.000	00:54
Control Test		0.079	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 05/20/2016 12:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON,THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PEREA BUITRAGO

DATE/TIME OF TEST: 05/08/2016 01:56:45

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.082	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:55	Interferent Detect
Breath Volume		3.019	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 05/20/2016 13:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON,THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PEREA BUITRAGO

DATE/TIME OF TEST: 05/08/2016 02:02:23

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.082	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		02:00	Interferent Detect
Breath Volume		1.742	02:00	
Air Blank		0.000	02:01	
Control Test		0.082	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 05/20/2016 13:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006028

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON,THOMAS,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: STEELE

DATE/TIME OF TEST: 05/09/2016 00:35:20

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.082	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.162	00:29
Breath Volume		1.320	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:31
Subject Sample	# 2	0.167	00:33
Breath Volume		1.851	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 05/20/2016 13:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MUJICA

DATE/TIME OF TEST: 05/13/2016 22:24:07

Diagnostics Check		OK	22:16
Air Blank		0.000	22:17
Control Test		0.081	22:17
Air Blank		0.000	22:18
Subject Sample	# 1	0.179	22:19
Breath Volume		4.503	22:19
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 2	0.183	22:22
Breath Volume		4.832	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 05/16/2016 11:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 04/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LUCIANI

DATE/TIME OF TEST: 05/14/2016 01:45:38

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.081	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.173	01:39
Breath Volume		1.312	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:42
Subject Sample	# 2	0.184	01:43
Breath Volume		1.925	01:43
Air Blank		0.000	01:44
Control Test		0.080	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 05/16/2016 11:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 05/28/2016 03:24:30

Diagnostics Check		OK	03:17
Air Blank		0.000	03:17
Control Test		0.078	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.109	03:19
Breath Volume		2.304	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.110	03:22
Breath Volume		1.683	03:22
Air Blank		0.000	03:23
Control Test		0.076	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 06/17/2016 23:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.