Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,CHASE,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DELLAPORTA,A
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: O DARE

DATE/TIME OF TEST: 06/01/2016 00:08:12

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.264	00:03
Breath Volume		1.503	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.258	00:06
Breath Volume		1.488	00:06
Air Blank		0.000	00:07
Control Test		0.081	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 06/29/2016 14:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAGY,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SASTRE,STACCI
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 13:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: EARNHART

DATE/TIME OF TEST: 06/04/2016 17:08:29

Diagnostics Check		OK	17:01
Air Blank		0.000	17:01
Control Test		0.081	17:01
Air Blank		0.000	17:02
Subject Sample	# 1	0.202	17:03
Breath Volume		1.675	17:03
Air Blank		0.000	17:04
Air Blank		0.000	17:05
Subject Sample	# 2	0.204	17:06
Breath Volume		1.753	17:06
Air Blank		0.000	17:07
Control Test		0.080	17:07
Air Blank		0.000	17:08
Diagnostics Check		OK	17:08

Data Download Date/Time: 06/29/2016 14:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLY,EZEKIEL,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: DRAKE,T
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 15:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 06/06/2016 15:45:26

Diagnostics Check		OK	15:42	
Air Blank		0.000	15:42	
Control Test		0.081	15:42	
Air Blank		0.000	15:43	
Subject Sample	# 1		15:44	Subject Test Refused
Breath Volume		0.000	15:44	
Air Blank		0.000	15:44	
Control Test		0.080	15:44	
Air Blank		0.000	15:45	
Diagnostics Check		OK	15:45	

Data Download Date/Time: 06/29/2016 14:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,BRUCE,L
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: CLINE,BRUCE
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 06/07/2016 23:36:06

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.158	23:30
Breath Volume		1.437	23:30
Air Blank		0.000	23:31
Air Blank		0.000	23:32
Subject Sample	# 2	0.157	23:34
Breath Volume		1.347	23:34
Air Blank		0.000	23:35
Control Test		0.079	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 06/29/2016 14:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMONDS,FRANKIE,D
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: ABLE,PAUL
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: HUBBARD

DATE/TIME OF TEST: 06/13/2016 22:15:26

Diagnostics Check		OK	22:06
Air Blank		0.000	22:07
Control Test		0.080	22:07
Air Blank		0.000	22:08
Subject Sample	# 1	0.368	22:09
Breath Volume		2.097	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.365	22:13
Breath Volume		2.195	22:13
Air Blank		0.000	22:14
Control Test		0.080	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 06/29/2016 14:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAGY,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: COOPER,NICHOLAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 13:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: MECHUM

DATE/TIME OF TEST: 06/14/2016 14:26:18

Diagnostics Check		OK	14:23	
Air Blank		0.000	14:23	
Control Test		0.080	14:24	
Air Blank		0.000	14:24	
Subject Sample	# 1		14:24	Subject Test Refused
Breath Volume		0.000	14:24	
Air Blank		0.000	14:25	
Control Test		0.081	14:25	
Air Blank		0.000	14:26	
Diagnostics Check		OK	14:26	

Data Download Date/Time: 06/29/2016 14:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,CHASE,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KING,CHASE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 17:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 06/19/2016 18:09:51

Diagnostics Check		OK	18:01	
Air Blank		0.000	18:01	
Control Test		0.081	18:01	
Air Blank		0.000	18:02	
Subject Sample	# 1	0.108	18:05	Volume Not Met
Breath Volume		0.722	18:05	
Air Blank		0.000	18:06	
Air Blank		0.000	18:08	
Subject Sample	# 2		18:08	Subject Test Refused
Breath Volume		0.000	18:08	
Air Blank		0.000	18:08	
Control Test		0.080	18:09	
Air Blank		0.000	18:09	
Diagnostics Check		OK	18:09	

Data Download Date/Time: 06/29/2016 14:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,ERIN,M
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: CORALUZZO,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 06:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:52

SUBJECT'S LAST NAME: SINGLER

DATE/TIME OF TEST: 06/25/2016 07:18:12

Diagnostics Check		OK	07:15	
Air Blank		0.000	07:15	
Control Test		0.081	07:16	
Air Blank		0.000	07:16	
Subject Sample	# 1		07:16	Subject Test Refused
Breath Volume		0.000	07:16	
Air Blank		0.000	07:17	
Control Test		0.080	07:17	
Air Blank		0.000	07:18	
Diagnostics Check		OK	07:18	

Data Download Date/Time: 06/29/2016 14:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SPARKES

DATE/TIME OF TEST: 06/27/2016 23:33:33

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.324	23:29
Breath Volume		1.449	23:29
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.322	23:32
Breath Volume		1.605	23:32
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 06/29/2016 14:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ABEL,PAUL
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1626932
DRY GAS EXPIRY DATE: 01/31/2017
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: CRAFTON

DATE/TIME OF TEST: 06/28/2016 22:36:35

Diagnostics Check		OK	22:30
Air Blank		0.000	22:31
Control Test		0.081	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.016	22:32
Breath Volume		1.363	22:32
Air Blank		0.000	22:32
Air Blank		0.000	22:34
Subject Sample	# 2	0.015	22:35
Breath Volume		1.433	22:35
Air Blank		0.000	22:35
Control Test		0.081	22:36
Air Blank		0.000	22:36
Diagnostics Check		OK	22:36

Data Download Date/Time: 06/29/2016 14:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: MAHER,C
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 06/15/2016 05:27:28

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.080	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.168	05:22
Breath Volume		1.863	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.165	05:25
Breath Volume		2.003	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 06/24/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CONLEY,SHAWNA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MEDIAK

DATE/TIME OF TEST: 06/19/2016 03:38:27

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.200	03:33
Breath Volume		4.000	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.208	03:36
Breath Volume		2.921	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 06/24/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: KAHN,DAVID
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: KAYE

DATE/TIME OF TEST: 06/21/2016 23:51:45

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.214	23:47
Breath Volume		2.175	23:47
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.212	23:50
Breath Volume		2.300	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:51
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 06/24/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: ZIGLER,R
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: LI

DATE/TIME OF TEST: 06/24/2016 04:21:14

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.135	04:16
Breath Volume		1.937	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:18
Subject Sample	# 2	0.142	04:19
Breath Volume		1.468	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 06/24/2016 15:48

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,MATTHEW,C
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: RAPOANO,EDWARD
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HOGSETTE 3

DATE/TIME OF TEST: 06/24/2016 20:16:02

Diagnostics Check		OK	20:08
Air Blank		0.000	20:09
Control Test		0.080	20:09
Air Blank		0.000	20:09
Subject Sample	# 1	0.336	20:11
Breath Volume		2.074	20:11
Air Blank		0.000	20:12
Air Blank		0.000	20:13
Subject Sample	# 2	0.340	20:14
Breath Volume		1.457	20:14
Air Blank		0.000	20:15
Control Test		0.079	20:15
Air Blank		0.000	20:15
Diagnostics Check		OK	20:16

Data Download Date/Time: 07/07/2016 12:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CALKINS,BRIAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 06/25/2016 01:05:46

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:00	
Control Test		0.080	01:01	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:04	
Control Test		0.080	01:05	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 07/07/2016 12:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001135

INSTRUMENT REGISTERED TO: DAYTONA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KILKER,J,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: KILKER,J
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/25/2016 14:30:47

Diagnostics Check		OK	14:24
Air Blank		0.000	14:24
Control Test		0.080	14:25
Air Blank		0.000	14:25
Subject Sample	# 1	0.007	14:26
Breath Volume		2.050	14:26
Air Blank		0.000	14:26
Air Blank		0.000	14:28
Subject Sample	# 2	0.000	14:29
Breath Volume		1.617	14:29
Air Blank		0.000	14:29
Control Test		0.079	14:30
Air Blank		0.000	14:30
Diagnostics Check		OK	14:30

Data Download Date/Time: 07/07/2016 12:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILLANT,THOMAS,C
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: ZIGLER,RICKIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: HOLMQUIST

DATE/TIME OF TEST: 06/08/2016 21:57:09

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:54	
Control Test		0.081	21:55	
Air Blank		0.000	21:55	
Subject Sample	# 1		21:55	Subject Test Refused
Breath Volume		0.000	21:55	
Air Blank		0.000	21:56	
Control Test		0.081	21:56	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 06/29/2016 19:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTER,C,K
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: WALTER,C
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 05/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DREGGORS

DATE/TIME OF TEST: 06/22/2016 02:48:11

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.100	02:43
Breath Volume		1.671	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.105	02:46
Breath Volume		1.851	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 06/29/2016 19:12

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIEMER,AMANDA,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DUENAS,JONANTHON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:18

SUBJECT'S LAST NAME: BURKLEY

DATE/TIME OF TEST: 06/26/2016 11:51:09

Diagnostics Check		OK	11:43
Air Blank		0.000	11:44
Control Test		0.079	11:44
Air Blank		0.000	11:45
Subject Sample	# 1	0.180	11:46
Breath Volume		1.500	11:46
Air Blank		0.000	11:47
Air Blank		0.000	11:49
Subject Sample	# 2	0.186	11:49
Breath Volume		1.546	11:49
Air Blank		0.000	11:50
Control Test		0.079	11:50
Air Blank		0.000	11:51
Diagnostics Check		OK	11:51

Data Download Date/Time: 07/28/2016 17:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRANEY,JOSEPH,S
BREATH TEST OPERATOR AGENCY: Flagler Beach PD
ARRESTING OFFICER: WEST,JACOB
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 16014080A2
DRY GAS EXPIRY DATE: 07/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: O'MEALLY

DATE/TIME OF TEST: 06/30/2016 03:02:25

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.219	02:57
Breath Volume		3.593	02:57
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.214	03:00
Breath Volume		3.722	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/28/2016 17:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 06/24/2016 00:48:16

Diagnostics Check		OK	00:41
Air Blank		0.000	00:42
Control Test		0.078	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.150	00:43
Breath Volume		3.726	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.156	00:46
Breath Volume		3.730	00:46
Air Blank		0.000	00:47
Control Test		0.077	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 07/26/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 06/25/2016 00:19:07

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.078	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:17	Subject Test Refused
Breath Volume		0.000	00:17	
Air Blank		0.000	00:18	
Control Test		0.077	00:18	
Air Blank		0.000	00:19	
Diagnostics Check		OK	00:19	

Data Download Date/Time: 07/26/2016 11:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 06/30/2016 00:34:17

Diagnostics Check		OK	00:28
Air Blank		0.000	00:28
Control Test		0.077	00:28
Air Blank		0.000	00:29
Subject Sample	# 1	0.194	00:29
Breath Volume		2.335	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.182	00:32
Breath Volume		2.218	00:32
Air Blank		0.000	00:33
Control Test		0.077	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 07/26/2016 11:24

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: BENITEZ

DATE/TIME OF TEST: 06/04/2016 05:48:01

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:39	
Control Test		0.085	05:40	
Air Blank		0.000	05:40	
Subject Sample	# 1		05:43	No Sample Provided
Breath Volume		0.000	05:43	
Air Blank		0.000	05:44	
Air Blank		0.000	05:46	
Subject Sample	# 2		05:46	Subject Test Refused
Breath Volume		0.000	05:46	
Air Blank		0.000	05:47	
Control Test		0.085	05:47	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:48	

Data Download Date/Time: 06/27/2016 06:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 06/08/2016 00:56:32

Diagnostics Check		OK	00:53	
Air Blank		0.000	00:53	
Control Test		0.084	00:54	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:55	Subject Test Refused
Breath Volume		0.000	00:55	
Air Blank		0.000	00:55	
Control Test		0.083	00:55	
Air Blank		0.000	00:56	
Diagnostics Check		OK	00:56	

Data Download Date/Time: 06/27/2016 06:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: MOGENSEN

DATE/TIME OF TEST: 06/21/2016 04:34:11

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.084	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.180	04:29
Breath Volume		3.585	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.182	04:32
Breath Volume		3.343	04:32
Air Blank		0.000	04:33
Control Test		0.084	04:33
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 06/27/2016 06:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 06/23/2016 02:04:53

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.083	02:02	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:04	Sequence Aborted
Breath Volume		0.000	02:04	
Air Blank		0.000	02:04	

Data Download Date/Time: 06/27/2016 06:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GLOVER

DATE/TIME OF TEST: 06/23/2016 02:46:28

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.084	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.217	02:41
Breath Volume		3.980	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.218	02:44
Breath Volume		4.000	02:44
Air Blank		0.000	02:45
Control Test		0.083	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 06/27/2016 06:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: PENA-RIVERA

DATE/TIME OF TEST: 06/26/2016 00:36:44

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.084	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.183	00:28
Breath Volume		2.531	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.160	00:32
Breath Volume		2.136	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 3	0.169	00:35
Breath Volume		2.757	00:35
Air Blank		0.000	00:35
Control Test		0.083	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 06/27/2016 06:52

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ANDREW,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: DAVIS,ANDREW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MCFADDIN

DATE/TIME OF TEST: 06/01/2016 20:03:23

Diagnostics Check		OK	19:56
Air Blank		0.000	19:57
Control Test		0.081	19:57
Air Blank		0.000	19:58
Subject Sample	# 1	0.000	19:58
Breath Volume		1.722	19:58
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.000	20:02
Breath Volume		1.695	20:02
Air Blank		0.000	20:02
Control Test		0.081	20:02
Air Blank		0.000	20:03
Diagnostics Check		OK	20:03

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: SCHOBERT

DATE/TIME OF TEST: 06/03/2016 23:21:58

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.081	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.090	23:16
Breath Volume		1.425	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.084	23:20
Breath Volume		1.703	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TURNEY,JOEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: LAURITSEN-MCCAY

DATE/TIME OF TEST: 06/04/2016 01:43:39

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.080	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.171	01:39
Breath Volume		4.093	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.167	01:42
Breath Volume		3.441	01:42
Air Blank		0.000	01:42
Control Test		0.081	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HIRSCH,STEVEN
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/06/2016 00:02:01

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.081	23:53	
Air Blank		0.000	23:53	
Subject Sample	# 2	0.103	23:56	Volume Not Met
Breath Volume		0.105	23:56	
Air Blank		0.000	23:57	
Air Blank		0.000	23:59	
Subject Sample	# 1		00:00	Subject Test Refused
Breath Volume		0.046	00:00	
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKNIGHT,OMAR,J
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: IGO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/12/2016 22:28:09

Diagnostics Check		OK	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.214	22:23
Breath Volume		2.089	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.215	22:26
Breath Volume		1.792	22:26
Air Blank		0.000	22:27
Control Test		0.080	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECATES,C,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,K
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 06/18/2016 02:05:58

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.080	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.163	02:01
Breath Volume		2.886	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.157	02:04
Breath Volume		1.406	02:04
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRYAN,STEVEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: VIGNEAUX

DATE/TIME OF TEST: 06/18/2016 21:02:31

Diagnostics Check		OK	20:55
Air Blank		0.000	20:55
Control Test		0.081	20:56
Air Blank		0.000	20:56
Subject Sample	# 1	0.242	20:57
Breath Volume		2.832	20:57
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.238	21:00
Breath Volume		2.953	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 06/20/2016 19:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: DELAGARDELLE

DATE/TIME OF TEST: 06/19/2016 01:04:05

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.080	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.229	00:59
Breath Volume		3.609	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.221	01:02
Breath Volume		3.796	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRYAN,STEVEN,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TURNEY,JOEL
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 06/19/2016 01:21:59

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.154	01:17
Breath Volume		2.925	01:17
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.153	01:20
Breath Volume		3.285	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CALKINS,BRIAN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 05/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 06/19/2016 19:51:10

Diagnostics Check		OK	19:44
Air Blank		0.000	19:45
Control Test		0.080	19:45
Air Blank		0.000	19:45
Subject Sample	# 1	0.230	19:46
Breath Volume		2.414	19:46
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.227	19:49
Breath Volume		2.101	19:49
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 06/20/2016 19:38

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECATES,C,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: TATUM,T
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WARNER

DATE/TIME OF TEST: 06/25/2016 01:37:25

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.080	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.109	01:32
Breath Volume		1.359	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.106	01:35
Breath Volume		1.628	01:35
Air Blank		0.000	01:36
Control Test		0.080	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 07/11/2016 07:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMAN,JOHN,S
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRAMAN,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ACOPINE

DATE/TIME OF TEST: 06/26/2016 22:55:16

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.010	22:50
Breath Volume		1.695	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.010	22:53
Breath Volume		2.070	22:53
Air Blank		0.000	22:54
Control Test		0.082	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 07/11/2016 07:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMAN,JOHN,S
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BRAMAN,JOHN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: GARZIA

DATE/TIME OF TEST: 06/30/2016 02:48:17

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.080	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.230	02:43
Breath Volume		2.101	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.232	02:46
Breath Volume		1.945	02:46
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 07/11/2016 07:28

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RENELL,BENITA,L
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MACKER

DATE/TIME OF TEST: 06/03/2016 03:02:08

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:58	
Control Test		0.080	02:59	
Air Blank		0.000	02:59	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:01	
Control Test		0.081	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 06/25/2016 07:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MICHAELIS,AMBERLY
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/05/2016 19:37:42

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:35	
Control Test		0.080	19:35	
Air Blank		0.000	19:36	
Subject Sample	# 1		19:36	Subject Test Refused
Breath Volume		0.000	19:36	
Air Blank		0.000	19:36	
Control Test		0.080	19:37	
Air Blank		0.000	19:37	
Diagnostics Check		OK	19:37	

Data Download Date/Time: 06/25/2016 07:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 06/11/2016 02:38:18

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.234	02:33
Breath Volume		4.003	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.223	02:36
Breath Volume		2.648	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 06/25/2016 07:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: SECHRIST

DATE/TIME OF TEST: 06/17/2016 23:59:58

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.080	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.260	23:55
Breath Volume		1.859	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.250	23:58
Breath Volume		2.066	23:58
Air Blank		0.000	23:59
Control Test		0.079	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 06/25/2016 07:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: PICCOLA,DANIEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 06/19/2016 00:18:46

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:13
Air Blank		0.000	00:13
Subject Sample	# 1	0.301	00:14
Breath Volume		3.316	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.308	00:17
Breath Volume		3.246	00:17
Air Blank		0.000	00:17
Control Test		0.081	00:18
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 06/25/2016 07:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARMINGTON,WILLIAM,G
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MEDINA,RAFAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: DUFRESNE

DATE/TIME OF TEST: 06/19/2016 00:45:03

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.080	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:44	
Control Test		0.080	00:44	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:45	

Data Download Date/Time: 06/25/2016 07:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOSS,JEREMY,A
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: HYATT,JUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: DOOBE

DATE/TIME OF TEST: 06/20/2016 00:16:42

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.144	00:12
Breath Volume		1.980	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.144	00:15
Breath Volume		1.171	00:15
Air Blank		0.000	00:15
Control Test		0.080	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 06/25/2016 07:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER,MICHAEL,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER,MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: VICK

DATE/TIME OF TEST: 06/22/2016 02:53:20

Diagnostics Check		OK	02:38	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:42	No Sample Provided
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 2	0.214	02:46	
Breath Volume		2.242	02:46	
Air Blank		0.000	02:47	
Air Blank		0.000	02:49	
Subject Sample	# 3	0.215	02:51	
Breath Volume		1.492	02:51	
Air Blank		0.000	02:52	
Control Test		0.080	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 06/25/2016 07:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001159

INSTRUMENT REGISTERED TO: ORANGE CITY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRALSON,SAMUEL,E
BREATH TEST OPERATOR AGENCY: Orange City PD
ARRESTING OFFICER: NESBITT,STEPHEN
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: VAUGHN

DATE/TIME OF TEST: 06/30/2016 21:24:31

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.078	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.000	21:20
Breath Volume		1.449	21:20
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.000	21:23
Breath Volume		1.425	21:23
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 07/18/2016 09:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BISSOON,STEPHEN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LUNDE

DATE/TIME OF TEST: 06/04/2016 04:27:12

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.347	04:20
Breath Volume		2.023	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.356	04:25
Breath Volume		2.695	04:25
Air Blank		0.000	04:26
Control Test		0.079	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 06/09/2016 09:47

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCK,JOHN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: STOELKER

DATE/TIME OF TEST: 06/07/2016 10:56:45

Diagnostics Check		OK	10:49
Air Blank		0.000	10:50
Control Test		0.079	10:50
Air Blank		0.000	10:51
Subject Sample	# 1	0.000	10:51
Breath Volume		1.660	10:51
Air Blank		0.000	10:52
Air Blank		0.000	10:54
Subject Sample	# 2	0.000	10:55
Breath Volume		1.367	10:55
Air Blank		0.000	10:55
Control Test		0.080	10:56
Air Blank		0.000	10:56
Diagnostics Check		OK	10:56

Data Download Date/Time: 06/09/2016 09:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JESIONEK,ROBERT,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEVE
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 05/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: HABERL

DATE/TIME OF TEST: 06/09/2016 02:29:27

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:27	
Control Test		0.079	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.078	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 06/09/2016 09:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,BRITTINI,L
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/20/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 06/21/2016 23:34:06

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:31	
Control Test		0.079	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.079	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 07/26/2016 13:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALVAREZ,AMANDA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: D ADAMO

DATE/TIME OF TEST: 06/25/2016 22:29:34

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:27	
Control Test		0.079	22:27	
Air Blank		0.000	22:28	
Subject Sample	# 1		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Control Test		0.079	22:28	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 07/26/2016 13:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001183

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARNEY,DANIEL,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 06/28/2016 01:47:13

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.142	01:42
Breath Volume		3.851	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.145	01:45
Breath Volume		2.000	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 07/26/2016 13:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JANKE

DATE/TIME OF TEST: 06/01/2016 03:19:28

Diagnostics Check		OK	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.225	03:14
Breath Volume		2.496	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.225	03:17
Breath Volume		2.226	03:17
Air Blank		0.000	03:18
Control Test		0.079	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 06/02/2016 00:28:48

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.000	00:24
Breath Volume		2.781	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.000	00:27
Breath Volume		1.914	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEEDHAM,K,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEEDHAM,K
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:54

SUBJECT'S LAST NAME: ZUCHERO

DATE/TIME OF TEST: 06/02/2016 16:29:19

Diagnostics Check		OK	16:23
Air Blank		0.000	16:23
Control Test		0.080	16:23
Air Blank		0.000	16:24
Subject Sample	# 1	0.191	16:24
Breath Volume		2.101	16:24
Air Blank		0.000	16:25
Air Blank		0.000	16:27
Subject Sample	# 2	0.197	16:27
Breath Volume		2.113	16:27
Air Blank		0.000	16:28
Control Test		0.080	16:28
Air Blank		0.000	16:29
Diagnostics Check		OK	16:29

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CHRISTIE

DATE/TIME OF TEST: 06/03/2016 00:48:46

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.081	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.092	00:44
Breath Volume		1.683	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.093	00:47
Breath Volume		1.695	00:47
Air Blank		0.000	00:47
Control Test		0.080	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 06/03/2016 04:11:41

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.171	04:07
Breath Volume		2.015	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.170	04:10
Breath Volume		1.628	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/03/2016 07:38:53

Diagnostics Check		OK	07:30	
Air Blank		0.000	07:31	
Control Test		0.081	07:31	
Air Blank		0.000	07:32	
Subject Sample	# 1	0.174	07:33	
Breath Volume		1.417	07:33	
Air Blank		0.000	07:33	
Air Blank		0.000	07:35	
Subject Sample	# 2		07:37	Subject Test Refused
Breath Volume		0.000	07:37	
Air Blank		0.000	07:37	
Control Test		0.080	07:38	
Air Blank		0.000	07:38	
Diagnostics Check		OK	07:38	

Data Download Date/Time: 06/28/2016 00:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:36

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/03/2016 08:03:56

Diagnostics Check		OK	07:56
Air Blank		0.000	07:56
Control Test		0.080	07:57
Air Blank		0.000	07:57
Subject Sample	# 1	0.163	07:58
Breath Volume		1.300	07:58
Air Blank		0.000	07:58
Air Blank		0.000	08:00
Subject Sample	# 2	0.155	08:02
Breath Volume		1.378	08:02
Air Blank		0.000	08:03
Control Test		0.080	08:03
Air Blank		0.000	08:03
Diagnostics Check		OK	08:03

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: GOMES

DATE/TIME OF TEST: 06/04/2016 00:04:47

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:00
Breath Volume		2.136	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.000	00:03
Breath Volume		1.671	00:03
Air Blank		0.000	00:03
Control Test		0.080	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: DONATO

DATE/TIME OF TEST: 06/04/2016 00:57:23

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.206	00:52
Breath Volume		2.148	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.207	00:55
Breath Volume		2.007	00:55
Air Blank		0.000	00:56
Control Test		0.079	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 06/28/2016 00:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: RESSLER

DATE/TIME OF TEST: 06/05/2016 02:58:46

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Subject Sample	# 1	0.219	02:53
Breath Volume		1.128	02:53
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.202	02:57
Breath Volume		1.480	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANARO,RONNIE,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROOT,KATHY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:20

SUBJECT'S LAST NAME: THORSEN

DATE/TIME OF TEST: 06/05/2016 13:52:35

Diagnostics Check		OK	13:46
Air Blank		0.000	13:46
Control Test		0.079	13:46
Air Blank		0.000	13:47
Subject Sample	# 1	0.312	13:48
Breath Volume		2.335	13:48
Air Blank		0.000	13:48
Air Blank		0.000	13:50
Subject Sample	# 2	0.310	13:51
Breath Volume		2.296	13:51
Air Blank		0.000	13:51
Control Test		0.079	13:51
Air Blank		0.000	13:52
Diagnostics Check		OK	13:52

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: HAGGE

DATE/TIME OF TEST: 06/05/2016 23:48:07

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Subject Sample	# 1	0.168	23:43
Breath Volume		1.601	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.176	23:46
Breath Volume		1.441	23:46
Air Blank		0.000	23:47
Control Test		0.079	23:47
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: FASSBERG

DATE/TIME OF TEST: 06/06/2016 20:14:31

Diagnostics Check		OK	19:59	
Air Blank		0.000	20:00	
Control Test		0.080	20:00	
Air Blank		0.000	20:01	
Subject Sample	# 1	0.119	20:03	
Breath Volume		1.214	20:03	
Air Blank		0.000	20:03	
Air Blank		0.000	20:05	
Subject Sample	# 2	0.100	20:08	Volume Not Met
Breath Volume		0.089	20:08	
Air Blank		0.000	20:09	
Air Blank		0.000	20:11	
Subject Sample	# 3	0.115	20:13	
Breath Volume		1.234	20:13	
Air Blank		0.000	20:13	
Control Test		0.080	20:13	
Air Blank		0.000	20:14	
Diagnostics Check		OK	20:14	

Data Download Date/Time: 06/28/2016 00:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: MCKEON

DATE/TIME OF TEST: 06/08/2016 19:33:02

Diagnostics Check		OK	19:26
Air Blank		0.000	19:27
Control Test		0.079	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.081	19:28
Breath Volume		2.851	19:28
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.078	19:31
Breath Volume		2.847	19:31
Air Blank		0.000	19:32
Control Test		0.079	19:32
Air Blank		0.000	19:32
Diagnostics Check		OK	19:33

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MC MENAMAN

DATE/TIME OF TEST: 06/08/2016 22:27:44

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.079	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.142	22:23
Breath Volume		1.488	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.130	22:26
Breath Volume		1.222	22:26
Air Blank		0.000	22:26
Control Test		0.078	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 06/09/2016 18:36:27

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.079	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.098	18:31
Breath Volume		3.222	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.093	18:34
Breath Volume		2.625	18:34
Air Blank		0.000	18:35
Control Test		0.079	18:35
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 06/28/2016 00:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KIERNAN

DATE/TIME OF TEST: 06/10/2016 00:38:38

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.196	00:34
Breath Volume		1.292	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.208	00:37
Breath Volume		1.570	00:37
Air Blank		0.000	00:37
Control Test		0.079	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECI,ZACHARY,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,ZACH
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HELLMANN

DATE/TIME OF TEST: 06/10/2016 03:17:41

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.192	03:13
Breath Volume		1.703	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.181	03:16
Breath Volume		2.847	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:17
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,ANGERVIL,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,ANGERVIL
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: BYNUM

DATE/TIME OF TEST: 06/10/2016 23:57:09

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.079	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1	0.234	23:43	
Breath Volume		2.750	23:43	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 2	0.179	23:49	Volume Not Met
Breath Volume		0.144	23:49	
Air Blank		0.000	23:50	
Air Blank		0.000	23:52	
Subject Sample	# 3	0.212	23:55	Volume Not Met
Breath Volume		0.000	23:55	
Air Blank		0.000	23:55	
Control Test		0.077	23:56	
Air Blank			23:56	Improper Sample
Air Blank		0.000	23:57	

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: MILEY

DATE/TIME OF TEST: 06/11/2016 00:11:15

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.078	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.117	00:06
Breath Volume		3.230	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.110	00:09
Breath Volume		3.074	00:09
Air Blank		0.000	00:10
Control Test		0.078	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: PASCUAL-JIMENEZ

DATE/TIME OF TEST: 06/11/2016 18:35:57

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.080	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.185	18:31
Breath Volume		2.800	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.186	18:34
Breath Volume		2.371	18:34
Air Blank		0.000	18:35
Control Test		0.079	18:35
Air Blank		0.000	18:35
Diagnostics Check		OK	18:35

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: UBILLA

DATE/TIME OF TEST: 06/11/2016 18:48:11

Diagnostics Check		OK	18:41
Air Blank		0.000	18:42
Control Test		0.079	18:42
Air Blank		0.000	18:42
Subject Sample	# 1	0.104	18:43
Breath Volume		1.167	18:43
Air Blank		0.000	18:44
Air Blank		0.000	18:46
Subject Sample	# 2	0.094	18:46
Breath Volume		1.781	18:46
Air Blank		0.000	18:47
Control Test		0.079	18:47
Air Blank		0.000	18:48
Diagnostics Check		OK	18:48

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: PASQUALINO

DATE/TIME OF TEST: 06/11/2016 21:35:32

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.080	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.139	21:31
Breath Volume		2.863	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.137	21:34
Breath Volume		2.039	21:34
Air Blank		0.000	21:34
Control Test		0.080	21:34
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 06/28/2016 00:44

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,MARK,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: STEELE,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: ARIAS MANCHENO

DATE/TIME OF TEST: 06/12/2016 02:48:06

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.129	02:42
Breath Volume		2.128	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.148	02:46
Breath Volume		1.812	02:46
Air Blank		0.000	02:47
Control Test		0.078	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BRAVO VALDEZ

DATE/TIME OF TEST: 06/12/2016 04:30:27

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.079	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.178	04:25
Breath Volume		1.320	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.168	04:28
Breath Volume		1.210	04:28
Air Blank		0.000	04:29
Control Test		0.078	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 06/28/2016 00:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEELE,MARK,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: STEELE,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/12/2016 05:04:40

Diagnostics Check		OK	04:58
Air Blank		0.000	04:58
Control Test		0.080	04:59
Air Blank		0.000	04:59
Subject Sample	# 1	0.237	05:00
Breath Volume		3.636	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.235	05:03
Breath Volume		3.269	05:03
Air Blank		0.000	05:03
Control Test		0.080	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 06/28/2016 00:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: LOPEZ-LOPEZ

DATE/TIME OF TEST: 06/12/2016 21:54:29

Diagnostics Check		OK	21:48
Air Blank		0.000	21:49
Control Test		0.080	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.194	21:50
Breath Volume		1.667	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.187	21:52
Breath Volume		1.429	21:52
Air Blank		0.000	21:53
Control Test		0.079	21:53
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,ZACH,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,ZACH
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: SATRIANI

DATE/TIME OF TEST: 06/13/2016 22:12:48

Diagnostics Check		OK	22:06
Air Blank		0.000	22:07
Control Test		0.079	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.204	22:08
Breath Volume		1.835	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:10
Subject Sample	# 2	0.202	22:11
Breath Volume		1.128	22:11
Air Blank		0.000	22:11
Control Test		0.078	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: KEMPINSKI

DATE/TIME OF TEST: 06/14/2016 10:11:09

Diagnostics Check		OK	10:05
Air Blank		0.000	10:05
Control Test		0.079	10:06
Air Blank		0.000	10:06
Subject Sample	# 1	0.000	10:06
Breath Volume		3.234	10:06
Air Blank		0.000	10:07
Air Blank		0.000	10:09
Subject Sample	# 2	0.000	10:09
Breath Volume		2.382	10:09
Air Blank		0.000	10:10
Control Test		0.079	10:10
Air Blank		0.000	10:11
Diagnostics Check		OK	10:11

Data Download Date/Time: 06/28/2016 00:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,RONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: WEBER,RONALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: COLLIER

DATE/TIME OF TEST: 06/16/2016 16:07:22

Diagnostics Check		OK	16:00
Air Blank		0.000	16:01
Control Test		0.079	16:01
Air Blank		0.000	16:02
Subject Sample	# 1	0.193	16:02
Breath Volume		4.109	16:02
Air Blank		0.000	16:03
Air Blank		0.000	16:05
Subject Sample	# 2	0.203	16:05
Breath Volume		3.773	16:05
Air Blank		0.000	16:06
Control Test		0.079	16:06
Air Blank		0.000	16:07
Diagnostics Check		OK	16:07

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TANSKY

DATE/TIME OF TEST: 06/16/2016 19:27:59

Diagnostics Check		OK	19:21
Air Blank		0.000	19:21
Control Test		0.080	19:22
Air Blank		0.000	19:22
Subject Sample	# 1	0.104	19:23
Breath Volume		1.371	19:23
Air Blank		0.000	19:24
Air Blank		0.000	19:25
Subject Sample	# 2	0.105	19:26
Breath Volume		1.285	19:26
Air Blank		0.000	19:27
Control Test		0.079	19:27
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 06/17/2016 03:36:23

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.167	03:32
Breath Volume		3.496	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.173	03:34
Breath Volume		3.179	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDIGER,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: RUEDIGER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: GEEVER

DATE/TIME OF TEST: 06/18/2016 23:45:04

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.080	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:36	No Sample Provided
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2	0.193	23:40	
Breath Volume		2.632	23:40	
Air Blank		0.000	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 3	0.203	23:43	
Breath Volume		3.285	23:43	
Air Blank		0.000	23:44	
Control Test		0.079	23:44	
Air Blank		0.000	23:44	
Diagnostics Check		OK	23:45	

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 06/19/2016 01:02:03

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.142	00:57
Breath Volume		1.816	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.129	01:00
Breath Volume		1.375	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 06/28/2016 00:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BOUCHARD

DATE/TIME OF TEST: 06/22/2016 03:13:18

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.102	03:08
Breath Volume		2.855	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.097	03:11
Breath Volume		2.386	03:11
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 06/22/2016 21:27:08

Diagnostics Check		OK	21:20
Air Blank		0.000	21:21
Control Test		0.080	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.372	21:22
Breath Volume		3.523	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:25
Subject Sample	# 2	0.380	21:25
Breath Volume		3.402	21:25
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HAFNER

DATE/TIME OF TEST: 06/23/2016 00:05:58

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:01
Breath Volume		1.910	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.000	00:04
Breath Volume		1.519	00:04
Air Blank		0.000	00:05
Control Test		0.081	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEE,COREY,C
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: LEE,COREY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GLADWIN

DATE/TIME OF TEST: 06/23/2016 02:39:13

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.181	02:34
Breath Volume		1.707	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.182	02:37
Breath Volume		3.812	02:37
Air Blank		0.000	02:38
Control Test		0.081	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,ZACH,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,ZACH
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROSARIO JR

DATE/TIME OF TEST: 06/23/2016 03:53:07

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.081	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.195	03:47
Breath Volume		1.601	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.182	03:51
Breath Volume		2.214	03:51
Air Blank		0.000	03:52
Control Test		0.081	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 06/28/2016 00:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: DICKINSON

DATE/TIME OF TEST: 06/23/2016 23:30:41

Diagnostics Check		OK	23:24
Air Blank		0.000	23:24
Control Test		0.080	23:25
Air Blank		0.000	23:25
Subject Sample	# 1	0.237	23:26
Breath Volume		1.488	23:26
Air Blank		0.000	23:26
Air Blank		0.000	23:28
Subject Sample	# 2	0.247	23:29
Breath Volume		1.914	23:29
Air Blank		0.000	23:29
Control Test		0.080	23:30
Air Blank		0.000	23:30
Diagnostics Check		OK	23:30

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANTON,JON,E
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HANTON,JON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: SILVESTRE-MATEO

DATE/TIME OF TEST: 06/24/2016 21:51:32

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.080	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.136	21:47
Breath Volume		3.269	21:47
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.121	21:50
Breath Volume		2.488	21:50
Air Blank		0.000	21:50
Control Test		0.080	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/25/2016 01:47:12

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.131	01:42
Breath Volume		2.507	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.126	01:45
Breath Volume		2.277	01:45
Air Blank		0.000	01:46
Control Test		0.081	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: WORTHINGTON

DATE/TIME OF TEST: 06/26/2016 01:36:16

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.081	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.000	01:31
Breath Volume		1.316	01:31
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.000	01:34
Breath Volume		1.468	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SHELL

DATE/TIME OF TEST: 06/26/2016 02:21:59

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.081	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.166	02:17
Breath Volume		1.785	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.167	02:20
Breath Volume		1.984	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANARO,RONNIE,M
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROOT,KATHY
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: ROCHELEAU

DATE/TIME OF TEST: 06/26/2016 17:40:30

Diagnostics Check		OK	17:34
Air Blank		0.000	17:34
Control Test		0.081	17:34
Air Blank		0.000	17:35
Subject Sample	# 1	0.260	17:35
Breath Volume		2.250	17:35
Air Blank		0.000	17:36
Air Blank		0.000	17:38
Subject Sample	# 2	0.255	17:38
Breath Volume		2.320	17:38
Air Blank		0.000	17:39
Control Test		0.080	17:39
Air Blank		0.000	17:40
Diagnostics Check		OK	17:40

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: BOULET

DATE/TIME OF TEST: 06/26/2016 23:48:02

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.080	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.242	23:40
Breath Volume		2.390	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.265	23:43
Breath Volume		2.566	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:45
Subject Sample	# 3	0.254	23:46
Breath Volume		2.769	23:46
Air Blank		0.000	23:47
Control Test		0.080	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:48

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001189

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 06/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 06/27/2016 02:32:44

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.194	02:28
Breath Volume		2.136	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.196	02:31
Breath Volume		2.113	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 06/28/2016 00:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: STROUDE

DATE/TIME OF TEST: 06/22/2016 20:01:03

Diagnostics Check		OK	19:55	
Air Blank		0.000	19:55	
Control Test		0.079	19:56	
Air Blank		0.000	19:56	
Subject Sample	# 1		19:59	Slope Not Met
Breath Volume		1.421	19:59	
Air Blank		0.000	20:00	
Control Test		0.077	20:00	
Air Blank		0.000	20:00	
Diagnostics Check		OK	20:01	

Data Download Date/Time: 07/21/2016 09:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 19:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: STROUDE

DATE/TIME OF TEST: 06/22/2016 20:36:42

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:28	
Control Test		0.078	20:29	
Air Blank		0.000	20:29	
Subject Sample	# 1	0.178	20:32	Volume Not Met
Breath Volume		0.121	20:32	
Air Blank		0.000	20:33	
Air Blank		0.000	20:35	
Subject Sample	# 2		20:35	Subject Test Refused
Breath Volume		0.000	20:35	
Air Blank		0.000	20:35	
Control Test		0.077	20:36	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:36	

Data Download Date/Time: 07/21/2016 09:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIKEMAN,HOUSTON,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MULHOLLAND,SALLY
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: OXLAJITZEP

DATE/TIME OF TEST: 06/02/2016 00:04:40

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.077	23:59
Air Blank		0.000	23:59
Subject Sample	# 1	0.202	00:00
Breath Volume		3.195	00:00
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 2	0.211	00:03
Breath Volume		2.187	00:03
Air Blank		0.000	00:03
Control Test		0.077	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 06/06/2016 15:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY,M,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WEEKS,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/05/2016 19:57:43

Diagnostics Check		OK	19:50	
Air Blank		0.000	19:51	
Control Test		0.077	19:51	
Air Blank		0.000	19:52	
Subject Sample	# 1	0.241	19:54	
Breath Volume		1.742	19:54	
Air Blank		0.000	19:54	
Air Blank		0.000	19:56	
Subject Sample	# 2		19:57	Sequence Aborted
Breath Volume		0.000	19:57	
Air Blank		0.000	19:57	

Data Download Date/Time: 06/06/2016 15:34

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,MITCHELL,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BERNAGENE,FRITZLER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: LOWREY

DATE/TIME OF TEST: 06/26/2016 03:38:47

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.077	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.191	03:34
Breath Volume		3.175	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.192	03:37
Breath Volume		4.605	03:37
Air Blank		0.000	03:37
Control Test		0.077	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 07/12/2016 09:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABRAHAM,ERIC,D
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ABRAHAM,ERIC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 11:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:49

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 06/09/2016 12:15:01

Diagnostics Check		OK	12:11	
Air Blank		0.000	12:11	
Control Test		0.079	12:12	
Air Blank		0.000	12:12	
Subject Sample	# 1		12:13	Subject Test Refused
Breath Volume		0.000	12:13	
Air Blank		0.000	12:14	
Control Test		0.078	12:14	
Air Blank		0.000	12:14	
Diagnostics Check		OK	12:15	

Data Download Date/Time: 06/21/2016 13:07

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUBER,KEVIN,R
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: QUINN,C
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 05/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: BACHUS

DATE/TIME OF TEST: 06/14/2016 23:36:12

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.078	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.144	23:31
Breath Volume		2.285	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:33
Subject Sample	# 2	0.143	23:34
Breath Volume		2.757	23:34
Air Blank		0.000	23:35
Control Test		0.078	23:35
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 06/21/2016 13:07

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAND,CURTIS,L
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: ALDERMAN,JOSHEPH
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MC GILL

DATE/TIME OF TEST: 06/21/2016 18:36:35

Diagnostics Check		OK	18:29
Air Blank		0.000	18:30
Control Test		0.079	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.190	18:31
Breath Volume		3.152	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.194	18:35
Breath Volume		3.816	18:35
Air Blank		0.000	18:35
Control Test		0.079	18:35
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 07/24/2016 07:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,JAMES,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BRANCH,JOSHUA
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JUDAH

DATE/TIME OF TEST: 06/23/2016 22:23:36

Diagnostics Check		OK	22:20	
Air Blank		0.000	22:21	
Control Test		0.080	22:21	
Air Blank		0.000	22:21	
Subject Sample	# 1		22:22	Subject Test Refused
Breath Volume		0.000	22:22	
Air Blank		0.000	22:22	
Control Test		0.079	22:23	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 07/24/2016 07:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001204

INSTRUMENT REGISTERED TO: PANAMA CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOHENY,KEVIN,M
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: DOHENY,KEVIN
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: FALLS

DATE/TIME OF TEST: 06/26/2016 06:05:20

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.080	05:59
Air Blank		0.000	06:00
Subject Sample	# 1	0.165	06:00
Breath Volume		2.667	06:00
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.165	06:03
Breath Volume		1.921	06:03
Air Blank		0.000	06:04
Control Test		0.079	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 07/24/2016 07:45

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SOMMERS,EDWARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: REMO

DATE/TIME OF TEST: 06/17/2016 22:21:56

Diagnostics Check		OK	22:15
Air Blank		0.000	22:16
Control Test		0.077	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.000	22:17
Breath Volume		1.660	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.000	22:20
Breath Volume		1.390	22:20
Air Blank		0.000	22:21
Control Test		0.076	22:21
Air Blank		0.000	22:21
Diagnostics Check		OK	22:21

Data Download Date/Time: 06/23/2016 10:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,NATHAN,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 06/04/2016 05:43:25

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:37	
Control Test		0.080	05:37	
Air Blank		0.000	05:38	
Subject Sample	# 1	0.128	05:39	
Breath Volume		1.554	05:39	
Air Blank		0.000	05:40	
Air Blank		0.000	05:42	
Subject Sample	# 2	0.129	05:42	
Breath Volume		1.710	05:42	
Air Blank			05:42	RFI Detect
Air Blank		0.000	05:43	

Data Download Date/Time: 06/10/2016 10:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,NATHAN,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 06/04/2016 05:53:00

Diagnostics Check		OK	05:46
Air Blank		0.000	05:47
Control Test		0.079	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.127	05:48
Breath Volume		1.777	05:48
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.125	05:51
Breath Volume		2.199	05:51
Air Blank		0.000	05:52
Control Test		0.079	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:53

Data Download Date/Time: 06/10/2016 10:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MCCLARD

DATE/TIME OF TEST: 06/10/2016 01:07:02

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:06	
Control Test	0.000	01:06	Control Outside Tolerance
Air Blank	0.000	01:07	

Data Download Date/Time: 06/10/2016 10:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MCCLARD

DATE/TIME OF TEST: 06/10/2016 01:11:33

Diagnostics Check	OK	01:10	
Air Blank	0.000	01:10	
Control Test	0.000	01:11	Control Outside Tolerance
Air Blank	0.000	01:11	

Data Download Date/Time: 06/10/2016 10:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MCCLARD

DATE/TIME OF TEST: 06/10/2016 01:21:56

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.084	01:16
Breath Volume		1.738	01:16
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.088	01:20
Breath Volume		2.078	01:20
Air Blank		0.000	01:21
Control Test		0.079	01:21
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 06/10/2016 10:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 05/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: ALEMAN

DATE/TIME OF TEST: 06/10/2016 02:05:15

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.080	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.008	02:01
Breath Volume		1.718	02:01
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.007	02:03
Breath Volume		2.371	02:03
Air Blank		0.000	02:04
Control Test		0.079	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 06/10/2016 10:56

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: CENTERS,BRITTNEY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 06/15/2016 05:24:51

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.080	05:17
Air Blank		0.000	05:17
Subject Sample	# 1	0.174	05:19
Breath Volume		1.371	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.178	05:23
Breath Volume		2.476	05:23
Air Blank		0.000	05:23
Control Test		0.077	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 07/08/2016 13:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,NATHAN,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANEIL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 06/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 050
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MAXWELL

DATE/TIME OF TEST: 06/29/2016 04:40:13

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.000	04:36
Breath Volume		2.046	04:36
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.000	04:38
Breath Volume		1.992	04:38
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 07/08/2016 13:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: WALL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BENSON

DATE/TIME OF TEST: 06/04/2016 03:14:31

Diagnostics Check		OK	03:05	
Air Blank		0.000	03:05	
Control Test		0.076	03:06	
Air Blank		0.000	03:06	
Subject Sample	# 1	0.224	03:09	
Breath Volume		1.488	03:09	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 2	0.227	03:12	
Breath Volume		1.628	03:12	
Air Blank		0.000	03:13	
Control Test		0.074	03:13	Control Outside Tolerance
Air Blank		0.000	03:14	

Data Download Date/Time: 06/29/2016 03:15

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 06/01/2016 19:51:07

Diagnostics Check		OK	19:44
Air Blank		0.000	19:44
Control Test		0.079	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.125	19:46
Breath Volume		2.164	19:46
Air Blank		0.000	19:46
Air Blank		0.000	19:48
Subject Sample	# 2	0.129	19:49
Breath Volume		2.246	19:49
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LASSO,JAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: GASPER

DATE/TIME OF TEST: 06/02/2016 22:53:50

Diagnostics Check		OK	22:46
Air Blank		0.000	22:47
Control Test		0.079	22:47
Air Blank		0.000	22:47
Subject Sample	# 1	0.187	22:49
Breath Volume		3.074	22:49
Air Blank		0.000	22:49
Air Blank		0.000	22:51
Subject Sample	# 2	0.184	22:52
Breath Volume		3.539	22:52
Air Blank		0.000	22:52
Control Test		0.079	22:53
Air Blank		0.000	22:53
Diagnostics Check		OK	22:53

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CUESTA

DATE/TIME OF TEST: 06/05/2016 01:52:09

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.081	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.155	01:47
Breath Volume		1.910	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:49
Subject Sample	# 2	0.151	01:50
Breath Volume		1.613	01:50
Air Blank		0.000	01:51
Control Test		0.080	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MIRANDA,EDWIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: SANCHEZ MONTERO

DATE/TIME OF TEST: 06/06/2016 02:56:37

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.081	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.144	02:51
Breath Volume		1.386	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.144	02:55
Breath Volume		1.394	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GODDEN,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:30

SUBJECT'S LAST NAME: SCHMIDT

DATE/TIME OF TEST: 06/07/2016 12:16:56

Diagnostics Check		OK	12:10
Air Blank		0.000	12:11
Control Test		0.081	12:11
Air Blank		0.000	12:12
Subject Sample	# 1	0.000	12:12
Breath Volume		1.992	12:12
Air Blank		0.000	12:13
Air Blank		0.000	12:15
Subject Sample	# 2	0.000	12:15
Breath Volume		2.785	12:15
Air Blank		0.000	12:16
Control Test		0.081	12:16
Air Blank		0.000	12:16
Diagnostics Check		OK	12:16

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: GAYDOS

DATE/TIME OF TEST: 06/10/2016 23:46:25

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.081	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.180	23:41
Breath Volume		2.214	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.184	23:44
Breath Volume		3.230	23:44
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: GOMEZ ROLDAN

DATE/TIME OF TEST: 06/12/2016 01:04:43

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.081	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.300	00:56
Breath Volume		2.503	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.273	00:59
Breath Volume		1.207	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 3	0.297	01:03
Breath Volume		1.972	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MESSIER

DATE/TIME OF TEST: 06/15/2016 22:32:18

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.081	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.250	22:27
Breath Volume		2.492	22:27
Air Blank		0.000	22:28
Air Blank		0.000	22:30
Subject Sample	# 2	0.235	22:30
Breath Volume		1.980	22:30
Air Blank		0.000	22:31
Control Test		0.080	22:31
Air Blank		0.000	22:32
Diagnostics Check		OK	22:32

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: OLAH

DATE/TIME OF TEST: 06/16/2016 03:14:35

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.081	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.154	03:09
Breath Volume		1.710	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.155	03:13
Breath Volume		1.269	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 06/17/2016 12:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: HOBLIT

DATE/TIME OF TEST: 06/17/2016 03:30:04

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.138	03:25
Breath Volume		2.320	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.137	03:28
Breath Volume		2.027	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:30

Data Download Date/Time: 06/17/2016 12:21

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: VELASQUEZ-BARRIOS

DATE/TIME OF TEST: 06/18/2016 06:16:28

Diagnostics Check		OK	06:10
Air Blank		0.000	06:10
Control Test		0.081	06:10
Air Blank		0.000	06:11
Subject Sample	# 1	0.135	06:12
Breath Volume		2.507	06:12
Air Blank		0.000	06:12
Air Blank		0.000	06:14
Subject Sample	# 2	0.136	06:14
Breath Volume		3.566	06:14
Air Blank		0.000	06:15
Control Test		0.080	06:15
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ECKES,PHILIP
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:28

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 06/18/2016 18:12:09

Diagnostics Check		OK	17:59	
Air Blank		0.000	18:00	
Control Test		0.080	18:00	
Air Blank		0.000	18:01	
Subject Sample	# 1	0.238	18:01	No .020 agreement
Breath Volume		1.726	18:01	
Air Blank		0.000	18:02	
Air Blank		0.000	18:04	
Subject Sample	# 2	0.217	18:04	No .020 agreement
Breath Volume		1.617	18:04	
Air Blank		0.000	18:05	
Air Blank		0.000	18:07	
Subject Sample	# 3	0.149	18:10	Volume Not Met
Breath Volume		0.308	18:10	
Air Blank		0.000	18:11	
Control Test		0.079	18:11	
Air Blank		0.000	18:12	
Diagnostics Check		OK	18:12	

Data Download Date/Time: 07/15/2016 14:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ECKES,PHILIP
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 06/18/2016 18:50:11

Diagnostics Check		OK	18:40
Air Blank		0.000	18:40
Control Test		0.080	18:40
Air Blank		0.000	18:41
Subject Sample	# 1	0.219	18:41
Breath Volume		1.859	18:41
Air Blank		0.000	18:42
Air Blank		0.000	18:44
Subject Sample	# 2	0.187	18:45
Breath Volume		1.515	18:45
Air Blank		0.000	18:46
Air Blank		0.000	18:48
Subject Sample	# 3	0.233	18:48
Breath Volume		2.410	18:48
Air Blank		0.000	18:49
Control Test		0.079	18:49
Air Blank		0.000	18:50
Diagnostics Check		OK	18:50

Data Download Date/Time: 07/15/2016 14:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROCHA,LUIS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: PETRASEK

DATE/TIME OF TEST: 06/21/2016 00:05:19

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:01
Breath Volume		1.273	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.000	00:03
Breath Volume		1.425	00:03
Air Blank		0.000	00:04
Control Test		0.081	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 07/15/2016 14:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KOVAK,EVAN
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: MENJIVAR

DATE/TIME OF TEST: 06/23/2016 02:29:21

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.212	02:24
Breath Volume		3.621	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.207	02:27
Breath Volume		3.066	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: NOEL,KENNETH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: JULES

DATE/TIME OF TEST: 06/25/2016 03:53:06

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.081	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.211	03:48
Breath Volume		1.859	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.216	03:51
Breath Volume		1.511	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STUBBLEBINE,LIVIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: KINSLER IV

DATE/TIME OF TEST: 06/25/2016 05:59:21

Diagnostics Check		OK	05:52
Air Blank		0.000	05:53
Control Test		0.080	05:53
Air Blank		0.000	05:54
Subject Sample	# 1	0.140	05:54
Breath Volume		2.035	05:54
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.144	05:57
Breath Volume		2.757	05:57
Air Blank		0.000	05:58
Control Test		0.080	05:58
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LASSO,JAVIER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 06/25/2016 21:30:55

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:24	
Control Test		0.080	21:24	
Air Blank		0.000	21:24	
Subject Sample	# 1	0.206	21:26	
Breath Volume		2.648	21:26	
Air Blank		0.000	21:27	
Air Blank		0.000	21:29	
Subject Sample	# 2		21:29	Subject Test Refused
Breath Volume		0.000	21:29	
Air Blank		0.000	21:30	
Control Test		0.081	21:30	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:30	

Data Download Date/Time: 07/15/2016 14:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,GARY,W
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DEANGELES

DATE/TIME OF TEST: 06/26/2016 03:03:13

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.000	02:58
Breath Volume		2.042	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.000	03:01
Breath Volume		1.941	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,GARY,W
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: TORRES CHICO

DATE/TIME OF TEST: 06/26/2016 05:17:40

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.080	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.205	05:13
Breath Volume		3.644	05:13
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.194	05:16
Breath Volume		3.863	05:16
Air Blank		0.000	05:16
Control Test		0.080	05:17
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,GARY,W
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOWARD,MONTOUTE
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: RICHARDS

DATE/TIME OF TEST: 06/26/2016 05:59:39

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.080	05:54
Air Blank		0.000	05:54
Subject Sample	# 1	0.130	05:55
Breath Volume		2.402	05:55
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.128	05:58
Breath Volume		2.972	05:58
Air Blank		0.000	05:58
Control Test		0.079	05:59
Air Blank		0.000	05:59
Diagnostics Check		OK	05:59

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001234

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: FRELING

DATE/TIME OF TEST: 06/26/2016 23:33:30

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.227	23:28
Breath Volume		1.101	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.231	23:31
Breath Volume		1.324	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 07/15/2016 14:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ANNESE,MORRIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: KIEDEL

DATE/TIME OF TEST: 06/04/2016 01:31:29

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.080	01:18	
Air Blank		0.000	01:19	
Subject Sample	# 1	0.117	01:22	Volume Not Met
Breath Volume		0.074	01:22	
Air Blank		0.000	01:23	
Air Blank		0.000	01:24	
Subject Sample	# 2	0.125	01:27	
Breath Volume		1.148	01:27	
Air Blank		0.000	01:27	
Air Blank		0.000	01:29	
Subject Sample	# 3	0.133	01:29	
Breath Volume		1.847	01:29	
Air Blank		0.000	01:30	
Control Test		0.079	01:30	
Air Blank		0.000	01:31	
Diagnostics Check		OK	01:31	

Data Download Date/Time: 07/01/2016 11:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/19/2016 00:32:03

Diagnostics Check		OK	00:16	
Air Blank		0.000	00:16	
Control Test		0.080	00:16	
Air Blank		0.000	00:17	
Subject Sample	# 1	0.000	00:20	Volume Not Met
Breath Volume		0.070	00:20	
Air Blank		0.000	00:21	
Air Blank		0.000	00:22	
Subject Sample	# 2	0.144	00:24	
Breath Volume		1.578	00:24	
Air Blank		0.000	00:25	
Air Blank		0.000	00:27	
Subject Sample	# 3	0.146	00:30	Volume Not Met
Breath Volume		0.195	00:30	
Air Blank		0.000	00:31	
Control Test		0.079	00:31	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 07/01/2016 11:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/19/2016 00:48:59

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:34	
Control Test		0.080	00:35	
Air Blank		0.000	00:35	
Subject Sample	# 1	0.135	00:38	
Breath Volume		1.968	00:38	
Air Blank		0.000	00:39	
Air Blank		0.000	00:41	
Subject Sample	# 2	0.141	00:44	Volume Not Met
Breath Volume		0.046	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 3		00:47	Subject Test Refused
Breath Volume		0.000	00:47	
Air Blank		0.000	00:48	
Control Test		0.077	00:48	
Air Blank		0.000	00:48	
Diagnostics Check		OK	00:48	

Data Download Date/Time: 07/01/2016 11:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,D
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 06/25/2016 03:44:00

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.081	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:43	
Control Test		0.080	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 07/01/2016 11:18

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1752856
DRY GAS EXPIRY DATE: 10/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 06/26/2016 03:08:36

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.081	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.035	03:04
Breath Volume		1.765	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.034	03:07
Breath Volume		1.949	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 07/01/2016 11:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALINSKI,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: CORMAN,G
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: FULGENCIO

DATE/TIME OF TEST: 06/04/2016 22:27:55

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Subject Sample	# 1	0.129	22:23
Breath Volume		1.328	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.130	22:26
Breath Volume		1.343	22:26
Air Blank		0.000	22:27
Control Test		0.080	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 06/27/2016 13:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALINSKI,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: STEWART,D
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: STAPLETON

DATE/TIME OF TEST: 06/06/2016 00:22:50

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.127	00:18
Breath Volume		1.765	00:18
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.120	00:21
Breath Volume		1.515	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 06/27/2016 13:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,MELISSA,A
BREATH TEST OPERATOR AGENCY: Tavares PD
ARRESTING OFFICER: GRIEVES,MATTHEW
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MUSCANELL

DATE/TIME OF TEST: 06/09/2016 23:12:34

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.132	23:08
Breath Volume		1.437	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.133	23:11
Breath Volume		1.523	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 06/27/2016 13:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOOTH,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DOUGLAS,AMANDA
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: ELLISON-CRUTCHFIELD

DATE/TIME OF TEST: 06/19/2016 01:49:50

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.256	01:44
Breath Volume		1.125	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.251	01:48
Breath Volume		1.367	01:48
Air Blank		0.000	01:48
Control Test		0.077	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 06/27/2016 13:27

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,TRAVIS,V
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: COOK,TRAVIS
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: IMPERATRICE

DATE/TIME OF TEST: 06/12/2016 03:18:26

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.158	03:13
Breath Volume		2.281	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.161	03:16
Breath Volume		1.796	03:16
Air Blank		0.000	03:17
Control Test		0.078	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 06/15/2016 04:01

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GINTZ,BRIAN,W
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: POTTER,KYLE
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: KRUSE

DATE/TIME OF TEST: 06/11/2016 06:03:43

Diagnostics Check		OK	06:00	
Air Blank		0.000	06:01	
Control Test		0.076	06:01	
Air Blank		0.000	06:02	
Subject Sample	# 1		06:02	Subject Test Refused
Breath Volume		0.000	06:02	
Air Blank		0.000	06:02	
Control Test		0.076	06:03	
Air Blank		0.000	06:03	
Diagnostics Check		OK	06:03	

Data Download Date/Time: 06/12/2016 08:17

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: WETHERELL,TYLER
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 06/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 06/16/2016 21:50:05

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:41	
Control Test		0.075	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1	0.217	21:43	
Breath Volume		1.617	21:43	
Air Blank		0.000	21:43	
Air Blank		0.000	21:45	
Subject Sample	# 2	0.232	21:48	
Breath Volume		1.988	21:48	
Air Blank		0.000	21:49	
Control Test		0.074	21:49	Control Outside Tolerance
Air Blank		0.000	21:50	

Data Download Date/Time: 07/23/2016 07:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,STEPHEN,D
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: KIRKLAND,RICHARD
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 06/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 06/16/2016 23:54:58

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:45	
Control Test		0.075	23:45	
Air Blank		0.000	23:46	
Subject Sample	# 1	0.281	23:49	Volume Not Met
Breath Volume		0.164	23:49	
Air Blank		0.000	23:49	
Air Blank		0.000	23:51	
Subject Sample	# 2		23:53	Subject Test Refused
Breath Volume		0.062	23:53	
Air Blank		0.000	23:54	
Control Test		0.072	23:54	Control Outside Tolerance
Air Blank		0.000	23:54	

Data Download Date/Time: 07/23/2016 07:43

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,ERIC,R
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: FIELDS,ERIC
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: BROZIO

DATE/TIME OF TEST: 06/01/2016 19:47:29

Diagnostics Check		OK	19:41
Air Blank		0.000	19:42
Control Test		0.080	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.000	19:43
Breath Volume		2.242	19:43
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.000	19:46
Breath Volume		2.015	19:46
Air Blank		0.000	19:46
Control Test		0.082	19:46
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: GUILLERM,MICHAEL
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:07

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/03/2016 06:36:19

Diagnostics Check		OK	06:28
Air Blank		0.000	06:28
Control Test		0.081	06:28
Air Blank		0.000	06:29
Subject Sample	# 1	0.056	06:31
Breath Volume		3.535	06:31
Air Blank		0.000	06:31
Air Blank		0.000	06:33
Subject Sample	# 2	0.058	06:34
Breath Volume		1.941	06:34
Air Blank		0.000	06:35
Control Test		0.082	06:35
Air Blank		0.000	06:36
Diagnostics Check		OK	06:36

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: SYDESKI,BLAKE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GATCH

DATE/TIME OF TEST: 06/11/2016 20:19:03

Diagnostics Check		OK	20:11
Air Blank		0.000	20:12
Control Test		0.082	20:12
Air Blank		0.000	20:13
Subject Sample	# 1	0.209	20:14
Breath Volume		1.933	20:14
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.216	20:17
Breath Volume		1.902	20:17
Air Blank		0.000	20:18
Control Test		0.080	20:18
Air Blank		0.000	20:18
Diagnostics Check		OK	20:19

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: LANK,GABRIELLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: SAED

DATE/TIME OF TEST: 06/22/2016 00:29:45

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.082	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.119	00:25
Breath Volume		1.828	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.130	00:28
Breath Volume		1.746	00:28
Air Blank		0.000	00:28
Control Test		0.082	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 06/23/2016 10:14:11

Diagnostics Check		OK	10:08
Air Blank		0.000	10:08
Control Test		0.082	10:08
Air Blank		0.000	10:09
Subject Sample	# 1	0.000	10:09
Breath Volume		1.980	10:09
Air Blank		0.000	10:10
Air Blank		0.000	10:12
Subject Sample	# 2	0.000	10:12
Breath Volume		1.773	10:12
Air Blank		0.000	10:13
Control Test		0.082	10:13
Air Blank		0.000	10:14
Diagnostics Check		OK	10:14

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: WEIR

DATE/TIME OF TEST: 06/23/2016 12:42:21

Diagnostics Check		OK	12:35
Air Blank		0.000	12:36
Control Test		0.082	12:36
Air Blank		0.000	12:37
Subject Sample	# 1	0.000	12:37
Breath Volume		3.003	12:37
Air Blank		0.000	12:38
Air Blank		0.000	12:40
Subject Sample	# 2	0.000	12:40
Breath Volume		2.886	12:40
Air Blank		0.000	12:41
Control Test		0.083	12:41
Air Blank		0.000	12:42
Diagnostics Check		OK	12:42

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STRAIGHT,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: STRAIGHT,KAYLA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 12:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MIALKI

DATE/TIME OF TEST: 06/27/2016 13:27:58

Diagnostics Check		OK	13:22
Air Blank		0.000	13:22
Control Test		0.082	13:23
Air Blank		0.000	13:23
Subject Sample	# 1	0.000	13:23
Breath Volume		2.175	13:23
Air Blank		0.000	13:24
Air Blank		0.000	13:26
Subject Sample	# 2	0.000	13:26
Breath Volume		2.093	13:26
Air Blank		0.000	13:27
Control Test		0.081	13:27
Air Blank		0.000	13:27
Diagnostics Check		OK	13:27

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: NEWBORN

DATE/TIME OF TEST: 06/28/2016 09:43:13

Diagnostics Check		OK	09:36
Air Blank		0.000	09:37
Control Test		0.081	09:37
Air Blank		0.000	09:38
Subject Sample	# 1	0.000	09:38
Breath Volume		1.820	09:38
Air Blank		0.000	09:39
Air Blank		0.000	09:41
Subject Sample	# 2	0.000	09:41
Breath Volume		3.855	09:41
Air Blank		0.000	09:42
Control Test		0.082	09:42
Air Blank		0.000	09:43
Diagnostics Check		OK	09:43

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 08:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:30

SUBJECT'S LAST NAME: WILGERS

DATE/TIME OF TEST: 06/29/2016 08:59:03

Diagnostics Check		OK	08:53
Air Blank		0.000	08:53
Control Test		0.081	08:53
Air Blank		0.000	08:54
Subject Sample	# 1	0.000	08:54
Breath Volume		2.871	08:54
Air Blank		0.000	08:55
Air Blank		0.000	08:57
Subject Sample	# 2	0.000	08:57
Breath Volume		3.140	08:57
Air Blank		0.000	08:58
Control Test		0.081	08:58
Air Blank		0.000	08:58
Diagnostics Check		OK	08:59

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001245

INSTRUMENT REGISTERED TO: PORT ORANGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 06/29/2016 13:39:31

Diagnostics Check		OK	13:33
Air Blank		0.000	13:33
Control Test		0.082	13:33
Air Blank		0.000	13:34
Subject Sample	# 1	0.000	13:35
Breath Volume		2.457	13:35
Air Blank		0.000	13:35
Air Blank		0.000	13:37
Subject Sample	# 2	0.000	13:38
Breath Volume		2.644	13:38
Air Blank		0.000	13:38
Control Test		0.082	13:38
Air Blank		0.000	13:39
Diagnostics Check		OK	13:39

Data Download Date/Time: 06/30/2016 13:49

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: LANK,GABRIELLE
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 06/15/2016 22:49:44

Diagnostics Check		OK	22:46	
Air Blank		0.000	22:47	
Control Test		0.080	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 1		22:48	Subject Test Refused
Breath Volume		0.000	22:48	
Air Blank		0.000	22:48	
Control Test		0.080	22:49	
Air Blank		0.000	22:49	
Diagnostics Check		OK	22:49	

Data Download Date/Time: 06/29/2016 10:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCRACKEN,KEVIN,R
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: CHARLA,MICHAEL
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 15:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: MZINGO

DATE/TIME OF TEST: 06/25/2016 16:10:31

Diagnostics Check		OK	16:07	
Air Blank		0.000	16:07	
Control Test		0.080	16:08	
Air Blank		0.000	16:08	
Subject Sample	# 1		16:09	Subject Test Refused
Breath Volume		0.000	16:09	
Air Blank		0.000	16:09	
Control Test		0.079	16:09	
Air Blank		0.000	16:10	
Diagnostics Check		OK	16:10	

Data Download Date/Time: 06/29/2016 10:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: POPICK,SCOTT
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MATHURIN

DATE/TIME OF TEST: 06/01/2016 05:08:48

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.000	05:04
Breath Volume		1.500	05:04
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.000	05:07
Breath Volume		1.230	05:07
Air Blank		0.000	05:07
Control Test		0.080	05:08
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERNABE,DAVID
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: BERNABE

DATE/TIME OF TEST: 06/01/2016 20:48:18

Diagnostics Check		OK	20:41
Air Blank		0.000	20:41
Control Test		0.080	20:42
Air Blank		0.000	20:42
Subject Sample	# 1	0.162	20:43
Breath Volume		1.164	20:43
Air Blank		0.000	20:43
Air Blank		0.000	20:45
Subject Sample	# 2	0.163	20:46
Breath Volume		1.988	20:46
Air Blank		0.000	20:47
Control Test		0.079	20:47
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ-BUENO,YENIER
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: ORFANEDES

DATE/TIME OF TEST: 06/03/2016 05:25:59

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.079	05:18
Air Blank		0.000	05:19
Subject Sample	# 1	0.021	05:21
Breath Volume		1.523	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.021	05:24
Breath Volume		1.367	05:24
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KOMP,CHRIS
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 06/04/2016 20:41:42

Diagnostics Check		OK	20:34
Air Blank		0.000	20:34
Control Test		0.079	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.207	20:36
Breath Volume		1.863	20:36
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.206	20:40
Breath Volume		1.445	20:40
Air Blank		0.000	20:40
Control Test		0.080	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 06/23/2016 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ARGUETA

DATE/TIME OF TEST: 06/05/2016 05:13:05

Diagnostics Check		OK	05:03
Air Blank		0.000	05:04
Control Test		0.080	05:04
Air Blank		0.000	05:05
Subject Sample	# 1	0.130	05:05
Breath Volume		1.132	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.128	05:11
Breath Volume		1.437	05:11
Air Blank		0.000	05:12
Control Test		0.079	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 06/23/2016 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPROUSE,MICHAEL
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: DUBAR

DATE/TIME OF TEST: 06/05/2016 06:18:50

Diagnostics Check		OK	06:12
Air Blank		0.000	06:13
Control Test		0.079	06:13
Air Blank		0.000	06:13
Subject Sample	# 1	0.182	06:14
Breath Volume		2.386	06:14
Air Blank		0.000	06:15
Air Blank		0.000	06:16
Subject Sample	# 2	0.180	06:17
Breath Volume		1.917	06:17
Air Blank		0.000	06:17
Control Test		0.078	06:18
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: LOPEZ GUZMAN

DATE/TIME OF TEST: 06/05/2016 23:33:21

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:21	
Control Test		0.079	23:21	
Air Blank		0.000	23:22	
Subject Sample	# 1	0.226	23:22	
Breath Volume		2.222	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 2		23:28	No Sample Provided
Breath Volume		0.000	23:28	
Air Blank		0.000	23:28	
Air Blank		0.000	23:30	
Subject Sample	# 3	0.224	23:31	
Breath Volume		1.468	23:31	
Air Blank		0.000	23:32	
Control Test		0.080	23:32	
Air Blank		0.000	23:33	
Diagnostics Check		OK	23:33	

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: THOMPSON,BRETT
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BARAHONA

DATE/TIME OF TEST: 06/11/2016 05:29:02

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.079	05:21
Air Blank		0.000	05:21
Subject Sample	# 1	0.062	05:24
Breath Volume		2.097	05:24
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.057	05:27
Breath Volume		1.179	05:27
Air Blank		0.000	05:28
Control Test		0.078	05:28
Air Blank		0.000	05:28
Diagnostics Check		OK	05:29

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUENO-DIAZ,YENIER
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 06/12/2016 04:13:07

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.152	04:08
Breath Volume		2.617	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.156	04:11
Breath Volume		2.375	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHARLES,WOODNEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 06/13/2016 03:14:01

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.154	03:09
Breath Volume		1.183	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.155	03:12
Breath Volume		1.562	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:14

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: KENT

DATE/TIME OF TEST: 06/16/2016 21:36:35

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.080	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.235	21:32
Breath Volume		2.253	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.243	21:35
Breath Volume		2.265	21:35
Air Blank		0.000	21:35
Control Test		0.078	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 06/23/2016 15:53

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: CAMPOSECO-MENDEZ

DATE/TIME OF TEST: 06/18/2016 00:05:55

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.079	23:59
Air Blank		0.000	23:59
Subject Sample	# 1	0.162	00:01
Breath Volume		3.453	00:01
Air Blank		0.000	00:02
Air Blank		0.000	00:03
Subject Sample	# 2	0.153	00:04
Breath Volume		1.832	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: LILJA

DATE/TIME OF TEST: 06/18/2016 03:50:11

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.232	03:45
Breath Volume		2.277	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.224	03:48
Breath Volume		1.386	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001248

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 15615080A3
DRY GAS EXPIRY DATE: 07/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: JEAN GILLES

DATE/TIME OF TEST: 06/19/2016 05:10:03

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.166	05:05
Breath Volume		1.523	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:07
Subject Sample	# 2	0.170	05:08
Breath Volume		1.546	05:08
Air Blank		0.000	05:09
Control Test		0.078	05:09
Air Blank		0.000	05:09
Diagnostics Check		OK	05:10

Data Download Date/Time: 06/23/2016 15:53

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEENE,PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: CONTE II

DATE/TIME OF TEST: 06/02/2016 02:45:53

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:43	
Control Test		0.079	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1		02:44	Subject Test Refused
Breath Volume		0.000	02:44	
Air Blank		0.000	02:44	
Control Test		0.079	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 06/03/2016 09:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,EDWARD
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: MERCER

DATE/TIME OF TEST: 06/03/2016 02:02:11

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.194	01:53
Breath Volume		1.398	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.216	01:57
Breath Volume		1.722	01:57
Air Blank		0.000	01:58
Air Blank		0.000	01:59
Subject Sample	# 3	0.219	02:00
Breath Volume		4.390	02:00
Air Blank		0.000	02:01
Control Test		0.077	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 06/03/2016 09:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: CHIRRE

DATE/TIME OF TEST: 06/03/2016 03:08:37

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.079	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.123	03:04
Breath Volume		1.652	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.120	03:07
Breath Volume		1.308	03:07
Air Blank		0.000	03:07
Control Test		0.077	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 06/03/2016 09:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMPSON,MAURICE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MESSEGUER

DATE/TIME OF TEST: 06/04/2016 03:45:42

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.079	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.082	03:41
Breath Volume		3.167	03:41
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.085	03:44
Breath Volume		3.699	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 06/09/2016 15:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BECTON,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/05/2016 03:42:37

Diagnostics Check		OK	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.134	03:38
Breath Volume		1.679	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.138	03:41
Breath Volume		1.984	03:41
Air Blank		0.000	03:41
Control Test		0.079	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 06/09/2016 15:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOTES,ANDREW
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BERNARD

DATE/TIME OF TEST: 06/06/2016 01:38:42

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.078	01:30	
Air Blank		0.000	01:31	
Subject Sample	# 1		01:34	No Sample Provided
Breath Volume		0.000	01:34	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.078	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 06/09/2016 15:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERR,BRANDEN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GRAVER

DATE/TIME OF TEST: 06/06/2016 03:28:01

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.079	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:26	Subject Test Refused
Breath Volume		0.000	03:26	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 06/09/2016 15:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS,C.F.
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/06/2016 04:28:20

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.281	04:23
Breath Volume		1.183	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.288	04:26
Breath Volume		2.292	04:26
Air Blank		0.000	04:27
Control Test		0.079	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 06/09/2016 15:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 06/08/2016 02:56:38

Diagnostics Check		OK	02:50
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.287	02:51
Breath Volume		2.699	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.281	02:54
Breath Volume		2.761	02:54
Air Blank		0.000	02:55
Control Test		0.078	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/09/2016 15:16

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 06/11/2016 02:25:34

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.079	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.055	02:21
Breath Volume		1.230	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.052	02:24
Breath Volume		1.125	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 06/15/2016 12:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHISEN,LEIGH
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: BIDIYANATH

DATE/TIME OF TEST: 06/12/2016 01:14:26

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.187	01:09
Breath Volume		2.351	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.185	01:12
Breath Volume		2.328	01:12
Air Blank		0.000	01:13
Control Test		0.078	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 06/15/2016 12:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZITO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: HANSON

DATE/TIME OF TEST: 06/12/2016 03:26:25

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.267	03:21
Breath Volume		2.601	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.267	03:24
Breath Volume		1.683	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/15/2016 12:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: VERDUGO

DATE/TIME OF TEST: 06/12/2016 23:03:42

Diagnostics Check		OK	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Subject Sample	# 1	0.264	22:59
Breath Volume		1.968	22:59
Air Blank		0.000	22:59
Air Blank		0.000	23:01
Subject Sample	# 2	0.248	23:02
Breath Volume		1.136	23:02
Air Blank		0.000	23:02
Control Test		0.079	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 06/15/2016 12:33

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MULLEN,NATHAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 06/19/2016 10:13:03

Diagnostics Check		OK	10:06
Air Blank		0.000	10:06
Control Test		0.080	10:07
Air Blank		0.000	10:07
Subject Sample	# 1	0.253	10:08
Breath Volume		1.988	10:08
Air Blank		0.000	10:09
Air Blank		0.000	10:11
Subject Sample	# 2	0.248	10:11
Breath Volume		3.292	10:11
Air Blank		0.000	10:12
Control Test		0.078	10:12
Air Blank		0.000	10:12
Diagnostics Check		OK	10:13

Data Download Date/Time: 06/23/2016 14:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUTTO,GERALD
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: METZGER

DATE/TIME OF TEST: 06/25/2016 04:29:16

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.157	04:24
Breath Volume		1.949	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.160	04:27
Breath Volume		1.632	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 07/07/2016 11:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELFORD,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:46

SUBJECT'S LAST NAME: VALCARCEL

DATE/TIME OF TEST: 06/25/2016 08:15:25

Diagnostics Check		OK	08:09
Air Blank		0.000	08:09
Control Test		0.078	08:09
Air Blank		0.000	08:10
Subject Sample	# 1	0.217	08:10
Breath Volume		2.808	08:10
Air Blank		0.000	08:11
Air Blank		0.000	08:13
Subject Sample	# 2	0.213	08:13
Breath Volume		2.011	08:13
Air Blank		0.000	08:14
Control Test		0.078	08:14
Air Blank		0.000	08:15
Diagnostics Check		OK	08:15

Data Download Date/Time: 07/07/2016 11:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAM,ALEXANDER
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SANTOSGARCIA

DATE/TIME OF TEST: 06/01/2016 00:58:28

Diagnostics Check		OK	00:52
Air Blank		0.000	00:52
Control Test		0.080	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.185	00:54
Breath Volume		2.406	00:54
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.183	00:57
Breath Volume		1.792	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 06/03/2016 09:25

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REARDON,SARAH
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: WASKOWSKY

DATE/TIME OF TEST: 06/03/2016 20:16:47

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.081	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.000	20:12
Breath Volume		1.699	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.000	20:15
Breath Volume		1.574	20:15
Air Blank		0.000	20:15
Control Test		0.081	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 06/09/2016 15:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WING,RYAN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: PETRIE

DATE/TIME OF TEST: 06/04/2016 03:40:15

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.166	03:35
Breath Volume		2.855	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.160	03:38
Breath Volume		3.046	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIILS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: DOMINGUEZGONZALEZ

DATE/TIME OF TEST: 06/04/2016 06:31:31

Diagnostics Check		OK	06:23
Air Blank		0.000	06:23
Control Test		0.080	06:24
Air Blank		0.000	06:24
Subject Sample	# 1	0.201	06:26
Breath Volume		1.406	06:26
Air Blank		0.000	06:26
Air Blank		0.000	06:28
Subject Sample	# 2	0.195	06:30
Breath Volume		1.550	06:30
Air Blank		0.000	06:30
Control Test		0.080	06:30
Air Blank		0.000	06:31
Diagnostics Check		OK	06:31

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RANDO,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: ASHLEY

DATE/TIME OF TEST: 06/04/2016 10:43:36

Diagnostics Check		OK	10:37
Air Blank		0.000	10:38
Control Test		0.080	10:38
Air Blank		0.000	10:39
Subject Sample	# 1	0.000	10:39
Breath Volume		3.445	10:39
Air Blank		0.000	10:39
Air Blank		0.000	10:41
Subject Sample	# 2	0.000	10:42
Breath Volume		3.304	10:42
Air Blank		0.000	10:42
Control Test		0.080	10:43
Air Blank		0.000	10:43
Diagnostics Check		OK	10:43

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/05/2016 02:57:06

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.081	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1	0.101	02:45	
Breath Volume		1.109	02:45	
Air Blank		0.000	02:46	
Air Blank		0.000	02:48	
Subject Sample	# 2	0.081	02:51	Volume Not Met
Breath Volume		0.425	02:51	
Air Blank		0.000	02:51	
Air Blank		0.000	02:53	
Subject Sample	# 3	0.105	02:55	
Breath Volume		1.589	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CAMPI

DATE/TIME OF TEST: 06/05/2016 03:43:44

Diagnostics Check		OK	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.139	03:39
Breath Volume		2.535	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.140	03:42
Breath Volume		3.238	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GATTON,BRADY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: SCHUSTER

DATE/TIME OF TEST: 06/06/2016 00:07:59

Diagnostics Check		OK	23:56	
Air Blank		0.000	23:57	
Control Test		0.080	23:57	
Air Blank		0.000	23:57	
Subject Sample	# 1		00:01	No Sample Provided
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Air Blank		0.000	00:03	
Subject Sample	# 2		00:06	No Sample Provided
Breath Volume		0.000	00:06	
Air Blank		0.000	00:07	
Control Test		0.080	00:07	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 06/09/2016 15:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,ELVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: JUSTE

DATE/TIME OF TEST: 06/06/2016 02:36:39

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:24	
Control Test		0.079	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1	0.262	02:28	Volume Not Met
Breath Volume		0.753	02:28	
Air Blank		0.000	02:28	
Air Blank		0.000	02:30	
Subject Sample	# 2	0.254	02:32	
Breath Volume		2.308	02:32	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 3	0.272	02:35	
Breath Volume		1.265	02:35	
Air Blank		0.000	02:35	
Control Test		0.077	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:36	

Data Download Date/Time: 06/09/2016 15:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHAPLIN,CHRIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: STARR

DATE/TIME OF TEST: 06/06/2016 10:52:47

Diagnostics Check		OK	10:46
Air Blank		0.000	10:47
Control Test		0.079	10:47
Air Blank		0.000	10:48
Subject Sample	# 1	0.219	10:48
Breath Volume		2.187	10:48
Air Blank		0.000	10:49
Air Blank		0.000	10:50
Subject Sample	# 2	0.218	10:51
Breath Volume		2.027	10:51
Air Blank		0.000	10:51
Control Test		0.079	10:52
Air Blank		0.000	10:52
Diagnostics Check		OK	10:52

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROY,BRIAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: EITH

DATE/TIME OF TEST: 06/07/2016 03:44:59

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.124	03:40
Breath Volume		2.121	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.127	03:43
Breath Volume		1.972	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ODIERNO,TROY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: KORN

DATE/TIME OF TEST: 06/09/2016 03:26:01

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:15	
Control Test		0.081	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:19	No Sample Provided
Breath Volume		0.000	03:19	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 2		03:24	No Sample Provided
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.080	03:25	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 06/09/2016 15:09

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 06/10/2016 03:59:03

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Subject Sample	# 1	0.145	03:54
Breath Volume		1.472	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.144	03:57
Breath Volume		1.820	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:59

Data Download Date/Time: 06/15/2016 09:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELFORD,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 06/11/2016 02:04:30

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.211	02:00
Breath Volume		2.960	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.191	02:03
Breath Volume		1.585	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 06/15/2016 09:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILSON,RUDY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 06/11/2016 04:54:15

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.079	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.170	04:46
Breath Volume		1.492	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:48
Subject Sample	# 2	0.192	04:49
Breath Volume		1.648	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:52
Subject Sample	# 3	0.174	04:52
Breath Volume		1.464	04:52
Air Blank		0.000	04:53
Control Test		0.079	04:53
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 06/15/2016 09:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JASAT,HAMZA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: SERRANORAMOS

DATE/TIME OF TEST: 06/12/2016 00:15:27

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.080	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.179	00:10
Breath Volume		1.457	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.176	00:14
Breath Volume		1.328	00:14
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 06/15/2016 09:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,GEAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: BUCKLAND

DATE/TIME OF TEST: 06/14/2016 01:40:18

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.081	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.135	01:35
Breath Volume		1.867	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.130	01:38
Breath Volume		2.382	01:38
Air Blank		0.000	01:39
Control Test		0.080	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 06/15/2016 09:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,FOREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: AYENI

DATE/TIME OF TEST: 06/14/2016 03:52:20

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.000	03:48
Breath Volume		1.175	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.000	03:51
Breath Volume		1.730	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 06/15/2016 09:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BIRDSONG,KEVIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 10:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:07

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 06/16/2016 11:40:10

Diagnostics Check		OK	11:33
Air Blank		0.000	11:33
Control Test		0.078	11:34
Air Blank		0.000	11:34
Subject Sample	# 1	0.380	11:35
Breath Volume		2.035	11:35
Air Blank		0.000	11:35
Air Blank		0.000	11:37
Subject Sample	# 2	0.398	11:38
Breath Volume		3.375	11:38
Air Blank		0.000	11:39
Control Test		0.078	11:39
Air Blank		0.000	11:40
Diagnostics Check		OK	11:40

Data Download Date/Time: 06/23/2016 15:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE,PETER
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: PISILLO

DATE/TIME OF TEST: 06/17/2016 01:18:49

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.079	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.171	01:14
Breath Volume		2.093	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.183	01:17
Breath Volume		2.675	01:17
Air Blank		0.000	01:17
Control Test		0.078	01:18
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 06/23/2016 15:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAL,RAJIV
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 06/18/2016 03:26:23

Diagnostics Check		OK	03:19
Air Blank		0.000	03:20
Control Test		0.079	03:20
Air Blank		0.000	03:20
Subject Sample	# 1	0.078	03:21
Breath Volume		1.265	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.069	03:24
Breath Volume		1.148	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 06/23/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:46

SUBJECT'S LAST NAME: ALLI

DATE/TIME OF TEST: 06/18/2016 06:14:55

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.078	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.141	06:10
Breath Volume		2.648	06:10
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.144	06:13
Breath Volume		1.289	06:13
Air Blank		0.000	06:14
Control Test		0.079	06:14
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 06/23/2016 15:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ,EDUARDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BONDS

DATE/TIME OF TEST: 06/20/2016 05:28:32

Diagnostics Check		OK	05:22
Air Blank		0.000	05:23
Control Test		0.078	05:23
Air Blank		0.000	05:23
Subject Sample	# 1	0.088	05:24
Breath Volume		2.003	05:24
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.084	05:27
Breath Volume		1.750	05:27
Air Blank		0.000	05:27
Control Test		0.078	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 06/23/2016 15:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 06/23/2016 02:04:38

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.078	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.171	02:00
Breath Volume		2.207	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.170	02:03
Breath Volume		2.656	02:03
Air Blank		0.000	02:03
Control Test		0.078	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 06/23/2016 15:04

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BRANSDORF

DATE/TIME OF TEST: 06/24/2016 02:14:51

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.194	02:10
Breath Volume		3.976	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:12
Subject Sample	# 2	0.199	02:13
Breath Volume		4.269	02:13
Air Blank		0.000	02:13
Control Test		0.077	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 07/07/2016 11:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSA,SHAYMARIE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 06/24/2016 03:57:00

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.000	03:52
Breath Volume		1.542	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.000	03:55
Breath Volume		1.570	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:57

Data Download Date/Time: 07/07/2016 11:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: MALLIA

DATE/TIME OF TEST: 06/25/2016 03:34:50

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:29
Air Blank		0.000	03:29
Subject Sample	# 1	0.115	03:30
Breath Volume		3.125	03:30
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.112	03:33
Breath Volume		2.121	03:33
Air Blank		0.000	03:33
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 07/07/2016 11:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MCGRATH

DATE/TIME OF TEST: 06/30/2016 02:51:02

Diagnostics Check		OK	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.257	02:46
Breath Volume		2.332	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.260	02:49
Breath Volume		1.242	02:49
Air Blank		0.000	02:50
Control Test		0.078	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:51

Data Download Date/Time: 07/07/2016 11:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 06/10/2016 00:37:21

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.080	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.174	00:33
Breath Volume		2.054	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.168	00:35
Breath Volume		1.906	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 06/15/2016 12:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CULLUM

DATE/TIME OF TEST: 06/10/2016 04:05:18

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.213	04:00
Breath Volume		2.800	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.220	04:03
Breath Volume		3.203	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 06/15/2016 12:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GABRUS,KYLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: MASTERSON

DATE/TIME OF TEST: 06/10/2016 19:55:35

Diagnostics Check		OK	19:48
Air Blank		0.000	19:49
Control Test		0.080	19:49
Air Blank		0.000	19:50
Subject Sample	# 1	0.244	19:50
Breath Volume		2.765	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.248	19:53
Breath Volume		2.964	19:53
Air Blank		0.000	19:54
Control Test		0.078	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 06/15/2016 12:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHERRY,ADAM
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 06/11/2016 01:01:46

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.079	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:57	No Sample Provided
Breath Volume		0.000	00:57	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2		01:00	Subject Test Refused
Breath Volume		0.000	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 06/15/2016 12:29

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MERKS,ANDREW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: EMORY

DATE/TIME OF TEST: 06/11/2016 04:00:38

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 06/15/2016 12:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REMBISLICARI,JAZMIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: CRUZMARTINEZ

DATE/TIME OF TEST: 06/11/2016 22:02:06

Diagnostics Check		OK	21:55
Air Blank		0.000	21:56
Control Test		0.080	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.009	21:57
Breath Volume		1.187	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.009	22:00
Breath Volume		1.121	22:00
Air Blank		0.000	22:01
Control Test		0.080	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 06/15/2016 12:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROLAND,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HALLADAY

DATE/TIME OF TEST: 06/12/2016 00:58:03

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.134	00:53
Breath Volume		4.042	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.125	00:56
Breath Volume		4.167	00:56
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:58

Data Download Date/Time: 06/15/2016 12:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,BRETT,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GERARDI,GREGORY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BEKDERNAH

DATE/TIME OF TEST: 06/12/2016 01:50:37

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.015	01:46
Breath Volume		1.156	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.015	01:49
Breath Volume		1.175	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 06/15/2016 12:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCREWS,DARCY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BARTON

DATE/TIME OF TEST: 06/12/2016 02:44:18

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.120	02:39
Breath Volume		2.589	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.109	02:42
Breath Volume		2.652	02:42
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 06/15/2016 12:30

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KING,GORDON
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 06/19/2016 07:19:25

Diagnostics Check		OK	07:13
Air Blank		0.000	07:13
Control Test		0.080	07:13
Air Blank		0.000	07:14
Subject Sample	# 1	0.000	07:15
Breath Volume		2.550	07:15
Air Blank		0.000	07:15
Air Blank		0.000	07:17
Subject Sample	# 2	0.000	07:18
Breath Volume		2.421	07:18
Air Blank		0.000	07:18
Control Test		0.080	07:18
Air Blank		0.000	07:19
Diagnostics Check		OK	07:19

Data Download Date/Time: 06/23/2016 15:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/22/2016 22:37:42

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.080	22:35	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.080	22:37	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 06/23/2016 15:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BAYLEY

DATE/TIME OF TEST: 06/23/2016 03:53:58

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:38	
Air Blank		0.000	03:39	
Subject Sample	# 1	0.110	03:42	
Breath Volume		2.000	03:42	
Air Blank		0.000	03:42	
Air Blank		0.000	03:44	
Subject Sample	# 2	0.084	03:47	Volume Not Met
Breath Volume		0.402	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:50	
Subject Sample	# 3	0.097	03:52	
Breath Volume		1.273	03:52	
Air Blank		0.000	03:53	
Control Test		0.078	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 06/23/2016 15:00

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WING,BRIAN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DABKOWSKI

DATE/TIME OF TEST: 06/24/2016 00:42:53

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.000	00:38
Breath Volume		1.664	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.000	00:41
Breath Volume		2.035	00:41
Air Blank		0.000	00:41
Control Test		0.080	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 07/07/2016 11:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DICARLO,JOE
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ROBERTSON III

DATE/TIME OF TEST: 06/24/2016 02:44:31

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.202	02:39
Breath Volume		3.753	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.206	02:42
Breath Volume		3.714	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 07/07/2016 11:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: D0BBINS,SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: EKRAM

DATE/TIME OF TEST: 06/25/2016 01:11:57

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.143	01:07
Breath Volume		3.414	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.127	01:10
Breath Volume		1.128	01:10
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 07/07/2016 11:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROGERS,MATTHEW
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 06/26/2016 02:38:31

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.079	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.079	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 07/07/2016 11:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROSARIO,MIGUEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 06/29/2016 19:22:43

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.080	19:17
Air Blank		0.000	19:17
Subject Sample	# 1	0.220	19:18
Breath Volume		1.613	19:18
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.218	19:21
Breath Volume		1.804	19:21
Air Blank		0.000	19:21
Control Test		0.079	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 07/07/2016 11:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PINO-PEREZ,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05515080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HERNANDEZ-PEREZ

DATE/TIME OF TEST: 06/30/2016 22:12:46

Diagnostics Check		OK	22:06
Air Blank		0.000	22:06
Control Test		0.080	22:07
Air Blank		0.000	22:07
Subject Sample	# 1	0.146	22:08
Breath Volume		2.625	22:08
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.146	22:11
Breath Volume		2.867	22:11
Air Blank		0.000	22:11
Control Test		0.079	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 07/07/2016 11:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAVALIERI,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: RUSH,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: FRAHM

DATE/TIME OF TEST: 06/02/2016 15:48:01

Diagnostics Check		OK	15:38	
Air Blank		0.000	15:38	
Control Test		0.080	15:39	
Air Blank		0.000	15:39	
Subject Sample	# 1	0.314	15:40	
Breath Volume		2.058	15:40	
Air Blank		0.000	15:41	
Air Blank		0.000	15:42	
Subject Sample	# 2	0.293	15:43	
Breath Volume		1.296	15:43	
Air Blank		0.000	15:44	
Air Blank		0.000	15:45	
Subject Sample	# 3		15:46	Slope Not Met
Breath Volume		0.871	15:46	
Air Blank		0.000	15:47	
Control Test		0.079	15:47	
Air Blank		0.000	15:47	
Diagnostics Check		OK	15:48	

Data Download Date/Time: 06/23/2016 23:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAVALIERI,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: RUSH,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:18

SUBJECT'S LAST NAME: FRAHM

DATE/TIME OF TEST: 06/02/2016 15:58:17

Diagnostics Check		OK	15:50
Air Blank		0.000	15:51
Control Test		0.078	15:51
Air Blank		0.000	15:51
Subject Sample	# 1	0.301	15:53
Breath Volume		2.605	15:53
Air Blank		0.000	15:54
Air Blank		0.000	15:55
Subject Sample	# 2	0.291	15:56
Breath Volume		2.386	15:56
Air Blank		0.000	15:57
Control Test		0.077	15:57
Air Blank		0.000	15:58
Diagnostics Check		OK	15:58

Data Download Date/Time: 06/23/2016 23:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAVALIERI,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: RUSH,MATTHEW
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A2
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/04/2016 21:09:02

Diagnostics Check		OK	21:01
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.188	21:03
Breath Volume		1.246	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.191	21:07
Breath Volume		2.394	21:07
Air Blank		0.000	21:08
Control Test		0.079	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:09

Data Download Date/Time: 06/23/2016 23:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: ANDERSON,HUNTER
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: FOHY

DATE/TIME OF TEST: 06/04/2016 21:48:04

Diagnostics Check		OK	21:44	
Air Blank		0.000	21:45	
Control Test		0.079	21:45	
Air Blank		0.000	21:45	
Subject Sample	# 1		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:47	
Control Test		0.079	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 06/28/2016 12:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: STRICKLAND,WINFRED
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: STAPP

DATE/TIME OF TEST: 06/05/2016 23:38:49

Diagnostics Check		OK	23:31
Air Blank		0.000	23:32
Control Test		0.079	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.140	23:33
Breath Volume		1.750	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.135	23:37
Breath Volume		1.472	23:37
Air Blank		0.000	23:37
Control Test		0.078	23:38
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 06/28/2016 12:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SIMMONS,BRIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 06/09/2016 02:43:19

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:32	
Control Test		0.078	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.145	02:36	Volume Not Met
Breath Volume		0.417	02:36	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.164	02:41	Volume Not Met
Breath Volume		0.390	02:41	
Air Blank		0.000	02:42	
Control Test		0.077	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 06/28/2016 12:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SIMMONS,BRIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 06/09/2016 03:02:15

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:49	
Control Test		0.077	02:49	
Air Blank		0.000	02:50	
Subject Sample	# 1	0.163	02:51	
Breath Volume		1.824	02:51	
Air Blank		0.000	02:52	
Air Blank		0.000	02:53	
Subject Sample	# 2	0.158	02:56	Volume Not Met
Breath Volume		0.343	02:56	
Air Blank		0.000	02:57	
Air Blank		0.000	02:59	
Subject Sample	# 3	0.159	03:00	
Breath Volume		2.007	03:00	
Air Blank		0.000	03:01	
Control Test		0.077	03:01	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 06/28/2016 12:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: ANDERSON,HUNTER
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: SAVAGE

DATE/TIME OF TEST: 06/12/2016 03:10:59

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.222	03:06
Breath Volume		1.859	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.217	03:09
Breath Volume		1.867	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 06/28/2016 12:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSE,DAVID,A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: JOHNSON,SHAWN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PEREZ PALENCIA

DATE/TIME OF TEST: 06/19/2016 03:52:20

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.193	03:47
Breath Volume		2.468	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.183	03:50
Breath Volume		1.421	03:50
Air Blank		0.000	03:51
Control Test		0.077	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 06/28/2016 12:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSE,DAVID,A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: WOOD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/19/2016 21:55:20

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:46	
Control Test		0.079	21:47	
Air Blank		0.000	21:47	
Subject Sample	# 1	0.270	21:50	
Breath Volume		1.445	21:50	
Air Blank		0.000	21:51	
Air Blank		0.000	21:52	
Subject Sample	# 2	0.251	21:53	
Breath Volume		1.890	21:53	
Air Blank		0.000	21:54	
Control Test		0.074	21:54	Control Outside Tolerance
Air Blank		0.000	21:55	

Data Download Date/Time: 06/28/2016 12:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSE,DAVID,A
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: WOOD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 06/19/2016 22:04:16

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.078	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.259	21:59
Breath Volume		2.035	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.262	22:02
Breath Volume		2.203	22:02
Air Blank		0.000	22:03
Control Test		0.077	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 06/28/2016 12:37

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BURTTRAM

DATE/TIME OF TEST: 06/01/2016 18:11:43

Diagnostics Check		OK	18:05
Air Blank		0.000	18:05
Control Test		0.079	18:06
Air Blank		0.000	18:06
Subject Sample	# 1	0.186	18:07
Breath Volume		1.699	18:07
Air Blank		0.000	18:07
Air Blank		0.000	18:09
Subject Sample	# 2	0.187	18:10
Breath Volume		2.277	18:10
Air Blank		0.000	18:10
Control Test		0.081	18:11
Air Blank		0.000	18:11
Diagnostics Check		OK	18:11

Data Download Date/Time: 06/28/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: HEESTAND

DATE/TIME OF TEST: 06/05/2016 02:44:27

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.073	02:39
Breath Volume		1.488	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.070	02:42
Breath Volume		1.410	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 06/28/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELF,BRETT,R
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: CARLSON,EDWIN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MONTES DE OCA

DATE/TIME OF TEST: 06/18/2016 02:56:38

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.080	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.061	02:51
Breath Volume		2.589	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.058	02:55
Breath Volume		2.027	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 06/28/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: GOLDMAN

DATE/TIME OF TEST: 06/20/2016 00:25:24

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.129	00:20
Breath Volume		2.656	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.130	00:23
Breath Volume		2.804	00:23
Air Blank		0.000	00:24
Control Test		0.080	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 06/28/2016 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,MATTHEW,G
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: MILLER,MATTHEW
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 06/25/2016 03:28:39

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.080	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.132	03:23
Breath Volume		1.156	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.131	03:27
Breath Volume		1.527	03:27
Air Blank		0.000	03:27
Control Test		0.080	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 06/28/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,MATTHEW,G
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: ROCHE,MARK
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: SIMONES

DATE/TIME OF TEST: 06/25/2016 05:07:06

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.155	05:02
Breath Volume		1.863	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:04
Subject Sample	# 2	0.153	05:05
Breath Volume		2.332	05:05
Air Blank		0.000	05:06
Control Test		0.079	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 06/28/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SELF,BRETT,R
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BRANDEL,NATHANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 133849
DRY GAS EXPIRY DATE: 10/12/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GOWINS

DATE/TIME OF TEST: 06/28/2016 00:25:35

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.080	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.132	00:19
Breath Volume		2.074	00:19
Air Blank		0.000	00:20
Air Blank		0.000	00:21
Subject Sample	# 2	0.123	00:23
Breath Volume		1.855	00:23
Air Blank		0.000	00:24
Control Test		0.081	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 06/28/2016 11:41

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,B,S
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: WILLIAMS,B
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 05/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MAHDESIAN

DATE/TIME OF TEST: 06/11/2016 03:38:14

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.217	03:33
Breath Volume		1.187	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.213	03:36
Breath Volume		1.542	03:36
Air Blank		0.000	03:37
Control Test		0.077	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 06/23/2016 08:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: WEST,CHRISTINA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: MCDERMOTT

DATE/TIME OF TEST: 06/01/2016 00:26:43

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.079	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.030	00:21
Breath Volume		1.871	00:21
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.029	00:25
Breath Volume		1.832	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 06/27/2016 16:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 06/02/2016 21:21:30

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.079	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.098	21:16
Breath Volume		1.113	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.090	21:20
Breath Volume		1.488	21:20
Air Blank		0.000	21:20
Control Test		0.078	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: WOOTEN

DATE/TIME OF TEST: 06/03/2016 01:05:56

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.116	01:01
Breath Volume		2.445	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.116	01:04
Breath Volume		2.113	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 06/07/2016 00:53:16

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.078	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.150	00:47
Breath Volume		1.453	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.144	00:51
Breath Volume		1.515	00:51
Air Blank		0.000	00:52
Control Test		0.076	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 06/27/2016 16:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SVINAREFF

DATE/TIME OF TEST: 06/12/2016 02:31:29

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.078	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.030	02:27
Breath Volume		1.593	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.030	02:30
Breath Volume		1.628	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ - NON-ARREST,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: OTHER
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 06/18/2016 21:25:01

Diagnostics Check		OK	21:18
Air Blank		0.000	21:19
Control Test		0.079	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.167	21:20
Breath Volume		1.480	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.171	21:23
Breath Volume		1.773	21:23
Air Blank		0.000	21:24
Control Test		0.078	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:25

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: TILLER

DATE/TIME OF TEST: 06/24/2016 02:16:07

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:10	
Control Test		0.079	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 1	0.136	02:11	
Breath Volume		2.250	02:11	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 2		02:14	Interferent Detect
Breath Volume		1.683	02:14	
Air Blank		0.000	02:15	
Control Test		0.077	02:15	Interferent Detect
Air Blank		0.000	02:16	

Data Download Date/Time: 06/27/2016 16:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: TILLER

DATE/TIME OF TEST: 06/24/2016 02:23:32

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.078	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.136	02:18
Breath Volume		1.394	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.135	02:22
Breath Volume		1.550	02:22
Air Blank		0.000	02:22
Control Test		0.078	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PELLOT SANTIAGO

DATE/TIME OF TEST: 06/25/2016 02:37:42

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.079	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.134	02:33
Breath Volume		2.476	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.133	02:36
Breath Volume		2.648	02:36
Air Blank		0.000	02:36
Control Test		0.078	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SIDORAVICIUS EIDRIGE

DATE/TIME OF TEST: 06/25/2016 21:45:40

Diagnostics Check		OK	21:39
Air Blank		0.000	21:40
Control Test		0.079	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.144	21:41
Breath Volume		1.671	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.137	21:44
Breath Volume		2.121	21:44
Air Blank		0.000	21:44
Control Test		0.078	21:45
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MOREHEAD

DATE/TIME OF TEST: 06/26/2016 02:26:33

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.175	02:22
Breath Volume		1.789	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.192	02:25
Breath Volume		1.699	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 05/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MULCAHEY

DATE/TIME OF TEST: 06/27/2016 02:31:30

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.138	02:27
Breath Volume		2.902	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.136	02:30
Breath Volume		2.371	02:30
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 06/27/2016 16:26

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ROLEK

DATE/TIME OF TEST: 06/03/2016 02:55:36

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:53	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:54	
Control Test		0.081	02:54	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BARLOW

DATE/TIME OF TEST: 06/03/2016 20:28:27

Diagnostics Check		OK	20:25	
Air Blank		0.000	20:25	
Control Test		0.081	20:26	
Air Blank		0.000	20:26	
Subject Sample	# 1		20:27	Subject Test Refused
Breath Volume		0.000	20:27	
Air Blank		0.000	20:27	
Control Test		0.080	20:27	
Air Blank		0.000	20:28	
Diagnostics Check		OK	20:28	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 06/04/2016 01:47:04

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.081	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.249	01:42
Breath Volume		2.320	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:44
Subject Sample	# 2	0.261	01:45
Breath Volume		2.730	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:47

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SHABEL

DATE/TIME OF TEST: 06/04/2016 03:19:11

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.126	03:14
Breath Volume		1.707	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.127	03:17
Breath Volume		1.988	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MCGRAINE

DATE/TIME OF TEST: 06/04/2016 04:15:14

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.081	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.000	04:10
Breath Volume		1.710	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.000	04:13
Breath Volume		1.906	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 06/04/2016 21:37:58

Diagnostics Check		OK	21:34	
Air Blank		0.000	21:34	
Control Test		0.081	21:34	
Air Blank		0.000	21:35	
Subject Sample	# 1		21:36	Subject Test Refused
Breath Volume		0.000	21:36	
Air Blank		0.000	21:37	
Control Test		0.081	21:37	
Air Blank		0.000	21:37	
Diagnostics Check		OK	21:37	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: DAVIDSON

DATE/TIME OF TEST: 06/04/2016 23:06:06

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.081	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.114	23:01
Breath Volume		2.378	23:01
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.116	23:04
Breath Volume		2.460	23:04
Air Blank		0.000	23:05
Control Test		0.081	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SOCIAS

DATE/TIME OF TEST: 06/05/2016 02:10:42

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.080	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.000	02:06
Breath Volume		3.792	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.000	02:09
Breath Volume		3.992	02:09
Air Blank		0.000	02:09
Control Test		0.080	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: GILFILLAN

DATE/TIME OF TEST: 06/05/2016 03:26:15

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.081	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:24	Subject Test Refused
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Control Test		0.081	03:25	
Air Blank		0.000	03:26	
Diagnostics Check		OK	03:26	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/05/2016 04:22:34

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.196	04:17
Breath Volume		2.523	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.197	04:20
Breath Volume		2.679	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POULIN',DANIELLE,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: POULIN,DANIELLE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: HALLER

DATE/TIME OF TEST: 06/05/2016 17:54:14

Diagnostics Check		OK	17:47
Air Blank		0.000	17:47
Control Test		0.080	17:47
Air Blank		0.000	17:48
Subject Sample	# 1	0.319	17:49
Breath Volume		2.933	17:49
Air Blank		0.000	17:50
Air Blank		0.000	17:51
Subject Sample	# 2	0.322	17:52
Breath Volume		3.058	17:52
Air Blank		0.000	17:53
Control Test		0.079	17:53
Air Blank		0.000	17:54
Diagnostics Check		OK	17:54

Data Download Date/Time: 07/28/2016 22:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEIL,ROBERT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: SOLVERSON

DATE/TIME OF TEST: 06/05/2016 20:10:19

Diagnostics Check		OK	20:03
Air Blank		0.000	20:04
Control Test		0.081	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.044	20:05
Breath Volume		1.464	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.041	20:08
Breath Volume		1.800	20:08
Air Blank		0.000	20:09
Control Test		0.080	20:09
Air Blank		0.000	20:10
Diagnostics Check		OK	20:10

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: ALDRICH

DATE/TIME OF TEST: 06/07/2016 02:44:17

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:41	
Control Test		0.081	02:42	
Air Blank		0.000	02:42	
Subject Sample	# 1		02:42	Subject Test Refused
Breath Volume		0.000	02:42	
Air Blank		0.000	02:43	
Control Test		0.082	02:43	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PIECK

DATE/TIME OF TEST: 06/08/2016 23:19:18

Diagnostics Check		OK	23:15	
Air Blank		0.000	23:16	
Control Test		0.081	23:16	
Air Blank		0.000	23:16	
Subject Sample	# 1		23:17	Subject Test Refused
Breath Volume		0.000	23:17	
Air Blank		0.000	23:18	
Control Test		0.081	23:18	
Air Blank		0.000	23:19	
Diagnostics Check		OK	23:19	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONNELLY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONNELLY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: SARTOS

DATE/TIME OF TEST: 06/09/2016 01:07:51

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.082	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1		01:00	No Sample Provided
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Control Test		0.079	01:07	
Air Blank		0.000	01:07	
Diagnostics Check		OK	01:07	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONNELLY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONNELLY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: SARTO

DATE/TIME OF TEST: 06/09/2016 01:13:19

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.081	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:11	Subject Test Refused
Breath Volume		0.000	01:11	
Air Blank		0.000	01:12	
Control Test		0.081	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 06/09/2016 19:23:45

Diagnostics Check		OK	19:14	
Air Blank		0.000	19:14	
Control Test		0.081	19:14	
Air Blank		0.000	19:15	
Subject Sample	# 1		19:18	No Sample Provided
Breath Volume		0.000	19:18	
Air Blank		0.000	19:19	
Air Blank		0.000	19:21	
Subject Sample	# 2		19:22	Subject Test Refused
Breath Volume		0.000	19:22	
Air Blank		0.000	19:22	
Control Test		0.081	19:23	
Air Blank		0.000	19:23	
Diagnostics Check		OK	19:23	

Data Download Date/Time: 07/28/2016 22:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VINSON,DAVID,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: VINSON,DAVID
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: THIBODEAU

DATE/TIME OF TEST: 06/10/2016 23:10:48

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.081	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.218	23:06
Breath Volume		2.277	23:06
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.218	23:09
Breath Volume		2.636	23:09
Air Blank		0.000	23:09
Control Test		0.080	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:10

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HORNING,PETER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 06/11/2016 00:29:24

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:26	
Control Test		0.082	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Control Test		0.081	00:28	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 06/15/2016 19:23:10

Diagnostics Check		OK	19:16
Air Blank		0.000	19:17
Control Test		0.081	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.318	19:18
Breath Volume		2.542	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.315	19:21
Breath Volume		2.531	19:21
Air Blank		0.000	19:22
Control Test		0.081	19:22
Air Blank		0.000	19:23
Diagnostics Check		OK	19:23

Data Download Date/Time: 07/28/2016 22:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 06/16/2016 22:46:20

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.081	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.184	22:41
Breath Volume		1.933	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.180	22:44
Breath Volume		1.742	22:44
Air Blank		0.000	22:45
Control Test		0.079	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 06/16/2016 23:31:18

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.080	23:29	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:29	Subject Test Refused
Breath Volume		0.000	23:29	
Air Blank		0.000	23:30	
Control Test		0.080	23:30	
Air Blank		0.000	23:31	
Diagnostics Check		OK	23:31	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: RAJAGOPALAN

DATE/TIME OF TEST: 06/17/2016 04:27:04

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.081	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.105	04:22
Breath Volume		2.039	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.106	04:25
Breath Volume		1.281	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:27

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: OLESUK

DATE/TIME OF TEST: 06/17/2016 23:25:42

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.187	23:20
Breath Volume		1.289	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.193	23:24
Breath Volume		1.500	23:24
Air Blank		0.000	23:24
Control Test		0.081	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 07/28/2016 22:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 06/18/2016 00:53:40

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.135	00:48
Breath Volume		1.324	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.140	00:52
Breath Volume		1.660	00:52
Air Blank		0.000	00:52
Control Test		0.082	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: LINEBAUGH

DATE/TIME OF TEST: 06/18/2016 02:26:28

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.089	02:22
Breath Volume		1.742	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.082	02:24
Breath Volume		2.011	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEELER,BRYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/18/2016 03:30:37

Diagnostics Check		OK	03:27	
Air Blank		0.000	03:28	
Control Test		0.081	03:28	
Air Blank		0.000	03:28	
Subject Sample	# 1		03:29	RFI Detect
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank			03:30	RFI Detect
Air Blank		0.000	03:30	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEELER,BRYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 06/18/2016 03:41:09

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:34	
Control Test		0.080	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1	0.213	03:37	
Breath Volume		2.144	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.222	03:40	
Breath Volume		2.960	03:40	
Air Blank			03:41	RFI Detect
Air Blank			03:41	RFI Detect

Data Download Date/Time: 07/28/2016 22:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: FLANNIGAN

DATE/TIME OF TEST: 06/18/2016 04:27:30

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.081	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.174	04:22
Breath Volume		2.523	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.173	04:25
Breath Volume		2.828	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIL,ROBERT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WHITEHURST,WYETH
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 06/18/2016 18:06:35

Diagnostics Check		OK	18:00
Air Blank		0.000	18:00
Control Test		0.081	18:00
Air Blank		0.000	18:01
Subject Sample	# 1	0.000	18:01
Breath Volume		1.808	18:01
Air Blank		0.000	18:02
Air Blank		0.000	18:04
Subject Sample	# 2	0.000	18:05
Breath Volume		1.503	18:05
Air Blank		0.000	18:05
Control Test		0.081	18:05
Air Blank		0.000	18:06
Diagnostics Check		OK	18:06

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: SEMENAS

DATE/TIME OF TEST: 06/18/2016 20:19:03

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.081	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:18	
Control Test		0.081	20:18	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:19	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 06/18/2016 22:57:02

Diagnostics Check		OK	22:54	
Air Blank		0.000	22:54	
Control Test		0.082	22:54	
Air Blank		0.000	22:55	
Subject Sample	# 1		22:55	Subject Test Refused
Breath Volume		0.000	22:55	
Air Blank		0.000	22:56	
Control Test		0.082	22:56	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:57	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: WARDEN

DATE/TIME OF TEST: 06/19/2016 00:59:37

Diagnostics Check		OK	00:56	
Air Blank		0.000	00:56	
Control Test		0.081	00:57	
Air Blank		0.000	00:57	
Subject Sample	# 1		00:58	Subject Test Refused
Breath Volume		0.000	00:58	
Air Blank		0.000	00:58	
Control Test		0.082	00:59	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: HARABIN

DATE/TIME OF TEST: 06/19/2016 02:04:51

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.082	01:59
Air Blank		0.000	01:59
Subject Sample	# 1	0.176	02:00
Breath Volume		1.625	02:00
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.161	02:03
Breath Volume		1.617	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: KALAJDZIC

DATE/TIME OF TEST: 06/19/2016 04:36:38

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.196	04:31
Breath Volume		3.355	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.196	04:35
Breath Volume		3.320	04:35
Air Blank		0.000	04:35
Control Test		0.081	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/28/2016 22:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FALLAHEE,TERRANCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:54

SUBJECT'S LAST NAME: DREYER

DATE/TIME OF TEST: 06/21/2016 12:26:00

Diagnostics Check		OK	12:19
Air Blank		0.000	12:19
Control Test		0.082	12:20
Air Blank		0.000	12:20
Subject Sample	# 1	0.000	12:21
Breath Volume		1.453	12:21
Air Blank		0.000	12:21
Air Blank		0.000	12:23
Subject Sample	# 2	0.000	12:24
Breath Volume		2.113	12:24
Air Blank		0.000	12:25
Control Test		0.082	12:25
Air Blank		0.000	12:25
Diagnostics Check		OK	12:26

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE,SCOTT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOVE,PHILLIP
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: STILES

DATE/TIME OF TEST: 06/21/2016 14:33:37

Diagnostics Check		OK	14:25
Air Blank		0.000	14:25
Control Test		0.082	14:25
Air Blank		0.000	14:26
Subject Sample	# 1	0.000	14:28
Breath Volume		1.699	14:28
Air Blank		0.000	14:28
Air Blank		0.000	14:30
Subject Sample	# 2	0.000	14:32
Breath Volume		1.785	14:32
Air Blank		0.000	14:32
Control Test		0.081	14:33
Air Blank		0.000	14:33
Diagnostics Check		OK	14:33

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BAKER,CRAIG
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 06/22/2016 01:33:31

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.082	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.144	01:28
Breath Volume		2.132	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.142	01:31
Breath Volume		2.136	01:31
Air Blank		0.000	01:32
Control Test		0.081	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHIEBNER,JAMIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: POUCHER

DATE/TIME OF TEST: 06/22/2016 04:00:25

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.082	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1		03:59	Subject Test Refused
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.082	03:59	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE,RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE,RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: WOOMER

DATE/TIME OF TEST: 06/22/2016 17:57:59

Diagnostics Check		OK	17:51
Air Blank		0.000	17:51
Control Test		0.081	17:52
Air Blank		0.000	17:52
Subject Sample	# 1	0.190	17:53
Breath Volume		1.472	17:53
Air Blank		0.000	17:53
Air Blank		0.000	17:55
Subject Sample	# 2	0.192	17:56
Breath Volume		2.105	17:56
Air Blank		0.000	17:57
Control Test		0.080	17:57
Air Blank		0.000	17:57
Diagnostics Check		OK	17:57

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: GREALISH

DATE/TIME OF TEST: 06/23/2016 00:24:36

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.000	00:20
Breath Volume		2.070	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.000	00:23
Breath Volume		1.929	00:23
Air Blank		0.000	00:23
Control Test		0.082	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: SICKEL

DATE/TIME OF TEST: 06/23/2016 01:18:39

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.081	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Control Test		0.082	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/23/2016 01:34:23

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.082	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.082	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 06/23/2016 05:57:12

Diagnostics Check		OK	05:54	
Air Blank		0.000	05:54	
Control Test		0.081	05:55	
Air Blank		0.000	05:55	
Subject Sample	# 1		05:55	Subject Test Refused
Breath Volume		0.000	05:55	
Air Blank		0.000	05:56	
Control Test		0.081	05:56	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 06/23/2016 22:42:20

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:39	
Control Test		0.082	22:40	
Air Blank		0.000	22:40	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:41	
Control Test		0.082	22:41	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 07/28/2016 22:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,STEVEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: JAMES,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: SALAY

DATE/TIME OF TEST: 06/23/2016 23:54:45

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:49
Subject Sample	# 1	0.171	23:50
Breath Volume		2.804	23:50
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.169	23:53
Breath Volume		3.511	23:53
Air Blank		0.000	23:53
Control Test		0.081	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GOTTWALD

DATE/TIME OF TEST: 06/24/2016 00:53:05

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.082	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.157	00:48
Breath Volume		1.781	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.158	00:51
Breath Volume		2.292	00:51
Air Blank		0.000	00:52
Control Test		0.081	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 06/24/2016 02:27:35

Diagnostics Check		OK	02:19
Air Blank		0.000	02:20
Control Test		0.081	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.160	02:22
Breath Volume		1.632	02:22
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.159	02:25
Breath Volume		1.445	02:25
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 07/28/2016 22:50

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 06/24/2016 02:43:23

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.082	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.183	02:38
Breath Volume		3.343	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.200	02:41
Breath Volume		2.980	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: KINSMAN

DATE/TIME OF TEST: 06/24/2016 04:25:56

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.082	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.265	04:21
Breath Volume		3.753	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.249	04:24
Breath Volume		2.988	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: HASANI

DATE/TIME OF TEST: 06/24/2016 21:40:52

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.081	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.077	21:36
Breath Volume		2.359	21:36
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.074	21:39
Breath Volume		2.273	21:39
Air Blank		0.000	21:39
Control Test		0.081	21:40
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/24/2016 22:26:06

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.081	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.000	22:21
Breath Volume		1.769	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.000	22:24
Breath Volume		2.023	22:24
Air Blank		0.000	22:25
Control Test		0.082	22:25
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 07/28/2016 22:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 06/25/2016 00:13:38

Diagnostics Check		OK	00:06
Air Blank		0.000	00:07
Control Test		0.081	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.135	00:09
Breath Volume		1.660	00:09
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 2	0.130	00:12
Breath Volume		1.449	00:12
Air Blank		0.000	00:12
Control Test		0.081	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: FERRAND

DATE/TIME OF TEST: 06/25/2016 00:52:36

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.082	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.197	00:47
Breath Volume		1.390	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.192	00:51
Breath Volume		2.281	00:51
Air Blank		0.000	00:51
Control Test		0.081	00:52
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: BAYS

DATE/TIME OF TEST: 06/25/2016 01:42:39

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.026	01:38
Breath Volume		2.671	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.024	01:41
Breath Volume		2.597	01:41
Air Blank		0.000	01:41
Control Test		0.081	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: DIPASQUALE

DATE/TIME OF TEST: 06/25/2016 04:26:21

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.173	04:20
Breath Volume		2.660	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.172	04:24
Breath Volume		2.535	04:24
Air Blank		0.000	04:25
Control Test		0.081	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 07/28/2016 22:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SWISHER,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:51

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 06/26/2016 14:20:50

Diagnostics Check	OK	14:20	
Air Blank		14:20	RFI Detect
Air Blank	0.000	14:20	

Data Download Date/Time: 07/28/2016 22:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SWISHER,JEFFREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:51

SUBJECT'S LAST NAME: HUYNH

DATE/TIME OF TEST: 06/26/2016 14:34:10

Diagnostics Check		OK	14:22	
Air Blank		0.000	14:22	
Control Test		0.082	14:23	
Air Blank		0.000	14:23	
Subject Sample	# 1	0.182	14:26	Volume Not Met
Breath Volume		0.558	14:26	
Air Blank		0.000	14:27	
Air Blank		0.000	14:29	
Subject Sample	# 2	0.173	14:32	Volume Not Met
Breath Volume		0.789	14:32	
Air Blank		0.000	14:33	
Control Test		0.080	14:33	
Air Blank		0.000	14:34	
Diagnostics Check		OK	14:34	

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WELSHANS,DEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 06/27/2016 04:08:41

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.228	04:03
Breath Volume		2.031	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:05
Subject Sample	# 2	0.237	04:07
Breath Volume		1.175	04:07
Air Blank		0.000	04:07
Control Test		0.080	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELSHANS,DEAN,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WELSHANS,DEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 06/27/2016 07:56:11

Diagnostics Check		OK	07:49
Air Blank		0.000	07:49
Control Test		0.081	07:49
Air Blank		0.000	07:50
Subject Sample	# 1	0.165	07:51
Breath Volume		1.558	07:51
Air Blank		0.000	07:51
Air Blank		0.000	07:53
Subject Sample	# 2	0.175	07:54
Breath Volume		2.175	07:54
Air Blank		0.000	07:55
Control Test		0.081	07:55
Air Blank		0.000	07:56
Diagnostics Check		OK	07:56

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: LOFTIS-DORN

DATE/TIME OF TEST: 06/29/2016 01:48:18

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.082	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.090	01:43
Breath Volume		2.449	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.090	01:46
Breath Volume		1.546	01:46
Air Blank		0.000	01:47
Control Test		0.081	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOODNIGHT,MICHAEL,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GOODNIGHT,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 14:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:12

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 06/29/2016 15:30:37

Diagnostics Check		OK	15:22
Air Blank		0.000	15:23
Control Test		0.080	15:23
Air Blank		0.000	15:24
Subject Sample	# 1	0.306	15:25
Breath Volume		1.769	15:25
Air Blank		0.000	15:26
Air Blank		0.000	15:28
Subject Sample	# 2	0.296	15:28
Breath Volume		1.601	15:28
Air Blank		0.000	15:29
Control Test		0.079	15:30
Air Blank		0.000	15:30
Diagnostics Check		OK	15:30

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 06/29/2016 22:14:37

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.081	22:08
Air Blank		0.000	22:09
Subject Sample	# 1	0.207	22:09
Breath Volume		2.449	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.204	22:13
Breath Volume		2.699	22:13
Air Blank		0.000	22:13
Control Test		0.080	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: FLORO

DATE/TIME OF TEST: 06/30/2016 03:39:06

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.199	03:34
Breath Volume		3.406	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.200	03:37
Breath Volume		1.777	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/28/2016 22:51

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 06/04/2016 02:09:38

Diagnostics Check		OK	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.131	02:02
Breath Volume		2.046	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.124	02:08
Breath Volume		1.664	02:08
Air Blank		0.000	02:08
Control Test		0.078	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 07/28/2016 20:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ALLAIRE,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: HEFFNER

DATE/TIME OF TEST: 06/21/2016 02:10:08

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.079	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 07/28/2016 20:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LONGGOOD,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SHAFFER

DATE/TIME OF TEST: 06/21/2016 03:24:35

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.208	03:20
Breath Volume		2.410	03:20
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.211	03:23
Breath Volume		2.191	03:23
Air Blank		0.000	03:23
Control Test		0.079	03:24
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 07/28/2016 20:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLASINI,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: TOOME

DATE/TIME OF TEST: 06/23/2016 05:32:30

Diagnostics Check		OK	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:27
Subject Sample	# 1	0.098	05:27
Breath Volume		1.628	05:27
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.101	05:31
Breath Volume		2.031	05:31
Air Blank		0.000	05:31
Control Test		0.079	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 07/28/2016 20:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARR,JOHNATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 06/27/2016 03:28:57

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.079	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 07/28/2016 20:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCDOWELL,DEREK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: STREET

DATE/TIME OF TEST: 06/27/2016 04:59:26

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.131	04:55
Breath Volume		1.691	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.131	04:57
Breath Volume		1.296	04:57
Air Blank		0.000	04:58
Control Test		0.080	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 07/28/2016 20:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVARTS,JASON,R
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FALLAHEE,TERRANCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 10:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:31

SUBJECT'S LAST NAME: KEENE

DATE/TIME OF TEST: 06/30/2016 10:57:59

Diagnostics Check		OK	10:51
Air Blank		0.000	10:52
Control Test		0.080	10:52
Air Blank		0.000	10:52
Subject Sample	# 1	0.000	10:53
Breath Volume		2.437	10:53
Air Blank		0.000	10:54
Air Blank		0.000	10:56
Subject Sample	# 2	0.000	10:56
Breath Volume		2.398	10:56
Air Blank		0.000	10:57
Control Test		0.080	10:57
Air Blank		0.000	10:57
Diagnostics Check		OK	10:57

Data Download Date/Time: 07/28/2016 20:14

Printed: 11 June 2018

COBRA ©1997-2014 CMI, Inc.