Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN HELDEN,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MOYE,JOSEPH
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/04/2016 00:46:34

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.077	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.232	00:41
Breath Volume		2.550	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.229	00:44
Breath Volume		2.277	00:44
Air Blank		0.000	00:45
Control Test		0.076	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 07/06/2016 10:09

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TOSLLUKU,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: VASQUEZ-SANCHEZ

DATE/TIME OF TEST: 07/04/2016 17:22:08

Diagnostics Check		OK	17:15
Air Blank		0.000	17:15
Control Test		0.077	17:15
Air Blank		0.000	17:16
Subject Sample	# 1	0.225	17:17
Breath Volume		2.007	17:17
Air Blank		0.000	17:17
Air Blank		0.000	17:19
Subject Sample	# 2	0.216	17:20
Breath Volume		1.515	17:20
Air Blank		0.000	17:21
Control Test		0.077	17:21
Air Blank		0.000	17:22
Diagnostics Check		OK	17:22

Data Download Date/Time: 07/06/2016 10:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001722

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RIDENOUR,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 07/05/2016 21:50:08

Diagnostics Check		OK	21:43
Air Blank		0.000	21:44
Control Test		0.077	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.134	21:45
Breath Volume		1.968	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:47
Subject Sample	# 2	0.135	21:48
Breath Volume		2.605	21:48
Air Blank		0.000	21:49
Control Test		0.077	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 07/06/2016 10:09

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: ROBINETT

DATE/TIME OF TEST: 07/01/2016 23:47:01

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.078	23:44	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:45	Subject Test Refused
Breath Volume		0.000	23:45	
Air Blank		0.000	23:46	
Control Test		0.078	23:46	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 07/07/2016 08:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORICE,DOMINIC,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SORICE,DOMINIC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 07/02/2016 01:37:21

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:34	
Control Test		0.078	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:35	Subject Test Refused
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Control Test		0.079	01:36	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 07/07/2016 08:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLEY,CPL,S
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GOSIER,OFC
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: PALO

DATE/TIME OF TEST: 07/02/2016 05:41:20

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:37	
Control Test		0.078	05:37	
Air Blank		0.000	05:38	
Subject Sample	# 1		05:39	Subject Test Refused
Breath Volume		0.000	05:39	
Air Blank		0.000	05:40	
Control Test		0.077	05:40	
Air Blank		0.000	05:41	
Diagnostics Check		OK	05:41	

Data Download Date/Time: 07/07/2016 08:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: SZMAJSER

DATE/TIME OF TEST: 07/02/2016 22:15:11

Diagnostics Check		OK	22:08
Air Blank		0.000	22:08
Control Test		0.079	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.000	22:10
Breath Volume		2.386	22:10
Air Blank		0.000	22:11
Air Blank		0.000	22:13
Subject Sample	# 2	0.000	22:13
Breath Volume		2.359	22:13
Air Blank		0.000	22:14
Control Test		0.078	22:14
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 07/07/2016 08:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,JOHN,D
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 07/04/2016 02:19:10

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.141	02:14
Breath Volume		1.429	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.136	02:17
Breath Volume		1.378	02:17
Air Blank		0.000	02:18
Control Test		0.077	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 07/07/2016 08:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCHENEY,WESLEY,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DUCHENEY,WESLEY
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: WHITELEY

DATE/TIME OF TEST: 07/08/2016 23:51:34

Diagnostics Check		OK	23:44
Air Blank		0.000	23:44
Control Test		0.078	23:45
Air Blank		0.000	23:45
Subject Sample	# 1	0.311	23:46
Breath Volume		1.996	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.311	23:49
Breath Volume		2.578	23:49
Air Blank		0.000	23:50
Control Test		0.076	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: NAWROCKI

DATE/TIME OF TEST: 07/09/2016 02:29:21

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.103	02:24
Breath Volume		2.390	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.100	02:27
Breath Volume		1.878	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,RONALD,W
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: MCKENZIE,RONALD
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:57

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/13/2016 10:30:00

Diagnostics Check		OK	10:18	
Air Blank		0.000	10:19	
Control Test		0.076	10:19	
Air Blank		0.000	10:19	
Subject Sample	# 1	0.232	10:23	Volume Not Met
Breath Volume		0.421	10:23	
Air Blank		0.000	10:24	
Air Blank		0.000	10:25	
Subject Sample	# 2		10:28	No Sample Provided
Breath Volume		0.000	10:28	
Air Blank		0.000	10:29	
Control Test		0.071	10:29	Control Outside Tolerance
Air Blank		0.000	10:30	

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASSETT,STEVEN,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: NAUMANN,NATHAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: NIGRO

DATE/TIME OF TEST: 07/15/2016 23:06:24

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.078	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.000	23:02
Breath Volume		1.632	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.000	23:05
Breath Volume		1.875	23:05
Air Blank		0.000	23:05
Control Test		0.078	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORICE,DOMINIC,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: FICCO,CHAD
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 07/16/2016 05:13:04

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.078	05:10	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:11	Subject Test Refused
Breath Volume		0.000	05:11	
Air Blank		0.000	05:12	
Control Test		0.078	05:12	
Air Blank		0.000	05:12	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: OHM

DATE/TIME OF TEST: 07/18/2016 00:08:31

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.079	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.116	00:03
Breath Volume		1.109	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.112	00:06
Breath Volume		1.406	00:06
Air Blank		0.000	00:07
Control Test		0.077	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KOLBICKA,LAWRENCE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: PENNA

DATE/TIME OF TEST: 07/19/2016 13:46:49

Diagnostics Check		OK	13:38
Air Blank		0.000	13:39
Control Test		0.079	13:39
Air Blank		0.000	13:40
Subject Sample	# 1	0.033	13:42
Breath Volume		2.402	13:42
Air Blank		0.000	13:43
Air Blank		0.000	13:44
Subject Sample	# 2	0.032	13:45
Breath Volume		2.585	13:45
Air Blank		0.000	13:45
Control Test		0.078	13:46
Air Blank		0.000	13:46
Diagnostics Check		OK	13:46

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: O STEEN

DATE/TIME OF TEST: 07/22/2016 01:51:22

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:48	
Control Test		0.078	01:49	
Air Blank		0.000	01:49	
Subject Sample	# 1		01:50	Subject Test Refused
Breath Volume		0.000	01:50	
Air Blank		0.000	01:50	
Control Test		0.078	01:50	
Air Blank		0.000	01:51	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 08/05/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COUCH,WILLIAM,K
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: DELAY,A
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BRIZZIE

DATE/TIME OF TEST: 07/22/2016 19:22:13

Diagnostics Check		OK	19:13
Air Blank		0.000	19:14
Control Test		0.078	19:14
Air Blank		0.000	19:14
Subject Sample	# 1	0.183	19:17
Breath Volume		1.156	19:17
Air Blank		0.000	19:17
Air Blank		0.000	19:19
Subject Sample	# 2	0.170	19:20
Breath Volume		1.148	19:20
Air Blank		0.000	19:21
Control Test		0.075	19:21
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 08/05/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: FALTO

DATE/TIME OF TEST: 07/23/2016 02:43:43

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.078	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1		02:35	No Sample Provided
Breath Volume		0.183	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.089	02:38	
Breath Volume		2.015	02:38	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 3	0.095	02:42	
Breath Volume		1.570	02:42	
Air Blank		0.000	02:42	
Control Test		0.078	02:43	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSEN,STEPHEN,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JACOBSEN,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: JUDAH

DATE/TIME OF TEST: 07/23/2016 04:10:01

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.133	04:05
Breath Volume		3.078	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:07
Subject Sample	# 2	0.132	04:08
Breath Volume		2.488	04:08
Air Blank		0.000	04:09
Control Test		0.077	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:10

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/23/2016 23:55:36

Diagnostics Check		OK	23:47
Air Blank		0.000	23:47
Control Test		0.078	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.149	23:50
Breath Volume		1.570	23:50
Air Blank		0.000	23:51
Air Blank		0.000	23:52
Subject Sample	# 2	0.160	23:53
Breath Volume		1.855	23:53
Air Blank		0.000	23:54
Control Test		0.076	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 08/05/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,OFFICER,B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GILLIS,R
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: SCHUSTER

DATE/TIME OF TEST: 07/27/2016 02:59:00

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.077	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.077	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 08/05/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KOLBICKA,LAWRENCE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/29/2016 19:12:33

Diagnostics Check		OK	19:09	
Air Blank		0.000	19:10	
Control Test		0.078	19:10	
Air Blank		0.000	19:11	
Subject Sample	# 1		19:11	Subject Test Refused
Breath Volume		0.000	19:11	
Air Blank		0.000	19:11	
Control Test		0.077	19:11	
Air Blank		0.000	19:12	
Diagnostics Check		OK	19:12	

Data Download Date/Time: 08/05/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: VANDERVEER

DATE/TIME OF TEST: 07/03/2016 03:11:02

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.198	03:06
Breath Volume		3.273	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.192	03:09
Breath Volume		3.066	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:11

Data Download Date/Time: 07/28/2016 20:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MCAVOY

DATE/TIME OF TEST: 07/04/2016 03:45:21

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.087	03:40
Breath Volume		2.855	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.090	03:43
Breath Volume		3.550	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 07/28/2016 20:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREITAS,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 07/04/2016 18:51:28

Diagnostics Check		OK	18:40	
Air Blank		0.000	18:40	
Control Test		0.079	18:40	
Air Blank		0.000	18:41	
Subject Sample	# 1		18:44	No Sample Provided
Breath Volume		0.000	18:44	
Air Blank		0.000	18:44	
Air Blank		0.000	18:46	
Subject Sample	# 2		18:50	No Sample Provided
Breath Volume		0.000	18:50	
Air Blank		0.000	18:50	
Control Test		0.079	18:50	
Air Blank		0.000	18:51	
Diagnostics Check		OK	18:51	

Data Download Date/Time: 07/28/2016 20:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DANNER,CHASE
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BAMBER

DATE/TIME OF TEST: 07/06/2016 20:12:35

Diagnostics Check		OK	20:06
Air Blank		0.000	20:06
Control Test		0.079	20:06
Air Blank		0.000	20:07
Subject Sample	# 1	0.124	20:07
Breath Volume		2.359	20:07
Air Blank		0.000	20:08
Air Blank		0.000	20:10
Subject Sample	# 2	0.123	20:11
Breath Volume		1.906	20:11
Air Blank		0.000	20:11
Control Test		0.080	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 07/28/2016 20:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2016 05:42:05

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:36
Subject Sample	# 1	0.209	05:37
Breath Volume		3.433	05:37
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.206	05:40
Breath Volume		3.144	05:40
Air Blank		0.000	05:41
Control Test		0.079	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 07/28/2016 20:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURROWS,ALBERT,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2016 09:22:00

Diagnostics Check		OK	09:15
Air Blank		0.000	09:15
Control Test		0.079	09:16
Air Blank		0.000	09:16
Subject Sample	# 1	0.146	09:17
Breath Volume		2.644	09:17
Air Blank		0.000	09:17
Air Blank		0.000	09:19
Subject Sample	# 2	0.148	09:20
Breath Volume		3.105	09:20
Air Blank		0.000	09:21
Control Test		0.078	09:21
Air Blank		0.000	09:21
Diagnostics Check		OK	09:22

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURROWS,ALBERT,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2016 14:07:43

Diagnostics Check		OK	14:00
Air Blank		0.000	14:01
Control Test		0.079	14:01
Air Blank		0.000	14:02
Subject Sample	# 1	0.075	14:03
Breath Volume		1.746	14:03
Air Blank		0.000	14:03
Air Blank		0.000	14:05
Subject Sample	# 2	0.075	14:06
Breath Volume		2.085	14:06
Air Blank		0.000	14:06
Control Test		0.079	14:07
Air Blank		0.000	14:07
Diagnostics Check		OK	14:07

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURROWS,ALBERT,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2016 15:04:36

Diagnostics Check		OK	14:58
Air Blank		0.000	14:58
Control Test		0.079	14:59
Air Blank		0.000	14:59
Subject Sample	# 1	0.066	15:00
Breath Volume		1.855	15:00
Air Blank		0.000	15:00
Air Blank		0.000	15:02
Subject Sample	# 2	0.062	15:03
Breath Volume		2.312	15:03
Air Blank		0.000	15:03
Control Test		0.079	15:04
Air Blank		0.000	15:04
Diagnostics Check		OK	15:04

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURROWS,ALBERT,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2016 16:35:21

Diagnostics Check		OK	16:29
Air Blank		0.000	16:29
Control Test		0.080	16:29
Air Blank		0.000	16:30
Subject Sample	# 1	0.044	16:30
Breath Volume		1.875	16:30
Air Blank		0.000	16:31
Air Blank		0.000	16:33
Subject Sample	# 2	0.043	16:33
Breath Volume		2.550	16:33
Air Blank		0.000	16:34
Control Test		0.079	16:34
Air Blank		0.000	16:35
Diagnostics Check		OK	16:35

Data Download Date/Time: 07/28/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURROWS,ALBERT,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CCSO,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/07/2016 17:36:49

Diagnostics Check		OK	17:30
Air Blank		0.000	17:30
Control Test		0.079	17:31
Air Blank		0.000	17:31
Subject Sample	# 1	0.031	17:32
Breath Volume		2.191	17:32
Air Blank		0.000	17:32
Air Blank		0.000	17:34
Subject Sample	# 2	0.031	17:35
Breath Volume		2.117	17:35
Air Blank		0.000	17:35
Control Test		0.079	17:36
Air Blank		0.000	17:36
Diagnostics Check		OK	17:36

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAFFARAS,CELINA,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GIRARD,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: CONN

DATE/TIME OF TEST: 07/08/2016 20:59:11

Diagnostics Check		OK	20:53
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:54
Subject Sample	# 1	0.153	20:54
Breath Volume		2.660	20:54
Air Blank		0.000	20:55
Air Blank		0.000	20:57
Subject Sample	# 2	0.150	20:57
Breath Volume		2.945	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHIABLE,MARK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MCEWEN

DATE/TIME OF TEST: 07/10/2016 01:45:29

Diagnostics Check		OK	01:33	
Air Blank		0.000	01:34	
Control Test		0.080	01:34	
Air Blank		0.000	01:34	
Subject Sample	# 1	0.107	01:38	Volume Not Met
Breath Volume		0.515	01:38	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 2	0.118	01:40	
Breath Volume		2.429	01:40	
Air Blank		0.000	01:41	
Air Blank		0.000	01:43	
Subject Sample	# 3	0.120	01:44	
Breath Volume		1.363	01:44	
Air Blank		0.000	01:44	
Control Test		0.078	01:44	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 07/28/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTON,JOHN,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: URSITTI,VINCENT
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/13/2016 00:55:08

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.171	00:50
Breath Volume		1.472	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.158	00:53
Breath Volume		1.679	00:53
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATTON,JOHN,G
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PUTMAN,BRIAN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BOLDUC

DATE/TIME OF TEST: 07/13/2016 02:55:26

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.125	02:50
Breath Volume		1.718	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.122	02:53
Breath Volume		1.597	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: BALTIMORE,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/16/2016 02:19:28

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.195	02:14
Breath Volume		3.925	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.179	02:18
Breath Volume		1.808	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PUTMAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GILLEY

DATE/TIME OF TEST: 07/17/2016 01:11:54

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.073	01:07
Breath Volume		3.898	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.069	01:10
Breath Volume		3.621	01:10
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRELL,THOMAS,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GONZALEZ,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 07/17/2016 01:31:38

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.153	01:27
Breath Volume		1.863	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.137	01:30
Breath Volume		1.789	01:30
Air Blank		0.000	01:30
Control Test		0.079	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,J,J
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CAMPBELL,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/17/2016 15:17:53

Diagnostics Check		OK	15:09	
Air Blank		0.000	15:09	
Control Test		0.079	15:10	
Air Blank		0.000	15:10	
Subject Sample	# 1		15:13	No Sample Provided
Breath Volume		0.000	15:13	
Air Blank		0.000	15:14	
Air Blank		0.000	15:16	
Subject Sample	# 2		15:16	Subject Test Refused
Breath Volume		0.000	15:16	
Air Blank		0.000	15:16	
Control Test		0.080	15:17	
Air Blank		0.000	15:17	
Diagnostics Check		OK	15:17	

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COOPER,DAVID
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 07/20/2016 02:03:21

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:49	
Control Test		0.079	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1	0.236	01:53	Volume Not Met
Breath Volume		0.269	01:53	
Air Blank		0.000	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 2	0.201	01:56	
Breath Volume		2.632	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 3	0.205	02:01	
Breath Volume		2.191	02:01	
Air Blank		0.000	02:02	
Control Test		0.078	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARIAS,MONICA,S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TRUFAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: WENDT

DATE/TIME OF TEST: 07/20/2016 15:51:50

Diagnostics Check		OK	15:40	
Air Blank		0.000	15:40	
Control Test		0.077	15:41	
Air Blank		0.000	15:41	
Subject Sample	# 1	0.158	15:44	Volume Not Met
Breath Volume		0.125	15:44	
Air Blank		0.000	15:45	
Air Blank		0.000	15:47	
Subject Sample	# 2		15:50	No Sample Provided
Breath Volume		0.000	15:50	
Air Blank		0.000	15:50	
Control Test		0.076	15:51	
Air Blank		0.000	15:51	
Diagnostics Check		OK	15:51	

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: TRENTINI

DATE/TIME OF TEST: 07/21/2016 03:50:17

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:34	
Control Test		0.080	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.220	03:38	Volume Not Met
Breath Volume		0.085	03:38	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.228	03:43	
Breath Volume		1.167	03:43	
Air Blank		0.000	03:43	
Air Blank		0.000	03:45	
Subject Sample	# 3		03:48	No Sample Provided
Breath Volume		0.000	03:48	
Air Blank		0.000	03:49	
Control Test		0.078	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MONAHAN,J
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 07/22/2016 23:26:15

Diagnostics Check		OK	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Subject Sample	# 1	0.090	23:21
Breath Volume		1.156	23:21
Air Blank		0.000	23:21
Air Blank		0.000	23:23
Subject Sample	# 2	0.091	23:24
Breath Volume		2.007	23:24
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTOPHER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 07/23/2016 02:43:37

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.154	02:38
Breath Volume		1.523	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.140	02:42
Breath Volume		1.691	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RICHARDSON,B
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WEINER

DATE/TIME OF TEST: 07/23/2016 23:18:14

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.079	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.000	23:13
Breath Volume		2.117	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.000	23:16
Breath Volume		2.328	23:16
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ALLEN,MARTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: ALBERTSON

DATE/TIME OF TEST: 07/24/2016 20:11:06

Diagnostics Check		OK	20:03
Air Blank		0.000	20:04
Control Test		0.080	20:04
Air Blank		0.000	20:05
Subject Sample	# 1	0.323	20:06
Breath Volume		1.898	20:06
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.327	20:09
Breath Volume		2.082	20:09
Air Blank		0.000	20:10
Control Test		0.078	20:10
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 07/28/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001739

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COOPER,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/25/2016 01:20:49

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.079	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1	0.060	01:12	Volume Not Met
Breath Volume		0.863	01:12	
Air Blank		0.000	01:12	
Air Blank		0.000	01:14	
Subject Sample	# 2	0.065	01:15	
Breath Volume		1.847	01:15	
Air Blank		0.000	01:16	
Air Blank		0.000	01:18	
Subject Sample	# 3	0.063	01:19	
Breath Volume		1.570	01:19	
Air Blank		0.000	01:19	
Control Test		0.079	01:20	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 07/28/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: MENDOZA-BODINES

DATE/TIME OF TEST: 07/01/2016 01:46:13

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.117	01:41
Breath Volume		3.027	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.111	01:44
Breath Volume		2.906	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHEO,FRANCISCO,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: MICHEO,FRANCISCO
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ALBERTS

DATE/TIME OF TEST: 07/01/2016 21:28:57

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:26	
Control Test		0.081	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 1		21:27	RFI Detect
Breath Volume		0.000	21:27	
Air Blank		0.000	21:28	
Control Test		0.080	21:28	
Air Blank		0.000	21:28	
Diagnostics Check		OK	21:28	

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINKMAN,KELLY,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: BRINKMAN,KELLY
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: MIRDZIC

DATE/TIME OF TEST: 07/02/2016 23:37:44

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.169	23:29
Breath Volume		1.898	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.190	23:32
Breath Volume		2.472	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 3	0.192	23:36
Breath Volume		3.976	23:36
Air Blank		0.000	23:36
Control Test		0.078	23:37
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 08/07/2016 23:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/05/2016 21:51:53

Diagnostics Check	OK	21:51	
Air Blank		21:51	RFI Detect
Air Blank	0.000	21:51	

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ENGLAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/05/2016 22:05:05

Diagnostics Check		OK	21:53	
Air Blank		0.000	21:53	
Control Test		0.080	21:54	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:57	No Sample Provided
Breath Volume		0.000	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 2	0.198	22:03	Volume Not Met
Breath Volume		0.171	22:03	
Air Blank		0.000	22:04	
Control Test		0.078	22:04	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:05	

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MICHEO,FRANCISCO,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: VENEZIA,STEVEN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/06/2016 22:24:07

Diagnostics Check		OK	22:16
Air Blank		0.000	22:16
Control Test		0.081	22:17
Air Blank		0.000	22:17
Subject Sample	# 1	0.235	22:19
Breath Volume		1.941	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:21
Subject Sample	# 2	0.233	22:22
Breath Volume		1.628	22:22
Air Blank		0.000	22:23
Control Test		0.079	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMONDS,RICHARD,J
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: EDMONDS,RICHARD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/07/2016 01:07:23

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.175	01:02
Breath Volume		1.144	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.172	01:05
Breath Volume		1.199	01:05
Air Blank		0.000	01:06
Control Test		0.077	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RESLER,OFC.K.,G
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: GOSIER,OFC.M.
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ADANI

DATE/TIME OF TEST: 07/11/2016 03:32:55

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:23	
Control Test		0.080	03:24	
Air Blank		0.000	03:24	
Subject Sample	# 1		03:27	No Sample Provided
Breath Volume		0.000	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:30	
Subject Sample	# 2		03:31	Subject Test Refused
Breath Volume		0.000	03:31	
Air Blank		0.000	03:31	
Control Test		0.080	03:32	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: PLENSKE

DATE/TIME OF TEST: 07/15/2016 19:09:01

Diagnostics Check		OK	19:02
Air Blank		0.000	19:03
Control Test		0.080	19:03
Air Blank		0.000	19:04
Subject Sample	# 1	0.000	19:04
Breath Volume		2.609	19:04
Air Blank		0.000	19:05
Air Blank		0.000	19:07
Subject Sample	# 2	0.000	19:07
Breath Volume		2.937	19:07
Air Blank		0.000	19:08
Control Test		0.079	19:08
Air Blank		0.000	19:08
Diagnostics Check		OK	19:09

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUGENT,DAVID,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CARR

DATE/TIME OF TEST: 07/16/2016 01:04:03

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.080	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:57	No Sample Provided
Breath Volume		0.000	00:57	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2		01:02	No Sample Provided
Breath Volume		0.000	01:02	
Air Blank		0.000	01:03	
Control Test		0.080	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:04	

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001740

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,CHRIS,N
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HENNIS,J
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: DELEON

DATE/TIME OF TEST: 07/16/2016 18:19:05

Diagnostics Check		OK	18:10
Air Blank		0.000	18:10
Control Test		0.079	18:10
Air Blank		0.000	18:11
Subject Sample	# 1	0.365	18:13
Breath Volume		2.042	18:13
Air Blank		0.000	18:14
Air Blank		0.000	18:15
Subject Sample	# 2	0.366	18:17
Breath Volume		1.777	18:17
Air Blank		0.000	18:18
Control Test		0.076	18:18
Air Blank		0.000	18:19
Diagnostics Check		OK	18:19

Data Download Date/Time: 08/07/2016 23:10

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BIA,S.W.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: VICKERY

DATE/TIME OF TEST: 07/01/2016 00:02:31

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.080	23:57
Air Blank		0.000	23:57
Subject Sample	# 2	0.000	23:58
Breath Volume		3.558	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.000	00:01
Breath Volume		3.644	00:01
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 07/01/2016 05:41

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MARLOWE

DATE/TIME OF TEST: 07/01/2016 22:57:49

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.080	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.272	22:53
Breath Volume		1.296	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.253	22:56
Breath Volume		1.550	22:56
Air Blank		0.000	22:56
Control Test		0.078	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 07/08/2016 10:20

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/02/2016 03:16:26

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.080	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.229	03:11
Breath Volume		1.984	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.223	03:14
Breath Volume		1.531	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 07/08/2016 10:20

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: BALDWIN

DATE/TIME OF TEST: 07/03/2016 03:01:41

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.173	02:57
Breath Volume		2.390	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.157	03:00
Breath Volume		1.253	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 07/08/2016 10:20

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: LOWE

DATE/TIME OF TEST: 07/05/2016 23:55:41

Diagnostics Check		OK	23:48
Air Blank		0.000	23:48
Control Test		0.080	23:49
Air Blank		0.000	23:49
Subject Sample	# 1	0.035	23:51
Breath Volume		1.363	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 2	0.040	23:54
Breath Volume		2.226	23:54
Air Blank		0.000	23:54
Control Test		0.079	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 07/08/2016 10:20

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/06/2016 00:43:35

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.079	00:38
Air Blank		0.000	00:38
Subject Sample	# 1	0.191	00:39
Breath Volume		4.691	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.189	00:41
Breath Volume		4.636	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 07/08/2016 10:20

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PFORTMILLER,B.W.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 07/06/2016 06:56:07

Diagnostics Check		OK	06:52	
Air Blank		0.000	06:52	
Control Test		0.080	06:53	
Air Blank		0.000	06:53	
Subject Sample	# 1		06:54	Subject Test Refused
Breath Volume		0.000	06:54	
Air Blank		0.000	06:55	
Control Test		0.080	06:55	
Air Blank		0.000	06:56	
Diagnostics Check		OK	06:56	

Data Download Date/Time: 07/08/2016 10:21

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALL,C.A.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HILLARD

DATE/TIME OF TEST: 07/07/2016 02:01:32

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.080	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:00	
Control Test		0.080	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 07/08/2016 10:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 07/07/2016 20:31:34

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:26
Subject Sample	# 1	0.000	20:27
Breath Volume		2.777	20:27
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.000	20:30
Breath Volume		1.664	20:30
Air Blank		0.000	20:30
Control Test		0.080	20:30
Air Blank		0.000	20:31
Diagnostics Check		OK	20:31

Data Download Date/Time: 07/08/2016 10:21

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MONTGOMERY,K.R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MELTZER

DATE/TIME OF TEST: 07/10/2016 01:52:28

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.133	01:47
Breath Volume		1.347	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.137	01:50
Breath Volume		1.937	01:50
Air Blank		0.000	01:51
Control Test		0.079	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SUTHERLAND,R.A.
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/11/2016 01:53:04

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:49	
Control Test		0.080	01:49	
Air Blank		0.000	01:50	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Control Test		0.080	01:52	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J.T.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/11/2016 03:28:50

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.081	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.079	03:28	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TERRELL,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: STOCKS

DATE/TIME OF TEST: 07/15/2016 00:32:58

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:29	
Control Test		0.080	00:29	
Air Blank		0.000	00:29	
Subject Sample	# 1		00:31	Subject Test Refused
Breath Volume		0.000	00:31	
Air Blank		0.000	00:32	
Control Test		0.080	00:32	
Air Blank		0.000	00:32	
Diagnostics Check		OK	00:32	

Data Download Date/Time: 07/25/2016 09:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GERMANN,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: FAGGART

DATE/TIME OF TEST: 07/15/2016 20:24:54

Diagnostics Check		OK	20:22	
Air Blank		0.000	20:22	
Control Test		0.079	20:22	
Air Blank		0.000	20:23	
Subject Sample	# 1		20:23	Subject Test Refused
Breath Volume		0.000	20:23	
Air Blank		0.000	20:23	
Control Test		0.080	20:24	
Air Blank		0.000	20:24	
Diagnostics Check		OK	20:24	

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SPARKMAN,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/17/2016 03:03:09

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.189	02:58
Breath Volume		1.535	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.186	03:01
Breath Volume		1.339	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: SUHETSKIS

DATE/TIME OF TEST: 07/17/2016 04:27:52

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.211	04:23
Breath Volume		2.078	04:23
Air Blank		0.000	04:23
Air Blank		0.000	04:25
Subject Sample	# 2	0.220	04:26
Breath Volume		1.539	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SATTERWHITE,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/17/2016 06:15:28

Diagnostics Check		OK	06:11	
Air Blank		0.000	06:11	
Control Test		0.080	06:12	
Air Blank		0.000	06:12	
Subject Sample	# 1		06:14	Subject Test Refused
Breath Volume		0.000	06:14	
Air Blank		0.000	06:14	
Control Test		0.081	06:14	
Air Blank		0.000	06:15	
Diagnostics Check		OK	06:15	

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: PRADELLA

DATE/TIME OF TEST: 07/17/2016 23:37:28

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.104	23:33
Breath Volume		3.179	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.099	23:35
Breath Volume		2.535	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALLIDAY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BERRY,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PEEPLES

DATE/TIME OF TEST: 07/20/2016 01:57:42

Diagnostics Check		OK	01:50
Air Blank		0.000	01:50
Control Test		0.080	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.000	01:52
Breath Volume		1.316	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.000	01:56
Breath Volume		1.843	01:56
Air Blank		0.000	01:56
Control Test		0.080	01:57
Air Blank		0.000	01:57
Diagnostics Check		OK	01:57

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: TRENCH

DATE/TIME OF TEST: 07/20/2016 21:50:05

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.235	21:45
Breath Volume		1.839	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.232	21:48
Breath Volume		1.636	21:48
Air Blank		0.000	21:49
Control Test		0.078	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MAINES

DATE/TIME OF TEST: 07/21/2016 21:11:33

Diagnostics Check	OK	21:10	
Air Blank		21:11	RFI Detect
Air Blank	0.000	21:11	

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MAINES

DATE/TIME OF TEST: 07/21/2016 21:15:53

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:13	
Control Test		0.080	21:13	
Air Blank		0.000	21:14	
Subject Sample	# 1		21:14	Subject Test Refused
Breath Volume		0.000	21:14	
Air Blank		0.000	21:14	
Control Test		0.081	21:15	
Air Blank		0.000	21:15	
Diagnostics Check		OK	21:15	

Data Download Date/Time: 07/25/2016 09:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: GOLDSTEIN

DATE/TIME OF TEST: 07/26/2016 01:53:34

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Subject Sample	# 1	0.132	01:49
Breath Volume		4.320	01:49
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.120	01:52
Breath Volume		2.582	01:52
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 08/12/2016 10:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 07/29/2016 19:43:12

Diagnostics Check		OK	19:36
Air Blank		0.000	19:36
Control Test		0.079	19:37
Air Blank		0.000	19:37
Subject Sample	# 1	0.178	19:38
Breath Volume		2.421	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.183	19:41
Breath Volume		2.605	19:41
Air Blank		0.000	19:42
Control Test		0.079	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 08/12/2016 10:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RYAN,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TURKALJ

DATE/TIME OF TEST: 07/30/2016 03:56:38

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.179	03:52
Breath Volume		3.589	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.190	03:55
Breath Volume		3.882	03:55
Air Blank		0.000	03:55
Control Test		0.078	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 08/12/2016 10:23

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RHODES,BRIAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: BARRANCO

DATE/TIME OF TEST: 07/30/2016 23:22:40

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:19	
Control Test		0.080	23:20	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:21	
Control Test		0.080	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 08/25/2016 10:02

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: RHODES,BRIAN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: BARRANCO

DATE/TIME OF TEST: 07/30/2016 23:43:24

Diagnostics Check		OK	23:35
Air Blank		0.000	23:35
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.233	23:38
Breath Volume		1.558	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.237	23:41
Breath Volume		2.148	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 08/25/2016 10:02

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,E,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MULET,C
ARRESTING OFFICER AGENCY: Miami Springs PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: SARANGO DIAZ

DATE/TIME OF TEST: 07/29/2016 06:59:33

Diagnostics Check		OK	06:55	
Air Blank		0.000	06:56	
Control Test		0.076	06:56	
Air Blank		0.000	06:57	
Subject Sample	# 1	0.103	06:57	
Breath Volume		1.730	06:57	
Air Blank			06:58	RFI Detect
Air Blank		0.000	06:58	
Control Test		0.078	06:58	
Air Blank		0.000	06:59	
Diagnostics Check		OK	06:59	

Data Download Date/Time: 08/22/2016 12:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,E,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MULET,C
ARRESTING OFFICER AGENCY: Miami Springs PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: SARANGO DIAZ

DATE/TIME OF TEST: 07/29/2016 07:07:39

Diagnostics Check		OK	07:01
Air Blank		0.000	07:01
Control Test		0.076	07:02
Air Blank		0.000	07:02
Subject Sample	# 1	0.096	07:03
Breath Volume		1.871	07:03
Air Blank		0.000	07:03
Air Blank		0.000	07:05
Subject Sample	# 2	0.094	07:06
Breath Volume		2.242	07:06
Air Blank		0.000	07:06
Control Test		0.076	07:07
Air Blank		0.000	07:07
Diagnostics Check		OK	07:07

Data Download Date/Time: 08/22/2016 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVEZ,MAYRELIS,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: CHAVEZ,MAYRELIS
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:42

SUBJECT'S LAST NAME: PINO FUENTES

DATE/TIME OF TEST: 07/11/2016 13:08:47

Diagnostics Check		OK	13:02
Air Blank		0.000	13:03
Control Test		0.078	13:03
Air Blank		0.000	13:04
Subject Sample	# 1	0.000	13:04
Breath Volume		3.050	13:04
Air Blank		0.000	13:05
Air Blank		0.000	13:07
Subject Sample	# 2	0.000	13:07
Breath Volume		3.257	13:07
Air Blank		0.000	13:07
Control Test		0.076	13:08
Air Blank		0.000	13:08
Diagnostics Check		OK	13:08

Data Download Date/Time: 07/14/2016 13:13

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002464

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONSALVO,ESTEBAN,A
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: LINARES,CARLOS
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 07/02/2016 02:59:36

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Subject Sample	# 1	0.154	02:55
Breath Volume		3.003	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.152	02:57
Breath Volume		2.398	02:57
Air Blank		0.000	02:58
Control Test		0.076	02:58
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 07/14/2016 13:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRUKE,DOUGLAS,E
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HUTCHINSON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 16:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:07

SUBJECT'S LAST NAME: KRUKE

DATE/TIME OF TEST: 07/30/2016 17:02:16

Diagnostics Check		OK	16:50	
Air Blank		0.000	16:51	
Control Test		0.078	16:51	
Air Blank		0.000	16:52	
Subject Sample	# 1		16:55	No Sample Provided
Breath Volume		0.000	16:55	
Air Blank		0.000	16:55	
Air Blank		0.000	16:57	
Subject Sample	# 2		17:00	No Sample Provided
Breath Volume		0.000	17:00	
Air Blank		0.000	17:01	
Control Test		0.079	17:01	
Air Blank		0.000	17:02	
Diagnostics Check		OK	17:02	

Data Download Date/Time: 08/23/2016 17:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003197

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANNING,KENNETH,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ALLGOOD,JEFFEREY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: AMY

DATE/TIME OF TEST: 07/31/2016 19:51:43

Diagnostics Check		OK	19:45
Air Blank		0.000	19:45
Control Test		0.079	19:46
Air Blank		0.000	19:46
Subject Sample	# 1	0.043	19:47
Breath Volume		2.605	19:47
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.038	19:50
Breath Volume		2.046	19:50
Air Blank		0.000	19:50
Control Test		0.077	19:51
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 08/23/2016 17:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/01/2016 04:31:40

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.071	04:27
Breath Volume		1.121	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.061	04:30
Breath Volume		1.167	04:30
Air Blank		0.000	04:30
Control Test		0.079	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 07/13/2016 13:50

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/01/2016 21:49:32

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:36	
Control Test		0.079	21:37	
Air Blank		0.000	21:37	
Subject Sample	# 1	0.183	21:40	Volume Not Met
Breath Volume		0.453	21:40	
Air Blank		0.000	21:41	
Air Blank		0.000	21:43	
Subject Sample	# 2	0.188	21:44	
Breath Volume		1.375	21:44	
Air Blank		0.000	21:45	
Air Blank		0.000	21:47	
Subject Sample	# 3	0.208	21:47	
Breath Volume		1.753	21:47	
Air Blank		0.000	21:48	
Control Test		0.076	21:48	
Air Blank		0.000	21:49	
Diagnostics Check		OK	21:49	

Data Download Date/Time: 07/13/2016 13:50

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/02/2016 00:05:13

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.078	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.098	00:00
Breath Volume		2.335	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.098	00:03
Breath Volume		1.785	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 07/13/2016 13:50

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARNWELL,T
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JOINER

DATE/TIME OF TEST: 07/02/2016 04:32:07

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.078	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.049	04:27
Breath Volume		1.265	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.045	04:30
Breath Volume		1.457	04:30
Air Blank		0.000	04:31
Control Test		0.077	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: MYRE

DATE/TIME OF TEST: 07/02/2016 05:21:01

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:14
Subject Sample	# 1	0.157	05:15
Breath Volume		1.468	05:15
Air Blank		0.000	05:15
Air Blank		0.000	05:17
Subject Sample	# 2	0.144	05:19
Breath Volume		1.519	05:19
Air Blank		0.000	05:20
Control Test		0.079	05:20
Air Blank		0.000	05:20
Diagnostics Check		OK	05:21

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MCNAIR

DATE/TIME OF TEST: 07/03/2016 01:35:48

Diagnostics Check		OK	01:28
Air Blank		0.000	01:28
Control Test		0.078	01:28
Air Blank		0.000	01:29
Subject Sample	# 1	0.145	01:31
Breath Volume		1.582	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.140	01:34
Breath Volume		1.167	01:34
Air Blank		0.000	01:34
Control Test		0.077	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 07/13/2016 13:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 07/03/2016 04:07:40

Diagnostics Check		OK	04:00
Air Blank		0.000	04:00
Control Test		0.078	04:01
Air Blank		0.000	04:01
Subject Sample	# 1	0.125	04:02
Breath Volume		1.257	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.129	04:06
Breath Volume		1.429	04:06
Air Blank		0.000	04:06
Control Test		0.077	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: MEJIA-SANCHEZ

DATE/TIME OF TEST: 07/03/2016 21:08:42

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.079	21:02
Air Blank		0.000	21:03
Subject Sample	# 1	0.242	21:04
Breath Volume		2.195	21:04
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.248	21:07
Breath Volume		1.207	21:07
Air Blank		0.000	21:07
Control Test		0.078	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 07/04/2016 03:05:51

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.078	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.000	03:01
Breath Volume		1.367	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.000	03:04
Breath Volume		1.398	03:04
Air Blank		0.000	03:04
Control Test		0.078	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HOBSON,D
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/04/2016 04:44:52

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.078	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.154	04:40
Breath Volume		1.968	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.144	04:43
Breath Volume		1.109	04:43
Air Blank		0.000	04:43
Control Test		0.078	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GEHRI

DATE/TIME OF TEST: 07/05/2016 03:41:21

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.079	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.278	03:36
Breath Volume		4.058	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.270	03:39
Breath Volume		4.062	03:39
Air Blank		0.000	03:40
Control Test		0.079	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 07/08/2016 00:39:18

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.203	00:34
Breath Volume		3.628	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.210	00:37
Breath Volume		3.367	00:37
Air Blank		0.000	00:38
Control Test		0.078	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 07/13/2016 13:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 07/08/2016 04:21:35

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.078	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1	0.114	04:12	Volume Not Met
Breath Volume		0.843	04:12	
Air Blank		0.000	04:13	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.129	04:16	
Breath Volume		1.488	04:16	
Air Blank		0.000	04:17	
Air Blank		0.000	04:19	
Subject Sample	# 3	0.130	04:20	
Breath Volume		1.492	04:20	
Air Blank		0.000	04:20	
Control Test		0.077	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 07/13/2016 13:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 07/09/2016 04:32:28

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.078	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.087	04:26
Breath Volume		1.707	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.086	04:30
Breath Volume		1.707	04:30
Air Blank		0.000	04:31
Control Test		0.077	04:31
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 07/13/2016 13:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 07/10/2016 01:05:55

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.078	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.154	01:01
Breath Volume		2.496	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.149	01:04
Breath Volume		2.261	01:04
Air Blank		0.000	01:05
Control Test		0.078	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 07/13/2016 13:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ANTHNEY

DATE/TIME OF TEST: 07/10/2016 05:04:21

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.079	04:56
Air Blank		0.000	04:57
Subject Sample	# 1	0.135	04:58
Breath Volume		1.933	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:00
Subject Sample	# 2	0.127	05:02
Breath Volume		1.621	05:02
Air Blank		0.000	05:03
Control Test		0.078	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 07/13/2016 13:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GOFF,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:46

SUBJECT'S LAST NAME: HIPPEN

DATE/TIME OF TEST: 07/28/2016 17:13:49

Diagnostics Check		OK	17:07
Air Blank		0.000	17:07
Control Test		0.080	17:08
Air Blank		0.000	17:08
Subject Sample	# 1	0.254	17:09
Breath Volume		3.996	17:09
Air Blank		0.000	17:10
Air Blank		0.000	17:11
Subject Sample	# 2	0.248	17:12
Breath Volume		4.027	17:12
Air Blank		0.000	17:12
Control Test		0.078	17:13
Air Blank		0.000	17:13
Diagnostics Check		OK	17:13

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 07/28/2016 20:04:17

Diagnostics Check		OK	19:57
Air Blank		0.000	19:58
Control Test		0.080	19:58
Air Blank		0.000	19:58
Subject Sample	# 1	0.222	19:59
Breath Volume		1.125	19:59
Air Blank		0.000	20:00
Air Blank		0.000	20:01
Subject Sample	# 2	0.228	20:02
Breath Volume		2.476	20:02
Air Blank		0.000	20:03
Control Test		0.078	20:03
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/29/2016 04:00:51

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.078	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.176	03:56
Breath Volume		1.992	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.176	03:59
Breath Volume		2.132	03:59
Air Blank		0.000	03:59
Control Test		0.077	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HCSO,DONOVAN
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 07/29/2016 17:17:25

Diagnostics Check		OK	17:11
Air Blank		0.000	17:11
Control Test		0.078	17:11
Air Blank		0.000	17:12
Subject Sample	# 1	0.330	17:12
Breath Volume		1.835	17:12
Air Blank		0.000	17:13
Air Blank		0.000	17:15
Subject Sample	# 2	0.321	17:15
Breath Volume		1.582	17:15
Air Blank		0.000	17:16
Control Test		0.077	17:16
Air Blank		0.000	17:17
Diagnostics Check		OK	17:17

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: ALSADAH

DATE/TIME OF TEST: 07/30/2016 03:08:33

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.224	03:03
Breath Volume		2.183	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.235	03:06
Breath Volume		2.175	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 08/10/2016 18:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 07/30/2016 04:46:29

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.129	04:41
Breath Volume		2.382	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.134	04:44
Breath Volume		2.742	04:44
Air Blank		0.000	04:45
Control Test		0.078	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/30/2016 05:41:34

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.078	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.023	05:37
Breath Volume		1.980	05:37
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.022	05:40
Breath Volume		2.453	05:40
Air Blank		0.000	05:40
Control Test		0.078	05:40
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 08/10/2016 18:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LADE,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:21

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/30/2016 08:48:38

Diagnostics Check		OK	08:41
Air Blank		0.000	08:42
Control Test		0.079	08:42
Air Blank		0.000	08:43
Subject Sample	# 1	0.189	08:43
Breath Volume		1.664	08:43
Air Blank		0.000	08:44
Air Blank		0.000	08:46
Subject Sample	# 2	0.189	08:47
Breath Volume		2.304	08:47
Air Blank		0.000	08:47
Control Test		0.078	08:48
Air Blank		0.000	08:48
Diagnostics Check		OK	08:48

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: WASKO

DATE/TIME OF TEST: 07/31/2016 01:01:38

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.079	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:00	Subject Test Refused
Breath Volume		0.000	01:00	
Air Blank		0.000	01:00	
Control Test		0.079	01:01	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 08/10/2016 18:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: WASKO

DATE/TIME OF TEST: 07/31/2016 12:07:31

Diagnostics Check		OK	12:01
Air Blank		0.000	12:01
Control Test		0.078	12:02
Air Blank		0.000	12:02
Subject Sample	# 1	0.099	12:02
Breath Volume		2.371	12:02
Air Blank		0.000	12:03
Air Blank		0.000	12:05
Subject Sample	# 2	0.099	12:05
Breath Volume		2.335	12:05
Air Blank		0.000	12:06
Control Test		0.079	12:06
Air Blank		0.000	12:07
Diagnostics Check		OK	12:07

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: WASKO

DATE/TIME OF TEST: 07/31/2016 14:19:40

Diagnostics Check		OK	14:13	
Air Blank		0.000	14:14	
Control Test		0.079	14:14	
Air Blank		0.000	14:15	
Subject Sample	# 1	0.069	14:15	
Breath Volume		2.683	14:15	
Air Blank		0.000	14:16	
Air Blank		0.000	14:18	
Subject Sample	# 2		14:18	Subject Test Refused
Breath Volume		0.000	14:18	
Air Blank		0.000	14:18	
Control Test		0.079	14:19	
Air Blank		0.000	14:19	
Diagnostics Check		OK	14:19	

Data Download Date/Time: 08/10/2016 18:40

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/03/2016 22:42:17

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:34	
Control Test		0.081	22:34	
Air Blank		0.000	22:35	
Subject Sample	# 1	0.207	22:38	Volume Not Met
Breath Volume		0.082	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:40	
Subject Sample	# 2		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:41	
Control Test		0.080	22:41	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 07/12/2016 18:17

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/14/2016 04:13:40

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.081	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1	0.131	04:06	Volume Not Met
Breath Volume		0.050	04:06	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 2		04:12	No Sample Provided
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.080	04:13	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 08/10/2016 18:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 07/15/2016 01:42:09

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.082	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.199	01:36
Breath Volume		1.960	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:39
Subject Sample	# 2	0.193	01:40
Breath Volume		1.902	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 08/10/2016 18:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MORDINI

DATE/TIME OF TEST: 07/15/2016 02:02:21

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.215	01:57
Breath Volume		1.976	01:57
Air Blank		0.000	01:58
Air Blank		0.000	01:59
Subject Sample	# 2	0.222	02:00
Breath Volume		2.230	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 08/10/2016 18:33

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/15/2016 04:13:37

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.213	04:09
Breath Volume		2.050	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.204	04:12
Breath Volume		1.507	04:12
Air Blank		0.000	04:12
Control Test		0.078	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 08/10/2016 18:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: SANTIZ GOMEZ

DATE/TIME OF TEST: 07/16/2016 04:38:40

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.081	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.162	04:33
Breath Volume		1.386	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.175	04:37
Breath Volume		1.703	04:37
Air Blank		0.000	04:37
Control Test		0.079	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 08/10/2016 18:33

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ZACKERY

DATE/TIME OF TEST: 07/16/2016 21:47:19

Diagnostics Check		OK	21:40
Air Blank		0.000	21:41
Control Test		0.082	21:41
Air Blank		0.000	21:42
Subject Sample	# 1	0.161	21:42
Breath Volume		2.093	21:42
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.160	21:45
Breath Volume		1.632	21:45
Air Blank		0.000	21:46
Control Test		0.080	21:46
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 08/10/2016 18:33

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/17/2016 00:12:15

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.267	00:04
Breath Volume		2.242	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.289	00:07
Breath Volume		2.746	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 3	0.290	00:10
Breath Volume		2.835	00:10
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAVERAS,N
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 07/18/2016 02:28:51

Diagnostics Check		OK	02:22
Air Blank		0.000	02:23
Control Test		0.082	02:23
Air Blank		0.000	02:24
Subject Sample	# 1	0.102	02:24
Breath Volume		2.046	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.101	02:27
Breath Volume		2.222	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: NOVO PERAZA

DATE/TIME OF TEST: 07/19/2016 03:15:22

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.169	03:10
Breath Volume		2.789	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.176	03:13
Breath Volume		2.410	03:13
Air Blank		0.000	03:14
Control Test		0.080	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: LAPLANTE

DATE/TIME OF TEST: 07/20/2016 20:33:58

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.082	20:28
Air Blank		0.000	20:29
Subject Sample	# 1	0.000	20:29
Breath Volume		1.207	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.000	20:32
Breath Volume		1.480	20:32
Air Blank		0.000	20:33
Control Test		0.081	20:33
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAVIS,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: MENDEZ JR

DATE/TIME OF TEST: 07/21/2016 02:18:42

Diagnostics Check		OK	02:11
Air Blank		0.000	02:12
Control Test		0.082	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.098	02:14
Breath Volume		2.218	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.096	02:17
Breath Volume		1.281	02:17
Air Blank		0.000	02:17
Control Test		0.080	02:18
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SANCHEZ RODRIGUEZ

DATE/TIME OF TEST: 07/21/2016 04:16:05

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.081	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.158	04:11
Breath Volume		2.312	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.161	04:14
Breath Volume		2.378	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/22/2016 04:19:34

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.147	04:14
Breath Volume		1.164	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.150	04:17
Breath Volume		1.230	04:17
Air Blank		0.000	04:18
Control Test		0.079	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TAYLOR,S
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 07/22/2016 21:19:57

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.082	21:14
Breath Volume		1.953	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.085	21:18
Breath Volume		2.304	21:18
Air Blank		0.000	21:19
Control Test		0.081	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 07/23/2016 03:57:13

Diagnostics Check		OK	03:43
Air Blank		0.000	03:44
Control Test		0.081	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.166	03:46
Breath Volume		2.222	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:48
Subject Sample	# 2	0.138	03:51
Breath Volume		1.406	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 3	0.160	03:55
Breath Volume		1.691	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COHEN,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: IRIZARRY

DATE/TIME OF TEST: 07/23/2016 08:11:29

Diagnostics Check		OK	08:05
Air Blank		0.000	08:05
Control Test		0.080	08:05
Air Blank		0.000	08:06
Subject Sample	# 1	0.026	08:07
Breath Volume		2.660	08:07
Air Blank		0.000	08:07
Air Blank		0.000	08:09
Subject Sample	# 2	0.022	08:09
Breath Volume		1.777	08:09
Air Blank		0.000	08:10
Control Test		0.080	08:10
Air Blank		0.000	08:11
Diagnostics Check		OK	08:11

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ARSENEAU

DATE/TIME OF TEST: 07/24/2016 02:42:44

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.081	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.134	02:38
Breath Volume		3.617	02:38
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.138	02:41
Breath Volume		2.109	02:41
Air Blank		0.000	02:41
Control Test		0.080	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 08/10/2016 18:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/24/2016 04:45:28

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.081	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.105	04:40
Breath Volume		3.226	04:40
Air Blank		0.000	04:41
Air Blank		0.000	04:43
Subject Sample	# 2	0.101	04:43
Breath Volume		4.296	04:43
Air Blank		0.000	04:44
Control Test		0.080	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/25/2016 00:40:21

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.081	00:34
Air Blank		0.000	00:34
Subject Sample	# 1	0.154	00:35
Breath Volume		3.238	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.156	00:38
Breath Volume		2.988	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 08/10/2016 18:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GRACH

DATE/TIME OF TEST: 07/25/2016 03:27:56

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.082	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.130	03:23
Breath Volume		2.750	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.132	03:26
Breath Volume		1.523	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNSBERRY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:46

SUBJECT'S LAST NAME: SLUSMEYER

DATE/TIME OF TEST: 07/27/2016 10:13:23

Diagnostics Check		OK	10:06	
Air Blank		0.000	10:07	
Control Test		0.082	10:07	
Air Blank		0.000	10:07	
Subject Sample	# 1	0.132	10:08	
Breath Volume		1.417	10:08	
Air Blank		0.000	10:09	
Air Blank		0.000	10:11	
Subject Sample	# 2		10:11	Slope Not Met
Breath Volume		1.492	10:11	
Air Blank		0.000	10:12	
Control Test		0.081	10:12	
Air Blank		0.000	10:13	
Diagnostics Check		OK	10:13	

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: THORNSBERRY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:46

SUBJECT'S LAST NAME: SLUSMEYER

DATE/TIME OF TEST: 07/27/2016 10:48:45

Diagnostics Check		OK	10:42
Air Blank		0.000	10:42
Control Test		0.081	10:43
Air Blank		0.000	10:43
Subject Sample	# 1	0.128	10:44
Breath Volume		2.078	10:44
Air Blank		0.000	10:44
Air Blank		0.000	10:46
Subject Sample	# 2	0.126	10:47
Breath Volume		2.613	10:47
Air Blank		0.000	10:47
Control Test		0.080	10:48
Air Blank		0.000	10:48
Diagnostics Check		OK	10:48

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MOLINARES-BELTRAN

DATE/TIME OF TEST: 07/28/2016 02:47:26

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.169	02:42
Breath Volume		1.445	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.185	02:45
Breath Volume		2.703	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/10/2016 18:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 07/29/2016 00:01:17

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.000	23:56
Breath Volume		2.648	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.000	23:59
Breath Volume		1.359	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: SLONAKER

DATE/TIME OF TEST: 07/30/2016 04:14:23

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.134	04:09
Breath Volume		3.382	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.130	04:12
Breath Volume		2.078	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GRAY,S
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: CRAKER

DATE/TIME OF TEST: 07/30/2016 16:32:30

Diagnostics Check		OK	16:25
Air Blank		0.000	16:25
Control Test		0.082	16:26
Air Blank		0.000	16:26
Subject Sample	# 1	0.237	16:27
Breath Volume		2.585	16:27
Air Blank		0.000	16:28
Air Blank		0.000	16:30
Subject Sample	# 2	0.249	16:30
Breath Volume		2.503	16:30
Air Blank		0.000	16:31
Control Test		0.079	16:31
Air Blank		0.000	16:32
Diagnostics Check		OK	16:32

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/30/2016 23:43:41

Diagnostics Check		OK	23:36
Air Blank		0.000	23:37
Control Test		0.082	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.097	23:39
Breath Volume		2.125	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.101	23:42
Breath Volume		2.078	23:42
Air Blank		0.000	23:42
Control Test		0.081	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/31/2016 04:04:07

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.174	03:59
Breath Volume		1.710	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.178	04:02
Breath Volume		1.226	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GARRELL,Q
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/31/2016 04:30:18

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.184	04:25
Breath Volume		2.808	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.183	04:28
Breath Volume		2.394	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 07/31/2016 04:59:38

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.081	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.165	04:55
Breath Volume		3.453	04:55
Air Blank		0.000	04:55
Air Blank		0.000	04:57
Subject Sample	# 2	0.165	04:58
Breath Volume		3.382	04:58
Air Blank		0.000	04:58
Control Test		0.079	04:59
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 08/10/2016 18:34

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORSTRICK,DAKOTA,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TORSTRICK,DAKOTA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: WOLFF

DATE/TIME OF TEST: 07/26/2016 19:22:43

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.081	19:16
Air Blank		0.000	19:17
Subject Sample	# 1	0.234	19:17
Breath Volume		2.023	19:17
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.217	19:21
Breath Volume		1.550	19:21
Air Blank		0.000	19:21
Control Test		0.080	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 08/12/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003407

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 07/26/2016 23:58:36

Diagnostics Check		OK	23:52
Air Blank		0.000	23:52
Control Test		0.079	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.139	23:54
Breath Volume		2.828	23:54
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.132	23:57
Breath Volume		1.625	23:57
Air Blank		0.000	23:57
Control Test		0.078	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 08/31/2016 12:25

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTANA,YUDIEL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SANTANA,YUDIEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: KEYS

DATE/TIME OF TEST: 07/09/2016 01:10:56

Diagnostics Check		OK	00:58	
Air Blank		0.000	00:59	
Control Test		0.080	00:59	
Air Blank		0.000	01:00	
Subject Sample	# 1	0.269	01:00	
Breath Volume		1.781	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:03	
Subject Sample	# 2		01:06	No Sample Provided
Breath Volume		0.000	01:06	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 3	0.256	01:09	
Breath Volume		1.425	01:09	
Air Blank		0.000	01:10	
Control Test		0.081	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 07/12/2016 11:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: WASKOWITZ

DATE/TIME OF TEST: 07/23/2016 20:29:06

Diagnostics Check		OK	20:23
Air Blank		0.000	20:23
Control Test		0.079	20:23
Air Blank		0.000	20:24
Subject Sample	# 1	0.177	20:24
Breath Volume		2.464	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.172	20:27
Breath Volume		1.871	20:27
Air Blank		0.000	20:28
Control Test		0.075	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 08/02/2016 17:23

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURHAM,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DURHAM,JAMES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: CURA

DATE/TIME OF TEST: 07/24/2016 00:01:28

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.000	23:57
Breath Volume		1.617	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.000	00:00
Breath Volume		3.511	00:00
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 08/02/2016 17:23

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YASUDA,SHOHEI,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YASUDA,SHOHEI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: VERDECIA RODRIGUEZ

DATE/TIME OF TEST: 07/28/2016 05:27:26

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.079	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.108	05:21
Breath Volume		3.152	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.100	05:25
Breath Volume		2.218	05:25
Air Blank		0.000	05:26
Control Test		0.080	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 08/02/2016 17:23

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:41

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 07/28/2016 17:33:09

Diagnostics Check		OK	17:25
Air Blank		0.000	17:26
Control Test		0.079	17:26
Air Blank		0.000	17:27
Subject Sample	# 1	0.000	17:28
Breath Volume		1.339	17:28
Air Blank		0.000	17:28
Air Blank		0.000	17:30
Subject Sample	# 2	0.000	17:31
Breath Volume		1.109	17:31
Air Blank		0.000	17:32
Control Test		0.078	17:32
Air Blank		0.000	17:33
Diagnostics Check		OK	17:33

Data Download Date/Time: 08/02/2016 17:23

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURHAM,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DURHAM,JAMES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ETTERS

DATE/TIME OF TEST: 07/28/2016 23:42:33

Diagnostics Check		OK	23:36
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.135	23:38
Breath Volume		4.597	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.132	23:41
Breath Volume		3.843	23:41
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 08/02/2016 17:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VILORIO

DATE/TIME OF TEST: 07/30/2016 01:14:01

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.172	01:09
Breath Volume		1.269	01:09
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.169	01:12
Breath Volume		1.152	01:12
Air Blank		0.000	01:13
Control Test		0.077	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:14

Data Download Date/Time: 08/25/2016 09:51

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: WOODIN

DATE/TIME OF TEST: 07/01/2016 02:09:00

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.080	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.139	02:04
Breath Volume		1.457	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.132	02:07
Breath Volume		1.367	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:09

Data Download Date/Time: 07/13/2016 11:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 07/01/2016 05:53:02

Diagnostics Check		OK	05:43
Air Blank		0.000	05:44
Control Test		0.080	05:44
Air Blank		0.000	05:45
Subject Sample	# 1	0.117	05:45
Breath Volume		1.921	05:45
Air Blank		0.000	05:46
Air Blank		0.000	05:47
Subject Sample	# 2	0.139	05:48
Breath Volume		2.960	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 3	0.139	05:51
Breath Volume		4.523	05:51
Air Blank		0.000	05:52
Control Test		0.080	05:52
Air Blank		0.000	05:52
Diagnostics Check		OK	05:53

Data Download Date/Time: 07/13/2016 11:38

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,ERIC,R
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DIAZ,ERIC
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: SMALLEY

DATE/TIME OF TEST: 07/01/2016 20:52:16

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:44
Subject Sample	# 1	0.224	20:46
Breath Volume		1.960	20:46
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.212	20:50
Breath Volume		1.953	20:50
Air Blank		0.000	20:51
Control Test		0.077	20:51
Air Blank		0.000	20:52
Diagnostics Check		OK	20:52

Data Download Date/Time: 07/13/2016 11:38

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: MARTINS

DATE/TIME OF TEST: 07/02/2016 23:07:10

Diagnostics Check		OK	22:58	
Air Blank		0.000	22:59	
Control Test		0.080	22:59	
Air Blank		0.000	22:59	
Subject Sample	# 1		23:03	No Sample Provided
Breath Volume		0.000	23:03	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 2		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:06	
Control Test		0.079	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 07/13/2016 11:38

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCOY,RICHARD,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MCCOY,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 07/04/2016 19:50:35

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.079	19:44
Air Blank		0.000	19:44
Subject Sample	# 1	0.155	19:45
Breath Volume		1.195	19:45
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.150	19:49
Breath Volume		1.562	19:49
Air Blank		0.000	19:49
Control Test		0.079	19:49
Air Blank		0.000	19:50
Diagnostics Check		OK	19:50

Data Download Date/Time: 07/13/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MUNDT,BRANDON,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BOGGAN,MARLON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: LANE

DATE/TIME OF TEST: 07/05/2016 02:34:25

Diagnostics Check		OK	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.112	02:28
Breath Volume		1.375	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:30
Subject Sample	# 2	0.104	02:32
Breath Volume		2.972	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 07/13/2016 11:38

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JUSTISS,WILTON,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DUGRE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SEALE

DATE/TIME OF TEST: 07/07/2016 00:54:40

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.227	00:46
Breath Volume		1.320	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.267	00:49
Breath Volume		2.976	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 3	0.255	00:53
Breath Volume		2.031	00:53
Air Blank		0.000	00:53
Control Test		0.079	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 07/13/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: HOWLAND BUSHNELL

DATE/TIME OF TEST: 07/07/2016 03:41:34

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:37	No Sample Provided
Breath Volume		0.000	03:37	
Air Blank		0.000	03:37	
Air Blank		0.000	03:39	
Subject Sample	# 2		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:40	
Control Test		0.080	03:40	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 07/13/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,VICTOR,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JACKSON,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: WLOCH

DATE/TIME OF TEST: 07/07/2016 21:13:20

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.079	21:05
Air Blank		0.000	21:06
Subject Sample	# 1	0.212	21:07
Breath Volume		1.253	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.205	21:11
Breath Volume		1.378	21:11
Air Blank		0.000	21:12
Control Test		0.079	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 07/13/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COVENEY,PATRICK,F
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COVENEY,PATRICK
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: POGUE

DATE/TIME OF TEST: 07/09/2016 22:15:15

Diagnostics Check		OK	22:10	
Air Blank		0.000	22:10	
Control Test		0.080	22:10	
Air Blank		0.000	22:11	
Subject Sample	# 1	0.042	22:12	
Breath Volume		2.097	22:12	
Air Blank		0.000	22:12	
Air Blank			22:14	Ambient Fail
Air Blank		0.000	22:15	

Data Download Date/Time: 07/13/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COVENEY,PATRICK,F
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COVENEY,PATRICK
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: POGUE

DATE/TIME OF TEST: 07/09/2016 22:28:30

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.080	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.037	22:23
Breath Volume		1.976	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.035	22:27
Breath Volume		1.269	22:27
Air Blank		0.000	22:27
Control Test		0.079	22:27
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 07/13/2016 11:38

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAND,GLENN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: NIEVES

DATE/TIME OF TEST: 07/10/2016 23:15:01

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:00	
Control Test		0.080	23:00	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:04	No Sample Provided
Breath Volume		0.000	23:04	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 2	0.000	23:08	
Breath Volume		1.324	23:08	
Air Blank		0.000	23:08	
Air Blank		0.000	23:11	
Subject Sample	# 3	0.000	23:13	
Breath Volume		1.828	23:13	
Air Blank		0.000	23:14	
Control Test		0.079	23:14	
Air Blank		0.000	23:14	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 07/13/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT,ADAM,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: POGOZELSKI,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: DODD

DATE/TIME OF TEST: 07/11/2016 19:49:47

Diagnostics Check		OK	19:38	
Air Blank		0.000	19:38	
Control Test		0.079	19:39	
Air Blank		0.000	19:39	
Subject Sample	# 1		19:42	No Sample Provided
Breath Volume		0.000	19:42	
Air Blank		0.000	19:43	
Air Blank		0.000	19:45	
Subject Sample	# 2		19:48	No Sample Provided
Breath Volume		0.000	19:48	
Air Blank		0.000	19:48	
Control Test		0.079	19:49	
Air Blank		0.000	19:49	
Diagnostics Check		OK	19:49	

Data Download Date/Time: 07/13/2016 11:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYBRAND,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRISTOPHER
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 07/12/2016 22:31:45

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.079	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.152	22:26
Breath Volume		1.328	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.138	22:30
Breath Volume		1.140	22:30
Air Blank		0.000	22:30
Control Test		0.079	22:31
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 07/13/2016 11:38

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTWELL,JEFFREY,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: CANTWELL,JEFFREY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 07/16/2016 02:01:14

Diagnostics Check		OK	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Subject Sample	# 1	0.148	01:56
Breath Volume		1.742	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.153	01:59
Breath Volume		2.640	01:59
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 08/11/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROBINSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 07/18/2016 02:15:52

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.150	02:09
Breath Volume		1.160	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.156	02:14
Breath Volume		2.136	02:14
Air Blank		0.000	02:14
Control Test		0.078	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 08/11/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCFALL,CHARLES,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MCFALL,CHARLES
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:04

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 07/22/2016 18:31:49

Diagnostics Check		OK	18:25
Air Blank		0.000	18:25
Control Test		0.079	18:26
Air Blank		0.000	18:26
Subject Sample	# 1	0.237	18:27
Breath Volume		3.359	18:27
Air Blank		0.000	18:27
Air Blank		0.000	18:29
Subject Sample	# 2	0.240	18:30
Breath Volume		2.980	18:30
Air Blank		0.000	18:30
Control Test		0.079	18:31
Air Blank		0.000	18:31
Diagnostics Check		OK	18:31

Data Download Date/Time: 08/11/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/23/2016 04:21:05

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.081	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.151	04:16
Breath Volume		1.132	04:16
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.142	04:19
Breath Volume		2.367	04:19
Air Blank		0.000	04:20
Control Test		0.076	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,DUSTAN,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILDE,MATTHEW
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: JANSSEN

DATE/TIME OF TEST: 07/24/2016 18:55:37

Diagnostics Check		OK	18:52	
Air Blank		0.000	18:53	
Control Test		0.080	18:53	
Air Blank		0.000	18:53	
Subject Sample	# 1		18:54	Subject Test Refused
Breath Volume		0.000	18:54	
Air Blank		0.000	18:54	
Control Test		0.080	18:55	
Air Blank		0.000	18:55	
Diagnostics Check		OK	18:55	

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,DUSTAN,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WOODARD,TAYLOR
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/24/2016 19:32:04

Diagnostics Check		OK	19:25
Air Blank		0.000	19:26
Control Test		0.081	19:26
Air Blank		0.000	19:27
Subject Sample	# 1	0.200	19:27
Breath Volume		2.222	19:27
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.201	19:30
Breath Volume		2.929	19:30
Air Blank		0.000	19:31
Control Test		0.080	19:31
Air Blank		0.000	19:31
Diagnostics Check		OK	19:32

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,CASEY,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EMRICH,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: DULANEY

DATE/TIME OF TEST: 07/25/2016 01:17:57

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.081	01:12
Air Blank		0.000	01:12
Subject Sample	# 1	0.252	01:13
Breath Volume		3.070	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:15
Subject Sample	# 2	0.250	01:16
Breath Volume		3.109	01:16
Air Blank		0.000	01:17
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,CASEY,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DENNIS,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 07/25/2016 01:48:10

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.178	01:43
Breath Volume		3.164	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.173	01:46
Breath Volume		2.937	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COVENEY,PATRICK,F
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: COVENEY,PATRICK
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MORALES-HOYOS

DATE/TIME OF TEST: 07/25/2016 02:21:59

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:19	
Control Test		0.080	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: MOSHER

DATE/TIME OF TEST: 07/26/2016 02:42:44

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1		02:38	No Sample Provided
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.080	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 08/11/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,CASEY,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: EMRICH,IAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/27/2016 22:48:38

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:37	
Control Test		0.079	22:38	
Air Blank		0.000	22:38	
Subject Sample	# 1		22:41	No Sample Provided
Breath Volume		0.000	22:41	
Air Blank		0.000	22:42	
Air Blank		0.000	22:44	
Subject Sample	# 2		22:47	No Sample Provided
Breath Volume		0.000	22:47	
Air Blank		0.000	22:47	
Control Test		0.080	22:48	
Air Blank		0.000	22:48	
Diagnostics Check		OK	22:48	

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNIS,CASEY,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: DENNIS,CASEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: MARSKE

DATE/TIME OF TEST: 07/27/2016 23:23:41

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:12	
Control Test		0.081	23:12	
Air Blank		0.000	23:13	
Subject Sample	# 1	0.082	23:16	Volume Not Met
Breath Volume		0.199	23:16	
Air Blank		0.000	23:17	
Air Blank		0.000	23:19	
Subject Sample	# 2	0.079	23:22	Volume Not Met
Breath Volume		0.253	23:22	
Air Blank		0.000	23:22	
Control Test		0.079	23:23	
Air Blank		0.000	23:23	
Diagnostics Check		OK	23:23	

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ROTH

DATE/TIME OF TEST: 07/29/2016 01:27:09

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.131	01:22
Breath Volume		1.890	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.128	01:25
Breath Volume		2.597	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUDD,CHASE,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RUDD,CHASE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: LAMPTON

DATE/TIME OF TEST: 07/29/2016 06:18:26

Diagnostics Check		OK	06:15	
Air Blank		0.000	06:15	
Control Test		0.080	06:16	
Air Blank		0.000	06:16	
Subject Sample	# 1		06:17	Subject Test Refused
Breath Volume		0.000	06:17	
Air Blank		0.000	06:17	
Control Test		0.079	06:17	
Air Blank		0.000	06:18	
Diagnostics Check		OK	06:18	

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCOY,RICHARD,G
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MCCOY,RICHARD
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 07/30/2016 01:55:10

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Subject Sample	# 1	0.216	01:50
Breath Volume		2.746	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.208	01:53
Breath Volume		1.503	01:53
Air Blank		0.000	01:54
Control Test		0.079	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIAGET,PHILIP,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PIAGET,PHILIP
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CANTASANO

DATE/TIME OF TEST: 07/31/2016 04:04:09

Diagnostics Check		OK	04:01	
Air Blank		0.000	04:01	
Control Test		0.080	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 08/11/2016 12:18

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREIRA,JOSE,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: PEREIRA,JOSE
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: STANFIELD

DATE/TIME OF TEST: 07/31/2016 04:31:56

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.125	04:27
Breath Volume		2.476	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.124	04:30
Breath Volume		2.339	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 08/11/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SARMENT,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: THEODORO

DATE/TIME OF TEST: 07/31/2016 05:13:19

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1		05:11	Subject Test Refused
Breath Volume		0.000	05:11	
Air Blank		0.000	05:12	
Control Test		0.079	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 08/11/2016 12:18

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS,EVAN,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ANDREWS,MATT
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 07/05/2016 21:39:10

Diagnostics Check		OK	21:32	
Air Blank		0.000	21:32	
Control Test		0.081	21:33	
Air Blank		0.000	21:33	
Subject Sample	# 1		21:36	No Sample Provided
Breath Volume		0.000	21:36	
Air Blank		0.000	21:37	
Air Blank		0.000	21:39	

Data Download Date/Time: 08/02/2016 10:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS,EVAN,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ANDREWS,MATT
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: HUNTER JR

DATE/TIME OF TEST: 07/05/2016 21:53:22

Diagnostics Check		OK	21:46
Air Blank		0.000	21:47
Control Test		0.081	21:47
Air Blank		0.000	21:47
Subject Sample	# 1	0.333	21:48
Breath Volume		2.699	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:50
Subject Sample	# 2	0.339	21:51
Breath Volume		2.296	21:51
Air Blank		0.000	21:52
Control Test		0.080	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 08/02/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: DREADING

DATE/TIME OF TEST: 07/07/2016 04:46:14

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:43	
Control Test		0.082	04:44	
Air Blank		0.000	04:44	
Subject Sample	# 1		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:45	
Control Test		0.081	04:45	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 08/02/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEALS,JOSEPH,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SEALS,JOSEPH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/08/2016 20:56:07

Diagnostics Check		OK	20:49	
Air Blank		0.000	20:50	
Control Test		0.081	20:50	
Air Blank		0.000	20:51	
Subject Sample	# 1	0.034	20:51	
Breath Volume		1.816	20:51	
Air Blank		0.000	20:52	
Air Blank		0.000	20:54	
Subject Sample	# 2		20:54	Subject Test Refused
Breath Volume		0.000	20:54	
Air Blank		0.000	20:55	
Control Test		0.081	20:55	
Air Blank		0.000	20:56	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 08/02/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEALS,JOSEPH,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SEALS,JOSEPH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 07/09/2016 21:36:51

Diagnostics Check		OK	21:33	
Air Blank		0.000	21:34	
Control Test		0.080	21:34	
Air Blank		0.000	21:34	
Subject Sample	# 1		21:35	Subject Test Refused
Breath Volume		0.000	21:35	
Air Blank		0.000	21:35	
Control Test		0.080	21:36	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 08/02/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEALS,JOSEPH,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: SEALS,JOSEPH
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LOWTHER

DATE/TIME OF TEST: 07/10/2016 02:09:36

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.155	02:04
Breath Volume		2.726	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.156	02:08
Breath Volume		2.367	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 08/02/2016 10:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNOLDS,EVAN,D
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ANDREWS,MATT
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/11/2016 03:47:19

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.081	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.107	03:42
Breath Volume		1.906	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.096	03:45
Breath Volume		1.355	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 08/02/2016 10:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: DUNLAP,GABE
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 07/12/2016 22:54:53

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.026	22:50
Breath Volume		1.687	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:52
Subject Sample	# 2	0.022	22:53
Breath Volume		1.105	22:53
Air Blank		0.000	22:53
Control Test		0.080	22:54
Air Blank		0.000	22:54
Diagnostics Check		OK	22:54

Data Download Date/Time: 08/02/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SUTTER

DATE/TIME OF TEST: 07/20/2016 21:33:29

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.081	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.290	21:28
Breath Volume		1.214	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.289	21:31
Breath Volume		1.835	21:31
Air Blank		0.000	21:32
Control Test		0.078	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 08/02/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARMER,GLENCHARLES,J
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: FARMER,GLENCHARLES
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASE

DATE/TIME OF TEST: 07/21/2016 02:33:24

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:30	
Control Test		0.081	02:31	
Air Blank		0.000	02:31	
Subject Sample	# 1		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.081	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 08/02/2016 10:47

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: SANTAMARIA

DATE/TIME OF TEST: 07/03/2016 02:21:11

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.077	02:15
Air Blank		0.000	02:15
Subject Sample	# 1	0.139	02:16
Breath Volume		1.316	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:18
Subject Sample	# 2	0.144	02:19
Breath Volume		1.386	02:19
Air Blank		0.000	02:20
Control Test		0.077	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 07/29/2016 08:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BLYDEN

DATE/TIME OF TEST: 07/03/2016 04:22:45

Diagnostics Check		OK	04:16
Air Blank		0.000	04:17
Control Test		0.077	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.000	04:18
Breath Volume		2.664	04:18
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.000	04:21
Breath Volume		2.257	04:21
Air Blank		0.000	04:21
Control Test		0.077	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 07/29/2016 08:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/16/2016 04:13:43

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.077	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1	0.153	04:09	Volume Not Met
Breath Volume		0.417	04:09	
Air Blank		0.000	04:09	
Air Blank		0.000	04:11	
Subject Sample	# 2		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.075	04:13	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 07/29/2016 08:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: F0WLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/16/2016 04:28:58

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.077	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.178	04:21
Breath Volume		1.398	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.172	04:27
Breath Volume		3.746	04:27
Air Blank		0.000	04:28
Control Test		0.077	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 07/29/2016 08:35

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: SGARZI

DATE/TIME OF TEST: 07/17/2016 00:52:54

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:44	
Control Test		0.077	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:48	No Sample Provided
Breath Volume		0.000	00:48	
Air Blank		0.000	00:48	
Air Blank		0.000	00:50	
Subject Sample	# 2		00:51	Subject Test Refused
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Control Test		0.077	00:52	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 07/29/2016 08:35

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 07/02/2016 01:50:53

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.077	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.302	01:46
Breath Volume		2.589	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.303	01:49
Breath Volume		1.816	01:49
Air Blank		0.000	01:49
Control Test		0.077	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 07/12/2016 18:24

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: VICENTYRODRIGUEZ

DATE/TIME OF TEST: 07/02/2016 04:48:01

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:42
Subject Sample	# 1	0.111	04:43
Breath Volume		2.941	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.117	04:46
Breath Volume		3.601	04:46
Air Blank		0.000	04:47
Control Test		0.078	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:48

Data Download Date/Time: 07/13/2016 09:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: VALLIERE

DATE/TIME OF TEST: 07/03/2016 01:54:38

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.077	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.045	01:50
Breath Volume		2.324	01:50
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.044	01:53
Breath Volume		2.335	01:53
Air Blank		0.000	01:53
Control Test		0.078	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 07/13/2016 09:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 07/16/2016 05:30:30

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.078	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.169	05:25
Breath Volume		1.554	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.154	05:28
Breath Volume		1.742	05:28
Air Blank		0.000	05:29
Control Test		0.077	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 08/24/2016 14:57

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: MOMPEROUSSECORDONES

DATE/TIME OF TEST: 07/17/2016 18:02:33

Diagnostics Check		OK	17:56
Air Blank		0.000	17:56
Control Test		0.078	17:57
Air Blank		0.000	17:57
Subject Sample	# 1	0.300	17:57
Breath Volume		1.140	17:57
Air Blank		0.000	17:58
Air Blank		0.000	18:00
Subject Sample	# 2	0.294	18:00
Breath Volume		2.074	18:00
Air Blank		0.000	18:01
Control Test		0.078	18:01
Air Blank		0.000	18:02
Diagnostics Check		OK	18:02

Data Download Date/Time: 08/24/2016 14:57

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/18/2016 02:02:38

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:57	
Control Test		0.077	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.215	02:01	Volume Not Met
Breath Volume		0.171	02:01	
Air Blank		0.000	02:02	
Air Blank		0.000	02:02	

Data Download Date/Time: 08/24/2016 14:57

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: QASSM

DATE/TIME OF TEST: 07/22/2016 01:55:22

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.077	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.000	01:51
Breath Volume		1.597	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.000	01:54
Breath Volume		1.410	01:54
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 08/24/2016 14:57

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 07/30/2016 18:42:40

Diagnostics Check		OK	18:35
Air Blank		0.000	18:36
Control Test		0.077	18:36
Air Blank		0.000	18:37
Subject Sample	# 1	0.274	18:37
Breath Volume		2.093	18:37
Air Blank		0.000	18:38
Air Blank		0.000	18:40
Subject Sample	# 2	0.269	18:41
Breath Volume		2.031	18:41
Air Blank		0.000	18:41
Control Test		0.077	18:42
Air Blank		0.000	18:42
Diagnostics Check		OK	18:42

Data Download Date/Time: 08/24/2016 14:57

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MASS SALAS

DATE/TIME OF TEST: 07/31/2016 04:20:46

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.078	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:20	Sequence Aborted
Breath Volume		0.390	04:20	
Air Blank		0.000	04:20	

Data Download Date/Time: 08/24/2016 14:57

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALONE,CHARLES,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MALONE,CHARLES
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: GARLOBO

DATE/TIME OF TEST: 07/31/2016 18:54:55

Diagnostics Check		OK	18:47
Air Blank		0.000	18:47
Control Test		0.077	18:48
Air Blank		0.000	18:48
Subject Sample	# 1	0.225	18:50
Breath Volume		2.347	18:50
Air Blank		0.000	18:50
Air Blank		0.000	18:52
Subject Sample	# 2	0.229	18:53
Breath Volume		1.222	18:53
Air Blank		0.000	18:53
Control Test		0.076	18:54
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 08/24/2016 14:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNAUD,GEORGE,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: WASHINGTON,RONALD
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: NIETO

DATE/TIME OF TEST: 07/02/2016 22:26:20

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:23	
Control Test		0.079	22:24	
Air Blank		0.000	22:24	
Subject Sample	# 1		22:24	Subject Test Refused
Breath Volume		0.000	22:24	
Air Blank		0.000	22:25	
Control Test		0.079	22:25	
Air Blank		0.000	22:26	
Diagnostics Check		OK	22:26	

Data Download Date/Time: 07/22/2016 21:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HANEY,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: REGALADO

DATE/TIME OF TEST: 07/04/2016 18:12:54

Diagnostics Check		OK	18:06
Air Blank		0.000	18:06
Control Test		0.078	18:07
Air Blank		0.000	18:07
Subject Sample	# 1	0.155	18:08
Breath Volume		3.722	18:08
Air Blank		0.000	18:08
Air Blank		0.000	18:10
Subject Sample	# 2	0.145	18:11
Breath Volume		2.308	18:11
Air Blank		0.000	18:11
Control Test		0.078	18:12
Air Blank		0.000	18:12
Diagnostics Check		OK	18:12

Data Download Date/Time: 07/22/2016 21:26

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HANEY,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: BARZOLA SANCHEZ

DATE/TIME OF TEST: 07/04/2016 18:48:16

Diagnostics Check		OK	18:45	
Air Blank		0.000	18:45	
Control Test		0.079	18:46	
Air Blank		0.000	18:46	
Subject Sample	# 1		18:46	Subject Test Refused
Breath Volume		0.000	18:46	
Air Blank		0.000	18:47	
Control Test		0.079	18:47	
Air Blank		0.000	18:48	
Diagnostics Check		OK	18:48	

Data Download Date/Time: 07/22/2016 21:26

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FREDRICKSON,S
ARRESTING OFFICER AGENCY: Brooksville PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: MCNAIR

DATE/TIME OF TEST: 07/01/2016 03:44:52

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.084	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1		03:43	Subject Test Refused
Breath Volume		0.000	03:43	
Air Blank		0.000	03:43	
Control Test		0.083	03:44	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 07/19/2016 09:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LITTLE,J
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: BUI
ARREST TIME: 18:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WALCZAK

DATE/TIME OF TEST: 07/03/2016 20:03:40

Diagnostics Check		OK	20:00	
Air Blank		0.000	20:01	
Control Test		0.083	20:01	
Air Blank		0.000	20:02	
Subject Sample	# 1		20:02	Subject Test Refused
Breath Volume		0.000	20:02	
Air Blank		0.000	20:02	
Control Test		0.082	20:03	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:03	

Data Download Date/Time: 07/19/2016 09:36

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERY,TRICIA,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FLETCHER,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BOGUSLAWSKI

DATE/TIME OF TEST: 07/04/2016 02:33:35

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.084	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.029	02:29
Breath Volume		1.695	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.028	02:32
Breath Volume		1.628	02:32
Air Blank		0.000	02:32
Control Test		0.084	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 07/19/2016 09:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 07/04/2016 21:24:29

Diagnostics Check		OK	21:17
Air Blank		0.000	21:18
Control Test		0.083	21:18
Air Blank		0.000	21:18
Subject Sample	# 1	0.225	21:19
Breath Volume		3.417	21:19
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.215	21:22
Breath Volume		3.085	21:22
Air Blank		0.000	21:23
Control Test		0.081	21:23
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 07/19/2016 09:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ROWELL

DATE/TIME OF TEST: 07/08/2016 23:55:45

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.084	23:46
Air Blank		0.000	23:46
Subject Sample	# 1	0.159	23:47
Breath Volume		1.203	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:49
Subject Sample	# 2	0.182	23:51
Breath Volume		2.777	23:51
Air Blank		0.000	23:51
Air Blank		0.000	23:53
Subject Sample	# 3	0.164	23:54
Breath Volume		1.683	23:54
Air Blank		0.000	23:54
Control Test		0.083	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 07/19/2016 09:36

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 07/09/2016 01:11:46

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:09	
Control Test		0.083	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 1		01:10	Subject Test Refused
Breath Volume		0.000	01:10	
Air Blank		0.000	01:10	
Control Test		0.084	01:11	
Air Blank		0.000	01:11	
Diagnostics Check		OK	01:11	

Data Download Date/Time: 07/19/2016 09:36

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FREDRICKSON,S
ARRESTING OFFICER AGENCY: Brooksville PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: LEONARD

DATE/TIME OF TEST: 07/09/2016 04:22:16

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.084	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.000	04:17
Breath Volume		2.453	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.000	04:20
Breath Volume		2.128	04:20
Air Blank		0.000	04:21
Control Test		0.083	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 07/19/2016 09:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,A
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DORAN

DATE/TIME OF TEST: 07/09/2016 23:57:47

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.084	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.165	23:53
Breath Volume		4.164	23:53
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.165	23:56
Breath Volume		3.832	23:56
Air Blank		0.000	23:56
Control Test		0.083	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 07/19/2016 09:36

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: READO

DATE/TIME OF TEST: 07/10/2016 01:58:50

Diagnostics Check		OK	01:50
Air Blank		0.000	01:51
Control Test		0.083	01:51
Air Blank		0.000	01:51
Subject Sample	# 1	0.171	01:54
Breath Volume		1.398	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.181	01:57
Breath Volume		2.046	01:57
Air Blank		0.000	01:57
Control Test		0.084	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 07/19/2016 09:36

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: FREDRICKSON,S
ARRESTING OFFICER AGENCY: Brooksville PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WADDY

DATE/TIME OF TEST: 07/10/2016 02:38:35

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.083	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.082	02:37	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 07/19/2016 09:36

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,COLLEEN,E
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOAGLAN,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: MAZZA

DATE/TIME OF TEST: 07/10/2016 15:14:57

Diagnostics Check		OK	15:09
Air Blank		0.000	15:09
Control Test		0.084	15:09
Air Blank		0.000	15:10
Subject Sample	# 1	0.000	15:10
Breath Volume		2.906	15:10
Air Blank		0.000	15:11
Air Blank		0.000	15:13
Subject Sample	# 2	0.000	15:13
Breath Volume		2.007	15:13
Air Blank		0.000	15:14
Control Test		0.084	15:14
Air Blank		0.000	15:14
Diagnostics Check		OK	15:14

Data Download Date/Time: 07/19/2016 09:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: CARTER,KENNETH
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:35

SUBJECT'S LAST NAME: PAOLELLA

DATE/TIME OF TEST: 07/11/2016 12:02:52

Diagnostics Check		OK	11:55
Air Blank		0.000	11:55
Control Test		0.085	11:56
Air Blank		0.000	11:56
Subject Sample	# 1	0.000	11:58
Breath Volume		1.828	11:58
Air Blank		0.000	11:59
Air Blank		0.000	12:01
Subject Sample	# 2	0.000	12:01
Breath Volume		1.718	12:01
Air Blank		0.000	12:01
Control Test		0.083	12:02
Air Blank		0.000	12:02
Diagnostics Check		OK	12:02

Data Download Date/Time: 07/19/2016 09:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GENOVESE,L
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 07/12/2016 03:56:40

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.084	03:54	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.085	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 07/19/2016 09:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/12/2016 23:24:32

Diagnostics Check		OK	23:20	
Air Blank		0.000	23:20	
Control Test		0.084	23:21	
Air Blank		0.000	23:21	
Subject Sample	# 1		23:23	Subject Test Refused
Breath Volume		0.000	23:23	
Air Blank		0.000	23:23	
Control Test		0.085	23:23	
Air Blank		0.000	23:24	
Diagnostics Check		OK	23:24	

Data Download Date/Time: 07/19/2016 09:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HYLER,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: KAYE

DATE/TIME OF TEST: 07/13/2016 00:26:31

Diagnostics Check		OK	00:17
Air Blank		0.000	00:18
Control Test		0.084	00:18
Air Blank		0.000	00:19
Subject Sample	# 1	0.126	00:21
Breath Volume		1.539	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.120	00:24
Breath Volume		1.675	00:24
Air Blank		0.000	00:25
Control Test		0.083	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 07/19/2016 09:37

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: BARIKOS

DATE/TIME OF TEST: 07/21/2016 03:04:44

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.084	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.000	03:00
Breath Volume		2.945	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.000	03:03
Breath Volume		3.785	03:03
Air Blank		0.000	03:03
Control Test		0.084	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BOZARD

DATE/TIME OF TEST: 07/22/2016 01:08:11

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.084	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.153	01:02
Breath Volume		2.351	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.133	01:06
Breath Volume		1.660	01:06
Air Blank		0.000	01:07
Control Test		0.082	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MAZUR

DATE/TIME OF TEST: 07/23/2016 00:05:51

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.084	00:03	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:04	Subject Test Refused
Breath Volume		0.000	00:04	
Air Blank		0.000	00:04	
Control Test		0.084	00:05	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVENPORT,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MURPHY,JAVON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: WEIL

DATE/TIME OF TEST: 07/24/2016 01:24:57

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.084	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.000	01:18
Breath Volume		1.125	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.000	01:23
Breath Volume		1.136	01:23
Air Blank		0.000	01:24
Control Test		0.084	01:24
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,BEN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 07/24/2016 20:46:37

Diagnostics Check		OK	20:43	
Air Blank		0.000	20:44	
Control Test		0.084	20:44	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:45	Subject Test Refused
Breath Volume		0.000	20:45	
Air Blank		0.000	20:45	
Control Test		0.083	20:46	
Air Blank		0.000	20:46	
Diagnostics Check		OK	20:46	

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: GARTNER,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: VO

DATE/TIME OF TEST: 07/27/2016 01:12:00

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.083	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.159	01:07
Breath Volume		1.617	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.158	01:10
Breath Volume		1.445	01:10
Air Blank		0.000	01:11
Control Test		0.083	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:12

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO,I,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: HOGAN

DATE/TIME OF TEST: 07/27/2016 01:46:26

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.083	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.000	01:42
Breath Volume		3.773	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.000	01:45
Breath Volume		3.605	01:45
Air Blank		0.000	01:45
Control Test		0.083	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,DAVID,O
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: REID,W
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 07/27/2016 22:36:13

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:33	
Control Test		0.083	22:34	
Air Blank		0.000	22:34	
Subject Sample	# 1		22:34	Subject Test Refused
Breath Volume		0.000	22:34	
Air Blank		0.000	22:35	
Control Test		0.084	22:35	
Air Blank		0.000	22:36	
Diagnostics Check		OK	22:36	

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005249

INSTRUMENT REGISTERED TO: HERNANDO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAOLELLA,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,B
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 07/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: MEYER

DATE/TIME OF TEST: 07/29/2016 18:24:41

Diagnostics Check		OK	18:17
Air Blank		0.000	18:18
Control Test		0.084	18:18
Air Blank		0.000	18:19
Subject Sample	# 1	0.039	18:19
Breath Volume		1.460	18:19
Air Blank		0.000	18:20
Air Blank		0.000	18:22
Subject Sample	# 2	0.038	18:23
Breath Volume		1.167	18:23
Air Blank		0.000	18:23
Control Test		0.083	18:24
Air Blank		0.000	18:24
Diagnostics Check		OK	18:24

Data Download Date/Time: 08/19/2016 05:39

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/01/2016 02:32:14

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.115	02:27
Breath Volume		1.667	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.114	02:30
Breath Volume		1.628	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONNELLY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONNELLY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: LURVEY

DATE/TIME OF TEST: 07/01/2016 04:31:19

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.172	04:26
Breath Volume		2.636	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.170	04:29
Breath Volume		2.078	04:29
Air Blank		0.000	04:30
Control Test		0.081	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NOBLES,W
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ORTIZ-CRUZ

DATE/TIME OF TEST: 07/02/2016 03:37:15

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.082	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.199	03:32
Breath Volume		1.695	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.186	03:35
Breath Volume		1.125	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: SHADE

DATE/TIME OF TEST: 07/02/2016 04:45:23

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:43	
Control Test		0.081	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:44	
Control Test		0.080	04:44	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: FRIESE

DATE/TIME OF TEST: 07/02/2016 22:40:57

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:38	
Control Test		0.081	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:39	Subject Test Refused
Breath Volume		0.000	22:39	
Air Blank		0.000	22:40	
Control Test		0.080	22:40	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLETON,KRISTEN,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MIDDLETON,KRISTEN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DONOHUE

DATE/TIME OF TEST: 07/03/2016 01:26:39

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:24	
Control Test		0.081	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:25	
Control Test		0.080	01:26	
Air Blank		0.000	01:26	
Diagnostics Check		OK	01:26	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BURROWS

DATE/TIME OF TEST: 07/03/2016 03:28:28

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.080	03:26	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:27	Subject Test Refused
Breath Volume		0.000	03:27	
Air Blank		0.000	03:27	
Control Test		0.081	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: BALASKAS

DATE/TIME OF TEST: 07/03/2016 04:55:50

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.207	04:51
Breath Volume		1.660	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.216	04:54
Breath Volume		2.023	04:54
Air Blank		0.000	04:54
Control Test		0.080	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLASINI,MIGUEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 07/03/2016 05:23:29

Diagnostics Check		OK	05:16
Air Blank		0.000	05:17
Control Test		0.080	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.161	05:18
Breath Volume		1.855	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.163	05:21
Breath Volume		1.414	05:21
Air Blank		0.000	05:22
Control Test		0.079	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 07/28/2016 23:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: TALMADGE

DATE/TIME OF TEST: 07/05/2016 04:59:31

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:56	
Control Test		0.081	04:57	
Air Blank		0.000	04:57	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:58	
Control Test		0.081	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 07/28/2016 23:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/07/2016 04:38:26

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:36	
Control Test		0.082	04:36	
Air Blank		0.000	04:36	
Subject Sample	# 1		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:37	
Control Test		0.080	04:37	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCELOWICZ,CHARLES,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WINCELOWICZ,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: FRANSON

DATE/TIME OF TEST: 07/07/2016 05:32:42

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.081	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.118	05:28
Breath Volume		3.304	05:28
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.127	05:31
Breath Volume		3.222	05:31
Air Blank		0.000	05:31
Control Test		0.080	05:32
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: REMINGA

DATE/TIME OF TEST: 07/08/2016 00:45:12

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.081	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.118	00:40
Breath Volume		2.222	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.126	00:43
Breath Volume		1.953	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: WILLS

DATE/TIME OF TEST: 07/09/2016 00:23:29

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:21	
Control Test		0.082	00:21	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:22	
Control Test		0.081	00:22	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 07/28/2016 23:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: RUTHERFORD

DATE/TIME OF TEST: 07/09/2016 03:01:18

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:58	
Control Test		0.081	02:58	
Air Blank		0.000	02:58	
Subject Sample	# 1		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	03:00	
Control Test		0.081	03:00	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARTZIG,BOBBY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 16:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: ST JOHN

DATE/TIME OF TEST: 07/09/2016 17:29:41

Diagnostics Check		OK	17:23
Air Blank		0.000	17:23
Control Test		0.081	17:23
Air Blank		0.000	17:24
Subject Sample	# 1	0.000	17:25
Breath Volume		1.375	17:25
Air Blank		0.000	17:25
Air Blank		0.000	17:27
Subject Sample	# 2	0.000	17:28
Breath Volume		1.503	17:28
Air Blank		0.000	17:28
Control Test		0.081	17:29
Air Blank		0.000	17:29
Diagnostics Check		OK	17:29

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: GWIN

DATE/TIME OF TEST: 07/10/2016 01:06:11

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.105	01:01
Breath Volume		2.664	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.102	01:04
Breath Volume		1.808	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: THORKELSON

DATE/TIME OF TEST: 07/10/2016 03:45:53

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.081	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.080	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 07/10/2016 04:45:49

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:43	
Control Test		0.080	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:44	Subject Test Refused
Breath Volume		0.000	04:44	
Air Blank		0.000	04:44	
Control Test		0.081	04:45	
Air Blank		0.000	04:45	
Diagnostics Check		OK	04:45	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONNELLY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONNELLY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HORN

DATE/TIME OF TEST: 07/14/2016 03:31:15

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.024	03:26
Breath Volume		3.289	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.022	03:29
Breath Volume		3.597	03:29
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOHR,K,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOHR,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: STADNICKI

DATE/TIME OF TEST: 07/14/2016 04:30:56

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.081	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.212	04:26
Breath Volume		1.601	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.195	04:29
Breath Volume		1.875	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/17/2016 02:16:12

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.082	02:14	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:14	Subject Test Refused
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Control Test		0.081	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MARTIN,V
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GUZMAN,MARTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:57

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/17/2016 08:24:16

Diagnostics Check		OK	08:18
Air Blank		0.000	08:18
Control Test		0.082	08:18
Air Blank		0.000	08:19
Subject Sample	# 1	0.132	08:19
Breath Volume		1.593	08:19
Air Blank		0.000	08:20
Air Blank		0.000	08:22
Subject Sample	# 2	0.134	08:22
Breath Volume		1.175	08:22
Air Blank		0.000	08:23
Control Test		0.080	08:23
Air Blank		0.000	08:24
Diagnostics Check		OK	08:24

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWANNER,SARAH,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LOCKE,REBECCA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 07/17/2016 19:24:04

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:21	
Control Test		0.082	19:21	
Air Blank		0.000	19:22	
Subject Sample	# 1		19:22	Subject Test Refused
Breath Volume		0.000	19:22	
Air Blank		0.000	19:23	
Control Test		0.081	19:23	
Air Blank		0.000	19:23	
Diagnostics Check		OK	19:24	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: COLLMAN,I
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: BEININGEN

DATE/TIME OF TEST: 07/20/2016 21:56:58

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.083	21:50
Air Blank		0.000	21:50
Subject Sample	# 1	0.156	21:52
Breath Volume		2.246	21:52
Air Blank		0.000	21:52
Air Blank		0.000	21:54
Subject Sample	# 2	0.149	21:55
Breath Volume		2.175	21:55
Air Blank		0.000	21:56
Control Test		0.081	21:56
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CHIVAS

DATE/TIME OF TEST: 07/21/2016 00:54:10

Diagnostics Check		OK	00:50	
Air Blank		0.000	00:51	
Control Test		0.082	00:51	
Air Blank		0.000	00:51	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:53	
Control Test		0.081	00:53	
Air Blank		0.000	00:54	
Diagnostics Check		OK	00:54	

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLASINI,M
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PENZIN

DATE/TIME OF TEST: 07/21/2016 03:17:13

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.082	03:11
Air Blank		0.000	03:12
Subject Sample	# 1	0.176	03:12
Breath Volume		1.894	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.168	03:15
Breath Volume		1.347	03:15
Air Blank		0.000	03:16
Control Test		0.081	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG505803
DRY GAS EXPIRY DATE: 02/27/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: SHELTO

DATE/TIME OF TEST: 07/22/2016 00:57:37

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.082	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.170	00:52
Breath Volume		1.773	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.175	00:56
Breath Volume		1.253	00:56
Air Blank		0.000	00:56
Control Test		0.081	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: OMALLY

DATE/TIME OF TEST: 07/22/2016 03:58:59

Diagnostics Check		OK	03:49
Air Blank		0.000	03:50
Control Test		0.081	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.024	03:52
Breath Volume		1.722	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:54
Subject Sample	# 2	0.024	03:57
Breath Volume		1.246	03:57
Air Blank		0.000	03:58
Control Test		0.082	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: COMI

DATE/TIME OF TEST: 07/22/2016 04:21:35

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.200	04:16
Breath Volume		2.175	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.206	04:19
Breath Volume		3.281	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: BLOCHER

DATE/TIME OF TEST: 07/23/2016 03:15:40

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.082	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.164	03:11
Breath Volume		1.500	03:11
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.153	03:14
Breath Volume		1.964	03:14
Air Blank		0.000	03:14
Control Test		0.080	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 07/28/2016 23:27

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/23/2016 03:47:29

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.081	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:46	
Control Test		0.081	03:46	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 07/28/2016 23:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: MAVROMATIS

DATE/TIME OF TEST: 07/23/2016 05:00:49

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:58	
Control Test		0.081	04:58	
Air Blank		0.000	04:59	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	04:59	
Control Test		0.081	05:00	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 07/28/2016 23:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SPARKS,J
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: VOLPE

DATE/TIME OF TEST: 07/24/2016 02:50:08

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.082	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.258	02:45
Breath Volume		1.882	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.258	02:48
Breath Volume		1.687	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 07/28/2016 23:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: KOVAC

DATE/TIME OF TEST: 07/27/2016 03:34:00

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.082	03:30	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:32	Subject Test Refused
Breath Volume		0.000	03:32	
Air Blank		0.000	03:33	
Control Test		0.082	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 07/28/2016 23:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 14:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: DURBIN

DATE/TIME OF TEST: 07/27/2016 15:49:49

Diagnostics Check		OK	15:46	
Air Blank		0.000	15:46	
Control Test		0.081	15:46	
Air Blank		0.000	15:47	
Subject Sample	# 1		15:48	Subject Test Refused
Breath Volume		0.000	15:48	
Air Blank		0.000	15:48	
Control Test		0.081	15:49	
Air Blank		0.000	15:49	
Diagnostics Check		OK	15:49	

Data Download Date/Time: 07/28/2016 23:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: COLMERY

DATE/TIME OF TEST: 07/28/2016 01:34:52

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.082	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.081	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 07/28/2016 23:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/28/2016 06:14:08

Diagnostics Check		OK	06:08
Air Blank		0.000	06:08
Control Test		0.082	06:08
Air Blank		0.000	06:09
Subject Sample	# 1	0.089	06:09
Breath Volume		1.707	06:09
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.089	06:12
Breath Volume		2.710	06:12
Air Blank		0.000	06:13
Control Test		0.081	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 07/28/2016 23:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 07/29/2016 02:02:46

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.082	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.257	01:58
Breath Volume		1.628	01:58
Air Blank		0.000	01:58
Air Blank		0.000	02:00
Subject Sample	# 2	0.256	02:01
Breath Volume		1.578	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 08/27/2016 22:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 06:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: ESKALD

DATE/TIME OF TEST: 07/29/2016 07:33:22

Diagnostics Check		OK	07:27
Air Blank		0.000	07:27
Control Test		0.082	07:27
Air Blank		0.000	07:28
Subject Sample	# 1	0.029	07:29
Breath Volume		1.660	07:29
Air Blank		0.000	07:29
Air Blank		0.000	07:31
Subject Sample	# 2	0.027	07:31
Breath Volume		1.585	07:31
Air Blank		0.000	07:32
Control Test		0.081	07:32
Air Blank		0.000	07:33
Diagnostics Check		OK	07:33

Data Download Date/Time: 08/27/2016 22:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HERNANDEZ,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: PEREZ-GOMEZ

DATE/TIME OF TEST: 07/30/2016 00:20:36

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.082	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.257	00:12
Breath Volume		2.238	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.234	00:15
Breath Volume		2.375	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 3	0.241	00:19
Breath Volume		2.246	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:20
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 08/27/2016 22:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR,RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BUCKNER

DATE/TIME OF TEST: 07/30/2016 03:11:51

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.081	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 08/27/2016 22:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,WAYNE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WAGNER,WAYNE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: BROWNLOW

DATE/TIME OF TEST: 07/30/2016 04:00:05

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.082	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:59	
Control Test		0.081	03:59	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 08/27/2016 22:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CHARSLEY

DATE/TIME OF TEST: 07/31/2016 01:42:08

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.082	01:35
Air Blank		0.000	01:36
Subject Sample	# 1	0.376	01:37
Breath Volume		3.878	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.375	01:40
Breath Volume		2.828	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 08/27/2016 22:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG605301
DRY GAS EXPIRY DATE: 02/22/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: ITEHUA

DATE/TIME OF TEST: 07/31/2016 03:32:42

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.081	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.113	03:27
Breath Volume		1.613	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.108	03:31
Breath Volume		1.636	03:31
Air Blank		0.000	03:31
Control Test		0.081	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 08/27/2016 22:55

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,BRUCE,H
BREATH TEST OPERATOR AGENCY: Perry PD
ARRESTING OFFICER: GRIFFIN,BRUCE
ARRESTING OFFICER AGENCY: Perry PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: BARROW

DATE/TIME OF TEST: 07/28/2016 21:01:02

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.077	20:55
Air Blank		0.000	20:56
Subject Sample	# 1	0.088	20:56
Breath Volume		2.359	20:56
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.095	20:59
Breath Volume		2.351	20:59
Air Blank		0.000	21:00
Control Test		0.078	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:01

Data Download Date/Time: 08/31/2016 16:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31113080A1
DRY GAS EXPIRY DATE: 12/01/2015
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: CARBAUGH

DATE/TIME OF TEST: 07/01/2016 00:48:50

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.247	00:44
Breath Volume		1.449	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.228	00:47
Breath Volume		1.550	00:47
Air Blank		0.000	00:47
Control Test		0.077	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 07/14/2016 17:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ARKOVICH,ROBBIE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/01/2016 02:50:59

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.077	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:50	
Control Test		0.077	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 07/14/2016 17:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31113080A1
DRY GAS EXPIRY DATE: 12/01/2015
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: CANADY

DATE/TIME OF TEST: 07/08/2016 02:35:25

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.077	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.170	02:30
Breath Volume		2.238	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.167	02:33
Breath Volume		1.953	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 07/14/2016 17:28

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31113080A1
DRY GAS EXPIRY DATE: 12/01/2015
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 07/09/2016 02:39:10

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.077	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.263	02:34
Breath Volume		1.296	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.262	02:37
Breath Volume		1.816	02:37
Air Blank		0.000	02:38
Control Test		0.077	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 07/14/2016 17:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARKOVICH,ROBBIE,S
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 06/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 07/14/2016 02:06:06

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.078	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1		02:04	Subject Test Refused
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Control Test		0.078	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 07/14/2016 17:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/15/2016 22:33:52

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:31	
Control Test		0.078	22:31	
Air Blank		0.000	22:32	
Subject Sample	# 1		22:32	Subject Test Refused
Breath Volume		0.000	22:32	
Air Blank		0.000	22:32	
Control Test		0.078	22:33	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:33	

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JENKINS,DANE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SANCHEZ-BUSTAMANTE

DATE/TIME OF TEST: 07/16/2016 00:53:42

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.183	00:49
Breath Volume		1.671	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.178	00:52
Breath Volume		1.339	00:52
Air Blank		0.000	00:52
Control Test		0.078	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: KOPEC

DATE/TIME OF TEST: 07/16/2016 19:18:36

Diagnostics Check		OK	19:10
Air Blank		0.000	19:11
Control Test		0.078	19:11
Air Blank		0.000	19:12
Subject Sample	# 1	0.196	19:14
Breath Volume		1.835	19:14
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.193	19:17
Breath Volume		1.394	19:17
Air Blank		0.000	19:17
Control Test		0.076	19:17
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: COMSTOCK

DATE/TIME OF TEST: 07/16/2016 23:52:49

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:47	
Subject Sample	# 1	0.238	23:48	
Breath Volume		3.425	23:48	
Air Blank		0.000	23:49	
Air Blank		0.000	23:50	
Subject Sample	# 2	0.235	23:51	
Breath Volume		2.734	23:51	
Air Blank		0.000	23:51	
Control Test		0.073	23:52	Control Outside Tolerance
Air Blank		0.000	23:52	

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: COMSTOCK

DATE/TIME OF TEST: 07/17/2016 00:01:30

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.076	23:55
Air Blank		0.000	23:56
Subject Sample	# 1	0.242	23:56
Breath Volume		1.710	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.248	23:59
Breath Volume		3.125	23:59
Air Blank		0.000	00:00
Control Test		0.079	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JENKINS,DANE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SKALET

DATE/TIME OF TEST: 07/17/2016 02:28:14

Diagnostics Check		OK	02:23	
Air Blank		0.000	02:23	
Control Test		0.079	02:24	
Air Blank		0.000	02:24	
Subject Sample	# 1	0.214	02:25	
Breath Volume		2.031	02:25	
Air Blank		0.000	02:25	
Air Blank			02:27	RFI Detect
Air Blank			02:27	RFI Detect
Air Blank		0.000	02:28	

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JENKINS,DANE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SKALET

DATE/TIME OF TEST: 07/17/2016 02:35:48

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.231	02:31
Breath Volume		3.285	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.226	02:34
Breath Volume		2.867	02:34
Air Blank		0.000	02:34
Control Test		0.078	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 07/17/2016 20:32:30

Diagnostics Check		OK	20:25	
Air Blank		0.000	20:26	
Control Test		0.078	20:26	
Air Blank		0.000	20:26	
Subject Sample	# 1		20:29	No Sample Provided
Breath Volume		0.000	20:29	
Air Blank		0.000	20:30	
Air Blank			20:32	Improper Sample
Air Blank			20:32	Improper Sample
Air Blank			20:32	Improper Sample

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,CALVIN,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 07/17/2016 20:38:33

Diagnostics Check	OK	20:37	
Air Blank		20:38	Ambient Fail
Air Blank	0.000	20:38	

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: WYSONG

DATE/TIME OF TEST: 07/17/2016 23:38:24

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.079	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.178	23:33
Breath Volume		2.105	23:33
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.168	23:36
Breath Volume		2.062	23:36
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 07/19/2016 01:46:09

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.079	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.198	01:41
Breath Volume		2.964	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:43
Subject Sample	# 2	0.195	01:44
Breath Volume		2.593	01:44
Air Blank		0.000	01:45
Control Test		0.079	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAGLE,TERRI,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE,SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:04

SUBJECT'S LAST NAME: PALM

DATE/TIME OF TEST: 07/20/2016 10:31:36

Diagnostics Check		OK	10:25
Air Blank		0.000	10:25
Control Test		0.080	10:26
Air Blank		0.000	10:26
Subject Sample	# 1	0.000	10:27
Breath Volume		1.261	10:27
Air Blank		0.000	10:27
Air Blank		0.000	10:29
Subject Sample	# 2	0.000	10:30
Breath Volume		1.304	10:30
Air Blank		0.000	10:30
Control Test		0.078	10:31
Air Blank		0.000	10:31
Diagnostics Check		OK	10:31

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: JOHANSSON

DATE/TIME OF TEST: 07/21/2016 03:59:51

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:57	
Control Test		0.079	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JENSEN,BROCK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: PRIOR

DATE/TIME OF TEST: 07/22/2016 01:57:48

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.079	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:56	
Control Test		0.078	01:57	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: BOORD

DATE/TIME OF TEST: 07/22/2016 03:12:09

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.232	03:07
Breath Volume		2.078	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.220	03:10
Breath Volume		2.343	03:10
Air Blank		0.000	03:11
Control Test		0.078	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JENSEN,BROCK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: FOUNTAINE

DATE/TIME OF TEST: 07/23/2016 00:22:36

Diagnostics Check		OK	00:15
Air Blank		0.000	00:15
Control Test		0.078	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.260	00:17
Breath Volume		2.132	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.248	00:21
Breath Volume		2.792	00:21
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KOVAR

DATE/TIME OF TEST: 07/23/2016 00:47:36

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.078	00:46	
Air Blank		00:47	RFI Detect
Air Blank	0.000	00:47	

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KOVAR

DATE/TIME OF TEST: 07/23/2016 00:52:13

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:50	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Control Test		0.078	00:51	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JENSEN,BROCK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: YALACKI

DATE/TIME OF TEST: 07/23/2016 01:20:32

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.248	01:15
Breath Volume		2.160	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:17
Subject Sample	# 2	0.237	01:18
Breath Volume		2.187	01:18
Air Blank		0.000	01:19
Control Test		0.077	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: COROUTHERS

DATE/TIME OF TEST: 07/23/2016 04:42:25

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.079	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank			04:41	RFI Detect
Air Blank		0.000	04:42	

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: COROUTHERS

DATE/TIME OF TEST: 07/23/2016 04:46:42

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.079	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1		04:45	Subject Test Refused
Breath Volume		0.000	04:45	
Air Blank		0.000	04:45	
Control Test		0.078	04:46	
Air Blank		0.000	04:46	
Diagnostics Check		OK	04:46	

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 07/23/2016 22:04:32

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.077	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.264	21:59
Breath Volume		2.210	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.270	22:02
Breath Volume		1.636	22:02
Air Blank		0.000	22:03
Control Test		0.078	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JENSEN,BROCK
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 07/24/2016 03:14:06

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.220	03:09
Breath Volume		1.277	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.224	03:12
Breath Volume		1.640	03:12
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMEU,ANTONIO,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: NAGLE,TERI
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:53

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 07/26/2016 18:18:58

Diagnostics Check		OK	18:13	
Air Blank		0.000	18:14	
Control Test		0.076	18:14	
Air Blank		0.000	18:15	
Subject Sample	# 1		18:17	Subject Test Refused
Breath Volume		0.140	18:17	
Air Blank		0.000	18:18	
Control Test		0.078	18:18	
Air Blank		0.000	18:18	
Diagnostics Check		OK	18:18	

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMEU,ANTONIO,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WILLIAMS,LEIGHTON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:19

SUBJECT'S LAST NAME: CEHRS

DATE/TIME OF TEST: 07/26/2016 19:48:13

Diagnostics Check		OK	19:39
Air Blank		0.000	19:40
Control Test		0.075	19:40
Air Blank		0.000	19:41
Subject Sample	# 1	0.195	19:43
Breath Volume		2.082	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:46
Subject Sample	# 2	0.213	19:46
Breath Volume		3.796	19:46
Air Blank		0.000	19:47
Control Test		0.075	19:47
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WARRICK

DATE/TIME OF TEST: 07/27/2016 23:03:00

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.079	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.190	22:58
Breath Volume		1.261	22:58
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.185	23:01
Breath Volume		1.222	23:01
Air Blank		0.000	23:02
Control Test		0.078	23:02
Air Blank		0.000	23:02
Diagnostics Check		OK	23:03

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOCKERS,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE JR,S
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/28/2016 12:51:49

Diagnostics Check		OK	12:44
Air Blank		0.000	12:44
Control Test		0.079	12:44
Air Blank		0.000	12:45
Subject Sample	# 1	0.181	12:46
Breath Volume		1.179	12:46
Air Blank		0.000	12:47
Air Blank		0.000	12:49
Subject Sample	# 2	0.172	12:50
Breath Volume		1.144	12:50
Air Blank		0.000	12:50
Control Test		0.079	12:51
Air Blank		0.000	12:51
Diagnostics Check		OK	12:51

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAGLE,TERRI,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ADMINISTRATIVE,ADMIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: ADMIN
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: LASLO

DATE/TIME OF TEST: 07/29/2016 21:53:35

Diagnostics Check		OK	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:48
Air Blank		0.000	21:48
Subject Sample	# 1	0.025	21:49
Breath Volume		2.707	21:49
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.024	21:52
Breath Volume		3.179	21:52
Air Blank		0.000	21:52
Control Test		0.080	21:52
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 08/30/2016 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: VEATCH,KYLE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RATLIFF

DATE/TIME OF TEST: 07/29/2016 23:53:13

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.079	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.151	23:48
Breath Volume		2.203	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:50
Subject Sample	# 2	0.154	23:51
Breath Volume		2.769	23:51
Air Blank		0.000	23:52
Control Test		0.079	23:52
Air Blank		0.000	23:53
Diagnostics Check		OK	23:53

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: VEATCH,KYLE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GOVAN

DATE/TIME OF TEST: 07/30/2016 03:10:33

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.079	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.055	03:01	Volume Not Met
Breath Volume		0.613	03:01	
Air Blank		0.000	03:01	
Air Blank		0.000	03:03	
Subject Sample	# 2	0.068	03:04	
Breath Volume		1.832	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 3	0.056	03:08	
Breath Volume		1.152	03:08	
Air Blank		0.000	03:09	
Control Test		0.075	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005338

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: VEATCH,KYLE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CROMPTON

DATE/TIME OF TEST: 07/31/2016 03:41:27

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:38	
Control Test		0.080	03:39	
Air Blank		0.000	03:39	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:40	
Control Test		0.079	03:40	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 08/30/2016 20:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: POLLEY

DATE/TIME OF TEST: 07/15/2016 04:18:01

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.077	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1	0.000	04:11	Volume Not Met
Breath Volume		0.472	04:11	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 2	0.000	04:16	Volume Not Met
Breath Volume		0.429	04:16	
Air Blank		0.000	04:17	
Control Test		0.078	04:17	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 08/12/2016 21:22

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 07/16/2016 03:00:20

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.078	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.149	02:55
Breath Volume		1.378	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.161	02:58
Breath Volume		2.683	02:58
Air Blank		0.000	02:59
Control Test		0.077	02:59
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 08/12/2016 21:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 07/23/2016 02:49:29

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.225	02:44
Breath Volume		2.468	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.217	02:47
Breath Volume		1.433	02:47
Air Blank		0.000	02:48
Control Test		0.076	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 08/12/2016 21:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005454

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CAPPELLANO

DATE/TIME OF TEST: 07/24/2016 01:33:14

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.077	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1	0.242	01:23	
Breath Volume		1.242	01:23	
Air Blank		0.000	01:24	
Air Blank		0.000	01:26	
Subject Sample	# 2	0.215	01:29	Volume Not Met
Breath Volume		0.558	01:29	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 3		01:31	Subject Test Refused
Breath Volume		0.000	01:31	
Air Blank		0.000	01:32	
Control Test		0.075	01:32	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 08/12/2016 21:22

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,R,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SMITH,R
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/02/2016 00:12:56

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.076	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.057	00:08
Breath Volume		3.515	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:10
Subject Sample	# 2	0.056	00:11
Breath Volume		2.765	00:11
Air Blank		0.000	00:11
Control Test		0.076	00:12
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 07/06/2016 14:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,R,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: NERONI,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/02/2016 00:29:48

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.077	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.170	00:24
Breath Volume		1.527	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:26
Subject Sample	# 2	0.156	00:27
Breath Volume		1.953	00:27
Air Blank		0.000	00:28
Control Test		0.075	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 07/06/2016 14:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,R,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 07/02/2016 01:57:34

Diagnostics Check	OK	01:56	
Air Blank	0.000	01:56	
Control Test	0.074	01:56	Control Outside Tolerance
Air Blank	0.000	01:57	

Data Download Date/Time: 07/06/2016 14:38

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,R,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 07/02/2016 02:07:08

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.075	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1	0.217	02:02	
Breath Volume		1.351	02:02	
Air Blank		0.000	02:03	
Air Blank		0.000	02:05	
Subject Sample	# 2	0.210	02:05	
Breath Volume		1.718	02:05	
Air Blank			02:06	Purge Fail
Air Blank		0.000	02:07	

Data Download Date/Time: 07/06/2016 14:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEINER,R,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SANDERS,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 07/02/2016 02:26:12

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.075	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.202	02:24	
Breath Volume		2.113	02:24	
Air Blank			02:25	Purge Fail
Air Blank		0.000	02:26	

Data Download Date/Time: 07/06/2016 14:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005468

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,RAYMOND,P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER,RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29614080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BERESTECKI

DATE/TIME OF TEST: 07/29/2016 23:28:34

Diagnostics Check		OK	23:20
Air Blank		0.000	23:21
Control Test		0.078	23:21
Air Blank		0.000	23:22
Subject Sample	# 1	0.058	23:23
Breath Volume		1.347	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:25
Subject Sample	# 2	0.053	23:26
Breath Volume		1.343	23:26
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 08/12/2016 14:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:44

SUBJECT'S LAST NAME: POND

DATE/TIME OF TEST: 07/04/2016 10:11:00

Diagnostics Check		OK	10:04
Air Blank		0.000	10:04
Control Test		0.079	10:05
Air Blank		0.000	10:05
Subject Sample	# 1	0.046	10:06
Breath Volume		1.921	10:06
Air Blank		0.000	10:07
Air Blank		0.000	10:09
Subject Sample	# 2	0.042	10:09
Breath Volume		2.421	10:09
Air Blank		0.000	10:10
Control Test		0.078	10:10
Air Blank		0.000	10:10
Diagnostics Check		OK	10:11

Data Download Date/Time: 07/22/2016 09:25

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAVIS,CHARLES,B
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: TRAVIS,CHARLES
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: COMM VEHIC
ARREST TIME: 10:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: JAIKARAN

DATE/TIME OF TEST: 07/15/2016 11:16:56

Diagnostics Check		OK	11:10
Air Blank		0.000	11:11
Control Test		0.079	11:11
Air Blank		0.000	11:12
Subject Sample	# 1	0.152	11:12
Breath Volume		2.335	11:12
Air Blank		0.000	11:13
Air Blank		0.000	11:14
Subject Sample	# 2	0.149	11:15
Breath Volume		1.609	11:15
Air Blank		0.000	11:16
Control Test		0.079	11:16
Air Blank		0.000	11:16
Diagnostics Check		OK	11:16

Data Download Date/Time: 07/22/2016 09:25

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAVISC,CHARLES,B
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: TRAVIS,CHARLES
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/16/2016 07:43:23

Diagnostics Check		OK	07:40	
Air Blank		0.000	07:41	
Control Test		0.079	07:41	
Air Blank		0.000	07:41	
Subject Sample	# 1		07:42	Subject Test Refused
Breath Volume		0.000	07:42	
Air Blank		0.000	07:42	
Control Test		0.080	07:42	
Air Blank		0.000	07:43	
Diagnostics Check		OK	07:43	

Data Download Date/Time: 07/22/2016 09:25

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,RICARDO,J
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: PEREZ,EMILIO
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 06/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/18/2016 05:56:10

Diagnostics Check		OK	05:49
Air Blank		0.000	05:50
Control Test		0.079	05:50
Air Blank		0.000	05:51
Subject Sample	# 1	0.199	05:51
Breath Volume		3.757	05:51
Air Blank		0.000	05:52
Air Blank		0.000	05:54
Subject Sample	# 2	0.196	05:54
Breath Volume		3.746	05:54
Air Blank		0.000	05:55
Control Test		0.079	05:55
Air Blank		0.000	05:56
Diagnostics Check		OK	05:56

Data Download Date/Time: 07/22/2016 09:25

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTANO,DANIEL,A
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: MONTANO,DANIEL
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 07/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: NOA NODA

DATE/TIME OF TEST: 07/31/2016 18:25:07

Diagnostics Check		OK	18:21	
Air Blank		0.000	18:22	
Control Test		0.078	18:22	
Air Blank		0.000	18:23	
Subject Sample	# 1		18:23	Subject Test Refused
Breath Volume		0.015	18:23	
Air Blank		0.000	18:24	
Control Test		0.079	18:24	
Air Blank		0.000	18:25	
Diagnostics Check		OK	18:25	

Data Download Date/Time: 08/10/2016 10:20

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUTLEDGE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ARCIA TEJEDA

DATE/TIME OF TEST: 07/02/2016 23:46:41

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.078	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.206	23:42
Breath Volume		2.269	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.216	23:45
Breath Volume		2.675	23:45
Air Blank		0.000	23:45
Control Test		0.079	23:46
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ARCILA HERNANDEZ

DATE/TIME OF TEST: 07/03/2016 00:15:40

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.156	00:10
Breath Volume		2.054	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.161	00:13
Breath Volume		1.718	00:13
Air Blank		0.000	00:14
Control Test		0.078	00:15
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,FEDERICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: LOPEZ II

DATE/TIME OF TEST: 07/03/2016 02:01:07

Diagnostics Check		OK	01:58	
Air Blank		0.000	01:58	
Control Test		0.079	01:59	
Air Blank		0.000	01:59	
Subject Sample	# 1		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	02:00	
Control Test		0.080	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TORRES,DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 07/03/2016 03:56:07

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.079	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:55	
Control Test		0.080	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 07/27/2016 09:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MAIR

DATE/TIME OF TEST: 07/03/2016 04:30:13

Diagnostics Check		OK	04:23
Air Blank		0.000	04:23
Control Test		0.079	04:24
Air Blank		0.000	04:24
Subject Sample	# 1	0.149	04:25
Breath Volume		1.742	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:27
Subject Sample	# 2	0.144	04:28
Breath Volume		1.535	04:28
Air Blank		0.000	04:29
Control Test		0.079	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BERENYI

DATE/TIME OF TEST: 07/03/2016 23:02:34

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:57
Subject Sample	# 1	0.215	22:57
Breath Volume		3.105	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.205	23:00
Breath Volume		1.277	23:00
Air Blank		0.000	23:01
Control Test		0.078	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SOCA

DATE/TIME OF TEST: 07/04/2016 00:42:45

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:40	
Control Test		0.079	00:40	
Air Blank		0.000	00:41	
Subject Sample	# 1		00:41	Subject Test Refused
Breath Volume		0.000	00:41	
Air Blank		0.000	00:41	
Control Test		0.080	00:42	
Air Blank		0.000	00:42	
Diagnostics Check		OK	00:42	

Data Download Date/Time: 07/27/2016 09:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 07/04/2016 01:43:35

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.098	01:39
Breath Volume		2.562	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.098	01:41
Breath Volume		3.335	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: PEREZ JARRIN

DATE/TIME OF TEST: 07/10/2016 02:49:58

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.139	02:44
Breath Volume		2.097	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.157	02:48
Breath Volume		2.906	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: O'DONOVAN

DATE/TIME OF TEST: 07/10/2016 04:17:43

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.079	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.128	04:12
Breath Volume		2.007	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.133	04:16
Breath Volume		2.812	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LOPEZ,FEDERICO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: PEREZ BERNAL

DATE/TIME OF TEST: 07/10/2016 04:47:00

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.169	04:42
Breath Volume		1.484	04:42
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.167	04:45
Breath Volume		1.382	04:45
Air Blank		0.000	04:46
Control Test		0.080	04:46
Air Blank		0.000	04:46
Diagnostics Check		OK	04:47

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHEAL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 07/16/2016 02:10:22

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/24/2016 02:49:11

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.189	02:44
Breath Volume		1.187	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.179	02:47
Breath Volume		1.132	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 07/27/2016 09:52

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TORRES,DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: TUCKLER

DATE/TIME OF TEST: 07/30/2016 23:58:06

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.222	23:52
Breath Volume		1.308	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:54
Subject Sample	# 2	0.219	23:56
Breath Volume		1.441	23:56
Air Blank		0.000	23:57
Control Test		0.079	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 08/02/2016 12:01

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: MENDEZ REYES

DATE/TIME OF TEST: 07/31/2016 01:35:42

Diagnostics Check		OK	01:29
Air Blank		0.000	01:29
Control Test		0.079	01:30
Air Blank		0.000	01:30
Subject Sample	# 1	0.034	01:31
Breath Volume		2.222	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.033	01:34
Breath Volume		2.531	01:34
Air Blank		0.000	01:34
Control Test		0.080	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 08/02/2016 12:02

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 07/31/2016 02:38:00

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.079	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.131	02:33
Breath Volume		1.296	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:35
Subject Sample	# 2	0.135	02:36
Breath Volume		1.175	02:36
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:37
Diagnostics Check		OK	02:38

Data Download Date/Time: 08/02/2016 12:02

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 07/31/2016 04:23:04

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.000	04:17
Breath Volume		1.164	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.000	04:21
Breath Volume		1.363	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:22
Diagnostics Check		OK	04:23

Data Download Date/Time: 08/02/2016 12:02

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: TRIANA

DATE/TIME OF TEST: 07/02/2016 00:21:19

Diagnostics Check		OK	00:14
Air Blank		0.000	00:15
Control Test		0.077	00:15
Air Blank		0.000	00:16
Subject Sample	# 1	0.165	00:16
Breath Volume		2.214	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:18
Subject Sample	# 2	0.150	00:19
Breath Volume		2.558	00:19
Air Blank		0.000	00:20
Control Test		0.077	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 07/07/2016 14:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: ROJAS RODRIGUEZ

DATE/TIME OF TEST: 07/02/2016 23:59:52

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.077	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.154	23:55
Breath Volume		1.636	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.154	23:58
Breath Volume		2.179	23:58
Air Blank		0.000	23:58
Control Test		0.076	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 07/07/2016 14:43

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: GUERRERO-ALVAREZ

DATE/TIME OF TEST: 07/07/2016 03:22:53

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.077	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.216	03:18
Breath Volume		2.058	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.205	03:21
Breath Volume		2.042	03:21
Air Blank		0.000	03:21
Control Test		0.076	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 07/07/2016 14:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: TOLEDO

DATE/TIME OF TEST: 07/09/2016 07:04:12

Diagnostics Check		OK	07:01	
Air Blank		0.000	07:01	
Control Test		0.077	07:02	
Air Blank		0.000	07:02	
Subject Sample	# 1		07:02	Subject Test Refused
Breath Volume		0.000	07:02	
Air Blank		0.000	07:03	
Control Test		0.077	07:03	
Air Blank		0.000	07:04	
Diagnostics Check		OK	07:04	

Data Download Date/Time: 08/02/2016 12:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 07/10/2016 02:34:57

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.077	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.117	02:30
Breath Volume		1.421	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.112	02:33
Breath Volume		1.566	02:33
Air Blank		0.000	02:34
Control Test		0.076	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 08/02/2016 12:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MILIAN,WANDA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: DELGADOALVAREZ

DATE/TIME OF TEST: 07/10/2016 02:44:50

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.139	02:40
Breath Volume		2.683	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:42
Subject Sample	# 2	0.143	02:43
Breath Volume		2.789	02:43
Air Blank		0.000	02:43
Control Test		0.078	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 08/02/2016 12:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/13/2016 06:18:25

Diagnostics Check		OK	06:10
Air Blank		0.000	06:11
Control Test		0.077	06:11
Air Blank		0.000	06:12
Subject Sample	# 1	0.192	06:12
Breath Volume		1.441	06:12
Air Blank		0.000	06:13
Air Blank		0.000	06:15
Subject Sample	# 2	0.198	06:16
Breath Volume		2.070	06:16
Air Blank		0.000	06:17
Control Test		0.076	06:17
Air Blank		0.000	06:18
Diagnostics Check		OK	06:18

Data Download Date/Time: 08/02/2016 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ALFONSO

DATE/TIME OF TEST: 07/27/2016 03:37:12

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.077	03:31
Air Blank		0.000	03:32
Subject Sample	# 1	0.135	03:32
Breath Volume		2.328	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.129	03:35
Breath Volume		1.738	03:35
Air Blank		0.000	03:36
Control Test		0.077	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 08/02/2016 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ENRIQUEZ

DATE/TIME OF TEST: 07/28/2016 04:51:39

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.077	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.136	04:46
Breath Volume		1.718	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.144	04:50
Breath Volume		2.847	04:50
Air Blank		0.000	04:50
Control Test		0.077	04:51
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 08/02/2016 12:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 07/30/2016 01:27:18

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.119	01:22
Breath Volume		3.773	01:22
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.113	01:25
Breath Volume		3.492	01:25
Air Blank		0.000	01:26
Control Test		0.076	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 08/02/2016 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: TRONCOSO

DATE/TIME OF TEST: 07/30/2016 04:54:44

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:53	
Control Test	0.069	04:54	Control Outside Tolerance
Air Blank	0.000	04:54	

Data Download Date/Time: 08/02/2016 12:11

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005550

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: TRONCOSO

DATE/TIME OF TEST: 07/30/2016 05:06:35

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.077	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.116	04:58
Breath Volume		2.195	04:58
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.093	05:01
Breath Volume		1.328	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 3	0.109	05:04
Breath Volume		1.925	05:04
Air Blank		0.000	05:05
Control Test		0.077	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 08/02/2016 12:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 07/01/2016 03:51:05

Diagnostics Check		OK	03:35	
Air Blank		0.000	03:36	
Control Test		0.078	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.103	03:40	Volume Not Met
Breath Volume		0.015	03:40	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 2	0.149	03:44	
Breath Volume		1.140	03:44	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 3	0.143	03:49	
Breath Volume		1.351	03:49	
Air Blank		0.000	03:50	
Control Test		0.077	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 07/07/2016 10:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: MONSALVE

DATE/TIME OF TEST: 07/02/2016 01:35:38

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.077	01:29
Air Blank		0.000	01:29
Subject Sample	# 1	0.229	01:31
Breath Volume		1.101	01:31
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 2	0.239	01:34
Breath Volume		1.386	01:34
Air Blank		0.000	01:34
Control Test		0.075	01:35
Air Blank		0.000	01:35
Diagnostics Check		OK	01:35

Data Download Date/Time: 07/07/2016 10:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MELO

DATE/TIME OF TEST: 07/03/2016 00:26:22

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.078	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.125	00:22
Breath Volume		1.308	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.136	00:24
Breath Volume		1.941	00:24
Air Blank		0.000	00:25
Control Test		0.078	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 07/07/2016 10:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 07/06/2016 05:27:25

Diagnostics Check		OK	05:21
Air Blank		0.000	05:21
Control Test		0.079	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.172	05:22
Breath Volume		2.042	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.174	05:25
Breath Volume		2.757	05:25
Air Blank		0.000	05:26
Control Test		0.077	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 07/07/2016 10:04

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CAPPE

DATE/TIME OF TEST: 07/07/2016 02:37:06

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.167	02:32
Breath Volume		1.183	02:32
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.175	02:35
Breath Volume		1.441	02:35
Air Blank		0.000	02:36
Control Test		0.076	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:37

Data Download Date/Time: 07/07/2016 10:04

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAMBOA VOLOJ,JUAN,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: GAMBOA VOLOJ

DATE/TIME OF TEST: 07/09/2016 05:18:23

Diagnostics Check		OK	05:11
Air Blank		0.000	05:12
Control Test		0.077	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.172	05:13
Breath Volume		4.378	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.165	05:16
Breath Volume		4.937	05:16
Air Blank		0.000	05:17
Control Test		0.076	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 08/17/2016 09:45

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GARCIA JR

DATE/TIME OF TEST: 07/14/2016 03:56:38

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.078	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.078	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 08/17/2016 09:45

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GARCIA JR

DATE/TIME OF TEST: 07/14/2016 05:22:43

Diagnostics Check		OK	05:14
Air Blank		0.000	05:15
Control Test		0.077	05:15
Air Blank		0.000	05:16
Subject Sample	# 1	0.113	05:17
Breath Volume		1.449	05:17
Air Blank		0.000	05:18
Air Blank		0.000	05:19
Subject Sample	# 2	0.113	05:21
Breath Volume		1.527	05:21
Air Blank		0.000	05:21
Control Test		0.077	05:22
Air Blank		0.000	05:22
Diagnostics Check		OK	05:22

Data Download Date/Time: 08/17/2016 09:45

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/15/2016 03:21:02

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:06	
Control Test		0.078	03:07	
Air Blank		0.000	03:07	
Subject Sample	# 1		03:10	No Sample Provided
Breath Volume		0.000	03:10	
Air Blank		0.000	03:11	
Air Blank		0.000	03:13	
Subject Sample	# 2	0.221	03:16	
Breath Volume		1.679	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:18	
Subject Sample	# 3	0.225	03:19	
Breath Volume		1.609	03:19	
Air Blank		0.000	03:20	
Control Test		0.078	03:20	
Air Blank		0.000	03:20	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 08/17/2016 09:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PONCE-OLIVARES

DATE/TIME OF TEST: 07/16/2016 01:32:08

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.078	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.063	01:27
Breath Volume		2.960	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.067	01:30
Breath Volume		2.937	01:30
Air Blank		0.000	01:31
Control Test		0.077	01:31
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 08/17/2016 09:45

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/27/2016 05:43:22

Diagnostics Check		OK	05:36
Air Blank		0.000	05:37
Control Test		0.078	05:37
Air Blank		0.000	05:38
Subject Sample	# 1	0.157	05:38
Breath Volume		2.320	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.156	05:41
Breath Volume		1.125	05:41
Air Blank		0.000	05:42
Control Test		0.077	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 08/17/2016 09:45

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BATISTA MARTINEZ

DATE/TIME OF TEST: 07/28/2016 02:34:58

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.076	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.202	02:30
Breath Volume		2.414	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.207	02:33
Breath Volume		2.402	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 08/17/2016 09:45

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 07/30/2016 03:36:42

Diagnostics Check		OK	03:30
Air Blank		0.000	03:31
Control Test		0.079	03:31
Air Blank		0.000	03:31
Subject Sample	# 1	0.153	03:32
Breath Volume		1.164	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.153	03:35
Breath Volume		1.992	03:35
Air Blank		0.000	03:35
Control Test		0.081	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 08/17/2016 09:45

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: MAGNOTTA

DATE/TIME OF TEST: 07/03/2016 05:46:30

Diagnostics Check		OK	05:35	
Air Blank		0.000	05:35	
Control Test		0.078	05:35	
Air Blank		0.000	05:36	
Subject Sample	# 1		05:39	No Sample Provided
Breath Volume		0.000	05:39	
Air Blank		0.000	05:39	
Air Blank		0.000	05:41	
Subject Sample	# 2		05:45	No Sample Provided
Breath Volume		0.000	05:45	
Air Blank		0.000	05:45	
Control Test		0.079	05:45	
Air Blank		0.000	05:46	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 07/25/2016 08:32

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: DELRIO,ADAM
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: ESCALANTE ROBLERO

DATE/TIME OF TEST: 07/10/2016 04:10:30

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.078	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.087	04:05
Breath Volume		1.308	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.083	04:08
Breath Volume		1.367	04:08
Air Blank		0.000	04:09
Control Test		0.077	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 07/25/2016 08:48

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: AYCOCK

DATE/TIME OF TEST: 07/18/2016 00:56:06

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.078	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank			00:55	RFI Detect
Air Blank		0.000	00:56	

Data Download Date/Time: 07/25/2016 08:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: AYCOCK

DATE/TIME OF TEST: 07/18/2016 01:25:16

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.078	01:23	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:24	
Control Test		0.079	01:24	
Air Blank		0.000	01:25	
Diagnostics Check		OK	01:25	

Data Download Date/Time: 07/25/2016 08:48

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KULCHER,KENT,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:59

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 07/18/2016 17:45:31

Diagnostics Check		OK	17:38
Air Blank		0.000	17:38
Control Test		0.078	17:39
Air Blank		0.000	17:39
Subject Sample	# 1	0.185	17:40
Breath Volume		1.328	17:40
Air Blank		0.000	17:40
Air Blank		0.000	17:42
Subject Sample	# 2	0.176	17:43
Breath Volume		2.511	17:43
Air Blank		0.000	17:44
Control Test		0.079	17:44
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 07/25/2016 08:49

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 07/23/2016 01:06:46

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.152	01:02
Breath Volume		2.550	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.155	01:05
Breath Volume		2.722	01:05
Air Blank		0.000	01:05
Control Test		0.078	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 07/25/2016 08:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BULLINGTON

DATE/TIME OF TEST: 07/23/2016 03:41:28

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.076	03:40	
Air Blank		03:40	RFI Detect
Air Blank	0.000	03:41	

Data Download Date/Time: 07/25/2016 08:49

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BULLINGTON

DATE/TIME OF TEST: 07/23/2016 03:49:33

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.163	03:45
Breath Volume		2.832	03:45
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.166	03:47
Breath Volume		2.351	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 07/25/2016 08:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: PASSLEY

DATE/TIME OF TEST: 07/28/2016 00:57:43

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.079	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.079	00:57	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 08/22/2016 08:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: PASSLEY

DATE/TIME OF TEST: 07/28/2016 01:14:52

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.137	01:06	Slope Not Level
Breath Volume		3.050	01:06	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.142	01:09	
Breath Volume		2.113	01:09	
Air Blank		0.000	01:09	
Air Blank		0.000	01:11	
Subject Sample	# 3		01:13	Subject Test Refused
Breath Volume		0.000	01:13	
Air Blank		0.000	01:13	
Control Test		0.079	01:14	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 08/22/2016 08:41

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MELANIE,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SULLIVAN,MELANIE
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CUEVAS

DATE/TIME OF TEST: 07/31/2016 02:52:05

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1	0.148	02:48	
Breath Volume		3.054	02:48	
Air Blank		0.000	02:49	
Air Blank			02:50	RFI Detect
Air Blank		0.000	02:51	
Control Test		0.079	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 08/22/2016 08:41

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,MELANIE,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SULLIVAN,MELANIE
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CUEVAS

DATE/TIME OF TEST: 07/31/2016 03:01:28

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.143	02:56
Breath Volume		1.937	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.139	02:59
Breath Volume		2.101	02:59
Air Blank		0.000	03:00
Control Test		0.078	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 08/22/2016 08:41

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PAZ

DATE/TIME OF TEST: 07/12/2016 03:23:02

Diagnostics Check		OK	03:14
Air Blank		0.000	03:14
Control Test		0.079	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.200	03:17
Breath Volume		2.437	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.194	03:21
Breath Volume		3.089	03:21
Air Blank		0.000	03:22
Control Test		0.077	03:22
Air Blank		0.000	03:22
Diagnostics Check		OK	03:23

Data Download Date/Time: 07/26/2016 14:59

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: READ

DATE/TIME OF TEST: 07/22/2016 02:02:36

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.080	01:58	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.000	01:59	
Breath Volume		1.644	01:59	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2	0.000	02:02	
Breath Volume		1.265	02:02	
Air Blank			02:02	RFI Detect
Air Blank		0.000	02:02	

Data Download Date/Time: 07/26/2016 14:59

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005935

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: READ

DATE/TIME OF TEST: 07/22/2016 02:10:25

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.000	02:06
Breath Volume		1.796	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.000	02:09
Breath Volume		1.367	02:09
Air Blank		0.000	02:09
Control Test		0.077	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 07/26/2016 14:59

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: BRISENO

DATE/TIME OF TEST: 07/30/2016 00:37:59

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.178	00:33
Breath Volume		2.164	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:35
Subject Sample	# 2	0.177	00:36
Breath Volume		1.679	00:36
Air Blank		0.000	00:37
Control Test		0.077	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 08/30/2016 10:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 07/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GARCIA SAMAYOA

DATE/TIME OF TEST: 07/31/2016 01:46:42

Diagnostics Check		OK	01:40
Air Blank		0.000	01:40
Control Test		0.078	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.184	01:42
Breath Volume		2.535	01:42
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.181	01:45
Breath Volume		2.886	01:45
Air Blank		0.000	01:45
Control Test		0.078	01:46
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 08/30/2016 10:41

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DELEEUW,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13314080A3
DRY GAS EXPIRY DATE: 06/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 07/15/2016 22:57:58

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.082	22:50	
Air Blank		0.000	22:50	
Subject Sample	# 1		22:53	No Sample Provided
Breath Volume		0.000	22:53	
Air Blank		0.000	22:54	
Air Blank		0.000	22:56	
Subject Sample	# 2		22:56	Subject Test Refused
Breath Volume		0.000	22:56	
Air Blank		0.000	22:57	
Control Test		0.081	22:57	
Air Blank		0.000	22:57	
Diagnostics Check		OK	22:57	

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLASCO,LUIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:17

SUBJECT'S LAST NAME: PEREZ-NOLASCO

DATE/TIME OF TEST: 07/17/2016 18:04:39

Diagnostics Check		OK	17:57
Air Blank		0.000	17:58
Control Test		0.082	17:58
Air Blank		0.000	17:58
Subject Sample	# 1	0.189	17:59
Breath Volume		1.601	17:59
Air Blank		0.000	18:00
Air Blank		0.000	18:02
Subject Sample	# 2	0.185	18:03
Breath Volume		1.632	18:03
Air Blank		0.000	18:03
Control Test		0.080	18:04
Air Blank		0.000	18:04
Diagnostics Check		OK	18:04

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/17/2016 22:08:38

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.082	22:00
Air Blank		0.000	22:01
Subject Sample	# 1	0.169	22:02
Breath Volume		1.714	22:02
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.171	22:06
Breath Volume		3.015	22:06
Air Blank		0.000	22:07
Control Test		0.083	22:08
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CORDERO-PEREZ

DATE/TIME OF TEST: 07/18/2016 00:20:18

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.082	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.118	00:15
Breath Volume		1.847	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.116	00:18
Breath Volume		1.812	00:18
Air Blank		0.000	00:19
Control Test		0.081	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VARONE,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 16:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: RAMOS-PEREIRA

DATE/TIME OF TEST: 07/20/2016 17:34:47

Diagnostics Check		OK	17:27
Air Blank		0.000	17:28
Control Test		0.082	17:28
Air Blank		0.000	17:28
Subject Sample	# 1	0.248	17:30
Breath Volume		1.863	17:30
Air Blank		0.000	17:30
Air Blank		0.000	17:32
Subject Sample	# 2	0.247	17:33
Breath Volume		1.296	17:33
Air Blank		0.000	17:33
Control Test		0.081	17:34
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MESIATSEV

DATE/TIME OF TEST: 07/21/2016 01:22:52

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:09	
Control Test		0.082	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 1	0.151	01:13	Volume Not Met
Breath Volume		0.464	01:13	
Air Blank		0.000	01:14	
Air Blank		0.000	01:15	
Subject Sample	# 2	0.183	01:16	
Breath Volume		1.535	01:16	
Air Blank		0.000	01:17	
Air Blank		0.000	01:19	
Subject Sample	# 3	0.182	01:21	
Breath Volume		1.632	01:21	
Air Blank		0.000	01:21	
Control Test		0.081	01:22	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 07/22/2016 03:35:51

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.082	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.164	03:31
Breath Volume		2.507	03:31
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.168	03:34
Breath Volume		2.703	03:34
Air Blank		0.000	03:34
Control Test		0.082	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HUGHES,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RAMIREZ LUCAS

DATE/TIME OF TEST: 07/22/2016 23:58:18

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.083	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.220	23:52
Breath Volume		1.394	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.220	23:56
Breath Volume		1.667	23:56
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHAPARRO,CARLOS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MATEOANTONIO

DATE/TIME OF TEST: 07/23/2016 02:23:31

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.190	02:18
Breath Volume		1.425	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:20
Subject Sample	# 2	0.206	02:21
Breath Volume		1.671	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PIMENTEL,LOISE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: NICARAGUA DE SPANYOL

DATE/TIME OF TEST: 07/23/2016 04:19:32

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.082	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.180	04:14
Breath Volume		1.714	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.177	04:17
Breath Volume		1.492	04:17
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: PARDEW

DATE/TIME OF TEST: 07/23/2016 22:25:18

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.082	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.111	22:20
Breath Volume		3.589	22:20
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.110	22:23
Breath Volume		3.437	22:23
Air Blank		0.000	22:24
Control Test		0.081	22:24
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON,ADAM
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 14:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:19

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/24/2016 16:37:17

Diagnostics Check		OK	16:30
Air Blank		0.000	16:30
Control Test		0.082	16:30
Air Blank		0.000	16:31
Subject Sample	# 1	0.321	16:32
Breath Volume		2.246	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.321	16:35
Breath Volume		1.835	16:35
Air Blank		0.000	16:36
Control Test		0.080	16:36
Air Blank		0.000	16:37
Diagnostics Check		OK	16:37

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHAPARRO,CARLOS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/24/2016 17:20:20

Diagnostics Check		OK	17:13
Air Blank		0.000	17:13
Control Test		0.082	17:13
Air Blank		0.000	17:14
Subject Sample	# 1	0.172	17:15
Breath Volume		3.605	17:15
Air Blank		0.000	17:16
Air Blank		0.000	17:18
Subject Sample	# 2	0.169	17:18
Breath Volume		3.468	17:18
Air Blank		0.000	17:19
Control Test		0.080	17:19
Air Blank		0.000	17:20
Diagnostics Check		OK	17:20

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 07/25/2016 02:19:06

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.082	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.211	02:14
Breath Volume		1.164	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.211	02:17
Breath Volume		1.613	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ESCALANTE ROBLERO

DATE/TIME OF TEST: 07/25/2016 04:23:50

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.082	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.170	04:19
Breath Volume		1.144	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:21
Subject Sample	# 2	0.184	04:22
Breath Volume		1.777	04:22
Air Blank		0.000	04:22
Control Test		0.078	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PERALTA

DATE/TIME OF TEST: 07/27/2016 01:56:33

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.082	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Control Test		0.082	01:55	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FURNESS,SCOTT
ARRESTING OFFICER AGENCY: Atlantis PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 07/27/2016 02:26:25

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.083	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.212	02:20
Breath Volume		1.960	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.194	02:24
Breath Volume		2.046	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: ROCHA IBARRA

DATE/TIME OF TEST: 07/27/2016 22:42:17

Diagnostics Check		OK	22:35
Air Blank		0.000	22:36
Control Test		0.082	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.165	22:37
Breath Volume		3.589	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.165	22:40
Breath Volume		3.570	22:40
Air Blank		0.000	22:41
Control Test		0.081	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCMAHON,GEORGE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:27

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 07/28/2016 07:06:20

Diagnostics Check		OK	06:59
Air Blank		0.000	07:00
Control Test		0.082	07:00
Air Blank		0.000	07:00
Subject Sample	# 1	0.214	07:01
Breath Volume		3.777	07:01
Air Blank		0.000	07:02
Air Blank		0.000	07:04
Subject Sample	# 2	0.199	07:04
Breath Volume		1.199	07:04
Air Blank		0.000	07:05
Control Test		0.081	07:05
Air Blank		0.000	07:06
Diagnostics Check		OK	07:06

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: VENTO

DATE/TIME OF TEST: 07/29/2016 02:19:55

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.082	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.082	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,5
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CROY,PAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 07:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:41

SUBJECT'S LAST NAME: FORTUNE

DATE/TIME OF TEST: 07/29/2016 08:11:31

Diagnostics Check		OK	08:05
Air Blank		0.000	08:05
Control Test		0.082	08:05
Air Blank		0.000	08:06
Subject Sample	# 1	0.086	08:07
Breath Volume		1.453	08:07
Air Blank		0.000	08:07
Air Blank		0.000	08:09
Subject Sample	# 2	0.088	08:09
Breath Volume		1.941	08:09
Air Blank		0.000	08:10
Control Test		0.082	08:10
Air Blank		0.000	08:11
Diagnostics Check		OK	08:11

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLASCO,LUIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 07/29/2016 12:56:40

Diagnostics Check		OK	12:50
Air Blank		0.000	12:50
Control Test		0.083	12:51
Air Blank		0.000	12:51
Subject Sample	# 1	0.000	12:52
Breath Volume		1.339	12:52
Air Blank		0.000	12:52
Air Blank		0.000	12:54
Subject Sample	# 2	0.000	12:55
Breath Volume		2.007	12:55
Air Blank		0.000	12:55
Control Test		0.081	12:56
Air Blank		0.000	12:56
Diagnostics Check		OK	12:56

Data Download Date/Time: 08/05/2016 12:31

Printed: 12 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 14:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/29/2016 15:04:23

Diagnostics Check		OK	14:57
Air Blank		0.000	14:57
Control Test		0.082	14:57
Air Blank		0.000	14:58
Subject Sample	# 1	0.191	14:59
Breath Volume		1.308	14:59
Air Blank		0.000	15:00
Air Blank		0.000	15:01
Subject Sample	# 2	0.184	15:02
Breath Volume		1.441	15:02
Air Blank		0.000	15:03
Control Test		0.081	15:03
Air Blank		0.000	15:04
Diagnostics Check		OK	15:04

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DIEGO MIGUEL

DATE/TIME OF TEST: 07/30/2016 00:17:03

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.081	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.288	00:12
Breath Volume		2.167	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:14
Subject Sample	# 2	0.289	00:15
Breath Volume		2.320	00:15
Air Blank		0.000	00:16
Control Test		0.081	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:17

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: ORDONES PEREZ

DATE/TIME OF TEST: 07/30/2016 01:02:55

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.082	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.261	00:58
Breath Volume		2.390	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.260	01:01
Breath Volume		2.445	01:01
Air Blank		0.000	01:01
Control Test		0.081	01:02
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WEBER,RONALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 06:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: BISONO

DATE/TIME OF TEST: 07/30/2016 07:03:56

Diagnostics Check		OK	06:57
Air Blank		0.000	06:58
Control Test		0.083	06:58
Air Blank		0.000	06:58
Subject Sample	# 1	0.136	06:59
Breath Volume		1.125	06:59
Air Blank		0.000	07:00
Air Blank		0.000	07:01
Subject Sample	# 2	0.140	07:02
Breath Volume		1.554	07:02
Air Blank		0.000	07:03
Control Test		0.082	07:03
Air Blank		0.000	07:03
Diagnostics Check		OK	07:03

Data Download Date/Time: 08/05/2016 12:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006085

INSTRUMENT REGISTERED TO: US COAST GUARD MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURKOWSKI,PAUL,J
BREATH TEST OPERATOR AGENCY: USCG Miami
ARRESTING OFFICER: SARMIENTO,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: BUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: SCANDRETH

DATE/TIME OF TEST: 07/16/2016 19:34:31

Diagnostics Check		OK	19:31	
Air Blank		0.000	19:32	
Control Test		0.080	19:32	
Air Blank		0.000	19:32	
Subject Sample	# 1		19:33	Subject Test Refused
Breath Volume		0.000	19:33	
Air Blank		0.000	19:33	
Control Test		0.079	19:33	
Air Blank		0.000	19:34	
Diagnostics Check		OK	19:34	

Data Download Date/Time: 07/26/2016 11:59

COBRA ©1997-2014 CMI, Inc.