Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MURGUIDO,ALEJANDRO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: BORRERO

DATE/TIME OF TEST: 07/01/2016 18:18:09

Diagnostics Check		OK	18:11
Air Blank		0.000	18:11
Control Test		0.081	18:12
Air Blank		0.000	18:12
Subject Sample	# 1	0.092	18:13
Breath Volume		1.835	18:13
Air Blank		0.000	18:14
Air Blank		0.000	18:16
Subject Sample	# 2	0.085	18:16
Breath Volume		1.644	18:16
Air Blank		0.000	18:17
Control Test		0.078	18:17
Air Blank		0.000	18:18
Diagnostics Check		OK	18:18

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: STEPHENSON,JENNIFER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: TUM US

DATE/TIME OF TEST: 07/03/2016 05:11:58

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.082	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.148	05:07
Breath Volume		2.578	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.148	05:10
Breath Volume		3.160	05:10
Air Blank		0.000	05:11
Control Test		0.081	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,JORGE,E
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: DAVIS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: MARTIN AGUILAR

DATE/TIME OF TEST: 07/10/2016 19:26:46

Diagnostics Check		OK	19:23	
Air Blank		0.000	19:23	
Control Test		0.080	19:23	
Air Blank		0.000	19:24	
Subject Sample	# 1	0.215	19:25	
Breath Volume		2.937	19:25	
Air Blank			19:25	RFI Detect
Air Blank		0.000	19:25	
Control Test		0.080	19:26	
Air Blank		0.000	19:26	
Diagnostics Check		OK	19:26	

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,JORGE,E
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: DAVIS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: MARTIN AGUILAR

DATE/TIME OF TEST: 07/10/2016 19:36:54

Diagnostics Check		OK	19:30
Air Blank		0.000	19:30
Control Test		0.081	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.217	19:32
Breath Volume		2.730	19:32
Air Blank		0.000	19:32
Air Blank		0.000	19:34
Subject Sample	# 2	0.214	19:35
Breath Volume		2.761	19:35
Air Blank		0.000	19:35
Control Test		0.081	19:36
Air Blank		0.000	19:36
Diagnostics Check		OK	19:36

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RAMIREZ VASQUEZ

DATE/TIME OF TEST: 07/16/2016 00:51:42

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.302	00:45
Breath Volume		1.421	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.291	00:49
Breath Volume		1.265	00:49
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,JORGE,E
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,CARLOS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: RICHARD

DATE/TIME OF TEST: 07/17/2016 22:41:28

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:33	
Control Test		0.082	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 1		22:37	No Sample Provided
Breath Volume		0.000	22:37	
Air Blank		0.000	22:37	
Air Blank		0.000	22:39	
Subject Sample	# 2		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:40	
Control Test		0.081	22:40	
Air Blank		0.000	22:41	
Diagnostics Check		OK	22:41	

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEWTON,JOSHUA,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: DUCKSWORTH,D.
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 07/19/2016 00:42:59

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.081	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.182	00:38
Breath Volume		2.585	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.180	00:41
Breath Volume		2.976	00:41
Air Blank		0.000	00:42
Control Test		0.081	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: PENA ALONZO

DATE/TIME OF TEST: 07/23/2016 03:27:13

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:23	No Sample Provided
Breath Volume		0.000	03:23	
Air Blank		0.000	03:23	
Air Blank		0.000	03:25	
Subject Sample	# 2		03:25	Subject Test Refused
Breath Volume		0.000	03:25	
Air Blank		0.000	03:26	
Control Test		0.082	03:26	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 07/28/2016 07:21

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENSON,JENNIFER,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: STEPHENSON,JENNIFER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 07/31/2016 04:55:56

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.142	04:51
Breath Volume		1.363	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.141	04:54
Breath Volume		1.394	04:54
Air Blank		0.000	04:55
Control Test		0.081	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 08/11/2016 10:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MARTINEZ RUPERTO

DATE/TIME OF TEST: 07/03/2016 03:00:31

Diagnostics Check	OK	03:00	
Air Blank		03:00	RFI Detect
Air Blank		03:00	RFI Detect

Data Download Date/Time: 07/13/2016 22:03

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MARTINEZ RUPERTO

DATE/TIME OF TEST: 07/03/2016 03:19:31

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.076	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:12	No Sample Provided
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Control Test		0.076	03:18	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 07/13/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MARTINEZ RUPERTO

DATE/TIME OF TEST: 07/03/2016 03:28:45

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.076	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.123	03:24
Breath Volume		1.304	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.123	03:27
Breath Volume		2.460	03:27
Air Blank		0.000	03:27
Control Test		0.076	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 07/13/2016 22:03

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PALACIOS CASTILLO

DATE/TIME OF TEST: 07/17/2016 02:58:21

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:52	
Control Test		0.076	02:53	
Air Blank		0.000	02:53	
Subject Sample	# 1	0.192	02:54	
Breath Volume		2.980	02:54	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2		02:57	RFI Detect
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.076	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 08/12/2016 21:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: PALACIOS CASTILLO

DATE/TIME OF TEST: 07/17/2016 03:06:33

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.076	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.198	03:02
Breath Volume		2.671	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.194	03:05
Breath Volume		1.554	03:05
Air Blank		0.000	03:05
Control Test		0.075	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 08/12/2016 21:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 07/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: TACKETT

DATE/TIME OF TEST: 07/28/2016 01:47:46

Diagnostics Check		OK	01:41
Air Blank		0.000	01:41
Control Test		0.077	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.194	01:43
Breath Volume		1.375	01:43
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.177	01:46
Breath Volume		1.121	01:46
Air Blank		0.000	01:46
Control Test		0.076	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 08/12/2016 21:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: ASTATULA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALINSKI,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: PALINSKI,TIMOTHY
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00816080A1
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 07/07/2016 21:54:47

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:53	
Control Test		0.080	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:54	Sequence Aborted
Breath Volume		0.000	21:54	
Air Blank		0.000	21:54	

Data Download Date/Time: 07/22/2016 03:20

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: ASTATULA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALINSKI,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: STEWART,D
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00816080A1
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 07/17/2016 21:45:00

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.080	21:38
Air Blank		0.000	21:38
Subject Sample	# 1	0.068	21:40
Breath Volume		1.292	21:40
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.060	21:43
Breath Volume		1.527	21:43
Air Blank		0.000	21:44
Control Test		0.079	21:44
Air Blank		0.000	21:44
Diagnostics Check		OK	21:45

Data Download Date/Time: 07/22/2016 03:20

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006222

INSTRUMENT REGISTERED TO: ASTATULA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALINSKI,TIMOTHY,A
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: PALINSKI,TIMOTHY
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 07/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 00816080A1
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/31/2016 00:39:09

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.134	00:34
Breath Volume		2.679	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.124	00:37
Breath Volume		1.605	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 08/29/2016 03:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY,THOMAS,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/02/2016 02:52:46

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:47
Subject Sample	# 1	0.068	02:48
Breath Volume		2.230	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.061	02:51
Breath Volume		1.660	02:51
Air Blank		0.000	02:51
Control Test		0.076	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: MATTINGLY

DATE/TIME OF TEST: 07/02/2016 21:42:37

Diagnostics Check		OK	21:34
Air Blank		0.000	21:34
Control Test		0.077	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.210	21:37
Breath Volume		2.667	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:39
Subject Sample	# 2	0.212	21:40
Breath Volume		2.058	21:40
Air Blank		0.000	21:41
Control Test		0.076	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: BURKHARDT

DATE/TIME OF TEST: 07/07/2016 01:01:47

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.078	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.241	00:57
Breath Volume		3.273	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.237	01:00
Breath Volume		3.250	01:00
Air Blank		0.000	01:00
Control Test		0.077	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 07/21/2016 06:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CAWTHON,EVAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 07/07/2016 02:16:02

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:07	
Control Test		0.077	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1	0.221	02:10	
Breath Volume		1.308	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:13	
Subject Sample	# 2	0.224	02:14	
Breath Volume		1.656	02:14	
Air Blank		0.000	02:15	
Control Test		0.074	02:15	Control Outside Tolerance
Air Blank		0.000	02:16	

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CAWTHON,EVAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 07/07/2016 02:33:17

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.075	02:19	
Air Blank		0.000	02:19	
Subject Sample	# 1	0.199	02:22	Volume Not Met
Breath Volume		0.632	02:22	
Air Blank		0.000	02:23	
Air Blank		0.000	02:25	
Subject Sample	# 2	0.202	02:26	
Breath Volume		1.671	02:26	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 3	0.204	02:31	
Breath Volume		1.308	02:31	
Air Blank		0.000	02:32	
Control Test		0.075	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZYCKI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: BARNEY

DATE/TIME OF TEST: 07/09/2016 01:40:58

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:24	
Control Test		0.078	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1	0.125	01:28	Volume Not Met
Breath Volume		0.050	01:28	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 2	0.110	01:33	
Breath Volume		1.394	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 3	0.113	01:39	Volume Not Met
Breath Volume		0.613	01:39	
Air Blank		0.000	01:40	
Control Test		0.075	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZYCKI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HEUSSY

DATE/TIME OF TEST: 07/10/2016 04:26:43

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.077	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.160	04:22
Breath Volume		1.867	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.157	04:25
Breath Volume		2.496	04:25
Air Blank		0.000	04:25
Control Test		0.076	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,JOSH,G
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: REDFERN

DATE/TIME OF TEST: 07/13/2016 04:00:36

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.078	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1	0.160	03:55	
Breath Volume		3.261	03:55	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2	0.155	03:58	
Breath Volume		3.144	03:58	
Air Blank		0.000	03:59	
Control Test		0.073	04:00	Control Outside Tolerance
Air Blank		0.000	04:00	

Data Download Date/Time: 07/21/2016 06:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,JOSH,G
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: REDFERN

DATE/TIME OF TEST: 07/13/2016 04:10:27

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.076	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.158	04:05
Breath Volume		3.046	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.155	04:08
Breath Volume		3.085	04:08
Air Blank		0.000	04:09
Control Test		0.076	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY,THOMAS,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: REYNOLDS

DATE/TIME OF TEST: 07/17/2016 20:15:24

Diagnostics Check		OK	20:06
Air Blank		0.000	20:07
Control Test		0.077	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.167	20:08
Breath Volume		1.312	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:10
Subject Sample	# 2	0.164	20:13
Breath Volume		1.468	20:13
Air Blank		0.000	20:14
Control Test		0.077	20:14
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASE,THOMAS,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARBER,AMANDA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/19/2016 23:05:50

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.077	23:02	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:05	Sequence Aborted
Breath Volume		0.117	23:05	
Air Blank		0.000	23:05	

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: KIMBROUGH

DATE/TIME OF TEST: 07/21/2016 02:05:55

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.077	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1	0.127	01:58	Volume Not Met
Breath Volume		0.078	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2	0.128	02:04	Volume Not Met
Breath Volume		0.101	02:04	
Air Blank		0.000	02:04	
Control Test		0.076	02:05	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 07/21/2016 06:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CAWTHON,EVAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ROMANELLI

DATE/TIME OF TEST: 07/21/2016 02:25:28

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:09	
Control Test		0.075	02:09	
Air Blank		0.000	02:10	
Subject Sample	# 1	0.224	02:12	
Breath Volume		1.195	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 2	0.180	02:18	Volume Not Met
Breath Volume		0.464	02:18	
Air Blank		0.000	02:18	
Air Blank		0.000	02:20	
Subject Sample	# 3	0.222	02:23	Volume Not Met
Breath Volume		0.019	02:23	
Air Blank		0.000	02:24	
Control Test		0.074	02:24	Control Outside Tolerance
Air Blank		0.000	02:25	

Data Download Date/Time: 07/21/2016 06:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CAWTHON,EVAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ROMANELLI

DATE/TIME OF TEST: 07/21/2016 02:30:41

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.073	02:30	Control Outside Tolerance
Air Blank	0.000	02:30	

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CAWTHON,EVAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: ROMANELLI

DATE/TIME OF TEST: 07/21/2016 02:45:26

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:33	
Control Test		0.076	02:34	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.161	02:37	Volume Not Met
Breath Volume		0.347	02:37	
Air Blank		0.000	02:38	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.200	02:43	Volume Not Met
Breath Volume		0.300	02:43	
Air Blank		0.000	02:44	
Control Test		0.075	02:44	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT,TYLER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG425202
DRY GAS EXPIRY DATE: 09/09/2016
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 07/21/2016 03:57:28

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.075	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.037	03:50
Breath Volume		1.250	03:50
Air Blank		0.000	03:51
Air Blank		0.000	03:53
Subject Sample	# 2	0.036	03:55
Breath Volume		1.367	03:55
Air Blank		0.000	03:56
Control Test		0.076	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 07/21/2016 06:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: EMERSON

DATE/TIME OF TEST: 07/24/2016 03:06:38

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.078	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.134	03:00
Breath Volume		1.203	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:02
Subject Sample	# 2	0.124	03:05
Breath Volume		2.167	03:05
Air Blank		0.000	03:05
Control Test		0.077	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 08/24/2016 19:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: CABRERA MOURE

DATE/TIME OF TEST: 07/30/2016 02:26:32

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.078	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.000	02:21
Breath Volume		1.828	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.000	02:25
Breath Volume		2.050	02:25
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 08/24/2016 19:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: SOLIZ

DATE/TIME OF TEST: 07/31/2016 05:02:07

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.078	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.135	04:57
Breath Volume		2.062	04:57
Air Blank		0.000	04:58
Air Blank		0.000	04:59
Subject Sample	# 2	0.149	05:00
Breath Volume		1.964	05:00
Air Blank		0.000	05:01
Control Test		0.078	05:01
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 08/24/2016 19:35

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/02/2016 04:12:30

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.108	04:07
Breath Volume		2.406	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.095	04:10
Breath Volume		1.695	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 07/03/2016 00:09:03

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.081	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Control Test		0.082	00:08	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: ELVER,GEORGE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/03/2016 05:49:58

Diagnostics Check		OK	05:42
Air Blank		0.000	05:43
Control Test		0.081	05:43
Air Blank		0.000	05:44
Subject Sample	# 1	0.188	05:44
Breath Volume		1.976	05:44
Air Blank		0.000	05:45
Air Blank		0.000	05:47
Subject Sample	# 2	0.197	05:48
Breath Volume		1.542	05:48
Air Blank		0.000	05:49
Control Test		0.079	05:49
Air Blank		0.000	05:49
Diagnostics Check		OK	05:49

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: HORVATH

DATE/TIME OF TEST: 07/04/2016 02:09:26

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.081	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.081	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 07/30/2016 20:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: HODGSON

DATE/TIME OF TEST: 07/07/2016 04:31:31

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.218	04:23
Breath Volume		2.179	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.195	04:26
Breath Volume		1.738	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:28
Subject Sample	# 3	0.211	04:29
Breath Volume		2.507	04:29
Air Blank		0.000	04:30
Control Test		0.078	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,DANIEL,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: CRUZ,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: MCGOVERN

DATE/TIME OF TEST: 07/12/2016 00:29:45

Diagnostics Check		OK	00:23
Air Blank		0.000	00:23
Control Test		0.081	00:23
Air Blank		0.000	00:24
Subject Sample	# 1	0.075	00:25
Breath Volume		1.421	00:25
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.069	00:28
Breath Volume		1.718	00:28
Air Blank		0.000	00:28
Control Test		0.079	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: COREY,GAVIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/16/2016 00:31:43

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.081	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.061	00:26
Breath Volume		1.691	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.053	00:30
Breath Volume		1.785	00:30
Air Blank		0.000	00:30
Control Test		0.078	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: SETTLE

DATE/TIME OF TEST: 07/16/2016 02:54:35

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.180	02:49
Breath Volume		2.554	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.180	02:52
Breath Volume		2.000	02:52
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,JOSHUA,H
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CAMERON,D
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/19/2016 03:52:17

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.081	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.134	03:47
Breath Volume		1.625	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.136	03:50
Breath Volume		1.355	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELVER,GEORGE,T
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,ANDREW
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: BANNING

DATE/TIME OF TEST: 07/21/2016 21:18:37

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.081	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.087	21:14
Breath Volume		3.835	21:14
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.083	21:17
Breath Volume		3.484	21:17
Air Blank		0.000	21:17
Control Test		0.079	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 07/30/2016 20:27

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 07/01/2016 00:01:59

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.080	23:59	
Air Blank		0.000	23:59	
Subject Sample	# 1	0.191	00:00	
Breath Volume		2.562	00:00	
Air Blank			00:00	RFI Detect
Air Blank		0.000	00:01	
Control Test		0.080	00:01	
Air Blank		0.000	00:01	
Diagnostics Check		OK	00:01	

Data Download Date/Time: 07/03/2016 13:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 07/01/2016 00:11:20

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.079	00:05
Air Blank		0.000	00:06
Subject Sample	# 1	0.199	00:06
Breath Volume		1.781	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:09
Subject Sample	# 2	0.197	00:09
Breath Volume		1.714	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 07/03/2016 13:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 07/01/2016 05:13:58

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.179	05:09
Breath Volume		4.027	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:11
Subject Sample	# 2	0.179	05:12
Breath Volume		3.164	05:12
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 07/03/2016 13:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GLENN,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 07/07/2016 01:29:48

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.146	01:24
Breath Volume		2.480	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.149	01:28
Breath Volume		3.929	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 08/24/2016 17:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLF,B,S
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SWIM,G
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: BRANTLEY

DATE/TIME OF TEST: 07/07/2016 10:37:35

Diagnostics Check		OK	10:30
Air Blank		0.000	10:31
Control Test		0.079	10:31
Air Blank		0.000	10:32
Subject Sample	# 1	0.217	10:32
Breath Volume		2.542	10:32
Air Blank		0.000	10:33
Air Blank		0.000	10:35
Subject Sample	# 2	0.233	10:35
Breath Volume		1.980	10:35
Air Blank		0.000	10:36
Control Test		0.080	10:36
Air Blank		0.000	10:37
Diagnostics Check		OK	10:37

Data Download Date/Time: 08/24/2016 17:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANUDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: KAY

DATE/TIME OF TEST: 07/07/2016 17:39:10

Diagnostics Check		OK	17:32
Air Blank		0.000	17:33
Control Test		0.080	17:33
Air Blank		0.000	17:33
Subject Sample	# 1	0.000	17:34
Breath Volume		2.640	17:34
Air Blank		0.000	17:35
Air Blank		0.000	17:37
Subject Sample	# 2	0.000	17:37
Breath Volume		2.132	17:37
Air Blank		0.000	17:38
Control Test		0.080	17:38
Air Blank		0.000	17:39
Diagnostics Check		OK	17:39

Data Download Date/Time: 08/24/2016 17:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CLOUSER,BRANT
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 07/08/2016 00:50:27

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.080	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.197	00:45
Breath Volume		1.738	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:47
Subject Sample	# 2	0.187	00:48
Breath Volume		1.554	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 08/24/2016 17:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,CHARLES,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARWOOD,CHARLES
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 07/09/2016 09:38:39

Diagnostics Check		OK	09:31
Air Blank		0.000	09:32
Control Test		0.079	09:32
Air Blank		0.000	09:33
Subject Sample	# 1	0.173	09:33
Breath Volume		2.718	09:33
Air Blank		0.000	09:34
Air Blank		0.000	09:36
Subject Sample	# 2	0.177	09:36
Breath Volume		2.679	09:36
Air Blank		0.000	09:37
Control Test		0.080	09:38
Air Blank		0.000	09:38
Diagnostics Check		OK	09:38

Data Download Date/Time: 08/24/2016 17:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARWOOD,C,M
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HARWOOD,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: VELLECA

DATE/TIME OF TEST: 07/14/2016 16:13:27

Diagnostics Check		OK	16:07
Air Blank		0.000	16:07
Control Test		0.079	16:08
Air Blank		0.000	16:08
Subject Sample	# 1	0.000	16:09
Breath Volume		4.011	16:09
Air Blank		0.000	16:09
Air Blank		0.000	16:11
Subject Sample	# 2	0.000	16:12
Breath Volume		3.132	16:12
Air Blank		0.000	16:12
Control Test		0.079	16:12
Air Blank		0.000	16:13
Diagnostics Check		OK	16:13

Data Download Date/Time: 08/24/2016 17:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: LABRADA

DATE/TIME OF TEST: 07/22/2016 21:20:09

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.080	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.086	21:14
Breath Volume		1.218	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.091	21:18
Breath Volume		2.070	21:18
Air Blank		0.000	21:19
Control Test		0.079	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 08/24/2016 17:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BARBER,J
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: BUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CANTRELL

DATE/TIME OF TEST: 07/24/2016 22:46:40

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.080	22:36	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:39	No Sample Provided
Breath Volume		0.000	22:39	
Air Blank		0.000	22:40	
Air Blank		0.000	22:42	
Subject Sample	# 2		22:45	Subject Test Refused
Breath Volume		0.000	22:45	
Air Blank		0.000	22:45	
Control Test		0.081	22:46	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 08/24/2016 17:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZIER,TAMMY,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,CHASE
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GREER

DATE/TIME OF TEST: 07/27/2016 01:50:16

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.000	01:45
Breath Volume		1.242	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.000	01:48
Breath Volume		1.390	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 08/24/2016 17:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male

OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: LORENZO

DATE/TIME OF TEST: 07/02/2016 02:13:25

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:10	
Control Test		0.082	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:12	Subject Test Refused
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Control Test		0.081	02:12	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:13	

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CRUZ,GARDENIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 10:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: MARTINEZ VILLATORO

DATE/TIME OF TEST: 07/02/2016 11:49:38

Diagnostics Check		OK	11:42
Air Blank		0.000	11:43
Control Test		0.082	11:43
Air Blank		0.000	11:44
Subject Sample	# 1	0.137	11:45
Breath Volume		3.093	11:45
Air Blank		0.000	11:45
Air Blank		0.000	11:47
Subject Sample	# 2	0.137	11:48
Breath Volume		4.492	11:48
Air Blank		0.000	11:48
Control Test		0.081	11:49
Air Blank		0.000	11:49
Diagnostics Check		OK	11:49

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITA-RODRIGUEZ,LENDDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/03/2016 04:38:56

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:28	
Control Test		0.082	04:28	
Air Blank		0.000	04:29	
Subject Sample	# 1	0.236	04:32	Volume Not Met
Breath Volume		0.152	04:32	
Air Blank		0.000	04:33	
Air Blank		0.000	04:34	
Subject Sample	# 2		04:37	Subject Test Refused
Breath Volume		0.355	04:37	
Air Blank		0.000	04:37	
Control Test		0.080	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GODDEN,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 07/03/2016 05:26:06

Diagnostics Check		OK	05:19
Air Blank		0.000	05:20
Control Test		0.081	05:20
Air Blank		0.000	05:21
Subject Sample	# 1	0.045	05:21
Breath Volume		2.519	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.047	05:24
Breath Volume		2.511	05:24
Air Blank		0.000	05:25
Control Test		0.082	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,JOSE
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 07/03/2016 06:23:48

Diagnostics Check		OK	06:17
Air Blank		0.000	06:17
Control Test		0.082	06:18
Air Blank		0.000	06:18
Subject Sample	# 1	0.140	06:19
Breath Volume		2.320	06:19
Air Blank		0.000	06:20
Air Blank		0.000	06:21
Subject Sample	# 2	0.138	06:22
Breath Volume		2.277	06:22
Air Blank		0.000	06:22
Control Test		0.081	06:23
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PALEVODA,GARY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 07/04/2016 00:50:37

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.082	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.085	00:46
Breath Volume		1.117	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.089	00:49
Breath Volume		1.800	00:49
Air Blank		0.000	00:49
Control Test		0.082	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PANCZAK,BRIAN
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 07/04/2016 22:18:42

Diagnostics Check		OK	22:12
Air Blank		0.000	22:12
Control Test		0.082	22:12
Air Blank		0.000	22:13
Subject Sample	# 1	0.054	22:13
Breath Volume		1.363	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:16
Subject Sample	# 2	0.057	22:17
Breath Volume		1.320	22:17
Air Blank		0.000	22:17
Control Test		0.082	22:18
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HILL,JEREMY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: FORERO

DATE/TIME OF TEST: 07/07/2016 05:39:24

Diagnostics Check		OK	05:32
Air Blank		0.000	05:33
Control Test		0.082	05:33
Air Blank		0.000	05:34
Subject Sample	# 1	0.086	05:34
Breath Volume		2.617	05:34
Air Blank		0.000	05:35
Air Blank		0.000	05:37
Subject Sample	# 2	0.086	05:37
Breath Volume		2.593	05:37
Air Blank		0.000	05:38
Control Test		0.082	05:38
Air Blank		0.000	05:39
Diagnostics Check		OK	05:39

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCDONOUGH,MARK
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MATHEWSON BROWN

DATE/TIME OF TEST: 07/10/2016 05:15:16

Diagnostics Check		OK	05:08
Air Blank		0.000	05:08
Control Test		0.082	05:09
Air Blank		0.000	05:09
Subject Sample	# 1	0.078	05:10
Breath Volume		3.949	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.070	05:13
Breath Volume		1.269	05:13
Air Blank		0.000	05:14
Control Test		0.080	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CUMMINGS,PETER
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: COLEMON

DATE/TIME OF TEST: 07/12/2016 00:00:17

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.082	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.000	23:55
Breath Volume		1.765	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:58
Breath Volume		1.898	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ISHAM,MATTHEW
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 07/12/2016 01:23:30

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.082	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.164	01:19
Breath Volume		2.535	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.163	01:21
Breath Volume		2.511	01:21
Air Blank		0.000	01:22
Control Test		0.082	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ARCE,MOISES
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 07/14/2016 23:11:02

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.083	23:05
Air Blank		0.000	23:06
Subject Sample	# 1	0.000	23:06
Breath Volume		1.351	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.000	23:09
Breath Volume		1.601	23:09
Air Blank		0.000	23:10
Control Test		0.082	23:10
Air Blank		0.000	23:10
Diagnostics Check		OK	23:11

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: CHAMPANERIA

DATE/TIME OF TEST: 07/15/2016 03:45:02

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.083	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.010	03:40
Breath Volume		2.734	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.011	03:43
Breath Volume		3.347	03:43
Air Blank		0.000	03:44
Control Test		0.082	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:45

Data Download Date/Time: 07/15/2016 12:33

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHAPARRO,CARLOS
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: TEJADA-ALMANZAR

DATE/TIME OF TEST: 07/15/2016 19:16:26

Diagnostics Check		OK	19:10
Air Blank		0.000	19:10
Control Test		0.082	19:10
Air Blank		0.000	19:11
Subject Sample	# 1	0.175	19:11
Breath Volume		2.375	19:11
Air Blank		0.000	19:12
Air Blank		0.000	19:14
Subject Sample	# 2	0.177	19:14
Breath Volume		2.546	19:14
Air Blank		0.000	19:15
Control Test		0.081	19:15
Air Blank		0.000	19:16
Diagnostics Check		OK	19:16

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 12:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:27

SUBJECT'S LAST NAME: BUSSEY

DATE/TIME OF TEST: 07/16/2016 13:59:48

Diagnostics Check		OK	13:53
Air Blank		0.000	13:54
Control Test		0.082	13:54
Air Blank		0.000	13:55
Subject Sample	# 1	0.000	13:55
Breath Volume		2.707	13:55
Air Blank		0.000	13:56
Air Blank		0.000	13:58
Subject Sample	# 2	0.000	13:58
Breath Volume		1.718	13:58
Air Blank		0.000	13:58
Control Test		0.082	13:59
Air Blank		0.000	13:59
Diagnostics Check		OK	13:59

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DELEEUW,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: NITZ

DATE/TIME OF TEST: 07/16/2016 20:43:38

Diagnostics Check		OK	20:37
Air Blank		0.000	20:37
Control Test		0.082	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.208	20:39
Breath Volume		3.324	20:39
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.211	20:42
Breath Volume		2.609	20:42
Air Blank		0.000	20:42
Control Test		0.082	20:43
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: VAIL LOPEZ

DATE/TIME OF TEST: 07/16/2016 22:59:15

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.082	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.207	22:54
Breath Volume		1.296	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.200	22:57
Breath Volume		1.167	22:57
Air Blank		0.000	22:58
Control Test		0.081	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PAJON,JOSE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 08:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: CANALES JR.

DATE/TIME OF TEST: 07/17/2016 09:49:55

Diagnostics Check		OK	09:43
Air Blank		0.000	09:43
Control Test		0.082	09:43
Air Blank		0.000	09:44
Subject Sample	# 1	0.202	09:45
Breath Volume		1.484	09:45
Air Blank		0.000	09:45
Air Blank		0.000	09:47
Subject Sample	# 2	0.193	09:48
Breath Volume		1.796	09:48
Air Blank		0.000	09:49
Control Test		0.082	09:49
Air Blank		0.000	09:49
Diagnostics Check		OK	09:49

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KWAKU,EDWARD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: EL-MAHDAWY

DATE/TIME OF TEST: 07/17/2016 10:25:45

Diagnostics Check		OK	10:19
Air Blank		0.000	10:20
Control Test		0.082	10:20
Air Blank		0.000	10:20
Subject Sample	# 1	0.000	10:21
Breath Volume		4.093	10:21
Air Blank		0.000	10:21
Air Blank		0.000	10:23
Subject Sample	# 2	0.000	10:24
Breath Volume		4.054	10:24
Air Blank		0.000	10:24
Control Test		0.081	10:25
Air Blank		0.000	10:25
Diagnostics Check		OK	10:25

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LAMPE,ERIC
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: LEIB

DATE/TIME OF TEST: 07/19/2016 23:54:19

Diagnostics Check		OK	23:46
Air Blank		0.000	23:46
Control Test		0.082	23:47
Air Blank		0.000	23:47
Subject Sample	# 1	0.156	23:48
Breath Volume		1.386	23:48
Air Blank		0.000	23:49
Air Blank		0.000	23:51
Subject Sample	# 2	0.164	23:52
Breath Volume		1.539	23:52
Air Blank		0.000	23:53
Control Test		0.082	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COE,S
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: WERNER

DATE/TIME OF TEST: 07/20/2016 11:31:59

Diagnostics Check		OK	11:24
Air Blank		0.000	11:25
Control Test		0.082	11:25
Air Blank		0.000	11:26
Subject Sample	# 1	0.000	11:27
Breath Volume		4.000	11:27
Air Blank		0.000	11:27
Air Blank		0.000	11:29
Subject Sample	# 2	0.000	11:30
Breath Volume		3.207	11:30
Air Blank		0.000	11:31
Control Test		0.082	11:31
Air Blank		0.000	11:31
Diagnostics Check		OK	11:31

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: KILLGORE

DATE/TIME OF TEST: 07/21/2016 04:38:32

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.082	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.166	04:33
Breath Volume		2.074	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:35
Subject Sample	# 2	0.169	04:36
Breath Volume		2.449	04:36
Air Blank		0.000	04:37
Control Test		0.081	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: HOUZE

DATE/TIME OF TEST: 07/21/2016 19:46:04

Diagnostics Check		OK	19:38
Air Blank		0.000	19:38
Control Test		0.082	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.234	19:40
Breath Volume		1.707	19:40
Air Blank		0.000	19:41
Air Blank		0.000	19:43
Subject Sample	# 2	0.222	19:44
Breath Volume		1.167	19:44
Air Blank		0.000	19:45
Control Test		0.081	19:45
Air Blank		0.000	19:45
Diagnostics Check		OK	19:46

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RIEGER,MITHCELL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: OGDEN

DATE/TIME OF TEST: 07/22/2016 13:02:54

Diagnostics Check		OK	12:57
Air Blank		0.000	12:57
Control Test		0.082	12:57
Air Blank		0.000	12:58
Subject Sample	# 1	0.000	12:58
Breath Volume		1.117	12:58
Air Blank		0.000	12:59
Air Blank		0.000	13:01
Subject Sample	# 2	0.000	13:01
Breath Volume		1.636	13:01
Air Blank		0.000	13:01
Control Test		0.082	13:02
Air Blank		0.000	13:02
Diagnostics Check		OK	13:02

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: REYNOLDS,PETER
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: BIRDSALL

DATE/TIME OF TEST: 07/22/2016 23:57:05

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.082	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:52
Breath Volume		3.207	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		3.652	23:55
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 07/23/2016 02:04:45

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.082	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.225	01:59
Breath Volume		2.503	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:02
Subject Sample	# 2	0.225	02:03
Breath Volume		1.765	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,GARY,W
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CADRERA

DATE/TIME OF TEST: 07/24/2016 03:30:32

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.082	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.118	03:26
Breath Volume		2.535	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.123	03:28
Breath Volume		2.988	03:28
Air Blank		0.000	03:29
Control Test		0.081	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,GARY,W
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZEITZ,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: LOPEZ RIVAS

DATE/TIME OF TEST: 07/24/2016 04:27:02

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.228	04:22
Breath Volume		2.308	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.214	04:25
Breath Volume		1.375	04:25
Air Blank		0.000	04:26
Control Test		0.081	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:27

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/26/2016 00:04:36

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.082	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		2.992	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:03
Breath Volume		2.582	00:03
Air Blank		0.000	00:03
Control Test		0.082	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WILSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: UHIA

DATE/TIME OF TEST: 07/27/2016 19:21:00

Diagnostics Check		OK	19:14
Air Blank		0.000	19:14
Control Test		0.082	19:15
Air Blank		0.000	19:15
Subject Sample	# 1	0.173	19:16
Breath Volume		2.375	19:16
Air Blank		0.000	19:16
Air Blank		0.000	19:18
Subject Sample	# 2	0.173	19:19
Breath Volume		2.171	19:19
Air Blank		0.000	19:20
Control Test		0.083	19:20
Air Blank		0.000	19:20
Diagnostics Check		OK	19:21

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BEESLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/28/2016 05:05:08

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.082	04:58
Air Blank		0.000	04:58
Subject Sample	# 1	0.160	05:00
Breath Volume		2.203	05:00
Air Blank		0.000	05:00
Air Blank		0.000	05:02
Subject Sample	# 2	0.154	05:03
Breath Volume		1.601	05:03
Air Blank		0.000	05:04
Control Test		0.082	05:04
Air Blank		0.000	05:05
Diagnostics Check		OK	05:05

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: BUITRAGO GOMEZ

DATE/TIME OF TEST: 07/28/2016 06:49:32

Diagnostics Check		OK	06:42
Air Blank		0.000	06:42
Control Test		0.081	06:43
Air Blank		0.000	06:43
Subject Sample	# 1	0.143	06:44
Breath Volume		1.402	06:44
Air Blank		0.000	06:45
Air Blank		0.000	06:47
Subject Sample	# 2	0.150	06:47
Breath Volume		2.007	06:47
Air Blank		0.000	06:48
Control Test		0.081	06:48
Air Blank		0.000	06:49
Diagnostics Check		OK	06:49

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GODDEN,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 11:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:17

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 07/28/2016 12:51:08

Diagnostics Check		OK	12:44
Air Blank		0.000	12:45
Control Test		0.082	12:45
Air Blank		0.000	12:46
Subject Sample	# 1	0.036	12:46
Breath Volume		1.492	12:46
Air Blank		0.000	12:47
Air Blank		0.000	12:49
Subject Sample	# 2	0.040	12:49
Breath Volume		1.582	12:49
Air Blank		0.000	12:50
Control Test		0.081	12:50
Air Blank		0.000	12:51
Diagnostics Check		OK	12:51

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BOUCHER

DATE/TIME OF TEST: 07/28/2016 23:12:14

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.082	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.152	23:07
Breath Volume		4.246	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:09
Subject Sample	# 2	0.154	23:10
Breath Volume		3.292	23:10
Air Blank		0.000	23:11
Control Test		0.081	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,E.K.
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RAMOS JUAREZ

DATE/TIME OF TEST: 07/29/2016 01:26:07

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.081	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.186	01:21
Breath Volume		2.261	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.181	01:24
Breath Volume		1.730	01:24
Air Blank		0.000	01:25
Control Test		0.081	01:25
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ACUNA

DATE/TIME OF TEST: 07/29/2016 04:30:09

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:20	
Control Test		0.082	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1	0.194	04:24	Volume Not Met
Breath Volume		1.078	04:24	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 2		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:29	
Control Test		0.080	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ORGAZ MESA

DATE/TIME OF TEST: 07/30/2016 01:11:58

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.266	01:07
Breath Volume		2.816	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.271	01:10
Breath Volume		2.792	01:10
Air Blank		0.000	01:11
Control Test		0.082	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MALDONADO,MARK
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 07/30/2016 07:44:03

Diagnostics Check		OK	07:32	
Air Blank		0.000	07:32	
Control Test		0.082	07:33	
Air Blank		0.000	07:33	
Subject Sample	# 1	0.216	07:36	Volume Not Met
Breath Volume		0.968	07:36	
Air Blank		0.000	07:37	
Air Blank		0.000	07:39	
Subject Sample	# 2	0.202	07:42	Volume Not Met
Breath Volume		0.203	07:42	
Air Blank		0.000	07:43	
Control Test		0.080	07:43	
Air Blank		0.000	07:43	
Diagnostics Check		OK	07:44	

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MALDONADO,MARK
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:54

SUBJECT'S LAST NAME: IVERY

DATE/TIME OF TEST: 07/30/2016 07:58:34

Diagnostics Check		OK	07:46	
Air Blank		0.000	07:47	
Control Test		0.080	07:47	
Air Blank		0.000	07:48	
Subject Sample	# 1	0.189	07:51	Volume Not Met
Breath Volume		0.304	07:51	
Air Blank		0.000	07:52	
Air Blank		0.000	07:53	
Subject Sample	# 2	0.179	07:56	Volume Not Met
Breath Volume		0.121	07:56	
Air Blank		0.000	07:57	
Control Test		0.080	07:57	
Air Blank		0.000	07:58	
Diagnostics Check		OK	07:58	

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RANGEL

DATE/TIME OF TEST: 07/30/2016 20:34:31

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.081	20:28
Air Blank		0.000	20:29
Subject Sample	# 1	0.271	20:29
Breath Volume		3.449	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.278	20:32
Breath Volume		2.734	20:32
Air Blank		0.000	20:33
Control Test		0.081	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 08/05/2016 14:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,GARY,W
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PUGA-DIAZ

DATE/TIME OF TEST: 07/31/2016 02:22:07

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.082	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.206	02:17
Breath Volume		2.316	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.214	02:20
Breath Volume		2.375	02:20
Air Blank		0.000	02:21
Control Test		0.081	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/05/2016 14:49

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: CORBIN,STEPHEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:41

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 07/12/2016 19:33:01

Diagnostics Check		OK	19:26
Air Blank		0.000	19:26
Control Test		0.080	19:27
Air Blank		0.000	19:27
Subject Sample	# 1	0.203	19:28
Breath Volume		1.773	19:28
Air Blank		0.000	19:28
Air Blank		0.000	19:30
Subject Sample	# 2	0.196	19:31
Breath Volume		1.757	19:31
Air Blank		0.000	19:32
Control Test		0.079	19:32
Air Blank		0.000	19:32
Diagnostics Check		OK	19:33

Data Download Date/Time: 08/24/2016 23:50

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 07/13/2016 01:06:04

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.078	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.284	00:59
Breath Volume		2.652	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.275	01:04
Breath Volume		2.929	01:04
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:06

Data Download Date/Time: 08/24/2016 23:50

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,JASON,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: ESCOBAR,RUBEN
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: FEENEY

DATE/TIME OF TEST: 07/24/2016 02:40:02

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.127	02:35
Breath Volume		3.671	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.128	02:38
Breath Volume		2.726	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:40

Data Download Date/Time: 08/24/2016 23:50

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BROGAN-WEHNER

DATE/TIME OF TEST: 07/31/2016 00:52:26

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.273	00:47
Breath Volume		2.066	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.267	00:50
Breath Volume		2.253	00:50
Air Blank		0.000	00:51
Control Test		0.080	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 08/24/2016 23:50

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: RONCHETTO

DATE/TIME OF TEST: 07/08/2016 01:42:26

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.224	01:37
Breath Volume		3.031	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.216	01:40
Breath Volume		1.781	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/20/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: RAINES

DATE/TIME OF TEST: 07/09/2016 01:06:52

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:51	
Control Test		0.080	00:51	
Air Blank		0.000	00:52	
Subject Sample	# 1		00:55	No Sample Provided
Breath Volume		0.000	00:55	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2	0.141	01:00	
Breath Volume		1.792	01:00	
Air Blank		0.000	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 3	0.151	01:05	
Breath Volume		1.906	01:05	
Air Blank		0.000	01:05	
Control Test		0.080	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 07/20/2016 22:02

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 07/09/2016 23:40:37

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.081	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:36	No Sample Provided
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.081	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 07/20/2016 22:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: KAUFMAN

DATE/TIME OF TEST: 07/01/2016 01:26:54

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.079	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.186	01:21
Breath Volume		1.265	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.190	01:25
Breath Volume		2.585	01:25
Air Blank		0.000	01:25
Control Test		0.078	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: HERNANDEZ VEGA

DATE/TIME OF TEST: 07/01/2016 05:43:21

Diagnostics Check		OK	05:36
Air Blank		0.000	05:36
Control Test		0.079	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.191	05:38
Breath Volume		1.328	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.191	05:41
Breath Volume		1.890	05:41
Air Blank		0.000	05:42
Control Test		0.078	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: YURICH

DATE/TIME OF TEST: 07/02/2016 00:27:18

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Subject Sample	# 1	0.179	00:22
Breath Volume		2.402	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:24
Subject Sample	# 2	0.165	00:25
Breath Volume		1.410	00:25
Air Blank		0.000	00:26
Control Test		0.078	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BODAMER,E
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/02/2016 03:40:46

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.078	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.176	03:35
Breath Volume		2.136	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.176	03:39
Breath Volume		1.609	03:39
Air Blank		0.000	03:39
Control Test		0.077	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: HEANEY

DATE/TIME OF TEST: 07/02/2016 05:00:21

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.078	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.158	04:55
Breath Volume		2.269	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:57
Subject Sample	# 2	0.159	04:58
Breath Volume		2.707	04:58
Air Blank		0.000	04:59
Control Test		0.078	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: MARTINEZ-MARTINEZ

DATE/TIME OF TEST: 07/03/2016 01:40:35

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.225	01:32
Breath Volume		1.410	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.246	01:35
Breath Volume		2.019	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 3	0.247	01:38
Breath Volume		3.039	01:38
Air Blank		0.000	01:39
Control Test		0.077	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: WILLIAMS JR

DATE/TIME OF TEST: 07/03/2016 03:44:50

Diagnostics Check		OK	03:38
Air Blank		0.000	03:38
Control Test		0.076	03:38
Air Blank		0.000	03:39
Subject Sample	# 1	0.149	03:40
Breath Volume		2.792	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.158	03:43
Breath Volume		3.449	03:43
Air Blank		0.000	03:43
Control Test		0.077	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: UNO,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 07/03/2016 04:58:12

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.079	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.169	04:53
Breath Volume		2.503	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.175	04:56
Breath Volume		1.910	04:56
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 07/03/2016 18:46:11

Diagnostics Check		OK	18:39
Air Blank		0.000	18:39
Control Test		0.079	18:40
Air Blank		0.000	18:40
Subject Sample	# 1	0.000	18:41
Breath Volume		1.289	18:41
Air Blank		0.000	18:42
Air Blank		0.000	18:44
Subject Sample	# 2	0.000	18:44
Breath Volume		2.816	18:44
Air Blank		0.000	18:45
Control Test		0.078	18:45
Air Blank		0.000	18:46
Diagnostics Check		OK	18:46

Data Download Date/Time: 07/13/2016 09:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CESPEDES JR

DATE/TIME OF TEST: 07/04/2016 01:48:34

Diagnostics Check		OK	01:37
Air Blank		0.000	01:38
Control Test		0.079	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.167	01:42
Breath Volume		3.207	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:44
Subject Sample	# 2	0.159	01:46
Breath Volume		1.164	01:46
Air Blank		0.000	01:47
Control Test		0.077	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: HUERTAS

DATE/TIME OF TEST: 07/04/2016 04:43:01

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.079	04:37
Air Blank		0.000	04:37
Subject Sample	# 1	0.261	04:38
Breath Volume		1.515	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:40
Subject Sample	# 2	0.260	04:41
Breath Volume		1.566	04:41
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:43

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 07/04/2016 22:55:18

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.079	22:48
Air Blank		0.000	22:49
Subject Sample	# 1	0.125	22:50
Breath Volume		2.062	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.137	22:53
Breath Volume		2.406	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FOWLER,J
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: IBARRA

DATE/TIME OF TEST: 07/05/2016 05:47:32

Diagnostics Check		OK	05:43	
Air Blank		0.000	05:44	
Control Test		0.079	05:44	
Air Blank		0.000	05:45	
Subject Sample	# 1		05:46	Subject Test Refused
Breath Volume		0.000	05:46	
Air Blank		0.000	05:46	
Control Test		0.079	05:46	
Air Blank		0.000	05:47	
Diagnostics Check		OK	05:47	

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LONG II

DATE/TIME OF TEST: 07/06/2016 02:56:23

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.079	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.161	02:51
Breath Volume		1.328	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.157	02:54
Breath Volume		1.722	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 07/08/2016 00:10:07

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.000	00:05
Breath Volume		1.406	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.000	00:08
Breath Volume		1.476	00:08
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 07/09/2016 02:21:40

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.080	02:14
Air Blank		0.000	02:14
Subject Sample	# 1	0.116	02:16
Breath Volume		3.429	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.121	02:20
Breath Volume		3.511	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 07/13/2016 09:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: NICODEMUS

DATE/TIME OF TEST: 07/09/2016 04:49:31

Diagnostics Check		OK	04:42
Air Blank		0.000	04:42
Control Test		0.079	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.129	04:44
Breath Volume		2.023	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.125	04:47
Breath Volume		1.242	04:47
Air Blank		0.000	04:48
Control Test		0.078	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: ANGELES GARRIDO

DATE/TIME OF TEST: 07/09/2016 05:36:03

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.170	05:31
Breath Volume		1.316	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:33
Subject Sample	# 2	0.181	05:34
Breath Volume		1.199	05:34
Air Blank		0.000	05:35
Control Test		0.078	05:35
Air Blank		0.000	05:35
Diagnostics Check		OK	05:36

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: GARCIA JUAREZ

DATE/TIME OF TEST: 07/09/2016 21:44:23

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:40	
Control Test		0.079	21:41	
Air Blank		0.000	21:41	
Subject Sample	# 1		21:42	Subject Test Refused
Breath Volume		0.000	21:42	
Air Blank		0.000	21:43	
Control Test		0.079	21:43	
Air Blank		0.000	21:44	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DIAZ HERNANDEZ

DATE/TIME OF TEST: 07/10/2016 02:30:28

Diagnostics Check		OK	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:18	
Subject Sample	# 1	0.261	02:22	Slope Not Level
Breath Volume		1.882	02:22	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 2	0.272	02:25	
Breath Volume		2.222	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 3	0.284	02:28	
Breath Volume		1.808	02:28	
Air Blank		0.000	02:29	
Control Test		0.077	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: MADRIGALES

DATE/TIME OF TEST: 07/10/2016 05:09:10

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.078	05:03
Air Blank		0.000	05:03
Subject Sample	# 1	0.236	05:04
Breath Volume		2.503	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.232	05:07
Breath Volume		2.695	05:07
Air Blank		0.000	05:08
Control Test		0.078	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRITO,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:32

SUBJECT'S LAST NAME: BORJA

DATE/TIME OF TEST: 07/10/2016 15:06:40

Diagnostics Check		OK	15:00
Air Blank		0.000	15:00
Control Test		0.079	15:00
Air Blank		0.000	15:01
Subject Sample	# 1	0.306	15:01
Breath Volume		2.855	15:01
Air Blank		0.000	15:02
Air Blank		0.000	15:04
Subject Sample	# 2	0.302	15:04
Breath Volume		1.691	15:04
Air Blank		0.000	15:05
Control Test		0.078	15:06
Air Blank		0.000	15:06
Diagnostics Check		OK	15:06

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: LEATHERWOOD

DATE/TIME OF TEST: 07/11/2016 04:36:54

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.164	04:32
Breath Volume		2.230	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:34
Subject Sample	# 2	0.156	04:35
Breath Volume		1.355	04:35
Air Blank		0.000	04:35
Control Test		0.076	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/13/2016 09:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: JUDGE THARPE,N/A
ARRESTING OFFICER AGENCY: OTHER County
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: PATE

DATE/TIME OF TEST: 07/11/2016 10:19:09

Diagnostics Check		OK	10:09
Air Blank		0.000	10:09
Control Test		0.079	10:10
Air Blank		0.000	10:10
Subject Sample	# 1	0.224	10:11
Breath Volume		3.062	10:11
Air Blank		0.000	10:11
Air Blank		0.000	10:13
Subject Sample	# 2	0.201	10:14
Breath Volume		1.160	10:14
Air Blank		0.000	10:15
Air Blank		0.000	10:16
Subject Sample	# 3	0.213	10:17
Breath Volume		2.164	10:17
Air Blank		0.000	10:18
Control Test		0.078	10:18
Air Blank		0.000	10:19
Diagnostics Check		OK	10:19

Data Download Date/Time: 07/13/2016 09:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD,ANTHONY,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HOWARD,ANTHONY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 07/14/2016 01:55:49

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.080	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.158	01:50
Breath Volume		1.539	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:52
Subject Sample	# 2	0.154	01:54
Breath Volume		1.328	01:54
Air Blank		0.000	01:54
Control Test		0.078	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOWRY,BRYAN,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LINDBLOM,ROBERT
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 07/23/2016 02:01:22

Diagnostics Check		OK	01:56	
Air Blank		0.000	01:56	
Control Test		0.081	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:00	
Control Test		0.081	02:00	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: BRAKE

DATE/TIME OF TEST: 07/23/2016 22:28:39

Diagnostics Check		OK	22:22
Air Blank		0.000	22:23
Control Test		0.081	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.000	22:24
Breath Volume		1.191	22:24
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.000	22:27
Breath Volume		1.949	22:27
Air Blank		0.000	22:27
Control Test		0.079	22:28
Air Blank		0.000	22:28
Diagnostics Check		OK	22:28

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROOP,CHAD,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ROOP,CHAD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/29/2016 21:31:53

Diagnostics Check		OK	21:24
Air Blank		0.000	21:25
Control Test		0.080	21:25
Air Blank		0.000	21:26
Subject Sample	# 1	0.201	21:27
Breath Volume		1.128	21:27
Air Blank		0.000	21:27
Air Blank		0.000	21:29
Subject Sample	# 2	0.204	21:30
Breath Volume		1.707	21:30
Air Blank		0.000	21:30
Control Test		0.080	21:31
Air Blank		0.000	21:31
Diagnostics Check		OK	21:31

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: POOL,HAROLD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 07/30/2016 01:41:23

Diagnostics Check		OK	01:38	
Air Blank		0.000	01:38	
Control Test		0.080	01:39	
Air Blank		0.000	01:39	
Subject Sample	# 1		01:40	Subject Test Refused
Breath Volume		0.000	01:40	
Air Blank		0.000	01:40	
Control Test		0.080	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: POOL,HAROLD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 07/30/2016 03:33:29

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.081	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.109	03:28
Breath Volume		4.082	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.109	03:31
Breath Volume		1.757	03:31
Air Blank		0.000	03:32
Control Test		0.081	03:32
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: ALVAREZ,ERIC
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 07/30/2016 04:08:48

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.081	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Control Test		0.081	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 07/31/2016 01:13:16

Diagnostics Check		OK	01:06
Air Blank		0.000	01:07
Control Test		0.081	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.171	01:08
Breath Volume		2.402	01:08
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.170	01:11
Breath Volume		1.976	01:11
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 08/23/2016 10:34

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,MITCHELL,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: STEVENS,MITCHELL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/01/2016 00:06:54

Diagnostics Check		OK	23:58	
Air Blank		0.000	23:58	
Control Test		0.078	23:59	
Air Blank		0.000	23:59	
Subject Sample	# 1	0.185	00:01	
Breath Volume		1.273	00:01	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2	0.178	00:05	
Breath Volume		1.527	00:05	
Air Blank		0.000	00:05	
Control Test		0.072	00:06	Control Outside Tolerance
Air Blank		0.000	00:06	

Data Download Date/Time: 07/12/2016 08:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,MITCHELL,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: STEVENS,MITCHELL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/01/2016 00:19:02

Diagnostics Check		OK	00:11	
Air Blank		0.000	00:12	
Control Test		0.077	00:12	
Air Blank		0.000	00:13	
Subject Sample	# 1	0.186	00:14	
Breath Volume		1.343	00:14	
Air Blank		0.000	00:14	
Air Blank		0.000	00:16	
Subject Sample	# 2	0.174	00:17	
Breath Volume		1.398	00:17	
Air Blank		0.000	00:18	
Control Test		0.074	00:18	Control Outside Tolerance
Air Blank		0.000	00:19	

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 07/01/2016 16:04:22

Diagnostics Check		OK	15:58
Air Blank		0.000	15:58
Control Test		0.078	15:59
Air Blank		0.000	15:59
Subject Sample	# 1	0.000	16:00
Breath Volume		3.718	16:00
Air Blank		0.000	16:00
Air Blank		0.000	16:02
Subject Sample	# 2	0.000	16:02
Breath Volume		4.039	16:02
Air Blank		0.000	16:03
Control Test		0.077	16:03
Air Blank		0.000	16:04
Diagnostics Check		OK	16:04

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMBERT,BRIAN,K
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAMBERT,BRIAN
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LIND

DATE/TIME OF TEST: 07/03/2016 00:41:00

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.078	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.024	00:36
Breath Volume		1.964	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.023	00:39
Breath Volume		1.417	00:39
Air Blank		0.000	00:40
Control Test		0.078	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:41

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: RESTIFO,CHARLES
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: TARBUCK

DATE/TIME OF TEST: 07/03/2016 21:26:43

Diagnostics Check		OK	21:19
Air Blank		0.000	21:20
Control Test		0.079	21:20
Air Blank		0.000	21:20
Subject Sample	# 1	0.000	21:21
Breath Volume		1.246	21:21
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.000	21:25
Breath Volume		2.757	21:25
Air Blank		0.000	21:25
Control Test		0.078	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 07/04/2016 00:02:39

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:51	
Control Test		0.077	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 2		23:55	No Sample Provided
Breath Volume		0.000	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:58	
Subject Sample	# 1		00:01	No Sample Provided
Breath Volume		0.000	00:01	
Air Blank		0.000	00:01	
Control Test		0.077	00:02	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 07/12/2016 08:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HOLCOMBE,JOHN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03415080A2
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BUTCHER

DATE/TIME OF TEST: 07/04/2016 20:38:54

Diagnostics Check		OK	20:32
Air Blank		0.000	20:33
Control Test		0.078	20:33
Air Blank		0.000	20:34
Subject Sample	# 1	0.064	20:34
Breath Volume		2.230	20:34
Air Blank		0.000	20:35
Air Blank		0.000	20:36
Subject Sample	# 2	0.063	20:37
Breath Volume		2.351	20:37
Air Blank		0.000	20:37
Control Test		0.076	20:38
Air Blank		0.000	20:38
Diagnostics Check		OK	20:38

Data Download Date/Time: 07/12/2016 08:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 07/05/2016 20:35:10

Diagnostics Check		OK	20:28
Air Blank		0.000	20:29
Control Test		0.079	20:29
Air Blank		0.000	20:29
Subject Sample	# 1	0.000	20:30
Breath Volume		3.304	20:30
Air Blank		0.000	20:31
Air Blank		0.000	20:33
Subject Sample	# 2	0.000	20:33
Breath Volume		2.187	20:33
Air Blank		0.000	20:34
Control Test		0.078	20:34
Air Blank		0.000	20:35
Diagnostics Check		OK	20:35

Data Download Date/Time: 07/12/2016 08:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HOBBY

DATE/TIME OF TEST: 07/06/2016 01:57:47

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.079	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1	0.084	01:54	
Breath Volume		1.144	01:54	
Air Blank		0.000	01:54	
Air Blank		0.000	01:56	
Subject Sample	# 2	0.082	01:56	
Breath Volume		1.449	01:56	
Air Blank			01:57	RFI Detect
Air Blank		0.000	01:57	

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HOBBY

DATE/TIME OF TEST: 07/06/2016 02:07:48

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.078	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.083	02:02
Breath Volume		1.300	02:02
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.079	02:06
Breath Volume		1.769	02:06
Air Blank		0.000	02:06
Control Test		0.077	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 07/12/2016 08:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,KRISTAVIA,P
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEE,MIKEL
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/08/2016 01:18:01

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:07	
Control Test		0.078	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 1	0.147	01:11	Volume Not Met
Breath Volume		0.035	01:11	
Air Blank		0.000	01:11	
Air Blank		0.000	01:13	
Subject Sample	# 2		01:16	Subject Test Refused
Breath Volume		0.035	01:16	
Air Blank		0.000	01:17	
Control Test		0.078	01:17	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 07/08/2016 03:02:10

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.084	02:57
Breath Volume		1.792	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.080	03:00
Breath Volume		1.187	03:00
Air Blank		0.000	03:01
Control Test		0.076	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUDSON,JOSHUA,K
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HUDSON,JOSHUA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WINDLEY

DATE/TIME OF TEST: 07/09/2016 03:08:28

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.078	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.183	03:03
Breath Volume		1.335	03:03
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.185	03:06
Breath Volume		2.480	03:06
Air Blank		0.000	03:07
Control Test		0.077	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEARD,JOHN,K
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HADDOCKS,KEVIN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: LOFTON

DATE/TIME OF TEST: 07/11/2016 03:46:01

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.077	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.181	03:40
Breath Volume		3.351	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.179	03:44
Breath Volume		3.707	03:44
Air Blank		0.000	03:45
Control Test		0.077	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:46

Data Download Date/Time: 07/12/2016 08:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VASQUEZ,GERBER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: MC LEAN

DATE/TIME OF TEST: 07/02/2016 05:52:46

Diagnostics Check		OK	05:43	
Air Blank		0.000	05:43	
Control Test		0.080	05:44	
Air Blank		0.000	05:44	
Subject Sample	# 1		05:47	No Sample Provided
Breath Volume		0.000	05:47	
Air Blank		0.000	05:48	
Air Blank		0.000	05:50	
Subject Sample	# 2		05:51	Subject Test Refused
Breath Volume		0.000	05:51	
Air Blank		0.000	05:51	
Control Test		0.080	05:52	
Air Blank		0.000	05:52	
Diagnostics Check		OK	05:52	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,A,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,F
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: MATUTE

DATE/TIME OF TEST: 07/03/2016 02:59:40

Diagnostics Check		OK	02:53
Air Blank		0.000	02:53
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.198	02:55
Breath Volume		2.785	02:55
Air Blank		0.000	02:55
Air Blank		0.000	02:57
Subject Sample	# 2	0.200	02:58
Breath Volume		1.691	02:58
Air Blank		0.000	02:58
Control Test		0.079	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MOISES,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/03/2016 05:40:23

Diagnostics Check		OK	05:31
Air Blank		0.000	05:31
Control Test		0.079	05:32
Air Blank		0.000	05:32
Subject Sample	# 1	0.274	05:33
Breath Volume		2.367	05:33
Air Blank		0.000	05:34
Air Blank		0.000	05:36
Subject Sample	# 2	0.272	05:38
Breath Volume		1.816	05:38
Air Blank		0.000	05:39
Control Test		0.079	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ALVARADO-HABER,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ARCE

DATE/TIME OF TEST: 07/04/2016 00:50:00

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.208	00:45
Breath Volume		1.160	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:47
Subject Sample	# 2	0.208	00:48
Breath Volume		1.316	00:48
Air Blank		0.000	00:49
Control Test		0.079	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:50

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MENA,ALEXANDER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 07/04/2016 01:53:17

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.080	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.179	01:48
Breath Volume		1.750	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.188	01:51
Breath Volume		1.121	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ALVARADO-HABER,ROYLAND
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: TORUNO

DATE/TIME OF TEST: 07/09/2016 02:03:18

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.080	02:00	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:02	
Control Test		0.080	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 07/28/2016 19:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,LEONARDO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,LEONARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 11:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/10/2016 12:17:59

Diagnostics Check		OK	12:08
Air Blank		0.000	12:08
Control Test		0.080	12:09
Air Blank		0.000	12:09
Subject Sample	# 1	0.243	12:10
Breath Volume		3.761	12:10
Air Blank		0.000	12:11
Air Blank		0.000	12:13
Subject Sample	# 2	0.216	12:13
Breath Volume		2.078	12:13
Air Blank		0.000	12:14
Air Blank		0.000	12:15
Subject Sample	# 3	0.226	12:16
Breath Volume		3.722	12:16
Air Blank		0.000	12:17
Control Test		0.078	12:17
Air Blank		0.000	12:17
Diagnostics Check		OK	12:17

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTANA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: QUINTANA,MICHAEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: ASTORGA

DATE/TIME OF TEST: 07/14/2016 06:47:26

Diagnostics Check		OK	06:43	
Air Blank		0.000	06:43	
Control Test		0.080	06:44	
Air Blank		0.000	06:44	
Subject Sample	# 1		06:46	Subject Test Refused
Breath Volume		0.000	06:46	
Air Blank		0.000	06:46	
Control Test		0.079	06:46	
Air Blank		0.000	06:47	
Diagnostics Check		OK	06:47	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,LUIS,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,GUILLERMO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: LORIGA

DATE/TIME OF TEST: 07/14/2016 21:42:24

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:39	
Control Test		0.081	21:40	
Air Blank		0.000	21:40	
Subject Sample	# 1		21:41	Subject Test Refused
Breath Volume		0.000	21:41	
Air Blank		0.000	21:41	
Control Test		0.080	21:41	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SAMS,MARKELL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: WYMAN

DATE/TIME OF TEST: 07/15/2016 05:11:53

Diagnostics Check		OK	05:04
Air Blank		0.000	05:04
Control Test		0.080	05:05
Air Blank		0.000	05:05
Subject Sample	# 1	0.140	05:06
Breath Volume		2.496	05:06
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.137	05:10
Breath Volume		2.054	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIL,JORGE,Y
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GIL,JORGE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 07/16/2016 02:31:37

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Control Test		0.079	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FONSECA,JOHNNY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 11:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:25

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/17/2016 11:59:26

Diagnostics Check		OK	11:51
Air Blank		0.000	11:52
Control Test		0.080	11:52
Air Blank		0.000	11:53
Subject Sample	# 1	0.189	11:54
Breath Volume		2.917	11:54
Air Blank		0.000	11:55
Air Blank		0.000	11:57
Subject Sample	# 2	0.192	11:57
Breath Volume		2.785	11:57
Air Blank		0.000	11:58
Control Test		0.077	11:58
Air Blank		0.000	11:59
Diagnostics Check		OK	11:59

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUKES,ANTONIO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VELLADARES,JAREL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: JENKINS

DATE/TIME OF TEST: 07/19/2016 07:18:42

Diagnostics Check		OK	07:15	
Air Blank		0.000	07:16	
Control Test		0.080	07:16	
Air Blank		0.000	07:17	
Subject Sample	# 1		07:17	Subject Test Refused
Breath Volume		0.000	07:17	
Air Blank		0.000	07:17	
Control Test		0.080	07:18	
Air Blank		0.000	07:18	
Diagnostics Check		OK	07:18	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSMAN,JAY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GROSSMAN,JAY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: RODRIGUEZ-HERNANDEZ

DATE/TIME OF TEST: 07/19/2016 19:11:56

Diagnostics Check		OK	19:05
Air Blank		0.000	19:06
Control Test		0.080	19:06
Air Blank		0.000	19:07
Subject Sample	# 1	0.158	19:07
Breath Volume		2.402	19:07
Air Blank		0.000	19:08
Air Blank		0.000	19:09
Subject Sample	# 2	0.154	19:10
Breath Volume		2.777	19:10
Air Blank		0.000	19:11
Control Test		0.078	19:11
Air Blank		0.000	19:11
Diagnostics Check		OK	19:11

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BOHNENBLUST,LESTER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: TRIANA

DATE/TIME OF TEST: 07/23/2016 08:20:36

Diagnostics Check		OK	08:17	
Air Blank		0.000	08:18	
Control Test		0.079	08:18	
Air Blank		0.000	08:19	
Subject Sample	# 1		08:19	Subject Test Refused
Breath Volume		0.000	08:19	
Air Blank		0.000	08:19	
Control Test		0.079	08:19	
Air Blank		0.000	08:20	
Diagnostics Check		OK	08:20	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DOVALE

DATE/TIME OF TEST: 07/25/2016 04:29:50

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.079	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:22	No Sample Provided
Breath Volume		0.000	04:22	
Air Blank		0.000	04:23	
Air Blank		0.000	04:25	
Subject Sample	# 2		04:28	No Sample Provided
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.081	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DOVALE

DATE/TIME OF TEST: 07/25/2016 04:42:34

Diagnostics Check		OK	04:32	
Air Blank		0.000	04:32	
Control Test		0.079	04:32	
Air Blank		0.000	04:33	
Subject Sample	# 1	0.150	04:36	Volume Not Met
Breath Volume		0.031	04:36	
Air Blank		0.000	04:37	
Air Blank		0.000	04:39	
Subject Sample	# 2		04:41	Subject Test Refused
Breath Volume		0.050	04:41	
Air Blank		0.000	04:41	
Control Test		0.079	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DOVALE

DATE/TIME OF TEST: 07/25/2016 04:59:08

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:47	
Control Test		0.079	04:48	
Air Blank		0.000	04:48	
Subject Sample	# 1	0.141	04:51	Volume Not Met
Breath Volume		0.140	04:51	
Air Blank		0.000	04:52	
Air Blank		0.000	04:54	
Subject Sample	# 2	0.183	04:57	Volume Not Met
Breath Volume		0.085	04:57	
Air Blank		0.000	04:58	
Control Test		0.078	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DOVALE

DATE/TIME OF TEST: 07/25/2016 05:13:44

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:02	
Control Test		0.078	05:03	
Air Blank		0.000	05:03	
Subject Sample	# 1	0.162	05:06	Volume Not Met
Breath Volume		0.050	05:06	
Air Blank		0.000	05:07	
Air Blank		0.000	05:09	
Subject Sample	# 2		05:12	Subject Test Refused
Breath Volume		0.359	05:12	
Air Blank		0.000	05:12	
Control Test		0.079	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,RUDDENS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: DOVALE

DATE/TIME OF TEST: 07/25/2016 05:51:03

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:39	
Control Test		0.078	05:39	
Air Blank		0.000	05:40	
Subject Sample	# 1	0.148	05:43	Volume Not Met
Breath Volume		1.117	05:43	
Air Blank		0.000	05:44	
Air Blank		0.000	05:46	
Subject Sample	# 2	0.146	05:49	Volume Not Met
Breath Volume		0.750	05:49	
Air Blank		0.000	05:50	
Control Test		0.078	05:50	
Air Blank		0.000	05:50	
Diagnostics Check		OK	05:51	

Data Download Date/Time: 07/28/2016 19:46

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,A,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ROMERO,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 07/29/2016 00:36:02

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:34	Subject Test Refused
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Control Test		0.080	00:35	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 08/11/2016 12:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HOLLANDER

DATE/TIME OF TEST: 07/30/2016 02:45:10

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:42	
Control Test		0.080	02:43	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank		0.000	02:44	
Control Test		0.080	02:44	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 08/11/2016 12:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HOLLANDER

DATE/TIME OF TEST: 07/30/2016 03:17:03

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:12	
Control Test		0.080	03:12	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.160	03:15	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 08/11/2016 12:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,A,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HOLLANDER

DATE/TIME OF TEST: 07/30/2016 04:07:26

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.077	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.133	03:57	
Breath Volume		1.617	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2	0.120	04:03	Volume Not Met
Breath Volume		0.542	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 3		04:06	Subject Test Refused
Breath Volume		0.000	04:06	
Air Blank		0.000	04:06	
Control Test		0.078	04:06	
Air Blank		0.000	04:07	
Diagnostics Check		OK	04:07	

Data Download Date/Time: 08/11/2016 12:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUKSA,ANDREW
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: PORRAS PAIZ

DATE/TIME OF TEST: 07/31/2016 06:22:18

Diagnostics Check		OK	06:14
Air Blank		0.000	06:14
Control Test		0.080	06:14
Air Blank		0.000	06:15
Subject Sample	# 1	0.015	06:17
Breath Volume		1.445	06:17
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.011	06:20
Breath Volume		1.214	06:20
Air Blank		0.000	06:21
Control Test		0.078	06:21
Air Blank		0.000	06:22
Diagnostics Check		OK	06:22

Data Download Date/Time: 08/11/2016 12:52

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LABRADOR,GABRIEL,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LABRADOR,GABRIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/03/2016 05:01:31

Diagnostics Check		OK	04:54
Air Blank		0.000	04:54
Control Test		0.077	04:55
Air Blank		0.000	04:55
Subject Sample	# 1	0.213	04:56
Breath Volume		2.941	04:56
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.216	04:59
Breath Volume		2.492	04:59
Air Blank		0.000	05:00
Control Test		0.076	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 07/28/2016 19:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GROSSMAN,JAY,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GROSSMAN,JAY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 07/13/2016 22:29:23

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:27	
Control Test		0.080	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:28	Subject Test Refused
Breath Volume		0.000	22:28	
Air Blank		0.000	22:28	
Control Test		0.080	22:28	
Air Blank		0.000	22:29	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 07/28/2016 19:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCE,DOMINUSE-CAIN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTRO,JUAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: AARON IV

DATE/TIME OF TEST: 07/16/2016 08:26:46

Diagnostics Check		OK	08:23	
Air Blank		0.000	08:24	
Control Test		0.080	08:24	
Air Blank		0.000	08:25	
Subject Sample	# 1		08:25	Subject Test Refused
Breath Volume		0.000	08:25	
Air Blank		0.000	08:25	
Control Test		0.080	08:26	
Air Blank		0.000	08:26	
Diagnostics Check		OK	08:26	

Data Download Date/Time: 07/28/2016 19:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOLCE,DOMINUSE-CAIN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASTRO,JUAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: AARON IV

DATE/TIME OF TEST: 07/16/2016 09:00:12

Diagnostics Check		OK	08:51	
Air Blank		0.000	08:52	
Control Test		0.080	08:52	
Air Blank		0.000	08:53	
Subject Sample	# 1		08:56	No Sample Provided
Breath Volume		0.000	08:56	
Air Blank		0.000	08:56	
Air Blank		0.000	08:58	
Subject Sample	# 2		08:58	Subject Test Refused
Breath Volume		0.000	08:58	
Air Blank		0.000	08:59	
Control Test		0.081	08:59	
Air Blank		0.000	09:00	
Diagnostics Check		OK	09:00	

Data Download Date/Time: 07/28/2016 19:59

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JASIEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERBELLO,JAVIER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/30/2016 04:08:44

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:55	
Control Test		0.080	03:55	
Air Blank		0.000	03:56	
Subject Sample	# 1	0.229	03:59	Volume Not Met
Breath Volume		0.679	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 2	0.278	04:02	
Breath Volume		1.917	04:02	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 3	0.284	04:07	
Breath Volume		2.183	04:07	
Air Blank		0.000	04:07	
Control Test		0.079	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 08/11/2016 13:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PROANO

DATE/TIME OF TEST: 07/03/2016 19:27:44

Diagnostics Check		OK	19:20
Air Blank		0.000	19:21
Control Test		0.076	19:21
Air Blank		0.000	19:21
Subject Sample	# 1	0.080	19:23
Breath Volume		2.015	19:23
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.086	19:26
Breath Volume		3.085	19:26
Air Blank		0.000	19:26
Control Test		0.075	19:27
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 07/28/2016 09:22

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 07/20/2016 01:12:10

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.077	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.209	01:07
Breath Volume		1.632	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.199	01:10
Breath Volume		1.660	01:10
Air Blank		0.000	01:11
Control Test		0.077	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 07/28/2016 09:22

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RIVERA,RAUL
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:37

SUBJECT'S LAST NAME: OSOKINA

DATE/TIME OF TEST: 07/26/2016 16:13:44

Diagnostics Check		OK	16:06
Air Blank		0.000	16:07
Control Test		0.080	16:07
Air Blank		0.000	16:07
Subject Sample	# 1	0.226	16:09
Breath Volume		2.261	16:09
Air Blank		0.000	16:09
Air Blank		0.000	16:11
Subject Sample	# 2	0.213	16:12
Breath Volume		1.300	16:12
Air Blank		0.000	16:12
Control Test		0.078	16:13
Air Blank		0.000	16:13
Diagnostics Check		OK	16:13

Data Download Date/Time: 08/25/2016 14:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GALULLO

DATE/TIME OF TEST: 07/28/2016 01:28:37

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.222	01:21
Breath Volume		1.550	01:21
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.231	01:26
Breath Volume		1.910	01:26
Air Blank		0.000	01:27
Control Test		0.080	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 08/25/2016 14:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 07/28/2016 21:49:37

Diagnostics Check		OK	21:43
Air Blank		0.000	21:43
Control Test		0.080	21:43
Air Blank		0.000	21:44
Subject Sample	# 1	0.223	21:45
Breath Volume		2.250	21:45
Air Blank		0.000	21:45
Air Blank		0.000	21:47
Subject Sample	# 2	0.232	21:47
Breath Volume		1.945	21:47
Air Blank		0.000	21:48
Control Test		0.079	21:49
Air Blank		0.000	21:49
Diagnostics Check		OK	21:49

Data Download Date/Time: 08/25/2016 14:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: DE LA FUENTE

DATE/TIME OF TEST: 07/30/2016 04:07:15

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.015	04:02
Breath Volume		1.812	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.013	04:05
Breath Volume		1.730	04:05
Air Blank		0.000	04:06
Control Test		0.080	04:06
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 08/25/2016 14:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAMZE,NEZAR
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 10:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 07/30/2016 11:54:03

Diagnostics Check		OK	11:47
Air Blank		0.000	11:48
Control Test		0.080	11:48
Air Blank		0.000	11:48
Subject Sample	# 1	0.121	11:49
Breath Volume		2.425	11:49
Air Blank		0.000	11:49
Air Blank		0.000	11:51
Subject Sample	# 2	0.117	11:52
Breath Volume		1.902	11:52
Air Blank		0.000	11:53
Control Test		0.080	11:53
Air Blank		0.000	11:53
Diagnostics Check		OK	11:54

Data Download Date/Time: 08/25/2016 14:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MAGUIRE,MATTHEW
ARRESTING OFFICER AGENCY: Lauderhill PD
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: MCKIE

DATE/TIME OF TEST: 07/31/2016 20:33:36

Diagnostics Check		OK	20:24	
Air Blank		0.000	20:24	
Control Test		0.079	20:24	
Air Blank		0.000	20:25	
Subject Sample	# 1	0.156	20:28	Volume Not Met
Breath Volume		0.148	20:28	
Air Blank		0.000	20:29	
Air Blank		0.000	20:30	
Subject Sample	# 2		20:32	Subject Test Refused
Breath Volume		0.000	20:32	
Air Blank		0.000	20:32	
Control Test		0.078	20:33	
Air Blank		0.000	20:33	
Diagnostics Check		OK	20:33	

Data Download Date/Time: 08/25/2016 14:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: SAINT LOUIS

DATE/TIME OF TEST: 07/01/2016 01:22:14

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.079	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:21	
Control Test		0.078	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 07/18/2016 10:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 07/01/2016 03:16:39

Diagnostics Check		OK	03:13	
Air Blank		0.000	03:14	
Control Test		0.079	03:14	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:15	Subject Test Refused
Breath Volume		0.000	03:15	
Air Blank		0.000	03:15	
Control Test		0.079	03:16	
Air Blank		0.000	03:16	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 07/18/2016 10:16

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CHISHOLM II

DATE/TIME OF TEST: 07/01/2016 23:11:33

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.079	23:09	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:10	
Control Test		0.079	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: AMBRIATI,NICHOLAS
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SUAREZ-ALEJANDRO

DATE/TIME OF TEST: 07/03/2016 03:31:23

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.245	03:26
Breath Volume		1.648	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.239	03:29
Breath Volume		1.886	03:29
Air Blank		0.000	03:30
Control Test		0.076	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 07/03/2016 22:19:55

Diagnostics Check		OK	22:13
Air Blank		0.000	22:14
Control Test		0.079	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.000	22:15
Breath Volume		3.714	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.000	22:18
Breath Volume		1.859	22:18
Air Blank		0.000	22:19
Control Test		0.079	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DONALD,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BECKLEY,JEFF
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 07/04/2016 20:40:28

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:38	
Control Test		0.079	20:38	
Air Blank		0.000	20:38	
Subject Sample	# 1		20:39	Subject Test Refused
Breath Volume		0.000	20:39	
Air Blank		0.000	20:39	
Control Test		0.078	20:39	
Air Blank		0.000	20:40	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 07/08/2016 23:16:55

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.079	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.316	23:12
Breath Volume		2.332	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.322	23:15
Breath Volume		2.980	23:15
Air Blank		0.000	23:15
Control Test		0.077	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 10:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: ELLANA

DATE/TIME OF TEST: 07/09/2016 23:26:42

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.079	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.131	23:19
Breath Volume		3.289	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.110	23:22
Breath Volume		1.773	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:24
Subject Sample	# 3	0.111	23:24
Breath Volume		1.410	23:24
Air Blank		0.000	23:25
Control Test		0.077	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: DECOLA

DATE/TIME OF TEST: 07/10/2016 03:16:00

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.236	03:07	Volume Not Met
Breath Volume		0.546	03:07	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 2	0.256	03:11	
Breath Volume		2.292	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:13	
Subject Sample	# 3	0.249	03:14	
Breath Volume		1.484	03:14	
Air Blank		0.000	03:15	
Control Test		0.076	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:16	

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,AARON,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BOWEN,AARON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 07/12/2016 20:48:37

Diagnostics Check		OK	20:40
Air Blank		0.000	20:40
Control Test		0.078	20:40
Air Blank		0.000	20:41
Subject Sample	# 1	0.216	20:43
Breath Volume		1.925	20:43
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.217	20:46
Breath Volume		1.542	20:46
Air Blank		0.000	20:47
Control Test		0.076	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO-LOPEZ,HELBER,F
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,BARBARA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: CARRILLO-LOPEZ

DATE/TIME OF TEST: 07/15/2016 05:52:23

Diagnostics Check		OK	05:46	
Air Blank		0.000	05:47	
Control Test		0.079	05:47	
Air Blank		0.000	05:48	
Subject Sample	# 1		05:51	Subject Test Refused
Breath Volume		0.000	05:51	
Air Blank		0.000	05:51	
Control Test		0.078	05:51	
Air Blank		0.000	05:52	
Diagnostics Check		OK	05:52	

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,BARBARA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: CARRILLO-LOPEZ

DATE/TIME OF TEST: 07/15/2016 05:59:43

Diagnostics Check		OK	05:56	
Air Blank		0.000	05:57	
Control Test		0.078	05:57	
Air Blank		0.000	05:58	
Subject Sample	# 1		05:58	Subject Test Refused
Breath Volume		0.000	05:58	
Air Blank		0.000	05:58	
Control Test		0.079	05:59	
Air Blank		0.000	05:59	
Diagnostics Check		OK	05:59	

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: LEVERETTE

DATE/TIME OF TEST: 07/15/2016 23:32:22

Diagnostics Check		OK	23:27	
Air Blank		0.000	23:27	
Control Test		0.079	23:28	
Air Blank		0.000	23:28	
Subject Sample	# 1	0.234	23:29	
Breath Volume		2.546	23:29	
Air Blank		0.000	23:30	
Air Blank			23:31	Ambient Fail
Air Blank		0.000	23:32	

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: LEVERETTE

DATE/TIME OF TEST: 07/15/2016 23:41:17

Diagnostics Check		OK	23:34
Air Blank		0.000	23:35
Control Test		0.076	23:35
Air Blank		0.000	23:36
Subject Sample	# 1	0.233	23:36
Breath Volume		1.382	23:36
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.217	23:39
Breath Volume		1.554	23:39
Air Blank		0.000	23:40
Control Test		0.076	23:40
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 07/16/2016 02:37:18

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.078	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.217	02:32
Breath Volume		3.117	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.218	02:35
Breath Volume		4.097	02:35
Air Blank		0.000	02:36
Control Test		0.076	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: THIERINGER

DATE/TIME OF TEST: 07/16/2016 04:09:42

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.139	04:05
Breath Volume		3.175	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.140	04:08
Breath Volume		3.117	04:08
Air Blank		0.000	04:08
Control Test		0.077	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 07/18/2016 10:17

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: GOMEZ-GOMEZ

DATE/TIME OF TEST: 07/18/2016 22:52:18

Diagnostics Check		OK	22:49	
Air Blank		0.000	22:49	
Control Test		0.080	22:50	
Air Blank		0.000	22:50	
Subject Sample	# 1		22:50	Subject Test Refused
Breath Volume		0.000	22:50	
Air Blank		0.000	22:51	
Control Test		0.079	22:51	
Air Blank		0.000	22:52	
Diagnostics Check		OK	22:52	

Data Download Date/Time: 08/29/2016 11:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MORTELLO II

DATE/TIME OF TEST: 07/19/2016 03:06:31

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:05	
Control Test		0.078	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 08/29/2016 11:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALLACE,JENNIFER,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WALLACE,JENNIFER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 12:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: VILLANUEVA

DATE/TIME OF TEST: 07/19/2016 13:56:57

Diagnostics Check		OK	13:50
Air Blank		0.000	13:51
Control Test		0.079	13:51
Air Blank		0.000	13:52
Subject Sample	# 1	0.000	13:52
Breath Volume		3.468	13:52
Air Blank		0.000	13:53
Air Blank		0.000	13:55
Subject Sample	# 2	0.000	13:55
Breath Volume		3.582	13:55
Air Blank		0.000	13:56
Control Test		0.079	13:56
Air Blank		0.000	13:56
Diagnostics Check		OK	13:56

Data Download Date/Time: 08/29/2016 11:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CONNELLY

DATE/TIME OF TEST: 07/21/2016 03:06:09

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.129	03:00
Breath Volume		1.796	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.134	03:04
Breath Volume		3.718	03:04
Air Blank		0.000	03:05
Control Test		0.076	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ANDERS

DATE/TIME OF TEST: 07/22/2016 01:21:17

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.079	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:19	Subject Test Refused
Breath Volume		0.000	01:19	
Air Blank		0.000	01:20	
Control Test		0.078	01:20	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/23/2016 02:17:47

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.078	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.078	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MCELFRESH

DATE/TIME OF TEST: 07/24/2016 01:10:42

Diagnostics Check		OK	01:07	
Air Blank		0.000	01:08	
Control Test		0.078	01:08	
Air Blank		0.000	01:08	
Subject Sample	# 1		01:09	Subject Test Refused
Breath Volume		0.000	01:09	
Air Blank		0.000	01:09	
Control Test		0.078	01:10	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BOK,RANDY
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MALDONADO-GOMEZ

DATE/TIME OF TEST: 07/24/2016 01:57:15

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:53	
Control Test		0.079	01:53	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HERNANDEZ-CRUZ

DATE/TIME OF TEST: 07/25/2016 01:31:37

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.181	01:27
Breath Volume		1.109	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.187	01:30
Breath Volume		2.160	01:30
Air Blank		0.000	01:30
Control Test		0.078	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: GAMSKY

DATE/TIME OF TEST: 07/25/2016 19:12:58

Diagnostics Check		OK	19:06
Air Blank		0.000	19:07
Control Test		0.079	19:07
Air Blank		0.000	19:08
Subject Sample	# 1	0.000	19:08
Breath Volume		2.609	19:08
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.000	19:11
Breath Volume		3.148	19:11
Air Blank		0.000	19:12
Control Test		0.076	19:12
Air Blank		0.000	19:12
Diagnostics Check		OK	19:12

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: STETHEN

DATE/TIME OF TEST: 07/25/2016 22:41:37

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.079	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.033	22:37
Breath Volume		2.449	22:37
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.030	22:40
Breath Volume		2.296	22:40
Air Blank		0.000	22:40
Control Test		0.076	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: SMITHBURGER

DATE/TIME OF TEST: 07/27/2016 01:10:18

Diagnostics Check		OK	01:03
Air Blank		0.000	01:04
Control Test		0.079	01:04
Air Blank		0.000	01:05
Subject Sample	# 1	0.148	01:05
Breath Volume		1.203	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:08
Subject Sample	# 2	0.161	01:08
Breath Volume		2.265	01:08
Air Blank		0.000	01:09
Control Test		0.076	01:09
Air Blank		0.000	01:10
Diagnostics Check		OK	01:10

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 07/27/2016 23:06:52

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.071	23:06	Control Outside Tolerance
Air Blank	0.000	23:06	

Data Download Date/Time: 08/29/2016 11:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/30/2016 01:57:29

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.079	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:56	Subject Test Refused
Breath Volume		0.000	01:56	
Air Blank		0.000	01:56	
Control Test		0.078	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: ZEHR

DATE/TIME OF TEST: 07/30/2016 22:03:15

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.079	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1	0.000	21:53	Volume Not Met
Breath Volume		0.093	21:53	
Air Blank		0.000	21:53	
Air Blank		0.000	21:55	
Subject Sample	# 2	0.000	21:57	
Breath Volume		1.554	21:57	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 3		22:01	Subject Test Refused
Breath Volume		0.066	22:01	
Air Blank		0.000	22:02	
Control Test		0.077	22:02	
Air Blank		0.000	22:03	
Diagnostics Check		OK	22:03	

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/31/2016 01:05:38

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.078	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.225	00:57
Breath Volume		3.699	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.203	01:00
Breath Volume		1.968	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 3	0.213	01:03
Breath Volume		1.480	01:03
Air Blank		0.000	01:04
Control Test		0.076	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:05

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MENDEZ ESCOBAR

DATE/TIME OF TEST: 07/31/2016 03:31:42

Diagnostics Check		OK	03:25
Air Blank		0.000	03:26
Control Test		0.079	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.157	03:27
Breath Volume		2.949	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.154	03:30
Breath Volume		2.675	03:30
Air Blank		0.000	03:30
Control Test		0.077	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 08/29/2016 11:06

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 12:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MCLELAND

DATE/TIME OF TEST: 07/27/2016 01:50:22

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.158	01:45
Breath Volume		1.382	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.162	01:48
Breath Volume		1.812	01:48
Air Blank		0.000	01:49
Control Test		0.077	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 08/29/2016 11:01

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: HOOVER

DATE/TIME OF TEST: 07/27/2016 23:16:01

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:13	
Control Test		0.077	23:13	
Air Blank		0.000	23:14	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Control Test		0.077	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 08/29/2016 11:01

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: JEWETT,KENNETH
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LOPEZ-JIMINEZ

DATE/TIME OF TEST: 07/28/2016 01:14:54

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.191	01:10
Breath Volume		2.582	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.195	01:13
Breath Volume		2.699	01:13
Air Blank		0.000	01:13
Control Test		0.078	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 08/29/2016 11:01

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GENGLER CAPPARELLI

DATE/TIME OF TEST: 07/28/2016 23:46:34

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:39	No Sample Provided
Breath Volume		0.000	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:42	
Subject Sample	# 2		23:45	No Sample Provided
Breath Volume		0.000	23:45	
Air Blank		0.000	23:45	
Control Test		0.078	23:45	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 08/29/2016 11:01

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: ST LOUIS

DATE/TIME OF TEST: 07/01/2016 03:38:17

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.079	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.189	03:33
Breath Volume		2.308	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.181	03:36
Breath Volume		2.265	03:36
Air Blank		0.000	03:37
Control Test		0.077	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: HINDS

DATE/TIME OF TEST: 07/01/2016 23:09:37

Diagnostics Check		OK	23:03
Air Blank		0.000	23:03
Control Test		0.079	23:04
Air Blank		0.000	23:04
Subject Sample	# 1	0.039	23:05
Breath Volume		1.902	23:05
Air Blank		0.000	23:05
Air Blank		0.000	23:07
Subject Sample	# 2	0.039	23:08
Breath Volume		2.273	23:08
Air Blank		0.000	23:08
Control Test		0.078	23:09
Air Blank		0.000	23:09
Diagnostics Check		OK	23:09

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 07/02/2016 02:57:50

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.180	02:53
Breath Volume		3.414	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.172	02:56
Breath Volume		3.464	02:56
Air Blank		0.000	02:56
Control Test		0.077	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 07/02/2016 04:59:58

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:51	
Control Test		0.078	04:52	
Air Blank		0.000	04:52	
Subject Sample	# 1		04:55	No Sample Provided
Breath Volume		0.000	04:55	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 2		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:59	
Control Test		0.079	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: MELENDEZ

DATE/TIME OF TEST: 07/03/2016 01:53:13

Diagnostics Check		OK	01:41	
Air Blank		0.000	01:42	
Control Test		0.078	01:42	
Air Blank		0.000	01:43	
Subject Sample	# 1		01:46	No Sample Provided
Breath Volume		0.000	01:46	
Air Blank		0.000	01:46	
Air Blank		0.000	01:48	
Subject Sample	# 2		01:51	No Sample Provided
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Control Test		0.079	01:52	
Air Blank		0.000	01:53	
Diagnostics Check		OK	01:53	

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VAZQUEZ MIRANDA

DATE/TIME OF TEST: 07/03/2016 04:02:01

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.077	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.072	03:57
Breath Volume		2.457	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.075	04:00
Breath Volume		3.718	04:00
Air Blank		0.000	04:01
Control Test		0.077	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:02

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: GRAYER

DATE/TIME OF TEST: 07/03/2016 23:56:04

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.078	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.266	23:51
Breath Volume		1.347	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:53
Subject Sample	# 2	0.270	23:54
Breath Volume		1.199	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:56

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHLEEN,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: FLETCHER III

DATE/TIME OF TEST: 07/04/2016 00:08:30

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.077	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.149	00:03
Breath Volume		3.019	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.147	00:06
Breath Volume		2.609	00:06
Air Blank		0.000	00:07
Control Test		0.078	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MARSHALL JR

DATE/TIME OF TEST: 07/04/2016 01:06:20

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.078	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.000	01:02
Breath Volume		2.468	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.000	01:04
Breath Volume		2.281	01:04
Air Blank		0.000	01:05
Control Test		0.078	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 07/04/2016 03:54:03

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.078	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.161	03:49
Breath Volume		2.875	03:49
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.158	03:52
Breath Volume		3.417	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 07/05/2016 04:28:31

Diagnostics Check		OK	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.156	04:23
Breath Volume		1.839	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.153	04:26
Breath Volume		1.746	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 07/05/2016 23:55:01

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:43	
Control Test		0.079	23:43	
Air Blank		0.000	23:44	
Subject Sample	# 1	0.168	23:44	
Breath Volume		1.898	23:44	
Air Blank		0.000	23:45	
Air Blank		0.000	23:47	
Subject Sample	# 2	0.133	23:50	Volume Not Met
Breath Volume		0.628	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 3	0.152	23:53	
Breath Volume		1.546	23:53	
Air Blank		0.000	23:54	
Control Test		0.076	23:54	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/08/2016 02:07:57

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.078	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.119	02:02
Breath Volume		1.199	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.120	02:06
Breath Volume		1.433	02:06
Air Blank		0.000	02:07
Control Test		0.077	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 07/09/2016 04:36:38

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.078	04:30
Air Blank		0.000	04:30
Subject Sample	# 1	0.078	04:32
Breath Volume		2.269	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.078	04:35
Breath Volume		2.968	04:35
Air Blank		0.000	04:35
Control Test		0.077	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: TAHO

DATE/TIME OF TEST: 07/09/2016 22:12:16

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.078	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.162	22:07
Breath Volume		4.171	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.151	22:10
Breath Volume		3.359	22:10
Air Blank		0.000	22:11
Control Test		0.078	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: LADNER

DATE/TIME OF TEST: 07/10/2016 04:41:43

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.078	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.183	04:36
Breath Volume		4.042	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.173	04:40
Breath Volume		2.953	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARNHART,H
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: VEGA MORALES

DATE/TIME OF TEST: 07/12/2016 00:15:09

Diagnostics Check		OK	00:08
Air Blank		0.000	00:09
Control Test		0.079	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.175	00:10
Breath Volume		2.730	00:10
Air Blank		0.000	00:11
Air Blank		0.000	00:13
Subject Sample	# 2	0.188	00:13
Breath Volume		2.515	00:13
Air Blank		0.000	00:14
Control Test		0.078	00:14
Air Blank		0.000	00:15
Diagnostics Check		OK	00:15

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SAVELLI

DATE/TIME OF TEST: 07/13/2016 02:19:31

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.115	02:14
Breath Volume		3.675	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.115	02:17
Breath Volume		2.359	02:17
Air Blank		0.000	02:18
Control Test		0.077	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 07/13/2016 09:05

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: AFFLERBACH

DATE/TIME OF TEST: 07/14/2016 04:15:09

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.240	04:10
Breath Volume		2.042	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:12
Subject Sample	# 2	0.242	04:13
Breath Volume		1.972	04:13
Air Blank		0.000	04:14
Control Test		0.077	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 08/10/2016 20:07

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: BARBER

DATE/TIME OF TEST: 07/14/2016 21:20:53

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.078	21:14
Air Blank		0.000	21:15
Subject Sample	# 1	0.199	21:16
Breath Volume		2.062	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.199	21:19
Breath Volume		2.179	21:19
Air Blank		0.000	21:19
Control Test		0.077	21:20
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 08/10/2016 20:07

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SMETANA,K
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: KOONCE

DATE/TIME OF TEST: 07/15/2016 18:06:13

Diagnostics Check		OK	18:00
Air Blank		0.000	18:00
Control Test		0.079	18:00
Air Blank		0.000	18:01
Subject Sample	# 1	0.000	18:02
Breath Volume		1.617	18:02
Air Blank		0.000	18:02
Air Blank		0.000	18:04
Subject Sample	# 2	0.000	18:04
Breath Volume		1.531	18:04
Air Blank		0.000	18:05
Control Test		0.078	18:05
Air Blank		0.000	18:06
Diagnostics Check		OK	18:06

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: ANGON-MARTINEZ

DATE/TIME OF TEST: 07/16/2016 03:29:10

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.078	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.219	03:24
Breath Volume		2.433	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:26
Subject Sample	# 2	0.219	03:27
Breath Volume		1.589	03:27
Air Blank		0.000	03:28
Control Test		0.077	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: MACIA PEREZ

DATE/TIME OF TEST: 07/16/2016 05:33:11

Diagnostics Check		OK	05:24	
Air Blank		0.000	05:25	
Control Test		0.078	05:25	
Air Blank		0.000	05:26	
Subject Sample	# 1		05:29	No Sample Provided
Breath Volume		0.000	05:29	
Air Blank		0.000	05:29	
Air Blank		0.000	05:31	
Subject Sample	# 2		05:31	Subject Test Refused
Breath Volume		0.000	05:31	
Air Blank		0.000	05:32	
Control Test		0.078	05:32	
Air Blank		0.000	05:33	
Diagnostics Check		OK	05:33	

Data Download Date/Time: 08/10/2016 20:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: LOPEZ ORTEGA

DATE/TIME OF TEST: 07/16/2016 23:21:13

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.125	23:16
Breath Volume		1.941	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.126	23:19
Breath Volume		1.832	23:19
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KROSLAK

DATE/TIME OF TEST: 07/17/2016 00:53:13

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.078	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.233	00:48
Breath Volume		2.492	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.234	00:51
Breath Volume		3.695	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 07/17/2016 05:04:05

Diagnostics Check		OK	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.106	04:59
Breath Volume		2.667	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.104	05:02
Breath Volume		2.894	05:02
Air Blank		0.000	05:03
Control Test		0.078	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: CERNA

DATE/TIME OF TEST: 07/17/2016 05:51:48

Diagnostics Check		OK	05:44
Air Blank		0.000	05:45
Control Test		0.079	05:45
Air Blank		0.000	05:46
Subject Sample	# 1	0.060	05:47
Breath Volume		1.250	05:47
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.059	05:50
Breath Volume		1.562	05:50
Air Blank		0.000	05:50
Control Test		0.079	05:51
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 07/17/2016 22:12:46

Diagnostics Check		OK	22:03
Air Blank		0.000	22:04
Control Test		0.078	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.197	22:06
Breath Volume		2.074	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:08
Subject Sample	# 2	0.197	22:11
Breath Volume		1.484	22:11
Air Blank		0.000	22:11
Control Test		0.077	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: NANDU

DATE/TIME OF TEST: 07/19/2016 04:22:01

Diagnostics Check		OK	04:10	
Air Blank		0.000	04:11	
Control Test		0.078	04:11	
Air Blank		0.000	04:11	
Subject Sample	# 1		04:15	No Sample Provided
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Air Blank		0.000	04:17	
Subject Sample	# 2		04:20	No Sample Provided
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.078	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 08/10/2016 20:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS-COLON,S
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/21/2016 00:54:04

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.078	00:48
Air Blank		0.000	00:48
Subject Sample	# 1	0.164	00:49
Breath Volume		2.453	00:49
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.158	00:52
Breath Volume		2.609	00:52
Air Blank		0.000	00:53
Control Test		0.078	00:53
Air Blank		0.000	00:53
Diagnostics Check		OK	00:54

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: DAVID

DATE/TIME OF TEST: 07/21/2016 02:17:55

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.078	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.078	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SHEA

DATE/TIME OF TEST: 07/22/2016 02:55:16

Diagnostics Check		OK	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:46
Subject Sample	# 1	0.241	02:47
Breath Volume		2.410	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.263	02:50
Breath Volume		2.199	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 3	0.267	02:53
Breath Volume		2.691	02:53
Air Blank		0.000	02:54
Control Test		0.077	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 07/22/2016 04:42:55

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.078	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.207	04:37
Breath Volume		1.500	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.209	04:41
Breath Volume		1.722	04:41
Air Blank		0.000	04:41
Control Test		0.078	04:42
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/23/2016 01:54:37

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.124	01:49
Breath Volume		1.132	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.134	01:53
Breath Volume		1.781	01:53
Air Blank		0.000	01:53
Control Test		0.078	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: GRAFF

DATE/TIME OF TEST: 07/23/2016 02:44:47

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.211	02:40
Breath Volume		1.480	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.227	02:43
Breath Volume		2.210	02:43
Air Blank		0.000	02:43
Control Test		0.078	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GEORGE,M
ARRESTING OFFICER AGENCY: U of So FL Campus PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SANTANA JR

DATE/TIME OF TEST: 07/23/2016 03:26:50

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.078	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.146	03:21
Breath Volume		2.105	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.146	03:25
Breath Volume		2.410	03:25
Air Blank		0.000	03:25
Control Test		0.077	03:26
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: BALBIN

DATE/TIME OF TEST: 07/23/2016 05:03:36

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.077	04:57
Air Blank		0.000	04:58
Subject Sample	# 1	0.170	04:58
Breath Volume		1.429	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.163	05:02
Breath Volume		2.242	05:02
Air Blank		0.000	05:02
Control Test		0.077	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: MOLINA LOPEZ

DATE/TIME OF TEST: 07/23/2016 23:24:34

Diagnostics Check		OK	23:17
Air Blank		0.000	23:18
Control Test		0.078	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.180	23:19
Breath Volume		1.734	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.171	23:22
Breath Volume		1.628	23:22
Air Blank		0.000	23:23
Control Test		0.077	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 08/10/2016 20:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/24/2016 02:14:56

Diagnostics Check		OK	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:08
Subject Sample	# 1	0.268	02:10
Breath Volume		1.535	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.266	02:13
Breath Volume		1.441	02:13
Air Blank		0.000	02:14
Control Test		0.076	02:14
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: BAUCHAM

DATE/TIME OF TEST: 07/24/2016 06:10:20

Diagnostics Check		OK	06:02
Air Blank		0.000	06:03
Control Test		0.078	06:03
Air Blank		0.000	06:04
Subject Sample	# 1	0.235	06:05
Breath Volume		1.343	06:05
Air Blank		0.000	06:05
Air Blank		0.000	06:07
Subject Sample	# 2	0.235	06:08
Breath Volume		1.843	06:08
Air Blank		0.000	06:09
Control Test		0.078	06:09
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HALE,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: MARTINEZ-ARRIBA

DATE/TIME OF TEST: 07/24/2016 07:55:53

Diagnostics Check		OK	07:49
Air Blank		0.000	07:49
Control Test		0.078	07:49
Air Blank		0.000	07:50
Subject Sample	# 1	0.201	07:51
Breath Volume		2.277	07:51
Air Blank		0.000	07:52
Air Blank		0.000	07:53
Subject Sample	# 2	0.191	07:54
Breath Volume		1.710	07:54
Air Blank		0.000	07:54
Control Test		0.077	07:55
Air Blank		0.000	07:55
Diagnostics Check		OK	07:55

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EHRHART,M
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: LANEY

DATE/TIME OF TEST: 07/25/2016 00:58:37

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.162	00:52
Breath Volume		1.648	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.163	00:57
Breath Volume		1.789	00:57
Air Blank		0.000	00:57
Control Test		0.077	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: QOSAJ

DATE/TIME OF TEST: 07/25/2016 02:45:03

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.078	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.078	02:39
Breath Volume		1.230	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.074	02:43
Breath Volume		1.476	02:43
Air Blank		0.000	02:44
Control Test		0.078	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:45

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FIPPS

DATE/TIME OF TEST: 07/25/2016 03:58:12

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.258	03:53
Breath Volume		3.066	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:55
Subject Sample	# 2	0.269	03:56
Breath Volume		2.679	03:56
Air Blank		0.000	03:57
Control Test		0.078	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: LLERA

DATE/TIME OF TEST: 07/27/2016 23:28:37

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:22
Air Blank		0.000	23:23
Subject Sample	# 1	0.198	23:24
Breath Volume		1.343	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.206	23:27
Breath Volume		1.839	23:27
Air Blank		0.000	23:27
Control Test		0.078	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 07/28/2016 06:22:47

Diagnostics Check		OK	06:14	
Air Blank		0.000	06:14	
Control Test		0.078	06:15	
Air Blank		0.000	06:15	
Subject Sample	# 1		06:18	No Sample Provided
Breath Volume		0.000	06:18	
Air Blank		0.000	06:19	
Air Blank		0.000	06:21	
Subject Sample	# 2		06:21	Subject Test Refused
Breath Volume		0.000	06:21	
Air Blank		0.000	06:21	
Control Test		0.078	06:22	
Air Blank		0.000	06:22	
Diagnostics Check		OK	06:22	

Data Download Date/Time: 08/10/2016 20:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 07/29/2016 01:39:34

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.118	01:34
Breath Volume		1.105	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.118	01:37
Breath Volume		2.398	01:37
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 07/30/2016 01:08:30

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.078	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.137	01:03
Breath Volume		2.089	01:03
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.134	01:06
Breath Volume		2.468	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWANDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 07/30/2016 04:22:26

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.078	04:16
Air Blank		0.000	04:16
Subject Sample	# 1	0.203	04:17
Breath Volume		1.640	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.206	04:20
Breath Volume		2.929	04:20
Air Blank		0.000	04:21
Control Test		0.077	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: TILLETT JR

DATE/TIME OF TEST: 07/30/2016 05:02:41

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.077	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.147	04:58
Breath Volume		2.546	04:58
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.153	05:01
Breath Volume		3.183	05:01
Air Blank		0.000	05:01
Control Test		0.077	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: WELDON

DATE/TIME OF TEST: 07/31/2016 01:39:47

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.078	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.134	01:35
Breath Volume		3.214	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.146	01:38
Breath Volume		4.105	01:38
Air Blank		0.000	01:38
Control Test		0.078	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HAIR

DATE/TIME OF TEST: 07/31/2016 02:47:25

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.078	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.042	02:42
Breath Volume		2.550	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.039	02:45
Breath Volume		1.957	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: RIVERA RIVERA

DATE/TIME OF TEST: 07/31/2016 03:57:02

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.078	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.272	03:52
Breath Volume		1.667	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.268	03:55
Breath Volume		3.128	03:55
Air Blank		0.000	03:56
Control Test		0.077	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:57

Data Download Date/Time: 08/10/2016 20:08

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SHUMAN

DATE/TIME OF TEST: 07/31/2016 04:45:57

Diagnostics Check		OK	04:39
Air Blank		0.000	04:39
Control Test		0.078	04:40
Air Blank		0.000	04:40
Subject Sample	# 1	0.213	04:41
Breath Volume		1.781	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:43
Subject Sample	# 2	0.203	04:44
Breath Volume		1.835	04:44
Air Blank		0.000	04:45
Control Test		0.077	04:45
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 08/10/2016 20:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PETKOVSEK,RICHARD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/05/2016 03:15:03

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.077	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.166	03:10
Breath Volume		2.164	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:12
Subject Sample	# 2	0.156	03:13
Breath Volume		1.656	03:13
Air Blank		0.000	03:14
Control Test		0.077	03:14
Air Blank		0.000	03:14
Diagnostics Check		OK	03:15

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEHAN,SHANE,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: DAWSON,M
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: CARBRELLO

DATE/TIME OF TEST: 07/05/2016 22:26:19

Diagnostics Check		OK	22:19
Air Blank		0.000	22:20
Control Test		0.078	22:20
Air Blank		0.000	22:21
Subject Sample	# 1	0.246	22:21
Breath Volume		1.207	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.249	22:24
Breath Volume		1.867	22:24
Air Blank		0.000	22:25
Control Test		0.077	22:25
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 07/28/2016 17:31

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TAYLOR,GRANT
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: COUBROUGH

DATE/TIME OF TEST: 07/09/2016 23:38:22

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.079	23:36	
Air Blank		0.000	23:36	
Subject Sample	# 1		23:36	Subject Test Refused
Breath Volume		0.000	23:36	
Air Blank		0.000	23:37	
Control Test		0.078	23:37	
Air Blank		0.000	23:38	
Diagnostics Check		OK	23:38	

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBAGALLO,FRANK,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TAYLOR,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 07/16/2016 23:46:34

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.078	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.280	23:42
Breath Volume		1.714	23:42
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.284	23:44
Breath Volume		1.167	23:44
Air Blank		0.000	23:45
Control Test		0.077	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: JORDAHL

DATE/TIME OF TEST: 07/19/2016 01:09:21

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.078	01:03
Air Blank		0.000	01:03
Subject Sample	# 1	0.227	01:04
Breath Volume		2.277	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.226	01:07
Breath Volume		1.167	01:07
Air Blank		0.000	01:08
Control Test		0.078	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GONZALEZ,DIEGO
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: FLEFIL

DATE/TIME OF TEST: 07/19/2016 21:52:51

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.078	21:46
Air Blank		0.000	21:47
Subject Sample	# 1	0.018	21:48
Breath Volume		2.445	21:48
Air Blank		0.000	21:48
Air Blank		0.000	21:50
Subject Sample	# 2	0.018	21:51
Breath Volume		2.441	21:51
Air Blank		0.000	21:51
Control Test		0.079	21:52
Air Blank		0.000	21:52
Diagnostics Check		OK	21:52

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHEWNING,A
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 07/20/2016 23:55:40

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:53	
Control Test		0.078	23:53	
Air Blank		0.000	23:53	
Subject Sample	# 1		23:54	Subject Test Refused
Breath Volume		0.000	23:54	
Air Blank		0.000	23:54	
Control Test		0.078	23:55	
Air Blank		0.000	23:55	
Diagnostics Check		OK	23:55	

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEAUSOLEIL,AARON,F
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: STOGDON,BRAD
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 07/22/2016 18:08:10

Diagnostics Check		OK	18:01
Air Blank		0.000	18:02
Control Test		0.078	18:02
Air Blank		0.000	18:03
Subject Sample	# 1	0.096	18:03
Breath Volume		2.351	18:03
Air Blank		0.000	18:04
Air Blank		0.000	18:06
Subject Sample	# 2	0.096	18:06
Breath Volume		2.582	18:06
Air Blank		0.000	18:07
Control Test		0.076	18:07
Air Blank		0.000	18:08
Diagnostics Check		OK	18:08

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARBAGALLO,FRANK,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WEST,JACOB
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: DENOTO

DATE/TIME OF TEST: 07/23/2016 01:43:14

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.078	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.151	01:38
Breath Volume		1.937	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.165	01:41
Breath Volume		2.578	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRAY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHEWNING,AUSTIN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/23/2016 04:58:15

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.077	04:51
Air Blank		0.000	04:51
Subject Sample	# 1	0.140	04:52
Breath Volume		1.503	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.143	04:56
Breath Volume		2.003	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: RAINEY,C
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 07/23/2016 12:55:04

Diagnostics Check		OK	12:48
Air Blank		0.000	12:49
Control Test		0.078	12:49
Air Blank		0.000	12:49
Subject Sample	# 1	0.000	12:50
Breath Volume		1.628	12:50
Air Blank		0.000	12:51
Air Blank		0.000	12:53
Subject Sample	# 2	0.000	12:53
Breath Volume		1.378	12:53
Air Blank		0.000	12:54
Control Test		0.079	12:54
Air Blank		0.000	12:54
Diagnostics Check		OK	12:55

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WEST,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: INFANTE MACHADO

DATE/TIME OF TEST: 07/24/2016 03:16:54

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.145	03:11
Breath Volume		2.304	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.136	03:15
Breath Volume		1.964	03:15
Air Blank		0.000	03:15
Control Test		0.077	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 07/28/2016 17:31

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: FIGUEROA PINTO

DATE/TIME OF TEST: 07/17/2016 07:51:56

Diagnostics Check		OK	07:45
Air Blank		0.000	07:45
Control Test		0.076	07:46
Air Blank		0.000	07:46
Subject Sample	# 1	0.065	07:47
Breath Volume		1.546	07:47
Air Blank		0.000	07:47
Air Blank		0.000	07:49
Subject Sample	# 2	0.065	07:50
Breath Volume		2.726	07:50
Air Blank		0.000	07:50
Control Test		0.078	07:51
Air Blank		0.000	07:51
Diagnostics Check		OK	07:51

Data Download Date/Time: 08/26/2016 11:23

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/27/2016 11:17:33

Diagnostics Check	OK	11:15	
Air Blank	0.000	11:16	
Control Test	0.076	11:16	
Air Blank		11:17	RFI Detect
Air Blank	0.000	11:17	

Data Download Date/Time: 08/26/2016 11:23

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 07/27/2016 11:23:53

Diagnostics Check		OK	11:21	
Air Blank		0.000	11:21	
Control Test		0.077	11:21	
Air Blank		0.000	11:22	
Subject Sample	# 1		11:22	Subject Test Refused
Breath Volume		0.000	11:22	
Air Blank		0.000	11:23	
Control Test		0.000	11:23	RFI Detect
Air Blank		0.000	11:23	

Data Download Date/Time: 08/26/2016 11:23

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006628

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAUGHN,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BORGES

DATE/TIME OF TEST: 07/31/2016 01:26:41

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.077	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.110	01:22
Breath Volume		1.933	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.109	01:25
Breath Volume		2.191	01:25
Air Blank		0.000	01:25
Control Test		0.078	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 08/31/2016 22:15

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: VINCUR

DATE/TIME OF TEST: 07/17/2016 03:36:29

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.078	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1		03:29	No Sample Provided
Breath Volume		0.000	03:29	
Air Blank		0.000	03:29	
Air Blank		0.000	03:31	
Subject Sample	# 2		03:35	No Sample Provided
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Control Test		0.077	03:35	
Air Blank		0.000	03:36	
Diagnostics Check		OK	03:36	

Data Download Date/Time: 08/31/2016 22:07

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: HERNANDEZ PEREZ

DATE/TIME OF TEST: 07/31/2016 01:16:19

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.077	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.181	01:11
Breath Volume		2.816	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.186	01:14
Breath Volume		2.832	01:14
Air Blank		0.000	01:15
Control Test		0.076	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 08/31/2016 22:07

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKNIS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: PICON

DATE/TIME OF TEST: 07/31/2016 05:18:24

Diagnostics Check		OK	05:11
Air Blank		0.000	05:12
Control Test		0.078	05:12
Air Blank		0.000	05:13
Subject Sample	# 1	0.201	05:13
Breath Volume		1.703	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:16
Subject Sample	# 2	0.189	05:16
Breath Volume		1.402	05:16
Air Blank		0.000	05:17
Control Test		0.076	05:17
Air Blank		0.000	05:18
Diagnostics Check		OK	05:18

Data Download Date/Time: 08/31/2016 22:07

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,RICARDO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: MENDOZA ROBLEDO

DATE/TIME OF TEST: 07/01/2016 19:55:50

Diagnostics Check	Fail	19:55	Voltage/Current Test
Air Blank	0.000	19:55	

Data Download Date/Time: 07/12/2016 11:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,RICARDO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: MENDOZA ROBLEDO

DATE/TIME OF TEST: 07/01/2016 20:05:15

Diagnostics Check		OK	19:58
Air Blank		0.000	19:59
Control Test		0.079	19:59
Air Blank		0.000	20:00
Subject Sample	# 1	0.181	20:00
Breath Volume		2.789	20:00
Air Blank		0.000	20:01
Air Blank		0.000	20:03
Subject Sample	# 2	0.179	20:03
Breath Volume		4.183	20:03
Air Blank		0.000	20:04
Control Test		0.078	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 07/12/2016 11:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ROSARIO,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/01/2016 23:27:08

Diagnostics Check		OK	23:18	
Air Blank		0.000	23:19	
Control Test		0.079	23:19	
Air Blank		0.000	23:19	
Subject Sample	# 1		23:23	No Sample Provided
Breath Volume		0.000	23:23	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 2		23:25	Subject Test Refused
Breath Volume		0.000	23:25	
Air Blank		0.000	23:26	
Control Test		0.079	23:26	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 07/12/2016 11:14

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 07/01/2016 01:45:40

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:44	
Control Test		0.078	01:45	
Air Blank		0.000	01:45	
Diagnostics Check		OK	01:45	

Data Download Date/Time: 07/18/2016 09:23

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 07/01/2016 22:59:34

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.078	22:55	
Air Blank		0.000	22:56	
Subject Sample	# 1		22:58	Subject Test Refused
Breath Volume		0.000	22:58	
Air Blank		0.000	22:58	
Control Test		0.079	22:58	
Air Blank		0.000	22:59	
Diagnostics Check		OK	22:59	

Data Download Date/Time: 07/18/2016 09:23

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 07/02/2016 02:06:35

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.079	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.164	02:01
Breath Volume		1.871	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.172	02:04
Breath Volume		3.144	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 07/18/2016 09:23

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLEMAN,WILLIAM,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: COLEMAN,WILLIAM
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: OBRIEN

DATE/TIME OF TEST: 07/02/2016 23:26:40

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.135	23:22
Breath Volume		2.574	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.131	23:25
Breath Volume		1.925	23:25
Air Blank		0.000	23:25
Control Test		0.078	23:26
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 07/18/2016 09:24

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SARCHI

DATE/TIME OF TEST: 07/31/2016 23:13:15

Diagnostics Check		OK	23:10	
Air Blank		0.000	23:10	
Control Test		0.080	23:11	
Air Blank		0.000	23:11	
Subject Sample	# 1		23:11	Subject Test Refused
Breath Volume		0.000	23:11	
Air Blank		0.000	23:12	
Control Test		0.078	23:12	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 08/02/2016 20:09

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: GREKO

DATE/TIME OF TEST: 07/16/2016 20:36:06

Diagnostics Check		OK	20:29
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.000	20:31
Breath Volume		1.566	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.000	20:34
Breath Volume		2.003	20:34
Air Blank		0.000	20:35
Control Test		0.079	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 07/18/2016 10:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: GAGE

DATE/TIME OF TEST: 07/24/2016 20:51:46

Diagnostics Check		OK	20:45
Air Blank		0.000	20:45
Control Test		0.080	20:46
Air Blank		0.000	20:46
Subject Sample	# 1	0.143	20:47
Breath Volume		1.656	20:47
Air Blank		0.000	20:47
Air Blank		0.000	20:49
Subject Sample	# 2	0.145	20:50
Breath Volume		1.691	20:50
Air Blank		0.000	20:50
Control Test		0.077	20:51
Air Blank		0.000	20:51
Diagnostics Check		OK	20:51

Data Download Date/Time: 08/29/2016 10:42

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONDA,SHUAIB,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3RTG
DRY GAS EXPIRY DATE: 11/11/2022
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: DIEUJUSTE

DATE/TIME OF TEST: 07/17/2016 03:25:20

Diagnostics Check		OK	03:18
Air Blank		0.000	03:18
Control Test		0.077	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.062	03:20
Breath Volume		1.394	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.062	03:23
Breath Volume		1.285	03:23
Air Blank		0.000	03:24
Control Test		0.077	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 08/31/2016 22:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONDA,SHUAIB,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3RTG
DRY GAS EXPIRY DATE: 11/11/2022
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ROUSSEAU

DATE/TIME OF TEST: 07/22/2016 04:25:20

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.078	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.056	04:20
Breath Volume		1.718	04:20
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.055	04:23
Breath Volume		1.347	04:23
Air Blank		0.000	04:24
Control Test		0.077	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 08/31/2016 22:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONDA,SHUAIB,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CURADO,GLAUDSON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3RTG
DRY GAS EXPIRY DATE: 11/11/2022
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: DELANG

DATE/TIME OF TEST: 07/29/2016 00:30:18

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.079	00:20
Air Blank		0.000	00:21
Subject Sample	# 1	0.198	00:23
Breath Volume		1.976	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.196	00:28
Breath Volume		1.769	00:28
Air Blank		0.000	00:29
Control Test		0.078	00:29
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 08/31/2016 22:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONDA,SHUAIB,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ZITO,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3RTG
DRY GAS EXPIRY DATE: 11/11/2022
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/29/2016 04:06:14

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.079	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.166	04:01
Breath Volume		2.824	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.169	04:04
Breath Volume		2.796	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 08/31/2016 22:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006635

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOONDA,SHUAIB,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 3RTG
DRY GAS EXPIRY DATE: 11/11/2022
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BIEBER

DATE/TIME OF TEST: 07/31/2016 02:42:26

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.080	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.137	02:37
Breath Volume		1.488	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.143	02:40
Breath Volume		2.289	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 08/31/2016 22:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123458
DRY GAS EXPIRY DATE: 11/11/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PEREZ ORDONEZ

DATE/TIME OF TEST: 07/21/2016 00:22:17

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.161	00:17
Breath Volume		2.296	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.151	00:20
Breath Volume		1.613	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 08/29/2016 03:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123458
DRY GAS EXPIRY DATE: 11/11/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 07/22/2016 03:03:41

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.113	02:59
Breath Volume		1.781	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.112	03:02
Breath Volume		1.593	03:02
Air Blank		0.000	03:02
Control Test		0.078	03:03
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 08/29/2016 03:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123458
DRY GAS EXPIRY DATE: 11/11/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: HORAN

DATE/TIME OF TEST: 07/26/2016 03:09:22

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.144	03:04
Breath Volume		1.718	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.145	03:07
Breath Volume		1.820	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 08/29/2016 03:11

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: TURANO

DATE/TIME OF TEST: 07/23/2016 22:57:44

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.079	22:51
Air Blank		0.000	22:51
Subject Sample	# 1	0.099	22:52
Breath Volume		2.906	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.097	22:56
Breath Volume		3.042	22:56
Air Blank		0.000	22:56
Control Test		0.078	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 08/02/2016 20:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/24/2016 04:09:13

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.078	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.197	04:04
Breath Volume		2.285	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.185	04:07
Breath Volume		1.992	04:07
Air Blank		0.000	04:08
Control Test		0.078	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 08/02/2016 20:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: VELEZ MALDONADO

DATE/TIME OF TEST: 07/26/2016 03:28:56

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.078	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.035	03:24
Breath Volume		1.179	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.033	03:27
Breath Volume		2.042	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 08/02/2016 20:19

Printed: 06 June 2018

COBRA ©1997-2014 CMI, Inc.