

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1816818
DRY GAS EXPIRY DATE: 02/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CARSON

DATE/TIME OF TEST: 07/01/2016 23:26:30

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.080	23:21
Air Blank	0.000	23:21
Subject Sample # 1	0.125	23:22
Breath Volume	1.757	23:22
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.121	23:24
Breath Volume	1.976	23:24
Air Blank	0.000	23:25
Control Test	0.080	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/04/2016 07:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: GODOY

DATE/TIME OF TEST: 07/18/2016 02:27:54

Diagnostics Check	OK	02:25	
Air Blank	0.000	02:25	
Control Test	0.081	02:25	
Air Blank	0.000	02:26	
Subject Sample	# 1	02:26	Subject Test Refused
Breath Volume	0.000	02:26	
Air Blank	0.000	02:26	
Control Test	0.082	02:27	
Air Blank	0.000	02:27	
Diagnostics Check	OK	02:27	

Data Download Date/Time: 08/14/2016 07:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: PLUSA,KYLA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BLANCO

DATE/TIME OF TEST: 07/20/2016 23:33:00

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.081	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.092	23:28
Breath Volume	1.214	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.096	23:31
Breath Volume	1.464	23:31
Air Blank	0.000	23:32
Control Test	0.081	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:33

Data Download Date/Time: 08/14/2016 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 07/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: DEL CASTILLO

DATE/TIME OF TEST: 07/23/2016 00:28:25

Diagnostics Check	OK	00:21
Air Blank	0.000	00:22
Control Test	0.080	00:22
Air Blank	0.000	00:23
Subject Sample # 1	0.000	00:24
Breath Volume	1.210	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.000	00:27
Breath Volume	1.113	00:27
Air Blank	0.000	00:27
Control Test	0.081	00:27
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 08/14/2016 07:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/23/2016 05:12:07

Diagnostics Check	OK	05:05
Air Blank	0.000	05:05
Control Test	0.080	05:06
Air Blank	0.000	05:06
Subject Sample # 1	0.154	05:06
Breath Volume	1.324	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.148	05:10
Breath Volume	1.570	05:10
Air Blank	0.000	05:11
Control Test	0.081	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 08/14/2016 07:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: STERN

DATE/TIME OF TEST: 07/01/2016 01:46:17

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.158	01:41
Breath Volume	3.082	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.166	01:44
Breath Volume	3.386	01:44
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 07/03/2016 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/03/2016 22:26:25

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:21	
Control Test	0.081	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:25	
Control Test	0.081	22:25	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERLMAN,GREGG,B
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PERLMAN,GREGG
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: BENTLEY

DATE/TIME OF TEST: 07/07/2016 17:03:52

Diagnostics Check	OK	16:57
Air Blank	0.000	16:58
Control Test	0.078	16:58
Air Blank	0.000	16:59
Subject Sample # 1	0.323	16:59
Breath Volume	2.046	16:59
Air Blank	0.000	17:00
Air Blank	0.000	17:01
Subject Sample # 2	0.312	17:02
Breath Volume	1.578	17:02
Air Blank	0.000	17:02
Control Test	0.078	17:03
Air Blank	0.000	17:03
Diagnostics Check	OK	17:03

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: BETZ,FRANK
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: HILLARD

DATE/TIME OF TEST: 07/09/2016 09:50:37

Diagnostics Check		OK	09:43
Air Blank		0.000	09:44
Control Test		0.080	09:44
Air Blank		0.000	09:44
Subject Sample	# 1	0.182	09:45
Breath Volume		2.847	09:45
Air Blank		0.000	09:46
Air Blank		0.000	09:48
Subject Sample	# 2	0.180	09:49
Breath Volume		2.847	09:49
Air Blank		0.000	09:49
Control Test		0.080	09:50
Air Blank		0.000	09:50
Diagnostics Check		OK	09:50

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'CONNELL,KEVIN,M
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: O'CONNELL,KEVIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: CHANDLER

DATE/TIME OF TEST: 07/09/2016 12:57:18

Diagnostics Check	OK	12:54	
Air Blank	0.000	12:54	
Control Test	0.080	12:55	
Air Blank	0.000	12:55	
Subject Sample	# 1	12:55	Subject Test Refused
Breath Volume	0.000	12:55	
Air Blank	0.000	12:56	
Control Test	0.081	12:56	
Air Blank	0.000	12:57	
Diagnostics Check	OK	12:57	

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,K,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: KEOHANE,K
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:40

SUBJECT'S LAST NAME: WIESMAN

DATE/TIME OF TEST: 07/13/2016 14:14:36

Diagnostics Check	OK	14:08
Air Blank	0.000	14:09
Control Test	0.080	14:09
Air Blank	0.000	14:09
Subject Sample # 1	0.000	14:10
Breath Volume	1.914	14:10
Air Blank	0.000	14:10
Air Blank	0.000	14:12
Subject Sample # 2	0.000	14:13
Breath Volume	2.394	14:13
Air Blank	0.000	14:13
Control Test	0.080	14:14
Air Blank	0.000	14:14
Diagnostics Check	OK	14:14

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: HAMSHIRE

DATE/TIME OF TEST: 07/17/2016 00:58:18

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:56	Subject Test Refused
Breath Volume	0.000	00:56	
Air Blank	0.000	00:57	
Control Test	0.080	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/17/2016 20:41:01

Diagnostics Check	OK	20:37	
Air Blank	0.000	20:38	
Control Test	0.080	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:39	Subject Test Refused
Breath Volume	0.000	20:39	
Air Blank	0.000	20:40	
Control Test	0.080	20:40	
Air Blank	0.000	20:40	
Diagnostics Check	OK	20:41	

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/17/2016 23:21:31

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.079	23:19	
Air Blank	0.000	23:19	
Subject Sample	# 1	23:20	Subject Test Refused
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.080	23:20	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: BOUCHARD

DATE/TIME OF TEST: 07/22/2016 04:29:49

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.216	04:25
Breath Volume	2.136	04:25
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.203	04:28
Breath Volume	2.023	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/23/2016 00:21:44

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.279	00:17
Breath Volume	2.898	00:17
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.284	00:20
Breath Volume	2.683	00:20
Air Blank	0.000	00:20
Control Test	0.079	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERLMAN,GREGG,B
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PERLMAN,GREGG
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: HAMILTON JR

DATE/TIME OF TEST: 07/23/2016 12:44:54

Diagnostics Check	OK	12:42	
Air Blank	0.000	12:42	
Control Test	0.078	12:42	
Air Blank	0.000	12:43	
Subject Sample	# 1	12:43	Subject Test Refused
Breath Volume	0.000	12:43	
Air Blank	0.000	12:44	
Control Test	0.078	12:44	
Air Blank	0.000	12:44	
Diagnostics Check	OK	12:44	

Data Download Date/Time: 08/15/2016 08:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGUDO,NATALIE,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RIVERA QUINONES,IGNACIO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: AGUDO

DATE/TIME OF TEST: 07/02/2016 23:51:18

Diagnostics Check		OK	23:35	
Air Blank		0.000	23:35	
Control Test		0.080	23:35	
Air Blank		0.000	23:36	
Subject Sample	# 1	0.167	23:39	Volume Not Met
Breath Volume		0.914	23:39	
Air Blank		0.000	23:40	
Air Blank		0.000	23:41	
Subject Sample	# 2	0.176	23:44	
Breath Volume		2.437	23:44	
Air Blank		0.000	23:44	
Air Blank		0.000	23:46	
Subject Sample	# 3	0.156	23:49	Volume Not Met
Breath Volume		0.824	23:49	
Air Blank		0.000	23:50	
Control Test		0.080	23:50	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:51	

Data Download Date/Time: 07/29/2016 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DE LA MILERA, KARLA,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RIVERA, IGNACIO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: PINEYRO FERNANDEZ

DATE/TIME OF TEST: 07/03/2016 19:37:02

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.080	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.136	19:32
Breath Volume	1.394	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:34
Subject Sample # 2	0.137	19:35
Breath Volume	1.253	19:35
Air Blank	0.000	19:36
Control Test	0.080	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:37

Data Download Date/Time: 07/29/2016 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BHOJWANI

DATE/TIME OF TEST: 07/13/2016 02:39:05

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:33	No Sample Provided
Breath Volume	0.000	02:33	
Air Blank	0.000	02:34	
Air Blank	0.000	02:36	
Subject Sample	# 2	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

Data Download Date/Time: 07/29/2016 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RIVERA QUINONES,IGNACIO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 07/15/2016 23:28:52

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.080	23:22
Air Blank	0.000	23:23
Subject Sample # 1	0.190	23:23
Breath Volume	1.683	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.187	23:27
Breath Volume	3.082	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 07/29/2016 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NARCISSE, ANORCE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: FUENTES, ADALBERTO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: VILLACORTA

DATE/TIME OF TEST: 07/16/2016 04:51:03

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.153	04:46
Breath Volume	3.136	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.148	04:49
Breath Volume	2.550	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:50
Diagnostics Check	OK	04:51

Data Download Date/Time: 07/29/2016 14:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: FUENTES,ALEJANDRO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: AGUIRRE COTRINO

DATE/TIME OF TEST: 07/24/2016 02:22:14

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.147	02:17
Breath Volume	3.679	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.138	02:20
Breath Volume	3.703	02:20
Air Blank	0.000	02:21
Control Test	0.079	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 07/29/2016 14:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDOSO,E,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: MORGADO,Y
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 07/16/2016 03:52:53

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.076	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.243	03:46
Breath Volume	1.433	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:48
Subject Sample # 2	0.228	03:51
Breath Volume	1.972	03:51
Air Blank	0.000	03:51
Control Test	0.076	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 08/22/2016 12:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,ALEXIS,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: DUQUE,ALBERTO
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 07/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: DAIZI

DATE/TIME OF TEST: 07/29/2016 22:12:13

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.077	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.062	22:07
Breath Volume	1.761	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.064	22:10
Breath Volume	1.570	22:10
Air Blank	0.000	22:11
Control Test	0.076	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 08/22/2016 12:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: TAYLOR,DANIEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 07/04/2016 16:39:36

Diagnostics Check	OK	16:38	
Air Blank		16:39	RFI Detect
Air Blank	0.000	16:39	

Data Download Date/Time: 07/26/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: TAYLOR,DANIEL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 07/04/2016 16:43:47

Diagnostics Check	OK	16:40	
Air Blank	0.000	16:41	
Control Test	0.080	16:41	
Air Blank	0.000	16:42	
Subject Sample	# 1	16:42	Subject Test Refused
Breath Volume	0.000	16:42	
Air Blank	0.000	16:42	
Control Test	0.080	16:43	
Air Blank	0.000	16:43	
Diagnostics Check	OK	16:43	

Data Download Date/Time: 07/26/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOCARRAS,PEDRO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:50

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 07/05/2016 10:16:57

Diagnostics Check	OK	10:14	
Air Blank	0.000	10:14	
Control Test	0.080	10:14	
Air Blank	0.000	10:15	
Subject Sample	# 1	10:15	Subject Test Refused
Breath Volume	0.000	10:15	
Air Blank	0.000	10:16	
Control Test	0.081	10:16	
Air Blank	0.000	10:16	
Diagnostics Check	OK	10:16	

Data Download Date/Time: 07/26/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,PEDRO,A
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOCARRAS,PEDRO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 08:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:58

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 07/11/2016 09:26:04

Diagnostics Check	OK	09:19
Air Blank	0.000	09:19
Control Test	0.080	09:19
Air Blank	0.000	09:20
Subject Sample # 1	0.208	09:21
Breath Volume	2.195	09:21
Air Blank	0.000	09:22
Air Blank	0.000	09:23
Subject Sample # 2	0.207	09:24
Breath Volume	1.773	09:24
Air Blank	0.000	09:25
Control Test	0.080	09:25
Air Blank	0.000	09:25
Diagnostics Check	OK	09:26

Data Download Date/Time: 07/26/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COLON,ISSAC
ARRESTING OFFICER AGENCY: Bay Harbor Island PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: GONZALEZ CARRASQUILL

DATE/TIME OF TEST: 07/11/2016 14:24:15

Diagnostics Check	OK	14:16
Air Blank	0.000	14:17
Control Test	0.080	14:17
Air Blank	0.000	14:18
Subject Sample # 1	0.337	14:19
Breath Volume	1.171	14:19
Air Blank	0.000	14:20
Air Blank	0.000	14:21
Subject Sample # 2	0.320	14:22
Breath Volume	1.289	14:22
Air Blank	0.000	14:23
Control Test	0.078	14:23
Air Blank	0.000	14:24
Diagnostics Check	OK	14:24

Data Download Date/Time: 07/26/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGODA,JOSHUA,S
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: GAVALLAS,ELIZABETH
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: ABDULRAHMAN

DATE/TIME OF TEST: 07/11/2016 20:31:59

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.146	20:27
Breath Volume	2.351	20:27
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.149	20:30
Breath Volume	1.980	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 07/26/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN, ANTHONY,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: POMARES, MAKIN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 06/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 14:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:15

SUBJECT'S LAST NAME: MARCONI

DATE/TIME OF TEST: 07/15/2016 14:47:45

Diagnostics Check	OK	14:40
Air Blank	0.000	14:40
Control Test	0.081	14:41
Air Blank	0.000	14:41
Subject Sample # 1	0.125	14:42
Breath Volume	1.265	14:42
Air Blank	0.000	14:42
Air Blank	0.000	14:44
Subject Sample # 2	0.123	14:46
Breath Volume	1.433	14:46
Air Blank	0.000	14:46
Control Test	0.081	14:47
Air Blank	0.000	14:47
Diagnostics Check	OK	14:47

Data Download Date/Time: 07/26/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 07/02/2016 23:21:45

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.081	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:20	Subject Test Refused
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.081	23:21	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: RUDAS-HERNANDEZ

DATE/TIME OF TEST: 07/04/2016 01:18:00

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:09	
Control Test		0.081	01:10	
Air Blank		0.000	01:10	
Subject Sample	# 1	0.106	01:13	Volume Not Met
Breath Volume		0.406	01:13	
Air Blank		0.000	01:14	
Air Blank		0.000	01:16	
Subject Sample	# 2		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.079	01:17	
Air Blank		0.000	01:17	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 09:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: SOACHE-DIAZ

DATE/TIME OF TEST: 07/04/2016 10:01:13

Diagnostics Check	OK	09:53
Air Blank	0.000	09:54
Control Test	0.081	09:54
Air Blank	0.000	09:55
Subject Sample # 1	0.199	09:56
Breath Volume	1.875	09:56
Air Blank	0.000	09:56
Air Blank	0.000	09:58
Subject Sample # 2	0.188	09:59
Breath Volume	1.816	09:59
Air Blank	0.000	10:00
Control Test	0.080	10:00
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JESSE,N
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: EXCEUS,YANICK
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 07/06/2016 07:57:01

Diagnostics Check	OK	07:43	
Air Blank	0.000	07:44	
Control Test	0.081	07:44	
Air Blank	0.000	07:44	
Subject Sample	# 1 0.179	07:47	
Breath Volume	1.179	07:47	
Air Blank	0.000	07:47	
Air Blank	0.000	07:49	
Subject Sample	# 2 0.157	07:52	Volume Not Met
Breath Volume	0.289	07:52	
Air Blank	0.000	07:53	
Air Blank	0.000	07:55	
Subject Sample	# 3	07:55	Subject Test Refused
Breath Volume	0.000	07:55	
Air Blank	0.000	07:56	
Control Test	0.078	07:56	
Air Blank	0.000	07:56	
Diagnostics Check	OK	07:57	

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 07/06/2016 23:26:32

Diagnostics Check	OK	23:18
Air Blank	0.000	23:19
Control Test	0.081	23:19
Air Blank	0.000	23:19
Subject Sample # 1	0.139	23:21
Breath Volume	2.406	23:21
Air Blank	0.000	23:22
Air Blank	0.000	23:24
Subject Sample # 2	0.137	23:24
Breath Volume	2.015	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Diagnostics Check	OK	23:26

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: AQUINO

DATE/TIME OF TEST: 07/07/2016 02:18:46

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.079	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1	0.157	02:10	Volume Not Met
Breath Volume		0.617	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.166	02:13	
Breath Volume		1.367	02:13	
Air Blank		0.000	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 3	0.173	02:17	
Breath Volume		2.101	02:17	
Air Blank		0.000	02:17	
Control Test		0.078	02:18	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: ARTOLA

DATE/TIME OF TEST: 07/09/2016 22:59:13

Diagnostics Check	OK	22:56	
Air Blank	0.000	22:56	
Control Test	0.081	22:57	
Air Blank	0.000	22:57	
Subject Sample	# 1	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:58	
Control Test	0.081	22:58	
Air Blank	0.000	22:59	
Diagnostics Check	OK	22:59	

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,JASON,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WAITE,JASON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/10/2016 21:19:39

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.148	21:15
Breath Volume	2.144	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.142	21:18
Breath Volume	2.234	21:18
Air Blank	0.000	21:18
Control Test	0.080	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:19

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 07/10/2016 23:56:20

Diagnostics Check	OK	23:50
Air Blank	0.000	23:50
Control Test	0.081	23:51
Air Blank	0.000	23:51
Subject Sample # 1	0.174	23:51
Breath Volume	2.335	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.172	23:54
Breath Volume	2.742	23:54
Air Blank	0.000	23:55
Control Test	0.080	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENKINS,JESSICA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JENKINS,JESSICA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: ORTOLA

DATE/TIME OF TEST: 07/19/2016 20:47:21

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.081	20:39	
Air Blank	0.000	20:40	
Subject Sample	# 1	20:43	No Sample Provided
Breath Volume	0.000	20:43	
Air Blank	0.000	20:43	
Air Blank	0.000	20:45	
Subject Sample	# 2	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:46	
Control Test	0.081	20:46	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,AMY,J
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CLERVIL,JUNIOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: MELTON

DATE/TIME OF TEST: 07/20/2016 20:55:46

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:44	
Control Test	0.077	20:45	
Air Blank	0.000	20:45	
Subject Sample	# 1	20:48	No Sample Provided
Breath Volume	0.000	20:48	
Air Blank	0.000	20:49	
Air Blank	0.000	20:51	
Subject Sample	# 2	20:54	No Sample Provided
Breath Volume	0.000	20:54	
Air Blank	0.000	20:54	
Control Test	0.078	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:55	

Data Download Date/Time: 07/22/2016 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAMAYO,KEVIN,X
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TAMAYO,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CHENSKIY

DATE/TIME OF TEST: 07/22/2016 00:57:27

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.077	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.105	00:55	
Air Blank	0.000	00:56	
Control Test	0.077	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 08/13/2016 00:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIELDS,REASHA,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FIELDS,REASHA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/24/2016 20:17:28

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:15	
Control Test	0.078	20:15	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:16	Subject Test Refused
Breath Volume	0.000	20:16	
Air Blank	0.000	20:16	
Control Test	0.078	20:16	
Air Blank	0.000	20:17	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 08/13/2016 00:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HART,MARK,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 07/29/2016 02:55:00

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.079	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.087	02:50
Breath Volume	1.257	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.092	02:53
Breath Volume	1.562	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/13/2016 00:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: ZEREWORCK

DATE/TIME OF TEST: 07/01/2016 00:37:13

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:36	
Control Test	0.078	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SANCHEZ HERNANDEZ

DATE/TIME OF TEST: 07/02/2016 00:48:35

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:43
Air Blank	0.000	00:43
Subject Sample # 1	0.119	00:43
Breath Volume	2.027	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:46
Subject Sample # 2	0.113	00:46
Breath Volume	1.734	00:46
Air Blank	0.000	00:47
Control Test	0.078	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,CHRISTINA,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MARTIN,CHRISTINA
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PELAEZ

DATE/TIME OF TEST: 07/02/2016 03:47:34

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.167	03:42
Breath Volume	4.332	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.166	03:45
Breath Volume	4.460	03:45
Air Blank	0.000	03:46
Control Test	0.076	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 07/03/2016 02:12:59

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.175	02:08
Breath Volume	1.695	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.178	02:11
Breath Volume	1.511	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: ANDRE

DATE/TIME OF TEST: 07/04/2016 01:13:08

Diagnostics Check	OK	01:03
Air Blank	0.000	01:04
Control Test	0.078	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.099	01:07
Breath Volume	2.160	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.102	01:11
Breath Volume	2.164	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 07/04/2016 04:38:55

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.192	04:33
Breath Volume	1.183	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.195	04:37
Breath Volume	1.296	04:37
Air Blank	0.000	04:38
Control Test	0.078	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: O'DONNELL

DATE/TIME OF TEST: 07/04/2016 16:59:27

Diagnostics Check	OK	16:56	
Air Blank	0.000	16:56	
Control Test	0.079	16:57	
Air Blank	0.000	16:57	
Subject Sample	# 1	16:58	Subject Test Refused
Breath Volume	0.000	16:58	
Air Blank	0.000	16:58	
Control Test	0.078	16:58	
Air Blank	0.000	16:59	
Diagnostics Check	OK	16:59	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: MARTINEZ,ANA
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/06/2016 22:42:36

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.096	22:38
Breath Volume	1.167	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.098	22:41
Breath Volume	1.507	22:41
Air Blank	0.000	22:41
Control Test	0.078	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MUSTEDANAGIC

DATE/TIME OF TEST: 07/07/2016 22:37:37

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.120	22:32
Breath Volume	1.808	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.126	22:36
Breath Volume	1.914	22:36
Air Blank	0.000	22:36
Control Test	0.077	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: VALDERRAMA

DATE/TIME OF TEST: 07/08/2016 01:09:19

Diagnostics Check	OK	01:06	
Air Blank	0.000	01:06	
Control Test	0.079	01:07	
Air Blank	0.000	01:07	
Subject Sample	# 1	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.078	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ARENCIBIA

DATE/TIME OF TEST: 07/08/2016 01:30:06

Diagnostics Check	OK	01:21
Air Blank	0.000	01:21
Control Test	0.078	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.281	01:24
Breath Volume	1.359	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.267	01:28
Breath Volume	1.496	01:28
Air Blank	0.000	01:29
Control Test	0.076	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CARMENATES

DATE/TIME OF TEST: 07/08/2016 03:09:39

Diagnostics Check		OK	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.279	03:04
Breath Volume		1.378	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.280	03:07
Breath Volume		1.953	03:07
Air Blank		0.000	03:08
Control Test		0.077	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ABELL

DATE/TIME OF TEST: 07/13/2016 02:45:49

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:43	
Control Test	0.079	02:43	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:44	Subject Test Refused
Breath Volume	0.000	02:44	
Air Blank	0.000	02:44	
Control Test	0.078	02:45	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/14/2016 22:20:51

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.078	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:10	No Sample Provided
Breath Volume	0.000	22:10	
Air Blank	0.000	22:10	
Air Blank	0.000	22:12	
Subject Sample	# 2	0.122	22:14
Breath Volume	1.460	22:14	
Air Blank	0.000	22:15	
Air Blank	0.000	22:17	
Subject Sample	# 3	0.113	22:19
Breath Volume	2.421	22:19	
Air Blank	0.000	22:19	
Control Test	0.077	22:20	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/16/2016 03:49:03

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.079	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.078	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUESTA,DANIEL,A
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUESTA,DANIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 07/16/2016 06:46:08

Diagnostics Check	OK	06:34	
Air Blank	0.000	06:35	
Control Test	0.079	06:35	
Air Blank	0.000	06:36	
Subject Sample	# 1	06:39	No Sample Provided
Breath Volume	0.000	06:39	
Air Blank	0.000	06:39	
Air Blank	0.000	06:41	
Subject Sample	# 2	06:44	No Sample Provided
Breath Volume	0.000	06:44	
Air Blank	0.000	06:45	
Control Test	0.079	06:45	
Air Blank	0.000	06:46	
Diagnostics Check	OK	06:46	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MILLER,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 07/20/2016 05:06:45

Diagnostics Check	OK	05:00
Air Blank	0.000	05:00
Control Test	0.078	05:00
Air Blank	0.000	05:01
Subject Sample # 1	0.060	05:02
Breath Volume	3.093	05:02
Air Blank	0.000	05:02
Air Blank	0.000	05:04
Subject Sample # 2	0.061	05:05
Breath Volume	3.175	05:05
Air Blank	0.000	05:05
Control Test	0.078	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:10

SUBJECT'S LAST NAME: SPRINGER

DATE/TIME OF TEST: 07/21/2016 17:38:54

Diagnostics Check	OK	17:32
Air Blank	0.000	17:33
Control Test	0.079	17:33
Air Blank	0.000	17:34
Subject Sample # 1	0.000	17:34
Breath Volume	1.890	17:34
Air Blank	0.000	17:35
Air Blank	0.000	17:37
Subject Sample # 2	0.000	17:37
Breath Volume	2.464	17:37
Air Blank	0.000	17:37
Control Test	0.077	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY, LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO, LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: AGUAS

DATE/TIME OF TEST: 07/24/2016 00:52:00

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.079	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.016	00:47
Breath Volume	2.894	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.014	00:50
Breath Volume	1.914	00:50
Air Blank	0.000	00:51
Control Test	0.078	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:52

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 07/27/2016 04:42:40

Diagnostics Check	OK	04:39	
Air Blank	0.000	04:40	
Control Test	0.079	04:40	
Air Blank	0.000	04:40	
Subject Sample	# 1	04:41	Subject Test Refused
Breath Volume	0.000	04:41	
Air Blank	0.000	04:41	
Control Test	0.078	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 07/27/2016 09:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: YEPES

DATE/TIME OF TEST: 07/29/2016 06:07:09

Diagnostics Check	OK	05:58	
Air Blank	0.000	05:59	
Control Test	0.079	05:59	
Air Blank	0.000	05:59	
Subject Sample	# 1	06:03	No Sample Provided
Breath Volume	0.000	06:03	
Air Blank	0.000	06:03	
Air Blank	0.000	06:05	
Subject Sample	# 2	06:05	Subject Test Refused
Breath Volume	0.000	06:05	
Air Blank	0.000	06:06	
Control Test	0.077	06:06	
Air Blank	0.000	06:07	
Diagnostics Check	OK	06:07	

Data Download Date/Time: 08/03/2016 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUTLEDGE,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RUTLEDGE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 12:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: SANCHEZ MACHADO

DATE/TIME OF TEST: 07/29/2016 13:39:58

Diagnostics Check	OK	13:30
Air Blank	0.000	13:31
Control Test	0.078	13:31
Air Blank	0.000	13:32
Subject Sample # 1	0.139	13:35
Breath Volume	1.687	13:35
Air Blank	0.000	13:36
Air Blank	0.000	13:37
Subject Sample # 2	0.127	13:38
Breath Volume	1.246	13:38
Air Blank	0.000	13:39
Control Test	0.075	13:39
Air Blank	0.000	13:39
Diagnostics Check	OK	13:39

Data Download Date/Time: 08/03/2016 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/30/2016 01:55:29

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.079	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.078	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 08/03/2016 09:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUAN,RALPH,B
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUAN,RALPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: DA CUNA

DATE/TIME OF TEST: 07/30/2016 18:47:06

Diagnostics Check	OK	18:39
Air Blank	0.000	18:40
Control Test	0.078	18:40
Air Blank	0.000	18:41
Subject Sample # 1	0.160	18:42
Breath Volume	1.558	18:42
Air Blank	0.000	18:43
Air Blank	0.000	18:45
Subject Sample # 2	0.159	18:45
Breath Volume	1.343	18:45
Air Blank	0.000	18:46
Control Test	0.076	18:46
Air Blank	0.000	18:47
Diagnostics Check	OK	18:47

Data Download Date/Time: 08/03/2016 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CALERO ROMERO

DATE/TIME OF TEST: 07/31/2016 03:18:27

Diagnostics Check	OK	03:08
Air Blank	0.000	03:08
Control Test	0.079	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.120	03:11
Breath Volume	1.308	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.121	03:16
Breath Volume	1.269	03:16
Air Blank	0.000	03:17
Control Test	0.076	03:17
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 08/03/2016 09:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: AJIN

DATE/TIME OF TEST: 07/04/2016 02:21:05

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.080	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.160	02:16
Breath Volume	3.046	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.153	02:19
Breath Volume	2.128	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 07/21/2016 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: BEAUBRUN

DATE/TIME OF TEST: 07/07/2016 04:51:11

Diagnostics Check	OK	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.189	04:46
Breath Volume	1.910	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.192	04:49
Breath Volume	1.394	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 07/21/2016 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/09/2016 02:55:49

Diagnostics Check	OK	02:49
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.170	02:51
Breath Volume	1.218	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.174	02:54
Breath Volume	1.546	02:54
Air Blank	0.000	02:54
Control Test	0.080	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 07/21/2016 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 06:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:48

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 07/09/2016 07:20:55

Diagnostics Check	OK	07:14
Air Blank	0.000	07:15
Control Test	0.080	07:15
Air Blank	0.000	07:15
Subject Sample # 1	0.181	07:16
Breath Volume	1.539	07:16
Air Blank	0.000	07:17
Air Blank	0.000	07:18
Subject Sample # 2	0.189	07:19
Breath Volume	1.394	07:19
Air Blank	0.000	07:20
Control Test	0.079	07:20
Air Blank	0.000	07:20
Diagnostics Check	OK	07:20

Data Download Date/Time: 07/21/2016 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SONG,J
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:38

SUBJECT'S LAST NAME: FOOTMAN

DATE/TIME OF TEST: 07/20/2016 09:14:07

Diagnostics Check	OK	09:07
Air Blank	0.000	09:08
Control Test	0.081	09:08
Air Blank	0.000	09:09
Subject Sample # 1	0.139	09:09
Breath Volume	2.449	09:09
Air Blank	0.000	09:10
Air Blank	0.000	09:12
Subject Sample # 2	0.131	09:12
Breath Volume	1.902	09:12
Air Blank	0.000	09:13
Control Test	0.080	09:13
Air Blank	0.000	09:14
Diagnostics Check	OK	09:14

Data Download Date/Time: 07/21/2016 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/24/2016 14:53:13

Diagnostics Check	OK	14:52	
Air Blank		14:52	RFI Detect
Air Blank	0.000	14:53	

Data Download Date/Time: 08/18/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/24/2016 15:04:19

Diagnostics Check	OK	14:57	
Air Blank	0.000	14:58	
Control Test	0.082	14:58	
Air Blank	0.000	14:59	
Subject Sample	# 1 0.244	15:00	
Breath Volume	2.628	15:00	
Air Blank	0.000	15:00	
Air Blank	0.000	15:02	
Subject Sample	# 2 0.244	15:02	
Breath Volume	3.312	15:02	
Air Blank	0.000	15:03	
		15:03	RFI Detect
Air Blank	0.000	15:04	

Data Download Date/Time: 08/18/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/24/2016 15:19:38

Diagnostics Check	OK	15:19	
Air Blank		15:19	RFI Detect
Air Blank	0.000	15:19	

Data Download Date/Time: 08/18/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 07/24/2016 15:26:05

Diagnostics Check	OK	15:22	
Air Blank	0.000	15:23	
Control Test	0.081	15:23	
Air Blank	0.000	15:24	
Subject Sample	# 1	15:24	Subject Test Refused
Breath Volume	0.000	15:24	
Air Blank	0.000	15:25	
Control Test	0.081	15:25	
Air Blank	0.000	15:25	
Diagnostics Check	OK	15:26	

Data Download Date/Time: 08/18/2016 10:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: CHILMAZA

DATE/TIME OF TEST: 07/30/2016 04:45:20

Diagnostics Check	OK	04:39
Air Blank	0.000	04:39
Control Test	0.081	04:40
Air Blank	0.000	04:40
Subject Sample # 1	0.000	04:41
Breath Volume	3.898	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.000	04:43
Breath Volume	3.746	04:43
Air Blank	0.000	04:44
Control Test	0.082	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 08/18/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:53

SUBJECT'S LAST NAME: ORELLANO GONZALEZ

DATE/TIME OF TEST: 07/30/2016 06:25:24

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.081	06:18
Air Blank	0.000	06:18
Subject Sample # 1	0.137	06:20
Breath Volume	3.496	06:20
Air Blank	0.000	06:20
Air Blank	0.000	06:22
Subject Sample # 2	0.149	06:23
Breath Volume	1.949	06:23
Air Blank	0.000	06:24
Control Test	0.083	06:24
Air Blank	0.000	06:25
Diagnostics Check	OK	06:25

Data Download Date/Time: 08/18/2016 10:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GALA ESTEVEZ

DATE/TIME OF TEST: 07/02/2016 02:08:12

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.173	02:03
Breath Volume	1.968	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.176	02:06
Breath Volume	2.269	02:06
Air Blank	0.000	02:07
Control Test	0.077	02:07
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 07/21/2016 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: PUIG FONSECA

DATE/TIME OF TEST: 07/03/2016 01:23:21

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.078	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.248	01:19
Breath Volume		2.039	01:19
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.242	01:21
Breath Volume		2.277	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 07/21/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MORENO,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:31

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 07/07/2016 13:57:56

Diagnostics Check	OK	13:51
Air Blank	0.000	13:52
Control Test	0.079	13:52
Air Blank	0.000	13:53
Subject Sample # 1	0.000	13:53
Breath Volume	1.230	13:53
Air Blank	0.000	13:54
Air Blank	0.000	13:56
Subject Sample # 2	0.000	13:56
Breath Volume	1.675	13:56
Air Blank	0.000	13:57
Control Test	0.080	13:57
Air Blank	0.000	13:57
Diagnostics Check	OK	13:57

Data Download Date/Time: 07/21/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HELMICK

DATE/TIME OF TEST: 07/15/2016 04:03:51

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.078	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.180	03:59
Breath Volume	1.449	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.183	04:02
Breath Volume	1.457	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 07/21/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 07/17/2016 04:40:02

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.163	04:35
Breath Volume	1.484	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.157	04:38
Breath Volume	1.398	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:39
Diagnostics Check	OK	04:40

Data Download Date/Time: 07/21/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: URIAS

DATE/TIME OF TEST: 07/17/2016 08:13:06

Diagnostics Check	OK	08:05
Air Blank	0.000	08:05
Control Test	0.080	08:05
Air Blank	0.000	08:06
Subject Sample # 1	0.159	08:06
Breath Volume	1.328	08:06
Air Blank	0.000	08:07
Air Blank	0.000	08:09
Subject Sample # 2	0.163	08:11
Breath Volume	1.167	08:11
Air Blank	0.000	08:12
Control Test	0.082	08:12
Air Blank	0.000	08:13
Diagnostics Check	OK	08:13

Data Download Date/Time: 07/21/2016 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 08:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:52

SUBJECT'S LAST NAME: ALI

DATE/TIME OF TEST: 07/23/2016 09:15:27

Diagnostics Check	OK	09:12	
Air Blank	0.000	09:13	
Control Test	0.078	09:13	
Air Blank	0.000	09:13	
Subject Sample	# 1	09:14	Subject Test Refused
Breath Volume	0.000	09:14	
Air Blank	0.000	09:14	
Control Test	0.079	09:14	
Air Blank	0.000	09:15	
Diagnostics Check	OK	09:15	

Data Download Date/Time: 08/18/2016 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: CABRAL SMITH

DATE/TIME OF TEST: 07/30/2016 06:28:50

Diagnostics Check	OK	06:21
Air Blank	0.000	06:22
Control Test	0.079	06:22
Air Blank	0.000	06:23
Subject Sample # 1	0.115	06:24
Breath Volume	1.835	06:24
Air Blank	0.000	06:24
Air Blank	0.000	06:26
Subject Sample # 2	0.120	06:27
Breath Volume	1.574	06:27
Air Blank	0.000	06:27
Control Test	0.078	06:28
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 08/18/2016 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 07/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: FIGUEIRA

DATE/TIME OF TEST: 07/30/2016 08:31:26

Diagnostics Check	OK	08:25
Air Blank	0.000	08:25
Control Test	0.078	08:26
Air Blank	0.000	08:26
Subject Sample # 1	0.153	08:27
Breath Volume	1.472	08:27
Air Blank	0.000	08:27
Air Blank	0.000	08:29
Subject Sample # 2	0.145	08:29
Breath Volume	1.781	08:29
Air Blank	0.000	08:30
Control Test	0.079	08:30
Air Blank	0.000	08:31
Diagnostics Check	OK	08:31

Data Download Date/Time: 08/18/2016 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYA,JOSE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ALADIN,K
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/01/2016 04:23:10

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.079	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.000	04:18
Breath Volume	1.125	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.000	04:21
Breath Volume	1.375	04:21
Air Blank	0.000	04:22
Control Test	0.078	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 07/28/2016 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KELLNER,D
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: JULES

DATE/TIME OF TEST: 07/01/2016 06:24:28

Diagnostics Check	OK	06:17
Air Blank	0.000	06:18
Control Test	0.079	06:18
Air Blank	0.000	06:19
Subject Sample # 1	0.187	06:19
Breath Volume	1.125	06:19
Air Blank	0.000	06:20
Air Blank	0.000	06:21
Subject Sample # 2	0.183	06:22
Breath Volume	1.343	06:22
Air Blank	0.000	06:23
Control Test	0.077	06:23
Air Blank	0.000	06:24
Diagnostics Check	OK	06:24

Data Download Date/Time: 07/28/2016 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINCHESTER,AARON,A
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHEAL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 07/09/2016 22:13:16

Diagnostics Check	OK	22:03
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.175	22:07
Breath Volume	1.378	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.172	22:11
Breath Volume	1.191	22:11
Air Blank	0.000	22:12
Control Test	0.076	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 07/28/2016 13:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KIRKUS,JOSHUA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: SAEIH

DATE/TIME OF TEST: 07/23/2016 04:24:14

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.078	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.164	04:17
Breath Volume	1.242	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.160	04:22
Breath Volume	1.218	04:22
Air Blank	0.000	04:23
Control Test	0.076	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 07/28/2016 13:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: OKNYANSKY

DATE/TIME OF TEST: 07/29/2016 00:23:14

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:20	
Control Test	0.079	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1	00:21	Subject Test Refused
Breath Volume	0.000	00:21	
Air Blank	0.000	00:22	
Control Test	0.079	00:22	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 08/31/2016 01:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/29/2016 21:36:01

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.000	21:31
Breath Volume	1.839	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.000	21:34
Breath Volume	1.445	21:34
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:35
Diagnostics Check	OK	21:36

Data Download Date/Time: 08/31/2016 01:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ACEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: JEAN CHARLES

DATE/TIME OF TEST: 07/31/2016 03:56:54

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.079	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.000	03:51
Breath Volume	1.222	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.000	03:55
Breath Volume	1.277	03:55
Air Blank	0.000	03:55
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 08/31/2016 01:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 28214080A1
BREATH TEST OPERATOR: BARTLETT,SHELTON,T	DRY GAS EXPIRY DATE: 11/05/2016
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)	VIOLATION CODE: BUI
ARRESTING OFFICER: BARTLETT,SHELTON	ARREST TIME: 18:01
ARRESTING OFFICER AGENCY: FFWCC (Panama City)	GENDER: Male
LAST AGENCY INSPECTION: 06/21/2016	OBSERVATION PERIOD BEGAN: 18:20
SUBJECT'S LAST NAME: CRAIG	DATE/TIME OF TEST: 07/03/2016 18:52:21

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:48	
Control Test	0.079	18:48	
Air Blank	0.000	18:49	
Subject Sample	# 1	18:50	Subject Test Refused
Breath Volume	0.000	18:50	
Air Blank	0.000	18:51	
Control Test	0.079	18:51	
Air Blank	0.000	18:52	
Diagnostics Check	OK	18:52	

Data Download Date/Time: 07/28/2016 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIFER,JUSTIN,S
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: CORBIN,BRITTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 07/04/2016 17:57:22

Diagnostics Check	OK	17:50
Air Blank	0.000	17:51
Control Test	0.079	17:51
Air Blank	0.000	17:52
Subject Sample # 1	0.122	17:52
Breath Volume	1.878	17:52
Air Blank	0.000	17:53
Air Blank	0.000	17:55
Subject Sample # 2	0.127	17:55
Breath Volume	2.476	17:55
Air Blank	0.000	17:56
Control Test	0.078	17:56
Air Blank	0.000	17:57
Diagnostics Check	OK	17:57

Data Download Date/Time: 07/28/2016 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARTLETT,SHELTON,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MOLNAR,JARROD
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 07/04/2016 19:28:02

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.078	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.130	19:23
Breath Volume	3.367	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.142	19:26
Breath Volume	3.308	19:26
Air Blank	0.000	19:27
Control Test	0.077	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:28

Data Download Date/Time: 07/28/2016 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BARTLETT,SHELTON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 07/04/2016 23:03:45

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.079	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.079	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 07/28/2016 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HUTCHINSON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: BOYINGTON

DATE/TIME OF TEST: 07/16/2016 17:25:22

Diagnostics Check	OK	17:19
Air Blank	0.000	17:19
Control Test	0.078	17:19
Air Blank	0.000	17:20
Subject Sample # 1	0.115	17:20
Breath Volume	1.613	17:20
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.117	17:23
Breath Volume	3.777	17:23
Air Blank	0.000	17:24
Control Test	0.077	17:24
Air Blank	0.000	17:25
Diagnostics Check	OK	17:25

Data Download Date/Time: 07/28/2016 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LAND,MATTHEW
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: HOGUE

DATE/TIME OF TEST: 07/16/2016 17:41:49

Diagnostics Check	OK	17:38	
Air Blank	0.000	17:39	
Control Test	0.079	17:39	
Air Blank	0.000	17:40	
Subject Sample	# 1	17:40	Subject Test Refused
Breath Volume	0.000	17:40	
Air Blank	0.000	17:40	
Control Test	0.078	17:41	
Air Blank	0.000	17:41	
Diagnostics Check	OK	17:41	

Data Download Date/Time: 07/28/2016 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 07/16/2016 19:08:33

Diagnostics Check	OK	18:59	
Air Blank	0.000	18:59	
Control Test	0.078	18:59	
Air Blank	0.000	19:00	
Subject Sample	# 1 0.176	19:03	
Breath Volume	2.410	19:03	
Air Blank	0.000	19:03	
Air Blank	0.000	19:05	
Subject Sample	# 2	19:07	Subject Test Refused
Breath Volume	0.000	19:07	
Air Blank	0.000	19:07	
Control Test	0.076	19:07	
Air Blank	0.000	19:08	
Diagnostics Check	OK	19:08	

Data Download Date/Time: 07/28/2016 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MOLNAR,JARROD
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 16:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 07/31/2016 17:35:47

Diagnostics Check	OK	17:29
Air Blank	0.000	17:30
Control Test	0.080	17:30
Air Blank	0.000	17:31
Subject Sample # 1	0.017	17:31
Breath Volume	3.136	17:31
Air Blank	0.000	17:32
Air Blank	0.000	17:34
Subject Sample # 2	0.013	17:34
Breath Volume	3.230	17:34
Air Blank	0.000	17:34
Control Test	0.077	17:35
Air Blank	0.000	17:35
Diagnostics Check	OK	17:35

Data Download Date/Time: 08/23/2016 17:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLETT,KARL,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HELLETT,KARL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: PARTIN

DATE/TIME OF TEST: 07/02/2016 18:08:40

Diagnostics Check	OK	18:02
Air Blank	0.000	18:02
Control Test	0.076	18:03
Air Blank	0.000	18:03
Subject Sample # 1	0.174	18:04
Breath Volume	4.195	18:04
Air Blank	0.000	18:04
Air Blank	0.000	18:06
Subject Sample # 2	0.174	18:06
Breath Volume	4.464	18:06
Air Blank	0.000	18:07
Control Test	0.077	18:08
Air Blank	0.000	18:08
Diagnostics Check	OK	18:08

Data Download Date/Time: 07/28/2016 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BASFORD,TRAVIS,L
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: BASFORD,TRAVIS
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28214080A1
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: BUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 07/02/2016 22:41:24

Diagnostics Check	OK	22:38	
Air Blank	0.000	22:39	
Control Test	0.078	22:39	
Air Blank	0.000	22:39	
Subject Sample	# 1	22:40	Subject Test Refused
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.078	22:40	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 07/28/2016 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000900

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSTON,ROBERT,B
BREATH TEST OPERATOR AGENCY: FFWCC (Lake City)
ARRESTING OFFICER: JOHNSTON,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 06/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: BUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: TAUNTON

DATE/TIME OF TEST: 07/04/2016 21:50:40

Diagnostics Check	OK	21:44
Air Blank	0.000	21:44
Control Test	0.077	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.147	21:46
Breath Volume	2.191	21:46
Air Blank	0.000	21:46
Air Blank	0.000	21:48
Subject Sample # 2	0.150	21:49
Breath Volume	2.175	21:49
Air Blank	0.000	21:49
Control Test	0.077	21:50
Air Blank	0.000	21:50
Diagnostics Check	OK	21:50

Data Download Date/Time: 07/27/2016 13:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANEY, MICHAEL, P
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: HANEY, MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: BRINDLE

DATE/TIME OF TEST: 07/02/2016 18:13:39

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.081	18:08
Air Blank	0.000	18:08
Subject Sample # 1	0.083	18:09
Breath Volume	2.121	18:09
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.082	18:12
Breath Volume	1.769	18:12
Air Blank	0.000	18:12
Control Test	0.080	18:13
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 07/22/2016 21:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000904

INSTRUMENT REGISTERED TO: FFWCC SOUTH REGION
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MATTSON,CHRIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 07/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: VALLADARES

DATE/TIME OF TEST: 07/30/2016 19:24:17

Diagnostics Check	OK	19:21	
Air Blank	0.000	19:21	
Control Test	0.081	19:22	
Air Blank	0.000	19:22	
Subject Sample	# 1	19:22	Subject Test Refused
Breath Volume	0.000	19:22	
Air Blank	0.000	19:23	
Control Test	0.080	19:23	
Air Blank	0.000	19:24	
Diagnostics Check	OK	19:24	

Data Download Date/Time: 08/22/2016 21:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTH,DUSTIN,L
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: SHERWOOD,JOHN
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32114080A2
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: BAIZE

DATE/TIME OF TEST: 07/17/2016 23:39:31

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.127	23:34
Breath Volume	1.101	23:34
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.119	23:37
Breath Volume	1.636	23:37
Air Blank	0.000	23:38
Control Test	0.082	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 07/27/2016 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROMAN, ANTHONY, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FLORES, S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BRAGG

DATE/TIME OF TEST: 07/02/2016 01:01:30

Diagnostics Check	OK	00:55
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:56
Subject Sample # 1	0.000	00:57
Breath Volume	4.226	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.000	01:00
Breath Volume	2.738	01:00
Air Blank	0.000	01:00
Control Test	0.079	01:00
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 07/06/2016 15:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHILKE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 22:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MILTHALER

DATE/TIME OF TEST: 07/06/2016 23:46:58

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.114	23:42
Breath Volume	3.160	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:44
Subject Sample # 2	0.106	23:45
Breath Volume	1.101	23:45
Air Blank	0.000	23:46
Control Test	0.079	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WILLIAMS,H
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DE SOUZA

DATE/TIME OF TEST: 07/06/2016 23:55:16

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.079	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.000	23:51
Breath Volume	1.316	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.000	23:53
Breath Volume	1.128	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,ALEX,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOTELLO,ALEX
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: HEIL

DATE/TIME OF TEST: 07/08/2016 11:59:47

Diagnostics Check	OK	11:53
Air Blank	0.000	11:53
Control Test	0.079	11:54
Air Blank	0.000	11:54
Subject Sample # 1	0.000	11:55
Breath Volume	2.972	11:55
Air Blank	0.000	11:55
Air Blank	0.000	11:57
Subject Sample # 2	0.000	11:58
Breath Volume	2.882	11:58
Air Blank	0.000	11:58
Control Test	0.078	11:59
Air Blank	0.000	11:59
Diagnostics Check	OK	11:59

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFER,KENDALL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:27

SUBJECT'S LAST NAME: ELOI

DATE/TIME OF TEST: 07/12/2016 20:57:31

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.079	20:51
Air Blank	0.000	20:52
Subject Sample # 1	0.263	20:52
Breath Volume	1.992	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.248	20:55
Breath Volume	2.308	20:55
Air Blank	0.000	20:56
Control Test	0.078	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MAMALIS,N
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: POLARA

DATE/TIME OF TEST: 07/12/2016 22:45:54

Diagnostics Check	OK	22:38
Air Blank	0.000	22:39
Control Test	0.079	22:39
Air Blank	0.000	22:40
Subject Sample # 1	0.120	22:40
Breath Volume	2.203	22:40
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 2	0.125	22:44
Breath Volume	2.054	22:44
Air Blank	0.000	22:44
Control Test	0.078	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: NERONI, KRISTI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:44

SUBJECT'S LAST NAME: SHUMATE

DATE/TIME OF TEST: 07/16/2016 11:27:33

Diagnostics Check	OK	11:20
Air Blank	0.000	11:20
Control Test	0.079	11:21
Air Blank	0.000	11:21
Subject Sample # 1	0.022	11:22
Breath Volume	1.203	11:22
Air Blank	0.000	11:23
Air Blank	0.000	11:25
Subject Sample # 2	0.024	11:26
Breath Volume	1.558	11:26
Air Blank	0.000	11:26
Control Test	0.079	11:26
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWLEY,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ROWLEY,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: OLSON

DATE/TIME OF TEST: 07/17/2016 04:22:23

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.078	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.202	04:17
Breath Volume	1.875	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:19
Subject Sample # 2	0.191	04:20
Breath Volume	1.214	04:20
Air Blank	0.000	04:21
Control Test	0.077	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:37

SUBJECT'S LAST NAME: BLANKENSHIP

DATE/TIME OF TEST: 07/18/2016 19:05:01

Diagnostics Check	OK	19:01	
Air Blank	0.000	19:02	
Control Test	0.079	19:02	
Air Blank	0.000	19:02	
Subject Sample	# 1	19:03	Interferent Detect
Breath Volume	2.535	19:03	
Air Blank	0.000	19:04	
Control Test	0.079	19:04	
Air Blank	0.000	19:04	
Diagnostics Check	OK	19:05	

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ROMAN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GAIDE,DAVID
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 10:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:55

SUBJECT'S LAST NAME: HANNON

DATE/TIME OF TEST: 07/19/2016 11:42:23

Diagnostics Check	OK	11:35
Air Blank	0.000	11:36
Control Test	0.079	11:36
Air Blank	0.000	11:37
Subject Sample # 1	0.000	11:37
Breath Volume	1.195	11:37
Air Blank	0.000	11:38
Air Blank	0.000	11:40
Subject Sample # 2	0.000	11:41
Breath Volume	1.796	11:41
Air Blank	0.000	11:41
Control Test	0.079	11:41
Air Blank	0.000	11:42
Diagnostics Check	OK	11:42

Data Download Date/Time: 08/12/2016 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTIE, EDWARD,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CUTIE, EDWARD
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: MULLEN

DATE/TIME OF TEST: 07/19/2016 12:21:19

Diagnostics Check	OK	12:15
Air Blank	0.000	12:15
Control Test	0.079	12:16
Air Blank	0.000	12:16
Subject Sample # 1	0.000	12:17
Breath Volume	2.261	12:17
Air Blank	0.000	12:17
Air Blank	0.000	12:19
Subject Sample # 2	0.000	12:19
Breath Volume	2.937	12:19
Air Blank	0.000	12:20
Control Test	0.080	12:20
Air Blank	0.000	12:21
Diagnostics Check	OK	12:21

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BARRINGER, B
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 05:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: BILLINGSLEY

DATE/TIME OF TEST: 07/23/2016 06:28:25

Diagnostics Check	OK	06:24	
Air Blank	0.000	06:25	
Control Test	0.078	06:25	
Air Blank	0.000	06:26	
Subject Sample	# 1	06:26	Subject Test Refused
Breath Volume	1.035	06:26	
Air Blank	0.000	06:27	
Control Test	0.078	06:27	
Air Blank	0.000	06:28	
Diagnostics Check	OK	06:28	

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SUTPHIN, S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ST MARTIN

DATE/TIME OF TEST: 07/24/2016 02:01:04

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.078	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.169	01:54
Breath Volume	1.968	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.163	01:59
Breath Volume	1.933	01:59
Air Blank	0.000	02:00
Control Test	0.078	02:00
Air Blank	0.000	02:00
Diagnostics Check	OK	02:01

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: NYBERG

DATE/TIME OF TEST: 07/24/2016 20:53:37

Diagnostics Check	OK	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:48
Subject Sample # 1	0.168	20:49
Breath Volume	2.949	20:49
Air Blank	0.000	20:49
Air Blank	0.000	20:51
Subject Sample # 2	0.156	20:52
Breath Volume	2.187	20:52
Air Blank	0.000	20:52
Control Test	0.078	20:53
Air Blank	0.000	20:53
Diagnostics Check	OK	20:53

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFFER,KENDALL
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MEINHARDT

DATE/TIME OF TEST: 07/25/2016 23:19:24

Diagnostics Check	OK	23:12
Air Blank	0.000	23:12
Control Test	0.078	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.071	23:14
Breath Volume	1.527	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:16
Subject Sample # 2	0.069	23:17
Breath Volume	1.839	23:17
Air Blank	0.000	23:18
Control Test	0.078	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: NERONI, KRISTI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 07/26/2016 13:33:28

Diagnostics Check	OK	13:24
Air Blank	0.000	13:25
Control Test	0.078	13:25
Air Blank	0.000	13:26
Subject Sample # 1	0.000	13:26
Breath Volume	1.488	13:26
Air Blank	0.000	13:27
Air Blank	0.000	13:29
Subject Sample # 2	0.000	13:32
Breath Volume	4.019	13:32
Air Blank	0.000	13:32
Control Test	0.079	13:32
Air Blank	0.000	13:33
Diagnostics Check	OK	13:33

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFER,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: VRAHAS

DATE/TIME OF TEST: 07/26/2016 23:58:09

Diagnostics Check	OK	23:52
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.000	23:53
Breath Volume	3.757	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:56
Breath Volume	3.167	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLURG,JEREMIAH,D
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: MCCLURG,JEREMIAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: WILLETT

DATE/TIME OF TEST: 07/28/2016 14:09:30

Diagnostics Check	OK	14:03
Air Blank	0.000	14:03
Control Test	0.078	14:03
Air Blank	0.000	14:04
Subject Sample # 1	0.231	14:04
Breath Volume	2.320	14:04
Air Blank	0.000	14:05
Air Blank	0.000	14:07
Subject Sample # 2	0.222	14:07
Breath Volume	2.359	14:07
Air Blank	0.000	14:08
Control Test	0.077	14:08
Air Blank	0.000	14:09
Diagnostics Check	OK	14:09

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 21:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: FINN

DATE/TIME OF TEST: 07/28/2016 23:04:47

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.078	23:02	
Air Blank	0.000	23:03	
Subject Sample	# 1	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Control Test	0.079	23:04	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CATES,CURTIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 07/30/2016 02:06:09

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.079	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.218	02:01
Breath Volume		1.421	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:03
Subject Sample	# 2	0.216	02:04
Breath Volume		1.449	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CONDOR

DATE/TIME OF TEST: 07/30/2016 04:25:39

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.126	04:20
Breath Volume	2.644	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.118	04:24
Breath Volume	2.476	04:24
Air Blank	0.000	04:24
Control Test	0.078	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHAEFER,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 04:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: DEAS

DATE/TIME OF TEST: 07/30/2016 05:26:51

Diagnostics Check	OK	05:20
Air Blank	0.000	05:21
Control Test	0.079	05:21
Air Blank	0.000	05:21
Subject Sample # 1	0.128	05:22
Breath Volume	1.707	05:22
Air Blank	0.000	05:23
Air Blank	0.000	05:24
Subject Sample # 2	0.125	05:25
Breath Volume	1.429	05:25
Air Blank	0.000	05:25
Control Test	0.079	05:26
Air Blank	0.000	05:26
Diagnostics Check	OK	05:26

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CATES,CURTIS,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: CATES,CURTIS
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 07/30/2016 05:35:42

Diagnostics Check	OK	05:29
Air Blank	0.000	05:29
Control Test	0.078	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.105	05:31
Breath Volume	1.449	05:31
Air Blank	0.000	05:31
Air Blank	0.000	05:33
Subject Sample # 2	0.102	05:34
Breath Volume	1.195	05:34
Air Blank	0.000	05:34
Control Test	0.079	05:35
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,JUSTIN,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MARSHALL,BRANDON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: REDDICK

DATE/TIME OF TEST: 07/30/2016 11:47:50

Diagnostics Check	OK	11:41
Air Blank	0.000	11:41
Control Test	0.079	11:41
Air Blank	0.000	11:42
Subject Sample # 1	0.000	11:43
Breath Volume	2.644	11:43
Air Blank	0.000	11:43
Air Blank	0.000	11:45
Subject Sample # 2	0.000	11:46
Breath Volume	3.628	11:46
Air Blank	0.000	11:46
Control Test	0.078	11:47
Air Blank	0.000	11:47
Diagnostics Check	OK	11:47

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HAWLEY

DATE/TIME OF TEST: 07/30/2016 22:57:51

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.079	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.139	22:52
Breath Volume	1.160	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.145	22:56
Breath Volume	1.558	22:56
Air Blank	0.000	22:56
Control Test	0.078	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 07/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FRECHETTE

DATE/TIME OF TEST: 07/31/2016 02:27:27

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.078	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.169	02:22
Breath Volume	2.292	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.165	02:25
Breath Volume	1.296	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 08/12/2016 14:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,CLINTON,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMB,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: FOOTS

DATE/TIME OF TEST: 07/23/2016 04:05:38

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.095	04:01
Breath Volume	1.332	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.083	04:04
Breath Volume	2.035	04:04
Air Blank	0.000	04:04
Control Test	0.075	04:05
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NODARSE,MARIO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TUFF,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MARTINEZ MUNOZ

DATE/TIME OF TEST: 07/23/2016 04:15:51

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.134	04:10
Breath Volume	3.066	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.128	04:14
Breath Volume	2.343	04:14
Air Blank	0.000	04:14
Control Test	0.078	04:15
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEMONBREUN,ANTHONY,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,KURT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 13:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: DESAUTELS

DATE/TIME OF TEST: 07/23/2016 14:03:48

Diagnostics Check	OK	13:57
Air Blank	0.000	13:58
Control Test	0.078	13:58
Air Blank	0.000	13:58
Subject Sample # 1	0.092	13:59
Breath Volume	2.789	13:59
Air Blank	0.000	14:00
Air Blank	0.000	14:01
Subject Sample # 2	0.092	14:02
Breath Volume	2.148	14:02
Air Blank	0.000	14:02
Control Test	0.078	14:03
Air Blank	0.000	14:03
Diagnostics Check	OK	14:03

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,SHANE,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLOMON,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/24/2016 00:16:16

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:10
Subject Sample	# 1	0.053	00:11
Breath Volume		2.261	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:13
Subject Sample	# 2	0.052	00:14
Breath Volume		3.203	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON,MITCHELL,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KLEIS,K
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SANKIN

DATE/TIME OF TEST: 07/24/2016 00:47:47

Diagnostics Check	OK	00:41
Air Blank	0.000	00:41
Control Test	0.078	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.100	00:43
Breath Volume	1.937	00:43
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.091	00:46
Breath Volume	1.761	00:46
Air Blank	0.000	00:46
Control Test	0.079	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VADNEY, MICHELE, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMBS, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CLAIR

DATE/TIME OF TEST: 07/24/2016 02:13:22

Diagnostics Check	OK	02:04
Air Blank	0.000	02:04
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.169	02:08
Breath Volume	1.863	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.170	02:11
Breath Volume	1.269	02:11
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROLIGA,ALEXANDRA,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAFFORD,J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ZARIPOV

DATE/TIME OF TEST: 07/25/2016 21:48:03

Diagnostics Check	OK	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:44	
Air Blank	0.000	21:44	
Subject Sample	# 1 0.190	21:46	
Breath Volume	1.464	21:46	
Air Blank		21:47	Purge Fail
Air Blank	0.000	21:48	

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROLIGA,ALEXANDRA,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAFFORD,J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ZARIPOV

DATE/TIME OF TEST: 07/25/2016 21:51:58

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:51	
Control Test	0.073	21:51	Control Outside Tolerance
Air Blank	0.000	21:51	

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROLIGA,ALEXANDRA,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAFFORD,J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ZARIPOV

DATE/TIME OF TEST: 07/25/2016 22:14:22

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:04	
Control Test	0.075	22:05	
Air Blank	0.000	22:05	
Subject Sample	# 1 0.189	22:08	
Breath Volume	1.695	22:08	
Air Blank	0.000	22:09	
Air Blank	0.000	22:10	
Subject Sample	# 2 0.200	22:12	
Breath Volume	2.851	22:12	
Air Blank		22:13	Purge Fail
Air Blank	0.000	22:14	

Data Download Date/Time: 08/02/2016 13:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SILL,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/01/2016 02:52:11

Diagnostics Check	OK	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.199	02:47
Breath Volume	2.011	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.196	02:50
Breath Volume	1.863	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 07/02/2016 02:18:33

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.078	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.054	02:14
Breath Volume	2.734	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.052	02:17
Breath Volume	2.183	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BJAZEVIC

DATE/TIME OF TEST: 07/02/2016 04:36:12

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.079	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:29	No Sample Provided
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2	04:34	No Sample Provided
Breath Volume	0.000	04:34	
Air Blank	0.000	04:35	
Control Test	0.079	04:35	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/03/2016 00:00:11

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.201	23:55
Breath Volume	1.160	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.218	23:58
Breath Volume	1.585	23:58
Air Blank	0.000	23:59
Control Test	0.078	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: RODRIGUEZ, R
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/03/2016 04:27:11

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.078	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.193	04:20
Breath Volume	1.562	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:22
Subject Sample # 2	0.184	04:25
Breath Volume	1.679	04:25
Air Blank	0.000	04:26
Control Test	0.080	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN, DAVID, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN, DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 07/05/2016 21:55:27

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:40	
Control Test		0.079	21:40	
Air Blank		0.000	21:41	
Subject Sample	# 1	0.178	21:44	Volume Not Met
Breath Volume		0.675	21:44	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 2	0.191	21:48	
Breath Volume		1.546	21:48	
Air Blank		0.000	21:48	
Air Blank		0.000	21:50	
Subject Sample	# 3	0.170	21:53	Volume Not Met
Breath Volume		0.718	21:53	
Air Blank		0.000	21:54	
Control Test		0.078	21:54	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ, M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: EASTMAN

DATE/TIME OF TEST: 07/06/2016 04:05:10

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.177	04:00
Breath Volume	2.652	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.160	04:03
Breath Volume	1.398	04:03
Air Blank	0.000	04:04
Control Test	0.079	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOLIEK,P,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BOLIEK,P
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:26

SUBJECT'S LAST NAME: UNTEA

DATE/TIME OF TEST: 07/06/2016 16:00:20

Diagnostics Check	OK	15:51
Air Blank	0.000	15:52
Control Test	0.079	15:52
Air Blank	0.000	15:53
Subject Sample # 1	0.179	15:55
Breath Volume	1.152	15:55
Air Blank	0.000	15:56
Air Blank	0.000	15:57
Subject Sample # 2	0.199	15:58
Breath Volume	1.652	15:58
Air Blank	0.000	15:59
Control Test	0.076	15:59
Air Blank	0.000	16:00
Diagnostics Check	OK	16:00

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,HANSER,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DARREN,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/09/2016 17:56:06

Diagnostics Check	OK	17:49
Air Blank	0.000	17:50
Control Test	0.079	17:50
Air Blank	0.000	17:51
Subject Sample # 1	0.094	17:51
Breath Volume	1.796	17:51
Air Blank	0.000	17:52
Air Blank	0.000	17:54
Subject Sample # 2	0.095	17:54
Breath Volume	2.203	17:54
Air Blank	0.000	17:55
Control Test	0.078	17:55
Air Blank	0.000	17:56
Diagnostics Check	OK	17:56

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: TRESSLER

DATE/TIME OF TEST: 07/10/2016 03:39:01

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.224	03:34
Breath Volume		1.632	03:34
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.214	03:37
Breath Volume		1.246	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:39

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 06/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: FINA

DATE/TIME OF TEST: 07/10/2016 21:44:28

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.079	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.178	21:40
Breath Volume	3.605	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.176	21:42
Breath Volume	3.367	21:42
Air Blank	0.000	21:43
Control Test	0.079	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 07/11/2016 12:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: FOOTE

DATE/TIME OF TEST: 07/15/2016 22:24:06

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.078	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.101	22:19
Breath Volume	2.824	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.097	22:22
Breath Volume	3.132	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEPANSKI,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,REBECCA
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MILLER-FOOTE

DATE/TIME OF TEST: 07/15/2016 22:53:14

Diagnostics Check	OK	22:49	
Air Blank	0.000	22:49	
Control Test	0.079	22:49	
Air Blank	0.000	22:50	
Subject Sample	# 1	22:51	Subject Test Refused
Breath Volume	0.898	22:51	
Air Blank	0.000	22:52	
Control Test	0.078	22:52	
Air Blank	0.000	22:53	
Diagnostics Check	OK	22:53	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEPANSKI,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MILLER-FOOTE

DATE/TIME OF TEST: 07/15/2016 22:57:03

Diagnostics Check	OK	22:56	
Air Blank		22:56	RFI Detect
Air Blank	0.000	22:57	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEPANSKI,JEFFREY,W
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: MILLER-FOOTE

DATE/TIME OF TEST: 07/15/2016 23:03:17

Diagnostics Check	OK	22:59	
Air Blank	0.000	22:59	
Control Test	0.079	22:59	
Air Blank	0.000	23:00	
Subject Sample	# 1	23:01	Subject Test Refused
Breath Volume	1.085	23:01	
Air Blank	0.000	23:02	
Control Test	0.079	23:02	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,DALE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/16/2016 18:11:27

Diagnostics Check	OK	18:04
Air Blank	0.000	18:05
Control Test	0.078	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.179	18:06
Breath Volume	3.902	18:06
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.179	18:09
Breath Volume	3.687	18:09
Air Blank	0.000	18:10
Control Test	0.078	18:10
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/17/2016 03:21:56

Diagnostics Check	OK	03:15
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.193	03:17
Breath Volume	1.324	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.191	03:20
Breath Volume	1.980	03:20
Air Blank	0.000	03:21
Control Test	0.078	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,SAMUEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: OFFICER,FHP
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:32

SUBJECT'S LAST NAME: DUMOND

DATE/TIME OF TEST: 07/17/2016 10:07:14

Diagnostics Check		OK	09:52	
Air Blank		0.000	09:53	
Control Test		0.079	09:53	
Air Blank		0.000	09:53	
Subject Sample	# 1	0.162	09:54	
Breath Volume		1.312	09:54	
Air Blank		0.000	09:54	
Air Blank		0.000	09:56	
Subject Sample	# 2	0.133	09:59	Volume Not Met
Breath Volume		0.625	09:59	
Air Blank		0.000	10:00	
Air Blank		0.000	10:02	
Subject Sample	# 3	0.140	10:05	Volume Not Met
Breath Volume		0.679	10:05	
Air Blank		0.000	10:06	
Control Test		0.077	10:06	
Air Blank		0.000	10:07	
Diagnostics Check		OK	10:07	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: BATCHER

DATE/TIME OF TEST: 07/17/2016 19:53:12

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.078	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.199	19:48
Breath Volume	2.128	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.205	19:51
Breath Volume	2.007	19:51
Air Blank	0.000	19:52
Control Test	0.078	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERFERT,RANDY,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ARREOLA MORALES

DATE/TIME OF TEST: 07/18/2016 03:41:51

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1 0.145	03:34	Volume Not Met
Breath Volume	0.316	03:34	
Air Blank	0.000	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 2 0.147	03:40	Volume Not Met
Breath Volume	0.105	03:40	
Air Blank	0.000	03:40	
Control Test	0.078	03:41	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERFERT,RANDY,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEAR,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ARREOLA MORALES

DATE/TIME OF TEST: 07/18/2016 03:58:16

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.077	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1 0.161	03:46	
Breath Volume	1.621	03:46	
Air Blank	0.000	03:47	
Air Blank	0.000	03:48	
Subject Sample	# 2 0.141	03:52	Volume Not Met
Breath Volume	0.339	03:52	
Air Blank	0.000	03:52	
Air Blank	0.000	03:54	
Subject Sample	# 3 0.145	03:56	
Breath Volume	2.035	03:56	
Air Blank	0.000	03:57	
Control Test	0.076	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: AUSTIN,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: KEEFE

DATE/TIME OF TEST: 07/18/2016 23:31:11

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.000	23:26
Breath Volume	2.101	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.000	23:29
Breath Volume	1.925	23:29
Air Blank	0.000	23:30
Control Test	0.078	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: MASOUD

DATE/TIME OF TEST: 07/21/2016 22:53:32

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Subject Sample # 1	0.079	22:49
Breath Volume	2.460	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.083	22:52
Breath Volume	2.437	22:52
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN,DAVID,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN,DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: IGNERI

DATE/TIME OF TEST: 07/23/2016 02:57:37

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Subject Sample # 1	0.180	02:52
Breath Volume	1.679	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.179	02:56
Breath Volume	2.644	02:56
Air Blank	0.000	02:56
Control Test	0.078	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NODARSE,MARIO,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TUFF,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: MARTINEZ MUNOZ

DATE/TIME OF TEST: 07/23/2016 03:45:01

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.078	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1 0.146	03:37	
Breath Volume	2.703	03:37	
Air Blank	0.000	03:37	
Air Blank	0.000	03:39	
Subject Sample	# 2 0.115	03:40	
Breath Volume	3.917	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 3 0.127	03:43	
Breath Volume	2.171	03:43	
Air Blank	0.000	03:44	
Control Test	0.071	03:44	Control Outside Tolerance
Air Blank	0.000	03:45	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,CLINTON,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMB,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: FOOTS

DATE/TIME OF TEST: 07/23/2016 03:54:24

Diagnostics Check	OK	03:52	
Air Blank	0.000	03:53	
Control Test	0.069	03:53	Control Outside Tolerance
Air Blank	0.000	03:54	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SOLOMON,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 07/24/2016 00:32:46

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.078	00:27
Air Blank	0.000	00:27
Subject Sample # 1	0.120	00:28
Breath Volume	2.695	00:28
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.118	00:31
Breath Volume	2.847	00:31
Air Blank	0.000	00:31
Control Test	0.078	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VECIANA, YASSIEL, A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: PEREZ SANCHEZ

DATE/TIME OF TEST: 07/25/2016 03:19:39

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Subject Sample	# 1	0.206	03:15
Breath Volume		1.726	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.205	03:18
Breath Volume		1.265	03:18
Air Blank		0.000	03:18
Control Test		0.078	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROLIGA,ALEXANDRA,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAFFORD,J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ZARIPOV

DATE/TIME OF TEST: 07/25/2016 21:38:24

Diagnostics Check	OK	21:25	
Air Blank	0.000	21:26	
Control Test	0.079	21:26	
Air Blank	0.000	21:26	
Subject Sample	# 1 0.170	21:30	Volume Not Met
Breath Volume	0.339	21:30	
Air Blank	0.000	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 2 0.197	21:33	
Breath Volume	1.753	21:33	
Air Blank	0.000	21:34	
Air Blank	0.000	21:36	
Subject Sample	# 3 0.214	21:36	
Breath Volume	1.621	21:36	
Air Blank	0.000	21:37	
Control Test	0.074	21:37	Control Outside Tolerance
Air Blank	0.000	21:38	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARROLIGA,ALEXANDRA,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAFFORD,J
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: ZARIPOV

DATE/TIME OF TEST: 07/25/2016 22:39:07

Diagnostics Check		OK	22:23	
Air Blank		0.000	22:24	
Control Test		0.078	22:24	
Air Blank		0.000	22:25	
Subject Sample	# 1	0.171	22:28	Volume Not Met
Breath Volume		0.523	22:28	
Air Blank		0.000	22:29	
Air Blank		0.000	22:30	
Subject Sample	# 2	0.190	22:31	
Breath Volume		1.957	22:31	
Air Blank		0.000	22:32	
Air Blank		0.000	22:34	
Subject Sample	# 3	0.177	22:37	Volume Not Met
Breath Volume		0.054	22:37	
Air Blank		0.000	22:38	
Control Test		0.075	22:38	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HOUWARI

DATE/TIME OF TEST: 07/26/2016 00:50:21

Diagnostics Check	OK	00:44
Air Blank	0.000	00:44
Control Test	0.077	00:45
Air Blank	0.000	00:45
Subject Sample # 1	0.120	00:45
Breath Volume	3.132	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:48
Subject Sample # 2	0.112	00:48
Breath Volume	3.097	00:48
Air Blank	0.000	00:49
Control Test	0.077	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 08/02/2016 13:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICARLO,J,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEPASQUALE,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: KORBEL

DATE/TIME OF TEST: 07/27/2016 03:33:45

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.078	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.126	03:29
Breath Volume	2.566	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.118	03:32
Breath Volume	1.113	03:32
Air Blank	0.000	03:32
Control Test	0.077	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 08/02/2016 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,NICKOLAUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/28/2016 22:57:55

Diagnostics Check		OK	22:50
Air Blank		0.000	22:50
Control Test		0.079	22:50
Air Blank		0.000	22:51
Subject Sample	# 1	0.251	22:52
Breath Volume		1.398	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.261	22:56
Breath Volume		1.222	22:56
Air Blank		0.000	22:57
Control Test		0.077	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 08/02/2016 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SAVEL,MICHAEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 07/29/2016 21:49:33

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.078	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.196	21:45
Breath Volume	3.210	21:45
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.194	21:47
Breath Volume	3.082	21:47
Air Blank	0.000	21:48
Control Test	0.078	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 08/02/2016 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 07/30/2016 02:35:01

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.200	02:30
Breath Volume		2.546	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.196	02:33
Breath Volume		2.328	02:33
Air Blank		0.000	02:34
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:35

Data Download Date/Time: 08/02/2016 13:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN, DAVID, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN, DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 07/30/2016 04:11:57

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.078	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.154	04:07
Breath Volume	2.339	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:09
Subject Sample # 2	0.149	04:10
Breath Volume	2.093	04:10
Air Blank	0.000	04:11
Control Test	0.077	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 08/02/2016 13:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ACEVEDO

DATE/TIME OF TEST: 07/30/2016 01:44:36

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.077	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 1	0.000	01:35	Volume Not Met
Breath Volume		0.253	01:35	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 2	0.000	01:39	
Breath Volume		2.703	01:39	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 3	0.000	01:43	
Breath Volume		1.546	01:43	
Air Blank		0.000	01:43	
Control Test		0.077	01:44	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

Data Download Date/Time: 08/31/2016 21:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLENDENING,JAMES,T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MCQUAID,DEWAYNE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 07/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 07/17/2016 02:17:12

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.080	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:16	
Control Test	0.080	02:16	
Air Blank	0.000	02:17	
Diagnostics Check	OK	02:17	

Data Download Date/Time: 07/29/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: G,T,
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MCQUAID,DWAYNE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 07/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: G

DATE/TIME OF TEST: 07/29/2016 01:26:55

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:22
Subject Sample # 1	0.120	01:22
Breath Volume	2.007	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.118	01:25
Breath Volume	1.773	01:25
Air Blank	0.000	01:25
Control Test	0.079	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 07/29/2016 12:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: STUMP

DATE/TIME OF TEST: 07/05/2016 01:45:47

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.080	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.278	01:41
Breath Volume		2.703	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.276	01:44
Breath Volume		2.449	01:44
Air Blank		0.000	01:44
Control Test		0.077	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 07/22/2016 15:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 07/29/2016 00:39:14

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.082	00:37	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Sequence Aborted
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	

Data Download Date/Time: 08/20/2016 23:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 07/29/2016 00:54:17

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:47	No Sample Provided
Breath Volume	0.000	00:47	
Air Blank	0.000	00:47	
Air Blank	0.000	00:49	
Subject Sample	# 2	00:52	No Sample Provided
Breath Volume	0.000	00:52	
Air Blank	0.000	00:53	
Control Test	0.083	00:53	
Air Blank	0.000	00:54	
Diagnostics Check	OK	00:54	

Data Download Date/Time: 08/20/2016 23:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VOLLERTSEN,CRAIG,W
BREATH TEST OPERATOR AGENCY: Jefferson County SO
ARRESTING OFFICER: VOLLERTSEN,CRAIG
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CLARK'

DATE/TIME OF TEST: 07/16/2016 00:38:18

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:35	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 08/25/2016 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JAMES,D
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: HIRES

DATE/TIME OF TEST: 07/17/2016 02:00:18

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.080	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.000	01:56
Breath Volume	2.042	01:56
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.000	01:58
Breath Volume	1.593	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 08/25/2016 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,STEWART,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH,STEWART
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 07/23/2016 23:01:59

Diagnostics Check	OK	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:56
Subject Sample # 1	0.196	22:57
Breath Volume	3.484	22:57
Air Blank	0.000	22:58
Air Blank	0.000	22:59
Subject Sample # 2	0.190	23:00
Breath Volume	2.535	23:00
Air Blank	0.000	23:01
Control Test	0.079	23:01
Air Blank	0.000	23:01
Diagnostics Check	OK	23:01

Data Download Date/Time: 08/25/2016 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARRAMORE,BRENT,J
BREATH TEST OPERATOR AGENCY: Monticello PD
ARRESTING OFFICER: PARRAMORE,BRENT
ARRESTING OFFICER AGENCY: Monticello PD
LAST AGENCY INSPECTION: 07/11/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/30/2016 20:22:37

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:11	
Control Test	0.079	20:11	
Air Blank	0.000	20:11	
Subject Sample	# 1 0.212	20:15	Volume Not Met
Breath Volume	0.578	20:15	
Air Blank	0.000	20:15	
Air Blank	0.000	20:17	
Subject Sample	# 2 0.271	20:18	
Breath Volume	2.160	20:18	
Air Blank	0.000	20:18	
Air Blank	0.000	20:20	
Subject Sample	# 3 0.275	20:21	
Breath Volume	2.203	20:21	
Air Blank	0.000	20:22	
		20:22	RFI Detect
Air Blank	0.000	20:22	

Data Download Date/Time: 08/25/2016 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SANDERS,EVON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ADAIR

DATE/TIME OF TEST: 07/04/2016 04:04:04

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.143	03:59
Breath Volume	1.347	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.139	04:02
Breath Volume	1.203	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:04

Data Download Date/Time: 07/20/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,JAMES,H
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SHAW,EMILY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/09/2016 00:43:57

Diagnostics Check	OK	00:37
Air Blank	0.000	00:37
Control Test	0.080	00:38
Air Blank	0.000	00:38
Subject Sample # 1	0.149	00:39
Breath Volume	2.171	00:39
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.153	00:42
Breath Volume	2.273	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 07/20/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIDAY,CHAD,E
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: OCONNER,STEPHEN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:11

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 07/11/2016 12:41:30

Diagnostics Check	OK	12:36	
Air Blank	0.000	12:36	
Control Test	0.080	12:36	
Air Blank	0.000	12:37	
Subject Sample	# 1	12:39	Subject Test Refused
Breath Volume	1.039	12:39	
Air Blank	0.000	12:40	
Control Test	0.080	12:40	
Air Blank	0.000	12:41	
Diagnostics Check	OK	12:41	

Data Download Date/Time: 07/20/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FREDRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FREDRICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/17/2016 02:01:36

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.242	01:56
Breath Volume	3.566	01:56
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.243	02:00
Breath Volume	3.683	02:00
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/20/2016 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODS,CHRIS,D
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WOODS,CHRIS
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: DELAVALLE

DATE/TIME OF TEST: 07/17/2016 04:41:11

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.079	04:35
Air Blank	0.000	04:35
Subject Sample # 1	0.097	04:36
Breath Volume	1.730	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.116	04:39
Breath Volume	2.519	04:39
Air Blank	0.000	04:40
Control Test	0.078	04:40
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 07/20/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,EMILY,P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SHAW,EMILY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/22/2016 22:44:37

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:44	RFI Detect
Breath Volume	0.000	22:44	
Air Blank		22:44	RFI Detect
Air Blank	0.000	22:44	

Data Download Date/Time: 08/11/2016 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAW,EMILY,P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SHAW,EMILY
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/22/2016 22:52:04

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.022	22:47
Breath Volume	2.265	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.021	22:50
Breath Volume	2.175	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:52

Data Download Date/Time: 08/11/2016 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DILLS,JAMES,M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: FLOYD,JEVON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: EPSTEIN

DATE/TIME OF TEST: 07/02/2016 00:12:57

Diagnostics Check	OK	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.212	00:08
Breath Volume	3.746	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:10
Subject Sample # 2	0.215	00:11
Breath Volume	3.742	00:11
Air Blank	0.000	00:12
Control Test	0.077	00:12
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 07/20/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 07/02/2016 23:21:20

Diagnostics Check	OK	23:15
Air Blank	0.000	23:15
Control Test	0.081	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.000	23:17
Breath Volume	2.667	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.000	23:19
Breath Volume	2.648	23:19
Air Blank	0.000	23:20
Control Test	0.079	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 07/20/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,STEWART,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMITH,STEWART
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 07/03/2016 22:15:13

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.080	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.181	22:10
Breath Volume	2.222	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:12
Subject Sample # 2	0.179	22:13
Breath Volume	1.976	22:13
Air Blank	0.000	22:14
Control Test	0.078	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 07/20/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BAKER,MARK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: SANTURRI

DATE/TIME OF TEST: 07/04/2016 05:55:37

Diagnostics Check	OK	05:49
Air Blank	0.000	05:49
Control Test	0.080	05:50
Air Blank	0.000	05:50
Subject Sample # 1	0.067	05:51
Breath Volume	4.023	05:51
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.067	05:54
Breath Volume	4.589	05:54
Air Blank	0.000	05:54
Control Test	0.079	05:55
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 07/20/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,MARK,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BAKER,MARK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 07/04/2016 06:21:44

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.080	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.037	06:17
Breath Volume	4.429	06:17
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.030	06:20
Breath Volume	1.570	06:20
Air Blank	0.000	06:20
Control Test	0.079	06:21
Air Blank	0.000	06:21
Diagnostics Check	OK	06:21

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BRUTON

DATE/TIME OF TEST: 07/07/2016 03:05:40

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.136	03:01
Breath Volume	2.535	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.134	03:04
Breath Volume	1.789	03:04
Air Blank	0.000	03:04
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: POMEROY

DATE/TIME OF TEST: 07/08/2016 03:23:44

Diagnostics Check	OK	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:18
Subject Sample # 1	0.133	03:19
Breath Volume	1.699	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.128	03:22
Breath Volume	2.011	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Diagnostics Check	OK	03:23

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,JAMES,H
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PITTMAN,JAMES
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/15/2016 02:53:33

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.090	02:48	
Breath Volume	1.296	02:48	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2 0.090	02:51	
Breath Volume	2.972	02:51	
Air Blank	0.000	02:52	
Control Test	0.079	02:52	
Air Blank		02:53	RFI Detect
Air Blank	0.000	02:53	

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,JAMES,H
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PITTMAN,JAMES
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/15/2016 03:01:36

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.088	02:57
Breath Volume	2.921	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.086	03:00
Breath Volume	2.019	03:00
Air Blank	0.000	03:00
Control Test	0.079	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: CHAFIN

DATE/TIME OF TEST: 07/15/2016 03:42:55

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.079	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.074	03:38
Breath Volume	2.609	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 2	0.070	03:41
Breath Volume	2.308	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FREDRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FREDRICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/17/2016 02:31:37

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
		02:31	RFI Detect
Air Blank	0.000	02:31	

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FREDRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FREDRICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BRADLEY

DATE/TIME OF TEST: 07/17/2016 02:40:03

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.224	02:35
Breath Volume	3.511	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:37
Subject Sample # 2	0.243	02:38
Breath Volume	3.738	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:40

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALLARD SR,JEFFERY,J
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROEDER,DUSTIN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:51

SUBJECT'S LAST NAME: SCOTT-ROUISE

DATE/TIME OF TEST: 07/19/2016 08:09:02

Diagnostics Check	OK	07:57	
Air Blank	0.000	07:58	
Control Test	0.079	07:58	
Air Blank	0.000	07:58	
Subject Sample	# 1	08:02	No Sample Provided
Breath Volume	0.000	08:02	
Air Blank	0.000	08:02	
Air Blank	0.000	08:04	
Subject Sample	# 2	08:07	No Sample Provided
Breath Volume	0.000	08:07	
Air Blank	0.000	08:08	
Control Test	0.080	08:08	
Air Blank	0.000	08:08	
Diagnostics Check	OK	08:09	

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 06/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/20/2016 03:12:56

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.080	03:07
Air Blank	0.000	03:07
Subject Sample # 1	0.178	03:08
Breath Volume	1.695	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:10
Subject Sample # 2	0.176	03:11
Breath Volume	1.152	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:12

Data Download Date/Time: 07/20/2016 11:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 07/21/2016 03:54:04

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.024	03:49
Breath Volume	1.339	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:51
Subject Sample # 2	0.024	03:52
Breath Volume	1.207	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:53
Diagnostics Check	OK	03:54

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/23/2016 04:17:18

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.080	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:10	No Sample Provided
Breath Volume	0.000	04:10	
Air Blank	0.000	04:10	
Air Blank	0.000	04:12	
Subject Sample	# 2	04:15	No Sample Provided
Breath Volume	0.000	04:15	
Air Blank	0.000	04:16	
Control Test	0.080	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, ANDREW, L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: HANCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: YAZDIYA

DATE/TIME OF TEST: 07/24/2016 02:57:58

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.228	02:53
Breath Volume	3.816	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:55
Subject Sample # 2	0.234	02:56
Breath Volume	4.382	02:56
Air Blank	0.000	02:57
Control Test	0.079	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FISCHER,ANDREW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: BALLARD

DATE/TIME OF TEST: 07/25/2016 04:07:35

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.178	04:02
Breath Volume	3.492	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.192	04:05
Breath Volume	2.203	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEGRAW,JACK,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SHERER,DANIEL
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: STANDRIDGE

DATE/TIME OF TEST: 07/27/2016 03:30:44

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.116	03:25
Breath Volume	1.359	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.125	03:29
Breath Volume	2.695	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SIAM

DATE/TIME OF TEST: 07/27/2016 03:59:58

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.080	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.161	03:55
Breath Volume	1.933	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:57
Subject Sample # 2	0.160	03:58
Breath Volume	1.515	03:58
Air Blank	0.000	03:59
Control Test	0.078	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: FINCH

DATE/TIME OF TEST: 07/29/2016 22:22:39

Diagnostics Check	OK	22:16
Air Blank	0.000	22:16
Control Test	0.079	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.186	22:18
Breath Volume	2.941	22:18
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.189	22:21
Breath Volume	1.453	22:21
Air Blank	0.000	22:21
Control Test	0.079	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SHERER,DANIEL
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: S360165878630

DATE/TIME OF TEST: 07/31/2016 02:02:14

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.235	01:57
Breath Volume	1.617	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.239	02:00
Breath Volume	2.078	02:00
Air Blank	0.000	02:01
Control Test	0.079	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEGRAW,JACK,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: DEGRAW,JACK
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 07/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BRASHER

DATE/TIME OF TEST: 07/31/2016 02:37:58

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1	02:36	Subject Test Refused
Breath Volume	0.000	02:36	
Air Blank	0.000	02:37	
Control Test	0.080	02:37	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 08/11/2016 11:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JONATHAN, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BIGGS, JOSHUA
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 07/07/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: EARNEST

DATE/TIME OF TEST: 07/16/2016 22:35:03

Diagnostics Check	OK	22:28
Air Blank	0.000	22:29
Control Test	0.081	22:29
Air Blank	0.000	22:30
Subject Sample # 1	0.144	22:30
Breath Volume	3.105	22:30
Air Blank	0.000	22:31
Air Blank	0.000	22:32
Subject Sample # 2	0.138	22:33
Breath Volume	2.679	22:33
Air Blank	0.000	22:34
Control Test	0.081	22:34
Air Blank	0.000	22:34
Diagnostics Check	OK	22:35

Data Download Date/Time: 08/10/2016 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRAN,JEREMY,H
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ROBB,MITCHELL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: BUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: COCHRAN

DATE/TIME OF TEST: 07/02/2016 19:42:05

Diagnostics Check	OK	19:36
Air Blank	0.000	19:36
Control Test	0.079	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.218	19:37
Breath Volume	2.015	19:37
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.213	19:40
Breath Volume	2.203	19:40
Air Blank	0.000	19:41
Control Test	0.078	19:41
Air Blank	0.000	19:41
Diagnostics Check	OK	19:42

Data Download Date/Time: 07/22/2016 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TALLMAN,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TALLMAN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SLAGLE

DATE/TIME OF TEST: 07/03/2016 03:14:23

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:11	
Control Test	0.079	03:12	
Air Blank	0.000	03:12	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.079	03:13	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 07/22/2016 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TALLMAN,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TALLMAN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: CHURCHILL

DATE/TIME OF TEST: 07/11/2016 21:06:23

Diagnostics Check	OK	21:00
Air Blank	0.000	21:00
Control Test	0.079	21:00
Air Blank	0.000	21:01
Subject Sample # 1	0.250	21:01
Breath Volume	1.539	21:01
Air Blank	0.000	21:02
Air Blank	0.000	21:04
Subject Sample # 2	0.234	21:04
Breath Volume	2.332	21:04
Air Blank	0.000	21:05
Control Test	0.078	21:05
Air Blank	0.000	21:06
Diagnostics Check	OK	21:06

Data Download Date/Time: 07/22/2016 09:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIEN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BRIEN,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 06/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 07/01/2016 00:56:08

Diagnostics Check		OK	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.153	00:51
Breath Volume		2.703	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.148	00:54
Breath Volume		2.371	00:54
Air Blank		0.000	00:55
Control Test		0.078	00:55
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 07/12/2016 11:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: MOLINA-AGUSTIN

DATE/TIME OF TEST: 07/16/2016 16:18:21

Diagnostics Check	OK	16:08
Air Blank	0.000	16:09
Control Test	0.080	16:09
Air Blank	0.000	16:09
Subject Sample # 1	0.252	16:12
Breath Volume	2.773	16:12
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.244	16:16
Breath Volume	2.777	16:16
Air Blank	0.000	16:17
Control Test	0.080	16:17
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 08/08/2016 11:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BRASHER

DATE/TIME OF TEST: 07/17/2016 00:04:16

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.080	23:59	
Air Blank	0.000	23:59	
Subject Sample	# 1 0.132	00:02	Volume Not Met
Breath Volume	0.421	00:02	
Air Blank		00:03	Purge Fail
Air Blank	0.000	00:04	

Data Download Date/Time: 08/08/2016 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BRASHER

DATE/TIME OF TEST: 07/17/2016 00:14:37

Diagnostics Check	OK	00:08
Air Blank	0.000	00:08
Control Test	0.079	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.126	00:10
Breath Volume	2.628	00:10
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.118	00:13
Breath Volume	2.542	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 08/08/2016 11:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURNETTE,JAVON,A

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: BURNETTE,JAVON

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/26/2016 20:48:01

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:45	
Control Test	0.080	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:46	Subject Test Refused
Breath Volume	0.000	20:46	
Air Blank	0.000	20:47	
Control Test	0.080	20:47	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 08/08/2016 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DEON,L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SANDERS,DEON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/12/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 07/31/2016 02:52:37

Diagnostics Check	OK	02:44
Air Blank	0.000	02:44
Control Test	0.080	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.172	02:47
Breath Volume	1.816	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.166	02:51
Breath Volume	1.765	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 08/08/2016 11:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEHL, EDWARD,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DIEHL, EDWARD
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: FIGUEROA RIVERA

DATE/TIME OF TEST: 07/01/2016 23:04:26

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.219	23:00
Breath Volume	1.871	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.221	23:02
Breath Volume	1.937	23:02
Air Blank	0.000	23:03
Control Test	0.079	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOUBERLEY,GREGORY,
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DOUBERLEY,GREGORY
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SOOKRAM

DATE/TIME OF TEST: 07/03/2016 21:32:04

Diagnostics Check	OK	21:26
Air Blank	0.000	21:26
Control Test	0.081	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.201	21:27
Breath Volume	1.589	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.201	21:30
Breath Volume	2.082	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:32

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD,GLENN,W
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: FORD,G
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: OCAMPO VARGAS

DATE/TIME OF TEST: 07/08/2016 01:46:16

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.080	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1	0.212	01:40	Volume Not Met
Breath Volume		0.253	01:40	
Air Blank		0.000	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 2		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.077	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,AYLER,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CRUZ,AYLER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CHEVERE GARCIA

DATE/TIME OF TEST: 07/08/2016 01:58:59

Diagnostics Check	OK	01:51
Air Blank	0.000	01:52
Control Test	0.079	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.134	01:53
Breath Volume	1.878	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.120	01:57
Breath Volume	1.179	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:58
Diagnostics Check	OK	01:58

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: STINSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: TORRES RIVERA

DATE/TIME OF TEST: 07/09/2016 04:32:04

Diagnostics Check	OK	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.108	04:27
Breath Volume	4.316	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.095	04:30
Breath Volume	1.730	04:30
Air Blank	0.000	04:31
Control Test	0.078	04:31
Air Blank	0.000	04:31
Diagnostics Check	OK	04:32

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOW,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BLOW,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/11/2016 01:36:58

Diagnostics Check	OK	01:30
Air Blank	0.000	01:30
Control Test	0.080	01:30
Air Blank	0.000	01:31
Subject Sample # 1	0.226	01:32
Breath Volume	2.042	01:32
Air Blank	0.000	01:33
Air Blank	0.000	01:34
Subject Sample # 2	0.212	01:35
Breath Volume	1.675	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:36
Diagnostics Check	OK	01:36

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: STINSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BORRELL

DATE/TIME OF TEST: 07/11/2016 03:09:47

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.079	03:03
Air Blank	0.000	03:03
Subject Sample # 1	0.193	03:04
Breath Volume	1.687	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.207	03:08
Breath Volume	1.339	03:08
Air Blank	0.000	03:08
Control Test	0.078	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYLE,OWEN,D
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GAYLE,OWEN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:10

SUBJECT'S LAST NAME: FERRAO-SANTIAGO

DATE/TIME OF TEST: 07/19/2016 13:57:18

Diagnostics Check	OK	13:49
Air Blank	0.000	13:50
Control Test	0.081	13:50
Air Blank	0.000	13:51
Subject Sample # 1	0.000	13:52
Breath Volume	1.226	13:52
Air Blank	0.000	13:52
Air Blank	0.000	13:54
Subject Sample # 2	0.000	13:55
Breath Volume	2.347	13:55
Air Blank	0.000	13:56
Control Test	0.079	13:56
Air Blank	0.000	13:57
Diagnostics Check	OK	13:57

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: SOLANO

DATE/TIME OF TEST: 07/24/2016 07:02:04

Diagnostics Check	OK	06:55
Air Blank	0.000	06:55
Control Test	0.081	06:56
Air Blank	0.000	06:56
Subject Sample # 1	0.101	06:57
Breath Volume	2.136	06:57
Air Blank	0.000	06:58
Air Blank	0.000	07:00
Subject Sample # 2	0.104	07:00
Breath Volume	1.417	07:00
Air Blank	0.000	07:01
Control Test	0.081	07:01
Air Blank	0.000	07:01
Diagnostics Check	OK	07:02

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: SMITH,P
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: LEBRON

DATE/TIME OF TEST: 07/24/2016 22:48:09

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.207	22:43
Breath Volume	2.125	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.223	22:46
Breath Volume	2.156	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 07/26/2016 14:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SEAWRIGHT

DATE/TIME OF TEST: 07/04/2016 01:53:15

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.076	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.000	01:49
Breath Volume	1.460	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.000	01:51
Breath Volume	1.464	01:51
Air Blank	0.000	01:52
Control Test	0.076	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 07/15/2016 22:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: SNUGGS

DATE/TIME OF TEST: 07/24/2016 02:29:35

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.075	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.307	02:25
Breath Volume		2.289	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.304	02:27
Breath Volume		2.386	02:27
Air Blank		0.000	02:28
Control Test		0.076	02:28
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 08/26/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: SORRELLS,BRIAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HERRICK

DATE/TIME OF TEST: 07/25/2016 23:13:05

Diagnostics Check	OK	23:01	
Air Blank	0.000	23:02	
Control Test	0.077	23:02	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:06	No Sample Provided
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Air Blank	0.000	23:08	
Subject Sample	# 2	23:11	No Sample Provided
Breath Volume	0.000	23:11	
Air Blank	0.000	23:12	
Control Test	0.077	23:12	
Air Blank	0.000	23:13	
Diagnostics Check	OK	23:13	

Data Download Date/Time: 08/26/2016 21:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: SORRELLS,BRIAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 07/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HERRICK

DATE/TIME OF TEST: 07/25/2016 23:33:49

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.077	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.211	23:29
Breath Volume	1.824	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.210	23:32
Breath Volume	1.976	23:32
Air Blank	0.000	23:32
Control Test	0.076	23:33
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 08/26/2016 21:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 07/02/2016 03:28:50

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.081	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.179	03:24
Breath Volume	1.730	03:24
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.181	03:27
Breath Volume	1.925	03:27
Air Blank	0.000	03:27
Control Test	0.080	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,M
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: LOOSEMOORE

DATE/TIME OF TEST: 07/03/2016 00:32:24

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.188	00:27
Breath Volume	1.847	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.170	00:30
Breath Volume	1.152	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MARK,A
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ANTHONY,MARK
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/03/2016 03:46:51

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.113	03:41
Breath Volume	1.488	03:41
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.104	03:45
Breath Volume	1.878	03:45
Air Blank	0.000	03:45
Control Test	0.081	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: HUSSAIN

DATE/TIME OF TEST: 07/07/2016 00:44:49

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.082	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.000	00:40
Breath Volume	2.687	00:40
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:43
Breath Volume	2.593	00:43
Air Blank	0.000	00:43
Control Test	0.081	00:44
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MARTINEZ-GUEVARA

DATE/TIME OF TEST: 07/16/2016 03:31:50

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.081	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1 0.225	03:18	
Breath Volume	1.339	03:18	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2 0.201	03:24	Volume Not Met
Breath Volume	0.218	03:24	
Air Blank	0.000	03:25	
Air Blank	0.000	03:26	
Subject Sample	# 3 0.233	03:30	
Breath Volume	3.242	03:30	
Air Blank	0.000	03:30	
Control Test	0.077	03:31	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: ESTRADA-MERCEDES

DATE/TIME OF TEST: 07/17/2016 01:12:48

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.081	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.153	01:07
Breath Volume	1.921	01:07
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.137	01:11
Breath Volume	1.109	01:11
Air Blank	0.000	01:11
Control Test	0.081	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:12

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/17/2016 22:07:15

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:02
Subject Sample # 1	0.108	22:02
Breath Volume	1.632	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.120	22:05
Breath Volume	2.136	22:05
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 06/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WORDEN

DATE/TIME OF TEST: 07/21/2016 22:12:20

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.000	22:08
Breath Volume	2.195	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.000	22:10
Breath Volume	2.355	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 07/26/2016 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLERIA,TROY,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: REDDITT,MATTHEW
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 07/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: BUSKER

DATE/TIME OF TEST: 07/29/2016 18:24:59

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:11	
Control Test	0.081	18:11	
Air Blank	0.000	18:12	
Subject Sample	# 1	18:15	No Sample Provided
Breath Volume	0.000	18:15	
Air Blank	0.000	18:15	
Air Blank	0.000	18:17	
Subject Sample	# 2	0.184	18:19
Breath Volume	2.757	18:19	
Air Blank	0.000	18:20	
Air Blank	0.000	18:21	
Subject Sample	# 3	0.171	18:23
Breath Volume	2.289	18:23	
Air Blank	0.000	18:24	
Control Test	0.080	18:24	
Air Blank	0.000	18:24	
Diagnostics Check	OK	18:24	

Data Download Date/Time: 08/25/2016 10:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRBANKS, ANTHONY, M
BREATH TEST OPERATOR AGENCY: Winter Park PD
ARRESTING OFFICER: FAIRBANKS, ANTHONY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 14:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/07/2016 15:11:08

Diagnostics Check	OK	15:04
Air Blank	0.000	15:05
Control Test	0.077	15:05
Air Blank	0.000	15:06
Subject Sample # 1	0.343	15:06
Breath Volume	3.035	15:06
Air Blank	0.000	15:07
Air Blank	0.000	15:09
Subject Sample # 2	0.345	15:09
Breath Volume	3.007	15:09
Air Blank	0.000	15:10
Control Test	0.077	15:10
Air Blank	0.000	15:11
Diagnostics Check	OK	15:11

Data Download Date/Time: 07/19/2016 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000994

INSTRUMENT REGISTERED TO: WINTER PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICARLO,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Winter Park PD
ARRESTING OFFICER: DICARLO,JOSEPH
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/07/2016 22:46:55

Diagnostics Check	OK	22:44	
Air Blank	0.000	22:44	
Control Test	0.078	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:46	
Control Test	0.078	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 07/19/2016 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 06/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MOREY

DATE/TIME OF TEST: 07/17/2016 03:17:51

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:07	
Control Test		0.079	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1	0.152	03:11	Volume Not Met
Breath Volume		0.144	03:11	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2		03:17	Sequence Aborted
Breath Volume		0.000	03:17	
Air Blank		0.000	03:17	

Data Download Date/Time: 07/22/2016 09:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,OSCAR,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ARMENT

DATE/TIME OF TEST: 07/04/2016 02:29:43

Diagnostics Check	OK	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:25	
Subject Sample	# 1 0.177	02:28	
Breath Volume	1.636	02:28	
Air Blank		02:28	Purge Fail
Air Blank		02:29	Purge Fail

Data Download Date/Time: 07/28/2016 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,OSCAR,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ARMENT

DATE/TIME OF TEST: 07/04/2016 02:41:44

Diagnostics Check	OK	02:37	
Air Blank	0.000	02:37	
Control Test	0.078	02:37	
Air Blank	0.000	02:38	
Subject Sample	# 1 0.171	02:40	
Breath Volume	1.640	02:40	
Air Blank		02:41	Purge Fail
Air Blank	0.000	02:41	

Data Download Date/Time: 07/28/2016 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,OSCAR,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: ARMENT

DATE/TIME OF TEST: 07/04/2016 02:58:15

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.078	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.159	02:53
Breath Volume	2.273	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.171	02:56
Breath Volume	1.714	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 07/28/2016 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,OSCAR,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: FATU

DATE/TIME OF TEST: 07/07/2016 01:20:51

Diagnostics Check	OK	01:14
Air Blank	0.000	01:14
Control Test	0.079	01:15
Air Blank	0.000	01:15
Subject Sample # 1	0.165	01:16
Breath Volume	3.152	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.161	01:19
Breath Volume	3.199	01:19
Air Blank	0.000	01:19
Control Test	0.080	01:20
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 07/28/2016 12:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: GLEASON,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FAIRCHILD

DATE/TIME OF TEST: 07/23/2016 01:57:52

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:52
Air Blank	0.000	01:52
Subject Sample # 1	0.110	01:53
Breath Volume	1.855	01:53
Air Blank	0.000	01:54
Air Blank	0.000	01:55
Subject Sample # 2	0.106	01:56
Breath Volume	1.742	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 07/28/2016 12:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,OSCAR,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LUGO,OSCAR
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 07/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: PRATT

DATE/TIME OF TEST: 07/31/2016 03:20:22

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.135	03:15
Breath Volume	2.554	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.131	03:18
Breath Volume	2.250	03:18
Air Blank	0.000	03:19
Control Test	0.078	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 08/28/2016 10:00