

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HOPP,SETH  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: LOPEZ GUERRA

DATE/TIME OF TEST: 09/24/2016 03:30:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.081 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 1 | 0.166 | 03:25 |
| Breath Volume      | 2.773 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Air Blank          | 0.000 | 03:28 |
| Subject Sample # 2 | 0.165 | 03:28 |
| Breath Volume      | 1.996 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Control Test       | 0.080 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Diagnostics Check  | OK    | 03:30 |

---

Data Download Date/Time: 10/24/2016 07:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LONES,SHAUN,G  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: LONES,SHAUN  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:11

SUBJECT'S LAST NAME: HASKINS

DATE/TIME OF TEST: 09/30/2016 19:40:03

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:31 |                      |
| Air Blank         | 0.000 | 19:32 |                      |
| Control Test      | 0.079 | 19:32 |                      |
| Air Blank         | 0.000 | 19:32 |                      |
| Subject Sample    | # 1   | 19:36 | No Sample Provided   |
| Breath Volume     | 0.000 | 19:36 |                      |
| Air Blank         | 0.000 | 19:36 |                      |
| Air Blank         | 0.000 | 19:38 |                      |
| Subject Sample    | # 2   | 19:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:38 |                      |
| Air Blank         | 0.000 | 19:39 |                      |
| Control Test      | 0.081 | 19:39 |                      |
| Air Blank         | 0.000 | 19:39 |                      |
| Diagnostics Check | OK    | 19:40 |                      |

---

Data Download Date/Time: 10/24/2016 07:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: BRAGG,GARRETT  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 08/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: KELLER

DATE/TIME OF TEST: 09/01/2016 04:21:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:12 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.083 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 1 | 0.163 | 04:14 |
| Breath Volume      | 1.152 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Air Blank          | 0.000 | 04:17 |
| Subject Sample # 2 | 0.165 | 04:19 |
| Breath Volume      | 1.636 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Control Test       | 0.084 | 04:20 |
| Air Blank          | 0.000 | 04:20 |
| Diagnostics Check  | OK    | 04:21 |

---

Data Download Date/Time: 09/15/2016 01:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HANSEN,DYLLON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 08/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: SANCHEZ RIVERA

DATE/TIME OF TEST: 09/02/2016 04:41:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:38 |                      |
| Air Blank         | 0.000 | 04:38 |                      |
| Control Test      | 0.081 | 04:38 |                      |
| Air Blank         | 0.000 | 04:39 |                      |
| Subject Sample    | # 1   | 04:39 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:39 |                      |
| Air Blank         | 0.000 | 04:40 |                      |
| Control Test      | 0.081 | 04:40 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Diagnostics Check | OK    | 04:41 |                      |

---

Data Download Date/Time: 09/15/2016 01:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORELL,JONAS,R  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 08/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1737087  
DRY GAS EXPIRY DATE: 09/12/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: METTS

DATE/TIME OF TEST: 09/04/2016 05:14:58

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 05:08 |                |
| Air Blank         | 0.000     | 05:09 |                |
| Control Test      | 0.082     | 05:09 |                |
| Air Blank         | 0.000     | 05:10 |                |
| Subject Sample    | # 1 0.211 | 05:13 | Volume Not Met |
| Breath Volume     | 0.410     | 05:13 |                |
| Air Blank         |           | 05:14 | Purge Fail     |
| Air Blank         |           | 05:14 | Purge Fail     |

---

Data Download Date/Time: 09/15/2016 01:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: OCONNELL,SHANNON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 08/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/09/2016 19:23:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:20 |                      |
| Air Blank         | 0.000 | 19:21 |                      |
| Control Test      | 0.083 | 19:21 |                      |
| Air Blank         | 0.000 | 19:22 |                      |
| Subject Sample    | # 1   | 19:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:22 |                      |
| Air Blank         | 0.000 | 19:22 |                      |
| Control Test      | 0.084 | 19:23 |                      |
| Air Blank         | 0.000 | 19:23 |                      |
| Diagnostics Check | OK    | 19:23 |                      |

---

Data Download Date/Time: 09/15/2016 01:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HOPP,SETH  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 08/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: OLIVO SANCHEZ

DATE/TIME OF TEST: 09/09/2016 21:28:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:25 |                      |
| Air Blank         | 0.000 | 21:25 |                      |
| Control Test      | 0.082 | 21:26 |                      |
| Air Blank         | 0.000 | 21:26 |                      |
| Subject Sample    | # 1   | 21:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:26 |                      |
| Air Blank         | 0.000 | 21:27 |                      |
| Control Test      | 0.083 | 21:27 |                      |
| Air Blank         | 0.000 | 21:28 |                      |
| Diagnostics Check | OK    | 21:28 |                      |

---

Data Download Date/Time: 09/15/2016 01:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HILL,DANIEL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 08/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 09/10/2016 03:32:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.082 | 03:24 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 1 | 0.150 | 03:27 |
| Breath Volume      | 1.691 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.153 | 03:31 |
| Breath Volume      | 1.800 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.082 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 09/15/2016 01:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SIMO,ANNETTE  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 08/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: RYERSON

DATE/TIME OF TEST: 09/12/2016 01:51:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:48 |                      |
| Air Blank         | 0.000 | 01:49 |                      |
| Control Test      | 0.083 | 01:49 |                      |
| Air Blank         | 0.000 | 01:50 |                      |
| Subject Sample    | # 1   | 01:50 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:50 |                      |
| Air Blank         | 0.000 | 01:50 |                      |
| Control Test      | 0.084 | 01:51 |                      |
| Air Blank         | 0.000 | 01:51 |                      |
| Diagnostics Check | OK    | 01:51 |                      |

---

Data Download Date/Time: 09/15/2016 01:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: RODRIGUEZ,ALEX  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 091602  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: ESPOSITO

DATE/TIME OF TEST: 09/16/2016 23:39:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:33 |
| Air Blank          | 0.000 | 23:33 |
| Control Test       | 0.079 | 23:33 |
| Air Blank          | 0.000 | 23:34 |
| Subject Sample # 1 | 0.179 | 23:34 |
| Breath Volume      | 1.222 | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Air Blank          | 0.000 | 23:37 |
| Subject Sample # 2 | 0.179 | 23:37 |
| Breath Volume      | 1.425 | 23:37 |
| Air Blank          | 0.000 | 23:38 |
| Control Test       | 0.078 | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Diagnostics Check  | OK    | 23:39 |

---

Data Download Date/Time: 10/23/2016 08:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'CONNELL,KEVIN,M  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: DAVIS,ALEXANDRA  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: ANGELOTTI

DATE/TIME OF TEST: 09/18/2016 07:23:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:20 |                      |
| Air Blank         | 0.000 | 07:21 |                      |
| Control Test      | 0.080 | 07:21 |                      |
| Air Blank         | 0.000 | 07:22 |                      |
| Subject Sample    | # 1   | 07:22 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:22 |                      |
| Air Blank         | 0.000 | 07:22 |                      |
| Control Test      | 0.080 | 07:23 |                      |
| Air Blank         | 0.000 | 07:23 |                      |
| Diagnostics Check | OK    | 07:23 |                      |

---

Data Download Date/Time: 10/23/2016 08:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPEICKER,TINA,M  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: SPEICKER,TINA  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: BEJNA

DATE/TIME OF TEST: 09/20/2016 23:25:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:19 |
| Air Blank          | 0.000 | 23:19 |
| Control Test       | 0.080 | 23:20 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 1 | 0.145 | 23:21 |
| Breath Volume      | 3.070 | 23:21 |
| Air Blank          | 0.000 | 23:21 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 2 | 0.138 | 23:23 |
| Breath Volume      | 1.750 | 23:23 |
| Air Blank          | 0.000 | 23:24 |
| Control Test       | 0.078 | 23:24 |
| Air Blank          | 0.000 | 23:25 |
| Diagnostics Check  | OK    | 23:25 |

---

Data Download Date/Time: 10/23/2016 08:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: RODRIGUEZ,ALEX  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 091602  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: RARICK

DATE/TIME OF TEST: 09/22/2016 02:04:18

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:01 |                      |
| Air Blank         | 0.000 | 02:01 |                      |
| Control Test      | 0.079 | 02:01 |                      |
| Air Blank         | 0.000 | 02:02 |                      |
| Subject Sample    | # 1   | 02:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:02 |                      |
| Air Blank         | 0.000 | 02:03 |                      |
| Control Test      | 0.079 | 02:03 |                      |
| Air Blank         | 0.000 | 02:04 |                      |
| Diagnostics Check | OK    | 02:04 |                      |

---

Data Download Date/Time: 10/23/2016 08:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGEE,JW,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: MCGEE,J  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: ADMIN  
ARREST TIME: 09:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: SUBJECT

DATE/TIME OF TEST: 09/22/2016 10:24:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:18 |
| Air Blank          | 0.000 | 10:19 |
| Control Test       | 0.080 | 10:19 |
| Air Blank          | 0.000 | 10:20 |
| Subject Sample # 1 | 0.000 | 10:20 |
| Breath Volume      | 2.707 | 10:20 |
| Air Blank          | 0.000 | 10:21 |
| Air Blank          | 0.000 | 10:23 |
| Subject Sample # 2 | 0.000 | 10:23 |
| Breath Volume      | 1.718 | 10:23 |
| Air Blank          | 0.000 | 10:23 |
| Control Test       | 0.078 | 10:24 |
| Air Blank          | 0.000 | 10:24 |
| Diagnostics Check  | OK    | 10:24 |

---

Data Download Date/Time: 10/23/2016 08:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETTEE,MICHAEL,E  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: PETTEE,MICHAEL  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 12:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:14

SUBJECT'S LAST NAME: MAURER

DATE/TIME OF TEST: 09/27/2016 12:53:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 12:47 |
| Air Blank          | 0.000 | 12:47 |
| Control Test       | 0.080 | 12:48 |
| Air Blank          | 0.000 | 12:48 |
| Subject Sample # 1 | 0.250 | 12:49 |
| Breath Volume      | 4.925 | 12:49 |
| Air Blank          | 0.000 | 12:50 |
| Air Blank          | 0.000 | 12:51 |
| Subject Sample # 2 | 0.264 | 12:52 |
| Breath Volume      | 4.812 | 12:52 |
| Air Blank          | 0.000 | 12:53 |
| Control Test       | 0.078 | 12:53 |
| Air Blank          | 0.000 | 12:53 |
| Diagnostics Check  | OK    | 12:53 |

---

Data Download Date/Time: 10/23/2016 08:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PETTEE,MICHAEL,E  
BREATH TEST OPERATOR AGENCY: Key West PD  
ARRESTING OFFICER: PETTEE,MICHAEL  
ARRESTING OFFICER AGENCY: Key West PD  
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: BEHMAN

DATE/TIME OF TEST: 09/30/2016 16:44:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:38 |
| Air Blank          | 0.000 | 16:38 |
| Control Test       | 0.080 | 16:38 |
| Air Blank          | 0.000 | 16:39 |
| Subject Sample # 1 | 0.119 | 16:39 |
| Breath Volume      | 4.398 | 16:39 |
| Air Blank          | 0.000 | 16:40 |
| Air Blank          | 0.000 | 16:42 |
| Subject Sample # 2 | 0.110 | 16:42 |
| Breath Volume      | 2.093 | 16:42 |
| Air Blank          | 0.000 | 16:43 |
| Control Test       | 0.078 | 16:43 |
| Air Blank          | 0.000 | 16:44 |
| Diagnostics Check  | OK    | 16:44 |

---

Data Download Date/Time: 10/23/2016 08:08



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,JESSICA,Y  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: RODRIGUEZ,JESSICA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/03/2016 03:23:03

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:13 |
| Air Blank         |     | 0.000 | 03:13 |
| Control Test      |     | 0.080 | 03:14 |
| Air Blank         |     | 0.000 | 03:14 |
| Subject Sample    | # 1 | 0.123 | 03:17 |
| Breath Volume     |     | 1.222 | 03:17 |
| Air Blank         |     | 0.000 | 03:17 |
| Air Blank         |     | 0.000 | 03:19 |
| Subject Sample    | # 2 | 0.117 | 03:21 |
| Breath Volume     |     | 1.277 | 03:21 |
| Air Blank         |     | 0.000 | 03:22 |
| Control Test      |     | 0.079 | 03:22 |
| Air Blank         |     | 0.000 | 03:22 |
| Diagnostics Check |     | OK    | 03:23 |

---

Data Download Date/Time: 09/23/2016 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMANN,ANDRE,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TAHA,KHALED  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BOW JR

DATE/TIME OF TEST: 09/08/2016 02:48:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.080 | 02:41 |
| Air Blank          | 0.000 | 02:42 |
| Subject Sample # 1 | 0.214 | 02:43 |
| Breath Volume      | 2.031 | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 2 | 0.207 | 02:47 |
| Breath Volume      | 1.777 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Control Test       | 0.079 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Diagnostics Check  | OK    | 02:48 |

---

Data Download Date/Time: 09/23/2016 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMANN,ANDRE,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: IZQUIERDO,LEANDRO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: APARICIO

DATE/TIME OF TEST: 09/09/2016 04:50:29

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 04:37 |                |
| Air Blank         |     | 0.000 | 04:38 |                |
| Control Test      |     | 0.079 | 04:38 |                |
| Air Blank         |     | 0.000 | 04:38 |                |
| Subject Sample    | # 1 | 0.165 | 04:42 | Volume Not Met |
| Breath Volume     |     | 0.285 | 04:42 |                |
| Air Blank         |     | 0.000 | 04:42 |                |
| Air Blank         |     | 0.000 | 04:44 |                |
| Subject Sample    | # 2 | 0.201 | 04:46 |                |
| Breath Volume     |     | 1.445 | 04:46 |                |
| Air Blank         |     | 0.000 | 04:46 |                |
| Air Blank         |     | 0.000 | 04:48 |                |
| Subject Sample    | # 3 | 0.196 | 04:49 |                |
| Breath Volume     |     | 1.460 | 04:49 |                |
| Air Blank         |     | 0.000 | 04:49 |                |
| Control Test      |     | 0.079 | 04:49 |                |
| Air Blank         |     | 0.000 | 04:50 |                |
| Diagnostics Check |     | OK    | 04:50 |                |

---

Data Download Date/Time: 09/23/2016 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAHA,KHALED,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: BECKER,ALEXA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 09/11/2016 02:39:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:30 |                      |
| Air Blank         | 0.000 | 02:31 |                      |
| Control Test      | 0.079 | 02:31 |                      |
| Air Blank         | 0.000 | 02:31 |                      |
| Subject Sample    | # 1   | 02:35 | No Sample Provided   |
| Breath Volume     | 0.000 | 02:35 |                      |
| Air Blank         | 0.000 | 02:35 |                      |
| Air Blank         | 0.000 | 02:37 |                      |
| Subject Sample    | # 2   | 02:38 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:38 |                      |
| Air Blank         | 0.000 | 02:38 |                      |
| Control Test      | 0.079 | 02:39 |                      |
| Air Blank         | 0.000 | 02:39 |                      |
| Diagnostics Check | OK    | 02:39 |                      |

---

Data Download Date/Time: 09/23/2016 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAHA,KHALED,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: BECKER,ALEXA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 09/11/2016 02:47:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:44 |                      |
| Air Blank         | 0.000 | 02:44 |                      |
| Control Test      | 0.080 | 02:44 |                      |
| Air Blank         | 0.000 | 02:45 |                      |
| Subject Sample    | # 1   | 02:45 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:45 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Control Test      | 0.080 | 02:46 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Diagnostics Check | OK    | 02:47 |                      |

---

Data Download Date/Time: 09/23/2016 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA QUINONES,IGNACIO,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: RIVERA QUINONES,IGNACIO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PAREDES

DATE/TIME OF TEST: 09/17/2016 02:14:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:11 |                      |
| Air Blank         | 0.000 | 02:11 |                      |
| Control Test      | 0.079 | 02:12 |                      |
| Air Blank         | 0.000 | 02:12 |                      |
| Subject Sample    | # 1   | 02:13 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:13 |                      |
| Air Blank         | 0.000 | 02:13 |                      |
| Control Test      | 0.079 | 02:13 |                      |
| Air Blank         | 0.000 | 02:14 |                      |
| Diagnostics Check | OK    | 02:14 |                      |

---

Data Download Date/Time: 09/23/2016 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMANN,ANDRE,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: GRIGG,DAVID  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SANCHEZ HALL

DATE/TIME OF TEST: 09/19/2016 04:35:08

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:30 |                      |
| Air Blank         | 0.000 | 04:31 |                      |
| Control Test      | 0.080 | 04:31 |                      |
| Air Blank         | 0.000 | 04:32 |                      |
| Subject Sample    | # 1   | 04:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:33 |                      |
| Air Blank         | 0.000 | 04:34 |                      |
| Control Test      | 0.080 | 04:34 |                      |
| Air Blank         | 0.000 | 04:35 |                      |
| Diagnostics Check | OK    | 04:35 |                      |

---

Data Download Date/Time: 09/23/2016 12:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAHA,KHALED,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: SOBALVARRO,PAOLA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: COLAVINI

DATE/TIME OF TEST: 09/20/2016 03:13:49

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 03:02 |                |
| Air Blank         |     | 0.000 | 03:02 |                |
| Control Test      |     | 0.081 | 03:02 |                |
| Air Blank         |     | 0.000 | 03:03 |                |
| Subject Sample    | # 1 | 0.165 | 03:06 | Volume Not Met |
| Breath Volume     |     | 0.097 | 03:06 |                |
| Air Blank         |     | 0.000 | 03:07 |                |
| Air Blank         |     | 0.000 | 03:09 |                |
| Subject Sample    | # 2 | 0.148 | 03:12 | Volume Not Met |
| Breath Volume     |     | 0.035 | 03:12 |                |
| Air Blank         |     | 0.000 | 03:12 |                |
| Control Test      |     | 0.079 | 03:13 |                |
| Air Blank         |     | 0.000 | 03:13 |                |
| Diagnostics Check |     | OK    | 03:13 |                |

---

Data Download Date/Time: 09/23/2016 12:53



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAHA,KHALED,P  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: SOBALVARRO,PAOLA  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: COLAVINI

DATE/TIME OF TEST: 09/20/2016 03:23:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.079 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 1 | 0.155 | 03:19 |
| Breath Volume      | 1.265 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 2 | 0.159 | 03:22 |
| Breath Volume      | 1.687 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.078 | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Diagnostics Check  | OK    | 03:23 |

---

Data Download Date/Time: 09/23/2016 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAHA,KHALED,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: RIVERA,LUIS  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:14

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 09/21/2016 05:37:22

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:34 |                      |
| Air Blank         | 0.000 | 05:35 |                      |
| Control Test      | 0.080 | 05:35 |                      |
| Air Blank         | 0.000 | 05:35 |                      |
| Subject Sample    | # 1   | 05:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:36 |                      |
| Air Blank         | 0.000 | 05:36 |                      |
| Control Test      | 0.081 | 05:36 |                      |
| Air Blank         | 0.000 | 05:37 |                      |
| Diagnostics Check | OK    | 05:37 |                      |

---

Data Download Date/Time: 09/23/2016 12:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TANIS,DAREU  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 09/24/2016 01:40:49

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 01:23 |                |
| Air Blank         |     | 0.000 | 01:24 |                |
| Control Test      |     | 0.079 | 01:24 |                |
| Air Blank         |     | 0.000 | 01:25 |                |
| Subject Sample    | # 1 | 0.193 | 01:28 | Volume Not Met |
| Breath Volume     |     | 0.039 | 01:28 |                |
| Air Blank         |     | 0.000 | 01:29 |                |
| Air Blank         |     | 0.000 | 01:30 |                |
| Subject Sample    | # 2 | 0.216 | 01:33 |                |
| Breath Volume     |     | 1.371 | 01:33 |                |
| Air Blank         |     | 0.000 | 01:34 |                |
| Air Blank         |     | 0.000 | 01:35 |                |
| Subject Sample    | # 3 | 0.224 | 01:39 | Volume Not Met |
| Breath Volume     |     | 0.085 | 01:39 |                |
| Air Blank         |     | 0.000 | 01:39 |                |
| Control Test      |     | 0.079 | 01:40 |                |
| Air Blank         |     | 0.000 | 01:40 |                |
| Diagnostics Check |     | OK    | 01:40 |                |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TANIS,DAREU  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 09/24/2016 02:02:29

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 01:56 |                |
| Air Blank         | 0.000     | 01:56 |                |
| Control Test      | 0.078     | 01:57 |                |
| Air Blank         | 0.000     | 01:57 |                |
| Subject Sample    | # 1 0.192 | 02:00 | Volume Not Met |
| Breath Volume     | 0.113     | 02:00 |                |
| Air Blank         |           | 02:01 | Purge Fail     |
| Air Blank         |           | 02:02 | Purge Fail     |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TANIS,DAREU  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 09/24/2016 02:17:49

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 02:16 |                           |
| Air Blank         | 0.000 | 02:16 |                           |
| Control Test      | 0.066 | 02:17 | Control Outside Tolerance |
| Air Blank         | 0.000 | 02:17 |                           |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TANIS,DAREU  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 09/24/2016 02:21:41

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 02:20 |                           |
| Air Blank         | 0.000 | 02:20 |                           |
| Control Test      | 0.073 | 02:21 | Control Outside Tolerance |
| Air Blank         | 0.000 | 02:21 |                           |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TANIS,DAREU  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 09/24/2016 02:25:58

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 02:24 |                           |
| Air Blank         | 0.000 | 02:25 |                           |
| Control Test      | 0.073 | 02:25 | Control Outside Tolerance |
| Air Blank         | 0.000 | 02:25 |                           |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TANIS,DAREU  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 09/24/2016 02:30:45

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 02:29 |                           |
| Air Blank         | 0.000 | 02:29 |                           |
| Control Test      | 0.071 | 02:30 | Control Outside Tolerance |
| Air Blank         | 0.000 | 02:30 |                           |

---

Data Download Date/Time: 10/28/2016 15:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUJAN,ARTURO,D  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: TANIS,DAREU  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ECHEVARRIA

DATE/TIME OF TEST: 09/24/2016 02:36:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:34 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Control Test      | 0.075 | 02:34 |                      |
| Air Blank         | 0.000 | 02:35 |                      |
| Subject Sample    | # 1   | 02:35 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:35 |                      |
| Air Blank         | 0.000 | 02:36 |                      |
| Control Test      | 0.076 | 02:36 |                      |
| Air Blank         | 0.000 | 02:36 |                      |
| Diagnostics Check | OK    | 02:36 |                      |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAHA,KHALED,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: NAYA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: CARBAJAL PEREZ

DATE/TIME OF TEST: 09/26/2016 04:21:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Control Test       | 0.081 | 04:15 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 1 | 0.141 | 04:16 |
| Breath Volume      | 1.734 | 04:16 |
| Air Blank          | 0.000 | 04:17 |
| Air Blank          | 0.000 | 04:19 |
| Subject Sample # 2 | 0.145 | 04:19 |
| Breath Volume      | 2.359 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Control Test       | 0.080 | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Diagnostics Check  | OK    | 04:21 |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEREZ,CRISTINO,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: PEREZ,CRISTINO  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 15:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 09/26/2016 16:43:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:36 |
| Air Blank          | 0.000 | 16:37 |
| Control Test       | 0.079 | 16:37 |
| Air Blank          | 0.000 | 16:38 |
| Subject Sample # 1 | 0.234 | 16:38 |
| Breath Volume      | 1.960 | 16:38 |
| Air Blank          | 0.000 | 16:39 |
| Air Blank          | 0.000 | 16:41 |
| Subject Sample # 2 | 0.234 | 16:41 |
| Breath Volume      | 1.777 | 16:41 |
| Air Blank          | 0.000 | 16:42 |
| Control Test       | 0.079 | 16:42 |
| Air Blank          | 0.000 | 16:43 |
| Diagnostics Check  | OK    | 16:43 |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMANN,ANDRE,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: HAMMANN,ANDRE  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/27/2016 02:40:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.080 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Subject Sample # 1 | 0.163 | 02:36 |
| Breath Volume      | 1.640 | 02:36 |
| Air Blank          | 0.000 | 02:36 |
| Air Blank          | 0.000 | 02:38 |
| Subject Sample # 2 | 0.159 | 02:39 |
| Breath Volume      | 1.496 | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.080 | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TAHA,KHALED,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: NAYA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PHILIDOR

DATE/TIME OF TEST: 09/28/2016 02:41:34

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:35 |
| Air Blank         |     | 0.000 | 02:35 |
| Control Test      |     | 0.079 | 02:36 |
| Air Blank         |     | 0.000 | 02:36 |
| Subject Sample    | # 1 | 0.118 | 02:37 |
| Breath Volume     |     | 2.265 | 02:37 |
| Air Blank         |     | 0.000 | 02:37 |
| Air Blank         |     | 0.000 | 02:39 |
| Subject Sample    | # 2 | 0.121 | 02:40 |
| Breath Volume     |     | 2.628 | 02:40 |
| Air Blank         |     | 0.000 | 02:40 |
| Control Test      |     | 0.080 | 02:40 |
| Air Blank         |     | 0.000 | 02:41 |
| Diagnostics Check |     | OK    | 02:41 |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMANN,ANDRE,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: HAMMANN,ANDRE  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/30/2016 02:12:41

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 02:01 |                |
| Air Blank         |     | 0.000 | 02:01 |                |
| Control Test      |     | 0.079 | 02:01 |                |
| Air Blank         |     | 0.000 | 02:02 |                |
| Subject Sample    | # 1 | 0.213 | 02:05 | Volume Not Met |
| Breath Volume     |     | 0.945 | 02:05 |                |
| Air Blank         |     | 0.000 | 02:06 |                |
| Air Blank         |     | 0.000 | 02:08 |                |
| Subject Sample    | # 2 | 0.188 | 02:11 | Volume Not Met |
| Breath Volume     |     | 0.511 | 02:11 |                |
| Air Blank         |     | 0.000 | 02:11 |                |
| Control Test      |     | 0.080 | 02:12 |                |
| Air Blank         |     | 0.000 | 02:12 |                |
| Diagnostics Check |     | OK    | 02:12 |                |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMANN,ANDRE,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: HAMMANN,ANDRE  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/30/2016 03:18:19

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:15 |                      |
| Air Blank         | 0.000 | 03:15 |                      |
| Control Test      | 0.079 | 03:16 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Subject Sample    | # 1   | 03:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:16 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Control Test      | 0.079 | 03:17 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Diagnostics Check | OK    | 03:18 |                      |

---

Data Download Date/Time: 10/28/2016 15:30

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMMANN,ANDRE,  
BREATH TEST OPERATOR AGENCY: Coral Gables PD  
ARRESTING OFFICER: NAYA,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Coral Gables PD  
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG515306  
DRY GAS EXPIRY DATE: 06/02/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GUTIERREZ CONTRERAS

DATE/TIME OF TEST: 09/30/2016 03:32:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:27 |                      |
| Air Blank         | 0.000 | 03:27 |                      |
| Control Test      | 0.079 | 03:27 |                      |
| Air Blank         | 0.000 | 03:28 |                      |
| Subject Sample    | # 1   | 03:31 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:31 |                      |
| Air Blank         | 0.000 | 03:31 |                      |
| Control Test      | 0.079 | 03:32 |                      |
| Air Blank         | 0.000 | 03:32 |                      |
| Diagnostics Check | OK    | 03:32 |                      |

---

Data Download Date/Time: 10/28/2016 15:30



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYNAUD,GEORGE,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: DESPIAN,MIGUEL  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 08/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1789170  
DRY GAS EXPIRY DATE: 01/14/2018  
VIOLATION CODE: BUI  
ARREST TIME: 20:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 09/03/2016 21:32:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:28 |                      |
| Air Blank         | 0.000 | 21:28 |                      |
| Control Test      | 0.078 | 21:29 |                      |
| Air Blank         | 0.000 | 21:29 |                      |
| Subject Sample    | # 1   | 21:30 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:30 |                      |
| Air Blank         | 0.000 | 21:31 |                      |
| Control Test      | 0.078 | 21:31 |                      |
| Air Blank         | 0.000 | 21:32 |                      |
| Diagnostics Check | OK    | 21:32 |                      |

---

Data Download Date/Time: 09/22/2016 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO,GREGORY,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1789170  
DRY GAS EXPIRY DATE: 01/14/2018  
VIOLATION CODE: BUI  
ARREST TIME: 17:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: PLAZA

DATE/TIME OF TEST: 09/04/2016 18:22:37

---

|                   |       |       |                    |                |
|-------------------|-------|-------|--------------------|----------------|
| Diagnostics Check | OK    | 18:11 |                    |                |
| Air Blank         | 0.000 | 18:11 |                    |                |
| Control Test      | 0.077 | 18:11 |                    |                |
| Air Blank         | 0.000 | 18:12 |                    |                |
| Subject Sample    | # 1   | 18:15 | No Sample Provided |                |
| Breath Volume     | 0.000 | 18:15 |                    |                |
| Air Blank         | 0.000 | 18:15 |                    |                |
| Air Blank         | 0.000 | 18:17 |                    |                |
| Subject Sample    | # 2   | 0.154 | 18:21              | Volume Not Met |
| Breath Volume     | 0.914 | 18:21 |                    |                |
| Air Blank         | 0.000 | 18:21 |                    |                |
| Control Test      | 0.077 | 18:22 |                    |                |
| Air Blank         | 0.000 | 18:22 |                    |                |
| Diagnostics Check | OK    | 18:22 |                    |                |

---

Data Download Date/Time: 09/22/2016 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANCO,GREGORY,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1789170  
DRY GAS EXPIRY DATE: 01/14/2018  
VIOLATION CODE: BUI  
ARREST TIME: 17:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: PLAZA

DATE/TIME OF TEST: 09/04/2016 18:32:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:24 |
| Air Blank          | 0.000 | 18:25 |
| Control Test       | 0.077 | 18:25 |
| Air Blank          | 0.000 | 18:26 |
| Subject Sample # 1 | 0.151 | 18:27 |
| Breath Volume      | 2.066 | 18:27 |
| Air Blank          | 0.000 | 18:28 |
| Air Blank          | 0.000 | 18:30 |
| Subject Sample # 2 | 0.167 | 18:31 |
| Breath Volume      | 1.269 | 18:31 |
| Air Blank          | 0.000 | 18:31 |
| Control Test       | 0.076 | 18:31 |
| Air Blank          | 0.000 | 18:32 |
| Diagnostics Check  | OK    | 18:32 |

---

Data Download Date/Time: 09/22/2016 11:45

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,CHRISTINA,A  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: MARTIN,CHRISTINA  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 08/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1789170  
DRY GAS EXPIRY DATE: 01/14/2018  
VIOLATION CODE: BUI  
ARREST TIME: 19:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: FORTUNY

DATE/TIME OF TEST: 09/04/2016 20:08:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:02 |
| Air Blank          | 0.000 | 20:02 |
| Control Test       | 0.078 | 20:02 |
| Air Blank          | 0.000 | 20:03 |
| Subject Sample # 1 | 0.128 | 20:04 |
| Breath Volume      | 3.664 | 20:04 |
| Air Blank          | 0.000 | 20:04 |
| Air Blank          | 0.000 | 20:06 |
| Subject Sample # 2 | 0.128 | 20:07 |
| Breath Volume      | 3.796 | 20:07 |
| Air Blank          | 0.000 | 20:07 |
| Control Test       | 0.078 | 20:08 |
| Air Blank          | 0.000 | 20:08 |
| Diagnostics Check  | OK    | 20:08 |

---

Data Download Date/Time: 09/22/2016 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOAGLAND,JAMES,J  
BREATH TEST OPERATOR AGENCY: Blountstown PD  
ARRESTING OFFICER: KELLER,WILLIAM  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 08/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 09/12/2016 22:36:41

---

|                   |       |       |                             |
|-------------------|-------|-------|-----------------------------|
| Diagnostics Check | Fail  | 22:36 | Temperature Regulation Test |
| Air Blank         | 0.000 | 22:36 |                             |

---

Data Download Date/Time: 09/26/2016 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOAGLAND,JAMES,J  
BREATH TEST OPERATOR AGENCY: Blountstown PD  
ARRESTING OFFICER: KELLER,WILLIAM  
ARRESTING OFFICER AGENCY: Calhoun County SO  
LAST AGENCY INSPECTION: 08/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 09/12/2016 22:46:59

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 22:39 |                           |
| Air Blank         | 0.000     | 22:39 |                           |
| Control Test      | 0.077     | 22:40 |                           |
| Air Blank         | 0.000     | 22:40 |                           |
| Subject Sample    | # 1 0.153 | 22:42 |                           |
| Breath Volume     | 1.683     | 22:42 |                           |
| Air Blank         | 0.000     | 22:43 |                           |
| Air Blank         | 0.000     | 22:45 |                           |
| Subject Sample    | # 2 0.156 | 22:45 |                           |
| Breath Volume     | 1.277     | 22:45 |                           |
| Air Blank         | 0.000     | 22:46 |                           |
| Control Test      | 0.074     | 22:46 | Control Outside Tolerance |
| Air Blank         | 0.000     | 22:46 |                           |

---

Data Download Date/Time: 09/26/2016 10:48

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VENNELL,JOHN,B  
BREATH TEST OPERATOR AGENCY: FHP Troop A  
ARRESTING OFFICER: VENNELL,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop A  
LAST AGENCY INSPECTION: 08/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A3  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 09/14/2016 01:34:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.077 | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Subject Sample # 1 | 0.090 | 01:29 |
| Breath Volume      | 2.433 | 01:29 |
| Air Blank          | 0.000 | 01:30 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 2 | 0.078 | 01:33 |
| Breath Volume      | 1.343 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Control Test       | 0.077 | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Diagnostics Check  | OK    | 01:34 |

---

Data Download Date/Time: 09/26/2016 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTERO,JOSE,  
BREATH TEST OPERATOR AGENCY: Hialeah PD  
ARRESTING OFFICER: ABREU,A  
ARRESTING OFFICER AGENCY: Hialeah PD  
LAST AGENCY INSPECTION: 08/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MONDEJA GALINDO

DATE/TIME OF TEST: 09/03/2016 22:18:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:08 |
| Air Blank          | 0.000 | 22:09 |
| Control Test       | 0.076 | 22:09 |
| Air Blank          | 0.000 | 22:10 |
| Subject Sample # 1 | 0.158 | 22:10 |
| Breath Volume      | 1.406 | 22:10 |
| Air Blank          | 0.000 | 22:11 |
| Air Blank          | 0.000 | 22:13 |
| Subject Sample # 2 | 0.184 | 22:14 |
| Breath Volume      | 1.691 | 22:14 |
| Air Blank          | 0.000 | 22:14 |
| Air Blank          | 0.000 | 22:16 |
| Subject Sample # 3 | 0.183 | 22:17 |
| Breath Volume      | 2.230 | 22:17 |
| Air Blank          | 0.000 | 22:17 |
| Control Test       | 0.076 | 22:18 |
| Air Blank          | 0.000 | 22:18 |
| Diagnostics Check  | OK    | 22:18 |

---

Data Download Date/Time: 09/21/2016 11:45



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTERO,JOSE,  
BREATH TEST OPERATOR AGENCY: Hialeah PD  
ARRESTING OFFICER: MUI,K  
ARRESTING OFFICER AGENCY: Hialeah PD  
LAST AGENCY INSPECTION: 08/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 06415080A1  
DRY GAS EXPIRY DATE: 04/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: VERDECIA

DATE/TIME OF TEST: 09/03/2016 23:26:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:23 |                      |
| Air Blank         | 0.000 | 23:24 |                      |
| Control Test      | 0.077 | 23:24 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Subject Sample    | # 1   | 23:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:25 |                      |
| Air Blank         | 0.000 | 23:25 |                      |
| Control Test      | 0.077 | 23:26 |                      |
| Air Blank         | 0.000 | 23:26 |                      |
| Diagnostics Check | OK    | 23:26 |                      |

---

Data Download Date/Time: 09/21/2016 11:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLAN,KEVIN,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SOCARRAS,PEDRO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: TIMOTHEE

DATE/TIME OF TEST: 09/03/2016 04:42:37

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:39 |                      |
| Air Blank         | 0.000 | 04:40 |                      |
| Control Test      | 0.082 | 04:40 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Subject Sample    | # 1   | 04:41 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:41 |                      |
| Air Blank         | 0.000 | 04:41 |                      |
| Control Test      | 0.081 | 04:42 |                      |
| Air Blank         | 0.000 | 04:42 |                      |
| Diagnostics Check | OK    | 04:42 |                      |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUMPHRY,KERRY,R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: DOMINGUEZ,ERIC  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/03/2016 23:12:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 23:09 |                      |
| Air Blank         | 0.000 | 23:09 |                      |
| Control Test      | 0.081 | 23:10 |                      |
| Air Blank         | 0.000 | 23:10 |                      |
| Subject Sample    | # 1   | 23:10 | Subject Test Refused |
| Breath Volume     | 0.000 | 23:10 |                      |
| Air Blank         | 0.000 | 23:11 |                      |
| Control Test      | 0.081 | 23:11 |                      |
| Air Blank         | 0.000 | 23:12 |                      |
| Diagnostics Check | OK    | 23:12 |                      |

---

Data Download Date/Time: 09/26/2016 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: FERNANDEZ,JAVIER  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: LETONA-OLIVEROS

DATE/TIME OF TEST: 09/10/2016 01:02:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:55 |
| Air Blank          | 0.000 | 00:56 |
| Control Test       | 0.081 | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Subject Sample # 1 | 0.163 | 00:57 |
| Breath Volume      | 1.414 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Air Blank          | 0.000 | 01:00 |
| Subject Sample # 2 | 0.161 | 01:00 |
| Breath Volume      | 2.101 | 01:00 |
| Air Blank          | 0.000 | 01:01 |
| Control Test       | 0.081 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Diagnostics Check  | OK    | 01:02 |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MILIAN,KEVIN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: JUNCO

DATE/TIME OF TEST: 09/10/2016 01:11:27

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 01:10 |            |
| Air Blank         |    | 01:11 | RFI Detect |
| Air Blank         |    | 01:11 | RFI Detect |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MILIAN,KEVIN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: JUNCO

DATE/TIME OF TEST: 09/10/2016 01:22:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:14 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.082 | 01:15 |
| Air Blank          | 0.000 | 01:15 |
| Subject Sample # 1 | 0.000 | 01:18 |
| Breath Volume      | 1.140 | 01:18 |
| Air Blank          | 0.000 | 01:18 |
| Air Blank          | 0.000 | 01:20 |
| Subject Sample # 2 | 0.000 | 01:21 |
| Breath Volume      | 1.250 | 01:21 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.081 | 01:21 |
| Air Blank          | 0.000 | 01:22 |
| Diagnostics Check  | OK    | 01:22 |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MILLAN,KEVIN  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/10/2016 04:27:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.081 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Subject Sample # 1 | 0.176 | 04:22 |
| Breath Volume      | 1.589 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Subject Sample # 2 | 0.165 | 04:26 |
| Breath Volume      | 1.488 | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.080 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Diagnostics Check  | OK    | 04:27 |

---

Data Download Date/Time: 09/26/2016 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE,STEVE,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: COSNER,STEVE  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: LOAZA LOAZA

DATE/TIME OF TEST: 09/15/2016 04:41:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:32 |
| Air Blank          | 0.000 | 04:33 |
| Control Test       | 0.081 | 04:33 |
| Air Blank          | 0.000 | 04:33 |
| Subject Sample # 1 | 0.129 | 04:35 |
| Breath Volume      | 2.535 | 04:35 |
| Air Blank          | 0.000 | 04:35 |
| Air Blank          | 0.000 | 04:37 |
| Subject Sample # 2 | 0.131 | 04:39 |
| Breath Volume      | 1.683 | 04:39 |
| Air Blank          | 0.000 | 04:40 |
| Control Test       | 0.080 | 04:40 |
| Air Blank          | 0.000 | 04:41 |
| Diagnostics Check  | OK    | 04:41 |

---

Data Download Date/Time: 09/26/2016 11:47



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONFESSORE,STEVE,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: MAROTTA,DANIELLE  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: OLIVAS

DATE/TIME OF TEST: 09/16/2016 03:26:51

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:23 |                      |
| Air Blank         | 0.000 | 03:24 |                      |
| Control Test      | 0.081 | 03:24 |                      |
| Air Blank         | 0.000 | 03:25 |                      |
| Subject Sample    | # 1   | 03:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:25 |                      |
| Air Blank         | 0.000 | 03:25 |                      |
| Control Test      | 0.081 | 03:26 |                      |
| Air Blank         | 0.000 | 03:26 |                      |
| Diagnostics Check | OK    | 03:26 |                      |

---

Data Download Date/Time: 09/26/2016 11:47

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: FERNANDEZ,JAVIER  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/16/2016 22:38:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:32 |
| Air Blank          | 0.000 | 22:33 |
| Control Test       | 0.082 | 22:33 |
| Air Blank          | 0.000 | 22:33 |
| Subject Sample # 1 | 0.000 | 22:34 |
| Breath Volume      | 1.832 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Air Blank          | 0.000 | 22:37 |
| Subject Sample # 2 | 0.000 | 22:37 |
| Breath Volume      | 2.199 | 22:37 |
| Air Blank          | 0.000 | 22:37 |
| Control Test       | 0.081 | 22:38 |
| Air Blank          | 0.000 | 22:38 |
| Diagnostics Check  | OK    | 22:38 |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLAN,KEVIN,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SOCARRAS,PEDRO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: KERWIN

DATE/TIME OF TEST: 09/17/2016 03:20:24

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:17 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Control Test      | 0.081 | 03:18 |                      |
| Air Blank         | 0.000 | 03:18 |                      |
| Subject Sample    | # 1   | 03:19 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:19 |                      |
| Air Blank         | 0.000 | 03:19 |                      |
| Control Test      | 0.082 | 03:19 |                      |
| Air Blank         | 0.000 | 03:20 |                      |
| Diagnostics Check | OK    | 03:20 |                      |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: RODRIGUEZ,RICHARD  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: ROCA

DATE/TIME OF TEST: 09/17/2016 03:38:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:35 |                      |
| Air Blank         | 0.000 | 03:35 |                      |
| Control Test      | 0.081 | 03:35 |                      |
| Air Blank         | 0.000 | 03:36 |                      |
| Subject Sample    | # 1   | 03:36 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:36 |                      |
| Air Blank         | 0.000 | 03:37 |                      |
| Control Test      | 0.082 | 03:37 |                      |
| Air Blank         | 0.000 | 03:38 |                      |
| Diagnostics Check | OK    | 03:38 |                      |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: GUASTO,NICOLAS  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 09/17/2016 04:18:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Control Test       | 0.082 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Subject Sample # 1 | 0.028 | 04:13 |
| Breath Volume      | 1.472 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.027 | 04:17 |
| Breath Volume      | 1.859 | 04:17 |
| Air Blank          | 0.000 | 04:18 |
| Control Test       | 0.081 | 04:18 |
| Air Blank          | 0.000 | 04:18 |
| Diagnostics Check  | OK    | 04:18 |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUMPHRY,KERRY,R  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: HUMPHRY,KERRY  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GOLDSCHMIDT

DATE/TIME OF TEST: 09/18/2016 00:45:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.082 | 00:39 |
| Air Blank          | 0.000 | 00:40 |
| Subject Sample # 1 | 0.066 | 00:40 |
| Breath Volume      | 1.281 | 00:40 |
| Air Blank          | 0.000 | 00:41 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 2 | 0.064 | 00:43 |
| Breath Volume      | 1.691 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Control Test       | 0.081 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Diagnostics Check  | OK    | 00:45 |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,RICHARD,  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: SOCARRAS,PEDRO  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 09/18/2016 03:48:31

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:45 |                      |
| Air Blank         | 0.000 | 03:45 |                      |
| Control Test      | 0.081 | 03:46 |                      |
| Air Blank         | 0.000 | 03:46 |                      |
| Subject Sample    | # 1   | 03:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:47 |                      |
| Air Blank         | 0.000 | 03:47 |                      |
| Control Test      | 0.081 | 03:47 |                      |
| Air Blank         | 0.000 | 03:48 |                      |
| Diagnostics Check | OK    | 03:48 |                      |

---

Data Download Date/Time: 09/26/2016 11:47

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,ADRIAN,F  
BREATH TEST OPERATOR AGENCY: Miami Beach PD  
ARRESTING OFFICER: BAUMER,WERNER  
ARRESTING OFFICER AGENCY: Miami Beach PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1826916  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: URBINA MOSCOSO

DATE/TIME OF TEST: 09/25/2016 07:55:45

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 07:52 |                      |
| Air Blank         | 0.000 | 07:53 |                      |
| Control Test      | 0.082 | 07:53 |                      |
| Air Blank         | 0.000 | 07:54 |                      |
| Subject Sample    | # 1   | 07:54 | Subject Test Refused |
| Breath Volume     | 0.000 | 07:54 |                      |
| Air Blank         | 0.000 | 07:54 |                      |
| Control Test      | 0.082 | 07:55 |                      |
| Air Blank         | 0.000 | 07:55 |                      |
| Diagnostics Check | OK    | 07:55 |                      |

---

Data Download Date/Time: 09/26/2016 11:47



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,STEVEN,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ADAMS,STEVEN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 09/07/2016 03:37:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.081 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 1 | 0.164 | 03:32 |
| Breath Volume      | 1.503 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Air Blank          | 0.000 | 03:34 |
| Subject Sample # 2 | 0.163 | 03:36 |
| Breath Volume      | 1.453 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.080 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 09/12/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: MADERA

DATE/TIME OF TEST: 09/10/2016 03:05:53

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:58 |
| Air Blank         |     | 0.000 | 02:58 |
| Control Test      |     | 0.078 | 02:58 |
| Air Blank         |     | 0.000 | 02:59 |
| Subject Sample    | # 1 | 0.202 | 03:00 |
| Breath Volume     |     | 2.566 | 03:00 |
| Air Blank         |     | 0.000 | 03:00 |
| Air Blank         |     | 0.000 | 03:02 |
| Subject Sample    | # 2 | 0.210 | 03:04 |
| Breath Volume     |     | 1.769 | 03:04 |
| Air Blank         |     | 0.000 | 03:04 |
| Control Test      |     | 0.079 | 03:05 |
| Air Blank         |     | 0.000 | 03:05 |
| Diagnostics Check |     | OK    | 03:05 |

---

Data Download Date/Time: 10/13/2016 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 04:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:04

SUBJECT'S LAST NAME: PICHARDO-ARIAS

DATE/TIME OF TEST: 09/10/2016 05:34:55

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:32 |                      |
| Air Blank         | 0.000 | 05:32 |                      |
| Control Test      | 0.078 | 05:32 |                      |
| Air Blank         | 0.000 | 05:33 |                      |
| Subject Sample    | # 1   | 05:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:33 |                      |
| Air Blank         | 0.000 | 05:33 |                      |
| Control Test      | 0.078 | 05:34 |                      |
| Air Blank         | 0.000 | 05:34 |                      |
| Diagnostics Check | OK    | 05:34 |                      |

---

Data Download Date/Time: 10/13/2016 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DE LA FE MARI

DATE/TIME OF TEST: 09/14/2016 04:08:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.079 | 04:00 |
| Air Blank          | 0.000 | 04:00 |
| Subject Sample # 1 | 0.189 | 04:02 |
| Breath Volume      | 1.218 | 04:02 |
| Air Blank          | 0.000 | 04:03 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 2 | 0.185 | 04:06 |
| Breath Volume      | 1.500 | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Control Test       | 0.077 | 04:07 |
| Air Blank          | 0.000 | 04:07 |
| Diagnostics Check  | OK    | 04:08 |

---

Data Download Date/Time: 10/13/2016 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: PINO

DATE/TIME OF TEST: 09/15/2016 02:23:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Control Test       | 0.077 | 02:18 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 1 | 0.177 | 02:19 |
| Breath Volume      | 1.746 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Air Blank          | 0.000 | 02:21 |
| Subject Sample # 2 | 0.179 | 02:21 |
| Breath Volume      | 2.472 | 02:21 |
| Air Blank          | 0.000 | 02:22 |
| Control Test       | 0.077 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Diagnostics Check  | OK    | 02:23 |

---

Data Download Date/Time: 10/13/2016 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 05:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: ANCTIL

DATE/TIME OF TEST: 09/16/2016 06:19:26

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 06:16 |                      |
| Air Blank         | 0.000 | 06:17 |                      |
| Control Test      | 0.079 | 06:17 |                      |
| Air Blank         | 0.000 | 06:17 |                      |
| Subject Sample    | # 1   | 06:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 06:18 |                      |
| Air Blank         | 0.000 | 06:18 |                      |
| Control Test      | 0.078 | 06:18 |                      |
| Air Blank         | 0.000 | 06:19 |                      |
| Diagnostics Check | OK    | 06:19 |                      |

---

Data Download Date/Time: 10/13/2016 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: AYALA-LEAL

DATE/TIME OF TEST: 09/17/2016 00:57:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:55 |                      |
| Air Blank         | 0.000 | 00:55 |                      |
| Control Test      | 0.077 | 00:55 |                      |
| Air Blank         | 0.000 | 00:56 |                      |
| Subject Sample    | # 1   | 00:56 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:56 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Control Test      | 0.077 | 00:57 |                      |
| Air Blank         | 0.000 | 00:57 |                      |
| Diagnostics Check | OK    | 00:57 |                      |

---

Data Download Date/Time: 10/13/2016 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RODOLI

DATE/TIME OF TEST: 09/17/2016 02:01:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:55 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.078 | 01:56 |
| Air Blank          | 0.000 | 01:57 |
| Subject Sample # 1 | 0.150 | 01:57 |
| Breath Volume      | 2.523 | 01:57 |
| Air Blank          | 0.000 | 01:58 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 2 | 0.150 | 02:00 |
| Breath Volume      | 2.562 | 02:00 |
| Air Blank          | 0.000 | 02:00 |
| Control Test       | 0.077 | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Diagnostics Check  | OK    | 02:01 |

---

Data Download Date/Time: 10/13/2016 10:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: NAZCO RODRIGUEZ

DATE/TIME OF TEST: 09/21/2016 01:00:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Control Test       | 0.078 | 00:54 |
| Air Blank          | 0.000 | 00:55 |
| Subject Sample # 1 | 0.131 | 00:55 |
| Breath Volume      | 3.765 | 00:55 |
| Air Blank          | 0.000 | 00:56 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 2 | 0.134 | 00:58 |
| Breath Volume      | 3.757 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Control Test       | 0.078 | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Diagnostics Check  | OK    | 01:00 |

---

Data Download Date/Time: 10/13/2016 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: MANCILLA-PERTUZ

DATE/TIME OF TEST: 09/28/2016 01:22:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Control Test       | 0.079 | 01:16 |
| Air Blank          | 0.000 | 01:17 |
| Subject Sample # 1 | 0.190 | 01:18 |
| Breath Volume      | 3.355 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Subject Sample # 2 | 0.182 | 01:21 |
| Breath Volume      | 3.097 | 01:21 |
| Air Blank          | 0.000 | 01:21 |
| Control Test       | 0.078 | 01:22 |
| Air Blank          | 0.000 | 01:22 |
| Diagnostics Check  | OK    | 01:22 |

---

Data Download Date/Time: 10/13/2016 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 01:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: OGANDO

DATE/TIME OF TEST: 09/29/2016 02:39:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Control Test       | 0.078 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 1 | 0.141 | 02:34 |
| Breath Volume      | 2.199 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Air Blank          | 0.000 | 02:37 |
| Subject Sample # 2 | 0.147 | 02:37 |
| Breath Volume      | 3.234 | 02:37 |
| Air Blank          | 0.000 | 02:38 |
| Control Test       | 0.076 | 02:38 |
| Air Blank          | 0.000 | 02:39 |
| Diagnostics Check  | OK    | 02:39 |

---

Data Download Date/Time: 10/13/2016 10:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19014080A3  
DRY GAS EXPIRY DATE: 08/05/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/30/2016 03:12:01

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:00 |                |
| Air Blank         | 0.000     | 03:00 |                |
| Control Test      | 0.078     | 03:01 |                |
| Air Blank         | 0.000     | 03:01 |                |
| Subject Sample    | # 1 0.193 | 03:02 |                |
| Breath Volume     | 1.511     | 03:02 |                |
| Air Blank         | 0.000     | 03:02 |                |
| Air Blank         | 0.000     | 03:04 |                |
| Subject Sample    | # 2 0.159 | 03:07 | Volume Not Met |
| Breath Volume     | 0.605     | 03:07 |                |
| Air Blank         | 0.000     | 03:08 |                |
| Air Blank         | 0.000     | 03:10 |                |
| Subject Sample    | # 3 0.174 | 03:10 |                |
| Breath Volume     | 1.562     | 03:10 |                |
| Air Blank         | 0.000     | 03:11 |                |
| Control Test      | 0.077     | 03:11 |                |
| Air Blank         | 0.000     | 03:11 |                |
| Diagnostics Check | OK        | 03:12 |                |

---

Data Download Date/Time: 10/13/2016 10:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/03/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CORDERO

DATE/TIME OF TEST: 09/02/2016 02:27:50

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:24 |                      |
| Air Blank         | 0.000 | 02:25 |                      |
| Control Test      | 0.078 | 02:25 |                      |
| Air Blank         | 0.000 | 02:26 |                      |
| Subject Sample    | # 1   | 02:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:26 |                      |
| Air Blank         | 0.000 | 02:26 |                      |
| Control Test      | 0.077 | 02:27 |                      |
| Air Blank         | 0.000 | 02:27 |                      |
| Diagnostics Check | OK    | 02:27 |                      |

---

Data Download Date/Time: 09/06/2016 12:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CLOSIUS,RICHARD  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/03/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 09/02/2016 04:26:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:23 |                      |
| Air Blank         | 0.000 | 04:24 |                      |
| Control Test      | 0.079 | 04:24 |                      |
| Air Blank         | 0.000 | 04:24 |                      |
| Subject Sample    | # 1   | 04:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:25 |                      |
| Air Blank         | 0.000 | 04:25 |                      |
| Control Test      | 0.078 | 04:25 |                      |
| Air Blank         | 0.000 | 04:26 |                      |
| Diagnostics Check | OK    | 04:26 |                      |

---

Data Download Date/Time: 09/06/2016 12:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VALENCIA,ALEJANDRO,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: SIDDO,RAAJU  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 08/03/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: MUNOZ HERNANDEZ

DATE/TIME OF TEST: 09/03/2016 01:06:33

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 00:59 |
| Air Blank         |     | 0.000 | 01:00 |
| Control Test      |     | 0.079 | 01:00 |
| Air Blank         |     | 0.000 | 01:01 |
| Subject Sample    | # 1 | 0.275 | 01:01 |
| Breath Volume     |     | 2.640 | 01:01 |
| Air Blank         |     | 0.000 | 01:02 |
| Air Blank         |     | 0.000 | 01:04 |
| Subject Sample    | # 2 | 0.272 | 01:04 |
| Breath Volume     |     | 1.847 | 01:04 |
| Air Blank         |     | 0.000 | 01:05 |
| Control Test      |     | 0.075 | 01:05 |
| Air Blank         |     | 0.000 | 01:06 |
| Diagnostics Check |     | OK    | 01:06 |

---

Data Download Date/Time: 09/06/2016 12:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 09/07/2016 03:43:18

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:36 |
| Air Blank         |     | 0.000 | 03:36 |
| Control Test      |     | 0.079 | 03:37 |
| Air Blank         |     | 0.000 | 03:37 |
| Subject Sample    | # 1 | 0.214 | 03:38 |
| Breath Volume     |     | 2.316 | 03:38 |
| Air Blank         |     | 0.000 | 03:39 |
| Air Blank         |     | 0.000 | 03:41 |
| Subject Sample    | # 2 | 0.200 | 03:41 |
| Breath Volume     |     | 1.441 | 03:41 |
| Air Blank         |     | 0.000 | 03:42 |
| Control Test      |     | 0.077 | 03:42 |
| Air Blank         |     | 0.000 | 03:43 |
| Diagnostics Check |     | OK    | 03:43 |

---

Data Download Date/Time: 10/10/2016 10:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARDELAS,MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: BARDELAS,MAYLEE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/07/2016 22:00:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 21:56 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Control Test      | 0.078 | 21:57 |                      |
| Air Blank         | 0.000 | 21:57 |                      |
| Subject Sample    | # 1   | 21:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 21:59 |                      |
| Air Blank         | 0.000 | 21:59 |                      |
| Control Test      | 0.078 | 22:00 |                      |
| Air Blank         | 0.000 | 22:00 |                      |
| Diagnostics Check | OK    | 22:00 |                      |

---

Data Download Date/Time: 10/10/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 09/08/2016 04:11:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:04 |
| Air Blank          | 0.000 | 04:05 |
| Control Test       | 0.079 | 04:05 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 1 | 0.109 | 04:06 |
| Breath Volume      | 2.250 | 04:06 |
| Air Blank          | 0.000 | 04:07 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 2 | 0.108 | 04:09 |
| Breath Volume      | 2.132 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.077 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Diagnostics Check  | OK    | 04:11 |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: ORRIOLS

DATE/TIME OF TEST: 09/08/2016 04:40:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:34 |
| Air Blank          | 0.000 | 04:34 |
| Control Test       | 0.077 | 04:35 |
| Air Blank          | 0.000 | 04:35 |
| Subject Sample # 1 | 0.173 | 04:35 |
| Breath Volume      | 2.714 | 04:35 |
| Air Blank          | 0.000 | 04:36 |
| Air Blank          | 0.000 | 04:38 |
| Subject Sample # 2 | 0.156 | 04:38 |
| Breath Volume      | 1.941 | 04:38 |
| Air Blank          | 0.000 | 04:39 |
| Control Test       | 0.077 | 04:39 |
| Air Blank          | 0.000 | 04:40 |
| Diagnostics Check  | OK    | 04:40 |

---

Data Download Date/Time: 10/10/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PACHECO,GEORGE,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 09/08/2016 06:39:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:33 |
| Air Blank          | 0.000 | 06:33 |
| Control Test       | 0.078 | 06:34 |
| Air Blank          | 0.000 | 06:34 |
| Subject Sample # 1 | 0.149 | 06:35 |
| Breath Volume      | 2.933 | 06:35 |
| Air Blank          | 0.000 | 06:35 |
| Air Blank          | 0.000 | 06:37 |
| Subject Sample # 2 | 0.154 | 06:38 |
| Breath Volume      | 2.671 | 06:38 |
| Air Blank          | 0.000 | 06:38 |
| Control Test       | 0.077 | 06:39 |
| Air Blank          | 0.000 | 06:39 |
| Diagnostics Check  | OK    | 06:39 |

---

Data Download Date/Time: 10/10/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARDELAS,MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: BARDELAS,MAYLEE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/08/2016 20:30:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:23 |
| Air Blank          | 0.000 | 20:23 |
| Control Test       | 0.078 | 20:24 |
| Air Blank          | 0.000 | 20:24 |
| Subject Sample # 1 | 0.254 | 20:25 |
| Breath Volume      | 1.871 | 20:25 |
| Air Blank          | 0.000 | 20:25 |
| Air Blank          | 0.000 | 20:27 |
| Subject Sample # 2 | 0.234 | 20:28 |
| Breath Volume      | 1.734 | 20:28 |
| Air Blank          | 0.000 | 20:29 |
| Control Test       | 0.077 | 20:29 |
| Air Blank          | 0.000 | 20:30 |
| Diagnostics Check  | OK    | 20:30 |

---

Data Download Date/Time: 10/10/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: GOMEZ VICENTE

DATE/TIME OF TEST: 09/10/2016 02:51:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.078 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Subject Sample # 1 | 0.266 | 02:46 |
| Breath Volume      | 1.886 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Air Blank          | 0.000 | 02:49 |
| Subject Sample # 2 | 0.265 | 02:49 |
| Breath Volume      | 2.542 | 02:49 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.078 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Diagnostics Check  | OK    | 02:51 |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,DAEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: DE POOL

DATE/TIME OF TEST: 09/10/2016 04:13:58

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:10 |                      |
| Air Blank         | 0.000 | 04:11 |                      |
| Control Test      | 0.078 | 04:11 |                      |
| Air Blank         | 0.000 | 04:12 |                      |
| Subject Sample    | # 1   | 04:12 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:12 |                      |
| Air Blank         | 0.000 | 04:13 |                      |
| Control Test      | 0.078 | 04:13 |                      |
| Air Blank         | 0.000 | 04:13 |                      |
| Diagnostics Check | OK    | 04:13 |                      |

---

Data Download Date/Time: 10/10/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIMM,HUNTER,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRIMM,HUNTER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITE CARVAJAL

DATE/TIME OF TEST: 09/12/2016 23:47:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:38 |
| Air Blank          | 0.000 | 23:39 |
| Control Test       | 0.078 | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 1 | 0.315 | 23:42 |
| Breath Volume      | 1.585 | 23:42 |
| Air Blank          | 0.000 | 23:43 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 2 | 0.317 | 23:45 |
| Breath Volume      | 1.984 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Control Test       | 0.077 | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Diagnostics Check  | OK    | 23:47 |

---

Data Download Date/Time: 10/10/2016 10:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 09/14/2016 01:44:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.078 | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Subject Sample # 1 | 0.184 | 01:39 |
| Breath Volume      | 1.367 | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 2 | 0.178 | 01:42 |
| Breath Volume      | 1.429 | 01:42 |
| Air Blank          | 0.000 | 01:43 |
| Control Test       | 0.077 | 01:43 |
| Air Blank          | 0.000 | 01:44 |
| Diagnostics Check  | OK    | 01:44 |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LOPEZ,PAOLI,  
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD  
ARRESTING OFFICER: LOPEZ,PAOLI  
ARRESTING OFFICER AGENCY: Hialeah Gardens PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 10:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 11:45

SUBJECT'S LAST NAME: LLANES

DATE/TIME OF TEST: 09/14/2016 12:08:57

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 12:06 |                      |
| Air Blank         | 0.000 | 12:06 |                      |
| Control Test      | 0.078 | 12:06 |                      |
| Air Blank         | 0.000 | 12:07 |                      |
| Subject Sample    | # 1   | 12:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 12:07 |                      |
| Air Blank         | 0.000 | 12:08 |                      |
| Control Test      | 0.077 | 12:08 |                      |
| Air Blank         | 0.000 | 12:08 |                      |
| Diagnostics Check | OK    | 12:08 |                      |

---

Data Download Date/Time: 10/10/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WHITTINGHAM

DATE/TIME OF TEST: 09/15/2016 03:44:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Control Test       | 0.079 | 03:38 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 1 | 0.182 | 03:39 |
| Breath Volume      | 2.078 | 03:39 |
| Air Blank          | 0.000 | 03:40 |
| Air Blank          | 0.000 | 03:42 |
| Subject Sample # 2 | 0.188 | 03:42 |
| Breath Volume      | 2.410 | 03:42 |
| Air Blank          | 0.000 | 03:43 |
| Control Test       | 0.077 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Diagnostics Check  | OK    | 03:44 |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,DAEL  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: COOMBS

DATE/TIME OF TEST: 09/16/2016 02:19:48

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:16 |                      |
| Air Blank         | 0.000 | 02:17 |                      |
| Control Test      | 0.078 | 02:17 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Subject Sample    | # 1   | 02:18 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:18 |                      |
| Air Blank         | 0.000 | 02:18 |                      |
| Control Test      | 0.077 | 02:19 |                      |
| Air Blank         | 0.000 | 02:19 |                      |
| Diagnostics Check | OK    | 02:19 |                      |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: GONZALEZ RODRIGUEZ

DATE/TIME OF TEST: 09/16/2016 04:36:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:29 |
| Air Blank          | 0.000 | 04:29 |
| Control Test       | 0.078 | 04:30 |
| Air Blank          | 0.000 | 04:30 |
| Subject Sample # 1 | 0.083 | 04:31 |
| Breath Volume      | 1.914 | 04:31 |
| Air Blank          | 0.000 | 04:32 |
| Air Blank          | 0.000 | 04:34 |
| Subject Sample # 2 | 0.078 | 04:34 |
| Breath Volume      | 1.773 | 04:34 |
| Air Blank          | 0.000 | 04:35 |
| Control Test       | 0.077 | 04:35 |
| Air Blank          | 0.000 | 04:36 |
| Diagnostics Check  | OK    | 04:36 |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ESCONTELA,CRISTINA,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: ESCONTELA,CRISTINA  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 08:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:32

SUBJECT'S LAST NAME: ROMERO DIAZ

DATE/TIME OF TEST: 09/17/2016 08:59:34

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:53 |
| Air Blank          | 0.000 | 08:53 |
| Control Test       | 0.078 | 08:53 |
| Air Blank          | 0.000 | 08:54 |
| Subject Sample # 1 | 0.183 | 08:54 |
| Breath Volume      | 2.261 | 08:54 |
| Air Blank          | 0.000 | 08:55 |
| Air Blank          | 0.000 | 08:57 |
| Subject Sample # 2 | 0.178 | 08:57 |
| Breath Volume      | 2.250 | 08:57 |
| Air Blank          | 0.000 | 08:58 |
| Control Test       | 0.076 | 08:58 |
| Air Blank          | 0.000 | 08:59 |
| Diagnostics Check  | OK    | 08:59 |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BONILLA,JOSE,I  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: BONILLA,JOSE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: BISBE HERNANDEZ

DATE/TIME OF TEST: 09/18/2016 06:17:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:10 |
| Air Blank          | 0.000 | 06:11 |
| Control Test       | 0.079 | 06:11 |
| Air Blank          | 0.000 | 06:11 |
| Subject Sample # 1 | 0.151 | 06:12 |
| Breath Volume      | 4.093 | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Air Blank          | 0.000 | 06:14 |
| Subject Sample # 2 | 0.151 | 06:15 |
| Breath Volume      | 2.941 | 06:15 |
| Air Blank          | 0.000 | 06:16 |
| Control Test       | 0.075 | 06:16 |
| Air Blank          | 0.000 | 06:17 |
| Diagnostics Check  | OK    | 06:17 |

---

Data Download Date/Time: 10/10/2016 10:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VALENCIA,KEVIN,A  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: VALENCIA,KEVIN  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: ZAMORANO

DATE/TIME OF TEST: 09/18/2016 19:26:23

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 19:21 |                      |
| Air Blank         | 0.000 | 19:22 |                      |
| Control Test      | 0.079 | 19:22 |                      |
| Air Blank         | 0.000 | 19:22 |                      |
| Subject Sample    | # 1   | 19:25 | Subject Test Refused |
| Breath Volume     | 0.000 | 19:25 |                      |
| Air Blank         | 0.000 | 19:25 |                      |
| Control Test      | 0.078 | 19:25 |                      |
| Air Blank         | 0.000 | 19:26 |                      |
| Diagnostics Check | OK    | 19:26 |                      |

---

Data Download Date/Time: 10/10/2016 10:10



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUERRA,FRANCISCO,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GUERRA,FRANCISCO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 09/19/2016 04:28:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.078 | 04:21 |
| Air Blank          | 0.000 | 04:22 |
| Subject Sample # 1 | 0.213 | 04:23 |
| Breath Volume      | 2.882 | 04:23 |
| Air Blank          | 0.000 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 2 | 0.212 | 04:26 |
| Breath Volume      | 3.023 | 04:26 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.075 | 04:27 |
| Air Blank          | 0.000 | 04:28 |
| Diagnostics Check  | OK    | 04:28 |

---

Data Download Date/Time: 10/10/2016 10:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUESTA,DANIEL,A  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: CUESTA,DANIEL  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 07:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: VANEGAS

DATE/TIME OF TEST: 09/20/2016 07:37:13

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 07:25 |                    |
| Air Blank         | 0.000 | 07:26 |                    |
| Control Test      | 0.078 | 07:26 |                    |
| Air Blank         | 0.000 | 07:27 |                    |
| Subject Sample    | # 1   | 07:30 | No Sample Provided |
| Breath Volume     | 0.000 | 07:30 |                    |
| Air Blank         | 0.000 | 07:30 |                    |
| Air Blank         | 0.000 | 07:32 |                    |
| Subject Sample    | # 2   | 07:35 | No Sample Provided |
| Breath Volume     | 0.000 | 07:35 |                    |
| Air Blank         | 0.000 | 07:36 |                    |
| Control Test      | 0.079 | 07:36 |                    |
| Air Blank         | 0.000 | 07:37 |                    |
| Diagnostics Check | OK    | 07:37 |                    |

---

Data Download Date/Time: 10/10/2016 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/21/2016 04:27:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:20 |
| Air Blank          | 0.000 | 04:21 |
| Control Test       | 0.078 | 04:21 |
| Air Blank          | 0.000 | 04:21 |
| Subject Sample # 1 | 0.000 | 04:22 |
| Breath Volume      | 1.496 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Air Blank          | 0.000 | 04:25 |
| Subject Sample # 2 | 0.000 | 04:25 |
| Breath Volume      | 1.636 | 04:25 |
| Air Blank          | 0.000 | 04:26 |
| Control Test       | 0.077 | 04:26 |
| Air Blank          | 0.000 | 04:26 |
| Diagnostics Check  | OK    | 04:27 |

---

Data Download Date/Time: 10/10/2016 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARDELAS,MAYLEE,  
BREATH TEST OPERATOR AGENCY: OTHER Local  
ARRESTING OFFICER: BARDELAS,MAYLEE  
ARRESTING OFFICER AGENCY: OTHER Local  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/21/2016 19:00:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:54 |
| Air Blank          | 0.000 | 18:55 |
| Control Test       | 0.078 | 18:55 |
| Air Blank          | 0.000 | 18:56 |
| Subject Sample # 1 | 0.000 | 18:56 |
| Breath Volume      | 2.285 | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Air Blank          | 0.000 | 18:58 |
| Subject Sample # 2 | 0.000 | 18:59 |
| Breath Volume      | 1.777 | 18:59 |
| Air Blank          | 0.000 | 19:00 |
| Control Test       | 0.077 | 19:00 |
| Air Blank          | 0.000 | 19:00 |
| Diagnostics Check  | OK    | 19:00 |

---

Data Download Date/Time: 10/10/2016 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: GUTIERREZ-FLORES

DATE/TIME OF TEST: 09/23/2016 04:20:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.079 | 04:09 |
| Air Blank          | 0.000 | 04:09 |
| Subject Sample # 1 | 0.149 | 04:12 |
| Breath Volume      | 1.218 | 04:12 |
| Air Blank          | 0.000 | 04:13 |
| Air Blank          | 0.000 | 04:15 |
| Subject Sample # 2 | 0.145 | 04:18 |
| Breath Volume      | 1.460 | 04:18 |
| Air Blank          | 0.000 | 04:19 |
| Control Test       | 0.076 | 04:19 |
| Air Blank          | 0.000 | 04:20 |
| Diagnostics Check  | OK    | 04:20 |

---

Data Download Date/Time: 10/10/2016 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: TUNIDOR

DATE/TIME OF TEST: 09/23/2016 04:35:43

---

|                    |       |       |                           |
|--------------------|-------|-------|---------------------------|
| Diagnostics Check  | OK    | 04:28 |                           |
| Air Blank          | 0.000 | 04:28 |                           |
| Control Test       | 0.076 | 04:29 |                           |
| Air Blank          | 0.000 | 04:29 |                           |
| Subject Sample # 1 | 0.093 | 04:30 |                           |
| Breath Volume      | 1.796 | 04:30 |                           |
| Air Blank          | 0.000 | 04:31 |                           |
| Air Blank          | 0.000 | 04:33 |                           |
| Subject Sample # 2 | 0.089 | 04:34 |                           |
| Breath Volume      | 1.820 | 04:34 |                           |
| Air Blank          | 0.000 | 04:34 |                           |
| Control Test       | 0.070 | 04:35 | Control Outside Tolerance |
| Air Blank          | 0.000 | 04:35 |                           |

---

Data Download Date/Time: 10/10/2016 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: TUNDIDOR

DATE/TIME OF TEST: 09/23/2016 04:46:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:39 |
| Air Blank          | 0.000 | 04:40 |
| Control Test       | 0.075 | 04:40 |
| Air Blank          | 0.000 | 04:41 |
| Subject Sample # 1 | 0.099 | 04:41 |
| Breath Volume      | 2.511 | 04:41 |
| Air Blank          | 0.000 | 04:42 |
| Air Blank          | 0.000 | 04:44 |
| Subject Sample # 2 | 0.098 | 04:44 |
| Breath Volume      | 2.363 | 04:44 |
| Air Blank          | 0.000 | 04:45 |
| Control Test       | 0.076 | 04:45 |
| Air Blank          | 0.000 | 04:46 |
| Diagnostics Check  | OK    | 04:46 |

---

Data Download Date/Time: 10/10/2016 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: LEON VELARDE

DATE/TIME OF TEST: 09/23/2016 05:19:10

---

|                   |           |       |                           |
|-------------------|-----------|-------|---------------------------|
| Diagnostics Check | OK        | 05:10 |                           |
| Air Blank         | 0.000     | 05:11 |                           |
| Control Test      | 0.076     | 05:11 |                           |
| Air Blank         | 0.000     | 05:11 |                           |
| Subject Sample    | # 1 0.110 | 05:14 |                           |
| Breath Volume     | 2.089     | 05:14 |                           |
| Air Blank         | 0.000     | 05:14 |                           |
| Air Blank         | 0.000     | 05:16 |                           |
| Subject Sample    | # 2 0.098 | 05:17 |                           |
| Breath Volume     | 2.093     | 05:17 |                           |
| Air Blank         | 0.000     | 05:18 |                           |
| Control Test      | 0.074     | 05:18 | Control Outside Tolerance |
| Air Blank         | 0.000     | 05:19 |                           |

---

Data Download Date/Time: 10/10/2016 10:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: FRANCO,GREGORY  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: LEON VELARDE

DATE/TIME OF TEST: 09/23/2016 05:33:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:27 |
| Air Blank          | 0.000 | 05:27 |
| Control Test       | 0.076 | 05:28 |
| Air Blank          | 0.000 | 05:28 |
| Subject Sample # 1 | 0.093 | 05:29 |
| Breath Volume      | 1.343 | 05:29 |
| Air Blank          | 0.000 | 05:29 |
| Air Blank          | 0.000 | 05:31 |
| Subject Sample # 2 | 0.093 | 05:31 |
| Breath Volume      | 1.273 | 05:31 |
| Air Blank          | 0.000 | 05:32 |
| Control Test       | 0.076 | 05:33 |
| Air Blank          | 0.000 | 05:33 |
| Diagnostics Check  | OK    | 05:33 |

---

Data Download Date/Time: 10/10/2016 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,F  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: AVEILLE RODRIGUEZ

DATE/TIME OF TEST: 09/24/2016 03:19:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.079 | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 1 | 0.127 | 03:14 |
| Breath Volume      | 2.867 | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 2 | 0.125 | 03:17 |
| Breath Volume      | 3.445 | 03:17 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.077 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Diagnostics Check  | OK    | 03:19 |

---

Data Download Date/Time: 10/10/2016 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 09/28/2016 04:56:40

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:53 |                      |
| Air Blank         | 0.000 | 04:54 |                      |
| Control Test      | 0.078 | 04:54 |                      |
| Air Blank         | 0.000 | 04:55 |                      |
| Subject Sample    | # 1   | 04:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:55 |                      |
| Air Blank         | 0.000 | 04:55 |                      |
| Control Test      | 0.078 | 04:56 |                      |
| Air Blank         | 0.000 | 04:56 |                      |
| Diagnostics Check | OK    | 04:56 |                      |

---

Data Download Date/Time: 10/10/2016 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: PACHECO,GEORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 09/30/2016 04:14:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:08 |
| Air Blank          | 0.000 | 04:09 |
| Control Test       | 0.079 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Subject Sample # 1 | 0.142 | 04:10 |
| Breath Volume      | 1.667 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 2 | 0.136 | 04:13 |
| Breath Volume      | 1.617 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Control Test       | 0.078 | 04:14 |
| Air Blank          | 0.000 | 04:14 |
| Diagnostics Check  | OK    | 04:14 |

---

Data Download Date/Time: 10/10/2016 10:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYRTIL,REGINALD,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ALFARO

DATE/TIME OF TEST: 09/30/2016 05:38:44

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:35 |                      |
| Air Blank         | 0.000 | 05:36 |                      |
| Control Test      | 0.078 | 05:36 |                      |
| Air Blank         | 0.000 | 05:37 |                      |
| Subject Sample    | # 1   | 05:37 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:37 |                      |
| Air Blank         | 0.000 | 05:37 |                      |
| Control Test      | 0.078 | 05:38 |                      |
| Air Blank         | 0.000 | 05:38 |                      |
| Diagnostics Check | OK    | 05:38 |                      |

---

Data Download Date/Time: 10/10/2016 10:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: BARQUERO,J  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: AGIL-CANCIO

DATE/TIME OF TEST: 09/03/2016 22:49:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:42 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.082 | 22:43 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 1 | 0.164 | 22:44 |
| Breath Volume      | 1.394 | 22:44 |
| Air Blank          | 0.000 | 22:45 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 2 | 0.158 | 22:48 |
| Breath Volume      | 1.847 | 22:48 |
| Air Blank          | 0.000 | 22:49 |
| Control Test       | 0.083 | 22:49 |
| Air Blank          | 0.000 | 22:49 |
| Diagnostics Check  | OK    | 22:49 |

---

Data Download Date/Time: 09/05/2016 11:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIANCHI,MARIO,D  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: BIANCHI,MARIO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: CAMINOS RAMIREZ

DATE/TIME OF TEST: 09/21/2016 06:32:17

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 06:32 |            |
| Air Blank         |    | 06:32 | RFI Detect |
| Air Blank         |    | 06:32 | RFI Detect |

---

Data Download Date/Time: 10/12/2016 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BIANCHI,MARIO,D  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: BIANCHI,MARIO  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:08

SUBJECT'S LAST NAME: CAMINOS RAMIREZ

DATE/TIME OF TEST: 09/21/2016 06:43:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:37 |
| Air Blank          | 0.000 | 06:37 |
| Control Test       | 0.082 | 06:37 |
| Air Blank          | 0.000 | 06:38 |
| Subject Sample # 1 | 0.221 | 06:38 |
| Breath Volume      | 1.156 | 06:38 |
| Air Blank          | 0.000 | 06:39 |
| Air Blank          | 0.000 | 06:41 |
| Subject Sample # 2 | 0.218 | 06:42 |
| Breath Volume      | 1.546 | 06:42 |
| Air Blank          | 0.000 | 06:42 |
| Control Test       | 0.081 | 06:43 |
| Air Blank          | 0.000 | 06:43 |
| Diagnostics Check  | OK    | 06:43 |

---

Data Download Date/Time: 10/12/2016 11:04



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: TASE

DATE/TIME OF TEST: 09/22/2016 02:35:17

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:31 |                      |
| Air Blank         | 0.000 | 02:31 |                      |
| Control Test      | 0.082 | 02:32 |                      |
| Air Blank         | 0.000 | 02:32 |                      |
| Subject Sample    | # 1   | 02:33 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:33 |                      |
| Air Blank         | 0.000 | 02:34 |                      |
| Control Test      | 0.082 | 02:34 |                      |
| Air Blank         | 0.000 | 02:35 |                      |
| Diagnostics Check | OK    | 02:35 |                      |

---

Data Download Date/Time: 10/12/2016 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CASTILLO,CHRISTIAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: MIDGETTE

DATE/TIME OF TEST: 09/24/2016 04:52:51

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:49 |            |
| Air Blank         | 0.000 | 04:50 |            |
| Control Test      | 0.082 | 04:50 |            |
| Air Blank         | 0.000 | 04:51 |            |
| Subject Sample    | # 1   | 04:51 | RFI Detect |
| Breath Volume     | 0.000 | 04:51 |            |
| Air Blank         | 0.000 | 04:51 |            |
| Control Test      | 0.082 | 04:52 |            |
| Air Blank         | 0.000 | 04:52 |            |
| Diagnostics Check | OK    | 04:52 |            |

---

Data Download Date/Time: 10/12/2016 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CASTILLO,CHRISTIAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: MIDGETTE

DATE/TIME OF TEST: 09/24/2016 05:21:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:13 |
| Air Blank          | 0.000 | 05:14 |
| Control Test       | 0.082 | 05:14 |
| Air Blank          | 0.000 | 05:15 |
| Subject Sample # 1 | 0.122 | 05:16 |
| Breath Volume      | 3.554 | 05:16 |
| Air Blank          | 0.000 | 05:16 |
| Air Blank          | 0.000 | 05:18 |
| Subject Sample # 2 | 0.123 | 05:19 |
| Breath Volume      | 3.777 | 05:19 |
| Air Blank          | 0.000 | 05:20 |
| Control Test       | 0.082 | 05:20 |
| Air Blank          | 0.000 | 05:20 |
| Diagnostics Check  | OK    | 05:21 |

---

Data Download Date/Time: 10/12/2016 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CASTILLO,CHRISTIAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NADAL BORGES

DATE/TIME OF TEST: 09/25/2016 03:34:15

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:26 |
| Air Blank         |     | 0.000 | 03:27 |
| Control Test      |     | 0.082 | 03:27 |
| Air Blank         |     | 0.000 | 03:27 |
| Subject Sample    | # 1 | 0.216 | 03:29 |
| Breath Volume     |     | 3.207 | 03:29 |
| Air Blank         |     | 0.000 | 03:30 |
| Air Blank         |     | 0.000 | 03:31 |
| Subject Sample    | # 2 | 0.208 | 03:32 |
| Breath Volume     |     | 2.585 | 03:32 |
| Air Blank         |     | 0.000 | 03:33 |
| Control Test      |     | 0.082 | 03:33 |
| Air Blank         |     | 0.000 | 03:34 |
| Diagnostics Check |     | OK    | 03:34 |

---

Data Download Date/Time: 10/12/2016 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTEGA,PEDRO,L  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: BRUTTO,ESTRELLA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 08:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 09/25/2016 08:46:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 08:43 |                      |
| Air Blank         | 0.000 | 08:44 |                      |
| Control Test      | 0.082 | 08:44 |                      |
| Air Blank         | 0.000 | 08:45 |                      |
| Subject Sample    | # 1   | 08:45 | Subject Test Refused |
| Breath Volume     | 0.000 | 08:45 |                      |
| Air Blank         | 0.000 | 08:45 |                      |
| Control Test      | 0.082 | 08:46 |                      |
| Air Blank         | 0.000 | 08:46 |                      |
| Diagnostics Check | OK    | 08:46 |                      |

---

Data Download Date/Time: 10/12/2016 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CASTILLO,CHRISTIAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MASIP TOURIZ

DATE/TIME OF TEST: 09/26/2016 00:40:50

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 00:35 |            |
| Air Blank         | 0.000     | 00:35 |            |
| Control Test      | 0.082     | 00:36 |            |
| Air Blank         | 0.000     | 00:36 |            |
| Subject Sample    | # 1 0.309 | 00:39 |            |
| Breath Volume     | 1.519     | 00:39 |            |
| Air Blank         |           | 00:40 | Purge Fail |
| Air Blank         |           | 00:40 | Purge Fail |

---

Data Download Date/Time: 10/12/2016 11:04

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: CASTILLO,CHRISTIAN  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: MASIP TOURIZ

DATE/TIME OF TEST: 09/26/2016 01:13:55

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 01:02 |                    |
| Air Blank         | 0.000 | 01:02 |                    |
| Control Test      | 0.081 | 01:03 |                    |
| Air Blank         | 0.000 | 01:03 |                    |
| Subject Sample    | # 1   | 01:06 | No Sample Provided |
| Breath Volume     | 0.000 | 01:06 |                    |
| Air Blank         | 0.000 | 01:07 |                    |
| Air Blank         | 0.000 | 01:09 |                    |
| Subject Sample    | # 2   | 01:12 | No Sample Provided |
| Breath Volume     | 0.000 | 01:12 |                    |
| Air Blank         | 0.000 | 01:13 |                    |
| Control Test      | 0.082 | 01:13 |                    |
| Air Blank         | 0.000 | 01:13 |                    |
| Diagnostics Check | OK    | 01:13 |                    |

---

Data Download Date/Time: 10/12/2016 11:04

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VARGAS,D,  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: VARGAS,D  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 08/18/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: MARIN

DATE/TIME OF TEST: 09/03/2016 19:58:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:52 |
| Air Blank          | 0.000 | 19:52 |
| Control Test       | 0.080 | 19:53 |
| Air Blank          | 0.000 | 19:53 |
| Subject Sample # 1 | 0.076 | 19:54 |
| Breath Volume      | 1.585 | 19:54 |
| Air Blank          | 0.000 | 19:54 |
| Air Blank          | 0.000 | 19:56 |
| Subject Sample # 2 | 0.070 | 19:57 |
| Breath Volume      | 1.433 | 19:57 |
| Air Blank          | 0.000 | 19:57 |
| Control Test       | 0.080 | 19:58 |
| Air Blank          | 0.000 | 19:58 |
| Diagnostics Check  | OK    | 19:58 |

---

Data Download Date/Time: 09/05/2016 12:02



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIMM,HUNTER,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRIMM,HUNTER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITE CARVAJAL

DATE/TIME OF TEST: 09/12/2016 21:57:21

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 21:51 |            |
| Air Blank         | 0.000     | 21:51 |            |
| Control Test      | 0.079     | 21:51 |            |
| Air Blank         | 0.000     | 21:52 |            |
| Subject Sample    | # 1 0.343 | 21:55 |            |
| Breath Volume     | 1.207     | 21:55 |            |
| Air Blank         |           | 21:56 | Purge Fail |
| Air Blank         | 0.000     | 21:57 |            |

---

Data Download Date/Time: 10/12/2016 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIMM,HUNTER,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRIMM,HUNTER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITE CARVAJAL

DATE/TIME OF TEST: 09/12/2016 22:10:31

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 22:03 |            |
| Air Blank         | 0.000     | 22:03 |            |
| Control Test      | 0.077     | 22:03 |            |
| Air Blank         | 0.000     | 22:04 |            |
| Subject Sample    | # 1 0.353 | 22:05 |            |
| Breath Volume     | 1.343     | 22:05 |            |
| Air Blank         | 0.000     | 22:06 |            |
| Air Blank         | 0.000     | 22:07 |            |
| Subject Sample    | # 2 0.351 | 22:08 |            |
| Breath Volume     | 1.835     | 22:08 |            |
| Air Blank         |           | 22:09 | Purge Fail |
| Air Blank         | 0.000     | 22:10 |            |

---

Data Download Date/Time: 10/12/2016 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIMM,HUNTER,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRIMM,HUNTER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITE CARVAJAL

DATE/TIME OF TEST: 09/12/2016 22:19:56

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 22:18 |                           |
| Air Blank         | 0.000 | 22:19 |                           |
| Control Test      | 0.067 | 22:19 | Control Outside Tolerance |
| Air Blank         | 0.000 | 22:19 |                           |

---

Data Download Date/Time: 10/12/2016 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIMM,HUNTER,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRIMM,HUNTER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITE CARVAJAL

DATE/TIME OF TEST: 09/12/2016 22:29:25

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 22:28 |                           |
| Air Blank         | 0.000 | 22:28 |                           |
| Control Test      | 0.072 | 22:28 | Control Outside Tolerance |
| Air Blank         | 0.000 | 22:29 |                           |

---

Data Download Date/Time: 10/12/2016 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRIMM,HUNTER,F  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: GRIMM,HUNTER  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: WHITE CARVAJAL

DATE/TIME OF TEST: 09/12/2016 22:54:36

---

|                   |       |       |                           |
|-------------------|-------|-------|---------------------------|
| Diagnostics Check | OK    | 22:53 |                           |
| Air Blank         | 0.000 | 22:53 |                           |
| Control Test      | 0.074 | 22:54 | Control Outside Tolerance |
| Air Blank         | 0.000 | 22:54 |                           |

---

Data Download Date/Time: 10/12/2016 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOIVU,DAVID,J  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: ENCINOSA,JORGE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: FERNANDEZ CASTILLO

DATE/TIME OF TEST: 09/15/2016 20:05:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:59 |
| Air Blank          | 0.000 | 19:59 |
| Control Test       | 0.079 | 19:59 |
| Air Blank          | 0.000 | 20:00 |
| Subject Sample # 1 | 0.223 | 20:00 |
| Breath Volume      | 2.519 | 20:00 |
| Air Blank          | 0.000 | 20:01 |
| Air Blank          | 0.000 | 20:03 |
| Subject Sample # 2 | 0.220 | 20:03 |
| Breath Volume      | 2.820 | 20:03 |
| Air Blank          | 0.000 | 20:04 |
| Control Test       | 0.079 | 20:04 |
| Air Blank          | 0.000 | 20:05 |
| Diagnostics Check  | OK    | 20:05 |

---

Data Download Date/Time: 10/12/2016 11:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KINSEY-SMITH,FELCIA,R  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: KINSEY-SMITH,FELICIA  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 07:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 09/24/2016 08:54:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:47 |
| Air Blank          | 0.000 | 08:47 |
| Control Test       | 0.078 | 08:48 |
| Air Blank          | 0.000 | 08:48 |
| Subject Sample # 1 | 0.090 | 08:49 |
| Breath Volume      | 4.167 | 08:49 |
| Air Blank          | 0.000 | 08:50 |
| Air Blank          | 0.000 | 08:52 |
| Subject Sample # 2 | 0.084 | 08:52 |
| Breath Volume      | 2.902 | 08:52 |
| Air Blank          | 0.000 | 08:53 |
| Control Test       | 0.079 | 08:53 |
| Air Blank          | 0.000 | 08:54 |
| Diagnostics Check  | OK    | 08:54 |

---

Data Download Date/Time: 10/12/2016 11:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENRY, MAURICE, M  
BREATH TEST OPERATOR AGENCY: Miami-Dade PD  
ARRESTING OFFICER: HENRY, MAURICE  
ARRESTING OFFICER AGENCY: Miami-Dade PD  
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 22315080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: ULLOA

DATE/TIME OF TEST: 09/25/2016 02:01:32

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:58 |                      |
| Air Blank         | 0.000 | 01:59 |                      |
| Control Test      | 0.079 | 01:59 |                      |
| Air Blank         | 0.000 | 01:59 |                      |
| Subject Sample    | # 1   | 02:00 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:00 |                      |
| Air Blank         | 0.000 | 02:00 |                      |
| Control Test      | 0.078 | 02:00 |                      |
| Air Blank         | 0.000 | 02:01 |                      |
| Diagnostics Check | OK    | 02:01 |                      |

---

Data Download Date/Time: 10/12/2016 11:13



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYA,JOSE,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: AJO,M  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: HEMINGWAY

DATE/TIME OF TEST: 09/03/2016 03:33:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Control Test       | 0.080 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 1 | 0.152 | 03:29 |
| Breath Volume      | 1.710 | 03:29 |
| Air Blank          | 0.000 | 03:29 |
| Air Blank          | 0.000 | 03:31 |
| Subject Sample # 2 | 0.150 | 03:32 |
| Breath Volume      | 1.320 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.077 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Diagnostics Check  | OK    | 03:33 |

---

Data Download Date/Time: 09/29/2016 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,KENNETH,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: KIRKUS,JOSHUA  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 09/12/2016 03:31:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.079 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 1 | 0.174 | 03:27 |
| Breath Volume      | 1.417 | 03:27 |
| Air Blank          | 0.000 | 03:27 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.177 | 03:29 |
| Breath Volume      | 1.203 | 03:29 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.080 | 03:30 |
| Air Blank          | 0.000 | 03:31 |
| Diagnostics Check  | OK    | 03:31 |

---

Data Download Date/Time: 09/29/2016 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINCHESTER,AARON,A  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: KELLNER,DOUGLAS  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: HILL JR

DATE/TIME OF TEST: 09/17/2016 03:11:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:03 |
| Air Blank          | 0.000 | 03:03 |
| Control Test       | 0.080 | 03:03 |
| Air Blank          | 0.000 | 03:04 |
| Subject Sample # 1 | 0.037 | 03:06 |
| Breath Volume      | 1.476 | 03:06 |
| Air Blank          | 0.000 | 03:06 |
| Air Blank          | 0.000 | 03:08 |
| Subject Sample # 2 | 0.038 | 03:09 |
| Breath Volume      | 1.378 | 03:09 |
| Air Blank          | 0.000 | 03:10 |
| Control Test       | 0.079 | 03:10 |
| Air Blank          | 0.000 | 03:11 |
| Diagnostics Check  | OK    | 03:11 |

---

Data Download Date/Time: 09/29/2016 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,KENNETH,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: PERIGNY,J  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: WHITAKER

DATE/TIME OF TEST: 09/19/2016 00:21:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:12 |
| Air Blank          | 0.000 | 00:13 |
| Control Test       | 0.079 | 00:13 |
| Air Blank          | 0.000 | 00:13 |
| Subject Sample # 1 | 0.000 | 00:16 |
| Breath Volume      | 1.652 | 00:16 |
| Air Blank          | 0.000 | 00:16 |
| Air Blank          | 0.000 | 00:18 |
| Subject Sample # 2 | 0.000 | 00:19 |
| Breath Volume      | 1.886 | 00:19 |
| Air Blank          | 0.000 | 00:20 |
| Control Test       | 0.078 | 00:20 |
| Air Blank          | 0.000 | 00:20 |
| Diagnostics Check  | OK    | 00:21 |

---

Data Download Date/Time: 09/29/2016 12:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,KENNETH,  
BREATH TEST OPERATOR AGENCY: North Miami Beach PD  
ARRESTING OFFICER: BAKER,KENNETH  
ARRESTING OFFICER AGENCY: North Miami Beach PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ARENASSANCHEZ

DATE/TIME OF TEST: 09/26/2016 02:49:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:46 |                      |
| Air Blank         | 0.000 | 02:46 |                      |
| Control Test      | 0.080 | 02:47 |                      |
| Air Blank         | 0.000 | 02:47 |                      |
| Subject Sample    | # 1   | 02:47 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:47 |                      |
| Air Blank         | 0.000 | 02:48 |                      |
| Control Test      | 0.079 | 02:48 |                      |
| Air Blank         | 0.000 | 02:49 |                      |
| Diagnostics Check | OK    | 02:49 |                      |

---

Data Download Date/Time: 09/29/2016 12:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000887

INSTRUMENT REGISTERED TO: CHATTAHOOCHEE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: POSTON, WILLIAM, A  
BREATH TEST OPERATOR AGENCY: Chattahoochee PD  
ARRESTING OFFICER: POSTON, WILLIAM  
ARRESTING OFFICER AGENCY: Chattahoochee PD  
LAST AGENCY INSPECTION: 08/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:36  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: MATHIS

DATE/TIME OF TEST: 09/03/2016 21:46:02

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 21:34 |                |
| Air Blank         |     | 0.000 | 21:34 |                |
| Control Test      |     | 0.077 | 21:35 |                |
| Air Blank         |     | 0.000 | 21:35 |                |
| Subject Sample    | # 1 | 0.182 | 21:38 | Volume Not Met |
| Breath Volume     |     | 0.050 | 21:38 |                |
| Air Blank         |     | 0.000 | 21:39 |                |
| Air Blank         |     | 0.000 | 21:41 |                |
| Subject Sample    | # 2 | 0.170 | 21:44 | Volume Not Met |
| Breath Volume     |     | 0.335 | 21:44 |                |
| Air Blank         |     | 0.000 | 21:45 |                |
| Control Test      |     | 0.077 | 21:45 |                |
| Air Blank         |     | 0.000 | 21:45 |                |
| Diagnostics Check |     | OK    | 21:46 |                |

---

Data Download Date/Time: 09/15/2016 19:42

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000891

INSTRUMENT REGISTERED TO: FWC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORBIN,BRITTON,L  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: PIFER,JUSTIN  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 08/23/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 28214080A1  
DRY GAS EXPIRY DATE: 11/05/2016  
VIOLATION CODE: BUI  
ARREST TIME: 16:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: SOMMERICH

DATE/TIME OF TEST: 09/20/2016 16:57:12

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 16:54 |                      |
| Air Blank         | 0.000 | 16:54 |                      |
| Control Test      | 0.079 | 16:55 |                      |
| Air Blank         | 0.000 | 16:55 |                      |
| Subject Sample    | # 1   | 16:55 | Subject Test Refused |
| Breath Volume     | 0.000 | 16:55 |                      |
| Air Blank         | 0.000 | 16:56 |                      |
| Control Test      | 0.080 | 16:56 |                      |
| Air Blank         | 0.000 | 16:57 |                      |
| Diagnostics Check | OK    | 16:57 |                      |

---

Data Download Date/Time: 09/22/2016 16:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,REID,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: ARBOGAST,JOSHUA  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1568991  
DRY GAS EXPIRY DATE: 10/01/2016  
VIOLATION CODE: BUI  
ARREST TIME: 15:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:27

SUBJECT'S LAST NAME: GALLEGO GIL

DATE/TIME OF TEST: 09/05/2016 16:00:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:53 |
| Air Blank          | 0.000 | 15:54 |
| Control Test       | 0.077 | 15:54 |
| Air Blank          | 0.000 | 15:55 |
| Subject Sample # 1 | 0.025 | 15:55 |
| Breath Volume      | 3.218 | 15:55 |
| Air Blank          | 0.000 | 15:56 |
| Air Blank          | 0.000 | 15:58 |
| Subject Sample # 2 | 0.025 | 15:58 |
| Breath Volume      | 3.847 | 15:58 |
| Air Blank          | 0.000 | 15:59 |
| Control Test       | 0.077 | 15:59 |
| Air Blank          | 0.000 | 16:00 |
| Diagnostics Check  | OK    | 16:00 |

---

Data Download Date/Time: 10/03/2016 12:19



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,REID,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: WHITE,REID  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1568991  
DRY GAS EXPIRY DATE: 10/01/2016  
VIOLATION CODE: BUI  
ARREST TIME: 18:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:36

SUBJECT'S LAST NAME: FRIEND

DATE/TIME OF TEST: 09/10/2016 19:02:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:56 |
| Air Blank          | 0.000 | 18:57 |
| Control Test       | 0.079 | 18:57 |
| Air Blank          | 0.000 | 18:57 |
| Subject Sample # 1 | 0.148 | 18:58 |
| Breath Volume      | 2.488 | 18:58 |
| Air Blank          | 0.000 | 18:59 |
| Air Blank          | 0.000 | 19:01 |
| Subject Sample # 2 | 0.151 | 19:01 |
| Breath Volume      | 2.375 | 19:01 |
| Air Blank          | 0.000 | 19:02 |
| Control Test       | 0.078 | 19:02 |
| Air Blank          | 0.000 | 19:02 |
| Diagnostics Check  | OK    | 19:02 |

---

Data Download Date/Time: 10/03/2016 12:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CLAFLIN,DASVID,S  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TEST,TEST  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26012080A1  
DRY GAS EXPIRY DATE: 11/01/2014  
VIOLATION CODE: OTHER  
ARREST TIME: 10:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 09/20/2016 21:20:31

---

|                    |       |       |                      |
|--------------------|-------|-------|----------------------|
| Diagnostics Check  | OK    | 21:13 |                      |
| Air Blank          | 0.000 | 21:14 |                      |
| Control Test       | 0.078 | 21:14 |                      |
| Air Blank          | 0.000 | 21:15 |                      |
| Subject Sample # 1 | 0.000 | 21:15 |                      |
| Breath Volume      | 3.585 | 21:15 |                      |
| Air Blank          | 0.000 | 21:16 |                      |
| Air Blank          | 0.000 | 21:18 |                      |
| Subject Sample # 2 |       | 21:19 | Subject Test Refused |
| Breath Volume      | 0.000 | 21:19 |                      |
| Air Blank          | 0.000 | 21:19 |                      |
| Control Test       | 0.077 | 21:19 |                      |
| Air Blank          | 0.000 | 21:20 |                      |
| Diagnostics Check  | OK    | 21:20 |                      |

---

Data Download Date/Time: 10/11/2016 12:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHERWOOD,JOHN,M  
BREATH TEST OPERATOR AGENCY: Sanibel PD  
ARRESTING OFFICER: HUBATCH,NICHOLAS  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05514080A1  
DRY GAS EXPIRY DATE: 04/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HARDER

DATE/TIME OF TEST: 09/02/2016 23:11:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:05 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.081 | 23:05 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 1 | 0.221 | 23:06 |
| Breath Volume      | 2.507 | 23:06 |
| Air Blank          | 0.000 | 23:07 |
| Air Blank          | 0.000 | 23:09 |
| Subject Sample # 2 | 0.215 | 23:09 |
| Breath Volume      | 3.750 | 23:09 |
| Air Blank          | 0.000 | 23:10 |
| Control Test       | 0.080 | 23:10 |
| Air Blank          | 0.000 | 23:10 |
| Diagnostics Check  | OK    | 23:11 |

---

Data Download Date/Time: 10/10/2016 17:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SHERWOOD,JOHN,M  
BREATH TEST OPERATOR AGENCY: Sanibel PD  
ARRESTING OFFICER: SHERWOOD,JOHN  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 05514080A1  
DRY GAS EXPIRY DATE: 04/01/2016  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CIUFFETTI

DATE/TIME OF TEST: 09/26/2016 23:25:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:19 |
| Air Blank          | 0.000 | 23:19 |
| Control Test       | 0.081 | 23:20 |
| Air Blank          | 0.000 | 23:20 |
| Subject Sample # 1 | 0.098 | 23:21 |
| Breath Volume      | 2.347 | 23:21 |
| Air Blank          | 0.000 | 23:22 |
| Air Blank          | 0.000 | 23:23 |
| Subject Sample # 2 | 0.091 | 23:24 |
| Breath Volume      | 3.644 | 23:24 |
| Air Blank          | 0.000 | 23:24 |
| Control Test       | 0.082 | 23:25 |
| Air Blank          | 0.000 | 23:25 |
| Diagnostics Check  | OK    | 23:25 |

---

Data Download Date/Time: 10/10/2016 17:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GALLAGHER, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GARROW

DATE/TIME OF TEST: 09/02/2016 03:22:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:16 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.079 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 1 | 0.118 | 03:18 |
| Breath Volume      | 1.519 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Subject Sample # 2 | 0.118 | 03:21 |
| Breath Volume      | 1.378 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.078 | 03:22 |
| Air Blank          | 0.000 | 03:22 |
| Diagnostics Check  | OK    | 03:22 |

---

Data Download Date/Time: 09/16/2016 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERONI, KRISTI, A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: NERONI, KRISTI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LAMOUR

DATE/TIME OF TEST: 09/03/2016 00:50:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Control Test       | 0.079 | 00:43 |
| Air Blank          | 0.000 | 00:44 |
| Subject Sample # 1 | 0.190 | 00:45 |
| Breath Volume      | 1.773 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 2 | 0.188 | 00:48 |
| Breath Volume      | 1.703 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Control Test       | 0.079 | 00:49 |
| Air Blank          | 0.000 | 00:49 |
| Diagnostics Check  | OK    | 00:50 |

---

Data Download Date/Time: 09/16/2016 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: O'GRADY,PATRICK,  
BREATH TEST OPERATOR AGENCY: Cape Coral PD  
ARRESTING OFFICER: O'GRADY,PATRICK  
ARRESTING OFFICER AGENCY: Cape Coral PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BLACKSTONE

DATE/TIME OF TEST: 09/03/2016 02:31:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:24 |
| Air Blank          | 0.000 | 02:25 |
| Control Test       | 0.079 | 02:25 |
| Air Blank          | 0.000 | 02:26 |
| Subject Sample # 1 | 0.040 | 02:26 |
| Breath Volume      | 2.312 | 02:26 |
| Air Blank          | 0.000 | 02:27 |
| Air Blank          | 0.000 | 02:29 |
| Subject Sample # 2 | 0.039 | 02:29 |
| Breath Volume      | 2.468 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Control Test       | 0.079 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Diagnostics Check  | OK    | 02:31 |

---

Data Download Date/Time: 09/16/2016 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONCALVES,GUSTAVO,R  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: GONCALVES,GUSTAVO  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 09/03/2016 02:47:52

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 02:36 |                    |
| Air Blank         | 0.000     | 02:36 |                    |
| Control Test      | 0.078     | 02:37 |                    |
| Air Blank         | 0.000     | 02:37 |                    |
| Subject Sample    | # 1 0.000 | 02:40 | Volume Not Met     |
| Breath Volume     | 1.085     | 02:40 |                    |
| Air Blank         | 0.000     | 02:41 |                    |
| Air Blank         | 0.000     | 02:43 |                    |
| Subject Sample    | # 2       | 02:46 | No Sample Provided |
| Breath Volume     | 0.000     | 02:46 |                    |
| Air Blank         | 0.000     | 02:46 |                    |
| Control Test      | 0.078     | 02:47 |                    |
| Air Blank         | 0.000     | 02:47 |                    |
| Diagnostics Check | OK        | 02:47 |                    |

---

Data Download Date/Time: 09/16/2016 10:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMYRNIOS,MANUEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 08:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: PEREZ GONZALEZ

DATE/TIME OF TEST: 09/03/2016 10:27:50

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:21 |
| Air Blank          | 0.000 | 10:21 |
| Control Test       | 0.078 | 10:21 |
| Air Blank          | 0.000 | 10:22 |
| Subject Sample # 1 | 0.187 | 10:23 |
| Breath Volume      | 1.531 | 10:23 |
| Air Blank          | 0.000 | 10:24 |
| Air Blank          | 0.000 | 10:25 |
| Subject Sample # 2 | 0.182 | 10:26 |
| Breath Volume      | 1.363 | 10:26 |
| Air Blank          | 0.000 | 10:26 |
| Control Test       | 0.077 | 10:27 |
| Air Blank          | 0.000 | 10:27 |
| Diagnostics Check  | OK    | 10:27 |

---

Data Download Date/Time: 09/16/2016 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,ANDREW,R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMITH,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/03/2016 23:19:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:12 |
| Air Blank          | 0.000 | 23:13 |
| Control Test       | 0.079 | 23:13 |
| Air Blank          | 0.000 | 23:13 |
| Subject Sample # 1 | 0.197 | 23:14 |
| Breath Volume      | 2.625 | 23:14 |
| Air Blank          | 0.000 | 23:15 |
| Air Blank          | 0.000 | 23:16 |
| Subject Sample # 2 | 0.193 | 23:17 |
| Breath Volume      | 2.351 | 23:17 |
| Air Blank          | 0.000 | 23:18 |
| Control Test       | 0.078 | 23:18 |
| Air Blank          | 0.000 | 23:19 |
| Diagnostics Check  | OK    | 23:19 |

---

Data Download Date/Time: 09/16/2016 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOULD,LEONARD,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,CARL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VEST

DATE/TIME OF TEST: 09/05/2016 03:07:14

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:00 |
| Air Blank         |     | 0.000 | 03:01 |
| Control Test      |     | 0.079 | 03:01 |
| Air Blank         |     | 0.000 | 03:01 |
| Subject Sample    | # 1 | 0.135 | 03:02 |
| Breath Volume     |     | 2.316 | 03:02 |
| Air Blank         |     | 0.000 | 03:02 |
| Air Blank         |     | 0.000 | 03:04 |
| Subject Sample    | # 2 | 0.135 | 03:05 |
| Breath Volume     |     | 1.992 | 03:05 |
| Air Blank         |     | 0.000 | 03:06 |
| Control Test      |     | 0.079 | 03:06 |
| Air Blank         |     | 0.000 | 03:07 |
| Diagnostics Check |     | OK    | 03:07 |

---

Data Download Date/Time: 09/16/2016 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,LISA,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MILLER,A  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 09/05/2016 05:25:44

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:16 |
| Air Blank          | 0.000 | 05:17 |
| Control Test       | 0.079 | 05:17 |
| Air Blank          | 0.000 | 05:18 |
| Subject Sample # 1 | 0.206 | 05:20 |
| Breath Volume      | 1.898 | 05:20 |
| Air Blank          | 0.000 | 05:21 |
| Air Blank          | 0.000 | 05:22 |
| Subject Sample # 2 | 0.210 | 05:24 |
| Breath Volume      | 1.230 | 05:24 |
| Air Blank          | 0.000 | 05:24 |
| Control Test       | 0.078 | 05:25 |
| Air Blank          | 0.000 | 05:25 |
| Diagnostics Check  | OK    | 05:25 |

---

Data Download Date/Time: 09/16/2016 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNCAN,N,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUNCAN,N  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FREZZA

DATE/TIME OF TEST: 09/07/2016 02:49:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:43 |
| Air Blank          | 0.000 | 02:43 |
| Control Test       | 0.079 | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 1 | 0.158 | 02:44 |
| Breath Volume      | 1.820 | 02:44 |
| Air Blank          | 0.000 | 02:45 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 2 | 0.160 | 02:47 |
| Breath Volume      | 1.347 | 02:47 |
| Air Blank          | 0.000 | 02:48 |
| Control Test       | 0.079 | 02:48 |
| Air Blank          | 0.000 | 02:49 |
| Diagnostics Check  | OK    | 02:49 |

---

Data Download Date/Time: 09/16/2016 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERONI, KRISTI, A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: NERONI, KRISTI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 13:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: MC BRIDE

DATE/TIME OF TEST: 09/07/2016 13:59:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:53 |
| Air Blank          | 0.000 | 13:53 |
| Control Test       | 0.080 | 13:53 |
| Air Blank          | 0.000 | 13:54 |
| Subject Sample # 1 | 0.000 | 13:55 |
| Breath Volume      | 3.468 | 13:55 |
| Air Blank          | 0.000 | 13:55 |
| Air Blank          | 0.000 | 13:57 |
| Subject Sample # 2 | 0.000 | 13:58 |
| Breath Volume      | 2.636 | 13:58 |
| Air Blank          | 0.000 | 13:58 |
| Control Test       | 0.080 | 13:59 |
| Air Blank          | 0.000 | 13:59 |
| Diagnostics Check  | OK    | 13:59 |

---

Data Download Date/Time: 09/16/2016 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MALDONADO,RYAN,A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: MALDONADO,RYAN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 09/09/2016 00:11:00

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:08 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Control Test      | 0.079 | 00:08 |                      |
| Air Blank         | 0.000 | 00:09 |                      |
| Subject Sample    | # 1   | 00:09 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:09 |                      |
| Air Blank         | 0.000 | 00:10 |                      |
| Control Test      | 0.079 | 00:10 |                      |
| Air Blank         | 0.000 | 00:10 |                      |
| Diagnostics Check | OK    | 00:11 |                      |

---

Data Download Date/Time: 09/16/2016 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNCAN,N,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUNCAN,N  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: KINT

DATE/TIME OF TEST: 09/10/2016 02:35:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:29 |
| Air Blank          | 0.000 | 02:29 |
| Control Test       | 0.079 | 02:29 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 1 | 0.100 | 02:31 |
| Breath Volume      | 3.125 | 02:31 |
| Air Blank          | 0.000 | 02:31 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 2 | 0.099 | 02:33 |
| Breath Volume      | 3.304 | 02:33 |
| Air Blank          | 0.000 | 02:34 |
| Control Test       | 0.079 | 02:34 |
| Air Blank          | 0.000 | 02:35 |
| Diagnostics Check  | OK    | 02:35 |

---

Data Download Date/Time: 09/16/2016 10:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNCAN,N,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUNCAN,N  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:52

SUBJECT'S LAST NAME: HEINS

DATE/TIME OF TEST: 09/14/2016 18:19:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:13 |
| Air Blank          | 0.000 | 18:13 |
| Control Test       | 0.079 | 18:14 |
| Air Blank          | 0.000 | 18:14 |
| Subject Sample # 1 | 0.000 | 18:15 |
| Breath Volume      | 1.457 | 18:15 |
| Air Blank          | 0.000 | 18:15 |
| Air Blank          | 0.000 | 18:17 |
| Subject Sample # 2 | 0.000 | 18:18 |
| Breath Volume      | 1.933 | 18:18 |
| Air Blank          | 0.000 | 18:18 |
| Control Test       | 0.079 | 18:18 |
| Air Blank          | 0.000 | 18:19 |
| Diagnostics Check  | OK    | 18:19 |

---

Data Download Date/Time: 09/16/2016 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNCAN,N,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUNCAN,N  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FAJARDO

DATE/TIME OF TEST: 09/15/2016 01:38:31

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Control Test       | 0.080 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Subject Sample # 1 | 0.194 | 01:34 |
| Breath Volume      | 1.527 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Air Blank          | 0.000 | 01:36 |
| Subject Sample # 2 | 0.193 | 01:36 |
| Breath Volume      | 1.734 | 01:36 |
| Air Blank          | 0.000 | 01:37 |
| Control Test       | 0.079 | 01:37 |
| Air Blank          | 0.000 | 01:38 |
| Diagnostics Check  | OK    | 01:38 |

---

Data Download Date/Time: 09/16/2016 10:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GALLAGHER, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 08/12/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RIVER RIOS

DATE/TIME OF TEST: 09/15/2016 02:47:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.080 | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.108 | 02:42 |
| Breath Volume      | 1.750 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.100 | 02:45 |
| Breath Volume      | 2.277 | 02:45 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.078 | 02:46 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 09/16/2016 10:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNCAN,N,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SCHILKE,C  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 09/17/2016 03:59:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:53 |
| Air Blank          | 0.000 | 03:53 |
| Control Test       | 0.079 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Subject Sample # 1 | 0.150 | 03:55 |
| Breath Volume      | 1.449 | 03:55 |
| Air Blank          | 0.000 | 03:55 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 2 | 0.153 | 03:58 |
| Breath Volume      | 1.289 | 03:58 |
| Air Blank          | 0.000 | 03:59 |
| Control Test       | 0.082 | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 10/11/2016 11:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYS,BRANDON,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MAYS,BRANDON  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: THELUS

DATE/TIME OF TEST: 09/19/2016 03:27:58

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 03:14 |                |
| Air Blank         |     | 0.000 | 03:14 |                |
| Control Test      |     | 0.079 | 03:15 |                |
| Air Blank         |     | 0.000 | 03:15 |                |
| Subject Sample    | # 1 | 0.114 | 03:16 |                |
| Breath Volume     |     | 1.394 | 03:16 |                |
| Air Blank         |     | 0.000 | 03:16 |                |
| Air Blank         |     | 0.000 | 03:18 |                |
| Subject Sample    | # 2 | 0.098 | 03:21 | Volume Not Met |
| Breath Volume     |     | 0.039 | 03:21 |                |
| Air Blank         |     | 0.000 | 03:22 |                |
| Air Blank         |     | 0.000 | 03:24 |                |
| Subject Sample    | # 3 | 0.116 | 03:26 |                |
| Breath Volume     |     | 2.980 | 03:26 |                |
| Air Blank         |     | 0.000 | 03:27 |                |
| Control Test      |     | 0.077 | 03:27 |                |
| Air Blank         |     | 0.000 | 03:27 |                |
| Diagnostics Check |     | OK    | 03:27 |                |

---

Data Download Date/Time: 10/11/2016 11:02

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAYS,BRANDON,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: MACEDO,KAREN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 09/19/2016 03:54:18

---

|                   |           |       |                      |
|-------------------|-----------|-------|----------------------|
| Diagnostics Check | OK        | 03:43 |                      |
| Air Blank         | 0.000     | 03:43 |                      |
| Control Test      | 0.079     | 03:43 |                      |
| Air Blank         | 0.000     | 03:44 |                      |
| Subject Sample    | # 1 0.021 | 03:47 | Volume Not Met       |
| Breath Volume     | 0.402     | 03:47 |                      |
| Air Blank         | 0.000     | 03:48 |                      |
| Air Blank         | 0.000     | 03:50 |                      |
| Subject Sample    | # 2       | 03:52 | Subject Test Refused |
| Breath Volume     | 1.039     | 03:52 |                      |
| Air Blank         | 0.000     | 03:52 |                      |
| Control Test      | 0.078     | 03:53 |                      |
| Air Blank         | 0.000     | 03:53 |                      |
| Diagnostics Check | Fail      | 03:53 | Voltage/Current Test |
| Air Blank         | 0.000     | 03:54 |                      |

---

Data Download Date/Time: 10/11/2016 11:02

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,ANDREW,R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: SMITH,ANDREW  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 09/19/2016 16:18:36

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 16:12 |
| Air Blank         |     | 0.000 | 16:12 |
| Control Test      |     | 0.079 | 16:13 |
| Air Blank         |     | 0.000 | 16:13 |
| Subject Sample    | # 1 | 0.234 | 16:14 |
| Breath Volume     |     | 2.316 | 16:14 |
| Air Blank         |     | 0.000 | 16:14 |
| Air Blank         |     | 0.000 | 16:16 |
| Subject Sample    | # 2 | 0.226 | 16:17 |
| Breath Volume     |     | 2.300 | 16:17 |
| Air Blank         |     | 0.000 | 16:17 |
| Control Test      |     | 0.079 | 16:17 |
| Air Blank         |     | 0.000 | 16:18 |
| Diagnostics Check |     | OK    | 16:18 |

---

Data Download Date/Time: 10/11/2016 11:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: TUBOIS, MICHAEL  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/19/2016 23:06:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:00 |
| Air Blank          | 0.000 | 23:01 |
| Control Test       | 0.080 | 23:01 |
| Air Blank          | 0.000 | 23:01 |
| Subject Sample # 1 | 0.000 | 23:02 |
| Breath Volume      | 1.109 | 23:02 |
| Air Blank          | 0.000 | 23:02 |
| Air Blank          | 0.000 | 23:04 |
| Subject Sample # 2 | 0.000 | 23:05 |
| Breath Volume      | 1.304 | 23:05 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.080 | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Diagnostics Check  | OK    | 23:06 |

---

Data Download Date/Time: 10/11/2016 11:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DUNCAN,N,  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DUNCAN,N  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: WHATLEY

DATE/TIME OF TEST: 09/20/2016 05:23:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:17 |
| Air Blank          | 0.000 | 05:17 |
| Control Test       | 0.080 | 05:18 |
| Air Blank          | 0.000 | 05:18 |
| Subject Sample # 1 | 0.000 | 05:19 |
| Breath Volume      | 2.867 | 05:19 |
| Air Blank          | 0.000 | 05:19 |
| Air Blank          | 0.000 | 05:21 |
| Subject Sample # 2 | 0.000 | 05:22 |
| Breath Volume      | 2.289 | 05:22 |
| Air Blank          | 0.000 | 05:22 |
| Control Test       | 0.079 | 05:22 |
| Air Blank          | 0.000 | 05:23 |
| Diagnostics Check  | OK    | 05:23 |

---

Data Download Date/Time: 10/11/2016 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GALLAGHER, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 09/21/2016 02:46:48

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:41 |
| Control Test       | 0.080 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.103 | 02:42 |
| Breath Volume      | 2.550 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:44 |
| Subject Sample # 2 | 0.106 | 02:45 |
| Breath Volume      | 1.789 | 02:45 |
| Air Blank          | 0.000 | 02:45 |
| Control Test       | 0.079 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Diagnostics Check  | OK    | 02:46 |

---

Data Download Date/Time: 10/11/2016 11:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROMAN, ANTHONY, C  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: RODRIGUEZ, C  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: SANCHEZ-PEREZ

DATE/TIME OF TEST: 09/23/2016 21:52:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:45 |
| Air Blank          | 0.000 | 21:46 |
| Control Test       | 0.080 | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Subject Sample # 1 | 0.237 | 21:47 |
| Breath Volume      | 1.972 | 21:47 |
| Air Blank          | 0.000 | 21:48 |
| Air Blank          | 0.000 | 21:50 |
| Subject Sample # 2 | 0.230 | 21:50 |
| Breath Volume      | 2.113 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Control Test       | 0.079 | 21:51 |
| Air Blank          | 0.000 | 21:52 |
| Diagnostics Check  | OK    | 21:52 |

---

Data Download Date/Time: 10/11/2016 11:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRENEMAN,LESA,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BRENEMAN,LESA  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROZYCKI

DATE/TIME OF TEST: 09/24/2016 01:39:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:33 |
| Air Blank          | 0.000 | 01:34 |
| Control Test       | 0.080 | 01:34 |
| Air Blank          | 0.000 | 01:34 |
| Subject Sample # 1 | 0.000 | 01:35 |
| Breath Volume      | 2.453 | 01:35 |
| Air Blank          | 0.000 | 01:35 |
| Air Blank          | 0.000 | 01:37 |
| Subject Sample # 2 | 0.000 | 01:38 |
| Breath Volume      | 2.312 | 01:38 |
| Air Blank          | 0.000 | 01:38 |
| Control Test       | 0.079 | 01:39 |
| Air Blank          | 0.000 | 01:39 |
| Diagnostics Check  | OK    | 01:39 |

---

Data Download Date/Time: 10/11/2016 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BOTELLO,ALEX,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: BOTELLO,ALEX  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 09/24/2016 13:20:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:14 |
| Air Blank          | 0.000 | 13:14 |
| Control Test       | 0.079 | 13:15 |
| Air Blank          | 0.000 | 13:15 |
| Subject Sample # 1 | 0.000 | 13:16 |
| Breath Volume      | 2.187 | 13:16 |
| Air Blank          | 0.000 | 13:16 |
| Air Blank          | 0.000 | 13:18 |
| Subject Sample # 2 | 0.000 | 13:19 |
| Breath Volume      | 1.933 | 13:19 |
| Air Blank          | 0.000 | 13:19 |
| Control Test       | 0.079 | 13:19 |
| Air Blank          | 0.000 | 13:20 |
| Diagnostics Check  | OK    | 13:20 |

---

Data Download Date/Time: 10/11/2016 11:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GALLAGHER, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: JUAN

DATE/TIME OF TEST: 09/26/2016 23:42:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:36 |
| Air Blank          | 0.000 | 23:36 |
| Control Test       | 0.080 | 23:37 |
| Air Blank          | 0.000 | 23:37 |
| Subject Sample # 1 | 0.164 | 23:38 |
| Breath Volume      | 1.703 | 23:38 |
| Air Blank          | 0.000 | 23:38 |
| Air Blank          | 0.000 | 23:40 |
| Subject Sample # 2 | 0.171 | 23:41 |
| Breath Volume      | 2.207 | 23:41 |
| Air Blank          | 0.000 | 23:41 |
| Control Test       | 0.080 | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Diagnostics Check  | OK    | 23:42 |

---

Data Download Date/Time: 10/11/2016 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERONI, KRISTIA  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: BARRINGER, BRIAN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 09216080A2  
DRY GAS EXPIRY DATE: 05/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:07  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: AYEMPEROUMAL

DATE/TIME OF TEST: 09/30/2016 08:29:18

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 08:16 |                |
| Air Blank         |     | 0.000 | 08:16 |                |
| Control Test      |     | 0.080 | 08:17 |                |
| Air Blank         |     | 0.000 | 08:17 |                |
| Subject Sample    | # 1 | 0.034 | 08:20 | Volume Not Met |
| Breath Volume     |     | 0.218 | 08:20 |                |
| Air Blank         |     | 0.000 | 08:21 |                |
| Air Blank         |     | 0.000 | 08:23 |                |
| Subject Sample    | # 2 | 0.031 | 08:24 |                |
| Breath Volume     |     | 1.972 | 08:24 |                |
| Air Blank         |     | 0.000 | 08:24 |                |
| Air Blank         |     | 0.000 | 08:26 |                |
| Subject Sample    | # 3 | 0.030 | 08:27 |                |
| Breath Volume     |     | 1.734 | 08:27 |                |
| Air Blank         |     | 0.000 | 08:28 |                |
| Control Test      |     | 0.080 | 08:28 |                |
| Air Blank         |     | 0.000 | 08:29 |                |
| Diagnostics Check |     | OK    | 08:29 |                |

---

Data Download Date/Time: 10/11/2016 11:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL, MITCHELL, B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MANSHADI, TAINA  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 09/02/2016 22:48:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.079 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 1 | 0.142 | 22:43 |
| Breath Volume      | 2.585 | 22:43 |
| Air Blank          | 0.000 | 22:44 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 2 | 0.142 | 22:47 |
| Breath Volume      | 3.503 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.080 | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Diagnostics Check  | OK    | 22:48 |

---

Data Download Date/Time: 10/03/2016 16:34



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL, MITCHELL, B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BELL, JOHN  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: PUENTE

DATE/TIME OF TEST: 09/05/2016 01:31:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Control Test       | 0.079 | 01:24 |
| Air Blank          | 0.000 | 01:24 |
| Subject Sample # 1 | 0.175 | 01:25 |
| Breath Volume      | 1.359 | 01:25 |
| Air Blank          | 0.000 | 01:26 |
| Air Blank          | 0.000 | 01:28 |
| Subject Sample # 2 | 0.183 | 01:30 |
| Breath Volume      | 1.910 | 01:30 |
| Air Blank          | 0.000 | 01:30 |
| Control Test       | 0.080 | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Diagnostics Check  | OK    | 01:31 |

---

Data Download Date/Time: 10/03/2016 16:34

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL, MITCHELL, B  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: VAZQUEZ, C  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/18/2016 20:55:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:49 |
| Air Blank          | 0.000 | 20:49 |
| Control Test       | 0.079 | 20:50 |
| Air Blank          | 0.000 | 20:50 |
| Subject Sample # 1 | 0.222 | 20:51 |
| Breath Volume      | 1.746 | 20:51 |
| Air Blank          | 0.000 | 20:51 |
| Air Blank          | 0.000 | 20:53 |
| Subject Sample # 2 | 0.202 | 20:54 |
| Breath Volume      | 1.156 | 20:54 |
| Air Blank          | 0.000 | 20:54 |
| Control Test       | 0.079 | 20:55 |
| Air Blank          | 0.000 | 20:55 |
| Diagnostics Check  | OK    | 20:55 |

---

Data Download Date/Time: 10/03/2016 16:34

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEVIS,KAYLA,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: GROENDYK,R  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROESLER

DATE/TIME OF TEST: 09/02/2016 03:00:03

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 02:45 |                    |
| Air Blank         | 0.000     | 02:45 |                    |
| Control Test      | 0.079     | 02:46 |                    |
| Air Blank         | 0.000     | 02:46 |                    |
| Subject Sample    | # 1 0.174 | 02:47 |                    |
| Breath Volume     | 1.335     | 02:47 |                    |
| Air Blank         | 0.000     | 02:48 |                    |
| Air Blank         | 0.000     | 02:49 |                    |
| Subject Sample    | # 2       | 02:53 | No Sample Provided |
| Breath Volume     | 0.000     | 02:53 |                    |
| Air Blank         | 0.000     | 02:53 |                    |
| Air Blank         | 0.000     | 02:55 |                    |
| Subject Sample    | # 3       | 02:58 | No Sample Provided |
| Breath Volume     | 0.000     | 02:58 |                    |
| Air Blank         | 0.000     | 02:59 |                    |
| Control Test      | 0.078     | 02:59 |                    |
| Air Blank         | 0.000     | 02:59 |                    |
| Diagnostics Check | OK        | 03:00 |                    |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VADNEY, MICHELE, M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CCSO, A  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: OTHER  
ARREST TIME: 18:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WAHLUND

DATE/TIME OF TEST: 09/02/2016 19:00:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:54 |
| Air Blank          | 0.000 | 18:54 |
| Control Test       | 0.078 | 18:55 |
| Air Blank          | 0.000 | 18:55 |
| Subject Sample # 1 | 0.230 | 18:56 |
| Breath Volume      | 2.355 | 18:56 |
| Air Blank          | 0.000 | 18:56 |
| Air Blank          | 0.000 | 18:58 |
| Subject Sample # 2 | 0.225 | 18:59 |
| Breath Volume      | 2.558 | 18:59 |
| Air Blank          | 0.000 | 18:59 |
| Control Test       | 0.077 | 19:00 |
| Air Blank          | 0.000 | 19:00 |
| Diagnostics Check  | OK    | 19:00 |

---

Data Download Date/Time: 10/03/2016 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARP,RYAN,M  
BREATH TEST OPERATOR AGENCY: Naples PD  
ARRESTING OFFICER: HARP,RYAN  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 09/03/2016 01:03:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:56 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.078 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Subject Sample # 1 | 0.084 | 00:58 |
| Breath Volume      | 2.781 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 2 | 0.083 | 01:01 |
| Breath Volume      | 2.058 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.078 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Diagnostics Check  | OK    | 01:03 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTILLO,NICKOLAUS,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MCCOY,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: GONZALEZ STUART

DATE/TIME OF TEST: 09/03/2016 03:54:21

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:46 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.078 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.157 | 03:49 |
| Breath Volume      | 2.089 | 03:49 |
| Air Blank          | 0.000 | 03:49 |
| Air Blank          | 0.000 | 03:51 |
| Subject Sample # 2 | 0.153 | 03:52 |
| Breath Volume      | 2.058 | 03:52 |
| Air Blank          | 0.000 | 03:53 |
| Control Test       | 0.077 | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Diagnostics Check  | OK    | 03:54 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUGGINS, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: HUGGINS, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: REGO

DATE/TIME OF TEST: 09/03/2016 18:39:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:30 |
| Air Blank          | 0.000 | 18:31 |
| Control Test       | 0.078 | 18:31 |
| Air Blank          | 0.000 | 18:32 |
| Subject Sample # 1 | 0.142 | 18:34 |
| Breath Volume      | 2.500 | 18:34 |
| Air Blank          | 0.000 | 18:34 |
| Air Blank          | 0.000 | 18:36 |
| Subject Sample # 2 | 0.129 | 18:37 |
| Breath Volume      | 2.656 | 18:37 |
| Air Blank          | 0.000 | 18:38 |
| Control Test       | 0.077 | 18:38 |
| Air Blank          | 0.000 | 18:39 |
| Diagnostics Check  | OK    | 18:39 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MENDOZA,URIEL,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MARTINEZ,E  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 09/03/2016 22:41:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.078 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Subject Sample # 1 | 0.038 | 22:36 |
| Breath Volume      | 2.289 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Air Blank          | 0.000 | 22:39 |
| Subject Sample # 2 | 0.042 | 22:39 |
| Breath Volume      | 2.160 | 22:39 |
| Air Blank          | 0.000 | 22:40 |
| Control Test       | 0.078 | 22:40 |
| Air Blank          | 0.000 | 22:40 |
| Diagnostics Check  | OK    | 22:41 |

---

Data Download Date/Time: 10/03/2016 12:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CHRIS,DAVID  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 12:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 09/04/2016 02:24:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:17 |
| Air Blank          | 0.000 | 02:18 |
| Control Test       | 0.077 | 02:18 |
| Air Blank          | 0.000 | 02:19 |
| Subject Sample # 1 | 0.162 | 02:19 |
| Breath Volume      | 1.867 | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Air Blank          | 0.000 | 02:22 |
| Subject Sample # 2 | 0.153 | 02:22 |
| Breath Volume      | 1.628 | 02:22 |
| Air Blank          | 0.000 | 02:23 |
| Control Test       | 0.078 | 02:23 |
| Air Blank          | 0.000 | 02:24 |
| Diagnostics Check  | OK    | 02:24 |

---

Data Download Date/Time: 10/03/2016 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASHLINE,C,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CARWIN,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: BERNSTEIN

DATE/TIME OF TEST: 09/04/2016 02:35:13

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 02:28 |
| Air Blank         |     | 0.000 | 02:29 |
| Control Test      |     | 0.078 | 02:29 |
| Air Blank         |     | 0.000 | 02:30 |
| Subject Sample    | # 1 | 0.180 | 02:30 |
| Breath Volume     |     | 2.269 | 02:30 |
| Air Blank         |     | 0.000 | 02:31 |
| Air Blank         |     | 0.000 | 02:33 |
| Subject Sample    | # 2 | 0.178 | 02:33 |
| Breath Volume     |     | 1.750 | 02:33 |
| Air Blank         |     | 0.000 | 02:34 |
| Control Test      |     | 0.077 | 02:34 |
| Air Blank         |     | 0.000 | 02:35 |
| Diagnostics Check |     | OK    | 02:35 |

---

Data Download Date/Time: 10/03/2016 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARP,RYAN,M  
BREATH TEST OPERATOR AGENCY: Naples PD  
ARRESTING OFFICER: HARP,RYAN  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 09/04/2016 03:19:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:13 |
| Air Blank          | 0.000 | 03:13 |
| Control Test       | 0.078 | 03:14 |
| Air Blank          | 0.000 | 03:14 |
| Subject Sample # 1 | 0.102 | 03:15 |
| Breath Volume      | 2.312 | 03:15 |
| Air Blank          | 0.000 | 03:15 |
| Air Blank          | 0.000 | 03:17 |
| Subject Sample # 2 | 0.100 | 03:18 |
| Breath Volume      | 2.503 | 03:18 |
| Air Blank          | 0.000 | 03:18 |
| Control Test       | 0.077 | 03:19 |
| Air Blank          | 0.000 | 03:19 |
| Diagnostics Check  | OK    | 03:19 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,REID,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BARRINGER,BRIAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: BUI  
ARREST TIME: 17:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: LEMOINE

DATE/TIME OF TEST: 09/04/2016 19:32:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:26 |
| Air Blank          | 0.000 | 19:26 |
| Control Test       | 0.078 | 19:27 |
| Air Blank          | 0.000 | 19:27 |
| Subject Sample # 1 | 0.191 | 19:28 |
| Breath Volume      | 2.824 | 19:28 |
| Air Blank          | 0.000 | 19:29 |
| Air Blank          | 0.000 | 19:30 |
| Subject Sample # 2 | 0.176 | 19:31 |
| Breath Volume      | 1.945 | 19:31 |
| Air Blank          | 0.000 | 19:32 |
| Control Test       | 0.078 | 19:32 |
| Air Blank          | 0.000 | 19:32 |
| Diagnostics Check  | OK    | 19:32 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WHITE,REID,B  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: KLEIS,KEVIN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: BUI  
ARREST TIME: 19:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HARDING

DATE/TIME OF TEST: 09/04/2016 21:06:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:00 |
| Air Blank          | 0.000 | 21:00 |
| Control Test       | 0.078 | 21:01 |
| Air Blank          | 0.000 | 21:01 |
| Subject Sample # 1 | 0.070 | 21:02 |
| Breath Volume      | 3.378 | 21:02 |
| Air Blank          | 0.000 | 21:02 |
| Air Blank          | 0.000 | 21:04 |
| Subject Sample # 2 | 0.071 | 21:05 |
| Breath Volume      | 4.371 | 21:05 |
| Air Blank          | 0.000 | 21:05 |
| Control Test       | 0.079 | 21:06 |
| Air Blank          | 0.000 | 21:06 |
| Diagnostics Check  | OK    | 21:06 |

---

Data Download Date/Time: 10/03/2016 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARSH, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: FOLDEN, THOMAS  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 09:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:53

SUBJECT'S LAST NAME: TIRELLA

DATE/TIME OF TEST: 09/05/2016 10:25:17

---

|                    |       |       |                 |
|--------------------|-------|-------|-----------------|
| Diagnostics Check  | OK    | 10:19 |                 |
| Air Blank          | 0.000 | 10:19 |                 |
| Control Test       | 0.078 | 10:20 |                 |
| Air Blank          | 0.000 | 10:20 |                 |
| Subject Sample # 1 | 0.000 | 10:21 |                 |
| Breath Volume      | 1.476 | 10:21 |                 |
| Air Blank          | 0.000 | 10:22 |                 |
| Air Blank          |       | 10:23 | Improper Sample |
| Air Blank          | 0.000 | 10:24 |                 |
| Control Test       | 0.078 | 10:24 |                 |
| Air Blank          | 0.000 | 10:25 |                 |
| Diagnostics Check  | OK    | 10:25 |                 |

---

Data Download Date/Time: 10/03/2016 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARSH,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: FOLDEN,THOMAS  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 09:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 09:55

SUBJECT'S LAST NAME: TIRELLA

DATE/TIME OF TEST: 09/05/2016 10:35:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:29 |
| Air Blank          | 0.000 | 10:29 |
| Control Test       | 0.077 | 10:30 |
| Air Blank          | 0.000 | 10:30 |
| Subject Sample # 1 | 0.000 | 10:31 |
| Breath Volume      | 1.148 | 10:31 |
| Air Blank          | 0.000 | 10:31 |
| Air Blank          | 0.000 | 10:33 |
| Subject Sample # 2 | 0.000 | 10:34 |
| Breath Volume      | 1.500 | 10:34 |
| Air Blank          | 0.000 | 10:34 |
| Control Test       | 0.078 | 10:34 |
| Air Blank          | 0.000 | 10:35 |
| Diagnostics Check  | OK    | 10:35 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUONPANE,VINCENT,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SACKMANN,A  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: ESMELDER

DATE/TIME OF TEST: 09/05/2016 22:20:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:13 |
| Air Blank          | 0.000 | 22:13 |
| Control Test       | 0.078 | 22:14 |
| Air Blank          | 0.000 | 22:14 |
| Subject Sample # 1 | 0.200 | 22:15 |
| Breath Volume      | 3.800 | 22:15 |
| Air Blank          | 0.000 | 22:16 |
| Air Blank          | 0.000 | 22:18 |
| Subject Sample # 2 | 0.187 | 22:18 |
| Breath Volume      | 3.027 | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Control Test       | 0.078 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Diagnostics Check  | OK    | 22:20 |

---

Data Download Date/Time: 10/03/2016 12:27



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PATTERSON,D,S  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PROBATION,OFFICER  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 00:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:13

SUBJECT'S LAST NAME: OTLOWSKI

DATE/TIME OF TEST: 09/06/2016 10:42:39

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 10:36 |
| Air Blank         |     | 0.000 | 10:36 |
| Control Test      |     | 0.078 | 10:37 |
| Air Blank         |     | 0.000 | 10:37 |
| Subject Sample    | # 1 | 0.208 | 10:38 |
| Breath Volume     |     | 2.496 | 10:38 |
| Air Blank         |     | 0.000 | 10:38 |
| Air Blank         |     | 0.000 | 10:40 |
| Subject Sample    | # 2 | 0.197 | 10:41 |
| Breath Volume     |     | 2.148 | 10:41 |
| Air Blank         |     | 0.000 | 10:41 |
| Control Test      |     | 0.079 | 10:42 |
| Air Blank         |     | 0.000 | 10:42 |
| Diagnostics Check |     | OK    | 10:42 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARRILLO,XAVIER,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DIXON,P  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: YOUNGS

DATE/TIME OF TEST: 09/06/2016 22:42:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Control Test       | 0.078 | 22:36 |
| Air Blank          | 0.000 | 22:37 |
| Subject Sample # 1 | 0.276 | 22:37 |
| Breath Volume      | 1.746 | 22:37 |
| Air Blank          | 0.000 | 22:38 |
| Air Blank          | 0.000 | 22:40 |
| Subject Sample # 2 | 0.273 | 22:40 |
| Breath Volume      | 2.382 | 22:40 |
| Air Blank          | 0.000 | 22:41 |
| Control Test       | 0.078 | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Diagnostics Check  | OK    | 22:42 |

---

Data Download Date/Time: 10/03/2016 12:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS,TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS,TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/08/2016 03:31:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:24 |
| Control Test       | 0.077 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 1 | 0.170 | 03:27 |
| Breath Volume      | 1.144 | 03:27 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Subject Sample # 2 | 0.150 | 03:30 |
| Breath Volume      | 1.261 | 03:30 |
| Air Blank          | 0.000 | 03:30 |
| Control Test       | 0.076 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Diagnostics Check  | OK    | 03:31 |

---

Data Download Date/Time: 10/03/2016 12:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPOLO,JOSH,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PUKA,M  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 09/08/2016 03:59:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:53 |
| Air Blank          | 0.000 | 03:54 |
| Control Test       | 0.077 | 03:54 |
| Air Blank          | 0.000 | 03:54 |
| Subject Sample # 1 | 0.384 | 03:55 |
| Breath Volume      | 1.554 | 03:55 |
| Air Blank          | 0.000 | 03:56 |
| Air Blank          | 0.000 | 03:57 |
| Subject Sample # 2 | 0.385 | 03:58 |
| Breath Volume      | 3.335 | 03:58 |
| Air Blank          | 0.000 | 03:58 |
| Control Test       | 0.077 | 03:59 |
| Air Blank          | 0.000 | 03:59 |
| Diagnostics Check  | OK    | 03:59 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCDUFFIE,B.,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: PROBATION,N/A  
ARRESTING OFFICER AGENCY: OTHER County  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 10:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: LESTERNAJARIAN

DATE/TIME OF TEST: 09/08/2016 10:44:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:37 |
| Air Blank          | 0.000 | 10:38 |
| Control Test       | 0.078 | 10:38 |
| Air Blank          | 0.000 | 10:39 |
| Subject Sample # 1 | 0.079 | 10:40 |
| Breath Volume      | 2.007 | 10:40 |
| Air Blank          | 0.000 | 10:40 |
| Air Blank          | 0.000 | 10:42 |
| Subject Sample # 2 | 0.084 | 10:43 |
| Breath Volume      | 1.867 | 10:43 |
| Air Blank          | 0.000 | 10:43 |
| Control Test       | 0.078 | 10:44 |
| Air Blank          | 0.000 | 10:44 |
| Diagnostics Check  | OK    | 10:44 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARMENT,KERRY,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: COMBS,S  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LEONA

DATE/TIME OF TEST: 09/08/2016 21:31:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:24 |
| Air Blank          | 0.000 | 21:24 |
| Control Test       | 0.078 | 21:24 |
| Air Blank          | 0.000 | 21:25 |
| Subject Sample # 1 | 0.024 | 21:27 |
| Breath Volume      | 1.660 | 21:27 |
| Air Blank          | 0.000 | 21:27 |
| Air Blank          | 0.000 | 21:29 |
| Subject Sample # 2 | 0.022 | 21:30 |
| Breath Volume      | 1.742 | 21:30 |
| Air Blank          | 0.000 | 21:30 |
| Control Test       | 0.077 | 21:30 |
| Air Blank          | 0.000 | 21:31 |
| Diagnostics Check  | OK    | 21:31 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: DUGENER

DATE/TIME OF TEST: 09/09/2016 02:40:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:32 |
| Air Blank          | 0.000 | 02:32 |
| Control Test       | 0.078 | 02:32 |
| Air Blank          | 0.000 | 02:33 |
| Subject Sample # 1 | 0.174 | 02:34 |
| Breath Volume      | 1.687 | 02:34 |
| Air Blank          | 0.000 | 02:34 |
| Air Blank          | 0.000 | 02:36 |
| Subject Sample # 2 | 0.173 | 02:39 |
| Breath Volume      | 2.113 | 02:39 |
| Air Blank          | 0.000 | 02:39 |
| Control Test       | 0.078 | 02:39 |
| Air Blank          | 0.000 | 02:40 |
| Diagnostics Check  | OK    | 02:40 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:39

SUBJECT'S LAST NAME: MIERENDORFF

DATE/TIME OF TEST: 09/09/2016 19:09:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:03 |
| Air Blank          | 0.000 | 19:03 |
| Control Test       | 0.079 | 19:03 |
| Air Blank          | 0.000 | 19:04 |
| Subject Sample # 1 | 0.286 | 19:05 |
| Breath Volume      | 2.257 | 19:05 |
| Air Blank          | 0.000 | 19:05 |
| Air Blank          | 0.000 | 19:07 |
| Subject Sample # 2 | 0.269 | 19:08 |
| Breath Volume      | 2.605 | 19:08 |
| Air Blank          | 0.000 | 19:09 |
| Control Test       | 0.078 | 19:09 |
| Air Blank          | 0.000 | 19:09 |
| Diagnostics Check  | OK    | 19:09 |

---

Data Download Date/Time: 10/03/2016 12:28



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURTSINGER,CODY,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SUDANO,B  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: SYFRETT

DATE/TIME OF TEST: 09/10/2016 21:04:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:58 |
| Air Blank          | 0.000 | 20:58 |
| Control Test       | 0.078 | 20:58 |
| Air Blank          | 0.000 | 20:59 |
| Subject Sample # 1 | 0.127 | 20:59 |
| Breath Volume      | 3.203 | 20:59 |
| Air Blank          | 0.000 | 21:00 |
| Air Blank          | 0.000 | 21:02 |
| Subject Sample # 2 | 0.144 | 21:02 |
| Breath Volume      | 2.886 | 21:02 |
| Air Blank          | 0.000 | 21:03 |
| Control Test       | 0.079 | 21:03 |
| Air Blank          | 0.000 | 21:04 |
| Diagnostics Check  | OK    | 21:04 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SALDIVAR

DATE/TIME OF TEST: 09/13/2016 00:50:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Control Test       | 0.078 | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 1 | 0.191 | 00:46 |
| Breath Volume      | 2.917 | 00:46 |
| Air Blank          | 0.000 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Subject Sample # 2 | 0.192 | 00:49 |
| Breath Volume      | 2.933 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Control Test       | 0.078 | 00:50 |
| Air Blank          | 0.000 | 00:50 |
| Diagnostics Check  | OK    | 00:50 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELMORE,D,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: ELMORE,D  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 15:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: VALENTINE JR

DATE/TIME OF TEST: 09/13/2016 16:17:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:10 |
| Air Blank          | 0.000 | 16:11 |
| Control Test       | 0.079 | 16:11 |
| Air Blank          | 0.000 | 16:12 |
| Subject Sample # 1 | 0.131 | 16:12 |
| Breath Volume      | 3.203 | 16:12 |
| Air Blank          | 0.000 | 16:13 |
| Air Blank          | 0.000 | 16:14 |
| Subject Sample # 2 | 0.121 | 16:15 |
| Breath Volume      | 2.863 | 16:15 |
| Air Blank          | 0.000 | 16:16 |
| Control Test       | 0.079 | 16:16 |
| Air Blank          | 0.000 | 16:16 |
| Diagnostics Check  | OK    | 16:17 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DECKER,DAVID,P  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: FORTUNE,T  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: KERCHNER

DATE/TIME OF TEST: 09/13/2016 22:24:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:17 |
| Air Blank          | 0.000 | 22:18 |
| Control Test       | 0.078 | 22:18 |
| Air Blank          | 0.000 | 22:19 |
| Subject Sample # 1 | 0.086 | 22:19 |
| Breath Volume      | 1.515 | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Air Blank          | 0.000 | 22:22 |
| Subject Sample # 2 | 0.083 | 22:23 |
| Breath Volume      | 1.503 | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.079 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Diagnostics Check  | OK    | 22:24 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KAPLAN,S,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KAPLAN,S  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NUTTER

DATE/TIME OF TEST: 09/14/2016 01:57:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:51 |
| Air Blank          | 0.000 | 01:51 |
| Control Test       | 0.078 | 01:51 |
| Air Blank          | 0.000 | 01:52 |
| Subject Sample # 1 | 0.265 | 01:53 |
| Breath Volume      | 2.300 | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Air Blank          | 0.000 | 01:55 |
| Subject Sample # 2 | 0.247 | 01:56 |
| Breath Volume      | 1.355 | 01:56 |
| Air Blank          | 0.000 | 01:56 |
| Control Test       | 0.077 | 01:57 |
| Air Blank          | 0.000 | 01:57 |
| Diagnostics Check  | OK    | 01:57 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS,TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS,TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: SHEARN

DATE/TIME OF TEST: 09/14/2016 04:28:49

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:22 |
| Air Blank          | 0.000 | 04:22 |
| Control Test       | 0.078 | 04:22 |
| Air Blank          | 0.000 | 04:23 |
| Subject Sample # 1 | 0.000 | 04:24 |
| Breath Volume      | 2.437 | 04:24 |
| Air Blank          | 0.000 | 04:25 |
| Air Blank          | 0.000 | 04:27 |
| Subject Sample # 2 | 0.000 | 04:27 |
| Breath Volume      | 1.777 | 04:27 |
| Air Blank          | 0.000 | 04:27 |
| Control Test       | 0.078 | 04:28 |
| Air Blank          | 0.000 | 04:28 |
| Diagnostics Check  | OK    | 04:28 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DECKER,STEPHEN,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DECKER,S  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: EGALITE

DATE/TIME OF TEST: 09/16/2016 18:46:19

---

|                    |       |       |                  |
|--------------------|-------|-------|------------------|
| Diagnostics Check  | OK    | 18:40 |                  |
| Air Blank          | 0.000 | 18:41 |                  |
| Control Test       | 0.078 | 18:41 |                  |
| Air Blank          | 0.000 | 18:41 |                  |
| Subject Sample # 1 | 0.239 | 18:42 |                  |
| Breath Volume      | 3.792 | 18:42 |                  |
| Air Blank          | 0.000 | 18:43 |                  |
| Air Blank          | 0.000 | 18:44 |                  |
| Subject Sample # 2 |       | 18:45 | Slope Not Met    |
| Breath Volume      | 1.785 | 18:45 |                  |
| Air Blank          |       | 18:45 | Sequence Aborted |
| Air Blank          | 0.000 | 18:46 |                  |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHRISTIANSEN, DAVID, M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CHRISTIANSEN, DAVID  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: FINNEY

DATE/TIME OF TEST: 09/17/2016 18:45:43

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:39 |
| Air Blank          | 0.000 | 18:39 |
| Control Test       | 0.078 | 18:40 |
| Air Blank          | 0.000 | 18:40 |
| Subject Sample # 1 | 0.000 | 18:41 |
| Breath Volume      | 3.140 | 18:41 |
| Air Blank          | 0.000 | 18:41 |
| Air Blank          | 0.000 | 18:43 |
| Subject Sample # 2 | 0.000 | 18:44 |
| Breath Volume      | 3.312 | 18:44 |
| Air Blank          | 0.000 | 18:44 |
| Control Test       | 0.078 | 18:45 |
| Air Blank          | 0.000 | 18:45 |
| Diagnostics Check  | OK    | 18:45 |

---

Data Download Date/Time: 10/03/2016 12:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LYONS,SHANE,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SMITH,L  
ARRESTING OFFICER AGENCY: Marco Island PD  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 09/17/2016 23:59:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:53 |
| Air Blank          | 0.000 | 23:53 |
| Control Test       | 0.078 | 23:53 |
| Air Blank          | 0.000 | 23:54 |
| Subject Sample # 1 | 0.107 | 23:55 |
| Breath Volume      | 2.093 | 23:55 |
| Air Blank          | 0.000 | 23:55 |
| Air Blank          | 0.000 | 23:57 |
| Subject Sample # 2 | 0.113 | 23:57 |
| Breath Volume      | 2.699 | 23:57 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.078 | 23:58 |
| Air Blank          | 0.000 | 23:59 |
| Diagnostics Check  | OK    | 23:59 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARP,RYAN,M  
BREATH TEST OPERATOR AGENCY: Naples PD  
ARRESTING OFFICER: HARP,RYAN  
ARRESTING OFFICER AGENCY: Naples PD  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: JONTZA ANEZ

DATE/TIME OF TEST: 09/18/2016 01:01:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:55 |
| Air Blank          | 0.000 | 00:55 |
| Control Test       | 0.078 | 00:56 |
| Air Blank          | 0.000 | 00:56 |
| Subject Sample # 1 | 0.118 | 00:57 |
| Breath Volume      | 1.808 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Air Blank          | 0.000 | 00:59 |
| Subject Sample # 2 | 0.116 | 01:00 |
| Breath Volume      | 1.812 | 01:00 |
| Air Blank          | 0.000 | 01:00 |
| Control Test       | 0.078 | 01:01 |
| Air Blank          | 0.000 | 01:01 |
| Diagnostics Check  | OK    | 01:01 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CERVANTES HERNANDEZ

DATE/TIME OF TEST: 09/18/2016 03:30:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:23 |
| Air Blank          | 0.000 | 03:23 |
| Control Test       | 0.078 | 03:24 |
| Air Blank          | 0.000 | 03:24 |
| Subject Sample # 1 | 0.259 | 03:25 |
| Breath Volume      | 3.335 | 03:25 |
| Air Blank          | 0.000 | 03:25 |
| Air Blank          | 0.000 | 03:27 |
| Subject Sample # 2 | 0.260 | 03:28 |
| Breath Volume      | 3.457 | 03:28 |
| Air Blank          | 0.000 | 03:29 |
| Control Test       | 0.079 | 03:29 |
| Air Blank          | 0.000 | 03:29 |
| Diagnostics Check  | OK    | 03:30 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NATSON,ANDRAE,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BUTLER,DAVID  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: ZDINAK

DATE/TIME OF TEST: 09/18/2016 03:49:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:43 |
| Air Blank          | 0.000 | 03:43 |
| Control Test       | 0.078 | 03:44 |
| Air Blank          | 0.000 | 03:44 |
| Subject Sample # 1 | 0.244 | 03:45 |
| Breath Volume      | 1.421 | 03:45 |
| Air Blank          | 0.000 | 03:45 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 2 | 0.249 | 03:48 |
| Breath Volume      | 1.300 | 03:48 |
| Air Blank          | 0.000 | 03:48 |
| Control Test       | 0.078 | 03:48 |
| Air Blank          | 0.000 | 03:49 |
| Diagnostics Check  | OK    | 03:49 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARP,K,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BORCHERS,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: MC FADDEN

DATE/TIME OF TEST: 09/18/2016 21:00:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:54 |
| Air Blank          | 0.000 | 20:54 |
| Control Test       | 0.078 | 20:55 |
| Air Blank          | 0.000 | 20:55 |
| Subject Sample # 1 | 0.319 | 20:55 |
| Breath Volume      | 1.386 | 20:55 |
| Air Blank          | 0.000 | 20:56 |
| Air Blank          | 0.000 | 20:58 |
| Subject Sample # 2 | 0.318 | 20:59 |
| Breath Volume      | 3.097 | 20:59 |
| Air Blank          | 0.000 | 20:59 |
| Control Test       | 0.078 | 21:00 |
| Air Blank          | 0.000 | 21:00 |
| Diagnostics Check  | OK    | 21:00 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: QUIGLEY, MICHAEL, R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: HENDRICKS, L  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 16:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: KNIEBBE

DATE/TIME OF TEST: 09/20/2016 17:24:00

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 17:16 |
| Air Blank         |     | 0.000 | 17:17 |
| Control Test      |     | 0.079 | 17:17 |
| Air Blank         |     | 0.000 | 17:17 |
| Subject Sample    | # 1 | 0.185 | 17:19 |
| Breath Volume     |     | 2.410 | 17:19 |
| Air Blank         |     | 0.000 | 17:20 |
| Air Blank         |     | 0.000 | 17:22 |
| Subject Sample    | # 2 | 0.190 | 17:22 |
| Breath Volume     |     | 2.289 | 17:22 |
| Air Blank         |     | 0.000 | 17:23 |
| Control Test      |     | 0.079 | 17:23 |
| Air Blank         |     | 0.000 | 17:23 |
| Diagnostics Check |     | OK    | 17:24 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,T,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORWIN,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 09/20/2016 22:25:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Control Test       | 0.078 | 22:20 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 1 | 0.000 | 22:21 |
| Breath Volume      | 2.601 | 22:21 |
| Air Blank          | 0.000 | 22:21 |
| Air Blank          | 0.000 | 22:23 |
| Subject Sample # 2 | 0.000 | 22:24 |
| Breath Volume      | 2.722 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Control Test       | 0.078 | 22:25 |
| Air Blank          | 0.000 | 22:25 |
| Diagnostics Check  | OK    | 22:25 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS,TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS,TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: BRANSTRATOR

DATE/TIME OF TEST: 09/22/2016 22:07:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:00 |
| Air Blank          | 0.000 | 22:01 |
| Control Test       | 0.078 | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Subject Sample # 1 | 0.096 | 22:02 |
| Breath Volume      | 3.878 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Air Blank          | 0.000 | 22:05 |
| Subject Sample # 2 | 0.097 | 22:05 |
| Breath Volume      | 4.121 | 22:05 |
| Air Blank          | 0.000 | 22:06 |
| Control Test       | 0.078 | 22:06 |
| Air Blank          | 0.000 | 22:06 |
| Diagnostics Check  | OK    | 22:07 |

---

Data Download Date/Time: 10/03/2016 12:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GEORGE,SHANNON,N  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: LONG,M  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/22/2016 23:46:37

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:39 |
| Air Blank          | 0.000 | 23:40 |
| Control Test       | 0.078 | 23:40 |
| Air Blank          | 0.000 | 23:41 |
| Subject Sample # 1 | 0.049 | 23:42 |
| Breath Volume      | 1.304 | 23:42 |
| Air Blank          | 0.000 | 23:42 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 2 | 0.045 | 23:45 |
| Breath Volume      | 1.207 | 23:45 |
| Air Blank          | 0.000 | 23:45 |
| Control Test       | 0.078 | 23:46 |
| Air Blank          | 0.000 | 23:46 |
| Diagnostics Check  | OK    | 23:46 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DICARLO,JACOB,T  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MCCORMICK,PATRICK  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: CANCINO

DATE/TIME OF TEST: 09/25/2016 04:56:05

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:47 |
| Air Blank         |     | 0.000 | 04:48 |
| Control Test      |     | 0.078 | 04:48 |
| Air Blank         |     | 0.000 | 04:49 |
| Subject Sample    | # 1 | 0.203 | 04:50 |
| Breath Volume     |     | 2.765 | 04:50 |
| Air Blank         |     | 0.000 | 04:51 |
| Air Blank         |     | 0.000 | 04:53 |
| Subject Sample    | # 2 | 0.192 | 04:54 |
| Breath Volume     |     | 2.015 | 04:54 |
| Air Blank         |     | 0.000 | 04:55 |
| Control Test      |     | 0.075 | 04:55 |
| Air Blank         |     | 0.000 | 04:56 |
| Diagnostics Check |     | OK    | 04:56 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRATT,NICHOLAS,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: RODRIGUEZ,VICTOR  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: STURZENEGGER

DATE/TIME OF TEST: 09/25/2016 18:48:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:42 |
| Air Blank          | 0.000 | 18:42 |
| Control Test       | 0.078 | 18:42 |
| Air Blank          | 0.000 | 18:43 |
| Subject Sample # 1 | 0.186 | 18:43 |
| Breath Volume      | 2.589 | 18:43 |
| Air Blank          | 0.000 | 18:44 |
| Air Blank          | 0.000 | 18:46 |
| Subject Sample # 2 | 0.189 | 18:46 |
| Breath Volume      | 2.269 | 18:46 |
| Air Blank          | 0.000 | 18:47 |
| Control Test       | 0.077 | 18:47 |
| Air Blank          | 0.000 | 18:48 |
| Diagnostics Check  | OK    | 18:48 |

---

Data Download Date/Time: 10/03/2016 12:28

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORWIN,JONATHAN  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 19615080A4  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: SANJUAN-SANTIAGO

DATE/TIME OF TEST: 09/26/2016 05:04:15

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:57 |
| Air Blank          | 0.000 | 04:58 |
| Control Test       | 0.078 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Subject Sample # 1 | 0.064 | 04:59 |
| Breath Volume      | 1.593 | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Air Blank          | 0.000 | 05:02 |
| Subject Sample # 2 | 0.064 | 05:02 |
| Breath Volume      | 2.125 | 05:02 |
| Air Blank          | 0.000 | 05:03 |
| Control Test       | 0.078 | 05:03 |
| Air Blank          | 0.000 | 05:04 |
| Diagnostics Check  | OK    | 05:04 |

---

Data Download Date/Time: 10/03/2016 12:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DECKER,STEPHEN,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CCSO,CCSO  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 11315080A1  
DRY GAS EXPIRY DATE: 05/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 18:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: EGALITE

DATE/TIME OF TEST: 09/16/2016 18:57:23

---

|                   |           |       |               |
|-------------------|-----------|-------|---------------|
| Diagnostics Check | OK        | 18:50 |               |
| Air Blank         | 0.000     | 18:51 |               |
| Control Test      | 0.079     | 18:51 |               |
| Air Blank         | 0.000     | 18:52 |               |
| Subject Sample    | # 1 0.180 | 18:52 |               |
| Breath Volume     | 3.000     | 18:52 |               |
| Air Blank         | 0.000     | 18:53 |               |
| Air Blank         | 0.000     | 18:55 |               |
| Subject Sample    | # 2       | 18:55 | Slope Not Met |
| Breath Volume     | 0.890     | 18:55 |               |
| Air Blank         | 0.000     | 18:56 |               |
| Control Test      | 0.079     | 18:56 |               |
| Air Blank         | 0.000     | 18:57 |               |
| Diagnostics Check | OK        | 18:57 |               |

---

Data Download Date/Time: 10/03/2016 12:21

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GONZALEZ,OFFICER  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: ADMIN  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/01/2016 01:37:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.079 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.199 | 01:33 |
| Breath Volume      | 3.171 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.190 | 01:36 |
| Breath Volume      | 3.542 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.079 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLOS,LINDSAY,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GONZALEZ,OFFICER  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: ADMIN  
ARREST TIME: 22:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/01/2016 05:07:32

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 05:01 |
| Air Blank          | 0.000 | 05:01 |
| Control Test       | 0.080 | 05:02 |
| Air Blank          | 0.000 | 05:02 |
| Subject Sample # 1 | 0.142 | 05:03 |
| Breath Volume      | 2.472 | 05:03 |
| Air Blank          | 0.000 | 05:03 |
| Air Blank          | 0.000 | 05:05 |
| Subject Sample # 2 | 0.131 | 05:06 |
| Breath Volume      | 2.062 | 05:06 |
| Air Blank          | 0.000 | 05:06 |
| Control Test       | 0.079 | 05:06 |
| Air Blank          | 0.000 | 05:07 |
| Diagnostics Check  | OK    | 05:07 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GONZALEZ,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: ADMIN  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/01/2016 08:11:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 08:05 |
| Air Blank          | 0.000 | 08:05 |
| Control Test       | 0.079 | 08:06 |
| Air Blank          | 0.000 | 08:06 |
| Subject Sample # 1 | 0.074 | 08:07 |
| Breath Volume      | 1.406 | 08:07 |
| Air Blank          | 0.000 | 08:07 |
| Air Blank          | 0.000 | 08:09 |
| Subject Sample # 2 | 0.086 | 08:10 |
| Breath Volume      | 2.382 | 08:10 |
| Air Blank          | 0.000 | 08:11 |
| Control Test       | 0.080 | 08:11 |
| Air Blank          | 0.000 | 08:11 |
| Diagnostics Check  | OK    | 08:11 |

---

Data Download Date/Time: 09/30/2016 02:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GONZALEZ,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: ADMIN  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:11

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/01/2016 10:38:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:32 |
| Air Blank          | 0.000 | 10:32 |
| Control Test       | 0.079 | 10:32 |
| Air Blank          | 0.000 | 10:33 |
| Subject Sample # 1 | 0.051 | 10:34 |
| Breath Volume      | 1.921 | 10:34 |
| Air Blank          | 0.000 | 10:34 |
| Air Blank          | 0.000 | 10:36 |
| Subject Sample # 2 | 0.051 | 10:37 |
| Breath Volume      | 2.132 | 10:37 |
| Air Blank          | 0.000 | 10:37 |
| Control Test       | 0.079 | 10:37 |
| Air Blank          | 0.000 | 10:38 |
| Diagnostics Check  | OK    | 10:38 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GONZALEZ,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: ADMIN  
ARREST TIME: 20:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 09/01/2016 11:51:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:45 |
| Air Blank          | 0.000 | 11:45 |
| Control Test       | 0.079 | 11:46 |
| Air Blank          | 0.000 | 11:46 |
| Subject Sample # 1 | 0.039 | 11:46 |
| Breath Volume      | 2.394 | 11:46 |
| Air Blank          | 0.000 | 11:47 |
| Air Blank          | 0.000 | 11:49 |
| Subject Sample # 2 | 0.036 | 11:49 |
| Breath Volume      | 2.113 | 11:49 |
| Air Blank          | 0.000 | 11:50 |
| Control Test       | 0.079 | 11:50 |
| Air Blank          | 0.000 | 11:51 |
| Diagnostics Check  | OK    | 11:51 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,DEPUTY,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: URSITTI,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 15:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 09/01/2016 16:22:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 16:15 |
| Air Blank          | 0.000 | 16:16 |
| Control Test       | 0.080 | 16:16 |
| Air Blank          | 0.000 | 16:16 |
| Subject Sample # 1 | 0.201 | 16:17 |
| Breath Volume      | 2.625 | 16:17 |
| Air Blank          | 0.000 | 16:18 |
| Air Blank          | 0.000 | 16:20 |
| Subject Sample # 2 | 0.193 | 16:20 |
| Breath Volume      | 2.031 | 16:20 |
| Air Blank          | 0.000 | 16:21 |
| Control Test       | 0.079 | 16:21 |
| Air Blank          | 0.000 | 16:22 |
| Diagnostics Check  | OK    | 16:22 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: LUTHNER,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: LANIER

DATE/TIME OF TEST: 09/02/2016 02:20:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:14 |
| Air Blank          | 0.000 | 02:14 |
| Control Test       | 0.080 | 02:14 |
| Air Blank          | 0.000 | 02:15 |
| Subject Sample # 1 | 0.000 | 02:16 |
| Breath Volume      | 1.972 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Air Blank          | 0.000 | 02:18 |
| Subject Sample # 2 | 0.000 | 02:19 |
| Breath Volume      | 2.460 | 02:19 |
| Air Blank          | 0.000 | 02:19 |
| Control Test       | 0.080 | 02:19 |
| Air Blank          | 0.000 | 02:20 |
| Diagnostics Check  | OK    | 02:20 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLOS,LINDSAY,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JIMENEZ,OFFICER  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: INCITTI

DATE/TIME OF TEST: 09/02/2016 03:27:26

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:20 |
| Air Blank          | 0.000 | 03:21 |
| Control Test       | 0.080 | 03:21 |
| Air Blank          | 0.000 | 03:21 |
| Subject Sample # 1 | 0.257 | 03:22 |
| Breath Volume      | 1.523 | 03:22 |
| Air Blank          | 0.000 | 03:23 |
| Air Blank          | 0.000 | 03:25 |
| Subject Sample # 2 | 0.239 | 03:25 |
| Breath Volume      | 1.628 | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.080 | 03:26 |
| Air Blank          | 0.000 | 03:27 |
| Diagnostics Check  | OK    | 03:27 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SHIVELY,DOMENICA  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: EYSTER

DATE/TIME OF TEST: 09/02/2016 22:29:54

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:23 |
| Air Blank          | 0.000 | 22:23 |
| Control Test       | 0.080 | 22:24 |
| Air Blank          | 0.000 | 22:24 |
| Subject Sample # 1 | 0.000 | 22:25 |
| Breath Volume      | 2.976 | 22:25 |
| Air Blank          | 0.000 | 22:25 |
| Air Blank          | 0.000 | 22:27 |
| Subject Sample # 2 | 0.000 | 22:28 |
| Breath Volume      | 2.980 | 22:28 |
| Air Blank          | 0.000 | 22:29 |
| Control Test       | 0.080 | 22:29 |
| Air Blank          | 0.000 | 22:29 |
| Diagnostics Check  | OK    | 22:29 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: DIAZ,ELIO  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: PAYNE

DATE/TIME OF TEST: 09/02/2016 23:20:52

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:14 |
| Air Blank          | 0.000 | 23:14 |
| Control Test       | 0.079 | 23:15 |
| Air Blank          | 0.000 | 23:15 |
| Subject Sample # 1 | 0.000 | 23:16 |
| Breath Volume      | 2.347 | 23:16 |
| Air Blank          | 0.000 | 23:16 |
| Air Blank          | 0.000 | 23:18 |
| Subject Sample # 2 | 0.000 | 23:19 |
| Breath Volume      | 2.628 | 23:19 |
| Air Blank          | 0.000 | 23:19 |
| Control Test       | 0.080 | 23:20 |
| Air Blank          | 0.000 | 23:20 |
| Diagnostics Check  | OK    | 23:20 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: BAILEY,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: HIGBEE

DATE/TIME OF TEST: 09/03/2016 00:11:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.079 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.176 | 00:07 |
| Breath Volume      | 2.117 | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 2 | 0.181 | 00:09 |
| Breath Volume      | 2.328 | 00:09 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.079 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 09/30/2016 02:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HEALEY,D,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: RICHARDSON,B  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 09:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 09/03/2016 12:42:50

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 12:36 |
| Air Blank         |     | 0.000 | 12:36 |
| Control Test      |     | 0.079 | 12:36 |
| Air Blank         |     | 0.000 | 12:37 |
| Subject Sample    | # 1 | 0.208 | 12:37 |
| Breath Volume     |     | 1.214 | 12:37 |
| Air Blank         |     | 0.000 | 12:38 |
| Air Blank         |     | 0.000 | 12:40 |
| Subject Sample    | # 2 | 0.208 | 12:41 |
| Breath Volume     |     | 2.261 | 12:41 |
| Air Blank         |     | 0.000 | 12:41 |
| Control Test      |     | 0.079 | 12:42 |
| Air Blank         |     | 0.000 | 12:42 |
| Diagnostics Check |     | OK    | 12:42 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GIRARD,JOSH  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:46

SUBJECT'S LAST NAME: ATKINSON

DATE/TIME OF TEST: 09/04/2016 21:13:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:06 |
| Air Blank          | 0.000 | 21:07 |
| Control Test       | 0.079 | 21:07 |
| Air Blank          | 0.000 | 21:08 |
| Subject Sample # 1 | 0.169 | 21:09 |
| Breath Volume      | 1.285 | 21:09 |
| Air Blank          | 0.000 | 21:09 |
| Air Blank          | 0.000 | 21:11 |
| Subject Sample # 2 | 0.175 | 21:12 |
| Breath Volume      | 2.175 | 21:12 |
| Air Blank          | 0.000 | 21:12 |
| Control Test       | 0.079 | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Diagnostics Check  | OK    | 21:13 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: KLEIVER,J  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 09/09/2016 01:03:25

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.079 | 00:57 |
| Air Blank          | 0.000 | 00:58 |
| Subject Sample # 1 | 0.176 | 00:58 |
| Breath Volume      | 2.421 | 00:58 |
| Air Blank          | 0.000 | 00:59 |
| Air Blank          | 0.000 | 01:01 |
| Subject Sample # 2 | 0.176 | 01:01 |
| Breath Volume      | 2.597 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Control Test       | 0.079 | 01:02 |
| Air Blank          | 0.000 | 01:03 |
| Diagnostics Check  | OK    | 01:03 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: JIMENEZ,OFFICER  
ARRESTING OFFICER AGENCY: Punta Gorda PD  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GRIZZ

DATE/TIME OF TEST: 09/10/2016 00:49:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:42 |
| Air Blank          | 0.000 | 00:42 |
| Control Test       | 0.079 | 00:43 |
| Air Blank          | 0.000 | 00:43 |
| Subject Sample # 1 | 0.261 | 00:44 |
| Breath Volume      | 1.941 | 00:44 |
| Air Blank          | 0.000 | 00:45 |
| Air Blank          | 0.000 | 00:47 |
| Subject Sample # 2 | 0.276 | 00:47 |
| Breath Volume      | 2.707 | 00:47 |
| Air Blank          | 0.000 | 00:48 |
| Control Test       | 0.079 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Diagnostics Check  | OK    | 00:49 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: ROMAN,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 09/10/2016 02:32:49

---

|                    |       |       |                 |
|--------------------|-------|-------|-----------------|
| Diagnostics Check  | OK    | 02:27 |                 |
| Air Blank          | 0.000 | 02:27 |                 |
| Control Test       | 0.079 | 02:27 |                 |
| Air Blank          | 0.000 | 02:28 |                 |
| Subject Sample # 1 | 0.153 | 02:29 |                 |
| Breath Volume      | 3.222 | 02:29 |                 |
| Air Blank          | 0.000 | 02:29 |                 |
| Air Blank          | 0.000 | 02:31 |                 |
| Subject Sample # 2 | 0.145 | 02:32 |                 |
| Breath Volume      | 2.203 | 02:32 |                 |
| Air Blank          |       | 02:32 | Improper Sample |
| Air Blank          | 0.000 | 02:32 |                 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: SARVER,SCOTT  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: JENISON

DATE/TIME OF TEST: 09/11/2016 02:01:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:53 |
| Air Blank          | 0.000 | 01:53 |
| Control Test       | 0.079 | 01:54 |
| Air Blank          | 0.000 | 01:54 |
| Subject Sample # 1 | 0.184 | 01:57 |
| Breath Volume      | 1.492 | 01:57 |
| Air Blank          | 0.000 | 01:57 |
| Air Blank          | 0.000 | 01:59 |
| Subject Sample # 2 | 0.181 | 01:59 |
| Breath Volume      | 1.582 | 01:59 |
| Air Blank          | 0.000 | 02:00 |
| Control Test       | 0.078 | 02:00 |
| Air Blank          | 0.000 | 02:01 |
| Diagnostics Check  | OK    | 02:01 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLOS,LINDSAY,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HAUSCHILD,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DIRRIGLE

DATE/TIME OF TEST: 09/15/2016 00:07:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:58 |
| Air Blank          | 0.000 | 23:58 |
| Control Test       | 0.079 | 23:59 |
| Air Blank          | 0.000 | 23:59 |
| Subject Sample # 1 | 0.071 | 00:02 |
| Breath Volume      | 1.585 | 00:02 |
| Air Blank          | 0.000 | 00:03 |
| Air Blank          | 0.000 | 00:05 |
| Subject Sample # 2 | 0.073 | 00:06 |
| Breath Volume      | 1.164 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Control Test       | 0.080 | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Diagnostics Check  | OK    | 00:07 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLOS,LINDSAY,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: HAUSCHILD,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: ADMIN  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: DIRRIGLE

DATE/TIME OF TEST: 09/15/2016 04:18:06

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 04:03 |                    |
| Air Blank         | 0.000 | 04:04 |                    |
| Control Test      | 0.079 | 04:04 |                    |
| Air Blank         | 0.000 | 04:05 |                    |
| Subject Sample    | # 1   | 04:08 | No Sample Provided |
| Breath Volume     | 0.000 | 04:08 |                    |
| Air Blank         | 0.000 | 04:08 |                    |
| Air Blank         | 0.000 | 04:10 |                    |
| Subject Sample    | # 2   | 0.014 | 04:12              |
| Breath Volume     | 2.472 | 04:12 |                    |
| Air Blank         | 0.000 | 04:12 |                    |
| Air Blank         | 0.000 | 04:14 |                    |
| Subject Sample    | # 3   | 0.012 | 04:16              |
| Breath Volume     | 2.105 | 04:16 |                    |
| Air Blank         | 0.000 | 04:17 |                    |
| Control Test      | 0.080 | 04:17 |                    |
| Air Blank         | 0.000 | 04:18 |                    |
| Diagnostics Check | OK    | 04:18 |                    |

---

Data Download Date/Time: 09/30/2016 02:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,J,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: CAMPBELL,J  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 07:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:30

SUBJECT'S LAST NAME: DIRRIGLE

DATE/TIME OF TEST: 09/15/2016 08:04:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:57 |
| Air Blank          | 0.000 | 07:58 |
| Control Test       | 0.080 | 07:58 |
| Air Blank          | 0.000 | 07:58 |
| Subject Sample # 1 | 0.000 | 07:59 |
| Breath Volume      | 2.609 | 07:59 |
| Air Blank          | 0.000 | 08:00 |
| Air Blank          | 0.000 | 08:02 |
| Subject Sample # 2 | 0.000 | 08:02 |
| Breath Volume      | 2.851 | 08:02 |
| Air Blank          | 0.000 | 08:03 |
| Control Test       | 0.079 | 08:03 |
| Air Blank          | 0.000 | 08:04 |
| Diagnostics Check  | OK    | 08:04 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: MURRAY,JAMES  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 09/17/2016 01:15:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.080 | 01:10 |
| Air Blank          | 0.000 | 01:10 |
| Subject Sample # 1 | 0.159 | 01:11 |
| Breath Volume      | 2.652 | 01:11 |
| Air Blank          | 0.000 | 01:11 |
| Air Blank          | 0.000 | 01:13 |
| Subject Sample # 2 | 0.161 | 01:14 |
| Breath Volume      | 2.625 | 01:14 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.079 | 01:14 |
| Air Blank          | 0.000 | 01:15 |
| Diagnostics Check  | OK    | 01:15 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FREITAS,ANTHONY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: REINECK

DATE/TIME OF TEST: 09/22/2016 03:37:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:31 |
| Air Blank          | 0.000 | 03:32 |
| Control Test       | 0.079 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Subject Sample # 1 | 0.170 | 03:33 |
| Breath Volume      | 1.257 | 03:33 |
| Air Blank          | 0.000 | 03:33 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 2 | 0.171 | 03:36 |
| Breath Volume      | 1.253 | 03:36 |
| Air Blank          | 0.000 | 03:36 |
| Control Test       | 0.079 | 03:37 |
| Air Blank          | 0.000 | 03:37 |
| Diagnostics Check  | OK    | 03:37 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLOS,LINDSAY,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GRAVEL,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 09/24/2016 04:11:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:04 |
| Air Blank          | 0.000 | 04:04 |
| Control Test       | 0.079 | 04:05 |
| Air Blank          | 0.000 | 04:05 |
| Subject Sample # 1 | 0.151 | 04:06 |
| Breath Volume      | 1.859 | 04:06 |
| Air Blank          | 0.000 | 04:06 |
| Air Blank          | 0.000 | 04:08 |
| Subject Sample # 2 | 0.150 | 04:09 |
| Breath Volume      | 1.203 | 04:09 |
| Air Blank          | 0.000 | 04:10 |
| Control Test       | 0.080 | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Diagnostics Check  | OK    | 04:11 |

---

Data Download Date/Time: 09/30/2016 02:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FISHER,SCOTT,R  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: GALLEG0,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 09/24/2016 22:10:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:03 |
| Air Blank          | 0.000 | 22:03 |
| Control Test       | 0.079 | 22:04 |
| Air Blank          | 0.000 | 22:04 |
| Subject Sample # 1 | 0.156 | 22:05 |
| Breath Volume      | 1.871 | 22:05 |
| Air Blank          | 0.000 | 22:05 |
| Air Blank          | 0.000 | 22:07 |
| Subject Sample # 2 | 0.166 | 22:08 |
| Breath Volume      | 2.519 | 22:08 |
| Air Blank          | 0.000 | 22:09 |
| Control Test       | 0.080 | 22:09 |
| Air Blank          | 0.000 | 22:10 |
| Diagnostics Check  | OK    | 22:10 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KLOS,LINDSAY,M  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: FREITAS,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 09/25/2016 21:17:25

---

|                   |           |       |                    |
|-------------------|-----------|-------|--------------------|
| Diagnostics Check | OK        | 21:00 |                    |
| Air Blank         | 0.000     | 21:01 |                    |
| Control Test      | 0.079     | 21:01 |                    |
| Air Blank         | 0.000     | 21:01 |                    |
| Subject Sample    | # 1 0.202 | 21:04 |                    |
| Breath Volume     | 1.457     | 21:04 |                    |
| Air Blank         | 0.000     | 21:05 |                    |
| Air Blank         | 0.000     | 21:07 |                    |
| Subject Sample    | # 2       | 21:10 | No Sample Provided |
| Breath Volume     | 0.000     | 21:10 |                    |
| Air Blank         | 0.000     | 21:10 |                    |
| Air Blank         | 0.000     | 21:12 |                    |
| Subject Sample    | # 3       | 21:16 | No Sample Provided |
| Breath Volume     | 0.000     | 21:16 |                    |
| Air Blank         | 0.000     | 21:16 |                    |
| Control Test      | 0.076     | 21:16 |                    |
| Air Blank         | 0.000     | 21:17 |                    |
| Diagnostics Check | OK        | 21:17 |                    |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EGAN,MELISSA,  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: KNAPP,ERIC  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 09/28/2016 03:00:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.079 | 02:54 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 1 | 0.074 | 02:55 |
| Breath Volume      | 1.152 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Subject Sample # 2 | 0.081 | 02:58 |
| Breath Volume      | 1.621 | 02:58 |
| Air Blank          | 0.000 | 02:59 |
| Control Test       | 0.079 | 02:59 |
| Air Blank          | 0.000 | 02:59 |
| Diagnostics Check  | OK    | 03:00 |

---

Data Download Date/Time: 09/30/2016 02:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BURROWS,ALBERT,L  
BREATH TEST OPERATOR AGENCY: Charlotte County SO  
ARRESTING OFFICER: STAUB,DEPUTY  
ARRESTING OFFICER AGENCY: Charlotte County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 19615080A2  
DRY GAS EXPIRY DATE: 09/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 12:59  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:32

SUBJECT'S LAST NAME: MCCARTHY

DATE/TIME OF TEST: 09/28/2016 13:59:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 13:52 |
| Air Blank          | 0.000 | 13:53 |
| Control Test       | 0.078 | 13:53 |
| Air Blank          | 0.000 | 13:54 |
| Subject Sample # 1 | 0.000 | 13:54 |
| Breath Volume      | 1.378 | 13:54 |
| Air Blank          | 0.000 | 13:55 |
| Air Blank          | 0.000 | 13:57 |
| Subject Sample # 2 | 0.000 | 13:58 |
| Breath Volume      | 1.238 | 13:58 |
| Air Blank          | 0.000 | 13:58 |
| Control Test       | 0.080 | 13:58 |
| Air Blank          | 0.000 | 13:59 |
| Diagnostics Check  | OK    | 13:59 |

---

Data Download Date/Time: 09/30/2016 02:29



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKUPSKI,RYAN,T  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: OKUPSKI,R  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04216080A3  
DRY GAS EXPIRY DATE: 03/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MILLER-WALDRON

DATE/TIME OF TEST: 09/03/2016 02:47:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:40 |
| Air Blank          | 0.000 | 02:40 |
| Control Test       | 0.080 | 02:41 |
| Air Blank          | 0.000 | 02:41 |
| Subject Sample # 1 | 0.139 | 02:42 |
| Breath Volume      | 1.757 | 02:42 |
| Air Blank          | 0.000 | 02:43 |
| Air Blank          | 0.000 | 02:45 |
| Subject Sample # 2 | 0.129 | 02:46 |
| Breath Volume      | 1.519 | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.080 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Diagnostics Check  | OK    | 02:47 |

---

Data Download Date/Time: 09/30/2016 15:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000948

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OKUPSKI,RYAN,T  
BREATH TEST OPERATOR AGENCY: Glades County SO  
ARRESTING OFFICER: OKUPSKI,R  
ARRESTING OFFICER AGENCY: Glades County SO  
LAST AGENCY INSPECTION: 08/31/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 04216080A3  
DRY GAS EXPIRY DATE: 03/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WALDORF

DATE/TIME OF TEST: 09/13/2016 20:47:56

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:45 |                      |
| Air Blank         | 0.000 | 20:45 |                      |
| Control Test      | 0.079 | 20:45 |                      |
| Air Blank         | 0.000 | 20:46 |                      |
| Subject Sample    | # 1   | 20:46 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:46 |                      |
| Air Blank         | 0.000 | 20:47 |                      |
| Control Test      | 0.079 | 20:47 |                      |
| Air Blank         | 0.000 | 20:47 |                      |
| Diagnostics Check | OK    | 20:47 |                      |

---

Data Download Date/Time: 09/30/2016 15:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LITTLE,RODERIAL,A

BREATH TEST OPERATOR AGENCY: Havana PD

ARRESTING OFFICER: LITTLE,RODERIAL

ARRESTING OFFICER AGENCY: Havana PD

LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 22:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 09/26/2016 00:08:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:04 |                      |
| Air Blank         | 0.000 | 00:04 |                      |
| Control Test      | 0.077 | 00:05 |                      |
| Air Blank         | 0.000 | 00:05 |                      |
| Subject Sample    | # 1   | 00:07 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:07 |                      |
| Air Blank         | 0.000 | 00:07 |                      |
| Control Test      | 0.076 | 00:08 |                      |
| Air Blank         | 0.000 | 00:08 |                      |
| Diagnostics Check | OK    | 00:08 |                      |

---

Data Download Date/Time: 10/03/2016 11:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,RYAN,A  
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)  
ARRESTING OFFICER: LEWIS,PAMELA  
ARRESTING OFFICER AGENCY: Apalachicola PD  
LAST AGENCY INSPECTION: 08/20/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: OTHER  
ARREST TIME: 11:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: VOYLES

DATE/TIME OF TEST: 09/20/2016 11:54:26

---

|                   |       |       |                    |
|-------------------|-------|-------|--------------------|
| Diagnostics Check | OK    | 11:43 |                    |
| Air Blank         | 0.000 | 11:43 |                    |
| Control Test      | 0.081 | 11:43 |                    |
| Air Blank         | 0.000 | 11:44 |                    |
| Subject Sample    | # 1   | 11:47 | No Sample Provided |
| Breath Volume     | 0.000 | 11:47 |                    |
| Air Blank         | 0.000 | 11:47 |                    |
| Air Blank         | 0.000 | 11:49 |                    |
| Subject Sample    | # 2   | 11:53 | No Sample Provided |
| Breath Volume     | 0.000 | 11:53 |                    |
| Air Blank         | 0.000 | 11:53 |                    |
| Control Test      | 0.081 | 11:53 |                    |
| Air Blank         | 0.000 | 11:54 |                    |
| Diagnostics Check | OK    | 11:54 |                    |

---

Data Download Date/Time: 09/21/2016 17:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARD,JAMES,H  
BREATH TEST OPERATOR AGENCY: Franklin County SO  
ARRESTING OFFICER: ANDERSON,KASEY  
ARRESTING OFFICER AGENCY: FFWCC (Panama City)  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: LOCKWOOD

DATE/TIME OF TEST: 09/26/2016 18:44:10

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:38 |
| Air Blank          | 0.000 | 18:38 |
| Control Test       | 0.078 | 18:38 |
| Air Blank          | 0.000 | 18:39 |
| Subject Sample # 1 | 0.195 | 18:39 |
| Breath Volume      | 3.457 | 18:39 |
| Air Blank          | 0.000 | 18:40 |
| Air Blank          | 0.000 | 18:42 |
| Subject Sample # 2 | 0.188 | 18:42 |
| Breath Volume      | 2.914 | 18:42 |
| Air Blank          | 0.000 | 18:43 |
| Control Test       | 0.078 | 18:43 |
| Air Blank          | 0.000 | 18:44 |
| Diagnostics Check  | OK    | 18:44 |

---

Data Download Date/Time: 09/27/2016 15:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REVELL,DENNIS,E  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: REVELL,DENNIS  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1841081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 09:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:46

SUBJECT'S LAST NAME: GAY

DATE/TIME OF TEST: 09/30/2016 11:15:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 11:08 |
| Air Blank          | 0.000 | 11:09 |
| Control Test       | 0.077 | 11:09 |
| Air Blank          | 0.000 | 11:10 |
| Subject Sample # 1 | 0.000 | 11:10 |
| Breath Volume      | 3.683 | 11:10 |
| Air Blank          | 0.000 | 11:11 |
| Air Blank          | 0.000 | 11:13 |
| Subject Sample # 2 | 0.000 | 11:14 |
| Breath Volume      | 1.531 | 11:14 |
| Air Blank          | 0.000 | 11:14 |
| Control Test       | 0.078 | 11:14 |
| Air Blank          | 0.000 | 11:15 |
| Diagnostics Check  | OK    | 11:15 |

---

Data Download Date/Time: 10/29/2016 02:44

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,BERNABE,  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: HERNANDEZ,BERNABE  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 06:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/02/2016 07:44:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:38 |
| Air Blank          | 0.000 | 07:38 |
| Control Test       | 0.078 | 07:38 |
| Air Blank          | 0.000 | 07:39 |
| Subject Sample # 1 | 0.217 | 07:39 |
| Breath Volume      | 1.117 | 07:39 |
| Air Blank          | 0.000 | 07:40 |
| Air Blank          | 0.000 | 07:42 |
| Subject Sample # 2 | 0.218 | 07:43 |
| Breath Volume      | 1.156 | 07:43 |
| Air Blank          | 0.000 | 07:43 |
| Control Test       | 0.078 | 07:44 |
| Air Blank          | 0.000 | 07:44 |
| Diagnostics Check  | OK    | 07:44 |

---

Data Download Date/Time: 09/08/2016 10:57

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,BERNABE,  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: HERNANDEZ,BERNABE  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 06:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 09/02/2016 08:01:01

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 07:52 |            |
| Air Blank         | 0.000     | 07:52 |            |
| Control Test      | 0.077     | 07:53 |            |
| Air Blank         | 0.000     | 07:53 |            |
| Subject Sample    | # 1 0.202 | 07:55 |            |
| Breath Volume     | 1.152     | 07:55 |            |
| Air Blank         | 0.000     | 07:56 |            |
| Air Blank         | 0.000     | 07:58 |            |
| Subject Sample    | # 2 0.201 | 07:59 |            |
| Breath Volume     | 1.148     | 07:59 |            |
| Air Blank         | 0.000     | 07:59 |            |
| Control Test      | 0.076     | 08:00 |            |
| Air Blank         |           | 08:00 | RFI Detect |
| Air Blank         | 0.000     | 08:01 |            |

---

Data Download Date/Time: 09/08/2016 10:57



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,BERNABE,  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: HERNANDEZ,BERNABE  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 13:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 09/08/2016 14:09:37

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 14:05 |            |
| Air Blank          | 0.000 | 14:06 |            |
| Control Test       | 0.078 | 14:06 |            |
| Air Blank          | 0.000 | 14:06 |            |
| Subject Sample # 1 | 0.134 | 14:07 |            |
| Breath Volume      | 1.636 | 14:07 |            |
| Air Blank          |       | 14:08 | RFI Detect |
| Air Blank          | 0.000 | 14:08 |            |
| Control Test       | 0.079 | 14:09 |            |
| Air Blank          | 0.000 | 14:09 |            |
| Diagnostics Check  | OK    | 14:09 |            |

---

Data Download Date/Time: 01/04/2017 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HERNANDEZ,BERNABE,  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: HERNANDEZ,BERNABE  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 13:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 13:44

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 09/08/2016 14:18:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:11 |
| Air Blank          | 0.000 | 14:12 |
| Control Test       | 0.078 | 14:12 |
| Air Blank          | 0.000 | 14:13 |
| Subject Sample # 1 | 0.133 | 14:13 |
| Breath Volume      | 1.539 | 14:13 |
| Air Blank          | 0.000 | 14:14 |
| Air Blank          | 0.000 | 14:15 |
| Subject Sample # 2 | 0.127 | 14:16 |
| Breath Volume      | 1.648 | 14:16 |
| Air Blank          | 0.000 | 14:17 |
| Control Test       | 0.078 | 14:17 |
| Air Blank          | 0.000 | 14:18 |
| Diagnostics Check  | OK    | 14:18 |

---

Data Download Date/Time: 01/04/2017 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITTMAN,JAMES,H  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: PITTMAN,JAMES  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 17:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 09/08/2016 18:54:14

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:47 |
| Air Blank          | 0.000 | 18:48 |
| Control Test       | 0.080 | 18:48 |
| Air Blank          | 0.000 | 18:48 |
| Subject Sample # 1 | 0.141 | 18:49 |
| Breath Volume      | 3.695 | 18:49 |
| Air Blank          | 0.000 | 18:50 |
| Air Blank          | 0.000 | 18:52 |
| Subject Sample # 2 | 0.137 | 18:52 |
| Breath Volume      | 3.792 | 18:52 |
| Air Blank          | 0.000 | 18:53 |
| Control Test       | 0.078 | 18:53 |
| Air Blank          | 0.000 | 18:54 |
| Diagnostics Check  | OK    | 18:54 |

---

Data Download Date/Time: 01/04/2017 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: SODERHOLM,CHELSEA  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HENN

DATE/TIME OF TEST: 09/09/2016 02:08:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:01 |
| Air Blank          | 0.000 | 02:01 |
| Control Test       | 0.080 | 02:02 |
| Air Blank          | 0.000 | 02:02 |
| Subject Sample # 1 | 0.101 | 02:03 |
| Breath Volume      | 1.324 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Air Blank          | 0.000 | 02:06 |
| Subject Sample # 2 | 0.095 | 02:06 |
| Breath Volume      | 1.191 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Control Test       | 0.079 | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Diagnostics Check  | OK    | 02:08 |

---

Data Download Date/Time: 01/04/2017 10:17

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HENING,FREDRICK,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: HENING,FREDRICK  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MCMILLAN

DATE/TIME OF TEST: 09/10/2016 04:56:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:50 |
| Air Blank          | 0.000 | 04:50 |
| Control Test       | 0.080 | 04:50 |
| Air Blank          | 0.000 | 04:51 |
| Subject Sample # 1 | 0.203 | 04:51 |
| Breath Volume      | 1.546 | 04:51 |
| Air Blank          | 0.000 | 04:52 |
| Air Blank          | 0.000 | 04:54 |
| Subject Sample # 2 | 0.193 | 04:55 |
| Breath Volume      | 1.589 | 04:55 |
| Air Blank          | 0.000 | 04:55 |
| Control Test       | 0.079 | 04:56 |
| Air Blank          | 0.000 | 04:56 |
| Diagnostics Check  | OK    | 04:56 |

---

Data Download Date/Time: 01/04/2017 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULZE,JON,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCHULZE,JON  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:15  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: HOLTON

DATE/TIME OF TEST: 09/11/2016 00:39:07

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.080 | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.234 | 00:34 |
| Breath Volume      | 1.910 | 00:34 |
| Air Blank          | 0.000 | 00:35 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 2 | 0.235 | 00:37 |
| Breath Volume      | 1.355 | 00:37 |
| Air Blank          | 0.000 | 00:38 |
| Control Test       | 0.079 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Diagnostics Check  | OK    | 00:39 |

---

Data Download Date/Time: 01/04/2017 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: CUTCLIFFE,BRIAN  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HARRIGAN

DATE/TIME OF TEST: 09/11/2016 02:53:58

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:46 |
| Air Blank          | 0.000 | 02:46 |
| Control Test       | 0.079 | 02:47 |
| Air Blank          | 0.000 | 02:47 |
| Subject Sample # 1 | 0.121 | 02:48 |
| Breath Volume      | 1.375 | 02:48 |
| Air Blank          | 0.000 | 02:48 |
| Air Blank          | 0.000 | 02:50 |
| Subject Sample # 2 | 0.121 | 02:52 |
| Breath Volume      | 2.222 | 02:52 |
| Air Blank          | 0.000 | 02:53 |
| Control Test       | 0.079 | 02:53 |
| Air Blank          | 0.000 | 02:53 |
| Diagnostics Check  | OK    | 02:53 |

---

Data Download Date/Time: 01/04/2017 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GRAY,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: GRAY,NICHOLAS  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: CORNET

DATE/TIME OF TEST: 09/11/2016 03:01:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:58 |                      |
| Air Blank         | 0.000 | 02:58 |                      |
| Control Test      | 0.079 | 02:59 |                      |
| Air Blank         | 0.000 | 02:59 |                      |
| Subject Sample    | # 1   | 02:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:59 |                      |
| Air Blank         | 0.000 | 03:00 |                      |
| Control Test      | 0.079 | 03:00 |                      |
| Air Blank         | 0.000 | 03:01 |                      |
| Diagnostics Check | OK    | 03:01 |                      |

---

Data Download Date/Time: 01/04/2017 10:17



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON, BRANDON, J  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: SHERER, DANIEL  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 09/12/2016 23:12:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:05 |
| Air Blank          | 0.000 | 23:05 |
| Control Test       | 0.080 | 23:06 |
| Air Blank          | 0.000 | 23:06 |
| Subject Sample # 1 | 0.000 | 23:07 |
| Breath Volume      | 2.218 | 23:07 |
| Air Blank          | 0.000 | 23:07 |
| Air Blank          | 0.000 | 23:09 |
| Subject Sample # 2 | 0.000 | 23:10 |
| Breath Volume      | 2.261 | 23:10 |
| Air Blank          | 0.000 | 23:11 |
| Control Test       | 0.080 | 23:11 |
| Air Blank          | 0.000 | 23:12 |
| Diagnostics Check  | OK    | 23:12 |

---

Data Download Date/Time: 01/04/2017 10:17

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KELLER,D,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: KELLER,D  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 09/13/2016 00:42:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:37 |                      |
| Air Blank         | 0.000 | 00:37 |                      |
| Control Test      | 0.080 | 00:38 |                      |
| Air Blank         | 0.000 | 00:38 |                      |
| Subject Sample    | # 1   | 00:41 | Subject Test Refused |
| Breath Volume     | 0.265 | 00:41 |                      |
| Air Blank         | 0.000 | 00:41 |                      |
| Control Test      | 0.081 | 00:42 |                      |
| Air Blank         | 0.000 | 00:42 |                      |
| Diagnostics Check | OK    | 00:42 |                      |

---

Data Download Date/Time: 01/04/2017 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BAKER,MARK,J  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BAKER,MARK  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 14:49  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 09/13/2016 15:48:24

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 15:42 |
| Air Blank          | 0.000 | 15:42 |
| Control Test       | 0.080 | 15:42 |
| Air Blank          | 0.000 | 15:43 |
| Subject Sample # 1 | 0.000 | 15:44 |
| Breath Volume      | 1.519 | 15:44 |
| Air Blank          | 0.000 | 15:44 |
| Air Blank          | 0.000 | 15:46 |
| Subject Sample # 2 | 0.000 | 15:47 |
| Breath Volume      | 1.816 | 15:47 |
| Air Blank          | 0.000 | 15:47 |
| Control Test       | 0.080 | 15:47 |
| Air Blank          | 0.000 | 15:48 |
| Diagnostics Check  | OK    | 15:48 |

---

Data Download Date/Time: 01/04/2017 10:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,G,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,G  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: MARCELIN

DATE/TIME OF TEST: 09/14/2016 02:21:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:15 |
| Air Blank          | 0.000 | 02:16 |
| Control Test       | 0.080 | 02:16 |
| Air Blank          | 0.000 | 02:16 |
| Subject Sample # 1 | 0.188 | 02:17 |
| Breath Volume      | 1.152 | 02:17 |
| Air Blank          | 0.000 | 02:17 |
| Air Blank          | 0.000 | 02:19 |
| Subject Sample # 2 | 0.172 | 02:20 |
| Breath Volume      | 1.164 | 02:20 |
| Air Blank          | 0.000 | 02:21 |
| Control Test       | 0.080 | 02:21 |
| Air Blank          | 0.000 | 02:21 |
| Diagnostics Check  | OK    | 02:21 |

---

Data Download Date/Time: 01/04/2017 10:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRAXTON,D,L  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BRAXTON,D  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:27  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 09/15/2016 01:29:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.080 | 01:23 |
| Air Blank          | 0.000 | 01:23 |
| Subject Sample # 1 | 0.279 | 01:24 |
| Breath Volume      | 3.074 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 2 | 0.280 | 01:27 |
| Breath Volume      | 2.558 | 01:27 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.079 | 01:28 |
| Air Blank          | 0.000 | 01:29 |
| Diagnostics Check  | OK    | 01:29 |

---

Data Download Date/Time: 01/04/2017 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORAN,CORY,P  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MORAN,CORY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/17/2016 00:07:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:04 |                      |
| Air Blank         | 0.000 | 00:05 |                      |
| Control Test      | 0.079 | 00:05 |                      |
| Air Blank         | 0.000 | 00:05 |                      |
| Subject Sample    | # 1   | 00:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:06 |                      |
| Air Blank         | 0.000 | 00:06 |                      |
| Control Test      | 0.079 | 00:06 |                      |
| Air Blank         | 0.000 | 00:07 |                      |
| Diagnostics Check | OK    | 00:07 |                      |

---

Data Download Date/Time: 01/04/2017 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIDSON,CECELIA,D  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: DAVIDSON,CECELIA  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 06:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: GODBOLT

DATE/TIME OF TEST: 09/17/2016 07:07:33

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 07:07 |            |
| Air Blank         |    | 07:07 | RFI Detect |
| Air Blank         |    | 07:07 | RFI Detect |

---

Data Download Date/Time: 01/04/2017 10:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIDSON,CECELIA,D  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: DAVIDSON,CECELIA  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 06:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: GODBOLT

DATE/TIME OF TEST: 09/17/2016 07:10:52

---

|                   |    |       |            |
|-------------------|----|-------|------------|
| Diagnostics Check | OK | 07:10 |            |
| Air Blank         |    | 07:10 | RFI Detect |
| Air Blank         |    | 07:10 | RFI Detect |

---

Data Download Date/Time: 01/04/2017 10:18



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRAXTON,D,L  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BRAXTON,D  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: GUILER

DATE/TIME OF TEST: 09/04/2016 03:01:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:55 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.080 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Subject Sample # 1 | 0.111 | 02:56 |
| Breath Volume      | 3.871 | 02:56 |
| Air Blank          | 0.000 | 02:57 |
| Air Blank          | 0.000 | 02:59 |
| Subject Sample # 2 | 0.109 | 02:59 |
| Breath Volume      | 3.109 | 02:59 |
| Air Blank          | 0.000 | 03:00 |
| Control Test       | 0.079 | 03:00 |
| Air Blank          | 0.000 | 03:01 |
| Diagnostics Check  | OK    | 03:01 |

---

Data Download Date/Time: 09/08/2016 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROY,G,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: ROY,G  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 08/11/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 09/07/2016 00:58:45

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Control Test       | 0.079 | 00:53 |
| Air Blank          | 0.000 | 00:54 |
| Subject Sample # 1 | 0.000 | 00:54 |
| Breath Volume      | 3.179 | 00:54 |
| Air Blank          | 0.000 | 00:55 |
| Air Blank          | 0.000 | 00:57 |
| Subject Sample # 2 | 0.000 | 00:57 |
| Breath Volume      | 2.792 | 00:57 |
| Air Blank          | 0.000 | 00:57 |
| Control Test       | 0.079 | 00:58 |
| Air Blank          | 0.000 | 00:58 |
| Diagnostics Check  | OK    | 00:58 |

---

Data Download Date/Time: 09/08/2016 10:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DILLS,JAMES,M  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: DILLS,JAMES  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HOLGUIN

DATE/TIME OF TEST: 09/10/2016 03:48:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:40 |
| Air Blank          | 0.000 | 03:40 |
| Control Test       | 0.080 | 03:41 |
| Air Blank          | 0.000 | 03:41 |
| Subject Sample # 1 | 0.124 | 03:43 |
| Breath Volume      | 1.960 | 03:43 |
| Air Blank          | 0.000 | 03:44 |
| Air Blank          | 0.000 | 03:46 |
| Subject Sample # 2 | 0.125 | 03:46 |
| Breath Volume      | 2.187 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Control Test       | 0.078 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Diagnostics Check  | OK    | 03:48 |

---

Data Download Date/Time: 10/18/2016 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RABUN,MATTHEW,K  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: RABUN,MATTHEW  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/10/2016 04:12:36

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:11 |            |
| Air Blank         |       | 04:12 | RFI Detect |
| Air Blank         | 0.000 | 04:12 |            |

---

Data Download Date/Time: 10/18/2016 11:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLLIDAY,CHAD,E  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: GLADDEN,JOHN  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:32

SUBJECT'S LAST NAME: GRUENEWALD

DATE/TIME OF TEST: 09/10/2016 17:59:03

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 17:53 |
| Air Blank          | 0.000 | 17:53 |
| Control Test       | 0.079 | 17:53 |
| Air Blank          | 0.000 | 17:54 |
| Subject Sample # 1 | 0.063 | 17:54 |
| Breath Volume      | 1.464 | 17:54 |
| Air Blank          | 0.000 | 17:55 |
| Air Blank          | 0.000 | 17:57 |
| Subject Sample # 2 | 0.059 | 17:57 |
| Breath Volume      | 1.199 | 17:57 |
| Air Blank          | 0.000 | 17:58 |
| Control Test       | 0.078 | 17:58 |
| Air Blank          | 0.000 | 17:58 |
| Diagnostics Check  | OK    | 17:59 |

---

Data Download Date/Time: 10/18/2016 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MORAN,CORY,P  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: MORAN,CORY  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 09/17/2016 00:03:44

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 00:00 |                      |
| Air Blank         | 0.000 | 00:01 |                      |
| Control Test      | 0.079 | 00:01 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Subject Sample    | # 1   | 00:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 00:02 |                      |
| Air Blank         | 0.000 | 00:02 |                      |
| Control Test      | 0.079 | 00:03 |                      |
| Air Blank         | 0.000 | 00:03 |                      |
| Diagnostics Check | OK    | 00:03 |                      |

---

Data Download Date/Time: 10/18/2016 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DAVIDSON,CECELIA,D  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: DAVIDSON,CECELIA  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: GODBOLT

DATE/TIME OF TEST: 09/17/2016 07:23:09

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:16 |
| Air Blank          | 0.000 | 07:17 |
| Control Test       | 0.080 | 07:17 |
| Air Blank          | 0.000 | 07:18 |
| Subject Sample # 1 | 0.159 | 07:18 |
| Breath Volume      | 1.406 | 07:18 |
| Air Blank          | 0.000 | 07:19 |
| Air Blank          | 0.000 | 07:20 |
| Subject Sample # 2 | 0.155 | 07:21 |
| Breath Volume      | 1.355 | 07:21 |
| Air Blank          | 0.000 | 07:22 |
| Control Test       | 0.079 | 07:22 |
| Air Blank          | 0.000 | 07:23 |
| Diagnostics Check  | OK    | 07:23 |

---

Data Download Date/Time: 10/18/2016 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BULLOCK, JOSHUA, R  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: FISCHER, ANDREW  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 09/18/2016 02:13:06

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Control Test       | 0.080 | 02:07 |
| Air Blank          | 0.000 | 02:08 |
| Subject Sample # 1 | 0.315 | 02:08 |
| Breath Volume      | 2.585 | 02:08 |
| Air Blank          | 0.000 | 02:09 |
| Air Blank          | 0.000 | 02:11 |
| Subject Sample # 2 | 0.306 | 02:11 |
| Breath Volume      | 2.437 | 02:11 |
| Air Blank          | 0.000 | 02:12 |
| Control Test       | 0.078 | 02:12 |
| Air Blank          | 0.000 | 02:13 |
| Diagnostics Check  | OK    | 02:13 |

---

Data Download Date/Time: 10/18/2016 11:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,PARISE,L  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: SHERER,DANIEL  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: STALLONE

DATE/TIME OF TEST: 09/18/2016 04:27:54

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:25 |                      |
| Air Blank         | 0.000 | 04:25 |                      |
| Control Test      | 0.080 | 04:25 |                      |
| Air Blank         | 0.000 | 04:26 |                      |
| Subject Sample    | # 1   | 04:26 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:26 |                      |
| Air Blank         | 0.000 | 04:26 |                      |
| Control Test      | 0.079 | 04:27 |                      |
| Air Blank         | 0.000 | 04:27 |                      |
| Diagnostics Check | OK    | 04:27 |                      |

---

Data Download Date/Time: 10/18/2016 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUSCITTO,C,E  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: RUSCITTO,C  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: TIDWELL

DATE/TIME OF TEST: 09/19/2016 23:50:22

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:43 |
| Air Blank          | 0.000 | 23:43 |
| Control Test       | 0.078 | 23:44 |
| Air Blank          | 0.000 | 23:44 |
| Subject Sample # 1 | 0.147 | 23:45 |
| Breath Volume      | 2.796 | 23:45 |
| Air Blank          | 0.000 | 23:46 |
| Air Blank          | 0.000 | 23:48 |
| Subject Sample # 2 | 0.147 | 23:48 |
| Breath Volume      | 2.910 | 23:48 |
| Air Blank          | 0.000 | 23:49 |
| Control Test       | 0.078 | 23:49 |
| Air Blank          | 0.000 | 23:50 |
| Diagnostics Check  | OK    | 23:50 |

---

Data Download Date/Time: 10/18/2016 11:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON, BRANDON, J  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: WILSON, BRANDON  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:23  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: O BRIEN

DATE/TIME OF TEST: 09/22/2016 01:58:27

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:55 |                      |
| Air Blank         | 0.000 | 01:56 |                      |
| Control Test      | 0.080 | 01:56 |                      |
| Air Blank         | 0.000 | 01:56 |                      |
| Subject Sample    | # 1   | 01:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:57 |                      |
| Air Blank         | 0.000 | 01:57 |                      |
| Control Test      | 0.079 | 01:57 |                      |
| Air Blank         | 0.000 | 01:58 |                      |
| Diagnostics Check | OK    | 01:58 |                      |

---

Data Download Date/Time: 10/18/2016 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHULZE,JON,  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SCHULZE,JON  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 09/22/2016 02:08:00

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:01 |
| Air Blank          | 0.000 | 02:02 |
| Control Test       | 0.080 | 02:02 |
| Air Blank          | 0.000 | 02:03 |
| Subject Sample # 1 | 0.160 | 02:03 |
| Breath Volume      | 2.269 | 02:03 |
| Air Blank          | 0.000 | 02:04 |
| Air Blank          | 0.000 | 02:05 |
| Subject Sample # 2 | 0.158 | 02:06 |
| Breath Volume      | 1.753 | 02:06 |
| Air Blank          | 0.000 | 02:07 |
| Control Test       | 0.080 | 02:07 |
| Air Blank          | 0.000 | 02:07 |
| Diagnostics Check  | OK    | 02:08 |

---

Data Download Date/Time: 10/18/2016 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON, BRANDON, J  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: WILSON, BRANDON  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: SAINTIL

DATE/TIME OF TEST: 09/22/2016 03:58:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:54 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Control Test      | 0.079 | 03:55 |                      |
| Air Blank         | 0.000 | 03:55 |                      |
| Subject Sample    | # 1   | 03:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:57 |                      |
| Air Blank         | 0.000 | 03:57 |                      |
| Control Test      | 0.079 | 03:58 |                      |
| Air Blank         | 0.000 | 03:58 |                      |
| Diagnostics Check | OK    | 03:58 |                      |

---

Data Download Date/Time: 10/18/2016 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RAYBORN,GEORGE,K  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: RAYBORN,GK  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: ALVARADO RESENDIZ

DATE/TIME OF TEST: 09/23/2016 19:50:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:42 |
| Air Blank          | 0.000 | 19:43 |
| Control Test       | 0.078 | 19:43 |
| Air Blank          | 0.000 | 19:44 |
| Subject Sample # 1 | 0.122 | 19:45 |
| Breath Volume      | 1.710 | 19:45 |
| Air Blank          | 0.000 | 19:45 |
| Air Blank          | 0.000 | 19:47 |
| Subject Sample # 2 | 0.128 | 19:48 |
| Breath Volume      | 2.070 | 19:48 |
| Air Blank          | 0.000 | 19:49 |
| Control Test       | 0.080 | 19:49 |
| Air Blank          | 0.000 | 19:50 |
| Diagnostics Check  | OK    | 19:50 |

---

Data Download Date/Time: 10/18/2016 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ZIMBA,MIKE,J  
BREATH TEST OPERATOR AGENCY: Leon County SO  
ARRESTING OFFICER: MICHAELS,RYAN  
ARRESTING OFFICER AGENCY: Leon County SO  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 09/24/2016 22:32:18

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:26 |
| Air Blank          | 0.000 | 22:26 |
| Control Test       | 0.080 | 22:26 |
| Air Blank          | 0.000 | 22:27 |
| Subject Sample # 1 | 0.152 | 22:28 |
| Breath Volume      | 2.117 | 22:28 |
| Air Blank          | 0.000 | 22:28 |
| Air Blank          | 0.000 | 22:30 |
| Subject Sample # 2 | 0.143 | 22:30 |
| Breath Volume      | 1.421 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Control Test       | 0.078 | 22:31 |
| Air Blank          | 0.000 | 22:32 |
| Diagnostics Check  | OK    | 22:32 |

---

Data Download Date/Time: 10/18/2016 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SODERHOLM,CHELSEA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HENEREY

DATE/TIME OF TEST: 09/25/2016 02:47:41

---

|                   |       |       |                    |                |
|-------------------|-------|-------|--------------------|----------------|
| Diagnostics Check | OK    | 02:35 |                    |                |
| Air Blank         | 0.000 | 02:36 |                    |                |
| Control Test      | 0.080 | 02:36 |                    |                |
| Air Blank         | 0.000 | 02:37 |                    |                |
| Subject Sample    | # 1   | 02:40 | No Sample Provided |                |
| Breath Volume     | 0.000 | 02:40 |                    |                |
| Air Blank         | 0.000 | 02:40 |                    |                |
| Air Blank         | 0.000 | 02:42 |                    |                |
| Subject Sample    | # 2   | 0.178 | 02:45              | Volume Not Met |
| Breath Volume     | 0.117 | 02:45 |                    |                |
| Air Blank         | 0.000 | 02:46 |                    |                |
| Control Test      | 0.079 | 02:47 |                    |                |
| Air Blank         | 0.000 | 02:47 |                    |                |
| Diagnostics Check | OK    | 02:47 |                    |                |

---

Data Download Date/Time: 10/18/2016 11:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: SODERHOLM,CHELSEA  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HENEREY

DATE/TIME OF TEST: 09/25/2016 02:55:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:49 |                      |
| Air Blank         | 0.000 | 02:50 |                      |
| Control Test      | 0.078 | 02:50 |                      |
| Air Blank         | 0.000 | 02:51 |                      |
| Subject Sample    | # 1   | 02:53 | Subject Test Refused |
| Breath Volume     | 0.308 | 02:53 |                      |
| Air Blank         | 0.000 | 02:54 |                      |
| Control Test      | 0.079 | 02:54 |                      |
| Air Blank         | 0.000 | 02:55 |                      |
| Diagnostics Check | OK    | 02:55 |                      |

---

Data Download Date/Time: 10/18/2016 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AZRAK,PAUL,M  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: AZRAK,PAUL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 09/25/2016 03:17:29

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:13 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Control Test      | 0.079 | 03:14 |                      |
| Air Blank         | 0.000 | 03:14 |                      |
| Subject Sample    | # 1   | 03:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:16 |                      |
| Air Blank         | 0.000 | 03:16 |                      |
| Control Test      | 0.079 | 03:16 |                      |
| Air Blank         | 0.000 | 03:17 |                      |
| Diagnostics Check | OK    | 03:17 |                      |

---

Data Download Date/Time: 10/18/2016 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRAXTON,D,L  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: BRAXTON,D  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/29/2016 01:37:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:31 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.080 | 01:32 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.128 | 01:33 |
| Breath Volume      | 1.582 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.137 | 01:36 |
| Breath Volume      | 2.175 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.079 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 10/18/2016 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,CATHY,A  
BREATH TEST OPERATOR AGENCY: Tallahassee PD  
ARRESTING OFFICER: TTERLIKKIS,STEPHANOS  
ARRESTING OFFICER AGENCY: Tallahassee PD  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 09:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: GIORDANO

DATE/TIME OF TEST: 09/29/2016 10:30:42

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 10:24 |
| Air Blank          | 0.000 | 10:25 |
| Control Test       | 0.079 | 10:25 |
| Air Blank          | 0.000 | 10:25 |
| Subject Sample # 1 | 0.012 | 10:26 |
| Breath Volume      | 4.257 | 10:26 |
| Air Blank          | 0.000 | 10:26 |
| Air Blank          | 0.000 | 10:28 |
| Subject Sample # 2 | 0.012 | 10:29 |
| Breath Volume      | 4.589 | 10:29 |
| Air Blank          | 0.000 | 10:29 |
| Control Test       | 0.079 | 10:30 |
| Air Blank          | 0.000 | 10:30 |
| Diagnostics Check  | OK    | 10:30 |

---

Data Download Date/Time: 10/18/2016 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON, BRANDON, J  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: WILSON, BRANDON  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: MCGOVERN

DATE/TIME OF TEST: 09/30/2016 19:44:20

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 19:43 |            |
| Air Blank         |       | 19:43 | RFI Detect |
| Air Blank         | 0.000 | 19:44 |            |

---

Data Download Date/Time: 10/18/2016 11:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILSON, BRANDON, J  
BREATH TEST OPERATOR AGENCY: FL State University  
ARRESTING OFFICER: WILSON, BRANDON  
ARRESTING OFFICER AGENCY: FL State University  
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:22

SUBJECT'S LAST NAME: MCGOVERN

DATE/TIME OF TEST: 09/30/2016 19:48:27

---

|                    |       |       |            |
|--------------------|-------|-------|------------|
| Diagnostics Check  | OK    | 19:45 |            |
| Air Blank          | 0.000 | 19:46 |            |
| Control Test       | 0.080 | 19:46 |            |
| Air Blank          | 0.000 | 19:47 |            |
| Subject Sample # 1 | 0.039 | 19:47 |            |
| Breath Volume      | 3.488 | 19:47 |            |
| Air Blank          |       | 19:47 | RFI Detect |
| Air Blank          |       | 19:47 | RFI Detect |
| Air Blank          | 0.000 | 19:48 |            |

---

Data Download Date/Time: 10/18/2016 11:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BARROW,JESHUA,C  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: BARROW,JESHUA  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MCCREA

DATE/TIME OF TEST: 09/11/2016 05:32:11

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 05:25 |            |
| Air Blank         | 0.000     | 05:26 |            |
| Control Test      | 0.081     | 05:26 |            |
| Air Blank         | 0.000     | 05:27 |            |
| Subject Sample    | # 1 0.243 | 05:27 |            |
| Breath Volume     | 1.960     | 05:27 |            |
| Air Blank         | 0.000     | 05:28 |            |
| Air Blank         | 0.000     | 05:29 |            |
| Subject Sample    | # 2 0.237 | 05:30 |            |
| Breath Volume     | 2.140     | 05:30 |            |
| Air Blank         | 0.000     | 05:30 |            |
| Control Test      | 0.081     | 05:31 |            |
| Air Blank         |           | 05:31 | RFI Detect |
| Air Blank         | 0.000     | 05:32 |            |

---

Data Download Date/Time: 09/26/2016 10:46

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 09/13/2016 20:23:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 20:17 |
| Air Blank          | 0.000 | 20:17 |
| Control Test       | 0.082 | 20:17 |
| Air Blank          | 0.000 | 20:18 |
| Subject Sample # 1 | 0.072 | 20:18 |
| Breath Volume      | 3.238 | 20:18 |
| Air Blank          | 0.000 | 20:19 |
| Air Blank          | 0.000 | 20:21 |
| Subject Sample # 2 | 0.068 | 20:21 |
| Breath Volume      | 2.257 | 20:21 |
| Air Blank          | 0.000 | 20:22 |
| Control Test       | 0.081 | 20:22 |
| Air Blank          | 0.000 | 20:23 |
| Diagnostics Check  | OK    | 20:23 |

---

Data Download Date/Time: 09/26/2016 10:46



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/10/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: MAKOWIECKI

DATE/TIME OF TEST: 09/07/2016 22:48:35

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:41 |
| Air Blank          | 0.000 | 22:42 |
| Control Test       | 0.080 | 22:42 |
| Air Blank          | 0.000 | 22:43 |
| Subject Sample # 1 | 0.102 | 22:44 |
| Breath Volume      | 1.523 | 22:44 |
| Air Blank          | 0.000 | 22:44 |
| Air Blank          | 0.000 | 22:46 |
| Subject Sample # 2 | 0.101 | 22:47 |
| Breath Volume      | 2.285 | 22:47 |
| Air Blank          | 0.000 | 22:47 |
| Control Test       | 0.079 | 22:48 |
| Air Blank          | 0.000 | 22:48 |
| Diagnostics Check  | OK    | 22:48 |

---

Data Download Date/Time: 09/28/2016 09:23

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON,TRAVIS,W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DANJOU,SCOTT  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: FOLTZ

DATE/TIME OF TEST: 09/28/2016 22:07:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:01 |
| Air Blank          | 0.000 | 22:02 |
| Control Test       | 0.079 | 22:02 |
| Air Blank          | 0.000 | 22:03 |
| Subject Sample # 1 | 0.000 | 22:03 |
| Breath Volume      | 2.796 | 22:03 |
| Air Blank          | 0.000 | 22:04 |
| Air Blank          | 0.000 | 22:06 |
| Subject Sample # 2 | 0.000 | 22:06 |
| Breath Volume      | 2.917 | 22:06 |
| Air Blank          | 0.000 | 22:06 |
| Control Test       | 0.079 | 22:07 |
| Air Blank          | 0.000 | 22:07 |
| Diagnostics Check  | OK    | 22:07 |

---

Data Download Date/Time: 09/29/2016 08:51

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLEMAN,BRETT,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: GATTON,BRADY  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CROSSLEY

DATE/TIME OF TEST: 09/29/2016 22:36:19

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:30 |
| Air Blank          | 0.000 | 22:30 |
| Control Test       | 0.079 | 22:30 |
| Air Blank          | 0.000 | 22:31 |
| Subject Sample # 1 | 0.000 | 22:32 |
| Breath Volume      | 1.867 | 22:32 |
| Air Blank          | 0.000 | 22:32 |
| Air Blank          | 0.000 | 22:34 |
| Subject Sample # 2 | 0.000 | 22:34 |
| Breath Volume      | 3.234 | 22:34 |
| Air Blank          | 0.000 | 22:35 |
| Control Test       | 0.079 | 22:35 |
| Air Blank          | 0.000 | 22:36 |
| Diagnostics Check  | OK    | 22:36 |

---

Data Download Date/Time: 10/06/2016 11:59

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DENSON, TRAVIS, W  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ZITO, ANTHONY  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 09/29/2016 23:20:56

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:14 |
| Air Blank          | 0.000 | 23:14 |
| Control Test       | 0.078 | 23:14 |
| Air Blank          | 0.000 | 23:15 |
| Subject Sample # 1 | 0.174 | 23:16 |
| Breath Volume      | 1.558 | 23:16 |
| Air Blank          | 0.000 | 23:16 |
| Air Blank          | 0.000 | 23:18 |
| Subject Sample # 2 | 0.166 | 23:19 |
| Breath Volume      | 1.113 | 23:19 |
| Air Blank          | 0.000 | 23:20 |
| Control Test       | 0.078 | 23:20 |
| Air Blank          | 0.000 | 23:20 |
| Diagnostics Check  | OK    | 23:20 |

---

Data Download Date/Time: 10/06/2016 11:59

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:32  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: BLACK

DATE/TIME OF TEST: 09/09/2016 02:32:17

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:26 |
| Air Blank          | 0.000 | 02:26 |
| Control Test       | 0.080 | 02:27 |
| Air Blank          | 0.000 | 02:27 |
| Subject Sample # 1 | 0.201 | 02:27 |
| Breath Volume      | 2.621 | 02:27 |
| Air Blank          | 0.000 | 02:28 |
| Air Blank          | 0.000 | 02:30 |
| Subject Sample # 2 | 0.209 | 02:30 |
| Breath Volume      | 2.906 | 02:30 |
| Air Blank          | 0.000 | 02:31 |
| Control Test       | 0.080 | 02:31 |
| Air Blank          | 0.000 | 02:32 |
| Diagnostics Check  | OK    | 02:32 |

---

Data Download Date/Time: 09/29/2016 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 09/10/2016 02:54:16

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:51 |                      |
| Air Blank         | 0.000 | 02:51 |                      |
| Control Test      | 0.080 | 02:52 |                      |
| Air Blank         | 0.000 | 02:52 |                      |
| Subject Sample    | # 1   | 02:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 02:52 |                      |
| Air Blank         | 0.000 | 02:53 |                      |
| Control Test      | 0.079 | 02:53 |                      |
| Air Blank         | 0.000 | 02:54 |                      |
| Diagnostics Check | OK    | 02:54 |                      |

---

Data Download Date/Time: 09/29/2016 12:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/11/2016 02:56:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 02:50 |
| Air Blank          | 0.000 | 02:50 |
| Control Test       | 0.080 | 02:51 |
| Air Blank          | 0.000 | 02:51 |
| Subject Sample # 1 | 0.113 | 02:52 |
| Breath Volume      | 2.003 | 02:52 |
| Air Blank          | 0.000 | 02:52 |
| Air Blank          | 0.000 | 02:54 |
| Subject Sample # 2 | 0.111 | 02:54 |
| Breath Volume      | 1.453 | 02:54 |
| Air Blank          | 0.000 | 02:55 |
| Control Test       | 0.081 | 02:55 |
| Air Blank          | 0.000 | 02:56 |
| Diagnostics Check  | OK    | 02:56 |

---

Data Download Date/Time: 09/29/2016 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEKLE, MATTHEW, L

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: CENTENO, A

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 17:52

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: MESSERVY

DATE/TIME OF TEST: 09/13/2016 18:49:04

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 18:48 |            |
| Air Blank         |       | 18:48 | RFI Detect |
| Air Blank         | 0.000 | 18:49 |            |

---

Data Download Date/Time: 09/29/2016 12:06



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEKLE, MATTHEW, L

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: CENTENO, ANA

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 17:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: MESSERVY

DATE/TIME OF TEST: 09/13/2016 18:54:05

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:51 |                      |
| Air Blank         | 0.000 | 18:51 |                      |
| Control Test      | 0.079 | 18:51 |                      |
| Air Blank         | 0.000 | 18:52 |                      |
| Subject Sample    | # 1   | 18:52 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:52 |                      |
| Air Blank         | 0.000 | 18:53 |                      |
| Control Test      | 0.079 | 18:53 |                      |
| Air Blank         | 0.000 | 18:54 |                      |
| Diagnostics Check | OK    | 18:54 |                      |

---

Data Download Date/Time: 09/29/2016 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEKLE, MATTHEW, L

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: CENTENO, ANA

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 17:57

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: MESSERVY

DATE/TIME OF TEST: 09/13/2016 19:00:37

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 18:57 |                      |
| Air Blank         | 0.000 | 18:58 |                      |
| Control Test      | 0.079 | 18:58 |                      |
| Air Blank         | 0.000 | 18:59 |                      |
| Subject Sample    | # 1   | 18:59 | Subject Test Refused |
| Breath Volume     | 0.000 | 18:59 |                      |
| Air Blank         | 0.000 | 18:59 |                      |
| Control Test      | 0.080 | 19:00 |                      |
| Air Blank         | 0.000 | 19:00 |                      |
| Diagnostics Check | OK    | 19:00 |                      |

---

Data Download Date/Time: 09/29/2016 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 09/22/2016 23:57:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:51 |
| Air Blank          | 0.000 | 23:52 |
| Control Test       | 0.080 | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Subject Sample # 1 | 0.000 | 23:53 |
| Breath Volume      | 1.652 | 23:53 |
| Air Blank          | 0.000 | 23:53 |
| Air Blank          | 0.000 | 23:55 |
| Subject Sample # 2 | 0.000 | 23:56 |
| Breath Volume      | 1.726 | 23:56 |
| Air Blank          | 0.000 | 23:56 |
| Control Test       | 0.081 | 23:56 |
| Air Blank          | 0.000 | 23:57 |
| Diagnostics Check  | OK    | 23:57 |

---

Data Download Date/Time: 09/29/2016 12:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEKLE, MATTHEW, L  
BREATH TEST OPERATOR AGENCY: Quincy PD  
ARRESTING OFFICER: DEKLE, MATTHEW  
ARRESTING OFFICER AGENCY: Quincy PD  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HOVER

DATE/TIME OF TEST: 09/23/2016 03:27:52

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 03:20 |
| Air Blank         |     | 0.000 | 03:21 |
| Control Test      |     | 0.080 | 03:21 |
| Air Blank         |     | 0.000 | 03:21 |
| Subject Sample    | # 1 | 0.176 | 03:22 |
| Breath Volume     |     | 1.945 | 03:22 |
| Air Blank         |     | 0.000 | 03:23 |
| Air Blank         |     | 0.000 | 03:25 |
| Subject Sample    | # 2 | 0.181 | 03:26 |
| Breath Volume     |     | 1.363 | 03:26 |
| Air Blank         |     | 0.000 | 03:26 |
| Control Test      |     | 0.080 | 03:27 |
| Air Blank         |     | 0.000 | 03:27 |
| Diagnostics Check |     | OK    | 03:27 |

---

Data Download Date/Time: 09/29/2016 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:55  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: PARRISH

DATE/TIME OF TEST: 09/25/2016 00:38:39

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:32 |
| Air Blank          | 0.000 | 00:32 |
| Control Test       | 0.080 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.132 | 00:34 |
| Breath Volume      | 2.976 | 00:34 |
| Air Blank          | 0.000 | 00:34 |
| Air Blank          | 0.000 | 00:36 |
| Subject Sample # 2 | 0.132 | 00:37 |
| Breath Volume      | 3.953 | 00:37 |
| Air Blank          | 0.000 | 00:37 |
| Control Test       | 0.080 | 00:38 |
| Air Blank          | 0.000 | 00:38 |
| Diagnostics Check  | OK    | 00:38 |

---

Data Download Date/Time: 09/29/2016 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FINUFF,DANIEL,W  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: FINUFF,DANIEL  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 09/25/2016 03:40:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:33 |
| Air Blank          | 0.000 | 03:34 |
| Control Test       | 0.080 | 03:34 |
| Air Blank          | 0.000 | 03:35 |
| Subject Sample # 1 | 0.165 | 03:35 |
| Breath Volume      | 1.316 | 03:35 |
| Air Blank          | 0.000 | 03:36 |
| Air Blank          | 0.000 | 03:38 |
| Subject Sample # 2 | 0.150 | 03:38 |
| Breath Volume      | 1.957 | 03:38 |
| Air Blank          | 0.000 | 03:39 |
| Control Test       | 0.080 | 03:39 |
| Air Blank          | 0.000 | 03:40 |
| Diagnostics Check  | OK    | 03:40 |

---

Data Download Date/Time: 09/29/2016 12:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GENTRY, JONATHAN, P  
BREATH TEST OPERATOR AGENCY: FHP Troop H  
ARRESTING OFFICER: GENTRY, JONATHAN  
ARRESTING OFFICER AGENCY: FHP Troop H  
LAST AGENCY INSPECTION: 08/08/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34015080A2  
DRY GAS EXPIRY DATE: 01/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: NEWBURN

DATE/TIME OF TEST: 09/28/2016 21:51:05

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:44 |
| Air Blank          | 0.000 | 21:44 |
| Control Test       | 0.080 | 21:45 |
| Air Blank          | 0.000 | 21:45 |
| Subject Sample # 1 | 0.125 | 21:46 |
| Breath Volume      | 2.902 | 21:46 |
| Air Blank          | 0.000 | 21:46 |
| Air Blank          | 0.000 | 21:48 |
| Subject Sample # 2 | 0.118 | 21:49 |
| Breath Volume      | 4.019 | 21:49 |
| Air Blank          | 0.000 | 21:50 |
| Control Test       | 0.080 | 21:50 |
| Air Blank          | 0.000 | 21:51 |
| Diagnostics Check  | OK    | 21:51 |

---

Data Download Date/Time: 09/29/2016 12:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTONE,L,M  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: ROBLES,S  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/05/2016 21:20:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 21:13 |
| Air Blank          | 0.000 | 21:13 |
| Control Test       | 0.080 | 21:14 |
| Air Blank          | 0.000 | 21:14 |
| Subject Sample # 1 | 0.144 | 21:15 |
| Breath Volume      | 2.277 | 21:15 |
| Air Blank          | 0.000 | 21:16 |
| Air Blank          | 0.000 | 21:17 |
| Subject Sample # 2 | 0.135 | 21:18 |
| Breath Volume      | 2.125 | 21:18 |
| Air Blank          | 0.000 | 21:19 |
| Control Test       | 0.078 | 21:19 |
| Air Blank          | 0.000 | 21:20 |
| Diagnostics Check  | OK    | 21:20 |

---

Data Download Date/Time: 09/29/2016 08:53



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: MONDELLO, ANTHONY  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GATES

DATE/TIME OF TEST: 09/06/2016 01:37:41

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:30 |
| Air Blank          | 0.000 | 01:31 |
| Control Test       | 0.081 | 01:31 |
| Air Blank          | 0.000 | 01:32 |
| Subject Sample # 1 | 0.000 | 01:33 |
| Breath Volume      | 2.710 | 01:33 |
| Air Blank          | 0.000 | 01:33 |
| Air Blank          | 0.000 | 01:35 |
| Subject Sample # 2 | 0.000 | 01:36 |
| Breath Volume      | 2.351 | 01:36 |
| Air Blank          | 0.000 | 01:36 |
| Control Test       | 0.079 | 01:37 |
| Air Blank          | 0.000 | 01:37 |
| Diagnostics Check  | OK    | 01:37 |

---

Data Download Date/Time: 09/29/2016 08:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: OWENS,JOHN  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 09/07/2016 06:38:36

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:31 |
| Air Blank          | 0.000 | 06:32 |
| Control Test       | 0.080 | 06:32 |
| Air Blank          | 0.000 | 06:32 |
| Subject Sample # 1 | 0.118 | 06:33 |
| Breath Volume      | 2.210 | 06:33 |
| Air Blank          | 0.000 | 06:34 |
| Air Blank          | 0.000 | 06:36 |
| Subject Sample # 2 | 0.110 | 06:37 |
| Breath Volume      | 1.398 | 06:37 |
| Air Blank          | 0.000 | 06:37 |
| Control Test       | 0.079 | 06:38 |
| Air Blank          | 0.000 | 06:38 |
| Diagnostics Check  | OK    | 06:38 |

---

Data Download Date/Time: 09/29/2016 08:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLERIA,TROY,R  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: TELLERIA,TROY  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 09/09/2016 23:53:38

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:46 |
| Air Blank          | 0.000 | 23:47 |
| Control Test       | 0.080 | 23:47 |
| Air Blank          | 0.000 | 23:48 |
| Subject Sample # 1 | 0.078 | 23:49 |
| Breath Volume      | 2.875 | 23:49 |
| Air Blank          | 0.000 | 23:49 |
| Air Blank          | 0.000 | 23:51 |
| Subject Sample # 2 | 0.073 | 23:52 |
| Breath Volume      | 1.746 | 23:52 |
| Air Blank          | 0.000 | 23:52 |
| Control Test       | 0.079 | 23:53 |
| Air Blank          | 0.000 | 23:53 |
| Diagnostics Check  | OK    | 23:53 |

---

Data Download Date/Time: 09/29/2016 08:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SANTONE,LACEY,M  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: ROBLES,S  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: ESPINOZA-CALDERON

DATE/TIME OF TEST: 09/10/2016 18:49:29

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:42 |
| Air Blank          | 0.000 | 18:42 |
| Control Test       | 0.080 | 18:43 |
| Air Blank          | 0.000 | 18:43 |
| Subject Sample # 1 | 0.203 | 18:45 |
| Breath Volume      | 1.898 | 18:45 |
| Air Blank          | 0.000 | 18:45 |
| Air Blank          | 0.000 | 18:47 |
| Subject Sample # 2 | 0.202 | 18:47 |
| Breath Volume      | 1.675 | 18:47 |
| Air Blank          | 0.000 | 18:48 |
| Control Test       | 0.077 | 18:48 |
| Air Blank          | 0.000 | 18:49 |
| Diagnostics Check  | OK    | 18:49 |

---

Data Download Date/Time: 09/29/2016 08:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDELLO,ANTHNY,J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: MONDELLO,ANTHONY  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: FEBRES

DATE/TIME OF TEST: 09/11/2016 01:14:59

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:08 |
| Air Blank          | 0.000 | 01:09 |
| Control Test       | 0.080 | 01:09 |
| Air Blank          | 0.000 | 01:09 |
| Subject Sample # 1 | 0.000 | 01:10 |
| Breath Volume      | 3.488 | 01:10 |
| Air Blank          | 0.000 | 01:11 |
| Air Blank          | 0.000 | 01:13 |
| Subject Sample # 2 | 0.000 | 01:13 |
| Breath Volume      | 3.792 | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Control Test       | 0.078 | 01:14 |
| Air Blank          | 0.000 | 01:14 |
| Diagnostics Check  | OK    | 01:14 |

---

Data Download Date/Time: 09/29/2016 08:53

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: CARLO, NATHANIEL  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: FODOR

DATE/TIME OF TEST: 09/12/2016 00:40:04

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:32 |
| Air Blank          | 0.000 | 00:33 |
| Control Test       | 0.081 | 00:33 |
| Air Blank          | 0.000 | 00:33 |
| Subject Sample # 1 | 0.339 | 00:35 |
| Breath Volume      | 3.566 | 00:35 |
| Air Blank          | 0.000 | 00:35 |
| Air Blank          | 0.000 | 00:37 |
| Subject Sample # 2 | 0.343 | 00:38 |
| Breath Volume      | 3.445 | 00:38 |
| Air Blank          | 0.000 | 00:39 |
| Control Test       | 0.078 | 00:39 |
| Air Blank          | 0.000 | 00:39 |
| Diagnostics Check  | OK    | 00:40 |

---

Data Download Date/Time: 09/29/2016 08:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TELLERIA,TROY,R  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: TELLERIA,TROY  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 09/20/2016 01:20:12

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:12 |
| Air Blank          | 0.000 | 01:13 |
| Control Test       | 0.080 | 01:13 |
| Air Blank          | 0.000 | 01:14 |
| Subject Sample # 1 | 0.143 | 01:15 |
| Breath Volume      | 2.289 | 01:15 |
| Air Blank          | 0.000 | 01:16 |
| Air Blank          | 0.000 | 01:18 |
| Subject Sample # 2 | 0.136 | 01:18 |
| Breath Volume      | 2.257 | 01:18 |
| Air Blank          | 0.000 | 01:19 |
| Control Test       | 0.079 | 01:19 |
| Air Blank          | 0.000 | 01:20 |
| Diagnostics Check  | OK    | 01:20 |

---

Data Download Date/Time: 09/29/2016 08:53

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BEEKMAN,JACOB,C  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: BEEKMAN,JACOB  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 163437  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: BARAJAS

DATE/TIME OF TEST: 09/30/2016 04:46:04

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 04:38 |
| Air Blank         |     | 0.000 | 04:39 |
| Control Test      |     | 0.080 | 04:39 |
| Air Blank         |     | 0.000 | 04:39 |
| Subject Sample    | # 1 | 0.156 | 04:41 |
| Breath Volume     |     | 2.296 | 04:41 |
| Air Blank         |     | 0.000 | 04:42 |
| Air Blank         |     | 0.000 | 04:43 |
| Subject Sample    | # 2 | 0.154 | 04:44 |
| Breath Volume     |     | 2.238 | 04:44 |
| Air Blank         |     | 0.000 | 04:45 |
| Control Test      |     | 0.080 | 04:45 |
| Air Blank         |     | 0.000 | 04:45 |
| Diagnostics Check |     | OK    | 04:46 |

---

Data Download Date/Time: 10/14/2016 09:20



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: GERARDI,AARON  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: COULTER

DATE/TIME OF TEST: 09/01/2016 04:09:52

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 04:08 |            |
| Air Blank         | 0.000 | 04:08 |            |
| Control Test      | 0.076 | 04:09 |            |
| Air Blank         |       | 04:09 | RFI Detect |
| Air Blank         | 0.000 | 04:09 |            |

---

Data Download Date/Time: 09/29/2016 21:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: GERARDI,AARON  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: COULTER

DATE/TIME OF TEST: 09/01/2016 04:17:46

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:10 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.076 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 1 | 0.000 | 04:13 |
| Breath Volume      | 1.382 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Air Blank          | 0.000 | 04:16 |
| Subject Sample # 2 | 0.000 | 04:16 |
| Breath Volume      | 1.390 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.076 | 04:17 |
| Air Blank          | 0.000 | 04:17 |
| Diagnostics Check  | OK    | 04:17 |

---

Data Download Date/Time: 09/29/2016 21:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: GERARDI,AARON  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 09/03/2016 00:53:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:45 |
| Air Blank          | 0.000 | 00:46 |
| Control Test       | 0.076 | 00:46 |
| Air Blank          | 0.000 | 00:46 |
| Subject Sample # 1 | 0.180 | 00:48 |
| Breath Volume      | 1.636 | 00:48 |
| Air Blank          | 0.000 | 00:49 |
| Air Blank          | 0.000 | 00:50 |
| Subject Sample # 2 | 0.179 | 00:51 |
| Breath Volume      | 1.492 | 00:51 |
| Air Blank          | 0.000 | 00:52 |
| Control Test       | 0.075 | 00:52 |
| Air Blank          | 0.000 | 00:53 |
| Diagnostics Check  | OK    | 00:53 |

---

Data Download Date/Time: 09/29/2016 21:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: ELLER,ASHLEY  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: GUERRIER

DATE/TIME OF TEST: 09/03/2016 03:08:06

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 02:59 |                      |
| Air Blank         | 0.000 | 03:00 |                      |
| Control Test      | 0.075 | 03:00 |                      |
| Air Blank         | 0.000 | 03:00 |                      |
| Subject Sample    | # 1   | 03:04 | No Sample Provided   |
| Breath Volume     | 0.000 | 03:04 |                      |
| Air Blank         | 0.000 | 03:04 |                      |
| Air Blank         | 0.000 | 03:06 |                      |
| Subject Sample    | # 2   | 03:06 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:06 |                      |
| Air Blank         | 0.000 | 03:07 |                      |
| Control Test      | 0.075 | 03:07 |                      |
| Air Blank         | 0.000 | 03:08 |                      |
| Diagnostics Check | OK    | 03:08 |                      |

---

Data Download Date/Time: 09/29/2016 21:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: VELAZQUEZ,SAMUEL  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 04:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ELIZONDOCARRILLO

DATE/TIME OF TEST: 09/03/2016 05:04:21

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 05:00 |                      |
| Air Blank         | 0.000 | 05:01 |                      |
| Control Test      | 0.076 | 05:01 |                      |
| Air Blank         | 0.000 | 05:02 |                      |
| Subject Sample    | # 1   | 05:02 | Subject Test Refused |
| Breath Volume     | 0.000 | 05:02 |                      |
| Air Blank         | 0.000 | 05:03 |                      |
| Control Test      | 0.076 | 05:03 |                      |
| Air Blank         | 0.000 | 05:04 |                      |
| Diagnostics Check | OK    | 05:04 |                      |

---

Data Download Date/Time: 09/29/2016 21:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: SORRELLS,BRIAN  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: SHATZER

DATE/TIME OF TEST: 09/04/2016 03:58:42

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 03:55 |                      |
| Air Blank         | 0.000 | 03:56 |                      |
| Control Test      | 0.077 | 03:56 |                      |
| Air Blank         | 0.000 | 03:57 |                      |
| Subject Sample    | # 1   | 03:57 | Subject Test Refused |
| Breath Volume     | 0.000 | 03:57 |                      |
| Air Blank         | 0.000 | 03:57 |                      |
| Control Test      | 0.077 | 03:58 |                      |
| Air Blank         | 0.000 | 03:58 |                      |
| Diagnostics Check | OK    | 03:58 |                      |

---

Data Download Date/Time: 09/29/2016 21:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: HENRY,FRITZ  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 03:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GONZALES-LOPEZ

DATE/TIME OF TEST: 09/07/2016 04:05:20

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 04:00 |                      |
| Air Blank         | 0.000 | 04:01 |                      |
| Control Test      | 0.076 | 04:01 |                      |
| Air Blank         | 0.000 | 04:01 |                      |
| Subject Sample    | # 1   | 04:03 | Subject Test Refused |
| Breath Volume     | 0.000 | 04:03 |                      |
| Air Blank         | 0.000 | 04:04 |                      |
| Control Test      | 0.076 | 04:04 |                      |
| Air Blank         | 0.000 | 04:05 |                      |
| Diagnostics Check | OK    | 04:05 |                      |

---

Data Download Date/Time: 09/29/2016 21:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: ELLER,ASHLEY  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LAFABER

DATE/TIME OF TEST: 09/09/2016 23:36:20

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 23:30 |
| Air Blank          | 0.000 | 23:30 |
| Control Test       | 0.075 | 23:31 |
| Air Blank          | 0.000 | 23:31 |
| Subject Sample # 1 | 0.129 | 23:31 |
| Breath Volume      | 3.199 | 23:31 |
| Air Blank          | 0.000 | 23:32 |
| Air Blank          | 0.000 | 23:34 |
| Subject Sample # 2 | 0.129 | 23:34 |
| Breath Volume      | 3.132 | 23:34 |
| Air Blank          | 0.000 | 23:35 |
| Control Test       | 0.076 | 23:35 |
| Air Blank          | 0.000 | 23:36 |
| Diagnostics Check  | OK    | 23:36 |

---

Data Download Date/Time: 09/29/2016 21:10



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: PYBURN,JENNIFER  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 09/11/2016 01:06:01

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:59 |
| Air Blank          | 0.000 | 00:59 |
| Control Test       | 0.075 | 00:59 |
| Air Blank          | 0.000 | 01:00 |
| Subject Sample # 1 | 0.000 | 01:01 |
| Breath Volume      | 1.121 | 01:01 |
| Air Blank          | 0.000 | 01:02 |
| Air Blank          | 0.000 | 01:04 |
| Subject Sample # 2 | 0.000 | 01:04 |
| Breath Volume      | 1.617 | 01:04 |
| Air Blank          | 0.000 | 01:05 |
| Control Test       | 0.075 | 01:05 |
| Air Blank          | 0.000 | 01:05 |
| Diagnostics Check  | OK    | 01:06 |

---

Data Download Date/Time: 09/29/2016 21:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: HENRY,FRITZ  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: KAUFMAN

DATE/TIME OF TEST: 09/14/2016 01:17:52

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 01:14 |                      |
| Air Blank         | 0.000 | 01:15 |                      |
| Control Test      | 0.077 | 01:15 |                      |
| Air Blank         | 0.000 | 01:16 |                      |
| Subject Sample    | # 1   | 01:16 | Subject Test Refused |
| Breath Volume     | 0.000 | 01:16 |                      |
| Air Blank         | 0.000 | 01:16 |                      |
| Control Test      | 0.077 | 01:17 |                      |
| Air Blank         | 0.000 | 01:17 |                      |
| Diagnostics Check | OK    | 01:17 |                      |

---

Data Download Date/Time: 09/29/2016 21:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: ELLER,ASHLEY  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 09/17/2016 01:46:13

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:39 |
| Air Blank          | 0.000 | 01:40 |
| Control Test       | 0.076 | 01:40 |
| Air Blank          | 0.000 | 01:41 |
| Subject Sample # 1 | 0.000 | 01:41 |
| Breath Volume      | 3.949 | 01:41 |
| Air Blank          | 0.000 | 01:42 |
| Air Blank          | 0.000 | 01:44 |
| Subject Sample # 2 | 0.000 | 01:44 |
| Breath Volume      | 3.984 | 01:44 |
| Air Blank          | 0.000 | 01:45 |
| Control Test       | 0.076 | 01:45 |
| Air Blank          | 0.000 | 01:46 |
| Diagnostics Check  | OK    | 01:46 |

---

Data Download Date/Time: 09/29/2016 21:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLER,ASHLEY,B  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: HENRY,FRITZ  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 05:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: QUINONES-OLSON

DATE/TIME OF TEST: 09/17/2016 07:08:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:01 |
| Air Blank          | 0.000 | 07:01 |
| Control Test       | 0.076 | 07:01 |
| Air Blank          | 0.000 | 07:02 |
| Subject Sample # 1 | 0.113 | 07:04 |
| Breath Volume      | 1.500 | 07:04 |
| Air Blank          | 0.000 | 07:05 |
| Air Blank          | 0.000 | 07:06 |
| Subject Sample # 2 | 0.113 | 07:07 |
| Breath Volume      | 1.367 | 07:07 |
| Air Blank          | 0.000 | 07:08 |
| Control Test       | 0.076 | 07:08 |
| Air Blank          | 0.000 | 07:08 |
| Diagnostics Check  | OK    | 07:08 |

---

Data Download Date/Time: 09/29/2016 21:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: PYBURN,JENNIFER  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 09/18/2016 17:31:42

---

|                   |       |       |               |
|-------------------|-------|-------|---------------|
| Diagnostics Check | OK    | 17:28 |               |
| Air Blank         | 0.000 | 17:28 |               |
| Control Test      | 0.076 | 17:29 |               |
| Air Blank         | 0.000 | 17:29 |               |
| Subject Sample    | # 1   | 17:30 | Slope Not Met |
| Breath Volume     | 1.558 | 17:30 |               |
| Air Blank         | 0.000 | 17:30 |               |
| Control Test      | 0.075 | 17:31 |               |
| Air Blank         | 0.000 | 17:31 |               |
| Diagnostics Check | OK    | 17:31 |               |

---

Data Download Date/Time: 09/29/2016 21:10

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J  
BREATH TEST OPERATOR AGENCY: Apopka PD  
ARRESTING OFFICER: PYBURN,JENNIFER  
ARRESTING OFFICER AGENCY: Apopka PD  
LAST AGENCY INSPECTION: 08/26/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 13115080A1  
DRY GAS EXPIRY DATE: 06/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 16:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: CERVANTES

DATE/TIME OF TEST: 09/18/2016 18:13:28

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 18:07 |
| Air Blank          | 0.000 | 18:07 |
| Control Test       | 0.076 | 18:08 |
| Air Blank          | 0.000 | 18:08 |
| Subject Sample # 1 | 0.294 | 18:09 |
| Breath Volume      | 3.019 | 18:09 |
| Air Blank          | 0.000 | 18:09 |
| Air Blank          | 0.000 | 18:11 |
| Subject Sample # 2 | 0.299 | 18:11 |
| Breath Volume      | 3.078 | 18:11 |
| Air Blank          | 0.000 | 18:12 |
| Control Test       | 0.076 | 18:12 |
| Air Blank          | 0.000 | 18:13 |
| Diagnostics Check  | OK    | 18:13 |

---

Data Download Date/Time: 09/29/2016 21:10

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICHOLSON,JOEL,R  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: FISHER,JARED  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: BONET

DATE/TIME OF TEST: 09/02/2016 00:11:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.082 | 00:06 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.198 | 00:07 |
| Breath Volume      | 2.714 | 00:07 |
| Air Blank          | 0.000 | 00:07 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 2 | 0.188 | 00:10 |
| Breath Volume      | 2.257 | 00:10 |
| Air Blank          | 0.000 | 00:10 |
| Control Test       | 0.080 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 09/29/2016 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: ESH,JOSHUA  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 18:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HILFERS

DATE/TIME OF TEST: 09/02/2016 19:18:21

---

|                    |       |       |                 |
|--------------------|-------|-------|-----------------|
| Diagnostics Check  | OK    | 19:11 |                 |
| Air Blank          | 0.000 | 19:12 |                 |
| Control Test       | 0.081 | 19:12 |                 |
| Air Blank          | 0.000 | 19:13 |                 |
| Subject Sample # 1 | 0.371 | 19:14 |                 |
| Breath Volume      | 1.714 | 19:14 |                 |
| Air Blank          | 0.000 | 19:14 |                 |
| Air Blank          | 0.000 | 19:16 |                 |
|                    |       | 19:16 | Improper Sample |
| Air Blank          | 0.000 | 19:17 |                 |
| Control Test       | 0.082 | 19:17 |                 |
| Air Blank          | 0.000 | 19:18 |                 |
| Diagnostics Check  | OK    | 19:18 |                 |

---

Data Download Date/Time: 09/29/2016 09:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: ESH,JOSHUA  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 18:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: HILFERS

DATE/TIME OF TEST: 09/02/2016 19:31:53

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 19:22 |
| Air Blank          | 0.000 | 19:23 |
| Control Test       | 0.083 | 19:23 |
| Air Blank          | 0.000 | 19:24 |
| Subject Sample # 1 | 0.333 | 19:24 |
| Breath Volume      | 1.144 | 19:24 |
| Air Blank          | 0.000 | 19:25 |
| Air Blank          | 0.000 | 19:26 |
| Subject Sample # 2 | 0.366 | 19:27 |
| Breath Volume      | 2.089 | 19:27 |
| Air Blank          | 0.000 | 19:28 |
| Air Blank          | 0.000 | 19:29 |
| Subject Sample # 3 | 0.375 | 19:30 |
| Breath Volume      | 2.617 | 19:30 |
| Air Blank          | 0.000 | 19:30 |
| Control Test       | 0.080 | 19:31 |
| Air Blank          | 0.000 | 19:31 |
| Diagnostics Check  | OK    | 19:31 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HANSELL,ROBERT,E  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: SAYERS,MICHELLE  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 20:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CORTES

DATE/TIME OF TEST: 09/02/2016 20:44:38

---

|                   |       |       |                      |
|-------------------|-------|-------|----------------------|
| Diagnostics Check | OK    | 20:40 |                      |
| Air Blank         | 0.000 | 20:40 |                      |
| Control Test      | 0.080 | 20:41 |                      |
| Air Blank         | 0.000 | 20:41 |                      |
| Subject Sample    | # 1   | 20:43 | Subject Test Refused |
| Breath Volume     | 0.000 | 20:43 |                      |
| Air Blank         | 0.000 | 20:43 |                      |
| Control Test      | 0.081 | 20:44 |                      |
| Air Blank         | 0.000 | 20:44 |                      |
| Diagnostics Check | OK    | 20:44 |                      |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUGO,NEFTALIF  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: BAXTER,MATTHEW  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: VELEZ-ARCE

DATE/TIME OF TEST: 09/03/2016 03:52:30

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:45 |
| Air Blank          | 0.000 | 03:46 |
| Control Test       | 0.081 | 03:46 |
| Air Blank          | 0.000 | 03:47 |
| Subject Sample # 1 | 0.117 | 03:47 |
| Breath Volume      | 2.578 | 03:47 |
| Air Blank          | 0.000 | 03:48 |
| Air Blank          | 0.000 | 03:50 |
| Subject Sample # 2 | 0.113 | 03:51 |
| Breath Volume      | 2.230 | 03:51 |
| Air Blank          | 0.000 | 03:51 |
| Control Test       | 0.081 | 03:51 |
| Air Blank          | 0.000 | 03:52 |
| Diagnostics Check  | OK    | 03:52 |

---

Data Download Date/Time: 09/29/2016 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FORD, GLENN, W  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: FORD, GLENN  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 09/06/2016 04:47:57

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:41 |
| Air Blank          | 0.000 | 04:42 |
| Control Test       | 0.081 | 04:42 |
| Air Blank          | 0.000 | 04:42 |
| Subject Sample # 1 | 0.301 | 04:43 |
| Breath Volume      | 1.398 | 04:43 |
| Air Blank          | 0.000 | 04:43 |
| Air Blank          | 0.000 | 04:45 |
| Subject Sample # 2 | 0.301 | 04:46 |
| Breath Volume      | 1.503 | 04:46 |
| Air Blank          | 0.000 | 04:47 |
| Control Test       | 0.080 | 04:47 |
| Air Blank          | 0.000 | 04:47 |
| Diagnostics Check  | OK    | 04:47 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,JEREMY,R  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: OWENS,J  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: ADKINS

DATE/TIME OF TEST: 09/07/2016 05:20:20

---

|                   |           |       |            |
|-------------------|-----------|-------|------------|
| Diagnostics Check | OK        | 05:15 |            |
| Air Blank         | 0.000     | 05:15 |            |
| Control Test      | 0.081     | 05:16 |            |
| Air Blank         | 0.000     | 05:16 |            |
| Subject Sample    | # 1 0.149 | 05:18 |            |
| Breath Volume     | 1.242     | 05:18 |            |
| Air Blank         |           | 05:19 | Purge Fail |
| Air Blank         | 0.000     | 05:20 |            |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: CAMPOS,MICHAEL  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 04:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: LORENZ

DATE/TIME OF TEST: 09/10/2016 05:00:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:50 |
| Air Blank          | 0.000 | 04:51 |
| Control Test       | 0.081 | 04:51 |
| Air Blank          | 0.000 | 04:51 |
| Subject Sample # 1 | 0.229 | 04:54 |
| Breath Volume      | 1.285 | 04:54 |
| Air Blank          | 0.000 | 04:54 |
| Air Blank          | 0.000 | 04:56 |
| Subject Sample # 2 | 0.230 | 04:58 |
| Breath Volume      | 1.402 | 04:58 |
| Air Blank          | 0.000 | 04:59 |
| Control Test       | 0.079 | 04:59 |
| Air Blank          | 0.000 | 05:00 |
| Diagnostics Check  | OK    | 05:00 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEDULLA,VINCENT,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: ZUNINO,LUIZ  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: LOPEZ-DELGADO

DATE/TIME OF TEST: 09/11/2016 03:14:11

---

|                   |           |       |                |
|-------------------|-----------|-------|----------------|
| Diagnostics Check | OK        | 03:00 |                |
| Air Blank         | 0.000     | 03:01 |                |
| Control Test      | 0.081     | 03:01 |                |
| Air Blank         | 0.000     | 03:02 |                |
| Subject Sample    | # 1 0.113 | 03:03 |                |
| Breath Volume     | 1.171     | 03:03 |                |
| Air Blank         | 0.000     | 03:04 |                |
| Air Blank         | 0.000     | 03:06 |                |
| Subject Sample    | # 2 0.119 | 03:09 | Volume Not Met |
| Breath Volume     | 0.000     | 03:09 |                |
| Air Blank         | 0.000     | 03:10 |                |
| Air Blank         | 0.000     | 03:12 |                |
| Subject Sample    | # 3 0.116 | 03:12 |                |
| Breath Volume     | 1.984     | 03:12 |                |
| Air Blank         | 0.000     | 03:13 |                |
| Control Test      | 0.079     | 03:13 |                |
| Air Blank         | 0.000     | 03:14 |                |
| Diagnostics Check | OK        | 03:14 |                |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: FREEBERN, ROBERT  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 09/12/2016 03:17:11

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:11 |
| Air Blank          | 0.000 | 03:11 |
| Control Test       | 0.081 | 03:11 |
| Air Blank          | 0.000 | 03:12 |
| Subject Sample # 1 | 0.194 | 03:12 |
| Breath Volume      | 2.292 | 03:12 |
| Air Blank          | 0.000 | 03:13 |
| Air Blank          | 0.000 | 03:15 |
| Subject Sample # 2 | 0.182 | 03:15 |
| Breath Volume      | 1.218 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Control Test       | 0.080 | 03:16 |
| Air Blank          | 0.000 | 03:17 |
| Diagnostics Check  | OK    | 03:17 |

---

Data Download Date/Time: 09/29/2016 09:00



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUGO,NEFTALIF  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: BAXTER,MATTHEW  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: HUNTS

DATE/TIME OF TEST: 09/14/2016 01:29:40

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 01:22 |
| Air Blank          | 0.000 | 01:23 |
| Control Test       | 0.081 | 01:23 |
| Air Blank          | 0.000 | 01:24 |
| Subject Sample # 1 | 0.026 | 01:24 |
| Breath Volume      | 1.214 | 01:24 |
| Air Blank          | 0.000 | 01:25 |
| Air Blank          | 0.000 | 01:27 |
| Subject Sample # 2 | 0.029 | 01:28 |
| Breath Volume      | 1.855 | 01:28 |
| Air Blank          | 0.000 | 01:28 |
| Control Test       | 0.082 | 01:29 |
| Air Blank          | 0.000 | 01:29 |
| Diagnostics Check  | OK    | 01:29 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PEDULLA,VINCENT,A  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: PEDULLA,VINCENT  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 21:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MOSQUEDA-NAHUATH

DATE/TIME OF TEST: 09/15/2016 22:27:08

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 22:19 |
| Air Blank          | 0.000 | 22:20 |
| Control Test       | 0.081 | 22:20 |
| Air Blank          | 0.000 | 22:20 |
| Subject Sample # 1 | 0.205 | 22:22 |
| Breath Volume      | 3.097 | 22:22 |
| Air Blank          | 0.000 | 22:23 |
| Air Blank          | 0.000 | 22:24 |
| Subject Sample # 2 | 0.202 | 22:25 |
| Breath Volume      | 3.000 | 22:25 |
| Air Blank          | 0.000 | 22:26 |
| Control Test       | 0.080 | 22:26 |
| Air Blank          | 0.000 | 22:27 |
| Diagnostics Check  | OK    | 22:27 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIZ,JOAQUIN,G  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: MARCELLA,MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MARTINEZ LOPEZ

DATE/TIME OF TEST: 09/17/2016 01:28:43

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:23 |            |
| Air Blank         | 0.000 | 01:23 |            |
| Control Test      | 0.081 | 01:24 |            |
| Air Blank         | 0.000 | 01:24 |            |
| Subject Sample    | # 1   | 01:27 | RFI Detect |
| Breath Volume     | 0.054 | 01:27 |            |
| Air Blank         | 0.000 | 01:27 |            |
| Control Test      | 0.080 | 01:28 | RFI Detect |
| Air Blank         | 0.000 | 01:28 |            |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GIZ, JOAQUIN, G

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: MARCELLA, MICHAEL

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1578837

DRY GAS EXPIRY DATE: 10/16/2016

VIOLATION CODE: DUI

ARREST TIME: 00:09

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MARTINEZ LOPEZ

DATE/TIME OF TEST: 09/17/2016 01:32:11

---

|                   |       |       |            |
|-------------------|-------|-------|------------|
| Diagnostics Check | OK    | 01:30 |            |
| Air Blank         | 0.000 | 01:31 |            |
| Control Test      | 0.081 | 01:31 |            |
| Air Blank         |       | 01:31 | RFI Detect |
| Air Blank         | 0.000 | 01:32 |            |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GIZ, JOAQUIN, G  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: MARCELLA, MICHAEL  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 00:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MARTINEZ LOPEZ

DATE/TIME OF TEST: 09/17/2016 01:45:58

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 01:34 |                |
| Air Blank         |     | 0.000 | 01:34 |                |
| Control Test      |     | 0.081 | 01:35 |                |
| Air Blank         |     | 0.000 | 01:35 |                |
| Subject Sample    | # 1 | 0.122 | 01:38 | Volume Not Met |
| Breath Volume     |     | 0.738 | 01:38 |                |
| Air Blank         |     | 0.000 | 01:39 |                |
| Air Blank         |     | 0.000 | 01:41 |                |
| Subject Sample    | # 2 | 0.139 | 01:44 | Volume Not Met |
| Breath Volume     |     | 0.019 | 01:44 |                |
| Air Blank         |     | 0.000 | 01:45 |                |
| Control Test      |     | 0.080 | 01:45 |                |
| Air Blank         |     | 0.000 | 01:45 |                |
| Diagnostics Check |     | OK    | 01:45 |                |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STINSON,ANDREW,L  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: STINSON,ANDREW  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MERCADO COLON

DATE/TIME OF TEST: 09/17/2016 04:17:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 04:11 |
| Air Blank          | 0.000 | 04:11 |
| Control Test       | 0.081 | 04:11 |
| Air Blank          | 0.000 | 04:12 |
| Subject Sample # 1 | 0.110 | 04:13 |
| Breath Volume      | 2.937 | 04:13 |
| Air Blank          | 0.000 | 04:14 |
| Air Blank          | 0.000 | 04:15 |
| Subject Sample # 2 | 0.106 | 04:16 |
| Breath Volume      | 1.500 | 04:16 |
| Air Blank          | 0.000 | 04:16 |
| Control Test       | 0.081 | 04:17 |
| Air Blank          | 0.000 | 04:17 |
| Diagnostics Check  | OK    | 04:17 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STINSON,ANDREW,L  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: STINSON,ANDREW  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: VENTURA CASTRO

DATE/TIME OF TEST: 09/18/2016 03:53:42

---

|                   |     |       |       |                |
|-------------------|-----|-------|-------|----------------|
| Diagnostics Check |     | OK    | 03:42 |                |
| Air Blank         |     | 0.000 | 03:42 |                |
| Control Test      |     | 0.082 | 03:42 |                |
| Air Blank         |     | 0.000 | 03:43 |                |
| Subject Sample    | # 1 | 0.092 | 03:46 | Volume Not Met |
| Breath Volume     |     | 0.015 | 03:46 |                |
| Air Blank         |     | 0.000 | 03:47 |                |
| Air Blank         |     | 0.000 | 03:48 |                |
| Subject Sample    | # 2 | 0.104 | 03:52 | Volume Not Met |
| Breath Volume     |     | 0.074 | 03:52 |                |
| Air Blank         |     | 0.000 | 03:52 |                |
| Control Test      |     | 0.080 | 03:53 |                |
| Air Blank         |     | 0.000 | 03:53 |                |
| Diagnostics Check |     | OK    | 03:53 |                |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ADAMS,BRIAN,F  
BREATH TEST OPERATOR AGENCY: Osceola County SO  
ARRESTING OFFICER: LAMB,JAMES  
ARRESTING OFFICER AGENCY: Osceola County SO  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 07:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:53

SUBJECT'S LAST NAME: BONILLA ALVAREZ

DATE/TIME OF TEST: 09/25/2016 07:19:23

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 07:13 |
| Air Blank          | 0.000 | 07:13 |
| Control Test       | 0.081 | 07:14 |
| Air Blank          | 0.000 | 07:14 |
| Subject Sample # 1 | 0.071 | 07:14 |
| Breath Volume      | 2.039 | 07:14 |
| Air Blank          | 0.000 | 07:15 |
| Air Blank          | 0.000 | 07:17 |
| Subject Sample # 2 | 0.065 | 07:17 |
| Breath Volume      | 1.765 | 07:17 |
| Air Blank          | 0.000 | 07:18 |
| Control Test       | 0.081 | 07:18 |
| Air Blank          | 0.000 | 07:19 |
| Diagnostics Check  | OK    | 07:19 |

---

Data Download Date/Time: 09/29/2016 09:00



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUGO,NEFTALIF  
BREATH TEST OPERATOR AGENCY: Kissimmee PD  
ARRESTING OFFICER: BAXTER,MATTHEW  
ARRESTING OFFICER AGENCY: Kissimmee PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: FARRISH

DATE/TIME OF TEST: 09/28/2016 03:32:47

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:25 |
| Air Blank          | 0.000 | 03:26 |
| Control Test       | 0.082 | 03:26 |
| Air Blank          | 0.000 | 03:26 |
| Subject Sample # 1 | 0.151 | 03:28 |
| Breath Volume      | 1.214 | 03:28 |
| Air Blank          | 0.000 | 03:28 |
| Air Blank          | 0.000 | 03:30 |
| Subject Sample # 2 | 0.168 | 03:31 |
| Breath Volume      | 1.652 | 03:31 |
| Air Blank          | 0.000 | 03:31 |
| Control Test       | 0.082 | 03:32 |
| Air Blank          | 0.000 | 03:32 |
| Diagnostics Check  | OK    | 03:32 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORTIZ JR,DAMASO,  
BREATH TEST OPERATOR AGENCY: FHP Troop K  
ARRESTING OFFICER: ORTIZ JR,DAMASO  
ARRESTING OFFICER AGENCY: FHP Troop K  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 05:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: TOLEDO

DATE/TIME OF TEST: 09/28/2016 06:13:27

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 06:05 |
| Air Blank          | 0.000 | 06:05 |
| Control Test       | 0.082 | 06:06 |
| Air Blank          | 0.000 | 06:06 |
| Subject Sample # 1 | 0.209 | 06:08 |
| Breath Volume      | 1.414 | 06:08 |
| Air Blank          | 0.000 | 06:09 |
| Air Blank          | 0.000 | 06:11 |
| Subject Sample # 2 | 0.197 | 06:11 |
| Breath Volume      | 1.914 | 06:11 |
| Air Blank          | 0.000 | 06:12 |
| Control Test       | 0.080 | 06:12 |
| Air Blank          | 0.000 | 06:13 |
| Diagnostics Check  | OK    | 06:13 |

---

Data Download Date/Time: 09/29/2016 09:00

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: MONDELLO, ANTHONY  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 08/25/2016

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 23:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 09/29/2016 00:11:55

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 00:05 |
| Air Blank          | 0.000 | 00:05 |
| Control Test       | 0.081 | 00:05 |
| Air Blank          | 0.000 | 00:06 |
| Subject Sample # 1 | 0.000 | 00:06 |
| Breath Volume      | 3.562 | 00:06 |
| Air Blank          | 0.000 | 00:07 |
| Air Blank          | 0.000 | 00:09 |
| Subject Sample # 2 | 0.000 | 00:10 |
| Breath Volume      | 3.714 | 00:10 |
| Air Blank          | 0.000 | 00:11 |
| Control Test       | 0.081 | 00:11 |
| Air Blank          | 0.000 | 00:11 |
| Diagnostics Check  | OK    | 00:11 |

---

Data Download Date/Time: 09/29/2016 09:00

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J  
BREATH TEST OPERATOR AGENCY: St. Cloud PD  
ARRESTING OFFICER: MONDELLO, ANTHONY  
ARRESTING OFFICER AGENCY: St. Cloud PD  
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1578837  
DRY GAS EXPIRY DATE: 10/16/2016  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 09/30/2016 03:20:33

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 03:13 |
| Air Blank          | 0.000 | 03:14 |
| Control Test       | 0.081 | 03:14 |
| Air Blank          | 0.000 | 03:15 |
| Subject Sample # 1 | 0.103 | 03:15 |
| Breath Volume      | 5.050 | 03:15 |
| Air Blank          | 0.000 | 03:16 |
| Air Blank          | 0.000 | 03:18 |
| Subject Sample # 2 | 0.102 | 03:18 |
| Breath Volume      | 5.050 | 03:18 |
| Air Blank          | 0.000 | 03:19 |
| Control Test       | 0.081 | 03:19 |
| Air Blank          | 0.000 | 03:20 |
| Diagnostics Check  | OK    | 03:20 |

---

Data Download Date/Time: 10/14/2016 07:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHEINER,JAMES,R  
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD  
ARRESTING OFFICER: HOBUS,CAVIN  
ARRESTING OFFICER AGENCY: Cocoa Beach PD  
LAST AGENCY INSPECTION: 08/29/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 36314080A3  
DRY GAS EXPIRY DATE: 02/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 23:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 09/05/2016 00:06:05

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 23:58 |
| Air Blank         |     | 0.000 | 23:58 |
| Control Test      |     | 0.078 | 23:59 |
| Air Blank         |     | 0.000 | 23:59 |
| Subject Sample    | # 1 | 0.284 | 00:01 |
| Breath Volume     |     | 2.078 | 00:01 |
| Air Blank         |     | 0.000 | 00:01 |
| Air Blank         |     | 0.000 | 00:03 |
| Subject Sample    | # 2 | 0.280 | 00:04 |
| Breath Volume     |     | 1.410 | 00:04 |
| Air Blank         |     | 0.000 | 00:05 |
| Control Test      |     | 0.077 | 00:05 |
| Air Blank         |     | 0.000 | 00:06 |
| Diagnostics Check |     | OK    | 00:06 |

---

Data Download Date/Time: 10/06/2016 10:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SYPIEN,MARK,A  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: SANCHEZ,JASON  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 08/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01316080A2  
DRY GAS EXPIRY DATE: 02/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 13:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: MEUSEL

DATE/TIME OF TEST: 09/08/2016 14:28:02

---

|                    |       |       |
|--------------------|-------|-------|
| Diagnostics Check  | OK    | 14:21 |
| Air Blank          | 0.000 | 14:21 |
| Control Test       | 0.078 | 14:22 |
| Air Blank          | 0.000 | 14:22 |
| Subject Sample # 1 | 0.261 | 14:23 |
| Breath Volume      | 2.265 | 14:23 |
| Air Blank          | 0.000 | 14:24 |
| Air Blank          | 0.000 | 14:25 |
| Subject Sample # 2 | 0.268 | 14:26 |
| Breath Volume      | 2.113 | 14:26 |
| Air Blank          | 0.000 | 14:27 |
| Control Test       | 0.078 | 14:27 |
| Air Blank          | 0.000 | 14:27 |
| Diagnostics Check  | OK    | 14:28 |

---

Data Download Date/Time: 09/30/2016 09:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOLDEN,ROBERT,A  
BREATH TEST OPERATOR AGENCY: Titusville PD  
ARRESTING OFFICER: SANTIAGO,FRANCES  
ARRESTING OFFICER AGENCY: Titusville PD  
LAST AGENCY INSPECTION: 08/28/2016

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01316080A2  
DRY GAS EXPIRY DATE: 02/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 10:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: MOINE

DATE/TIME OF TEST: 09/17/2016 11:20:48

---

|                   |     |       |       |
|-------------------|-----|-------|-------|
| Diagnostics Check |     | OK    | 11:14 |
| Air Blank         |     | 0.000 | 11:14 |
| Control Test      |     | 0.079 | 11:14 |
| Air Blank         |     | 0.000 | 11:15 |
| Subject Sample    | # 1 | 0.202 | 11:16 |
| Breath Volume     |     | 4.132 | 11:16 |
| Air Blank         |     | 0.000 | 11:16 |
| Air Blank         |     | 0.000 | 11:18 |
| Subject Sample    | # 2 | 0.202 | 11:19 |
| Breath Volume     |     | 4.468 | 11:19 |
| Air Blank         |     | 0.000 | 11:19 |
| Control Test      |     | 0.079 | 11:20 |
| Air Blank         |     | 0.000 | 11:20 |
| Diagnostics Check |     | OK    | 11:20 |

---

Data Download Date/Time: 09/30/2016 09:07