

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HOLDING

DATE/TIME OF TEST: 10/01/2016 01:03:25

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:01	
Control Test	0.082	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:02	
Control Test	0.081	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ GARCIA, REINIER,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HOPP, SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: GOLDBLATT

DATE/TIME OF TEST: 10/03/2016 21:18:22

Diagnostics Check	OK	21:14	
Air Blank	0.000	21:15	
Control Test	0.080	21:15	
Air Blank	0.000	21:15	
Subject Sample	# 1	21:17	Subject Test Refused
Breath Volume	0.000	21:17	
Air Blank	0.000	21:17	
Control Test	0.081	21:17	
Air Blank	0.000	21:18	
Diagnostics Check	OK	21:18	

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SCHWARTZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: YONSON

DATE/TIME OF TEST: 10/04/2016 23:52:59

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:51	
Control Test	0.082	23:52	
Air Blank		23:52	RFI Detect
Air Blank	0.000	23:52	

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: SCHWARTZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: YONSON

DATE/TIME OF TEST: 10/05/2016 00:00:12

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.081	23:58	
Air Blank	0.000	23:58	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 10/09/2016 22:22:58

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.199	22:18
Breath Volume	2.132	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:20
Subject Sample # 2	0.194	22:21
Breath Volume	2.769	22:21
Air Blank	0.000	22:22
Control Test	0.081	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RIVERA HERNANDEZ

DATE/TIME OF TEST: 10/10/2016 04:06:40

Diagnostics Check	OK	04:03	
Air Blank	0.000	04:04	
Control Test	0.081	04:04	
Air Blank	0.000	04:05	
Subject Sample	# 1	04:05	Subject Test Refused
Breath Volume	0.000	04:05	
Air Blank	0.000	04:05	
Control Test	0.080	04:06	
Air Blank	0.000	04:06	
Diagnostics Check	OK	04:06	

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: SPANULO

DATE/TIME OF TEST: 10/13/2016 00:43:00

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:38	
Control Test	0.081	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANTON, TYLER, A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS, WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 10/16/2016 06:13:03

Diagnostics Check	OK	06:08	
Air Blank	0.000	06:09	
Control Test	0.081	06:09	
Air Blank	0.000	06:10	
Subject Sample	# 1	06:11	Subject Test Refused
Breath Volume	0.000	06:11	
Air Blank	0.000	06:12	
Control Test	0.081	06:12	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:13	

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANTON, TYLER, A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS, WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 10/16/2016 06:21:00

Diagnostics Check	OK	06:17	
Air Blank	0.000	06:18	
Control Test	0.081	06:18	
Air Blank	0.000	06:19	
Subject Sample	# 1	06:20	Sequence Aborted
Breath Volume	1.789	06:20	
Air Blank		06:21	RFI Detect

Data Download Date/Time: 10/24/2016 07:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DANIELS,WILLIAM
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/24/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: ALDRIDGE

DATE/TIME OF TEST: 10/29/2016 21:38:44

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.081	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.197	21:33
Breath Volume	1.542	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.180	21:37
Breath Volume	2.257	21:37
Air Blank	0.000	21:37
Control Test	0.081	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 11/16/2016 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: GIROD

DATE/TIME OF TEST: 10/04/2016 22:45:14

Diagnostics Check	OK	22:38
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.235	22:40
Breath Volume	1.792	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.218	22:43
Breath Volume	1.390	22:43
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 10/06/2016 04:55:23

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.080	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.150	04:50
Breath Volume	1.937	04:50
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.153	04:53
Breath Volume	2.546	04:53
Air Blank	0.000	04:54
Control Test	0.078	04:54
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERLMAN,GREGG,B
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: PERLMAN,GREGG
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: SCHILLING

DATE/TIME OF TEST: 10/09/2016 05:06:57

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.157	04:58
Breath Volume	1.644	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:00
Subject Sample # 2	0.136	05:02
Breath Volume	2.816	05:02
Air Blank	0.000	05:03
Air Blank	0.000	05:04
Subject Sample # 3	0.143	05:05
Breath Volume	2.320	05:05
Air Blank	0.000	05:06
Control Test	0.077	05:06
Air Blank	0.000	05:06
Diagnostics Check	OK	05:06

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: JAROSZEWSKI

DATE/TIME OF TEST: 10/10/2016 05:13:16

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:05	
Control Test	0.080	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:09	No Sample Provided
Breath Volume	0.000	05:09	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:12	
Control Test	0.079	05:12	
Air Blank	0.000	05:13	
Diagnostics Check	OK	05:13	

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,AUSTIN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 091602
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 10/13/2016 06:35:30

Diagnostics Check	OK	06:25
Air Blank	0.000	06:25
Control Test	0.081	06:25
Air Blank	0.000	06:26
Subject Sample # 1	0.209	06:29
Breath Volume	1.515	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.204	06:33
Breath Volume	1.789	06:33
Air Blank	0.000	06:34
Control Test	0.078	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: DURBIN

DATE/TIME OF TEST: 10/14/2016 05:44:43

Diagnostics Check	OK	05:41	
Air Blank	0.000	05:42	
Control Test	0.079	05:42	
Air Blank	0.000	05:43	
Subject Sample	# 1	05:43	Subject Test Refused
Breath Volume	0.000	05:43	
Air Blank	0.000	05:43	
Control Test	0.079	05:44	
Air Blank	0.000	05:44	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CALHOUN,EVAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: COLLETTI

DATE/TIME OF TEST: 10/16/2016 01:56:00

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.080	01:44	
Air Blank	0.000	01:44	
Subject Sample	# 1 0.156	01:47	Volume Not Met
Breath Volume	0.515	01:47	
Air Blank	0.000	01:48	
Air Blank	0.000	01:50	
Subject Sample	# 2 0.178	01:50	
Breath Volume	1.363	01:50	
Air Blank	0.000	01:51	
Air Blank	0.000	01:53	
Subject Sample	# 3 0.195	01:54	
Breath Volume	1.562	01:54	
Air Blank	0.000	01:55	
Control Test	0.079	01:55	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:56	

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BETHEL

DATE/TIME OF TEST: 10/18/2016 18:41:44

Diagnostics Check	OK	18:38	
Air Blank	0.000	18:39	
Control Test	0.080	18:39	
Air Blank	0.000	18:39	
Subject Sample	# 1	18:40	Subject Test Refused
Breath Volume	0.000	18:40	
Air Blank	0.000	18:40	
Control Test	0.080	18:41	
Air Blank	0.000	18:41	
Diagnostics Check	OK	18:41	

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DAVIS,ALEXANDRA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ST LOUIS

DATE/TIME OF TEST: 10/21/2016 01:37:44

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.081	01:33	
Air Blank	0.000	01:33	
Subject Sample # 1	0.163	01:34	
Breath Volume	1.804	01:34	
Air Blank	0.000	01:34	
Air Blank		01:36	Improper Sample
Air Blank	0.000	01:36	
Control Test	0.080	01:37	
Air Blank	0.000	01:37	
Diagnostics Check	OK	01:37	

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,ALEXANDRA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DAVIS,ALEXANDRA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ST LOUIS

DATE/TIME OF TEST: 10/21/2016 01:46:00

Diagnostics Check	OK	01:40
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.165	01:41
Breath Volume	1.425	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.161	01:44
Breath Volume	1.445	01:44
Air Blank	0.000	01:45
Control Test	0.080	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:46

Data Download Date/Time: 10/23/2016 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAUFILLET,ALEXANDRE,J
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: GAUFILLET,ALEXANDRE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 15:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:26

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 10/25/2016 16:56:33

Diagnostics Check	OK	16:53	
Air Blank	0.000	16:54	
Control Test	0.083	16:54	
Air Blank	0.000	16:55	
Subject Sample	# 1	16:55	Subject Test Refused
Breath Volume	0.000	16:55	
Air Blank	0.000	16:55	
Control Test	0.081	16:55	
Air Blank	0.000	16:56	
Diagnostics Check	OK	16:56	

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CHIGAS

DATE/TIME OF TEST: 10/27/2016 03:40:33

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.082	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.165	03:33
Breath Volume	3.718	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.144	03:36
Breath Volume	2.367	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 3	0.138	03:39
Breath Volume	1.785	03:39
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: HELMS

DATE/TIME OF TEST: 10/27/2016 19:59:31

Diagnostics Check	OK	19:56	
Air Blank	0.000	19:57	
Control Test	0.082	19:57	
Air Blank	0.000	19:57	
Subject Sample	# 1	19:58	Subject Test Refused
Breath Volume	0.000	19:58	
Air Blank	0.000	19:58	
Control Test	0.081	19:58	
Air Blank	0.000	19:59	
Diagnostics Check	OK	19:59	

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: BASHAM

DATE/TIME OF TEST: 10/28/2016 03:12:21

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:09	
Control Test	0.081	03:10	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:10	Subject Test Refused
Breath Volume	0.000	03:10	
Air Blank	0.000	03:11	
Control Test	0.082	03:11	
Air Blank	0.000	03:12	
Diagnostics Check	OK	03:12	

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CURRUL,ROBERT
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: CANIZARES

DATE/TIME OF TEST: 10/28/2016 04:30:36

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.081	04:26	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.081	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: DEL RIO

DATE/TIME OF TEST: 10/29/2016 00:10:00

Diagnostics Check	OK	00:07	
Air Blank	0.000	00:07	
Control Test	0.081	00:07	
Air Blank	0.000	00:08	
Subject Sample	# 1	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:09	
Control Test	0.081	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALHOUN, EVAN, R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: CALHOUN, EVAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: VENEGAS

DATE/TIME OF TEST: 10/29/2016 00:21:09

Diagnostics Check	OK	00:16	
Air Blank	0.000	00:17	
Control Test	0.082	00:17	
Air Blank	0.000	00:18	
Subject Sample	# 1	00:19	Subject Test Refused
Breath Volume	0.000	00:19	
Air Blank	0.000	00:20	
Control Test	0.081	00:20	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KORZEN,GREG,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KORZEN,GREG
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TENGCO

DATE/TIME OF TEST: 10/30/2016 02:05:46

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:03	
Control Test	0.082	02:03	
Air Blank	0.000	02:04	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:04	
Control Test	0.081	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 11/11/2016 08:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,DANIEL,J
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/06/2016 02:45:14

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.000	02:40
Breath Volume	2.625	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.000	02:43
Breath Volume	2.191	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 10/28/2016 15:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMANN, ANDRE,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HAMMANN, ANDRE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MONTANO BOSQUE

DATE/TIME OF TEST: 10/10/2016 05:08:13

Diagnostics Check	OK	05:05	
Air Blank	0.000	05:05	
Control Test	0.080	05:06	
Air Blank	0.000	05:06	
Subject Sample	# 1	05:06	Subject Test Refused
Breath Volume	0.000	05:06	
Air Blank	0.000	05:07	
Control Test	0.080	05:07	
Air Blank	0.000	05:08	
Diagnostics Check	OK	05:08	

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ST AMAND,STEVE,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUACES,THERESA
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: RAMS

DATE/TIME OF TEST: 10/10/2016 21:36:32

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.080	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.179	21:32
Breath Volume	1.648	21:32
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.183	21:35
Breath Volume	1.871	21:35
Air Blank	0.000	21:35
Control Test	0.080	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/11/2016 02:09:38

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.081	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.000	02:05
Breath Volume	1.140	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.000	02:08
Breath Volume	1.359	02:08
Air Blank	0.000	02:08
Control Test	0.080	02:09
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HAMMANN,ANDRE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: CASTANEDA QUINTANA

DATE/TIME OF TEST: 10/11/2016 02:55:29

Diagnostics Check	OK	02:49
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:50
Subject Sample # 1	0.241	02:51
Breath Volume	2.160	02:51
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.222	02:53
Breath Volume	1.902	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: RUIZMARTINEZ

DATE/TIME OF TEST: 10/13/2016 02:37:20

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RORIGUEZ,JESSICA,Y
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: LUKSA,ANDREW
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: PINZON

DATE/TIME OF TEST: 10/14/2016 03:56:07

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.169	03:49
Breath Volume	1.199	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.165	03:54
Breath Volume	1.507	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ARISTIZABAL RAMIREZ

DATE/TIME OF TEST: 10/17/2016 00:16:37

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.242	00:11
Breath Volume	1.398	00:11
Air Blank	0.000	00:12
Air Blank	0.000	00:13
Subject Sample # 2	0.240	00:14
Breath Volume	1.589	00:14
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:16

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMANN,ANDRE,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HAMMANN,ANDRE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DAVARO

DATE/TIME OF TEST: 10/17/2016 02:57:46

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.080	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.080	02:57	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMANN,ANDRE,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: FUNTES,ALEJANDRO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: CASAS

DATE/TIME OF TEST: 10/21/2016 03:08:07

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:05	
Control Test	0.080	03:06	
Air Blank	0.000	03:06	
Subject Sample	# 1	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:07	
Control Test	0.080	03:07	
Air Blank	0.000	03:08	
Diagnostics Check	OK	03:08	

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMANN,ANDRE,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HAMMANN,ANDRE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/28/2016 02:53:54

Diagnostics Check	OK	02:50	
Air Blank	0.000	02:51	
Control Test	0.080	02:51	
Air Blank	0.000	02:52	
Subject Sample	# 1	02:52	Subject Test Refused
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Control Test	0.079	02:53	
Air Blank	0.000	02:53	
Diagnostics Check	OK	02:53	

Data Download Date/Time: 10/28/2016 15:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000871

INSTRUMENT REGISTERED TO: CORAL GABLES PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORES, EMILIO, E
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: CHAVEZ, NESTOR
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 10:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:48

SUBJECT'S LAST NAME: HERNANDEZ ROMERO

DATE/TIME OF TEST: 10/30/2016 11:12:22

Diagnostics Check	OK	11:09	
Air Blank	0.000	11:09	
Control Test	0.080	11:10	
Air Blank	0.000	11:10	
Subject Sample	# 1	11:11	Subject Test Refused
Breath Volume	0.000	11:11	
Air Blank	0.000	11:11	
Control Test	0.080	11:11	
Air Blank	0.000	11:12	
Diagnostics Check	OK	11:12	

Data Download Date/Time: 11/28/2016 14:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 10/17/2016 04:00:18

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.079	03:57	
Air Blank	0.000	03:58	
Subject Sample	# 1	03:58	Subject Test Refused
Breath Volume	0.000	03:58	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 11/29/2016 16:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTEAGA, GUILLERMO, J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MILLER, VINCENT
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 10/17/2016 10:37:29

Diagnostics Check	OK	10:34	
Air Blank	0.000	10:35	
Control Test	0.079	10:35	
Air Blank	0.000	10:35	
Subject Sample	# 1	10:36	Subject Test Refused
Breath Volume	0.000	10:36	
Air Blank	0.000	10:36	
Control Test	0.079	10:36	
Air Blank	0.000	10:37	
Diagnostics Check	OK	10:37	

Data Download Date/Time: 11/29/2016 16:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERRELUS,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CHERRELUS,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: RAMS

DATE/TIME OF TEST: 10/29/2016 01:51:52

Diagnostics Check	OK	01:40	
Air Blank	0.000	01:41	
Control Test	0.077	01:41	
Air Blank	0.000	01:42	
Subject Sample	# 1 0.182	01:43	
Breath Volume	2.386	01:43	
Air Blank	0.000	01:43	
Air Blank	0.000	01:45	
Subject Sample	# 2 0.186	01:46	
Breath Volume	2.753	01:46	
Air Blank	0.000	01:51	
Control Test	0.068	01:51	Control Outside Tolerance
Air Blank		01:51	Improper Sample

Data Download Date/Time: 11/29/2016 16:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SANCHEZ,DANIL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/14/2016 22:28:24

Diagnostics Check	OK	22:24	
Air Blank	0.000	22:25	
Control Test	0.078	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Slope Not Met
Breath Volume	1.277	22:26	
Air Blank	0.000	22:27	
Control Test	0.078	22:27	
Air Blank	0.000	22:28	
Diagnostics Check	OK	22:28	

Data Download Date/Time: 11/29/2016 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RAMIREZ,DANIL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/14/2016 22:55:41

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:49
Subject Sample	# 1 0.187	22:50
Breath Volume	1.789	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample	# 2 0.191	22:54
Breath Volume	2.617	22:54
Air Blank	0.000	22:54
Control Test	0.079	22:55
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 11/29/2016 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARDELAS,MAYLEE,

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: BARDELAS,MAYLEE

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 10/14/2016 23:06:36

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.079	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:06	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 11/29/2016 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: JOSE

DATE/TIME OF TEST: 10/15/2016 00:14:15

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.079	00:12	
Air Blank	0.000	00:12	
Subject Sample	# 1	00:12	Subject Test Refused
Breath Volume	0.000	00:12	
Air Blank	0.000	00:13	
Control Test	0.079	00:13	
Air Blank	0.000	00:14	
Diagnostics Check	OK	00:14	

Data Download Date/Time: 11/29/2016 16:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARDELAS,MAYLE,

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: BARDELAS,MAYLEE

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/15/2016 00:56:04

Diagnostics Check	OK	00:55	
Air Blank		00:55	RFI Detect
Air Blank	0.000	00:56	

Data Download Date/Time: 11/29/2016 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARDELAS,MAYLEE,

BREATH TEST OPERATOR AGENCY: OTHER Local

ARRESTING OFFICER: BARDELAS,MAYLE

ARRESTING OFFICER AGENCY: OTHER Local

LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/15/2016 01:08:21

Diagnostics Check	OK	00:58	
Air Blank	0.000	00:58	
Control Test	0.079	00:59	
Air Blank	0.000	00:59	
Subject Sample	# 1 0.131	01:02	Volume Not Met
Breath Volume	0.410	01:02	
Air Blank	0.000	01:03	
Air Blank	0.000	01:05	
Subject Sample	# 2	01:06	Subject Test Refused
Breath Volume	1.070	01:06	
Air Blank	0.000	01:07	
Control Test	0.080	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 11/29/2016 16:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000876

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUQUE,ALBERTO,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: DUQUE,ALBERTO
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 09/21/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 10/09/2016 22:38:18

Diagnostics Check	OK	22:29	
Air Blank	0.000	22:29	
Control Test	0.077	22:30	
Air Blank	0.000	22:30	
Subject Sample	# 1	22:33	No Sample Provided
Breath Volume	0.000	22:33	
Air Blank	0.000	22:34	
Air Blank	0.000	22:36	
Subject Sample	# 2	22:36	Subject Test Refused
Breath Volume	0.000	22:36	
Air Blank	0.000	22:37	
Control Test	0.076	22:37	
Air Blank	0.000	22:38	
Diagnostics Check	OK	22:38	

Data Download Date/Time: 10/12/2016 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: NERI

DATE/TIME OF TEST: 10/28/2016 00:40:23

Diagnostics Check	OK	00:32
Air Blank	0.000	00:32
Control Test	0.080	00:33
Air Blank	0.000	00:33
Subject Sample # 1	0.135	00:35
Breath Volume	3.488	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.142	00:38
Breath Volume	3.234	00:38
Air Blank	0.000	00:39
Control Test	0.083	00:39
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 12/02/2016 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: DIEGO

DATE/TIME OF TEST: 10/29/2016 15:36:59

Diagnostics Check	OK	15:29
Air Blank	0.000	15:30
Control Test	0.080	15:30
Air Blank	0.000	15:31
Subject Sample # 1	0.257	15:32
Breath Volume	1.968	15:32
Air Blank	0.000	15:32
Air Blank	0.000	15:34
Subject Sample # 2	0.252	15:35
Breath Volume	1.667	15:35
Air Blank	0.000	15:36
Control Test	0.079	15:36
Air Blank	0.000	15:36
Diagnostics Check	OK	15:36

Data Download Date/Time: 12/02/2016 08:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,JOSE,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CHEVALIER,J.
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1673300
DRY GAS EXPIRY DATE: 04/28/2017
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:49

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/30/2016 05:20:54

Diagnostics Check	OK	05:18	
Air Blank	0.000	05:18	
Control Test	0.080	05:18	
Air Blank	0.000	05:19	
Subject Sample	# 1	05:19	Subject Test Refused
Breath Volume	0.000	05:19	
Air Blank	0.000	05:19	
Control Test	0.080	05:20	
Air Blank	0.000	05:20	
Diagnostics Check	OK	05:20	

Data Download Date/Time: 12/02/2016 08:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: FABREGAT

DATE/TIME OF TEST: 10/18/2016 04:03:11

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:00	
Control Test	0.081	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:01	Subject Test Refused
Breath Volume	0.000	04:01	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 11/17/2016 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,JASON,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WAITE,JASON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:54

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 10/23/2016 06:21:42

Diagnostics Check	OK	06:15	
Air Blank	0.000	06:15	
Control Test	0.080	06:16	
Air Blank	0.000	06:16	
Subject Sample	# 1 0.191	06:17	
Breath Volume	2.703	06:17	
Air Blank	0.000	06:17	
Air Blank	0.000	06:19	
Subject Sample	# 2 0.181	06:20	
Breath Volume	3.105	06:20	
Air Blank	0.000	06:20	
Control Test	0.080	06:21	
Air Blank		06:21	RFI Detect
Air Blank	0.000	06:21	

Data Download Date/Time: 11/17/2016 13:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAITE,JASON,B
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: WAITE,JASON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:52

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 10/23/2016 06:29:28

Diagnostics Check	OK	06:23
Air Blank	0.000	06:23
Control Test	0.080	06:24
Air Blank	0.000	06:24
Subject Sample # 1	0.186	06:25
Breath Volume	3.214	06:25
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.181	06:27
Breath Volume	2.886	06:27
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 11/17/2016 13:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 05:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/01/2016 06:16:28

Diagnostics Check	OK	06:13	
Air Blank	0.000	06:14	
Control Test	0.079	06:14	
Air Blank	0.000	06:14	
Subject Sample	# 1	06:15	Subject Test Refused
Breath Volume	0.000	06:15	
Air Blank	0.000	06:15	
Control Test	0.078	06:15	
Air Blank	0.000	06:16	
Diagnostics Check	OK	06:16	

Data Download Date/Time: 10/13/2016 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: PUAC-MALDONADO

DATE/TIME OF TEST: 10/06/2016 02:05:06

Diagnostics Check	OK	01:54
Air Blank	0.000	01:55
Control Test	0.077	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.321	01:57
Breath Volume	1.351	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.322	02:03
Breath Volume	1.214	02:03
Air Blank	0.000	02:04
Control Test	0.077	02:04
Air Blank	0.000	02:04
Diagnostics Check	OK	02:05

Data Download Date/Time: 10/13/2016 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: PEOPLES

DATE/TIME OF TEST: 10/06/2016 04:36:15

Diagnostics Check	OK	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Subject Sample # 1	0.157	04:31
Breath Volume	1.367	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.160	04:34
Breath Volume	1.699	04:34
Air Blank	0.000	04:35
Control Test	0.078	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 10/13/2016 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: VASALLO

DATE/TIME OF TEST: 10/08/2016 01:34:44

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.077	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.127	01:30
Breath Volume	1.894	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.137	01:33
Breath Volume	1.964	01:33
Air Blank	0.000	01:33
Control Test	0.077	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 10/13/2016 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 10/08/2016 03:16:53

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.078	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.203	03:11
Breath Volume	1.589	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.214	03:15
Breath Volume	2.460	03:15
Air Blank	0.000	03:15
Control Test	0.077	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 10/13/2016 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: SAN JUAN TEJEDA

DATE/TIME OF TEST: 10/12/2016 03:49:06

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.079	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.078	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 10/13/2016 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/13/2016 00:24:37

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:22	
Control Test	0.078	00:22	
Air Blank	0.000	00:23	
Subject Sample	# 1	00:23	Subject Test Refused
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Control Test	0.078	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 10/13/2016 10:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/14/2016 02:57:11

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.078	02:43	
Air Blank	0.000	02:44	
Subject Sample	# 1 0.158	02:45	
Breath Volume	1.917	02:45	
Air Blank	0.000	02:45	
Air Blank	0.000	02:47	
Subject Sample	# 2 0.120	02:50	Volume Not Met
Breath Volume	0.535	02:50	
Air Blank	0.000	02:51	
Air Blank	0.000	02:53	
Subject Sample	# 3 0.143	02:55	
Breath Volume	1.507	02:55	
Air Blank	0.000	02:56	
Control Test	0.077	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 11/09/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: HERRERA-GRACIA

DATE/TIME OF TEST: 10/15/2016 01:51:07

Diagnostics Check	OK	01:44
Air Blank	0.000	01:45
Control Test	0.078	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.157	01:46
Breath Volume	2.179	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.158	01:49
Breath Volume	2.394	01:49
Air Blank	0.000	01:50
Control Test	0.076	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 11/09/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: GOMEZDEROSAS

DATE/TIME OF TEST: 10/15/2016 04:16:26

Diagnostics Check	OK	04:13	
Air Blank	0.000	04:14	
Control Test	0.078	04:14	
Air Blank	0.000	04:14	
Subject Sample	# 1	04:15	Subject Test Refused
Breath Volume	0.000	04:15	
Air Blank	0.000	04:15	
Control Test	0.078	04:15	
Air Blank	0.000	04:16	
Diagnostics Check	OK	04:16	

Data Download Date/Time: 11/09/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: JUAREZ REY

DATE/TIME OF TEST: 10/19/2016 04:00:57

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:58	
Control Test	0.079	03:58	
Air Blank	0.000	03:59	
Subject Sample	# 1	03:59	Subject Test Refused
Breath Volume	0.000	03:59	
Air Blank	0.000	04:00	
Control Test	0.078	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 11/09/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 10/20/2016 02:11:33

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.078	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.000	02:07
Breath Volume	1.855	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.000	02:10
Breath Volume	1.574	02:10
Air Blank	0.000	02:10
Control Test	0.077	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 11/09/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 10/21/2016 02:30:09

Diagnostics Check	OK	02:27	
Air Blank	0.000	02:27	
Control Test	0.078	02:28	
Air Blank	0.000	02:28	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.078	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 11/09/2016 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19014080A3
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: UGARTE

DATE/TIME OF TEST: 10/29/2016 01:36:34

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:34	
Control Test	0.079	01:34	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:35	Subject Test Refused
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Control Test	0.078	01:35	
Air Blank	0.000	01:36	
Diagnostics Check	OK	01:36	

Data Download Date/Time: 11/09/2016 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLARDO,JORGE,L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GALLARDO,JORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GAMARRA

DATE/TIME OF TEST: 10/03/2016 00:40:07

Diagnostics Check	OK	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:38	
Air Blank	0.000	00:38	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.078	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

Data Download Date/Time: 10/10/2016 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 10/05/2016 22:31:05

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.078	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.158	22:26
Breath Volume	1.199	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:28
Subject Sample # 2	0.152	22:29
Breath Volume	2.226	22:29
Air Blank	0.000	22:30
Control Test	0.077	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:31

Data Download Date/Time: 10/10/2016 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DE VELASCO

DATE/TIME OF TEST: 10/08/2016 03:01:05

Diagnostics Check	OK	02:54
Air Blank	0.000	02:54
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.247	02:56
Breath Volume	1.632	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.248	02:59
Breath Volume	1.468	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:01

Data Download Date/Time: 10/10/2016 10:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/13/2016 01:03:21

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.079	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.078	01:02	
Air Blank	0.000	01:03	
Diagnostics Check	OK	01:03	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CARBALLOSA ROJAS

DATE/TIME OF TEST: 10/15/2016 04:03:08

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.155	03:58
Breath Volume	2.003	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.156	04:01
Breath Volume	2.039	04:01
Air Blank	0.000	04:02
Control Test	0.077	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RODRIGUEZ,YOEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SAVEDRA

DATE/TIME OF TEST: 10/18/2016 02:34:46

Diagnostics Check	OK	02:27
Air Blank	0.000	02:27
Control Test	0.079	02:28
Air Blank	0.000	02:28
Subject Sample # 1	0.091	02:29
Breath Volume	1.285	02:29
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.100	02:33
Breath Volume	2.128	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: COLVENBACK

DATE/TIME OF TEST: 10/19/2016 04:25:31

Diagnostics Check	OK	04:18
Air Blank	0.000	04:19
Control Test	0.079	04:19
Air Blank	0.000	04:20
Subject Sample # 1	0.186	04:20
Breath Volume	3.339	04:20
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.184	04:23
Breath Volume	2.582	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:42

SUBJECT'S LAST NAME: ADRIANO AVILES

DATE/TIME OF TEST: 10/19/2016 06:12:17

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.078	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.132	06:07
Breath Volume	2.023	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:09
Subject Sample # 2	0.132	06:10
Breath Volume	2.742	06:10
Air Blank	0.000	06:11
Control Test	0.078	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: FRANCHI

DATE/TIME OF TEST: 10/20/2016 02:27:04

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.072	02:22
Breath Volume	2.343	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.074	02:25
Breath Volume	3.152	02:25
Air Blank	0.000	02:26
Control Test	0.078	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:27

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: HOLTZCLAW

DATE/TIME OF TEST: 10/20/2016 05:04:56

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.143	04:58
Breath Volume	1.316	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:01
Subject Sample # 2	0.145	05:03
Breath Volume	1.406	05:03
Air Blank	0.000	05:04
Control Test	0.077	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 10/21/2016 00:36:48

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:35	Subject Test Refused
Breath Volume	0.000	00:35	
Air Blank	0.000	00:35	
Control Test	0.078	00:36	
Air Blank	0.000	00:36	
Diagnostics Check	OK	00:36	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,OMAIRA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 10/21/2016 03:18:31

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.078	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUAN,RALPH,B
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: CUAN,RALPH
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 10/22/2016 02:47:41

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:45	
Control Test	0.079	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:46	Subject Test Refused
Breath Volume	0.000	02:46	
Air Blank	0.000	02:46	
Control Test	0.078	02:47	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: NAVAS

DATE/TIME OF TEST: 10/22/2016 03:46:57

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.172	03:42
Breath Volume	2.097	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.183	03:45
Breath Volume	1.925	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 10/23/2016 00:01:13

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.079	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.165	23:55
Breath Volume	1.632	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.160	23:59
Breath Volume	1.250	23:59
Air Blank	0.000	00:00
Control Test	0.077	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLARDO,JORGE,L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SOTO,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: PEKHTEREV

DATE/TIME OF TEST: 10/26/2016 02:16:46

Diagnostics Check	OK	02:13	
Air Blank	0.000	02:14	
Control Test	0.079	02:14	
Air Blank	0.000	02:15	
Subject Sample	# 1	02:15	Subject Test Refused
Breath Volume	0.000	02:15	
Air Blank	0.000	02:15	
Control Test	0.077	02:16	
Air Blank	0.000	02:16	
Diagnostics Check	OK	02:16	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 10/26/2016 03:05:56

Diagnostics Check	OK	03:03	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.078	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:05	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: PELEGRIN CRUZ

DATE/TIME OF TEST: 10/27/2016 02:35:52

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:27	
Control Test	0.078	02:27	
Air Blank	0.000	02:27	
Subject Sample	# 1 0.178	02:31	Volume Not Met
Breath Volume	0.269	02:31	
Air Blank	0.000	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 2 0.183	02:34	
Breath Volume	3.378	02:34	
Air Blank		02:35	Purge Fail
Air Blank	0.000	02:35	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: PELEGIN CRUZ

DATE/TIME OF TEST: 10/27/2016 03:04:29

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.075	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.179	02:59
Breath Volume	1.949	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.172	03:02
Breath Volume	1.890	03:02
Air Blank	0.000	03:03
Control Test	0.076	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: SALCEDO

DATE/TIME OF TEST: 10/28/2016 03:51:26

Diagnostics Check	OK	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.110	03:46
Breath Volume	1.281	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.111	03:49
Breath Volume	1.683	03:49
Air Blank	0.000	03:50
Control Test	0.077	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: DOMINGUEZ LAZONCET

DATE/TIME OF TEST: 10/28/2016 07:21:34

Diagnostics Check	OK	07:12
Air Blank	0.000	07:13
Control Test	0.078	07:13
Air Blank	0.000	07:14
Subject Sample # 1	0.140	07:15
Breath Volume	1.472	07:15
Air Blank	0.000	07:16
Air Blank	0.000	07:18
Subject Sample # 2	0.134	07:19
Breath Volume	2.421	07:19
Air Blank	0.000	07:20
Control Test	0.079	07:20
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 10/29/2016 02:04:15

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.185	01:58
Breath Volume	1.234	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.179	02:02
Breath Volume	1.562	02:02
Air Blank	0.000	02:03
Control Test	0.075	02:03
Air Blank	0.000	02:04
Diagnostics Check	OK	02:04

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:29

SUBJECT'S LAST NAME: BRUNO FORNERETO

DATE/TIME OF TEST: 10/29/2016 06:56:58

Diagnostics Check	OK	06:50	
Air Blank	0.000	06:50	
Control Test	0.078	06:51	
Air Blank	0.000	06:51	
Subject Sample	# 1 0.128	06:52	
Breath Volume	3.871	06:52	
Air Blank	0.000	06:52	
Air Blank	0.000	06:54	
Subject Sample	# 2 0.135	06:55	
Breath Volume	3.597	06:55	
Air Blank	0.000	06:56	
Control Test	0.074	06:56	Control Outside Tolerance
Air Blank	0.000	06:56	

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: ALVAREZ ACOSTA

DATE/TIME OF TEST: 10/29/2016 08:56:25

Diagnostics Check	OK	08:49
Air Blank	0.000	08:49
Control Test	0.078	08:50
Air Blank	0.000	08:50
Subject Sample # 1	0.141	08:51
Breath Volume	1.847	08:51
Air Blank	0.000	08:52
Air Blank	0.000	08:54
Subject Sample # 2	0.141	08:54
Breath Volume	1.335	08:54
Air Blank	0.000	08:55
Control Test	0.077	08:55
Air Blank	0.000	08:56
Diagnostics Check	OK	08:56

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BONILLA,JOSE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/10/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: CHAVARRO CUBILLOS

DATE/TIME OF TEST: 10/30/2016 05:51:46

Diagnostics Check	OK	05:45
Air Blank	0.000	05:45
Control Test	0.078	05:46
Air Blank	0.000	05:46
Subject Sample # 1	0.206	05:47
Breath Volume	2.867	05:47
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.207	05:50
Breath Volume	3.042	05:50
Air Blank	0.000	05:50
Control Test	0.076	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 11/08/2016 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MUNOZ,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/02/2016 00:46:17

Diagnostics Check	OK	00:38
Air Blank	0.000	00:39
Control Test	0.082	00:39
Air Blank	0.000	00:40
Subject Sample # 1	0.279	00:41
Breath Volume	2.210	00:41
Air Blank	0.000	00:42
Air Blank	0.000	00:43
Subject Sample # 2	0.281	00:44
Breath Volume	2.031	00:44
Air Blank	0.000	00:45
Control Test	0.082	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 10/12/2016 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 10/02/2016 05:23:04

Diagnostics Check	OK	05:22	
Air Blank		05:22	Ambient Fail
Air Blank	0.000	05:23	

Data Download Date/Time: 10/12/2016 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CASTILLO,CHRISTIAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: JAIMES

DATE/TIME OF TEST: 10/02/2016 05:31:01

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.082	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.070	05:26
Breath Volume	2.082	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.066	05:29
Breath Volume	1.886	05:29
Air Blank	0.000	05:30
Control Test	0.082	05:30
Air Blank	0.000	05:30
Diagnostics Check	OK	05:31

Data Download Date/Time: 10/12/2016 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: DENSON

DATE/TIME OF TEST: 10/15/2016 03:42:56

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.082	03:42	
Air Blank	0.000	03:42	
Diagnostics Check	OK	03:42	

Data Download Date/Time: 11/11/2016 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDEZ,MICHAEL,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:18

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/17/2016 13:45:51

Diagnostics Check	OK	13:39
Air Blank	0.000	13:40
Control Test	0.082	13:40
Air Blank	0.000	13:40
Subject Sample # 1	0.294	13:41
Breath Volume	1.863	13:41
Air Blank	0.000	13:41
Air Blank	0.000	13:43
Subject Sample # 2	0.287	13:44
Breath Volume	1.441	13:44
Air Blank	0.000	13:44
Control Test	0.082	13:45
Air Blank	0.000	13:45
Diagnostics Check	OK	13:45

Data Download Date/Time: 11/11/2016 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: ACOSTA BANKS

DATE/TIME OF TEST: 10/26/2016 02:28:19

Diagnostics Check	OK	02:21	
Air Blank	0.000	02:21	
Control Test	0.082	02:21	
Air Blank	0.000	02:22	
Subject Sample	# 1 0.280	02:22	
Breath Volume	2.046	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:25	
Subject Sample	# 2 0.259	02:26	
Breath Volume	1.867	02:26	
Air Blank		02:27	Purge Fail
Air Blank	0.000	02:28	

Data Download Date/Time: 11/11/2016 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ACOSTA BANKS

DATE/TIME OF TEST: 10/26/2016 02:52:35

Diagnostics Check	OK	02:51	
Air Blank	0.000	02:51	
Control Test	0.070	02:52	Control Outside Tolerance
Air Blank		02:52	RFI Detect

Data Download Date/Time: 11/11/2016 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: ACOSTA BANKS

DATE/TIME OF TEST: 10/26/2016 03:02:47

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.259	02:57
Breath Volume	2.343	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.265	03:00
Breath Volume	1.781	03:00
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/11/2016 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHANG,BRADLEY,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CHANG,BRADLEY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: WINGO

DATE/TIME OF TEST: 10/29/2016 03:51:24

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.082	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:50	Subject Test Refused
Breath Volume	0.000	03:50	
Air Blank	0.000	03:50	
Control Test	0.083	03:50	
Air Blank	0.000	03:51	
Diagnostics Check	OK	03:51	

Data Download Date/Time: 11/11/2016 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIMM,HUNTER,F
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GRIMM,HUNTER
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: GODINEZ

DATE/TIME OF TEST: 10/03/2016 18:19:48

Diagnostics Check	OK	18:12
Air Blank	0.000	18:13
Control Test	0.080	18:13
Air Blank	0.000	18:13
Subject Sample # 1	0.172	18:15
Breath Volume	3.531	18:15
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.170	18:18
Breath Volume	2.808	18:18
Air Blank	0.000	18:18
Control Test	0.079	18:19
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

Data Download Date/Time: 10/12/2016 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,ANTONIO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SOCARRAS,ANTONIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CURIEL

DATE/TIME OF TEST: 10/07/2016 00:34:27

Diagnostics Check	OK	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Subject Sample	# 1	00:33	RFI Detect
Breath Volume	0.250	00:33	
Air Blank		00:34	RFI Detect
Air Blank		00:34	RFI Detect

Data Download Date/Time: 10/12/2016 11:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,ANTONIO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SOCARRAS,ANTONIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CURIEL

DATE/TIME OF TEST: 10/07/2016 00:37:02

Diagnostics Check	OK	00:36	
Air Blank		00:36	RFI Detect
Air Blank	0.000	00:37	

Data Download Date/Time: 10/12/2016 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOCARRAS,ANTONIO,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SOCARRAS,ANTONIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/05/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CURIEL

DATE/TIME OF TEST: 10/07/2016 00:44:27

Diagnostics Check	OK	00:38
Air Blank	0.000	00:38
Control Test	0.079	00:39
Air Blank	0.000	00:39
Subject Sample # 1	0.132	00:40
Breath Volume	1.187	00:40
Air Blank	0.000	00:40
Air Blank	0.000	00:42
Subject Sample # 2	0.149	00:42
Breath Volume	1.980	00:42
Air Blank	0.000	00:43
Control Test	0.079	00:43
Air Blank	0.000	00:44
Diagnostics Check	OK	00:44

Data Download Date/Time: 10/12/2016 11:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/15/2016 10:49:44

Diagnostics Check	OK	10:43	
Air Blank	0.000	10:43	
Control Test	0.079	10:43	
Air Blank	0.000	10:44	
Subject Sample # 1	0.000	10:45	
Breath Volume	1.250	10:45	
Air Blank	0.000	10:45	
Air Blank	0.000	10:47	
Subject Sample # 2	0.000	10:48	
Breath Volume	2.578	10:48	
Air Blank	0.000	10:48	
Control Test	0.079	10:49	
Air Blank		10:49	RFI Detect
Air Blank	0.000	10:49	

Data Download Date/Time: 11/11/2016 10:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/15/2016 11:00:35

Diagnostics Check	OK	10:52
Air Blank	0.000	10:53
Control Test	0.079	10:53
Air Blank	0.000	10:54
Subject Sample # 1	0.000	10:55
Breath Volume	1.421	10:55
Air Blank	0.000	10:56
Air Blank	0.000	10:58
Subject Sample # 2	0.000	10:59
Breath Volume	1.910	10:59
Air Blank	0.000	10:59
Control Test	0.078	11:00
Air Blank	0.000	11:00
Diagnostics Check	OK	11:00

Data Download Date/Time: 11/11/2016 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MAURICE, M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HENRY, MAURICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/23/2016 03:44:21

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:41	
Control Test	0.078	03:41	
Air Blank	0.000	03:41	
Subject Sample	# 1 0.189	03:42	
Breath Volume	1.640	03:42	
Air Blank		03:43	Purge Fail
Air Blank		03:44	Purge Fail

Data Download Date/Time: 11/11/2016 10:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MAURICE, M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HENRY, MAURICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/23/2016 03:54:32

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.079	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1 0.185	03:52	
Breath Volume	1.910	03:52	
Air Blank		03:53	Purge Fail
Air Blank		03:54	Purge Fail

Data Download Date/Time: 11/11/2016 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOIVU,DAVID,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: COUTO,ADRIEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:26

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 10/23/2016 18:55:24

Diagnostics Check	OK	18:46	
Air Blank	0.000	18:47	
Control Test	0.079	18:47	
Air Blank	0.000	18:47	
Subject Sample	# 1 0.217	18:50	Volume Not Met
Breath Volume	0.019	18:50	
Air Blank	0.000	18:51	
Air Blank	0.000	18:53	
Subject Sample	# 2	18:53	Subject Test Refused
Breath Volume	0.039	18:53	
Air Blank	0.000	18:54	
Control Test	0.078	18:54	
Air Blank	0.000	18:55	
Diagnostics Check	OK	18:55	

Data Download Date/Time: 11/11/2016 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIERRA,LUIS,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SIERRA,L
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/25/2016 00:25:11

Diagnostics Check	OK	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:19
Air Blank	0.000	00:19
Subject Sample # 1	0.287	00:20
Breath Volume	1.828	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:22
Subject Sample # 2	0.286	00:23
Breath Volume	2.324	00:23
Air Blank	0.000	00:24
Control Test	0.078	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 11/11/2016 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOIVU,DAVID,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BENNETT,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/12/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: REIGOSA FERNANDEZ

DATE/TIME OF TEST: 10/25/2016 19:46:27

Diagnostics Check	OK	19:43	
Air Blank	0.000	19:44	
Control Test	0.080	19:44	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:45	
Control Test	0.079	19:45	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:46	

Data Download Date/Time: 11/11/2016 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JETRO,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: BAKER,KENNETH
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 10/04/2016 01:05:52

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.079	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.241	01:00
Breath Volume	2.812	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.249	01:04
Breath Volume	2.910	01:04
Air Blank	0.000	01:04
Control Test	0.078	01:05
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 10/26/2016 01:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: PERIGNY,JONATHAN
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: LUIS

DATE/TIME OF TEST: 10/06/2016 03:52:47

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.078	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.079	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 10/26/2016 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAESTRE,HANS,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BESTARD,BRANDON
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 10/07/2016 19:23:40

Diagnostics Check	OK	19:17
Air Blank	0.000	19:17
Control Test	0.079	19:18
Air Blank	0.000	19:18
Subject Sample # 1	0.183	19:19
Breath Volume	1.535	19:19
Air Blank	0.000	19:19
Air Blank	0.000	19:21
Subject Sample # 2	0.183	19:22
Breath Volume	1.664	19:22
Air Blank	0.000	19:22
Control Test	0.078	19:23
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 10/26/2016 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,KENNETH,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: BAKER,KENNETH
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/13/2016 03:24:37

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.080	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.045	03:20
Breath Volume	1.414	03:20
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.045	03:23
Breath Volume	1.535	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:24
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 10/26/2016 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,JETRO,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: BAKER,KENNETH
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MAXIME

DATE/TIME OF TEST: 10/25/2016 03:18:38

Diagnostics Check	OK	03:11
Air Blank	0.000	03:11
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.192	03:13
Breath Volume	2.023	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.190	03:17
Breath Volume	2.074	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 10/26/2016 01:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLIVARES,AJ,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: OLIVARES,AJ
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 10/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: LACASA

DATE/TIME OF TEST: 10/29/2016 04:47:35

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.133	04:43
Breath Volume	2.234	04:43
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.134	04:46
Breath Volume	2.144	04:46
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 11/29/2016 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,BRIAN,A
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: DICKEY,LARRY
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 09/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28814080A3
DRY GAS EXPIRY DATE: 11/05/2016
VIOLATION CODE: OTHER
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 10/21/2016 01:09:13

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.078	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.266	01:04
Breath Volume	2.843	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.249	01:07
Breath Volume	1.601	01:07
Air Blank	0.000	01:08
Control Test	0.077	01:08
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 10/24/2016 22:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BOWMAN,JOANNA,N

BREATH TEST OPERATOR AGENCY: FHP Troop A

ARRESTING OFFICER: BOWMAN,JOANNA

ARRESTING OFFICER AGENCY: FHP Troop A

LAST AGENCY INSPECTION: 10/24/2016

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 20016080A3

DRY GAS EXPIRY DATE: 09/05/2018

VIOLATION CODE: DUI

ARREST TIME: 22:19

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 10/27/2016 00:53:13

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.081	00:41
Air Blank	0.000	00:42
Subject Sample # 1	0.175	00:44
Breath Volume	1.265	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:46
Subject Sample # 2	0.198	00:48
Breath Volume	2.246	00:48
Air Blank	0.000	00:48
Air Blank	0.000	00:50
Subject Sample # 3	0.185	00:51
Breath Volume	1.328	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

Data Download Date/Time: 11/09/2016 10:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/12/2016 02:50:12

Diagnostics Check	OK	02:44	
Air Blank	0.000	02:44	
Control Test	0.078	02:45	
Air Blank	0.000	02:45	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.039	02:48	
Air Blank	0.000	02:49	
Control Test	0.079	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SEAY, JOHNNY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:16

SUBJECT'S LAST NAME: LAKERS

DATE/TIME OF TEST: 10/12/2016 10:47:47

Diagnostics Check	OK	10:40
Air Blank	0.000	10:41
Control Test	0.078	10:41
Air Blank	0.000	10:42
Subject Sample # 1	0.000	10:42
Breath Volume	1.792	10:42
Air Blank	0.000	10:43
Air Blank	0.000	10:45
Subject Sample # 2	0.000	10:46
Breath Volume	1.707	10:46
Air Blank	0.000	10:46
Control Test	0.079	10:47
Air Blank	0.000	10:47
Diagnostics Check	OK	10:47

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFER,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/13/2016 04:11:46

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Subject Sample # 1	0.151	04:01
Breath Volume	2.789	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.173	04:06
Breath Volume	4.421	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 3	0.177	04:10
Breath Volume	3.972	04:10
Air Blank	0.000	04:10
Control Test	0.076	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/13/2016 22:00:21

Diagnostics Check	OK	21:53
Air Blank	0.000	21:54
Control Test	0.078	21:54
Air Blank	0.000	21:54
Subject Sample # 1	0.156	21:55
Breath Volume	1.195	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:57
Subject Sample # 2	0.158	21:58
Breath Volume	1.226	21:58
Air Blank	0.000	21:59
Control Test	0.078	21:59
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: RONGISH, SHERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: BARTHA

DATE/TIME OF TEST: 10/14/2016 09:04:54

Diagnostics Check	OK	08:57
Air Blank	0.000	08:57
Control Test	0.079	08:58
Air Blank	0.000	08:58
Subject Sample # 1	0.000	08:59
Breath Volume	1.820	08:59
Air Blank	0.000	09:00
Air Blank	0.000	09:02
Subject Sample # 2	0.000	09:03
Breath Volume	2.351	09:03
Air Blank	0.000	09:03
Control Test	0.079	09:04
Air Blank	0.000	09:04
Diagnostics Check	OK	09:04

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 10/15/2016 15:28:18

Diagnostics Check	OK	15:21
Air Blank	0.000	15:22
Control Test	0.078	15:22
Air Blank	0.000	15:22
Subject Sample # 1	0.161	15:23
Breath Volume	2.656	15:23
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.151	15:26
Breath Volume	1.546	15:26
Air Blank	0.000	15:27
Control Test	0.077	15:27
Air Blank	0.000	15:28
Diagnostics Check	OK	15:28

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:04

SUBJECT'S LAST NAME: DEANGELUS

DATE/TIME OF TEST: 10/16/2016 11:30:31

Diagnostics Check	OK	11:24
Air Blank	0.000	11:25
Control Test	0.079	11:25
Air Blank	0.000	11:25
Subject Sample # 1	0.000	11:26
Breath Volume	2.507	11:26
Air Blank	0.000	11:26
Air Blank	0.000	11:28
Subject Sample # 2	0.000	11:29
Breath Volume	1.910	11:29
Air Blank	0.000	11:29
Control Test	0.079	11:29
Air Blank	0.000	11:30
Diagnostics Check	OK	11:30

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,LISA,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SANTOS,E
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: TRECHE MARTINEZ

DATE/TIME OF TEST: 10/17/2016 05:38:18

Diagnostics Check	OK	05:35	
Air Blank	0.000	05:35	
Control Test	0.080	05:35	
Air Blank	0.000	05:36	
Subject Sample	# 1	05:36	Subject Test Refused
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Control Test	0.079	05:37	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: POPOLI

DATE/TIME OF TEST: 10/17/2016 16:23:41

Diagnostics Check	OK	16:13
Air Blank	0.000	16:14
Control Test	0.079	16:14
Air Blank	0.000	16:15
Subject Sample # 1	0.114	16:17
Breath Volume	1.105	16:17
Air Blank	0.000	16:18
Air Blank	0.000	16:19
Subject Sample # 2	0.110	16:22
Breath Volume	1.109	16:22
Air Blank	0.000	16:22
Control Test	0.079	16:23
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: GRIJALVA

DATE/TIME OF TEST: 10/17/2016 20:41:14

Diagnostics Check	OK	20:35
Air Blank	0.000	20:35
Control Test	0.079	20:36
Air Blank	0.000	20:36
Subject Sample # 1	0.000	20:37
Breath Volume	1.667	20:37
Air Blank	0.000	20:37
Air Blank	0.000	20:39
Subject Sample # 2	0.000	20:39
Breath Volume	1.605	20:39
Air Blank	0.000	20:40
Control Test	0.078	20:40
Air Blank	0.000	20:41
Diagnostics Check	OK	20:41

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: FYKES

DATE/TIME OF TEST: 10/18/2016 02:18:50

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.000	02:14
Breath Volume	1.500	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.000	02:17
Breath Volume	1.128	02:17
Air Blank	0.000	02:17
Control Test	0.079	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: BURGESS

DATE/TIME OF TEST: 10/19/2016 03:00:01

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.079	02:54
Air Blank	0.000	02:55
Subject Sample # 1	0.219	02:55
Breath Volume	1.679	02:55
Air Blank	0.000	02:56
Air Blank	0.000	02:58
Subject Sample # 2	0.213	02:58
Breath Volume	1.246	02:58
Air Blank	0.000	02:59
Control Test	0.078	02:59
Air Blank	0.000	02:59
Diagnostics Check	OK	03:00

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI, A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: NERONI, KRISTI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:49

SUBJECT'S LAST NAME: MEADORS

DATE/TIME OF TEST: 10/19/2016 13:18:00

Diagnostics Check	OK	13:11
Air Blank	0.000	13:11
Control Test	0.079	13:11
Air Blank	0.000	13:12
Subject Sample # 1	0.000	13:13
Breath Volume	1.882	13:13
Air Blank	0.000	13:14
Air Blank	0.000	13:16
Subject Sample # 2	0.000	13:16
Breath Volume	1.636	13:16
Air Blank	0.000	13:17
Control Test	0.079	13:17
Air Blank	0.000	13:17
Diagnostics Check	OK	13:18

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRANO,ROMAN,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LANGLEY,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 10/20/2016 10:12:04

Diagnostics Check	OK	10:05
Air Blank	0.000	10:06
Control Test	0.079	10:06
Air Blank	0.000	10:07
Subject Sample # 1	0.000	10:07
Breath Volume	2.546	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:10
Subject Sample # 2	0.000	10:10
Breath Volume	2.230	10:10
Air Blank	0.000	10:11
Control Test	0.079	10:11
Air Blank	0.000	10:11
Diagnostics Check	OK	10:12

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALEJO,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: BROCKA III

DATE/TIME OF TEST: 10/20/2016 11:46:24

Diagnostics Check	OK	11:40
Air Blank	0.000	11:40
Control Test	0.079	11:41
Air Blank	0.000	11:41
Subject Sample # 1	0.000	11:42
Breath Volume	2.070	11:42
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.000	11:45
Breath Volume	2.562	11:45
Air Blank	0.000	11:45
Control Test	0.078	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: SLEIGHTER

DATE/TIME OF TEST: 10/21/2016 20:57:24

Diagnostics Check	OK	20:52	
Air Blank	0.000	20:53	
Control Test	0.079	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:56	Sequence Aborted
Breath Volume	0.000	20:56	
Air Blank	0.000	20:57	

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: SLEIGHTER

DATE/TIME OF TEST: 10/21/2016 21:26:17

Diagnostics Check	OK	21:20	
Air Blank	0.000	21:21	
Control Test	0.080	21:21	
Air Blank	0.000	21:21	
Subject Sample	# 1 0.236	21:25	Slope Not Level
Breath Volume	1.343	21:25	
Air Blank		21:25	Purge Fail
Air Blank	0.000	21:26	

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: TRASK,J
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DESSALINES

DATE/TIME OF TEST: 10/22/2016 02:37:23

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.080	02:32	
Air Blank	0.000	02:33	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.972	02:35	
Air Blank	0.000	02:36	
Control Test	0.081	02:36	
Air Blank	0.000	02:37	
Diagnostics Check	OK	02:37	

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFER,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 10/22/2016 03:44:52

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.000	03:40
Breath Volume	2.066	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.000	03:43
Breath Volume	1.843	03:43
Air Blank	0.000	03:43
Control Test	0.079	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: MIRKOWSKA

DATE/TIME OF TEST: 10/22/2016 04:10:16

Diagnostics Check	OK	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.173	04:05
Breath Volume	1.109	04:05
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.159	04:08
Breath Volume	1.335	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CERVANTES-GARCIA

DATE/TIME OF TEST: 10/23/2016 01:20:17

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.079	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.142	01:15
Breath Volume	1.691	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.139	01:18
Breath Volume	1.511	01:18
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTIA
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN, LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MORALES CANO

DATE/TIME OF TEST: 10/23/2016 03:41:12

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.142	03:36
Breath Volume	2.437	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:38
Subject Sample # 2	0.137	03:39
Breath Volume	2.242	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FLORES,SAMUEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 10/24/2016 02:57:31

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.266	02:52
Breath Volume	3.062	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.261	02:55
Breath Volume	2.214	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTIA
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN, LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: MEADORS

DATE/TIME OF TEST: 10/25/2016 15:31:37

Diagnostics Check	OK	15:25
Air Blank	0.000	15:25
Control Test	0.080	15:25
Air Blank	0.000	15:26
Subject Sample # 1	0.000	15:27
Breath Volume	1.156	15:27
Air Blank	0.000	15:27
Air Blank	0.000	15:29
Subject Sample # 2	0.000	15:30
Breath Volume	1.390	15:30
Air Blank	0.000	15:30
Control Test	0.078	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:31

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 10/25/2016 22:31:05

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.136	22:26
Breath Volume	2.117	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.128	22:29
Breath Volume	2.351	22:29
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, RAYMOND, P
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLAGHER, RAYMOND
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: ROCKIE

DATE/TIME OF TEST: 10/28/2016 02:03:05

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.146	01:58
Breath Volume	1.789	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:00
Subject Sample # 2	0.147	02:01
Breath Volume	1.136	02:01
Air Blank	0.000	02:02
Control Test	0.078	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GALLOWAY,E
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: KRUGER

DATE/TIME OF TEST: 10/29/2016 00:05:24

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.128	00:00
Breath Volume	2.500	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.125	00:03
Breath Volume	2.671	00:03
Air Blank	0.000	00:04
Control Test	0.078	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: DEARCIA GUILLEN

DATE/TIME OF TEST: 10/29/2016 00:24:36

Diagnostics Check	OK	00:11	
Air Blank	0.000	00:11	
Control Test	0.079	00:11	
Air Blank	0.000	00:12	
Subject Sample	# 1 0.154	00:14	
Breath Volume	1.468	00:14	
Air Blank	0.000	00:14	
Air Blank	0.000	00:16	
Subject Sample	# 2 0.134	00:19	Volume Not Met
Breath Volume	0.156	00:19	
Air Blank	0.000	00:20	
Air Blank	0.000	00:22	
Subject Sample	# 3	00:23	Subject Test Refused
Breath Volume	0.296	00:23	
Air Blank	0.000	00:23	
Control Test	0.076	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/30/2016 01:41:59

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.247	01:37
Breath Volume	2.011	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.244	01:40
Breath Volume	1.550	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MAHLER

DATE/TIME OF TEST: 10/30/2016 01:57:58

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.078	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.141	01:53
Breath Volume	1.882	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.136	01:56
Breath Volume	1.652	01:56
Air Blank	0.000	01:57
Control Test	0.078	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR, MAURICE, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NAYLOR, MAURICE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DE DIOS

DATE/TIME OF TEST: 10/30/2016 04:28:38

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.092	04:24
Breath Volume	2.214	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.086	04:27
Breath Volume	2.339	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: BRETON

DATE/TIME OF TEST: 10/30/2016 05:15:47

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.079	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.144	05:11
Breath Volume	1.949	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.135	05:14
Breath Volume	2.371	05:14
Air Blank	0.000	05:14
Control Test	0.078	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: LALLIER

DATE/TIME OF TEST: 10/30/2016 19:52:52

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.080	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.000	19:48
Breath Volume	2.003	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:51
Subject Sample # 2	0.000	19:51
Breath Volume	1.285	19:51
Air Blank	0.000	19:51
Control Test	0.079	19:52
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 11/11/2016 14:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: CARTWRIGHT

DATE/TIME OF TEST: 10/01/2016 16:23:39

Diagnostics Check	OK	16:17
Air Blank	0.000	16:17
Control Test	0.080	16:18
Air Blank	0.000	16:18
Subject Sample # 1	0.107	16:18
Breath Volume	1.671	16:18
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.105	16:22
Breath Volume	1.980	16:22
Air Blank	0.000	16:22
Control Test	0.080	16:23
Air Blank	0.000	16:23
Diagnostics Check	OK	16:23

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 10/03/2016 02:07:58

Diagnostics Check	OK	02:01
Air Blank	0.000	02:02
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.133	02:03
Breath Volume	2.679	02:03
Air Blank	0.000	02:04
Air Blank	0.000	02:05
Subject Sample # 2	0.125	02:06
Breath Volume	1.972	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 10/05/2016 23:33:25

Diagnostics Check	OK	23:26
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.257	23:28
Breath Volume	2.058	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.249	23:31
Breath Volume	2.074	23:31
Air Blank	0.000	23:32
Control Test	0.079	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: CONDON

DATE/TIME OF TEST: 10/07/2016 02:56:48

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.186	02:51
Breath Volume	1.449	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.187	02:55
Breath Volume	2.363	02:55
Air Blank	0.000	02:55
Control Test	0.079	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 10/07/2016 05:17:32

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.154	05:12
Breath Volume	2.984	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:14
Subject Sample # 2	0.152	05:15
Breath Volume	1.726	05:15
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENEMAN,LESA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN,LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: TIERNEY

DATE/TIME OF TEST: 10/07/2016 22:07:27

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.080	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.190	22:02
Breath Volume	2.855	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:05
Subject Sample # 2	0.182	22:05
Breath Volume	2.562	22:05
Air Blank	0.000	22:06
Control Test	0.078	22:06
Air Blank	0.000	22:07
Diagnostics Check	OK	22:07

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NERONI, KRISTI, A
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: NERONI, KRISTI
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: PLACA

DATE/TIME OF TEST: 10/08/2016 00:58:09

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.249	00:53
Breath Volume	1.226	00:53
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.256	00:56
Breath Volume	2.003	00:56
Air Blank	0.000	00:57
Control Test	0.079	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENEMAN,LESA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN,LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: STETZNER

DATE/TIME OF TEST: 10/08/2016 02:28:23

Diagnostics Check	OK	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.090	02:23
Breath Volume	1.277	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.104	02:26
Breath Volume	2.710	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 10/08/2016 02:37:34

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.198	02:33
Breath Volume	2.136	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.200	02:35
Breath Volume	2.609	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ROBERT,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELLIS,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 10/08/2016 04:07:41

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.151	04:01
Breath Volume	2.445	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.143	04:06
Breath Volume	1.937	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SANTOS,E
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/08/2016 20:56:40

Diagnostics Check	OK	20:50
Air Blank	0.000	20:50
Control Test	0.079	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.241	20:52
Breath Volume	1.738	20:52
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.232	20:55
Breath Volume	1.296	20:55
Air Blank	0.000	20:55
Control Test	0.079	20:56
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENEMAN,LESA,
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: BRENEMAN,LESA
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 10/08/2016 23:48:53

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.079	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.052	23:44
Breath Volume	2.125	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:46
Subject Sample # 2	0.052	23:47
Breath Volume	2.000	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ANDREW,R
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 10/09/2016 02:45:00

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:40	
Control Test	0.079	02:40	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.503	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:44	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRASK,J,N
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: SCHAEFR,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: TOOMEY

DATE/TIME OF TEST: 10/09/2016 23:04:41

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.079	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.028	23:00
Breath Volume	2.746	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.029	23:03
Breath Volume	2.890	23:03
Air Blank	0.000	23:03
Control Test	0.078	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GOULD, LEONARD
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:09

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 10/10/2016 14:35:47

Diagnostics Check	OK	14:29
Air Blank	0.000	14:29
Control Test	0.079	14:30
Air Blank	0.000	14:30
Subject Sample # 1	0.247	14:30
Breath Volume	1.375	14:30
Air Blank	0.000	14:31
Air Blank	0.000	14:33
Subject Sample # 2	0.243	14:34
Breath Volume	1.402	14:34
Air Blank	0.000	14:34
Control Test	0.079	14:35
Air Blank	0.000	14:35
Diagnostics Check	OK	14:35

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOULD, LEONARD,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ZAMMIT, DOMINIC
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: GALEK

DATE/TIME OF TEST: 10/10/2016 20:09:10

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.079	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.000	20:04
Breath Volume	1.402	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.000	20:07
Breath Volume	1.851	20:07
Air Blank	0.000	20:08
Control Test	0.079	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000939

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/16/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: NORDLING

DATE/TIME OF TEST: 10/11/2016 03:42:43

Diagnostics Check	OK	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Subject Sample # 1	0.112	03:37
Breath Volume	3.093	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.110	03:41
Breath Volume	2.746	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:42
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 10/11/2016 11:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,BRIAN,K
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SUDANO,BRIAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/04/2016 20:10:06

Diagnostics Check	OK	20:03
Air Blank	0.000	20:03
Control Test	0.079	20:04
Air Blank	0.000	20:04
Subject Sample # 1	0.101	20:05
Breath Volume	2.078	20:05
Air Blank	0.000	20:06
Air Blank	0.000	20:07
Subject Sample # 2	0.098	20:08
Breath Volume	2.003	20:08
Air Blank	0.000	20:09
Control Test	0.079	20:09
Air Blank	0.000	20:10
Diagnostics Check	OK	20:10

Data Download Date/Time: 11/01/2016 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,JACOB,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARVER,JERROD
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 10/13/2016 05:51:47

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.079	05:46
Air Blank	0.000	05:46
Subject Sample # 1	0.160	05:47
Breath Volume	1.480	05:47
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.157	05:50
Breath Volume	1.593	05:50
Air Blank	0.000	05:50
Control Test	0.080	05:51
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 11/01/2016 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,NICKOLAUS,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PICCOLI,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: STIDHAM

DATE/TIME OF TEST: 10/01/2016 00:57:16

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.078	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1 0.185	00:48	Volume Not Met
Breath Volume	0.882	00:48	
Air Blank	0.000	00:49	
Air Blank	0.000	00:51	
Subject Sample	# 2 0.190	00:51	
Breath Volume	2.074	00:51	
Air Blank	0.000	00:52	
Air Blank	0.000	00:54	
Subject Sample	# 3 0.179	00:55	
Breath Volume	1.476	00:55	
Air Blank	0.000	00:56	
Control Test	0.075	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,RASHAAD,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SMITH,RASHAAD
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: WILSON GONZALEZ

DATE/TIME OF TEST: 10/01/2016 03:20:54

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.078	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.220	03:14
Breath Volume	1.929	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.219	03:19
Breath Volume	1.722	03:19
Air Blank	0.000	03:20
Control Test	0.078	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AXELSON,B,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAUFFER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: EMILE

DATE/TIME OF TEST: 10/01/2016 05:16:13

Diagnostics Check	OK	05:09
Air Blank	0.000	05:09
Control Test	0.078	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.211	05:11
Breath Volume	2.136	05:11
Air Blank	0.000	05:12
Air Blank	0.000	05:14
Subject Sample # 2	0.206	05:14
Breath Volume	2.140	05:14
Air Blank	0.000	05:15
Control Test	0.077	05:15
Air Blank	0.000	05:16
Diagnostics Check	OK	05:16

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON,MITCHELL,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FREDRICKS,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:26

SUBJECT'S LAST NAME: SANTIAGO RIVERA

DATE/TIME OF TEST: 10/01/2016 20:55:24

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:50
Subject Sample # 1	0.140	20:50
Breath Volume	1.492	20:50
Air Blank	0.000	20:51
Air Blank	0.000	20:53
Subject Sample # 2	0.144	20:53
Breath Volume	1.875	20:53
Air Blank	0.000	20:54
Control Test	0.078	20:54
Air Blank	0.000	20:55
Diagnostics Check	OK	20:55

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 10/02/2016 02:49:53

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Subject Sample # 1	0.137	02:45
Breath Volume	1.613	02:45
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.136	02:48
Breath Volume	2.132	02:48
Air Blank	0.000	02:48
Control Test	0.077	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOLO,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CASCIANO,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: DREXALL

DATE/TIME OF TEST: 10/02/2016 03:41:20

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.136	03:36
Breath Volume	1.953	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.145	03:39
Breath Volume	2.335	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: MIRANDA

DATE/TIME OF TEST: 10/02/2016 03:53:21

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.078	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.148	03:48
Breath Volume	2.367	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.148	03:51
Breath Volume	1.566	03:51
Air Blank	0.000	03:52
Control Test	0.078	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,HANSER,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NEGRON,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:43

SUBJECT'S LAST NAME: DELVALLEFERNANDE

DATE/TIME OF TEST: 10/02/2016 12:24:33

Diagnostics Check	OK	12:17
Air Blank	0.000	12:18
Control Test	0.078	12:18
Air Blank	0.000	12:18
Subject Sample # 1	0.175	12:19
Breath Volume	2.500	12:19
Air Blank	0.000	12:20
Air Blank	0.000	12:22
Subject Sample # 2	0.163	12:23
Breath Volume	1.929	12:23
Air Blank	0.000	12:23
Control Test	0.077	12:23
Air Blank	0.000	12:24
Diagnostics Check	OK	12:24

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,DAN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 10/03/2016 00:10:52

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.237	00:03	Volume Not Met
Breath Volume	0.109	00:03	
Air Blank	0.000	00:04	
Air Blank	0.000	00:06	
Subject Sample	# 2 0.166	00:09	Volume Not Met
Breath Volume	0.164	00:09	
Air Blank	0.000	00:09	
Control Test	0.077	00:10	
Air Blank	0.000	00:10	
Diagnostics Check	OK	00:10	

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN, DAVID, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN, DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/03/2016 03:49:42

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.079	03:38	
Air Blank	0.000	03:39	
Subject Sample	# 1 0.129	03:42	Volume Not Met
Breath Volume	0.019	03:42	
Air Blank	0.000	03:43	
Air Blank	0.000	03:45	
Subject Sample	# 2 0.161	03:48	Volume Not Met
Breath Volume	0.492	03:48	
Air Blank	0.000	03:48	
Control Test	0.078	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTIANSEN, DAVID, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN, DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/01/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A4
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 10/03/2016 04:02:44

Diagnostics Check	OK	03:51	
Air Blank	0.000	03:51	
Control Test	0.077	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1 0.143	03:55	Volume Not Met
Breath Volume	0.062	03:55	
Air Blank	0.000	03:56	
Air Blank	0.000	03:58	
Subject Sample	# 2 0.153	04:01	Volume Not Met
Breath Volume	0.425	04:01	
Air Blank	0.000	04:01	
Control Test	0.076	04:02	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 10/03/2016 12:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JESSE,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KLEIS,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DUBRUELER

DATE/TIME OF TEST: 10/17/2016 01:25:57

Diagnostics Check	OK	01:19
Air Blank	0.000	01:19
Control Test	0.079	01:19
Air Blank	0.000	01:20
Subject Sample # 1	0.170	01:21
Breath Volume	1.273	01:21
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.164	01:24
Breath Volume	1.292	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:25
Diagnostics Check	OK	01:25

Data Download Date/Time: 11/01/2016 11:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ROCA CHAVEZ

DATE/TIME OF TEST: 10/09/2016 03:52:58

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.080	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.198	03:48
Breath Volume	1.394	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.192	03:51
Breath Volume	1.992	03:51
Air Blank	0.000	03:52
Control Test	0.081	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLANEDA ROMO,KARLA,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,DEMETRIS
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: MARANDA SIMARD

DATE/TIME OF TEST: 10/13/2016 01:39:01

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.202	01:34
Breath Volume	1.941	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:36
Subject Sample # 2	0.197	01:37
Breath Volume	2.164	01:37
Air Blank	0.000	01:38
Control Test	0.080	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:39

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: Naples PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: CROSS

DATE/TIME OF TEST: 10/14/2016 23:44:32

Diagnostics Check	OK	23:38
Air Blank	0.000	23:38
Control Test	0.081	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.101	23:39
Breath Volume	2.070	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:42
Subject Sample # 2	0.104	23:43
Breath Volume	2.421	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Diagnostics Check	OK	23:44

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: TRUDNAK

DATE/TIME OF TEST: 10/15/2016 03:35:11

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.121	03:30
Breath Volume	2.515	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.120	03:33
Breath Volume	1.613	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA, URIEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE, B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: VELASQUEZ

DATE/TIME OF TEST: 10/15/2016 05:07:29

Diagnostics Check	OK	05:01
Air Blank	0.000	05:01
Control Test	0.080	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.181	05:03
Breath Volume	1.671	05:03
Air Blank	0.000	05:03
Air Blank	0.000	05:05
Subject Sample # 2	0.173	05:05
Breath Volume	2.730	05:05
Air Blank	0.000	05:06
Control Test	0.081	05:06
Air Blank	0.000	05:07
Diagnostics Check	OK	05:07

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SILL,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/15/2016 20:27:50

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.170	20:22
Breath Volume	1.960	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.169	20:26
Breath Volume	2.066	20:26
Air Blank	0.000	20:26
Control Test	0.081	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VECIANA, YASSIEL, A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN, D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: VICTORERO GREGORES

DATE/TIME OF TEST: 10/16/2016 01:13:43

Diagnostics Check	OK	01:04
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Subject Sample # 1	0.241	01:06
Breath Volume	2.210	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.218	01:09
Breath Volume	1.195	01:09
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 3	0.228	01:12
Breath Volume	1.632	01:12
Air Blank	0.000	01:12
Control Test	0.079	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: Naples PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/16/2016 01:35:34

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.080	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.219	01:30
Breath Volume	3.734	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.217	01:34
Breath Volume	4.234	01:34
Air Blank	0.000	01:34
Control Test	0.080	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CARPENTER, CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: LE CLAIR

DATE/TIME OF TEST: 10/16/2016 03:30:31

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.080	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.113	03:26
Breath Volume	1.640	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.110	03:29
Breath Volume	1.871	03:29
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMENT,KERRY,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KAPLAN,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 10/17/2016 01:09:25

Diagnostics Check	OK	00:57	
Air Blank	0.000	00:57	
Control Test	0.081	00:58	
Air Blank	0.000	00:58	
Subject Sample	# 1 0.170	01:01	Volume Not Met
Breath Volume	0.050	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 2 0.227	01:07	Volume Not Met
Breath Volume	0.410	01:07	
Air Blank	0.000	01:08	
Control Test	0.080	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMENT,KERRY,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KAPLAN,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: CORREA

DATE/TIME OF TEST: 10/17/2016 01:41:54

Diagnostics Check	OK	01:29	
Air Blank	0.000	01:29	
Control Test	0.078	01:30	
Air Blank	0.000	01:30	
Subject Sample	# 1 0.224	01:33	Volume Not Met
Breath Volume	0.839	01:33	
Air Blank	0.000	01:34	
Air Blank	0.000	01:36	
Subject Sample	# 2 0.216	01:36	
Breath Volume	1.191	01:36	
Air Blank	0.000	01:37	
Air Blank	0.000	01:39	
Subject Sample	# 3 0.210	01:40	
Breath Volume	1.675	01:40	
Air Blank	0.000	01:40	
Control Test	0.076	01:41	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBEY,AMANDA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEROSA,A
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: REGALADO

DATE/TIME OF TEST: 10/19/2016 04:36:05

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.081	04:30
Air Blank	0.000	04:30
Subject Sample # 1	0.087	04:31
Breath Volume	1.953	04:31
Air Blank	0.000	04:32
Air Blank	0.000	04:33
Subject Sample # 2	0.084	04:34
Breath Volume	1.714	04:34
Air Blank	0.000	04:35
Control Test	0.079	04:35
Air Blank	0.000	04:35
Diagnostics Check	OK	04:36

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TIRADO

DATE/TIME OF TEST: 10/21/2016 00:48:22

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.133	00:43
Breath Volume	4.335	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.142	00:46
Breath Volume	3.136	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:48
Diagnostics Check	OK	00:48

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: JACKSON,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: AVILA HERNANDEZ

DATE/TIME OF TEST: 10/22/2016 03:29:28

Diagnostics Check	OK	03:22
Air Blank	0.000	03:23
Control Test	0.080	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.230	03:24
Breath Volume	1.371	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.245	03:27
Breath Volume	1.976	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHRISTIANSEN,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GUTIERREZ ALDANA

DATE/TIME OF TEST: 10/23/2016 00:54:48

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.183	00:49
Breath Volume	1.542	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.175	00:53
Breath Volume	1.332	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTSINGER,CODY,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DIAZ,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/23/2016 01:05:13

Diagnostics Check	OK	00:59
Air Blank	0.000	00:59
Control Test	0.080	00:59
Air Blank	0.000	01:00
Subject Sample # 1	0.199	01:00
Breath Volume	1.906	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:03
Subject Sample # 2	0.187	01:03
Breath Volume	1.238	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICARLO,J,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WILLIAMS,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: DIMAS TORRES

DATE/TIME OF TEST: 10/23/2016 05:48:11

Diagnostics Check	OK	05:41
Air Blank	0.000	05:41
Control Test	0.081	05:42
Air Blank	0.000	05:42
Subject Sample # 1	0.160	05:43
Breath Volume	1.859	05:43
Air Blank	0.000	05:44
Air Blank	0.000	05:46
Subject Sample # 2	0.165	05:46
Breath Volume	2.191	05:46
Air Blank	0.000	05:47
Control Test	0.079	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMBS,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: SHUMWAY

DATE/TIME OF TEST: 10/24/2016 22:58:02

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.023	22:53
Breath Volume	1.488	22:53
Air Blank	0.000	22:54
Air Blank	0.000	22:55
Subject Sample # 2	0.025	22:56
Breath Volume	2.050	22:56
Air Blank	0.000	22:57
Control Test	0.081	22:57
Air Blank	0.000	22:57
Diagnostics Check	OK	22:58

Data Download Date/Time: 11/01/2016 11:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SILL,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 10/25/2016 06:02:19

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.080	05:55
Air Blank	0.000	05:55
Subject Sample # 1	0.214	05:57
Breath Volume	1.230	05:57
Air Blank	0.000	05:58
Air Blank	0.000	05:59
Subject Sample # 2	0.230	06:00
Breath Volume	1.441	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: MATHEWSON

DATE/TIME OF TEST: 10/26/2016 02:32:55

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.081	02:27
Air Blank	0.000	02:27
Subject Sample # 1	0.170	02:28
Breath Volume	2.902	02:28
Air Blank	0.000	02:28
Air Blank	0.000	02:30
Subject Sample # 2	0.168	02:31
Breath Volume	2.191	02:31
Air Blank	0.000	02:32
Control Test	0.080	02:32
Air Blank	0.000	02:32
Diagnostics Check	OK	02:32

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA, URIEL,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMBS, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: CAMPOS

DATE/TIME OF TEST: 10/26/2016 03:56:08

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.117	03:51
Breath Volume	1.457	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.122	03:54
Breath Volume	1.660	03:54
Air Blank	0.000	03:55
Control Test	0.080	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FLOWERS,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: LARA

DATE/TIME OF TEST: 10/27/2016 02:17:21

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.185	02:12
Breath Volume	2.406	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.189	02:15
Breath Volume	2.480	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN, TIFFANY, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN, J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: AMARO

DATE/TIME OF TEST: 10/28/2016 03:42:33

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.133	03:38
Breath Volume	1.359	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.139	03:41
Breath Volume	1.597	03:41
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HARDYMON

DATE/TIME OF TEST: 10/28/2016 04:04:47

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.160	03:57
Breath Volume	1.886	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.183	04:00
Breath Volume	2.582	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 3	0.188	04:03
Breath Volume	1.234	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VADNEY, MICHELE, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WEEKENDER, WEEKENDER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: OTHER
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:11

SUBJECT'S LAST NAME: BICKFORD

DATE/TIME OF TEST: 10/28/2016 19:32:55

Diagnostics Check	OK	19:18	
Air Blank	0.000	19:19	
Control Test	0.080	19:19	
Air Blank	0.000	19:20	
Subject Sample	# 1 0.042	19:23	Volume Not Met
Breath Volume	0.511	19:23	
Air Blank	0.000	19:24	
Air Blank	0.000	19:25	
Subject Sample	# 2 0.042	19:28	
Breath Volume	1.945	19:28	
Air Blank	0.000	19:29	
Air Blank	0.000	19:30	
Subject Sample	# 3 0.037	19:31	
Breath Volume	2.347	19:31	
Air Blank	0.000	19:32	
Control Test	0.078	19:32	
Air Blank	0.000	19:32	
Diagnostics Check	OK	19:32	

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BUTLER,DAVID
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: KREAMER

DATE/TIME OF TEST: 10/28/2016 21:49:08

Diagnostics Check	OK	21:43
Air Blank	0.000	21:43
Control Test	0.080	21:44
Air Blank	0.000	21:44
Subject Sample # 1	0.000	21:44
Breath Volume	2.125	21:44
Air Blank	0.000	21:45
Air Blank	0.000	21:47
Subject Sample # 2	0.000	21:47
Breath Volume	1.726	21:47
Air Blank	0.000	21:48
Control Test	0.080	21:48
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: Naples PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: LAUGHERTY

DATE/TIME OF TEST: 10/28/2016 23:57:52

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.164	23:53
Breath Volume	4.941	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:55
Subject Sample # 2	0.172	23:56
Breath Volume	3.390	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: Naples PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: LAUGHERTY

DATE/TIME OF TEST: 10/29/2016 00:08:04

Diagnostics Check	OK	00:01
Air Blank	0.000	00:02
Control Test	0.080	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.167	00:03
Breath Volume	1.476	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.169	00:06
Breath Volume	3.777	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:08

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMBS,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MOSSER

DATE/TIME OF TEST: 10/29/2016 03:37:02

Diagnostics Check	OK	03:30
Air Blank	0.000	03:30
Control Test	0.080	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.105	03:32
Breath Volume	1.671	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.097	03:35
Breath Volume	2.339	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:37

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,RICHARD,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEDOUX,CHRISTOPHER
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 10/29/2016 08:07:52

Diagnostics Check	OK	07:56	
Air Blank	0.000	07:56	
Control Test	0.080	07:57	
Air Blank	0.000	07:57	
Subject Sample	# 1 0.157	08:00	Volume Not Met
Breath Volume	0.070	08:00	
Air Blank	0.000	08:01	
Air Blank	0.000	08:03	
Subject Sample	# 2 0.130	08:06	Volume Not Met
Breath Volume	0.144	08:06	
Air Blank	0.000	08:06	
Control Test	0.079	08:07	
Air Blank	0.000	08:07	
Diagnostics Check	OK	08:07	

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: Naples PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: ROBINS

DATE/TIME OF TEST: 10/30/2016 22:31:14

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.080	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.381	22:26
Breath Volume	2.285	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.379	22:29
Breath Volume	2.699	22:29
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JESSE,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PICCOLI,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: MELLIEN

DATE/TIME OF TEST: 10/30/2016 22:42:46

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.144	22:37
Breath Volume	1.324	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.144	22:41
Breath Volume	1.855	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,DAVID,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: FORSYTH

DATE/TIME OF TEST: 10/31/2016 04:20:44

Diagnostics Check	OK	04:13
Air Blank	0.000	04:14
Control Test	0.080	04:14
Air Blank	0.000	04:14
Subject Sample # 1	0.176	04:15
Breath Volume	1.402	04:15
Air Blank	0.000	04:16
Air Blank	0.000	04:17
Subject Sample # 2	0.166	04:19
Breath Volume	1.539	04:19
Air Blank	0.000	04:19
Control Test	0.080	04:20
Air Blank	0.000	04:20
Diagnostics Check	OK	04:20

Data Download Date/Time: 11/01/2016 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: QUEGAN,R
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: CHERY

DATE/TIME OF TEST: 10/01/2016 03:20:47

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.080	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.226	03:15
Breath Volume	1.433	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.230	03:19
Breath Volume	1.339	03:19
Air Blank	0.000	03:19
Control Test	0.080	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,BRADLEY,W
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COOPER,DAVID
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:37

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/02/2016 13:06:39

Diagnostics Check	OK	13:00
Air Blank	0.000	13:00
Control Test	0.081	13:01
Air Blank	0.000	13:01
Subject Sample # 1	0.000	13:02
Breath Volume	1.367	13:02
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.000	13:05
Breath Volume	1.402	13:05
Air Blank	0.000	13:05
Control Test	0.080	13:06
Air Blank	0.000	13:06
Diagnostics Check	OK	13:06

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREITAS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: KUTIL

DATE/TIME OF TEST: 10/04/2016 22:52:35

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.082	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.189	22:47
Breath Volume	3.492	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.184	22:51
Breath Volume	2.753	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MARRONE-
CABRERA,ANTHONY,J

BREATH TEST OPERATOR AGENCY: Charlotte County SO

ARRESTING OFFICER: MORAGHAN,JOHN

ARRESTING OFFICER AGENCY: Charlotte County SO

LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 19615080A2

DRY GAS EXPIRY DATE: 09/05/2017

VIOLATION CODE: DUI

ARREST TIME: 15:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: BODEKER

DATE/TIME OF TEST: 10/06/2016 16:18:00

Diagnostics Check	OK	16:11
Air Blank	0.000	16:12
Control Test	0.081	16:12
Air Blank	0.000	16:12
Subject Sample # 1	0.007	16:13
Breath Volume	4.589	16:13
Air Blank	0.000	16:13
Air Blank	0.000	16:15
Subject Sample # 2	0.007	16:16
Breath Volume	4.765	16:16
Air Blank	0.000	16:17
Control Test	0.081	16:17
Air Blank	0.000	16:17
Diagnostics Check	OK	16:18

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHAIBLE,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: WHARWOOD

DATE/TIME OF TEST: 10/14/2016 21:12:21

Diagnostics Check	OK	21:02
Air Blank	0.000	21:03
Control Test	0.082	21:03
Air Blank	0.000	21:04
Subject Sample # 1	0.183	21:04
Breath Volume	1.593	21:04
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.204	21:07
Breath Volume	2.046	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 3	0.203	21:10
Breath Volume	1.953	21:10
Air Blank	0.000	21:11
Control Test	0.081	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HERREY,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: CHRISTOPHER

DATE/TIME OF TEST: 10/14/2016 23:56:56

Diagnostics Check	OK	23:40	
Air Blank	0.000	23:41	
Control Test	0.082	23:41	
Air Blank	0.000	23:42	
Subject Sample	# 1 0.171	23:45	Volume Not Met
Breath Volume	1.093	23:45	
Air Blank	0.000	23:45	
Air Blank	0.000	23:47	
Subject Sample	# 2 0.213	23:49	No .020 agreement
Breath Volume	1.828	23:49	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 3 0.184	23:55	No .020 agreement
Breath Volume	2.242	23:55	
Air Blank	0.000	23:56	
Control Test	0.080	23:56	
Air Blank	0.000	23:56	
Diagnostics Check	OK	23:56	

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MINGS,TERESA
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HATLEY

DATE/TIME OF TEST: 10/15/2016 22:17:44

Diagnostics Check	OK	22:06	
Air Blank	0.000	22:06	
Control Test	0.081	22:07	
Air Blank	0.000	22:07	
Subject Sample	# 1 0.239	22:10	Volume Not Met
Breath Volume	0.636	22:10	
Air Blank	0.000	22:11	
Air Blank	0.000	22:13	
Subject Sample	# 2	22:16	No Sample Provided
Breath Volume	0.000	22:16	
Air Blank	0.000	22:16	
Control Test	0.080	22:17	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VAUGHAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/17/2016 18:16:02

Diagnostics Check	OK	18:09
Air Blank	0.000	18:09
Control Test	0.082	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.221	18:11
Breath Volume	2.230	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:13
Subject Sample # 2	0.218	18:14
Breath Volume	1.992	18:14
Air Blank	0.000	18:15
Control Test	0.081	18:15
Air Blank	0.000	18:15
Diagnostics Check	OK	18:16

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SMITH,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MALLOUS

DATE/TIME OF TEST: 10/19/2016 00:43:13

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.081	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.135	00:38
Breath Volume	1.914	00:38
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.141	00:41
Breath Volume	2.082	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: FREITAS,ANTHONY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 10/20/2016 04:49:53

Diagnostics Check	OK	04:38	
Air Blank	0.000	04:38	
Control Test	0.081	04:39	
Air Blank	0.000	04:39	
Subject Sample	# 1 0.256	04:42	Volume Not Met
Breath Volume	0.074	04:42	
Air Blank	0.000	04:43	
Air Blank	0.000	04:45	
Subject Sample	# 2	04:48	No Sample Provided
Breath Volume	0.000	04:48	
Air Blank	0.000	04:48	
Control Test	0.080	04:49	
Air Blank	0.000	04:49	
Diagnostics Check	OK	04:49	

Data Download Date/Time: 10/27/2016 10:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,CELESTE,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: COOPER,D
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:35

SUBJECT'S LAST NAME: DOTSON

DATE/TIME OF TEST: 10/20/2016 15:01:34

Diagnostics Check	OK	14:55
Air Blank	0.000	14:55
Control Test	0.081	14:56
Air Blank	0.000	14:56
Subject Sample # 1	0.225	14:57
Breath Volume	3.230	14:57
Air Blank	0.000	14:57
Air Blank	0.000	14:59
Subject Sample # 2	0.221	15:00
Breath Volume	3.593	15:00
Air Blank	0.000	15:00
Control Test	0.081	15:00
Air Blank	0.000	15:01
Diagnostics Check	OK	15:01

Data Download Date/Time: 10/27/2016 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GALLEGO,SAM
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: GRIBBIN

DATE/TIME OF TEST: 10/21/2016 20:59:14

Diagnostics Check	OK	20:46	
Air Blank	0.000	20:46	
Control Test	0.081	20:47	
Air Blank	0.000	20:47	
Subject Sample	# 1 0.187	20:50	Volume Not Met
Breath Volume	0.265	20:50	
Air Blank	0.000	20:51	
Air Blank	0.000	20:53	
Subject Sample	# 2 0.244	20:54	
Breath Volume	1.722	20:54	
Air Blank	0.000	20:55	
Air Blank	0.000	20:57	
Subject Sample	# 3 0.256	20:57	
Breath Volume	1.562	20:57	
Air Blank	0.000	20:58	
Control Test	0.079	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 10/27/2016 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,CHRIS
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MC PHERSON

DATE/TIME OF TEST: 10/24/2016 21:18:42

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.082	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.255	21:13
Breath Volume	1.632	21:13
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.237	21:17
Breath Volume	1.191	21:17
Air Blank	0.000	21:17
Control Test	0.082	21:18
Air Blank	0.000	21:18
Diagnostics Check	OK	21:18

Data Download Date/Time: 10/27/2016 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: HARTMANSTORFER,C
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MAUGERI

DATE/TIME OF TEST: 10/25/2016 20:20:38

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.082	20:15
Air Blank	0.000	20:15
Subject Sample # 1	0.000	20:16
Breath Volume	2.144	20:16
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.000	20:19
Breath Volume	1.925	20:19
Air Blank	0.000	20:19
Control Test	0.082	20:20
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 10/27/2016 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKUPSKI, RYAN, T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALVADOR, E
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MICO

DATE/TIME OF TEST: 10/11/2016 23:06:34

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.077	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Control Test	0.078	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: LANGDALE,JYLES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: TROXEL

DATE/TIME OF TEST: 10/12/2016 12:55:52

Diagnostics Check	OK	12:49
Air Blank	0.000	12:49
Control Test	0.077	12:50
Air Blank	0.000	12:50
Subject Sample # 1	0.252	12:51
Breath Volume	1.945	12:51
Air Blank	0.000	12:52
Air Blank	0.000	12:53
Subject Sample # 2	0.244	12:54
Breath Volume	1.617	12:54
Air Blank	0.000	12:54
Control Test	0.076	12:55
Air Blank	0.000	12:55
Diagnostics Check	OK	12:55

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 10:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 10/16/2016 00:57:16

Diagnostics Check	OK	00:50
Air Blank	0.000	00:51
Control Test	0.079	00:51
Air Blank	0.000	00:52
Subject Sample # 1	0.139	00:52
Breath Volume	1.226	00:52
Air Blank	0.000	00:53
Air Blank	0.000	00:55
Subject Sample # 2	0.151	00:55
Breath Volume	1.460	00:55
Air Blank	0.000	00:56
Control Test	0.078	00:56
Air Blank	0.000	00:57
Diagnostics Check	OK	00:57

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOEL,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 10/16/2016 07:25:36

Diagnostics Check	OK	07:19
Air Blank	0.000	07:20
Control Test	0.079	07:20
Air Blank	0.000	07:20
Subject Sample # 1	0.022	07:21
Breath Volume	1.636	07:21
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample # 2	0.020	07:24
Breath Volume	1.628	07:24
Air Blank	0.000	07:24
Control Test	0.078	07:24
Air Blank	0.000	07:25
Diagnostics Check	OK	07:25

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKUPSKI, RYAN, T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI, R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: MENDOZA-VALIENTE

DATE/TIME OF TEST: 10/16/2016 19:38:13

Diagnostics Check	OK	19:32
Air Blank	0.000	19:32
Control Test	0.078	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.136	19:33
Breath Volume	1.656	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.142	19:36
Breath Volume	1.609	19:36
Air Blank	0.000	19:37
Control Test	0.078	19:37
Air Blank	0.000	19:38
Diagnostics Check	OK	19:38

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKUPSKI, RYAN, T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALVADOR, E
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: OTHER
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: CAVASOS-HERERA

DATE/TIME OF TEST: 10/16/2016 20:04:08

Diagnostics Check	OK	19:57
Air Blank	0.000	19:58
Control Test	0.078	19:58
Air Blank	0.000	19:59
Subject Sample # 1	0.072	19:59
Breath Volume	1.863	19:59
Air Blank	0.000	20:00
Air Blank	0.000	20:02
Subject Sample # 2	0.063	20:02
Breath Volume	1.792	20:02
Air Blank	0.000	20:03
Control Test	0.078	20:03
Air Blank	0.000	20:04
Diagnostics Check	OK	20:04

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CROSSWELL,DONALD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/16/2016 23:39:55

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.268	23:35
Breath Volume	2.437	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:37
Subject Sample # 2	0.265	23:38
Breath Volume	2.261	23:38
Air Blank	0.000	23:38
Control Test	0.077	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CAVASOS

DATE/TIME OF TEST: 10/17/2016 02:46:45

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.000	02:42
Breath Volume	1.410	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.000	02:45
Breath Volume	1.519	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CROSSWELL,DONALD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/17/2016 08:23:20

Diagnostics Check	OK	08:17
Air Blank	0.000	08:17
Control Test	0.078	08:17
Air Blank	0.000	08:18
Subject Sample # 1	0.109	08:18
Breath Volume	1.886	08:18
Air Blank	0.000	08:19
Air Blank	0.000	08:21
Subject Sample # 2	0.111	08:21
Breath Volume	1.945	08:21
Air Blank	0.000	08:22
Control Test	0.076	08:22
Air Blank	0.000	08:23
Diagnostics Check	OK	08:23

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CROSSWELL,DONALD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:25

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/17/2016 10:07:08

Diagnostics Check	OK	10:01
Air Blank	0.000	10:01
Control Test	0.079	10:01
Air Blank	0.000	10:02
Subject Sample # 1	0.080	10:02
Breath Volume	2.160	10:02
Air Blank	0.000	10:03
Air Blank	0.000	10:05
Subject Sample # 2	0.082	10:05
Breath Volume	1.925	10:05
Air Blank	0.000	10:06
Control Test	0.079	10:06
Air Blank	0.000	10:07
Diagnostics Check	OK	10:07

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CROSSWELL,DONALD
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/17/2016 11:21:20

Diagnostics Check	OK	11:15
Air Blank	0.000	11:15
Control Test	0.078	11:16
Air Blank	0.000	11:16
Subject Sample # 1	0.056	11:16
Breath Volume	2.566	11:16
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.053	11:19
Breath Volume	1.847	11:19
Air Blank	0.000	11:20
Control Test	0.078	11:20
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKUPSKI, RYAN, T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI, R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WOODWARD

DATE/TIME OF TEST: 10/20/2016 03:32:49

Diagnostics Check	OK	03:26
Air Blank	0.000	03:27
Control Test	0.079	03:27
Air Blank	0.000	03:27
Subject Sample # 1	0.110	03:28
Breath Volume	3.273	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.099	03:31
Breath Volume	1.917	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKUPSKI, RYAN, T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SOTO, M
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 10/25/2016 00:57:43

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:54	
Control Test	0.079	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:56	Subject Test Refused
Breath Volume	0.308	00:56	
Air Blank	0.000	00:56	
Control Test	0.080	00:57	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKUPSKI,RYAN,T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SOTO,M
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ORTIZ-CASTELL

DATE/TIME OF TEST: 10/26/2016 00:07:38

Diagnostics Check	OK	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:01
Subject Sample # 1	0.178	00:02
Breath Volume	1.117	00:02
Air Blank	0.000	00:03
Air Blank	0.000	00:04
Subject Sample # 2	0.169	00:05
Breath Volume	1.207	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:07

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 08:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:12

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 10/28/2016 10:01:06

Diagnostics Check	OK	09:54
Air Blank	0.000	09:55
Control Test	0.079	09:55
Air Blank	0.000	09:56
Subject Sample # 1	0.255	09:56
Breath Volume	1.425	09:56
Air Blank	0.000	09:57
Air Blank	0.000	09:59
Subject Sample # 2	0.259	09:59
Breath Volume	1.246	09:59
Air Blank	0.000	10:00
Control Test	0.079	10:00
Air Blank	0.000	10:01
Diagnostics Check	OK	10:01

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: ALVAREZ,VICTOR
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 08:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:55

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 10/28/2016 16:31:32

Diagnostics Check	OK	16:24
Air Blank	0.000	16:25
Control Test	0.079	16:25
Air Blank	0.000	16:26
Subject Sample # 1	0.078	16:26
Breath Volume	1.425	16:26
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.076	16:29
Breath Volume	1.707	16:29
Air Blank	0.000	16:30
Control Test	0.077	16:30
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

Data Download Date/Time: 11/04/2016 11:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: NORWOOD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 10/07/2016 18:30:05

Diagnostics Check	OK	18:27	
Air Blank	0.000	18:27	
Control Test	0.080	18:27	
Air Blank	0.000	18:28	
Subject Sample	# 1	18:28	Subject Test Refused
Breath Volume	0.000	18:28	
Air Blank	0.000	18:29	
Control Test	0.080	18:29	
Air Blank	0.000	18:30	
Diagnostics Check	OK	18:30	

Data Download Date/Time: 10/26/2016 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,STEVEN,M
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: CHAMBERLAIN,STEVEN
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: ROBLERO-RAMIREZ

DATE/TIME OF TEST: 10/09/2016 03:47:14

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.082	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.194	03:42
Breath Volume	1.792	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.181	03:45
Breath Volume	1.957	03:45
Air Blank	0.000	03:46
Control Test	0.082	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 10/26/2016 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,DARRIN,J
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: HODGE,KYLER
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1451362
DRY GAS EXPIRY DATE: 01/31/2016
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: BARRET

DATE/TIME OF TEST: 10/09/2016 06:47:08

Diagnostics Check	OK	06:37
Air Blank	0.000	06:38
Control Test	0.081	06:38
Air Blank	0.000	06:39
Subject Sample # 1	0.082	06:40
Breath Volume	1.984	06:40
Air Blank	0.000	06:41
Air Blank	0.000	06:42
Subject Sample # 2	0.078	06:45
Breath Volume	1.828	06:45
Air Blank	0.000	06:46
Control Test	0.083	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 10/26/2016 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BEEZLEY,THOMAS
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 10/23/2016 00:00:02

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.081	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.122	23:55
Breath Volume	2.652	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.125	23:58
Breath Volume	2.148	23:58
Air Blank	0.000	23:59
Control Test	0.082	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	00:00

Data Download Date/Time: 10/26/2016 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000951

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: BEEZLEY,THOMAS
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 10/24/2016 01:30:14

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.081	01:24
Air Blank	0.000	01:25
Subject Sample # 1	0.097	01:25
Breath Volume	1.312	01:25
Air Blank	0.000	01:26
Air Blank	0.000	01:28
Subject Sample # 2	0.094	01:28
Breath Volume	2.003	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Diagnostics Check	OK	01:30

Data Download Date/Time: 10/26/2016 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/22/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: LYSTON

DATE/TIME OF TEST: 10/10/2016 03:50:11

Diagnostics Check	OK	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.205	03:44
Breath Volume	1.980	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.201	03:48
Breath Volume	1.886	03:48
Air Blank	0.000	03:49
Control Test	0.077	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 10/29/2016 02:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,N,

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: LAROSA,R

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 10/20/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1737087

DRY GAS EXPIRY DATE: 09/12/2017

VIOLATION CODE: DUI

ARREST TIME: 22:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: FUSSELL

DATE/TIME OF TEST: 10/25/2016 23:27:39

Diagnostics Check	OK	23:20
Air Blank	0.000	23:20
Control Test	0.076	23:20
Air Blank	0.000	23:21
Subject Sample # 1	0.111	23:23
Breath Volume	2.757	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.110	23:26
Breath Volume	1.738	23:26
Air Blank	0.000	23:26
Control Test	0.075	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 11/21/2016 16:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SHERER,DANIEL
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: NORIEGA

DATE/TIME OF TEST: 10/01/2016 01:55:52

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.137	01:51
Breath Volume	1.582	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:53
Subject Sample # 2	0.145	01:54
Breath Volume	1.968	01:54
Air Blank	0.000	01:54
Control Test	0.077	01:55
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: ADAMS,PARISE
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HENREY

DATE/TIME OF TEST: 10/02/2016 02:44:37

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.078	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.254	02:40
Breath Volume	2.425	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.258	02:43
Breath Volume	4.585	02:43
Air Blank	0.000	02:43
Control Test	0.078	02:44
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,D,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: FITE,B
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: BINGER

DATE/TIME OF TEST: 10/02/2016 21:57:41

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.079	21:52
Air Blank	0.000	21:52
Subject Sample # 1	0.264	21:53
Breath Volume	3.847	21:53
Air Blank	0.000	21:53
Air Blank	0.000	21:55
Subject Sample # 2	0.256	21:56
Breath Volume	3.445	21:56
Air Blank	0.000	21:56
Control Test	0.076	21:57
Air Blank	0.000	21:57
Diagnostics Check	OK	21:57

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTHEWS,JESSE,L
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: BATEMAN,JASON
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 10/03/2016 02:13:55

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.079	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.136	02:09
Breath Volume	2.566	02:09
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.132	02:12
Breath Volume	1.664	02:12
Air Blank	0.000	02:13
Control Test	0.078	02:13
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,K,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: NORRIS,R
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:10

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/03/2016 12:50:47

Diagnostics Check	OK	12:44
Air Blank	0.000	12:45
Control Test	0.079	12:45
Air Blank	0.000	12:46
Subject Sample # 1	0.000	12:46
Breath Volume	4.664	12:46
Air Blank	0.000	12:46
Air Blank	0.000	12:48
Subject Sample # 2	0.000	12:49
Breath Volume	3.484	12:49
Air Blank	0.000	12:49
Control Test	0.078	12:50
Air Blank	0.000	12:50
Diagnostics Check	OK	12:50

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/03/2016 22:29:52

Diagnostics Check	OK	22:22
Air Blank	0.000	22:23
Control Test	0.080	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.161	22:25
Breath Volume	1.570	22:25
Air Blank	0.000	22:25
Air Blank	0.000	22:27
Subject Sample # 2	0.157	22:28
Breath Volume	1.546	22:28
Air Blank	0.000	22:28
Control Test	0.079	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/05/2016 03:28:12

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:25	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:26	Subject Test Refused
Breath Volume	0.000	03:26	
Air Blank	0.000	03:27	
Control Test	0.079	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/05/2016 03:34:32

Diagnostics Check	OK	03:30	
Air Blank	0.000	03:30	
Control Test	0.080	03:30	
Air Blank	0.000	03:31	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.183	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CORNELIUS

DATE/TIME OF TEST: 10/07/2016 00:52:21

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:50	
Air Blank	0.000	00:50	
Subject Sample	# 1	00:50	Subject Test Refused
Breath Volume	0.000	00:50	
Air Blank	0.000	00:51	
Control Test	0.079	00:51	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMIDT,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: EVERK

DATE/TIME OF TEST: 10/08/2016 01:51:55

Diagnostics Check	OK	01:45
Air Blank	0.000	01:45
Control Test	0.079	01:45
Air Blank	0.000	01:46
Subject Sample # 1	0.188	01:47
Breath Volume	1.632	01:47
Air Blank	0.000	01:48
Air Blank	0.000	01:49
Subject Sample # 2	0.191	01:50
Breath Volume	1.476	01:50
Air Blank	0.000	01:50
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 10/08/2016 02:22:21

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.286	02:17
Breath Volume	1.289	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.302	02:20
Breath Volume	1.203	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT JUSTE,MILTON,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SAINT JUSTE,MILTON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DEVANE

DATE/TIME OF TEST: 10/08/2016 03:25:50

Diagnostics Check	OK	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:18	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:21	No Sample Provided
Breath Volume	0.000	03:21	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 2	03:24	Subject Test Refused
Breath Volume	0.000	03:24	
Air Blank	0.000	03:24	
Control Test	0.080	03:25	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT JUSTE, MILTON,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SAINT JUSTE, MILTON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: DEVANE

DATE/TIME OF TEST: 10/08/2016 03:41:06

Diagnostics Check	OK	03:38	
Air Blank	0.000	03:38	
Control Test	0.080	03:39	
Air Blank	0.000	03:39	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.080	03:40	
Air Blank	0.000	03:41	
Diagnostics Check	OK	03:41	

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,D,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: URANE

DATE/TIME OF TEST: 10/08/2016 04:51:10

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Subject Sample # 1	0.206	04:46
Breath Volume	1.542	04:46
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.207	04:49
Breath Volume	1.371	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: BYRNE

DATE/TIME OF TEST: 10/09/2016 21:23:30

Diagnostics Check	OK	21:16
Air Blank	0.000	21:17
Control Test	0.080	21:17
Air Blank	0.000	21:17
Subject Sample # 1	0.181	21:18
Breath Volume	3.222	21:18
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.181	21:21
Breath Volume	2.117	21:21
Air Blank	0.000	21:22
Control Test	0.078	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYBORN,GEORGE,K
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: RAYBORN,GEORGE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 10/09/2016 22:23:49

Diagnostics Check	OK	22:12	
Air Blank	0.000	22:12	
Control Test	0.078	22:13	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:16	No Sample Provided
Breath Volume	0.000	22:16	
Air Blank	0.000	22:17	
Air Blank	0.000	22:19	
Subject Sample	# 2	22:22	No Sample Provided
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.076	22:23	
Air Blank		22:23	RFI Detect
Air Blank	0.000	22:23	

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORSTRICK, DAKOTA, S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 10/12/2016 14:28:37

Diagnostics Check	OK	14:20
Air Blank	0.000	14:21
Control Test	0.080	14:21
Air Blank	0.000	14:22
Subject Sample # 1	0.000	14:23
Breath Volume	2.464	14:23
Air Blank	0.000	14:24
Air Blank	0.000	14:26
Subject Sample # 2	0.000	14:27
Breath Volume	1.273	14:27
Air Blank	0.000	14:27
Control Test	0.079	14:28
Air Blank	0.000	14:28
Diagnostics Check	OK	14:28

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: JOHNSON,ANDY
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 10/12/2016 23:36:35

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.000	23:32
Breath Volume	1.406	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	1.367	23:35
Air Blank	0.000	23:35
Control Test	0.079	23:35
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 10/13/2016 00:51:29

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.154	00:46
Breath Volume	2.632	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.158	00:49
Breath Volume	2.425	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: HAIRE

DATE/TIME OF TEST: 10/14/2016 00:02:42

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.079	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.211	23:58
Breath Volume	1.355	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.209	00:01
Breath Volume	1.910	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,JAMES,H
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PITTMAN,JAMES
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: BLUE

DATE/TIME OF TEST: 10/14/2016 01:48:39

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.080	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.181	01:44
Breath Volume	2.257	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.182	01:47
Breath Volume	2.574	01:47
Air Blank	0.000	01:47
Control Test	0.079	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK, JOSHUA, R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BULLOCK, JOSHUA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: CROUCH

DATE/TIME OF TEST: 10/14/2016 03:47:25

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.088	03:42
Breath Volume	1.570	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.086	03:45
Breath Volume	1.617	03:45
Air Blank	0.000	03:46
Control Test	0.079	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: SHERER,DANIEL
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 10/15/2016 03:55:45

Diagnostics Check	OK	03:49
Air Blank	0.000	03:49
Control Test	0.080	03:50
Air Blank	0.000	03:50
Subject Sample # 1	0.162	03:51
Breath Volume	2.203	03:51
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.182	03:54
Breath Volume	3.878	03:54
Air Blank	0.000	03:54
Control Test	0.077	03:55
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: HOSPINAL

DATE/TIME OF TEST: 10/15/2016 21:08:25

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Subject Sample # 1	0.131	21:03
Breath Volume	3.398	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.126	21:06
Breath Volume	2.777	21:06
Air Blank	0.000	21:07
Control Test	0.079	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LECKINGER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 10/15/2016 21:22:02

Diagnostics Check	OK	21:16
Air Blank	0.000	21:16
Control Test	0.077	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.186	21:17
Breath Volume	2.691	21:17
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.190	21:20
Breath Volume	2.632	21:20
Air Blank	0.000	21:21
Control Test	0.080	21:21
Air Blank	0.000	21:21
Diagnostics Check	OK	21:22

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OHARA,K
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 10/16/2016 04:34:49

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.080	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1 0.205	04:30	
Breath Volume	3.203	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	04:33	RFI Detect
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: OHARA,K
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 09/08/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 10/16/2016 04:46:21

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.080	04:37
Air Blank	0.000	04:37
Subject Sample # 1	0.169	04:38
Breath Volume	1.238	04:38
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.197	04:41
Breath Volume	1.359	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 3	0.180	04:44
Breath Volume	1.101	04:44
Air Blank	0.000	04:45
Control Test	0.079	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 10/18/2016 11:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 10/18/2016 22:42:22

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Subject Sample # 1	0.229	22:38
Breath Volume	2.222	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.216	22:40
Breath Volume	2.320	22:40
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: AYERS JR

DATE/TIME OF TEST: 10/20/2016 01:17:49

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.135	01:13
Breath Volume	1.773	01:13
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.143	01:16
Breath Volume	1.742	01:16
Air Blank	0.000	01:16
Control Test	0.080	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: FORBES

DATE/TIME OF TEST: 10/20/2016 02:47:18

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.163	02:42
Breath Volume	1.664	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.161	02:45
Breath Volume	1.687	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/21/2016 02:41:51

Diagnostics Check	OK	02:39	
Air Blank	0.000	02:39	
Control Test	0.078	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Control Test	0.079	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITTMAN,JAMES,H
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: PITTMAN,JAMES
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/21/2016 03:59:12

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.132	03:54
Breath Volume	2.457	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.129	03:57
Breath Volume	2.691	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 10/23/2016 03:53:25

Diagnostics Check	OK	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Subject Sample # 1	0.059	03:48
Breath Volume	2.824	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.058	03:51
Breath Volume	3.035	03:51
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 10/24/2016 21:58:07

Diagnostics Check	OK	21:51
Air Blank	0.000	21:52
Control Test	0.080	21:52
Air Blank	0.000	21:53
Subject Sample # 1	0.168	21:53
Breath Volume	1.707	21:53
Air Blank	0.000	21:54
Air Blank	0.000	21:56
Subject Sample # 2	0.168	21:56
Breath Volume	1.738	21:56
Air Blank	0.000	21:57
Control Test	0.079	21:57
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PETERSON IV

DATE/TIME OF TEST: 10/25/2016 23:59:42

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.000	23:55
Breath Volume	2.335	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.000	23:58
Breath Volume	3.148	23:58
Air Blank	0.000	23:58
Control Test	0.080	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULZE,JON,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SCHULZE,JON
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: PETTIBONE

DATE/TIME OF TEST: 10/26/2016 21:39:55

Diagnostics Check	OK	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:30
Air Blank	0.000	21:30
Subject Sample # 1	0.294	21:31
Breath Volume	2.269	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.258	21:34
Breath Volume	1.289	21:34
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 3	0.280	21:38
Breath Volume	1.589	21:38
Air Blank	0.000	21:38
Control Test	0.078	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/27/2016 00:15:29

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.080	00:09
Air Blank	0.000	00:10
Subject Sample # 1	0.108	00:10
Breath Volume	2.292	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.105	00:13
Breath Volume	1.679	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,BRANDON,J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON,BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 10/29/2016 02:41:08

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:39	Subject Test Refused
Breath Volume	0.000	02:39	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000958

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPRADLIN,BRETT,J
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: BAYER,ADAM
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: KIDWELL-BOZEMAN

DATE/TIME OF TEST: 10/30/2016 07:54:37

Diagnostics Check	OK	07:48
Air Blank	0.000	07:48
Control Test	0.078	07:49
Air Blank	0.000	07:49
Subject Sample # 1	0.183	07:50
Breath Volume	3.421	07:50
Air Blank	0.000	07:50
Air Blank	0.000	07:52
Subject Sample # 2	0.192	07:52
Breath Volume	3.195	07:52
Air Blank	0.000	07:53
Control Test	0.078	07:54
Air Blank	0.000	07:54
Diagnostics Check	OK	07:54

Data Download Date/Time: 11/08/2016 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: LAMSON

DATE/TIME OF TEST: 10/15/2016 04:48:00

Diagnostics Check	OK	04:41
Air Blank	0.000	04:41
Control Test	0.080	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.129	04:43
Breath Volume	3.148	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:45
Subject Sample # 2	0.127	04:46
Breath Volume	3.273	04:46
Air Blank	0.000	04:47
Control Test	0.080	04:47
Air Blank	0.000	04:47
Diagnostics Check	OK	04:48

Data Download Date/Time: 10/17/2016 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:20

SUBJECT'S LAST NAME: ULYATT

DATE/TIME OF TEST: 10/21/2016 11:53:26

Diagnostics Check	OK	11:47
Air Blank	0.000	11:47
Control Test	0.080	11:47
Air Blank	0.000	11:48
Subject Sample # 1	0.243	11:48
Breath Volume	2.660	11:48
Air Blank	0.000	11:49
Air Blank	0.000	11:51
Subject Sample # 2	0.242	11:51
Breath Volume	2.175	11:51
Air Blank	0.000	11:52
Control Test	0.079	11:52
Air Blank	0.000	11:53
Diagnostics Check	OK	11:53

Data Download Date/Time: 11/04/2016 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FREDRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FREDRICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/17/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BETTINGER

DATE/TIME OF TEST: 10/31/2016 01:14:40

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.080	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.149	01:10
Breath Volume	1.390	01:10
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.146	01:13
Breath Volume	1.132	01:13
Air Blank	0.000	01:13
Control Test	0.079	01:14
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 11/04/2016 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS, BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 10/01/2016 04:13:42

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.161	04:09
Breath Volume	1.457	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.166	04:12
Breath Volume	2.585	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/06/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE, JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: ELLINGSGARD

DATE/TIME OF TEST: 10/02/2016 03:28:40

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.079	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.138	03:22
Breath Volume	1.429	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.129	03:27
Breath Volume	1.144	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 10/06/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RITTER, ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: BANNER

DATE/TIME OF TEST: 10/03/2016 02:54:33

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.193	02:50
Breath Volume	2.089	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.197	02:53
Breath Volume	2.644	02:53
Air Blank	0.000	02:53
Control Test	0.078	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 10/06/2016 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO, GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: KOWNACKI

DATE/TIME OF TEST: 10/05/2016 04:33:46

Diagnostics Check	OK	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:28
Subject Sample # 1	0.155	04:29
Breath Volume	3.179	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.148	04:32
Breath Volume	2.449	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:33
Air Blank	0.000	04:33
Diagnostics Check	OK	04:33

Data Download Date/Time: 10/06/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SAYLER, ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 10/06/2016 02:21:16

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.222	02:16
Breath Volume	2.480	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.222	02:19
Breath Volume	2.300	02:19
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/06/2016 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA, EDDIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:09

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/06/2016 06:41:15

Diagnostics Check	OK	06:35
Air Blank	0.000	06:35
Control Test	0.079	06:36
Air Blank	0.000	06:36
Subject Sample # 1	0.007	06:37
Breath Volume	1.621	06:37
Air Blank	0.000	06:37
Air Blank	0.000	06:39
Subject Sample # 2	0.000	06:39
Breath Volume	1.863	06:39
Air Blank	0.000	06:40
Control Test	0.079	06:40
Air Blank	0.000	06:41
Diagnostics Check	OK	06:41

Data Download Date/Time: 10/06/2016 11:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JAMES,JOHN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 10/06/2016 08:19:36

Diagnostics Check	OK	08:13
Air Blank	0.000	08:14
Control Test	0.079	08:14
Air Blank	0.000	08:14
Subject Sample # 1	0.164	08:15
Breath Volume	2.058	08:15
Air Blank	0.000	08:15
Air Blank	0.000	08:17
Subject Sample # 2	0.153	08:18
Breath Volume	2.410	08:18
Air Blank	0.000	08:18
Control Test	0.078	08:19
Air Blank	0.000	08:19
Diagnostics Check	OK	08:19

Data Download Date/Time: 10/06/2016 11:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA, EDDIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: STOKES

DATE/TIME OF TEST: 10/07/2016 20:47:28

Diagnostics Check	OK	20:39
Air Blank	0.000	20:40
Control Test	0.079	20:40
Air Blank	0.000	20:41
Subject Sample # 1	0.141	20:42
Breath Volume	1.261	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.141	20:46
Breath Volume	1.847	20:46
Air Blank	0.000	20:46
Control Test	0.079	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOUGHTON, WELBIE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/07/2016 22:01:34

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.079	21:53	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:57	No Sample Provided
Breath Volume	0.000	21:57	
Air Blank	0.000	21:58	
Air Blank	0.000	22:00	
Subject Sample	# 2	22:00	Subject Test Refused
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.079	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JONES, EUGENE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: CUTLIFF

DATE/TIME OF TEST: 10/08/2016 23:13:19

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.079	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.131	23:09
Breath Volume	1.109	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.124	23:11
Breath Volume	1.339	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ZITO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PEROZO

DATE/TIME OF TEST: 10/09/2016 02:39:35

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.122	02:35
Breath Volume	3.542	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.122	02:38
Breath Volume	3.160	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORSO, ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: ORTIZ LIRIANO

DATE/TIME OF TEST: 10/09/2016 03:40:15

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.079	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.150	03:35
Breath Volume	2.511	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.155	03:38
Breath Volume	2.308	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE, FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 10/09/2016 04:12:45

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.200	04:07
Breath Volume	1.371	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.203	04:11
Breath Volume	2.281	04:11
Air Blank	0.000	04:11
Control Test	0.078	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RUSSAK, KEVIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: BAEZCAPELES

DATE/TIME OF TEST: 10/09/2016 05:24:34

Diagnostics Check	OK	05:13	
Air Blank	0.000	05:13	
Control Test	0.079	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1	05:17	No Sample Provided
Breath Volume	0.000	05:17	
Air Blank	0.000	05:18	
Air Blank	0.000	05:20	
Subject Sample	# 2	05:23	No Sample Provided
Breath Volume	0.000	05:23	
Air Blank	0.000	05:23	
Control Test	0.079	05:23	
Air Blank	0.000	05:24	
Diagnostics Check	OK	05:24	

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCSPADDEN, JASON
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: CABALAMECA

DATE/TIME OF TEST: 10/10/2016 00:40:53

Diagnostics Check	OK	00:34
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.133	00:36
Breath Volume	3.320	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.137	00:39
Breath Volume	3.300	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:40
Diagnostics Check	OK	00:40

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SALCEDO, NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MARCELIN

DATE/TIME OF TEST: 10/10/2016 03:14:03

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.079	03:03	
Air Blank	0.000	03:03	
Subject Sample	# 1 0.010	03:07	Volume Not Met
Breath Volume	0.285	03:07	
Air Blank	0.000	03:07	
Air Blank	0.000	03:09	
Subject Sample	# 2 0.010	03:12	Volume Not Met
Breath Volume	1.039	03:12	
Air Blank	0.000	03:13	
Control Test	0.080	03:13	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SALCEDO, NICHOLAS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MARCELIN

DATE/TIME OF TEST: 10/10/2016 03:22:55

Diagnostics Check	OK	03:14
Air Blank	0.000	03:15
Control Test	0.078	03:15
Air Blank	0.000	03:16
Subject Sample # 1	0.008	03:17
Breath Volume	1.812	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.008	03:21
Breath Volume	1.351	03:21
Air Blank	0.000	03:22
Control Test	0.079	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE, SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CORA SIERRA

DATE/TIME OF TEST: 10/11/2016 04:10:49

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.078	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.044	04:06
Breath Volume	1.421	04:06
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.050	04:09
Breath Volume	2.292	04:09
Air Blank	0.000	04:09
Control Test	0.079	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ, EDUARDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 10/12/2016 04:36:18

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:25	
Air Blank	0.000	04:25	
Subject Sample	# 1 0.184	04:28	Volume Not Met
Breath Volume	0.679	04:28	
Air Blank	0.000	04:29	
Air Blank	0.000	04:31	
Subject Sample	# 2 0.216	04:31	
Breath Volume	1.253	04:31	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 3 0.220	04:34	
Breath Volume	1.679	04:34	
Air Blank	0.000	04:35	
Control Test	0.079	04:35	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:36	

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RANDO,STEPHEN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: HIGH

DATE/TIME OF TEST: 10/12/2016 08:52:26

Diagnostics Check	OK	08:46
Air Blank	0.000	08:46
Control Test	0.079	08:46
Air Blank	0.000	08:47
Subject Sample # 1	0.080	08:48
Breath Volume	2.183	08:48
Air Blank	0.000	08:48
Air Blank	0.000	08:50
Subject Sample # 2	0.076	08:51
Breath Volume	2.738	08:51
Air Blank	0.000	08:51
Control Test	0.078	08:51
Air Blank	0.000	08:52
Diagnostics Check	OK	08:52

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK, STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:09

SUBJECT'S LAST NAME: GIORDANO

DATE/TIME OF TEST: 10/13/2016 00:37:07

Diagnostics Check	OK	00:30
Air Blank	0.000	00:30
Control Test	0.078	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.211	00:32
Breath Volume	1.132	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample # 2	0.227	00:35
Breath Volume	2.355	00:35
Air Blank	0.000	00:36
Control Test	0.078	00:36
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 10/13/2016 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DOBBINS, SEAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: KENDALL

DATE/TIME OF TEST: 10/15/2016 03:53:27

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.078	03:48
Air Blank	0.000	03:48
Subject Sample # 1	0.086	03:49
Breath Volume	2.585	03:49
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.091	03:52
Breath Volume	2.324	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 10/19/2016 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA, EDDIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: CORBETT

DATE/TIME OF TEST: 10/15/2016 22:58:25

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.078	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.204	22:54
Breath Volume	1.589	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.189	22:56
Breath Volume	1.765	22:56
Air Blank	0.000	22:57
Control Test	0.078	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 10/19/2016 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ, FRANKIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/16/2016 00:35:00

Diagnostics Check	OK	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.172	00:29
Breath Volume	1.644	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.181	00:33
Breath Volume	1.171	00:33
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:35

Data Download Date/Time: 10/19/2016 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPEER, ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DE ACETIS

DATE/TIME OF TEST: 10/16/2016 01:59:45

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.132	01:55
Breath Volume	1.867	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:57
Subject Sample # 2	0.129	01:58
Breath Volume	1.167	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 10/19/2016 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WINBURN, TAYLOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LUM

DATE/TIME OF TEST: 10/16/2016 23:27:07

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.184	23:22
Breath Volume	2.316	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.191	23:25
Breath Volume	1.902	23:25
Air Blank	0.000	23:26
Control Test	0.080	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 10/19/2016 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IV, ANDREW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/17/2016 01:49:53

Diagnostics Check	OK	01:43
Air Blank	0.000	01:44
Control Test	0.079	01:44
Air Blank	0.000	01:45
Subject Sample # 1	0.023	01:45
Breath Volume	2.292	01:45
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.022	01:48
Breath Volume	2.554	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:49
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/19/2016 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE, FORREST
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 10/18/2016 03:51:05

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.079	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.113	03:46
Breath Volume	2.878	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.113	03:49
Breath Volume	3.031	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/19/2016 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA, HECTOR, A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS, CHARLES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 10/19/2016 01:51:23

Diagnostics Check	OK	01:42
Air Blank	0.000	01:42
Control Test	0.079	01:43
Air Blank	0.000	01:43
Subject Sample # 1	0.147	01:44
Breath Volume	1.269	01:44
Air Blank	0.000	01:44
Air Blank	0.000	01:46
Subject Sample # 2	0.174	01:46
Breath Volume	1.996	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 3	0.171	01:49
Breath Volume	2.144	01:49
Air Blank	0.000	01:50
Control Test	0.079	01:50
Air Blank	0.000	01:51
Diagnostics Check	OK	01:51

Data Download Date/Time: 10/19/2016 12:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU, SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 10/20/2016 20:03:40

Diagnostics Check	OK	19:57
Air Blank	0.000	19:57
Control Test	0.079	19:58
Air Blank	0.000	19:58
Subject Sample # 1	0.199	19:59
Breath Volume	2.437	19:59
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.185	20:02
Breath Volume	1.796	20:02
Air Blank	0.000	20:02
Control Test	0.078	20:03
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 10/27/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SORRELLS, BRIAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 10/24/2016 01:45:40

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.157	01:41
Breath Volume	2.085	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.156	01:44
Breath Volume	1.800	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/27/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SCOTT, ASHLEY
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BRUDER

DATE/TIME OF TEST: 10/25/2016 01:41:18

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.167	01:36
Breath Volume	3.101	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.163	01:39
Breath Volume	3.730	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 10/27/2016 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE, JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: RODRIGUEZ JR.

DATE/TIME OF TEST: 10/30/2016 04:32:49

Diagnostics Check	OK	04:26
Air Blank	0.000	04:27
Control Test	0.079	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.071	04:28
Breath Volume	1.851	04:28
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.072	04:31
Breath Volume	1.363	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

Data Download Date/Time: 11/03/2016 17:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE, JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 10/31/2016 02:56:36

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.080	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:52	No Sample Provided
Breath Volume	0.000	02:52	
Air Blank	0.000	02:53	
Air Blank	0.000	02:55	
Subject Sample	# 2	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:56	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 11/03/2016 17:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BURNETTE,JAVON,A

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: BURNETTE,JAVON

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 21:23

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: HIGGINS

DATE/TIME OF TEST: 10/03/2016 22:04:55

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.222	22:00
Breath Volume	1.726	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:02
Subject Sample # 2	0.206	22:03
Breath Volume	1.738	22:03
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:04
Diagnostics Check	OK	22:04

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: KAUFFMAN, MITCHELL

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VELASCO GIRON

DATE/TIME OF TEST: 10/08/2016 23:51:08

Diagnostics Check	OK	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:45
Air Blank	0.000	23:45
Subject Sample # 1	0.196	23:46
Breath Volume	2.347	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 2	0.185	23:49
Breath Volume	3.347	23:49
Air Blank	0.000	23:50
Control Test	0.080	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: DEKLE,MATTHEW,L

BREATH TEST OPERATOR AGENCY: Quincy PD

ARRESTING OFFICER: BROXSIE,SHAYNE

ARRESTING OFFICER AGENCY: Quincy PD

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 12:39

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/16/2016 01:29:31

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.173	01:25
Breath Volume	1.816	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.168	01:28
Breath Volume	1.781	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,CHARLES,E

BREATH TEST OPERATOR AGENCY: Gretna PD

ARRESTING OFFICER: SANDER,M

ARRESTING OFFICER AGENCY: Gretna PD

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/20/2016 00:18:00

Diagnostics Check	OK	00:11
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:12
Subject Sample # 1	0.153	00:13
Breath Volume	2.000	00:13
Air Blank	0.000	00:14
Air Blank	0.000	00:15
Subject Sample # 2	0.141	00:16
Breath Volume	1.945	00:16
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:17
Diagnostics Check	OK	00:17

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDESMA,LEOPOLDO,E
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: COPELAND,MICHAEL
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 10/22/2016 04:29:27

Diagnostics Check	OK	04:21	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1 0.165	04:23	
Breath Volume	1.468	04:23	
Air Blank	0.000	04:23	
Air Blank	0.000	04:25	
Subject Sample	# 2 0.167	04:28	
Breath Volume	1.121	04:28	
Air Blank		04:28	Purge Fail
Air Blank	0.000	04:29	

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDESMA,LEOPOLDO,E
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: ARMSTRONG,JOSHUA
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 10/22/2016 06:30:48

Diagnostics Check	OK	06:27	
Air Blank	0.000	06:27	
Control Test	0.080	06:27	
Air Blank	0.000	06:28	
Subject Sample	# 1	06:29	Subject Test Refused
Breath Volume	0.000	06:29	
Air Blank	0.000	06:29	
Control Test	0.080	06:30	
Air Blank	0.000	06:30	
Diagnostics Check	OK	06:30	

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDESMA,LEOPOLDO,E
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: ARMSTRONG,JOSHUA
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 10/22/2016 06:36:55

Diagnostics Check	OK	06:34	
Air Blank	0.000	06:34	
Control Test	0.081	06:34	
Air Blank	0.000	06:35	
Subject Sample	# 1	06:35	Subject Test Refused
Breath Volume	0.000	06:35	
Air Blank	0.000	06:36	
Control Test	0.080	06:36	
Air Blank	0.000	06:36	
Diagnostics Check	OK	06:36	

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDESMA,LEOPOLDO,E
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: ARMSTRONG,JOSHUA
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 10/22/2016 06:51:20

Diagnostics Check	OK	06:48	
Air Blank	0.000	06:48	
Control Test	0.080	06:49	
Air Blank	0.000	06:49	
Subject Sample	# 1	06:49	Subject Test Refused
Breath Volume	0.000	06:49	
Air Blank	0.000	06:50	
Control Test	0.080	06:50	
Air Blank	0.000	06:51	
Diagnostics Check	OK	06:51	

Data Download Date/Time: 10/25/2016 11:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: ATKINS II

DATE/TIME OF TEST: 10/03/2016 04:59:18

Diagnostics Check	OK	04:52
Air Blank	0.000	04:53
Control Test	0.080	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.058	04:54
Breath Volume	1.464	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.059	04:57
Breath Volume	1.671	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 10/14/2016 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SANTONE,LACEY,M

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: BYWATER,K

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 163437

DRY GAS EXPIRY DATE: 10/30/2018

VIOLATION CODE: DUI

ARREST TIME: 03:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SEGREST

DATE/TIME OF TEST: 10/08/2016 03:39:07

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.199	03:31
Breath Volume	1.707	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.166	03:34
Breath Volume	1.609	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 3	0.185	03:37
Breath Volume	2.882	03:37
Air Blank	0.000	03:38
Control Test	0.078	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 10/14/2016 09:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONTRERAS,BRYANT,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: CONTRERAS,BRYANT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: MELELENDEZ

DATE/TIME OF TEST: 10/08/2016 22:36:12

Diagnostics Check	OK	22:27
Air Blank	0.000	22:27
Control Test	0.080	22:28
Air Blank	0.000	22:28
Subject Sample # 1	0.166	22:29
Breath Volume	1.105	22:29
Air Blank	0.000	22:30
Air Blank	0.000	22:32
Subject Sample # 2	0.150	22:34
Breath Volume	1.113	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 10/14/2016 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: VELEZ,JULIO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:51

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 10/10/2016 01:24:14

Diagnostics Check	OK	01:16
Air Blank	0.000	01:16
Control Test	0.079	01:16
Air Blank	0.000	01:17
Subject Sample # 1	0.148	01:18
Breath Volume	3.054	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.133	01:22
Breath Volume	3.351	01:22
Air Blank	0.000	01:23
Control Test	0.078	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 10/14/2016 09:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 10/14/2016 20:52:07

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:44
Air Blank	0.000	20:45
Subject Sample # 1	0.036	20:47
Breath Volume	1.269	20:47
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 2	0.032	20:50
Breath Volume	1.851	20:50
Air Blank	0.000	20:51
Control Test	0.078	20:51
Air Blank	0.000	20:52
Diagnostics Check	OK	20:52

Data Download Date/Time: 11/18/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ESH,JOSHUA
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: RODRIGUEZ APONTE

DATE/TIME OF TEST: 10/15/2016 22:53:02

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.136	22:48
Breath Volume	2.953	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:50
Subject Sample # 2	0.131	22:51
Breath Volume	2.058	22:51
Air Blank	0.000	22:52
Control Test	0.078	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:53

Data Download Date/Time: 11/18/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: STINSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 10/16/2016 04:15:33

Diagnostics Check	OK	04:08
Air Blank	0.000	04:08
Control Test	0.080	04:09
Air Blank	0.000	04:09
Subject Sample # 1	0.175	04:10
Breath Volume	3.503	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.174	04:13
Breath Volume	3.304	04:13
Air Blank	0.000	04:14
Control Test	0.077	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

Data Download Date/Time: 11/18/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ESH,J
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CHASE

DATE/TIME OF TEST: 10/17/2016 01:41:09

Diagnostics Check	OK	01:32
Air Blank	0.000	01:32
Control Test	0.081	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.144	01:34
Breath Volume	1.429	01:34
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.150	01:39
Breath Volume	1.890	01:39
Air Blank	0.000	01:40
Control Test	0.079	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 11/18/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,C,B
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: BYWATER,K
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 17:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: TEMPEs

DATE/TIME OF TEST: 10/18/2016 18:49:00

Diagnostics Check	OK	18:41
Air Blank	0.000	18:41
Control Test	0.081	18:41
Air Blank	0.000	18:42
Subject Sample # 1	0.169	18:44
Breath Volume	1.847	18:44
Air Blank	0.000	18:44
Air Blank	0.000	18:46
Subject Sample # 2	0.153	18:47
Breath Volume	1.101	18:47
Air Blank	0.000	18:48
Control Test	0.078	18:48
Air Blank	0.000	18:48
Diagnostics Check	OK	18:49

Data Download Date/Time: 11/18/2016 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ESH, JOSHUA
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: DEEGAN

DATE/TIME OF TEST: 10/19/2016 21:14:19

Diagnostics Check	OK	21:08
Air Blank	0.000	21:08
Control Test	0.081	21:09
Air Blank	0.000	21:09
Subject Sample # 1	0.007	21:10
Breath Volume	2.679	21:10
Air Blank	0.000	21:10
Air Blank	0.000	21:12
Subject Sample # 2	0.000	21:12
Breath Volume	2.882	21:12
Air Blank	0.000	21:13
Control Test	0.078	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 11/18/2016 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JORGE, JUNIOR,

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: JORGE, JUNIOR

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 163437

DRY GAS EXPIRY DATE: 10/30/2018

VIOLATION CODE: DUI

ARREST TIME: 04:09

GENDER: Female

OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/20/2016 05:03:55

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.080	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.136	04:59
Breath Volume	1.136	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.142	05:02
Breath Volume	1.289	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: STINSON,ANDREW,L

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: STINSON,ANDREW

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 163437

DRY GAS EXPIRY DATE: 10/30/2018

VIOLATION CODE: DUI

ARREST TIME: 00:22

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/21/2016 01:05:22

Diagnostics Check	OK	00:58
Air Blank	0.000	00:58
Control Test	0.081	00:59
Air Blank	0.000	00:59
Subject Sample # 1	0.055	01:00
Breath Volume	3.039	01:00
Air Blank	0.000	01:01
Air Blank	0.000	01:02
Subject Sample # 2	0.054	01:03
Breath Volume	3.472	01:03
Air Blank	0.000	01:04
Control Test	0.079	01:04
Air Blank	0.000	01:05
Diagnostics Check	OK	01:05

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON, ANDREW, L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: BAXTER, M
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MALDONADO LOVATO

DATE/TIME OF TEST: 10/21/2016 02:55:27

Diagnostics Check	OK	02:42
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.130	02:46
Breath Volume	3.531	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:48
Subject Sample # 2	0.105	02:50
Breath Volume	2.152	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 3	0.130	02:53
Breath Volume	2.769	02:53
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:26

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 10/23/2016 05:55:21

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.081	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.163	05:49
Breath Volume	1.613	05:49
Air Blank	0.000	05:50
Air Blank	0.000	05:52
Subject Sample # 2	0.176	05:53
Breath Volume	2.902	05:53
Air Blank	0.000	05:54
Control Test	0.080	05:54
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,C,B
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ANTHONY,C
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BERRIOS VAZQUEZ

DATE/TIME OF TEST: 10/24/2016 03:27:21

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.080	03:21
Air Blank	0.000	03:22
Subject Sample # 1	0.222	03:22
Breath Volume	1.789	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.230	03:25
Breath Volume	2.832	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGER, WILLIAM, W
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: REDDITT, MATTHEW
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:41

SUBJECT'S LAST NAME: WHEAT

DATE/TIME OF TEST: 10/27/2016 12:08:26

Diagnostics Check	OK	12:02
Air Blank	0.000	12:02
Control Test	0.079	12:03
Air Blank	0.000	12:03
Subject Sample # 1	0.000	12:04
Breath Volume	2.671	12:04
Air Blank	0.000	12:04
Air Blank	0.000	12:06
Subject Sample # 2	0.000	12:07
Breath Volume	2.714	12:07
Air Blank	0.000	12:07
Control Test	0.078	12:07
Air Blank	0.000	12:08
Diagnostics Check	OK	12:08

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,NEFTALI,F
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LATCHMAN,AJYAI
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: GARAY RAMOS

DATE/TIME OF TEST: 10/29/2016 03:40:11

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.226	03:35
Breath Volume	1.457	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.230	03:38
Breath Volume	1.882	03:38
Air Blank	0.000	03:39
Control Test	0.080	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSAM, BERNARD, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: ESH, JOSHUA
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/30/2016 02:29:27

Diagnostics Check	OK	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Subject Sample # 1	0.000	02:25
Breath Volume	2.589	02:25
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.000	02:28
Breath Volume	1.832	02:28
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRISWOLD,DAVID,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GRISWOLD,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/30/2016 07:23:03

Diagnostics Check	OK	07:16
Air Blank	0.000	07:16
Control Test	0.080	07:16
Air Blank	0.000	07:17
Subject Sample # 1	0.000	07:18
Breath Volume	1.644	07:18
Air Blank	0.000	07:18
Air Blank	0.000	07:20
Subject Sample # 2	0.000	07:21
Breath Volume	1.949	07:21
Air Blank	0.000	07:22
Control Test	0.078	07:22
Air Blank	0.000	07:22
Diagnostics Check	OK	07:23

Data Download Date/Time: 11/18/2016 08:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,ROBERT,J
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: CAMPBELL,ROBERT
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: MICELI

DATE/TIME OF TEST: 10/17/2016 00:32:44

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.075	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.178	00:27
Breath Volume	1.335	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:30
Subject Sample # 2	0.181	00:31
Breath Volume	2.285	00:31
Air Blank	0.000	00:31
Control Test	0.075	00:32
Air Blank	0.000	00:32
Diagnostics Check	OK	00:32

Data Download Date/Time: 10/23/2016 23:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ELLER,ASHLEY,B

BREATH TEST OPERATOR AGENCY: Apopka PD

ARRESTING OFFICER: ELLER,ASHLEY

ARRESTING OFFICER AGENCY: Apopka PD

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 13115080A1

DRY GAS EXPIRY DATE: 06/05/2017

VIOLATION CODE: DUI

ARREST TIME: 21:23

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 10/20/2016 22:15:08

Diagnostics Check	OK	22:09
Air Blank	0.000	22:09
Control Test	0.076	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.000	22:10
Breath Volume	3.207	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.000	22:13
Breath Volume	3.242	22:13
Air Blank	0.000	22:14
Control Test	0.076	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 10/23/2016 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: SANCHEZ,JOSE
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 10/20/2016 22:26:38

Diagnostics Check	OK	22:20
Air Blank	0.000	22:20
Control Test	0.077	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.203	22:22
Breath Volume	3.109	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.202	22:25
Breath Volume	1.945	22:25
Air Blank	0.000	22:25
Control Test	0.075	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 10/23/2016 23:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000967

INSTRUMENT REGISTERED TO: APOPKA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLER,ASHLEY,B
BREATH TEST OPERATOR AGENCY: Apopka PD
ARRESTING OFFICER: SANCHEZ,JOSE
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/25/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 10/27/2016 00:01:16

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.076	23:55
Air Blank	0.000	23:56
Subject Sample # 1	0.000	23:56
Breath Volume	3.371	23:56
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 2	0.000	23:59
Breath Volume	1.796	23:59
Air Blank	0.000	00:00
Control Test	0.077	00:00
Air Blank	0.000	00:01
Diagnostics Check	OK	00:01

Data Download Date/Time: 11/28/2016 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON, ANDREW, L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO, PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PENA-ORTIZ

DATE/TIME OF TEST: 10/01/2016 04:59:47

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:52
Subject Sample # 1	0.185	04:55
Breath Volume	2.281	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.183	04:58
Breath Volume	2.410	04:58
Air Blank	0.000	04:58
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: WILKES

DATE/TIME OF TEST: 10/02/2016 01:16:14

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.081	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.064	01:11
Breath Volume	2.253	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.063	01:14
Breath Volume	2.367	01:14
Air Blank	0.000	01:15
Control Test	0.081	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: ORTIZ,D
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 10/02/2016 05:32:47

Diagnostics Check	OK	05:20	
Air Blank	0.000	05:21	
Control Test	0.082	05:21	
Air Blank	0.000	05:22	
Subject Sample	# 1 0.288	05:25	Volume Not Met
Breath Volume	0.152	05:25	
Air Blank	0.000	05:26	
Air Blank	0.000	05:27	
Subject Sample	# 2 0.268	05:30	Volume Not Met
Breath Volume	0.371	05:30	
Air Blank	0.000	05:31	
Control Test	0.079	05:32	
Air Blank	0.000	05:32	
Diagnostics Check	OK	05:32	

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEHL, EDWARD,
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: DIEHL, EDWARD
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:57

SUBJECT'S LAST NAME: POUERIET

DATE/TIME OF TEST: 10/02/2016 15:26:27

Diagnostics Check	OK	15:19
Air Blank	0.000	15:19
Control Test	0.081	15:20
Air Blank	0.000	15:20
Subject Sample # 1	0.185	15:21
Breath Volume	2.390	15:21
Air Blank	0.000	15:22
Air Blank	0.000	15:24
Subject Sample # 2	0.181	15:24
Breath Volume	1.394	15:24
Air Blank	0.000	15:25
Control Test	0.082	15:25
Air Blank	0.000	15:26
Diagnostics Check	OK	15:26

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO,ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 10/03/2016 18:37:11

Diagnostics Check	OK	18:30
Air Blank	0.000	18:30
Control Test	0.082	18:30
Air Blank	0.000	18:31
Subject Sample # 1	0.211	18:32
Breath Volume	2.273	18:32
Air Blank	0.000	18:32
Air Blank	0.000	18:34
Subject Sample # 2	0.218	18:35
Breath Volume	1.753	18:35
Air Blank	0.000	18:36
Control Test	0.081	18:36
Air Blank	0.000	18:37
Diagnostics Check	OK	18:37

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO,ANTHONY,J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO,ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FLUKE

DATE/TIME OF TEST: 10/04/2016 03:50:44

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1 0.237	03:43	Volume Not Met
Breath Volume	0.128	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2 0.228	03:49	Volume Not Met
Breath Volume	0.019	03:49	
Air Blank		03:50	Purge Fail
Air Blank	0.000	03:50	

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYLE,OWEN,D
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: GAYLE,OWEN
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 09:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: PINHO

DATE/TIME OF TEST: 10/04/2016 10:34:38

Diagnostics Check	OK	10:28
Air Blank	0.000	10:28
Control Test	0.081	10:29
Air Blank	0.000	10:29
Subject Sample # 1	0.063	10:30
Breath Volume	3.152	10:30
Air Blank	0.000	10:30
Air Blank	0.000	10:32
Subject Sample # 2	0.060	10:33
Breath Volume	3.125	10:33
Air Blank	0.000	10:33
Control Test	0.081	10:34
Air Blank	0.000	10:34
Diagnostics Check	OK	10:34

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,NEFTALI,F
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: LUGO,NEFTALI
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CASTILLA VARGAS

DATE/TIME OF TEST: 10/06/2016 01:59:49

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.188	01:55
Breath Volume	2.863	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.195	01:58
Breath Volume	2.855	01:58
Air Blank	0.000	01:58
Control Test	0.082	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WEST,BRIAN,J

BREATH TEST OPERATOR AGENCY: Kissimmee PD

ARRESTING OFFICER: WEST,BRIAN

ARRESTING OFFICER AGENCY: Kissimmee PD

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1578837

DRY GAS EXPIRY DATE: 10/16/2016

VIOLATION CODE: DUI

ARREST TIME: 06:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: MCNEAL

DATE/TIME OF TEST: 10/06/2016 07:27:59

Diagnostics Check	OK	07:21
Air Blank	0.000	07:21
Control Test	0.081	07:22
Air Blank	0.000	07:22
Subject Sample # 1	0.230	07:23
Breath Volume	2.500	07:23
Air Blank	0.000	07:23
Air Blank	0.000	07:25
Subject Sample # 2	0.216	07:26
Breath Volume	2.191	07:26
Air Blank	0.000	07:27
Control Test	0.081	07:27
Air Blank	0.000	07:27
Diagnostics Check	OK	07:27

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO, ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 10/07/2016 22:15:27

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.110	22:10
Breath Volume	2.289	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.115	22:13
Breath Volume	2.531	22:13
Air Blank	0.000	22:14
Control Test	0.082	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SILVA, EMILIANO, R

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: SILVA, EMILIANO

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1578837

DRY GAS EXPIRY DATE: 10/16/2016

VIOLATION CODE: DUI

ARREST TIME: 03:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 10/10/2016 04:42:59

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.082	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:36	No Sample Provided
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Air Blank	0.000	04:38	
Subject Sample	# 2	04:41	No Sample Provided
Breath Volume	0.000	04:41	
Air Blank	0.000	04:42	
Control Test	0.082	04:42	
Air Blank	0.000	04:42	
Diagnostics Check	OK	04:42	

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,JASON,R
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: COLLEY,JONATHAN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 15:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:55

SUBJECT'S LAST NAME: ALLRED

DATE/TIME OF TEST: 10/10/2016 17:34:52

Diagnostics Check	OK	17:28
Air Blank	0.000	17:29
Control Test	0.082	17:29
Air Blank	0.000	17:30
Subject Sample # 1	0.000	17:30
Breath Volume	2.398	17:30
Air Blank	0.000	17:31
Air Blank	0.000	17:33
Subject Sample # 2	0.000	17:33
Breath Volume	2.535	17:33
Air Blank	0.000	17:33
Control Test	0.083	17:34
Air Blank	0.000	17:34
Diagnostics Check	OK	17:34

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1578837
DRY GAS EXPIRY DATE: 10/16/2016
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SANTOS GOMEZ

DATE/TIME OF TEST: 10/14/2016 02:21:06

Diagnostics Check	OK	02:11
Air Blank	0.000	02:12
Control Test	0.082	02:12
Air Blank	0.000	02:13
Subject Sample # 1	0.134	02:15
Breath Volume	1.894	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.116	02:19
Breath Volume	1.945	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/14/2016 07:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLEWELLYN,A,
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: LLEWELLYN,A
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/18/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: GORCH

DATE/TIME OF TEST: 10/27/2016 15:43:52

Diagnostics Check	OK	15:37
Air Blank	0.000	15:38
Control Test	0.080	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.344	15:39
Breath Volume	2.628	15:39
Air Blank	0.000	15:39
Air Blank	0.000	15:41
Subject Sample # 2	0.345	15:42
Breath Volume	1.488	15:42
Air Blank	0.000	15:42
Control Test	0.079	15:43
Air Blank	0.000	15:43
Diagnostics Check	OK	15:43

Data Download Date/Time: 11/23/2016 13:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO, OSCAR, J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CABRANES, JORGE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: AYERS

DATE/TIME OF TEST: 10/04/2016 22:40:51

Diagnostics Check	OK	22:34
Air Blank	0.000	22:34
Control Test	0.078	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.130	22:36
Breath Volume	2.460	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.128	22:39
Breath Volume	3.187	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 10/26/2016 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLDEN,ROBERT,A
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SANTIAGO,FRANCES
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:10

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 10/05/2016 09:16:17

Diagnostics Check	OK	09:09
Air Blank	0.000	09:10
Control Test	0.079	09:10
Air Blank	0.000	09:11
Subject Sample # 1	0.256	09:11
Breath Volume	2.203	09:11
Air Blank	0.000	09:12
Air Blank	0.000	09:14
Subject Sample # 2	0.240	09:14
Breath Volume	2.062	09:14
Air Blank	0.000	09:15
Control Test	0.079	09:15
Air Blank	0.000	09:16
Diagnostics Check	OK	09:16

Data Download Date/Time: 10/26/2016 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEDY,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SANTIAGO,FRANCES
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: ZIEGLER

DATE/TIME OF TEST: 10/19/2016 22:21:10

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.078	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.247	22:16
Breath Volume	1.539	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.239	22:19
Breath Volume	1.632	22:19
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 10/26/2016 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO, OSCAR, J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CABRANES, JORGE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 10/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/27/2016 23:15:03

Diagnostics Check	OK	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.084	23:10
Breath Volume	1.796	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.084	23:13
Breath Volume	2.394	23:13
Air Blank	0.000	23:14
Control Test	0.079	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:15

Data Download Date/Time: 11/30/2016 03:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 10/26/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: BRITO

DATE/TIME OF TEST: 10/29/2016 04:42:20

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.078	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.228	04:37
Breath Volume	1.664	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.226	04:40
Breath Volume	2.007	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 11/30/2016 03:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCGRORY,DIARMUID,J

BREATH TEST OPERATOR AGENCY: Titusville PD

ARRESTING OFFICER: MCGRORY,DIARMUID

ARRESTING OFFICER AGENCY: Titusville PD

LAST AGENCY INSPECTION: 10/26/2016

Date or Time Changed: No

DRY GAS STANDARD LOT #: 01316080A2

DRY GAS EXPIRY DATE: 02/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:21

GENDER: Female

OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 10/29/2016 20:48:37

Diagnostics Check	OK	20:42
Air Blank	0.000	20:42
Control Test	0.079	20:43
Air Blank	0.000	20:43
Subject Sample # 1	0.077	20:44
Breath Volume	2.378	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.072	20:47
Breath Volume	1.988	20:47
Air Blank	0.000	20:47
Control Test	0.078	20:48
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

Data Download Date/Time: 11/30/2016 03:03