Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOROTIZA,JESUS,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: GOROTIZA,JESUS
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HERNANDEZ ENRIQUEZ

DATE/TIME OF TEST: 01/04/2017 02:32:14

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.158	02:26
Breath Volume		2.121	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.177	02:30
Breath Volume		1.468	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 01/26/2017 13:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOROTIZA,JESUS,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: GOROTIZA,JESUS
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: CANTERO HARO

DATE/TIME OF TEST: 01/06/2017 22:34:39

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.078	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.185	22:29
Breath Volume		1.683	22:29
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.181	22:33
Breath Volume		2.121	22:33
Air Blank		0.000	22:33
Control Test		0.076	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 01/26/2017 13:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOROTIZA,JESUS,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: GOROTIZA,JESUS
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: NAPOLES CONSUEGRA

DATE/TIME OF TEST: 01/07/2017 21:51:32

Diagnostics Check		OK	21:47	
Air Blank		0.000	21:48	
Control Test		0.079	21:48	
Air Blank		0.000	21:48	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:50	
Control Test		0.079	21:50	
Air Blank		0.000	21:51	
Diagnostics Check		OK	21:51	

Data Download Date/Time: 01/26/2017 13:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOROTIZA,JESUS,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: GOROTIZA,JESUS
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: TRIANA ESPINOSA

DATE/TIME OF TEST: 01/07/2017 23:35:17

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.079	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:34	
Control Test		0.078	23:34	
Air Blank		0.000	23:35	
Diagnostics Check		OK	23:35	

Data Download Date/Time: 01/26/2017 13:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUERTA,EDWIN,M
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: HUERTA,EDWIN
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WILLIAMS BOYCE

DATE/TIME OF TEST: 01/09/2017 04:43:54

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:41	
Control Test		0.078	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:42	Subject Test Refused
Breath Volume		0.000	04:42	
Air Blank		0.000	04:42	
Control Test		0.078	04:43	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 01/26/2017 13:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOROTIZA,JESUS,
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: GOROTIZA,JESUS
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: CAMACHO HERNANDEZ

DATE/TIME OF TEST: 01/11/2017 23:15:09

Diagnostics Check		OK	23:02
Air Blank		0.000	23:03
Control Test		0.078	23:03
Air Blank		0.000	23:04
Subject Sample	# 1	0.252	23:05
Breath Volume		2.203	23:05
Air Blank		0.000	23:06
Air Blank		0.000	23:08
Subject Sample	# 2	0.230	23:09
Breath Volume		1.859	23:09
Air Blank		0.000	23:09
Air Blank		0.000	23:11
Subject Sample	# 3	0.232	23:13
Breath Volume		1.671	23:13
Air Blank		0.000	23:14
Control Test		0.079	23:14
Air Blank		0.000	23:15
Diagnostics Check		OK	23:15

Data Download Date/Time: 01/26/2017 13:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001275

INSTRUMENT REGISTERED TO: HIALEAH GARDENS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CISAR,DANIEL,G
BREATH TEST OPERATOR AGENCY: Hialeah Gardens PD
ARRESTING OFFICER: CISAR,DANIEL
ARRESTING OFFICER AGENCY: Hialeah Gardens PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PEDROSO RIVERO

DATE/TIME OF TEST: 01/26/2017 01:15:23

Diagnostics Check		OK	01:12	
Air Blank		0.000	01:12	
Control Test		0.078	01:13	
Air Blank		0.000	01:13	
Subject Sample	# 1		01:14	Subject Test Refused
Breath Volume		0.000	01:14	
Air Blank		0.000	01:14	
Control Test		0.078	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 01/26/2017 13:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 01/22/2017 01:29:25

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.075	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.093	01:24
Breath Volume		1.960	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.086	01:27
Breath Volume		1.167	01:27
Air Blank		0.000	01:28
Control Test		0.078	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 02/14/2017 13:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEUDONNE,K,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RODRIGUEZ,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 01/26/2017 21:25:59

Diagnostics Check		OK	21:23	
Air Blank		0.000	21:23	
Control Test		0.080	21:23	
Air Blank		0.000	21:24	
Subject Sample	# 1		21:24	Subject Test Refused
Breath Volume		0.000	21:24	
Air Blank		0.000	21:25	
Control Test		0.079	21:25	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 02/14/2017 13:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001276

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GUILLAUME

DATE/TIME OF TEST: 01/29/2017 01:23:43

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.080	01:21	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.079	01:23	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 02/14/2017 13:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUETT,RICHARD,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: FODEN

DATE/TIME OF TEST: 01/01/2017 23:44:10

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:41	
Control Test		0.080	23:42	
Air Blank		0.000	23:42	
Subject Sample	# 1		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:43	
Control Test		0.080	23:43	
Air Blank		0.000	23:44	
Diagnostics Check		OK	23:44	

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 01/03/2017 20:41:48

Diagnostics Check		OK	20:35
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:36
Subject Sample	# 1	0.124	20:37
Breath Volume		4.992	20:37
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.121	20:40
Breath Volume		5.253	20:40
Air Blank		0.000	20:40
Control Test		0.080	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMTOHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C.E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BUSTOS

DATE/TIME OF TEST: 01/04/2017 03:12:33

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.000	03:08
Breath Volume		2.144	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.000	03:11
Breath Volume		1.433	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ARKIN,P
ARRESTING OFFICER AGENCY: FFWCC (Lake City)
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: GRAEME

DATE/TIME OF TEST: 01/05/2017 19:00:31

Diagnostics Check		OK	18:53
Air Blank		0.000	18:53
Control Test		0.080	18:54
Air Blank		0.000	18:54
Subject Sample	# 1	0.275	18:55
Breath Volume		3.183	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:58
Subject Sample	# 2	0.268	18:58
Breath Volume		3.125	18:58
Air Blank		0.000	18:59
Control Test		0.080	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: BOBBIN

DATE/TIME OF TEST: 01/06/2017 02:09:53

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.279	02:04
Breath Volume		2.007	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.282	02:08
Breath Volume		3.480	02:08
Air Blank		0.000	02:08
Control Test		0.082	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ABBOTT

DATE/TIME OF TEST: 01/06/2017 04:27:55

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:25	
Control Test		0.080	04:25	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:26	Subject Test Refused
Breath Volume		0.000	04:26	
Air Blank		0.000	04:27	
Control Test		0.080	04:27	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BEGLEY

DATE/TIME OF TEST: 01/07/2017 03:04:38

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.080	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:03	
Control Test		0.080	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GUERRERO LIEVA

DATE/TIME OF TEST: 01/07/2017 03:50:16

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.080	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.096	03:45
Breath Volume		3.203	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.090	03:48
Breath Volume		2.054	03:48
Air Blank		0.000	03:49
Control Test		0.080	03:49
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSEL,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 01/07/2017 04:36:40

Diagnostics Check		OK	04:29
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Subject Sample	# 1	0.237	04:32
Breath Volume		1.714	04:32
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.239	04:35
Breath Volume		1.621	04:35
Air Blank		0.000	04:35
Control Test		0.079	04:36
Air Blank		0.000	04:36
Diagnostics Check		OK	04:36

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: VARNADORE

DATE/TIME OF TEST: 01/10/2017 03:18:50

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:17	
Control Test		0.081	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHRISTOFFERSON,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: STEELMAN

DATE/TIME OF TEST: 01/11/2017 05:01:06

Diagnostics Check		OK	04:58	
Air Blank		0.000	04:58	
Control Test		0.081	04:59	
Air Blank		0.000	04:59	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.000	04:59	
Air Blank		0.000	05:00	
Control Test		0.080	05:00	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 01/12/2017 11:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAZZO,P
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: O SHIELDS

DATE/TIME OF TEST: 01/12/2017 18:18:31

Diagnostics Check		OK	18:06	
Air Blank		0.000	18:07	
Control Test		0.080	18:07	
Air Blank		0.000	18:08	
Subject Sample	# 1	0.185	18:11	Volume Not Met
Breath Volume		0.613	18:11	
Air Blank		0.000	18:12	
Air Blank		0.000	18:13	
Subject Sample	# 2	0.165	18:16	Volume Not Met
Breath Volume		0.351	18:16	
Air Blank		0.000	18:17	
Control Test		0.078	18:17	
Air Blank		0.000	18:18	
Diagnostics Check		OK	18:18	

Data Download Date/Time: 01/27/2017 11:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/12/2017 21:57:45

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:55	
Control Test		0.080	21:55	
Air Blank		0.000	21:56	
Subject Sample	# 1		21:56	Subject Test Refused
Breath Volume		0.000	21:56	
Air Blank		0.000	21:56	
Control Test		0.079	21:57	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 01/27/2017 11:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAI,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: LINDSEY

DATE/TIME OF TEST: 01/13/2017 01:38:35

Diagnostics Check		OK	01:32
Air Blank		0.000	01:32
Control Test		0.080	01:33
Air Blank		0.000	01:33
Subject Sample	# 1	0.231	01:34
Breath Volume		1.796	01:34
Air Blank		0.000	01:34
Air Blank		0.000	01:36
Subject Sample	# 2	0.215	01:37
Breath Volume		1.152	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Diagnostics Check		OK	01:38

Data Download Date/Time: 01/27/2017 11:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADGETT,C,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,P
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 01/14/2017 02:38:34

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.110	02:34
Breath Volume		1.886	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.108	02:37
Breath Volume		1.949	02:37
Air Blank		0.000	02:37
Control Test		0.079	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/27/2017 11:15

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADGETT,C,N
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BRADY

DATE/TIME OF TEST: 01/14/2017 03:58:06

Diagnostics Check		OK	03:51
Air Blank		0.000	03:51
Control Test		0.080	03:52
Air Blank		0.000	03:52
Subject Sample	# 1	0.157	03:53
Breath Volume		2.750	03:53
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.155	03:56
Breath Volume		2.574	03:56
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 01/15/2017 02:08:14

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.208	02:03
Breath Volume		1.875	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.214	02:06
Breath Volume		2.015	02:06
Air Blank		0.000	02:07
Control Test		0.080	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WILLIAMS,D
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: PAYRET

DATE/TIME OF TEST: 01/16/2017 01:28:31

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.081	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: FLOREZ

DATE/TIME OF TEST: 01/16/2017 04:23:23

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.080	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.178	04:19
Breath Volume		1.308	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.176	04:21
Breath Volume		1.656	04:21
Air Blank		0.000	04:22
Control Test		0.076	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S.B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 01/17/2017 04:07:36

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.160	04:02
Breath Volume		2.863	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.160	04:06
Breath Volume		2.929	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 01/17/2017 23:48:38

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.204	23:43
Breath Volume		2.238	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.209	23:47
Breath Volume		1.964	23:47
Air Blank		0.000	23:47
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/19/2017 02:06:34

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.000	02:02
Breath Volume		3.335	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.000	02:05
Breath Volume		1.238	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GUILLROY

DATE/TIME OF TEST: 01/22/2017 03:46:29

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.079	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.151	03:42
Breath Volume		1.421	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.152	03:44
Breath Volume		1.503	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 01/27/2017 11:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: SELLS

DATE/TIME OF TEST: 01/27/2017 03:22:01

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.016	03:17
Breath Volume		1.632	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.017	03:20
Breath Volume		1.585	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:22

Data Download Date/Time: 01/27/2017 11:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BAGWELL,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: MORGAN

DATE/TIME OF TEST: 01/27/2017 18:25:15

Diagnostics Check		OK	18:22	
Air Blank		0.000	18:22	
Control Test		0.081	18:23	
Air Blank		0.000	18:23	
Subject Sample	# 1		18:23	Subject Test Refused
Breath Volume		0.000	18:23	
Air Blank		0.000	18:24	
Control Test		0.080	18:24	
Air Blank		0.000	18:25	
Diagnostics Check		OK	18:25	

Data Download Date/Time: 02/15/2017 10:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORRISON,M
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DIZE

DATE/TIME OF TEST: 01/28/2017 02:54:10

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.081	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.186	02:49
Breath Volume		5.437	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.185	02:52
Breath Volume		4.765	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 02/15/2017 10:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: GASTON

DATE/TIME OF TEST: 01/28/2017 04:06:27

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.081	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.167	04:01
Breath Volume		2.410	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:03
Subject Sample	# 2	0.162	04:04
Breath Volume		2.394	04:04
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 01/28/2017 23:17:29

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.080	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.151	23:13
Breath Volume		3.171	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.150	23:15
Breath Volume		3.148	23:15
Air Blank		0.000	23:16
Control Test		0.081	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: SPILKER

DATE/TIME OF TEST: 01/29/2017 01:03:51

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.225	00:59
Breath Volume		2.613	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.220	01:02
Breath Volume		2.781	01:02
Air Blank		0.000	01:02
Control Test		0.080	01:03
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LOSCO

DATE/TIME OF TEST: 01/29/2017 02:38:42

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.178	02:33
Breath Volume		1.246	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.185	02:37
Breath Volume		1.546	02:37
Air Blank		0.000	02:37
Control Test		0.081	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,EUGENE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 01/29/2017 18:56:47

Diagnostics Check		OK	18:50
Air Blank		0.000	18:51
Control Test		0.081	18:51
Air Blank		0.000	18:51
Subject Sample	# 1	0.000	18:52
Breath Volume		1.308	18:52
Air Blank		0.000	18:52
Air Blank		0.000	18:54
Subject Sample	# 2	0.000	18:55
Breath Volume		1.863	18:55
Air Blank		0.000	18:55
Control Test		0.081	18:56
Air Blank		0.000	18:56
Diagnostics Check		OK	18:56

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,EUGENE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SAMUELS,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 01/29/2017 21:40:08

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:37	
Control Test		0.081	21:37	
Air Blank		0.000	21:38	
Subject Sample	# 1		21:38	Subject Test Refused
Breath Volume		0.000	21:38	
Air Blank		0.000	21:39	
Control Test		0.080	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 01/30/2017 02:52:16

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.053	02:47
Breath Volume		1.597	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.052	02:50
Breath Volume		1.726	02:50
Air Blank		0.000	02:51
Control Test		0.081	02:51
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,EUGENE,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/30/2017 22:14:39

Diagnostics Check		OK	22:08
Air Blank		0.000	22:09
Control Test		0.081	22:09
Air Blank		0.000	22:09
Subject Sample	# 1	0.208	22:10
Breath Volume		3.214	22:10
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 2	0.206	22:13
Breath Volume		3.546	22:13
Air Blank		0.000	22:13
Control Test		0.079	22:14
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MIRELES

DATE/TIME OF TEST: 01/31/2017 03:43:55

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.081	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.225	03:35	
Breath Volume		1.351	03:35	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2	0.188	03:39	
Breath Volume		2.550	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:42	
Subject Sample	# 3	0.213	03:42	
Breath Volume		3.769	03:42	
Air Blank		0.000	03:43	
Control Test		0.000	03:43	RFI Detect
Air Blank		0.000	03:43	

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001278

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY,SHANON,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMCHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MIRELES

DATE/TIME OF TEST: 01/31/2017 03:51:41

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.232	03:47
Breath Volume		3.097	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.227	03:50
Breath Volume		2.667	03:50
Air Blank		0.000	03:50
Control Test		0.081	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 02/15/2017 10:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARTILUCCI,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: MILYANOVICH

DATE/TIME OF TEST: 01/09/2017 02:55:50

Diagnostics Check		OK	02:52	
Air Blank		0.000	02:53	
Control Test		0.081	02:53	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:54	Subject Test Refused
Breath Volume		0.000	02:54	
Air Blank		0.000	02:54	
Control Test		0.081	02:55	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 01/12/2017 10:57

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/09/2017 05:02:42

Diagnostics Check		OK	04:55
Air Blank		0.000	04:56
Control Test		0.081	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.062	04:58
Breath Volume		3.468	04:58
Air Blank		0.000	04:58
Air Blank		0.000	05:00
Subject Sample	# 2	0.061	05:01
Breath Volume		2.859	05:01
Air Blank		0.000	05:01
Control Test		0.081	05:02
Air Blank		0.000	05:02
Diagnostics Check		OK	05:02

Data Download Date/Time: 01/12/2017 10:57

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: ADDAIR

DATE/TIME OF TEST: 01/10/2017 03:51:38

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:49	
Control Test		0.081	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:50	Subject Test Refused
Breath Volume		0.000	03:50	
Air Blank		0.000	03:50	
Control Test		0.081	03:51	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 01/12/2017 10:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: GOOGE

DATE/TIME OF TEST: 01/13/2017 04:30:31

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:28	
Control Test		0.080	04:28	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:29	
Control Test		0.080	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/15/2017 07:18:57

Diagnostics Check		OK	07:11
Air Blank		0.000	07:12
Control Test		0.079	07:12
Air Blank		0.000	07:12
Subject Sample	# 1	0.132	07:14
Breath Volume		3.437	07:14
Air Blank		0.000	07:14
Air Blank		0.000	07:16
Subject Sample	# 2	0.127	07:17
Breath Volume		3.710	07:17
Air Blank		0.000	07:18
Control Test		0.079	07:18
Air Blank		0.000	07:18
Diagnostics Check		OK	07:18

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: LACERDA

DATE/TIME OF TEST: 01/15/2017 20:08:06

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.080	20:02
Air Blank		0.000	20:03
Subject Sample	# 1	0.232	20:03
Breath Volume		3.046	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:05
Subject Sample	# 2	0.230	20:06
Breath Volume		2.664	20:06
Air Blank		0.000	20:07
Control Test		0.080	20:07
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: GILBERT

DATE/TIME OF TEST: 01/15/2017 22:04:56

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.079	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.135	22:00
Breath Volume		2.238	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.131	22:03
Breath Volume		3.203	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,ERICKA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GRIFFIN,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:06

SUBJECT'S LAST NAME: JANICKI

DATE/TIME OF TEST: 01/17/2017 11:34:26

Diagnostics Check		OK	11:26
Air Blank		0.000	11:27
Control Test		0.080	11:27
Air Blank		0.000	11:28
Subject Sample	# 1	0.164	11:29
Breath Volume		3.839	11:29
Air Blank		0.000	11:30
Air Blank		0.000	11:32
Subject Sample	# 2	0.161	11:32
Breath Volume		3.183	11:32
Air Blank		0.000	11:33
Control Test		0.079	11:33
Air Blank		0.000	11:34
Diagnostics Check		OK	11:34

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JAMES,ALFONSO,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOWRD,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:55

SUBJECT'S LAST NAME: LOVETT

DATE/TIME OF TEST: 01/20/2017 14:21:27

Diagnostics Check		OK	14:15
Air Blank		0.000	14:15
Control Test		0.080	14:16
Air Blank		0.000	14:16
Subject Sample	# 1	0.000	14:17
Breath Volume		3.285	14:17
Air Blank		0.000	14:17
Air Blank		0.000	14:19
Subject Sample	# 2	0.000	14:20
Breath Volume		2.687	14:20
Air Blank		0.000	14:20
Control Test		0.080	14:20
Air Blank		0.000	14:21
Diagnostics Check		OK	14:21

Data Download Date/Time: 01/27/2017 11:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 01/24/2017 20:33:32

Diagnostics Check		OK	20:30	
Air Blank		0.000	20:30	
Control Test		0.080	20:31	
Air Blank		0.000	20:31	
Subject Sample	# 1		20:32	Subject Test Refused
Breath Volume		0.000	20:32	
Air Blank		0.000	20:32	
Control Test		0.081	20:32	
Air Blank		0.000	20:33	
Diagnostics Check		OK	20:33	

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: UNDA

DATE/TIME OF TEST: 01/24/2017 22:19:25

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:17	
Control Test		0.080	22:17	
Air Blank		0.000	22:17	
Subject Sample	# 1		22:18	Subject Test Refused
Breath Volume		0.000	22:18	
Air Blank		0.000	22:18	
Control Test		0.080	22:18	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAY,DANIEL,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: OVERFIELD

DATE/TIME OF TEST: 01/26/2017 22:49:12

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.190	22:43
Breath Volume		1.710	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.196	22:47
Breath Volume		1.929	22:47
Air Blank		0.000	22:48
Control Test		0.079	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 01/27/2017 02:17:01

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:11
Subject Sample	# 1	0.084	02:12
Breath Volume		3.246	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.073	02:15
Breath Volume		1.156	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/27/2017 11:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: PUGH

DATE/TIME OF TEST: 01/27/2017 03:48:19

Diagnostics Check		OK	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:42
Subject Sample	# 1	0.157	03:43
Breath Volume		1.300	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.166	03:46
Breath Volume		2.039	03:46
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 01/27/2017 11:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HUME

DATE/TIME OF TEST: 01/28/2017 23:42:31

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.125	23:37
Breath Volume		3.445	23:37
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.131	23:41
Breath Volume		4.109	23:41
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 02/15/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: HUANG

DATE/TIME OF TEST: 01/29/2017 01:55:25

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.081	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.032	01:50
Breath Volume		1.605	01:50
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.027	01:53
Breath Volume		1.316	01:53
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 02/15/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: ACOSTA DIAZ

DATE/TIME OF TEST: 01/29/2017 05:25:57

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.081	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.074	05:21
Breath Volume		2.714	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.070	05:24
Breath Volume		2.492	05:24
Air Blank		0.000	05:25
Control Test		0.081	05:25
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 02/15/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: HORVATH

DATE/TIME OF TEST: 01/30/2017 00:12:33

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.081	00:01	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.000	00:05	Volume Not Met
Breath Volume		0.132	00:05	
Air Blank		0.000	00:06	
Air Blank		0.000	00:08	
Subject Sample	# 2		00:11	No Sample Provided
Breath Volume		0.000	00:11	
Air Blank		0.000	00:11	
Control Test		0.080	00:11	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 02/15/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/30/2017 04:44:04

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:31	
Control Test		0.081	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.136	04:33	
Breath Volume		3.050	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2	0.091	04:39	Volume Not Met
Breath Volume		0.753	04:39	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 3	0.127	04:42	
Breath Volume		3.359	04:42	
Air Blank		0.000	04:43	
Control Test		0.080	04:43	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 02/15/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOLLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: POULOT CASTELLANOS

DATE/TIME OF TEST: 01/31/2017 03:39:48

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.049	03:35
Breath Volume		3.109	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.047	03:38
Breath Volume		2.890	03:38
Air Blank		0.000	03:38
Control Test		0.081	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/15/2017 09:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: ROEDIGER

DATE/TIME OF TEST: 01/04/2017 00:59:01

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.155	00:54
Breath Volume		2.273	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.157	00:57
Breath Volume		1.687	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:59

Data Download Date/Time: 01/12/2017 10:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JARRELL,K
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/05/2017 04:01:39

Diagnostics Check		OK	03:58	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1		04:00	Subject Test Refused
Breath Volume		0.000	04:00	
Air Blank		0.000	04:00	
Control Test		0.080	04:01	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 01/12/2017 10:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERK,CODY,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JOHNSON,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/11/2017 13:16:46

Diagnostics Check		OK	13:10
Air Blank		0.000	13:11
Control Test		0.079	13:11
Air Blank		0.000	13:11
Subject Sample	# 1	0.119	13:12
Breath Volume		3.625	13:12
Air Blank		0.000	13:12
Air Blank		0.000	13:14
Subject Sample	# 2	0.125	13:15
Breath Volume		3.976	13:15
Air Blank		0.000	13:15
Control Test		0.079	13:16
Air Blank		0.000	13:16
Diagnostics Check		OK	13:16

Data Download Date/Time: 01/12/2017 10:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOENIG,DUSTIN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 01/11/2017 21:13:10

Diagnostics Check		OK	21:06
Air Blank		0.000	21:07
Control Test		0.080	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.000	21:08
Breath Volume		1.671	21:08
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.000	21:11
Breath Volume		1.859	21:11
Air Blank		0.000	21:12
Control Test		0.080	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 01/12/2017 10:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MCDOUGAL

DATE/TIME OF TEST: 01/12/2017 02:09:33

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.080	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1		02:08	Subject Test Refused
Breath Volume		0.000	02:08	
Air Blank		0.000	02:08	
Control Test		0.080	02:08	
Air Blank		0.000	02:09	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 01/12/2017 10:50

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'STEEN,TIMOTHY,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BARNES,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:07

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/14/2017 07:38:18

Diagnostics Check		OK	07:31
Air Blank		0.000	07:32
Control Test		0.080	07:32
Air Blank		0.000	07:33
Subject Sample	# 1	0.000	07:33
Breath Volume		3.246	07:33
Air Blank		0.000	07:34
Air Blank		0.000	07:36
Subject Sample	# 2	0.000	07:36
Breath Volume		3.144	07:36
Air Blank		0.000	07:37
Control Test		0.080	07:37
Air Blank		0.000	07:38
Diagnostics Check		OK	07:38

Data Download Date/Time: 01/27/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/15/2017 03:12:29

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.079	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:11	Subject Test Refused
Breath Volume		0.000	03:11	
Air Blank		0.000	03:11	
Control Test		0.080	03:11	
Air Blank		0.000	03:12	
Diagnostics Check		OK	03:12	

Data Download Date/Time: 01/27/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 01/15/2017 04:47:40

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.188	04:43
Breath Volume		1.550	04:43
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.179	04:46
Breath Volume		1.136	04:46
Air Blank		0.000	04:46
Control Test		0.079	04:47
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 01/27/2017 11:05

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 01/16/2017 00:11:31

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.079	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1		00:10	Subject Test Refused
Breath Volume		0.000	00:10	
Air Blank		0.000	00:10	
Control Test		0.079	00:10	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 01/27/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASHER,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 01/21/2017 01:17:43

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.188	01:13
Breath Volume		2.851	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.184	01:16
Breath Volume		2.531	01:16
Air Blank		0.000	01:16
Control Test		0.077	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 01/27/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'STEEN JR,TIMOTHY,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:08

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 01/22/2017 07:40:46

Diagnostics Check		OK	07:34
Air Blank		0.000	07:34
Control Test		0.079	07:34
Air Blank		0.000	07:35
Subject Sample	# 1	0.000	07:36
Breath Volume		1.367	07:36
Air Blank		0.000	07:36
Air Blank		0.000	07:38
Subject Sample	# 2	0.000	07:39
Breath Volume		1.949	07:39
Air Blank		0.000	07:39
Control Test		0.080	07:40
Air Blank		0.000	07:40
Diagnostics Check		OK	07:40

Data Download Date/Time: 01/27/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIS,M,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUPUY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:28

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 01/30/2017 12:55:39

Diagnostics Check		OK	12:49
Air Blank		0.000	12:49
Control Test		0.081	12:50
Air Blank		0.000	12:50
Subject Sample	# 1	0.272	12:51
Breath Volume		2.921	12:51
Air Blank		0.000	12:51
Air Blank		0.000	12:53
Subject Sample	# 2	0.264	12:53
Breath Volume		2.628	12:53
Air Blank		0.000	12:54
Control Test		0.080	12:55
Air Blank		0.000	12:55
Diagnostics Check		OK	12:55

Data Download Date/Time: 02/15/2017 09:52

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIMON,SHANIESA,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: WALTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 01/03/2017 03:53:13

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:50	
Control Test		0.082	03:51	
Air Blank		0.000	03:51	
Subject Sample	# 1		03:51	Subject Test Refused
Breath Volume		0.000	03:51	
Air Blank		0.000	03:52	
Control Test		0.082	03:52	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,P,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALTON,C
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 01/03/2017 06:53:34

Diagnostics Check		OK	06:50	
Air Blank		0.000	06:51	
Control Test		0.081	06:51	
Air Blank		0.000	06:51	
Subject Sample	# 1		06:52	Subject Test Refused
Breath Volume		0.000	06:52	
Air Blank		0.000	06:52	
Control Test		0.082	06:52	
Air Blank		0.000	06:53	
Diagnostics Check		OK	06:53	

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,G,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELACRUZ,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: COBOURN

DATE/TIME OF TEST: 01/04/2017 00:36:17

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.082	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.164	00:31
Breath Volume		1.773	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.157	00:34
Breath Volume		3.054	00:34
Air Blank		0.000	00:35
Control Test		0.081	00:35
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FEDDERLY,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ENDICOTT

DATE/TIME OF TEST: 01/05/2017 01:39:59

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:36	
Control Test		0.082	01:36	
Air Blank		0.000	01:37	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Control Test		0.082	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,P,S
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: TURMAN,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:21

SUBJECT'S LAST NAME: ANSPACH

DATE/TIME OF TEST: 01/06/2017 17:48:26

Diagnostics Check		OK	17:41
Air Blank		0.000	17:42
Control Test		0.082	17:42
Air Blank		0.000	17:42
Subject Sample	# 1	0.287	17:44
Breath Volume		2.089	17:44
Air Blank		0.000	17:44
Air Blank		0.000	17:46
Subject Sample	# 2	0.287	17:46
Breath Volume		2.707	17:46
Air Blank		0.000	17:47
Control Test		0.079	17:47
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,G,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUBBARD,B
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: OTHER
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 01/06/2017 19:05:15

Diagnostics Check		OK	18:58
Air Blank		0.000	18:59
Control Test		0.081	18:59
Air Blank		0.000	19:00
Subject Sample	# 1	0.244	19:00
Breath Volume		1.527	19:00
Air Blank		0.000	19:01
Air Blank		0.000	19:03
Subject Sample	# 2	0.244	19:03
Breath Volume		1.824	19:03
Air Blank		0.000	19:04
Control Test		0.081	19:04
Air Blank		0.000	19:05
Diagnostics Check		OK	19:05

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIMON,SHANIESA,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PALECEK,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/12/2017 20:16:58

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.083	20:14	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:15	Subject Test Refused
Breath Volume		0.000	20:15	
Air Blank		0.000	20:16	
Control Test		0.083	20:16	
Air Blank		0.000	20:16	
Diagnostics Check		OK	20:16	

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 01/13/2017 02:08:15

Diagnostics Check		OK	02:01
Air Blank		0.000	02:02
Control Test		0.082	02:02
Air Blank		0.000	02:03
Subject Sample	# 1	0.101	02:03
Breath Volume		5.671	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.102	02:06
Breath Volume		6.019	02:06
Air Blank		0.000	02:07
Control Test		0.081	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: TALBIRD

DATE/TIME OF TEST: 01/14/2017 01:28:32

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.147	01:24
Breath Volume		2.648	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.149	01:27
Breath Volume		2.449	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JONES,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 01/14/2017 02:18:26

Diagnostics Check		OK	02:10
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.179	02:13
Breath Volume		1.308	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.179	02:16
Breath Volume		1.382	02:16
Air Blank		0.000	02:17
Control Test		0.082	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: FEDDERLY,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:29

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/15/2017 19:55:28

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:52	
Control Test		0.082	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 1		19:54	Subject Test Refused
Breath Volume		0.000	19:54	
Air Blank		0.000	19:54	
Control Test		0.082	19:54	
Air Blank		0.000	19:55	
Diagnostics Check		OK	19:55	

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,G,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: GALLION

DATE/TIME OF TEST: 01/18/2017 01:06:32

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.082	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.072	01:02
Breath Volume		3.789	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.070	01:05
Breath Volume		3.574	01:05
Air Blank		0.000	01:05
Control Test		0.082	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ANNO,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: FERRELL

DATE/TIME OF TEST: 01/18/2017 18:48:53

Diagnostics Check		OK	18:46	
Air Blank		0.000	18:46	
Control Test		0.082	18:46	
Air Blank		0.000	18:47	
Subject Sample	# 1		18:47	Subject Test Refused
Breath Volume		0.000	18:47	
Air Blank		0.000	18:47	
Control Test		0.082	18:48	
Air Blank		0.000	18:48	
Diagnostics Check		OK	18:48	

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIRSK,KEVIN,M
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 01/18/2017 23:07:40

Diagnostics Check		OK	23:02	
Air Blank		0.000	23:03	
Control Test		0.082	23:03	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:06	Subject Test Refused
Breath Volume		0.363	23:06	
Air Blank		0.000	23:06	
Control Test		0.083	23:07	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001281

INSTRUMENT REGISTERED TO: NASSAU COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: BALLARD,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: MARETT

DATE/TIME OF TEST: 01/19/2017 02:23:59

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.082	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.152	02:19
Breath Volume		2.390	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.135	02:22
Breath Volume		1.140	02:22
Air Blank		0.000	02:23
Control Test		0.081	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 01/23/2017 09:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: TRUJILLO

DATE/TIME OF TEST: 01/05/2017 03:56:27

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:48	
Control Test		0.077	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:52	No Sample Provided
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Air Blank		0.000	03:54	
Subject Sample	# 2		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.077	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: MEDINA OLSES

DATE/TIME OF TEST: 01/06/2017 03:05:06

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.078	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.125	03:00
Breath Volume		2.386	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.116	03:03
Breath Volume		2.101	03:03
Air Blank		0.000	03:04
Control Test		0.077	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLWAY,CHAD,S
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: JORDAN,ELIZABETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: LINTON

DATE/TIME OF TEST: 01/07/2017 00:23:41

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.078	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.174	00:18
Breath Volume		2.718	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.160	00:22
Breath Volume		1.179	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BECKHAM,MARCUS
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RADULESCU

DATE/TIME OF TEST: 01/15/2017 03:04:21

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.078	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.149	02:59
Breath Volume		1.121	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.151	03:02
Breath Volume		1.390	03:02
Air Blank		0.000	03:03
Control Test		0.078	03:03
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: BRIDWELL

DATE/TIME OF TEST: 01/19/2017 03:27:27

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.134	03:22
Breath Volume		1.726	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.132	03:25
Breath Volume		1.812	03:25
Air Blank		0.000	03:26
Control Test		0.077	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BECKHAM,MARCUS
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MARVAL MORENO

DATE/TIME OF TEST: 01/20/2017 04:18:15

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.077	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.000	04:14
Breath Volume		2.171	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.000	04:16
Breath Volume		1.441	04:16
Air Blank		0.000	04:17
Control Test		0.078	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: STEWART,DILLON
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: KALAP

DATE/TIME OF TEST: 01/21/2017 02:58:45

Diagnostics Check		OK	02:49	
Air Blank		0.000	02:50	
Control Test		0.075	02:50	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:54	No Sample Provided
Breath Volume		0.000	02:54	
Air Blank		0.000	02:54	
Air Blank		0.000	02:56	
Subject Sample	# 2		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.076	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: HOLWAY,CHAD
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: LATNER

DATE/TIME OF TEST: 01/21/2017 03:11:22

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:03	
Control Test		0.076	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:07	No Sample Provided
Breath Volume		0.000	03:07	
Air Blank		0.000	03:07	
Air Blank		0.000	03:09	
Subject Sample	# 2		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.077	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BIVENS

DATE/TIME OF TEST: 01/24/2017 23:17:24

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.076	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.173	23:12
Breath Volume		1.195	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.177	23:15
Breath Volume		1.617	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/25/2017 03:21:55

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.078	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.103	03:14	Volume Not Met
Breath Volume		0.183	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.116	03:20	Volume Not Met
Breath Volume		0.609	03:20	
Air Blank		0.000	03:20	
Control Test		0.078	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,GEOFFREY,A
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of Fl PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/25/2017 03:32:14

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.077	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.122	03:26
Breath Volume		2.347	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.108	03:30
Breath Volume		1.957	03:30
Air Blank		0.000	03:31
Control Test		0.075	03:31
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 02/02/2017 04:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 01/01/2017 03:06:12

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.079	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.079	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 01/30/2017 00:45

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THWEATT,CAMERON,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THWEATT,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/30/2017 07:44:20

Diagnostics Check		OK	07:37
Air Blank		0.000	07:38
Control Test		0.079	07:38
Air Blank		0.000	07:39
Subject Sample	# 1	0.115	07:39
Breath Volume		2.527	07:39
Air Blank		0.000	07:40
Air Blank		0.000	07:42
Subject Sample	# 2	0.109	07:42
Breath Volume		2.171	07:42
Air Blank		0.000	07:43
Control Test		0.079	07:43
Air Blank		0.000	07:44
Diagnostics Check		OK	07:44

Data Download Date/Time: 02/27/2017 01:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ROBINSON,KYLE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 01/09/2017 18:03:32

Diagnostics Check		OK	17:57
Air Blank		0.000	17:57
Control Test		0.082	17:58
Air Blank		0.000	17:58
Subject Sample	# 1	0.132	17:59
Breath Volume		1.410	17:59
Air Blank		0.000	17:59
Air Blank		0.000	18:01
Subject Sample	# 2	0.130	18:01
Breath Volume		2.988	18:01
Air Blank		0.000	18:02
Control Test		0.082	18:02
Air Blank		0.000	18:03
Diagnostics Check		OK	18:03

Data Download Date/Time: 01/31/2017 19:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,COREY,S
BREATH TEST OPERATOR AGENCY: Baker County SO
ARRESTING OFFICER: HEATH,MARC
ARRESTING OFFICER AGENCY: Baker County SO
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 01/17/2017 22:02:38

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.081	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.165	21:58
Breath Volume		2.070	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.172	22:01
Breath Volume		3.105	22:01
Air Blank		0.000	22:01
Control Test		0.081	22:02
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 01/31/2017 19:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:51

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/19/2017 20:17:25

Diagnostics Check		OK	20:11
Air Blank		0.000	20:11
Control Test		0.082	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.212	20:12
Breath Volume		1.972	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.213	20:15
Breath Volume		2.781	20:15
Air Blank		0.000	20:16
Control Test		0.081	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 01/31/2017 19:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FROST,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 01/21/2017 04:12:33

Diagnostics Check		OK	04:05
Air Blank		0.000	04:06
Control Test		0.081	04:06
Air Blank		0.000	04:07
Subject Sample	# 1	0.184	04:07
Breath Volume		2.886	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.176	04:10
Breath Volume		3.066	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 01/31/2017 19:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001291

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: MCREE,GARRET
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: CHILDRESS

DATE/TIME OF TEST: 01/14/2017 07:00:37

Diagnostics Check		OK	06:56	
Air Blank		0.000	06:56	
Control Test		0.080	06:56	
Air Blank		0.000	06:57	
Subject Sample	# 1		06:59	Subject Test Refused
Breath Volume		0.000	06:59	
Air Blank		0.000	06:59	
Control Test		0.080	07:00	
Air Blank		0.000	07:00	
Diagnostics Check		OK	07:00	

Data Download Date/Time: 01/30/2017 16:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SULLIVAN,RYAN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PHILIPPE

DATE/TIME OF TEST: 01/01/2017 03:21:27

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.081	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 01/06/2017 16:34

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: MCNIEL,MICHAEL
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: MAXFIELD

DATE/TIME OF TEST: 01/04/2017 19:54:15

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:51	
Control Test		0.079	19:52	
Air Blank		0.000	19:52	
Subject Sample	# 1		19:52	Subject Test Refused
Breath Volume		0.000	19:52	
Air Blank		0.000	19:53	
Control Test		0.079	19:53	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 01/06/2017 16:34

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SULLIVAN,BRANDON
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:23

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 01/08/2017 16:17:46

Diagnostics Check		OK	16:13	
Air Blank		0.000	16:13	
Control Test		0.082	16:13	
Air Blank		0.000	16:14	
Subject Sample	# 1		16:16	Subject Test Refused
Breath Volume		0.000	16:16	
Air Blank		0.000	16:16	
Control Test		0.082	16:17	
Air Blank		0.000	16:17	
Diagnostics Check		OK	16:17	

Data Download Date/Time: 01/30/2017 15:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: PARK,RYAN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ARON

DATE/TIME OF TEST: 01/09/2017 23:41:56

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.081	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 1	0.205	23:36	Volume Not Met
Breath Volume		0.312	23:36	
Air Blank		0.000	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 2		23:40	Subject Test Refused
Breath Volume		0.031	23:40	
Air Blank		0.000	23:41	
Control Test		0.079	23:41	
Air Blank		0.000	23:41	
Diagnostics Check		OK	23:41	

Data Download Date/Time: 01/30/2017 15:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001292

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERNST,WILLIAM,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ERNST,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CASWELL

DATE/TIME OF TEST: 01/14/2017 01:06:30

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.193	01:02
Breath Volume		1.394	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.184	01:05
Breath Volume		1.238	01:05
Air Blank		0.000	01:05
Control Test		0.080	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 01/30/2017 15:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: ROE,JEREMIAH
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 01/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: WILDER

DATE/TIME OF TEST: 01/08/2017 07:06:28

Diagnostics Check		OK	07:02	
Air Blank		0.000	07:02	
Control Test		0.080	07:03	
Air Blank		0.000	07:03	
Subject Sample	# 1		07:05	Subject Test Refused
Breath Volume		0.000	07:05	
Air Blank		0.000	07:05	
Control Test		0.081	07:05	
Air Blank		0.000	07:06	
Diagnostics Check		OK	07:06	

Data Download Date/Time: 02/27/2017 18:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LANDIS,BRYAN
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 01/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 01/26/2017 23:17:28

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.140	23:13
Breath Volume		4.113	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.136	23:15
Breath Volume		4.152	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/27/2017 18:42

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUSTIN,ARTHUR,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AUSTIN,ARTHUR
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 01/01/2017 02:32:37

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:21	
Control Test		0.079	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:25	No Sample Provided
Breath Volume		0.000	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 2		02:31	No Sample Provided
Breath Volume		0.000	02:31	
Air Blank		0.000	02:31	
Control Test		0.079	02:32	
Air Blank		0.000	02:32	
Diagnostics Check		OK	02:32	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RAMIREZ,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: MCCAULEY

DATE/TIME OF TEST: 01/01/2017 05:08:14

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.078	05:02
Air Blank		0.000	05:02
Subject Sample	# 1	0.117	05:03
Breath Volume		3.906	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.120	05:06
Breath Volume		3.878	05:06
Air Blank		0.000	05:07
Control Test		0.077	05:07
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHACHAREONVANICH,VANISSA,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LOHACHAREONVANICH,VANISSA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: JENKIN

DATE/TIME OF TEST: 01/02/2017 02:08:34

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.175	02:04
Breath Volume		2.863	02:04
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 2	0.175	02:06
Breath Volume		2.843	02:06
Air Blank		0.000	02:07
Control Test		0.078	02:07
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOZARD,CHRYSTAL,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: ZOMBORI

DATE/TIME OF TEST: 01/05/2017 17:39:57

Diagnostics Check		OK	17:28	
Air Blank		0.000	17:28	
Control Test		0.080	17:28	
Air Blank		0.000	17:29	
Subject Sample	# 1	0.049	17:32	Volume Not Met
Breath Volume		0.035	17:32	
Air Blank		0.000	17:33	
Air Blank		0.000	17:34	
Subject Sample	# 2	0.051	17:35	
Breath Volume		2.800	17:35	
Air Blank		0.000	17:36	
Air Blank		0.000	17:37	
Subject Sample	# 3	0.054	17:38	
Breath Volume		2.261	17:38	
Air Blank		0.000	17:39	
Control Test		0.078	17:39	
Air Blank		0.000	17:39	
Diagnostics Check		OK	17:39	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/07/2017 20:27:01

Diagnostics Check		OK	20:20
Air Blank		0.000	20:21
Control Test		0.080	20:21
Air Blank		0.000	20:22
Subject Sample	# 1	0.000	20:22
Breath Volume		2.882	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.000	20:25
Breath Volume		2.410	20:25
Air Blank		0.000	20:26
Control Test		0.079	20:26
Air Blank		0.000	20:26
Diagnostics Check		OK	20:27

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: CHAMBERLAIN

DATE/TIME OF TEST: 01/07/2017 22:33:23

Diagnostics Check		OK	22:27
Air Blank		0.000	22:27
Control Test		0.080	22:27
Air Blank		0.000	22:28
Subject Sample	# 1	0.082	22:28
Breath Volume		2.308	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.086	22:31
Breath Volume		2.281	22:31
Air Blank		0.000	22:32
Control Test		0.081	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: HOWELL

DATE/TIME OF TEST: 01/08/2017 02:06:15

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.080	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.226	02:01
Breath Volume		3.593	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.211	02:04
Breath Volume		2.578	02:04
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 01/08/2017 20:48:46

Diagnostics Check		OK	20:42
Air Blank		0.000	20:43
Control Test		0.080	20:43
Air Blank		0.000	20:44
Subject Sample	# 1	0.000	20:44
Breath Volume		2.242	20:44
Air Blank		0.000	20:44
Air Blank		0.000	20:46
Subject Sample	# 2	0.000	20:47
Breath Volume		2.320	20:47
Air Blank		0.000	20:47
Control Test		0.080	20:48
Air Blank		0.000	20:48
Diagnostics Check		OK	20:48

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOHACHAREONVANICH,VANISSA,
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: LOHACHAREOVANICH,VANISSA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SCHARTZER

DATE/TIME OF TEST: 01/09/2017 01:38:11

Diagnostics Check		OK	01:34	
Air Blank		0.000	01:35	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Subject Sample	# 1		01:36	Subject Test Refused
Breath Volume		0.000	01:36	
Air Blank		0.000	01:37	
Control Test		0.080	01:37	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 01/31/2017 23:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: ALLENDER

DATE/TIME OF TEST: 01/11/2017 23:26:10

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.096	23:21
Breath Volume		2.667	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.101	23:24
Breath Volume		3.058	23:24
Air Blank		0.000	23:25
Control Test		0.078	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,ROBERT,G
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: RAMOS,ROBERT
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: CARTE

DATE/TIME OF TEST: 01/13/2017 17:25:51

Diagnostics Check		OK	17:18
Air Blank		0.000	17:19
Control Test		0.079	17:19
Air Blank		0.000	17:20
Subject Sample	# 1	0.153	17:20
Breath Volume		3.488	17:20
Air Blank		0.000	17:21
Air Blank		0.000	17:23
Subject Sample	# 2	0.149	17:24
Breath Volume		2.554	17:24
Air Blank		0.000	17:24
Control Test		0.078	17:25
Air Blank		0.000	17:25
Diagnostics Check		OK	17:25

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,DANNY,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MILLER,DANNY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: HOSIER

DATE/TIME OF TEST: 01/13/2017 21:33:23

Diagnostics Check		OK	21:25
Air Blank		0.000	21:25
Control Test		0.079	21:26
Air Blank		0.000	21:26
Subject Sample	# 1	0.145	21:28
Breath Volume		1.804	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.138	21:31
Breath Volume		1.328	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: LOWERY

DATE/TIME OF TEST: 01/14/2017 22:24:33

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.079	22:19
Air Blank		0.000	22:19
Subject Sample	# 1	0.000	22:20
Breath Volume		1.937	22:20
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.000	22:23
Breath Volume		1.269	22:23
Air Blank		0.000	22:23
Control Test		0.080	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALABRO,ALFRED,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CALABRO,ALFRED
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: WISEMAN

DATE/TIME OF TEST: 01/16/2017 21:33:40

Diagnostics Check		OK	21:22
Air Blank		0.000	21:23
Control Test		0.079	21:23
Air Blank		0.000	21:24
Subject Sample	# 1	0.206	21:25
Breath Volume		1.332	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.231	21:29
Breath Volume		2.062	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 3	0.217	21:32
Breath Volume		2.289	21:32
Air Blank		0.000	21:32
Control Test		0.075	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: EIRAS

DATE/TIME OF TEST: 01/18/2017 04:22:15

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.116	04:17
Breath Volume		3.078	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.112	04:20
Breath Volume		1.671	04:20
Air Blank		0.000	04:21
Control Test		0.079	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEFIZ,TERRI,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELFORD,DERIC
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: NEFIZ

DATE/TIME OF TEST: 01/19/2017 00:03:37

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.079	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.000	23:59
Breath Volume		2.187	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:02
Breath Volume		1.746	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: FACKRELL

DATE/TIME OF TEST: 01/19/2017 03:13:39

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:02	
Control Test		0.080	03:03	
Air Blank		0.000	03:03	
Subject Sample	# 1		03:06	No Sample Provided
Breath Volume		0.000	03:06	
Air Blank		0.000	03:07	
Air Blank		0.000	03:09	
Subject Sample	# 2		03:12	No Sample Provided
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Control Test		0.079	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: FACKRELL

DATE/TIME OF TEST: 01/19/2017 03:35:14

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.160	03:30
Breath Volume		3.242	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.161	03:33
Breath Volume		2.683	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BASAN

DATE/TIME OF TEST: 01/19/2017 23:07:07

Diagnostics Check		OK	22:55	
Air Blank		0.000	22:55	
Control Test		0.079	22:56	
Air Blank		0.000	22:56	
Subject Sample	# 1	0.153	22:59	Volume Not Met
Breath Volume		0.695	22:59	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.172	23:02	
Breath Volume		1.808	23:02	
Air Blank		0.000	23:03	
Air Blank		0.000	23:05	
Subject Sample	# 3	0.161	23:05	
Breath Volume		1.339	23:05	
Air Blank		0.000	23:06	
Control Test		0.080	23:06	
Air Blank		0.000	23:07	
Diagnostics Check		OK	23:07	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/20/2017 23:22:53

Diagnostics Check		OK	23:16
Air Blank		0.000	23:16
Control Test		0.079	23:17
Air Blank		0.000	23:17
Subject Sample	# 1	0.175	23:18
Breath Volume		1.410	23:18
Air Blank		0.000	23:19
Air Blank		0.000	23:20
Subject Sample	# 2	0.179	23:21
Breath Volume		1.281	23:21
Air Blank		0.000	23:21
Control Test		0.078	23:22
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: ROSWELL

DATE/TIME OF TEST: 01/20/2017 23:59:47

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.079	23:54
Air Blank		0.000	23:54
Subject Sample	# 1	0.067	23:55
Breath Volume		4.812	23:55
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.068	23:58
Breath Volume		4.640	23:58
Air Blank		0.000	23:58
Control Test		0.079	23:59
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHMOND,BRENDAN,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RICHMOND,BRENDAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 01/21/2017 04:37:19

Diagnostics Check		OK	04:33	
Air Blank		0.000	04:33	
Control Test		0.079	04:34	
Air Blank		0.000	04:34	
Subject Sample	# 1		04:35	Subject Test Refused
Breath Volume		0.000	04:35	
Air Blank		0.000	04:36	
Control Test		0.079	04:36	
Air Blank		0.000	04:37	
Diagnostics Check		OK	04:37	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENNETT,EDGAR,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BENNETT,EDGAR
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: RUSHING

DATE/TIME OF TEST: 01/25/2017 03:59:42

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.272	03:55
Breath Volume		2.558	03:55
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.265	03:58
Breath Volume		2.984	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARTZOG,ROBERT,B
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARTZOG,ROBERT
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: NOWLING

DATE/TIME OF TEST: 01/26/2017 22:25:53

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.077	22:19
Air Blank		0.000	22:20
Subject Sample	# 1	0.194	22:21
Breath Volume		2.875	22:21
Air Blank		0.000	22:21
Air Blank		0.000	22:23
Subject Sample	# 2	0.192	22:24
Breath Volume		3.171	22:24
Air Blank		0.000	22:24
Control Test		0.079	22:25
Air Blank		0.000	22:25
Diagnostics Check		OK	22:25

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KIM

DATE/TIME OF TEST: 01/26/2017 23:23:43

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.081	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.047	23:19
Breath Volume		1.550	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.051	23:22
Breath Volume		1.460	23:22
Air Blank		0.000	23:22
Control Test		0.080	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: DABNEY

DATE/TIME OF TEST: 01/27/2017 00:41:26

Diagnostics Check		OK	00:38	
Air Blank		0.000	00:38	
Control Test		0.080	00:39	
Air Blank		0.000	00:39	
Subject Sample	# 1		00:40	Subject Test Refused
Breath Volume		0.000	00:40	
Air Blank		0.000	00:40	
Control Test		0.080	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: FENIMORE

DATE/TIME OF TEST: 01/27/2017 04:02:59

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.081	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:56	No Sample Provided
Breath Volume		0.000	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2		04:01	No Sample Provided
Breath Volume		0.000	04:01	
Air Blank		0.000	04:02	
Control Test		0.081	04:02	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELFORD,DERIC
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: BURMEISTER

DATE/TIME OF TEST: 01/27/2017 19:11:45

Diagnostics Check		OK	19:07	
Air Blank		0.000	19:07	
Control Test		0.080	19:08	
Air Blank		0.000	19:08	
Subject Sample	# 1		19:10	Subject Test Refused
Breath Volume		0.000	19:10	
Air Blank		0.000	19:10	
Control Test		0.080	19:11	
Air Blank		0.000	19:11	
Diagnostics Check		OK	19:11	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LEE JR

DATE/TIME OF TEST: 01/28/2017 00:51:45

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:46
Subject Sample	# 1	0.161	00:47
Breath Volume		3.808	00:47
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.158	00:50
Breath Volume		3.781	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITFIELD,HARLEY,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SHIVER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: WHITFIELD

DATE/TIME OF TEST: 01/28/2017 04:52:21

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:41	
Control Test		0.081	04:41	
Air Blank		0.000	04:42	
Subject Sample	# 1		04:45	No Sample Provided
Breath Volume		0.000	04:45	
Air Blank		0.000	04:45	
Air Blank		0.000	04:47	
Subject Sample	# 2		04:50	No Sample Provided
Breath Volume		0.000	04:50	
Air Blank		0.000	04:51	
Control Test		0.080	04:51	
Air Blank		0.000	04:52	
Diagnostics Check		OK	04:52	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELFORD,DERIC
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: HOPE

DATE/TIME OF TEST: 01/28/2017 17:41:44

Diagnostics Check		OK	17:38	
Air Blank		0.000	17:39	
Control Test		0.080	17:39	
Air Blank		0.000	17:40	
Subject Sample	# 1		17:40	Subject Test Refused
Breath Volume		0.000	17:40	
Air Blank		0.000	17:40	
Control Test		0.080	17:41	
Air Blank		0.000	17:41	
Diagnostics Check		OK	17:41	

Data Download Date/Time: 01/31/2017 23:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 01/29/2017 00:01:48

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:56
Subject Sample	# 2	0.182	23:57
Breath Volume		2.324	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 1	0.178	00:00
Breath Volume		2.328	00:00
Air Blank		0.000	00:00
Control Test		0.079	00:01
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 01/31/2017 23:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOZARD,CHRYSTAL,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: BOZARD,CHRYSTAL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:40

SUBJECT'S LAST NAME: POELLNITZ

DATE/TIME OF TEST: 01/29/2017 16:08:55

Diagnostics Check		OK	16:01
Air Blank		0.000	16:01
Control Test		0.079	16:02
Air Blank		0.000	16:02
Subject Sample	# 1	0.299	16:04
Breath Volume		2.500	16:04
Air Blank		0.000	16:05
Air Blank		0.000	16:06
Subject Sample	# 2	0.283	16:07
Breath Volume		1.476	16:07
Air Blank		0.000	16:08
Control Test		0.077	16:08
Air Blank		0.000	16:08
Diagnostics Check		OK	16:08

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELFORD,DERIC,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WELFORD,DERIC
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 17:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/29/2017 18:00:52

Diagnostics Check		OK	17:54
Air Blank		0.000	17:55
Control Test		0.080	17:55
Air Blank		0.000	17:56
Subject Sample	# 1	0.000	17:56
Breath Volume		3.992	17:56
Air Blank		0.000	17:57
Air Blank		0.000	17:59
Subject Sample	# 2	0.000	17:59
Breath Volume		3.453	17:59
Air Blank		0.000	17:59
Control Test		0.080	18:00
Air Blank		0.000	18:00
Diagnostics Check		OK	18:00

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOGGETTE,MASON,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DOGGETTE,MASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: DUNLAP

DATE/TIME OF TEST: 01/29/2017 19:33:50

Diagnostics Check		OK	19:24	
Air Blank		0.000	19:25	
Control Test		0.080	19:25	
Air Blank		0.000	19:26	
Subject Sample	# 1	0.195	19:27	
Breath Volume		1.394	19:27	
Air Blank		0.000	19:27	
Air Blank		0.000	19:29	
Subject Sample	# 2		19:32	Subject Test Refused
Breath Volume		0.199	19:32	
Air Blank		0.000	19:32	
Control Test		0.079	19:33	
Air Blank		0.000	19:33	
Diagnostics Check		OK	19:33	

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOGGETTE,MASON,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DOGGETTE,MASON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: DUNLAP

DATE/TIME OF TEST: 01/29/2017 19:44:35

Diagnostics Check	OK	19:44	
Air Blank		19:44	RFI Detect
Air Blank	0.000	19:44	

Data Download Date/Time: 01/31/2017 23:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 01/29/2017 21:10:09

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.080	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.000	21:05
Breath Volume		4.562	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.000	21:08
Breath Volume		2.253	21:08
Air Blank		0.000	21:09
Control Test		0.079	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNG,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 01/29/2017 22:31:02

Diagnostics Check		OK	22:19	
Air Blank		0.000	22:20	
Control Test		0.080	22:20	
Air Blank		0.000	22:20	
Subject Sample	# 1		22:24	No Sample Provided
Breath Volume		0.000	22:24	
Air Blank		0.000	22:24	
Air Blank		0.000	22:26	
Subject Sample	# 2		22:29	No Sample Provided
Breath Volume		0.000	22:29	
Air Blank		0.000	22:30	
Control Test		0.081	22:30	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:31	

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: MONTEITH

DATE/TIME OF TEST: 01/29/2017 23:38:05

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:32
Air Blank		0.000	23:32
Subject Sample	# 1	0.100	23:33
Breath Volume		4.136	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.094	23:36
Breath Volume		2.074	23:36
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DURKO

DATE/TIME OF TEST: 01/30/2017 03:21:53

Diagnostics Check		OK	03:15
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.194	03:17
Breath Volume		3.148	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.192	03:20
Breath Volume		3.140	03:20
Air Blank		0.000	03:20
Control Test		0.080	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VITALE,MICHAEL,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: VITALE,MICHAEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG606103
DRY GAS EXPIRY DATE: 03/01/2018
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: DAY-WELTY

DATE/TIME OF TEST: 01/30/2017 08:01:35

Diagnostics Check		OK	07:55
Air Blank		0.000	07:56
Control Test		0.080	07:56
Air Blank		0.000	07:56
Subject Sample	# 1	0.000	07:57
Breath Volume		1.308	07:57
Air Blank		0.000	07:57
Air Blank		0.000	07:59
Subject Sample	# 2	0.000	08:00
Breath Volume		1.707	08:00
Air Blank		0.000	08:00
Control Test		0.081	08:00
Air Blank		0.000	08:01
Diagnostics Check		OK	08:01

Data Download Date/Time: 01/31/2017 23:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNDON,DAVID,E
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: LEE,JERRY
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: GOINES

DATE/TIME OF TEST: 01/05/2017 04:50:50

Diagnostics Check		OK	04:38	
Air Blank		0.000	04:39	
Control Test		0.079	04:39	
Air Blank		0.000	04:40	
Subject Sample	# 1	0.079	04:43	Volume Not Met
Breath Volume		0.472	04:43	
Air Blank		0.000	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 2	0.076	04:46	
Breath Volume		1.945	04:46	
Air Blank		0.000	04:47	
Air Blank		0.000	04:48	
Subject Sample	# 3	0.077	04:49	
Breath Volume		2.207	04:49	
Air Blank		0.000	04:49	
Control Test		0.078	04:50	
Air Blank		0.000	04:50	
Diagnostics Check		OK	04:50	

Data Download Date/Time: 01/27/2017 10:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNDON,DAVID,E
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: LEE,JERRY
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CLEMMONS

DATE/TIME OF TEST: 01/06/2017 00:41:30

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:30	
Control Test		0.080	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1		00:34	No Sample Provided
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Air Blank		0.000	00:37	
Subject Sample	# 2		00:40	No Sample Provided
Breath Volume		0.000	00:40	
Air Blank		0.000	00:40	
Control Test		0.080	00:40	
Air Blank		0.000	00:41	
Diagnostics Check		OK	00:41	

Data Download Date/Time: 01/27/2017 10:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNDON,DAVID,E
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: LEE,JERRY
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: CLEMMONS

DATE/TIME OF TEST: 01/06/2017 00:59:34

Diagnostics Check		OK	00:45	
Air Blank		0.000	00:45	
Control Test		0.079	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 1	0.206	00:49	Volume Not Met
Breath Volume		0.171	00:49	
Air Blank		0.000	00:50	
Air Blank		0.000	00:51	
Subject Sample	# 2	0.221	00:53	No .020 agreement
Breath Volume		1.625	00:53	
Air Blank		0.000	00:53	
Air Blank		0.000	00:55	
Subject Sample	# 3	0.198	00:57	No .020 agreement
Breath Volume		1.421	00:57	
Air Blank		0.000	00:58	
Control Test		0.077	00:58	
Air Blank		0.000	00:59	
Diagnostics Check		OK	00:59	

Data Download Date/Time: 01/27/2017 10:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASSALLO,STEVEN,A
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: MITCHELL,WESLEY
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 01/09/2017 03:02:36

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.080	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1		03:01	Subject Test Refused
Breath Volume		0.000	03:01	
Air Blank		0.000	03:01	
Control Test		0.081	03:02	
Air Blank		0.000	03:02	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 01/27/2017 10:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOLDEN,ANTHONY,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GOLDEN,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SANCHEZ MENDEZ

DATE/TIME OF TEST: 01/10/2017 03:50:24

Diagnostics Check		OK	03:36	
Air Blank		0.000	03:36	
Control Test		0.080	03:37	
Air Blank		0.000	03:37	
Subject Sample	# 1	0.177	03:38	
Breath Volume		1.324	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2	0.154	03:44	Volume Not Met
Breath Volume		0.035	03:44	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 3	0.161	03:48	
Breath Volume		1.867	03:48	
Air Blank		0.000	03:49	
Control Test		0.078	03:49	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 01/27/2017 10:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: SHAMBLIN

DATE/TIME OF TEST: 01/10/2017 20:40:00

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:37	
Control Test		0.080	20:37	
Air Blank		0.000	20:38	
Subject Sample	# 1		20:38	Subject Test Refused
Breath Volume		0.000	20:38	
Air Blank		0.000	20:39	
Control Test		0.079	20:39	
Air Blank		0.000	20:39	
Diagnostics Check		OK	20:40	

Data Download Date/Time: 01/27/2017 10:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILL,BRIAN,K
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GRIFFITH,DALTON
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 15:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: LESTER

DATE/TIME OF TEST: 01/16/2017 16:29:08

Diagnostics Check		OK	16:22
Air Blank		0.000	16:22
Control Test		0.080	16:23
Air Blank		0.000	16:23
Subject Sample	# 1	0.202	16:24
Breath Volume		1.304	16:24
Air Blank		0.000	16:25
Air Blank		0.000	16:26
Subject Sample	# 2	0.197	16:27
Breath Volume		1.398	16:27
Air Blank		0.000	16:28
Control Test		0.079	16:28
Air Blank		0.000	16:29
Diagnostics Check		OK	16:29

Data Download Date/Time: 01/27/2017 10:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUD,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: LEE,JERRY
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/23/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HEATRICE

DATE/TIME OF TEST: 01/24/2017 01:22:08

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.080	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1	0.104	01:14	Volume Not Met
Breath Volume		0.082	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2	0.085	01:20	Volume Not Met
Breath Volume		0.144	01:20	
Air Blank		0.000	01:21	
Control Test		0.078	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 01/27/2017 10:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROUD,NICHOLAS,D
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: GUIDERA,FRANK
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1650824
DRY GAS EXPIRY DATE: 03/13/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: YATES

DATE/TIME OF TEST: 01/27/2017 19:17:40

Diagnostics Check		OK	19:11
Air Blank		0.000	19:11
Control Test		0.079	19:11
Air Blank		0.000	19:12
Subject Sample	# 1	0.212	19:13
Breath Volume		2.390	19:13
Air Blank		0.000	19:13
Air Blank		0.000	19:15
Subject Sample	# 2	0.219	19:16
Breath Volume		1.593	19:16
Air Blank		0.000	19:16
Control Test		0.078	19:17
Air Blank		0.000	19:17
Diagnostics Check		OK	19:17

Data Download Date/Time: 02/23/2017 15:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEST,KATHRINE,B
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: ROSENBERG,JEFF
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 01/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:25

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/24/2017 13:05:51

Diagnostics Check		OK	12:58
Air Blank		0.000	12:58
Control Test		0.078	12:59
Air Blank		0.000	12:59
Subject Sample	# 1	0.091	13:00
Breath Volume		1.750	13:00
Air Blank		0.000	13:01
Air Blank		0.000	13:03
Subject Sample	# 2	0.088	13:04
Breath Volume		2.457	13:04
Air Blank		0.000	13:04
Control Test		0.078	13:05
Air Blank		0.000	13:05
Diagnostics Check		OK	13:05

Data Download Date/Time: 02/08/2017 10:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001307

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEWEY,JARED,K
BREATH TEST OPERATOR AGENCY: Madison County SO
ARRESTING OFFICER: DEWEY,JARED
ARRESTING OFFICER AGENCY: Madison County SO
LAST AGENCY INSPECTION: 01/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:57

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/28/2017 19:54:25

Diagnostics Check		OK	19:48
Air Blank		0.000	19:48
Control Test		0.079	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.208	19:50
Breath Volume		1.863	19:50
Air Blank		0.000	19:50
Air Blank		0.000	19:52
Subject Sample	# 2	0.214	19:52
Breath Volume		1.437	19:52
Air Blank		0.000	19:53
Control Test		0.080	19:53
Air Blank		0.000	19:54
Diagnostics Check		OK	19:54

Data Download Date/Time: 02/08/2017 10:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SHELL,G
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: NORRED

DATE/TIME OF TEST: 01/01/2017 02:30:58

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.077	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Control Test		0.076	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 01/30/2017 12:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: SCHELL,G
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: NORRED

DATE/TIME OF TEST: 01/01/2017 02:50:18

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:47	
Control Test		0.075	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:49	
Control Test		0.076	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 01/30/2017 12:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REXFORD,CHRIS,M
BREATH TEST OPERATOR AGENCY: Springfield PD
ARRESTING OFFICER: REXFORD,C
ARRESTING OFFICER AGENCY: Springfield PD
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 01/16/2017 01:13:29

Diagnostics Check		OK	01:07
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:08
Subject Sample	# 1	0.131	01:09
Breath Volume		1.519	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.132	01:12
Breath Volume		1.476	01:12
Air Blank		0.000	01:12
Control Test		0.079	01:12
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 01/30/2017 12:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAMMONS,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: COOK,S
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/26/2017 01:29:58

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.079	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 1		01:28	Subject Test Refused
Breath Volume		0.000	01:28	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:29	
Diagnostics Check		OK	01:29	

Data Download Date/Time: 01/30/2017 12:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CUTCHIN,JAROD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/30/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 01/26/2017 05:57:11

Diagnostics Check		OK	05:53	
Air Blank		0.000	05:53	
Control Test		0.077	05:54	
Air Blank		0.000	05:54	
Subject Sample	# 1		05:55	Subject Test Refused
Breath Volume		0.000	05:55	
Air Blank		0.000	05:56	
Control Test		0.078	05:56	
Air Blank		0.000	05:57	
Diagnostics Check		OK	05:57	

Data Download Date/Time: 01/30/2017 12:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLES,JOSEPH,K
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: ALLSBROOK,GARY
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: ORLOSKY

DATE/TIME OF TEST: 01/20/2017 01:37:00

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:31	
Control Test		0.076	01:31	
Air Blank		0.000	01:32	
Subject Sample	# 1	0.109	01:33	
Breath Volume		1.296	01:33	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 2		01:35	RFI Detect
Breath Volume		0.000	01:35	
Air Blank		0.000	01:36	
Control Test		0.082	01:36	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 02/02/2017 09:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLES,JOSEPH,K
BREATH TEST OPERATOR AGENCY: Eglin AFB PD
ARRESTING OFFICER: ALLSBROOK,GARY
ARRESTING OFFICER AGENCY: Eglin AFB PD
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: O642555702810

DATE/TIME OF TEST: 01/20/2017 01:47:07

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.082	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.104	01:41
Breath Volume		1.609	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.103	01:45
Breath Volume		2.371	01:45
Air Blank		0.000	01:46
Control Test		0.082	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 02/02/2017 09:56

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,SHAWN,W
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/31/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 06:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:33

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 01/22/2017 08:04:43

Diagnostics Check		OK	07:58
Air Blank		0.000	07:58
Control Test		0.082	07:58
Air Blank		0.000	07:59
Subject Sample	# 1	0.164	08:00
Breath Volume		2.937	08:00
Air Blank		0.000	08:00
Air Blank		0.000	08:02
Subject Sample	# 2	0.154	08:03
Breath Volume		2.316	08:03
Air Blank		0.000	08:03
Control Test		0.082	08:04
Air Blank		0.000	08:04
Diagnostics Check		OK	08:04

Data Download Date/Time: 02/02/2017 09:56

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNING,LUKE,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: STEPHENS,JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BRASE

DATE/TIME OF TEST: 01/11/2017 05:31:43

Diagnostics Check		OK	05:25
Air Blank		0.000	05:25
Control Test		0.079	05:26
Air Blank		0.000	05:26
Subject Sample	# 1	0.112	05:27
Breath Volume		1.347	05:27
Air Blank		0.000	05:27
Air Blank		0.000	05:29
Subject Sample	# 2	0.107	05:30
Breath Volume		2.664	05:30
Air Blank		0.000	05:30
Control Test		0.077	05:31
Air Blank		0.000	05:31
Diagnostics Check		OK	05:31

Data Download Date/Time: 01/25/2017 16:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EARLEY,CHAD,D
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: MANZANO

DATE/TIME OF TEST: 01/17/2017 13:06:14

Diagnostics Check		OK	13:00
Air Blank		0.000	13:00
Control Test		0.081	13:01
Air Blank		0.000	13:01
Subject Sample	# 1	0.000	13:02
Breath Volume		1.753	13:02
Air Blank		0.000	13:02
Air Blank		0.000	13:04
Subject Sample	# 2	0.000	13:04
Breath Volume		2.656	13:04
Air Blank		0.000	13:05
Control Test		0.079	13:05
Air Blank		0.000	13:06
Diagnostics Check		OK	13:06

Data Download Date/Time: 01/25/2017 16:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASCHAL,THOMAS,C
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PASCHAL,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: ROSENHEIM

DATE/TIME OF TEST: 01/20/2017 07:21:57

Diagnostics Check		OK	07:15
Air Blank		0.000	07:15
Control Test		0.080	07:15
Air Blank		0.000	07:16
Subject Sample	# 1	0.150	07:16
Breath Volume		1.617	07:16
Air Blank		0.000	07:17
Air Blank		0.000	07:19
Subject Sample	# 2	0.143	07:20
Breath Volume		1.644	07:20
Air Blank		0.000	07:21
Control Test		0.080	07:21
Air Blank		0.000	07:21
Diagnostics Check		OK	07:21

Data Download Date/Time: 01/25/2017 16:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,STEVEN,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GRANDSTAFF,WAYNE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: STEVENSON

DATE/TIME OF TEST: 01/21/2017 04:55:38

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.080	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.121	04:51
Breath Volume		2.398	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.114	04:54
Breath Volume		1.296	04:54
Air Blank		0.000	04:54
Control Test		0.080	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 01/25/2017 16:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTTO,KYLE,K
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: HUTTO,KYL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: PROB/PAROL
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: MANZANO

DATE/TIME OF TEST: 01/24/2017 20:24:32

Diagnostics Check		OK	20:18
Air Blank		0.000	20:18
Control Test		0.081	20:19
Air Blank		0.000	20:19
Subject Sample	# 1	0.000	20:20
Breath Volume		2.425	20:20
Air Blank		0.000	20:20
Air Blank		0.000	20:22
Subject Sample	# 2	0.000	20:23
Breath Volume		2.031	20:23
Air Blank		0.000	20:23
Control Test		0.079	20:23
Air Blank		0.000	20:24
Diagnostics Check		OK	20:24

Data Download Date/Time: 01/25/2017 16:13

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TABITHA,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SALAS,LUIS
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163937
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 01/27/2017 19:19:14

Diagnostics Check		OK	19:12
Air Blank		0.000	19:12
Control Test		0.081	19:13
Air Blank		0.000	19:13
Subject Sample	# 1	0.146	19:14
Breath Volume		3.035	19:14
Air Blank		0.000	19:15
Air Blank		0.000	19:17
Subject Sample	# 2	0.146	19:17
Breath Volume		1.402	19:17
Air Blank		0.000	19:18
Control Test		0.080	19:18
Air Blank		0.000	19:19
Diagnostics Check		OK	19:19

Data Download Date/Time: 02/21/2017 16:35

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MOUNT,DEBORAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 01/05/2017 00:19:23

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.079	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.000	00:15
Breath Volume		2.816	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.000	00:18
Breath Volume		3.199	00:18
Air Blank		0.000	00:18
Control Test		0.077	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 01/25/2017 16:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: TICE,HUNTER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HABELL

DATE/TIME OF TEST: 01/08/2017 23:47:24

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.194	23:42
Breath Volume		1.742	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.182	23:45
Breath Volume		1.792	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/25/2017 16:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: DUBIE

DATE/TIME OF TEST: 01/12/2017 23:24:44

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.079	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.170	23:19
Breath Volume		1.140	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:21
Subject Sample	# 2	0.176	23:23
Breath Volume		1.210	23:23
Air Blank		0.000	23:23
Control Test		0.079	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 01/25/2017 16:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHULTZ,JEREMY,J
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCCORMICK,CHANTEL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: LAMMERS

DATE/TIME OF TEST: 01/13/2017 21:57:57

Diagnostics Check		OK	21:51
Air Blank		0.000	21:51
Control Test		0.080	21:51
Air Blank		0.000	21:52
Subject Sample	# 1	0.196	21:53
Breath Volume		1.617	21:53
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.190	21:56
Breath Volume		1.554	21:56
Air Blank		0.000	21:57
Control Test		0.078	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 01/25/2017 16:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BERBERT,HEATHER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: RACHALL

DATE/TIME OF TEST: 01/20/2017 22:30:35

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:25	
Control Test		0.078	22:26	
Air Blank		0.000	22:26	
Subject Sample	# 1		22:29	Subject Test Refused
Breath Volume		0.000	22:29	
Air Blank		0.000	22:29	
Control Test		0.079	22:29	
Air Blank		0.000	22:30	
Diagnostics Check		OK	22:30	

Data Download Date/Time: 01/25/2017 16:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERNACKI,CHAD,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BIERNACKI,CHAD
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: SIMMS

DATE/TIME OF TEST: 01/21/2017 19:55:07

Diagnostics Check		OK	19:48
Air Blank		0.000	19:48
Control Test		0.078	19:49
Air Blank		0.000	19:49
Subject Sample	# 1	0.126	19:50
Breath Volume		2.273	19:50
Air Blank		0.000	19:51
Air Blank		0.000	19:53
Subject Sample	# 2	0.127	19:53
Breath Volume		2.457	19:53
Air Blank		0.000	19:54
Control Test		0.076	19:54
Air Blank		0.000	19:55
Diagnostics Check		OK	19:55

Data Download Date/Time: 01/25/2017 16:10

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMMONS,RYANE,J
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: VASQUEZ,GERALDO
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 01/01/2017 22:46:53

Diagnostics Check		OK	22:40
Air Blank		0.000	22:40
Control Test		0.081	22:41
Air Blank		0.000	22:41
Subject Sample	# 1	0.075	22:42
Breath Volume		2.441	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.072	22:45
Breath Volume		1.824	22:45
Air Blank		0.000	22:45
Control Test		0.081	22:46
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 01/09/2017 14:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 01/06/2017 22:50:50

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:48	
Control Test		0.080	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:49	Subject Test Refused
Breath Volume		0.000	22:49	
Air Blank		0.000	22:49	
Control Test		0.080	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 01/09/2017 14:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 01/07/2017 00:04:12

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.239	23:59
Breath Volume		3.109	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.236	00:02
Breath Volume		2.152	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 01/09/2017 14:02

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/09/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: ACHESON

DATE/TIME OF TEST: 01/08/2017 03:12:05

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.081	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.126	03:07
Breath Volume		2.882	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:09
Subject Sample	# 2	0.125	03:10
Breath Volume		3.128	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:11
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/09/2017 14:02

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ-TAMAYO,TEODOMIRO,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GUERRERO,OMMAR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: GOMEZ-TAMAYO

DATE/TIME OF TEST: 01/14/2017 21:44:15

Diagnostics Check		OK	21:35
Air Blank		0.000	21:35
Control Test		0.081	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.269	21:39
Breath Volume		2.378	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.260	21:42
Breath Volume		1.578	21:42
Air Blank		0.000	21:43
Control Test		0.077	21:43
Air Blank		0.000	21:44
Diagnostics Check		OK	21:44

Data Download Date/Time: 02/03/2017 11:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ-TAMAYO,TEODOMIRO,
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GUERRERO,OMMAR
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: TAMAYO

DATE/TIME OF TEST: 01/14/2017 21:55:38

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:50	
Control Test		0.079	21:50	
Air Blank		0.000	21:51	
Subject Sample	# 1	0.256	21:54	
Breath Volume		2.351	21:54	
Air Blank			21:55	Purge Fail
Air Blank		0.000	21:55	

Data Download Date/Time: 02/03/2017 11:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: CRAWFORD,MATTHEW
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 01/24/2017 00:46:07

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.262	00:41
Breath Volume		3.191	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.273	00:44
Breath Volume		4.050	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/03/2017 11:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: JOLLY,WILLIAM
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 01/24/2017 19:15:43

Diagnostics Check		OK	19:13	
Air Blank		0.000	19:13	
Control Test		0.081	19:13	
Air Blank		0.000	19:14	
Subject Sample	# 1		19:14	RFI Detect
Breath Volume		0.000	19:14	
Air Blank			19:15	RFI Detect
Air Blank		0.000	19:15	

Data Download Date/Time: 02/03/2017 11:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: JOLLY,WILLIAM
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:52

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 01/24/2017 19:22:48

Diagnostics Check		OK	19:19	
Air Blank		0.000	19:19	
Control Test		0.081	19:20	
Air Blank		0.000	19:20	
Subject Sample	# 1		19:21	Subject Test Refused
Breath Volume		0.000	19:21	
Air Blank		0.000	19:21	
Control Test		0.080	19:22	
Air Blank		0.000	19:22	
Diagnostics Check		OK	19:22	

Data Download Date/Time: 02/03/2017 11:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RUEDE,KAYLA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: LOCKART

DATE/TIME OF TEST: 01/02/2017 00:51:25

Diagnostics Check		OK	00:43
Air Blank		0.000	00:44
Control Test		0.081	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.324	00:45
Breath Volume		2.636	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.320	00:49
Breath Volume		2.800	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 01/31/2017 15:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RUEDE,KAYLA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BURKETT

DATE/TIME OF TEST: 01/03/2017 03:18:48

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:17	
Control Test		0.081	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 01/31/2017 15:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RODRIGUEZ,RAFAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: STEHNEY

DATE/TIME OF TEST: 01/04/2017 20:17:39

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.081	20:14	
Air Blank		0.000	20:15	
Subject Sample	# 1		20:16	Subject Test Refused
Breath Volume		0.000	20:16	
Air Blank		0.000	20:16	
Control Test		0.081	20:17	
Air Blank		0.000	20:17	
Diagnostics Check		OK	20:17	

Data Download Date/Time: 01/31/2017 15:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WILLIAMS,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: CARVER

DATE/TIME OF TEST: 01/08/2017 02:46:14

Diagnostics Check		OK	02:39
Air Blank		0.000	02:40
Control Test		0.081	02:40
Air Blank		0.000	02:41
Subject Sample	# 1	0.153	02:41
Breath Volume		2.785	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.155	02:44
Breath Volume		2.281	02:44
Air Blank		0.000	02:45
Control Test		0.081	02:45
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 01/31/2017 15:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: MCINERNEY,JESSE
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/08/2017 23:44:08

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.082	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.155	23:39
Breath Volume		2.316	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.155	23:42
Breath Volume		2.152	23:42
Air Blank		0.000	23:43
Control Test		0.081	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 01/31/2017 15:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETO,JOSHUA,W
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: PETO,JOSHUA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: HILLS

DATE/TIME OF TEST: 01/11/2017 20:54:47

Diagnostics Check		OK	20:39	
Air Blank		0.000	20:40	
Control Test		0.081	20:40	
Air Blank		0.000	20:41	
Subject Sample	# 1	0.201	20:44	Volume Not Met
Breath Volume		0.757	20:44	
Air Blank		0.000	20:45	
Air Blank		0.000	20:46	
Subject Sample	# 2	0.231	20:49	
Breath Volume		1.777	20:49	
Air Blank		0.000	20:50	
Air Blank		0.000	20:52	
Subject Sample	# 3	0.226	20:53	
Breath Volume		1.519	20:53	
Air Blank		0.000	20:53	
Control Test		0.078	20:54	
Air Blank		0.000	20:54	
Diagnostics Check		OK	20:54	

Data Download Date/Time: 01/31/2017 15:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: RUEDE,KAYLA
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: HEITMANN

DATE/TIME OF TEST: 01/22/2017 00:51:50

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.081	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:50	
Control Test		0.081	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 01/31/2017 15:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUEDE,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: WILLIAMS,MICHAEL
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: FLYNCH

DATE/TIME OF TEST: 01/23/2017 05:40:49

Diagnostics Check		OK	05:38	
Air Blank		0.000	05:38	
Control Test		0.081	05:38	
Air Blank		0.000	05:39	
Subject Sample	# 1		05:39	Subject Test Refused
Breath Volume		0.000	05:39	
Air Blank		0.000	05:39	
Control Test		0.082	05:40	
Air Blank		0.000	05:40	
Diagnostics Check		OK	05:40	

Data Download Date/Time: 01/31/2017 15:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BROOKS,MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 01/21/2017 03:24:35

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:11	
Control Test		0.081	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1	0.147	03:15	Volume Not Met
Breath Volume		0.117	03:15	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.150	03:19	
Breath Volume		2.136	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 3	0.164	03:23	
Breath Volume		2.132	03:23	
Air Blank		0.000	03:23	
Control Test		0.080	03:23	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 01/25/2017 05:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BALLASH,BRIAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: DEFLORIO

DATE/TIME OF TEST: 01/21/2017 09:04:59

Diagnostics Check		OK	08:58
Air Blank		0.000	08:59
Control Test		0.082	08:59
Air Blank		0.000	08:59
Subject Sample	# 1	0.129	09:00
Breath Volume		2.621	09:00
Air Blank		0.000	09:01
Air Blank		0.000	09:02
Subject Sample	# 2	0.117	09:03
Breath Volume		1.199	09:03
Air Blank		0.000	09:04
Control Test		0.081	09:04
Air Blank		0.000	09:04
Diagnostics Check		OK	09:04

Data Download Date/Time: 01/25/2017 05:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MOORE,MARIAH
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 01/25/2017 00:58:31

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.082	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.177	00:53
Breath Volume		1.160	00:53
Air Blank		0.000	00:54
Air Blank		0.000	00:56
Subject Sample	# 2	0.174	00:57
Breath Volume		2.867	00:57
Air Blank		0.000	00:57
Control Test		0.081	00:57
Air Blank		0.000	00:58
Diagnostics Check		OK	00:58

Data Download Date/Time: 01/25/2017 05:03

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: SITES

DATE/TIME OF TEST: 01/26/2017 02:12:35

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:10	
Control Test		0.081	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 1		02:11	Subject Test Refused
Breath Volume		0.000	02:11	
Air Blank		0.000	02:11	
Control Test		0.081	02:12	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 02/17/2017 11:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: COLE,ROY
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MAYNARD

DATE/TIME OF TEST: 01/27/2017 02:12:10

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:09	
Control Test		0.081	02:10	
Air Blank		0.000	02:10	
Subject Sample	# 1		02:10	Subject Test Refused
Breath Volume		0.000	02:10	
Air Blank		0.000	02:11	
Control Test		0.081	02:11	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 02/17/2017 11:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER,MATTHEW,T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SCANLAN

DATE/TIME OF TEST: 01/28/2017 04:26:53

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.181	04:22
Breath Volume		2.039	04:22
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.180	04:25
Breath Volume		2.042	04:25
Air Blank		0.000	04:25
Control Test		0.082	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/17/2017 11:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULLER,B,B
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: VASQUEZ,C
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: ROMER

DATE/TIME OF TEST: 01/28/2017 22:27:45

Diagnostics Check		OK	22:24	
Air Blank		0.000	22:25	
Control Test		0.082	22:25	
Air Blank		0.000	22:26	
Subject Sample	# 1		22:26	Subject Test Refused
Breath Volume		0.000	22:26	
Air Blank		0.000	22:26	
Control Test		0.082	22:27	
Air Blank		0.000	22:27	
Diagnostics Check		OK	22:27	

Data Download Date/Time: 02/17/2017 11:29

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RISPOLI,MICHELLE
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: LAWHON

DATE/TIME OF TEST: 01/29/2017 00:27:47

Diagnostics Check		OK	00:21
Air Blank		0.000	00:21
Control Test		0.082	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.197	00:23
Breath Volume		2.738	00:23
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.189	00:26
Breath Volume		1.582	00:26
Air Blank		0.000	00:26
Control Test		0.082	00:27
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 02/17/2017 11:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUZMAN,MELISSA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: PITALLA

DATE/TIME OF TEST: 01/29/2017 01:14:12

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:11	
Control Test		0.082	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:13	
Control Test		0.082	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 02/17/2017 11:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWINDELL,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MACKENZIE,DOUG
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MANDELL

DATE/TIME OF TEST: 01/01/2017 04:02:34

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.175	03:57
Breath Volume		3.488	03:57
Air Blank		0.000	03:58
Air Blank		0.000	03:59
Subject Sample	# 2	0.168	04:00
Breath Volume		3.558	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: JOERGER,SHANE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: KLUTH

DATE/TIME OF TEST: 01/01/2017 04:38:49

Diagnostics Check		OK	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.206	04:34
Breath Volume		1.191	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.197	04:37
Breath Volume		2.933	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,Q,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 01/01/2017 23:51:50

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:40	
Control Test		0.081	23:41	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:44	No Sample Provided
Breath Volume		0.000	23:44	
Air Blank		0.000	23:45	
Air Blank		0.000	23:47	
Subject Sample	# 2		23:50	No Sample Provided
Breath Volume		0.000	23:50	
Air Blank		0.000	23:50	
Control Test		0.081	23:51	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:51	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,Q,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 01/02/2017 00:06:44

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.080	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 2		23:59	No Sample Provided
Breath Volume		0.000	23:59	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:05	No Sample Provided
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.080	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 01/25/2017 04:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 01/02/2017 00:16:17

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.081	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.054	00:11
Breath Volume		1.996	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:13
Subject Sample	# 2	0.053	00:14
Breath Volume		1.644	00:14
Air Blank		0.000	00:15
Control Test		0.080	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: TROISE

DATE/TIME OF TEST: 01/05/2017 22:07:00

Diagnostics Check		OK	21:59
Air Blank		0.000	22:00
Control Test		0.081	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.085	22:02
Breath Volume		1.375	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 2	0.084	22:05
Breath Volume		1.542	22:05
Air Blank		0.000	22:06
Control Test		0.080	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:07

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: RUDASILL

DATE/TIME OF TEST: 01/06/2017 21:17:52

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:15	
Control Test		0.081	21:15	
Air Blank		0.000	21:16	
Subject Sample	# 1		21:16	Subject Test Refused
Breath Volume		0.000	21:16	
Air Blank		0.000	21:16	
Control Test		0.081	21:17	
Air Blank		0.000	21:17	
Diagnostics Check		OK	21:17	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK,JIMMY,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,MICHAEL
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: BUSCEMI

DATE/TIME OF TEST: 01/07/2017 02:31:48

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Control Test		0.081	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BROOKS,MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: EMBINDER

DATE/TIME OF TEST: 01/07/2017 23:26:29

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.222	23:21
Breath Volume		3.273	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.212	23:24
Breath Volume		3.312	23:24
Air Blank		0.000	23:25
Control Test		0.081	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIST,SHARON,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BAKWELL,JAMES
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: BALOG

DATE/TIME OF TEST: 01/08/2017 09:20:59

Diagnostics Check		OK	09:12	
Air Blank		0.000	09:13	
Control Test		0.081	09:13	
Air Blank		0.000	09:14	
Subject Sample	# 1	0.125	09:15	
Breath Volume		1.164	09:15	
Air Blank		0.000	09:16	
Air Blank		0.000	09:17	
Subject Sample	# 2		09:19	Subject Test Refused
Breath Volume		0.000	09:19	
Air Blank		0.000	09:20	
Control Test		0.081	09:20	
Air Blank		0.000	09:20	
Diagnostics Check		OK	09:20	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSADO,L,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS,D/S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 01/09/2017 01:42:45

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.082	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.235	01:38
Breath Volume		1.191	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.241	01:41
Breath Volume		1.289	01:41
Air Blank		0.000	01:41
Control Test		0.081	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BROOKS,MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: HOUK

DATE/TIME OF TEST: 01/12/2017 03:51:02

Diagnostics Check		OK	03:40
Air Blank		0.000	03:40
Control Test		0.082	03:40
Air Blank		0.000	03:41
Subject Sample	# 1	0.052	03:44
Breath Volume		2.328	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.043	03:49
Breath Volume		2.554	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:51

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: OGONOSKI,JESSICA
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 14:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: IVEY

DATE/TIME OF TEST: 01/12/2017 15:08:36

Diagnostics Check		OK	15:02	
Air Blank		0.000	15:03	
Control Test		0.082	15:03	
Air Blank		0.000	15:03	
Subject Sample	# 1	0.000	15:04	
Breath Volume		2.847	15:04	
Air Blank		0.000	15:05	
Air Blank		0.000	15:07	
Subject Sample	# 2	0.000	15:07	
Breath Volume		2.222	15:07	
Air Blank			15:08	RFI Detect
Air Blank		0.000	15:08	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK,JIMMY,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CURBY,JACOB
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: THROM

DATE/TIME OF TEST: 01/13/2017 02:05:21

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.080	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1		02:03	Subject Test Refused
Breath Volume		0.000	02:03	
Air Blank		0.000	02:04	
Control Test		0.081	02:04	
Air Blank		0.000	02:05	
Diagnostics Check		OK	02:05	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 01/13/2017 03:26:26

Diagnostics Check		OK	03:20
Air Blank		0.000	03:20
Control Test		0.081	03:20
Air Blank		0.000	03:21
Subject Sample	# 1	0.087	03:21
Breath Volume		2.398	03:21
Air Blank		0.000	03:22
Air Blank		0.000	03:24
Subject Sample	# 2	0.085	03:24
Breath Volume		4.363	03:24
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Diagnostics Check		OK	03:26

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 01/13/2017 20:36:25

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.081	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.216	20:31
Breath Volume		1.757	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.226	20:34
Breath Volume		1.542	20:34
Air Blank		0.000	20:35
Control Test		0.079	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 01/14/2017 22:50:49

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.081	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:46	No Sample Provided
Breath Volume		0.000	22:46	
Air Blank		0.000	22:47	
Air Blank		0.000	22:49	
Subject Sample	# 2		22:49	Subject Test Refused
Breath Volume		0.000	22:49	
Air Blank		0.000	22:49	
Control Test		0.081	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:50	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FLORES,D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/15/2017 22:33:49

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:31	
Control Test		0.081	22:31	
Air Blank		0.000	22:32	
Subject Sample	# 1		22:32	Subject Test Refused
Breath Volume		0.000	22:32	
Air Blank		0.000	22:32	
Control Test		0.081	22:33	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:33	

Data Download Date/Time: 01/25/2017 04:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIGGOTT,MYRON,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: PIGGOTT,MYRON
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: COURT ORD
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: TALLEY

DATE/TIME OF TEST: 01/16/2017 19:39:35

Diagnostics Check		OK	19:33
Air Blank		0.000	19:34
Control Test		0.081	19:34
Air Blank		0.000	19:35
Subject Sample	# 1	0.000	19:35
Breath Volume		3.898	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.000	19:38
Breath Volume		3.355	19:38
Air Blank		0.000	19:38
Control Test		0.080	19:39
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 01/25/2017 05:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIGGOTT,MYRON,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS,ZACH
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SZAZ

DATE/TIME OF TEST: 01/16/2017 23:37:27

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.081	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.119	23:32
Breath Volume		2.890	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.116	23:35
Breath Volume		2.109	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 01/25/2017 05:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FISHMAN

DATE/TIME OF TEST: 01/19/2017 21:24:06

Diagnostics Check	OK	21:23	
Air Blank		21:23	RFI Detect
Air Blank	0.000	21:24	

Data Download Date/Time: 01/25/2017 05:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JAWORSKI,KEVIN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: FISHMAN

DATE/TIME OF TEST: 01/19/2017 21:33:37

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.079	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.164	21:27
Breath Volume		1.261	21:27
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.181	21:32
Breath Volume		1.792	21:32
Air Blank		0.000	21:32
Control Test		0.080	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 01/25/2017 05:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLE,MELINDA,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,CHAD
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/20/2017 22:00:32

Diagnostics Check		OK	21:50	
Air Blank		0.000	21:51	
Control Test		0.081	21:51	
Air Blank		0.000	21:51	
Subject Sample	# 1	0.110	21:55	Volume Not Met
Breath Volume		0.093	21:55	
Air Blank		0.000	21:55	
Air Blank		0.000	21:57	
Subject Sample	# 2		21:59	Subject Test Refused
Breath Volume		1.054	21:59	
Air Blank		0.000	21:59	
Control Test		0.080	21:59	
Air Blank		0.000	22:00	
Diagnostics Check		OK	22:00	

Data Download Date/Time: 01/25/2017 05:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROOKS,MATTHEW,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BROOKS,MATTHEW
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: HERNANDEZ GONZALEZ

DATE/TIME OF TEST: 01/21/2017 03:00:53

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.081	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1	0.158	02:55	Volume Not Met
Breath Volume		0.339	02:55	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 2	0.171	02:59	
Breath Volume		1.148	02:59	
Air Blank			03:00	Purge Fail
Air Blank			03:00	Purge Fail

Data Download Date/Time: 01/25/2017 05:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANGEL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BALLASH,BRIAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/28/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: DEFLORIO

DATE/TIME OF TEST: 01/21/2017 08:55:48

Diagnostics Check	OK	08:54	
Air Blank	0.000	08:54	
Control Test	0.055	08:54	Control Outside Tolerance
Air Blank		08:55	Purge Fail

Data Download Date/Time: 01/25/2017 05:00

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,EDMOND,W
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MCVEY,ROBERT
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ANGEL

DATE/TIME OF TEST: 01/21/2017 23:07:57

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.157	23:03
Breath Volume		3.531	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.145	23:06
Breath Volume		3.269	23:06
Air Blank		0.000	23:07
Control Test		0.078	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 01/22/2017 07:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTON,EDMOND,W
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: MATOS,FRANK
ARRESTING OFFICER AGENCY: Arcadia PD
LAST AGENCY INSPECTION: 12/20/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GUERRERO

DATE/TIME OF TEST: 01/22/2017 04:13:50

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:11	
Control Test		0.078	04:11	
Air Blank		0.000	04:12	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.079	04:13	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 01/22/2017 07:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEE,TIMOTHY,I
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MEE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GUTIERREZ CRUZ

DATE/TIME OF TEST: 01/31/2017 00:43:00

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.080	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.175	00:38
Breath Volume		1.707	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:40
Subject Sample	# 2	0.174	00:41
Breath Volume		1.921	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:43

Data Download Date/Time: 02/24/2017 00:02

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SGT.,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ARMAS DE LOYOLA

DATE/TIME OF TEST: 01/01/2017 02:36:26

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.077	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.158	02:31
Breath Volume		3.742	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.158	02:34
Breath Volume		3.562	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,JIMMY,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,JIMMY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: GAMBINO

DATE/TIME OF TEST: 01/01/2017 03:09:34

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.075	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.158	03:04
Breath Volume		3.097	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.163	03:07
Breath Volume		2.750	03:07
Air Blank		0.000	03:08
Control Test		0.077	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BELL,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PACE

DATE/TIME OF TEST: 01/01/2017 20:34:05

Diagnostics Check		OK	20:27
Air Blank		0.000	20:28
Control Test		0.078	20:28
Air Blank		0.000	20:29
Subject Sample	# 1	0.212	20:29
Breath Volume		2.945	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.196	20:32
Breath Volume		2.445	20:32
Air Blank		0.000	20:33
Control Test		0.078	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEPUTY,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY G.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: ESPINOZA-ARREOLA

DATE/TIME OF TEST: 01/01/2017 21:20:22

Diagnostics Check		OK	21:14
Air Blank		0.000	21:14
Control Test		0.078	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.201	21:16
Breath Volume		2.414	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.196	21:18
Breath Volume		2.285	21:18
Air Blank		0.000	21:19
Control Test		0.078	21:19
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: LINDSEY,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: REISH

DATE/TIME OF TEST: 01/02/2017 01:20:34

Diagnostics Check		OK	01:13
Air Blank		0.000	01:14
Control Test		0.078	01:14
Air Blank		0.000	01:15
Subject Sample	# 1	0.196	01:15
Breath Volume		4.253	01:15
Air Blank		0.000	01:16
Air Blank		0.000	01:18
Subject Sample	# 2	0.198	01:18
Breath Volume		4.582	01:18
Air Blank		0.000	01:19
Control Test		0.078	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOBAS,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KENNY,S
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: MCNIERNEY

DATE/TIME OF TEST: 01/04/2017 02:11:29

Diagnostics Check		OK	02:04
Air Blank		0.000	02:05
Control Test		0.078	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.150	02:06
Breath Volume		3.941	02:06
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.157	02:09
Breath Volume		4.113	02:09
Air Blank		0.000	02:10
Control Test		0.077	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHAFFER,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 01/04/2017 22:35:30

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.078	22:29
Air Blank		0.000	22:30
Subject Sample	# 1	0.000	22:31
Breath Volume		3.363	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.000	22:34
Breath Volume		3.000	22:34
Air Blank		0.000	22:34
Control Test		0.078	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/27/2017 07:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BRANDAU,S
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ALFARANO

DATE/TIME OF TEST: 01/06/2017 23:57:13

Diagnostics Check		OK	23:49
Air Blank		0.000	23:49
Control Test		0.078	23:50
Air Blank		0.000	23:50
Subject Sample	# 1	0.000	23:52
Breath Volume		2.402	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		2.042	23:55
Air Blank		0.000	23:56
Control Test		0.077	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: THERRIEN

DATE/TIME OF TEST: 01/07/2017 02:35:19

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.078	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.165	02:29
Breath Volume		1.531	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.166	02:33
Breath Volume		1.585	02:33
Air Blank		0.000	02:34
Control Test		0.077	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SGT.,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/07/2017 03:31:55

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.077	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.166	03:26
Breath Volume		2.480	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.155	03:30
Breath Volume		2.335	03:30
Air Blank		0.000	03:31
Control Test		0.077	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SGT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: SPICER

DATE/TIME OF TEST: 01/07/2017 21:01:39

Diagnostics Check		OK	20:54
Air Blank		0.000	20:55
Control Test		0.078	20:55
Air Blank		0.000	20:55
Subject Sample	# 1	0.124	20:57
Breath Volume		3.300	20:57
Air Blank		0.000	20:57
Air Blank		0.000	20:59
Subject Sample	# 2	0.108	21:00
Breath Volume		2.441	21:00
Air Blank		0.000	21:00
Control Test		0.077	21:01
Air Blank		0.000	21:01
Diagnostics Check		OK	21:01

Data Download Date/Time: 01/27/2017 07:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: 6L13,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NICK

DATE/TIME OF TEST: 01/08/2017 03:23:52

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.078	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1	0.155	03:14	Volume Not Met
Breath Volume		0.285	03:14	
Air Blank		0.000	03:15	
Air Blank		0.000	03:17	
Subject Sample	# 2	0.171	03:18	
Breath Volume		1.183	03:18	
Air Blank		0.000	03:19	
Air Blank		0.000	03:21	
Subject Sample	# 3	0.159	03:22	
Breath Volume		2.132	03:22	
Air Blank		0.000	03:22	
Control Test		0.069	03:23	Control Outside Tolerance
Air Blank		0.000	03:23	

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: 6L13,NICK,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/08/2017 03:42:52

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:30	
Control Test		0.077	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.163	03:32	
Breath Volume		2.234	03:32	
Air Blank		0.000	03:33	
Air Blank		0.000	03:35	
Subject Sample	# 2	0.171	03:38	Volume Not Met
Breath Volume		0.148	03:38	
Air Blank		0.000	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 3	0.155	03:41	
Breath Volume		1.652	03:41	
Air Blank		0.000	03:41	
Control Test		0.075	03:42	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SGT.,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 01/08/2017 03:52:52

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.173	03:48
Breath Volume		2.562	03:48
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.166	03:51
Breath Volume		2.578	03:51
Air Blank		0.000	03:51
Control Test		0.077	03:52
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 01/27/2017 07:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MOORE,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: STORGOFF

DATE/TIME OF TEST: 01/09/2017 18:36:15

Diagnostics Check		OK	18:29
Air Blank		0.000	18:30
Control Test		0.078	18:30
Air Blank		0.000	18:31
Subject Sample	# 1	0.187	18:31
Breath Volume		2.359	18:31
Air Blank		0.000	18:32
Air Blank		0.000	18:34
Subject Sample	# 2	0.177	18:34
Breath Volume		2.132	18:34
Air Blank		0.000	18:35
Control Test		0.078	18:35
Air Blank		0.000	18:36
Diagnostics Check		OK	18:36

Data Download Date/Time: 01/27/2017 07:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SGT.,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SGT.
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 01/13/2017 05:25:18

Diagnostics Check		OK	05:16
Air Blank		0.000	05:16
Control Test		0.077	05:16
Air Blank		0.000	05:17
Subject Sample	# 1	0.215	05:19
Breath Volume		1.179	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:21
Subject Sample	# 2	0.202	05:23
Breath Volume		2.093	05:23
Air Blank		0.000	05:24
Control Test		0.076	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MIKLUSCKA,ZACK
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: MANDUJANO

DATE/TIME OF TEST: 01/14/2017 03:42:07

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.078	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.048	03:36
Breath Volume		1.621	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.049	03:40
Breath Volume		2.488	03:40
Air Blank		0.000	03:41
Control Test		0.078	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICHTER,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DECKER,OFFICER
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 01/14/2017 06:24:52

Diagnostics Check		OK	06:18
Air Blank		0.000	06:18
Control Test		0.078	06:19
Air Blank		0.000	06:19
Subject Sample	# 1	0.261	06:20
Breath Volume		2.367	06:20
Air Blank		0.000	06:21
Air Blank		0.000	06:22
Subject Sample	# 2	0.250	06:23
Breath Volume		2.152	06:23
Air Blank		0.000	06:23
Control Test		0.076	06:24
Air Blank		0.000	06:24
Diagnostics Check		OK	06:24

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDERSON,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 01/16/2017 01:42:57

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.079	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1	0.119	01:35	Volume Not Met
Breath Volume		0.074	01:35	
Air Blank		0.000	01:36	
Air Blank		0.000	01:38	
Subject Sample	# 2	0.119	01:41	Volume Not Met
Breath Volume		0.429	01:41	
Air Blank		0.000	01:42	
Control Test		0.077	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 01/27/2017 07:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MIKLUSCAK,Z
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SHEPARD

DATE/TIME OF TEST: 01/16/2017 03:57:41

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.077	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.096	03:52
Breath Volume		1.972	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.086	03:56
Breath Volume		1.656	03:56
Air Blank		0.000	03:56
Control Test		0.078	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DELLER

DATE/TIME OF TEST: 01/19/2017 04:23:38

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.076	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.133	04:19
Breath Volume		2.546	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.136	04:22
Breath Volume		2.640	04:22
Air Blank		0.000	04:22
Control Test		0.077	04:23
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MINGS,DEPUTY G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MINGS,DEPUTY G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SALVATORE

DATE/TIME OF TEST: 01/19/2017 20:45:38

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.076	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.229	20:41
Breath Volume		2.195	20:41
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.223	20:44
Breath Volume		1.695	20:44
Air Blank		0.000	20:44
Control Test		0.076	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIEHL,LEE,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: DIEHL,LEE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 01/20/2017 05:29:30

Diagnostics Check		OK	05:23
Air Blank		0.000	05:23
Control Test		0.079	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.191	05:25
Breath Volume		1.320	05:25
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.185	05:27
Breath Volume		2.097	05:27
Air Blank		0.000	05:28
Control Test		0.078	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: NEFF,B
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 01/21/2017 01:51:26

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.078	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.079	01:46
Breath Volume		3.281	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.076	01:49
Breath Volume		3.687	01:49
Air Blank		0.000	01:50
Control Test		0.077	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: ORIOL

DATE/TIME OF TEST: 01/25/2017 04:27:59

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.079	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.100	04:23
Breath Volume		4.457	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:25
Subject Sample	# 2	0.099	04:26
Breath Volume		4.488	04:26
Air Blank		0.000	04:27
Control Test		0.078	04:27
Air Blank		0.000	04:27
Diagnostics Check		OK	04:27

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/27/2017 00:40:26

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.077	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.000	00:36
Breath Volume		2.738	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.000	00:39
Breath Volume		2.132	00:39
Air Blank		0.000	00:39
Control Test		0.077	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 01/27/2017 07:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BLOMGREN,J
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: HALLSTROM

DATE/TIME OF TEST: 01/27/2017 23:35:52

Diagnostics Check		OK	23:29
Air Blank		0.000	23:30
Control Test		0.078	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.023	23:31
Breath Volume		1.949	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:33
Subject Sample	# 2	0.022	23:34
Breath Volume		1.832	23:34
Air Blank		0.000	23:34
Control Test		0.079	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 02/27/2017 07:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURLEY,SHAUN,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,SHAUN
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: KURTZ

DATE/TIME OF TEST: 01/29/2017 04:42:18

Diagnostics Check		OK	04:36
Air Blank		0.000	04:36
Control Test		0.079	04:36
Air Blank		0.000	04:37
Subject Sample	# 1	0.202	04:37
Breath Volume		2.699	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.200	04:40
Breath Volume		3.160	04:40
Air Blank		0.000	04:41
Control Test		0.078	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 02/27/2017 07:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWELL,SHAWNA,M
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: CARPENTER,TRAVIS
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/02/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: GRIFFIS

DATE/TIME OF TEST: 01/05/2017 22:28:11

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:16	
Control Test		0.080	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 1		22:20	No Sample Provided
Breath Volume		0.000	22:20	
Air Blank		0.000	22:20	
Air Blank		0.000	22:22	
Subject Sample	# 2	0.252	22:23	
Breath Volume		2.691	22:23	
Air Blank		0.000	22:24	
Air Blank		0.000	22:25	
Subject Sample	# 3	0.247	22:26	
Breath Volume		2.289	22:26	
Air Blank		0.000	22:27	
Control Test		0.080	22:27	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 01/13/2017 09:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WESTCOTT

DATE/TIME OF TEST: 01/01/2017 03:17:32

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.080	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.116	03:13
Breath Volume		2.484	03:13
Air Blank		0.000	03:13
Air Blank		0.000	03:15
Subject Sample	# 2	0.107	03:16
Breath Volume		1.769	03:16
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:17
Diagnostics Check		OK	03:17

Data Download Date/Time: 01/27/2017 10:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 01/05/2017 01:04:36

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.080	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.118	01:00
Breath Volume		3.953	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.114	01:03
Breath Volume		3.003	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 01/27/2017 10:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,CHARLES
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 01/13/2017 19:14:09

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.080	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.000	19:09
Breath Volume		4.804	19:09
Air Blank		0.000	19:10
Air Blank		0.000	19:12
Subject Sample	# 2	0.000	19:12
Breath Volume		4.089	19:12
Air Blank		0.000	19:13
Control Test		0.080	19:13
Air Blank		0.000	19:14
Diagnostics Check		OK	19:14

Data Download Date/Time: 01/27/2017 10:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: SALIMI

DATE/TIME OF TEST: 01/13/2017 22:38:54

Diagnostics Check		OK	22:32
Air Blank		0.000	22:33
Control Test		0.079	22:33
Air Blank		0.000	22:34
Subject Sample	# 1	0.007	22:34
Breath Volume		2.011	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:37
Subject Sample	# 2	0.000	22:37
Breath Volume		1.417	22:37
Air Blank		0.000	22:38
Control Test		0.078	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 01/27/2017 10:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERN,DEP,B
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: BEATON

DATE/TIME OF TEST: 01/14/2017 19:02:03

Diagnostics Check		OK	18:55
Air Blank		0.000	18:56
Control Test		0.080	18:56
Air Blank		0.000	18:57
Subject Sample	# 1	0.000	18:57
Breath Volume		2.792	18:57
Air Blank		0.000	18:58
Air Blank		0.000	19:00
Subject Sample	# 2	0.000	19:00
Breath Volume		3.164	19:00
Air Blank		0.000	19:01
Control Test		0.079	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:02

Data Download Date/Time: 01/27/2017 10:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 01/15/2017 01:51:39

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:46
Subject Sample	# 1	0.165	01:47
Breath Volume		3.492	01:47
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.159	01:50
Breath Volume		2.699	01:50
Air Blank		0.000	01:50
Control Test		0.079	01:51
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 01/27/2017 10:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,M
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/13/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: CANAL

DATE/TIME OF TEST: 01/25/2017 21:15:33

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.079	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.203	21:09
Breath Volume		1.128	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.200	21:13
Breath Volume		1.273	21:13
Air Blank		0.000	21:14
Control Test		0.079	21:14
Air Blank		0.000	21:15
Diagnostics Check		OK	21:15

Data Download Date/Time: 01/27/2017 10:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001350

INSTRUMENT REGISTERED TO: FRUITLAND PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINS,HENRY,E
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: RAINS,HENRY
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 01/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A1
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: BYRNES

DATE/TIME OF TEST: 01/20/2017 23:39:27

Diagnostics Check		OK	23:29	
Air Blank		0.000	23:30	
Control Test		0.079	23:30	
Air Blank		0.000	23:30	
Subject Sample	# 1	0.124	23:34	Volume Not Met
Breath Volume		0.667	23:34	
Air Blank		0.000	23:34	
Air Blank		0.000	23:36	
Subject Sample	# 2		23:37	Subject Test Refused
Breath Volume		0.781	23:37	
Air Blank		0.000	23:38	
Control Test		0.078	23:38	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:39	

Data Download Date/Time: 02/03/2017 11:38

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DEAL

DATE/TIME OF TEST: 01/19/2017 02:41:52

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.082	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.057	02:37
Breath Volume		1.414	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.056	02:40
Breath Volume		1.714	02:40
Air Blank		0.000	02:40
Control Test		0.081	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 01/30/2017 02:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: BAGLEY

DATE/TIME OF TEST: 01/20/2017 02:35:08

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.038	02:30
Breath Volume		1.738	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.036	02:33
Breath Volume		1.605	02:33
Air Blank		0.000	02:34
Control Test		0.082	02:34
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 01/30/2017 02:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 12/19/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: TRIVETTE

DATE/TIME OF TEST: 01/27/2017 02:24:18

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.134	02:19
Breath Volume		3.371	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.121	02:22
Breath Volume		1.957	02:22
Air Blank		0.000	02:23
Control Test		0.082	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 01/30/2017 02:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,MEGAN,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: THWETT,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/03/2017 00:06:38

Diagnostics Check		OK	00:00	
Air Blank		0.000	00:01	
Control Test		0.079	00:01	
Air Blank		0.000	00:01	
Subject Sample	# 1		00:04	RFI Detect
Breath Volume		1.031	00:04	
Air Blank		0.000	00:05	
Control Test		0.078	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 01/21/2017 20:11

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,MEGAN,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: THWEATT,CAMERON
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/03/2017 00:23:37

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.078	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 1	0.208	00:13	Slope Not Level
Breath Volume		1.250	00:13	
Air Blank		0.000	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 2	0.205	00:17	
Breath Volume		2.851	00:17	
Air Blank		0.000	00:18	
Air Blank		0.000	00:19	
Subject Sample	# 3	0.197	00:21	
Breath Volume		1.261	00:21	
Air Blank		0.000	00:22	
Control Test		0.077	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 01/21/2017 20:11

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: JOHN,GUYTON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: RODRIGUEZ LUCIANO

DATE/TIME OF TEST: 01/06/2017 01:01:32

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.078	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.115	00:57
Breath Volume		2.453	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.124	01:00
Breath Volume		2.132	01:00
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/21/2017 20:11

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: SINGLETON

DATE/TIME OF TEST: 01/08/2017 04:55:36

Diagnostics Check		OK	04:49
Air Blank		0.000	04:50
Control Test		0.078	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.023	04:51
Breath Volume		1.562	04:51
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.022	04:54
Breath Volume		2.363	04:54
Air Blank		0.000	04:54
Control Test		0.079	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 01/21/2017 20:11

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,ROBERT,P
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BALLENGER,CASSANDRA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: TILLMAN

DATE/TIME OF TEST: 01/08/2017 21:22:17

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.000	21:18
Breath Volume		1.722	21:18
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.000	21:20
Breath Volume		1.351	21:20
Air Blank		0.000	21:21
Control Test		0.079	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 01/21/2017 20:11

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,MEGAN,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SINGLETARY,KENNETH
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: KIFFER

DATE/TIME OF TEST: 01/10/2017 19:32:11

Diagnostics Check		OK	19:20	
Air Blank		0.000	19:21	
Control Test		0.078	19:21	
Air Blank		0.000	19:22	
Subject Sample	# 1		19:25	No Sample Provided
Breath Volume		0.000	19:25	
Air Blank		0.000	19:25	
Air Blank		0.000	19:27	
Subject Sample	# 2		19:30	No Sample Provided
Breath Volume		0.000	19:30	
Air Blank		0.000	19:31	
Control Test		0.078	19:31	
Air Blank		0.000	19:32	
Diagnostics Check		OK	19:32	

Data Download Date/Time: 01/21/2017 20:11

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: RAMKA

DATE/TIME OF TEST: 01/13/2017 03:12:46

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.144	03:07
Breath Volume		1.902	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.141	03:11
Breath Volume		1.265	03:11
Air Blank		0.000	03:11
Control Test		0.078	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/21/2017 20:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 01/14/2017 01:20:18

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:10	
Control Test		0.078	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:14	No Sample Provided
Breath Volume		0.000	01:14	
Air Blank		0.000	01:15	
Air Blank		0.000	01:17	
Subject Sample	# 2		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:19	
Control Test		0.079	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 01/21/2017 20:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 01/28/2017 04:04:34

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.132	04:00
Breath Volume		2.445	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.128	04:03
Breath Volume		1.765	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 02/24/2017 22:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: JILES,DARRYL
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: JOYCE

DATE/TIME OF TEST: 01/29/2017 00:20:32

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.079	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.169	00:15
Breath Volume		1.875	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.168	00:19
Breath Volume		1.550	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 02/14/2017 10:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: JILES,DARRYL
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 01/29/2017 04:02:00

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.115	03:57
Breath Volume		1.906	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.112	04:00
Breath Volume		2.894	04:00
Air Blank		0.000	04:01
Control Test		0.080	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:02

Data Download Date/Time: 02/14/2017 10:29

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO,GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CROOKS

DATE/TIME OF TEST: 01/01/2017 03:03:08

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.082	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.340	02:58	
Breath Volume		2.339	02:58	
Air Blank		0.000	02:59	
Air Blank		0.000	03:01	
Subject Sample	# 2		03:01	Slope Not Met
Breath Volume		1.226	03:01	
Air Blank		0.000	03:02	
Control Test		0.080	03:02	
Air Blank		0.000	03:03	
Diagnostics Check		OK	03:03	

Data Download Date/Time: 01/05/2017 14:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,SHAWN,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO,GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: CROOKS

DATE/TIME OF TEST: 01/01/2017 03:35:50

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.081	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.306	03:31
Breath Volume		2.292	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:33
Subject Sample	# 2	0.314	03:34
Breath Volume		2.300	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 01/05/2017 14:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALEJANDRO,GENESIS
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: WOODS

DATE/TIME OF TEST: 01/16/2017 04:26:51

Diagnostics Check		OK	04:20
Air Blank		0.000	04:21
Control Test		0.083	04:21
Air Blank		0.000	04:22
Subject Sample	# 1	0.155	04:22
Breath Volume		3.246	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:24
Subject Sample	# 2	0.153	04:25
Breath Volume		2.414	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:26
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 01/18/2017 11:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 12/14/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LIN

DATE/TIME OF TEST: 01/18/2017 04:29:19

Diagnostics Check		OK	04:21
Air Blank		0.000	04:21
Control Test		0.083	04:22
Air Blank		0.000	04:22
Subject Sample	# 1	0.116	04:23
Breath Volume		1.144	04:23
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.119	04:27
Breath Volume		1.359	04:27
Air Blank		0.000	04:28
Control Test		0.082	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 01/18/2017 11:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,ELVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 01/19/2017 03:44:23

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:32	
Control Test		0.082	03:32	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:36	No Sample Provided
Breath Volume		0.000	03:36	
Air Blank		0.000	03:37	
Air Blank		0.000	03:38	
Subject Sample	# 2	0.190	03:39	
Breath Volume		2.664	03:39	
Air Blank		0.000	03:40	
Air Blank		0.000	03:41	
Subject Sample	# 3	0.195	03:42	
Breath Volume		2.281	03:42	
Air Blank		0.000	03:43	
Control Test		0.082	03:43	
Air Blank		0.000	03:44	
Diagnostics Check		OK	03:44	

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,MARCIE,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMPSON,MAURICE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:56

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 01/19/2017 08:27:52

Diagnostics Check		OK	08:20
Air Blank		0.000	08:20
Control Test		0.082	08:21
Air Blank		0.000	08:21
Subject Sample	# 1	0.142	08:22
Breath Volume		2.121	08:22
Air Blank		0.000	08:23
Air Blank		0.000	08:25
Subject Sample	# 2	0.140	08:26
Breath Volume		2.320	08:26
Air Blank		0.000	08:26
Control Test		0.080	08:27
Air Blank		0.000	08:27
Diagnostics Check		OK	08:27

Data Download Date/Time: 01/26/2017 13:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: HERNANDEZ PEREZ

DATE/TIME OF TEST: 01/21/2017 00:03:29

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.083	23:58
Air Blank		0.000	23:58
Subject Sample	# 2	0.058	23:59
Breath Volume		2.363	23:59
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.059	00:02
Breath Volume		3.300	00:02
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,A
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/21/2017 01:43:14

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.225	01:38
Breath Volume		2.089	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.229	01:41
Breath Volume		1.613	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THORNTON,BRANDEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BOBOLA

DATE/TIME OF TEST: 01/21/2017 03:27:20

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.133	03:22
Breath Volume		2.281	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.133	03:25
Breath Volume		3.898	03:25
Air Blank		0.000	03:26
Control Test		0.081	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GARCIA,EDDIE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: LA ROCCO

DATE/TIME OF TEST: 01/21/2017 06:24:50

Diagnostics Check		OK	06:17
Air Blank		0.000	06:17
Control Test		0.081	06:17
Air Blank		0.000	06:18
Subject Sample	# 1	0.118	06:19
Breath Volume		1.484	06:19
Air Blank		0.000	06:20
Air Blank		0.000	06:22
Subject Sample	# 2	0.120	06:23
Breath Volume		2.117	06:23
Air Blank		0.000	06:23
Control Test		0.081	06:24
Air Blank		0.000	06:24
Diagnostics Check		OK	06:24

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WINBURN,TAYLOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: MCKNIGHT

DATE/TIME OF TEST: 01/21/2017 22:19:40

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.082	22:14
Air Blank		0.000	22:14
Subject Sample	# 1	0.274	22:15
Breath Volume		1.324	22:15
Air Blank		0.000	22:15
Air Blank		0.000	22:17
Subject Sample	# 2	0.279	22:18
Breath Volume		1.324	22:18
Air Blank		0.000	22:18
Control Test		0.081	22:19
Air Blank		0.000	22:19
Diagnostics Check		OK	22:19

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,W
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RENDON

DATE/TIME OF TEST: 01/22/2017 02:50:55

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.082	02:44
Air Blank		0.000	02:45
Subject Sample	# 1	0.184	02:46
Breath Volume		2.390	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.182	02:49
Breath Volume		2.339	02:49
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PRICE,F
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: RICHBURG

DATE/TIME OF TEST: 01/23/2017 02:38:32

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.082	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.238	02:34
Breath Volume		1.351	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.238	02:36
Breath Volume		2.214	02:36
Air Blank		0.000	02:37
Control Test		0.081	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,C
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: NEVAREZ

DATE/TIME OF TEST: 01/24/2017 01:02:29

Diagnostics Check		OK	00:56
Air Blank		0.000	00:56
Control Test		0.082	00:57
Air Blank		0.000	00:57
Subject Sample	# 1	0.181	00:58
Breath Volume		2.324	00:58
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.176	01:00
Breath Volume		2.230	01:00
Air Blank		0.000	01:01
Control Test		0.081	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JUSTINA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 01/24/2017 06:43:46

Diagnostics Check		OK	06:32	
Air Blank		0.000	06:32	
Control Test		0.082	06:32	
Air Blank		0.000	06:33	
Subject Sample	# 1		06:36	No Sample Provided
Breath Volume		0.000	06:36	
Air Blank		0.000	06:36	
Air Blank		0.000	06:39	
Subject Sample	# 2	0.219	06:42	Volume Not Met
Breath Volume		0.101	06:42	
Air Blank		0.000	06:42	
Control Test		0.083	06:43	
Air Blank		0.000	06:43	
Diagnostics Check		OK	06:43	

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,JUSTINA
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 01/24/2017 06:57:29

Diagnostics Check		OK	06:46	
Air Blank		0.000	06:46	
Control Test		0.081	06:46	
Air Blank		0.000	06:47	
Subject Sample	# 1	0.219	06:50	Volume Not Met
Breath Volume		0.136	06:50	
Air Blank		0.000	06:51	
Air Blank		0.000	06:52	
Subject Sample	# 2		06:56	No Sample Provided
Breath Volume		0.000	06:56	
Air Blank		0.000	06:56	
Control Test		0.080	06:56	
Air Blank		0.000	06:57	
Diagnostics Check		OK	06:57	

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FIGUEROA,HECTOR,A
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,STEPHEN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: BEASLEY

DATE/TIME OF TEST: 01/25/2017 03:18:24

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.082	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.175	03:14
Breath Volume		1.890	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.167	03:16
Breath Volume		1.179	03:16
Air Blank		0.000	03:17
Control Test		0.082	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 01/26/2017 13:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRANNON,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: ABOUABBES

DATE/TIME OF TEST: 01/25/2017 17:19:02

Diagnostics Check		OK	17:12
Air Blank		0.000	17:13
Control Test		0.081	17:13
Air Blank		0.000	17:13
Subject Sample	# 1	0.393	17:14
Breath Volume		1.187	17:14
Air Blank		0.000	17:15
Air Blank		0.000	17:17
Subject Sample	# 2	0.391	17:17
Breath Volume		2.453	17:17
Air Blank		0.000	17:18
Control Test		0.081	17:18
Air Blank		0.000	17:18
Diagnostics Check		OK	17:19

Data Download Date/Time: 01/26/2017 13:05

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WINBURN,TAYLOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BUSCHBOM

DATE/TIME OF TEST: 01/26/2017 03:42:18

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.082	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.132	03:37
Breath Volume		4.246	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.127	03:40
Breath Volume		2.742	03:40
Air Blank		0.000	03:41
Control Test		0.082	03:41
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 01/26/2017 13:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE,SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 01/27/2017 03:53:20

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.082	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.166	03:48
Breath Volume		2.179	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.170	03:51
Breath Volume		3.027	03:51
Air Blank		0.000	03:52
Control Test		0.082	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 02/02/2017 13:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LAWRENCE,NICKOLAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 01/27/2017 22:38:58

Diagnostics Check		OK	22:31
Air Blank		0.000	22:32
Control Test		0.082	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.068	22:34
Breath Volume		2.460	22:34
Air Blank		0.000	22:35
Air Blank		0.000	22:36
Subject Sample	# 2	0.062	22:37
Breath Volume		1.191	22:37
Air Blank		0.000	22:38
Control Test		0.082	22:38
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BURKHEAD,KYLE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: KIRBY

DATE/TIME OF TEST: 01/28/2017 01:54:19

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.083	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.083	01:49
Breath Volume		2.914	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.082	01:52
Breath Volume		3.742	01:52
Air Blank		0.000	01:53
Control Test		0.082	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/28/2017 03:52:13

Diagnostics Check		OK	03:45
Air Blank		0.000	03:45
Control Test		0.083	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.174	03:47
Breath Volume		1.585	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.182	03:50
Breath Volume		2.984	03:50
Air Blank		0.000	03:51
Control Test		0.081	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOONDA,SHUAIB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:11

SUBJECT'S LAST NAME: MENNA

DATE/TIME OF TEST: 01/28/2017 07:43:55

Diagnostics Check		OK	07:37
Air Blank		0.000	07:37
Control Test		0.083	07:38
Air Blank		0.000	07:38
Subject Sample	# 1	0.137	07:39
Breath Volume		2.441	07:39
Air Blank		0.000	07:40
Air Blank		0.000	07:41
Subject Sample	# 2	0.142	07:42
Breath Volume		3.554	07:42
Air Blank		0.000	07:43
Control Test		0.083	07:43
Air Blank		0.000	07:43
Diagnostics Check		OK	07:43

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PETERSON,JOSEPH
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: LABOY

DATE/TIME OF TEST: 01/29/2017 02:19:39

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.082	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.181	02:15
Breath Volume		3.269	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.188	02:18
Breath Volume		2.261	02:18
Air Blank		0.000	02:18
Control Test		0.082	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAHLAU,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/29/2017 23:16:05

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:13	
Control Test		0.083	23:13	
Air Blank		0.000	23:13	
Subject Sample	# 1		23:14	Interferent Detect
Breath Volume		1.117	23:14	
Air Blank			23:15	Purge Fail
Air Blank		0.000	23:16	

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAHLAU,NICHOLAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 01/29/2017 23:24:04

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.082	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.157	23:19
Breath Volume		1.238	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:21
Subject Sample	# 2	0.154	23:22
Breath Volume		1.507	23:22
Air Blank		0.000	23:23
Control Test		0.081	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:24

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GLASS,A
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MOSES

DATE/TIME OF TEST: 01/31/2017 04:57:20

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.082	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.189	04:52
Breath Volume		1.390	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.196	04:55
Breath Volume		2.718	04:55
Air Blank		0.000	04:56
Control Test		0.082	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 02/02/2017 13:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GLENN,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: DONES

DATE/TIME OF TEST: 01/01/2017 03:42:47

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.081	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.215	03:38
Breath Volume		2.937	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.229	03:41
Breath Volume		2.816	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 01/21/2017 14:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GLENN,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 01/06/2017 03:54:14

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.082	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.259	03:49
Breath Volume		2.042	03:49
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.256	03:52
Breath Volume		2.164	03:52
Air Blank		0.000	03:53
Control Test		0.082	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 01/21/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHE,KEVIN,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: YOUNGBLOOD,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: POLM

DATE/TIME OF TEST: 01/11/2017 04:44:22

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.082	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.163	04:39
Breath Volume		2.324	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.161	04:42
Breath Volume		1.964	04:42
Air Blank		0.000	04:43
Control Test		0.082	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/21/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TENNENT,R
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: RADER

DATE/TIME OF TEST: 01/11/2017 22:52:25

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.081	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.113	22:47
Breath Volume		3.132	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:49
Subject Sample	# 2	0.113	22:50
Breath Volume		2.992	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 01/21/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TENNENT,R
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/13/2017 04:15:47

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.082	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:10	No Sample Provided
Breath Volume		0.000	04:10	
Air Blank		0.000	04:10	
Air Blank		0.000	04:12	
Subject Sample	# 2		04:14	Subject Test Refused
Breath Volume		0.000	04:14	
Air Blank		0.000	04:14	
Control Test		0.082	04:15	
Air Blank		0.000	04:15	
Diagnostics Check		OK	04:15	

Data Download Date/Time: 01/21/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GRACKI,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 01/14/2017 01:38:22

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:30	
Control Test		0.082	01:30	
Air Blank		0.000	01:31	
Subject Sample	# 1		01:34	No Sample Provided
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Air Blank		0.000	01:36	
Subject Sample	# 2		01:37	Subject Test Refused
Breath Volume		0.000	01:37	
Air Blank		0.000	01:37	
Control Test		0.082	01:37	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

Data Download Date/Time: 01/21/2017 14:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/11/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: MCLAUGHLIN

DATE/TIME OF TEST: 01/14/2017 04:06:35

Diagnostics Check		OK	03:59
Air Blank		0.000	03:59
Control Test		0.082	03:59
Air Blank		0.000	04:00
Subject Sample	# 1	0.000	04:01
Breath Volume		1.843	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:05
Breath Volume		1.406	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 01/21/2017 14:41

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BERDIN

DATE/TIME OF TEST: 01/21/2017 18:56:23

Diagnostics Check		OK	18:46
Air Blank		0.000	18:47
Control Test		0.081	18:47
Air Blank		0.000	18:48
Subject Sample	# 1	0.276	18:49
Breath Volume		1.578	18:49
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.259	18:54
Breath Volume		1.589	18:54
Air Blank		0.000	18:55
Control Test		0.080	18:55
Air Blank		0.000	18:56
Diagnostics Check		OK	18:56

Data Download Date/Time: 02/11/2017 14:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: PATRICK AFB SFS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JESSOP,ROBERT,W
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: JESSOP,JESSOP
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 01/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: OTHER
ARREST TIME: 18:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: GRABNER-DYSON

DATE/TIME OF TEST: 01/14/2017 18:40:38

Diagnostics Check		OK	18:37
Air Blank		0.000	18:37
Control Test		0.077	18:38
Air Blank		0.000	18:38
Subject Sample	# 1	0.127	18:39
Breath Volume		2.562	18:39
Air Blank		0.000	18:39
Air Blank		0.000	18:40

Data Download Date/Time: 02/06/2017 14:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001471

INSTRUMENT REGISTERED TO: PATRICK AFB SFS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLTON,RICHARD,A
BREATH TEST OPERATOR AGENCY: Patrick AFB PD
ARRESTING OFFICER: OLIVER,KYNDELL
ARRESTING OFFICER AGENCY: Patrick AFB PD
LAST AGENCY INSPECTION: 01/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GRABNERDYSON

DATE/TIME OF TEST: 01/14/2017 20:50:09

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.077	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.137	20:45
Breath Volume		2.480	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.131	20:48
Breath Volume		1.687	20:48
Air Blank		0.000	20:49
Control Test		0.076	20:49
Air Blank		0.000	20:50
Diagnostics Check		OK	20:50

Data Download Date/Time: 02/06/2017 14:01

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,NATHAN,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WILKINS,JEREMY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 001
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CRAVEN

DATE/TIME OF TEST: 01/01/2017 01:20:16

Diagnostics Check	OK	01:19	
Air Blank		01:19	RFI Detect
Air Blank	0.000	01:20	

Data Download Date/Time: 01/10/2017 13:08

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,NATHAN,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WILKINS,JEREMY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 001
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: CRAVEN

DATE/TIME OF TEST: 01/01/2017 01:29:19

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.160	01:24
Breath Volume		1.777	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.155	01:27
Breath Volume		1.828	01:27
Air Blank		0.000	01:28
Control Test		0.078	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 01/10/2017 13:08

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGNER,NATHAN,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SMITH,TERRANCE
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 001
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: BUSH

DATE/TIME OF TEST: 01/02/2017 03:21:15

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Subject Sample	# 1	0.210	03:16
Breath Volume		2.714	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.212	03:19
Breath Volume		2.640	03:19
Air Blank		0.000	03:20
Control Test		0.076	03:20
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/10/2017 13:08

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,KALEN,M
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 001
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 01/08/2017 04:57:45

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.111	04:53
Breath Volume		1.457	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:55
Subject Sample	# 2	0.104	04:56
Breath Volume		1.867	04:56
Air Blank		0.000	04:56
Control Test		0.077	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 01/10/2017 13:08

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,KALEN,M
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: FORD,FLOYD
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 001
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 01/08/2017 05:30:20

Diagnostics Check		OK	05:23
Air Blank		0.000	05:24
Control Test		0.079	05:24
Air Blank		0.000	05:24
Subject Sample	# 1	0.215	05:25
Breath Volume		2.261	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.197	05:28
Breath Volume		1.843	05:28
Air Blank		0.000	05:29
Control Test		0.075	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 01/10/2017 13:08

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,KALEN,M
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 001
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:44

SUBJECT'S LAST NAME: BRAZIER

DATE/TIME OF TEST: 01/08/2017 07:11:37

Diagnostics Check		OK	07:05
Air Blank		0.000	07:05
Control Test		0.080	07:06
Air Blank		0.000	07:06
Subject Sample	# 1	0.000	07:07
Breath Volume		1.281	07:07
Air Blank		0.000	07:07
Air Blank		0.000	07:09
Subject Sample	# 2	0.000	07:10
Breath Volume		1.207	07:10
Air Blank		0.000	07:10
Control Test		0.077	07:10
Air Blank		0.000	07:11
Diagnostics Check		OK	07:11

Data Download Date/Time: 01/10/2017 13:08

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMOS,MATHEW,T
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LAMANA,PHILIP
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:17

SUBJECT'S LAST NAME: HERNANDEZ III

DATE/TIME OF TEST: 01/06/2017 14:46:45

Diagnostics Check		OK	14:40
Air Blank		0.000	14:41
Control Test		0.080	14:41
Air Blank		0.000	14:41
Subject Sample	# 1	0.000	14:42
Breath Volume		4.050	14:42
Air Blank		0.000	14:42
Air Blank		0.000	14:44
Subject Sample	# 2	0.000	14:45
Breath Volume		4.000	14:45
Air Blank		0.000	14:45
Control Test		0.079	14:46
Air Blank		0.000	14:46
Diagnostics Check		OK	14:46

Data Download Date/Time: 01/12/2017 10:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FESSEL,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 01/07/2017 01:56:07

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:41	
Control Test		0.080	01:41	
Air Blank		0.000	01:42	
Subject Sample	# 1	0.153	01:45	Volume Not Met
Breath Volume		0.031	01:45	
Air Blank		0.000	01:45	
Air Blank		0.000	01:47	
Subject Sample	# 2	0.173	01:48	
Breath Volume		1.609	01:48	
Air Blank		0.000	01:49	
Air Blank		0.000	01:51	
Subject Sample	# 3	0.153	01:54	Volume Not Met
Breath Volume		0.039	01:54	
Air Blank		0.000	01:55	
Control Test		0.078	01:55	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 01/12/2017 10:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FESSEL,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TUCKER,HAILEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: FINI

DATE/TIME OF TEST: 01/07/2017 21:03:47

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.080	20:58
Air Blank		0.000	20:58
Subject Sample	# 1	0.241	20:59
Breath Volume		3.660	20:59
Air Blank		0.000	20:59
Air Blank		0.000	21:01
Subject Sample	# 2	0.250	21:02
Breath Volume		3.515	21:02
Air Blank		0.000	21:02
Control Test		0.079	21:03
Air Blank		0.000	21:03
Diagnostics Check		OK	21:03

Data Download Date/Time: 01/12/2017 10:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CIESLINSKI,JOSEPH,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HAMILTON,CORTNEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/06/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SHERNOWITZ

DATE/TIME OF TEST: 01/08/2017 04:30:47

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.190	04:26
Breath Volume		3.558	04:26
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.188	04:29
Breath Volume		2.871	04:29
Air Blank		0.000	04:29
Control Test		0.080	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 01/12/2017 10:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: COSLICK

DATE/TIME OF TEST: 01/12/2017 23:28:39

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.197	23:24
Breath Volume		3.164	23:24
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.202	23:27
Breath Volume		2.378	23:27
Air Blank		0.000	23:27
Control Test		0.079	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: FESSEL,DANIEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: FORREST

DATE/TIME OF TEST: 01/13/2017 20:01:03

Diagnostics Check		OK	19:55
Air Blank		0.000	19:55
Control Test		0.080	19:56
Air Blank		0.000	19:56
Subject Sample	# 1	0.000	19:56
Breath Volume		3.574	19:56
Air Blank		0.000	19:57
Air Blank		0.000	19:59
Subject Sample	# 2	0.000	19:59
Breath Volume		3.816	19:59
Air Blank		0.000	20:00
Control Test		0.080	20:00
Air Blank		0.000	20:00
Diagnostics Check		OK	20:01

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUFFY,JONATHAN,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ESTERLE,H
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: CESNER

DATE/TIME OF TEST: 01/17/2017 18:02:38

Diagnostics Check		OK	17:55
Air Blank		0.000	17:56
Control Test		0.079	17:56
Air Blank		0.000	17:57
Subject Sample	# 1	0.000	17:58
Breath Volume		1.996	17:58
Air Blank		0.000	17:58
Air Blank		0.000	18:00
Subject Sample	# 2	0.000	18:01
Breath Volume		1.785	18:01
Air Blank		0.000	18:01
Control Test		0.079	18:02
Air Blank		0.000	18:02
Diagnostics Check		OK	18:02

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK,MICHAEL,O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DEBERARDINIS,DREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 01/18/2017 23:21:28

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.080	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.132	23:16
Breath Volume		1.371	23:16
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.131	23:20
Breath Volume		1.550	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LEONARD,STEPHEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: FEIKERT

DATE/TIME OF TEST: 01/19/2017 01:11:49

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:06
Subject Sample	# 1	0.000	01:07
Breath Volume		2.828	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.000	01:10
Breath Volume		3.675	01:10
Air Blank		0.000	01:10
Control Test		0.080	01:11
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 02/22/2017 10:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,DANNY,W
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: GRAY,DANNY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 01/19/2017 03:20:44

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.209	03:16
Breath Volume		1.171	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.222	03:19
Breath Volume		2.433	03:19
Air Blank		0.000	03:19
Control Test		0.079	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMOS,MATHEW,T
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOOS,SAVANNAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: DELACEY

DATE/TIME OF TEST: 01/20/2017 16:16:12

Diagnostics Check		OK	16:10
Air Blank		0.000	16:10
Control Test		0.080	16:10
Air Blank		0.000	16:11
Subject Sample	# 1	0.000	16:11
Breath Volume		2.136	16:11
Air Blank		0.000	16:12
Air Blank		0.000	16:14
Subject Sample	# 2	0.000	16:14
Breath Volume		1.453	16:14
Air Blank		0.000	16:15
Control Test		0.080	16:15
Air Blank		0.000	16:16
Diagnostics Check		OK	16:16

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO,MARK,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WHIDDEN,TYLER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: EVEY

DATE/TIME OF TEST: 01/20/2017 17:49:54

Diagnostics Check		OK	17:38	
Air Blank		0.000	17:38	
Control Test		0.079	17:39	
Air Blank		0.000	17:39	
Subject Sample	# 1	0.149	17:42	Volume Not Met
Breath Volume		0.066	17:42	
Air Blank		0.000	17:43	
Air Blank		0.000	17:45	
Subject Sample	# 2	0.157	17:48	Slope Not Level
Breath Volume		1.164	17:48	
Air Blank		0.000	17:49	
Control Test		0.077	17:49	
Air Blank		0.000	17:49	
Diagnostics Check		OK	17:49	

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATO,MARK,P
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: WHIDDEN,TYLER
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:16

SUBJECT'S LAST NAME: EVEY

DATE/TIME OF TEST: 01/20/2017 18:09:12

Diagnostics Check		OK	17:55	
Air Blank		0.000	17:56	
Control Test		0.079	17:56	
Air Blank		0.000	17:57	
Subject Sample	# 1	0.154	18:00	Volume Not Met
Breath Volume		0.890	18:00	
Air Blank		0.000	18:00	
Air Blank		0.000	18:02	
Subject Sample	# 2	0.167	18:03	
Breath Volume		1.132	18:03	
Air Blank		0.000	18:04	
Air Blank		0.000	18:06	
Subject Sample	# 3	0.167	18:07	
Breath Volume		1.125	18:07	
Air Blank		0.000	18:08	
Control Test		0.078	18:08	
Air Blank		0.000	18:09	
Diagnostics Check		OK	18:09	

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 01/21/2017 01:44:34

Diagnostics Check		OK	01:35
Air Blank		0.000	01:35
Control Test		0.080	01:36
Air Blank		0.000	01:36
Subject Sample	# 1	0.185	01:37
Breath Volume		2.199	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.174	01:43
Breath Volume		2.464	01:43
Air Blank		0.000	01:43
Control Test		0.080	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 02/22/2017 10:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HAMILTON REYES,COURTNEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: SPIVEY

DATE/TIME OF TEST: 01/21/2017 05:04:08

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.079	04:58
Air Blank		0.000	04:58
Subject Sample	# 1	0.175	04:59
Breath Volume		1.597	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:01
Subject Sample	# 2	0.167	05:02
Breath Volume		1.730	05:02
Air Blank		0.000	05:03
Control Test		0.079	05:03
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: CRAGIN,S
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: HENRI

DATE/TIME OF TEST: 01/22/2017 03:18:29

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.205	03:13
Breath Volume		2.445	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:15
Subject Sample	# 2	0.214	03:17
Breath Volume		2.214	03:17
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUSER,HEATHER,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: JACKSON,M
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: POTTLE

DATE/TIME OF TEST: 01/23/2017 02:12:15

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Subject Sample	# 1		02:10	Subject Test Refused
Breath Volume		0.000	02:10	
Air Blank		0.000	02:11	
Control Test		0.079	02:11	
Air Blank		0.000	02:12	
Diagnostics Check		OK	02:12	

Data Download Date/Time: 02/22/2017 10:17

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELAY,ANDREW,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: KYNION,JASMINE
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: BECKNER

DATE/TIME OF TEST: 01/29/2017 04:39:23

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.077	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:38	
Control Test		0.079	04:38	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 02/22/2017 22:56

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/07/2017 04:02:32

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:56	
Control Test		0.077	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.185	03:58	
Breath Volume		1.394	03:58	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2		04:01	RFI Detect
Breath Volume		0.035	04:01	
Air Blank		0.000	04:01	
Control Test		0.077	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 01/14/2017 20:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 01/07/2017 04:10:20

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.076	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.193	04:05
Breath Volume		2.347	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.194	04:08
Breath Volume		1.304	04:08
Air Blank		0.000	04:09
Control Test		0.077	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 01/14/2017 20:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 01/07/2017 21:04:49

Diagnostics Check		OK	20:58
Air Blank		0.000	20:58
Control Test		0.077	20:59
Air Blank		0.000	20:59
Subject Sample	# 1	0.180	21:00
Breath Volume		3.675	21:00
Air Blank		0.000	21:01
Air Blank		0.000	21:02
Subject Sample	# 2	0.192	21:03
Breath Volume		3.496	21:03
Air Blank		0.000	21:03
Control Test		0.076	21:04
Air Blank		0.000	21:04
Diagnostics Check		OK	21:04

Data Download Date/Time: 01/14/2017 20:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: FLY

DATE/TIME OF TEST: 01/09/2017 00:49:44

Diagnostics Check		OK	00:40
Air Blank		0.000	00:41
Control Test		0.078	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.221	00:45
Breath Volume		2.808	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.218	00:48
Breath Volume		2.375	00:48
Air Blank		0.000	00:48
Control Test		0.079	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 01/14/2017 20:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SOUSA,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/15/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 01/11/2017 05:44:00

Diagnostics Check		OK	05:36	
Air Blank		0.000	05:36	
Control Test		0.076	05:36	
Air Blank		0.000	05:37	
Subject Sample	# 1	0.117	05:38	
Breath Volume		1.507	05:38	
Air Blank		0.000	05:39	
Air Blank		0.000	05:41	
Subject Sample	# 2		05:42	RFI Detect
Breath Volume		0.000	05:42	
Air Blank		0.000	05:43	
Control Test		0.078	05:43	
Air Blank		0.000	05:43	
Diagnostics Check		OK	05:44	

Data Download Date/Time: 01/14/2017 20:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KRONZ,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: MOYA

DATE/TIME OF TEST: 01/18/2017 01:55:23

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.079	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1	0.181	01:50	Volume Not Met
Breath Volume		0.000	01:50	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 2		01:53	Subject Test Refused
Breath Volume		0.375	01:53	
Air Blank		0.000	01:54	
Control Test		0.076	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 02/08/2017 21:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: NEUMANN

DATE/TIME OF TEST: 01/19/2017 05:04:42

Diagnostics Check		OK	04:57
Air Blank		0.000	04:57
Control Test		0.076	04:58
Air Blank		0.000	04:58
Subject Sample	# 1	0.231	04:59
Breath Volume		2.140	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.233	05:03
Breath Volume		2.992	05:03
Air Blank		0.000	05:03
Control Test		0.078	05:04
Air Blank		0.000	05:04
Diagnostics Check		OK	05:04

Data Download Date/Time: 02/08/2017 21:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 01/20/2017 05:07:43

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:03	
Control Test		0.078	05:04	
Air Blank		0.000	05:04	
Subject Sample	# 1		05:06	Subject Test Refused
Breath Volume		0.000	05:06	
Air Blank		0.000	05:06	
Control Test		0.078	05:07	
Air Blank		0.000	05:07	
Diagnostics Check		OK	05:07	

Data Download Date/Time: 02/08/2017 21:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 01/20/2017 23:56:05

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:52	
Control Test		0.078	23:53	
Air Blank		0.000	23:53	
Subject Sample	# 1	0.184	23:54	
Breath Volume		1.375	23:54	
Air Blank			23:54	RFI Detect
Air Blank		0.000	23:55	
Control Test		0.078	23:55	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 02/08/2017 21:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MCBRIDE

DATE/TIME OF TEST: 01/21/2017 00:04:15

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.078	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.196	23:59
Breath Volume		1.246	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.186	00:02
Breath Volume		1.375	00:02
Air Blank		0.000	00:03
Control Test		0.078	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 02/08/2017 21:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 01/21/2017 00:43:41

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.077	00:40	
Air Blank		0.000	00:40	
Subject Sample	# 1		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:42	
Control Test		0.077	00:43	
Air Blank		0.000	00:43	
Diagnostics Check		OK	00:43	

Data Download Date/Time: 02/08/2017 21:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: JOHNSON,COREY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: PATTATHAN

DATE/TIME OF TEST: 01/28/2017 05:54:06

Diagnostics Check		OK	05:50	
Air Blank		0.000	05:51	
Control Test		0.078	05:51	
Air Blank		0.000	05:51	
Subject Sample	# 1		05:52	Subject Test Refused
Breath Volume		0.000	05:52	
Air Blank		0.000	05:53	
Control Test		0.078	05:53	
Air Blank		0.000	05:54	
Diagnostics Check		OK	05:54	

Data Download Date/Time: 02/08/2017 21:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ACRI,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 01/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A2
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: KIRKWOOD

DATE/TIME OF TEST: 01/29/2017 04:56:06

Diagnostics Check		OK	04:50
Air Blank		0.000	04:50
Control Test		0.077	04:50
Air Blank		0.000	04:51
Subject Sample	# 1	0.161	04:51
Breath Volume		2.000	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:54
Subject Sample	# 2	0.168	04:54
Breath Volume		1.796	04:54
Air Blank		0.000	04:55
Control Test		0.078	04:55
Air Blank		0.000	04:56
Diagnostics Check		OK	04:56

Data Download Date/Time: 02/08/2017 21:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RINCON,SEBASTIAN,R
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TOBIN,EDWARD
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: VILLARD

DATE/TIME OF TEST: 01/15/2017 02:26:09

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:12	
Control Test		0.082	02:12	
Air Blank		0.000	02:13	
Subject Sample	# 1	0.188	02:15	
Breath Volume		2.289	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 2	0.148	02:20	Volume Not Met
Breath Volume		1.050	02:20	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 3	0.190	02:24	
Breath Volume		2.921	02:24	
Air Blank		0.000	02:25	
Control Test		0.074	02:25	Control Outside Tolerance
Air Blank		0.000	02:26	

Data Download Date/Time: 01/31/2017 18:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERNE,LUIS,N
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VERNE,LUIS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: MENZIES

DATE/TIME OF TEST: 01/15/2017 18:35:26

Diagnostics Check		OK	18:30	
Air Blank		0.000	18:30	
Control Test		0.081	18:30	
Air Blank		0.000	18:31	
Subject Sample	# 1		18:34	Subject Test Refused
Breath Volume		0.000	18:34	
Air Blank		0.000	18:34	
Control Test		0.081	18:34	
Air Blank		0.000	18:35	
Diagnostics Check		OK	18:35	

Data Download Date/Time: 01/31/2017 18:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEVENS,ALLIXEN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VOKATY,THOMAS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CARDOSO

DATE/TIME OF TEST: 01/26/2017 01:45:07

Diagnostics Check		OK	01:38
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.129	01:40
Breath Volume		2.062	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.132	01:43
Breath Volume		2.410	01:43
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 01/31/2017 18:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUZA,TANIA,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BAUZA,TANIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 01/29/2017 04:36:21

Diagnostics Check	OK	04:35	
Air Blank		04:35	RFI Detect
Air Blank	0.000	04:36	

Data Download Date/Time: 01/31/2017 18:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001654

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAUZA,TANIA,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BAUZA,TANIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/29/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 01/29/2017 04:43:16

Diagnostics Check		OK	04:40	
Air Blank		0.000	04:40	
Control Test		0.082	04:41	
Air Blank		0.000	04:41	
Subject Sample	# 1		04:41	Subject Test Refused
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Control Test		0.081	04:42	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 01/31/2017 18:54

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 12/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: KIROLOS

DATE/TIME OF TEST: 01/09/2017 05:13:39

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.079	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.107	05:09
Breath Volume		4.156	05:09
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.108	05:11
Breath Volume		3.566	05:11
Air Blank		0.000	05:12
Control Test		0.078	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 01/18/2017 20:14

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: ROACH,ANTHONY
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 12/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: ALUSMA

DATE/TIME OF TEST: 01/16/2017 07:05:34

Diagnostics Check		OK	07:02	
Air Blank		0.000	07:02	
Control Test		0.080	07:03	
Air Blank		0.000	07:03	
Subject Sample	# 1		07:04	Subject Test Refused
Breath Volume		0.000	07:04	
Air Blank		0.000	07:04	
Control Test		0.080	07:04	
Air Blank		0.000	07:05	
Diagnostics Check		OK	07:05	

Data Download Date/Time: 01/18/2017 20:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 12/27/2016

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 01/18/2017 06:58:42

Diagnostics Check		OK	06:55	
Air Blank		0.000	06:55	
Control Test		0.079	06:56	
Air Blank		0.000	06:56	
Subject Sample	# 1		06:57	Subject Test Refused
Breath Volume		0.000	06:57	
Air Blank		0.000	06:57	
Control Test		0.079	06:58	
Air Blank		0.000	06:58	
Diagnostics Check		OK	06:58	

Data Download Date/Time: 01/18/2017 20:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.