Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOLBICKA,LAWRENCE,R
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: KOLBICKA,LAWRENCE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 02/05/2017 23:25:54

Diagnostics Check		OK	23:17
Air Blank		0.000	23:18
Control Test		0.080	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.240	23:20
Breath Volume		3.066	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:22
Subject Sample	# 2	0.235	23:24
Breath Volume		1.914	23:24
Air Blank		0.000	23:24
Control Test		0.076	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 02/15/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: CIOPPA

DATE/TIME OF TEST: 02/06/2017 00:00:37

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.078	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.149	23:56
Breath Volume		1.886	23:56
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.142	23:59
Breath Volume		1.968	23:59
Air Blank		0.000	23:59
Control Test		0.078	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,OFFICER,B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SIMPKINS,OFC
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 02/06/2017 00:19:52

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.098	00:15
Breath Volume		2.015	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.100	00:18
Breath Volume		2.519	00:18
Air Blank		0.000	00:18
Control Test		0.078	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 02/06/2017 04:38:55

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.080	04:36	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:37	Subject Test Refused
Breath Volume		0.000	04:37	
Air Blank		0.000	04:38	
Control Test		0.079	04:38	
Air Blank		0.000	04:38	
Diagnostics Check		OK	04:38	

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: RASMUSSEN

DATE/TIME OF TEST: 02/07/2017 00:04:12

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.079	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.186	23:59
Breath Volume		3.246	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.190	00:02
Breath Volume		2.531	00:02
Air Blank		0.000	00:03
Control Test		0.078	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 02/10/2017 03:59:52

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.138	03:55
Breath Volume		1.906	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.132	03:58
Breath Volume		1.859	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: TOWLER

DATE/TIME OF TEST: 02/10/2017 22:56:07

Diagnostics Check		OK	22:53	
Air Blank		0.000	22:53	
Control Test		0.079	22:53	
Air Blank		0.000	22:54	
Subject Sample	# 1		22:54	Subject Test Refused
Breath Volume		0.000	22:54	
Air Blank		0.000	22:55	
Control Test		0.079	22:55	
Air Blank		0.000	22:56	
Diagnostics Check		OK	22:56	

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: KRONK

DATE/TIME OF TEST: 02/11/2017 04:54:51

Diagnostics Check		OK	04:51	
Air Blank		0.000	04:52	
Control Test		0.079	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:53	Subject Test Refused
Breath Volume		0.000	04:53	
Air Blank		0.000	04:53	
Control Test		0.079	04:54	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/12/2017 03:22:32

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.080	03:16
Air Blank		0.000	03:17
Subject Sample	# 1	0.117	03:17
Breath Volume		1.941	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.112	03:20
Breath Volume		2.261	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: FLETCHER

DATE/TIME OF TEST: 02/13/2017 01:19:45

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.079	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:18	
Control Test		0.079	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:19	

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CHESTINE

DATE/TIME OF TEST: 02/14/2017 02:01:49

Diagnostics Check		OK	01:54
Air Blank		0.000	01:55
Control Test		0.078	01:55
Air Blank		0.000	01:56
Subject Sample	# 1	0.087	01:57
Breath Volume		2.320	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.088	02:00
Breath Volume		2.695	02:00
Air Blank		0.000	02:00
Control Test		0.078	02:01
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: 4L29,BOYLE
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: IPPOLITO

DATE/TIME OF TEST: 02/14/2017 03:34:25

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.116	03:29
Breath Volume		3.324	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:32
Subject Sample	# 2	0.112	03:32
Breath Volume		2.906	03:32
Air Blank		0.000	03:33
Control Test		0.078	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DISCHER

DATE/TIME OF TEST: 02/15/2017 04:19:34

Diagnostics Check		OK	04:16	
Air Blank		0.000	04:17	
Control Test		0.080	04:17	
Air Blank		0.000	04:18	
Subject Sample	# 1		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:18	
Control Test		0.079	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 02/15/2017 08:47

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,OFFICER,B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SIMPKINS,OFFICER
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 02/17/2017 02:43:17

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.080	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.104	02:38
Breath Volume		4.378	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.101	02:41
Breath Volume		4.425	02:41
Air Blank		0.000	02:42
Control Test		0.079	02:42
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 03/14/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: STOCKWELL

DATE/TIME OF TEST: 02/21/2017 01:42:31

Diagnostics Check		OK	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Subject Sample	# 1	0.149	01:31	Volume Not Met
Breath Volume		0.128	01:31	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.179	01:35	
Breath Volume		1.710	01:35	
Air Blank		0.000	01:35	
Air Blank		0.000	01:37	
Subject Sample	# 3	0.173	01:40	
Breath Volume		2.343	01:40	
Air Blank		0.000	01:41	
Control Test		0.077	01:41	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 03/14/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEA,CPL,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SHEA,JOHN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 02/22/2017 18:47:01

Diagnostics Check		OK	18:41
Air Blank		0.000	18:41
Control Test		0.079	18:41
Air Blank		0.000	18:42
Subject Sample	# 1	0.000	18:42
Breath Volume		2.070	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.000	18:45
Breath Volume		1.234	18:45
Air Blank		0.000	18:46
Control Test		0.078	18:46
Air Blank		0.000	18:46
Diagnostics Check		OK	18:47

Data Download Date/Time: 03/14/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOYLE,THOMAS,E
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BOYLE,THOMAS
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SERRANO

DATE/TIME OF TEST: 02/23/2017 19:27:07

Diagnostics Check		OK	19:20
Air Blank		0.000	19:21
Control Test		0.078	19:21
Air Blank		0.000	19:22
Subject Sample	# 1	0.078	19:22
Breath Volume		1.371	19:22
Air Blank		0.000	19:23
Air Blank		0.000	19:25
Subject Sample	# 2	0.091	19:25
Breath Volume		2.667	19:25
Air Blank		0.000	19:26
Control Test		0.078	19:26
Air Blank		0.000	19:27
Diagnostics Check		OK	19:27

Data Download Date/Time: 03/14/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,OFFICER,B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCORDINO,OFFICER
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: PODLAS

DATE/TIME OF TEST: 02/25/2017 02:55:09

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.079	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.000	02:50
Breath Volume		2.152	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.000	02:53
Breath Volume		2.039	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 03/14/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,OFFICER,B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCORDINO,OFFICER
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FERRARI

DATE/TIME OF TEST: 02/26/2017 02:23:21

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.079	02:21	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:21	RFI Detect
Breath Volume		0.000	02:21	
Air Blank		0.000	02:22	
Control Test		0.079	02:22	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 03/14/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001732

INSTRUMENT REGISTERED TO: PINELLAS PARK P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPKINS,OFFICER,B
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: SCORDINO,OFFICER
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FERRARI

DATE/TIME OF TEST: 02/26/2017 02:29:49

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.079	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:28	
Control Test		0.078	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 03/14/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001743

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEVEREAUX,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KERSHAW,SEAN
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:01

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 02/20/2017 07:28:12

Diagnostics Check		OK	07:22
Air Blank		0.000	07:22
Control Test		0.080	07:22
Air Blank		0.000	07:23
Subject Sample	# 1	0.017	07:23
Breath Volume		3.324	07:23
Air Blank		0.000	07:24
Air Blank		0.000	07:26
Subject Sample	# 2	0.017	07:26
Breath Volume		3.222	07:26
Air Blank		0.000	07:27
Control Test		0.079	07:27
Air Blank		0.000	07:28
Diagnostics Check		OK	07:28

Data Download Date/Time: 02/24/2017 10:45

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: COSNER,STEVEN
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BRIDGEWATER JR

DATE/TIME OF TEST: 02/04/2017 03:02:12

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.169	02:57
Breath Volume		1.484	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.158	03:00
Breath Volume		1.222	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOCCARES,PEDRO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BERGAMASCO

DATE/TIME OF TEST: 02/10/2017 23:46:23

Diagnostics Check		OK	23:31	
Air Blank		0.000	23:32	
Control Test		0.080	23:32	
Air Blank		0.000	23:32	
Subject Sample	# 1	0.096	23:36	Volume Not Met
Breath Volume		0.269	23:36	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 2	0.113	23:39	
Breath Volume		1.152	23:39	
Air Blank		0.000	23:39	
Air Blank		0.000	23:41	
Subject Sample	# 3	0.110	23:44	Volume Not Met
Breath Volume		1.082	23:44	
Air Blank		0.000	23:45	
Control Test		0.078	23:45	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: SOCARRAS,PEDRO
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: BERGAMASCO

DATE/TIME OF TEST: 02/11/2017 00:02:05

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:49	
Control Test		0.078	23:49	
Air Blank		0.000	23:50	
Subject Sample	# 2	0.111	23:50	
Breath Volume		2.019	23:50	
Air Blank		0.000	23:51	
Air Blank		0.000	23:53	
Subject Sample	# 3	0.077	23:56	Volume Not Met
Breath Volume		0.363	23:56	
Air Blank		0.000	23:56	
Air Blank		0.000	23:58	
Subject Sample	# 1		00:00	Subject Test Refused
Breath Volume		0.039	00:00	
Air Blank		0.000	00:01	
Control Test		0.078	00:01	
Air Blank		0.000	00:02	
Diagnostics Check		OK	00:02	

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,DAVID,R
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PRICE,DAVID
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 07:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:38

SUBJECT'S LAST NAME: FONSECA CUAREZMA

DATE/TIME OF TEST: 02/18/2017 08:06:17

Diagnostics Check		OK	07:58
Air Blank		0.000	07:59
Control Test		0.080	07:59
Air Blank		0.000	08:00
Subject Sample	# 1	0.138	08:01
Breath Volume		2.089	08:01
Air Blank		0.000	08:02
Air Blank		0.000	08:04
Subject Sample	# 2	0.132	08:04
Breath Volume		1.738	08:04
Air Blank		0.000	08:05
Control Test		0.079	08:05
Air Blank		0.000	08:06
Diagnostics Check		OK	08:06

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,JAVIER,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: FERNANDEZ,JAVIER
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: PEREIRA

DATE/TIME OF TEST: 02/24/2017 20:29:25

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.080	20:22
Air Blank		0.000	20:22
Subject Sample	# 1	0.172	20:24
Breath Volume		1.449	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:27
Subject Sample	# 2	0.178	20:27
Breath Volume		1.460	20:27
Air Blank		0.000	20:28
Control Test		0.077	20:28
Air Blank		0.000	20:29
Diagnostics Check		OK	20:29

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ACEVEDO,YAMIL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ADER III

DATE/TIME OF TEST: 02/26/2017 02:22:55

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.194	02:10	Volume Not Met
Breath Volume		0.562	02:10	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.224	02:15	
Breath Volume		1.171	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 3	0.182	02:21	Volume Not Met
Breath Volume		0.316	02:21	
Air Blank		0.000	02:21	
Control Test		0.076	02:22	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:22	

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002232

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,CURTIS,L
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: ACEVEDO,YAMIL
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: ADER III

DATE/TIME OF TEST: 02/26/2017 02:36:01

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.077	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.214	02:28
Breath Volume		1.355	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.211	02:34
Breath Volume		1.433	02:34
Air Blank		0.000	02:35
Control Test		0.076	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:36

Data Download Date/Time: 02/28/2017 11:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002462

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTERO,JOSE,
BREATH TEST OPERATOR AGENCY: Hialeah PD
ARRESTING OFFICER: FABIEN,J
ARRESTING OFFICER AGENCY: Hialeah PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: NUNEZ CARDENAS

DATE/TIME OF TEST: 02/26/2017 09:12:50

Diagnostics Check		OK	09:10	
Air Blank		0.000	09:10	
Control Test		0.077	09:10	
Air Blank		0.000	09:11	
Subject Sample	# 1		09:11	Subject Test Refused
Breath Volume		0.000	09:11	
Air Blank		0.000	09:11	
Control Test		0.076	09:12	
Air Blank		0.000	09:12	
Diagnostics Check		OK	09:12	

Data Download Date/Time: 03/23/2017 11:53

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ,NORLAN,
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: GOMEZ,ANDREW
ARRESTING OFFICER AGENCY: Opa Locka PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SOLORZANO

DATE/TIME OF TEST: 02/07/2017 03:37:36

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.076	03:36	
Air Blank		03:37	RFI Detect
Air Blank	0.000	03:37	

Data Download Date/Time: 02/23/2017 12:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002463

INSTRUMENT REGISTERED TO: HIALEAH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENITEZ,NORLAN,
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: GOMEZ,ANDREW
ARRESTING OFFICER AGENCY: Opa Locka PD
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A1
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SOLORZANO

DATE/TIME OF TEST: 02/07/2017 03:43:37

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:40	
Control Test		0.075	03:41	
Air Blank		0.000	03:41	
Subject Sample	# 1		03:42	Subject Test Refused
Breath Volume		0.000	03:42	
Air Blank		0.000	03:42	
Control Test		0.076	03:43	
Air Blank		0.000	03:43	
Diagnostics Check		OK	03:43	

Data Download Date/Time: 02/23/2017 12:33

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 02/03/2017 04:52:54

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.079	04:47
Air Blank		0.000	04:47
Subject Sample	# 1	0.210	04:48
Breath Volume		1.191	04:48
Air Blank		0.000	04:49
Air Blank		0.000	04:50
Subject Sample	# 2	0.199	04:51
Breath Volume		1.777	04:51
Air Blank		0.000	04:51
Control Test		0.078	04:52
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 02/28/2017 11:40

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-002989

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 01/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: JOYNER

DATE/TIME OF TEST: 02/18/2017 03:08:54

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.182	03:04
Breath Volume		1.746	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.186	03:07
Breath Volume		2.152	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:08
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 02/28/2017 11:40

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,LEONARDO,J
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: REYES,LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 02/08/2017 23:11:07

Diagnostics Check	OK	23:10	
Air Blank		23:10	RFI Detect
Air Blank	0.000	23:11	

Data Download Date/Time: 02/28/2017 10:34

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,LEONARDO,J
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: REYES,LEONARDO
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 02/08/2017 23:21:19

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.078	23:15
Air Blank		0.000	23:15
Subject Sample	# 1	0.260	23:16
Breath Volume		1.417	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:18
Subject Sample	# 2	0.274	23:19
Breath Volume		1.750	23:19
Air Blank		0.000	23:20
Control Test		0.078	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 02/28/2017 10:34

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003348

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALLE,OWIN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: HAUGHIE,DARIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1878845
DRY GAS EXPIRY DATE: 06/22/2018
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:32

SUBJECT'S LAST NAME: VALLE

DATE/TIME OF TEST: 02/12/2017 07:16:31

Diagnostics Check		OK	07:13	
Air Blank		0.000	07:14	
Control Test		0.079	07:14	
Air Blank		0.000	07:14	
Subject Sample	# 1		07:15	Subject Test Refused
Breath Volume		0.000	07:15	
Air Blank		0.000	07:15	
Control Test		0.078	07:15	
Air Blank		0.000	07:16	
Diagnostics Check		OK	07:16	

Data Download Date/Time: 02/28/2017 10:34

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CRISWELL

DATE/TIME OF TEST: 02/01/2017 03:39:32

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.204	03:34
Breath Volume		1.804	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.197	03:37
Breath Volume		2.042	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NEIPERT,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:18

SUBJECT'S LAST NAME: MONTANEZ

DATE/TIME OF TEST: 02/01/2017 19:47:36

Diagnostics Check		OK	19:40
Air Blank		0.000	19:40
Control Test		0.079	19:40
Air Blank		0.000	19:41
Subject Sample	# 1	0.235	19:42
Breath Volume		1.277	19:42
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.227	19:45
Breath Volume		1.109	19:45
Air Blank		0.000	19:46
Control Test		0.078	19:46
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: PADILLA HERNANDEZ

DATE/TIME OF TEST: 02/02/2017 01:23:04

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.214	01:17
Breath Volume		1.304	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.214	01:21
Breath Volume		2.113	01:21
Air Blank		0.000	01:22
Control Test		0.077	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 02/02/2017 04:31:55

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.086	04:25
Breath Volume		1.566	04:25
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.088	04:30
Breath Volume		1.949	04:30
Air Blank		0.000	04:30
Control Test		0.078	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: ZALEWSKI JR

DATE/TIME OF TEST: 02/03/2017 00:18:08

Diagnostics Check		OK	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.000	00:13
Breath Volume		3.234	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.000	00:16
Breath Volume		2.222	00:16
Air Blank		0.000	00:17
Control Test		0.077	00:17
Air Blank		0.000	00:18
Diagnostics Check		OK	00:18

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PATTERSON,T
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: PETTUS

DATE/TIME OF TEST: 02/03/2017 15:27:23

Diagnostics Check		OK	15:20
Air Blank		0.000	15:21
Control Test		0.079	15:21
Air Blank		0.000	15:22
Subject Sample	# 1	0.153	15:22
Breath Volume		1.414	15:22
Air Blank		0.000	15:23
Air Blank		0.000	15:24
Subject Sample	# 2	0.164	15:25
Breath Volume		1.902	15:25
Air Blank		0.000	15:26
Control Test		0.081	15:26
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:49

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 02/03/2017 20:16:42

Diagnostics Check		OK	20:10
Air Blank		0.000	20:10
Control Test		0.079	20:10
Air Blank		0.000	20:11
Subject Sample	# 1	0.116	20:11
Breath Volume		1.535	20:11
Air Blank		0.000	20:12
Air Blank		0.000	20:14
Subject Sample	# 2	0.113	20:15
Breath Volume		1.816	20:15
Air Blank		0.000	20:15
Control Test		0.079	20:16
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/04/2017 03:49:48

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.163	03:45
Breath Volume		3.027	03:45
Air Blank		0.000	03:46
Air Blank		0.000	03:47
Subject Sample	# 2	0.163	03:48
Breath Volume		2.175	03:48
Air Blank		0.000	03:48
Control Test		0.077	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAYLOR,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: BOHAN

DATE/TIME OF TEST: 02/04/2017 08:48:06

Diagnostics Check		OK	08:41
Air Blank		0.000	08:41
Control Test		0.079	08:41
Air Blank		0.000	08:42
Subject Sample	# 1	0.175	08:43
Breath Volume		1.125	08:43
Air Blank		0.000	08:44
Air Blank		0.000	08:45
Subject Sample	# 2	0.184	08:46
Breath Volume		1.207	08:46
Air Blank		0.000	08:47
Control Test		0.078	08:47
Air Blank		0.000	08:48
Diagnostics Check		OK	08:48

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SESTO

DATE/TIME OF TEST: 02/05/2017 01:48:17

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:42
Subject Sample	# 1	0.242	01:43
Breath Volume		1.152	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:45
Subject Sample	# 2	0.231	01:46
Breath Volume		1.398	01:46
Air Blank		0.000	01:47
Control Test		0.078	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PERKINS

DATE/TIME OF TEST: 02/05/2017 04:57:51

Diagnostics Check		OK	04:51
Air Blank		0.000	04:51
Control Test		0.079	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.191	04:53
Breath Volume		3.613	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:55
Subject Sample	# 2	0.196	04:56
Breath Volume		2.851	04:56
Air Blank		0.000	04:56
Control Test		0.078	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ALAS SERRAT

DATE/TIME OF TEST: 02/05/2017 19:42:24

Diagnostics Check		OK	19:35
Air Blank		0.000	19:36
Control Test		0.079	19:36
Air Blank		0.000	19:37
Subject Sample	# 1	0.212	19:37
Breath Volume		1.222	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.215	19:40
Breath Volume		2.503	19:40
Air Blank		0.000	19:41
Control Test		0.078	19:41
Air Blank		0.000	19:42
Diagnostics Check		OK	19:42

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: FERNANDEZ PEREZ

DATE/TIME OF TEST: 02/06/2017 23:46:39

Diagnostics Check		OK	23:41	
Air Blank		0.000	23:42	
Control Test		0.079	23:42	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:45	Subject Test Refused
Breath Volume		0.000	23:45	
Air Blank		0.000	23:45	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003387

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: MICALE

DATE/TIME OF TEST: 02/08/2017 04:38:10

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.079	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.099	04:31
Breath Volume		1.117	04:31
Air Blank		0.000	04:32
Air Blank		0.000	04:34
Subject Sample	# 2	0.104	04:36
Breath Volume		1.558	04:36
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 02/08/2017 19:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 02/01/2017 04:01:16

Diagnostics Check		OK	03:54
Air Blank		0.000	03:55
Control Test		0.082	03:55
Air Blank		0.000	03:56
Subject Sample	# 1	0.123	03:56
Breath Volume		2.105	03:56
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.121	03:59
Breath Volume		2.093	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 02/08/2017 19:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 02/01/2017 04:56:00

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:48	
Control Test		0.082	04:48	
Air Blank		0.000	04:48	
Subject Sample	# 1	0.191	04:50	
Breath Volume		1.226	04:50	
Air Blank		0.000	04:50	
Air Blank		0.000	04:52	
Subject Sample	# 2		04:54	Subject Test Refused
Breath Volume		0.816	04:54	
Air Blank		0.000	04:55	
Control Test		0.081	04:55	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:56	

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 02/01/2017 05:04:03

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.081	04:58
Air Blank		0.000	04:59
Subject Sample	# 1	0.227	04:59
Breath Volume		1.289	04:59
Air Blank		0.000	05:00
Air Blank		0.000	05:01
Subject Sample	# 2	0.209	05:02
Breath Volume		1.449	05:02
Air Blank		0.000	05:03
Control Test		0.080	05:03
Air Blank		0.000	05:03
Diagnostics Check		OK	05:04

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BICKEL

DATE/TIME OF TEST: 02/01/2017 19:31:15

Diagnostics Check		OK	19:26	
Air Blank		0.000	19:26	
Control Test		0.082	19:26	
Air Blank		0.000	19:27	
Subject Sample	# 1		19:29	Subject Test Refused
Breath Volume		0.000	19:29	
Air Blank		0.000	19:30	
Control Test		0.082	19:30	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:31	

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: BICKEL

DATE/TIME OF TEST: 02/01/2017 19:42:44

Diagnostics Check		OK	19:34	
Air Blank		0.000	19:34	
Control Test		0.081	19:35	
Air Blank		0.000	19:35	
Subject Sample	# 1	0.000	19:38	Volume Not Met
Breath Volume		0.207	19:38	
Air Blank		0.000	19:39	
Air Blank		0.000	19:41	
Subject Sample	# 2		19:41	Subject Test Refused
Breath Volume		0.000	19:41	
Air Blank		0.000	19:41	
Control Test		0.081	19:42	
Air Blank		0.000	19:42	
Diagnostics Check		OK	19:42	

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/03/2017 00:31:00

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.082	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:30	
Control Test		0.081	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: CHANDLER JR

DATE/TIME OF TEST: 02/04/2017 01:08:59

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.082	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:08	
Control Test		0.082	01:08	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: PEURIFOY

DATE/TIME OF TEST: 02/04/2017 03:54:23

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.082	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.178	03:49
Breath Volume		1.636	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.162	03:52
Breath Volume		1.230	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 02/08/2017 19:16

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: OTERO

DATE/TIME OF TEST: 02/04/2017 05:19:54

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.081	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.113	05:15
Breath Volume		1.191	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:17
Subject Sample	# 2	0.127	05:18
Breath Volume		3.062	05:18
Air Blank		0.000	05:18
Control Test		0.081	05:19
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: BUBEL

DATE/TIME OF TEST: 02/04/2017 23:46:04

Diagnostics Check		OK	23:39
Air Blank		0.000	23:40
Control Test		0.082	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.103	23:41
Breath Volume		4.078	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:43
Subject Sample	# 2	0.103	23:44
Breath Volume		3.371	23:44
Air Blank		0.000	23:45
Control Test		0.082	23:45
Air Blank		0.000	23:45
Diagnostics Check		OK	23:46

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MUNDRA

DATE/TIME OF TEST: 02/05/2017 02:22:52

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.082	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.121	02:18
Breath Volume		2.921	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.122	02:21
Breath Volume		2.632	02:21
Air Blank		0.000	02:21
Control Test		0.081	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAMER,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/05/2017 04:26:13

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.082	04:20
Air Blank		0.000	04:21
Subject Sample	# 1	0.165	04:21
Breath Volume		1.171	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.158	04:24
Breath Volume		1.132	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 02/05/2017 05:06:56

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.082	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.122	05:02
Breath Volume		1.378	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.129	05:05
Breath Volume		1.972	05:05
Air Blank		0.000	05:06
Control Test		0.081	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003388

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/06/2017 04:21:20

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.252	04:16
Breath Volume		1.488	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:18
Subject Sample	# 2	0.246	04:19
Breath Volume		2.382	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 02/08/2017 19:16

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 02/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: HANSEL

DATE/TIME OF TEST: 02/25/2017 02:49:44

Diagnostics Check		OK	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:44
Air Blank		0.000	02:44
Subject Sample	# 1	0.189	02:45
Breath Volume		1.511	02:45
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.196	02:48
Breath Volume		1.707	02:48
Air Blank		0.000	02:48
Control Test		0.082	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 03/21/2017 11:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003406

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSS, SR.,MICHAEL,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CROSS, SR.,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 02/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: FENDER

DATE/TIME OF TEST: 02/25/2017 18:37:06

Diagnostics Check		OK	18:30
Air Blank		0.000	18:31
Control Test		0.083	18:31
Air Blank		0.000	18:32
Subject Sample	# 1	0.000	18:32
Breath Volume		1.562	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:35
Subject Sample	# 2	0.000	18:35
Breath Volume		1.496	18:35
Air Blank		0.000	18:36
Control Test		0.080	18:36
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 03/21/2017 11:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003408

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTERO,DANEIL,S
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SEGUI,YOSVANY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: Y
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 13:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:40

SUBJECT'S LAST NAME: WILKOTZ

DATE/TIME OF TEST: 02/27/2017 15:14:07

Diagnostics Check		OK	15:07
Air Blank		0.000	15:08
Control Test		0.081	15:08
Air Blank		0.000	15:09
Subject Sample	# 1	0.235	15:09
Breath Volume		1.906	15:09
Air Blank		0.000	15:10
Air Blank		0.000	15:12
Subject Sample	# 2	0.231	15:12
Breath Volume		1.781	15:12
Air Blank		0.000	15:13
Control Test		0.081	15:13
Air Blank		0.000	15:14
Diagnostics Check		OK	15:14

Data Download Date/Time: 02/28/2017 10:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003410

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINERA,ORESTE,I
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROBINSON,IVAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:07

SUBJECT'S LAST NAME: SHATTUCK

DATE/TIME OF TEST: 02/26/2017 17:38:05

Diagnostics Check		OK	17:28
Air Blank		0.000	17:28
Control Test		0.081	17:29
Air Blank		0.000	17:29
Subject Sample	# 1	0.195	17:30
Breath Volume		2.347	17:30
Air Blank		0.000	17:31
Air Blank		0.000	17:32
Subject Sample	# 2	0.163	17:33
Breath Volume		1.652	17:33
Air Blank		0.000	17:34
Air Blank		0.000	17:35
Subject Sample	# 3	0.191	17:36
Breath Volume		2.496	17:36
Air Blank		0.000	17:37
Control Test		0.082	17:37
Air Blank		0.000	17:38
Diagnostics Check		OK	17:38

Data Download Date/Time: 04/01/2017 00:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: PLAZA

DATE/TIME OF TEST: 02/04/2017 03:43:45

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Subject Sample	# 1	0.192	03:39
Breath Volume		2.382	03:39
Air Blank		0.000	03:39
Air Blank		0.000	03:41
Subject Sample	# 2	0.196	03:42
Breath Volume		2.328	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 02/27/2017 23:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 02/05/2017 02:45:50

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.147	02:41
Breath Volume		2.664	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.140	02:44
Breath Volume		2.425	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 02/27/2017 23:58

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003411

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,JOSHUA,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAMPBELL,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/09/2017 22:51:19

Diagnostics Check		OK	22:45
Air Blank		0.000	22:45
Control Test		0.080	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.000	22:47
Breath Volume		1.109	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.000	22:49
Breath Volume		1.289	22:49
Air Blank		0.000	22:50
Control Test		0.080	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 02/27/2017 23:58

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: POPOCA

DATE/TIME OF TEST: 02/04/2017 04:13:12

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.078	04:09	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:11	Subject Test Refused
Breath Volume		0.289	04:11	
Air Blank		0.000	04:12	
Control Test		0.078	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: SOCALL

DATE/TIME OF TEST: 02/05/2017 03:11:12

Diagnostics Check		OK	03:05
Air Blank		0.000	03:05
Control Test		0.080	03:06
Air Blank		0.000	03:06
Subject Sample	# 1	0.152	03:06
Breath Volume		2.578	03:06
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.154	03:09
Breath Volume		2.585	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:11
Diagnostics Check		OK	03:11

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NELSEN,G
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/07/2017 04:18:25

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.079	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.182	04:14
Breath Volume		1.371	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.183	04:16
Breath Volume		1.371	04:16
Air Blank		0.000	04:17
Control Test		0.079	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURHAM,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DURHAM,JAMES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:37

SUBJECT'S LAST NAME: BOWMAN

DATE/TIME OF TEST: 02/07/2017 11:03:45

Diagnostics Check		OK	10:57
Air Blank		0.000	10:58
Control Test		0.079	10:58
Air Blank		0.000	10:58
Subject Sample	# 1	0.052	10:59
Breath Volume		1.847	10:59
Air Blank		0.000	10:59
Air Blank		0.000	11:01
Subject Sample	# 2	0.052	11:02
Breath Volume		2.042	11:02
Air Blank		0.000	11:02
Control Test		0.079	11:03
Air Blank		0.000	11:03
Diagnostics Check		OK	11:03

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:46

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 02/10/2017 22:15:47

Diagnostics Check		OK	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:10
Subject Sample	# 1	0.243	22:11
Breath Volume		3.355	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:13
Subject Sample	# 2	0.242	22:14
Breath Volume		3.933	22:14
Air Blank		0.000	22:14
Control Test		0.078	22:15
Air Blank		0.000	22:15
Diagnostics Check		OK	22:15

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,ANNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: LEON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/12/2017 04:04:33

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.000	04:00
Breath Volume		1.734	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.000	04:03
Breath Volume		1.746	04:03
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 02/12/2017 21:20:44

Diagnostics Check		OK	21:14
Air Blank		0.000	21:15
Control Test		0.080	21:15
Air Blank		0.000	21:15
Subject Sample	# 1	0.225	21:16
Breath Volume		4.167	21:16
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.222	21:19
Breath Volume		3.847	21:19
Air Blank		0.000	21:19
Control Test		0.079	21:20
Air Blank		0.000	21:20
Diagnostics Check		OK	21:20

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: BOSWELL

DATE/TIME OF TEST: 02/18/2017 01:40:41

Diagnostics Check		OK	01:34
Air Blank		0.000	01:35
Control Test		0.080	01:35
Air Blank		0.000	01:35
Subject Sample	# 1	0.204	01:36
Breath Volume		4.187	01:36
Air Blank		0.000	01:37
Air Blank		0.000	01:38
Subject Sample	# 2	0.192	01:39
Breath Volume		2.824	01:39
Air Blank		0.000	01:39
Control Test		0.078	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:01

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 02/20/2017 17:28:22

Diagnostics Check		OK	17:22
Air Blank		0.000	17:22
Control Test		0.080	17:23
Air Blank		0.000	17:23
Subject Sample	# 1	0.000	17:24
Breath Volume		4.675	17:24
Air Blank		0.000	17:24
Air Blank		0.000	17:26
Subject Sample	# 2	0.000	17:27
Breath Volume		4.847	17:27
Air Blank		0.000	17:27
Control Test		0.080	17:27
Air Blank		0.000	17:28
Diagnostics Check		OK	17:28

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: QUERNEY

DATE/TIME OF TEST: 02/24/2017 02:07:03

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:03	
Control Test		0.079	02:03	
Air Blank		0.000	02:04	
Subject Sample	# 1	0.192	02:05	
Breath Volume		3.136	02:05	
Air Blank			02:05	RFI Detect
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: QUERNEY

DATE/TIME OF TEST: 02/24/2017 02:15:14

Diagnostics Check		OK	02:08
Air Blank		0.000	02:09
Control Test		0.079	02:09
Air Blank		0.000	02:10
Subject Sample	# 1	0.187	02:10
Breath Volume		2.582	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.185	02:13
Breath Volume		2.117	02:13
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHIPPLE,MATTHEW,A
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: WHIPPLE,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: PHAM

DATE/TIME OF TEST: 02/24/2017 22:17:24

Diagnostics Check		OK	22:10
Air Blank		0.000	22:11
Control Test		0.080	22:11
Air Blank		0.000	22:12
Subject Sample	# 1	0.167	22:12
Breath Volume		2.734	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:15
Subject Sample	# 2	0.165	22:15
Breath Volume		2.042	22:15
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 03/01/2017 15:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-003413

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: MCKAY

DATE/TIME OF TEST: 02/28/2017 00:14:31

Diagnostics Check		OK	00:08
Air Blank		0.000	00:08
Control Test		0.080	00:09
Air Blank		0.000	00:09
Subject Sample	# 1	0.056	00:10
Breath Volume		2.113	00:10
Air Blank		0.000	00:10
Air Blank		0.000	00:12
Subject Sample	# 2	0.047	00:13
Breath Volume		1.296	00:13
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 03/01/2017 15:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-004587

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 01/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 02/15/2017 00:30:03

Diagnostics Check		OK	00:22
Air Blank		0.000	00:22
Control Test		0.078	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.196	00:24
Breath Volume		1.832	00:24
Air Blank		0.000	00:25
Air Blank		0.000	00:27
Subject Sample	# 2	0.206	00:28
Breath Volume		2.167	00:28
Air Blank		0.000	00:29
Control Test		0.077	00:29
Air Blank		0.000	00:29
Diagnostics Check		OK	00:30

Data Download Date/Time: 02/18/2017 06:56

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SARMENT,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: CIVEY

DATE/TIME OF TEST: 02/05/2017 21:13:48

Diagnostics Check		OK	21:01	
Air Blank		0.000	21:02	
Control Test		0.077	21:02	
Air Blank		0.000	21:03	
Subject Sample	# 1		21:06	No Sample Provided
Breath Volume		0.000	21:06	
Air Blank		0.000	21:06	
Air Blank		0.000	21:08	
Subject Sample	# 2	0.055	21:09	
Breath Volume		2.496	21:09	
Air Blank		0.000	21:09	
Air Blank		0.000	21:11	
Subject Sample	# 3	0.054	21:12	
Breath Volume		3.281	21:12	
Air Blank		0.000	21:12	
Control Test		0.076	21:13	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 02/28/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SARMENT,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SARMENT,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: CIVEY

DATE/TIME OF TEST: 02/05/2017 21:22:45

Diagnostics Check		OK	21:16
Air Blank		0.000	21:17
Control Test		0.076	21:17
Air Blank		0.000	21:17
Subject Sample	# 1	0.052	21:18
Breath Volume		3.371	21:18
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.051	21:21
Breath Volume		3.480	21:21
Air Blank		0.000	21:21
Control Test		0.076	21:22
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 02/28/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOBLES,DARIUS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: NOBLES,DARIUS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 02/07/2017 22:18:50

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:16	
Control Test		0.077	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 1		22:17	Subject Test Refused
Breath Volume		0.000	22:17	
Air Blank		0.000	22:17	
Control Test		0.077	22:18	
Air Blank		0.000	22:18	
Diagnostics Check		OK	22:18	

Data Download Date/Time: 02/28/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,GLENN,H
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAND,GLENN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:38

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 02/16/2017 11:06:16

Diagnostics Check		OK	10:59
Air Blank		0.000	10:59
Control Test		0.078	11:00
Air Blank		0.000	11:00
Subject Sample	# 1	0.000	11:01
Breath Volume		5.007	11:01
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.000	11:04
Breath Volume		4.371	11:04
Air Blank		0.000	11:05
Control Test		0.077	11:05
Air Blank		0.000	11:06
Diagnostics Check		OK	11:06

Data Download Date/Time: 02/28/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEL,JESSICA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ROBINSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: MIHELICH

DATE/TIME OF TEST: 02/18/2017 01:40:42

Diagnostics Check		OK	01:34
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.193	01:36
Breath Volume		3.152	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.185	01:39
Breath Volume		3.164	01:39
Air Blank		0.000	01:39
Control Test		0.077	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 02/28/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005057

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SETH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILLIAMS,SETH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:52

SUBJECT'S LAST NAME: GUY

DATE/TIME OF TEST: 02/24/2017 10:19:28

Diagnostics Check		OK	10:12
Air Blank		0.000	10:13
Control Test		0.076	10:13
Air Blank		0.000	10:13
Subject Sample	# 1	0.154	10:14
Breath Volume		2.167	10:14
Air Blank		0.000	10:15
Air Blank		0.000	10:17
Subject Sample	# 2	0.156	10:17
Breath Volume		2.273	10:17
Air Blank		0.000	10:18
Control Test		0.076	10:18
Air Blank		0.000	10:19
Diagnostics Check		OK	10:19

Data Download Date/Time: 02/28/2017 17:44

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEROCHE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: LONDON,CHRISTIE
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:16

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 02/03/2017 06:43:13

Diagnostics Check		OK	06:36
Air Blank		0.000	06:36
Control Test		0.077	06:37
Air Blank		0.000	06:37
Subject Sample	# 1	0.254	06:38
Breath Volume		3.128	06:38
Air Blank		0.000	06:39
Air Blank		0.000	06:40
Subject Sample	# 2	0.257	06:41
Breath Volume		3.253	06:41
Air Blank		0.000	06:42
Control Test		0.076	06:42
Air Blank		0.000	06:43
Diagnostics Check		OK	06:43

Data Download Date/Time: 02/28/2017 17:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005058

INSTRUMENT REGISTERED TO: CRESTVIEW PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANDREW,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Crestview PD
ARRESTING OFFICER: ANDREW,MATTHEW
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 01/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 02/05/2017 04:29:33

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.078	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.161	04:24
Breath Volume		1.382	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.166	04:27
Breath Volume		1.578	04:27
Air Blank		0.000	04:28
Control Test		0.077	04:28
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 02/28/2017 17:49

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ADAMS,JIMMY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: BULANDA

DATE/TIME OF TEST: 02/15/2017 00:03:07

Diagnostics Check		OK	23:54
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:55
Subject Sample	# 2	0.271	23:58
Breath Volume		2.722	23:58
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.276	00:01
Breath Volume		2.703	00:01
Air Blank		0.000	00:02
Control Test		0.077	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 02/27/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005076

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: ABROMAITIS,MICHAEL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: ZAPPALA

DATE/TIME OF TEST: 02/18/2017 02:34:42

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.000	02:30
Breath Volume		1.425	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.000	02:33
Breath Volume		1.371	02:33
Air Blank		0.000	02:33
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 02/27/2017 08:46

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: JOSEPH,SHIRLEY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/07/2017 11:03:49

Diagnostics Check		OK	10:56	
Air Blank		0.000	10:56	
Control Test		0.078	10:57	
Air Blank		0.000	10:57	
Subject Sample	# 1	0.340	10:59	
Breath Volume		1.429	10:59	
Air Blank		0.000	11:00	
Air Blank		0.000	11:02	
			11:02	RFI Detect
Air Blank		0.000	11:02	
Control Test		0.079	11:03	
Air Blank		0.000	11:03	
Diagnostics Check		OK	11:03	

Data Download Date/Time: 02/14/2017 11:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005077

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: JOSEPH,SHIRLEY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06415080A2
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 09:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 02/07/2017 11:14:58

Diagnostics Check		OK	11:09
Air Blank		0.000	11:09
Control Test		0.077	11:09
Air Blank		0.000	11:10
Subject Sample	# 1	0.355	11:10
Breath Volume		1.289	11:10
Air Blank		0.000	11:11
Air Blank		0.000	11:13
Subject Sample	# 2	0.355	11:13
Breath Volume		1.351	11:13
Air Blank		0.000	11:14
Control Test		0.077	11:14
Air Blank		0.000	11:14
Diagnostics Check		OK	11:14

Data Download Date/Time: 02/14/2017 11:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: SHARP

DATE/TIME OF TEST: 02/09/2017 01:59:56

Diagnostics Check		OK	01:47	
Air Blank		0.000	01:47	
Control Test		0.080	01:48	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:51	No Sample Provided
Breath Volume		0.000	01:51	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 2	0.248	01:55	
Breath Volume		1.339	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 3	0.255	01:58	
Breath Volume		1.589	01:58	
Air Blank		0.000	01:59	
Control Test		0.079	01:59	
Air Blank		0.000	01:59	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: KURLAND

DATE/TIME OF TEST: 02/10/2017 00:46:10

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.190	00:41
Breath Volume		2.234	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.184	00:44
Breath Volume		1.781	00:44
Air Blank		0.000	00:45
Control Test		0.079	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 02/10/2017 23:07:18

Diagnostics Check		OK	23:00
Air Blank		0.000	23:01
Control Test		0.080	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.154	23:02
Breath Volume		1.312	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.156	23:05
Breath Volume		1.234	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MORAN OYUELA

DATE/TIME OF TEST: 02/11/2017 01:16:24

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.164	01:11
Breath Volume		1.824	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.161	01:14
Breath Volume		1.753	01:14
Air Blank		0.000	01:15
Control Test		0.079	01:15
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/11/2017 04:42:09

Diagnostics Check		OK	04:35
Air Blank		0.000	04:36
Control Test		0.080	04:36
Air Blank		0.000	04:36
Subject Sample	# 1	0.145	04:37
Breath Volume		2.195	04:37
Air Blank		0.000	04:38
Air Blank		0.000	04:40
Subject Sample	# 2	0.153	04:40
Breath Volume		3.144	04:40
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:42
Diagnostics Check		OK	04:42

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/12/2017 02:42:35

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Subject Sample	# 1	0.057	02:37
Breath Volume		1.726	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:40
Subject Sample	# 2	0.057	02:41
Breath Volume		1.644	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: YASSO

DATE/TIME OF TEST: 02/12/2017 03:44:45

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.079	03:39
Air Blank		0.000	03:39
Subject Sample	# 1	0.145	03:40
Breath Volume		2.566	03:40
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.144	03:43
Breath Volume		2.246	03:43
Air Blank		0.000	03:43
Control Test		0.079	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 02/12/2017 06:23:18

Diagnostics Check		OK	06:15
Air Blank		0.000	06:16
Control Test		0.079	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.031	06:18
Breath Volume		2.804	06:18
Air Blank		0.000	06:19
Air Blank		0.000	06:21
Subject Sample	# 2	0.031	06:21
Breath Volume		3.027	06:21
Air Blank		0.000	06:22
Control Test		0.079	06:22
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MONTEJO-PEREZ

DATE/TIME OF TEST: 02/12/2017 22:35:18

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.254	22:30
Breath Volume		1.230	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:32
Subject Sample	# 2	0.255	22:33
Breath Volume		2.855	22:33
Air Blank		0.000	22:34
Control Test		0.079	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 03/16/2017 19:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: ARSENAULT

DATE/TIME OF TEST: 02/13/2017 05:37:11

Diagnostics Check		OK	05:30
Air Blank		0.000	05:31
Control Test		0.079	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.159	05:32
Breath Volume		3.046	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.157	05:35
Breath Volume		3.289	05:35
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: PABON ROSADO

DATE/TIME OF TEST: 02/15/2017 03:10:52

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.079	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:03	No Sample Provided
Breath Volume		0.000	03:03	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:09	No Sample Provided
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.079	03:10	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAMER,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: DUNN

DATE/TIME OF TEST: 02/15/2017 03:48:46

Diagnostics Check		OK	03:42
Air Blank		0.000	03:42
Control Test		0.079	03:43
Air Blank		0.000	03:43
Subject Sample	# 1	0.098	03:44
Breath Volume		3.371	03:44
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.091	03:47
Breath Volume		1.769	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 02/18/2017 00:47:14

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.148	00:40
Breath Volume		1.542	00:40
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.151	00:45
Breath Volume		2.039	00:45
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: SUAREZ

DATE/TIME OF TEST: 02/18/2017 01:59:04

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:54	
Control Test		0.080	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:58	
Control Test		0.080	01:58	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:59	

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: MACLIN

DATE/TIME OF TEST: 02/19/2017 03:29:28

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.079	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:22	No Sample Provided
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Air Blank		0.000	03:24	
Subject Sample	# 2		03:28	No Sample Provided
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:28	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: INDUSTRIOSO

DATE/TIME OF TEST: 02/19/2017 04:28:48

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.080	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.147	04:24
Breath Volume		1.820	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.148	04:27
Breath Volume		1.984	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAMER,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: APARICIO

DATE/TIME OF TEST: 02/19/2017 05:15:20

Diagnostics Check		OK	05:08
Air Blank		0.000	05:09
Control Test		0.080	05:09
Air Blank		0.000	05:10
Subject Sample	# 1	0.123	05:10
Breath Volume		2.339	05:10
Air Blank		0.000	05:11
Air Blank		0.000	05:13
Subject Sample	# 2	0.125	05:13
Breath Volume		1.667	05:13
Air Blank		0.000	05:14
Control Test		0.079	05:14
Air Blank		0.000	05:15
Diagnostics Check		OK	05:15

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: BARNETT II

DATE/TIME OF TEST: 02/22/2017 04:48:54

Diagnostics Check		OK	04:41
Air Blank		0.000	04:42
Control Test		0.080	04:42
Air Blank		0.000	04:43
Subject Sample	# 1	0.205	04:43
Breath Volume		1.187	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.204	04:47
Breath Volume		2.035	04:47
Air Blank		0.000	04:47
Control Test		0.079	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: RUSSELL,K
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 02/24/2017 22:47:10

Diagnostics Check		OK	22:39
Air Blank		0.000	22:39
Control Test		0.079	22:39
Air Blank		0.000	22:40
Subject Sample	# 1	0.016	22:42
Breath Volume		1.972	22:42
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.015	22:45
Breath Volume		1.812	22:45
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: HERNANDEZ-HERNANDEZ

DATE/TIME OF TEST: 02/24/2017 23:44:23

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.080	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.162	23:39
Breath Volume		2.402	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.171	23:42
Breath Volume		3.000	23:42
Air Blank		0.000	23:43
Control Test		0.079	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BECERRA

DATE/TIME OF TEST: 02/25/2017 00:30:59

Diagnostics Check		OK	00:22
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:23
Subject Sample	# 1	0.169	00:26
Breath Volume		1.949	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:28
Subject Sample	# 2	0.179	00:29
Breath Volume		1.386	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:30

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HEIST,A
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 02/25/2017 03:00:50

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.079	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.177	02:56
Breath Volume		2.597	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.177	02:59
Breath Volume		1.535	02:59
Air Blank		0.000	02:59
Control Test		0.079	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: ERICKSON

DATE/TIME OF TEST: 02/25/2017 03:29:38

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.160	03:25
Breath Volume		1.492	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.171	03:28
Breath Volume		1.843	03:28
Air Blank		0.000	03:28
Control Test		0.080	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: VAN BUREN JR

DATE/TIME OF TEST: 02/26/2017 03:02:51

Diagnostics Check		OK	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.149	02:58
Breath Volume		4.472	02:58
Air Blank		0.000	02:58
Air Blank		0.000	03:00
Subject Sample	# 2	0.154	03:01
Breath Volume		4.558	03:01
Air Blank		0.000	03:01
Control Test		0.078	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: O'NEIL,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/26/2017 03:11:00

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.168	03:06
Breath Volume		1.359	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.177	03:09
Breath Volume		1.640	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:11

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/27/2017 01:14:46

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.079	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.209	01:10
Breath Volume		1.968	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.200	01:13
Breath Volume		2.152	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: COMMANDER

DATE/TIME OF TEST: 02/27/2017 04:31:06

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.080	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.036	04:26
Breath Volume		1.332	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.038	04:29
Breath Volume		1.738	04:29
Air Blank		0.000	04:30
Control Test		0.080	04:30
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: REDDING

DATE/TIME OF TEST: 02/27/2017 21:01:16

Diagnostics Check		OK	20:49	
Air Blank		0.000	20:49	
Control Test		0.079	20:50	
Air Blank		0.000	20:50	
Subject Sample	# 1	0.170	20:53	Volume Not Met
Breath Volume		0.414	20:53	
Air Blank		0.000	20:54	
Air Blank		0.000	20:56	
Subject Sample	# 2	0.170	20:59	Volume Not Met
Breath Volume		0.519	20:59	
Air Blank		0.000	21:00	
Control Test		0.081	21:00	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005113

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: REDDING

DATE/TIME OF TEST: 02/27/2017 21:15:03

Diagnostics Check		OK	21:03	
Air Blank		0.000	21:03	
Control Test		0.079	21:03	
Air Blank		0.000	21:04	
Subject Sample	# 1	0.151	21:07	Volume Not Met
Breath Volume		0.292	21:07	
Air Blank		0.000	21:08	
Air Blank		0.000	21:10	
Subject Sample	# 2	0.189	21:13	Volume Not Met
Breath Volume		0.253	21:13	
Air Blank		0.000	21:14	
Control Test		0.081	21:14	
Air Blank		0.000	21:14	
Diagnostics Check		OK	21:15	

Data Download Date/Time: 03/16/2017 19:59

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005248

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYNAUD,GEORGE,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: FARINAS,FRANK
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1757137
DRY GAS EXPIRY DATE: 10/15/2017
VIOLATION CODE: BUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: MARTINEZ SALGADO

DATE/TIME OF TEST: 02/12/2017 17:11:43

Diagnostics Check		OK	16:58	
Air Blank		0.000	16:59	
Control Test		0.080	16:59	
Air Blank		0.000	17:00	
Subject Sample	# 1	0.139	17:01	
Breath Volume		1.125	17:01	
Air Blank		0.000	17:02	
Air Blank		0.000	17:04	
Subject Sample	# 2	0.135	17:07	Slope Not Level
Breath Volume		1.140	17:07	
Air Blank		0.000	17:08	
Air Blank		0.000	17:09	
Subject Sample	# 3	0.130	17:10	
Breath Volume		1.527	17:10	
Air Blank		0.000	17:10	
Control Test		0.071	17:11	Control Outside Tolerance
Air Blank		0.000	17:11	

Data Download Date/Time: 02/15/2017 14:43

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RANDALL,EDMOND
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: STUEVE

DATE/TIME OF TEST: 02/01/2017 01:07:32

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.083	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.193	01:03
Breath Volume		2.300	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.188	01:05
Breath Volume		2.042	01:05
Air Blank		0.000	01:06
Control Test		0.081	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SARNAGO

DATE/TIME OF TEST: 02/02/2017 01:08:44

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.082	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.000	01:04
Breath Volume		2.335	01:04
Air Blank		0.000	01:04
Air Blank		0.000	01:06
Subject Sample	# 2	0.000	01:07
Breath Volume		2.523	01:07
Air Blank		0.000	01:07
Control Test		0.081	01:08
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMLIN,TODD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: PEREZ JR

DATE/TIME OF TEST: 02/02/2017 07:46:15

Diagnostics Check		OK	07:42	
Air Blank		0.000	07:43	
Control Test		0.083	07:43	
Air Blank		0.000	07:44	
Subject Sample	# 1	0.177	07:44	
Breath Volume		1.859	07:44	
Air Blank			07:44	RFI Detect
Air Blank		0.000	07:45	
Control Test		0.082	07:45	
Air Blank		0.000	07:46	
Diagnostics Check		OK	07:46	

Data Download Date/Time: 02/03/2017 00:09

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMLIN,TODD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 01/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 06:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:22

SUBJECT'S LAST NAME: PEREZ JR

DATE/TIME OF TEST: 02/02/2017 07:49:13

Diagnostics Check	OK	07:48	
Air Blank		07:48	RFI Detect
Air Blank	0.000	07:49	

Data Download Date/Time: 02/03/2017 00:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: KUFFER

DATE/TIME OF TEST: 02/03/2017 01:06:31

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.070	01:02
Breath Volume		1.648	01:02
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.069	01:04
Breath Volume		1.734	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 02/03/2017 02:59:28

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:57	
Control Test		0.081	02:57	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:58	Subject Test Refused
Breath Volume		0.000	02:58	
Air Blank		0.000	02:58	
Control Test		0.081	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LIPSHAW

DATE/TIME OF TEST: 02/03/2017 04:30:10

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.081	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.215	04:25
Breath Volume		2.054	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.214	04:28
Breath Volume		1.894	04:28
Air Blank		0.000	04:29
Control Test		0.081	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SAVIC

DATE/TIME OF TEST: 02/03/2017 05:18:46

Diagnostics Check		OK	05:15	
Air Blank		0.000	05:16	
Control Test		0.081	05:16	
Air Blank		0.000	05:16	
Subject Sample	# 1		05:17	Subject Test Refused
Breath Volume		0.000	05:17	
Air Blank		0.000	05:17	
Control Test		0.081	05:18	
Air Blank		0.000	05:18	
Diagnostics Check		OK	05:18	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: COY

DATE/TIME OF TEST: 02/03/2017 21:36:29

Diagnostics Check		OK	21:25	
Air Blank		0.000	21:25	
Control Test		0.081	21:25	
Air Blank		0.000	21:26	
Subject Sample	# 1		21:29	No Sample Provided
Breath Volume		0.000	21:29	
Air Blank		0.000	21:29	
Air Blank		0.000	21:31	
Subject Sample	# 2		21:35	No Sample Provided
Breath Volume		0.000	21:35	
Air Blank		0.000	21:35	
Control Test		0.082	21:35	
Air Blank		0.000	21:36	
Diagnostics Check		OK	21:36	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CAREY,K
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BANK

DATE/TIME OF TEST: 02/04/2017 00:28:39

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.083	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.054	00:23
Breath Volume		1.941	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.052	00:27
Breath Volume		1.925	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:28
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,JONATHAN,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: RICOTTILLI,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 02/04/2017 02:32:52

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.000	02:28
Breath Volume		2.230	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.000	02:31
Breath Volume		2.445	02:31
Air Blank		0.000	02:31
Control Test		0.081	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 02/05/2017 00:57:23

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.082	00:55	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.082	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUARTE,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DUARTE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/05/2017 04:30:30

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.081	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:29	Subject Test Refused
Breath Volume		0.000	04:29	
Air Blank		0.000	04:29	
Control Test		0.081	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 16:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MUNN

DATE/TIME OF TEST: 02/05/2017 17:50:42

Diagnostics Check		OK	17:42
Air Blank		0.000	17:42
Control Test		0.082	17:42
Air Blank		0.000	17:43
Subject Sample	# 1	0.132	17:46
Breath Volume		2.023	17:46
Air Blank		0.000	17:46
Air Blank		0.000	17:48
Subject Sample	# 2	0.124	17:49
Breath Volume		1.773	17:49
Air Blank		0.000	17:49
Control Test		0.079	17:50
Air Blank		0.000	17:50
Diagnostics Check		OK	17:50

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BICKFORD,B,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MCMILLAN,P
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:44

SUBJECT'S LAST NAME: FREEMAN

DATE/TIME OF TEST: 02/05/2017 19:07:05

Diagnostics Check		OK	19:04	
Air Blank		0.000	19:04	
Control Test		0.081	19:04	
Air Blank		0.000	19:05	
Subject Sample	# 1		19:05	Subject Test Refused
Breath Volume		0.000	19:05	
Air Blank		0.000	19:06	
Control Test		0.081	19:06	
Air Blank		0.000	19:07	
Diagnostics Check		OK	19:07	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: KONFORT

DATE/TIME OF TEST: 02/09/2017 01:56:42

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.082	01:54	
Air Blank		0.000	01:55	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ERNST

DATE/TIME OF TEST: 02/10/2017 04:46:25

Diagnostics Check		OK	04:38
Air Blank		0.000	04:39
Control Test		0.083	04:39
Air Blank		0.000	04:40
Subject Sample	# 1	0.130	04:41
Breath Volume		1.273	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.134	04:44
Breath Volume		2.070	04:44
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: WHITESIDE

DATE/TIME OF TEST: 02/10/2017 05:43:50

Diagnostics Check		OK	05:37
Air Blank		0.000	05:37
Control Test		0.082	05:38
Air Blank		0.000	05:38
Subject Sample	# 1	0.146	05:39
Breath Volume		1.101	05:39
Air Blank		0.000	05:40
Air Blank		0.000	05:41
Subject Sample	# 2	0.156	05:42
Breath Volume		1.300	05:42
Air Blank		0.000	05:42
Control Test		0.081	05:43
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: GERARDO URIBE

DATE/TIME OF TEST: 02/11/2017 01:52:18

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.083	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.147	01:47
Breath Volume		2.152	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.147	01:50
Breath Volume		1.761	01:50
Air Blank		0.000	01:51
Control Test		0.082	01:51
Air Blank		0.000	01:52
Diagnostics Check		OK	01:52

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: PEREZ-RODRIUEZ

DATE/TIME OF TEST: 02/11/2017 20:34:25

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.082	20:28
Air Blank		0.000	20:29
Subject Sample	# 1	0.237	20:29
Breath Volume		2.277	20:29
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.237	20:32
Breath Volume		1.582	20:32
Air Blank		0.000	20:33
Control Test		0.082	20:33
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUARTE,CHRISTOPHER,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DUARTE,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: DELOWREY

DATE/TIME OF TEST: 02/13/2017 00:31:46

Diagnostics Check		OK	00:28	
Air Blank		0.000	00:29	
Control Test		0.082	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 1		00:30	Subject Test Refused
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Control Test		0.082	00:31	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DONOHUE,STEVEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: MACCHIA

DATE/TIME OF TEST: 02/14/2017 04:14:40

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.082	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.156	04:10
Breath Volume		2.261	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.160	04:13
Breath Volume		1.726	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 16:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:06

SUBJECT'S LAST NAME: MOON JR

DATE/TIME OF TEST: 02/14/2017 17:31:14

Diagnostics Check		OK	17:28	
Air Blank		0.000	17:28	
Control Test		0.082	17:29	
Air Blank		0.000	17:29	
Subject Sample	# 1		17:29	Subject Test Refused
Breath Volume		0.000	17:29	
Air Blank		0.000	17:30	
Control Test		0.082	17:30	
Air Blank		0.000	17:31	
Diagnostics Check		OK	17:31	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOOD,COREY,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HOOD,COREY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GOOCH

DATE/TIME OF TEST: 02/15/2017 02:03:07

Diagnostics Check		OK	01:55
Air Blank		0.000	01:56
Control Test		0.082	01:56
Air Blank		0.000	01:57
Subject Sample	# 1	0.111	01:58
Breath Volume		2.058	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.107	02:01
Breath Volume		2.457	02:01
Air Blank		0.000	02:02
Control Test		0.081	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILEY,MELODY,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MORAN,SEAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: WILEY

DATE/TIME OF TEST: 02/15/2017 03:14:56

Diagnostics Check		OK	03:12	
Air Blank		0.000	03:12	
Control Test		0.082	03:12	
Air Blank		0.000	03:13	
Subject Sample	# 1		03:13	Subject Test Refused
Breath Volume		0.000	03:13	
Air Blank		0.000	03:14	
Control Test		0.082	03:14	
Air Blank		0.000	03:14	
Diagnostics Check		OK	03:14	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOVAK,RUSS,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: NOVAK,RUSS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HOLZHAUER

DATE/TIME OF TEST: 02/15/2017 04:29:41

Diagnostics Check		OK	04:26	
Air Blank		0.000	04:27	
Control Test		0.082	04:27	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:28	Subject Test Refused
Breath Volume		0.000	04:28	
Air Blank		0.000	04:28	
Control Test		0.081	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: MOSCARELLI

DATE/TIME OF TEST: 02/16/2017 00:32:13

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.082	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.000	00:27
Breath Volume		1.250	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.000	00:30
Breath Volume		1.515	00:30
Air Blank		0.000	00:31
Control Test		0.082	00:31
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: VRANICA

DATE/TIME OF TEST: 02/16/2017 02:27:21

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.082	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.000	02:23
Breath Volume		2.785	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.000	02:25
Breath Volume		3.007	02:25
Air Blank		0.000	02:26
Control Test		0.082	02:26
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 03/19/2017 00:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 02/16/2017 20:43:48

Diagnostics Check		OK	20:40	
Air Blank		0.000	20:41	
Control Test		0.082	20:41	
Air Blank		0.000	20:41	
Subject Sample	# 1		20:42	Subject Test Refused
Breath Volume		0.000	20:42	
Air Blank		0.000	20:42	
Control Test		0.081	20:43	
Air Blank		0.000	20:43	
Diagnostics Check		OK	20:43	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 02/17/2017 02:38:44

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:35	
Control Test		0.082	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:37	RFI Detect
Breath Volume		0.000	02:37	
Air Blank		0.000	02:37	
Control Test		0.081	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 02/17/2017 02:42:50

Diagnostics Check		OK	02:39	
Air Blank		0.000	02:40	
Control Test		0.081	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.081	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSADU,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/17/2017 03:25:19

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.081	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.154	03:19
Breath Volume		1.968	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.149	03:23
Breath Volume		1.554	03:23
Air Blank		0.000	03:24
Control Test		0.081	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: DAY

DATE/TIME OF TEST: 02/17/2017 04:10:51

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.081	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1		04:09	Subject Test Refused
Breath Volume		0.000	04:09	
Air Blank		0.000	04:09	
Control Test		0.081	04:10	
Air Blank		0.000	04:10	
Diagnostics Check		OK	04:10	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: JACK

DATE/TIME OF TEST: 02/17/2017 18:53:41

Diagnostics Check		OK	18:47
Air Blank		0.000	18:47
Control Test		0.083	18:48
Air Blank		0.000	18:48
Subject Sample	# 1	0.000	18:49
Breath Volume		4.593	18:49
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.000	18:52
Breath Volume		2.910	18:52
Air Blank		0.000	18:52
Control Test		0.082	18:53
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: EDMUNDS

DATE/TIME OF TEST: 02/17/2017 19:39:53

Diagnostics Check		OK	19:33
Air Blank		0.000	19:33
Control Test		0.082	19:34
Air Blank		0.000	19:34
Subject Sample	# 1	0.000	19:35
Breath Volume		1.917	19:35
Air Blank		0.000	19:35
Air Blank		0.000	19:37
Subject Sample	# 2	0.000	19:38
Breath Volume		2.414	19:38
Air Blank		0.000	19:38
Control Test		0.082	19:39
Air Blank		0.000	19:39
Diagnostics Check		OK	19:39

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GROMOLL

DATE/TIME OF TEST: 02/18/2017 00:46:04

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:43	
Control Test		0.082	00:43	
Air Blank		0.000	00:44	
Subject Sample	# 1		00:44	Subject Test Refused
Breath Volume		0.000	00:44	
Air Blank		0.000	00:45	
Control Test		0.081	00:45	
Air Blank		0.000	00:45	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GREENWELL

DATE/TIME OF TEST: 02/18/2017 03:00:49

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:55	
Control Test		0.083	02:55	
Air Blank		0.000	02:56	
Subject Sample	# 1	0.158	02:56	
Breath Volume		1.125	02:56	
Air Blank		0.000	02:57	
Air Blank		0.000	02:59	
Subject Sample	# 2		02:59	Subject Test Refused
Breath Volume		0.000	02:59	
Air Blank		0.000	02:59	
Control Test		0.081	03:00	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 03/19/2017 00:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 02/18/2017 03:11:41

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.082	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.083	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: DINGLER

DATE/TIME OF TEST: 02/18/2017 03:50:39

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.082	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.141	03:46
Breath Volume		2.097	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.138	03:49
Breath Volume		1.464	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:43

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 02/18/2017 05:06:56

Diagnostics Check		OK	05:03	
Air Blank		0.000	05:03	
Control Test		0.081	05:03	
Air Blank		0.000	05:04	
Subject Sample	# 1		05:05	Subject Test Refused
Breath Volume		0.000	05:05	
Air Blank		0.000	05:06	
Control Test		0.081	05:06	
Air Blank		0.000	05:06	
Diagnostics Check		OK	05:06	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/18/2017 16:23:36

Diagnostics Check		OK	16:20	
Air Blank		0.000	16:21	
Control Test		0.082	16:21	
Air Blank		0.000	16:22	
Subject Sample	# 1		16:22	Subject Test Refused
Breath Volume		0.000	16:22	
Air Blank		0.000	16:22	
Control Test		0.082	16:23	
Air Blank		0.000	16:23	
Diagnostics Check		OK	16:23	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 02/18/2017 21:55:31

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:52	
Control Test		0.082	21:53	
Air Blank		0.000	21:53	
Subject Sample	# 1		21:54	Subject Test Refused
Breath Volume		0.000	21:54	
Air Blank		0.000	21:54	
Control Test		0.082	21:54	
Air Blank		0.000	21:55	
Diagnostics Check		OK	21:55	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEX,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 02/18/2017 22:40:25

Diagnostics Check		OK	22:34
Air Blank		0.000	22:34
Control Test		0.082	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.000	22:36
Breath Volume		1.785	22:36
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.000	22:39
Breath Volume		2.320	22:39
Air Blank		0.000	22:39
Control Test		0.082	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 02/18/2017 23:36:27

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:24	
Control Test		0.080	23:24	
Air Blank		0.000	23:25	
Subject Sample	# 1	0.227	23:25	
Breath Volume		1.484	23:25	
Air Blank		0.000	23:26	
Air Blank		0.000	23:28	
Subject Sample	# 2	0.208	23:31	Volume Not Met
Breath Volume		0.031	23:31	
Air Blank		0.000	23:32	
Air Blank		0.000	23:33	
Subject Sample	# 3	0.234	23:34	
Breath Volume		4.316	23:34	
Air Blank		0.000	23:35	
Control Test		0.077	23:35	
Air Blank		0.000	23:36	
Diagnostics Check		OK	23:36	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SHONTERE

DATE/TIME OF TEST: 02/19/2017 00:34:02

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.000	00:29
Breath Volume		1.941	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.000	00:32
Breath Volume		1.488	00:32
Air Blank		0.000	00:33
Control Test		0.081	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:34

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: TACKA

DATE/TIME OF TEST: 02/19/2017 01:41:49

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.025	01:37
Breath Volume		2.441	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:39
Subject Sample	# 2	0.023	01:40
Breath Volume		2.164	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 02/19/2017 05:00:14

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.082	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.000	04:55
Breath Volume		1.140	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.000	04:58
Breath Volume		1.191	04:58
Air Blank		0.000	04:59
Control Test		0.081	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDALL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: NADERMANN

DATE/TIME OF TEST: 02/20/2017 01:54:20

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.173	01:49
Breath Volume		1.625	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.175	01:52
Breath Volume		1.484	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,NICOLE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: THOMPSON,NICOLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:46

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/20/2017 15:20:44

Diagnostics Check		OK	15:06	
Air Blank		0.000	15:06	
Control Test		0.081	15:07	
Air Blank		0.000	15:07	
Subject Sample	# 1	0.116	15:10	Volume Not Met
Breath Volume		0.847	15:10	
Air Blank		0.000	15:11	
Air Blank		0.000	15:13	
Subject Sample	# 2	0.126	15:14	
Breath Volume		1.230	15:14	
Air Blank		0.000	15:15	
Air Blank		0.000	15:16	
Subject Sample	# 3	0.117	15:19	
Breath Volume		1.113	15:19	
Air Blank		0.000	15:19	
Control Test		0.080	15:20	
Air Blank		0.000	15:20	
Diagnostics Check		OK	15:20	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ASHCOM,TIFFANY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: MORACK

DATE/TIME OF TEST: 02/20/2017 23:43:47

Diagnostics Check		OK	23:37
Air Blank		0.000	23:38
Control Test		0.081	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.083	23:39
Breath Volume		3.515	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:41
Subject Sample	# 2	0.079	23:42
Breath Volume		2.312	23:42
Air Blank		0.000	23:42
Control Test		0.081	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TAVERNIER,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ERKSON

DATE/TIME OF TEST: 02/22/2017 03:25:57

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.082	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.261	03:20
Breath Volume		1.800	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.248	03:24
Breath Volume		1.886	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: MCDOW

DATE/TIME OF TEST: 02/22/2017 22:58:24

Diagnostics Check		OK	22:51
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.215	22:53
Breath Volume		1.917	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.214	22:56
Breath Volume		1.847	22:56
Air Blank		0.000	22:57
Control Test		0.081	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: LANGENBURG

DATE/TIME OF TEST: 02/23/2017 02:33:13

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.082	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.024	02:28
Breath Volume		1.574	02:28
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.024	02:31
Breath Volume		1.125	02:31
Air Blank		0.000	02:32
Control Test		0.081	02:32
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LARACUENTE

DATE/TIME OF TEST: 02/23/2017 03:15:43

Diagnostics Check		OK	03:11	
Air Blank		0.000	03:11	
Control Test		0.081	03:11	
Air Blank		0.000	03:12	
Subject Sample	# 1		03:14	Subject Test Refused
Breath Volume		0.000	03:14	
Air Blank		0.000	03:14	
Control Test		0.081	03:15	
Air Blank		0.000	03:15	
Diagnostics Check		OK	03:15	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALATI,VINCENZO,F
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GALATI,VINCENZO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 02/24/2017 01:55:26

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.082	01:53	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:54	Subject Test Refused
Breath Volume		0.000	01:54	
Air Blank		0.000	01:54	
Control Test		0.081	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BOCON

DATE/TIME OF TEST: 02/24/2017 03:06:13

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.081	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.081	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOBSON,JON,B
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DOBSON,JON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: STEVES

DATE/TIME OF TEST: 02/24/2017 16:27:56

Diagnostics Check		OK	16:20
Air Blank		0.000	16:20
Control Test		0.081	16:20
Air Blank		0.000	16:21
Subject Sample	# 1	0.135	16:23
Breath Volume		1.472	16:23
Air Blank		0.000	16:23
Air Blank		0.000	16:25
Subject Sample	# 2	0.136	16:26
Breath Volume		1.156	16:26
Air Blank		0.000	16:27
Control Test		0.080	16:27
Air Blank		0.000	16:27
Diagnostics Check		OK	16:27

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: TEMPLE,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: DE CARVALHO

DATE/TIME OF TEST: 02/24/2017 23:20:09

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.082	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.126	23:15
Breath Volume		1.898	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:17
Subject Sample	# 2	0.124	23:18
Breath Volume		2.019	23:18
Air Blank		0.000	23:19
Control Test		0.080	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/25/2017 02:46:59

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.081	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:46	
Control Test		0.081	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 02/25/2017 18:46:27

Diagnostics Check		OK	18:35	
Air Blank		0.000	18:35	
Control Test		0.081	18:35	
Air Blank		0.000	18:36	
Subject Sample	# 1		18:39	No Sample Provided
Breath Volume		0.000	18:39	
Air Blank		0.000	18:39	
Air Blank		0.000	18:41	
Subject Sample	# 2		18:45	No Sample Provided
Breath Volume		0.000	18:45	
Air Blank		0.000	18:45	
Control Test		0.081	18:45	
Air Blank		0.000	18:46	
Diagnostics Check		OK	18:46	

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PREMRU,KIP,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PREMRU,KIP
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MONGUE

DATE/TIME OF TEST: 02/26/2017 22:47:55

Diagnostics Check		OK	22:41
Air Blank		0.000	22:41
Control Test		0.083	22:42
Air Blank		0.000	22:42
Subject Sample	# 1	0.313	22:43
Breath Volume		1.871	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:45
Subject Sample	# 2	0.317	22:46
Breath Volume		1.179	22:46
Air Blank		0.000	22:46
Control Test		0.081	22:47
Air Blank		0.000	22:47
Diagnostics Check		OK	22:47

Data Download Date/Time: 03/19/2017 00:21

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CURRY,RICHARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TOMKINS JR

DATE/TIME OF TEST: 02/28/2017 01:08:41

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.082	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.000	01:07	
Air Blank		0.000	01:07	
Control Test		0.082	01:08	
Air Blank			01:08	RFI Detect
Air Blank		0.000	01:08	

Data Download Date/Time: 03/19/2017 00:21

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CURRY,RICHARD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: TOMKINS JR

DATE/TIME OF TEST: 02/28/2017 01:18:10

Diagnostics Check		OK	01:11
Air Blank		0.000	01:12
Control Test		0.080	01:12
Air Blank		0.000	01:13
Subject Sample	# 1	0.172	01:13
Breath Volume		1.175	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.154	01:16
Breath Volume		1.292	01:16
Air Blank		0.000	01:17
Control Test		0.081	01:17
Air Blank		0.000	01:18
Diagnostics Check		OK	01:18

Data Download Date/Time: 03/19/2017 00:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: KIST

DATE/TIME OF TEST: 02/28/2017 18:31:24

Diagnostics Check		OK	18:24
Air Blank		0.000	18:25
Control Test		0.082	18:25
Air Blank		0.000	18:26
Subject Sample	# 1	0.187	18:26
Breath Volume		1.531	18:26
Air Blank		0.000	18:27
Air Blank		0.000	18:29
Subject Sample	# 2	0.202	18:29
Breath Volume		2.265	18:29
Air Blank		0.000	18:30
Control Test		0.081	18:30
Air Blank		0.000	18:31
Diagnostics Check		OK	18:31

Data Download Date/Time: 03/19/2017 00:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005290

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHMITTENDORF,JASON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG622905
DRY GAS EXPIRY DATE: 08/16/2018
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: THOMASON

DATE/TIME OF TEST: 02/28/2017 22:33:28

Diagnostics Check		OK	22:26
Air Blank		0.000	22:27
Control Test		0.082	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.182	22:28
Breath Volume		1.972	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.186	22:31
Breath Volume		2.496	22:31
Air Blank		0.000	22:32
Control Test		0.080	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 03/19/2017 00:22

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDLEY,RUDOLPH,A
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 12:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:21

SUBJECT'S LAST NAME: WOMACK

DATE/TIME OF TEST: 02/20/2017 13:58:57

Diagnostics Check		OK	13:50
Air Blank		0.000	13:50
Control Test		0.080	13:51
Air Blank		0.000	13:51
Subject Sample	# 1	0.015	13:54
Breath Volume		1.402	13:54
Air Blank		0.000	13:55
Air Blank		0.000	13:57
Subject Sample	# 2	0.009	13:57
Breath Volume		2.074	13:57
Air Blank		0.000	13:58
Control Test		0.078	13:58
Air Blank		0.000	13:58
Diagnostics Check		OK	13:58

Data Download Date/Time: 02/28/2017 11:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005337

INSTRUMENT REGISTERED TO: PERRY POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENR,JOSEPH,
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SWINDLE,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:14

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/25/2017 16:48:30

Diagnostics Check		OK	16:35	
Air Blank		0.000	16:35	
Control Test		0.081	16:35	
Air Blank		0.000	16:36	
Subject Sample	# 1	0.000	16:39	Volume Not Met
Breath Volume		0.054	16:39	
Air Blank		0.000	16:39	
Air Blank		0.000	16:41	
Subject Sample	# 2	0.000	16:44	
Breath Volume		2.011	16:44	
Air Blank		0.000	16:44	
Air Blank		0.000	16:46	
Subject Sample	# 3	0.000	16:47	
Breath Volume		2.207	16:47	
Air Blank		0.000	16:47	
Control Test		0.079	16:47	
Air Blank		0.000	16:48	
Diagnostics Check		OK	16:48	

Data Download Date/Time: 02/28/2017 11:39

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: MOUNT,JOSHUA
ARRESTING OFFICER AGENCY: South Daytona PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: BATES

DATE/TIME OF TEST: 02/01/2017 20:41:50

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.083	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.192	20:37
Breath Volume		3.089	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:39
Subject Sample	# 2	0.186	20:40
Breath Volume		1.992	20:40
Air Blank		0.000	20:40
Control Test		0.081	20:41
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAN,JONATHAN,W
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: RAYNOR,JASON
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 02/05/2017 04:01:38

Diagnostics Check		OK	03:55
Air Blank		0.000	03:55
Control Test		0.082	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.159	03:57
Breath Volume		2.250	03:57
Air Blank		0.000	03:57
Air Blank		0.000	03:59
Subject Sample	# 2	0.143	04:00
Breath Volume		2.285	04:00
Air Blank		0.000	04:00
Control Test		0.082	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KURTZ,THEODORE,F
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: KURTZ,THEODORE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: COSTNER

DATE/TIME OF TEST: 02/05/2017 21:54:45

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.082	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.119	21:49
Breath Volume		1.285	21:49
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.115	21:53
Breath Volume		1.652	21:53
Air Blank		0.000	21:53
Control Test		0.081	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANDESE,ANGELO,
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: KIMBERLY,KYLE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/16/2017 04:16:09

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.082	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.155	04:11
Breath Volume		2.105	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.142	04:14
Breath Volume		1.582	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKAY,ANDREW,G
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: CARELLI,PETER
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SAMPSON

DATE/TIME OF TEST: 02/20/2017 02:30:22

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.178	02:25
Breath Volume		1.117	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.191	02:28
Breath Volume		1.425	02:28
Air Blank		0.000	02:29
Control Test		0.082	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLAN,JONATHAN,W
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: NOLAN,JONATHAN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: KEENAN

DATE/TIME OF TEST: 02/23/2017 02:54:02

Diagnostics Check		OK	02:51	
Air Blank		0.000	02:51	
Control Test		0.081	02:52	
Air Blank		0.000	02:52	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:53	
Control Test		0.082	02:53	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:54	

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005402

INSTRUMENT REGISTERED TO: PORT ORANGE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOTT,KAYLA,M
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: REED,MICHAEL
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1836890
DRY GAS EXPIRY DATE: 03/26/2018
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: RAINERE

DATE/TIME OF TEST: 02/25/2017 11:32:47

Diagnostics Check		OK	11:26
Air Blank		0.000	11:27
Control Test		0.081	11:27
Air Blank		0.000	11:27
Subject Sample	# 1	0.192	11:28
Breath Volume		2.507	11:28
Air Blank		0.000	11:29
Air Blank		0.000	11:30
Subject Sample	# 2	0.194	11:31
Breath Volume		1.781	11:31
Air Blank		0.000	11:31
Control Test		0.081	11:32
Air Blank		0.000	11:32
Diagnostics Check		OK	11:32

Data Download Date/Time: 02/26/2017 13:20

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: RICHTER

DATE/TIME OF TEST: 02/01/2017 02:27:50

Diagnostics Check		OK	02:21
Air Blank		0.000	02:21
Control Test		0.080	02:22
Air Blank		0.000	02:22
Subject Sample	# 1	0.098	02:23
Breath Volume		1.300	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.110	02:26
Breath Volume		2.023	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 02/13/2017 14:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: CONLEY,SHAWNA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 02/01/2017 22:43:57

Diagnostics Check		OK	22:40	
Air Blank		0.000	22:41	
Control Test		0.080	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1		22:42	Subject Test Refused
Breath Volume		0.000	22:42	
Air Blank		0.000	22:43	
Control Test		0.080	22:43	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 02/13/2017 14:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 02/05/2017 01:22:09

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:18	
Control Test		0.080	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:21	
Control Test		0.081	01:21	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 02/13/2017 14:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ANTHONY

DATE/TIME OF TEST: 02/05/2017 01:40:07

Diagnostics Check		OK	01:28
Air Blank		0.000	01:29
Control Test		0.081	01:29
Air Blank		0.000	01:30
Subject Sample	# 1	0.199	01:32
Breath Volume		1.472	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.161	01:35
Breath Volume		1.886	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 3	0.192	01:38
Breath Volume		1.996	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 02/13/2017 14:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: DUNNE,JUSTIN
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: RAYNE

DATE/TIME OF TEST: 02/05/2017 18:53:04

Diagnostics Check		OK	18:50	
Air Blank		0.000	18:50	
Control Test		0.080	18:51	
Air Blank		0.000	18:51	
Subject Sample	# 1		18:51	Subject Test Refused
Breath Volume		0.000	18:51	
Air Blank		0.000	18:52	
Control Test		0.080	18:52	
Air Blank		0.000	18:52	
Diagnostics Check		OK	18:53	

Data Download Date/Time: 02/13/2017 14:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 01/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MASTROIANNI

DATE/TIME OF TEST: 02/09/2017 21:39:34

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.080	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.141	21:32
Breath Volume		1.453	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.126	21:37
Breath Volume		1.273	21:37
Air Blank		0.000	21:38
Control Test		0.080	21:38
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 02/13/2017 14:19

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNNE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: GREENE,BARBARA
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: CINICOLA

DATE/TIME OF TEST: 02/18/2017 02:46:42

Diagnostics Check		OK	02:40
Air Blank		0.000	02:40
Control Test		0.080	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.000	02:42
Breath Volume		1.382	02:42
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.000	02:45
Breath Volume		1.574	02:45
Air Blank		0.000	02:45
Control Test		0.080	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:46

Data Download Date/Time: 03/27/2017 14:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARD,JAMES,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: WARD,JAMES
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: FRISA

DATE/TIME OF TEST: 02/19/2017 22:01:17

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.080	21:55
Air Blank		0.000	21:55
Subject Sample	# 1	0.187	21:56
Breath Volume		1.582	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.187	21:59
Breath Volume		3.027	21:59
Air Blank		0.000	22:00
Control Test		0.080	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 03/27/2017 14:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: BOOTH,MARCUS
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 02/24/2017 23:56:23

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.080	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:51
Breath Volume		2.234	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		1.937	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 03/27/2017 14:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAMPE,CORTLAND,B
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: EDWARDS,LANDON
ARRESTING OFFICER AGENCY: Daytona Beach PD
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: BYNDLOSS

DATE/TIME OF TEST: 02/25/2017 04:38:05

Diagnostics Check		OK	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.141	04:33
Breath Volume		2.164	04:33
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.151	04:36
Breath Volume		2.011	04:36
Air Blank		0.000	04:37
Control Test		0.078	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 03/27/2017 14:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005537

INSTRUMENT REGISTERED TO: DAYTONA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONLEY,SHAWNA,A
BREATH TEST OPERATOR AGENCY: Daytona Beach PD
ARRESTING OFFICER: THORNSBERRY,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: MARLOWE JR

DATE/TIME OF TEST: 02/26/2017 20:27:59

Diagnostics Check		OK	20:25	
Air Blank		0.000	20:25	
Control Test		0.080	20:25	
Air Blank		0.000	20:26	
Subject Sample	# 1		20:26	Subject Test Refused
Breath Volume		0.000	20:26	
Air Blank		0.000	20:27	
Control Test		0.081	20:27	
Air Blank		0.000	20:27	
Diagnostics Check		OK	20:27	

Data Download Date/Time: 03/27/2017 14:55

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005548

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORENO,FRANCISCO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MEADORS,RACHEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/17/2017 08:40:59

Diagnostics Check		OK	08:34
Air Blank		0.000	08:35
Control Test		0.079	08:35
Air Blank		0.000	08:35
Subject Sample	# 1	0.018	08:36
Breath Volume		2.000	08:36
Air Blank		0.000	08:37
Air Blank		0.000	08:39
Subject Sample	# 2	0.015	08:39
Breath Volume		2.605	08:39
Air Blank		0.000	08:40
Control Test		0.076	08:40
Air Blank		0.000	08:40
Diagnostics Check		OK	08:40

Data Download Date/Time: 03/07/2017 10:18

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AGUIAR,JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 02/10/2017 23:25:55

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.079	23:19
Air Blank		0.000	23:19
Subject Sample	# 1	0.199	23:21
Breath Volume		3.367	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:23
Subject Sample	# 2	0.203	23:24
Breath Volume		3.183	23:24
Air Blank		0.000	23:24
Control Test		0.077	23:25
Air Blank		0.000	23:25
Diagnostics Check		OK	23:25

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 02/10/2017 23:50:09

Diagnostics Check		OK	23:42
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.057	23:45
Breath Volume		1.902	23:45
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.054	23:48
Breath Volume		2.058	23:48
Air Blank		0.000	23:49
Control Test		0.075	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CESPEDES DOMINGUEZ

DATE/TIME OF TEST: 02/11/2017 01:02:22

Diagnostics Check		OK	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Subject Sample	# 1	0.167	00:57
Breath Volume		2.332	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.168	01:00
Breath Volume		1.695	01:00
Air Blank		0.000	01:01
Control Test		0.078	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BEROVIDES MILIAN

DATE/TIME OF TEST: 02/11/2017 01:24:33

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.077	01:22	
Air Blank		0.000	01:22	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.077	01:23	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VALDEZ,MICHAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: CARDOSO

DATE/TIME OF TEST: 02/11/2017 02:52:44

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.078	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.073	02:48
Breath Volume		1.531	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.078	02:51
Breath Volume		2.488	02:51
Air Blank		0.000	02:51
Control Test		0.079	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TORRES,DAVID
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 02/11/2017 03:29:11

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.078	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.164	03:24
Breath Volume		2.230	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.167	03:27
Breath Volume		1.402	03:27
Air Blank		0.000	03:28
Control Test		0.078	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DAEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TABOADA VALDIVIA

DATE/TIME OF TEST: 02/11/2017 03:20:28

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.078	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.151	03:15
Breath Volume		1.824	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.140	03:18
Breath Volume		1.570	03:18
Air Blank		0.000	03:19
Control Test		0.077	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FERRO,LISA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: GOLATT

DATE/TIME OF TEST: 02/12/2017 02:31:37

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.080	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:30	Subject Test Refused
Breath Volume		0.000	02:30	
Air Blank		0.000	02:30	
Control Test		0.079	02:31	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 02/14/2017 13:48

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,MANUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 02/24/2017 21:54:45

Diagnostics Check		OK	21:51	
Air Blank		0.000	21:52	
Control Test		0.080	21:52	
Air Blank		0.000	21:53	
Subject Sample	# 1		21:53	Subject Test Refused
Breath Volume		0.000	21:53	
Air Blank		0.000	21:53	
Control Test		0.078	21:54	
Air Blank		0.000	21:54	
Diagnostics Check		OK	21:54	

Data Download Date/Time: 03/01/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: ALAIX

DATE/TIME OF TEST: 02/25/2017 01:53:06

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.276	01:47
Breath Volume		2.156	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.270	01:51
Breath Volume		2.386	01:51
Air Blank		0.000	01:52
Control Test		0.078	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 03/01/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEY,LAZARO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: STRINGER,ROBERT
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: VICTORIA DOMINGUEZ

DATE/TIME OF TEST: 02/25/2017 02:40:34

Diagnostics Check		OK	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.179	02:35
Breath Volume		3.117	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.184	02:38
Breath Volume		2.394	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 03/01/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005549

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESPINOZA,PABLO,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DE ROSA,MARCOS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A2
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: COLINA

DATE/TIME OF TEST: 02/25/2017 02:58:31

Diagnostics Check		OK	02:55	
Air Blank		0.000	02:56	
Control Test		0.078	02:56	
Air Blank		0.000	02:56	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:57	
Control Test		0.079	02:57	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:58	

Data Download Date/Time: 03/01/2017 11:05

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 02/10/2017 03:45:59

Diagnostics Check		OK	03:39
Air Blank		0.000	03:39
Control Test		0.078	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.166	03:41
Breath Volume		2.007	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.174	03:44
Breath Volume		1.871	03:44
Air Blank		0.000	03:45
Control Test		0.079	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 03/09/2017 12:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005551

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZRAK,JOSEPH,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: AZRAK,JOSEPH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 02/11/2017 01:37:02

Diagnostics Check		OK	01:30
Air Blank		0.000	01:31
Control Test		0.079	01:31
Air Blank		0.000	01:32
Subject Sample	# 1	0.141	01:32
Breath Volume		1.710	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.140	01:35
Breath Volume		1.539	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 03/09/2017 12:14

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITZGERALD,FRANK,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STAFFORD,BRIAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: STROUD

DATE/TIME OF TEST: 02/03/2017 03:31:18

Diagnostics Check		OK	03:24	
Air Blank		0.000	03:25	
Control Test		0.077	03:25	
Air Blank		0.000	03:26	
Subject Sample	# 1	0.225	03:27	
Breath Volume		1.449	03:27	
Air Blank		0.000	03:28	
Air Blank		0.000	03:29	
Subject Sample	# 2		03:30	RFI Detect
Breath Volume		1.085	03:30	
Air Blank			03:30	RFI Detect
Air Blank		0.000	03:31	

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITZGERALD,FRANK,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STAFFORD,BRIAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: STROUD

DATE/TIME OF TEST: 02/03/2017 03:40:04

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.076	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.204	03:35
Breath Volume		1.250	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.222	03:38
Breath Volume		1.781	03:38
Air Blank		0.000	03:39
Control Test		0.077	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:40

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: AULETTA

DATE/TIME OF TEST: 02/08/2017 00:08:17

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.077	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.000	00:03
Breath Volume		2.246	00:03
Air Blank		0.000	00:04
Air Blank		0.000	00:06
Subject Sample	# 2	0.000	00:06
Breath Volume		3.164	00:06
Air Blank		0.000	00:07
Control Test		0.077	00:07
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SALMERON ACEVEDO

DATE/TIME OF TEST: 02/12/2017 02:48:43

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.077	02:42
Air Blank		0.000	02:43
Subject Sample	# 1	0.130	02:43
Breath Volume		2.605	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.130	02:47
Breath Volume		1.878	02:47
Air Blank		0.000	02:47
Control Test		0.076	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: THOMAS JR

DATE/TIME OF TEST: 02/13/2017 04:05:30

Diagnostics Check		OK	03:54	
Air Blank		0.000	03:54	
Control Test		0.078	03:54	
Air Blank		0.000	03:55	
Subject Sample	# 1		03:58	No Sample Provided
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2		04:04	No Sample Provided
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	
Control Test		0.078	04:04	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: HARRISON,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ANARUMO

DATE/TIME OF TEST: 02/14/2017 17:26:51

Diagnostics Check	OK	17:24	
Air Blank	0.000	17:25	
Control Test	0.076	17:25	
Air Blank	0.000	17:26	
		17:26	RFI Detect
Air Blank	0.000	17:26	

Data Download Date/Time: 02/27/2017 09:31

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRISON,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: HARRISON,NICHOLAS
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: ANARUMO

DATE/TIME OF TEST: 02/14/2017 17:34:12

Diagnostics Check		OK	17:29	
Air Blank		0.000	17:29	
Control Test		0.077	17:29	
Air Blank		0.000	17:30	
Subject Sample	# 1		17:32	Subject Test Refused
Breath Volume		0.593	17:32	
Air Blank		0.000	17:33	
Control Test		0.078	17:33	
Air Blank		0.000	17:34	
Diagnostics Check		OK	17:34	

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ROSIER

DATE/TIME OF TEST: 02/15/2017 00:31:03

Diagnostics Check		OK	00:23
Air Blank		0.000	00:24
Control Test		0.076	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.287	00:25
Breath Volume		2.671	00:25
Air Blank		0.000	00:26
Air Blank		0.000	00:28
Subject Sample	# 2	0.279	00:29
Breath Volume		3.308	00:29
Air Blank		0.000	00:30
Control Test		0.076	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:31

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 02/17/2017 00:11:17

Diagnostics Check		OK	00:04
Air Blank		0.000	00:05
Control Test		0.077	00:05
Air Blank		0.000	00:05
Subject Sample	# 1	0.154	00:06
Breath Volume		2.695	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:08
Subject Sample	# 2	0.161	00:09
Breath Volume		1.941	00:09
Air Blank		0.000	00:10
Control Test		0.078	00:10
Air Blank		0.000	00:11
Diagnostics Check		OK	00:11

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PINEDA-ARIAS

DATE/TIME OF TEST: 02/19/2017 03:51:48

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.167	03:47
Breath Volume		2.921	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.164	03:50
Breath Volume		1.980	03:50
Air Blank		0.000	03:50
Control Test		0.077	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: WARE

DATE/TIME OF TEST: 02/20/2017 02:03:24

Diagnostics Check		OK	01:57
Air Blank		0.000	01:57
Control Test		0.078	01:58
Air Blank		0.000	01:58
Subject Sample	# 1	0.142	01:58
Breath Volume		3.980	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.136	02:01
Breath Volume		1.503	02:01
Air Blank		0.000	02:02
Control Test		0.077	02:02
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 02/21/2017 02:47:45

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.077	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.191	02:42
Breath Volume		2.214	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.192	02:46
Breath Volume		1.867	02:46
Air Blank		0.000	02:46
Control Test		0.077	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULDERRIG,SEAN,F
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: MULDERRIG,SEAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:03

SUBJECT'S LAST NAME: PALACIO

DATE/TIME OF TEST: 02/23/2017 06:30:01

Diagnostics Check		OK	06:23
Air Blank		0.000	06:23
Control Test		0.077	06:24
Air Blank		0.000	06:24
Subject Sample	# 1	0.123	06:25
Breath Volume		2.820	06:25
Air Blank		0.000	06:25
Air Blank		0.000	06:27
Subject Sample	# 2	0.121	06:28
Breath Volume		2.832	06:28
Air Blank		0.000	06:29
Control Test		0.076	06:29
Air Blank		0.000	06:29
Diagnostics Check		OK	06:30

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GAINOUS

DATE/TIME OF TEST: 02/25/2017 00:49:23

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.077	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.197	00:44
Breath Volume		2.042	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.197	00:47
Breath Volume		1.726	00:47
Air Blank		0.000	00:48
Control Test		0.075	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: STAFFORD,BRIAN
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: DIAZ II

DATE/TIME OF TEST: 02/25/2017 05:34:56

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.077	05:24	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:27	No Sample Provided
Breath Volume		0.000	05:27	
Air Blank		0.000	05:28	
Air Blank		0.000	05:30	
Subject Sample	# 2		05:33	No Sample Provided
Breath Volume		0.000	05:33	
Air Blank		0.000	05:33	
Control Test		0.078	05:34	
Air Blank		0.000	05:34	
Diagnostics Check		OK	05:34	

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LENTZ

DATE/TIME OF TEST: 02/26/2017 01:30:50

Diagnostics Check	OK	01:29	
Air Blank		01:30	RFI Detect
Air Blank	0.000	01:30	

Data Download Date/Time: 02/27/2017 09:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005810

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLNER,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: KELLNER,MICHAEL
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 01/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LENTZ

DATE/TIME OF TEST: 02/26/2017 01:46:45

Diagnostics Check		OK	01:35	
Air Blank		0.000	01:35	
Control Test		0.077	01:36	
Air Blank		0.000	01:36	
Subject Sample	# 1		01:39	No Sample Provided
Breath Volume		0.000	01:39	
Air Blank		0.000	01:40	
Air Blank		0.000	01:42	
Subject Sample	# 2		01:45	No Sample Provided
Breath Volume		0.000	01:45	
Air Blank		0.000	01:45	
Control Test		0.078	01:46	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 02/27/2017 09:32

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TRPR ABROMAITIS,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: KREBS

DATE/TIME OF TEST: 02/02/2017 23:45:03

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.081	23:39
Air Blank		0.000	23:39
Subject Sample	# 1	0.000	23:40
Breath Volume		2.394	23:40
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.000	23:43
Breath Volume		1.871	23:43
Air Blank		0.000	23:44
Control Test		0.081	23:44
Air Blank		0.000	23:44
Diagnostics Check		OK	23:45

Data Download Date/Time: 02/27/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BRAFFORD

DATE/TIME OF TEST: 02/08/2017 23:04:30

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.000	23:00
Breath Volume		3.445	23:00
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.000	23:03
Breath Volume		4.390	23:03
Air Blank		0.000	23:03
Control Test		0.080	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 02/27/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ADAM

DATE/TIME OF TEST: 02/18/2017 03:24:06

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.082	03:18
Air Blank		0.000	03:18
Subject Sample	# 1	0.124	03:19
Breath Volume		1.910	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.122	03:22
Breath Volume		2.062	03:22
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 02/27/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-005936

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUFAHL,G,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PUFAHL,G
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PARSONS

DATE/TIME OF TEST: 02/24/2017 02:20:45

Diagnostics Check		OK	02:12
Air Blank		0.000	02:12
Control Test		0.081	02:12
Air Blank		0.000	02:13
Subject Sample	# 1	0.213	02:14
Breath Volume		4.015	02:14
Air Blank		0.000	02:14
Air Blank		0.000	02:16
Subject Sample	# 2	0.211	02:18
Breath Volume		2.718	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:20
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 02/27/2017 08:23

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: GUTIERREZ LOPEZ

DATE/TIME OF TEST: 02/17/2017 21:00:39

Diagnostics Check		OK	20:52
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:53
Subject Sample	# 1	0.235	20:55
Breath Volume		2.375	20:55
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.220	20:59
Breath Volume		1.113	20:59
Air Blank		0.000	20:59
Control Test		0.079	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 03/24/2017 12:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: ROCKLAND

DATE/TIME OF TEST: 02/25/2017 01:04:55

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.080	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.207	00:59
Breath Volume		2.488	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:01
Subject Sample	# 2	0.200	01:03
Breath Volume		2.171	01:03
Air Blank		0.000	01:03
Control Test		0.078	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 03/24/2017 12:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: SANTIZ GOMEZ

DATE/TIME OF TEST: 02/26/2017 02:48:12

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.157	02:43
Breath Volume		2.292	02:43
Air Blank		0.000	02:44
Air Blank		0.000	02:45
Subject Sample	# 2	0.159	02:46
Breath Volume		1.562	02:46
Air Blank		0.000	02:47
Control Test		0.081	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 03/24/2017 12:30

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006027

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20314080A4
DRY GAS EXPIRY DATE: 08/05/2016
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 02/26/2017 04:10:49

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.081	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.070	04:06
Breath Volume		1.546	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.071	04:09
Breath Volume		1.539	04:09
Air Blank		0.000	04:09
Control Test		0.080	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 03/24/2017 12:30

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: HUDON

DATE/TIME OF TEST: 02/01/2017 01:04:53

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:58
Subject Sample	# 1	0.191	01:00
Breath Volume		2.351	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.190	01:03
Breath Volume		2.945	01:03
Air Blank		0.000	01:03
Control Test		0.080	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CONNOR,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:37

SUBJECT'S LAST NAME: MORDAN

DATE/TIME OF TEST: 02/01/2017 20:08:41

Diagnostics Check		OK	20:01
Air Blank		0.000	20:01
Control Test		0.082	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.162	20:03
Breath Volume		2.074	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:06
Subject Sample	# 2	0.162	20:07
Breath Volume		1.816	20:07
Air Blank		0.000	20:07
Control Test		0.081	20:08
Air Blank		0.000	20:08
Diagnostics Check		OK	20:08

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: D'IPPOLITO

DATE/TIME OF TEST: 02/02/2017 04:12:53

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.081	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.190	04:07
Breath Volume		2.058	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.189	04:11
Breath Volume		1.914	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EDWARDS,DEVON
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: HALLBERG

DATE/TIME OF TEST: 02/03/2017 02:07:04

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.082	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.177	02:02
Breath Volume		1.128	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.175	02:05
Breath Volume		1.308	02:05
Air Blank		0.000	02:06
Control Test		0.079	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:07

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: FREER

DATE/TIME OF TEST: 02/03/2017 22:55:48

Diagnostics Check		OK	22:52	
Air Blank		0.000	22:52	
Control Test		0.082	22:53	
Air Blank		0.000	22:53	
Subject Sample	# 1		22:54	Slope Not Met
Breath Volume		1.757	22:54	
Air Blank		0.000	22:54	
Control Test		0.082	22:55	
Air Blank		0.000	22:55	
Diagnostics Check		OK	22:55	

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: FREER

DATE/TIME OF TEST: 02/03/2017 23:23:49

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.081	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.169	23:19
Breath Volume		2.191	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.175	23:22
Breath Volume		2.023	23:22
Air Blank		0.000	23:22
Control Test		0.080	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STUBBLEBINE,LIVIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/05/2017 02:12:03

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.082	02:06
Air Blank		0.000	02:06
Subject Sample	# 1	0.226	02:07
Breath Volume		2.363	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.220	02:10
Breath Volume		1.441	02:10
Air Blank		0.000	02:11
Control Test		0.080	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:12

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: SANTANA LOZADA

DATE/TIME OF TEST: 02/05/2017 04:21:56

Diagnostics Check		OK	04:17	
Air Blank		0.000	04:18	
Control Test		0.081	04:18	
Air Blank		0.000	04:19	
Subject Sample	# 1		04:20	Subject Test Refused
Breath Volume		0.000	04:20	
Air Blank		0.000	04:21	
Control Test		0.082	04:21	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PICADO GONZALEZ

DATE/TIME OF TEST: 02/06/2017 02:16:35

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.082	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.153	02:11
Breath Volume		1.835	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.152	02:14
Breath Volume		1.753	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/07/2017 04:41:03

Diagnostics Check		OK	04:34
Air Blank		0.000	04:35
Control Test		0.082	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.142	04:36
Breath Volume		1.230	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:38
Subject Sample	# 2	0.146	04:39
Breath Volume		1.199	04:39
Air Blank		0.000	04:40
Control Test		0.081	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:41

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLOCK,RONALD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: OTTOLINO

DATE/TIME OF TEST: 02/07/2017 17:13:34

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.081	17:06
Air Blank		0.000	17:06
Subject Sample	# 1	0.151	17:08
Breath Volume		1.429	17:08
Air Blank		0.000	17:09
Air Blank		0.000	17:10
Subject Sample	# 2	0.164	17:11
Breath Volume		1.589	17:11
Air Blank		0.000	17:12
Control Test		0.081	17:12
Air Blank		0.000	17:13
Diagnostics Check		OK	17:13

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BALKISSOON,ANITA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: FRESE

DATE/TIME OF TEST: 02/09/2017 00:31:48

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.081	00:26
Air Blank		0.000	00:26
Subject Sample	# 1	0.163	00:27
Breath Volume		2.339	00:27
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.174	00:30
Breath Volume		3.277	00:30
Air Blank		0.000	00:30
Control Test		0.081	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Female

OBSERVATION PERIOD BEGAN: 19:39

SUBJECT'S LAST NAME: HOROWITZ

DATE/TIME OF TEST: 02/09/2017 20:16:18

Diagnostics Check		OK	20:08
Air Blank		0.000	20:08
Control Test		0.081	20:08
Air Blank		0.000	20:09
Subject Sample	# 1	0.160	20:11
Breath Volume		1.265	20:11
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.176	20:14
Breath Volume		1.992	20:14
Air Blank		0.000	20:15
Control Test		0.080	20:15
Air Blank		0.000	20:16
Diagnostics Check		OK	20:16

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STANG,FREDERICK
ARRESTING OFFICER AGENCY: Ocean Ridge PD
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: KOOTZ

DATE/TIME OF TEST: 02/09/2017 21:53:46

Diagnostics Check		OK	21:46
Air Blank		0.000	21:46
Control Test		0.081	21:47
Air Blank		0.000	21:47
Subject Sample	# 1	0.261	21:48
Breath Volume		1.433	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.253	21:52
Breath Volume		1.394	21:52
Air Blank		0.000	21:52
Control Test		0.081	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EVES,TYLER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PASCUAL-MATIAS

DATE/TIME OF TEST: 02/09/2017 23:59:41

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:48	
Control Test		0.082	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:52	No Sample Provided
Breath Volume		0.000	23:52	
Air Blank		0.000	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 2	0.164	23:57	Volume Not Met
Breath Volume		0.289	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:59	
Air Blank		0.000	23:59	
Diagnostics Check		OK	23:59	

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006029

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: EVES,TYLER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 01/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: PASCUAL-MATIAS

DATE/TIME OF TEST: 02/10/2017 00:14:11

Diagnostics Check		OK	00:02
Air Blank		0.000	00:02
Control Test		0.078	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.182	00:04
Breath Volume		1.378	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.157	00:08
Breath Volume		1.289	00:08
Air Blank		0.000	00:09
Air Blank		0.000	00:11
Subject Sample	# 3	0.166	00:12
Breath Volume		1.207	00:12
Air Blank		0.000	00:13
Control Test		0.078	00:13
Air Blank		0.000	00:14
Diagnostics Check		OK	00:14

Data Download Date/Time: 02/10/2017 14:04

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: RUTLAND

DATE/TIME OF TEST: 02/01/2017 22:09:33

Diagnostics Check		OK	22:02
Air Blank		0.000	22:02
Control Test		0.080	22:03
Air Blank		0.000	22:03
Subject Sample	# 1	0.125	22:04
Breath Volume		1.464	22:04
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.128	22:08
Breath Volume		1.343	22:08
Air Blank		0.000	22:08
Control Test		0.080	22:08
Air Blank		0.000	22:09
Diagnostics Check		OK	22:09

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 02/05/2017 01:29:42

Diagnostics Check		OK	01:22
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.163	01:24
Breath Volume		1.929	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.164	01:28
Breath Volume		1.859	01:28
Air Blank		0.000	01:28
Control Test		0.080	01:29
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LARA,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LARA,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: VAZQUEZ HERNANDEZ

DATE/TIME OF TEST: 02/07/2017 21:09:17

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:06	
Control Test		0.079	21:06	
Air Blank		0.000	21:06	
Subject Sample	# 1	0.188	21:08	
Breath Volume		2.046	21:08	
Air Blank			21:08	Ambient Fail
Air Blank		0.000	21:09	

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LARA,ALEJANDRO
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: VAZQUEZ HERNANDEZ

DATE/TIME OF TEST: 02/07/2017 21:19:16

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.077	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.186	21:14
Breath Volume		1.957	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:16
Subject Sample	# 2	0.183	21:17
Breath Volume		1.843	21:17
Air Blank		0.000	21:18
Control Test		0.077	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: POLANIA MARIN

DATE/TIME OF TEST: 02/11/2017 02:07:50

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.080	02:02
Air Blank		0.000	02:02
Subject Sample	# 1	0.159	02:03
Breath Volume		2.609	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.154	02:06
Breath Volume		2.093	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: CRABTREE

DATE/TIME OF TEST: 02/11/2017 03:46:45

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:41
Subject Sample	# 1	0.175	03:42
Breath Volume		2.289	03:42
Air Blank		0.000	03:42
Air Blank		0.000	03:44
Subject Sample	# 2	0.175	03:45
Breath Volume		1.902	03:45
Air Blank		0.000	03:45
Control Test		0.080	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 05:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:32

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 02/11/2017 06:00:04

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.081	05:53
Air Blank		0.000	05:54
Subject Sample	# 1	0.143	05:54
Breath Volume		1.636	05:54
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.147	05:58
Breath Volume		1.964	05:58
Air Blank		0.000	05:59
Control Test		0.080	05:59
Air Blank		0.000	05:59
Diagnostics Check		OK	06:00

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: CAUSEY

DATE/TIME OF TEST: 02/12/2017 01:35:00

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.194	01:29
Breath Volume		1.546	01:29
Air Blank		0.000	01:30
Air Blank		0.000	01:31
Subject Sample	# 2	0.183	01:33
Breath Volume		1.328	01:33
Air Blank		0.000	01:34
Control Test		0.079	01:34
Air Blank		0.000	01:34
Diagnostics Check		OK	01:35

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: RAULERSON

DATE/TIME OF TEST: 02/12/2017 02:32:21

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.079	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.164	02:28
Breath Volume		1.917	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.166	02:30
Breath Volume		1.714	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CESPEDES DIAZ

DATE/TIME OF TEST: 02/12/2017 03:54:48

Diagnostics Check		OK	03:51	
Air Blank		0.000	03:52	
Control Test		0.080	03:52	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:53	Subject Test Refused
Breath Volume		0.000	03:53	
Air Blank		0.000	03:53	
Control Test		0.080	03:54	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:54	

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: CESPEDES DIAZ

DATE/TIME OF TEST: 02/12/2017 03:59:55

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.080	03:56	
Air Blank		0.000	03:57	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:58	
Control Test		0.080	03:59	
Air Blank		0.000	03:59	
Diagnostics Check		OK	03:59	

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: BIKULIC

DATE/TIME OF TEST: 02/12/2017 22:01:21

Diagnostics Check		OK	21:54
Air Blank		0.000	21:55
Control Test		0.079	21:55
Air Blank		0.000	21:56
Subject Sample	# 1	0.043	21:56
Breath Volume		1.542	21:56
Air Blank		0.000	21:57
Air Blank		0.000	21:59
Subject Sample	# 2	0.045	21:59
Breath Volume		1.812	21:59
Air Blank		0.000	22:00
Control Test		0.080	22:00
Air Blank		0.000	22:01
Diagnostics Check		OK	22:01

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:03

SUBJECT'S LAST NAME: KEEN

DATE/TIME OF TEST: 02/16/2017 17:33:00

Diagnostics Check		OK	17:27
Air Blank		0.000	17:27
Control Test		0.080	17:27
Air Blank		0.000	17:28
Subject Sample	# 1	0.000	17:28
Breath Volume		2.933	17:28
Air Blank		0.000	17:29
Air Blank		0.000	17:31
Subject Sample	# 2	0.000	17:31
Breath Volume		3.015	17:31
Air Blank		0.000	17:32
Control Test		0.080	17:32
Air Blank		0.000	17:32
Diagnostics Check		OK	17:33

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RHOADES

DATE/TIME OF TEST: 02/17/2017 03:39:04

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.160	03:34
Breath Volume		2.574	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.153	03:37
Breath Volume		2.171	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: LOPEZ RODRIGUEZ

DATE/TIME OF TEST: 02/18/2017 03:32:35

Diagnostics Check		OK	03:28	
Air Blank		0.000	03:28	
Control Test		0.080	03:29	
Air Blank		0.000	03:29	
Subject Sample	# 1		03:31	Subject Test Refused
Breath Volume		0.000	03:31	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: RECIO

DATE/TIME OF TEST: 02/19/2017 22:42:39

Diagnostics Check		OK	22:34
Air Blank		0.000	22:35
Control Test		0.080	22:35
Air Blank		0.000	22:35
Subject Sample	# 1	0.080	22:36
Breath Volume		1.625	22:36
Air Blank		0.000	22:37
Air Blank		0.000	22:39
Subject Sample	# 2	0.076	22:41
Breath Volume		1.832	22:41
Air Blank		0.000	22:41
Control Test		0.079	22:42
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 02/25/2017 01:12:50

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.000	01:08
Breath Volume		1.957	01:08
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.000	01:11
Breath Volume		1.878	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:12
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 02/26/2017 03:12:56

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.079	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.194	03:08
Breath Volume		2.687	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.189	03:11
Breath Volume		2.472	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006046

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ANGULO

DATE/TIME OF TEST: 02/26/2017 04:12:40

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.078	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.079	04:08
Breath Volume		1.382	04:08
Air Blank		0.000	04:08
Air Blank		0.000	04:10
Subject Sample	# 2	0.075	04:11
Breath Volume		1.292	04:11
Air Blank		0.000	04:11
Control Test		0.079	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 02/28/2017 00:00

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 02/01/2017 00:55:33

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.082	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.191	00:51
Breath Volume		2.550	00:51
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.199	00:54
Breath Volume		1.246	00:54
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: SYLVIA

DATE/TIME OF TEST: 02/03/2017 23:22:00

Diagnostics Check		OK	23:15
Air Blank		0.000	23:15
Control Test		0.082	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.082	23:16
Breath Volume		1.300	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.088	23:20
Breath Volume		1.785	23:20
Air Blank		0.000	23:21
Control Test		0.081	23:21
Air Blank		0.000	23:21
Diagnostics Check		OK	23:22

Data Download Date/Time: 02/27/2017 13:36

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: LOREDO

DATE/TIME OF TEST: 02/04/2017 03:53:51

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.081	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.118	03:48
Breath Volume		1.265	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.129	03:52
Breath Volume		3.113	03:52
Air Blank		0.000	03:52
Control Test		0.080	03:53
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 02/04/2017 04:37:10

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.081	04:31
Air Blank		0.000	04:32
Subject Sample	# 1	0.133	04:32
Breath Volume		3.058	04:32
Air Blank		0.000	04:33
Air Blank		0.000	04:35
Subject Sample	# 2	0.142	04:35
Breath Volume		3.660	04:35
Air Blank		0.000	04:36
Control Test		0.081	04:36
Air Blank		0.000	04:37
Diagnostics Check		OK	04:37

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: WISE

DATE/TIME OF TEST: 02/04/2017 05:30:49

Diagnostics Check		OK	05:24
Air Blank		0.000	05:24
Control Test		0.080	05:25
Air Blank		0.000	05:25
Subject Sample	# 1	0.150	05:25
Breath Volume		1.843	05:25
Air Blank		0.000	05:26
Air Blank		0.000	05:28
Subject Sample	# 2	0.154	05:29
Breath Volume		3.574	05:29
Air Blank		0.000	05:29
Control Test		0.081	05:30
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSKO,DAVID,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROSKO,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:01

SUBJECT'S LAST NAME: MARTINELLI

DATE/TIME OF TEST: 02/04/2017 09:28:47

Diagnostics Check		OK	09:22
Air Blank		0.000	09:22
Control Test		0.081	09:23
Air Blank		0.000	09:23
Subject Sample	# 1	0.000	09:24
Breath Volume		3.835	09:24
Air Blank		0.000	09:24
Air Blank		0.000	09:26
Subject Sample	# 2	0.000	09:27
Breath Volume		3.960	09:27
Air Blank		0.000	09:27
Control Test		0.078	09:28
Air Blank		0.000	09:28
Diagnostics Check		OK	09:28

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSKO,DAVID,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROSKO,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: JABLONSKI

DATE/TIME OF TEST: 02/04/2017 19:08:48

Diagnostics Check		OK	19:00	
Air Blank		0.000	19:00	
Control Test		0.083	19:01	
Air Blank		0.000	19:01	
Subject Sample	# 1		19:04	No Sample Provided
Breath Volume		0.000	19:04	
Air Blank		0.000	19:05	
Air Blank		0.000	19:07	
Subject Sample	# 2		19:07	Subject Test Refused
Breath Volume		0.000	19:07	
Air Blank		0.000	19:07	
Control Test		0.083	19:08	
Air Blank		0.000	19:08	
Diagnostics Check		OK	19:08	

Data Download Date/Time: 02/27/2017 13:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: TROISE

DATE/TIME OF TEST: 02/05/2017 23:24:59

Diagnostics Check		OK	23:17
Air Blank		0.000	23:18
Control Test		0.082	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.078	23:20
Breath Volume		1.648	23:20
Air Blank		0.000	23:21
Air Blank		0.000	23:22
Subject Sample	# 2	0.079	23:23
Breath Volume		2.253	23:23
Air Blank		0.000	23:24
Control Test		0.082	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: TULLOCH

DATE/TIME OF TEST: 02/06/2017 01:22:23

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.081	01:16
Air Blank		0.000	01:16
Subject Sample	# 1	0.000	01:18
Breath Volume		1.675	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.000	01:21
Breath Volume		1.789	01:21
Air Blank		0.000	01:21
Control Test		0.081	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/06/2017 02:47:17

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.129	02:40	Volume Not Met
Breath Volume		0.363	02:40	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.126	02:45	Volume Not Met
Breath Volume		0.238	02:45	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 02/06/2017 03:08:03

Diagnostics Check		OK	02:59	
Air Blank		0.000	03:00	
Control Test		0.080	03:00	
Air Blank		0.000	03:00	
Subject Sample	# 1	0.133	03:04	Volume Not Met
Breath Volume		0.261	03:04	
Air Blank		0.000	03:04	
Air Blank		0.000	03:06	
Subject Sample	# 2		03:06	Subject Test Refused
Breath Volume		0.000	03:06	
Air Blank		0.000	03:07	
Control Test		0.078	03:07	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:08	

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLEMS,JASON,M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARROL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: BUI
ARREST TIME: 15:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: TIBOR

DATE/TIME OF TEST: 02/08/2017 17:09:58

Diagnostics Check		OK	16:59	
Air Blank		0.000	16:59	
Control Test		0.081	16:59	
Air Blank		0.000	17:00	
Subject Sample	# 1		17:03	No Sample Provided
Breath Volume		0.000	17:03	
Air Blank		0.000	17:04	
Air Blank		0.000	17:06	
Subject Sample	# 2		17:08	Subject Test Refused
Breath Volume		0.000	17:08	
Air Blank		0.000	17:09	
Control Test		0.081	17:09	
Air Blank		0.000	17:09	
Diagnostics Check		OK	17:09	

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STORY

DATE/TIME OF TEST: 02/08/2017 21:44:23

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:41	
Control Test		0.081	21:41	
Air Blank		0.000	21:42	
Subject Sample	# 1		21:42	Slope Not Met
Breath Volume		0.980	21:42	
Air Blank		0.000	21:43	
Control Test		0.082	21:43	
Air Blank		0.000	21:44	
Diagnostics Check		OK	21:44	

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STORY

DATE/TIME OF TEST: 02/08/2017 21:51:19

Diagnostics Check		OK	21:45
Air Blank		0.000	21:45
Control Test		0.081	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.132	21:46
Breath Volume		2.562	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.130	21:49
Breath Volume		2.570	21:49
Air Blank		0.000	21:50
Control Test		0.081	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: HIGHSTONE

DATE/TIME OF TEST: 02/10/2017 00:54:57

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.081	00:49
Air Blank		0.000	00:50
Subject Sample	# 1	0.000	00:50
Breath Volume		1.750	00:50
Air Blank		0.000	00:51
Air Blank		0.000	00:53
Subject Sample	# 2	0.000	00:53
Breath Volume		1.417	00:53
Air Blank		0.000	00:54
Control Test		0.082	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:08

SUBJECT'S LAST NAME: GONZALEZDOMINGO

DATE/TIME OF TEST: 02/10/2017 15:47:08

Diagnostics Check		OK	15:39
Air Blank		0.000	15:40
Control Test		0.081	15:40
Air Blank		0.000	15:40
Subject Sample	# 1	0.211	15:42
Breath Volume		1.824	15:42
Air Blank		0.000	15:43
Air Blank		0.000	15:44
Subject Sample	# 2	0.227	15:45
Breath Volume		1.750	15:45
Air Blank		0.000	15:46
Control Test		0.080	15:46
Air Blank		0.000	15:47
Diagnostics Check		OK	15:47

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 02/11/2017 03:24:42

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.223	03:19	
Breath Volume		1.187	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 2		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:23	
Control Test		0.081	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:24	

Data Download Date/Time: 02/27/2017 13:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINTO,LAWRENCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: QUINTO,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:37

SUBJECT'S LAST NAME: LABORDE

DATE/TIME OF TEST: 02/11/2017 15:07:31

Diagnostics Check		OK	14:58
Air Blank		0.000	14:58
Control Test		0.081	14:58
Air Blank		0.000	14:59
Subject Sample	# 1	0.116	15:01
Breath Volume		1.214	15:01
Air Blank		0.000	15:02
Air Blank		0.000	15:03
Subject Sample	# 2	0.120	15:05
Breath Volume		1.378	15:05
Air Blank		0.000	15:06
Control Test		0.080	15:06
Air Blank		0.000	15:07
Diagnostics Check		OK	15:07

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 02/11/2017 20:49:59

Diagnostics Check		OK	20:43
Air Blank		0.000	20:44
Control Test		0.081	20:44
Air Blank		0.000	20:45
Subject Sample	# 1	0.000	20:45
Breath Volume		2.949	20:45
Air Blank		0.000	20:46
Air Blank		0.000	20:48
Subject Sample	# 2	0.000	20:48
Breath Volume		1.210	20:48
Air Blank		0.000	20:49
Control Test		0.081	20:49
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VOGEL III

DATE/TIME OF TEST: 02/15/2017 00:38:58

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.081	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.000	00:34
Breath Volume		2.082	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.000	00:37
Breath Volume		1.855	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:38
Diagnostics Check		OK	00:38

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: TURNESA

DATE/TIME OF TEST: 02/15/2017 03:14:34

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.112	03:09
Breath Volume		1.699	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.112	03:13
Breath Volume		2.460	03:13
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGERVIL,CLAUDY,P
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: ANGERVIL,CLAUDY
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 02/16/2017 23:07:30

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:01	
Control Test		0.081	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:05	No Sample Provided
Breath Volume		0.000	23:05	
Air Blank		0.000	23:06	
Air Blank		0.000	23:07	

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: ECHEVERRIA

DATE/TIME OF TEST: 02/18/2017 02:10:34

Diagnostics Check		OK	01:56
Air Blank		0.000	01:56
Control Test		0.081	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.268	01:59
Breath Volume		1.320	01:59
Air Blank		0.000	01:59
Air Blank		0.000	02:01
Subject Sample	# 2	0.220	02:03
Breath Volume		1.593	02:03
Air Blank		0.000	02:04
Air Blank		0.000	02:06
Subject Sample	# 3	0.222	02:08
Breath Volume		2.578	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINES,LAUREN,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: HINES,LAUREN
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: WILSTERMANN

DATE/TIME OF TEST: 02/18/2017 02:29:16

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:26	
Control Test		0.079	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:27	Subject Test Refused
Breath Volume		0.000	02:27	
Air Blank		0.000	02:28	
Control Test		0.080	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/20/2017 21:12:03

Diagnostics Check		OK	21:05
Air Blank		0.000	21:06
Control Test		0.081	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.000	21:07
Breath Volume		3.054	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.000	21:10
Breath Volume		3.183	21:10
Air Blank		0.000	21:11
Control Test		0.082	21:11
Air Blank		0.000	21:11
Diagnostics Check		OK	21:12

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: KEMPINSKI

DATE/TIME OF TEST: 02/23/2017 23:02:12

Diagnostics Check		OK	22:56
Air Blank		0.000	22:56
Control Test		0.081	22:57
Air Blank		0.000	22:57
Subject Sample	# 1	0.000	22:57
Breath Volume		3.480	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.000	23:00
Breath Volume		3.421	23:00
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PITTMAN

DATE/TIME OF TEST: 02/24/2017 02:35:52

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.064	02:31
Breath Volume		2.925	02:31
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.064	02:34
Breath Volume		4.503	02:34
Air Blank		0.000	02:34
Control Test		0.080	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:35

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: BRUEGEMAN

DATE/TIME OF TEST: 02/24/2017 03:48:54

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.116	03:44
Breath Volume		1.253	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:46
Subject Sample	# 2	0.115	03:47
Breath Volume		1.847	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:48
Air Blank		0.000	03:48
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: JOSLIN

DATE/TIME OF TEST: 02/25/2017 03:35:20

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.212	03:30
Breath Volume		1.671	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.218	03:33
Breath Volume		2.964	03:33
Air Blank		0.000	03:34
Control Test		0.079	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,BRADFORD,B
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BAKER,BRADFORD
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/25/2017 04:43:19

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.189	04:38
Breath Volume		1.855	04:38
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.189	04:41
Breath Volume		2.605	04:41
Air Blank		0.000	04:42
Control Test		0.080	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 01/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: LAZAK

DATE/TIME OF TEST: 02/26/2017 01:21:23

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.257	01:13
Breath Volume		1.726	01:13
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.286	01:16
Breath Volume		1.250	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 3	0.281	01:19
Breath Volume		1.562	01:19
Air Blank		0.000	01:20
Control Test		0.077	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 02/27/2017 13:37

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSKO,DAVID,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: ROSKO,DAVID
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:01

SUBJECT'S LAST NAME: MONTH-VILLATORO

DATE/TIME OF TEST: 02/27/2017 16:35:08

Diagnostics Check		OK	16:30	
Air Blank		0.000	16:30	
Control Test		0.082	16:31	
Air Blank		0.000	16:31	
Subject Sample	# 1		16:33	Subject Test Refused
Breath Volume		0.000	16:33	
Air Blank		0.000	16:34	
Control Test		0.082	16:34	
Air Blank		0.000	16:35	
Diagnostics Check		OK	16:35	

Data Download Date/Time: 03/28/2017 12:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006168

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 02/28/2017 02:06:33

Diagnostics Check		OK	02:00
Air Blank		0.000	02:00
Control Test		0.081	02:01
Air Blank		0.000	02:01
Subject Sample	# 1	0.163	02:02
Breath Volume		1.730	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.157	02:05
Breath Volume		1.582	02:05
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 03/28/2017 12:12

Printed: 04 June 2018

COBRA ©1997-2014 CMI, Inc.