

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 03/01/2017 02:07:33

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Subject Sample # 1	0.189	02:02
Breath Volume	2.976	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:05
Subject Sample # 2	0.192	02:05
Breath Volume	3.281	02:05
Air Blank	0.000	02:06
Control Test	0.078	02:06
Air Blank	0.000	02:07
Diagnostics Check	OK	02:07

Data Download Date/Time: 03/28/2017 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTAMIRANDA,ANTONIO,G
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: ALTAMIRANDA,ANTONIO
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 03/04/2017 20:23:46

Diagnostics Check	OK	20:14
Air Blank	0.000	20:14
Control Test	0.078	20:14
Air Blank	0.000	20:15
Subject Sample # 1	0.268	20:15
Breath Volume	2.695	20:15
Air Blank	0.000	20:16
Air Blank	0.000	20:18
Subject Sample # 2	0.317	20:19
Breath Volume	2.593	20:19
Air Blank	0.000	20:19
Air Blank	0.000	20:21
Subject Sample # 3	0.305	20:22
Breath Volume	2.917	20:22
Air Blank	0.000	20:22
Control Test	0.077	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 03/28/2017 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECCI,Z,A
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: PECCI,Z
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HILLIARD

DATE/TIME OF TEST: 03/09/2017 00:33:17

Diagnostics Check	OK	00:21	
Air Blank	0.000	00:21	
Control Test	0.079	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1 0.170	00:23	
Breath Volume	1.750	00:23	
Air Blank	0.000	00:24	
Air Blank	0.000	00:25	
Subject Sample	# 2	00:29	No Sample Provided
Breath Volume	0.000	00:29	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 3	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:32	
Control Test	0.077	00:32	
Air Blank	0.000	00:33	
Diagnostics Check	OK	00:33	

Data Download Date/Time: 03/28/2017 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: GUTIERREZ DE LEON

DATE/TIME OF TEST: 03/23/2017 19:56:52

Diagnostics Check	OK	19:48
Air Blank	0.000	19:48
Control Test	0.079	19:49
Air Blank	0.000	19:49
Subject Sample # 1	0.225	19:51
Breath Volume	1.984	19:51
Air Blank	0.000	19:51
Air Blank	0.000	19:53
Subject Sample # 2	0.217	19:55
Breath Volume	1.984	19:55
Air Blank	0.000	19:55
Control Test	0.078	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:56

Data Download Date/Time: 03/28/2017 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WATTERS

DATE/TIME OF TEST: 03/24/2017 00:25:48

Diagnostics Check	OK	00:22	
Air Blank	0.000	00:22	
Control Test	0.078	00:23	
Air Blank	0.000	00:23	
Subject Sample # 1	0.176	00:24	
Breath Volume	2.953	00:24	
Air Blank		00:24	RFI Detect
Air Blank	0.000	00:24	
Control Test	0.079	00:25	
Air Blank	0.000	00:25	
Diagnostics Check	OK	00:25	

Data Download Date/Time: 03/28/2017 12:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006170

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31715080A3
DRY GAS EXPIRY DATE: 12/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:57

SUBJECT'S LAST NAME: WATTERS

DATE/TIME OF TEST: 03/24/2017 00:33:39

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.078	00:28
Air Blank	0.000	00:28
Subject Sample # 1	0.157	00:29
Breath Volume	1.925	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.158	00:32
Breath Volume	1.812	00:32
Air Blank	0.000	00:32
Control Test	0.078	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 03/28/2017 12:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CHEVALIER,JEAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:10

SUBJECT'S LAST NAME: DIAZ PEREZ

DATE/TIME OF TEST: 03/05/2017 14:39:30

Diagnostics Check	OK	14:32
Air Blank	0.000	14:33
Control Test	0.080	14:33
Air Blank	0.000	14:34
Subject Sample # 1	0.130	14:34
Breath Volume	1.335	14:34
Air Blank	0.000	14:35
Air Blank	0.000	14:36
Subject Sample # 2	0.125	14:37
Breath Volume	2.000	14:37
Air Blank	0.000	14:38
Control Test	0.081	14:38
Air Blank	0.000	14:39
Diagnostics Check	OK	14:39

Data Download Date/Time: 03/29/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,JOSE,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUIRRE,RICARDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 03/06/2017 01:57:08

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Subject Sample # 1	0.272	01:51
Breath Volume	1.917	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.259	01:55
Breath Volume	2.250	01:55
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 03/29/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 14:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 03/11/2017 15:34:38

Diagnostics Check	OK	15:26	
Air Blank	0.000	15:26	
Control Test	0.081	15:26	
Air Blank	0.000	15:27	
Subject Sample	# 1	15:30	No Sample Provided
Breath Volume	0.000	15:30	
Air Blank	0.000	15:30	
Air Blank	0.000	15:32	
Subject Sample	# 2	15:33	Subject Test Refused
Breath Volume	0.000	15:33	
Air Blank	0.000	15:33	
Control Test	0.079	15:34	
Air Blank	0.000	15:34	
Diagnostics Check	OK	15:34	

Data Download Date/Time: 03/29/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GILPIN,S
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: PEREZ PEREZ

DATE/TIME OF TEST: 03/12/2017 21:38:39

Diagnostics Check	OK	21:32
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Subject Sample # 1	0.212	21:33
Breath Volume	2.546	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.214	21:36
Breath Volume	2.671	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:38
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/29/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,JOSE,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: SINCORE,MICHAEL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DOUGLAS

DATE/TIME OF TEST: 03/13/2017 05:52:17

Diagnostics Check	OK	05:43	
Air Blank	0.000	05:43	
Control Test	0.079	05:44	
Air Blank	0.000	05:44	
Subject Sample	# 1 0.173	05:47	Volume Not Met
Breath Volume	0.339	05:47	
Air Blank	0.000	05:48	
Air Blank	0.000	05:50	
Subject Sample	# 2	05:50	Subject Test Refused
Breath Volume	0.542	05:50	
Air Blank	0.000	05:51	
Control Test	0.080	05:51	
Air Blank	0.000	05:52	
Diagnostics Check	OK	05:52	

Data Download Date/Time: 03/29/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON, LOUIS, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: TRUX, JOHN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 03/13/2017 17:20:23

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.081	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.182	17:15
Breath Volume	1.511	17:15
Air Blank	0.000	17:16
Air Blank	0.000	17:18
Subject Sample # 2	0.171	17:18
Breath Volume	1.496	17:18
Air Blank	0.000	17:19
Control Test	0.077	17:19
Air Blank	0.000	17:20
Diagnostics Check	OK	17:20

Data Download Date/Time: 03/29/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBISON, LOUIS, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: ROBISON, LOUIS
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: ADMIN
ARREST TIME: 09:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:45

SUBJECT'S LAST NAME: VAZQUEZ

DATE/TIME OF TEST: 03/26/2017 10:12:30

Diagnostics Check	OK	10:06
Air Blank	0.000	10:06
Control Test	0.081	10:07
Air Blank	0.000	10:07
Subject Sample # 1	0.000	10:08
Breath Volume	2.941	10:08
Air Blank	0.000	10:08
Air Blank	0.000	10:10
Subject Sample # 2	0.000	10:11
Breath Volume	2.144	10:11
Air Blank	0.000	10:11
Control Test	0.078	10:11
Air Blank	0.000	10:12
Diagnostics Check	OK	10:12

Data Download Date/Time: 03/29/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,JOEDICKE
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 03/10/2017 01:38:59

Diagnostics Check	OK	01:32
Air Blank	0.000	01:33
Control Test	0.078	01:33
Air Blank	0.000	01:33
Subject Sample # 1	0.184	01:34
Breath Volume	2.402	01:34
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.165	01:37
Breath Volume	1.214	01:37
Air Blank	0.000	01:38
Control Test	0.078	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 03/28/2017 23:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZYCKI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WEIMER,ADAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: KACHEL

DATE/TIME OF TEST: 03/04/2017 20:50:22

Diagnostics Check	OK	20:44
Air Blank	0.000	20:45
Control Test	0.078	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.000	20:46
Breath Volume	2.980	20:46
Air Blank	0.000	20:46
Air Blank	0.000	20:48
Subject Sample # 2	0.000	20:48
Breath Volume	3.570	20:48
Air Blank	0.000	20:49
Control Test	0.078	20:49
Air Blank	0.000	20:50
Diagnostics Check	OK	20:50

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRONIN,SEAN,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WEIMER,ADAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: KACHEL

DATE/TIME OF TEST: 03/04/2017 21:01:35

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:58	
Control Test	0.078	20:59	
Air Blank	0.000	20:59	
Subject Sample	# 1	21:01	Sequence Aborted
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS, HEATHER, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HAAS, GARY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 03/04/2017 21:33:29

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.078	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.161	21:26
Breath Volume	1.140	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:28
Subject Sample # 2	0.155	21:31
Breath Volume	1.929	21:31
Air Blank	0.000	21:32
Control Test	0.078	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART, PAULETTE,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT, TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: WOLYNSKI

DATE/TIME OF TEST: 03/05/2017 02:57:35

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.078	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.182	02:52
Breath Volume	4.343	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.186	02:55
Breath Volume	3.718	02:55
Air Blank	0.000	02:56
Control Test	0.077	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,LAUREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: BARRIOS

DATE/TIME OF TEST: 03/07/2017 03:31:41

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.078	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.265	03:26
Breath Volume	2.789	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.255	03:29
Breath Volume	2.984	03:29
Air Blank	0.000	03:30
Control Test	0.075	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LOUCEL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/07/2017 04:54:32

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.077	04:46	
Air Blank	0.000	04:46	
Subject Sample	# 1 0.182	04:49	
Breath Volume	1.164	04:49	
Air Blank	0.000	04:50	
Air Blank	0.000	04:51	
Subject Sample	# 2 0.179	04:52	
Breath Volume	1.277	04:52	
Air Blank	0.000	04:53	
Control Test	0.074	04:53	Control Outside Tolerance
Air Blank	0.000	04:54	

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LOUCEL,MICHAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 03/07/2017 05:07:07

Diagnostics Check	OK	04:57	
Air Blank	0.000	04:58	
Control Test	0.077	04:58	
Air Blank	0.000	04:58	
Subject Sample	# 1 0.172	05:01	
Breath Volume	1.531	05:01	
Air Blank	0.000	05:01	
Air Blank	0.000	05:03	
Subject Sample	# 2 0.176	05:05	
Breath Volume	1.757	05:05	
Air Blank	0.000	05:06	
Control Test	0.073	05:06	Control Outside Tolerance
Air Blank	0.000	05:07	

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 03/10/2017 23:41:37

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:32	
Control Test	0.078	23:32	
Air Blank	0.000	23:33	
Subject Sample	# 1 0.307	23:36	Volume Not Met
Breath Volume	0.140	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 2	23:39	Subject Test Refused
Breath Volume	0.347	23:39	
Air Blank	0.000	23:40	
Control Test	0.072	23:41	Control Outside Tolerance
Air Blank	0.000	23:41	

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM,
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: JEFFRIES

DATE/TIME OF TEST: 03/11/2017 00:02:44

Diagnostics Check	OK	23:57	
Air Blank	0.000	23:57	
Control Test	0.076	23:57	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.265	00:01	
Air Blank	0.000	00:01	
Control Test	0.077	00:02	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS, HEATHER, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS, HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 03/11/2017 02:36:16

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.077	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.100	02:31
Breath Volume	1.535	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.107	02:34
Breath Volume	1.406	02:34
Air Blank	0.000	02:35
Control Test	0.077	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WYNN,JOSEPH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 03/11/2017 07:21:53

Diagnostics Check	OK	07:15
Air Blank	0.000	07:16
Control Test	0.078	07:16
Air Blank	0.000	07:16
Subject Sample # 1	0.000	07:17
Breath Volume	1.761	07:17
Air Blank	0.000	07:17
Air Blank	0.000	07:19
Subject Sample # 2	0.000	07:20
Breath Volume	2.148	07:20
Air Blank	0.000	07:20
Control Test	0.076	07:21
Air Blank	0.000	07:21
Diagnostics Check	OK	07:21

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LAKEMAN,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:51

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 03/11/2017 22:26:24

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Subject Sample # 1	0.149	22:21
Breath Volume	2.441	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.132	22:24
Breath Volume	1.410	22:24
Air Blank	0.000	22:25
Control Test	0.078	22:25
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: FALERO

DATE/TIME OF TEST: 03/11/2017 22:39:16

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.077	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.028	22:34
Breath Volume	4.328	22:34
Air Blank	0.000	22:35
Air Blank	0.000	22:36
Subject Sample # 2	0.025	22:37
Breath Volume	4.539	22:37
Air Blank	0.000	22:38
Control Test	0.076	22:38
Air Blank	0.000	22:39
Diagnostics Check	OK	22:39

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: ENGEL

DATE/TIME OF TEST: 03/12/2017 20:45:05

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.078	20:38
Air Blank	0.000	20:38
Subject Sample # 1	0.135	20:40
Breath Volume	1.910	20:40
Air Blank	0.000	20:40
Air Blank	0.000	20:42
Subject Sample # 2	0.134	20:43
Breath Volume	1.480	20:43
Air Blank	0.000	20:44
Control Test	0.077	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LISTON,TIMOTHY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 03/16/2017 00:23:52

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.078	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.176	00:19
Breath Volume	2.179	00:19
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.181	00:22
Breath Volume	1.964	00:22
Air Blank	0.000	00:22
Control Test	0.077	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS, HEATHER, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS, HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: LONDAITZ

DATE/TIME OF TEST: 03/16/2017 02:25:06

Diagnostics Check	OK	02:18
Air Blank	0.000	02:18
Control Test	0.078	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.085	02:20
Breath Volume	2.886	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.082	02:23
Breath Volume	2.617	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILKIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: ARGUETA CORADO

DATE/TIME OF TEST: 03/16/2017 03:47:17

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.078	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.150	03:42
Breath Volume	2.468	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.148	03:45
Breath Volume	2.367	03:45
Air Blank	0.000	03:46
Control Test	0.078	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: ZENK

DATE/TIME OF TEST: 03/17/2017 01:21:27

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:15	
Control Test	0.078	01:15	
Air Blank	0.000	01:16	
Subject Sample # 1	0.147	01:17	
Breath Volume	1.222	01:17	
Air Blank	0.000	01:17	
Air Blank	0.000	01:19	
Subject Sample # 2	0.153	01:20	
Breath Volume	2.582	01:20	
Air Blank		01:20	RFI Detect
Air Blank	0.000	01:21	

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS, HEATHER, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS, HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: BEASTON

DATE/TIME OF TEST: 03/17/2017 01:34:59

Diagnostics Check	OK	01:26
Air Blank	0.000	01:27
Control Test	0.077	01:27
Air Blank	0.000	01:27
Subject Sample # 1	0.171	01:29
Breath Volume	1.988	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.163	01:33
Breath Volume	1.968	01:33
Air Blank	0.000	01:34
Control Test	0.077	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWARD, WILLIAM, J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HOWARD, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: ZENK

DATE/TIME OF TEST: 03/17/2017 01:46:54

Diagnostics Check	OK	01:39
Air Blank	0.000	01:40
Control Test	0.078	01:40
Air Blank	0.000	01:41
Subject Sample # 1	0.152	01:41
Breath Volume	2.085	01:41
Air Blank	0.000	01:42
Air Blank	0.000	01:44
Subject Sample # 2	0.150	01:45
Breath Volume	2.144	01:45
Air Blank	0.000	01:45
Control Test	0.076	01:46
Air Blank	0.000	01:46
Diagnostics Check	OK	01:46

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS, HEATHER, C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS, HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: HOLMQUIST

DATE/TIME OF TEST: 03/19/2017 00:43:00

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.078	00:29	
Air Blank	0.000	00:30	
Subject Sample	# 1 0.078	00:33	Volume Not Met
Breath Volume	0.273	00:33	
Air Blank	0.000	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 2 0.125	00:37	
Breath Volume	1.816	00:37	
Air Blank	0.000	00:38	
Air Blank	0.000	00:40	
Subject Sample	# 3 0.119	00:41	
Breath Volume	1.296	00:41	
Air Blank	0.000	00:42	
Control Test	0.077	00:42	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 03/22/2017 19:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 03/21/2017 03:06:05

Diagnostics Check	OK	03:02	
Air Blank	0.000	03:03	
Control Test	0.078	03:03	
Air Blank	0.000	03:04	
Subject Sample	# 1	03:04	Subject Test Refused
Breath Volume	0.000	03:04	
Air Blank	0.000	03:05	
Control Test	0.078	03:05	
Air Blank	0.000	03:05	
Diagnostics Check	OK	03:06	

Data Download Date/Time: 03/22/2017 19:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CANO,MYRA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: LEYVA

DATE/TIME OF TEST: 03/03/2017 21:13:29

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.081	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:12	Subject Test Refused
Breath Volume	0.000	21:12	
Air Blank	0.000	21:12	
Control Test	0.082	21:12	
Air Blank	0.000	21:13	
Diagnostics Check	OK	21:13	

Data Download Date/Time: 03/30/2017 04:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CAMERON,DAVID
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MARNEY

DATE/TIME OF TEST: 03/18/2017 21:31:18

Diagnostics Check	OK	21:24
Air Blank	0.000	21:25
Control Test	0.082	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.199	21:26
Breath Volume	4.265	21:26
Air Blank	0.000	21:27
Air Blank	0.000	21:28
Subject Sample # 2	0.189	21:29
Breath Volume	4.812	21:29
Air Blank	0.000	21:30
Control Test	0.078	21:30
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 03/30/2017 04:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,M,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: KEARNS,H
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1694353
DRY GAS EXPIRY DATE: 06/10/2017
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: ALMRAZRAFAEL

DATE/TIME OF TEST: 03/26/2017 00:35:54

Diagnostics Check	OK	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:29
Subject Sample # 1	0.199	00:30
Breath Volume	2.343	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.196	00:34
Breath Volume	1.875	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:35
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 03/30/2017 04:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: TOLSON

DATE/TIME OF TEST: 03/04/2017 21:20:47

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.082	21:14
Air Blank	0.000	21:15
Subject Sample # 1	0.176	21:15
Breath Volume	1.699	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.168	21:19
Breath Volume	1.824	21:19
Air Blank	0.000	21:19
Control Test	0.082	21:20
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 03/27/2017 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CHERVIOVSKIY

DATE/TIME OF TEST: 03/09/2017 23:18:30

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.081	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.196	23:13
Breath Volume	1.785	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.210	23:16
Breath Volume	3.019	23:16
Air Blank	0.000	23:17
Control Test	0.080	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 03/27/2017 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: MORANCY

DATE/TIME OF TEST: 03/15/2017 02:23:55

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.082	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.191	02:17
Breath Volume	1.222	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:19
Subject Sample # 2	0.181	02:22
Breath Volume	1.203	02:22
Air Blank	0.000	02:22
Control Test	0.078	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 03/27/2017 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GONZALEZ CAMARILLO

DATE/TIME OF TEST: 03/19/2017 03:25:26

Diagnostics Check	OK	03:17
Air Blank	0.000	03:17
Control Test	0.082	03:17
Air Blank	0.000	03:18
Subject Sample # 1	0.201	03:19
Breath Volume	1.742	03:19
Air Blank	0.000	03:20
Air Blank	0.000	03:22
Subject Sample # 2	0.203	03:23
Breath Volume	2.972	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 03/27/2017 09:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MORANDO,OFFICER
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: KUSHNER

DATE/TIME OF TEST: 03/01/2017 16:26:25

Diagnostics Check	OK	16:09	
Air Blank	0.000	16:10	
Control Test	0.080	16:10	
Air Blank	0.000	16:11	
Subject Sample	# 1 0.260	16:14	Volume Not Met
Breath Volume	0.035	16:14	
Air Blank	0.000	16:15	
Air Blank	0.000	16:16	
Subject Sample	# 2 0.306	16:18	
Breath Volume	1.214	16:18	
Air Blank	0.000	16:19	
Air Blank	0.000	16:21	
Subject Sample	# 3 0.215	16:24	Volume Not Met
Breath Volume	0.445	16:24	
Air Blank	0.000	16:25	
Control Test	0.077	16:25	
Air Blank	0.000	16:26	
Diagnostics Check	OK	16:26	

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,CODY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TENNANT,R
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: MAHER

DATE/TIME OF TEST: 03/04/2017 03:57:14

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.082	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.220	03:52
Breath Volume	3.484	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:54
Subject Sample # 2	0.223	03:55
Breath Volume	3.296	03:55
Air Blank	0.000	03:56
Control Test	0.081	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SENAK,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 03/04/2017 04:22:06

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.081	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.164	04:17
Breath Volume	2.261	04:17
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.161	04:20
Breath Volume	2.449	04:20
Air Blank	0.000	04:21
Control Test	0.081	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,CODY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MAHOOL,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LALLA

DATE/TIME OF TEST: 03/05/2017 04:10:38

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.113	04:05
Breath Volume	1.847	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.108	04:09
Breath Volume	3.613	04:09
Air Blank	0.000	04:09
Control Test	0.080	04:10
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,CODY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: MORALES

DATE/TIME OF TEST: 03/05/2017 22:12:42

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.080	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.193	22:07
Breath Volume	2.074	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:09
Subject Sample # 2	0.189	22:10
Breath Volume	1.593	22:10
Air Blank	0.000	22:11
Control Test	0.080	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,CODY,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CHECO,E
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/06/2017 00:37:54

Diagnostics Check	OK	00:31
Air Blank	0.000	00:31
Control Test	0.080	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.242	00:33
Breath Volume	2.812	00:33
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.232	00:36
Breath Volume	2.945	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:37
Air Blank	0.000	00:37
Diagnostics Check	OK	00:37

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON, MATTHEW, P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SEAGLE, C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MCDONALD

DATE/TIME OF TEST: 03/11/2017 04:01:47

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.158	03:56
Breath Volume	1.988	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.168	04:00
Breath Volume	2.433	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RANDOLPH,K,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: HANSELL,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 03/18/2017 07:26:13

Diagnostics Check	OK	07:19
Air Blank	0.000	07:19
Control Test	0.081	07:20
Air Blank	0.000	07:20
Subject Sample # 1	0.179	07:21
Breath Volume	1.355	07:21
Air Blank	0.000	07:21
Air Blank	0.000	07:23
Subject Sample # 2	0.187	07:24
Breath Volume	2.210	07:24
Air Blank	0.000	07:25
Control Test	0.081	07:25
Air Blank	0.000	07:26
Diagnostics Check	OK	07:26

Data Download Date/Time: 03/18/2017 11:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HIDALGO LOPEZ

DATE/TIME OF TEST: 03/01/2017 04:13:45

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.231	04:08
Breath Volume	2.425	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.224	04:12
Breath Volume	1.402	04:12
Air Blank	0.000	04:12
Control Test	0.078	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAFFER,NANCY,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: COMBS,B
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:47

SUBJECT'S LAST NAME: DOS SANTOS

DATE/TIME OF TEST: 03/01/2017 12:19:41

Diagnostics Check	OK	12:13
Air Blank	0.000	12:13
Control Test	0.080	12:13
Air Blank	0.000	12:14
Subject Sample # 1	0.135	12:15
Breath Volume	2.804	12:15
Air Blank	0.000	12:15
Air Blank	0.000	12:17
Subject Sample # 2	0.133	12:18
Breath Volume	2.222	12:18
Air Blank	0.000	12:18
Control Test	0.080	12:19
Air Blank	0.000	12:19
Diagnostics Check	OK	12:19

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WHITEFORD

DATE/TIME OF TEST: 03/03/2017 00:23:14

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:15
Air Blank	0.000	00:16
Subject Sample # 1	0.273	00:17
Breath Volume	3.644	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.254	00:21
Breath Volume	1.898	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: MORALES COLON

DATE/TIME OF TEST: 03/03/2017 04:07:41

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.025	04:02
Breath Volume	2.589	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.025	04:06
Breath Volume	2.800	04:06
Air Blank	0.000	04:06
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WILLIAMS,JULIE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:21

SUBJECT'S LAST NAME: BERRY

DATE/TIME OF TEST: 03/03/2017 06:54:15

Diagnostics Check	OK	06:47
Air Blank	0.000	06:48
Control Test	0.080	06:48
Air Blank	0.000	06:49
Subject Sample # 1	0.064	06:49
Breath Volume	2.917	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:52
Subject Sample # 2	0.061	06:52
Breath Volume	1.750	06:52
Air Blank	0.000	06:53
Control Test	0.081	06:53
Air Blank	0.000	06:54
Diagnostics Check	OK	06:54

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: GOGAN

DATE/TIME OF TEST: 03/04/2017 01:22:42

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.081	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:18
Breath Volume	1.699	01:18
Air Blank	0.000	01:18
Air Blank	0.000	01:20
Subject Sample # 2	0.000	01:21
Breath Volume	2.371	01:21
Air Blank	0.000	01:21
Control Test	0.080	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male

OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 03/04/2017 03:18:51

Diagnostics Check	OK	03:12
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.098	03:13
Breath Volume	1.566	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.094	03:17
Breath Volume	1.667	03:17
Air Blank	0.000	03:17
Control Test	0.080	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON,ROBERT
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: TAMPOSI

DATE/TIME OF TEST: 03/04/2017 04:57:50

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.118	04:53
Breath Volume	1.988	04:53
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.121	04:56
Breath Volume	1.285	04:56
Air Blank	0.000	04:56
Control Test	0.079	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GELINA,PHILIP
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:17

SUBJECT'S LAST NAME: IANNOTTI

DATE/TIME OF TEST: 03/04/2017 16:50:12

Diagnostics Check	OK	16:43
Air Blank	0.000	16:44
Control Test	0.081	16:44
Air Blank	0.000	16:45
Subject Sample # 1	0.176	16:45
Breath Volume	2.511	16:45
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 2	0.161	16:48
Breath Volume	1.378	16:48
Air Blank	0.000	16:49
Control Test	0.080	16:49
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TRENTMAN,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/04/2017 17:47:26

Diagnostics Check	OK	17:41
Air Blank	0.000	17:41
Control Test	0.080	17:42
Air Blank	0.000	17:42
Subject Sample # 1	0.000	17:43
Breath Volume	2.585	17:43
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.000	17:46
Breath Volume	2.355	17:46
Air Blank	0.000	17:46
Control Test	0.080	17:46
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/04/2017 22:26:45

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:15	
Control Test	0.081	22:16	
Air Blank	0.000	22:16	
Subject Sample	# 1	22:19	No Sample Provided
Breath Volume	0.000	22:19	
Air Blank	0.000	22:20	
Air Blank	0.000	22:22	
Subject Sample	# 2	22:25	No Sample Provided
Breath Volume	0.000	22:25	
Air Blank	0.000	22:25	
Control Test	0.080	22:26	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female

OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: TERRY

DATE/TIME OF TEST: 03/04/2017 22:52:36

Diagnostics Check	OK	22:46
Air Blank	0.000	22:46
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.124	22:48
Breath Volume	2.023	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.123	22:51
Breath Volume	1.160	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KIPPLE

DATE/TIME OF TEST: 03/05/2017 01:08:08

Diagnostics Check	OK	00:50	
Air Blank	0.000	00:51	
Control Test	0.080	00:51	
Air Blank	0.000	00:51	
Subject Sample	# 1 0.105	00:54	Volume Not Met
Breath Volume	0.214	00:54	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2 0.128	01:00	
Breath Volume	2.261	01:00	
Air Blank	0.000	01:01	
Air Blank	0.000	01:03	
Subject Sample	# 3 0.104	01:06	Volume Not Met
Breath Volume	1.074	01:06	
Air Blank	0.000	01:07	
Control Test	0.078	01:07	
Air Blank	0.000	01:08	
Diagnostics Check	OK	01:08	

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KIPPLE

DATE/TIME OF TEST: 03/05/2017 01:17:45

Diagnostics Check	OK	01:09
Air Blank	0.000	01:10
Control Test	0.078	01:10
Air Blank	0.000	01:11
Subject Sample # 1	0.117	01:11
Breath Volume	1.488	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.115	01:16
Breath Volume	1.160	01:16
Air Blank	0.000	01:16
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CROY,PAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: DELICE

DATE/TIME OF TEST: 03/06/2017 12:46:43

Diagnostics Check	OK	12:39
Air Blank	0.000	12:39
Control Test	0.080	12:39
Air Blank	0.000	12:40
Subject Sample # 1	0.179	12:41
Breath Volume	1.343	12:41
Air Blank	0.000	12:42
Air Blank	0.000	12:44
Subject Sample # 2	0.186	12:45
Breath Volume	1.644	12:45
Air Blank	0.000	12:45
Control Test	0.080	12:46
Air Blank	0.000	12:46
Diagnostics Check	OK	12:46

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MALARCZYK,JR.

DATE/TIME OF TEST: 03/08/2017 20:31:19

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.080	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.198	20:26
Breath Volume	3.691	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.205	20:29
Breath Volume	4.335	20:29
Air Blank	0.000	20:30
Control Test	0.080	20:30
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCNAB,TARA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RINCON

DATE/TIME OF TEST: 03/10/2017 03:48:25

Diagnostics Check	OK	03:36	
Air Blank	0.000	03:37	
Control Test	0.081	03:37	
Air Blank	0.000	03:38	
Subject Sample	# 1 0.126	03:41	Volume Not Met
Breath Volume	0.085	03:41	
Air Blank	0.000	03:41	
Air Blank	0.000	03:43	
Subject Sample	# 2 0.150	03:46	Volume Not Met
Breath Volume	0.218	03:46	
Air Blank	0.000	03:47	
Control Test	0.079	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCNAB,TARA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RINCON

DATE/TIME OF TEST: 03/10/2017 04:01:02

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.080	03:50	
Air Blank	0.000	03:50	
Subject Sample	# 1 0.125	03:53	Volume Not Met
Breath Volume	0.089	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2 0.110	03:59	Volume Not Met
Breath Volume	0.089	03:59	
Air Blank	0.000	04:00	
Control Test	0.079	04:00	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:01	

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GELINA,PHILIP
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/10/2017 11:02:12

Diagnostics Check	OK	10:55
Air Blank	0.000	10:55
Control Test	0.080	10:55
Air Blank	0.000	10:56
Subject Sample # 1	0.229	10:57
Breath Volume	1.753	10:57
Air Blank	0.000	10:58
Air Blank	0.000	11:00
Subject Sample # 2	0.237	11:00
Breath Volume	1.910	11:00
Air Blank	0.000	11:01
Control Test	0.080	11:01
Air Blank	0.000	11:02
Diagnostics Check	OK	11:02

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: FESLER

DATE/TIME OF TEST: 03/10/2017 21:09:47

Diagnostics Check	OK	20:54	
Air Blank	0.000	20:55	
Control Test	0.080	20:55	
Air Blank	0.000	20:56	
Subject Sample	# 1 0.171	20:59	Volume Not Met
Breath Volume	0.292	20:59	
Air Blank	0.000	20:59	
Air Blank	0.000	21:01	
Subject Sample	# 2 0.195	21:04	
Breath Volume	1.132	21:04	
Air Blank	0.000	21:05	
Air Blank	0.000	21:07	
Subject Sample	# 3 0.189	21:08	
Breath Volume	1.109	21:08	
Air Blank	0.000	21:08	
Control Test	0.077	21:09	
Air Blank	0.000	21:09	
Diagnostics Check	OK	21:09	

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,E.K.
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 03/11/2017 02:34:24

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.255	02:29
Breath Volume	2.781	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.266	02:32
Breath Volume	2.480	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GREEN,JR.

DATE/TIME OF TEST: 03/11/2017 03:16:03

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.080	03:09
Air Blank	0.000	03:09
Subject Sample # 1	0.236	03:10
Breath Volume	1.847	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.229	03:14
Breath Volume	2.019	03:14
Air Blank	0.000	03:15
Control Test	0.079	03:15
Air Blank	0.000	03:15
Diagnostics Check	OK	03:16

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LOUIS,WALEX
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/11/2017 09:43:27

Diagnostics Check	OK	09:36
Air Blank	0.000	09:36
Control Test	0.080	09:37
Air Blank	0.000	09:37
Subject Sample # 1	0.129	09:38
Breath Volume	2.574	09:38
Air Blank	0.000	09:39
Air Blank	0.000	09:40
Subject Sample # 2	0.130	09:41
Breath Volume	3.222	09:41
Air Blank	0.000	09:42
Control Test	0.079	09:42
Air Blank	0.000	09:43
Diagnostics Check	OK	09:43

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LUFKIN,AIREL
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 03/11/2017 20:44:38

Diagnostics Check	OK	20:32	
Air Blank	0.000	20:33	
Control Test	0.080	20:33	
Air Blank	0.000	20:34	
Subject Sample	# 1 0.123	20:37	Volume Not Met
Breath Volume	0.195	20:37	
Air Blank	0.000	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 2 0.116	20:42	Volume Not Met
Breath Volume	0.113	20:42	
Air Blank	0.000	20:43	
Control Test	0.078	20:44	
Air Blank	0.000	20:44	
Diagnostics Check	OK	20:44	

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 03/12/2017 01:29:51

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.000	01:25
Breath Volume	1.140	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.000	01:28
Breath Volume	1.492	01:28
Air Blank	0.000	01:28
Control Test	0.080	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITH,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: SALES VELASQUEZ

DATE/TIME OF TEST: 03/12/2017 04:46:10

Diagnostics Check	OK	04:39
Air Blank	0.000	04:40
Control Test	0.080	04:40
Air Blank	0.000	04:41
Subject Sample # 1	0.174	04:41
Breath Volume	2.355	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.172	04:44
Breath Volume	1.953	04:44
Air Blank	0.000	04:45
Control Test	0.080	04:45
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MEDINA BARAHONA

DATE/TIME OF TEST: 03/12/2017 05:18:44

Diagnostics Check	OK	05:10
Air Blank	0.000	05:10
Control Test	0.080	05:10
Air Blank	0.000	05:11
Subject Sample # 1	0.113	05:12
Breath Volume	1.519	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.115	05:17
Breath Volume	1.683	05:17
Air Blank	0.000	05:17
Control Test	0.079	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IOZZIO,GINA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 03/12/2017 22:56:24

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.080	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.230	22:50
Breath Volume	3.003	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.235	22:54
Breath Volume	4.109	22:54
Air Blank	0.000	22:55
Control Test	0.080	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREG
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: PARASMO

DATE/TIME OF TEST: 03/14/2017 01:33:28

Diagnostics Check	OK	01:25
Air Blank	0.000	01:26
Control Test	0.081	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.298	01:28
Breath Volume	1.253	01:28
Air Blank	0.000	01:29
Air Blank	0.000	01:30
Subject Sample # 2	0.295	01:31
Breath Volume	1.363	01:31
Air Blank	0.000	01:32
Control Test	0.079	01:32
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:48

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 03/16/2017 08:18:26

Diagnostics Check	OK	08:11
Air Blank	0.000	08:12
Control Test	0.081	08:12
Air Blank	0.000	08:13
Subject Sample # 1	0.112	08:13
Breath Volume	3.281	08:13
Air Blank	0.000	08:14
Air Blank	0.000	08:16
Subject Sample # 2	0.111	08:16
Breath Volume	2.984	08:16
Air Blank	0.000	08:17
Control Test	0.081	08:17
Air Blank	0.000	08:18
Diagnostics Check	OK	08:18

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANTON,MELINDA
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: JEAN-LOUIS

DATE/TIME OF TEST: 03/18/2017 02:13:14

Diagnostics Check	OK	02:06
Air Blank	0.000	02:07
Control Test	0.081	02:07
Air Blank	0.000	02:08
Subject Sample # 1	0.167	02:08
Breath Volume	1.648	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:11
Subject Sample # 2	0.179	02:11
Breath Volume	1.363	02:11
Air Blank	0.000	02:12
Control Test	0.080	02:12
Air Blank	0.000	02:13
Diagnostics Check	OK	02:13

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CAPASSO

DATE/TIME OF TEST: 03/18/2017 04:07:58

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.153	04:03
Breath Volume	2.542	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:05
Subject Sample # 2	0.158	04:06
Breath Volume	3.300	04:06
Air Blank	0.000	04:07
Control Test	0.080	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: RIVERO GARCIA

DATE/TIME OF TEST: 03/18/2017 06:09:25

Diagnostics Check	OK	06:02
Air Blank	0.000	06:03
Control Test	0.080	06:03
Air Blank	0.000	06:04
Subject Sample # 1	0.191	06:04
Breath Volume	2.089	06:04
Air Blank	0.000	06:05
Air Blank	0.000	06:07
Subject Sample # 2	0.186	06:07
Breath Volume	1.679	06:07
Air Blank	0.000	06:08
Control Test	0.079	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: MURILLO MANZANARES

DATE/TIME OF TEST: 03/18/2017 07:58:19

Diagnostics Check	OK	07:51
Air Blank	0.000	07:52
Control Test	0.081	07:52
Air Blank	0.000	07:53
Subject Sample # 1	0.124	07:53
Breath Volume	1.656	07:53
Air Blank	0.000	07:54
Air Blank	0.000	07:56
Subject Sample # 2	0.122	07:56
Breath Volume	1.574	07:56
Air Blank	0.000	07:57
Control Test	0.080	07:57
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,ELIZA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:19

SUBJECT'S LAST NAME: PINEDA

DATE/TIME OF TEST: 03/18/2017 09:01:07

Diagnostics Check	OK	08:51
Air Blank	0.000	08:52
Control Test	0.080	08:52
Air Blank	0.000	08:52
Subject Sample # 1	0.182	08:55
Breath Volume	1.609	08:55
Air Blank	0.000	08:55
Air Blank	0.000	08:57
Subject Sample # 2	0.180	08:59
Breath Volume	1.441	08:59
Air Blank	0.000	09:00
Control Test	0.079	09:00
Air Blank	0.000	09:01
Diagnostics Check	OK	09:01

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TFRANK,ALEXANDER
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: UZHIKOVSKI

DATE/TIME OF TEST: 03/19/2017 02:00:23

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.081	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.238	01:55
Breath Volume	1.308	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.245	01:58
Breath Volume	1.101	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HERRERA CHACLAN

DATE/TIME OF TEST: 03/19/2017 04:06:12

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.081	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.202	04:01
Breath Volume	3.468	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:03
Subject Sample # 2	0.201	04:04
Breath Volume	3.601	04:04
Air Blank	0.000	04:05
Control Test	0.079	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FRANCIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/19/2017 23:26:53

Diagnostics Check	OK	23:11	
Air Blank	0.000	23:12	
Control Test	0.080	23:12	
Air Blank	0.000	23:13	
Subject Sample	# 1 0.332	23:13	No .020 agreement
Breath Volume	1.648	23:13	
Air Blank	0.000	23:14	
Air Blank	0.000	23:16	
Subject Sample	# 2 0.268	23:19	Volume Not Met
Breath Volume	0.820	23:19	
Air Blank	0.000	23:20	
Air Blank	0.000	23:22	
Subject Sample	# 3 0.307	23:24	No .020 agreement
Breath Volume	1.714	23:24	
Air Blank	0.000	23:25	
Control Test	0.076	23:26	
Air Blank	0.000	23:26	
Diagnostics Check	OK	23:26	

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FRANCIS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Riviera Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 03/19/2017 23:38:34

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.076	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.292	23:33
Breath Volume	1.136	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:35
Subject Sample # 2	0.309	23:36
Breath Volume	2.128	23:36
Air Blank	0.000	23:37
Control Test	0.077	23:37
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 03/24/2017 15:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: REYES REYES

DATE/TIME OF TEST: 03/20/2017 01:27:01

Diagnostics Check	OK	01:18
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.236	01:20
Breath Volume	1.980	01:20
Air Blank	0.000	01:21
Air Blank	0.000	01:23
Subject Sample # 2	0.246	01:25
Breath Volume	3.039	01:25
Air Blank	0.000	01:26
Control Test	0.080	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:27

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AIME,DIANA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: FLORIL

DATE/TIME OF TEST: 03/20/2017 02:00:07

Diagnostics Check	OK	01:52
Air Blank	0.000	01:52
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.256	01:54
Breath Volume	1.511	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.251	01:58
Breath Volume	2.171	01:58
Air Blank	0.000	01:59
Control Test	0.080	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 03/20/2017 03:27:50

Diagnostics Check	OK	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Subject Sample # 1	0.257	03:23
Breath Volume	1.937	03:23
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.260	03:26
Breath Volume	3.347	03:26
Air Blank	0.000	03:26
Control Test	0.080	03:27
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DISTEFANO,BRITTANY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 03/22/2017 03:17:18

Diagnostics Check	OK	03:09
Air Blank	0.000	03:10
Control Test	0.081	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.157	03:12
Breath Volume	2.085	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.152	03:15
Breath Volume	2.097	03:15
Air Blank	0.000	03:16
Control Test	0.080	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLEARY,GUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 03/22/2017 13:55:39

Diagnostics Check	OK	13:48
Air Blank	0.000	13:49
Control Test	0.080	13:49
Air Blank	0.000	13:50
Subject Sample # 1	0.010	13:51
Breath Volume	4.421	13:51
Air Blank	0.000	13:51
Air Blank	0.000	13:53
Subject Sample # 2	0.008	13:54
Breath Volume	4.347	13:54
Air Blank	0.000	13:54
Control Test	0.079	13:55
Air Blank	0.000	13:55
Diagnostics Check	OK	13:55

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IOZZIO,GINA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: ADAMO

DATE/TIME OF TEST: 03/22/2017 21:46:36

Diagnostics Check	OK	21:40
Air Blank	0.000	21:40
Control Test	0.080	21:40
Air Blank	0.000	21:41
Subject Sample # 1	0.000	21:42
Breath Volume	2.109	21:42
Air Blank	0.000	21:42
Air Blank	0.000	21:44
Subject Sample # 2	0.000	21:45
Breath Volume	1.933	21:45
Air Blank	0.000	21:45
Control Test	0.080	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: NASH

DATE/TIME OF TEST: 03/23/2017 02:03:23

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.081	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.293	01:58
Breath Volume	1.667	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.302	02:01
Breath Volume	1.218	02:01
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 03/23/2017 22:11:53

Diagnostics Check	OK	22:05
Air Blank	0.000	22:05
Control Test	0.079	22:06
Air Blank	0.000	22:06
Subject Sample # 1	0.166	22:07
Breath Volume	2.082	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:09
Subject Sample # 2	0.160	22:10
Breath Volume	1.695	22:10
Air Blank	0.000	22:10
Control Test	0.079	22:11
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VASQUEZ,MATTHEW
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GRADOLPH

DATE/TIME OF TEST: 03/24/2017 01:59:58

Diagnostics Check	OK	01:46	
Air Blank	0.000	01:46	
Control Test	0.080	01:47	
Air Blank	0.000	01:47	
Subject Sample	# 1 0.178	01:49	
Breath Volume	1.457	01:49	
Air Blank	0.000	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 2 0.168	01:55	Volume Not Met
Breath Volume	0.832	01:55	
Air Blank	0.000	01:55	
Air Blank	0.000	01:57	
Subject Sample	# 3 0.173	01:58	
Breath Volume	1.929	01:58	
Air Blank	0.000	01:59	
Control Test	0.077	01:59	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 03/24/2017 15:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CALLARI

DATE/TIME OF TEST: 03/23/2017 00:17:30

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.080	00:16	
Air Blank		00:17	Ambient Fail
Air Blank	0.000	00:17	

Data Download Date/Time: 04/03/2017 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CALLARI

DATE/TIME OF TEST: 03/23/2017 00:29:33

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.084	00:21	
Air Blank	0.000	00:21	
Subject Sample	# 1 0.069	00:23	
Breath Volume	1.519	00:23	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 2	00:27	Interferent Detect
Breath Volume	3.097	00:27	
Air Blank	0.000	00:28	
Control Test	0.080	00:28	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:29	

Data Download Date/Time: 04/03/2017 16:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006260

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: CALLARI

DATE/TIME OF TEST: 03/23/2017 00:41:51

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.082	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.076	00:37
Breath Volume	2.839	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.076	00:40
Breath Volume	2.617	00:40
Air Blank	0.000	00:40
Control Test	0.081	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 04/03/2017 16:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 03/10/2017 21:03:16

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.228	20:58
Breath Volume	2.441	20:58
Air Blank	0.000	20:58
Air Blank	0.000	21:00
Subject Sample # 2	0.229	21:01
Breath Volume	2.304	21:01
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WARMINGTON,WILLIAM
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 03/11/2017 02:02:42

Diagnostics Check	OK	01:55
Air Blank	0.000	01:56
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.319	01:57
Breath Volume	2.789	01:57
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.314	02:01
Breath Volume	2.273	02:01
Air Blank	0.000	02:01
Control Test	0.078	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOFIELD,FRANK,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BAYNE,ANDREW
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: BRESTER

DATE/TIME OF TEST: 03/12/2017 19:25:37

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.080	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.210	19:20
Breath Volume	1.457	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.219	19:23
Breath Volume	1.660	19:23
Air Blank	0.000	19:24
Control Test	0.079	19:25
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOFIELD,FRANK,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: ORTIZ,CARLOS
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: BESSETE

DATE/TIME OF TEST: 03/13/2017 22:44:16

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.175	22:39
Breath Volume	2.253	22:39
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.179	22:42
Breath Volume	2.105	22:42
Air Blank	0.000	22:43
Control Test	0.078	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCOFIELD,FRANK,E
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: GRAVES,AUSTIN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: DONAHUE

DATE/TIME OF TEST: 03/14/2017 00:38:33

Diagnostics Check	OK	00:35	
Air Blank	0.000	00:36	
Control Test	0.080	00:36	
Air Blank	0.000	00:36	
Subject Sample	# 1	00:37	Subject Test Refused
Breath Volume	0.000	00:37	
Air Blank	0.000	00:37	
Control Test	0.079	00:37	
Air Blank	0.000	00:38	
Diagnostics Check	OK	00:38	

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACKEL,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: YACKEL,NICHOLAS
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: MOREA

DATE/TIME OF TEST: 03/14/2017 19:57:08

Diagnostics Check	OK	19:50
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.144	19:52
Breath Volume	2.601	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.133	19:55
Breath Volume	1.472	19:55
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YACKEL,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SCOFIELD,FRANK
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HARNEY

DATE/TIME OF TEST: 03/15/2017 01:19:22

Diagnostics Check	OK	01:11
Air Blank	0.000	01:11
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.232	01:14
Breath Volume	1.128	01:14
Air Blank	0.000	01:14
Air Blank	0.000	01:16
Subject Sample # 2	0.242	01:17
Breath Volume	1.343	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Diagnostics Check	OK	01:19

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKENS,SHAWN,C
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: ACKERT

DATE/TIME OF TEST: 03/16/2017 20:17:32

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.080	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.117	20:12
Breath Volume	2.378	20:12
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.116	20:15
Breath Volume	3.535	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUKENS,SHAWN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: SWOBODA

DATE/TIME OF TEST: 03/16/2017 22:53:32

Diagnostics Check	OK	22:46
Air Blank	0.000	22:47
Control Test	0.081	22:47
Air Blank	0.000	22:47
Subject Sample # 1	0.171	22:48
Breath Volume	3.207	22:48
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.163	22:51
Breath Volume	3.425	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUKENS,SHAWN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: HOWE

DATE/TIME OF TEST: 03/17/2017 21:05:24

Diagnostics Check	OK	20:58
Air Blank	0.000	20:59
Control Test	0.081	20:59
Air Blank	0.000	20:59
Subject Sample # 1	0.060	21:00
Breath Volume	2.921	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.058	21:03
Breath Volume	2.781	21:03
Air Blank	0.000	21:04
Control Test	0.081	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUKENS,SHAWN
ARRESTING OFFICER AGENCY: Volusia County Beach
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: GARDNER

DATE/TIME OF TEST: 03/17/2017 23:14:39

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.081	23:07
Air Blank	0.000	23:08
Subject Sample # 1	0.169	23:09
Breath Volume	3.226	23:09
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.171	23:13
Breath Volume	3.031	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Diagnostics Check	OK	23:14

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SANDERS,LAUREN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: DELGATTO

DATE/TIME OF TEST: 03/17/2017 23:34:07

Diagnostics Check	OK	23:31	
Air Blank	0.000	23:31	
Control Test	0.081	23:32	
Air Blank	0.000	23:32	
Subject Sample	# 1	23:32	Subject Test Refused
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Control Test	0.080	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKENS,SHAWN,C
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: WHITEHOUSE

DATE/TIME OF TEST: 03/18/2017 00:14:20

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.137	00:09
Breath Volume	1.550	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:12
Subject Sample # 2	0.139	00:12
Breath Volume	1.964	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUPPINO,SETH,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,DARRYL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: LOCKWOOD

DATE/TIME OF TEST: 03/18/2017 21:37:01

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.081	21:31
Air Blank	0.000	21:31
Subject Sample # 1	0.153	21:32
Breath Volume	2.808	21:32
Air Blank	0.000	21:33
Air Blank	0.000	21:34
Subject Sample # 2	0.154	21:35
Breath Volume	2.621	21:35
Air Blank	0.000	21:36
Control Test	0.080	21:36
Air Blank	0.000	21:36
Diagnostics Check	OK	21:37

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SCOFIELD,FRANK
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 03/18/2017 23:19:48

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.082	23:12
Air Blank	0.000	23:13
Subject Sample # 1	0.178	23:13
Breath Volume	3.269	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.161	23:18
Breath Volume	3.433	23:18
Air Blank	0.000	23:18
Control Test	0.081	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUKENS,SHAWN,C
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,S
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 03/19/2017 00:20:51

Diagnostics Check	OK	00:13
Air Blank	0.000	00:14
Control Test	0.081	00:14
Air Blank	0.000	00:15
Subject Sample # 1	0.161	00:15
Breath Volume	3.445	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.150	00:19
Breath Volume	1.640	00:19
Air Blank	0.000	00:19
Control Test	0.080	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUPPINO,SETH
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/19/2017 00:54:43

Diagnostics Check	OK	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:49
Subject Sample # 1	0.009	00:50
Breath Volume	2.578	00:50
Air Blank	0.000	00:50
Air Blank	0.000	00:52
Subject Sample # 2	0.008	00:53
Breath Volume	2.628	00:53
Air Blank	0.000	00:53
Control Test	0.081	00:54
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 04/11/2017 10:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/02/2017 02:26:32

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.079	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.138	02:21
Breath Volume	1.472	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.134	02:24
Breath Volume	1.773	02:24
Air Blank	0.000	02:25
Control Test	0.078	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 03/22/2017 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: SHRIGLEY,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 02/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LOUIS

DATE/TIME OF TEST: 03/18/2017 03:43:28

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Subject Sample # 1	0.163	03:38
Breath Volume	1.484	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.144	03:41
Breath Volume	1.425	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 03/22/2017 20:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY, B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: BARRETO

DATE/TIME OF TEST: 03/01/2017 02:45:35

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.186	02:40
Breath Volume	1.656	02:40
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.190	02:43
Breath Volume	2.027	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM, R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SCOULLIS

DATE/TIME OF TEST: 03/01/2017 03:29:04

Diagnostics Check	OK	03:25	
Air Blank	0.000	03:26	
Control Test	0.079	03:26	
Air Blank	0.000	03:26	
Subject Sample	# 1	03:27	Subject Test Refused
Breath Volume	0.000	03:27	
Air Blank	0.000	03:28	
Control Test	0.079	03:28	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:29	

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: EDGE,S
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/01/2017 22:41:19

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Subject Sample # 1	0.193	22:36
Breath Volume	2.472	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.206	22:39
Breath Volume	2.593	22:39
Air Blank	0.000	22:40
Control Test	0.078	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: AGUAYO,J
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: ARNOLD

DATE/TIME OF TEST: 03/02/2017 00:13:09

Diagnostics Check	OK	00:07
Air Blank	0.000	00:07
Control Test	0.079	00:07
Air Blank	0.000	00:08
Subject Sample # 1	0.000	00:08
Breath Volume	1.640	00:08
Air Blank	0.000	00:09
Air Blank	0.000	00:11
Subject Sample # 2	0.000	00:11
Breath Volume	1.656	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Diagnostics Check	OK	00:13

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:17

SUBJECT'S LAST NAME: DUMORNEY

DATE/TIME OF TEST: 03/03/2017 21:43:52

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.185	21:38
Breath Volume	1.625	21:38
Air Blank	0.000	21:39
Air Blank	0.000	21:41
Subject Sample # 2	0.182	21:42
Breath Volume	1.742	21:42
Air Blank	0.000	21:42
Control Test	0.080	21:43
Air Blank	0.000	21:43
Diagnostics Check	OK	21:43

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CASANOVA

DATE/TIME OF TEST: 03/04/2017 00:15:35

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.080	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.167	00:11
Breath Volume	1.511	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.173	00:13
Breath Volume	1.671	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: FADILA ALSABBAGH

DATE/TIME OF TEST: 03/04/2017 02:47:32

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.110	02:42
Breath Volume	2.160	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.100	02:45
Breath Volume	1.320	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 03/04/2017 03:38:28

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.170	03:33
Breath Volume	1.914	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.173	03:36
Breath Volume	1.964	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: JOSE JOSE

DATE/TIME OF TEST: 03/04/2017 20:23:14

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.080	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.210	20:18
Breath Volume	1.726	20:18
Air Blank	0.000	20:19
Air Blank	0.000	20:20
Subject Sample # 2	0.206	20:21
Breath Volume	1.730	20:21
Air Blank	0.000	20:22
Control Test	0.079	20:22
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: RAMIREZ SARGENTO

DATE/TIME OF TEST: 03/05/2017 02:10:28

Diagnostics Check	OK	02:03
Air Blank	0.000	02:04
Control Test	0.080	02:04
Air Blank	0.000	02:04
Subject Sample # 1	0.201	02:05
Breath Volume	1.582	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 2	0.199	02:08
Breath Volume	2.320	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN, R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: FRY

DATE/TIME OF TEST: 03/05/2017 02:26:44

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.129	02:22
Breath Volume	3.734	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.128	02:25
Breath Volume	3.570	02:25
Air Blank	0.000	02:25
Control Test	0.079	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: DUDLEY-MC SORLEY

DATE/TIME OF TEST: 03/05/2017 03:58:13

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.188	03:52
Breath Volume	1.703	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.194	03:56
Breath Volume	2.066	03:56
Air Blank	0.000	03:57
Control Test	0.077	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 03/05/2017 04:07:55

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.078	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.166	04:03
Breath Volume	1.429	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.160	04:06
Breath Volume	2.257	04:06
Air Blank	0.000	04:06
Control Test	0.078	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ODABAS

DATE/TIME OF TEST: 03/06/2017 02:51:40

Diagnostics Check	OK	02:43
Air Blank	0.000	02:44
Control Test	0.080	02:44
Air Blank	0.000	02:45
Subject Sample # 1	0.173	02:46
Breath Volume	1.574	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.178	02:49
Breath Volume	1.535	02:49
Air Blank	0.000	02:50
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER, J
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: KUHNLEIN

DATE/TIME OF TEST: 03/06/2017 04:38:41

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.207	04:34
Breath Volume	2.429	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.208	04:36
Breath Volume	1.886	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: KANAVOS

DATE/TIME OF TEST: 03/06/2017 04:59:34

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.185	04:54
Breath Volume	2.464	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.184	04:57
Breath Volume	2.410	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: KOSNOW

DATE/TIME OF TEST: 03/06/2017 05:35:22

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.193	05:30
Breath Volume	1.292	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample # 2	0.190	05:33
Breath Volume	1.230	05:33
Air Blank	0.000	05:34
Control Test	0.078	05:34
Air Blank	0.000	05:35
Diagnostics Check	OK	05:35

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ORTEGA

DATE/TIME OF TEST: 03/08/2017 03:32:17

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.124	03:27
Breath Volume	1.671	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.110	03:30
Breath Volume	2.148	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 03/09/2017 01:08:56

Diagnostics Check	OK	01:02
Air Blank	0.000	01:03
Control Test	0.079	01:03
Air Blank	0.000	01:04
Subject Sample # 1	0.152	01:04
Breath Volume	1.875	01:04
Air Blank	0.000	01:05
Air Blank	0.000	01:06
Subject Sample # 2	0.155	01:07
Breath Volume	3.003	01:07
Air Blank	0.000	01:08
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 03/09/2017 04:02:07

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.079	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAMER,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HATHAWAY

DATE/TIME OF TEST: 03/10/2017 04:00:55

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.163	03:56
Breath Volume	2.867	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:58
Subject Sample # 2	0.156	03:59
Breath Volume	2.531	03:59
Air Blank	0.000	03:59
Control Test	0.078	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS, D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: NIXON

DATE/TIME OF TEST: 03/11/2017 01:29:43

Diagnostics Check	OK	01:23
Air Blank	0.000	01:23
Control Test	0.080	01:23
Air Blank	0.000	01:24
Subject Sample # 1	0.061	01:25
Breath Volume	1.847	01:25
Air Blank	0.000	01:25
Air Blank	0.000	01:27
Subject Sample # 2	0.061	01:28
Breath Volume	1.761	01:28
Air Blank	0.000	01:28
Control Test	0.079	01:29
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/11/2017 04:56:14

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:50
Subject Sample # 1	0.141	04:51
Breath Volume	3.011	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:53
Subject Sample # 2	0.154	04:54
Breath Volume	2.773	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PACHECO FLORES

DATE/TIME OF TEST: 03/12/2017 01:56:51

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.129	01:52
Breath Volume	1.746	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.129	01:55
Breath Volume	1.839	01:55
Air Blank	0.000	01:55
Control Test	0.078	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 03/12/2017 04:01:16

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.000	03:56
Breath Volume	2.761	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.000	03:59
Breath Volume	2.378	03:59
Air Blank	0.000	04:00
Control Test	0.079	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: SCHWAB

DATE/TIME OF TEST: 03/12/2017 05:23:15

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.079	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.278	05:18
Breath Volume	1.449	05:18
Air Blank	0.000	05:19
Air Blank	0.000	05:20
Subject Sample # 2	0.264	05:21
Breath Volume	1.296	05:21
Air Blank	0.000	05:22
Control Test	0.078	05:22
Air Blank	0.000	05:23
Diagnostics Check	OK	05:23

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: O'NEILL,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: KURTZMAN

DATE/TIME OF TEST: 03/12/2017 06:31:03

Diagnostics Check	OK	06:24
Air Blank	0.000	06:24
Control Test	0.078	06:25
Air Blank	0.000	06:25
Subject Sample # 1	0.186	06:26
Breath Volume	3.242	06:26
Air Blank	0.000	06:27
Air Blank	0.000	06:28
Subject Sample # 2	0.201	06:29
Breath Volume	3.902	06:29
Air Blank	0.000	06:30
Control Test	0.079	06:30
Air Blank	0.000	06:30
Diagnostics Check	OK	06:31

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 03/13/2017 00:28:52

Diagnostics Check	OK	00:22
Air Blank	0.000	00:23
Control Test	0.079	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.124	00:24
Breath Volume	1.859	00:24
Air Blank	0.000	00:25
Air Blank	0.000	00:26
Subject Sample # 2	0.123	00:27
Breath Volume	1.894	00:27
Air Blank	0.000	00:27
Control Test	0.078	00:28
Air Blank	0.000	00:28
Diagnostics Check	OK	00:28

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 03/13/2017 23:13:22

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.198	23:08
Breath Volume	2.445	23:08
Air Blank	0.000	23:09
Air Blank	0.000	23:11
Subject Sample # 2	0.196	23:11
Breath Volume	2.039	23:11
Air Blank	0.000	23:12
Control Test	0.078	23:12
Air Blank	0.000	23:13
Diagnostics Check	OK	23:13

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAMER,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SARDINAS

DATE/TIME OF TEST: 03/14/2017 03:08:18

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.096	03:03
Breath Volume	1.945	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:06
Subject Sample # 2	0.086	03:06
Breath Volume	3.062	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FRANDSEN

DATE/TIME OF TEST: 03/15/2017 01:17:59

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.173	01:12
Breath Volume	2.914	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.174	01:16
Breath Volume	2.945	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/16/2017 20:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/16/2017 04:19:26

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.081	04:12
Air Blank	0.000	04:13
Subject Sample # 1	0.126	04:14
Breath Volume	1.785	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.127	04:17
Breath Volume	1.347	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 03/16/2017 20:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY, WILLIAM, C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY, WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: GULLEY

DATE/TIME OF TEST: 03/02/2017 05:18:45

Diagnostics Check	OK	05:11
Air Blank	0.000	05:12
Control Test	0.080	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.177	05:14
Breath Volume	1.156	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.174	05:17
Breath Volume	3.156	05:17
Air Blank	0.000	05:17
Control Test	0.078	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 03/21/2017 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 02/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: KENDRA

DATE/TIME OF TEST: 03/18/2017 03:55:15

Diagnostics Check	OK	03:48
Air Blank	0.000	03:48
Control Test	0.082	03:49
Air Blank	0.000	03:49
Subject Sample # 1	0.206	03:50
Breath Volume	2.320	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:52
Subject Sample # 2	0.209	03:53
Breath Volume	2.070	03:53
Air Blank	0.000	03:54
Control Test	0.077	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 03/21/2017 09:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREENE,ROBERT,R
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GREENE,ROBERT
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: STANSELL

DATE/TIME OF TEST: 03/23/2017 01:16:38

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.157	01:12
Breath Volume	4.085	01:12
Air Blank	0.000	01:12
Air Blank	0.000	01:14
Subject Sample # 2	0.146	01:15
Breath Volume	4.097	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:16
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 04/13/2017 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: POOL,HAROLD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 03/26/2017 15:45:45

Diagnostics Check	OK	15:41	
Air Blank	0.000	15:42	
Control Test	0.081	15:42	
Air Blank	0.000	15:42	
Subject Sample	# 1	15:44	Subject Test Refused
Breath Volume	0.000	15:44	
Air Blank	0.000	15:44	
Control Test	0.082	15:45	
Air Blank	0.000	15:45	
Diagnostics Check	OK	15:45	

Data Download Date/Time: 04/13/2017 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POOL,HAROLD,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: POOL,HAROLD
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: BARNES

DATE/TIME OF TEST: 03/26/2017 16:12:34

Diagnostics Check	OK	16:04
Air Blank	0.000	16:05
Control Test	0.082	16:05
Air Blank	0.000	16:06
Subject Sample # 1	0.155	16:06
Breath Volume	1.410	16:06
Air Blank	0.000	16:07
Air Blank	0.000	16:09
Subject Sample # 2	0.159	16:10
Breath Volume	2.347	16:10
Air Blank	0.000	16:11
Control Test	0.080	16:11
Air Blank	0.000	16:12
Diagnostics Check	OK	16:12

Data Download Date/Time: 04/13/2017 08:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: BANDURSKI,JAKE
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 03/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/31/2017 22:16:43

Diagnostics Check	OK	22:10
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.048	22:11
Breath Volume	1.488	22:11
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.048	22:15
Breath Volume	1.148	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 04/13/2017 08:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADDY,STEVEN,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUERRIER,KATLY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:11

SUBJECT'S LAST NAME: VALVERDE

DATE/TIME OF TEST: 03/05/2017 08:52:47

Diagnostics Check	OK	08:44	
Air Blank	0.000	08:44	
Control Test	0.080	08:45	
Air Blank	0.000	08:45	
Subject Sample	# 1	08:48	No Sample Provided
Breath Volume	0.000	08:48	
Air Blank	0.000	08:49	
Air Blank	0.000	08:51	
Subject Sample	# 2	08:51	Subject Test Refused
Breath Volume	0.000	08:51	
Air Blank	0.000	08:51	
Control Test	0.080	08:52	
Air Blank	0.000	08:52	
Diagnostics Check	OK	08:52	

Data Download Date/Time: 03/30/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUAREZ, BERNICE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ, JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 03/15/2017 03:52:12

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.080	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.000	03:47
Breath Volume	3.179	03:47
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.000	03:50
Breath Volume	1.789	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/30/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,ANDY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ALVARADO-HABER,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:19

SUBJECT'S LAST NAME: BRITO

DATE/TIME OF TEST: 03/26/2017 06:43:27

Diagnostics Check	OK	06:39	
Air Blank	0.000	06:40	
Control Test	0.079	06:40	
Air Blank	0.000	06:41	
Subject Sample	# 1	06:42	Subject Test Refused
Breath Volume	0.000	06:42	
Air Blank		06:42	Ambient Fail
Air Blank	0.000	06:43	

Data Download Date/Time: 03/30/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACOSTA,CAROLINA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HERNANDEZ,JASIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/20/2018

VIOLATION CODE: DUI

ARREST TIME: 03:39

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: ALVAYERO

DATE/TIME OF TEST: 03/03/2017 04:05:29

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:01	
Control Test	0.077	04:01	
Air Blank	0.000	04:02	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.079	04:04	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:39

SUBJECT'S LAST NAME: MONTEJO

DATE/TIME OF TEST: 03/03/2017 06:12:13

Diagnostics Check	OK	05:59	
Air Blank	0.000	05:59	
Control Test	0.080	05:59	
Air Blank	0.000	06:00	
Subject Sample	# 1	06:03	No Sample Provided
Breath Volume	0.000	06:03	
Air Blank	0.000	06:04	
Air Blank	0.000	06:06	
Subject Sample	# 2	0.187	06:07
Breath Volume	3.191	06:07	
Air Blank	0.000	06:08	
Air Blank	0.000	06:10	
Subject Sample	# 3	0.189	06:10
Breath Volume	3.464	06:10	
Air Blank	0.000	06:11	
Control Test	0.077	06:11	
Air Blank	0.000	06:12	
Diagnostics Check	OK	06:12	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CHADWELL, WENDY, B

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CHADWELL, WENDY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/20/2018

VIOLATION CODE: DUI

ARREST TIME: 21:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/03/2017 22:37:05

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.188	22:32
Breath Volume	1.847	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.185	22:35
Breath Volume	2.351	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:36
Diagnostics Check	OK	22:37

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: GODOY

DATE/TIME OF TEST: 03/04/2017 04:56:40

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.080	04:41	
Air Blank	0.000	04:41	
Subject Sample	# 1 0.159	04:44	
Breath Volume	1.320	04:44	
Air Blank	0.000	04:45	
Air Blank	0.000	04:47	
Subject Sample	# 2 0.147	04:50	Volume Not Met
Breath Volume	0.035	04:50	
Air Blank	0.000	04:51	
Air Blank	0.000	04:52	
Subject Sample	# 3	04:55	Subject Test Refused
Breath Volume	0.492	04:55	
Air Blank	0.000	04:55	
Control Test	0.078	04:56	
Air Blank	0.000	04:56	
Diagnostics Check	OK	04:56	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BAHAMONDE,JORDANY,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BAHAMONDE,JORDANY

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 22:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SEQUEIRA

DATE/TIME OF TEST: 03/04/2017 23:05:24

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.080	22:58
Air Blank	0.000	22:59
Subject Sample	# 1 0.176	23:00
Breath Volume	1.207	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:02
Subject Sample	# 2 0.170	23:03
Breath Volume	1.726	23:03
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,DIYANGO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,DIYANGO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/20/2018

VIOLATION CODE: DUI

ARREST TIME: 22:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HOLDER

DATE/TIME OF TEST: 03/04/2017 23:48:35

Diagnostics Check	OK	23:47	
Air Blank		23:48	RFI Detect
Air Blank	0.000	23:48	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FERNANDEZ,DIYANGO,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: FERNANDEZ,DIYANGO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/20/2018

VIOLATION CODE: DUI

ARREST TIME: 22:51

GENDER: Male

OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: HOLDER

DATE/TIME OF TEST: 03/04/2017 23:57:48

Diagnostics Check	OK	23:50
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.150	23:52
Breath Volume	1.613	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.158	23:56
Breath Volume	1.917	23:56
Air Blank	0.000	23:56
Control Test	0.077	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BARRIOS,MARYURI,O

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: BARRIOS,MARYURI

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/20/2018

VIOLATION CODE: DUI

ARREST TIME: 06:44

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:23

SUBJECT'S LAST NAME: VALLADARES

DATE/TIME OF TEST: 03/05/2017 07:52:18

Diagnostics Check	OK	07:43
Air Blank	0.000	07:43
Control Test	0.079	07:44
Air Blank	0.000	07:44
Subject Sample	# 1 0.228	07:46
Breath Volume	2.605	07:46
Air Blank	0.000	07:47
Air Blank	0.000	07:48
Subject Sample	# 2 0.211	07:50
Breath Volume	2.039	07:50
Air Blank	0.000	07:51
Control Test	0.077	07:51
Air Blank	0.000	07:52
Diagnostics Check	OK	07:52

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOANELLE,V
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JOANELLE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: OCHOA

DATE/TIME OF TEST: 03/05/2017 08:06:31

Diagnostics Check	OK	08:00
Air Blank	0.000	08:00
Control Test	0.078	08:01
Air Blank	0.000	08:01
Subject Sample # 1	0.199	08:01
Breath Volume	1.628	08:01
Air Blank	0.000	08:02
Air Blank	0.000	08:04
Subject Sample # 2	0.190	08:04
Breath Volume	1.828	08:04
Air Blank	0.000	08:05
Control Test	0.077	08:05
Air Blank	0.000	08:06
Diagnostics Check	OK	08:06

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SOSA,ALEJANDRO,A

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: SAINTFLUER,RANDOLPH

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/20/2018

VIOLATION CODE: DUI

ARREST TIME: 20:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: MYSLAKOWSKI

DATE/TIME OF TEST: 03/05/2017 21:02:23

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:59	
Control Test	0.080	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.079	21:01	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANK,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/20/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/08/2017 04:11:44

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.159	04:05
Breath Volume	1.894	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:07
Subject Sample # 2	0.167	04:10
Breath Volume	2.023	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,BRIAN,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LAFONT,WILLIAM
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/20/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: ANTROBUS

DATE/TIME OF TEST: 03/09/2017 02:04:41

Diagnostics Check	OK	01:54	
Air Blank	0.000	01:55	
Control Test	0.080	01:55	
Air Blank	0.000	01:55	
Subject Sample	# 1	01:59	No Sample Provided
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Air Blank	0.000	02:01	
Subject Sample	# 2	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Control Test	0.080	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA-PUJOLS,SARAH,P
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MENDOZA-PUJOLS,SARAH
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/20/2018
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: MORFA

DATE/TIME OF TEST: 03/11/2017 20:51:34

Diagnostics Check	OK	20:44
Air Blank	0.000	20:44
Control Test	0.080	20:45
Air Blank	0.000	20:45
Subject Sample # 1	0.227	20:46
Breath Volume	2.816	20:46
Air Blank	0.000	20:47
Air Blank	0.000	20:49
Subject Sample # 2	0.223	20:49
Breath Volume	2.003	20:49
Air Blank	0.000	20:50
Control Test	0.078	20:50
Air Blank	0.000	20:51
Diagnostics Check	OK	20:51

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,FABIANA,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLARREAL,FABIANA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/20/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: ROMO

DATE/TIME OF TEST: 03/12/2017 04:29:47

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.194	04:25
Breath Volume	2.300	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.187	04:28
Breath Volume	2.730	04:28
Air Blank	0.000	04:28
Control Test	0.079	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABELLO,ADRIAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PEREZ,JOSHUA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/20/2018
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/12/2017 20:53:34

Diagnostics Check	OK	20:52	
Air Blank		20:53	RFI Detect
Air Blank	0.000	20:53	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABELLO,ADRIAN,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PEREZ,JOSHUA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/20/2018
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 03/12/2017 20:59:34

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.078	20:57	
Air Blank	0.000	20:57	
Subject Sample	# 1	20:58	Subject Test Refused
Breath Volume	0.000	20:58	
Air Blank	0.000	20:58	
Control Test	0.079	20:58	
Air Blank	0.000	20:59	
Diagnostics Check	OK	20:59	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,EUGENE,H
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LOO,JORGE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/20/2018
VIOLATION CODE: DUI
ARREST TIME: 07:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:49

SUBJECT'S LAST NAME: SAN MIGUEL

DATE/TIME OF TEST: 03/14/2017 08:16:33

Diagnostics Check	OK	08:10
Air Blank	0.000	08:10
Control Test	0.080	08:10
Air Blank	0.000	08:11
Subject Sample # 1	0.204	08:11
Breath Volume	2.062	08:11
Air Blank	0.000	08:12
Air Blank	0.000	08:14
Subject Sample # 2	0.208	08:14
Breath Volume	1.949	08:14
Air Blank	0.000	08:15
Control Test	0.078	08:15
Air Blank	0.000	08:16
Diagnostics Check	OK	08:16

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GARCIA,VANESSA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: HEREDIA RUBIO,ROBERTO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 11:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:16

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 03/14/2017 12:58:49

Diagnostics Check	OK	12:51
Air Blank	0.000	12:51
Control Test	0.079	12:52
Air Blank	0.000	12:52
Subject Sample # 1	0.107	12:54
Breath Volume	3.351	12:54
Air Blank	0.000	12:54
Air Blank	0.000	12:56
Subject Sample # 2	0.106	12:57
Breath Volume	3.644	12:57
Air Blank	0.000	12:57
Control Test	0.077	12:58
Air Blank	0.000	12:58
Diagnostics Check	OK	12:58

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: ESCORCIA

DATE/TIME OF TEST: 03/15/2017 02:05:55

Diagnostics Check	OK	02:02	
Air Blank	0.000	02:02	
Control Test	0.079	02:03	
Air Blank	0.000	02:03	
Subject Sample	# 1	02:04	Subject Test Refused
Breath Volume	0.000	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MENEGAZZO,MARIO,J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: MENEGAZZO,MARIO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 11:55

GENDER: Male

OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 03/16/2017 13:13:11

Diagnostics Check	OK	13:04	
Air Blank	0.000	13:04	
Control Test	0.079	13:04	
Air Blank	0.000	13:05	
Subject Sample	# 1 0.210	13:08	Volume Not Met
Breath Volume	0.609	13:08	
Air Blank	0.000	13:09	
Air Blank	0.000	13:11	
Subject Sample	# 2	13:11	Subject Test Refused
Breath Volume	0.000	13:11	
Air Blank	0.000	13:12	
Control Test	0.078	13:12	
Air Blank	0.000	13:13	
Diagnostics Check	OK	13:13	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ, MERCEDES, S

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CRUZ, NATANAEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/20/2018

VIOLATION CODE: DUI

ARREST TIME: 18:25

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: RECIO

DATE/TIME OF TEST: 03/16/2017 19:26:18

Diagnostics Check	OK	19:17	
Air Blank	0.000	19:17	
Control Test	0.080	19:18	
Air Blank	0.000	19:18	
Subject Sample	# 1 0.094	19:21	Volume Not Met
Breath Volume	0.050	19:21	
Air Blank	0.000	19:22	
Air Blank	0.000	19:24	
Subject Sample	# 2	19:24	Subject Test Refused
Breath Volume	0.000	19:24	
Air Blank	0.000	19:25	
Control Test	0.078	19:25	
Air Blank	0.000	19:26	
Diagnostics Check	OK	19:26	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNANDEZ,JASIEL,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: PALOMINO,JULIO

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 04:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: LOACES GONZALEZ

DATE/TIME OF TEST: 03/17/2017 05:44:04

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:32	
Control Test	0.080	05:32	
Air Blank	0.000	05:32	
Subject Sample	# 1 0.118	05:35	Volume Not Met
Breath Volume	0.085	05:35	
Air Blank	0.000	05:36	
Air Blank	0.000	05:38	
Subject Sample	# 2 0.129	05:39	
Breath Volume	1.277	05:39	
Air Blank	0.000	05:39	
Air Blank	0.000	05:41	
Subject Sample	# 3 0.126	05:42	
Breath Volume	2.093	05:42	
Air Blank	0.000	05:43	
Control Test	0.080	05:43	
Air Blank	0.000	05:43	
Diagnostics Check	OK	05:44	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACOSTA,CAROLINA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: TARGIA,FRANK

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/19/2017 04:09:08

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.080	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.079	04:04
Breath Volume	1.128	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.082	04:07
Breath Volume	1.695	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:09
Diagnostics Check	OK	04:09

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PASTOR,JOSE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 03/22/2017 20:58:59

Diagnostics Check	OK	20:56	
Air Blank	0.000	20:56	
Control Test	0.080	20:56	
Air Blank	0.000	20:57	
Subject Sample	# 1	20:57	Subject Test Refused
Breath Volume	0.000	20:57	
Air Blank	0.000	20:58	
Control Test	0.080	20:58	
Air Blank	0.000	20:58	
Diagnostics Check	OK	20:58	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ACOSTA,CAROLINA,

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: TARGIA,FRANK

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 02:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: COLEMAN

DATE/TIME OF TEST: 03/25/2017 03:00:17

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.080	02:57	
Air Blank	0.000	02:57	
Subject Sample	# 1	02:58	Subject Test Refused
Breath Volume	0.000	02:58	
Air Blank	0.000	02:59	
Control Test	0.080	02:59	
Air Blank	0.000	03:00	
Diagnostics Check	OK	03:00	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,F,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MENA,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/26/2017 04:27:10

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.196	04:21
Breath Volume	1.406	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:24
Subject Sample # 2	0.190	04:25
Breath Volume	1.410	04:25
Air Blank	0.000	04:26
Control Test	0.077	04:26
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: CANTERO,DANIEL,J

BREATH TEST OPERATOR AGENCY: Miami PD

ARRESTING OFFICER: CANTERO,DANIEL

ARRESTING OFFICER AGENCY: Miami PD

LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 19:46

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: FERRER

DATE/TIME OF TEST: 03/26/2017 20:13:23

Diagnostics Check	OK	20:10	
Air Blank	0.000	20:10	
Control Test	0.080	20:11	
Air Blank	0.000	20:11	
Subject Sample	# 1	20:12	Subject Test Refused
Breath Volume	0.000	20:12	
Air Blank	0.000	20:12	
Control Test	0.080	20:12	
Air Blank	0.000	20:13	
Diagnostics Check	OK	20:13	

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,ARIEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LOPEZ,ARIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 02/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: MENENDEZ DOCAMPO

DATE/TIME OF TEST: 03/27/2017 03:19:10

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.103	03:14
Breath Volume	3.984	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.100	03:17
Breath Volume	3.238	03:17
Air Blank	0.000	03:18
Control Test	0.078	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 03/29/2017 17:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALOMINO,JULIO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PALOMINO,JULIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 03/30/2017 05:38:12

Diagnostics Check	OK	05:22	
Air Blank	0.000	05:22	
Control Test	0.080	05:22	
Air Blank	0.000	05:23	
Subject Sample	# 1	05:26	No Sample Provided
Breath Volume	0.000	05:26	
Air Blank	0.000	05:27	
Air Blank	0.000	05:29	
Subject Sample	# 2	0.049	05:31
Breath Volume	1.152	05:31	
Air Blank	0.000	05:32	
Air Blank	0.000	05:34	
Subject Sample	# 3	0.043	05:36
Breath Volume	1.238	05:36	
Air Blank	0.000	05:37	
Control Test	0.078	05:37	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 04/26/2017 19:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TARGIA,FRANK,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: PERDOMO

DATE/TIME OF TEST: 03/31/2017 00:46:03

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.080	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.000	00:41
Breath Volume	2.718	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.000	00:44
Breath Volume	2.781	00:44
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:46

Data Download Date/Time: 04/26/2017 19:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/31/2017 22:35:02

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:33	
Control Test	0.080	22:33	
Air Blank		22:34	Purge Fail
Air Blank		22:35	Purge Fail

Data Download Date/Time: 04/26/2017 19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/31/2017 22:37:52

Diagnostics Check	OK	22:36	
Air Blank		22:37	Ambient Fail
Air Blank	0.000	22:37	

Data Download Date/Time: 04/26/2017 19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TADLOCK,J,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TADLOCK,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 03/31/2017 22:48:39

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.080	22:39	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:43	No Sample Provided
Breath Volume	0.000	22:43	
Air Blank	0.000	22:44	
Air Blank	0.000	22:46	
Subject Sample	# 2	22:47	Subject Test Refused
Breath Volume	0.000	22:47	
Air Blank	0.000	22:47	
Control Test	0.077	22:48	
Air Blank	0.000	22:48	
Diagnostics Check	OK	22:48	

Data Download Date/Time: 04/26/2017 19:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN, INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP, JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 03/24/2017 05:36:49

Diagnostics Check	OK	05:30
Air Blank	0.000	05:30
Control Test	0.076	05:31
Air Blank	0.000	05:31
Subject Sample # 1	0.209	05:32
Breath Volume	2.625	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.205	05:35
Breath Volume	1.539	05:35
Air Blank	0.000	05:35
Control Test	0.077	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 03/25/2017 02:42:27

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.077	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.092	02:37
Breath Volume	2.609	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.091	02:40
Breath Volume	2.429	02:40
Air Blank	0.000	02:41
Control Test	0.078	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 03/25/2017 05:25:17

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.077	05:17
Air Blank	0.000	05:17
Subject Sample # 1	0.175	05:20
Breath Volume	2.265	05:20
Air Blank	0.000	05:20
Air Blank	0.000	05:22
Subject Sample # 2	0.177	05:23
Breath Volume	1.828	05:23
Air Blank	0.000	05:24
Control Test	0.077	05:24
Air Blank	0.000	05:25
Diagnostics Check	OK	05:25

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: SILVESTREQUINONEZ

DATE/TIME OF TEST: 03/25/2017 22:02:56

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.077	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.260	21:58
Breath Volume	3.214	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.257	22:01
Breath Volume	2.492	22:01
Air Blank	0.000	22:02
Control Test	0.076	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCGILLICUDDY,STEVEN
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: CLERMONT

DATE/TIME OF TEST: 03/26/2017 05:38:18

Diagnostics Check	OK	05:31
Air Blank	0.000	05:32
Control Test	0.078	05:32
Air Blank	0.000	05:33
Subject Sample # 1	0.145	05:33
Breath Volume	1.964	05:33
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.144	05:36
Breath Volume	2.097	05:36
Air Blank	0.000	05:37
Control Test	0.078	05:37
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: RODRIGUEZ,ALEXANDER
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: NODA

DATE/TIME OF TEST: 03/27/2017 05:30:12

Diagnostics Check	OK	05:22
Air Blank	0.000	05:22
Control Test	0.077	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.168	05:25
Breath Volume	2.351	05:25
Air Blank	0.000	05:25
Air Blank	0.000	05:27
Subject Sample # 2	0.171	05:28
Breath Volume	1.667	05:28
Air Blank	0.000	05:29
Control Test	0.076	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: COTTRELL,JOHN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 03/30/2017 10:28:40

Diagnostics Check	OK	10:22
Air Blank	0.000	10:22
Control Test	0.078	10:22
Air Blank	0.000	10:23
Subject Sample # 1	0.234	10:24
Breath Volume	1.355	10:24
Air Blank	0.000	10:24
Air Blank	0.000	10:26
Subject Sample # 2	0.242	10:27
Breath Volume	2.121	10:27
Air Blank	0.000	10:27
Control Test	0.078	10:28
Air Blank	0.000	10:28
Diagnostics Check	OK	10:28

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIZENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: VINCENT,BERNEX
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: ROMAN ROMERO

DATE/TIME OF TEST: 03/30/2017 19:25:24

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.078	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.000	19:20
Breath Volume	2.410	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.000	19:24
Breath Volume	1.914	19:24
Air Blank	0.000	19:24
Control Test	0.078	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 04/10/2017 15:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ISHAM,MATTHEW
ARRESTING OFFICER AGENCY: Juno Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: DESTEFANO

DATE/TIME OF TEST: 03/02/2017 02:21:11

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:14
Air Blank	0.000	02:14
Subject Sample # 1	0.168	02:16
Breath Volume	1.718	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:18
Subject Sample # 2	0.166	02:19
Breath Volume	1.816	02:19
Air Blank	0.000	02:20
Control Test	0.078	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 03/02/2017 04:34:25

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:31	
Control Test	0.080	04:32	
Air Blank	0.000	04:32	
Subject Sample	# 1	04:32	Subject Test Refused
Breath Volume	0.000	04:32	
Air Blank	0.000	04:33	
Control Test	0.080	04:33	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/03/2017 04:59:15

Diagnostics Check	OK	04:52
Air Blank	0.000	04:52
Control Test	0.081	04:53
Air Blank	0.000	04:53
Subject Sample # 1	0.159	04:54
Breath Volume	3.675	04:54
Air Blank	0.000	04:55
Air Blank	0.000	04:56
Subject Sample # 2	0.159	04:57
Breath Volume	1.812	04:57
Air Blank	0.000	04:58
Control Test	0.080	04:58
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: SONTAY

DATE/TIME OF TEST: 03/03/2017 07:11:02

Diagnostics Check	OK	07:04
Air Blank	0.000	07:04
Control Test	0.081	07:04
Air Blank	0.000	07:05
Subject Sample # 1	0.158	07:06
Breath Volume	3.125	07:06
Air Blank	0.000	07:07
Air Blank	0.000	07:08
Subject Sample # 2	0.161	07:09
Breath Volume	3.019	07:09
Air Blank	0.000	07:10
Control Test	0.079	07:10
Air Blank	0.000	07:10
Diagnostics Check	OK	07:11

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 03/03/2017 22:24:09

Diagnostics Check	OK	22:17
Air Blank	0.000	22:18
Control Test	0.081	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.246	22:19
Breath Volume	2.542	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:21
Subject Sample # 2	0.230	22:22
Breath Volume	2.335	22:22
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON,ROBERT
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: HOECKER

DATE/TIME OF TEST: 03/03/2017 23:48:12

Diagnostics Check	OK	23:38
Air Blank	0.000	23:39
Control Test	0.081	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.192	23:40
Breath Volume	2.113	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:42
Subject Sample # 2	0.171	23:43
Breath Volume	1.152	23:43
Air Blank	0.000	23:44
Air Blank	0.000	23:45
Subject Sample # 3	0.192	23:46
Breath Volume	2.300	23:46
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MECHALY

DATE/TIME OF TEST: 03/04/2017 01:21:17

Diagnostics Check	OK	01:15
Air Blank	0.000	01:15
Control Test	0.080	01:15
Air Blank	0.000	01:16
Subject Sample # 1	0.000	01:16
Breath Volume	3.472	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.000	01:19
Breath Volume	2.703	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MAZZA,ANTHONY
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: CUEVAS SANDOVAL

DATE/TIME OF TEST: 03/05/2017 01:27:07

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.196	01:22
Breath Volume	3.105	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.195	01:25
Breath Volume	3.371	01:25
Air Blank	0.000	01:26
Control Test	0.079	01:26
Air Blank	0.000	01:27
Diagnostics Check	OK	01:27

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MCCLUSKEY,RYAN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/05/2017 06:04:38

Diagnostics Check	OK	05:57
Air Blank	0.000	05:58
Control Test	0.081	05:58
Air Blank	0.000	05:59
Subject Sample # 1	0.139	05:59
Breath Volume	2.796	05:59
Air Blank	0.000	06:00
Air Blank	0.000	06:02
Subject Sample # 2	0.132	06:02
Breath Volume	1.968	06:02
Air Blank	0.000	06:03
Control Test	0.078	06:03
Air Blank	0.000	06:04
Diagnostics Check	OK	06:04

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHURCH,RYAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: CORADO-BATRES

DATE/TIME OF TEST: 03/05/2017 17:15:35

Diagnostics Check	OK	17:07
Air Blank	0.000	17:08
Control Test	0.081	17:08
Air Blank	0.000	17:08
Subject Sample # 1	0.155	17:09
Breath Volume	1.574	17:09
Air Blank	0.000	17:10
Air Blank	0.000	17:12
Subject Sample # 2	0.166	17:13
Breath Volume	1.382	17:13
Air Blank	0.000	17:14
Control Test	0.079	17:14
Air Blank	0.000	17:15
Diagnostics Check	OK	17:15

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DONNINO,ANTONIO
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: ROY

DATE/TIME OF TEST: 03/05/2017 21:08:22

Diagnostics Check	OK	21:00
Air Blank	0.000	21:01
Control Test	0.081	21:01
Air Blank	0.000	21:02
Subject Sample # 1	0.169	21:03
Breath Volume	2.398	21:03
Air Blank	0.000	21:04
Air Blank	0.000	21:05
Subject Sample # 2	0.176	21:06
Breath Volume	1.937	21:06
Air Blank	0.000	21:07
Control Test	0.080	21:07
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 03/24/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IOZZIO,GINA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 03/05/2017 21:38:09

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.080	21:31
Air Blank	0.000	21:32
Subject Sample # 1	0.238	21:33
Breath Volume	2.527	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:35
Subject Sample # 2	0.242	21:36
Breath Volume	2.609	21:36
Air Blank	0.000	21:37
Control Test	0.079	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:38

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MENDEZ CHACCHI

DATE/TIME OF TEST: 03/06/2017 03:41:19

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.080	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.114	03:35
Breath Volume	1.125	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.111	03:39
Breath Volume	1.929	03:39
Air Blank	0.000	03:40
Control Test	0.078	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ANTONETTI,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/06/2017 08:30:22

Diagnostics Check	OK	08:23
Air Blank	0.000	08:24
Control Test	0.080	08:24
Air Blank	0.000	08:25
Subject Sample # 1	0.131	08:25
Breath Volume	1.785	08:25
Air Blank	0.000	08:26
Air Blank	0.000	08:28
Subject Sample # 2	0.131	08:28
Breath Volume	2.054	08:28
Air Blank	0.000	08:29
Control Test	0.080	08:29
Air Blank	0.000	08:30
Diagnostics Check	OK	08:30

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GIANOUTSOS

DATE/TIME OF TEST: 03/08/2017 01:44:21

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:34
Air Blank	0.000	01:35
Subject Sample # 1	0.172	01:36
Breath Volume	1.531	01:36
Air Blank	0.000	01:36
Air Blank	0.000	01:38
Subject Sample # 2	0.202	01:39
Breath Volume	3.386	01:39
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 3	0.164	01:42
Breath Volume	1.515	01:42
Air Blank	0.000	01:43
Control Test	0.078	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SARAPOCHILLO

DATE/TIME OF TEST: 03/08/2017 21:46:47

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.163	21:40
Breath Volume	1.574	21:40
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.157	21:44
Breath Volume	2.093	21:44
Air Blank	0.000	21:45
Control Test	0.079	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MILLER,KARLY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: BRUNO

DATE/TIME OF TEST: 03/08/2017 23:15:19

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.081	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.000	23:10
Breath Volume	1.164	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.000	23:13
Breath Volume	1.507	23:13
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Diagnostics Check	OK	23:15

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FEQUIERE,RALPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HODGES JR.

DATE/TIME OF TEST: 03/09/2017 04:29:39

Diagnostics Check	OK	04:21
Air Blank	0.000	04:22
Control Test	0.081	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.216	04:24
Breath Volume	2.976	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:26
Subject Sample # 2	0.223	04:27
Breath Volume	3.367	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SANTELLI,PETER
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TIERNEY

DATE/TIME OF TEST: 03/10/2017 03:19:27

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.081	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.174	03:14
Breath Volume	2.589	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.172	03:17
Breath Volume	3.015	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: LABELLA

DATE/TIME OF TEST: 03/10/2017 03:52:57

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.079	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.316	03:48
Breath Volume	1.996	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:50
Subject Sample # 2	0.322	03:51
Breath Volume	1.750	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:52
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DOMINAK

DATE/TIME OF TEST: 03/11/2017 02:59:31

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.224	02:54
Breath Volume	1.515	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.224	02:57
Breath Volume	1.917	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 03/11/2017 04:40:46

Diagnostics Check	OK	04:33
Air Blank	0.000	04:33
Control Test	0.080	04:33
Air Blank	0.000	04:34
Subject Sample # 1	0.157	04:36
Breath Volume	2.000	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:38
Subject Sample # 2	0.154	04:38
Breath Volume	1.246	04:38
Air Blank	0.000	04:39
Control Test	0.079	04:40
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VICKERY,MEGAN
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: JEAN-BAPTISTE

DATE/TIME OF TEST: 03/11/2017 22:13:32

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.081	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.180	22:08
Breath Volume	1.164	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.187	22:11
Breath Volume	1.992	22:11
Air Blank	0.000	22:12
Control Test	0.079	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WEBER,RONALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/12/2017 07:28:54

Diagnostics Check	OK	07:22
Air Blank	0.000	07:22
Control Test	0.081	07:23
Air Blank	0.000	07:23
Subject Sample # 1	0.077	07:24
Breath Volume	3.273	07:24
Air Blank	0.000	07:25
Air Blank	0.000	07:26
Subject Sample # 2	0.077	07:27
Breath Volume	3.269	07:27
Air Blank	0.000	07:27
Control Test	0.081	07:28
Air Blank	0.000	07:28
Diagnostics Check	OK	07:28

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 03/13/2017 03:02:25

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.162	02:57
Breath Volume	1.121	02:57
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.163	03:00
Breath Volume	1.281	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 03/13/2017 03:49:14

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.246	03:43
Breath Volume	2.191	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.232	03:47
Breath Volume	2.191	03:47
Air Blank	0.000	03:48
Control Test	0.078	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GEE,JOHN
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: BLOCK

DATE/TIME OF TEST: 03/14/2017 01:17:41

Diagnostics Check	OK	01:10
Air Blank	0.000	01:10
Control Test	0.081	01:11
Air Blank	0.000	01:11
Subject Sample # 1	0.189	01:12
Breath Volume	2.718	01:12
Air Blank	0.000	01:13
Air Blank	0.000	01:15
Subject Sample # 2	0.187	01:15
Breath Volume	2.695	01:15
Air Blank	0.000	01:16
Control Test	0.079	01:17
Air Blank	0.000	01:17
Diagnostics Check	OK	01:17

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRIOSTPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: PENNICOTT

DATE/TIME OF TEST: 03/18/2017 05:32:54

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.082	05:26
Air Blank	0.000	05:27
Subject Sample # 1	0.078	05:27
Breath Volume	1.128	05:27
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.078	05:31
Breath Volume	1.375	05:31
Air Blank	0.000	05:31
Control Test	0.081	05:32
Air Blank	0.000	05:32
Diagnostics Check	OK	05:32

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 03/18/2017 22:01:52

Diagnostics Check	OK	21:54
Air Blank	0.000	21:55
Control Test	0.082	21:55
Air Blank	0.000	21:56
Subject Sample # 1	0.201	21:56
Breath Volume	2.550	21:56
Air Blank	0.000	21:57
Air Blank	0.000	21:59
Subject Sample # 2	0.201	22:00
Breath Volume	2.609	22:00
Air Blank	0.000	22:00
Control Test	0.079	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:01

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: WIRKKALA

DATE/TIME OF TEST: 03/18/2017 22:38:03

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.080	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.000	22:33
Breath Volume	2.710	22:33
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.000	22:36
Breath Volume	1.980	22:36
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:37
Diagnostics Check	OK	22:38

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEW,GARY
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: OCHOA MUNOZ

DATE/TIME OF TEST: 03/19/2017 06:18:10

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.081	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.185	06:13
Breath Volume	1.761	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.187	06:16
Breath Volume	1.171	06:16
Air Blank	0.000	06:17
Control Test	0.080	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 03/19/2017 21:01:14

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.081	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.000	20:56
Breath Volume	3.257	20:56
Air Blank	0.000	20:57
Air Blank	0.000	20:59
Subject Sample # 2	0.000	20:59
Breath Volume	2.269	20:59
Air Blank	0.000	21:00
Control Test	0.080	21:00
Air Blank	0.000	21:01
Diagnostics Check	OK	21:01

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: IV,ANDREW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: BESSETTE

DATE/TIME OF TEST: 03/19/2017 23:43:22

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.171	23:38
Breath Volume	2.816	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:40
Subject Sample # 2	0.177	23:41
Breath Volume	2.902	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LOBO

DATE/TIME OF TEST: 03/20/2017 00:47:26

Diagnostics Check	OK	00:40
Air Blank	0.000	00:41
Control Test	0.080	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.010	00:42
Breath Volume	1.906	00:42
Air Blank	0.000	00:43
Air Blank	0.000	00:45
Subject Sample # 2	0.010	00:46
Breath Volume	3.363	00:46
Air Blank	0.000	00:46
Control Test	0.080	00:46
Air Blank	0.000	00:47
Diagnostics Check	OK	00:47

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AIME,DIANA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: FLORIL

DATE/TIME OF TEST: 03/20/2017 01:44:42

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1 0.261	01:43	Volume Not Met
Breath Volume	0.875	01:43	
Air Blank		01:44	Purge Fail
Air Blank	0.000	01:44	

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AIME,DIANA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: FLORIL

DATE/TIME OF TEST: 03/20/2017 01:48:52

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:47	
Control Test	0.074	01:48	Control Outside Tolerance
Air Blank	0.000	01:48	

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORSEY,PAMELA,F
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ALVARADO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: HENAO

DATE/TIME OF TEST: 03/20/2017 22:16:53

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.081	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.165	22:12
Breath Volume	2.796	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.162	22:15
Breath Volume	1.250	22:15
Air Blank	0.000	22:15
Control Test	0.077	22:16
Air Blank	0.000	22:16
Diagnostics Check	OK	22:16

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: LILLY

DATE/TIME OF TEST: 03/21/2017 00:33:43

Diagnostics Check	OK	00:25
Air Blank	0.000	00:26
Control Test	0.080	00:26
Air Blank	0.000	00:27
Subject Sample # 1	0.146	00:28
Breath Volume	1.347	00:28
Air Blank	0.000	00:29
Air Blank	0.000	00:30
Subject Sample # 2	0.144	00:31
Breath Volume	1.261	00:31
Air Blank	0.000	00:32
Control Test	0.079	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: MACHADO COLON

DATE/TIME OF TEST: 03/22/2017 04:42:17

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.009	04:37
Breath Volume	2.859	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.007	04:40
Breath Volume	1.828	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 02/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DI ANTONIO

DATE/TIME OF TEST: 03/23/2017 04:02:00

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.081	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.220	03:57
Breath Volume	1.968	03:57
Air Blank	0.000	03:58
Air Blank	0.000	03:59
Subject Sample # 2	0.218	04:00
Breath Volume	2.113	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:02

Data Download Date/Time: 03/24/2017 12:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: ZELTSER

DATE/TIME OF TEST: 03/24/2017 20:17:56

Diagnostics Check	OK	20:10
Air Blank	0.000	20:11
Control Test	0.081	20:11
Air Blank	0.000	20:12
Subject Sample # 1	0.081	20:13
Breath Volume	2.507	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:15
Subject Sample # 2	0.079	20:16
Breath Volume	1.234	20:16
Air Blank	0.000	20:16
Control Test	0.078	20:17
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CARRANZA RIVERA

DATE/TIME OF TEST: 03/25/2017 02:41:48

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:30	
Control Test	0.080	02:30	
Air Blank	0.000	02:31	
Subject Sample	# 1 0.109	02:34	Volume Not Met
Breath Volume	0.214	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 2 0.115	02:39	Volume Not Met
Breath Volume	0.503	02:39	
Air Blank	0.000	02:40	
Control Test	0.077	02:41	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: CARRANZA RIVERA

DATE/TIME OF TEST: 03/25/2017 02:47:13

Diagnostics Check	OK	02:43	
Air Blank	0.000	02:43	
Control Test	0.077	02:44	
Air Blank	0.000	02:44	
Subject Sample	# 1	02:45	Subject Test Refused
Breath Volume	0.179	02:45	
Air Blank	0.000	02:46	
Control Test	0.079	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,ISRAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 03/25/2017 04:43:55

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.080	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.159	04:39
Breath Volume	1.496	04:39
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.162	04:42
Breath Volume	1.722	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 03/25/2017 21:40:53

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:36	
Control Test	0.081	21:36	
Air Blank	0.000	21:37	
Subject Sample	# 1 0.361	21:39	
Breath Volume	2.066	21:39	
Air Blank		21:40	Purge Fail
Air Blank	0.000	21:40	

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 03/25/2017 22:12:28

Diagnostics Check	OK	22:03
Air Blank	0.000	22:03
Control Test	0.077	22:04
Air Blank	0.000	22:04
Subject Sample # 1	0.330	22:07
Breath Volume	1.781	22:07
Air Blank	0.000	22:08
Air Blank	0.000	22:09
Subject Sample # 2	0.334	22:10
Breath Volume	2.035	22:10
Air Blank	0.000	22:11
Control Test	0.077	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FALCON,JESSICA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: LOPEZ GREGORIO

DATE/TIME OF TEST: 03/26/2017 00:20:56

Diagnostics Check	OK	00:11
Air Blank	0.000	00:12
Control Test	0.079	00:12
Air Blank	0.000	00:13
Subject Sample # 1	0.207	00:14
Breath Volume	1.250	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.202	00:19
Breath Volume	1.835	00:19
Air Blank	0.000	00:19
Control Test	0.076	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:20

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEHENY,JACOB
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: TROJANOWSKI

DATE/TIME OF TEST: 03/26/2017 01:18:49

Diagnostics Check	OK	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Subject Sample # 1	0.229	01:14
Breath Volume	2.957	01:14
Air Blank	0.000	01:15
Air Blank	0.000	01:16
Subject Sample # 2	0.233	01:16
Breath Volume	1.750	01:16
Air Blank	0.000	01:17
Control Test	0.078	01:18
Air Blank	0.000	01:18
Diagnostics Check	OK	01:18

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARIL
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTEEL,MEGAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: RUIZ CASAS

DATE/TIME OF TEST: 03/26/2017 04:29:27

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:22
Air Blank	0.000	04:23
Subject Sample # 1	0.259	04:23
Breath Volume	2.128	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.258	04:27
Breath Volume	2.414	04:27
Air Blank	0.000	04:28
Control Test	0.078	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CORUM,AMANDA
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: NUZZI

DATE/TIME OF TEST: 03/27/2017 22:03:23

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:58	
Control Test	0.081	21:58	
Air Blank	0.000	21:59	
Subject Sample	# 1	22:01	Subject Test Refused
Breath Volume	0.000	22:01	
Air Blank	0.000	22:02	
Control Test	0.080	22:02	
Air Blank	0.000	22:03	
Diagnostics Check	OK	22:03	

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MOGO

DATE/TIME OF TEST: 03/28/2017 01:03:52

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.080	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.097	00:59
Breath Volume	1.410	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.091	01:02
Breath Volume	1.527	01:02
Air Blank	0.000	01:02
Control Test	0.077	01:03
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZEITZ,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GOLLADAY

DATE/TIME OF TEST: 03/29/2017 00:24:20

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.080	00:13	
Air Blank	0.000	00:13	
Subject Sample	# 1 0.211	00:16	Volume Not Met
Breath Volume	0.839	00:16	
Air Blank	0.000	00:17	
Air Blank	0.000	00:19	
Subject Sample	# 2 0.193	00:22	Volume Not Met
Breath Volume	0.640	00:22	
Air Blank	0.000	00:23	
Control Test	0.074	00:23	Control Outside Tolerance
Air Blank	0.000	00:24	

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZEITZ,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: GOLLADAY

DATE/TIME OF TEST: 03/29/2017 00:30:59

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.075	00:27	
Air Blank	0.000	00:28	
Subject Sample	# 1	00:29	Subject Test Refused
Breath Volume	0.000	00:29	
Air Blank	0.000	00:30	
Control Test	0.078	00:30	
Air Blank	0.000	00:30	
Diagnostics Check	OK	00:30	

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: YOCHUM,CRAIG
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ROSAS

DATE/TIME OF TEST: 03/29/2017 04:59:01

Diagnostics Check	OK	04:51
Air Blank	0.000	04:51
Control Test	0.080	04:52
Air Blank	0.000	04:52
Subject Sample # 1	0.183	04:53
Breath Volume	1.957	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.179	04:57
Breath Volume	2.023	04:57
Air Blank	0.000	04:58
Control Test	0.078	04:58
Air Blank	0.000	04:58
Diagnostics Check	OK	04:59

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CUMMINGS,PETER
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PAIZ

DATE/TIME OF TEST: 03/29/2017 22:03:36

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.194	21:58
Breath Volume	1.957	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.197	22:01
Breath Volume	2.074	22:01
Air Blank	0.000	22:02
Control Test	0.078	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: YILDIRIM

DATE/TIME OF TEST: 03/29/2017 23:22:59

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.194	23:18
Breath Volume	1.718	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:20
Subject Sample # 2	0.207	23:21
Breath Volume	2.542	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VARGAS,CHRISTIAN
ARRESTING OFFICER AGENCY: Manalapan Pub Safety
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: ALARCON ARELLANO

DATE/TIME OF TEST: 03/31/2017 01:08:28

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:02
Air Blank	0.000	01:02
Subject Sample # 1	0.156	01:03
Breath Volume	1.105	01:03
Air Blank	0.000	01:04
Air Blank	0.000	01:05
Subject Sample # 2	0.150	01:06
Breath Volume	1.300	01:06
Air Blank	0.000	01:07
Control Test	0.078	01:07
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 04/21/2017 15:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: RODGERS

DATE/TIME OF TEST: 03/25/2017 22:25:54

Diagnostics Check	OK	22:19
Air Blank	0.000	22:20
Control Test	0.080	22:20
Air Blank	0.000	22:20
Subject Sample # 1	0.172	22:21
Breath Volume	2.507	22:21
Air Blank	0.000	22:22
Air Blank	0.000	22:23
Subject Sample # 2	0.176	22:24
Breath Volume	2.488	22:24
Air Blank	0.000	22:24
Control Test	0.080	22:25
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 04/21/2017 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BLASCO,LUIS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:40

SUBJECT'S LAST NAME: SIDELLA

DATE/TIME OF TEST: 03/26/2017 10:11:51

Diagnostics Check	OK	10:05
Air Blank	0.000	10:06
Control Test	0.080	10:06
Air Blank	0.000	10:06
Subject Sample # 1	0.000	10:07
Breath Volume	1.765	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:10
Subject Sample # 2	0.000	10:10
Breath Volume	1.164	10:10
Air Blank	0.000	10:10
Control Test	0.080	10:11
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

Data Download Date/Time: 04/21/2017 13:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FALCON,JESSICA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BACILIO JUAN

DATE/TIME OF TEST: 03/26/2017 23:10:43

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.080	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.197	23:06
Breath Volume	2.316	23:06
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.196	23:09
Breath Volume	2.960	23:09
Air Blank	0.000	23:09
Control Test	0.079	23:10
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 04/21/2017 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: TAVAREZ

DATE/TIME OF TEST: 03/30/2017 02:14:36

Diagnostics Check	OK	02:07
Air Blank	0.000	02:08
Control Test	0.080	02:08
Air Blank	0.000	02:09
Subject Sample # 1	0.139	02:09
Breath Volume	1.308	02:09
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.144	02:12
Breath Volume	1.660	02:12
Air Blank	0.000	02:13
Control Test	0.079	02:13
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 04/21/2017 13:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASILVA,CARLOS,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DASILVA,CARLOS
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SCRIVNER

DATE/TIME OF TEST: 03/01/2017 23:07:27

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.179	23:02
Breath Volume	2.250	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.178	23:05
Breath Volume	1.480	23:05
Air Blank	0.000	23:06
Control Test	0.077	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 03/07/2017 01:12:38

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.078	01:05	
Air Blank	0.000	01:05	
Subject Sample	# 1 0.179	01:08	
Breath Volume	1.710	01:08	
Air Blank	0.000	01:09	
Air Blank	0.000	01:10	
Subject Sample	# 2	01:11	Subject Test Refused
Breath Volume	0.000	01:11	
Air Blank	0.000	01:11	
Control Test	0.073	01:12	Control Outside Tolerance
Air Blank	0.000	01:12	

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 03/07/2017 01:17:19

Diagnostics Check	OK	01:14	
Air Blank	0.000	01:14	
Control Test	0.077	01:15	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:15	Subject Test Refused
Breath Volume	0.000	01:15	
Air Blank	0.000	01:16	
Control Test	0.077	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: MCAULEY

DATE/TIME OF TEST: 03/07/2017 21:14:22

Diagnostics Check	OK	21:07
Air Blank	0.000	21:07
Control Test	0.078	21:08
Air Blank	0.000	21:08
Subject Sample # 1	0.136	21:09
Breath Volume	2.312	21:09
Air Blank	0.000	21:10
Air Blank	0.000	21:11
Subject Sample # 2	0.140	21:12
Breath Volume	3.300	21:12
Air Blank	0.000	21:13
Control Test	0.077	21:13
Air Blank	0.000	21:14
Diagnostics Check	OK	21:14

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/11/2017 01:24:10

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:21	
Control Test	0.078	01:21	
Air Blank	0.000	01:22	
Subject Sample	# 1	01:22	Slope Not Met
Breath Volume	1.570	01:22	
Air Blank	0.000	01:23	
Control Test	0.078	01:23	
Air Blank	0.000	01:24	
Diagnostics Check	OK	01:24	

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 03/11/2017 01:34:53

Diagnostics Check	OK	01:25
Air Blank	0.000	01:25
Control Test	0.078	01:26
Air Blank	0.000	01:26
Subject Sample # 1	0.248	01:27
Breath Volume	1.121	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:29
Subject Sample # 2	0.222	01:30
Breath Volume	1.156	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 3	0.246	01:33
Breath Volume	1.363	01:33
Air Blank	0.000	01:33
Control Test	0.075	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/11/2017 02:31:28

Diagnostics Check	OK	02:28	
Air Blank	0.000	02:29	
Control Test	0.078	02:29	
Air Blank	0.000	02:29	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:30	
Control Test	0.078	02:30	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:31	

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: PERUSEK II

DATE/TIME OF TEST: 03/15/2017 04:42:11

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.079	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.123	04:34
Breath Volume	1.160	04:34
Air Blank	0.000	04:35
Air Blank	0.000	04:36
Subject Sample # 2	0.149	04:37
Breath Volume	2.832	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 3	0.126	04:40
Breath Volume	1.480	04:40
Air Blank	0.000	04:41
Control Test	0.078	04:41
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: ULMER

DATE/TIME OF TEST: 03/17/2017 22:45:46

Diagnostics Check	OK	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:40
Subject Sample # 1	0.234	22:41
Breath Volume	3.113	22:41
Air Blank	0.000	22:42
Air Blank	0.000	22:43
Subject Sample # 2	0.232	22:44
Breath Volume	3.039	22:44
Air Blank	0.000	22:44
Control Test	0.077	22:45
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MERRILL,PERRY
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 03/23/2017 05:12:33

Diagnostics Check	OK	05:06
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.232	05:07
Breath Volume	3.621	05:07
Air Blank	0.000	05:08
Air Blank	0.000	05:10
Subject Sample # 2	0.228	05:10
Breath Volume	4.671	05:10
Air Blank	0.000	05:11
Control Test	0.077	05:11
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOONANAN,R,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: GWODZ,R
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: TAPIA

DATE/TIME OF TEST: 03/24/2017 03:50:50

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.076	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.256	03:46
Breath Volume	2.066	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.257	03:49
Breath Volume	2.515	03:49
Air Blank	0.000	03:49
Control Test	0.077	03:50
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SIERRA FUNEZ

DATE/TIME OF TEST: 03/25/2017 03:56:42

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.078	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.078	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 03/27/2017 08:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEVERSTON,JONATHAN,E
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: FEVERSTON,JONATHAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 03/31/2017 03:31:38

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.184	03:25
Breath Volume	1.343	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:27
Subject Sample # 2	0.195	03:29
Breath Volume	1.902	03:29
Air Blank	0.000	03:30
Control Test	0.077	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/24/2017 11:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PEREZ RAMIREZ

DATE/TIME OF TEST: 03/04/2017 03:28:20

Diagnostics Check	OK	03:22	
Air Blank	0.000	03:23	
Control Test	0.082	03:23	
Air Blank	0.000	03:23	
Subject Sample	# 1 0.175	03:27	Volume Not Met
Breath Volume	0.097	03:27	
Air Blank		03:27	Purge Fail
Air Blank	0.000	03:28	

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MENCOS

DATE/TIME OF TEST: 03/05/2017 00:54:24

Diagnostics Check	OK	00:47
Air Blank	0.000	00:47
Control Test	0.082	00:48
Air Blank	0.000	00:48
Subject Sample # 1	0.161	00:49
Breath Volume	3.078	00:49
Air Blank	0.000	00:50
Air Blank	0.000	00:51
Subject Sample # 2	0.157	00:52
Breath Volume	3.113	00:52
Air Blank	0.000	00:53
Control Test	0.079	00:53
Air Blank	0.000	00:54
Diagnostics Check	OK	00:54

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN, WILLIAM, B
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: CARADIMOS, MATTHEW
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 03/06/2017 01:40:29

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.081	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PETERSON,RONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: MCIVER

DATE/TIME OF TEST: 03/14/2017 01:33:39

Diagnostics Check	OK	01:27
Air Blank	0.000	01:27
Control Test	0.081	01:28
Air Blank	0.000	01:28
Subject Sample # 1	0.208	01:29
Breath Volume	1.332	01:29
Air Blank	0.000	01:29
Air Blank	0.000	01:31
Subject Sample # 2	0.210	01:31
Breath Volume	1.167	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: SPERANZA

DATE/TIME OF TEST: 03/14/2017 04:12:30

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.153	04:07
Breath Volume	2.480	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.155	04:10
Breath Volume	2.285	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGREDE,JOHN,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MCGREDE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JANTES JR

DATE/TIME OF TEST: 03/16/2017 00:32:11

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:28	
Control Test	0.081	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Control Test	0.081	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 03/18/2017 03:17:44

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.081	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.158	03:13
Breath Volume	2.062	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.148	03:15
Breath Volume	1.492	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: NAVA NAVA

DATE/TIME OF TEST: 03/19/2017 00:00:45

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.080	23:49	
Air Blank	0.000	23:50	
Subject Sample	# 1 0.228	23:51	No .020 agreement
Breath Volume	1.640	23:51	
Air Blank	0.000	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 2 0.255	23:55	No .020 agreement
Breath Volume	1.609	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 3 0.206	23:58	No .020 agreement
Breath Volume	1.949	23:58	
Air Blank	0.000	23:59	
Control Test	0.078	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: RAMOS LOPEZ

DATE/TIME OF TEST: 03/19/2017 21:42:15

Diagnostics Check	OK	21:34
Air Blank	0.000	21:35
Control Test	0.081	21:35
Air Blank	0.000	21:36
Subject Sample # 1	0.197	21:37
Breath Volume	2.359	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:39
Subject Sample # 2	0.186	21:40
Breath Volume	2.730	21:40
Air Blank	0.000	21:41
Control Test	0.078	21:41
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GOSSAGE

DATE/TIME OF TEST: 03/24/2017 03:21:57

Diagnostics Check	OK	03:19	
Air Blank	0.000	03:19	
Control Test	0.081	03:19	
Air Blank	0.000	03:20	
Subject Sample	# 1	03:20	Subject Test Refused
Breath Volume	0.000	03:20	
Air Blank	0.000	03:21	
Control Test	0.080	03:21	
Air Blank	0.000	03:21	
Diagnostics Check	OK	03:21	

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN, WILLIAM, B
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: REAUME, JASON
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: MENCHU

DATE/TIME OF TEST: 03/25/2017 01:40:26

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.077	01:38	
Air Blank	0.000	01:38	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.078	01:39	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SOLLEY,KATRYNA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: MURRAY

DATE/TIME OF TEST: 03/25/2017 05:33:11

Diagnostics Check	OK	05:25
Air Blank	0.000	05:26
Control Test	0.081	05:26
Air Blank	0.000	05:26
Subject Sample # 1	0.164	05:28
Breath Volume	1.460	05:28
Air Blank	0.000	05:28
Air Blank	0.000	05:30
Subject Sample # 2	0.162	05:31
Breath Volume	2.281	05:31
Air Blank	0.000	05:32
Control Test	0.080	05:32
Air Blank	0.000	05:33
Diagnostics Check	OK	05:33

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: DOWNING

DATE/TIME OF TEST: 03/26/2017 16:50:51

Diagnostics Check	OK	16:44
Air Blank	0.000	16:44
Control Test	0.080	16:44
Air Blank	0.000	16:45
Subject Sample # 1	0.326	16:46
Breath Volume	2.351	16:46
Air Blank	0.000	16:46
Air Blank	0.000	16:48
Subject Sample # 2	0.336	16:49
Breath Volume	2.679	16:49
Air Blank	0.000	16:49
Control Test	0.079	16:50
Air Blank	0.000	16:50
Diagnostics Check	OK	16:50

Data Download Date/Time: 03/27/2017 08:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER, A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: WILLIAMS II

DATE/TIME OF TEST: 03/01/2017 04:53:53

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.080	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.142	04:49
Breath Volume	2.105	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:51
Subject Sample # 2	0.137	04:52
Breath Volume	1.640	04:52
Air Blank	0.000	04:52
Control Test	0.080	04:53
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 03/16/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LARA, A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: FALLON

DATE/TIME OF TEST: 03/01/2017 22:23:01

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.080	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.144	22:18
Breath Volume	2.429	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:20
Subject Sample # 2	0.152	22:21
Breath Volume	2.179	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:22
Diagnostics Check	OK	22:23

Data Download Date/Time: 03/16/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: MURDOCK

DATE/TIME OF TEST: 03/02/2017 05:01:57

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:55
Subject Sample # 1	0.237	04:57
Breath Volume	3.437	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.234	05:00
Breath Volume	2.687	05:00
Air Blank	0.000	05:01
Control Test	0.079	05:01
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/16/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: COGLEY

DATE/TIME OF TEST: 03/03/2017 02:06:45

Diagnostics Check	OK	02:00
Air Blank	0.000	02:00
Control Test	0.080	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.114	02:02
Breath Volume	3.230	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.113	02:05
Breath Volume	2.789	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 03/16/2017 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: M,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 03/03/2017 23:54:20

Diagnostics Check	OK	23:48	
Air Blank	0.000	23:49	
Control Test	0.080	23:49	
Air Blank	0.000	23:49	
Subject Sample	# 1 0.000	23:50	
Breath Volume	3.195	23:50	
Air Blank	0.000	23:50	
Air Blank	0.000	23:52	
Subject Sample	# 2	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 03/16/2017 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: IVY JR

DATE/TIME OF TEST: 03/04/2017 02:51:54

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.055	02:47
Breath Volume	2.339	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:49
Subject Sample # 2	0.055	02:50
Breath Volume	2.304	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 03/16/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 03/04/2017 03:49:16

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:43
Air Blank	0.000	03:44
Subject Sample # 1	0.159	03:44
Breath Volume	1.535	03:44
Air Blank	0.000	03:45
Air Blank	0.000	03:47
Subject Sample # 2	0.161	03:47
Breath Volume	1.339	03:47
Air Blank	0.000	03:48
Control Test	0.080	03:48
Air Blank	0.000	03:49
Diagnostics Check	OK	03:49

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ANDERSON JR

DATE/TIME OF TEST: 03/04/2017 04:10:26

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Subject Sample # 1	0.151	04:05
Breath Volume	3.378	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.152	04:08
Breath Volume	3.472	04:08
Air Blank	0.000	04:09
Control Test	0.080	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: CASTANEDA JUAREZ

DATE/TIME OF TEST: 03/04/2017 20:21:54

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.080	20:15
Air Blank	0.000	20:16
Subject Sample # 1	0.263	20:17
Breath Volume	2.144	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.262	20:20
Breath Volume	2.117	20:20
Air Blank	0.000	20:20
Control Test	0.079	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE, TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 03/04/2017 21:42:35

Diagnostics Check	OK	21:31	
Air Blank	0.000	21:31	
Control Test	0.081	21:31	
Air Blank	0.000	21:32	
Subject Sample	# 1	21:35	No Sample Provided
Breath Volume	0.000	21:35	
Air Blank	0.000	21:36	
Air Blank	0.000	21:38	
Subject Sample	# 2	21:41	No Sample Provided
Breath Volume	0.000	21:41	
Air Blank	0.000	21:41	
Control Test	0.080	21:41	
Air Blank	0.000	21:42	
Diagnostics Check	OK	21:42	

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG, G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 03/05/2017 01:14:14

Diagnostics Check	OK	01:07
Air Blank	0.000	01:08
Control Test	0.080	01:08
Air Blank	0.000	01:08
Subject Sample # 1	0.231	01:09
Breath Volume	2.058	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.233	01:12
Breath Volume	2.007	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAMER,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 03/05/2017 03:08:07

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Subject Sample # 1	0.123	03:03
Breath Volume	3.156	03:03
Air Blank	0.000	03:04
Air Blank	0.000	03:05
Subject Sample # 2	0.121	03:06
Breath Volume	2.984	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Diagnostics Check	OK	03:08

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD, J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 03/05/2017 03:57:30

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.160	03:52
Breath Volume	1.812	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.154	03:55
Breath Volume	1.515	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: DELGADO DIAZ

DATE/TIME OF TEST: 03/05/2017 04:48:05

Diagnostics Check	OK	04:42	
Air Blank	0.000	04:43	
Control Test	0.079	04:43	
Air Blank	0.000	04:44	
Subject Sample	# 1	04:46	Subject Test Refused
Breath Volume	0.000	04:46	
Air Blank	0.000	04:47	
Control Test	0.080	04:47	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: LABOY

DATE/TIME OF TEST: 03/05/2017 23:19:27

Diagnostics Check	OK	23:12
Air Blank	0.000	23:13
Control Test	0.081	23:13
Air Blank	0.000	23:13
Subject Sample # 1	0.145	23:14
Breath Volume	3.488	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.140	23:17
Breath Volume	2.324	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BADEN,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 03/06/2017 02:44:01

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.113	02:38
Breath Volume	1.921	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:40
Subject Sample # 2	0.112	02:42
Breath Volume	2.285	02:42
Air Blank	0.000	02:43
Control Test	0.079	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:44

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARFIELD-CRAIG,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 03/07/2017 02:58:39

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.076	02:53
Breath Volume	1.457	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.077	02:57
Breath Volume	1.253	02:57
Air Blank	0.000	02:57
Control Test	0.079	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MORENO JR

DATE/TIME OF TEST: 03/08/2017 03:54:57

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:43	
Control Test	0.080	03:43	
Air Blank	0.000	03:44	
Subject Sample	# 1 0.163	03:47	Volume Not Met
Breath Volume	0.019	03:47	
Air Blank	0.000	03:48	
Air Blank	0.000	03:50	
Subject Sample	# 2 0.162	03:53	Volume Not Met
Breath Volume	0.796	03:53	
Air Blank	0.000	03:53	
Control Test	0.078	03:54	
Air Blank	0.000	03:54	
Diagnostics Check	OK	03:54	

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: PAPROCKI

DATE/TIME OF TEST: 03/09/2017 01:09:39

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:56	
Control Test	0.079	00:56	
Air Blank	0.000	00:57	
Subject Sample	# 1 0.168	01:00	Volume Not Met
Breath Volume	0.187	01:00	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2 0.173	01:03	
Breath Volume	1.230	01:03	
Air Blank	0.000	01:04	
Air Blank	0.000	01:06	
Subject Sample	# 3 0.183	01:08	
Breath Volume	2.031	01:08	
Air Blank	0.000	01:08	
Control Test	0.078	01:09	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 03/09/2017 04:28:55

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.147	04:23
Breath Volume	2.000	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:25
Subject Sample # 2	0.150	04:27
Breath Volume	2.585	04:27
Air Blank	0.000	04:27
Control Test	0.079	04:28
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SULLIVAN,T
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: GIRTMAN

DATE/TIME OF TEST: 03/09/2017 21:03:28

Diagnostics Check	OK	20:54
Air Blank	0.000	20:54
Control Test	0.080	20:55
Air Blank	0.000	20:55
Subject Sample # 1	0.000	20:58
Breath Volume	1.320	20:58
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.000	21:02
Breath Volume	1.183	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:02
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: HAYES III

DATE/TIME OF TEST: 03/10/2017 03:27:36

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:20
Subject Sample # 1	0.222	03:21
Breath Volume	2.179	03:21
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.226	03:25
Breath Volume	2.742	03:25
Air Blank	0.000	03:26
Control Test	0.078	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK, M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: LEVIERE

DATE/TIME OF TEST: 03/10/2017 22:57:21

Diagnostics Check	OK	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Subject Sample # 1	0.133	22:52
Breath Volume	2.035	22:52
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.135	22:55
Breath Volume	2.234	22:55
Air Blank	0.000	22:56
Control Test	0.079	22:56
Air Blank	0.000	22:57
Diagnostics Check	OK	22:57

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,PATRICIA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: O'NEILL,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: KAGAWA

DATE/TIME OF TEST: 03/11/2017 02:02:29

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.080	01:56
Air Blank	0.000	01:56
Subject Sample # 1	0.181	01:57
Breath Volume	1.152	01:57
Air Blank	0.000	01:58
Air Blank	0.000	01:59
Subject Sample # 2	0.182	02:00
Breath Volume	1.929	02:00
Air Blank	0.000	02:01
Control Test	0.078	02:01
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND, CHARLEENE, N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE, S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 03/11/2017 05:01:12

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.157	04:56
Breath Volume	3.148	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.157	04:59
Breath Volume	2.996	04:59
Air Blank	0.000	05:00
Control Test	0.078	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WARD,E
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: ROSA RIVERA

DATE/TIME OF TEST: 03/11/2017 20:59:57

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.233	20:54
Breath Volume	1.656	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.218	20:58
Breath Volume	2.699	20:58
Air Blank	0.000	20:59
Control Test	0.077	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARFIELD-CRAIG,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: SAEN NGARM

DATE/TIME OF TEST: 03/11/2017 23:27:26

Diagnostics Check	OK	23:19
Air Blank	0.000	23:19
Control Test	0.079	23:20
Air Blank	0.000	23:20
Subject Sample # 1	0.176	23:22
Breath Volume	1.953	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:24
Subject Sample # 2	0.174	23:25
Breath Volume	2.269	23:25
Air Blank	0.000	23:26
Control Test	0.078	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: DURAN

DATE/TIME OF TEST: 03/12/2017 03:10:37

Diagnostics Check	OK	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.116	03:05
Breath Volume	2.925	03:05
Air Blank	0.000	03:06
Air Blank	0.000	03:08
Subject Sample # 2	0.118	03:09
Breath Volume	4.031	03:09
Air Blank	0.000	03:09
Control Test	0.078	03:10
Air Blank	0.000	03:10
Diagnostics Check	OK	03:10

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:24

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 03/12/2017 21:52:22

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.080	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.207	21:47
Breath Volume	1.804	21:47
Air Blank	0.000	21:48
Air Blank	0.000	21:50
Subject Sample # 2	0.208	21:50
Breath Volume	1.640	21:50
Air Blank	0.000	21:51
Control Test	0.078	21:51
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: DUNPHY

DATE/TIME OF TEST: 03/13/2017 01:13:04

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.142	01:08
Breath Volume	2.542	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.137	01:11
Breath Volume	2.085	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:12
Diagnostics Check	OK	01:13

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WILLIAMS,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:57

SUBJECT'S LAST NAME: GOINS

DATE/TIME OF TEST: 03/13/2017 17:28:36

Diagnostics Check	OK	17:18
Air Blank	0.000	17:19
Control Test	0.080	17:19
Air Blank	0.000	17:20
Subject Sample # 1	0.281	17:20
Breath Volume	1.734	17:20
Air Blank	0.000	17:21
Air Blank	0.000	17:23
Subject Sample # 2	0.249	17:23
Breath Volume	1.453	17:23
Air Blank	0.000	17:24
Air Blank	0.000	17:26
Subject Sample # 3	0.259	17:26
Breath Volume	2.472	17:26
Air Blank	0.000	17:27
Control Test	0.078	17:27
Air Blank	0.000	17:28
Diagnostics Check	OK	17:28

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: ESCANDON FERRA

DATE/TIME OF TEST: 03/14/2017 20:46:53

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.080	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.000	20:42
Breath Volume	1.328	20:42
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.000	20:45
Breath Volume	1.941	20:45
Air Blank	0.000	20:45
Control Test	0.080	20:46
Air Blank	0.000	20:46
Diagnostics Check	OK	20:46

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DEJESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 03/15/2017 02:40:26

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.135	02:35
Breath Volume	2.910	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.137	02:38
Breath Volume	2.222	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 02/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: ROWLAND

DATE/TIME OF TEST: 03/16/2017 04:18:51

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.060	04:14
Breath Volume	4.003	04:14
Air Blank	0.000	04:14
Air Blank	0.000	04:16
Subject Sample # 2	0.062	04:17
Breath Volume	4.136	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Diagnostics Check	OK	04:18

Data Download Date/Time: 03/16/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESCARTES,PETER,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COZZONE,GAETANO
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 03/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 28515080A3
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: NIDAY

DATE/TIME OF TEST: 03/29/2017 05:17:55

Diagnostics Check	OK	05:15	
Air Blank	0.000	05:15	
Control Test	0.079	05:15	
Air Blank	0.000	05:16	
Subject Sample	# 1	05:16	Subject Test Refused
Breath Volume	0.000	05:16	
Air Blank	0.000	05:16	
Control Test	0.078	05:17	
Air Blank	0.000	05:17	
Diagnostics Check	OK	05:17	

Data Download Date/Time: 04/28/2017 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PARE,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PARE

DATE/TIME OF TEST: 03/11/2017 23:11:12

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1 0.000	23:07	
Breath Volume	1.976	23:07	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 2	23:09	Subject Test Refused
Breath Volume	0.000	23:09	
Air Blank	0.000	23:10	
Control Test	0.078	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENZ,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: QUETANT

DATE/TIME OF TEST: 03/13/2017 20:47:39

Diagnostics Check	OK	20:44	
Air Blank	0.000	20:45	
Control Test	0.076	20:45	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:46	Subject Test Refused
Breath Volume	0.000	20:46	
Air Blank	0.000	20:46	
Control Test	0.078	20:47	
Air Blank	0.000	20:47	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,A,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: REISNER,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 03/17/2017 16:04:31

Diagnostics Check	OK	15:57
Air Blank	0.000	15:57
Control Test	0.077	15:57
Air Blank	0.000	15:58
Subject Sample # 1	0.232	15:59
Breath Volume	2.437	15:59
Air Blank	0.000	16:00
Air Blank	0.000	16:01
Subject Sample # 2	0.230	16:02
Breath Volume	2.335	16:02
Air Blank	0.000	16:03
Control Test	0.077	16:03
Air Blank	0.000	16:04
Diagnostics Check	OK	16:04

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MARANO

DATE/TIME OF TEST: 03/17/2017 22:42:17

Diagnostics Check	OK	22:35
Air Blank	0.000	22:35
Control Test	0.076	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.162	22:37
Breath Volume	1.816	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:39
Subject Sample # 2	0.158	22:40
Breath Volume	1.968	22:40
Air Blank	0.000	22:41
Control Test	0.078	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: DEARMAS,DIANEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ALBERTS

DATE/TIME OF TEST: 03/18/2017 02:51:51

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:49	
Control Test	0.078	02:49	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.000	02:50	
Air Blank	0.000	02:50	
Control Test	0.078	02:51	
Air Blank	0.000	02:51	
Diagnostics Check	OK	02:51	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GEHR

DATE/TIME OF TEST: 03/18/2017 03:15:17

Diagnostics Check	OK	03:07
Air Blank	0.000	03:08
Control Test	0.077	03:08
Air Blank	0.000	03:09
Subject Sample # 1	0.189	03:09
Breath Volume	1.531	03:09
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.187	03:13
Breath Volume	1.523	03:13
Air Blank	0.000	03:14
Control Test	0.079	03:14
Air Blank	0.000	03:15
Diagnostics Check	OK	03:15

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VAN CAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:33

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 03/21/2017 19:18:58

Diagnostics Check	OK	19:10
Air Blank	0.000	19:11
Control Test	0.077	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.243	19:12
Breath Volume	1.812	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.226	19:16
Breath Volume	1.566	19:16
Air Blank	0.000	19:17
Control Test	0.077	19:18
Air Blank	0.000	19:18
Diagnostics Check	OK	19:18

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: FONG,K
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: FREY

DATE/TIME OF TEST: 03/23/2017 01:45:09

Diagnostics Check	OK	01:42	
Air Blank	0.000	01:42	
Control Test	0.077	01:43	
Air Blank	0.000	01:43	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.078	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: AURIEMMA

DATE/TIME OF TEST: 03/23/2017 19:35:16

Diagnostics Check	OK	19:32	
Air Blank	0.000	19:32	
Control Test	0.078	19:33	
Air Blank	0.000	19:33	
Subject Sample	# 1	19:33	Subject Test Refused
Breath Volume	0.000	19:33	
Air Blank	0.000	19:34	
Control Test	0.079	19:34	
Air Blank	0.000	19:35	
Diagnostics Check	OK	19:35	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALVAREZ,AMANDA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: APPLEBAUM

DATE/TIME OF TEST: 03/23/2017 22:55:29

Diagnostics Check	OK	22:52	
Air Blank	0.000	22:53	
Control Test	0.078	22:53	
Air Blank	0.000	22:53	
Subject Sample	# 1	22:54	Subject Test Refused
Breath Volume	0.000	22:54	
Air Blank	0.000	22:54	
Control Test	0.078	22:54	
Air Blank	0.000	22:55	
Diagnostics Check	OK	22:55	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: BRADWAY

DATE/TIME OF TEST: 03/24/2017 21:55:23

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:52	
Control Test	0.078	21:53	
Air Blank	0.000	21:53	
Subject Sample	# 1	21:53	Subject Test Refused
Breath Volume	0.000	21:53	
Air Blank	0.000	21:54	
Control Test	0.080	21:54	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRIGGER,ROHALIO,E
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: ESCOWITZ

DATE/TIME OF TEST: 03/26/2017 15:03:37

Diagnostics Check	OK	14:57
Air Blank	0.000	14:57
Control Test	0.078	14:58
Air Blank	0.000	14:58
Subject Sample # 1	0.000	14:59
Breath Volume	1.300	14:59
Air Blank	0.000	14:59
Air Blank	0.000	15:01
Subject Sample # 2	0.000	15:02
Breath Volume	1.421	15:02
Air Blank	0.000	15:02
Control Test	0.079	15:03
Air Blank	0.000	15:03
Diagnostics Check	OK	15:03

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCK,NATHAN,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BROCK,NATHAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: BUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LINS

DATE/TIME OF TEST: 03/29/2017 00:04:41

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.078	23:53	
Air Blank	0.000	23:54	
Subject Sample	# 2	23:57	No Sample Provided
Breath Volume	0.000	23:57	
Air Blank	0.000	23:57	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:02	Volume Not Met
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.076	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROCK,NATHAN,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BROCK,NATHAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: BUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: LINS

DATE/TIME OF TEST: 03/29/2017 00:16:04

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.077	00:08
Air Blank	0.000	00:09
Subject Sample # 1	0.145	00:10
Breath Volume	1.625	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.147	00:14
Breath Volume	2.000	00:14
Air Blank	0.000	00:15
Control Test	0.077	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:16

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:27

SUBJECT'S LAST NAME: FERA

DATE/TIME OF TEST: 03/29/2017 22:14:38

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:12	
Control Test	0.077	22:12	
Air Blank	0.000	22:13	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:13	
Control Test	0.077	22:14	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: REISSI,DANIEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: BARWELL

DATE/TIME OF TEST: 03/31/2017 01:35:38

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:33	
Control Test	0.078	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 04/26/2017 14:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CORONADO,ROBERT
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 03/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: ELDREDGE

DATE/TIME OF TEST: 03/31/2017 01:48:16

Diagnostics Check	OK	01:41
Air Blank	0.000	01:42
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.183	01:43
Breath Volume	3.093	01:43
Air Blank	0.000	01:44
Air Blank	0.000	01:45
Subject Sample # 2	0.186	01:46
Breath Volume	4.000	01:46
Air Blank	0.000	01:47
Control Test	0.079	01:47
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 04/26/2017 14:58

Subject Test Electronic Data

DATE/TIME OF TEST: 03/04/2017 01:17:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCGLYNN,JOHN,L

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MCGLYNN,JOHN

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 22:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: SUMMERALL

DATE/TIME OF TEST: 03/08/2017 22:27:04

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Subject Sample # 1	0.168	22:22
Breath Volume	1.722	22:22
Air Blank	0.000	22:23
Air Blank	0.000	22:24
Subject Sample # 2	0.165	22:25
Breath Volume	2.109	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:27

Data Download Date/Time: 03/28/2017 09:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 06616080A2
BREATH TEST OPERATOR: MCGLYNN,JOHN,L	DRY GAS EXPIRY DATE: 04/05/2018
BREATH TEST OPERATOR AGENCY: FHP Troop E	VIOLATION CODE: DUI
ARRESTING OFFICER: MCGLYNN,JOHN	ARREST TIME: 03:48
ARRESTING OFFICER AGENCY: FHP Troop E	GENDER: Male
LAST AGENCY INSPECTION: 02/27/2017	OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: APARCEDO DATE/TIME OF TEST: 03/16/2017 04:16:40

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.081	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.116	04:12
Breath Volume	1.640	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.125	04:15
Breath Volume	1.855	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 06616080A2
BREATH TEST OPERATOR: MCGLYNN,JOHN,L	DRY GAS EXPIRY DATE: 04/05/2018
BREATH TEST OPERATOR AGENCY: FHP Troop E	VIOLATION CODE: DUI
ARRESTING OFFICER: MCGLYNN,JOHN	ARREST TIME: 04:33
ARRESTING OFFICER AGENCY: FHP Troop E	GENDER: Male
LAST AGENCY INSPECTION: 02/27/2017	OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: GARCIA	DATE/TIME OF TEST: 03/18/2017 05:00:04
-----------------------------	--

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.110	04:55
Breath Volume	3.140	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.112	04:58
Breath Volume	2.613	04:58
Air Blank	0.000	04:59
Control Test	0.079	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: LASSO SILVA,NICOLAS

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:11

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MARTY PADILLA

DATE/TIME OF TEST: 03/19/2017 00:41:24

Diagnostics Check	OK	00:38	
Air Blank	0.000	00:39	
Control Test	0.081	00:39	
Air Blank	0.000	00:39	
Subject Sample	# 1	00:40	Subject Test Refused
Breath Volume	0.000	00:40	
Air Blank	0.000	00:40	
Control Test	0.079	00:40	
Air Blank	0.000	00:41	
Diagnostics Check	OK	00:41	

Data Download Date/Time: 03/28/2017 09:15

Subject Test Electronic Data

DATE/TIME OF TEST: 03/19/2017 02:49:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 06616080A2
BREATH TEST OPERATOR: MCGLYNN,JOHN,L	DRY GAS EXPIRY DATE: 04/05/2018
BREATH TEST OPERATOR AGENCY: FHP Troop E	VIOLATION CODE: DUI
ARRESTING OFFICER: MCGLYNN,JOHN	ARREST TIME: 21:40
ARRESTING OFFICER AGENCY: FHP Troop E	GENDER: Male
LAST AGENCY INSPECTION: 02/27/2017	OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: SAAVEDRA	DATE/TIME OF TEST: 03/19/2017 22:07:02
-------------------------------	--

Diagnostics Check	OK	22:00
Air Blank	0.000	22:01
Control Test	0.080	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.165	22:02
Breath Volume	2.332	22:02
Air Blank	0.000	22:03
Air Blank	0.000	22:04
Subject Sample # 2	0.156	22:05
Breath Volume	2.195	22:05
Air Blank	0.000	22:06
Control Test	0.079	22:06
Air Blank	0.000	22:06
Diagnostics Check	OK	22:07

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCGLYNN,JOHN,L

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MADDOX,KOESHA

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 02:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ASLAM

DATE/TIME OF TEST: 03/20/2017 03:12:22

Diagnostics Check	OK	03:09	
Air Blank	0.000	03:10	
Control Test	0.081	03:10	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:11	Subject Test Refused
Breath Volume	0.000	03:11	
Air Blank		03:11	RFI Detect
Air Blank	0.000	03:12	

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCGLYNN,JOHN,L

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MADDOX,KOESHA

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 02:41

GENDER: Female

OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: ASLAM

DATE/TIME OF TEST: 03/20/2017 03:18:52

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:16	
Control Test	0.081	03:16	
Air Blank	0.000	03:17	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.080	03:18	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 0661608A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: PEREZ BEATON

DATE/TIME OF TEST: 03/21/2017 03:39:54

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.164	03:35
Breath Volume	1.984	03:35
Air Blank	0.000	03:36
Air Blank	0.000	03:37
Subject Sample # 2	0.175	03:38
Breath Volume	2.753	03:38
Air Blank	0.000	03:38
Control Test	0.078	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCGLYNN,JOHN,L

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MCGLYNN,JOHN

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:40

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 03/23/2017 21:05:30

Diagnostics Check	OK	21:02	
Air Blank	0.000	21:03	
Control Test	0.080	21:03	
Air Blank	0.000	21:03	
Subject Sample	# 1	21:04	Subject Test Refused
Breath Volume	0.000	21:04	
Air Blank	0.000	21:04	
Control Test	0.079	21:04	
Air Blank	0.000	21:05	
Diagnostics Check	OK	21:05	

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: MCGLYNN,JOHN,L

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: MCGLYNN,JOHN

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:13

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SALOMON

DATE/TIME OF TEST: 03/24/2017 03:44:35

Diagnostics Check	OK	03:33	
Air Blank	0.000	03:34	
Control Test	0.079	03:34	
Air Blank	0.000	03:35	
Subject Sample	# 1 0.169	03:38	Volume Not Met
Breath Volume	0.402	03:38	
Air Blank	0.000	03:38	
Air Blank	0.000	03:40	
Subject Sample	# 2	03:42	Subject Test Refused
Breath Volume	0.390	03:42	
Air Blank	0.000	03:43	
Control Test	0.079	03:43	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 06616080A2
BREATH TEST OPERATOR: MCGLYNN,JOHN,L	DRY GAS EXPIRY DATE: 04/05/2018
BREATH TEST OPERATOR AGENCY: FHP Troop E	VIOLATION CODE: DUI
ARRESTING OFFICER: MCGLYNN,JOHN	ARREST TIME: 21:14
ARRESTING OFFICER AGENCY: FHP Troop E	GENDER: Female
LAST AGENCY INSPECTION: 02/27/2017	OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: PAIZ	DATE/TIME OF TEST: 03/25/2017 21:45:00
---------------------------	--

Diagnostics Check	OK	21:36	
Air Blank	0.000	21:37	
Control Test	0.079	21:37	
Air Blank	0.000	21:37	
Subject Sample	# 1	21:41	No Sample Provided
Breath Volume	0.000	21:41	
Air Blank	0.000	21:41	
Air Blank	0.000	21:43	
Subject Sample	# 2	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:44	
Control Test	0.080	21:44	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:45	

Data Download Date/Time: 03/28/2017 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ROMAN CORTEZ

DATE/TIME OF TEST: 03/02/2017 02:38:17

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.205	02:33
Breath Volume	2.316	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.201	02:36
Breath Volume	2.257	02:36
Air Blank	0.000	02:37
Control Test	0.077	02:37
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 03/30/2017 02:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PEREZ ACEVEDO

DATE/TIME OF TEST: 03/04/2017 01:34:47

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.082	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.149	01:30
Breath Volume	3.164	01:30
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.149	01:33
Breath Volume	2.632	01:33
Air Blank	0.000	01:33
Control Test	0.081	01:34
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 03/30/2017 02:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAWKINS,DEBORAH,

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: HAWKINS,DEBORAH

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 09216080A3

DRY GAS EXPIRY DATE: 05/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: PACHECO HERNANDEZ

DATE/TIME OF TEST: 03/12/2017 03:50:32

Diagnostics Check	OK	03:43
Air Blank	0.000	03:43
Control Test	0.080	03:44
Air Blank	0.000	03:44
Subject Sample # 1	0.230	03:46
Breath Volume	2.515	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.230	03:48
Breath Volume	2.109	03:48
Air Blank	0.000	03:49
Control Test	0.079	03:49
Air Blank	0.000	03:50
Diagnostics Check	OK	03:50

Data Download Date/Time: 03/30/2017 02:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAWKINS,DEBORAH,

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: HAWKINS,DEBORAH

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 09216080A3

DRY GAS EXPIRY DATE: 05/05/2018

VIOLATION CODE: DUI

ARREST TIME: 01:01

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 03/16/2017 01:35:19

Diagnostics Check	OK	01:28
Air Blank	0.000	01:28
Control Test	0.081	01:29
Air Blank	0.000	01:29
Subject Sample # 1	0.169	01:30
Breath Volume	2.046	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:32
Subject Sample # 2	0.165	01:33
Breath Volume	1.773	01:33
Air Blank	0.000	01:34
Control Test	0.082	01:34
Air Blank	0.000	01:35
Diagnostics Check	OK	01:35

Data Download Date/Time: 03/30/2017 02:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAWKINS,DEBORAH,

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: HAWKINS,DEBORAH

ARRESTING OFFICER AGENCY: FHP Troop D

LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 09216080A3

DRY GAS EXPIRY DATE: 05/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:18

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: RICCIARDI

DATE/TIME OF TEST: 03/18/2017 03:49:59

Diagnostics Check	OK	03:49	
Air Blank		03:49	RFI Detect
Air Blank	0.000	03:49	

Data Download Date/Time: 03/30/2017 02:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006629

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAWKINS,DEBORAH,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: RICCIARDI

DATE/TIME OF TEST: 03/18/2017 03:58:18

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.081	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.128	03:53
Breath Volume	2.757	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:55
Subject Sample # 2	0.120	03:56
Breath Volume	2.523	03:56
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 03/30/2017 02:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: JIMENEZ CALDERON

DATE/TIME OF TEST: 03/04/2017 02:47:23

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Subject Sample # 1	0.123	02:42
Breath Volume	1.277	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.125	02:45
Breath Volume	1.832	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: LASSO SILVA,NICOLAS

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 02:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MESTRE DREKE

DATE/TIME OF TEST: 03/04/2017 03:49:09

Diagnostics Check	OK	03:46	
Air Blank	0.000	03:46	
Control Test	0.081	03:47	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:48	
Control Test	0.079	03:48	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: AMAYA

DATE/TIME OF TEST: 03/04/2017 22:30:48

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.081	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.174	22:25
Breath Volume	1.847	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.159	22:29
Breath Volume	2.437	22:29
Air Blank	0.000	22:29
Control Test	0.081	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: LASSO SILVA,NICOLAS

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:26

SUBJECT'S LAST NAME: CESPED

DATE/TIME OF TEST: 03/06/2017 01:07:34

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.081	00:54	
Air Blank	0.000	00:55	
Subject Sample	# 1 0.143	00:58	Volume Not Met
Breath Volume	0.347	00:58	
Air Blank	0.000	00:59	
Air Blank	0.000	01:01	
Subject Sample	# 2 0.162	01:01	
Breath Volume	2.152	01:01	
Air Blank	0.000	01:02	
Air Blank	0.000	01:04	
Subject Sample	# 3 0.154	01:05	
Breath Volume	2.363	01:05	
Air Blank	0.000	01:06	
Control Test	0.080	01:06	
Air Blank	0.000	01:07	
Diagnostics Check	OK	01:07	

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: LEAL CAMEJO

DATE/TIME OF TEST: 03/12/2017 04:04:07

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.067	03:59
Breath Volume	3.558	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.068	04:02
Breath Volume	3.558	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 03/13/2017 01:26:59

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1 0.166	01:15	
Breath Volume	1.546	01:15	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample	# 2 0.164	01:20	Volume Not Met
Breath Volume	0.476	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 3 0.153	01:25	
Breath Volume	1.253	01:25	
Air Blank	0.000	01:26	
Control Test	0.078	01:26	
Air Blank	0.000	01:26	
Diagnostics Check	OK	01:26	

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 02:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: SOSA ESCOBAR

DATE/TIME OF TEST: 03/15/2017 02:41:35

Diagnostics Check	OK	02:38	
Air Blank	0.000	02:39	
Control Test	0.081	02:39	
Air Blank	0.000	02:40	
Subject Sample	# 1	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Control Test	0.080	02:40	
Air Blank	0.000	02:41	
Diagnostics Check	OK	02:41	

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 21:24

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 03/19/2017 22:13:59

Diagnostics Check	OK	22:10	
Air Blank	0.000	22:11	
Control Test	0.080	22:11	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:12	Subject Test Refused
Breath Volume	0.000	22:12	
Air Blank	0.000	22:13	
Control Test	0.079	22:13	
Air Blank	0.000	22:13	
Diagnostics Check	OK	22:13	

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:59

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: FONSECA

DATE/TIME OF TEST: 03/22/2017 01:30:18

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:28	
Air Blank	0.000	01:28	
Subject Sample	# 1	01:28	Subject Test Refused
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	
Control Test	0.080	01:29	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:30	

Data Download Date/Time: 03/28/2017 09:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 03/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 03/31/2017 00:44:31

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.081	00:42	
Air Blank	0.000	00:42	
Subject Sample	# 1	00:43	Subject Test Refused
Breath Volume	0.000	00:43	
Air Blank	0.000	00:43	
Control Test	0.080	00:43	
Air Blank	0.000	00:44	
Diagnostics Check	OK	00:44	

Data Download Date/Time: 04/30/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LASSO SILVA,NICOLAS,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: LASSO SILVA,NICOLAS

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 03/28/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 06616080A2

DRY GAS EXPIRY DATE: 04/05/2018

VIOLATION CODE: DUI

ARREST TIME: 23:50

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 03/31/2017 00:52:44

Diagnostics Check	OK	00:49	
Air Blank	0.000	00:50	
Control Test	0.080	00:50	
Air Blank	0.000	00:51	
Subject Sample	# 1	00:51	Subject Test Refused
Breath Volume	0.000	00:51	
Air Blank	0.000	00:51	
Control Test	0.079	00:52	
Air Blank	0.000	00:52	
Diagnostics Check	OK	00:52	

Data Download Date/Time: 04/30/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS, ADAM, M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MORRIS, ADAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: HOFMANN

DATE/TIME OF TEST: 03/03/2017 00:11:15

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.000	00:06
Breath Volume	2.328	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.000	00:09
Breath Volume	2.101	00:09
Air Blank	0.000	00:10
Control Test	0.078	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 03/04/2017 00:55:05

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:49	
Control Test	0.082	00:49	
Air Blank	0.000	00:50	
Subject Sample	# 1 0.199	00:50	
Breath Volume	2.199	00:50	
Air Blank	0.000	00:51	
Air Blank	0.000	00:52	
Subject Sample	# 2 0.206	00:53	
Breath Volume	2.582	00:53	
Air Blank		00:54	Purge Fail
Air Blank	0.000	00:55	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 03/04/2017 01:05:00

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:03	
Control Test	0.085	01:04	Interferent Detect
Air Blank		01:05	Purge Fail

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MOHAMED

DATE/TIME OF TEST: 03/04/2017 01:14:37

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.083	01:08	
Air Blank	0.000	01:08	
Subject Sample # 1	0.193	01:09	
Breath Volume	1.765	01:09	
Air Blank	0.000	01:09	
Air Blank	0.000	01:11	
Subject Sample # 2	0.188	01:12	
Breath Volume	1.644	01:12	
Air Blank	0.000	01:13	
Control Test	0.082	01:13	
Air Blank	0.000	01:14	
Diagnostics Check	Fail	01:14	Temperature Regulation Test
Air Blank	0.000	01:14	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/05/2017 02:18:55

Diagnostics Check	OK	02:11	
Air Blank	0.000	02:12	
Control Test	0.079	02:12	
Air Blank	0.000	02:13	
Subject Sample	# 1 0.135	02:14	
Breath Volume	2.000	02:14	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2	02:17	Interferent Detect
Breath Volume	3.257	02:17	
Air Blank	0.000	02:17	
Control Test	0.082	02:18	
Air Blank	0.000	02:18	
Diagnostics Check	OK	02:18	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/05/2017 02:22:09

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:21	
Control Test	0.084	02:21	Interferent Detect
Air Blank	0.000	02:22	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 03/05/2017 02:30:15

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:24
Subject Sample # 1	0.128	02:25
Breath Volume	4.042	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.119	02:28
Breath Volume	2.949	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 12:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PATNELLA

DATE/TIME OF TEST: 03/09/2017 00:43:30

Diagnostics Check	OK	00:31	
Air Blank	0.000	00:31	
Control Test	0.080	00:32	
Air Blank	0.000	00:32	
Subject Sample	# 1 0.138	00:33	
Breath Volume	1.304	00:33	
Air Blank	0.000	00:33	
Air Blank	0.000	00:35	
Subject Sample	# 2 0.117	00:38	Volume Not Met
Breath Volume	0.328	00:38	
Air Blank	0.000	00:39	
Air Blank	0.000	00:41	
Subject Sample	# 3 0.137	00:41	
Breath Volume	2.601	00:41	
Air Blank	0.000	00:42	
Control Test	0.079	00:42	
Air Blank	0.000	00:43	
Diagnostics Check	OK	00:43	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WEINANDT

DATE/TIME OF TEST: 03/10/2017 22:07:07

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.084	22:06	Interferent Detect
Air Blank	0.000	22:07	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: WEINANDT

DATE/TIME OF TEST: 03/10/2017 22:14:19

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.081	22:09
Air Blank	0.000	22:09
Subject Sample # 1	0.000	22:10
Breath Volume	3.863	22:10
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.000	22:12
Breath Volume	4.011	22:12
Air Blank	0.000	22:13
Control Test	0.081	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CHAMBERS

DATE/TIME OF TEST: 03/11/2017 22:52:41

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.081	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1 0.143	22:42	
Breath Volume	1.425	22:42	
Air Blank	0.000	22:43	
Air Blank	0.000	22:44	
Subject Sample	# 2 0.131	22:47	Volume Not Met
Breath Volume	0.078	22:47	
Air Blank	0.000	22:48	
Air Blank	0.000	22:50	
Subject Sample	# 3 0.137	22:51	
Breath Volume	1.320	22:51	
Air Blank	0.000	22:51	
Control Test	0.076	22:52	
Air Blank	0.000	22:52	
Diagnostics Check	OK	22:52	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 03/12/2017 03:21:24

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.109	03:14
Breath Volume	1.359	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.106	03:19
Breath Volume	1.218	03:19
Air Blank	0.000	03:20
Control Test	0.080	03:20
Air Blank	0.000	03:21
Diagnostics Check	OK	03:21

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: MCLELLAN

DATE/TIME OF TEST: 03/17/2017 22:40:09

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:36	
Control Test	0.083	22:37	
Air Blank	0.000	22:37	
Subject Sample	# 1	22:38	Interferent Detect
Breath Volume	2.753	22:38	
Air Blank	0.000	22:39	
Control Test	0.080	22:39	
Air Blank	0.000	22:40	
Diagnostics Check	OK	22:40	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MCLELLAN

DATE/TIME OF TEST: 03/17/2017 22:45:49

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.081	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:44	Interferent Detect
Breath Volume	2.910	22:44	
Air Blank	0.000	22:44	
Control Test	0.081	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:45	

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006632

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MCLELLAN

DATE/TIME OF TEST: 03/17/2017 22:55:14

Diagnostics Check	OK	22:48
Air Blank	0.000	22:49
Control Test	0.079	22:49
Air Blank	0.000	22:50
Subject Sample # 1	0.083	22:50
Breath Volume	2.687	22:50
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.082	22:53
Breath Volume	2.250	22:53
Air Blank	0.000	22:54
Control Test	0.081	22:54
Air Blank	0.000	22:55
Diagnostics Check	OK	22:55

Data Download Date/Time: 03/20/2017 21:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 03/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: MALLETT

DATE/TIME OF TEST: 03/31/2017 23:40:45

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.079	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.183	23:35
Breath Volume	2.242	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.184	23:38
Breath Volume	2.300	23:38
Air Blank	0.000	23:39
Control Test	0.080	23:40
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 04/28/2017 22:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: OBERLIN

DATE/TIME OF TEST: 03/08/2017 01:59:12

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:49	
Subject Sample	# 1	01:52	No Sample Provided
Breath Volume	0.000	01:52	
Air Blank	0.000	01:52	
Air Blank	0.000	01:54	
Subject Sample	# 2	01:57	No Sample Provided
Breath Volume	0.000	01:57	
Air Blank	0.000	01:58	
Control Test	0.082	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 03/22/2017 02:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: BELL

DATE/TIME OF TEST: 03/11/2017 05:31:44

Diagnostics Check	OK	05:24
Air Blank	0.000	05:25
Control Test	0.080	05:25
Air Blank	0.000	05:25
Subject Sample # 1	0.144	05:26
Breath Volume	2.074	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:28
Subject Sample # 2	0.135	05:30
Breath Volume	1.175	05:30
Air Blank	0.000	05:30
Control Test	0.081	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

Data Download Date/Time: 03/22/2017 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WENZEL, THOMAS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: BARRAZA

DATE/TIME OF TEST: 03/15/2017 00:51:30

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:45
Air Blank	0.000	00:46
Subject Sample # 1	0.119	00:47
Breath Volume	3.468	00:47
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.113	00:49
Breath Volume	3.417	00:49
Air Blank	0.000	00:50
Control Test	0.081	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 03/22/2017 02:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D

BREATH TEST OPERATOR AGENCY: FHP Troop D

ARRESTING OFFICER: SCHEINER, JAMES

ARRESTING OFFICER AGENCY: Cocoa Beach PD

LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 09216080A3

DRY GAS EXPIRY DATE: 05/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:24

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 03/18/2017 00:49:27

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.077	00:48	
Air Blank		00:48	Ambient Fail
Air Blank	0.000	00:49	

Data Download Date/Time: 03/22/2017 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 03/19/2017 05:35:54

Diagnostics Check	OK	05:34	
Air Blank		05:35	Ambient Fail
Air Blank	0.000	05:35	

Data Download Date/Time: 03/22/2017 02:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: HICKMAN

DATE/TIME OF TEST: 03/19/2017 05:42:37

Diagnostics Check	OK	05:41	
Air Blank		05:41	Ambient Fail
Air Blank	0.000	05:42	

Data Download Date/Time: 03/22/2017 02:38

Subject Test Electronic Data

DATE/TIME OF TEST: 03/07/2017 20:32:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 02/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 03/20/2017 00:22:35

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.078	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.000	00:18
Breath Volume	2.738	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:21
Breath Volume	2.679	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 03/27/2017 10:01