

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: JENNINGS

DATE/TIME OF TEST: 04/04/2017 23:21:38

Diagnostics Check	OK	23:18	
Air Blank	0.000	23:19	
Control Test	0.082	23:19	
Air Blank	0.000	23:20	
Subject Sample	# 1	23:20	Subject Test Refused
Breath Volume	0.000	23:20	
Air Blank	0.000	23:20	
Control Test	0.082	23:21	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:21	

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: RIDER

DATE/TIME OF TEST: 04/08/2017 21:24:54

Diagnostics Check	OK	21:23	
Air Blank	0.000	21:23	
Control Test	0.080	21:24	
Air Blank		21:24	RFI Detect
Air Blank		21:24	RFI Detect

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HILL,DANIEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: RIDER

DATE/TIME OF TEST: 04/08/2017 21:29:34

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:27	
Control Test	0.081	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:28	
Control Test	0.080	21:28	
Air Blank	0.000	21:29	
Diagnostics Check	OK	21:29	

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: KONSTANTINIDIS

DATE/TIME OF TEST: 04/13/2017 05:19:00

Diagnostics Check		OK	05:12
Air Blank		0.000	05:12
Control Test		0.081	05:13
Air Blank		0.000	05:13
Subject Sample	# 1	0.180	05:14
Breath Volume		4.230	05:14
Air Blank		0.000	05:15
Air Blank		0.000	05:16
Subject Sample	# 2	0.189	05:17
Breath Volume		1.320	05:17
Air Blank		0.000	05:18
Control Test		0.081	05:18
Air Blank		0.000	05:18
Diagnostics Check		OK	05:19

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: TENAGLIA

DATE/TIME OF TEST: 04/15/2017 20:42:54

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.081	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.165	20:38
Breath Volume	2.937	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.164	20:41
Breath Volume	2.347	20:41
Air Blank	0.000	20:41
Control Test	0.081	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEDETTI,MARIO,E
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BENEDETTI,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 05:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: BOURGEOIS

DATE/TIME OF TEST: 04/17/2017 06:10:38

Diagnostics Check	OK	06:04
Air Blank	0.000	06:04
Control Test	0.080	06:05
Air Blank	0.000	06:05
Subject Sample # 1	0.216	06:06
Breath Volume	2.593	06:06
Air Blank	0.000	06:06
Air Blank	0.000	06:08
Subject Sample # 2	0.216	06:09
Breath Volume	2.828	06:09
Air Blank	0.000	06:09
Control Test	0.079	06:10
Air Blank	0.000	06:10
Diagnostics Check	OK	06:10

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:07

SUBJECT'S LAST NAME: LYNE 2

DATE/TIME OF TEST: 04/28/2017 15:52:53

Diagnostics Check	OK	15:46
Air Blank	0.000	15:46
Control Test	0.082	15:47
Air Blank	0.000	15:47
Subject Sample # 1	0.388	15:48
Breath Volume	2.093	15:48
Air Blank	0.000	15:48
Air Blank	0.000	15:50
Subject Sample # 2	0.399	15:51
Breath Volume	1.691	15:51
Air Blank	0.000	15:51
Control Test	0.080	15:52
Air Blank	0.000	15:52
Diagnostics Check	OK	15:52

Data Download Date/Time: 04/29/2017 09:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LANDRY

DATE/TIME OF TEST: 04/01/2017 01:21:25

Diagnostics Check		OK	01:15
Air Blank		0.000	01:15
Control Test		0.082	01:15
Air Blank		0.000	01:16
Subject Sample	# 1	0.212	01:16
Breath Volume		2.550	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.209	01:19
Breath Volume		2.972	01:19
Air Blank		0.000	01:20
Control Test		0.079	01:20
Air Blank		0.000	01:21
Diagnostics Check		OK	01:21

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEOHANE,KUNIKO,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:01

SUBJECT'S LAST NAME: PALMERO

DATE/TIME OF TEST: 04/27/2017 13:32:14

Diagnostics Check	OK	13:29	
Air Blank	0.000	13:29	
Control Test	0.082	13:30	
Air Blank	0.000	13:30	
Subject Sample	# 1	13:30	Subject Test Refused
Breath Volume	0.000	13:30	
Air Blank	0.000	13:31	
Control Test	0.081	13:31	
Air Blank	0.000	13:32	
Diagnostics Check	OK	13:32	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: DERY,D
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: FREYDER

DATE/TIME OF TEST: 04/09/2017 03:37:34

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.082	03:35	
Air Blank	0.000	03:35	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.081	03:36	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CONNELLY

DATE/TIME OF TEST: 04/17/2017 03:19:51

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.082	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1	0.172	03:07	Volume Not Met
Breath Volume		0.085	03:07	
Air Blank		0.000	03:08	
Air Blank		0.000	03:10	
Subject Sample	# 2	0.158	03:12	
Breath Volume		1.273	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 3	0.161	03:18	Volume Not Met
Breath Volume		0.191	03:18	
Air Blank		0.000	03:18	
Control Test		0.077	03:19	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: CONNELLY

DATE/TIME OF TEST: 04/17/2017 03:25:34

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.079	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:24	Subject Test Refused
Breath Volume	0.343	03:24	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MOELLER,JENNA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MATE

DATE/TIME OF TEST: 04/22/2017 00:25:12

Diagnostics Check	OK	00:17
Air Blank	0.000	00:18
Control Test	0.079	00:18
Air Blank	0.000	00:18
Subject Sample # 1	0.338	00:20
Breath Volume	2.289	00:20
Air Blank	0.000	00:21
Air Blank	0.000	00:22
Subject Sample # 2	0.354	00:23
Breath Volume	2.511	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: CARIUS

DATE/TIME OF TEST: 04/22/2017 00:53:18

Diagnostics Check	OK	00:41	
Air Blank	0.000	00:42	
Control Test	0.080	00:42	
Air Blank	0.000	00:43	
Subject Sample	# 1	00:46	No Sample Provided
Breath Volume	0.000	00:46	
Air Blank	0.000	00:46	
Air Blank	0.000	00:48	
Subject Sample	# 2	00:51	No Sample Provided
Breath Volume	0.000	00:51	
Air Blank	0.000	00:52	
Control Test	0.080	00:52	
Air Blank	0.000	00:53	
Diagnostics Check	OK	00:53	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BOUVIER,KRISTOPHER
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: GUNDERSEN

DATE/TIME OF TEST: 04/22/2017 22:02:05

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:52	
Control Test		0.082	21:53	
Air Blank		0.000	21:53	
Subject Sample	# 1	0.377	21:54	No .020 agreement
Breath Volume		2.585	21:54	
Air Blank		0.000	21:55	
Air Blank		0.000	21:57	
Subject Sample	# 2	0.353	21:57	No .020 agreement
Breath Volume		1.820	21:57	
Air Blank		0.000	21:58	
Air Blank		0.000	22:00	
Subject Sample	# 3	0.323	22:00	No .020 agreement
Breath Volume		1.207	22:00	
Air Blank		0.000	22:01	
Control Test		0.081	22:01	
Air Blank		0.000	22:02	
Diagnostics Check		OK	22:02	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,AUSTIN,S
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BOUVIER,KRISTOPHER
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 21:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: GUNDERSEN

DATE/TIME OF TEST: 04/22/2017 22:28:37

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.081	22:23
Air Blank	0.000	22:23
Subject Sample # 1	0.329	22:24
Breath Volume	1.367	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.329	22:27
Breath Volume	1.804	22:27
Air Blank	0.000	22:27
Control Test	0.081	22:28
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: RODRIGUEZ,ALEX
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 03/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: CASTLE

DATE/TIME OF TEST: 04/25/2017 01:16:31

Diagnostics Check	OK	01:05	
Air Blank	0.000	01:05	
Control Test	0.081	01:05	
Air Blank	0.000	01:06	
Subject Sample	# 1 0.177	01:09	Volume Not Met
Breath Volume	0.546	01:09	
Air Blank	0.000	01:10	
Air Blank	0.000	01:12	
Subject Sample	# 2	01:15	No Sample Provided
Breath Volume	0.000	01:15	
Air Blank	0.000	01:15	
Control Test	0.082	01:15	
Air Blank	0.000	01:16	
Diagnostics Check	OK	01:16	

Data Download Date/Time: 04/27/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: GORDON,JOSHUA
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 15:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 04/29/2017 16:54:52

Diagnostics Check	OK	16:48
Air Blank	0.000	16:49
Control Test	0.081	16:49
Air Blank	0.000	16:49
Subject Sample # 1	0.000	16:50
Breath Volume	2.992	16:50
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 2	0.000	16:53
Breath Volume	2.515	16:53
Air Blank	0.000	16:53
Control Test	0.081	16:54
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 05/04/2017 13:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETZ,FRANK,L
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: FRICKE,KEN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 04/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14968
DRY GAS EXPIRY DATE: 08/06/2018
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: ABOY

DATE/TIME OF TEST: 04/30/2017 17:45:29

Diagnostics Check	OK	17:42	
Air Blank	0.000	17:43	
Control Test	0.081	17:43	
Air Blank	0.000	17:43	
Subject Sample	# 1	17:44	Subject Test Refused
Breath Volume	0.000	17:44	
Air Blank	0.000	17:44	
Control Test	0.082	17:44	
Air Blank	0.000	17:45	
Diagnostics Check	OK	17:45	

Data Download Date/Time: 05/04/2017 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOANELLE,V
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JOANELLE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 09:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: TERAN

DATE/TIME OF TEST: 04/02/2017 10:11:59

Diagnostics Check	OK	10:05
Air Blank	0.000	10:05
Control Test	0.078	10:05
Air Blank	0.000	10:06
Subject Sample # 1	0.138	10:07
Breath Volume	1.992	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:09
Subject Sample # 2	0.147	10:10
Breath Volume	2.023	10:10
Air Blank	0.000	10:11
Control Test	0.078	10:11
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

Data Download Date/Time: 04/26/2017 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,MERCEDES,S
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERRERA,JUAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: FUNEZ

DATE/TIME OF TEST: 04/06/2017 03:32:42

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.078	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1 0.201	03:28	Volume Not Met
Breath Volume	0.394	03:28	
Air Blank	0.000	03:29	
Air Blank	0.000	03:30	
Subject Sample	# 2	03:31	Subject Test Refused
Breath Volume	0.410	03:31	
Air Blank	0.000	03:31	
Control Test	0.076	03:32	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 04/26/2017 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000872

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COSS,ATILA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERNANDEZ,JASIEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1789170
DRY GAS EXPIRY DATE: 01/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TAMAYO BENEDETTI

DATE/TIME OF TEST: 04/13/2017 03:57:13

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.078	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.231	03:52
Breath Volume	2.832	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.226	03:55
Breath Volume	2.308	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

Data Download Date/Time: 04/26/2017 19:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/07/2017 19:48:19

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.080	19:44	
Air Blank	0.000	19:45	
Subject Sample	# 1 0.078	19:46	
Breath Volume	1.933	19:46	
Air Blank		19:47	Purge Fail
Air Blank	0.000	19:48	

Data Download Date/Time: 04/26/2017 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PASTOR,JOSE,F
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 04/07/2017 20:00:56

Diagnostics Check	OK	19:54
Air Blank	0.000	19:55
Control Test	0.080	19:55
Air Blank	0.000	19:56
Subject Sample # 1	0.077	19:56
Breath Volume	1.882	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:58
Subject Sample # 2	0.071	19:59
Breath Volume	1.308	19:59
Air Blank	0.000	20:00
Control Test	0.081	20:00
Air Blank	0.000	20:00
Diagnostics Check	OK	20:00

Data Download Date/Time: 04/26/2017 19:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ALEMAN,ANTONIO,

BREATH TEST OPERATOR AGENCY: Coral Gables PD

ARRESTING OFFICER: RIOS,JOEL

ARRESTING OFFICER AGENCY: Coral Gables PD

LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 203498

DRY GAS EXPIRY DATE: 12/10/2018

VIOLATION CODE: DUI

ARREST TIME: 20:36

GENDER: Male

OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/26/2017 22:11:24

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.078	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:09	Interferent Detect
Breath Volume	1.218	22:09	
Air Blank		22:10	Purge Fail
Air Blank		22:11	Purge Fail

Data Download Date/Time: 05/30/2017 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000873

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALEMAN,ANTONIO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: RIOS,JOEL
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 04/26/2017 22:20:29

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.079	22:14
Air Blank	0.000	22:14
Subject Sample # 1	0.048	22:15
Breath Volume	1.945	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:17
Subject Sample # 2	0.043	22:18
Breath Volume	1.183	22:18
Air Blank	0.000	22:19
Control Test	0.078	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

Data Download Date/Time: 05/30/2017 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YASUDA,SHOHEI,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YASUDA,SHOHEI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: PRIDGEON

DATE/TIME OF TEST: 04/22/2017 00:04:25

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.089	23:58
Breath Volume	1.117	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.085	00:02
Breath Volume	1.320	00:02
Air Blank	0.000	00:03
Control Test	0.080	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 05/15/2017 09:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 05:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: EVERHARD

DATE/TIME OF TEST: 04/29/2017 06:31:43

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.082	06:26
Air Blank	0.000	06:26
Subject Sample # 1	0.062	06:27
Breath Volume	2.539	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:29
Subject Sample # 2	0.051	06:30
Breath Volume	2.625	06:30
Air Blank	0.000	06:30
Control Test	0.081	06:31
Air Blank	0.000	06:31
Diagnostics Check	OK	06:31

Data Download Date/Time: 05/31/2017 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAIGE,LEON,I
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: PAIGE,LEON
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 14:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:00

SUBJECT'S LAST NAME: MUNOZ

DATE/TIME OF TEST: 04/30/2017 14:30:13

Diagnostics Check	OK	14:25	
Air Blank	0.000	14:25	
Control Test	0.080	14:25	
Air Blank	0.000	14:26	
Subject Sample	# 1	14:28	Subject Test Refused
Breath Volume	0.000	14:28	
Air Blank	0.000	14:29	
Control Test	0.081	14:29	
Air Blank	0.000	14:30	
Diagnostics Check	OK	14:30	

Data Download Date/Time: 05/31/2017 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000879

INSTRUMENT REGISTERED TO: MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAZONOV,VLADISLAV,
BREATH TEST OPERATOR AGENCY: Miami Beach PD
ARRESTING OFFICER: OCEJO,RICHARD
ARRESTING OFFICER AGENCY: Miami Beach PD
LAST AGENCY INSPECTION: 04/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1826916
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: PRATER

DATE/TIME OF TEST: 04/30/2017 16:23:01

Diagnostics Check	OK	16:16
Air Blank	0.000	16:16
Control Test	0.081	16:17
Air Blank	0.000	16:17
Subject Sample # 1	0.164	16:18
Breath Volume	2.953	16:18
Air Blank	0.000	16:19
Air Blank	0.000	16:21
Subject Sample # 2	0.173	16:21
Breath Volume	3.394	16:21
Air Blank	0.000	16:22
Control Test	0.080	16:22
Air Blank	0.000	16:22
Diagnostics Check	OK	16:23

Data Download Date/Time: 05/31/2017 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRAVIS,CHARLES,B
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: TRAVIS,CHARLES
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: DUTRA DAS NEVES

DATE/TIME OF TEST: 04/08/2017 09:06:32

Diagnostics Check	OK	09:03	
Air Blank	0.000	09:04	
Control Test	0.079	09:04	
Air Blank	0.000	09:05	
Subject Sample	# 1	09:05	Subject Test Refused
Breath Volume	0.000	09:05	
Air Blank	0.000	09:05	
Control Test	0.079	09:05	
Air Blank	0.000	09:06	
Diagnostics Check	OK	09:06	

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:50

SUBJECT'S LAST NAME: KONDLA

DATE/TIME OF TEST: 04/15/2017 14:16:37

Diagnostics Check	OK	14:10
Air Blank	0.000	14:11
Control Test	0.080	14:11
Air Blank	0.000	14:11
Subject Sample # 1	0.140	14:12
Breath Volume	1.464	14:12
Air Blank	0.000	14:12
Air Blank	0.000	14:14
Subject Sample # 2	0.138	14:15
Breath Volume	1.835	14:15
Air Blank	0.000	14:15
Control Test	0.080	14:16
Air Blank	0.000	14:16
Diagnostics Check	OK	14:16

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERALTA,CHRISTIAN,A
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: PERALTA,CHRISTIAN
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: LECHTER

DATE/TIME OF TEST: 04/17/2017 18:39:21

Diagnostics Check	OK	18:36	
Air Blank	0.000	18:36	
Control Test	0.076	18:37	
Air Blank	0.000	18:37	
Subject Sample	# 1	18:38	Subject Test Refused
Breath Volume	0.000	18:38	
Air Blank	0.000	18:38	
Control Test	0.076	18:38	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:39	

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAPIRO,SABINA,M
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BUCHANAN,MATTHEW
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: MOSTOVOY

DATE/TIME OF TEST: 04/20/2017 20:19:48

Diagnostics Check	OK	20:13	
Air Blank	0.000	20:14	
Control Test	0.078	20:14	
Air Blank	0.000	20:15	
Subject Sample	# 1 0.196	20:18	Volume Not Met
Breath Volume	0.050	20:18	
Air Blank		20:19	Purge Fail
Air Blank		20:19	Purge Fail

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGINS,LYLE,P
BREATH TEST OPERATOR AGENCY: Sunny Isles Beach PD
ARRESTING OFFICER: AGINS,LYLE
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:39

SUBJECT'S LAST NAME: MENALIEV

DATE/TIME OF TEST: 04/25/2017 09:05:47

Diagnostics Check	OK	08:59
Air Blank	0.000	09:00
Control Test	0.080	09:00
Air Blank	0.000	09:00
Subject Sample # 1	0.105	09:01
Breath Volume	3.820	09:01
Air Blank	0.000	09:01
Air Blank	0.000	09:03
Subject Sample # 2	0.099	09:04
Breath Volume	2.367	09:04
Air Blank	0.000	09:04
Control Test	0.079	09:05
Air Blank	0.000	09:05
Diagnostics Check	OK	09:05

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OGDEN,JOHN,L
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: RODRIGUEZ,KATINA
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: WOLF

DATE/TIME OF TEST: 04/28/2017 01:53:17

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:43	
Control Test	0.078	01:43	
Air Blank	0.000	01:44	
Subject Sample	# 1 0.256	01:45	
Breath Volume	2.148	01:45	
Air Blank	0.000	01:46	
Air Blank	0.000	01:48	
Subject Sample	# 2 0.232	01:49	
Breath Volume	3.156	01:49	
Air Blank	0.000	01:49	
Air Blank	0.000	01:51	
Subject Sample	# 3 0.197	01:52	
Breath Volume	2.949	01:52	
Air Blank		01:52	RFI Detect
Air Blank	0.000	01:53	

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KUNCAS,ANDREW,J
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: KUNCAS,ANDREW
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: NIKOLAIEVA

DATE/TIME OF TEST: 04/28/2017 21:58:49

Diagnostics Check	OK	21:55	
Air Blank	0.000	21:55	
Control Test	0.079	21:56	
Air Blank	0.000	21:56	
Subject Sample	# 1	21:57	Subject Test Refused
Breath Volume	0.000	21:57	
Air Blank	0.000	21:57	
Control Test	0.079	21:58	
Air Blank	0.000	21:58	
Diagnostics Check	OK	21:58	

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAUST, TIFFANY, N
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BARRETT, LAMART
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: FAUST

DATE/TIME OF TEST: 04/28/2017 23:44:21

Diagnostics Check	OK	23:39	
Air Blank	0.000	23:40	
Control Test	0.079	23:40	
Air Blank	0.000	23:41	
Subject Sample	# 1	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Control Test	0.079	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVAS,DAVID,E
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: NAVAS,DAVID
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 04/30/2017 02:07:37

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:56	
Control Test		0.079	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 1	0.144	02:00	Volume Not Met
Breath Volume		0.339	02:00	
Air Blank		0.000	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 2	0.134	02:05	Volume Not Met
Breath Volume		0.074	02:05	
Air Blank		0.000	02:06	
Control Test		0.079	02:07	
Air Blank		0.000	02:07	
Diagnostics Check		OK	02:07	

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVAS,DAVID,E
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: NAVAS,DAVID
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 04/30/2017 02:22:37

Diagnostics Check	OK	02:13
Air Blank	0.000	02:14
Control Test	0.078	02:14
Air Blank	0.000	02:15
Subject Sample # 1	0.148	02:15
Breath Volume	1.750	02:15
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.131	02:20
Breath Volume	1.277	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/16/2017 07:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,STEVEN,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: ADAMS,STEVEN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FRANCOIS

DATE/TIME OF TEST: 04/07/2017 02:52:27

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:49	
Control Test	0.080	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.079	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 05/16/2017 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: FREIRE

DATE/TIME OF TEST: 04/01/2017 04:31:28

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:29	
Control Test	0.078	04:29	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:30	Subject Test Refused
Breath Volume	0.000	04:30	
Air Blank	0.000	04:30	
Control Test	0.079	04:30	
Air Blank	0.000	04:31	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000882

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:36

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 04/01/2017 07:12:55

Diagnostics Check	OK	07:04
Air Blank	0.000	07:05
Control Test	0.079	07:05
Air Blank	0.000	07:06
Subject Sample # 1	0.176	07:07
Breath Volume	1.394	07:07
Air Blank	0.000	07:08
Air Blank	0.000	07:09
Subject Sample # 2	0.165	07:11
Breath Volume	2.035	07:11
Air Blank	0.000	07:11
Control Test	0.077	07:12
Air Blank	0.000	07:12
Diagnostics Check	OK	07:12

Data Download Date/Time: 04/06/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 04/01/2017 02:20:40

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.143	02:16
Breath Volume	1.886	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.145	02:19
Breath Volume	2.464	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLARDO,JORGE,L
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GALLARDO,JORGE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: ROMERO OVIEDO

DATE/TIME OF TEST: 04/01/2017 04:45:17

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.080	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.160	04:40
Breath Volume	2.687	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.157	04:43
Breath Volume	1.753	04:43
Air Blank	0.000	04:44
Control Test	0.079	04:44
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA,JOSE,I
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: HERRERA VIRGUEZ

DATE/TIME OF TEST: 04/01/2017 06:10:47

Diagnostics Check		OK	06:04
Air Blank		0.000	06:04
Control Test		0.080	06:05
Air Blank		0.000	06:05
Subject Sample	# 1	0.252	06:06
Breath Volume		1.226	06:06
Air Blank		0.000	06:06
Air Blank		0.000	06:08
Subject Sample	# 2	0.244	06:09
Breath Volume		1.152	06:09
Air Blank		0.000	06:09
Control Test		0.077	06:10
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: OLIVERO

DATE/TIME OF TEST: 04/06/2017 15:09:51

Diagnostics Check	OK	15:03
Air Blank	0.000	15:04
Control Test	0.081	15:04
Air Blank	0.000	15:05
Subject Sample # 1	0.000	15:05
Breath Volume	3.175	15:05
Air Blank	0.000	15:06
Air Blank	0.000	15:08
Subject Sample # 2	0.000	15:08
Breath Volume	2.480	15:08
Air Blank	0.000	15:08
Control Test	0.080	15:09
Air Blank	0.000	15:09
Diagnostics Check	OK	15:09

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOTO,CARLOS,D
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: SOTO,CARLOS
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: BRUZUAL TRUJILLO

DATE/TIME OF TEST: 04/07/2017 22:43:08

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.081	22:40	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:42	
Control Test	0.079	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TELLEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TELLEZ,ALBERTO
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: GARCIA FIGUEROA

DATE/TIME OF TEST: 04/08/2017 03:34:04

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.206	03:29
Breath Volume	4.140	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:31
Subject Sample # 2	0.205	03:32
Breath Volume	4.066	03:32
Air Blank	0.000	03:33
Control Test	0.077	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:34

Data Download Date/Time: 04/14/2017 12:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: BALKARAN

DATE/TIME OF TEST: 04/11/2017 18:05:17

Diagnostics Check	OK	18:01	
Air Blank	0.000	18:01	
Control Test	0.080	18:02	
Air Blank	0.000	18:02	
Subject Sample	# 1	18:03	Subject Test Refused
Breath Volume	0.000	18:03	
Air Blank	0.000	18:04	
Control Test	0.079	18:04	
Air Blank	0.000	18:05	
Diagnostics Check	OK	18:05	

Data Download Date/Time: 04/14/2017 12:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MONGE GUEVARA

DATE/TIME OF TEST: 04/02/2017 02:41:07

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.079	02:30	
Air Blank		0.000	02:30	
Subject Sample	# 1	0.095	02:33	Volume Not Met
Breath Volume		0.039	02:33	
Air Blank		0.000	02:34	
Air Blank		0.000	02:36	
Subject Sample	# 2	0.108	02:39	Volume Not Met
Breath Volume		0.203	02:39	
Air Blank		0.000	02:40	
Control Test		0.078	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 04/11/2017 13:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: MONGE GUEVARA

DATE/TIME OF TEST: 04/02/2017 02:54:40

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.077	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.103	02:49
Breath Volume	1.453	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.104	02:53
Breath Volume	1.378	02:53
Air Blank	0.000	02:53
Control Test	0.076	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 04/11/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 04/02/2017 04:36:14

Diagnostics Check	OK	04:29
Air Blank	0.000	04:29
Control Test	0.075	04:29
Air Blank	0.000	04:30
Subject Sample # 1	0.139	04:31
Breath Volume	2.269	04:31
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.137	04:34
Breath Volume	1.652	04:34
Air Blank	0.000	04:35
Control Test	0.076	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 04/11/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MADRIGAL

DATE/TIME OF TEST: 04/03/2017 03:31:55

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.079	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.164	03:27
Breath Volume	2.578	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.166	03:30
Breath Volume	1.507	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 04/11/2017 13:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: GARD

DATE/TIME OF TEST: 04/05/2017 04:59:32

Diagnostics Check	OK	04:56	
Air Blank	0.000	04:57	
Control Test	0.079	04:57	
Air Blank	0.000	04:58	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.000	04:58	
Air Blank	0.000	04:58	
Control Test	0.079	04:58	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 04/11/2017 13:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIANCHI,MARIO,D
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: BIANCHI,MARIO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/14/2017 05:42:48

Diagnostics Check	OK	05:39	
Air Blank	0.000	05:40	
Control Test	0.079	05:40	
Air Blank	0.000	05:41	
Subject Sample	# 1	05:41	Subject Test Refused
Breath Volume	0.000	05:41	
Air Blank	0.000	05:41	
Control Test	0.079	05:42	
Air Blank	0.000	05:42	
Diagnostics Check	OK	05:42	

Data Download Date/Time: 05/26/2017 08:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 04/16/2017 01:41:51

Diagnostics Check	OK	01:34
Air Blank	0.000	01:35
Control Test	0.079	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.155	01:37
Breath Volume	1.234	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.156	01:40
Breath Volume	2.652	01:40
Air Blank	0.000	01:40
Control Test	0.079	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 05/26/2017 08:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPANES,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 04/29/2017 00:05:35

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.136	23:59
Breath Volume	1.656	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.131	00:03
Breath Volume	1.671	00:03
Air Blank	0.000	00:04
Control Test	0.079	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 05/26/2017 08:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOO,BRANDON,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: GONZALEZ,ALFREDO
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: RASZINSKI

DATE/TIME OF TEST: 04/08/2017 11:53:02

Diagnostics Check	OK	11:52	
Air Blank		11:52	Ambient Fail
Air Blank		11:53	Improper Sample

Data Download Date/Time: 05/26/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOO,BRANDON,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HOO,BRANDON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: RASZINKI

DATE/TIME OF TEST: 04/08/2017 11:57:01

Diagnostics Check	OK	11:56	
Air Blank		11:56	Ambient Fail
Air Blank	0.000	11:57	

Data Download Date/Time: 05/26/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOO,BRANDON,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HOO,BRANDON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:31

SUBJECT'S LAST NAME: RASZINKI

DATE/TIME OF TEST: 04/08/2017 12:16:20

Diagnostics Check	OK	12:00	
Air Blank	0.000	12:00	
Control Test	0.078	12:00	
Air Blank	0.000	12:01	
Subject Sample	# 1 0.217	12:03	
Breath Volume	3.050	12:03	
Air Blank	0.000	12:04	
Air Blank	0.000	12:06	
Subject Sample	# 2	12:09	No Sample Provided
Breath Volume	0.000	12:09	
Air Blank	0.000	12:09	
Air Blank	0.000	12:11	
Subject Sample	# 3	12:14	No Sample Provided
Breath Volume	0.000	12:14	
Air Blank	0.000	12:15	
Control Test	0.078	12:15	
Air Blank	0.000	12:16	
Diagnostics Check	OK	12:16	

Data Download Date/Time: 05/26/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,EDWIN,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,EDWIN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: NADER

DATE/TIME OF TEST: 04/14/2017 00:17:12

Diagnostics Check	OK	00:13	
Air Blank	0.000	00:13	
Control Test	0.080	00:14	
Air Blank	0.000	00:14	
Subject Sample	# 1	00:15	Subject Test Refused
Breath Volume	0.000	00:15	
Air Blank	0.000	00:16	
Control Test	0.079	00:16	
Air Blank	0.000	00:17	
Diagnostics Check	OK	00:17	

Data Download Date/Time: 05/26/2017 08:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MAURICE, M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HENRY, MAURICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GORDILLO VALENCIA

DATE/TIME OF TEST: 04/27/2017 05:40:15

Diagnostics Check	OK	05:31	
Air Blank	0.000	05:32	
Control Test	0.080	05:32	
Air Blank	0.000	05:33	
Subject Sample # 1	0.136	05:33	
Breath Volume	1.441	05:33	
Air Blank	0.000	05:34	
Air Blank	0.000	05:36	
Subject Sample # 2	0.133	05:38	
Breath Volume	1.746	05:38	
Air Blank		05:39	Purge Fail
Air Blank		05:40	Purge Fail

Data Download Date/Time: 05/26/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY, MAURICE, M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HENRY, MAURICE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: GORDILLO VALENCIA

DATE/TIME OF TEST: 04/27/2017 05:51:14

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.119	05:46
Breath Volume	4.175	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:49
Subject Sample # 2	0.118	05:49
Breath Volume	2.851	05:49
Air Blank	0.000	05:50
Control Test	0.078	05:50
Air Blank	0.000	05:51
Diagnostics Check	OK	05:51

Data Download Date/Time: 05/26/2017 08:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: TANIS,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: WEISBRODT

DATE/TIME OF TEST: 04/11/2017 01:19:28

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:17	
Control Test	0.078	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:18	Subject Test Refused
Breath Volume	0.000	01:18	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:19	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 04/26/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHAPIRO,SABINA,M
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: BUCHANAN,MATTHEW
ARRESTING OFFICER AGENCY: Sunny Isles Beach PD
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: MOSTOVOY

DATE/TIME OF TEST: 04/20/2017 21:00:00

Diagnostics Check	OK	20:55	
Air Blank	0.000	20:55	
Control Test	0.078	20:56	
Air Blank	0.000	20:56	
Subject Sample	# 1	20:58	Subject Test Refused
Breath Volume	0.191	20:58	
Air Blank	0.000	20:59	
Control Test	0.077	20:59	
Air Blank	0.000	20:59	
Diagnostics Check	OK	21:00	

Data Download Date/Time: 04/26/2017 19:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: SIMON,SAMUEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 03/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: IOVU

DATE/TIME OF TEST: 04/26/2017 04:26:11

Diagnostics Check	OK	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:17
Subject Sample # 1	0.129	04:20
Breath Volume	1.570	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.126	04:24
Breath Volume	2.085	04:24
Air Blank	0.000	04:25
Control Test	0.076	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 04/26/2017 19:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: ST VIL

DATE/TIME OF TEST: 04/27/2017 00:23:30

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.078	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.126	00:18
Breath Volume	1.476	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.124	00:21
Breath Volume	1.710	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 05/27/2017 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: DUME

DATE/TIME OF TEST: 04/27/2017 02:54:42

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Subject Sample # 1	0.070	02:49
Breath Volume	1.390	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.072	02:52
Breath Volume	1.718	02:52
Air Blank	0.000	02:53
Control Test	0.076	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 05/27/2017 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DOSTALY,STEVEN
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: YANES

DATE/TIME OF TEST: 04/30/2017 01:47:25

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.077	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.201	01:42
Breath Volume	1.265	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:44
Subject Sample # 2	0.210	01:45
Breath Volume	1.226	01:45
Air Blank	0.000	01:46
Control Test	0.077	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 05/27/2017 19:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 04/03/2017 01:57:46

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.206	01:53
Breath Volume	1.101	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.216	01:56
Breath Volume	2.222	01:56
Air Blank	0.000	01:56
Control Test	0.078	01:57
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 04/24/2017 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/10/2017 23:02:44

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.211	22:58
Breath Volume	2.878	22:58
Air Blank	0.000	22:58
Air Blank	0.000	23:00
Subject Sample # 2	0.208	23:01
Breath Volume	3.269	23:01
Air Blank	0.000	23:01
Control Test	0.080	23:02
Air Blank	0.000	23:02
Diagnostics Check	OK	23:02

Data Download Date/Time: 04/24/2017 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: GIBEL

DATE/TIME OF TEST: 04/12/2017 01:37:48

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.138	01:33
Breath Volume	4.105	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.129	01:36
Breath Volume	3.855	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 04/24/2017 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 04/20/2017 02:26:34

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:20
Air Blank	0.000	02:20
Subject Sample # 1	0.210	02:21
Breath Volume	1.234	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.219	02:25
Breath Volume	1.246	02:25
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 04/24/2017 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 04/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: VIVAS SANTANA

DATE/TIME OF TEST: 04/29/2017 01:50:38

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.181	01:46
Breath Volume	2.878	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.184	01:49
Breath Volume	2.449	01:49
Air Blank	0.000	01:49
Control Test	0.080	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 05/22/2017 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000908

INSTRUMENT REGISTERED TO: PORT ST JOE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIPMAN,KAREN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SHIPMAN,KAREN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FRANZ

DATE/TIME OF TEST: 04/11/2017 00:00:00

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:56	
Control Test	0.081	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	23:58	Subject Test Refused
Breath Volume	0.000	23:58	
Air Blank	0.000	23:59	
Control Test	0.081	23:59	
Air Blank	0.000	23:59	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 04/16/2017 15:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COLES

DATE/TIME OF TEST: 04/01/2017 00:13:39

Diagnostics Check	OK	00:12	
Air Blank	0.000	00:12	
Control Test	0.083	00:13	Interferent Detect
Air Blank	0.000	00:13	

Data Download Date/Time: 04/12/2017 11:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: COLES

DATE/TIME OF TEST: 04/01/2017 00:22:52

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.076	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.081	00:18
Breath Volume	2.234	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.080	00:21
Breath Volume	2.175	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 04/12/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: HAMBRICK

DATE/TIME OF TEST: 04/01/2017 01:42:41

Diagnostics Check	OK	01:34
Air Blank	0.000	01:34
Control Test	0.081	01:35
Air Blank	0.000	01:35
Subject Sample # 1	0.157	01:36
Breath Volume	1.191	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.146	01:41
Breath Volume	2.187	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED, MICHAEL, K
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REED, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MITAS

DATE/TIME OF TEST: 04/01/2017 02:40:51

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.080	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.164	02:33
Breath Volume	1.667	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.175	02:38
Breath Volume	1.574	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAYLOR, MAURICE, E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR, JAMES
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: O'DEA

DATE/TIME OF TEST: 04/02/2017 23:04:21

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:57
Subject Sample # 1	0.144	22:59
Breath Volume	3.000	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.142	23:02
Breath Volume	1.707	23:02
Air Blank	0.000	23:03
Control Test	0.080	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ALGER, MICHAEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: AMAR

DATE/TIME OF TEST: 04/03/2017 02:25:28

Diagnostics Check	OK	02:19
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:20
Breath Volume	1.914	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:24
Breath Volume	1.847	02:24
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: NATAL, J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WARCHOL

DATE/TIME OF TEST: 04/03/2017 02:56:23

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.232	02:48
Breath Volume		1.425	02:48
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.253	02:51
Breath Volume		1.785	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 3	0.261	02:54
Breath Volume		2.339	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNN,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MATTHIS

DATE/TIME OF TEST: 04/03/2017 20:38:58

Diagnostics Check	OK	20:32
Air Blank	0.000	20:33
Control Test	0.080	20:33
Air Blank	0.000	20:34
Subject Sample # 1	0.260	20:34
Breath Volume	1.683	20:34
Air Blank	0.000	20:35
Air Blank	0.000	20:37
Subject Sample # 2	0.258	20:37
Breath Volume	1.546	20:37
Air Blank	0.000	20:38
Control Test	0.079	20:38
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: DUAN

DATE/TIME OF TEST: 04/06/2017 00:43:18

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:36
Air Blank	0.000	00:36
Subject Sample # 1	0.058	00:37
Breath Volume	2.515	00:37
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.065	00:41
Breath Volume	1.175	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FORNAL,DOMINIC
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:34

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 04/06/2017 06:03:35

Diagnostics Check		OK	05:57
Air Blank		0.000	05:57
Control Test		0.080	05:58
Air Blank		0.000	05:58
Subject Sample	# 1	0.129	05:59
Breath Volume		1.667	05:59
Air Blank		0.000	05:59
Air Blank		0.000	06:01
Subject Sample	# 2	0.139	06:01
Breath Volume		2.015	06:01
Air Blank		0.000	06:02
Control Test		0.078	06:02
Air Blank		0.000	06:03
Diagnostics Check		OK	06:03

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH,JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: ANCELMO

DATE/TIME OF TEST: 04/06/2017 19:32:17

Diagnostics Check	OK	19:25
Air Blank	0.000	19:26
Control Test	0.080	19:26
Air Blank	0.000	19:26
Subject Sample # 1	0.212	19:27
Breath Volume	3.171	19:27
Air Blank	0.000	19:28
Air Blank	0.000	19:29
Subject Sample # 2	0.200	19:30
Breath Volume	1.605	19:30
Air Blank	0.000	19:31
Control Test	0.079	19:31
Air Blank	0.000	19:32
Diagnostics Check	OK	19:32

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SMITH, JOSIAH
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: CARBALLO

DATE/TIME OF TEST: 04/06/2017 19:39:58

Diagnostics Check	OK	19:33
Air Blank	0.000	19:33
Control Test	0.079	19:34
Air Blank	0.000	19:34
Subject Sample # 1	0.191	19:34
Breath Volume	2.496	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.186	19:38
Breath Volume	1.703	19:38
Air Blank	0.000	19:39
Control Test	0.080	19:39
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUNCAN,N,
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: DUNCAN,N
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: MULVIHILL

DATE/TIME OF TEST: 04/07/2017 18:35:36

Diagnostics Check	OK	18:29
Air Blank	0.000	18:30
Control Test	0.081	18:30
Air Blank	0.000	18:31
Subject Sample # 1	0.000	18:31
Breath Volume	1.585	18:31
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.000	18:34
Breath Volume	1.257	18:34
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 04/08/2017 02:25:19

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:19
Subject Sample # 1	0.182	02:20
Breath Volume	2.437	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.186	02:23
Breath Volume	2.312	02:23
Air Blank	0.000	02:24
Control Test	0.079	02:24
Air Blank	0.000	02:25
Diagnostics Check	OK	02:25

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHAEFER,KENDALL,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: WITT,KELLY
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: GRENIER

DATE/TIME OF TEST: 04/10/2017 00:08:08

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 3 0.036	23:55	
Breath Volume	2.347	23:55	
Air Blank	0.000	23:56	
Air Blank	0.000	23:57	
Subject Sample	# 1	00:01	No Sample Provided
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Air Blank	0.000	00:03	
Subject Sample	# 2	00:06	No Sample Provided
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALGER, MICHAEL, C
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CALHOUN, K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: FATA

DATE/TIME OF TEST: 04/11/2017 04:32:07

Diagnostics Check	OK	04:20	
Air Blank	0.000	04:21	
Control Test	0.080	04:21	
Air Blank	0.000	04:22	
Subject Sample	# 1	04:25	No Sample Provided
Breath Volume	0.000	04:25	
Air Blank	0.000	04:25	
Air Blank	0.000	04:27	
Subject Sample	# 2	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Control Test	0.080	04:31	
Air Blank	0.000	04:32	
Diagnostics Check	OK	04:32	

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000938

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAFLIN,DAVID,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TESTER,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 03/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 23016080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 06:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:30

SUBJECT'S LAST NAME: TESTED

DATE/TIME OF TEST: 04/11/2017 07:23:29

Diagnostics Check	OK	07:17	
Air Blank	0.000	07:17	
Control Test	0.080	07:18	
Air Blank	0.000	07:18	
Subject Sample	# 1 0.000	07:19	
Breath Volume	2.085	07:19	
Air Blank	0.000	07:19	
Air Blank	0.000	07:21	
Subject Sample	# 2	07:22	Subject Test Refused
Breath Volume	0.000	07:22	
Air Blank	0.000	07:22	
Control Test	0.080	07:22	
Air Blank	0.000	07:23	
Diagnostics Check	OK	07:23	

Data Download Date/Time: 04/12/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIKEMAN,H,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BELL,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BERNABEBERNABE

DATE/TIME OF TEST: 04/02/2017 03:20:41

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.079	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.253	03:11
Breath Volume	1.531	03:11
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.231	03:15
Breath Volume	2.132	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 3	0.247	03:19
Breath Volume	1.832	03:19
Air Blank	0.000	03:19
Control Test	0.077	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 04/04/2017 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000941

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER, CHARLES, D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WATTS, J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 04/02/2017 06:28:51

Diagnostics Check	OK	06:22
Air Blank	0.000	06:22
Control Test	0.077	06:23
Air Blank	0.000	06:23
Subject Sample # 1	0.201	06:24
Breath Volume	2.175	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:26
Subject Sample # 2	0.188	06:27
Breath Volume	1.546	06:27
Air Blank	0.000	06:27
Control Test	0.078	06:28
Air Blank	0.000	06:28
Diagnostics Check	OK	06:28

Data Download Date/Time: 04/04/2017 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,YVETTE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: TAKASH

DATE/TIME OF TEST: 04/01/2017 04:04:24

Diagnostics Check	OK	03:56
Air Blank	0.000	03:56
Control Test	0.079	03:57
Air Blank	0.000	03:57
Subject Sample # 1	0.275	03:59
Breath Volume	1.593	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.292	04:02
Breath Volume	1.593	04:02
Air Blank	0.000	04:03
Control Test	0.078	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 04/04/2017 13:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDOZA,U,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SMITH,L
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 04/01/2017 23:01:02

Diagnostics Check	OK	22:55
Air Blank	0.000	22:55
Control Test	0.079	22:55
Air Blank	0.000	22:56
Subject Sample # 1	0.171	22:56
Breath Volume	2.089	22:56
Air Blank	0.000	22:57
Air Blank	0.000	22:59
Subject Sample # 2	0.167	22:59
Breath Volume	1.949	22:59
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:00
Diagnostics Check	OK	23:01

Data Download Date/Time: 04/04/2017 13:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONSOOSTEN,STEVEN,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VONSOOSTEN,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: PITEO

DATE/TIME OF TEST: 04/04/2017 01:33:31

Diagnostics Check	OK	01:27	
Air Blank	0.000	01:28	
Control Test	0.080	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1 0.045	01:29	
Breath Volume	1.824	01:29	
Air Blank	0.000	01:30	
Air Blank		01:32	Improper Sample
Air Blank	0.000	01:32	
Control Test	0.079	01:32	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONSOOSTEN,STEVEN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VONSOOSTEN,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: PITEO

DATE/TIME OF TEST: 04/04/2017 01:41:54

Diagnostics Check	OK	01:35
Air Blank	0.000	01:36
Control Test	0.079	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.046	01:37
Breath Volume	2.484	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.044	01:40
Breath Volume	2.628	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 04/04/2017 02:40:20

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.241	02:35
Breath Volume	1.933	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.237	02:38
Breath Volume	1.812	02:38
Air Blank	0.000	02:39
Control Test	0.078	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROENDYK,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: LARCHER

DATE/TIME OF TEST: 04/04/2017 03:29:30

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.080	03:23
Air Blank	0.000	03:23
Subject Sample # 1	0.000	03:25
Breath Volume	1.257	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.000	03:28
Breath Volume	1.339	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROHRAFF,HARRY,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,MIKE
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/06/2017 02:17:11

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.230	02:12
Breath Volume	1.996	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.230	02:15
Breath Volume	1.992	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEROSA,ANNAMARIE,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEROSA,ANNAMARIE
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: VELIZ

DATE/TIME OF TEST: 04/06/2017 03:41:15

Diagnostics Check	OK	03:34
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Subject Sample # 1	0.129	03:36
Breath Volume	1.394	03:36
Air Blank	0.000	03:36
Air Blank	0.000	03:38
Subject Sample # 2	0.134	03:39
Breath Volume	2.234	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VADNEY, MICHELE, M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ZERPA, E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: NYE

DATE/TIME OF TEST: 04/07/2017 00:21:56

Diagnostics Check	OK	00:15
Air Blank	0.000	00:15
Control Test	0.080	00:16
Air Blank	0.000	00:16
Subject Sample # 1	0.151	00:17
Breath Volume	2.542	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:19
Subject Sample # 2	0.143	00:20
Breath Volume	1.906	00:20
Air Blank	0.000	00:21
Control Test	0.080	00:21
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTSINGER,CODY,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: WEEKENDER,W
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: FITZPATRICK

DATE/TIME OF TEST: 04/07/2017 19:52:20

Diagnostics Check	OK	19:45
Air Blank	0.000	19:46
Control Test	0.080	19:46
Air Blank	0.000	19:46
Subject Sample # 1	0.036	19:47
Breath Volume	2.460	19:47
Air Blank	0.000	19:48
Air Blank	0.000	19:50
Subject Sample # 2	0.036	19:50
Breath Volume	2.359	19:50
Air Blank	0.000	19:51
Control Test	0.081	19:51
Air Blank	0.000	19:52
Diagnostics Check	OK	19:52

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: PETRACCA

DATE/TIME OF TEST: 04/08/2017 04:11:27

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.126	04:07
Breath Volume	1.859	04:07
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.116	04:09
Breath Volume	1.984	04:09
Air Blank	0.000	04:10
Control Test	0.078	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DEVENDORF

DATE/TIME OF TEST: 04/09/2017 04:36:12

Diagnostics Check	OK	04:28
Air Blank	0.000	04:28
Control Test	0.080	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.137	04:30
Breath Volume	1.648	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.135	04:34
Breath Volume	2.183	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Diagnostics Check	OK	04:36

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELANEY,PATRICK,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: TURI,T
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: DEAL

DATE/TIME OF TEST: 04/11/2017 10:59:37

Diagnostics Check	OK	10:52
Air Blank	0.000	10:53
Control Test	0.080	10:53
Air Blank	0.000	10:53
Subject Sample # 1	0.197	10:54
Breath Volume	1.625	10:54
Air Blank	0.000	10:55
Air Blank	0.000	10:57
Subject Sample # 2	0.194	10:58
Breath Volume	1.976	10:58
Air Blank	0.000	10:58
Control Test	0.080	10:59
Air Blank	0.000	10:59
Diagnostics Check	OK	10:59

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: USHER

DATE/TIME OF TEST: 04/12/2017 19:17:30

Diagnostics Check	OK	19:11
Air Blank	0.000	19:11
Control Test	0.080	19:11
Air Blank	0.000	19:12
Subject Sample # 1	0.179	19:12
Breath Volume	2.253	19:12
Air Blank	0.000	19:13
Air Blank	0.000	19:15
Subject Sample # 2	0.180	19:15
Breath Volume	2.214	19:15
Air Blank	0.000	19:16
Control Test	0.080	19:16
Air Blank	0.000	19:17
Diagnostics Check	OK	19:17

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NELSON,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: CORNELL

DATE/TIME OF TEST: 04/13/2017 21:04:05

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.232	20:59
Breath Volume	1.820	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.225	21:02
Breath Volume	1.660	21:02
Air Blank	0.000	21:03
Control Test	0.079	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:04

Data Download Date/Time: 05/01/2017 15:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DECKER,STEPHEN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: HILTON

DATE/TIME OF TEST: 04/15/2017 03:27:05

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:15	
Control Test	0.080	03:15	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:19	
Air Blank	0.000	03:21	
Subject Sample	# 2	0.249	03:21
Breath Volume	1.886	03:21	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 3	0.255	03:25
Breath Volume	1.789	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:27	
Diagnostics Check	OK	03:27	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ASHBY,N
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 04/15/2017 05:49:24

Diagnostics Check		OK	05:37	
Air Blank		0.000	05:37	
Control Test		0.080	05:38	
Air Blank		0.000	05:38	
Subject Sample	# 1	0.204	05:41	Volume Not Met
Breath Volume		0.222	05:41	
Air Blank		0.000	05:42	
Air Blank		0.000	05:44	
Subject Sample	# 2	0.208	05:44	
Breath Volume		1.257	05:44	
Air Blank		0.000	05:45	
Air Blank		0.000	05:47	
Subject Sample	# 3	0.204	05:47	
Breath Volume		1.886	05:47	
Air Blank		0.000	05:48	
Control Test		0.079	05:48	
Air Blank		0.000	05:49	
Diagnostics Check		OK	05:49	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDILL,ROBERT,S
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MULRY,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 13:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:26

SUBJECT'S LAST NAME: HERNANDEZ CERRATO

DATE/TIME OF TEST: 04/15/2017 15:08:26

Diagnostics Check	OK	14:59
Air Blank	0.000	15:00
Control Test	0.080	15:00
Air Blank	0.000	15:01
Subject Sample # 1	0.161	15:03
Breath Volume	1.386	15:03
Air Blank	0.000	15:04
Air Blank	0.000	15:05
Subject Sample # 2	0.146	15:06
Breath Volume	1.585	15:06
Air Blank	0.000	15:07
Control Test	0.081	15:07
Air Blank	0.000	15:08
Diagnostics Check	OK	15:08

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOWLIN, T
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 17:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: ALONSO

DATE/TIME OF TEST: 04/17/2017 18:56:39

Diagnostics Check	OK	18:53	
Air Blank	0.000	18:54	
Control Test	0.080	18:54	
Air Blank	0.000	18:55	
Subject Sample	# 1	18:55	Subject Test Refused
Breath Volume	0.000	18:55	
Air Blank	0.000	18:55	
Control Test	0.080	18:56	
Air Blank	0.000	18:56	
Diagnostics Check	OK	18:56	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: WATT

DATE/TIME OF TEST: 04/17/2017 20:48:04

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.080	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.186	20:43
Breath Volume	2.054	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.192	20:46
Breath Volume	2.148	20:46
Air Blank	0.000	20:47
Control Test	0.080	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:48

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: VIELMAN SALAZAR

DATE/TIME OF TEST: 04/17/2017 23:28:53

Diagnostics Check	OK	23:22
Air Blank	0.000	23:23
Control Test	0.080	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.234	23:24
Breath Volume	2.710	23:24
Air Blank	0.000	23:25
Air Blank	0.000	23:26
Subject Sample # 2	0.237	23:27
Breath Volume	2.714	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:28
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GALLAGHER,B
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CIMENO

DATE/TIME OF TEST: 04/18/2017 03:06:46

Diagnostics Check	OK	03:00
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.122	03:02
Breath Volume	2.867	03:02
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.117	03:05
Breath Volume	3.027	03:05
Air Blank	0.000	03:05
Control Test	0.080	03:06
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONSOOSTEN,STEVEN,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VONSOOSTEN,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 04/18/2017 04:16:08

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:09	
Control Test	0.080	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:13	No Sample Provided
Breath Volume	0.000	04:13	
Air Blank	0.000	04:13	
Air Blank		04:15	Ambient Fail
Air Blank	0.000	04:16	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VONSOOSTEN,STEVEN,S
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: VONSOOSTEN,STEVEN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: TRAVIS

DATE/TIME OF TEST: 04/18/2017 04:27:57

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.079	04:21
Air Blank	0.000	04:21
Subject Sample # 1	0.141	04:22
Breath Volume	1.488	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.149	04:26
Breath Volume	2.125	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:27
Diagnostics Check	OK	04:27

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 04/18/2017 21:17:26

Diagnostics Check	OK	21:09
Air Blank	0.000	21:10
Control Test	0.080	21:10
Air Blank	0.000	21:10
Subject Sample # 1	0.000	21:12
Breath Volume	1.406	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.000	21:16
Breath Volume	1.449	21:16
Air Blank	0.000	21:16
Control Test	0.080	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIBEY,AMANDA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: AVILES

DATE/TIME OF TEST: 04/19/2017 00:14:16

Diagnostics Check	OK	00:06
Air Blank	0.000	00:06
Control Test	0.080	00:07
Air Blank	0.000	00:07
Subject Sample # 1	0.054	00:09
Breath Volume	1.628	00:09
Air Blank	0.000	00:10
Air Blank	0.000	00:11
Subject Sample # 2	0.050	00:12
Breath Volume	1.203	00:12
Air Blank	0.000	00:13
Control Test	0.079	00:13
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: CORTIZA

DATE/TIME OF TEST: 04/20/2017 00:59:29

Diagnostics Check	OK	00:52
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.073	00:54
Breath Volume	1.519	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.072	00:57
Breath Volume	2.484	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLIN,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MAURER

DATE/TIME OF TEST: 04/21/2017 01:06:55

Diagnostics Check		OK	00:55	
Air Blank		0.000	00:55	
Control Test		0.080	00:56	
Air Blank		0.000	00:56	
Subject Sample	# 1	0.016	00:59	Volume Not Met
Breath Volume		0.382	00:59	
Air Blank		0.000	01:00	
Air Blank		0.000	01:02	
Subject Sample	# 2	0.120	01:05	Volume Not Met
Breath Volume		0.593	01:05	
Air Blank		0.000	01:06	
Control Test		0.080	01:06	
Air Blank		0.000	01:06	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DESIMONE,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MAURER

DATE/TIME OF TEST: 04/21/2017 01:16:17

Diagnostics Check	OK	01:09
Air Blank	0.000	01:09
Control Test	0.080	01:10
Air Blank	0.000	01:10
Subject Sample # 1	0.149	01:11
Breath Volume	2.535	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.145	01:14
Breath Volume	2.863	01:14
Air Blank	0.000	01:15
Control Test	0.078	01:15
Air Blank	0.000	01:16
Diagnostics Check	OK	01:16

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUIGLEY, MICHAEL, R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SHAW, .
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 12:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: BERGEN

DATE/TIME OF TEST: 04/22/2017 13:47:57

Diagnostics Check	OK	13:37
Air Blank	0.000	13:37
Control Test	0.080	13:38
Air Blank	0.000	13:38
Subject Sample # 1	0.322	13:40
Breath Volume	2.281	13:40
Air Blank	0.000	13:41
Air Blank	0.000	13:42
Subject Sample # 2	0.293	13:43
Breath Volume	1.621	13:43
Air Blank	0.000	13:44
Air Blank	0.000	13:45
Subject Sample # 3	0.306	13:46
Breath Volume	1.539	13:46
Air Blank	0.000	13:47
Control Test	0.079	13:47
Air Blank	0.000	13:47
Diagnostics Check	OK	13:47

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BUTRICO

DATE/TIME OF TEST: 04/22/2017 20:55:00

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:46	
Control Test	0.080	20:46	
Air Blank	0.000	20:47	
Subject Sample # 1	0.248	20:48	
Breath Volume	1.457	20:48	
Air Blank	0.000	20:48	
Air Blank	0.000	20:50	
Subject Sample # 2	0.232	20:53	Volume Not Met
Breath Volume	0.726	20:53	
Air Blank		20:54	Purge Fail
Air Blank	0.000	20:55	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: BUTRICO

DATE/TIME OF TEST: 04/22/2017 21:09:39

Diagnostics Check		OK	21:02
Air Blank		0.000	21:03
Control Test		0.077	21:03
Air Blank		0.000	21:03
Subject Sample	# 1	0.231	21:04
Breath Volume		1.367	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:07
Subject Sample	# 2	0.230	21:07
Breath Volume		2.378	21:07
Air Blank		0.000	21:08
Control Test		0.078	21:09
Air Blank		0.000	21:09
Diagnostics Check		OK	21:09

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LEONI,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: LOPEZ-RODRIGUEZ

DATE/TIME OF TEST: 04/23/2017 01:04:31

Diagnostics Check	OK	00:57
Air Blank	0.000	00:58
Control Test	0.080	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.141	00:59
Breath Volume	1.324	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:02
Subject Sample # 2	0.138	01:03
Breath Volume	1.160	01:03
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ONDREA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STUDZINSKI,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 04/23/2017 02:57:07

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:43	
Control Test		0.080	02:43	
Air Blank		0.000	02:44	
Subject Sample	# 1	0.151	02:47	Volume Not Met
Breath Volume		0.160	02:47	
Air Blank		0.000	02:47	
Air Blank		0.000	02:49	
Subject Sample	# 2	0.187	02:51	
Breath Volume		2.019	02:51	
Air Blank		0.000	02:51	
Air Blank		0.000	02:53	
Subject Sample	# 3	0.182	02:55	
Breath Volume		2.062	02:55	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MUNOZ LA O

DATE/TIME OF TEST: 04/23/2017 03:22:01

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:13
Subject Sample # 1	0.225	03:15
Breath Volume	1.371	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.215	03:20
Breath Volume	1.765	03:20
Air Blank	0.000	03:21
Control Test	0.079	03:21
Air Blank	0.000	03:21
Diagnostics Check	OK	03:22

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DELUCA,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ROCUYAN

DATE/TIME OF TEST: 04/23/2017 04:56:18

Diagnostics Check	OK	04:50
Air Blank	0.000	04:50
Control Test	0.080	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.105	04:51
Breath Volume	2.035	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.099	04:54
Breath Volume	2.566	04:54
Air Blank	0.000	04:55
Control Test	0.080	04:55
Air Blank	0.000	04:56
Diagnostics Check	OK	04:56

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGE,LYNN,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,KURT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 04/24/2017 00:04:55

Diagnostics Check	OK	23:59	
Air Blank	0.000	23:59	
Control Test	0.080	23:59	
Air Blank	0.000	00:00	
Subject Sample	# 1 0.139	00:01	
Breath Volume	2.613	00:01	
Air Blank	0.000	00:01	
Air Blank		00:03	Improper Sample
Air Blank	0.000	00:04	
Control Test	0.081	00:04	
Air Blank	0.000	00:04	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGE,LYNN,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,KURT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: GROVES

DATE/TIME OF TEST: 04/24/2017 00:17:02

Diagnostics Check	OK	00:10
Air Blank	0.000	00:11
Control Test	0.080	00:11
Air Blank	0.000	00:11
Subject Sample # 1	0.141	00:12
Breath Volume	2.976	00:12
Air Blank	0.000	00:13
Air Blank	0.000	00:14
Subject Sample # 2	0.137	00:15
Breath Volume	2.972	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:16
Diagnostics Check	OK	00:17

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SILL,,
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 04/25/2017 23:57:34

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.213	23:52
Breath Volume	1.511	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.209	23:55
Breath Volume	2.312	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GANGL,B,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: HORSMAN,TOM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: OBREGON

DATE/TIME OF TEST: 04/26/2017 11:45:07

Diagnostics Check	OK	11:38
Air Blank	0.000	11:39
Control Test	0.080	11:39
Air Blank	0.000	11:40
Subject Sample # 1	0.042	11:40
Breath Volume	2.128	11:40
Air Blank	0.000	11:41
Air Blank	0.000	11:43
Subject Sample # 2	0.041	11:43
Breath Volume	2.003	11:43
Air Blank	0.000	11:44
Control Test	0.080	11:44
Air Blank	0.000	11:45
Diagnostics Check	OK	11:45

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: JEAN

DATE/TIME OF TEST: 04/27/2017 00:27:33

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.000	00:23
Breath Volume	1.214	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.000	00:26
Breath Volume	1.484	00:26
Air Blank	0.000	00:26
Control Test	0.081	00:26
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: PARRITT

DATE/TIME OF TEST: 04/27/2017 02:26:42

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:18	
Control Test	0.081	02:18	
Air Blank	0.000	02:19	
Subject Sample	# 1	02:22	No Sample Provided
Breath Volume	0.000	02:22	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2	02:25	Subject Test Refused
Breath Volume	0.000	02:25	
Air Blank	0.000	02:25	
Control Test	0.081	02:26	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,T,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 04/27/2017 03:40:45

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.080	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:39	Slope Not Met
Breath Volume	1.191	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOGAN,T,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 04/27/2017 04:08:43

Diagnostics Check	OK	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.160	04:04
Breath Volume	1.429	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.163	04:07
Breath Volume	1.277	04:07
Air Blank	0.000	04:07
Control Test	0.081	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNORAT,ADELY,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COLLINS,J
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ALVA

DATE/TIME OF TEST: 04/28/2017 01:41:14

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:35
Air Blank	0.000	01:36
Subject Sample # 1	0.143	01:36
Breath Volume	2.429	01:36
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.140	01:39
Breath Volume	2.457	01:39
Air Blank	0.000	01:40
Control Test	0.081	01:40
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,KURT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: POLITE

DATE/TIME OF TEST: 04/28/2017 02:58:58

Diagnostics Check	OK	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:53
Subject Sample # 1	0.196	02:54
Breath Volume	2.652	02:54
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.199	02:57
Breath Volume	2.542	02:57
Air Blank	0.000	02:58
Control Test	0.080	02:58
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO, JONATHAN, B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MAHOLTZ, S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: ANTONI

DATE/TIME OF TEST: 04/29/2017 03:31:00

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.203	03:26
Breath Volume	1.480	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.189	03:29
Breath Volume	2.042	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:31

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOLO,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STAUFFER,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 04/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 11315080A1
DRY GAS EXPIRY DATE: 05/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: AMES

DATE/TIME OF TEST: 04/30/2017 02:22:31

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.259	02:17
Breath Volume	2.058	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.256	02:20
Breath Volume	2.503	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 05/01/2017 15:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SCHULTE,A
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: TRESER

DATE/TIME OF TEST: 04/01/2017 22:32:32

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.128	22:27
Breath Volume	1.574	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.135	22:31
Breath Volume	1.503	22:31
Air Blank	0.000	22:31
Control Test	0.080	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ANGELINI,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/08/2017 03:17:53

Diagnostics Check	OK	03:10
Air Blank	0.000	03:10
Control Test	0.080	03:10
Air Blank	0.000	03:11
Subject Sample # 1	0.205	03:12
Breath Volume	2.398	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.205	03:16
Breath Volume	2.742	03:16
Air Blank	0.000	03:16
Control Test	0.081	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MILLS,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MCCANN

DATE/TIME OF TEST: 04/09/2017 04:16:49

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.172	04:12
Breath Volume	4.394	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.180	04:15
Breath Volume	2.917	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: RESTY,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: FREIBURGER

DATE/TIME OF TEST: 04/12/2017 22:47:25

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.079	22:41
Air Blank	0.000	22:41
Subject Sample # 1	0.165	22:43
Breath Volume	1.648	22:43
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.172	22:45
Breath Volume	1.539	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: QUEGAN,TOM
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 04/17/2017 01:08:38

Diagnostics Check	OK	01:00
Air Blank	0.000	01:01
Control Test	0.079	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.051	01:04
Breath Volume	1.222	01:04
Air Blank	0.000	01:04
Air Blank	0.000	01:06
Subject Sample # 2	0.051	01:07
Breath Volume	1.480	01:07
Air Blank	0.000	01:07
Control Test	0.079	01:08
Air Blank	0.000	01:08
Diagnostics Check	OK	01:08

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VAUGHN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 04/18/2017 04:17:13

Diagnostics Check	OK	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Subject Sample # 1	0.163	04:12
Breath Volume	2.554	04:12
Air Blank	0.000	04:13
Air Blank	0.000	04:14
Subject Sample # 2	0.170	04:15
Breath Volume	1.957	04:15
Air Blank	0.000	04:16
Control Test	0.078	04:16
Air Blank	0.000	04:17
Diagnostics Check	OK	04:17

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: VAUGHAN,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BUTTON

DATE/TIME OF TEST: 04/18/2017 20:02:32

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.193	19:55
Breath Volume	1.328	19:55
Air Blank	0.000	19:56
Air Blank	0.000	19:58
Subject Sample # 2	0.195	20:01
Breath Volume	1.152	20:01
Air Blank	0.000	20:01
Control Test	0.079	20:01
Air Blank	0.000	20:02
Diagnostics Check	OK	20:02

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000946

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER, JEREMY, A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WRASSE, NICK
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 04/18/2017 21:20:25

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.079	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.185	21:16
Breath Volume	1.484	21:16
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.185	21:18
Breath Volume	3.058	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 04/19/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLENDENING,JAMES,T
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ESPINOSA

DATE/TIME OF TEST: 04/12/2017 23:03:45

Diagnostics Check	OK	23:00	
Air Blank	0.000	23:01	
Control Test	0.078	23:01	
Air Blank	0.000	23:02	
Subject Sample	# 1	23:02	Subject Test Refused
Breath Volume	0.000	23:02	
Air Blank	0.000	23:02	
Control Test	0.078	23:03	
Air Blank	0.000	23:03	
Diagnostics Check	OK	23:03	

Data Download Date/Time: 04/27/2017 16:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: MARTINEZ CABRERA

DATE/TIME OF TEST: 04/14/2017 18:17:53

Diagnostics Check	OK	18:12	
Air Blank	0.000	18:12	
Control Test	0.078	18:13	
Air Blank	0.000	18:13	
Subject Sample	# 1	18:16	Subject Test Refused
Breath Volume	0.019	18:16	
Air Blank	0.000	18:16	
Control Test	0.079	18:17	
Air Blank	0.000	18:17	
Diagnostics Check	OK	18:17	

Data Download Date/Time: 04/27/2017 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MURRAY,S
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: TOBIAS

DATE/TIME OF TEST: 04/17/2017 02:05:51

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.192	02:01
Breath Volume	2.933	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.187	02:04
Breath Volume	2.109	02:04
Air Blank	0.000	02:04
Control Test	0.078	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 04/27/2017 16:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PARKS,TRAVIS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 03/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:16

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 04/17/2017 10:20:12

Diagnostics Check	OK	10:17	
Air Blank	0.000	10:17	
Control Test	0.079	10:18	
Air Blank	0.000	10:18	
Subject Sample	# 1	10:18	Subject Test Refused
Breath Volume	0.000	10:18	
Air Blank	0.000	10:19	
Control Test	0.078	10:19	
Air Blank	0.000	10:20	
Diagnostics Check	OK	10:20	

Data Download Date/Time: 04/27/2017 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HICKS,CHRISTINA,L
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CROSWELL,DONNIE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 16:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 04/17/2017 17:19:42

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.078	17:14
Air Blank	0.000	17:14
Subject Sample # 1	0.127	17:15
Breath Volume	2.699	17:15
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.133	17:18
Breath Volume	1.542	17:18
Air Blank	0.000	17:18
Control Test	0.076	17:19
Air Blank	0.000	17:19
Diagnostics Check	OK	17:19

Data Download Date/Time: 04/27/2017 16:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: HERNANDEZ,D
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 03/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: CHOQUETTE

DATE/TIME OF TEST: 04/25/2017 00:27:51

Diagnostics Check	OK	00:21
Air Blank	0.000	00:21
Control Test	0.078	00:21
Air Blank	0.000	00:22
Subject Sample # 1	0.128	00:23
Breath Volume	2.031	00:23
Air Blank	0.000	00:23
Air Blank	0.000	00:25
Subject Sample # 2	0.137	00:26
Breath Volume	3.195	00:26
Air Blank	0.000	00:26
Control Test	0.079	00:27
Air Blank	0.000	00:27
Diagnostics Check	OK	00:27

Data Download Date/Time: 04/27/2017 16:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDRAGON, EDWIN, F
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MURRAY, STACIE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 04/28/2017 23:42:06

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.154	23:37
Breath Volume	1.500	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:39
Subject Sample # 2	0.157	23:40
Breath Volume	3.117	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:42

Data Download Date/Time: 05/30/2017 11:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXTER,AUDREY,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: MURRAY,STACIE
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 04/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:51

SUBJECT'S LAST NAME: TOVAR

DATE/TIME OF TEST: 04/29/2017 06:18:00

Diagnostics Check	OK	06:11
Air Blank	0.000	06:12
Control Test	0.078	06:12
Air Blank	0.000	06:13
Subject Sample # 1	0.030	06:13
Breath Volume	2.226	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.027	06:16
Breath Volume	1.750	06:16
Air Blank	0.000	06:17
Control Test	0.078	06:17
Air Blank	0.000	06:17
Diagnostics Check	OK	06:18

Data Download Date/Time: 05/30/2017 11:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: DONTIGNY

DATE/TIME OF TEST: 04/13/2017 18:24:29

Diagnostics Check	OK	18:22	
Air Blank	0.000	18:23	
Control Test	0.082	18:23	
Air Blank	0.000	18:23	
Subject Sample	# 1	18:24	RFI Detect
Breath Volume	0.000	18:24	
Air Blank		18:24	RFI Detect
Air Blank		18:24	RFI Detect

Data Download Date/Time: 04/28/2017 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: ROY,G
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: DONTIGNY

DATE/TIME OF TEST: 04/13/2017 18:28:11

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:26	
Control Test	0.081	18:26	
Air Blank	0.000	18:27	
Subject Sample	# 1	18:27	RFI Detect
Breath Volume	0.000	18:27	
Air Blank		18:27	RFI Detect
Air Blank	0.000	18:28	

Data Download Date/Time: 04/28/2017 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROY,G,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: WRIGHT,S
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 04/22/2017 01:23:48

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.081	01:18
Air Blank		0.000	01:18
Subject Sample	# 1	0.216	01:19
Breath Volume		2.710	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:21
Subject Sample	# 2	0.222	01:22
Breath Volume		2.484	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 04/28/2017 10:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 03/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HATTEN

DATE/TIME OF TEST: 04/22/2017 02:34:30

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:33	
Control Test	0.081	02:34	
Air Blank		02:34	RFI Detect
Air Blank		02:34	RFI Detect

Data Download Date/Time: 04/28/2017 10:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,ERIC,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: DAVIS,TIMMY
ARRESTING OFFICER AGENCY: Apalachicola PD
LAST AGENCY INSPECTION: 03/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 04/06/2017 23:59:37

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.155	23:53
Breath Volume	2.277	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.159	23:57
Breath Volume	1.949	23:57
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 04/20/2017 09:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 644476
BREATH TEST OPERATOR: MCNEIL,DARRIN,J	DRY GAS EXPIRY DATE: 12/01/2019
BREATH TEST OPERATOR AGENCY: Clewiston PD	VIOLATION CODE: DUI
ARRESTING OFFICER: MCNEIL,DARRIN	ARREST TIME: 19:04
ARRESTING OFFICER AGENCY: Clewiston PD	GENDER: Male
LAST AGENCY INSPECTION: 03/01/2017	OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: RANGEL CABRIELES DATE/TIME OF TEST: 04/06/2017 20:06:51

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.081	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.123	20:02
Breath Volume	3.160	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:04
Subject Sample # 2	0.125	20:05
Breath Volume	3.476	20:05
Air Blank	0.000	20:05
Control Test	0.081	20:06
Air Blank	0.000	20:06
Diagnostics Check	OK	20:06

Data Download Date/Time: 04/14/2017 03:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: HESTER,JAMES
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 03/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: LUIS PEREZ

DATE/TIME OF TEST: 04/09/2017 22:17:34

Diagnostics Check	OK	22:09
Air Blank	0.000	22:10
Control Test	0.081	22:10
Air Blank	0.000	22:11
Subject Sample # 1	0.086	22:12
Breath Volume	3.140	22:12
Air Blank	0.000	22:12
Air Blank	0.000	22:14
Subject Sample # 2	0.076	22:16
Breath Volume	3.542	22:16
Air Blank	0.000	22:16
Control Test	0.082	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 04/14/2017 03:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.	Date or Time Changed: No
SOFTWARE VERSION: 8100.27	DRY GAS STANDARD LOT #: 644476
BREATH TEST OPERATOR: LAROSA,RYAN,M	DRY GAS EXPIRY DATE: 12/01/2019
BREATH TEST OPERATOR AGENCY: Clewiston PD	VIOLATION CODE: DUI
ARRESTING OFFICER: LAROSA,RYAN	ARREST TIME: 01:43
ARRESTING OFFICER AGENCY: Clewiston PD	GENDER: Male
LAST AGENCY INSPECTION: 03/01/2017	OBSERVATION PERIOD BEGAN: 02:05
SUBJECT'S LAST NAME: STREETER	DATE/TIME OF TEST: 04/11/2017 02:36:15

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.075	02:31
Breath Volume	1.281	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.076	02:34
Breath Volume	3.187	02:34
Air Blank	0.000	02:35
Control Test	0.080	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 04/14/2017 03:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LAROSA,RYAN,M

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: LAROSA,RYAN

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 04/14/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 644476

DRY GAS EXPIRY DATE: 12/01/2019

VIOLATION CODE: DUI

ARREST TIME: 20:00

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 04/15/2017 20:30:19

Diagnostics Check	OK	20:27	
Air Blank	0.000	20:27	
Control Test	0.080	20:28	
Air Blank	0.000	20:28	
Subject Sample	# 1	20:28	Subject Test Refused
Breath Volume	0.000	20:28	
Air Blank	0.000	20:29	
Control Test	0.080	20:29	
Air Blank	0.000	20:30	
Diagnostics Check	OK	20:30	

Data Download Date/Time: 05/24/2017 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,N,

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: WILLIAMS,N

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 04/14/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 644476

DRY GAS EXPIRY DATE: 12/01/2019

VIOLATION CODE: DUI

ARREST TIME: 00:12

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: COPPEDGE

DATE/TIME OF TEST: 04/16/2017 01:02:04

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.081	00:53
Air Blank		0.000	00:53
Subject Sample	# 1	0.214	00:54
Breath Volume		2.167	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.191	00:57
Breath Volume		1.265	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 3	0.197	01:00
Breath Volume		1.281	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:02

Data Download Date/Time: 05/24/2017 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: MCLAIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 04/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: GUERRA

DATE/TIME OF TEST: 04/17/2017 19:56:58

Diagnostics Check	OK	19:53	
Air Blank	0.000	19:54	
Control Test	0.080	19:54	
Air Blank	0.000	19:55	
Subject Sample	# 1	19:55	Subject Test Refused
Breath Volume	0.000	19:55	
Air Blank	0.000	19:56	
Control Test	0.081	19:56	
Air Blank	0.000	19:56	
Diagnostics Check	OK	19:56	

Data Download Date/Time: 05/24/2017 22:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESTER,JAMES,R
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: WILLIAMS,NICK
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 04/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: SORCI

DATE/TIME OF TEST: 04/24/2017 04:27:42

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:25	
Control Test	0.081	04:25	
Air Blank	0.000	04:26	
Subject Sample	# 1	04:26	Subject Test Refused
Breath Volume	0.000	04:26	
Air Blank	0.000	04:26	
Control Test	0.080	04:27	
Air Blank	0.000	04:27	
Diagnostics Check	OK	04:27	

Data Download Date/Time: 05/24/2017 22:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONTRAGER, WILLIAM, L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CAMPER, ALLAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BRAMLETT

DATE/TIME OF TEST: 04/09/2017 23:31:45

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.078	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.097	23:27
Breath Volume	2.703	23:27
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.100	23:30
Breath Volume	2.339	23:30
Air Blank	0.000	23:30
Control Test	0.079	23:31
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

Data Download Date/Time: 05/18/2017 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, JOSHUA, N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MITCHELL, KENDRICK
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 04/20/2017 22:13:08

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.078	22:07
Air Blank	0.000	22:07
Subject Sample # 1	0.169	22:08
Breath Volume	3.597	22:08
Air Blank	0.000	22:09
Air Blank	0.000	22:10
Subject Sample # 2	0.167	22:11
Breath Volume	3.757	22:11
Air Blank	0.000	22:12
Control Test	0.077	22:12
Air Blank	0.000	22:13
Diagnostics Check	OK	22:13

Data Download Date/Time: 05/18/2017 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COHEN, VALARIE, L
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LLANES, GABRIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 04/23/2017 04:11:15

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.000	04:06
Breath Volume	1.703	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.000	04:09
Breath Volume	1.531	04:09
Air Blank	0.000	04:10
Control Test	0.077	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 05/18/2017 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000956

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,GUS,T
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SMYRNIOS,GUS
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 04/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 04/29/2017 00:58:15

Diagnostics Check	OK	00:51
Air Blank	0.000	00:51
Control Test	0.079	00:52
Air Blank	0.000	00:52
Subject Sample # 1	0.190	00:53
Breath Volume	2.242	00:53
Air Blank	0.000	00:54
Air Blank	0.000	00:55
Subject Sample # 2	0.197	00:56
Breath Volume	2.097	00:56
Air Blank	0.000	00:57
Control Test	0.076	00:57
Air Blank	0.000	00:58
Diagnostics Check	OK	00:58

Data Download Date/Time: 05/18/2017 11:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIOTT,RICHARD,J
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: STROHECKER,KYNDALL
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 04/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 10:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:55

SUBJECT'S LAST NAME: ATKINS

DATE/TIME OF TEST: 04/06/2017 12:28:26

Diagnostics Check	OK	12:22
Air Blank	0.000	12:22
Control Test	0.082	12:23
Air Blank	0.000	12:23
Subject Sample # 1	0.000	12:24
Breath Volume	2.871	12:24
Air Blank	0.000	12:24
Air Blank	0.000	12:26
Subject Sample # 2	0.000	12:27
Breath Volume	2.601	12:27
Air Blank	0.000	12:27
Control Test	0.082	12:27
Air Blank	0.000	12:28
Diagnostics Check	OK	12:28

Data Download Date/Time: 05/15/2017 09:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,STEWRT,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HUNT,GENE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/01/2017 18:18:49

Diagnostics Check	OK	18:18	
Air Blank		18:18	RFI Detect
Air Blank	0.000	18:18	

Data Download Date/Time: 04/13/2017 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,STEWART,G
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HUNT,GENE
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 04/01/2017 18:27:15

Diagnostics Check	OK	18:21
Air Blank	0.000	18:21
Control Test	0.080	18:21
Air Blank	0.000	18:22
Subject Sample # 1	0.000	18:23
Breath Volume	1.578	18:23
Air Blank	0.000	18:23
Air Blank	0.000	18:25
Subject Sample # 2	0.000	18:25
Breath Volume	1.320	18:25
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:27
Diagnostics Check	OK	18:27

Data Download Date/Time: 04/13/2017 09:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 03/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: SHORES

DATE/TIME OF TEST: 04/03/2017 21:54:18

Diagnostics Check	OK	21:48
Air Blank	0.000	21:48
Control Test	0.079	21:49
Air Blank	0.000	21:49
Subject Sample # 1	0.000	21:50
Breath Volume	2.128	21:50
Air Blank	0.000	21:50
Air Blank	0.000	21:52
Subject Sample # 2	0.000	21:52
Breath Volume	2.125	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Diagnostics Check	OK	21:54

Data Download Date/Time: 04/13/2017 09:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN, MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: NOWLING

DATE/TIME OF TEST: 04/14/2017 03:44:37

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Subject Sample	# 1	03:43	Subject Test Refused
Breath Volume	0.000	03:43	
Air Blank	0.000	03:43	
Control Test	0.079	03:44	
Air Blank	0.000	03:44	
Diagnostics Check	OK	03:44	

Data Download Date/Time: 05/09/2017 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN, MITCHELL, P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LESANE, ASHTON
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 04/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 04/16/2017 01:56:49

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.234	01:52
Breath Volume	2.882	01:52
Air Blank	0.000	01:53
Air Blank	0.000	01:54
Subject Sample # 2	0.224	01:55
Breath Volume	3.171	01:55
Air Blank	0.000	01:55
Control Test	0.079	01:56
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 05/09/2017 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WHITNEY

DATE/TIME OF TEST: 04/17/2017 02:32:09

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.128	02:27
Breath Volume		2.417	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.127	02:30
Breath Volume		2.898	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 05/09/2017 09:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAUFFMAN,MITCHELL,P
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: KAUFFMAN,MITCHELL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 04/30/2017 03:18:31

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:16	
Control Test	0.079	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:17	Subject Test Refused
Breath Volume	0.000	03:17	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 05/09/2017 09:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNETTE,JAVON,T
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: BURNETTE,JAVON
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: CHAPPA

DATE/TIME OF TEST: 04/13/2017 20:03:15

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.080	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.140	19:58
Breath Volume	2.445	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.147	20:01
Breath Volume	2.203	20:01
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 05/09/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUCKHALT, WILLIAM, H
BREATH TEST OPERATOR AGENCY: Gadsden County SO
ARRESTING OFFICER: BUCKHALT, WILLIAM
ARRESTING OFFICER AGENCY: Gadsden County SO
LAST AGENCY INSPECTION: 04/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 04/14/2017 23:14:00

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.078	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.222	23:09
Breath Volume	3.582	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.205	23:12
Breath Volume	4.121	23:12
Air Blank	0.000	23:13
Control Test	0.078	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:14

Data Download Date/Time: 05/09/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 04/18/2017 02:58:10

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.079	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.232	02:50
Breath Volume	1.859	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.207	02:53
Breath Volume	1.429	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 3	0.218	02:56
Breath Volume	1.464	02:56
Air Blank	0.000	02:57
Control Test	0.078	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 05/09/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LITTLE,RODERIAL,A
BREATH TEST OPERATOR AGENCY: Havana PD
ARRESTING OFFICER: FLINT,KILIAN
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 04/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: DUBOSE

DATE/TIME OF TEST: 04/22/2017 04:52:03

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.079	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.032	04:47
Breath Volume	2.230	04:47
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.032	04:50
Breath Volume	2.601	04:50
Air Blank	0.000	04:51
Control Test	0.078	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:52

Data Download Date/Time: 05/09/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 04/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: CANSINO-TERCERO

DATE/TIME OF TEST: 04/22/2017 20:21:26

Diagnostics Check	OK	20:15
Air Blank	0.000	20:15
Control Test	0.079	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.215	20:17
Breath Volume	1.632	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.216	20:19
Breath Volume	1.753	20:19
Air Blank	0.000	20:20
Control Test	0.079	20:20
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

Data Download Date/Time: 05/09/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPELAND,MICHAEL,C
BREATH TEST OPERATOR AGENCY: Quincy PD
ARRESTING OFFICER: COPELAND,MICHAEL
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 04/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: DARBY SR

DATE/TIME OF TEST: 04/26/2017 02:35:54

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.078	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.134	02:31
Breath Volume	1.714	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.127	02:34
Breath Volume	2.453	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:35
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 05/09/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SAMUELS

DATE/TIME OF TEST: 04/01/2017 01:26:31

Diagnostics Check	OK	01:20
Air Blank	0.000	01:20
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.093	01:22
Breath Volume	2.105	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.088	01:24
Breath Volume	1.722	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 04/26/2017 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,P
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: TORO

DATE/TIME OF TEST: 04/02/2017 04:00:30

Diagnostics Check	OK	03:53
Air Blank	0.000	03:54
Control Test	0.080	03:54
Air Blank	0.000	03:55
Subject Sample # 1	0.143	03:55
Breath Volume	2.902	03:55
Air Blank	0.000	03:56
Air Blank	0.000	03:58
Subject Sample # 2	0.147	03:58
Breath Volume	2.933	03:58
Air Blank	0.000	03:59
Control Test	0.079	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 04/26/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:22

SUBJECT'S LAST NAME: DAREUS

DATE/TIME OF TEST: 04/02/2017 06:53:46

Diagnostics Check	OK	06:46
Air Blank	0.000	06:47
Control Test	0.079	06:47
Air Blank	0.000	06:47
Subject Sample # 1	0.177	06:49
Breath Volume	1.265	06:49
Air Blank	0.000	06:50
Air Blank	0.000	06:51
Subject Sample # 2	0.172	06:52
Breath Volume	1.121	06:52
Air Blank	0.000	06:52
Control Test	0.077	06:53
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

Data Download Date/Time: 04/26/2017 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCHOA,YAMIANIS,
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: OCHOA,YAMIANIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: GAAL

DATE/TIME OF TEST: 04/02/2017 19:05:00

Diagnostics Check		OK	18:58
Air Blank		0.000	18:59
Control Test		0.081	18:59
Air Blank		0.000	19:00
Subject Sample	# 1	0.200	19:00
Breath Volume		1.382	19:00
Air Blank		0.000	19:01
Air Blank		0.000	19:02
Subject Sample	# 2	0.208	19:03
Breath Volume		2.441	19:03
Air Blank		0.000	19:04
Control Test		0.080	19:04
Air Blank		0.000	19:04
Diagnostics Check		OK	19:05

Data Download Date/Time: 04/26/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPOS, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: CAMPOS, MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 04/08/2017 03:41:54

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.081	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.149	03:36
Breath Volume	2.351	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.149	03:40
Breath Volume	1.949	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 04/26/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEDULLA,VINCENT,A
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PEDULLA,VINCENT
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: WICKLIFF

DATE/TIME OF TEST: 04/15/2017 02:40:49

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.081	02:35
Air Blank	0.000	02:35
Subject Sample # 1	0.101	02:36
Breath Volume	2.410	02:36
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.107	02:39
Breath Volume	3.671	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 04/26/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ROBLES,STEFEN
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: CALDERON MARRERO

DATE/TIME OF TEST: 04/15/2017 05:13:30

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.080	05:07
Air Blank	0.000	05:07
Subject Sample # 1	0.203	05:08
Breath Volume	1.906	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.219	05:11
Breath Volume	3.332	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 04/26/2017 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA,EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:12

SUBJECT'S LAST NAME: BEDARD

DATE/TIME OF TEST: 04/15/2017 19:35:14

Diagnostics Check	OK	19:34	
Air Blank	0.000	19:34	
Control Test	0.000	19:34	RFI Detect
Air Blank	0.000	19:35	

Data Download Date/Time: 04/26/2017 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA,EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 03/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: BEDARD

DATE/TIME OF TEST: 04/15/2017 19:44:58

Diagnostics Check	OK	19:37
Air Blank	0.000	19:38
Control Test	0.080	19:38
Air Blank	0.000	19:39
Subject Sample # 1	0.276	19:40
Breath Volume	3.718	19:40
Air Blank	0.000	19:41
Air Blank	0.000	19:42
Subject Sample # 2	0.262	19:43
Breath Volume	2.976	19:43
Air Blank	0.000	19:44
Control Test	0.078	19:44
Air Blank	0.000	19:44
Diagnostics Check	OK	19:44

Data Download Date/Time: 04/26/2017 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO, ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: FRYE

DATE/TIME OF TEST: 04/27/2017 02:46:10

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.081	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.171	02:41
Breath Volume	1.601	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.163	02:44
Breath Volume	1.238	02:44
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 05/24/2017 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STINSON,ANDREW,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: STINSON,ANDREW
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: HILARIO

DATE/TIME OF TEST: 04/29/2017 03:25:14

Diagnostics Check	OK	03:18
Air Blank	0.000	03:19
Control Test	0.079	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.216	03:20
Breath Volume	1.640	03:20
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.215	03:23
Breath Volume	1.148	03:23
Air Blank	0.000	03:24
Control Test	0.079	03:24
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 05/24/2017 16:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FORD, GLENN, W
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CABRAL, CHRISTOPHER
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 04/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: HERNANDEZ ORTEZ

DATE/TIME OF TEST: 04/29/2017 04:47:17

Diagnostics Check	OK	04:35
Air Blank	0.000	04:36
Control Test	0.080	04:36
Air Blank	0.000	04:37
Subject Sample # 1	0.227	04:37
Breath Volume	1.593	04:37
Air Blank	0.000	04:38
Air Blank	0.000	04:40
Subject Sample # 2	0.201	04:41
Breath Volume	1.726	04:41
Air Blank	0.000	04:42
Air Blank	0.000	04:43
Subject Sample # 3	0.208	04:45
Breath Volume	1.937	04:45
Air Blank	0.000	04:46
Control Test	0.077	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 05/24/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 04/01/2017 03:48:38

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:43
Air Blank	0.000	03:43
Subject Sample # 1	0.176	03:44
Breath Volume	3.933	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.173	03:47
Breath Volume	3.847	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 05/04/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARBON,ASHLEY,S
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: STEVENSON,KYLE
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: CARBON

DATE/TIME OF TEST: 04/08/2017 04:06:33

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.162	04:02
Breath Volume	2.046	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.150	04:05
Breath Volume	1.914	04:05
Air Blank	0.000	04:05
Control Test	0.076	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 05/04/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: BONNETT

DATE/TIME OF TEST: 04/11/2017 00:33:06

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.079	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.121	00:28
Breath Volume		1.695	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.109	00:31
Breath Volume		1.339	00:31
Air Blank		0.000	00:32
Control Test		0.079	00:32
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 05/04/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 04/17/2017 22:45:29

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.079	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.300	22:38
Breath Volume	2.101	22:38
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.321	22:41
Breath Volume	1.937	22:41
Air Blank	0.000	22:41
Air Blank	0.000	22:43
Subject Sample # 3	0.317	22:44
Breath Volume	1.339	22:44
Air Blank	0.000	22:44
Control Test	0.079	22:44
Air Blank	0.000	22:45
Diagnostics Check	OK	22:45

Data Download Date/Time: 05/04/2017 07:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 03/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 36314080A3
DRY GAS EXPIRY DATE: 02/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: ELLIOTT

DATE/TIME OF TEST: 04/21/2017 22:20:26

Diagnostics Check		OK	22:14
Air Blank		0.000	22:14
Control Test		0.078	22:15
Air Blank		0.000	22:15
Subject Sample	# 1	0.223	22:15
Breath Volume		2.445	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.219	22:18
Breath Volume		2.695	22:18
Air Blank		0.000	22:19
Control Test		0.078	22:19
Air Blank		0.000	22:20
Diagnostics Check		OK	22:20

Data Download Date/Time: 05/04/2017 07:54