

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:19

SUBJECT'S LAST NAME: FINN

DATE/TIME OF TEST: 06/29/2017 18:54:20

Diagnostics Check	OK	18:48
Air Blank	0.000	18:48
Control Test	0.079	18:49
Air Blank	0.000	18:49
Subject Sample # 1	0.228	18:49
Breath Volume	2.468	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:52
Subject Sample # 2	0.226	18:52
Breath Volume	2.632	18:52
Air Blank	0.000	18:53
Control Test	0.078	18:53
Air Blank	0.000	18:54
Diagnostics Check	OK	18:54

Data Download Date/Time: 07/26/2017 09:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: HOFFMAN

DATE/TIME OF TEST: 06/03/2017 03:43:27

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.110	03:38
Breath Volume	1.343	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.116	03:41
Breath Volume	1.132	03:41
Air Blank	0.000	03:42
Control Test	0.077	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 06/27/2017 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WINTER

DATE/TIME OF TEST: 06/04/2017 03:31:35

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.078	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.208	03:27
Breath Volume	1.335	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.204	03:30
Breath Volume	1.382	03:30
Air Blank	0.000	03:30
Control Test	0.078	03:30
Air Blank	0.000	03:31
Diagnostics Check	OK	03:31

Data Download Date/Time: 06/27/2017 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLES,RACHEL,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: OCONNELL,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SOLANA

DATE/TIME OF TEST: 06/04/2017 03:42:39

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.077	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.188	03:38
Breath Volume		2.417	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.191	03:41
Breath Volume		2.058	03:41
Air Blank		0.000	03:41
Control Test		0.077	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 06/27/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: OCONNEL,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 06/07/2017 20:52:14

Diagnostics Check	OK	20:43	
Air Blank	0.000	20:43	
Control Test	0.078	20:43	
Air Blank	0.000	20:44	
Subject Sample	# 1	20:47	No Sample Provided
Breath Volume	0.000	20:47	
Air Blank	0.000	20:47	
Air Blank	0.000	20:49	
Subject Sample	# 2	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:51	
Control Test	0.078	20:51	
Air Blank	0.000	20:52	
Diagnostics Check	OK	20:52	

Data Download Date/Time: 06/27/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CABRERA

DATE/TIME OF TEST: 06/09/2017 04:07:26

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.101	04:03
Breath Volume	2.128	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.101	04:05
Breath Volume	2.203	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 06/27/2017 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEPHENS,ZACHARY,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: STEPHENS,ZACHARY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 06/09/2017 04:53:31

Diagnostics Check	OK	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.154	04:48
Breath Volume	1.175	04:48
Air Blank	0.000	04:49
Air Blank	0.000	04:51
Subject Sample # 2	0.153	04:51
Breath Volume	1.144	04:51
Air Blank	0.000	04:52
Control Test	0.076	04:52
Air Blank	0.000	04:53
Diagnostics Check	OK	04:53

Data Download Date/Time: 06/27/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,SCOTT,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MILLER,SCOTT
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:25

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 06/09/2017 16:15:45

Diagnostics Check	OK	16:09
Air Blank	0.000	16:09
Control Test	0.079	16:10
Air Blank	0.000	16:10
Subject Sample # 1	0.000	16:11
Breath Volume	1.304	16:11
Air Blank	0.000	16:11
Air Blank	0.000	16:13
Subject Sample # 2	0.000	16:14
Breath Volume	1.410	16:14
Air Blank	0.000	16:14
Control Test	0.079	16:15
Air Blank	0.000	16:15
Diagnostics Check	OK	16:15

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINSKY,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SHIPP,FRANKIE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:26

SUBJECT'S LAST NAME: GILLINS

DATE/TIME OF TEST: 06/09/2017 19:53:15

Diagnostics Check	OK	19:46
Air Blank	0.000	19:46
Control Test	0.078	19:47
Air Blank	0.000	19:47
Subject Sample # 1	0.164	19:48
Breath Volume	1.519	19:48
Air Blank	0.000	19:49
Air Blank	0.000	19:50
Subject Sample # 2	0.158	19:51
Breath Volume	1.617	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:53
Diagnostics Check	OK	19:53

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN, BRADLEY, A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CHAPMAN, SONJA
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 06/09/2017 20:03:33

Diagnostics Check	OK	19:56
Air Blank	0.000	19:57
Control Test	0.077	19:57
Air Blank	0.000	19:58
Subject Sample # 1	0.179	19:58
Breath Volume	1.914	19:58
Air Blank	0.000	19:59
Air Blank	0.000	20:01
Subject Sample # 2	0.181	20:02
Breath Volume	1.871	20:02
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DROUIN,CAROLINE,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: DROUIN,CAROLINE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 06/15/2017 04:08:22

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.077	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.149	04:03
Breath Volume	2.269	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.141	04:06
Breath Volume	1.683	04:06
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HERTZ,ADAM
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: WEIGEL

DATE/TIME OF TEST: 06/16/2017 23:22:00

Diagnostics Check	OK	23:05	
Air Blank	0.000	23:05	
Control Test	0.078	23:06	
Air Blank	0.000	23:06	
Subject Sample	# 1 0.218	23:09	
Breath Volume	1.312	23:09	
Air Blank	0.000	23:10	
Air Blank	0.000	23:11	
Subject Sample	# 2 0.219	23:15	Slope Not Level
Breath Volume	1.113	23:15	
Air Blank	0.000	23:15	
Air Blank	0.000	23:17	
Subject Sample	# 3 0.224	23:20	
Breath Volume	1.421	23:20	
Air Blank	0.000	23:21	
Control Test	0.078	23:21	
Air Blank	0.000	23:21	
Diagnostics Check	OK	23:22	

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 06/18/2017 03:22:06

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Subject Sample # 1	0.210	03:17
Breath Volume	1.566	03:17
Air Blank	0.000	03:18
Air Blank	0.000	03:20
Subject Sample # 2	0.210	03:20
Breath Volume	1.648	03:20
Air Blank	0.000	03:21
Control Test	0.077	03:21
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: HELD

DATE/TIME OF TEST: 06/19/2017 02:13:14

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.079	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DROUIN,CAROLINE,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: DROUIN,CAROLINE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: MOULDS

DATE/TIME OF TEST: 06/21/2017 01:00:16

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Subject Sample # 1	0.162	00:48	
Breath Volume	1.109	00:48	
Air Blank	0.000	00:48	
Air Blank	0.000	00:50	
Subject Sample # 2	0.150	00:53	Volume Not Met
Breath Volume	0.878	00:53	
Air Blank	0.000	00:54	
Air Blank	0.000	00:56	
Subject Sample # 3	0.151	00:58	
Breath Volume	1.214	00:58	
Air Blank	0.000	00:59	
Control Test	0.077	00:59	
Air Blank	0.000	01:00	
Diagnostics Check	OK	01:00	

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: OCONNELL,SCOTT
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 06/21/2017 20:29:08

Diagnostics Check		OK	20:14	
Air Blank		0.000	20:14	
Control Test		0.079	20:14	
Air Blank		0.000	20:15	
Subject Sample	# 1	0.200	20:18	Volume Not Met
Breath Volume		0.593	20:18	
Air Blank		0.000	20:19	
Air Blank		0.000	20:21	
Subject Sample	# 2	0.228	20:23	
Breath Volume		1.453	20:23	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 3	0.225	20:27	
Breath Volume		1.328	20:27	
Air Blank		0.000	20:28	
Control Test		0.079	20:28	
Air Blank		0.000	20:29	
Diagnostics Check		OK	20:29	

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WHITE,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: GALVEZ ZAMORA

DATE/TIME OF TEST: 06/23/2017 00:07:38

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.078	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.163	00:02
Breath Volume		1.703	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.160	00:06
Breath Volume		1.281	00:06
Air Blank		0.000	00:06
Control Test		0.078	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ROJAS CUERDO

DATE/TIME OF TEST: 06/23/2017 03:57:00

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.080	03:52
Breath Volume	1.527	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.073	03:55
Breath Volume	1.371	03:55
Air Blank	0.000	03:56
Control Test	0.078	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DROUIN,CAROLINE,
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: DROUIN,CAROLINE
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GERARDINO

DATE/TIME OF TEST: 06/24/2017 03:18:43

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:12
Subject Sample # 1	0.105	03:13
Breath Volume	1.546	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:15
Subject Sample # 2	0.104	03:17
Breath Volume	3.011	03:17
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: TRUNCALE

DATE/TIME OF TEST: 06/27/2017 02:05:02

Diagnostics Check	OK	02:01	
Air Blank	0.000	02:01	
Control Test	0.079	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 1	02:03	Subject Test Refused
Breath Volume	0.000	02:03	
Air Blank	0.000	02:04	
Control Test	0.078	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:05	

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GLEESON

DATE/TIME OF TEST: 06/27/2017 04:12:44

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.077	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.149	04:08
Breath Volume	1.585	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:10
Subject Sample # 2	0.145	04:11
Breath Volume	1.542	04:11
Air Blank	0.000	04:11
Control Test	0.076	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 06/27/2017 14:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAPPAN,DANIEL,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: BLUM,WESLEY
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 06:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: TAPPAN

DATE/TIME OF TEST: 06/30/2017 07:50:28

Diagnostics Check	OK	07:44
Air Blank	0.000	07:44
Control Test	0.079	07:45
Air Blank	0.000	07:45
Subject Sample # 1	0.039	07:46
Breath Volume	2.261	07:46
Air Blank	0.000	07:46
Air Blank	0.000	07:48
Subject Sample # 2	0.037	07:49
Breath Volume	2.226	07:49
Air Blank	0.000	07:49
Control Test	0.079	07:49
Air Blank	0.000	07:50
Diagnostics Check	OK	07:50

Data Download Date/Time: 07/26/2017 09:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,FREDERICK,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PREVATT,JENNIFER
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 06/04/2017 03:53:01

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.079	03:51	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:52	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:53	

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHEWNING,AUSTIN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 06/08/2017 22:02:38

Diagnostics Check	OK	21:56
Air Blank	0.000	21:56
Control Test	0.080	21:57
Air Blank	0.000	21:57
Subject Sample # 1	0.053	21:58
Breath Volume	1.292	21:58
Air Blank	0.000	21:58
Air Blank	0.000	22:00
Subject Sample # 2	0.053	22:01
Breath Volume	1.441	22:01
Air Blank	0.000	22:01
Control Test	0.079	22:02
Air Blank	0.000	22:02
Diagnostics Check	OK	22:02

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,FREDERICK,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: SMITH,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: STOSE

DATE/TIME OF TEST: 06/13/2017 20:38:45

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.079	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:37	
Control Test	0.080	20:38	
Air Blank	0.000	20:38	
Diagnostics Check	OK	20:38	

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROSBEE,JAMES,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WEST,JACOB
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SABBAT

DATE/TIME OF TEST: 06/16/2017 03:32:49

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.080	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.204	03:28
Breath Volume		1.992	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.195	03:31
Breath Volume		1.523	03:31
Air Blank		0.000	03:31
Control Test		0.079	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: COZZONE,G
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: VERRONE

DATE/TIME OF TEST: 06/16/2017 21:25:41

Diagnostics Check	OK	21:19
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:20
Subject Sample # 1	0.000	21:21
Breath Volume	1.226	21:21
Air Blank	0.000	21:21
Air Blank	0.000	21:23
Subject Sample # 2	0.000	21:24
Breath Volume	1.316	21:24
Air Blank	0.000	21:24
Control Test	0.079	21:25
Air Blank	0.000	21:25
Diagnostics Check	OK	21:25

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,FREDERICK,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: BUTLER,DARREL
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: DRUMHEISER

DATE/TIME OF TEST: 06/19/2017 03:54:14

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.080	03:48
Air Blank		0.000	03:48
Subject Sample	# 1	0.139	03:49
Breath Volume		2.742	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.158	03:52
Breath Volume		2.394	03:52
Air Blank		0.000	03:53
Control Test		0.079	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FAIRCLOTH, EDWARD, W
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TOWNS, R
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HEAFER

DATE/TIME OF TEST: 06/24/2017 03:01:15

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:55
Air Blank	0.000	02:56
Subject Sample # 1	0.000	02:56
Breath Volume	1.632	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.000	02:59
Breath Volume	1.144	02:59
Air Blank	0.000	03:00
Control Test	0.080	03:00
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,FREDERICK,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LICHTY,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 06/28/2017 01:11:34

Diagnostics Check	OK	01:05
Air Blank	0.000	01:06
Control Test	0.080	01:06
Air Blank	0.000	01:06
Subject Sample # 1	0.049	01:07
Breath Volume	1.585	01:07
Air Blank	0.000	01:07
Air Blank	0.000	01:09
Subject Sample # 2	0.049	01:10
Breath Volume	1.140	01:10
Air Blank	0.000	01:10
Control Test	0.080	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 06/29/2017 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 06/04/2017 00:07:09

Diagnostics Check	OK	00:03	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Control Test	0.082	00:06	
Air Blank	0.000	00:07	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARGENTI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:51

SUBJECT'S LAST NAME: BAUER

DATE/TIME OF TEST: 06/04/2017 05:18:24

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.081	05:12
Air Blank	0.000	05:13
Subject Sample # 1	0.150	05:13
Breath Volume	3.480	05:13
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.150	05:16
Breath Volume	3.027	05:16
Air Blank	0.000	05:17
Control Test	0.081	05:17
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUERRERO,CASEY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: MELLO

DATE/TIME OF TEST: 06/05/2017 03:48:16

Diagnostics Check	OK	03:41
Air Blank	0.000	03:42
Control Test	0.081	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.231	03:43
Breath Volume	2.363	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.225	03:46
Breath Volume	2.156	03:46
Air Blank	0.000	03:47
Control Test	0.081	03:47
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CLAIRE,CASEY
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: AVILA

DATE/TIME OF TEST: 06/09/2017 04:43:51

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.082	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.153	04:38
Breath Volume	1.484	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.158	04:42
Breath Volume	2.046	04:42
Air Blank	0.000	04:42
Control Test	0.081	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:43

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/09/2017 05:16:22

Diagnostics Check		OK	05:02	
Air Blank		0.000	05:02	
Control Test		0.082	05:02	
Air Blank		0.000	05:03	
Subject Sample	# 1	0.041	05:06	Volume Not Met
Breath Volume		0.039	05:06	
Air Blank		0.000	05:07	
Air Blank		0.000	05:08	
Subject Sample	# 2	0.103	05:10	
Breath Volume		1.429	05:10	
Air Blank		0.000	05:10	
Air Blank		0.000	05:12	
Subject Sample	# 3	0.101	05:14	
Breath Volume		2.273	05:14	
Air Blank		0.000	05:15	
Control Test		0.081	05:15	
Air Blank		0.000	05:16	
Diagnostics Check		OK	05:16	

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 06/10/2017 04:09:29

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.081	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.207	04:05
Breath Volume		2.367	04:05
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.203	04:08
Breath Volume		2.101	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GONZALEZ HERRERA

DATE/TIME OF TEST: 06/18/2017 04:16:18

Diagnostics Check	OK	04:08
Air Blank	0.000	04:09
Control Test	0.082	04:09
Air Blank	0.000	04:10
Subject Sample # 1	0.230	04:11
Breath Volume	1.238	04:11
Air Blank	0.000	04:12
Air Blank	0.000	04:13
Subject Sample # 2	0.233	04:14
Breath Volume	1.515	04:14
Air Blank	0.000	04:15
Control Test	0.082	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SPEZIALE,ADAM
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 05/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: GALANTAI

DATE/TIME OF TEST: 06/18/2017 06:22:25

Diagnostics Check	OK	06:15
Air Blank	0.000	06:15
Control Test	0.081	06:16
Air Blank	0.000	06:16
Subject Sample # 1	0.112	06:17
Breath Volume	2.144	06:17
Air Blank	0.000	06:17
Air Blank	0.000	06:19
Subject Sample # 2	0.110	06:20
Breath Volume	1.929	06:20
Air Blank	0.000	06:21
Control Test	0.082	06:21
Air Blank	0.000	06:22
Diagnostics Check	OK	06:22

Data Download Date/Time: 06/19/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 17:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:37

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 06/02/2017 18:04:26

Diagnostics Check	OK	17:57
Air Blank	0.000	17:57
Control Test	0.080	17:57
Air Blank	0.000	17:58
Subject Sample # 1	0.302	17:59
Breath Volume	2.367	17:59
Air Blank	0.000	18:00
Air Blank	0.000	18:01
Subject Sample # 2	0.320	18:02
Breath Volume	2.914	18:02
Air Blank	0.000	18:03
Control Test	0.077	18:03
Air Blank	0.000	18:04
Diagnostics Check	OK	18:04

Data Download Date/Time: 06/26/2017 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/03/2017 01:03:34

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Subject Sample # 1	0.205	00:58	
Breath Volume	1.605	00:58	
Air Blank	0.000	00:59	
Air Blank	0.000	01:00	
Subject Sample # 2	0.192	01:02	
Breath Volume	2.031	01:02	
Air Blank		01:03	Purge Fail
Air Blank	0.000	01:03	

Data Download Date/Time: 06/26/2017 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 06/03/2017 01:13:47

Diagnostics Check	OK	01:05
Air Blank	0.000	01:05
Control Test	0.077	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.211	01:08
Breath Volume	1.195	01:08
Air Blank	0.000	01:08
Air Blank	0.000	01:10
Subject Sample # 2	0.217	01:11
Breath Volume	1.757	01:11
Air Blank	0.000	01:12
Control Test	0.076	01:13
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 06/26/2017 10:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 05/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 184081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 17:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:24

SUBJECT'S LAST NAME: BALLESTEROS

DATE/TIME OF TEST: 06/17/2017 17:51:43

Diagnostics Check	OK	17:45
Air Blank	0.000	17:45
Control Test	0.081	17:46
Air Blank	0.000	17:46
Subject Sample # 1	0.036	17:47
Breath Volume	2.894	17:47
Air Blank	0.000	17:48
Air Blank	0.000	17:49
Subject Sample # 2	0.034	17:50
Breath Volume	4.046	17:50
Air Blank	0.000	17:50
Control Test	0.079	17:51
Air Blank	0.000	17:51
Diagnostics Check	OK	17:51

Data Download Date/Time: 06/26/2017 10:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER, WILLIAM,
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: TUCKER, WILLIAM
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 05/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 06/11/2017 06:30:51

Diagnostics Check	OK	06:27	
Air Blank	0.000	06:28	
Control Test	0.083	06:28	
Air Blank	0.000	06:29	
Subject Sample	# 1	06:29	Subject Test Refused
Breath Volume	0.000	06:29	
Air Blank	0.000	06:29	
Control Test	0.083	06:30	
Air Blank	0.000	06:30	
Diagnostics Check	OK	06:30	

Data Download Date/Time: 06/18/2017 09:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SANDERS,LAUREN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 06/16/2017 04:03:11

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.081	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.248	03:58
Breath Volume		2.871	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.241	04:01
Breath Volume		1.296	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 06/20/2017 09:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MOODY,S
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: STRUTHERS

DATE/TIME OF TEST: 06/17/2017 02:26:09

Diagnostics Check	OK	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.105	02:21
Breath Volume	1.687	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.106	02:24
Breath Volume	2.644	02:24
Air Blank	0.000	02:25
Control Test	0.081	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 06/22/2017 10:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: MOODY,S
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: IVANCIK

DATE/TIME OF TEST: 06/18/2017 03:26:55

Diagnostics Check	OK	03:23	
Air Blank	0.000	03:24	
Control Test	0.080	03:24	
Air Blank	0.000	03:25	
Subject Sample	# 1	03:25	Subject Test Refused
Breath Volume	0.000	03:25	
Air Blank	0.000	03:26	
Control Test	0.080	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 06/22/2017 10:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARCIA,T
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: ZARECKI

DATE/TIME OF TEST: 06/28/2017 01:33:46

Diagnostics Check	OK	01:23
Air Blank	0.000	01:24
Control Test	0.080	01:24
Air Blank	0.000	01:24
Subject Sample # 1	0.157	01:27
Breath Volume	1.382	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.151	01:32
Breath Volume	1.511	01:32
Air Blank	0.000	01:32
Control Test	0.080	01:33
Air Blank	0.000	01:33
Diagnostics Check	OK	01:33

Data Download Date/Time: 07/28/2017 15:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: SANDERS,LAUREN
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 06/29/2017 03:04:03

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:56	
Subject Sample	# 1	03:00	No Sample Provided
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Air Blank	0.000	03:02	
Subject Sample	# 2	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:03	
Control Test	0.081	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:04	

Data Download Date/Time: 07/28/2017 15:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 06/02/2017 01:53:03

Diagnostics Check	OK	01:46
Air Blank	0.000	01:47
Control Test	0.085	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.181	01:48
Breath Volume	2.066	01:48
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.175	01:51
Breath Volume	1.710	01:51
Air Blank	0.000	01:52
Control Test	0.083	01:52
Air Blank	0.000	01:52
Diagnostics Check	OK	01:53

Data Download Date/Time: 06/21/2017 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CHAMBERLAND,PETER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: WALLINGFORD

DATE/TIME OF TEST: 06/02/2017 03:16:20

Diagnostics Check	OK	03:06
Air Blank	0.000	03:06
Control Test	0.085	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.223	03:07
Breath Volume	1.964	03:07
Air Blank	0.000	03:08
Air Blank	0.000	03:10
Subject Sample # 2	0.186	03:11
Breath Volume	1.394	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 3	0.186	03:14
Breath Volume	1.175	03:14
Air Blank	0.000	03:15
Control Test	0.082	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 06/21/2017 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: PETSOS,CHRISTINA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NORRIS

DATE/TIME OF TEST: 06/02/2017 03:32:50

Diagnostics Check	OK	03:26
Air Blank	0.000	03:26
Control Test	0.084	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.159	03:28
Breath Volume	1.679	03:28
Air Blank	0.000	03:29
Air Blank	0.000	03:30
Subject Sample # 2	0.168	03:31
Breath Volume	1.539	03:31
Air Blank	0.000	03:31
Control Test	0.082	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 06/21/2017 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLEASON,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GLEASON,WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: L300161931770

DATE/TIME OF TEST: 06/02/2017 23:52:46

Diagnostics Check	OK	23:45
Air Blank	0.000	23:46
Control Test	0.085	23:46
Air Blank	0.000	23:47
Subject Sample # 1	0.253	23:47
Breath Volume	2.746	23:47
Air Blank	0.000	23:48
Air Blank	0.000	23:50
Subject Sample # 2	0.241	23:51
Breath Volume	2.121	23:51
Air Blank	0.000	23:51
Control Test	0.084	23:52
Air Blank	0.000	23:52
Diagnostics Check	OK	23:52

Data Download Date/Time: 06/21/2017 20:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: JOHANSSON

DATE/TIME OF TEST: 06/03/2017 23:51:27

Diagnostics Check	OK	23:44
Air Blank	0.000	23:45
Control Test	0.085	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.242	23:46
Breath Volume	2.886	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.239	23:49
Breath Volume	2.554	23:49
Air Blank	0.000	23:50
Control Test	0.083	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 06/21/2017 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,ASHLEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAILEY,NALICIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ROSENBERGER

DATE/TIME OF TEST: 06/06/2017 02:11:49

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.086	02:11	Control Outside Tolerance
Air Blank	0.000	02:11	

Data Download Date/Time: 06/21/2017 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,ASHLEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAILEY,NALICIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ROSENBERGER

DATE/TIME OF TEST: 06/06/2017 02:15:35

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.086	02:14	Control Outside Tolerance
Air Blank	0.000	02:15	

Data Download Date/Time: 06/21/2017 20:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JENNINGS,ASHLEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BAILEY,NALICIA
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: ROSENBERGER

DATE/TIME OF TEST: 06/06/2017 02:23:30

Diagnostics Check	OK	02:16
Air Blank	0.000	02:17
Control Test	0.085	02:17
Air Blank	0.000	02:18
Subject Sample # 1	0.104	02:18
Breath Volume	2.734	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 2	0.103	02:21
Breath Volume	3.058	02:21
Air Blank	0.000	02:22
Control Test	0.085	02:22
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

Data Download Date/Time: 06/21/2017 20:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANTON,KEVIN,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STANTON,KEVIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 06/09/2017 20:10:16

Diagnostics Check	OK	20:08	
Air Blank	0.000	20:09	
Control Test	0.086	20:09	Control Outside Tolerance
Air Blank	0.000	20:10	

Data Download Date/Time: 06/21/2017 20:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001166

INSTRUMENT REGISTERED TO: BREVARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANTON,KEVIN,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STANTON,KEVIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:47

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 06/09/2017 20:18:18

Diagnostics Check	OK	20:12	
Air Blank	0.000	20:12	
Control Test	0.085	20:12	
Air Blank	0.000	20:13	
Subject Sample # 1	0.000	20:13	
Breath Volume	3.960	20:13	
Air Blank	0.000	20:14	
Air Blank	0.000	20:16	
Subject Sample # 2	0.000	20:16	
Breath Volume	3.382	20:16	
Air Blank	0.000	20:17	
Control Test	0.086	20:17	Control Outside Tolerance
Air Blank	0.000	20:18	

Data Download Date/Time: 06/21/2017 20:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARZATI,NATHAN,C

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: FARZATI,NATHAN

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 686331

DRY GAS EXPIRY DATE: 01/31/2020

VIOLATION CODE: COURT ORD

ARREST TIME: 16:15

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: SKEETE

DATE/TIME OF TEST: 06/01/2017 16:47:45

Diagnostics Check	OK	16:41
Air Blank	0.000	16:42
Control Test	0.081	16:42
Air Blank	0.000	16:43
Subject Sample # 1	0.000	16:43
Breath Volume	1.109	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.000	16:46
Breath Volume	1.222	16:46
Air Blank	0.000	16:46
Control Test	0.080	16:47
Air Blank	0.000	16:47
Diagnostics Check	OK	16:47

Data Download Date/Time: 06/27/2017 16:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,JIMMIE,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: WHEELER,CHAD

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 477627

DRY GAS EXPIRY DATE: 06/29/2019

VIOLATION CODE: DUI

ARREST TIME: 05:18

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 06/03/2017 05:53:02

Diagnostics Check	OK	05:45
Air Blank	0.000	05:46
Control Test	0.081	05:46
Air Blank	0.000	05:47
Subject Sample # 1	0.131	05:48
Breath Volume	2.257	05:48
Air Blank	0.000	05:48
Air Blank	0.000	05:50
Subject Sample # 2	0.142	05:51
Breath Volume	3.117	05:51
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:52
Diagnostics Check	OK	05:53

Data Download Date/Time: 06/27/2017 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLOOM,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BLOOM,NICHOLAS
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SKEETE

DATE/TIME OF TEST: 06/12/2017 17:37:00

Diagnostics Check	OK	17:31
Air Blank	0.000	17:31
Control Test	0.081	17:31
Air Blank	0.000	17:32
Subject Sample # 1	0.000	17:32
Breath Volume	1.996	17:32
Air Blank	0.000	17:33
Air Blank	0.000	17:35
Subject Sample # 2	0.000	17:35
Breath Volume	1.832	17:35
Air Blank	0.000	17:36
Control Test	0.081	17:36
Air Blank	0.000	17:36
Diagnostics Check	OK	17:37

Data Download Date/Time: 06/27/2017 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FARZATI,NATHAN,C

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: FARZATI,N

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 686331

DRY GAS EXPIRY DATE: 01/31/2020

VIOLATION CODE: COURT ORD

ARREST TIME: 16:30

GENDER: Female

OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: SKEETE

DATE/TIME OF TEST: 06/13/2017 17:18:11

Diagnostics Check	OK	17:12
Air Blank	0.000	17:12
Control Test	0.081	17:13
Air Blank	0.000	17:13
Subject Sample # 1	0.000	17:13
Breath Volume	1.484	17:13
Air Blank	0.000	17:14
Air Blank	0.000	17:16
Subject Sample # 2	0.000	17:16
Breath Volume	1.250	17:16
Air Blank	0.000	17:17
Control Test	0.082	17:17
Air Blank	0.000	17:18
Diagnostics Check	OK	17:18

Data Download Date/Time: 06/27/2017 16:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HARPER,JIMMIE,L

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: HARPER,JIMMIE

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 477627

DRY GAS EXPIRY DATE: 06/29/2019

VIOLATION CODE: COURT ORD

ARREST TIME: 18:20

GENDER: Female

OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: SKEETE

DATE/TIME OF TEST: 06/21/2017 18:52:40

Diagnostics Check	OK	18:46
Air Blank	0.000	18:47
Control Test	0.080	18:47
Air Blank	0.000	18:47
Subject Sample # 1	0.000	18:48
Breath Volume	2.421	18:48
Air Blank	0.000	18:48
Air Blank	0.000	18:50
Subject Sample # 2	0.000	18:51
Breath Volume	1.960	18:51
Air Blank	0.000	18:51
Control Test	0.080	18:52
Air Blank	0.000	18:52
Diagnostics Check	OK	18:52

Data Download Date/Time: 06/27/2017 16:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURYEAMATTHEW,
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: DURYEAMATTHEW
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 06/04/2017 01:35:24

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:33	Subject Test Refused
Breath Volume	0.089	01:33	
Air Blank	0.000	01:34	
Control Test	0.080	01:34	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 06/16/2017 05:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001177

INSTRUMENT REGISTERED TO: LADY LAKE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,SHANE,M
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: COUCH,HEATHER
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 06/12/2017 01:03:54

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.079	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1	0.131	00:56	Volume Not Met
Breath Volume		0.378	00:56	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2	0.116	01:02	Volume Not Met
Breath Volume		0.378	01:02	
Air Blank		0.000	01:02	
Control Test		0.080	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 06/16/2017 05:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, RONALD, M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO, JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 05/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/24/2019
VIOLATION CODE: BUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/21/2017 23:06:38

Diagnostics Check	OK	23:03	
Air Blank	0.000	23:04	
Control Test	0.081	23:04	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:05	Subject Test Refused
Breath Volume	0.000	23:05	
Air Blank	0.000	23:05	
Control Test	0.080	23:05	
Air Blank	0.000	23:06	
Diagnostics Check	OK	23:06	

Data Download Date/Time: 06/28/2017 11:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: HALPERN,ZAC
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 06/02/2017 00:10:55

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.078	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.284	00:02
Breath Volume		2.281	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.261	00:06
Breath Volume		1.582	00:06
Air Blank		0.000	00:07
Air Blank		0.000	00:08
Subject Sample	# 3	0.269	00:09
Breath Volume		1.738	00:09
Air Blank		0.000	00:10
Control Test		0.077	00:10
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 06/30/2017 14:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: MCGOWEN

DATE/TIME OF TEST: 06/14/2017 22:10:57

Diagnostics Check	OK	22:04
Air Blank	0.000	22:05
Control Test	0.079	22:05
Air Blank	0.000	22:06
Subject Sample # 1	0.117	22:06
Breath Volume	3.148	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:08
Subject Sample # 2	0.118	22:09
Breath Volume	2.847	22:09
Air Blank	0.000	22:10
Control Test	0.079	22:10
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 06/30/2017 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: GREGG

DATE/TIME OF TEST: 06/22/2017 22:27:53

Diagnostics Check	OK	22:25	
Air Blank	0.000	22:25	
Control Test	0.080	22:25	
Air Blank	0.000	22:26	
Subject Sample	# 1	22:26	Subject Test Refused
Breath Volume	0.000	22:26	
Air Blank	0.000	22:26	
Control Test	0.080	22:27	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 06/30/2017 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 06/25/2017 20:39:16

Diagnostics Check	OK	20:35	
Air Blank	0.000	20:36	
Control Test	0.079	20:36	
Air Blank	0.000	20:37	
Subject Sample	# 1	20:37	Subject Test Refused
Breath Volume	0.000	20:37	
Air Blank	0.000	20:38	
Control Test	0.080	20:38	
Air Blank	0.000	20:39	
Diagnostics Check	OK	20:39	

Data Download Date/Time: 06/30/2017 14:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001205

INSTRUMENT REGISTERED TO: NAS PENSACOLA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUGGER,DANIEL,J
BREATH TEST OPERATOR AGENCY: NAS Pensacola
ARRESTING OFFICER: MORRIS,CRAIG
ARRESTING OFFICER AGENCY: NAS Pensacola
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 07515080A4
DRY GAS EXPIRY DATE: 04/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: HILTS

DATE/TIME OF TEST: 06/05/2017 06:07:34

Diagnostics Check	OK	06:04	
Air Blank	0.000	06:04	
Control Test	0.079	06:05	
Air Blank	0.000	06:05	
Subject Sample	# 1	06:06	Subject Test Refused
Breath Volume	0.000	06:06	
Air Blank	0.000	06:06	
Control Test	0.080	06:06	
Air Blank	0.000	06:07	
Diagnostics Check	OK	06:07	

Data Download Date/Time: 06/28/2017 21:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MCLAIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/14/2017 02:26:45

Diagnostics Check	OK	02:20
Air Blank	0.000	02:21
Control Test	0.080	02:21
Air Blank	0.000	02:22
Subject Sample # 1	0.000	02:22
Breath Volume	2.468	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.000	02:25
Breath Volume	3.222	02:25
Air Blank	0.000	02:25
Control Test	0.082	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 06/14/2017 02:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,STEVEN,M
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: MCLAIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 83551
DRY GAS EXPIRY DATE: 10/02/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: RODRIGUEZ-ZUNIGA

DATE/TIME OF TEST: 06/28/2017 02:24:59

Diagnostics Check	OK	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.094	02:19
Breath Volume	1.687	02:19
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.100	02:23
Breath Volume	1.640	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/23/2017 23:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMYRNIOS,MANUEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMYRNIOS,MANUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 05/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/01/2017 03:54:47

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.038	03:50
Breath Volume		2.601	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.036	03:53
Breath Volume		2.636	03:53
Air Blank		0.000	03:53
Control Test		0.078	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 06/02/2017 09:45

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,J,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 06/06/2017 05:34:30

Diagnostics Check		OK	05:27
Air Blank		0.000	05:28
Control Test		0.078	05:28
Air Blank		0.000	05:29
Subject Sample	# 1	0.250	05:30
Breath Volume		2.492	05:30
Air Blank		0.000	05:30
Air Blank		0.000	05:32
Subject Sample	# 2	0.255	05:32
Breath Volume		2.031	05:32
Air Blank		0.000	05:33
Control Test		0.078	05:33
Air Blank		0.000	05:34
Diagnostics Check		OK	05:34

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CALHOUN,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 06/07/2017 03:19:14

Diagnostics Check	OK	03:12
Air Blank	0.000	03:13
Control Test	0.079	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.061	03:14
Breath Volume	1.757	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.060	03:17
Breath Volume	1.695	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRIS, MICHAEL, T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FARRIS, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CHARLESTIN

DATE/TIME OF TEST: 06/08/2017 02:08:39

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:06	
Control Test	0.079	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.079	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STROEDECKE,JUSTIN,J
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CARVALHO,ASHLEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: JENSEN

DATE/TIME OF TEST: 06/08/2017 23:37:34

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.079	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.050	23:33
Breath Volume	1.289	23:33
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.051	23:36
Breath Volume	1.332	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: ROBINSON,T
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: OBERT

DATE/TIME OF TEST: 06/10/2017 00:57:38

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.078	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.244	00:52
Breath Volume		1.640	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.250	00:56
Breath Volume		2.382	00:56
Air Blank		0.000	00:56
Control Test		0.079	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,RYAN,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: MALDONADO,RYAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PALMER

DATE/TIME OF TEST: 06/10/2017 03:45:47

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.079	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.000	03:41
Breath Volume	2.121	03:41
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.000	03:44
Breath Volume	2.785	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,J,B
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: FINE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:34

SUBJECT'S LAST NAME: VANAMAN

DATE/TIME OF TEST: 06/10/2017 16:58:10

Diagnostics Check	OK	16:55	
Air Blank	0.000	16:55	
Control Test	0.079	16:55	
Air Blank	0.000	16:56	
Subject Sample	# 1	16:56	Subject Test Refused
Breath Volume	0.000	16:56	
Air Blank	0.000	16:57	
Control Test	0.079	16:57	
Air Blank	0.000	16:58	
Diagnostics Check	OK	16:58	

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARSON,JAMES,A
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SABOE,EVAN
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: DEPPEN

DATE/TIME OF TEST: 06/10/2017 20:34:39

Diagnostics Check	OK	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:29
Air Blank	0.000	20:29
Subject Sample # 1	0.000	20:30
Breath Volume	2.996	20:30
Air Blank	0.000	20:30
Air Blank	0.000	20:32
Subject Sample # 2	0.000	20:33
Breath Volume	3.218	20:33
Air Blank	0.000	20:33
Control Test	0.078	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 06/12/2017 00:23:16

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.222	00:18
Breath Volume		2.640	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.225	00:21
Breath Volume		2.496	00:21
Air Blank		0.000	00:22
Control Test		0.079	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 06/12/2017 03:46:43

Diagnostics Check	OK	03:40
Air Blank	0.000	03:40
Control Test	0.079	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.162	03:42
Breath Volume	1.425	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.155	03:45
Breath Volume	1.691	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ALMAGUER

DATE/TIME OF TEST: 06/12/2017 18:32:35

Diagnostics Check	OK	18:26
Air Blank	0.000	18:26
Control Test	0.079	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.000	18:28
Breath Volume	1.320	18:28
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.000	18:31
Breath Volume	1.187	18:31
Air Blank	0.000	18:31
Control Test	0.079	18:31
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: CARRUTHERS

DATE/TIME OF TEST: 06/13/2017 22:43:02

Diagnostics Check	OK	22:36
Air Blank	0.000	22:36
Control Test	0.078	22:37
Air Blank	0.000	22:37
Subject Sample # 1	0.103	22:38
Breath Volume	1.957	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.101	22:41
Breath Volume	1.500	22:41
Air Blank	0.000	22:42
Control Test	0.078	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:43

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDMOND,COREY,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EDMOND,COREY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BURNS

DATE/TIME OF TEST: 06/14/2017 02:01:03

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:45	
Control Test		0.079	01:45	
Air Blank		0.000	01:46	
Subject Sample	# 1	0.182	01:48	
Breath Volume		1.398	01:48	
Air Blank		0.000	01:48	
Air Blank		0.000	01:50	
Subject Sample	# 2	0.160	01:53	Volume Not Met
Breath Volume		0.085	01:53	
Air Blank		0.000	01:54	
Air Blank		0.000	01:56	
Subject Sample	# 3	0.157	01:59	Volume Not Met
Breath Volume		0.968	01:59	
Air Blank		0.000	02:00	
Control Test		0.075	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:01	

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CALHOUN,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: WELLER

DATE/TIME OF TEST: 06/14/2017 23:00:21

Diagnostics Check	OK	22:51	
Air Blank	0.000	22:52	
Control Test	0.078	22:52	
Air Blank	0.000	22:52	
Subject Sample	# 1 0.195	22:54	
Breath Volume	1.628	22:54	
Air Blank	0.000	22:54	
Air Blank	0.000	22:56	
Subject Sample	# 2	22:59	Subject Test Refused
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Control Test	0.078	22:59	
Air Blank	0.000	23:00	
Diagnostics Check	OK	23:00	

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,J,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: BORTOLAMETTI

DATE/TIME OF TEST: 06/16/2017 02:37:07

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.079	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.078	02:32
Breath Volume	1.816	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:34
Subject Sample # 2	0.082	02:35
Breath Volume	2.109	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK, JOSHUA, P
BREATH TEST OPERATOR AGENCY: Lee Port Authority
ARRESTING OFFICER: ORIHUELA, JENNIFER
ARRESTING OFFICER AGENCY: Lee Port Authority
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: COOTS

DATE/TIME OF TEST: 06/17/2017 02:26:33

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.079	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.147	02:22
Breath Volume		1.125	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.157	02:25
Breath Volume		1.683	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRIS, MICHAEL, T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: REYES, GUSTAVO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:07

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 06/17/2017 05:35:16

Diagnostics Check		OK	05:27
Air Blank		0.000	05:28
Control Test		0.079	05:28
Air Blank		0.000	05:29
Subject Sample	# 1	0.209	05:30
Breath Volume		2.656	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.206	05:33
Breath Volume		1.562	05:33
Air Blank		0.000	05:34
Control Test		0.078	05:34
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BERG,BARBARA
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: HERCULES-ZETO

DATE/TIME OF TEST: 06/17/2017 20:35:05

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.079	20:28
Air Blank	0.000	20:28
Subject Sample # 1	0.118	20:29
Breath Volume	1.136	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.106	20:33
Breath Volume	1.636	20:33
Air Blank	0.000	20:34
Control Test	0.079	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:35

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARRIS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: RODRIGUEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 06/17/2017 22:23:43

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.088	22:19
Breath Volume	2.019	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.088	22:22
Breath Volume	2.136	22:22
Air Blank	0.000	22:22
Control Test	0.077	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: KITTLESON,MYLES
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:41

SUBJECT'S LAST NAME: RYBAK

DATE/TIME OF TEST: 06/18/2017 02:24:52

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:20
Breath Volume	2.796	02:20
Air Blank	0.000	02:21
Air Blank	0.000	02:23
Subject Sample # 2	0.000	02:23
Breath Volume	2.183	02:23
Air Blank	0.000	02:23
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CROSS,K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/20/2017 20:26:32

Diagnostics Check	OK	20:19
Air Blank	0.000	20:19
Control Test	0.078	20:20
Air Blank	0.000	20:20
Subject Sample # 1	0.024	20:21
Breath Volume	3.957	20:21
Air Blank	0.000	20:22
Air Blank	0.000	20:24
Subject Sample # 2	0.020	20:25
Breath Volume	2.398	20:25
Air Blank	0.000	20:25
Control Test	0.078	20:25
Air Blank	0.000	20:26
Diagnostics Check	OK	20:26

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: DOMINGO FRANCISCO

DATE/TIME OF TEST: 06/22/2017 23:08:25

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.078	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.208	23:03
Breath Volume	1.457	23:03
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.219	23:06
Breath Volume	1.445	23:06
Air Blank	0.000	23:07
Control Test	0.078	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVENDANO,JUAN,M
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CAIVANO,STACEY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SANTISTEBAN

DATE/TIME OF TEST: 06/23/2017 00:50:23

Diagnostics Check	OK	00:42
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.233	00:45
Breath Volume	1.582	00:45
Air Blank	0.000	00:46
Air Blank	0.000	00:47
Subject Sample # 2	0.215	00:48
Breath Volume	1.292	00:48
Air Blank	0.000	00:49
Control Test	0.077	00:49
Air Blank	0.000	00:50
Diagnostics Check	OK	00:50

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINK,C
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SCHAAF

DATE/TIME OF TEST: 06/23/2017 04:01:59

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.078	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.136	03:54
Breath Volume		3.585	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.113	03:57
Breath Volume		1.289	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 3	0.119	04:00
Breath Volume		1.429	04:00
Air Blank		0.000	04:01
Control Test		0.078	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WILLIAMS,T
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: TZOY SARAT

DATE/TIME OF TEST: 06/23/2017 21:12:08

Diagnostics Check	OK	21:05
Air Blank	0.000	21:06
Control Test	0.078	21:06
Air Blank	0.000	21:07
Subject Sample # 1	0.193	21:07
Breath Volume	1.945	21:07
Air Blank	0.000	21:08
Air Blank	0.000	21:10
Subject Sample # 2	0.210	21:10
Breath Volume	1.820	21:10
Air Blank	0.000	21:11
Control Test	0.078	21:11
Air Blank	0.000	21:12
Diagnostics Check	OK	21:12

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: RONGISH,S
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: SIDOCK

DATE/TIME OF TEST: 06/24/2017 01:28:46

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.078	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.118	01:24
Breath Volume	2.640	01:24
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.120	01:27
Breath Volume	2.742	01:27
Air Blank	0.000	01:27
Control Test	0.078	01:28
Air Blank	0.000	01:28
Diagnostics Check	OK	01:28

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WITT,K
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: DUDLEY

DATE/TIME OF TEST: 06/24/2017 01:42:49

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.173	01:38
Breath Volume		3.515	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.158	01:41
Breath Volume		1.558	01:41
Air Blank		0.000	01:41
Control Test		0.078	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: RAMBARRAN

DATE/TIME OF TEST: 06/26/2017 04:10:23

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.078	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.141	04:05
Breath Volume	2.054	04:05
Air Blank	0.000	04:06
Air Blank	0.000	04:08
Subject Sample # 2	0.137	04:08
Breath Volume	2.535	04:08
Air Blank	0.000	04:09
Control Test	0.079	04:09
Air Blank	0.000	04:10
Diagnostics Check	OK	04:10

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR,J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: COGHLAN

DATE/TIME OF TEST: 06/30/2017 03:35:55

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.219	03:31
Breath Volume	3.171	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.208	03:34
Breath Volume	4.402	03:34
Air Blank	0.000	03:35
Control Test	0.078	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 07/26/2017 08:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYERS,F,E
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: TROMETER,S
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PAGAN

DATE/TIME OF TEST: 06/30/2017 03:57:27

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.137	03:52
Breath Volume		2.296	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.138	03:55
Breath Volume		2.175	03:55
Air Blank		0.000	03:56
Control Test		0.078	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 07/26/2017 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALL,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: WALL,CHRISTOPHER
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 05/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 06/13/2017 20:29:48

Diagnostics Check	OK	20:22
Air Blank	0.000	20:23
Control Test	0.078	20:23
Air Blank	0.000	20:23
Subject Sample # 1	0.219	20:25
Breath Volume	3.441	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.223	20:28
Breath Volume	3.269	20:28
Air Blank	0.000	20:28
Control Test	0.077	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 06/28/2017 08:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: SCHAEFERLE

DATE/TIME OF TEST: 06/04/2017 00:56:38

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.080	00:54	
Air Blank	0.000	00:54	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank	0.000	00:55	
Control Test	0.080	00:56	
Air Blank	0.000	00:56	
Diagnostics Check	OK	00:56	

Data Download Date/Time: 06/29/2017 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SABUNCU

DATE/TIME OF TEST: 06/19/2017 00:29:15

Diagnostics Check	OK	00:22
Air Blank	0.000	00:22
Control Test	0.079	00:23
Air Blank	0.000	00:23
Subject Sample # 1	0.052	00:24
Breath Volume	3.500	00:24
Air Blank	0.000	00:24
Air Blank	0.000	00:26
Subject Sample # 2	0.052	00:27
Breath Volume	2.597	00:27
Air Blank	0.000	00:28
Control Test	0.080	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 06/29/2017 12:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SECORD,RYAN,T
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 05/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: KAHN

DATE/TIME OF TEST: 06/28/2017 15:32:03

Diagnostics Check	OK	15:23
Air Blank	0.000	15:23
Control Test	0.080	15:23
Air Blank	0.000	15:24
Subject Sample # 1	0.293	15:24
Breath Volume	1.726	15:24
Air Blank	0.000	15:25
Air Blank	0.000	15:27
Subject Sample # 2	0.314	15:27
Breath Volume	2.679	15:27
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 3	0.307	15:30
Breath Volume	2.226	15:30
Air Blank	0.000	15:31
Control Test	0.079	15:31
Air Blank	0.000	15:31
Diagnostics Check	OK	15:32

Data Download Date/Time: 06/29/2017 12:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001242

INSTRUMENT REGISTERED TO: DAYTONA BEACH SHORES
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTELLANO,VINCENT,J
BREATH TEST OPERATOR AGENCY: Daytona Beach Shores
ARRESTING OFFICER: CASTELLANO,VINCENT
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/08/2017 05:44:31

Diagnostics Check	OK	05:38
Air Blank	0.000	05:38
Control Test	0.078	05:38
Air Blank	0.000	05:39
Subject Sample # 1	0.000	05:40
Breath Volume	2.382	05:40
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.000	05:43
Breath Volume	1.804	05:43
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:44
Diagnostics Check	OK	05:44

Data Download Date/Time: 10/12/2017 14:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: WETHERELL, WILLIAM
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 06/04/2017 03:43:14

Diagnostics Check	OK	03:40	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:41	
Subject Sample	# 1	03:41	Subject Test Refused
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Control Test	0.080	03:42	
Air Blank	0.000	03:43	
Diagnostics Check	OK	03:43	

Data Download Date/Time: 06/25/2017 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER, MATTHEW, R
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: GEIGER, CHARLES
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 21:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: BITTLE

DATE/TIME OF TEST: 06/07/2017 22:33:28

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.080	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.074	22:27
Breath Volume	1.464	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.081	22:31
Breath Volume	1.847	22:31
Air Blank	0.000	22:32
Control Test	0.080	22:32
Air Blank	0.000	22:33
Diagnostics Check	OK	22:33

Data Download Date/Time: 06/25/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLEYHAN,JARETT,G
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: SPURLOCK,ANDREW
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 05/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: TOOLE

DATE/TIME OF TEST: 06/20/2017 00:50:31

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.234	00:45
Breath Volume		1.230	00:45
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.250	00:48
Breath Volume		1.718	00:48
Air Blank		0.000	00:49
Control Test		0.078	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 06/25/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRABER,KIMBERLY,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: GILL,WILL
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31311080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:24

SUBJECT'S LAST NAME: SMILEY

DATE/TIME OF TEST: 06/15/2017 09:02:43

Diagnostics Check	OK	08:56
Air Blank	0.000	08:57
Control Test	0.077	08:57
Air Blank	0.000	08:57
Subject Sample # 1	0.150	08:58
Breath Volume	2.621	08:58
Air Blank	0.000	08:58
Air Blank	0.000	09:00
Subject Sample # 2	0.147	09:01
Breath Volume	1.953	09:01
Air Blank	0.000	09:01
Control Test	0.078	09:02
Air Blank	0.000	09:02
Diagnostics Check	OK	09:02

Data Download Date/Time: 06/30/2017 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONRAD,TROOPER C,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: CONRAD,TROOPER C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 06/25/2017 01:11:05

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.308	01:06
Breath Volume		1.769	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.317	01:09
Breath Volume		1.660	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 06/30/2017 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BLICHER

DATE/TIME OF TEST: 06/02/2017 01:55:20

Diagnostics Check	OK	01:48
Air Blank	0.000	01:49
Control Test	0.080	01:49
Air Blank	0.000	01:49
Subject Sample # 1	0.052	01:50
Breath Volume	1.601	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:52
Subject Sample # 2	0.049	01:53
Breath Volume	1.394	01:53
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DALTON,RYAN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: AURBIN

DATE/TIME OF TEST: 06/02/2017 05:10:36

Diagnostics Check	OK	05:02
Air Blank	0.000	05:03
Control Test	0.081	05:03
Air Blank	0.000	05:04
Subject Sample # 1	0.186	05:05
Breath Volume	1.671	05:05
Air Blank	0.000	05:05
Air Blank	0.000	05:07
Subject Sample # 2	0.176	05:08
Breath Volume	1.339	05:08
Air Blank	0.000	05:09
Control Test	0.080	05:09
Air Blank	0.000	05:10
Diagnostics Check	OK	05:10

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SEAN,M
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PALANT,JASON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: NEWMAN

DATE/TIME OF TEST: 06/03/2017 18:58:01

Diagnostics Check	OK	18:51
Air Blank	0.000	18:51
Control Test	0.081	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.210	18:53
Breath Volume	2.148	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:55
Subject Sample # 2	0.191	18:56
Breath Volume	1.332	18:56
Air Blank	0.000	18:57
Control Test	0.081	18:57
Air Blank	0.000	18:57
Diagnostics Check	OK	18:58

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:34

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 06/03/2017 19:12:14

Diagnostics Check	OK	19:05
Air Blank	0.000	19:06
Control Test	0.081	19:06
Air Blank	0.000	19:07
Subject Sample # 1	0.041	19:07
Breath Volume	1.882	19:07
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.041	19:10
Breath Volume	2.093	19:10
Air Blank	0.000	19:11
Control Test	0.082	19:11
Air Blank	0.000	19:12
Diagnostics Check	OK	19:12

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PALANT,JASON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: COTTRELL JR

DATE/TIME OF TEST: 06/03/2017 23:51:56

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.081	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.138	23:47
Breath Volume	2.265	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.137	23:50
Breath Volume	1.632	23:50
Air Blank	0.000	23:50
Control Test	0.080	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: STOGNER,CODY
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: ROMERO SEMINO

DATE/TIME OF TEST: 06/04/2017 00:08:23

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:01
Air Blank	0.000	00:02
Subject Sample # 1	0.172	00:03
Breath Volume	1.187	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:06
Subject Sample # 2	0.171	00:06
Breath Volume	1.464	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:08
Diagnostics Check	OK	00:08

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JN LOUIS,FRANCKY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 06/04/2017 01:03:28

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.143	00:58
Breath Volume	1.769	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.139	01:01
Breath Volume	2.195	01:01
Air Blank	0.000	01:02
Control Test	0.080	01:02
Air Blank	0.000	01:03
Diagnostics Check	OK	01:03

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: GONZALEZ INFANTE

DATE/TIME OF TEST: 06/04/2017 02:54:29

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.081	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.156	02:49
Breath Volume	3.144	02:49
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.162	02:52
Breath Volume	1.292	02:52
Air Blank	0.000	02:53
Control Test	0.081	02:53
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KELLY,TAYLOR
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: BRENIC

DATE/TIME OF TEST: 06/04/2017 22:17:07

Diagnostics Check	OK	22:04	
Air Blank	0.000	22:04	
Control Test	0.081	22:04	
Air Blank	0.000	22:05	
Subject Sample	# 1 0.127	22:06	
Breath Volume	1.121	22:06	
Air Blank	0.000	22:07	
Air Blank	0.000	22:09	
Subject Sample	# 2 0.115	22:12	Volume Not Met
Breath Volume	0.890	22:12	
Air Blank	0.000	22:13	
Air Blank	0.000	22:15	
Subject Sample	# 3 0.117	22:15	
Breath Volume	1.285	22:15	
Air Blank	0.000	22:16	
Control Test	0.078	22:16	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICIENZO,FRANCA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: PIERCE

DATE/TIME OF TEST: 06/08/2017 17:48:22

Diagnostics Check	OK	17:41
Air Blank	0.000	17:42
Control Test	0.081	17:42
Air Blank	0.000	17:42
Subject Sample # 1	0.095	17:43
Breath Volume	2.085	17:43
Air Blank	0.000	17:44
Air Blank	0.000	17:46
Subject Sample # 2	0.089	17:46
Breath Volume	2.042	17:46
Air Blank	0.000	17:47
Control Test	0.081	17:47
Air Blank	0.000	17:48
Diagnostics Check	OK	17:48

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WASHINGTON,YVETTE
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 06/09/2017 04:58:04

Diagnostics Check		OK	04:45	
Air Blank		0.000	04:46	
Control Test		0.081	04:46	
Air Blank		0.000	04:46	
Subject Sample	# 1	0.000	04:50	Volume Not Met
Breath Volume		0.699	04:50	
Air Blank		0.000	04:50	
Air Blank		0.000	04:52	
Subject Sample	# 2	0.000	04:53	
Breath Volume		1.191	04:53	
Air Blank		0.000	04:54	
Air Blank		0.000	04:56	
Subject Sample	# 3	0.000	04:56	
Breath Volume		1.632	04:56	
Air Blank		0.000	04:57	
Control Test		0.081	04:57	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:58	

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CHARLES,WOODNEL
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PACHECO HERNANDEZ

DATE/TIME OF TEST: 06/10/2017 03:32:43

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.081	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1	0.191	03:20	Volume Not Met
Breath Volume		0.679	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2	0.194	03:25	
Breath Volume		1.328	03:25	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 3	0.191	03:30	
Breath Volume		1.652	03:30	
Air Blank		0.000	03:31	
Control Test		0.079	03:32	
Air Blank		0.000	03:32	
Diagnostics Check		OK	03:32	

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 05/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: BARREDO

DATE/TIME OF TEST: 06/17/2017 23:59:52

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.081	23:53
Air Blank	0.000	23:54
Subject Sample # 1	0.162	23:55
Breath Volume	1.117	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.166	23:58
Breath Volume	1.558	23:58
Air Blank	0.000	23:58
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 06/19/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GUTIERREZ,KRISTINA
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: COLL

DATE/TIME OF TEST: 06/05/2017 21:22:47

Diagnostics Check	OK	21:17	
Air Blank	0.000	21:17	
Control Test	0.081	21:17	
Air Blank	0.000	21:18	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:21	
Control Test	0.079	21:22	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 06/07/2017 02:50:17

Diagnostics Check	OK	02:46	
Air Blank	0.000	02:47	
Control Test	0.081	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1	02:48	Subject Test Refused
Breath Volume	0.000	02:48	
Air Blank	0.000	02:49	
Control Test	0.080	02:49	
Air Blank	0.000	02:50	
Diagnostics Check	OK	02:50	

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 06/08/2017 01:39:54

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.081	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:38	Subject Test Refused
Breath Volume	0.000	01:38	
Air Blank	0.000	01:39	
Control Test	0.080	01:39	
Air Blank	0.000	01:39	
Diagnostics Check	OK	01:39	

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 06/10/2017 02:37:25

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.153	02:33
Breath Volume	2.609	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.148	02:35
Breath Volume	2.136	02:35
Air Blank	0.000	02:36
Control Test	0.080	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:24

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 06/11/2017 07:47:44

Diagnostics Check	OK	07:44	
Air Blank	0.000	07:45	
Control Test	0.080	07:45	
Air Blank	0.000	07:46	
Subject Sample	# 1	07:46	Subject Test Refused
Breath Volume	0.000	07:46	
Air Blank	0.000	07:46	
Control Test	0.079	07:47	
Air Blank	0.000	07:47	
Diagnostics Check	OK	07:47	

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: BETANCOURT

DATE/TIME OF TEST: 06/13/2017 18:50:01

Diagnostics Check	OK	18:47	
Air Blank	0.000	18:47	
Control Test	0.079	18:47	
Air Blank	0.000	18:48	
Subject Sample	# 1	18:48	RFI Detect
Breath Volume	0.000	18:48	
Air Blank	0.000	18:49	
Control Test	0.079	18:49	
Air Blank	0.000	18:49	
Diagnostics Check	OK	18:50	

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: BETANCOURT

DATE/TIME OF TEST: 06/13/2017 18:58:12

Diagnostics Check	OK	18:51
Air Blank	0.000	18:52
Control Test	0.079	18:52
Air Blank	0.000	18:52
Subject Sample # 1	0.159	18:53
Breath Volume	2.875	18:53
Air Blank	0.000	18:54
Air Blank	0.000	18:56
Subject Sample # 2	0.160	18:56
Breath Volume	2.878	18:56
Air Blank	0.000	18:57
Control Test	0.080	18:57
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: BARUFALDI

DATE/TIME OF TEST: 06/15/2017 02:08:42

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:59	
Control Test	0.081	01:59	
Air Blank	0.000	02:00	
Subject Sample	# 1	02:03	No Sample Provided
Breath Volume	0.000	02:03	
Air Blank	0.000	02:03	
Air Blank	0.000	02:05	
Subject Sample	# 2	02:07	Subject Test Refused
Breath Volume	0.000	02:07	
Air Blank	0.000	02:07	
Control Test	0.080	02:08	
Air Blank	0.000	02:08	
Diagnostics Check	OK	02:08	

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: ORBETA,JORGE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: AIKEN

DATE/TIME OF TEST: 06/16/2017 02:40:46

Diagnostics Check	OK	02:33
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.082	02:35
Breath Volume	2.746	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.083	02:39
Breath Volume	3.500	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 06/17/2017 16:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORBETA,JORGE,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VIERGUTZ,ERIC
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CHIRILLO

DATE/TIME OF TEST: 06/18/2017 01:52:01

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.116	01:46
Breath Volume	1.175	01:46
Air Blank	0.000	01:47
Air Blank	0.000	01:49
Subject Sample # 2	0.123	01:50
Breath Volume	2.140	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:51
Diagnostics Check	OK	01:52

Data Download Date/Time: 07/19/2017 13:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: MOLINA,IGNACIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: HESTER

DATE/TIME OF TEST: 06/21/2017 04:11:57

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.171	04:03	Volume Not Met
Breath Volume		0.222	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 2	0.179	04:07	
Breath Volume		1.386	04:07	
Air Blank		0.000	04:07	
Air Blank		0.000	04:09	
Subject Sample	# 3	0.165	04:10	
Breath Volume		1.230	04:10	
Air Blank		0.000	04:11	
Control Test		0.079	04:11	
Air Blank		0.000	04:11	
Diagnostics Check		OK	04:11	

Data Download Date/Time: 07/19/2017 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,GIL,A
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: GONZALEZ,GIL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 06/21/2017 06:14:00

Diagnostics Check	OK	06:08
Air Blank	0.000	06:08
Control Test	0.080	06:08
Air Blank	0.000	06:09
Subject Sample # 1	0.000	06:09
Breath Volume	3.867	06:09
Air Blank	0.000	06:10
Air Blank	0.000	06:12
Subject Sample # 2	0.000	06:12
Breath Volume	2.511	06:12
Air Blank	0.000	06:13
Control Test	0.079	06:13
Air Blank	0.000	06:13
Diagnostics Check	OK	06:14

Data Download Date/Time: 07/19/2017 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,JON,B
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HUFF,JON
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: WILLIAMS JR

DATE/TIME OF TEST: 06/24/2017 21:02:52

Diagnostics Check	OK	20:58	
Air Blank	0.000	20:59	
Control Test	0.080	20:59	
Air Blank	0.000	21:00	
Subject Sample	# 1	21:01	Subject Test Refused
Breath Volume	0.000	21:01	
Air Blank	0.000	21:01	
Control Test	0.080	21:02	
Air Blank	0.000	21:02	
Diagnostics Check	OK	21:02	

Data Download Date/Time: 07/19/2017 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MATTSON,CHRIS,J
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MATTSON,CHRIS
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: BUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ALBURY

DATE/TIME OF TEST: 06/24/2017 23:11:33

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:09	
Control Test	0.080	23:09	
Air Blank	0.000	23:10	
Subject Sample	# 1	23:10	Subject Test Refused
Breath Volume	0.000	23:10	
Air Blank	0.000	23:10	
Control Test	0.080	23:10	
Air Blank	0.000	23:11	
Diagnostics Check	OK	23:11	

Data Download Date/Time: 07/19/2017 13:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CABRAL

DATE/TIME OF TEST: 06/01/2017 01:59:12

Diagnostics Check	OK	01:51
Air Blank	0.000	01:51
Control Test	0.082	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.058	01:54
Breath Volume	2.917	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.058	01:57
Breath Volume	2.804	01:57
Air Blank	0.000	01:58
Control Test	0.081	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 06/08/2017 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MOREIRA-PONSE

DATE/TIME OF TEST: 06/01/2017 21:44:08

Diagnostics Check	OK	21:37
Air Blank	0.000	21:38
Control Test	0.082	21:38
Air Blank	0.000	21:39
Subject Sample # 1	0.170	21:39
Breath Volume	1.921	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:41
Subject Sample # 2	0.184	21:42
Breath Volume	3.566	21:42
Air Blank	0.000	21:43
Control Test	0.082	21:43
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 06/08/2017 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: CUEVASNIEVES

DATE/TIME OF TEST: 06/03/2017 02:47:08

Diagnostics Check	OK	02:40
Air Blank	0.000	02:41
Control Test	0.082	02:41
Air Blank	0.000	02:42
Subject Sample # 1	0.156	02:42
Breath Volume	2.753	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.163	02:45
Breath Volume	2.593	02:45
Air Blank	0.000	02:46
Control Test	0.081	02:46
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 06/08/2017 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SALDANA DURAND

DATE/TIME OF TEST: 06/04/2017 03:40:04

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.083	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.152	03:33
Breath Volume	1.574	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.159	03:38
Breath Volume	3.796	03:38
Air Blank	0.000	03:39
Control Test	0.082	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:40

Data Download Date/Time: 06/08/2017 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 06/04/2017 04:35:32

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.081	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.198	04:30
Breath Volume	2.214	04:30
Air Blank	0.000	04:31
Air Blank	0.000	04:33
Subject Sample # 2	0.204	04:33
Breath Volume	2.078	04:33
Air Blank	0.000	04:34
Control Test	0.080	04:34
Air Blank	0.000	04:35
Diagnostics Check	OK	04:35

Data Download Date/Time: 06/08/2017 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DIB, JAMIL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 06/04/2017 09:43:58

Diagnostics Check	OK	09:35
Air Blank	0.000	09:36
Control Test	0.082	09:36
Air Blank	0.000	09:36
Subject Sample # 1	0.125	09:39
Breath Volume	1.265	09:39
Air Blank	0.000	09:39
Air Blank	0.000	09:41
Subject Sample # 2	0.130	09:42
Breath Volume	1.492	09:42
Air Blank	0.000	09:43
Control Test	0.080	09:43
Air Blank	0.000	09:43
Diagnostics Check	OK	09:43

Data Download Date/Time: 06/08/2017 14:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 06/09/2017 01:48:40

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.083	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.093	01:43
Breath Volume	1.285	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.101	01:47
Breath Volume	2.230	01:47
Air Blank	0.000	01:47
Control Test	0.082	01:48
Air Blank	0.000	01:48
Diagnostics Check	OK	01:48

Data Download Date/Time: 06/14/2017 16:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: SELDERS

DATE/TIME OF TEST: 06/10/2017 05:15:41

Diagnostics Check	OK	05:09
Air Blank	0.000	05:10
Control Test	0.083	05:10
Air Blank	0.000	05:10
Subject Sample # 1	0.000	05:11
Breath Volume	2.460	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.000	05:14
Breath Volume	3.285	05:14
Air Blank	0.000	05:14
Control Test	0.080	05:15
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 06/14/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE,PETER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: MATHER

DATE/TIME OF TEST: 06/11/2017 04:39:02

Diagnostics Check	OK	04:31
Air Blank	0.000	04:32
Control Test	0.082	04:32
Air Blank	0.000	04:33
Subject Sample # 1	0.157	04:33
Breath Volume	1.789	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.159	04:37
Breath Volume	3.046	04:37
Air Blank	0.000	04:38
Control Test	0.081	04:38
Air Blank	0.000	04:38
Diagnostics Check	OK	04:39

Data Download Date/Time: 06/14/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DIAS

DATE/TIME OF TEST: 06/14/2017 04:39:00

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:33	
Control Test	0.082	04:33	
Air Blank	0.000	04:34	
Subject Sample	# 1 0.121	04:37	Volume Not Met
Breath Volume	0.769	04:37	
Air Blank		04:38	Purge Fail
Air Blank		04:39	Purge Fail

Data Download Date/Time: 06/14/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001255

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK,S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DIAS

DATE/TIME OF TEST: 06/14/2017 04:41:31

Diagnostics Check	OK	04:40	
Air Blank	0.000	04:40	
Control Test	0.033	04:40	Interferent Detect
Air Blank	0.000	04:41	

Data Download Date/Time: 06/14/2017 16:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: ALEXANDER

DATE/TIME OF TEST: 06/01/2017 04:17:08

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.139	04:12
Breath Volume		2.210	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.141	04:15
Breath Volume		1.464	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DANJOU,SCOTT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CISNERO

DATE/TIME OF TEST: 06/02/2017 02:36:11

Diagnostics Check	OK	02:30
Air Blank	0.000	02:30
Control Test	0.081	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.202	02:31
Breath Volume	1.640	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.190	02:34
Breath Volume	1.265	02:34
Air Blank	0.000	02:35
Control Test	0.081	02:35
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 06/02/2017 03:38:24

Diagnostics Check	OK	03:31
Air Blank	0.000	03:32
Control Test	0.082	03:32
Air Blank	0.000	03:33
Subject Sample # 1	0.106	03:33
Breath Volume	2.433	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.105	03:36
Breath Volume	2.523	03:36
Air Blank	0.000	03:37
Control Test	0.081	03:37
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DIMITRY

DATE/TIME OF TEST: 06/02/2017 23:58:18

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.135	23:54
Breath Volume	3.433	23:54
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.129	23:56
Breath Volume	2.375	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SHAPIRO

DATE/TIME OF TEST: 06/03/2017 01:52:19

Diagnostics Check	OK	01:46
Air Blank	0.000	01:46
Control Test	0.081	01:47
Air Blank	0.000	01:47
Subject Sample # 1	0.151	01:48
Breath Volume	1.964	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.152	01:50
Breath Volume	2.046	01:50
Air Blank	0.000	01:51
Control Test	0.081	01:51
Air Blank	0.000	01:52
Diagnostics Check	OK	01:52

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,AGUSTIN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: GAYNUS

DATE/TIME OF TEST: 06/03/2017 03:29:54

Diagnostics Check	OK	03:18
Air Blank	0.000	03:18
Control Test	0.081	03:19
Air Blank	0.000	03:19
Subject Sample # 1	0.160	03:22
Breath Volume	1.214	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.171	03:28
Breath Volume	2.089	03:28
Air Blank	0.000	03:29
Control Test	0.080	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LATTIMORE

DATE/TIME OF TEST: 06/03/2017 04:14:41

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:09
Subject Sample # 1	0.165	04:10
Breath Volume	1.296	04:10
Air Blank	0.000	04:10
Air Blank	0.000	04:12
Subject Sample # 2	0.181	04:13
Breath Volume	2.359	04:13
Air Blank	0.000	04:13
Control Test	0.080	04:14
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,E
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GARCIAVILLAFUERTE

DATE/TIME OF TEST: 06/03/2017 23:34:58

Diagnostics Check	OK	23:28
Air Blank	0.000	23:28
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.213	23:30
Breath Volume	3.308	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.206	23:33
Breath Volume	2.449	23:33
Air Blank	0.000	23:34
Control Test	0.081	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, GEAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: ACOSTA-GONZALEZ

DATE/TIME OF TEST: 06/04/2017 01:14:14

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.081	01:08
Air Blank	0.000	01:09
Subject Sample # 1	0.071	01:09
Breath Volume	1.277	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:12
Subject Sample # 2	0.075	01:12
Breath Volume	3.980	01:12
Air Blank	0.000	01:13
Control Test	0.082	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: SALDANA DURAND

DATE/TIME OF TEST: 06/04/2017 03:30:36

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.081	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	Sequence Aborted
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 08:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: SYS CHECK TEST

DATE/TIME OF TEST: 06/05/2017 08:54:54

Diagnostics Check	OK	08:52	
Air Blank	0.000	08:52	
Control Test	0.081	08:52	
Air Blank	0.000	08:53	
Subject Sample	# 1	08:53	Subject Test Refused
Breath Volume	0.000	08:53	
Air Blank	0.000	08:54	
Control Test	0.081	08:54	
Air Blank	0.000	08:54	
Diagnostics Check	OK	08:54	

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: SYS CHECK
ARREST TIME: 08:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: SYS CHECK TEST2

DATE/TIME OF TEST: 06/05/2017 09:08:02

Diagnostics Check	OK	09:02
Air Blank	0.000	09:02
Control Test	0.082	09:03
Air Blank	0.000	09:03
Subject Sample # 1	0.000	09:03
Breath Volume	1.472	09:03
Air Blank	0.000	09:04
Air Blank	0.000	09:06
Subject Sample # 2	0.000	09:06
Breath Volume	1.136	09:06
Air Blank	0.000	09:07
Control Test	0.081	09:07
Air Blank	0.000	09:07
Diagnostics Check	OK	09:08

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BARTKO,WILLIAM
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: SINGH

DATE/TIME OF TEST: 06/05/2017 17:15:08

Diagnostics Check		OK	17:03	
Air Blank		0.000	17:04	
Control Test		0.081	17:04	
Air Blank		0.000	17:04	
Subject Sample	# 1	0.185	17:07	Volume Not Met
Breath Volume		0.597	17:07	
Air Blank		0.000	17:08	
Air Blank		0.000	17:10	
Subject Sample	# 2	0.198	17:13	Volume Not Met
Breath Volume		0.945	17:13	
Air Blank		0.000	17:14	
Control Test		0.079	17:14	
Air Blank		0.000	17:15	
Diagnostics Check		OK	17:15	

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ, ELVIN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: FLORES KHEMCHAND

DATE/TIME OF TEST: 06/07/2017 21:59:59

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:54
Air Blank	0.000	21:55
Subject Sample # 1	0.000	21:55
Breath Volume	1.250	21:55
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.000	21:58
Breath Volume	1.132	21:58
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

Data Download Date/Time: 06/08/2017 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RICHARDSON,JUSTIN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 06/09/2017 01:45:15

Diagnostics Check	OK	01:37
Air Blank	0.000	01:38
Control Test	0.081	01:38
Air Blank	0.000	01:39
Subject Sample # 1	0.164	01:40
Breath Volume	1.277	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.175	01:43
Breath Volume	1.335	01:43
Air Blank	0.000	01:44
Control Test	0.080	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THORNTON,BRANDEN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 06/09/2017 02:35:12

Diagnostics Check	OK	02:28
Air Blank	0.000	02:29
Control Test	0.080	02:29
Air Blank	0.000	02:30
Subject Sample # 1	0.010	02:30
Breath Volume	1.871	02:30
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.009	02:33
Breath Volume	2.246	02:33
Air Blank	0.000	02:34
Control Test	0.081	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 06/09/2017 04:03:39

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.160	03:59
Breath Volume	2.976	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.158	04:02
Breath Volume	3.613	04:02
Air Blank	0.000	04:02
Control Test	0.080	04:03
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS,C
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 06/09/2017 19:27:48

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.081	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.000	19:23
Breath Volume	1.300	19:23
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.000	19:26
Breath Volume	1.398	19:26
Air Blank	0.000	19:26
Control Test	0.081	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PRUDENCIOMILLAN

DATE/TIME OF TEST: 06/09/2017 22:12:35

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.227	22:07
Breath Volume	2.207	22:07
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.228	22:11
Breath Volume	1.371	22:11
Air Blank	0.000	22:11
Control Test	0.079	22:12
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 06/10/2017 01:05:01

Diagnostics Check		OK	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Subject Sample	# 1	0.209	01:00
Breath Volume		4.203	01:00
Air Blank		0.000	01:01
Air Blank		0.000	01:03
Subject Sample	# 2	0.226	01:03
Breath Volume		4.046	01:03
Air Blank		0.000	01:04
Control Test		0.081	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:05

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: BALLIET

DATE/TIME OF TEST: 06/10/2017 02:35:32

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.153	02:31
Breath Volume	2.515	02:31
Air Blank	0.000	02:31
Air Blank	0.000	02:33
Subject Sample # 2	0.154	02:34
Breath Volume	2.457	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Diagnostics Check	OK	02:35

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRANTLEY, JASON
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: PERRON

DATE/TIME OF TEST: 06/10/2017 04:12:03

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.081	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.145	04:07
Breath Volume	1.722	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.143	04:10
Breath Volume	1.292	04:10
Air Blank	0.000	04:11
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:12

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WILLIAMSON,BRANDEE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 06/10/2017 05:00:00

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.080	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.148	04:55
Breath Volume	5.277	04:55
Air Blank	0.000	04:56
Air Blank	0.000	04:57
Subject Sample # 2	0.145	04:58
Breath Volume	5.753	04:58
Air Blank	0.000	04:59
Control Test	0.080	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	05:00

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ORENGO

DATE/TIME OF TEST: 06/10/2017 22:47:08

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.126	22:42
Breath Volume	1.726	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.121	22:45
Breath Volume	1.937	22:45
Air Blank	0.000	22:46
Control Test	0.081	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:24

SUBJECT'S LAST NAME: MEJIAHERNANDEZ

DATE/TIME OF TEST: 06/10/2017 23:55:16

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.157	23:50
Breath Volume	2.945	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.155	23:53
Breath Volume	2.968	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ACEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: AMELO SUCETA

DATE/TIME OF TEST: 06/11/2017 03:04:11

Diagnostics Check	OK	02:57
Air Blank	0.000	02:58
Control Test	0.081	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.155	02:59
Breath Volume	1.453	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.148	03:02
Breath Volume	3.589	03:02
Air Blank	0.000	03:03
Control Test	0.080	03:03
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: INCHAUSTI

DATE/TIME OF TEST: 06/11/2017 05:01:27

Diagnostics Check	OK	04:54
Air Blank	0.000	04:55
Control Test	0.081	04:55
Air Blank	0.000	04:56
Subject Sample # 1	0.014	04:57
Breath Volume	1.464	04:57
Air Blank	0.000	04:57
Air Blank	0.000	04:59
Subject Sample # 2	0.013	05:00
Breath Volume	2.296	05:00
Air Blank	0.000	05:00
Control Test	0.080	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 06/14/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 06/11/2017 05:48:09

Diagnostics Check	OK	05:41
Air Blank	0.000	05:42
Control Test	0.081	05:42
Air Blank	0.000	05:42
Subject Sample # 1	0.047	05:43
Breath Volume	1.273	05:43
Air Blank	0.000	05:43
Air Blank	0.000	05:45
Subject Sample # 2	0.053	05:46
Breath Volume	1.925	05:46
Air Blank	0.000	05:47
Control Test	0.081	05:47
Air Blank	0.000	05:48
Diagnostics Check	OK	05:48

Data Download Date/Time: 06/14/2017 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WELLS,GRAHAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: PREVALUS

DATE/TIME OF TEST: 06/11/2017 07:29:32

Diagnostics Check	OK	07:22
Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:23
Subject Sample # 1	0.102	07:24
Breath Volume	1.875	07:24
Air Blank	0.000	07:25
Air Blank	0.000	07:27
Subject Sample # 2	0.111	07:28
Breath Volume	2.164	07:28
Air Blank	0.000	07:28
Control Test	0.080	07:28
Air Blank	0.000	07:29
Diagnostics Check	OK	07:29

Data Download Date/Time: 06/14/2017 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HACKEY,C
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/11/2017 22:03:16

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:57
Air Blank	0.000	21:58
Subject Sample # 1	0.183	21:58
Breath Volume	3.855	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.190	22:01
Breath Volume	3.992	22:01
Air Blank	0.000	22:02
Control Test	0.080	22:02
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 06/14/2017 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 06/12/2017 00:31:49

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.144	00:26
Breath Volume	2.652	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.140	00:30
Breath Volume	1.582	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 06/14/2017 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LOPEZ,A
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 06/12/2017 02:44:51

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:32	
Control Test		0.081	02:32	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.148	02:36	Volume Not Met
Breath Volume		0.078	02:36	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.181	02:40	
Breath Volume		2.078	02:40	
Air Blank		0.000	02:41	
Air Blank		0.000	02:42	
Subject Sample	# 3	0.183	02:43	
Breath Volume		1.207	02:43	
Air Blank		0.000	02:43	
Control Test		0.080	02:44	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 06/14/2017 12:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELUCA,ALEXANDER
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BREEN

DATE/TIME OF TEST: 06/13/2017 02:00:13

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.081	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.119	01:55
Breath Volume	1.621	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.124	01:58
Breath Volume	1.121	01:58
Air Blank	0.000	01:59
Control Test	0.081	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 06/14/2017 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HALLOCK, S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 05/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: DIAS

DATE/TIME OF TEST: 06/14/2017 04:57:02

Diagnostics Check	OK	04:49
Air Blank	0.000	04:49
Control Test	0.081	04:49
Air Blank	0.000	04:50
Subject Sample # 1	0.115	04:51
Breath Volume	1.187	04:51
Air Blank	0.000	04:51
Air Blank	0.000	04:53
Subject Sample # 2	0.109	04:55
Breath Volume	2.523	04:55
Air Blank	0.000	04:56
Control Test	0.081	04:56
Air Blank	0.000	04:56
Diagnostics Check	OK	04:57

Data Download Date/Time: 06/14/2017 12:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE, PETER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: MAYAN

DATE/TIME OF TEST: 06/16/2017 00:09:39

Diagnostics Check	OK	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:04
Air Blank	0.000	00:04
Subject Sample # 1	0.000	00:05
Breath Volume	2.695	00:05
Air Blank	0.000	00:05
Air Blank	0.000	00:07
Subject Sample # 2	0.000	00:08
Breath Volume	3.164	00:08
Air Blank	0.000	00:08
Control Test	0.081	00:09
Air Blank	0.000	00:09
Diagnostics Check	OK	00:09

Data Download Date/Time: 06/22/2017 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRELAND, D
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: MORA RAMIREZ

DATE/TIME OF TEST: 06/17/2017 01:54:23

Diagnostics Check		OK	01:47
Air Blank		0.000	01:48
Control Test		0.081	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.213	01:49
Breath Volume		2.480	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.223	01:52
Breath Volume		1.890	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 06/22/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KASSELMAN,SETH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:05

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 06/20/2017 05:38:28

Diagnostics Check	OK	05:29	
Air Blank	0.000	05:30	
Control Test	0.081	05:30	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:34	No Sample Provided
Breath Volume	0.000	05:34	
Air Blank	0.000	05:34	
Air Blank	0.000	05:36	
Subject Sample	# 2	05:37	Subject Test Refused
Breath Volume	0.000	05:37	
Air Blank	0.000	05:37	
Control Test	0.079	05:37	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 06/22/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHISEN,LEIGH
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CASTILLO-RIVERA

DATE/TIME OF TEST: 06/26/2017 02:10:06

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:54	
Control Test		0.080	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.232	01:58	Volume Not Met
Breath Volume		0.386	01:58	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2	0.257	02:03	
Breath Volume		1.363	02:03	
Air Blank		0.000	02:04	
Air Blank		0.000	02:06	
Subject Sample	# 3	0.260	02:08	
Breath Volume		1.457	02:08	
Air Blank		0.000	02:09	
Control Test		0.077	02:09	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 06/29/2017 14:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER, W
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 06/30/2017 03:43:28

Diagnostics Check	OK	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Subject Sample # 1	0.254	03:39
Breath Volume	2.402	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.243	03:42
Breath Volume	1.539	03:42
Air Blank	0.000	03:42
Control Test	0.080	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 07/07/2017 13:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: COATES,J
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: PESSARO

DATE/TIME OF TEST: 06/20/2017 19:46:58

Diagnostics Check	OK	19:44	
Air Blank	0.000	19:44	
Control Test	0.080	19:44	
Air Blank	0.000	19:45	
Subject Sample	# 1	19:45	Subject Test Refused
Breath Volume	0.000	19:45	
Air Blank	0.000	19:46	
Control Test	0.080	19:46	
Air Blank	0.000	19:46	
Diagnostics Check	OK	19:46	

Data Download Date/Time: 06/20/2017 20:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,JAMES,H
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: MENENDEZ,G
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 05/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 06/14/2017 20:31:45

Diagnostics Check	OK	20:28	
Air Blank	0.000	20:29	
Control Test	0.080	20:29	
Air Blank	0.000	20:30	
Subject Sample	# 1	20:30	Subject Test Refused
Breath Volume	0.000	20:30	
Air Blank	0.000	20:30	
Control Test	0.081	20:31	
Air Blank	0.000	20:31	
Diagnostics Check	OK	20:31	

Data Download Date/Time: 06/20/2017 20:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,HECTOR,L
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SIMMONS,BRIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 12:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 06/09/2017 15:22:28

Diagnostics Check	OK	15:19	
Air Blank	0.000	15:20	
Control Test	0.080	15:20	
Air Blank	0.000	15:20	
Subject Sample	# 1	15:21	Subject Test Refused
Breath Volume	0.000	15:21	
Air Blank	0.000	15:21	
Control Test	0.080	15:21	
Air Blank	0.000	15:22	
Diagnostics Check	OK	15:22	

Data Download Date/Time: 06/26/2017 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEVELAND,WILLIAM,S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARCIA,KELLY
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 05/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: PAULO

DATE/TIME OF TEST: 06/12/2017 22:24:52

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.000	22:20
Breath Volume	2.687	22:20
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.000	22:23
Breath Volume	2.757	22:23
Air Blank	0.000	22:23
Control Test	0.078	22:24
Air Blank	0.000	22:24
Diagnostics Check	OK	22:24

Data Download Date/Time: 06/26/2017 08:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, MATTHEW, G
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DUMMER, LUKE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: STABILITO

DATE/TIME OF TEST: 06/01/2017 03:05:02

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.080	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.055	03:00
Breath Volume	3.812	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:02
Subject Sample # 2	0.053	03:03
Breath Volume	3.597	03:03
Air Blank	0.000	03:04
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:05

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: VIGLIONE

DATE/TIME OF TEST: 06/07/2017 22:52:26

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.202	22:47
Breath Volume	3.140	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.208	22:50
Breath Volume	3.292	22:50
Air Blank	0.000	22:51
Control Test	0.079	22:51
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 17:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:18

SUBJECT'S LAST NAME: RIPOSTA

DATE/TIME OF TEST: 06/08/2017 18:57:20

Diagnostics Check	OK	18:50
Air Blank	0.000	18:51
Control Test	0.081	18:51
Air Blank	0.000	18:52
Subject Sample # 1	0.000	18:52
Breath Volume	3.964	18:52
Air Blank	0.000	18:53
Air Blank	0.000	18:55
Subject Sample # 2	0.000	18:55
Breath Volume	3.902	18:55
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Diagnostics Check	OK	18:57

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: AMORETTI

DATE/TIME OF TEST: 06/13/2017 17:59:28

Diagnostics Check	OK	17:53
Air Blank	0.000	17:53
Control Test	0.080	17:54
Air Blank	0.000	17:54
Subject Sample # 1	0.000	17:55
Breath Volume	2.070	17:55
Air Blank	0.000	17:55
Air Blank	0.000	17:57
Subject Sample # 2	0.000	17:58
Breath Volume	1.929	17:58
Air Blank	0.000	17:58
Control Test	0.081	17:58
Air Blank	0.000	17:59
Diagnostics Check	OK	17:59

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 17:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: TOPPE

DATE/TIME OF TEST: 06/14/2017 18:06:39

Diagnostics Check	OK	18:00
Air Blank	0.000	18:00
Control Test	0.081	18:01
Air Blank	0.000	18:01
Subject Sample # 1	0.000	18:02
Breath Volume	2.792	18:02
Air Blank	0.000	18:02
Air Blank	0.000	18:04
Subject Sample # 2	0.000	18:05
Breath Volume	3.074	18:05
Air Blank	0.000	18:05
Control Test	0.081	18:06
Air Blank	0.000	18:06
Diagnostics Check	OK	18:06

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,JEFFERY,L
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAIR,JEFFERY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: RAPPORT

DATE/TIME OF TEST: 06/15/2017 03:32:41

Diagnostics Check	OK	03:21
Air Blank	0.000	03:22
Control Test	0.080	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.160	03:25
Breath Volume	2.058	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.156	03:31
Breath Volume	3.558	03:31
Air Blank	0.000	03:31
Control Test	0.081	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: MARA

DATE/TIME OF TEST: 06/16/2017 23:57:31

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.162	23:53
Breath Volume	1.769	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.164	23:55
Breath Volume	1.906	23:55
Air Blank	0.000	23:56
Control Test	0.081	23:56
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: SEVILLA

DATE/TIME OF TEST: 06/18/2017 01:04:11

Diagnostics Check	OK	00:57
Air Blank	0.000	00:57
Control Test	0.080	00:58
Air Blank	0.000	00:58
Subject Sample # 1	0.120	00:59
Breath Volume	1.214	00:59
Air Blank	0.000	01:00
Air Blank	0.000	01:01
Subject Sample # 2	0.114	01:02
Breath Volume	1.410	01:02
Air Blank	0.000	01:03
Control Test	0.081	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, MATTHEW, G
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: FORNASH, GEORGE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 06/20/2017 23:57:43

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.079	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.218	23:53
Breath Volume	2.675	23:53
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.212	23:56
Breath Volume	2.773	23:56
Air Blank	0.000	23:56
Control Test	0.080	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:57

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 05/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 06/21/2017 19:30:38

Diagnostics Check	OK	19:23
Air Blank	0.000	19:24
Control Test	0.079	19:24
Air Blank	0.000	19:25
Subject Sample # 1	0.263	19:26
Breath Volume	1.257	19:26
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.252	19:29
Breath Volume	1.429	19:29
Air Blank	0.000	19:29
Control Test	0.080	19:30
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 06/22/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: HESSION,CHRISTOPHER
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:34

SUBJECT'S LAST NAME: CHIASSON

DATE/TIME OF TEST: 06/29/2017 18:18:26

Diagnostics Check	OK	18:07
Air Blank	0.000	18:08
Control Test	0.077	18:08
Air Blank	0.000	18:09
Subject Sample # 1	0.206	18:09
Breath Volume	2.261	18:09
Air Blank	0.000	18:10
Air Blank	0.000	18:12
Subject Sample # 2	0.184	18:13
Breath Volume	1.859	18:13
Air Blank	0.000	18:14
Air Blank	0.000	18:16
Subject Sample # 3	0.219	18:16
Breath Volume	1.464	18:16
Air Blank	0.000	18:17
Control Test	0.080	18:17
Air Blank	0.000	18:18
Diagnostics Check	OK	18:18

Data Download Date/Time: 07/31/2017 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,JEFFERY,L
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAIR,JEFFERY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 06/30/2017 11:20:12

Diagnostics Check	OK	11:16	
Air Blank	0.000	11:17	
Control Test	0.080	11:17	
Air Blank	0.000	11:17	
Subject Sample	# 1	11:18	Slope Not Met
Breath Volume	1.511	11:18	
Air Blank	0.000	11:19	
Control Test	0.080	11:19	
Air Blank	0.000	11:20	
Diagnostics Check	OK	11:20	

Data Download Date/Time: 07/31/2017 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,JEFFERY,L
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAIR,JEFFERY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 10:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: FOLSOM

DATE/TIME OF TEST: 06/30/2017 11:27:57

Diagnostics Check	OK	11:21
Air Blank	0.000	11:21
Control Test	0.080	11:22
Air Blank	0.000	11:22
Subject Sample # 1	0.338	11:23
Breath Volume	2.007	11:23
Air Blank	0.000	11:24
Air Blank	0.000	11:25
Subject Sample # 2	0.326	11:26
Breath Volume	2.078	11:26
Air Blank	0.000	11:27
Control Test	0.079	11:27
Air Blank	0.000	11:27
Diagnostics Check	OK	11:27

Data Download Date/Time: 07/31/2017 08:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINEHAN,DANIEL,J
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: LINEHAN,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 06/30/2017 21:53:38

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.152	21:49
Breath Volume	2.011	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.159	21:52
Breath Volume	2.566	21:52
Air Blank	0.000	21:52
Control Test	0.080	21:53
Air Blank	0.000	21:53
Diagnostics Check	OK	21:53

Data Download Date/Time: 07/31/2017 08:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANTON,KEVIN,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: STANTON,KEVIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 06/22/2017 02:38:59

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.080	02:38	
Air Blank	0.000	02:38	
Diagnostics Check	OK	02:38	

Data Download Date/Time: 07/26/2017 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EDWARDS,KIMONE,V
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: ROWELL,NATHANAEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 06/22/2017 15:57:10

Diagnostics Check	OK	15:50
Air Blank	0.000	15:51
Control Test	0.079	15:51
Air Blank	0.000	15:52
Subject Sample # 1	0.000	15:52
Breath Volume	1.433	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.000	15:55
Breath Volume	2.117	15:55
Air Blank	0.000	15:56
Control Test	0.079	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 07/26/2017 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLEASON, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GLEASON, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 06/22/2017 22:49:27

Diagnostics Check	OK	22:43
Air Blank	0.000	22:43
Control Test	0.080	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.000	22:45
Breath Volume	3.902	22:45
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.000	22:48
Breath Volume	3.953	22:48
Air Blank	0.000	22:48
Control Test	0.080	22:48
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 07/26/2017 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GLEASON, WILLIAM, L
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GLEASON, WILLIAM
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: MONFORT

DATE/TIME OF TEST: 06/24/2017 00:39:07

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.080	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.206	00:34
Breath Volume		2.503	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:36
Subject Sample	# 2	0.204	00:37
Breath Volume		2.214	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 07/26/2017 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOCKHART,BRETT,A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: EASTMAN,MATTHEW
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 06/25/2017 02:01:45

Diagnostics Check	OK	01:55
Air Blank	0.000	01:55
Control Test	0.076	01:55
Air Blank	0.000	01:56
Subject Sample # 1	0.176	01:57
Breath Volume	3.820	01:57
Air Blank	0.000	01:57
Air Blank	0.000	01:59
Subject Sample # 2	0.161	02:00
Breath Volume	2.457	02:00
Air Blank	0.000	02:00
Control Test	0.075	02:01
Air Blank	0.000	02:01
Diagnostics Check	OK	02:01

Data Download Date/Time: 07/26/2017 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BEATTY

DATE/TIME OF TEST: 06/30/2017 23:49:08

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.080	23:43
Air Blank	0.000	23:44
Subject Sample # 1	0.208	23:44
Breath Volume	1.488	23:44
Air Blank	0.000	23:45
Air Blank	0.000	23:47
Subject Sample # 2	0.222	23:47
Breath Volume	1.246	23:47
Air Blank	0.000	23:48
Control Test	0.080	23:48
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 07/26/2017 19:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FREY,ROBERT,P

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HOLLCROFT,D

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:03

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: DE-FURIA

DATE/TIME OF TEST: 06/24/2017 21:05:20

Diagnostics Check	OK	20:56
Air Blank	0.000	20:57
Control Test	0.080	20:57
Air Blank	0.000	20:58
Subject Sample # 1	0.175	21:00
Breath Volume	1.769	21:00
Air Blank	0.000	21:01
Air Blank	0.000	21:02
Subject Sample # 2	0.170	21:03
Breath Volume	2.347	21:03
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 07/12/2017 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: NEMETH,BEN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: SWIDERSKI

DATE/TIME OF TEST: 06/25/2017 20:17:16

Diagnostics Check	OK	20:11
Air Blank	0.000	20:11
Control Test	0.080	20:12
Air Blank	0.000	20:12
Subject Sample # 1	0.181	20:12
Breath Volume	1.945	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:15
Subject Sample # 2	0.180	20:15
Breath Volume	1.691	20:15
Air Blank	0.000	20:16
Control Test	0.080	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 07/12/2017 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: ROLPH,AMANDA
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MENJIVAR

DATE/TIME OF TEST: 06/25/2017 21:00:04

Diagnostics Check	OK	20:53
Air Blank	0.000	20:53
Control Test	0.080	20:54
Air Blank	0.000	20:54
Subject Sample # 1	0.189	20:55
Breath Volume	1.804	20:55
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.183	20:58
Breath Volume	1.855	20:58
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	20:59
Diagnostics Check	OK	21:00

Data Download Date/Time: 07/12/2017 14:48

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREY,ROBERT,P
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: SMITH,D
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: ORLENA

DATE/TIME OF TEST: 06/25/2017 21:44:42

Diagnostics Check	OK	21:37
Air Blank	0.000	21:37
Control Test	0.080	21:38
Air Blank	0.000	21:38
Subject Sample # 1	0.190	21:39
Breath Volume	1.558	21:39
Air Blank	0.000	21:40
Air Blank	0.000	21:41
Subject Sample # 2	0.173	21:43
Breath Volume	1.683	21:43
Air Blank	0.000	21:43
Control Test	0.080	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 07/12/2017 14:48

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 03:28

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MICK

DATE/TIME OF TEST: 06/29/2017 04:02:09

Diagnostics Check		OK	03:56
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:57
Subject Sample	# 1	0.154	03:57
Breath Volume		1.726	03:57
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.147	04:00
Breath Volume		1.519	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Diagnostics Check		OK	04:02

Data Download Date/Time: 07/12/2017 14:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: COLBURN, MEGAN
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: MONTGOMERY

DATE/TIME OF TEST: 06/01/2017 03:59:00

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.224	03:54
Breath Volume	2.011	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.208	03:57
Breath Volume	1.609	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DAIL, EDD
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: ALICEA

DATE/TIME OF TEST: 06/01/2017 04:29:53

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.080	04:24
Air Blank	0.000	04:25
Subject Sample # 1	0.000	04:25
Breath Volume	1.906	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.000	04:28
Breath Volume	2.175	04:28
Air Blank	0.000	04:28
Control Test	0.079	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: ARIAS,JONATHAN
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CHUMITA

DATE/TIME OF TEST: 06/01/2017 20:27:21

Diagnostics Check	OK	20:26	
Air Blank		20:26	Ambient Fail
Air Blank	0.000	20:27	

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLBURN,MEGAN,E
BREATH TEST OPERATOR AGENCY: Casselberry PD
ARRESTING OFFICER: ARIAS,JONATHAN
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CHUMITA

DATE/TIME OF TEST: 06/01/2017 20:47:07

Diagnostics Check	OK	20:29	
Air Blank	0.000	20:30	
Control Test	0.078	20:30	
Air Blank	0.000	20:30	
Subject Sample	# 1 0.223	20:33	
Breath Volume	1.718	20:33	
Air Blank	0.000	20:33	
Air Blank	0.000	20:35	
Subject Sample	# 2 0.216	20:38	Volume Not Met
Breath Volume	0.082	20:38	
Air Blank	0.000	20:39	
Air Blank	0.000	20:41	
Subject Sample	# 3 0.226	20:41	
Breath Volume	1.910	20:41	
Air Blank	0.000	20:42	
Control Test	0.078	20:42	
Air Blank	0.000	20:43	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER, MATTHEW, D
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: FOWLER, MATTHEW
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: CHEEK

DATE/TIME OF TEST: 06/01/2017 21:43:06

Diagnostics Check	OK	21:31
Air Blank	0.000	21:31
Control Test	0.079	21:32
Air Blank	0.000	21:32
Subject Sample # 1	0.277	21:33
Breath Volume	1.656	21:33
Air Blank	0.000	21:34
Air Blank	0.000	21:36
Subject Sample # 2	0.284	21:36
Breath Volume	2.179	21:36
Air Blank	0.000	21:37
Control Test	0.078	21:37
Air Blank	0.000	21:38
Diagnostics Check	OK	21:43

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HEIM,EDWARD
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: MORETZ

DATE/TIME OF TEST: 06/02/2017 04:45:39

Diagnostics Check	OK	04:38
Air Blank	0.000	04:38
Control Test	0.079	04:39
Air Blank	0.000	04:39
Subject Sample # 1	0.056	04:40
Breath Volume	3.328	04:40
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.056	04:44
Breath Volume	3.070	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DONNELLY, SHANE
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: HAROLD

DATE/TIME OF TEST: 06/18/2017 01:09:36

Diagnostics Check	OK	01:03
Air Blank	0.000	01:03
Control Test	0.078	01:04
Air Blank	0.000	01:04
Subject Sample # 1	0.280	01:05
Breath Volume	3.269	01:05
Air Blank	0.000	01:05
Air Blank	0.000	01:07
Subject Sample # 2	0.288	01:08
Breath Volume	2.953	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Diagnostics Check	OK	01:09

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MOSELEY, SCOTT
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: IGBO-NWOKE

DATE/TIME OF TEST: 06/23/2017 04:09:02

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.079	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.220	03:59
Breath Volume		1.910	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.199	04:03
Breath Volume		1.750	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:05
Subject Sample	# 3	0.231	04:07
Breath Volume		1.585	04:07
Air Blank		0.000	04:08
Control Test		0.078	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:09

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ZABALA, JOSE
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 06/23/2017 21:51:50

Diagnostics Check	OK	21:45
Air Blank	0.000	21:46
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.149	21:47
Breath Volume	1.535	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.150	21:50
Breath Volume	2.375	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REZNY, LAUREN
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 06/24/2017 02:57:38

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.080	02:52
Air Blank	0.000	02:52
Subject Sample # 1	0.000	02:52
Breath Volume	1.195	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.000	02:56
Breath Volume	1.460	02:56
Air Blank	0.000	02:56
Control Test	0.079	02:57
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DAIL, EDD
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 06/24/2017 04:10:33

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.185	04:02
Breath Volume		3.269	04:02
Air Blank		0.000	04:03
Air Blank		0.000	04:04
Subject Sample	# 2	0.157	04:06
Breath Volume		1.910	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 3	0.166	04:09
Breath Volume		2.917	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: SIMPSON, CARL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 05/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BLAKE

DATE/TIME OF TEST: 06/29/2017 02:53:21

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.079	02:48
Air Blank	0.000	02:48
Subject Sample # 1	0.000	02:49
Breath Volume	2.121	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.000	02:51
Breath Volume	2.992	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 07/03/2017 22:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SILVERSTEIN, CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: MANTIONE

DATE/TIME OF TEST: 06/30/2017 22:24:47

Diagnostics Check	OK	22:21	
Air Blank	0.000	22:22	
Control Test	0.081	22:22	
Air Blank	0.000	22:23	
Subject Sample	# 1	22:23	Subject Test Refused
Breath Volume	0.000	22:23	
Air Blank	0.000	22:23	
Control Test	0.080	22:24	
Air Blank	0.000	22:24	
Diagnostics Check	OK	22:24	

Data Download Date/Time: 07/28/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: KUBISIAK

DATE/TIME OF TEST: 06/30/2017 22:46:49

Diagnostics Check	OK	22:43	
Air Blank	0.000	22:44	
Control Test	0.078	22:44	
Air Blank	0.000	22:45	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.077	22:46	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 07/28/2017 23:20