

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS, MATTHEW, A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: GELSEY, RICHARD  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: WOLLENBERG

DATE/TIME OF TEST: 07/13/2017 00:53:18

---

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.080	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.181	00:48
Breath Volume	4.082	00:48
Air Blank	0.000	00:49
Air Blank	0.000	00:50
Subject Sample # 2	0.184	00:51
Breath Volume	3.246	00:51
Air Blank	0.000	00:52
Control Test	0.080	00:52
Air Blank	0.000	00:53
Diagnostics Check	OK	00:53

---

Data Download Date/Time: 07/26/2017 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MYERS,MATTHEW,A  
BREATH TEST OPERATOR AGENCY: Volusia County SO  
ARRESTING OFFICER: MYERS,MATTHEW  
ARRESTING OFFICER AGENCY: Volusia County SO  
LAST AGENCY INSPECTION: 06/20/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 203498  
DRY GAS EXPIRY DATE: 12/10/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/21/2017 21:58:54

---

Diagnostics Check	OK	21:47
Air Blank	0.000	21:47
Control Test	0.080	21:48
Air Blank	0.000	21:48
Subject Sample # 1	0.250	21:49
Breath Volume	2.117	21:49
Air Blank	0.000	21:49
Air Blank	0.000	21:51
Subject Sample # 2	0.222	21:52
Breath Volume	1.855	21:52
Air Blank	0.000	21:52
Air Blank	0.000	21:54
Subject Sample # 3	0.237	21:57
Breath Volume	1.816	21:57
Air Blank	0.000	21:57
Control Test	0.081	21:58
Air Blank	0.000	21:58
Diagnostics Check	OK	21:58

---

Data Download Date/Time: 07/26/2017 09:01

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCGINNIS,J.,D  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: MCGINNIS,J.  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: KENNY

DATE/TIME OF TEST: 07/24/2017 22:44:29

---

Diagnostics Check	OK	22:39	
Air Blank	0.000	22:39	
Control Test	0.078	22:40	
Air Blank	0.000	22:40	
Subject Sample	# 1	22:42	Subject Test Refused
Breath Volume	0.300	22:42	
Air Blank	0.000	22:43	
Control Test	0.080	22:43	
Air Blank	0.000	22:44	
Diagnostics Check	OK	22:44	

---

Data Download Date/Time: 08/28/2017 11:33

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MASTERS

DATE/TIME OF TEST: 07/25/2017 00:51:25

---

Diagnostics Check	OK	00:45
Air Blank	0.000	00:45
Control Test	0.078	00:46
Air Blank	0.000	00:46
Subject Sample # 1	0.000	00:46
Breath Volume	1.257	00:46
Air Blank	0.000	00:47
Air Blank	0.000	00:49
Subject Sample # 2	0.000	00:50
Breath Volume	1.460	00:50
Air Blank	0.000	00:50
Control Test	0.078	00:50
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

---

Data Download Date/Time: 09/25/2017 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPROUSE,JOHN,S  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: HERTZ,ADAM  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: REDDEN

DATE/TIME OF TEST: 07/27/2017 02:10:24

---

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.078	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.262	02:01
Breath Volume	1.343	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.237	02:05
Breath Volume	1.242	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:07
Subject Sample # 3	0.254	02:08
Breath Volume	1.121	02:08
Air Blank	0.000	02:09
Control Test	0.077	02:09
Air Blank	0.000	02:10
Diagnostics Check	OK	02:10

---

Data Download Date/Time: 09/25/2017 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SPROUSE,JOHN,S  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: SPROUSE,JOHN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: HERR

DATE/TIME OF TEST: 07/27/2017 23:05:08

---

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.077	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.185	22:59
Breath Volume	1.757	22:59
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.184	23:03
Breath Volume	1.734	23:03
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

---

Data Download Date/Time: 09/25/2017 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WINTERS,ERIN,M  
BREATH TEST OPERATOR AGENCY: St. John's County SO  
ARRESTING OFFICER: FLOYD,JOHN  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: ONDERDONK

DATE/TIME OF TEST: 07/28/2017 23:58:01

---

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:53
Subject Sample # 1	0.170	23:53
Breath Volume	1.738	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.163	23:56
Breath Volume	1.554	23:56
Air Blank	0.000	23:57
Control Test	0.077	23:57
Air Blank	0.000	23:57
Diagnostics Check	OK	23:58

---

Data Download Date/Time: 09/25/2017 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MITCHELL,DAVID,M  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BRIGGS,JASON  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 14:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 07/29/2017 15:06:49

---

Diagnostics Check	OK	15:02	
Air Blank	0.000	15:03	
Control Test	0.078	15:03	
Air Blank	0.000	15:04	
Subject Sample	# 1	15:06	Sequence Aborted
Breath Volume	0.000	15:06	
Air Blank	0.000	15:06	

---

Data Download Date/Time: 09/25/2017 11:09



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: MONTGOMERY,KENNETH  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:52  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BEGZADA

DATE/TIME OF TEST: 07/30/2017 03:07:16

---

Diagnostics Check	OK	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.000	03:03
Breath Volume	1.152	03:03
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.000	03:05
Breath Volume	1.742	03:05
Air Blank	0.000	03:06
Control Test	0.079	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

---

Data Download Date/Time: 09/25/2017 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRICKEY,M,  
BREATH TEST OPERATOR AGENCY: St. Augustine PD  
ARRESTING OFFICER: BRICKEY,M  
ARRESTING OFFICER AGENCY: St. Augustine PD  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: EPSTEIN

DATE/TIME OF TEST: 07/30/2017 04:19:53

---

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.079	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:18	
Control Test	0.079	04:19	
Air Blank	0.000	04:19	
Diagnostics Check	OK	04:19	

---

Data Download Date/Time: 09/25/2017 11:09

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CONRAD,CHRISTOPHER,  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: CUBBEDGE,KYLE  
ARRESTING OFFICER AGENCY: St. John's County SO  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 366849  
DRY GAS EXPIRY DATE: 03/30/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: STILWELL

DATE/TIME OF TEST: 07/31/2017 22:05:34

---

Diagnostics Check	OK	21:55
Air Blank	0.000	21:55
Control Test	0.078	21:56
Air Blank	0.000	21:56
Subject Sample # 1	0.175	21:58
Breath Volume	1.332	21:58
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.166	22:04
Breath Volume	1.125	22:04
Air Blank	0.000	22:04
Control Test	0.078	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

---

Data Download Date/Time: 09/25/2017 11:09

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001136

INSTRUMENT REGISTERED TO: DELAND PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MULERO,JOHAN,  
BREATH TEST OPERATOR AGENCY: Deland PD  
ARRESTING OFFICER: SANTOS,JOSHUA  
ARRESTING OFFICER AGENCY: Deland PD  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 07/16/2017 02:42:15

---

Diagnostics Check	OK	02:36
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.254	02:37
Breath Volume	1.933	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:40
Subject Sample # 2	0.239	02:40
Breath Volume	2.960	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:42
Diagnostics Check	OK	02:42

---

Data Download Date/Time: 07/26/2017 09:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KEMP,FREDERICK,R  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: WEST,JACOB  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: STEVENS

DATE/TIME OF TEST: 07/02/2017 02:22:40

---

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.086	02:17
Breath Volume	1.304	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.085	02:21
Breath Volume	1.132	02:21
Air Blank	0.000	02:21
Control Test	0.079	02:22
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

---

Data Download Date/Time: 07/06/2017 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FAIRCLOTH,EDWARD,W  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: WEST,J  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 10:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/03/2017 23:14:04

---

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.080	23:08
Air Blank	0.000	23:09
Subject Sample # 1	0.213	23:09
Breath Volume	1.921	23:09
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.220	23:12
Breath Volume	1.886	23:12
Air Blank	0.000	23:13
Control Test	0.080	23:13
Air Blank	0.000	23:13
Diagnostics Check	OK	23:14

---

Data Download Date/Time: 07/06/2017 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WEAVER,D,  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: WEAVER,D  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: FLYNN

DATE/TIME OF TEST: 07/05/2017 00:40:13

---

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:37	
Control Test	0.080	00:37	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:39	
Control Test	0.080	00:39	
Air Blank	0.000	00:40	
Diagnostics Check	OK	00:40	

---

Data Download Date/Time: 07/06/2017 16:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VINCI BUSTOS,ROSANNA,L  
BREATH TEST OPERATOR AGENCY: Flagler Beach PD  
ARRESTING OFFICER: VINCI BUSTOS,ROSANNA  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CRANE

DATE/TIME OF TEST: 07/05/2017 01:06:07

---

Diagnostics Check	OK	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Subject Sample # 1	0.196	01:01
Breath Volume	2.238	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.195	01:04
Breath Volume	2.187	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

---

Data Download Date/Time: 07/06/2017 16:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: RODENBORN,J  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:51  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: FOX

DATE/TIME OF TEST: 07/12/2017 14:14:11

---

Diagnostics Check	OK	14:06
Air Blank	0.000	14:07
Control Test	0.080	14:07
Air Blank	0.000	14:08
Subject Sample # 1	0.000	14:09
Breath Volume	1.824	14:09
Air Blank	0.000	14:09
Air Blank	0.000	14:11
Subject Sample # 2	0.000	14:12
Breath Volume	1.425	14:12
Air Blank	0.000	14:13
Control Test	0.080	14:13
Air Blank	0.000	14:14
Diagnostics Check	OK	14:14

---

Data Download Date/Time: 08/18/2017 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: EBRILL,FIONA  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:34  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 07/14/2017 00:45:45

---

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.079	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.251	00:41
Breath Volume		2.085	00:41
Air Blank		0.000	00:41
Air Blank		0.000	00:43
Subject Sample	# 2	0.251	00:44
Breath Volume		1.558	00:44
Air Blank		0.000	00:44
Control Test		0.079	00:45
Air Blank		0.000	00:45
Diagnostics Check		OK	00:45

---

Data Download Date/Time: 08/18/2017 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,JASON,B  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: WILLIAMS,JASON  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 06:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 07:50

SUBJECT'S LAST NAME: VOJNOVSKI

DATE/TIME OF TEST: 07/19/2017 08:27:54

---

Diagnostics Check	OK	08:21
Air Blank	0.000	08:21
Control Test	0.080	08:22
Air Blank	0.000	08:22
Subject Sample # 1	0.178	08:23
Breath Volume	1.390	08:23
Air Blank	0.000	08:24
Air Blank	0.000	08:26
Subject Sample # 2	0.174	08:26
Breath Volume	1.117	08:26
Air Blank	0.000	08:27
Control Test	0.080	08:27
Air Blank	0.000	08:27
Diagnostics Check	OK	08:27

---

Data Download Date/Time: 08/18/2017 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: YELVINGTON,AUSTIN,R  
BREATH TEST OPERATOR AGENCY: Flagler Beach PD  
ARRESTING OFFICER: SYLVESTER,TIMOTHY  
ARRESTING OFFICER AGENCY: Flagler Beach PD  
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 17:29

SUBJECT'S LAST NAME: WILMOT

DATE/TIME OF TEST: 07/23/2017 19:58:41

---

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.079	19:53
Air Blank	0.000	19:53
Subject Sample # 1	0.305	19:54
Breath Volume	1.406	19:54
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.303	19:57
Breath Volume	1.671	19:57
Air Blank	0.000	19:57
Control Test	0.080	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

---

Data Download Date/Time: 08/18/2017 16:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NICOLAS,MITA,  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: PETKOVSEK,RICHARD  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 09:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:03

SUBJECT'S LAST NAME: CESAR

DATE/TIME OF TEST: 07/25/2017 10:30:03

---

Diagnostics Check	OK	10:23
Air Blank	0.000	10:24
Control Test	0.079	10:24
Air Blank	0.000	10:24
Subject Sample # 1	0.055	10:25
Breath Volume	1.765	10:25
Air Blank	0.000	10:26
Air Blank	0.000	10:28
Subject Sample # 2	0.056	10:28
Breath Volume	1.484	10:28
Air Blank	0.000	10:29
Control Test	0.080	10:29
Air Blank	0.000	10:29
Diagnostics Check	OK	10:30

---

Data Download Date/Time: 08/18/2017 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PRIEMER,AMANDA,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: COLSON,BLAKE  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 14:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:41

SUBJECT'S LAST NAME: MARTHIA

DATE/TIME OF TEST: 07/26/2017 16:07:20

---

Diagnostics Check	OK	16:04	
Air Blank	0.000	16:05	
Control Test	0.080	16:05	
Air Blank	0.000	16:05	
Subject Sample	# 1	16:05	Subject Test Refused
Breath Volume	0.000	16:05	
Air Blank	0.000	16:06	
Control Test	0.080	16:06	
Air Blank	0.000	16:07	
Diagnostics Check	OK	16:07	

---

Data Download Date/Time: 08/18/2017 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001138

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M  
BREATH TEST OPERATOR AGENCY: Flagler County SO  
ARRESTING OFFICER: LICHTY,DAVID  
ARRESTING OFFICER AGENCY: Flagler County SO  
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A6  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 07/30/2017 03:20:02

---

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.241	03:10
Breath Volume		1.402	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.217	03:15
Breath Volume		1.976	03:15
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 3	0.239	03:18
Breath Volume		1.851	03:18
Air Blank		0.000	03:19
Control Test		0.076	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:20

---

Data Download Date/Time: 08/18/2017 16:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEDINA,MARMILLY,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: GREISHEIMER,JOHN  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: TOLEDO

DATE/TIME OF TEST: 07/18/2017 05:31:44

---

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.082	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.171	05:26
Breath Volume	1.699	05:26
Air Blank	0.000	05:27
Air Blank	0.000	05:29
Subject Sample # 2	0.172	05:30
Breath Volume	1.734	05:30
Air Blank	0.000	05:30
Control Test	0.081	05:31
Air Blank	0.000	05:31
Diagnostics Check	OK	05:31

---

Data Download Date/Time: 07/25/2017 14:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: DALTON,RYAN  
ARRESTING OFFICER AGENCY: OTHER State  
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:19  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: ETIENNE

DATE/TIME OF TEST: 07/21/2017 06:46:14

---

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.081	06:40
Air Blank	0.000	06:41
Subject Sample # 1	0.013	06:41
Breath Volume	1.203	06:41
Air Blank	0.000	06:42
Air Blank	0.000	06:44
Subject Sample # 2	0.013	06:44
Breath Volume	1.480	06:44
Air Blank	0.000	06:45
Control Test	0.082	06:45
Air Blank	0.000	06:46
Diagnostics Check	OK	06:46

---

Data Download Date/Time: 07/25/2017 14:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: VANOVER,ANDREW,P  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: VANOVER,ANDREW  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 184081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:49  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DEL CUETO

DATE/TIME OF TEST: 07/03/2017 00:27:04

---

Diagnostics Check	OK	00:20
Air Blank	0.000	00:20
Control Test	0.080	00:21
Air Blank	0.000	00:21
Subject Sample # 1	0.145	00:22
Breath Volume	2.785	00:22
Air Blank	0.000	00:23
Air Blank	0.000	00:24
Subject Sample # 2	0.143	00:25
Breath Volume	1.886	00:25
Air Blank	0.000	00:26
Control Test	0.079	00:26
Air Blank	0.000	00:26
Diagnostics Check	OK	00:27

---

Data Download Date/Time: 07/24/2017 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAKAI,MICHAEL,T  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MAKAI,MICHAEL  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 184081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: STIVERS

DATE/TIME OF TEST: 07/05/2017 19:58:10

---

Diagnostics Check	OK	19:49	
Air Blank	0.000	19:49	
Control Test	0.080	19:49	
Air Blank	0.000	19:50	
Subject Sample	# 1	19:53	No Sample Provided
Breath Volume	0.000	19:53	
Air Blank	0.000	19:54	
Air Blank	0.000	19:56	
Subject Sample	# 2	19:56	Subject Test Refused
Breath Volume	0.000	19:56	
Air Blank	0.000	19:57	
Control Test	0.080	19:57	
Air Blank	0.000	19:58	
Diagnostics Check	OK	19:58	

---

Data Download Date/Time: 07/24/2017 10:45

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENYAN,MA,A  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: KENYAN,M  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 184081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:16  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: BROCK

DATE/TIME OF TEST: 07/28/2017 03:07:11

---

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:57	
Subject Sample	# 1 0.131	03:00	Volume Not Met
Breath Volume	0.359	03:00	
Air Blank	0.000	03:01	
Air Blank	0.000	03:02	
Subject Sample	# 2 0.103	03:05	Volume Not Met
Breath Volume	0.605	03:05	
Air Blank		03:06	Purge Fail
Air Blank	0.000	03:07	

---

Data Download Date/Time: 08/21/2017 11:25

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001150

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MAKAI,MICHAEL,  
BREATH TEST OPERATOR AGENCY: Manatee County SO  
ARRESTING OFFICER: MAKAI,MICHAEL  
ARRESTING OFFICER AGENCY: Manatee County SO  
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 184081  
DRY GAS EXPIRY DATE: 04/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:59  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: CARRILLO-ESPARZA

DATE/TIME OF TEST: 07/30/2017 22:28:27

---

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.082	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.177	22:23
Breath Volume	1.765	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.174	22:26
Breath Volume	2.339	22:26
Air Blank	0.000	22:27
Control Test	0.080	22:27
Air Blank	0.000	22:28
Diagnostics Check	OK	22:28

---

Data Download Date/Time: 08/21/2017 11:25

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAGER,KEVIN,M  
BREATH TEST OPERATOR AGENCY: Haines City PD  
ARRESTING OFFICER: HAGER,KEVIN  
ARRESTING OFFICER AGENCY: Haines City PD  
LAST AGENCY INSPECTION: 07/16/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: HUERTA GONZALEZ

DATE/TIME OF TEST: 07/23/2017 06:35:19

---

Diagnostics Check	OK	06:25
Air Blank	0.000	06:26
Control Test	0.083	06:26
Air Blank	0.000	06:27
Subject Sample # 1	0.200	06:27
Breath Volume	3.503	06:27
Air Blank	0.000	06:28
Air Blank	0.000	06:30
Subject Sample # 2	0.178	06:30
Breath Volume	1.222	06:30
Air Blank	0.000	06:31
Air Blank	0.000	06:33
Subject Sample # 3	0.171	06:33
Breath Volume	1.542	06:33
Air Blank	0.000	06:34
Control Test	0.081	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

---

Data Download Date/Time: 08/12/2017 09:01

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001152

INSTRUMENT REGISTERED TO: NEW SMYRNA BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COPPOLA, TYSON, J  
BREATH TEST OPERATOR AGENCY: New Smyrna Beach PD  
ARRESTING OFFICER: JENKINS, ANDRE  
ARRESTING OFFICER AGENCY: New Smyrna Beach PD  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:27

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 07/22/2017 00:45:14

---

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:34	
Control Test	0.081	00:34	
Air Blank	0.000	00:35	
Subject Sample	# 1	00:38	No Sample Provided
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Air Blank	0.000	00:40	
Subject Sample	# 2	00:43	No Sample Provided
Breath Volume	0.000	00:43	
Air Blank	0.000	00:44	
Control Test	0.081	00:44	
Air Blank	0.000	00:45	
Diagnostics Check	OK	00:45	

---

Data Download Date/Time: 07/26/2017 13:23

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001155

INSTRUMENT REGISTERED TO: EGLIN AFB  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATHRE,PETER,M  
BREATH TEST OPERATOR AGENCY: Eglin AFB PD  
ARRESTING OFFICER: BARR,BRANDON  
ARRESTING OFFICER AGENCY: Eglin AFB PD  
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WAGES

DATE/TIME OF TEST: 07/25/2017 23:56:23

---

Diagnostics Check	OK	23:46
Air Blank	0.000	23:46
Control Test	0.079	23:47
Air Blank	0.000	23:47
Subject Sample # 1	0.177	23:50
Breath Volume	1.191	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.166	23:54
Breath Volume	1.964	23:54
Air Blank	0.000	23:55
Control Test	0.079	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

---

Data Download Date/Time: 11/01/2017 13:29



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: GARCIA,T  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: BUCKLEY

DATE/TIME OF TEST: 07/05/2017 19:24:27

---

Diagnostics Check	OK	19:18
Air Blank	0.000	19:18
Control Test	0.080	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.267	19:19
Breath Volume	1.808	19:19
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.273	19:22
Breath Volume	2.082	19:22
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

---

Data Download Date/Time: 07/28/2017 15:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARMINGTON, WILLIAM, G  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: RUBIN, RICHARD  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PARASO

DATE/TIME OF TEST: 07/06/2017 00:06:26

---

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.081	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.136	00:02
Breath Volume		1.324	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.131	00:04
Breath Volume		1.460	00:04
Air Blank		0.000	00:05
Control Test		0.080	00:05
Air Blank		0.000	00:06
Diagnostics Check		OK	00:06

---

Data Download Date/Time: 07/28/2017 15:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,BENITA,L  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: HAMILTON,BENITA  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 07:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/23/2017 08:05:15

---

Diagnostics Check	OK	07:59	
Air Blank	0.000	08:00	
Control Test	0.081	08:00	
Air Blank	0.000	08:00	
Subject Sample	# 1 0.200	08:04	Volume Not Met
Breath Volume	0.015	08:04	
Air Blank		08:04	Purge Fail
Air Blank	0.000	08:05	

---

Data Download Date/Time: 07/28/2017 15:26

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HAMILTON,BENITA,L  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: HAMILTON,BENITA  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 07:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/23/2017 08:10:58

---

Diagnostics Check	OK	08:07	
Air Blank	0.000	08:07	
Control Test	0.078	08:07	
Air Blank	0.000	08:08	
Subject Sample	# 1	08:09	Subject Test Refused
Breath Volume	0.000	08:09	
Air Blank	0.000	08:10	
Control Test	0.079	08:10	
Air Blank	0.000	08:10	
Diagnostics Check	OK	08:10	

---

Data Download Date/Time: 07/28/2017 15:26

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RUBIN,RICHARD,D  
BREATH TEST OPERATOR AGENCY: Ormond Beach PD  
ARRESTING OFFICER: MOODY,S  
ARRESTING OFFICER AGENCY: Ormond Beach PD  
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:28  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: NURELL

DATE/TIME OF TEST: 07/31/2017 03:45:26

---

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.254	03:40
Breath Volume		2.378	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.250	03:43
Breath Volume		2.367	03:43
Air Blank		0.000	03:44
Control Test		0.079	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

---

Data Download Date/Time: 08/27/2017 09:50

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BELL,ETHAN,C

BREATH TEST OPERATOR AGENCY: Clermont PD

ARRESTING OFFICER: BELL,ETHAN

ARRESTING OFFICER AGENCY: Clermont PD

LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 477627

DRY GAS EXPIRY DATE: 06/29/2019

VIOLATION CODE: DUI

ARREST TIME: 01:07

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BLINCO

DATE/TIME OF TEST: 07/22/2017 02:55:04

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:49
Control Test	0.080	02:49
Air Blank	0.000	02:50
Subject Sample # 1	0.229	02:50
Breath Volume	2.382	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:52
Subject Sample # 2	0.219	02:53
Breath Volume	1.382	02:53
Air Blank	0.000	02:54
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:55

---

Data Download Date/Time: 07/26/2017 11:27

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001174

INSTRUMENT REGISTERED TO: CLERMONT P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARPER,JIMMIE,L  
BREATH TEST OPERATOR AGENCY: Clermont PD  
ARRESTING OFFICER: HARPER,J  
ARRESTING OFFICER AGENCY: Clermont PD  
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 477627  
DRY GAS EXPIRY DATE: 06/29/2019  
VIOLATION CODE: COURT ORD  
ARREST TIME: 19:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SKEETE

DATE/TIME OF TEST: 07/25/2017 19:31:24

---

Diagnostics Check	OK	19:24
Air Blank	0.000	19:25
Control Test	0.081	19:25
Air Blank	0.000	19:26
Subject Sample # 1	0.000	19:26
Breath Volume	2.351	19:26
Air Blank	0.000	19:27
Air Blank	0.000	19:29
Subject Sample # 2	0.000	19:30
Breath Volume	2.320	19:30
Air Blank	0.000	19:30
Control Test	0.081	19:30
Air Blank	0.000	19:31
Diagnostics Check	OK	19:31

---

Data Download Date/Time: 07/26/2017 11:27

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYNAUD,GEORGE,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BLANCO,JUAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 00:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: BARTET

DATE/TIME OF TEST: 07/01/2017 00:30:05

---

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.079	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:28	Subject Test Refused
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Control Test	0.079	00:29	
Air Blank	0.000	00:29	
Diagnostics Check	OK	00:30	

---

Data Download Date/Time: 07/18/2017 12:16



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYNAUD,GEORGE,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BLANCO,JUAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 15:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:17

SUBJECT'S LAST NAME: GUANCHE TORIBIO

DATE/TIME OF TEST: 07/16/2017 15:44:44

---

Diagnostics Check	OK	15:37
Air Blank	0.000	15:37
Control Test	0.080	15:38
Air Blank	0.000	15:38
Subject Sample # 1	0.168	15:39
Breath Volume	1.503	15:39
Air Blank	0.000	15:40
Air Blank	0.000	15:41
Subject Sample # 2	0.164	15:43
Breath Volume	2.843	15:43
Air Blank	0.000	15:43
Control Test	0.078	15:44
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

---

Data Download Date/Time: 07/18/2017 12:16

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYNAUD,GEORGE,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: PEREZ,ADRIAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/18/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:03

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/23/2017 16:29:39

---

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.077	16:24
Air Blank	0.000	16:24
Subject Sample # 1	0.142	16:25
Breath Volume	3.269	16:25
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.135	16:28
Breath Volume	2.175	16:28
Air Blank	0.000	16:28
Control Test	0.078	16:29
Air Blank	0.000	16:29
Diagnostics Check	OK	16:29

---

Data Download Date/Time: 08/16/2017 11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: REYNAUD,GEORGE,  
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)  
ARRESTING OFFICER: BLANCO,JUAN  
ARRESTING OFFICER AGENCY: FFWCC (West Palm)  
LAST AGENCY INSPECTION: 07/18/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/24/2019  
VIOLATION CODE: BUI  
ARREST TIME: 16:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: GUERRA VICTORIA

DATE/TIME OF TEST: 07/29/2017 16:52:58

---

Diagnostics Check	OK	16:48	
Air Blank	0.000	16:49	
Control Test	0.079	16:49	
Air Blank	0.000	16:50	
Subject Sample	# 1	16:51	Subject Test Refused
Breath Volume	0.000	16:51	
Air Blank	0.000	16:52	
Control Test	0.079	16:52	
Air Blank	0.000	16:52	
Diagnostics Check	OK	16:52	

---

Data Download Date/Time: 08/16/2017 11:44

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTRO,DENNIS,  
BREATH TEST OPERATOR AGENCY: Boynton Beach PD  
ARRESTING OFFICER: CASTRO,DENNIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:52  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 07/04/2017 23:41:53

---

Diagnostics Check	OK	23:36
Air Blank	0.000	23:36
Control Test	0.079	23:36
Air Blank	0.000	23:37
Subject Sample # 1	0.186	23:37
Breath Volume	1.257	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.184	23:40
Breath Volume	1.601	23:40
Air Blank	0.000	23:41
Control Test	0.079	23:41
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

---

Data Download Date/Time: 07/20/2017 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTRO,DENNIS,  
BREATH TEST OPERATOR AGENCY: Boynton Beach PD  
ARRESTING OFFICER: CASTRO,DENNIS  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 18:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 07/12/2017 18:43:59

---

Diagnostics Check	OK	18:38
Air Blank	0.000	18:38
Control Test	0.080	18:38
Air Blank	0.000	18:39
Subject Sample # 1	0.000	18:39
Breath Volume	3.585	18:39
Air Blank	0.000	18:40
Air Blank	0.000	18:42
Subject Sample # 2	0.000	18:42
Breath Volume	4.480	18:42
Air Blank	0.000	18:43
Control Test	0.079	18:43
Air Blank	0.000	18:43
Diagnostics Check	OK	18:43

---

Data Download Date/Time: 07/20/2017 22:29

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CASTRO,DENNIS,  
BREATH TEST OPERATOR AGENCY: Boynton Beach PD  
ARRESTING OFFICER: HALPERN,ZACHARY  
ARRESTING OFFICER AGENCY: Boynton Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 390314  
DRY GAS EXPIRY DATE: 04/19/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:56  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: SEXTON

DATE/TIME OF TEST: 07/20/2017 00:37:21

---

Diagnostics Check	OK	00:33	
Air Blank	0.000	00:33	
Control Test	0.079	00:33	
Air Blank	0.000	00:34	
Subject Sample	# 1	00:36	Subject Test Refused
Breath Volume	0.000	00:36	
Air Blank	0.000	00:36	
Control Test	0.079	00:36	
Air Blank	0.000	00:37	
Diagnostics Check	OK	00:37	

---

Data Download Date/Time: 07/20/2017 22:29

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,JACOB,A  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MOORE,JACOB  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 12:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:35

SUBJECT'S LAST NAME: LIEPSHUTZ

DATE/TIME OF TEST: 07/01/2017 13:02:33

---

Diagnostics Check	OK	12:56
Air Blank	0.000	12:57
Control Test	0.081	12:57
Air Blank	0.000	12:57
Subject Sample # 1	0.000	12:58
Breath Volume	2.914	12:58
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.000	13:01
Breath Volume	2.496	13:01
Air Blank	0.000	13:01
Control Test	0.083	13:01
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

---

Data Download Date/Time: 08/07/2017 10:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BROOKS, BLAKE, T  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: BROOKS, BLAKE  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:24  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/13/2017 02:51:21

---

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.081	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.158	02:46
Breath Volume	3.800	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.161	02:49
Breath Volume	4.007	02:49
Air Blank	0.000	02:50
Control Test	0.082	02:50
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

---

Data Download Date/Time: 08/07/2017 10:07



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER,DARYL,R  
BREATH TEST OPERATOR AGENCY: Panama City Beach PD  
ARRESTING OFFICER: MILLER,DARYL  
ARRESTING OFFICER AGENCY: Panama City Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:20

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/18/2017 15:48:05

---

Diagnostics Check	OK	15:41
Air Blank	0.000	15:42
Control Test	0.081	15:42
Air Blank	0.000	15:42
Subject Sample # 1	0.223	15:43
Breath Volume	3.324	15:43
Air Blank	0.000	15:44
Air Blank	0.000	15:45
Subject Sample # 2	0.220	15:46
Breath Volume	4.730	15:46
Air Blank	0.000	15:47
Control Test	0.083	15:47
Air Blank	0.000	15:48
Diagnostics Check	OK	15:48

---

Data Download Date/Time: 08/07/2017 10:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WARGO,TIMOTHY,M  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: WARGO,TIMOTHY  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 14:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: ROBARE

DATE/TIME OF TEST: 07/07/2017 18:15:40

---

Diagnostics Check	OK	18:09
Air Blank	0.000	18:09
Control Test	0.079	18:09
Air Blank	0.000	18:10
Subject Sample # 1	0.040	18:11
Breath Volume	2.882	18:11
Air Blank	0.000	18:11
Air Blank	0.000	18:13
Subject Sample # 2	0.039	18:14
Breath Volume	2.597	18:14
Air Blank	0.000	18:14
Control Test	0.078	18:15
Air Blank	0.000	18:15
Diagnostics Check	OK	18:15

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEROSA,ANNAMARIE,C  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DEROSA,ANNAMARIE  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 07/08/2017 01:27:01

---

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.079	01:19
Air Blank		0.000	01:19
Subject Sample	# 1	0.141	01:22
Breath Volume		1.105	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.139	01:25
Breath Volume		1.292	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:27

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALES,Y,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MILLER,A  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 07/09/2017 00:00:53

---

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:53	
Control Test	0.079	23:53	
Air Blank	0.000	23:54	
Subject Sample # 1	0.153	23:57	
Breath Volume	1.371	23:57	
Air Blank	0.000	23:57	
Air Blank		23:59	Improper Sample
Air Blank	0.000	23:59	
Control Test	0.079	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALES,Y,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MILLER,A  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 07/09/2017 00:16:56

---

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:04	
Control Test		0.078	00:05	
Air Blank		0.000	00:05	
Subject Sample	# 1	0.128	00:08	Volume Not Met
Breath Volume		0.656	00:08	
Air Blank		0.000	00:09	
Air Blank		0.000	00:11	
Subject Sample	# 2	0.139	00:11	
Breath Volume		1.777	00:11	
Air Blank		0.000	00:12	
Air Blank		0.000	00:14	
Subject Sample	# 3	0.145	00:15	
Breath Volume		1.488	00:15	
Air Blank		0.000	00:16	
Control Test		0.078	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: HEMRICK

DATE/TIME OF TEST: 07/09/2017 19:58:53

---

Diagnostics Check	OK	19:52
Air Blank	0.000	19:52
Control Test	0.078	19:53
Air Blank	0.000	19:53
Subject Sample # 1	0.249	19:54
Breath Volume	3.832	19:54
Air Blank	0.000	19:55
Air Blank	0.000	19:56
Subject Sample # 2	0.246	19:57
Breath Volume	2.945	19:57
Air Blank	0.000	19:57
Control Test	0.078	19:58
Air Blank	0.000	19:58
Diagnostics Check	OK	19:58

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOTO,SAMUEL,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KAPLAN,S  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 12:01  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 12:55

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/11/2017 13:27:15

---

Diagnostics Check	OK	13:20
Air Blank	0.000	13:21
Control Test	0.079	13:21
Air Blank	0.000	13:22
Subject Sample # 1	0.210	13:22
Breath Volume	1.472	13:22
Air Blank	0.000	13:23
Air Blank	0.000	13:25
Subject Sample # 2	0.220	13:25
Breath Volume	1.855	13:25
Air Blank	0.000	13:26
Control Test	0.079	13:26
Air Blank	0.000	13:27
Diagnostics Check	OK	13:27

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: OUELLETTE,DIANA,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: GANGL,BRADLEY  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 11:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:40

SUBJECT'S LAST NAME: ARIAS

DATE/TIME OF TEST: 07/12/2017 12:14:32

---

Diagnostics Check	OK	12:08
Air Blank	0.000	12:08
Control Test	0.079	12:08
Air Blank	0.000	12:09
Subject Sample # 1	0.141	12:09
Breath Volume	4.195	12:09
Air Blank	0.000	12:10
Air Blank	0.000	12:12
Subject Sample # 2	0.143	12:12
Breath Volume	4.070	12:12
Air Blank	0.000	12:13
Control Test	0.078	12:13
Air Blank	0.000	12:14
Diagnostics Check	OK	12:14

---

Data Download Date/Time: 08/02/2017 08:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DEMONBREUN,ANTHONY,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BAXTER,JUSTIN  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 14:21  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 15:03

SUBJECT'S LAST NAME: PLAMONDON

DATE/TIME OF TEST: 07/13/2017 15:35:17

---

Diagnostics Check	OK	15:29
Air Blank	0.000	15:29
Control Test	0.079	15:30
Air Blank	0.000	15:30
Subject Sample # 1	0.000	15:30
Breath Volume	2.125	15:30
Air Blank	0.000	15:31
Air Blank	0.000	15:33
Subject Sample # 2	0.000	15:33
Breath Volume	1.855	15:33
Air Blank	0.000	15:34
Control Test	0.078	15:34
Air Blank	0.000	15:35
Diagnostics Check	OK	15:35

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRISCOLL,JAMES,P  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DRISCOLL,JAMES  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: KGANELA TIRO

DATE/TIME OF TEST: 07/13/2017 17:15:26

---

Diagnostics Check	OK	17:09	
Air Blank	0.000	17:10	
Control Test	0.078	17:10	
Air Blank	0.000	17:10	
Subject Sample	# 1 0.214	17:13	Volume Not Met
Breath Volume	0.527	17:13	
Air Blank		17:14	Purge Fail
Air Blank	0.000	17:15	

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DRISCOLL,JAMES,P  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DRISCOLL,JAMES  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:48

SUBJECT'S LAST NAME: KGANELA TIRO

DATE/TIME OF TEST: 07/13/2017 17:17:49

---

Diagnostics Check	OK	17:16	
Air Blank	0.000	17:16	
Control Test	0.071	17:17	Control Outside Tolerance
Air Blank	0.000	17:17	

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BUONPANE,V,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DUSCH,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: PROB/PAROL  
ARREST TIME: 18:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/14/2017 19:19:20

---

Diagnostics Check	OK	19:12
Air Blank	0.000	19:12
Control Test	0.079	19:12
Air Blank	0.000	19:13
Subject Sample # 1	0.059	19:14
Breath Volume	1.117	19:14
Air Blank	0.000	19:14
Air Blank	0.000	19:16
Subject Sample # 2	0.064	19:17
Breath Volume	1.261	19:17
Air Blank	0.000	19:18
Control Test	0.079	19:18
Air Blank	0.000	19:19
Diagnostics Check	OK	19:19

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,RASHAAD,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SMITH,RASHAAD  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: TOMMASI

DATE/TIME OF TEST: 07/14/2017 23:06:38

---

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.079	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.151	23:01
Breath Volume	1.683	23:01
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.140	23:05
Breath Volume	1.277	23:05
Air Blank	0.000	23:05
Control Test	0.079	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNORAT,ADELY,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DELOMBA,G  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: DEMATTEO

DATE/TIME OF TEST: 07/16/2017 00:19:43

---

Diagnostics Check	OK	00:15	
Air Blank	0.000	00:16	
Control Test	0.078	00:16	
Air Blank	0.000	00:16	
Subject Sample	# 1	00:19	Sequence Aborted
Breath Volume	0.117	00:19	
Air Blank	0.000	00:19	

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANNORAT,ADELY,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DELOMBA,G  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: DEMATTEO

DATE/TIME OF TEST: 07/16/2017 00:32:00

---

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.078	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.163	00:26
Breath Volume	2.406	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.147	00:30
Breath Volume	1.179	00:30
Air Blank	0.000	00:31
Control Test	0.077	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:32

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVAS,W,R  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MCCOY,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: GUINN

DATE/TIME OF TEST: 07/17/2017 21:36:17

---

Diagnostics Check	OK	21:30
Air Blank	0.000	21:30
Control Test	0.078	21:30
Air Blank	0.000	21:31
Subject Sample # 1	0.245	21:31
Breath Volume	2.914	21:31
Air Blank	0.000	21:32
Air Blank	0.000	21:34
Subject Sample # 2	0.231	21:34
Breath Volume	2.156	21:34
Air Blank	0.000	21:35
Control Test	0.078	21:35
Air Blank	0.000	21:36
Diagnostics Check	OK	21:36

---

Data Download Date/Time: 08/02/2017 08:38



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FUQUEN,JENNIFER,M  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KIRKLAND,R  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 15:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: BILLIE

DATE/TIME OF TEST: 07/19/2017 17:13:03

---

Diagnostics Check	OK	17:06
Air Blank	0.000	17:07
Control Test	0.079	17:07
Air Blank	0.000	17:07
Subject Sample # 1	0.228	17:08
Breath Volume	1.894	17:08
Air Blank	0.000	17:09
Air Blank	0.000	17:10
Subject Sample # 2	0.212	17:11
Breath Volume	1.769	17:11
Air Blank	0.000	17:12
Control Test	0.078	17:12
Air Blank	0.000	17:12
Diagnostics Check	OK	17:13

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GREISHEIMER,JOHN,V  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: GREISHEIMER,JOHN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: CARLSON JR

DATE/TIME OF TEST: 07/21/2017 23:17:26

---

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.079	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.130	23:12
Breath Volume	1.777	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.119	23:15
Breath Volume	1.773	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: VENTURA

DATE/TIME OF TEST: 07/22/2017 20:45:33

---

Diagnostics Check	OK	20:38
Air Blank	0.000	20:38
Control Test	0.079	20:39
Air Blank	0.000	20:39
Subject Sample # 1	0.163	20:41
Breath Volume	1.886	20:41
Air Blank	0.000	20:41
Air Blank	0.000	20:43
Subject Sample # 2	0.157	20:44
Breath Volume	2.035	20:44
Air Blank	0.000	20:44
Control Test	0.079	20:44
Air Blank	0.000	20:45
Diagnostics Check	OK	20:45

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ASHLINE,C,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MCGFIVREY,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 07/23/2017 02:46:51

---

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.154	02:42
Breath Volume	2.167	02:42
Air Blank	0.000	02:43
Air Blank	0.000	02:44
Subject Sample # 2	0.143	02:45
Breath Volume	1.894	02:45
Air Blank	0.000	02:45
Control Test	0.078	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 08/02/2017 08:38

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NOEL,Y,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: DERRIG,J  
ARRESTING OFFICER AGENCY: Marco Island PD  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CASTELLANOS

DATE/TIME OF TEST: 07/23/2017 03:59:48

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.160	03:54
Breath Volume	1.660	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.160	03:58
Breath Volume	1.957	03:58
Air Blank	0.000	03:58
Control Test	0.078	03:59
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DECKER,STEPHEN,J  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CRISP,D  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CORTES REINA

DATE/TIME OF TEST: 07/24/2017 02:10:06

---

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.079	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.140	02:05
Breath Volume		1.109	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.133	02:08
Breath Volume		1.300	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:10
Diagnostics Check		OK	02:10

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CORNELIUS,TIM,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORNELIUS,TIM  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: LEIB

DATE/TIME OF TEST: 07/24/2017 02:59:20

---

Diagnostics Check	OK	02:52
Air Blank	0.000	02:53
Control Test	0.079	02:53
Air Blank	0.000	02:54
Subject Sample # 1	0.108	02:54
Breath Volume	1.339	02:54
Air Blank	0.000	02:55
Air Blank	0.000	02:57
Subject Sample # 2	0.107	02:57
Breath Volume	2.324	02:57
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Diagnostics Check	OK	02:59

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BELL,ONDREA,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: BOTELLO,A  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: WEISZ

DATE/TIME OF TEST: 07/26/2017 05:48:39

---

Diagnostics Check		OK	05:42
Air Blank		0.000	05:42
Control Test		0.078	05:42
Air Blank		0.000	05:43
Subject Sample	# 1	0.260	05:44
Breath Volume		2.238	05:44
Air Blank		0.000	05:44
Air Blank		0.000	05:46
Subject Sample	# 2	0.260	05:47
Breath Volume		2.238	05:47
Air Blank		0.000	05:47
Control Test		0.078	05:48
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

---

Data Download Date/Time: 08/02/2017 08:38



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KENNEDY,DANIEL,L  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: KENNEDY,DANIEL  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:33  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 07/27/2017 23:31:13

---

Diagnostics Check	OK	23:25
Air Blank	0.000	23:25
Control Test	0.080	23:26
Air Blank	0.000	23:26
Subject Sample # 1	0.196	23:26
Breath Volume	1.472	23:26
Air Blank	0.000	23:27
Air Blank	0.000	23:29
Subject Sample # 2	0.186	23:29
Breath Volume	1.269	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Diagnostics Check	OK	23:31

---

Data Download Date/Time: 08/02/2017 08:38

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOGAN,T,  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: SOLIS,E  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 07/29/2017 03:48:50

---

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.138	03:44
Breath Volume	2.679	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.134	03:47
Breath Volume	2.726	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

---

Data Download Date/Time: 08/02/2017 08:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURTSINGER,CODY,A  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: CORWIN,J  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: FULOP

DATE/TIME OF TEST: 07/30/2017 04:57:58

---

Diagnostics Check	OK	04:51
Air Blank	0.000	04:52
Control Test	0.079	04:52
Air Blank	0.000	04:53
Subject Sample # 1	0.000	04:53
Breath Volume	3.574	04:53
Air Blank	0.000	04:54
Air Blank	0.000	04:56
Subject Sample # 2	0.000	04:56
Breath Volume	3.003	04:56
Air Blank	0.000	04:57
Control Test	0.078	04:57
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

---

Data Download Date/Time: 08/02/2017 08:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001201

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CURTO,THEODORE,D  
BREATH TEST OPERATOR AGENCY: Collier County SO  
ARRESTING OFFICER: MCCORMICK,PATRICK  
ARRESTING OFFICER AGENCY: Collier County SO  
LAST AGENCY INSPECTION: 07/05/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 17316080A3  
DRY GAS EXPIRY DATE: 08/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 07/31/2017 00:31:58

---

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.079	00:26
Air Blank	0.000	00:26
Subject Sample # 1	0.150	00:27
Breath Volume	3.671	00:27
Air Blank	0.000	00:28
Air Blank	0.000	00:29
Subject Sample # 2	0.150	00:30
Breath Volume	2.714	00:30
Air Blank	0.000	00:31
Control Test	0.078	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

---

Data Download Date/Time: 08/02/2017 08:39

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESTER,JAMES,R  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: HESTER,JAMES  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: MONTALVO ALVARADO

DATE/TIME OF TEST: 07/01/2017 04:32:18

---

Diagnostics Check	OK	04:22
Air Blank	0.000	04:22
Control Test	0.081	04:23
Air Blank	0.000	04:23
Subject Sample # 1	0.123	04:24
Breath Volume	1.390	04:24
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.101	04:27
Breath Volume	2.074	04:27
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 3	0.102	04:30
Breath Volume	1.441	04:30
Air Blank	0.000	04:31
Control Test	0.081	04:31
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time: 07/23/2017 23:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHAMBERLAIN,STEVEN,M  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: CHAMBERLAIN,STEVEN  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:33  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/02/2017 04:11:12

---

Diagnostics Check	OK	04:04
Air Blank	0.000	04:04
Control Test	0.082	04:05
Air Blank	0.000	04:05
Subject Sample # 1	0.245	04:06
Breath Volume	1.601	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.237	04:09
Breath Volume	2.164	04:09
Air Blank	0.000	04:10
Control Test	0.081	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

---

Data Download Date/Time: 07/23/2017 23:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WILLIAMS,N,  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: WILLIAMS,N  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:24  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: DOWNHAM

DATE/TIME OF TEST: 07/02/2017 05:14:31

---

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.081	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.276	05:09
Breath Volume		1.664	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.268	05:12
Breath Volume		1.882	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

---

Data Download Date/Time: 07/23/2017 23:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: NORWOOD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 17:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: KESLER

DATE/TIME OF TEST: 07/06/2017 18:19:32

---

Diagnostics Check	OK	18:13
Air Blank	0.000	18:13
Control Test	0.081	18:14
Air Blank	0.000	18:14
Subject Sample # 1	0.065	18:15
Breath Volume	2.507	18:15
Air Blank	0.000	18:15
Air Blank	0.000	18:17
Subject Sample # 2	0.064	18:18
Breath Volume	2.148	18:18
Air Blank	0.000	18:18
Control Test	0.082	18:18
Air Blank	0.000	18:19
Diagnostics Check	OK	18:19

---

Data Download Date/Time: 07/23/2017 23:35



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MCNEIL,DARRIN,J  
BREATH TEST OPERATOR AGENCY: Clewiston PD  
ARRESTING OFFICER: MCNEIL,DARRIN  
ARRESTING OFFICER AGENCY: Clewiston PD  
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: HOLLOWAY

DATE/TIME OF TEST: 07/11/2017 03:58:47

---

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.081	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.014	03:54
Breath Volume	3.660	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.014	03:57
Breath Volume	3.738	03:57
Air Blank	0.000	03:57
Control Test	0.082	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

---

Data Download Date/Time: 07/23/2017 23:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CAMERON,DAVID,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: MCLAIN,KYLE  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 06/13/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 07/15/2017 00:21:03

---

Diagnostics Check	OK	00:12
Air Blank	0.000	00:12
Control Test	0.081	00:13
Air Blank	0.000	00:13
Subject Sample # 1	0.112	00:15
Breath Volume	1.132	00:15
Air Blank	0.000	00:16
Air Blank	0.000	00:18
Subject Sample # 2	0.114	00:19
Breath Volume	1.703	00:19
Air Blank	0.000	00:20
Control Test	0.082	00:20
Air Blank	0.000	00:20
Diagnostics Check	OK	00:21

---

Data Download Date/Time: 07/23/2017 23:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: MATEU,JOAQUIN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: VAZQUEZ-SIERRA

DATE/TIME OF TEST: 07/22/2017 00:33:51

---

Diagnostics Check	OK	00:27
Air Blank	0.000	00:27
Control Test	0.082	00:27
Air Blank	0.000	00:28
Subject Sample # 1	0.205	00:29
Breath Volume	2.585	00:29
Air Blank	0.000	00:29
Air Blank	0.000	00:31
Subject Sample # 2	0.202	00:32
Breath Volume	1.527	00:32
Air Blank	0.000	00:32
Control Test	0.081	00:33
Air Blank	0.000	00:33
Diagnostics Check	OK	00:33

---

Data Download Date/Time: 07/23/2017 23:35

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: MATEU,JOAQUIN  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: MORTON

DATE/TIME OF TEST: 07/23/2017 00:05:39

---

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:03	
Control Test	0.081	00:03	
Air Blank	0.000	00:04	
Subject Sample	# 1	00:04	Subject Test Refused
Breath Volume	0.000	00:04	
Air Blank	0.000	00:04	
Control Test	0.081	00:05	
Air Blank	0.000	00:05	
Diagnostics Check	OK	00:05	

---

Data Download Date/Time: 07/23/2017 23:35

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: NORWOOD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 07/27/2017 00:50:58

---

Diagnostics Check	OK	00:47	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:50	
Control Test	0.082	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

---

Data Download Date/Time: 08/22/2017 20:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: NORWOOD,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:47  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: DEHOYOS

DATE/TIME OF TEST: 07/28/2017 03:22:30

---

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.081	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.164	03:18
Breath Volume		2.906	03:18
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.154	03:20
Breath Volume		2.410	03:20
Air Blank		0.000	03:21
Control Test		0.081	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

---

Data Download Date/Time: 08/22/2017 20:22

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001207

INSTRUMENT REGISTERED TO: HENDRY COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NORWOOD,CHRISTOPHER,J  
BREATH TEST OPERATOR AGENCY: Hendry County SO  
ARRESTING OFFICER: CAMPBELL,SHANITA  
ARRESTING OFFICER AGENCY: Hendry County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 83551  
DRY GAS EXPIRY DATE: 10/02/2018  
VIOLATION CODE: DUI  
ARREST TIME: 05:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 07/28/2017 06:07:09

---

Diagnostics Check		OK	06:00
Air Blank		0.000	06:01
Control Test		0.081	06:01
Air Blank		0.000	06:02
Subject Sample	# 1	0.179	06:02
Breath Volume		2.175	06:02
Air Blank		0.000	06:03
Air Blank		0.000	06:05
Subject Sample	# 2	0.188	06:05
Breath Volume		3.031	06:05
Air Blank		0.000	06:06
Control Test		0.081	06:06
Air Blank		0.000	06:07
Diagnostics Check		OK	06:07

---

Data Download Date/Time: 08/22/2017 20:22

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: SCHAEFFER,KENDALL  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 07/01/2017 00:02:08

---

Diagnostics Check	OK	23:55
Air Blank	0.000	23:55
Control Test	0.079	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:57
Breath Volume	3.800	23:57
Air Blank	0.000	23:57
Air Blank	0.000	23:59
Subject Sample # 1	0.000	00:00
Breath Volume	2.222	00:00
Air Blank	0.000	00:01
Control Test	0.079	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

---

Data Download Date/Time: 07/26/2017 08:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HOGART,DEE  
ARRESTING OFFICER AGENCY: Lee Port Authority  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:38  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 07/02/2017 02:38:51

---

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.079	02:33
Air Blank	0.000	02:34
Subject Sample # 1	0.169	02:34
Breath Volume	1.652	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 2	0.163	02:37
Breath Volume	1.562	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STEVENS,LAUREN  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: AUSSCHNITT

DATE/TIME OF TEST: 07/02/2017 05:08:30

---

Diagnostics Check	OK	05:02
Air Blank	0.000	05:02
Control Test	0.078	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.090	05:04
Breath Volume	2.074	05:04
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.087	05:07
Breath Volume	2.277	05:07
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: THABIT,S  
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: MUISE

DATE/TIME OF TEST: 07/04/2017 03:52:52

---

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:50	
Control Test	0.078	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.078	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,J,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: RIVERA,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: YAT-SAMBRANO

DATE/TIME OF TEST: 07/04/2017 04:15:24

---

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.191	04:10
Breath Volume	1.417	04:10
Air Blank	0.000	04:11
Air Blank	0.000	04:13
Subject Sample # 2	0.192	04:13
Breath Volume	1.507	04:13
Air Blank	0.000	04:14
Control Test	0.077	04:14
Air Blank	0.000	04:15
Diagnostics Check	OK	04:15

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CALHOUN,K  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 07/04/2017 22:37:15

---

Diagnostics Check	OK	22:30
Air Blank	0.000	22:31
Control Test	0.078	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.186	22:32
Breath Volume	2.023	22:32
Air Blank	0.000	22:33
Air Blank	0.000	22:35
Subject Sample # 2	0.183	22:35
Breath Volume	2.289	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:37
Diagnostics Check	OK	22:37

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: EDMOND,COREY,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: EDMOND,COREY  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:09

SUBJECT'S LAST NAME: GODOY

DATE/TIME OF TEST: 07/05/2017 17:38:44

---

Diagnostics Check	OK	17:32
Air Blank	0.000	17:32
Control Test	0.079	17:33
Air Blank	0.000	17:33
Subject Sample # 1	0.182	17:34
Breath Volume	2.878	17:34
Air Blank	0.000	17:34
Air Blank	0.000	17:36
Subject Sample # 2	0.173	17:37
Breath Volume	1.699	17:37
Air Blank	0.000	17:37
Control Test	0.077	17:38
Air Blank	0.000	17:38
Diagnostics Check	OK	17:38

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BROWN,E  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: GRINOLD

DATE/TIME OF TEST: 07/08/2017 00:34:39

---

Diagnostics Check	OK	00:28
Air Blank	0.000	00:28
Control Test	0.079	00:28
Air Blank	0.000	00:29
Subject Sample # 1	0.105	00:29
Breath Volume	4.160	00:29
Air Blank	0.000	00:30
Air Blank	0.000	00:32
Subject Sample # 2	0.104	00:33
Breath Volume	3.054	00:33
Air Blank	0.000	00:33
Control Test	0.078	00:34
Air Blank	0.000	00:34
Diagnostics Check	OK	00:34

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WITT,KELLY,  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: WITT,KELLY  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:54  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:14

SUBJECT'S LAST NAME: SIMPSON-HOLLOWAY

DATE/TIME OF TEST: 07/08/2017 01:45:46

---

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.079	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.229	01:41
Breath Volume	1.695	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.220	01:44
Breath Volume	1.859	01:44
Air Blank	0.000	01:44
Control Test	0.078	01:45
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

---

Data Download Date/Time: 07/26/2017 08:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BROWN,E  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:26  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 07/08/2017 03:51:12

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.078	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.173	03:46
Breath Volume	1.679	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.172	03:49
Breath Volume	1.304	03:49
Air Blank	0.000	03:50
Control Test	0.079	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: WATSON,J,B  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: HUBATCH,N  
ARRESTING OFFICER AGENCY: Sanibel PD  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:50  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: KYLE

DATE/TIME OF TEST: 07/08/2017 06:01:32

---

Diagnostics Check	OK	05:55
Air Blank	0.000	05:55
Control Test	0.078	05:55
Air Blank	0.000	05:56
Subject Sample # 1	0.212	05:57
Breath Volume	2.023	05:57
Air Blank	0.000	05:57
Air Blank	0.000	05:59
Subject Sample # 2	0.214	06:00
Breath Volume	2.046	06:00
Air Blank	0.000	06:00
Control Test	0.078	06:00
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: NATAL,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: MAGLOTHIN

DATE/TIME OF TEST: 07/10/2017 03:36:42

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:30
Control Test	0.079	03:30
Air Blank	0.000	03:31
Subject Sample # 1	0.182	03:31
Breath Volume	4.675	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:34
Subject Sample # 2	0.183	03:35
Breath Volume	3.433	03:35
Air Blank	0.000	03:35
Control Test	0.078	03:36
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: GARCIA,J  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/12/2017 22:08:37

---

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.079	22:02
Air Blank	0.000	22:03
Subject Sample # 1	0.022	22:04
Breath Volume	1.183	22:04
Air Blank	0.000	22:04
Air Blank	0.000	22:06
Subject Sample # 2	0.022	22:07
Breath Volume	1.859	22:07
Air Blank	0.000	22:07
Control Test	0.079	22:08
Air Blank	0.000	22:08
Diagnostics Check	OK	22:08

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,J,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: RONGISH,S  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:06  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: DISERIO

DATE/TIME OF TEST: 07/13/2017 05:11:12

---

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:05
Subject Sample # 1	0.113	05:06
Breath Volume	1.312	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:09
Subject Sample # 2	0.123	05:09
Breath Volume	1.503	05:09
Air Blank	0.000	05:10
Control Test	0.078	05:10
Air Blank	0.000	05:11
Diagnostics Check	OK	05:11

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: NERONI, KRISTI, A  
BREATH TEST OPERATOR AGENCY: Fort Myers PD  
ARRESTING OFFICER: NERONI, KRISTI  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MURRIETA DA COSTA

DATE/TIME OF TEST: 07/15/2017 02:27:13

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.043	02:22
Breath Volume	1.539	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:24
Subject Sample # 2	0.040	02:25
Breath Volume	1.914	02:25
Air Blank	0.000	02:26
Control Test	0.079	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MATEO,JUAN,D  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: GONZALEZ,CARLOS  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: MATEO

DATE/TIME OF TEST: 07/15/2017 23:29:52

---

Diagnostics Check	OK	23:26	
Air Blank	0.000	23:27	
Control Test	0.079	23:27	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:28	Subject Test Refused
Breath Volume	0.000	23:28	
Air Blank	0.000	23:28	
Control Test	0.078	23:29	
Air Blank	0.000	23:29	
Diagnostics Check	OK	23:29	

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STROBEL, RAYMOND, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STROBEL, RAYMOND  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/16/2017 11:37:23

---

Diagnostics Check		OK	11:25	
Air Blank		0.000	11:26	
Control Test		0.079	11:26	
Air Blank		0.000	11:27	
Subject Sample	# 1	0.127	11:30	Volume Not Met
Breath Volume		0.242	11:30	
Air Blank		0.000	11:30	
Air Blank		0.000	11:32	
Subject Sample	# 2	0.136	11:35	Volume Not Met
Breath Volume		0.148	11:35	
Air Blank		0.000	11:36	
Control Test		0.078	11:36	
Air Blank		0.000	11:37	
Diagnostics Check		OK	11:37	

---

Data Download Date/Time: 07/26/2017 08:58



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SCHAFFER, THEODORE, E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: STROBEL, R  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/16/2017 11:59:04

---

Diagnostics Check		OK	11:49	
Air Blank		0.000	11:49	
Control Test		0.078	11:50	
Air Blank		0.000	11:50	
Subject Sample	# 1	0.133	11:53	Volume Not Met
Breath Volume		0.015	11:53	
Air Blank		0.000	11:54	
Air Blank		0.000	11:56	
Subject Sample	# 2		11:57	Subject Test Refused
Breath Volume		0.808	11:57	
Air Blank		0.000	11:58	
Control Test		0.078	11:58	
Air Blank		0.000	11:58	
Diagnostics Check		OK	11:59	

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUGO,ADRIAN,  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LUGO,ADRIAN  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: BUI  
ARREST TIME: 02:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 07/17/2017 03:16:39

---

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.078	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1 0.199	03:12	
Breath Volume	1.460	03:12	
Air Blank	0.000	03:12	
Air Blank	0.000	03:14	
Subject Sample	# 2	03:15	Subject Test Refused
Breath Volume	0.000	03:15	
Air Blank	0.000	03:15	
Control Test	0.078	03:16	
Air Blank	0.000	03:16	
Diagnostics Check	OK	03:16	

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: NATAL,J  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 19:08  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: MURGA

DATE/TIME OF TEST: 07/17/2017 20:09:19

---

Diagnostics Check	OK	20:02
Air Blank	0.000	20:03
Control Test	0.078	20:03
Air Blank	0.000	20:03
Subject Sample # 1	0.149	20:04
Breath Volume	1.531	20:04
Air Blank	0.000	20:05
Air Blank	0.000	20:07
Subject Sample # 2	0.146	20:07
Breath Volume	1.585	20:07
Air Blank	0.000	20:08
Control Test	0.078	20:08
Air Blank	0.000	20:09
Diagnostics Check	OK	20:09

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: BROWN,E  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: HONORAT

DATE/TIME OF TEST: 07/22/2017 03:02:00

---

Diagnostics Check	OK	02:53
Air Blank	0.000	02:54
Control Test	0.078	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.203	02:56
Breath Volume	2.304	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.191	03:00
Breath Volume	1.519	03:00
Air Blank	0.000	03:01
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,J,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: RIVERA,J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: FORTUNATO-DASILVA

DATE/TIME OF TEST: 07/22/2017 03:53:37

---

Diagnostics Check		OK	03:44	
Air Blank		0.000	03:44	
Control Test		0.078	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 1	0.164	03:48	Volume Not Met
Breath Volume		0.015	03:48	
Air Blank		0.000	03:49	
Air Blank		0.000	03:50	
Subject Sample	# 2		03:52	Subject Test Refused
Breath Volume		0.054	03:52	
Air Blank		0.000	03:52	
Control Test		0.077	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,J,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: LOOS,S  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: SARCINELLI

DATE/TIME OF TEST: 07/22/2017 06:03:41

---

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.078	06:01	
Air Blank	0.000	06:02	
Subject Sample	# 1	06:02	Subject Test Refused
Breath Volume	0.000	06:02	
Air Blank	0.000	06:02	
Control Test	0.079	06:03	
Air Blank	0.000	06:03	
Diagnostics Check	OK	06:03	

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: NAYLOR,M  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:02

SUBJECT'S LAST NAME: BOYCE

DATE/TIME OF TEST: 07/22/2017 21:32:08

---

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Subject Sample # 1	0.009	21:27
Breath Volume	4.109	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.009	21:30
Breath Volume	3.515	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: THABIT,S  
ARRESTING OFFICER AGENCY: FL Gulf Coast UPD  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: THALER

DATE/TIME OF TEST: 07/23/2017 03:56:50

---

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.152	03:52
Breath Volume		5.125	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.152	03:55
Breath Volume		5.066	03:55
Air Blank		0.000	03:55
Control Test		0.077	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

---

Data Download Date/Time: 07/26/2017 08:58



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH,JOSIAH,R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: SMITH,JOSIAH  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: TOJIN

DATE/TIME OF TEST: 07/23/2017 22:58:18

---

Diagnostics Check	OK	22:51
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.150	22:53
Breath Volume	1.253	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.162	22:56
Breath Volume	1.210	22:56
Air Blank	0.000	22:57
Control Test	0.079	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

---

Data Download Date/Time: 07/26/2017 08:58

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SMITH, JOSIAH, R  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: RIVERA, J  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/24/2017 04:00:14

---

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.219	03:55
Breath Volume	1.531	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.212	03:58
Breath Volume	1.445	03:58
Air Blank	0.000	03:59
Control Test	0.077	03:59
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

---

Data Download Date/Time: 07/26/2017 08:58

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RIVERA,J,A  
BREATH TEST OPERATOR AGENCY: FHP Troop F  
ARRESTING OFFICER: RONGISH,S  
ARRESTING OFFICER AGENCY: FHP Troop F  
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:36

SUBJECT'S LAST NAME: MARSALA

DATE/TIME OF TEST: 07/27/2017 18:17:47

---

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.167	18:12
Breath Volume	2.558	18:12
Air Blank	0.000	18:13
Air Blank	0.000	18:14
Subject Sample # 2	0.164	18:16
Breath Volume	2.285	18:16
Air Blank	0.000	18:16
Control Test	0.078	18:17
Air Blank	0.000	18:17
Diagnostics Check	OK	18:17

---

Data Download Date/Time: 08/09/2017 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MEYERS,F,E  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: CALHOUN,K  
ARRESTING OFFICER AGENCY: Lee County SO  
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 07/28/2017 01:44:32

---

Diagnostics Check	OK	01:38
Air Blank	0.000	01:38
Control Test	0.079	01:39
Air Blank	0.000	01:39
Subject Sample # 1	0.000	01:40
Breath Volume	3.058	01:40
Air Blank	0.000	01:40
Air Blank	0.000	01:42
Subject Sample # 2	0.000	01:43
Breath Volume	3.128	01:43
Air Blank	0.000	01:43
Control Test	0.079	01:43
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

---

Data Download Date/Time: 08/09/2017 09:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001208

INSTRUMENT REGISTERED TO: LEE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: AVENDANO,JUAN,M  
BREATH TEST OPERATOR AGENCY: Lee County SO  
ARRESTING OFFICER: DAVILA,MARTIN  
ARRESTING OFFICER AGENCY: Fort Myers PD  
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 34416080A2  
DRY GAS EXPIRY DATE: 02/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: UGARTE

DATE/TIME OF TEST: 07/30/2017 00:05:05

---

Diagnostics Check		OK	23:52	
Air Blank		0.000	23:53	
Control Test		0.079	23:53	
Air Blank		0.000	23:54	
Subject Sample	# 3	0.142	23:57	Volume Not Met
Breath Volume		0.398	23:57	
Air Blank		0.000	23:57	
Air Blank		0.000	23:59	
Subject Sample	# 1	0.152	00:00	
Breath Volume		1.285	00:00	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 2	0.139	00:03	
Breath Volume		1.390	00:03	
Air Blank		0.000	00:04	
Control Test		0.079	00:04	
Air Blank		0.000	00:05	
Diagnostics Check		OK	00:05	

---

Data Download Date/Time: 08/09/2017 09:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KING,DONALD,E  
BREATH TEST OPERATOR AGENCY: Auburndale PD  
ARRESTING OFFICER: SKINNER,CODY  
ARRESTING OFFICER AGENCY: Auburndale PD  
LAST AGENCY INSPECTION: 07/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A3  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: WEIAND

DATE/TIME OF TEST: 07/31/2017 22:20:53

---

Diagnostics Check	OK	22:09	
Air Blank	0.000	22:09	
Control Test	0.078	22:10	
Air Blank	0.000	22:10	
Subject Sample	# 1	22:13	No Sample Provided
Breath Volume	0.000	22:13	
Air Blank	0.000	22:14	
Air Blank	0.000	22:16	
Subject Sample	# 2	22:19	No Sample Provided
Breath Volume	0.000	22:19	
Air Blank	0.000	22:19	
Control Test	0.079	22:20	
Air Blank	0.000	22:20	
Diagnostics Check	OK	22:20	

---

Data Download Date/Time: 08/30/2017 22:39

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DONDE,J,  
BREATH TEST OPERATOR AGENCY: West Palm Beach PD  
ARRESTING OFFICER: DONDE,J  
ARRESTING OFFICER AGENCY: West Palm Beach PD  
LAST AGENCY INSPECTION: 07/20/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 152169  
DRY GAS EXPIRY DATE: 10/30/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 16:45

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/30/2017 17:31:09

---

Diagnostics Check	OK	17:18	
Air Blank	0.000	17:19	
Control Test	0.081	17:19	
Air Blank	0.000	17:20	
Subject Sample	# 1	17:23	No Sample Provided
Breath Volume	0.000	17:23	
Air Blank	0.000	17:23	
Air Blank	0.000	17:25	
Subject Sample	# 2	0.000	17:26
Breath Volume	1.519	17:26	
Air Blank	0.000	17:27	
Air Blank	0.000	17:29	
Subject Sample	# 3	0.000	17:29
Breath Volume	1.695	17:29	
Air Blank	0.000	17:30	
Control Test	0.080	17:30	
Air Blank	0.000	17:31	
Diagnostics Check	OK	17:31	

---

Data Download Date/Time: 09/01/2017 08:05

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARRELL,JEREMY,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: FARRELL,JEREMY  
ARRESTING OFFICER AGENCY: Polk County SO  
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 13:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 07/15/2017 16:31:45

---

Diagnostics Check	OK	16:25
Air Blank	0.000	16:26
Control Test	0.081	16:26
Air Blank	0.000	16:26
Subject Sample # 1	0.000	16:27
Breath Volume	1.148	16:27
Air Blank	0.000	16:27
Air Blank	0.000	16:29
Subject Sample # 2	0.000	16:30
Breath Volume	3.789	16:30
Air Blank	0.000	16:30
Control Test	0.079	16:31
Air Blank	0.000	16:31
Diagnostics Check	OK	16:31

---

Data Download Date/Time: 08/04/2017 13:28



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001237

INSTRUMENT REGISTERED TO: POLK COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FARRELL,JEREMY,D  
BREATH TEST OPERATOR AGENCY: Polk County SO  
ARRESTING OFFICER: YOUNG,JOHNATHAN  
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)  
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A3  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 19:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: KINARD

DATE/TIME OF TEST: 07/15/2017 20:47:52

---

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.079	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.260	20:39
Breath Volume	1.871	20:39
Air Blank	0.000	20:40
Air Blank	0.000	20:41
Subject Sample # 2	0.222	20:43
Breath Volume	1.617	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 3	0.228	20:46
Breath Volume	1.523	20:46
Air Blank	0.000	20:46
Control Test	0.076	20:47
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

---

Data Download Date/Time: 08/04/2017 13:28

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: PITCHER,DARRYL  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/07/2017 21:40:17

---

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.079	21:34
Air Blank	0.000	21:35
Subject Sample # 1	0.026	21:35
Breath Volume	3.656	21:35
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.024	21:38
Breath Volume	1.585	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

---

Data Download Date/Time: 07/21/2017 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PITCHER,DARRYL,P  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: CHESSHER,SANDRA  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:09  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PONTON

DATE/TIME OF TEST: 07/09/2017 01:38:41

---

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.077	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1	0.088	01:30	Volume Not Met
Breath Volume		0.472	01:30	
Air Blank		0.000	01:30	
Air Blank		0.000	01:32	
Subject Sample	# 2	0.094	01:33	
Breath Volume		2.339	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 3	0.095	01:37	
Breath Volume		2.089	01:37	
Air Blank		0.000	01:37	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Diagnostics Check		OK	01:38	

---

Data Download Date/Time: 07/21/2017 18:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: LUPPINO,SETH  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:48  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LAFAVE

DATE/TIME OF TEST: 07/09/2017 04:13:18

---

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.077	04:07	
Air Blank	0.000	04:08	
Subject Sample # 1	0.051	04:09	
Breath Volume	1.367	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:11	
Subject Sample # 2	0.054	04:12	
Breath Volume	2.042	04:12	
Air Blank		04:12	RFI Detect
Air Blank	0.000	04:13	

---

Data Download Date/Time: 07/21/2017 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: LUPPINO,SETH  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LAFAVE

DATE/TIME OF TEST: 07/09/2017 04:21:16

---

Diagnostics Check	OK	04:15	
Air Blank	0.000	04:15	
Control Test	0.078	04:16	
Air Blank	0.000	04:16	
Subject Sample # 1	0.049	04:17	
Breath Volume	2.070	04:17	
Air Blank	0.000	04:18	
Air Blank	0.000	04:20	
Subject Sample # 2	0.047	04:20	
Breath Volume	1.894	04:20	
Air Blank		04:20	RFI Detect
Air Blank	0.000	04:21	

---

Data Download Date/Time: 07/21/2017 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CHESSHER,SANDRA,W  
BREATH TEST OPERATOR AGENCY: Lake County SO  
ARRESTING OFFICER: LUPPINO,SETH  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: LAFAVE

DATE/TIME OF TEST: 07/09/2017 04:32:39

---

Diagnostics Check	OK	04:26
Air Blank	0.000	04:26
Control Test	0.080	04:27
Air Blank	0.000	04:27
Subject Sample # 1	0.046	04:28
Breath Volume	1.941	04:28
Air Blank	0.000	04:28
Air Blank	0.000	04:30
Subject Sample # 2	0.046	04:31
Breath Volume	1.773	04:31
Air Blank	0.000	04:31
Control Test	0.081	04:32
Air Blank	0.000	04:32
Diagnostics Check	OK	04:32

---

Data Download Date/Time: 07/21/2017 18:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001238

INSTRUMENT REGISTERED TO: LAKE COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROBERTS,DOUG,J  
BREATH TEST OPERATOR AGENCY: Tavares PD  
ARRESTING OFFICER: CHESSHER,S  
ARRESTING OFFICER AGENCY: Lake County SO  
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A7  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:51  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: LACEY

DATE/TIME OF TEST: 07/18/2017 03:02:40

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.139	02:58
Breath Volume	2.035	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.128	03:01
Breath Volume	1.296	03:01
Air Blank	0.000	03:01
Control Test	0.085	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 07/21/2017 18:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MATTHEW, R  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: SNYDER, MATTHEW  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 765462  
DRY GAS EXPIRY DATE: 04/13/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: MANN

DATE/TIME OF TEST: 07/25/2017 23:15:33

---

Diagnostics Check	OK	23:12	
Air Blank	0.000	23:13	
Control Test	0.080	23:13	
Air Blank	0.000	23:13	
Subject Sample	# 1	23:14	Subject Test Refused
Breath Volume	0.000	23:14	
Air Blank	0.000	23:14	
Control Test	0.081	23:14	
Air Blank	0.000	23:15	
Diagnostics Check	OK	23:15	

---

Data Download Date/Time: 08/28/2017 07:03



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SNYDER, MATTHEW, R  
BREATH TEST OPERATOR AGENCY: Edgewater PD  
ARRESTING OFFICER: SNYDER, MATTHEW  
ARRESTING OFFICER AGENCY: Edgewater PD  
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 765462  
DRY GAS EXPIRY DATE: 04/13/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ZASTAWRNY

DATE/TIME OF TEST: 07/26/2017 02:19:24

---

Diagnostics Check	OK	02:13
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.141	02:14
Breath Volume	1.648	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.139	02:17
Breath Volume	2.167	02:17
Air Blank	0.000	02:18
Control Test	0.083	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

---

Data Download Date/Time: 08/28/2017 07:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARMAN, MARIA, E  
BREATH TEST OPERATOR AGENCY: Palatka PD  
ARRESTING OFFICER: HARMAN, MARIA  
ARRESTING OFFICER AGENCY: Palatka PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: Yes  
DRY GAS STANDARD LOT #: 01316080A2  
DRY GAS EXPIRY DATE: 02/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:41  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PAPOOSHA

DATE/TIME OF TEST: 07/04/2017 02:34:47

---

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.079	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 2	02:33	No Sample Provided
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

---

Data Download Date/Time: 08/01/2017 14:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DASILVA,ANDRE,F  
BREATH TEST OPERATOR AGENCY: FHP Troop G  
ARRESTING OFFICER: DASILVA,ANDRE  
ARRESTING OFFICER AGENCY: FHP Troop G  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01316080A2  
DRY GAS EXPIRY DATE: 02/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:14  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CLAYTON MALLORY

DATE/TIME OF TEST: 07/05/2017 02:28:13

---

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.085	02:23
Breath Volume	1.238	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:26
Subject Sample # 2	0.083	02:26
Breath Volume	1.566	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Diagnostics Check	OK	02:28

---

Data Download Date/Time: 08/01/2017 14:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MELTON,W,  
BREATH TEST OPERATOR AGENCY: Welaka PD  
ARRESTING OFFICER: SNIPES,K  
ARRESTING OFFICER AGENCY: Welaka PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 01316080A2  
DRY GAS EXPIRY DATE: 02/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: CABEZA

DATE/TIME OF TEST: 07/07/2017 04:01:56

---

Diagnostics Check	OK	03:59	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank	0.000	04:01	
Diagnostics Check	OK	04:01	

---

Data Download Date/Time: 08/01/2017 14:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMA, MICHAEL, A  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: 3L08, MICHAEL  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1824329  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:17  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CARVER

DATE/TIME OF TEST: 07/07/2017 23:27:24

---

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:21
Air Blank	0.000	23:22
Subject Sample # 1	0.231	23:22
Breath Volume	1.503	23:22
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.222	23:25
Breath Volume	1.562	23:25
Air Blank	0.000	23:26
Control Test	0.079	23:26
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

---

Data Download Date/Time: 07/28/2017 09:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LIMA,MICHAEL,A  
BREATH TEST OPERATOR AGENCY: Daytona Beach PD  
ARRESTING OFFICER: LIMA,MICHAEL  
ARRESTING OFFICER AGENCY: Daytona Beach PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1824329  
DRY GAS EXPIRY DATE: 03/09/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VEINOT

DATE/TIME OF TEST: 07/10/2017 02:25:14

---

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.079	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.249	02:20
Breath Volume		1.203	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.262	02:23
Breath Volume		1.269	02:23
Air Blank		0.000	02:24
Control Test		0.079	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

---

Data Download Date/Time: 07/28/2017 09:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001247

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,  
BREATH TEST OPERATOR AGENCY: Broward County SO  
ARRESTING OFFICER: SERPHOS,STEVEN  
ARRESTING OFFICER AGENCY: Broward County SO  
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 562838  
DRY GAS EXPIRY DATE: 09/06/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 07/21/2017 03:02:58

---

Diagnostics Check	OK	02:56
Air Blank	0.000	02:56
Control Test	0.081	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.110	02:58
Breath Volume	2.218	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:00
Subject Sample # 2	0.111	03:01
Breath Volume	2.035	03:01
Air Blank	0.000	03:01
Control Test	0.081	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

---

Data Download Date/Time: 07/21/2017 10:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HUFF,JON,B  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HUFF,JON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: QUESADA

DATE/TIME OF TEST: 07/08/2017 01:42:50

---

Diagnostics Check	OK	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Subject Sample # 1	0.059	01:38
Breath Volume	2.718	01:38
Air Blank	0.000	01:39
Air Blank	0.000	01:40
Subject Sample # 2	0.059	01:41
Breath Volume	2.433	01:41
Air Blank	0.000	01:41
Control Test	0.080	01:42
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

---

Data Download Date/Time: 07/19/2017 13:20



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: OKEEFE,PATRICIA  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: RUIZ-GUTIERREZ

DATE/TIME OF TEST: 07/09/2017 02:39:14

---

Diagnostics Check	OK	02:36	
Air Blank	0.000	02:36	
Control Test	0.080	02:37	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:37	Subject Test Refused
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Control Test	0.081	02:38	
Air Blank	0.000	02:39	
Diagnostics Check	OK	02:39	

---

Data Download Date/Time: 07/19/2017 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTIERREZ,KRISTINA,L  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: GUTIERREZ,KRISTINA  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LIBERTY

DATE/TIME OF TEST: 07/11/2017 02:06:54

---

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.080	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.180	02:02
Breath Volume	3.625	02:02
Air Blank	0.000	02:03
Air Blank	0.000	02:04
Subject Sample # 2	0.196	02:05
Breath Volume	2.757	02:05
Air Blank	0.000	02:06
Control Test	0.079	02:06
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

---

Data Download Date/Time: 07/19/2017 13:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: ALVAREZ,ORLANDO  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SHENERY

DATE/TIME OF TEST: 07/13/2017 04:25:29

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.081	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.000	04:20
Breath Volume	1.382	04:20
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.000	04:24
Breath Volume	1.625	04:24
Air Blank	0.000	04:24
Control Test	0.081	04:24
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

---

Data Download Date/Time: 07/19/2017 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 07/14/2017 02:54:51

---

Diagnostics Check	OK	02:48
Air Blank	0.000	02:48
Control Test	0.080	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.173	02:50
Breath Volume	2.949	02:50
Air Blank	0.000	02:50
Air Blank	0.000	02:52
Subject Sample # 2	0.173	02:53
Breath Volume	3.128	02:53
Air Blank	0.000	02:53
Control Test	0.080	02:54
Air Blank	0.000	02:54
Diagnostics Check	OK	02:54

---

Data Download Date/Time: 07/19/2017 13:20

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SOLORZANO,DYLAN,A  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: SOLORZANO,DYLAN  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 15:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 15:35

SUBJECT'S LAST NAME: MUNGUIA

DATE/TIME OF TEST: 07/16/2017 16:00:50

---

Diagnostics Check	OK	15:55	
Air Blank	0.000	15:56	
Control Test	0.081	15:56	
Air Blank	0.000	15:56	
Subject Sample	# 1	15:59	Subject Test Refused
Breath Volume	0.000	15:59	
Air Blank	0.000	15:59	
Control Test	0.080	16:00	
Air Blank	0.000	16:00	
Diagnostics Check	OK	16:00	

---

Data Download Date/Time: 07/19/2017 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KERNS,CODY,J  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: FRAIOLL,JAY  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 06/17/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 279870  
DRY GAS EXPIRY DATE: 02/04/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: SPAULDING

DATE/TIME OF TEST: 07/16/2017 21:06:38

---

Diagnostics Check	OK	21:03	
Air Blank	0.000	21:04	
Control Test	0.080	21:04	
Air Blank	0.000	21:05	
Subject Sample	# 1	21:05	Subject Test Refused
Breath Volume	0.000	21:05	
Air Blank	0.000	21:05	
Control Test	0.079	21:06	
Air Blank	0.000	21:06	
Diagnostics Check	OK	21:06	

---

Data Download Date/Time: 07/19/2017 13:20

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,GIL,A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GONZALEZ,GIL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: CASCO

DATE/TIME OF TEST: 07/22/2017 23:49:07

---

Diagnostics Check	OK	23:42	
Air Blank	0.000	23:42	
Control Test	0.081	23:43	
Air Blank	0.000	23:43	
Subject Sample # 1	0.259	23:44	
Breath Volume	1.332	23:44	
Air Blank	0.000	23:45	
Air Blank	0.000	23:46	
Subject Sample # 2		23:47	RFI Detect
Breath Volume	1.250	23:47	
Air Blank	0.000	23:48	
Control Test	0.082	23:48	
Air Blank	0.000	23:49	
Diagnostics Check	OK	23:49	

---

Data Download Date/Time: 08/02/2017 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GONZALEZ,GIL,A  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: GONZALEZ,GIL  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: CASCO

DATE/TIME OF TEST: 07/22/2017 23:59:59

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:53
Control Test	0.081	23:54
Air Blank	0.000	23:54
Subject Sample # 1	0.257	23:55
Breath Volume	1.523	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 2	0.246	23:58
Breath Volume	2.671	23:58
Air Blank	0.000	23:59
Control Test	0.081	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

---

Data Download Date/Time: 08/02/2017 15:11



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HODGES,TITUS,R  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HODGES,TITUS  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 22:47  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: TISDALE

DATE/TIME OF TEST: 07/23/2017 23:40:03

---

Diagnostics Check	OK	23:36	
Air Blank	0.000	23:37	
Control Test	0.082	23:37	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:38	Subject Test Refused
Breath Volume	0.000	23:38	
Air Blank	0.000	23:39	
Control Test	0.081	23:39	
Air Blank	0.000	23:39	
Diagnostics Check	OK	23:40	

---

Data Download Date/Time: 08/02/2017 15:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:28  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/27/2017 02:10:11

---

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:06	
Control Test	0.082	02:06	
Air Blank	0.000	02:07	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:09	
Control Test	0.081	02:09	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

---

Data Download Date/Time: 08/02/2017 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ORBETA,JORGE,  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: ORBETA,JORGE  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 07/28/2017 02:01:38

---

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:53	
Control Test	0.082	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1	01:57	No Sample Provided
Breath Volume	0.000	01:57	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.081	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

---

Data Download Date/Time: 08/02/2017 15:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAWN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MARTIN,SHAWN  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:50  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 07/29/2017 00:49:26

---

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.083	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

---

Data Download Date/Time: 08/02/2017 15:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAWN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: MARTIN,SHAWN  
ARRESTING OFFICER AGENCY: FHP Troop E  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:27  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 07/29/2017 04:17:34

---

Diagnostics Check	OK	04:14	
Air Blank	0.000	04:15	
Control Test	0.082	04:15	
Air Blank	0.000	04:15	
Subject Sample	# 1	04:16	Subject Test Refused
Breath Volume	0.000	04:16	
Air Blank	0.000	04:16	
Control Test	0.082	04:16	
Air Blank	0.000	04:17	
Diagnostics Check	OK	04:17	

---

Data Download Date/Time: 08/02/2017 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: TORRES,DEANNA,N  
BREATH TEST OPERATOR AGENCY: Monroe County SO  
ARRESTING OFFICER: HANSEN,DYLLON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 11:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 11:21

SUBJECT'S LAST NAME: SALAZAR TORRES

DATE/TIME OF TEST: 07/29/2017 11:47:39

---

Diagnostics Check	OK	11:41
Air Blank	0.000	11:41
Control Test	0.081	11:42
Air Blank	0.000	11:42
Subject Sample # 1	0.183	11:43
Breath Volume	3.687	11:43
Air Blank	0.000	11:43
Air Blank	0.000	11:45
Subject Sample # 2	0.183	11:46
Breath Volume	3.316	11:46
Air Blank	0.000	11:46
Control Test	0.080	11:47
Air Blank	0.000	11:47
Diagnostics Check	OK	11:47

---

Data Download Date/Time: 08/02/2017 15:11

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAWN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: LAPRADD,ADDISON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GAHAGEN

DATE/TIME OF TEST: 07/30/2017 02:47:59

---

Diagnostics Check	OK	02:47	
Air Blank		02:47	RFI Detect
Air Blank	0.000	02:47	

---

Data Download Date/Time: 08/02/2017 15:11

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MARTIN,SHAWN,W  
BREATH TEST OPERATOR AGENCY: FHP Troop E  
ARRESTING OFFICER: LAPRADD,ADDISON  
ARRESTING OFFICER AGENCY: Monroe County SO  
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 686331  
DRY GAS EXPIRY DATE: 01/31/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: GAHAGEN

DATE/TIME OF TEST: 07/30/2017 02:52:27

---

Diagnostics Check	OK	02:49	
Air Blank	0.000	02:50	
Control Test	0.081	02:50	
Air Blank	0.000	02:50	
Subject Sample	# 1	02:51	Subject Test Refused
Breath Volume	0.000	02:51	
Air Blank	0.000	02:51	
Control Test	0.081	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

---

Data Download Date/Time: 08/02/2017 15:11



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: YUKNIEWICZ, B  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:14  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: GALVAN

DATE/TIME OF TEST: 07/01/2017 04:47:01

---

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.079	04:41
Air Blank	0.000	04:42
Subject Sample # 1	0.178	04:42
Breath Volume	3.769	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:45
Subject Sample # 2	0.168	04:45
Breath Volume	1.476	04:45
Air Blank	0.000	04:46
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:47

---

Data Download Date/Time: 07/07/2017 13:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WEAVER,WILLIAM  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 01:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: CANEPA

DATE/TIME OF TEST: 07/02/2017 02:23:47

---

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.174	02:15
Breath Volume	1.441	02:15
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.197	02:18
Breath Volume	3.417	02:18
Air Blank	0.000	02:19
Air Blank	0.000	02:21
Subject Sample # 3	0.195	02:22
Breath Volume	3.878	02:22
Air Blank	0.000	02:22
Control Test	0.080	02:23
Air Blank	0.000	02:23
Diagnostics Check	OK	02:23

---

Data Download Date/Time: 07/07/2017 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS,DEBORAH  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: O BRIEN

DATE/TIME OF TEST: 07/02/2017 04:23:45

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.162	04:19
Breath Volume	2.960	04:19
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.158	04:22
Breath Volume	2.601	04:22
Air Blank	0.000	04:22
Control Test	0.079	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

---

Data Download Date/Time: 07/07/2017 13:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: ANDREWS,GARY  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:12  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: LEARY

DATE/TIME OF TEST: 07/02/2017 23:06:17

---

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.158	23:02
Breath Volume	2.753	23:02
Air Blank	0.000	23:02
Air Blank	0.000	23:04
Subject Sample # 2	0.150	23:04
Breath Volume	2.679	23:04
Air Blank	0.000	23:05
Control Test	0.079	23:05
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

---

Data Download Date/Time: 07/07/2017 13:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELBON,NICHOLAS,C  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: KNIGHT,RYAN  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: PRATS FANTAUZZI

DATE/TIME OF TEST: 07/03/2017 00:55:13

---

Diagnostics Check	OK	00:48
Air Blank	0.000	00:48
Control Test	0.080	00:49
Air Blank	0.000	00:49
Subject Sample # 1	0.187	00:50
Breath Volume	1.992	00:50
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.192	00:53
Breath Volume	2.628	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

---

Data Download Date/Time: 07/07/2017 13:03

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WINBURN,T  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: DELGADOMEDINA

DATE/TIME OF TEST: 07/04/2017 22:05:34

---

Diagnostics Check	OK	21:58
Air Blank	0.000	21:59
Control Test	0.081	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.167	22:01
Breath Volume	2.000	22:01
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.180	22:04
Breath Volume	2.097	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

---

Data Download Date/Time: 07/07/2017 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MATHISEN,L  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:36  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: LOPEZDELAGUARDIA

DATE/TIME OF TEST: 07/05/2017 03:41:00

---

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:35
Subject Sample # 1	0.012	03:36
Breath Volume	1.421	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.011	03:39
Breath Volume	1.175	03:39
Air Blank	0.000	03:40
Control Test	0.080	03:40
Air Blank	0.000	03:40
Diagnostics Check	OK	03:41

---

Data Download Date/Time: 07/07/2017 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HALLOCK,STEPHEN  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: MITIL

DATE/TIME OF TEST: 07/06/2017 03:14:32

---

Diagnostics Check	OK	03:05
Air Blank	0.000	03:06
Control Test	0.081	03:06
Air Blank	0.000	03:07
Subject Sample # 1	0.237	03:09
Breath Volume	3.054	03:09
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.248	03:13
Breath Volume	3.316	03:13
Air Blank	0.000	03:13
Control Test	0.080	03:13
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

---

Data Download Date/Time: 07/07/2017 13:03



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAM,ALEXANDER  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 07/06/2017 04:19:41

---

Diagnostics Check	OK	04:13
Air Blank	0.000	04:13
Control Test	0.079	04:13
Air Blank	0.000	04:14
Subject Sample # 1	0.126	04:15
Breath Volume	2.226	04:15
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.128	04:18
Breath Volume	2.484	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:19
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

---

Data Download Date/Time: 07/07/2017 13:03

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAWKINS, D  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: COMLY

DATE/TIME OF TEST: 07/08/2017 03:47:09

---

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.080	03:41
Air Blank	0.000	03:42
Subject Sample # 1	0.161	03:42
Breath Volume	1.691	03:42
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.162	03:45
Breath Volume	2.128	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

---

Data Download Date/Time: 07/12/2017 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: DRACH, S  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: CHABLE GIL

DATE/TIME OF TEST: 07/09/2017 01:38:17

---

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.080	01:32
Air Blank	0.000	01:33
Subject Sample # 1	0.181	01:33
Breath Volume	1.816	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.189	01:36
Breath Volume	1.808	01:36
Air Blank	0.000	01:37
Control Test	0.080	01:37
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

---

Data Download Date/Time: 07/12/2017 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MERKS, A  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: RAMIREZARDON

DATE/TIME OF TEST: 07/09/2017 04:42:01

---

Diagnostics Check		OK	04:31
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.222	04:34
Breath Volume		2.375	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.197	04:37
Breath Volume		1.492	04:37
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 3	0.222	04:40
Breath Volume		1.226	04:40
Air Blank		0.000	04:41
Control Test		0.079	04:41
Air Blank		0.000	04:41
Diagnostics Check		OK	04:42

---

Data Download Date/Time: 07/12/2017 12:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AUSTIN,A  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 07/09/2017 05:22:55

---

Diagnostics Check	OK	05:15
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:16
Subject Sample # 1	0.144	05:17
Breath Volume	1.273	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:19
Subject Sample # 2	0.140	05:21
Breath Volume	1.343	05:21
Air Blank	0.000	05:22
Control Test	0.080	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

---

Data Download Date/Time: 07/12/2017 12:24

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BLITHE, W  
ARRESTING OFFICER AGENCY: Orange County SO  
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/12/2017 01:14:25

---

Diagnostics Check	OK	01:07
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:08
Subject Sample # 1	0.168	01:09
Breath Volume	3.832	01:09
Air Blank	0.000	01:10
Air Blank	0.000	01:11
Subject Sample # 2	0.162	01:12
Breath Volume	1.625	01:12
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Diagnostics Check	OK	01:14

---

Data Download Date/Time: 07/12/2017 12:24

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: MORRIS,STEVE  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: SCHMELZER

DATE/TIME OF TEST: 07/13/2017 23:20:41

---

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:14
Air Blank	0.000	23:15
Subject Sample # 1	0.000	23:15
Breath Volume	1.429	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.000	23:19
Breath Volume	2.210	23:19
Air Blank	0.000	23:19
Control Test	0.080	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:20

---

Data Download Date/Time: 07/19/2017 15:15

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: AZEVEDO,FABIO  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 04:56  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/15/2017 06:02:28

---

Diagnostics Check	OK	05:56
Air Blank	0.000	05:57
Control Test	0.081	05:57
Air Blank	0.000	05:57
Subject Sample # 1	0.113	05:58
Breath Volume	2.007	05:58
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.113	06:01
Breath Volume	2.242	06:01
Air Blank	0.000	06:01
Control Test	0.081	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

---

Data Download Date/Time: 07/19/2017 15:15



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: BECK,KURTIS  
ARRESTING OFFICER AGENCY: Ocoee PD  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:41  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/22/2017 03:44:28

---

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.090	03:40
Breath Volume	2.121	03:40
Air Blank	0.000	03:40
Air Blank	0.000	03:42
Subject Sample # 2	0.088	03:43
Breath Volume	2.195	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:43
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

---

Data Download Date/Time: 07/27/2017 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HAIGHT,MONTAQUE  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 09:46  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:15

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/22/2017 10:57:45

---

Diagnostics Check	OK	10:51
Air Blank	0.000	10:51
Control Test	0.080	10:52
Air Blank	0.000	10:52
Subject Sample # 1	0.113	10:53
Breath Volume	2.722	10:53
Air Blank	0.000	10:53
Air Blank	0.000	10:55
Subject Sample # 2	0.116	10:56
Breath Volume	2.636	10:56
Air Blank	0.000	10:56
Control Test	0.080	10:57
Air Blank	0.000	10:57
Diagnostics Check	OK	10:57

---

Data Download Date/Time: 07/27/2017 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: WONG SHUE, ANTHONY  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 07:45  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 08:03

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 07/23/2017 08:32:05

---

Diagnostics Check	OK	08:25
Air Blank	0.000	08:26
Control Test	0.080	08:26
Air Blank	0.000	08:27
Subject Sample # 1	0.109	08:27
Breath Volume	1.375	08:27
Air Blank	0.000	08:28
Air Blank	0.000	08:30
Subject Sample # 2	0.113	08:30
Breath Volume	1.195	08:30
Air Blank	0.000	08:31
Control Test	0.081	08:31
Air Blank	0.000	08:31
Diagnostics Check	OK	08:32

---

Data Download Date/Time: 07/27/2017 17:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KOSH,TOBI,J  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: HUDSON,CODY  
ARRESTING OFFICER AGENCY: Winter Garden PD  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:25  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: VIEIRA

DATE/TIME OF TEST: 07/28/2017 03:27:24

---

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.144	03:22
Breath Volume	1.726	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:24
Subject Sample # 2	0.137	03:25
Breath Volume	1.960	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

---

Data Download Date/Time: 08/03/2017 14:54

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001256

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: THOMPSON,CALEB,N  
BREATH TEST OPERATOR AGENCY: Orange County SO  
ARRESTING OFFICER: NUTTER,MICAH  
ARRESTING OFFICER AGENCY: Orlando PD  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A1  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: LLUBERES

DATE/TIME OF TEST: 07/29/2017 04:13:23

---

Diagnostics Check	OK	04:03
Air Blank	0.000	04:04
Control Test	0.081	04:04
Air Blank	0.000	04:04
Subject Sample # 1	0.256	04:06
Breath Volume	1.351	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.263	04:11
Breath Volume	2.058	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

---

Data Download Date/Time: 08/03/2017 14:54

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GOMEZ,IVETTE,  
BREATH TEST OPERATOR AGENCY: Rockledge PD  
ARRESTING OFFICER: RUSH,MARK  
ARRESTING OFFICER AGENCY: Cocoa PD  
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 572920  
DRY GAS EXPIRY DATE: 09/13/2019  
VIOLATION CODE: DUI  
ARREST TIME: 23:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 07/24/2017 00:03:37

---

Diagnostics Check	OK	23:57
Air Blank	0.000	23:57
Control Test	0.079	23:58
Air Blank	0.000	23:58
Subject Sample # 2	0.157	23:59
Breath Volume	3.164	23:59
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.158	00:02
Breath Volume	2.792	00:02
Air Blank	0.000	00:02
Control Test	0.079	00:03
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

---

Data Download Date/Time: 07/28/2017 04:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: WHITE,TAYLOR  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: MEJIA ULLOA

DATE/TIME OF TEST: 07/02/2017 03:35:58

---

Diagnostics Check		OK	03:25	
Air Blank		0.000	03:26	
Control Test		0.080	03:26	
Air Blank		0.000	03:27	
Subject Sample	# 1	0.145	03:30	Volume Not Met
Breath Volume		0.160	03:30	
Air Blank		0.000	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 2		03:34	Subject Test Refused
Breath Volume		0.156	03:34	
Air Blank		0.000	03:35	
Control Test		0.080	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

---

Data Download Date/Time: 07/31/2017 08:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HESSION,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 16:42  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: FERFORT

DATE/TIME OF TEST: 07/04/2017 17:21:32

---

Diagnostics Check	OK	17:14
Air Blank	0.000	17:15
Control Test	0.079	17:15
Air Blank	0.000	17:16
Subject Sample # 1	0.141	17:17
Breath Volume	2.542	17:17
Air Blank	0.000	17:17
Air Blank	0.000	17:19
Subject Sample # 2	0.138	17:20
Breath Volume	1.929	17:20
Air Blank	0.000	17:20
Control Test	0.079	17:20
Air Blank	0.000	17:21
Diagnostics Check	OK	17:21

---

Data Download Date/Time: 07/31/2017 08:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HESSION,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 13:57  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: FERDINAND

DATE/TIME OF TEST: 07/07/2017 14:40:20

---

Diagnostics Check	OK	14:32
Air Blank	0.000	14:33
Control Test	0.079	14:33
Air Blank	0.000	14:34
Subject Sample # 1	0.217	14:34
Breath Volume	1.394	14:34
Air Blank	0.000	14:35
Air Blank	0.000	14:37
Subject Sample # 2	0.215	14:38
Breath Volume	1.484	14:38
Air Blank	0.000	14:39
Control Test	0.081	14:39
Air Blank	0.000	14:40
Diagnostics Check	OK	14:40

---

Data Download Date/Time: 07/31/2017 08:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRACKETT,MARK,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: NOLES,GABE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:26  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PFEIFFER

DATE/TIME OF TEST: 07/08/2017 03:57:46

---

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Subject Sample # 1	0.087	03:52
Breath Volume	1.667	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.078	03:56
Breath Volume	1.488	03:56
Air Blank	0.000	03:56
Control Test	0.081	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:57

---

Data Download Date/Time: 07/31/2017 08:18

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HESSION,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 07/14/2017 23:04:54

---

Diagnostics Check	OK	22:54	
Air Blank	0.000	22:55	
Control Test	0.080	22:55	
Air Blank	0.000	22:56	
Subject Sample	# 1	22:59	No Sample Provided
Breath Volume	0.000	22:59	
Air Blank	0.000	22:59	
Air Blank	0.000	23:01	
Subject Sample	# 2	23:03	Subject Test Refused
Breath Volume	0.000	23:03	
Air Blank	0.000	23:03	
Control Test	0.079	23:04	
Air Blank	0.000	23:04	
Diagnostics Check	OK	23:04	

---

Data Download Date/Time: 07/31/2017 08:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HESSION,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 07/16/2017 02:32:53

---

Diagnostics Check	OK	02:20	
Air Blank	0.000	02:20	
Control Test	0.081	02:21	
Air Blank	0.000	02:21	
Subject Sample	# 1 0.136	02:22	
Breath Volume	1.230	02:22	
Air Blank	0.000	02:23	
Air Blank	0.000	02:24	
Subject Sample	# 2 0.000	02:28	Volume Not Met
Breath Volume	0.253	02:28	
Air Blank	0.000	02:28	
Air Blank	0.000	02:30	
Subject Sample	# 3	02:31	Subject Test Refused
Breath Volume	0.000	02:31	
Air Blank	0.000	02:31	
Control Test	0.081	02:32	
Air Blank	0.000	02:32	
Diagnostics Check	OK	02:32	

---

Data Download Date/Time: 07/31/2017 08:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANQUI,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: FRANQUI,DANIEL  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: DIJAN

DATE/TIME OF TEST: 07/16/2017 23:35:20

---

Diagnostics Check	OK	23:29	
Air Blank	0.000	23:29	
Control Test	0.080	23:30	
Air Blank	0.000	23:30	
Subject Sample	# 1 0.209	23:33	
Breath Volume	1.261	23:33	
Air Blank		23:34	Purge Fail
Air Blank	0.000	23:35	

---

Data Download Date/Time: 07/31/2017 08:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: FRANQUI,DANIEL,C  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: FRANQUI,DANIEL  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 539031  
DRY GAS EXPIRY DATE: 08/25/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:11  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: DIJAN

DATE/TIME OF TEST: 07/16/2017 23:45:11

---

Diagnostics Check	OK	23:37
Air Blank	0.000	23:38
Control Test	0.078	23:38
Air Blank	0.000	23:39
Subject Sample # 1	0.191	23:40
Breath Volume	1.222	23:40
Air Blank	0.000	23:41
Air Blank	0.000	23:42
Subject Sample # 2	0.183	23:43
Breath Volume	1.652	23:43
Air Blank	0.000	23:44
Control Test	0.079	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

---

Data Download Date/Time: 07/31/2017 08:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MILLER, MATTHEW, G  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: NOLES, GABE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 19:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: COLLINS WARE

DATE/TIME OF TEST: 07/21/2017 20:21:50

---

Diagnostics Check	OK	20:15
Air Blank	0.000	20:16
Control Test	0.081	20:16
Air Blank	0.000	20:16
Subject Sample # 1	0.000	20:17
Breath Volume	2.964	20:17
Air Blank	0.000	20:17
Air Blank	0.000	20:19
Subject Sample # 2	0.000	20:20
Breath Volume	2.753	20:20
Air Blank	0.000	20:20
Control Test	0.081	20:21
Air Blank	0.000	20:21
Diagnostics Check	OK	20:21

---

Data Download Date/Time: 07/31/2017 08:18

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRACKETT,MARK,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: FERGUSON,STEVEN  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: FOSBRINK

DATE/TIME OF TEST: 07/23/2017 01:19:37

---

Diagnostics Check		OK	01:13
Air Blank		0.000	01:13
Control Test		0.081	01:14
Air Blank		0.000	01:14
Subject Sample	# 1	0.202	01:15
Breath Volume		2.621	01:15
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.200	01:18
Breath Volume		1.871	01:18
Air Blank		0.000	01:18
Control Test		0.081	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

---

Data Download Date/Time: 07/31/2017 08:18



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRACKETT,MARK,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: GIL,DANIEL  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BROTHERS

DATE/TIME OF TEST: 07/23/2017 01:44:23

---

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1	0.224	01:37	Volume Not Met
Breath Volume		0.343	01:37	
Air Blank		0.000	01:38	
Air Blank		0.000	01:39	
Subject Sample	# 2	0.228	01:42	Volume Not Met
Breath Volume		0.082	01:42	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Diagnostics Check		OK	01:44	

---

Data Download Date/Time: 07/31/2017 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRACKETT,MARK,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: GIL,DANIEL  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 01:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: BROTHERS

DATE/TIME OF TEST: 07/23/2017 02:29:24

---

Diagnostics Check	OK	02:21
Air Blank	0.000	02:22
Control Test	0.080	02:22
Air Blank	0.000	02:22
Subject Sample # 1	0.220	02:23
Breath Volume	1.691	02:23
Air Blank	0.000	02:24
Air Blank	0.000	02:25
Subject Sample # 2	0.210	02:27
Breath Volume	1.667	02:27
Air Blank	0.000	02:28
Control Test	0.080	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 07/31/2017 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRACKETT,MARK,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: MALLON,MARIE  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 07/28/2017 01:00:31

---

Diagnostics Check		OK	00:51	
Air Blank		0.000	00:52	
Control Test		0.080	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1	0.000	00:56	Volume Not Met
Breath Volume		0.039	00:56	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 2		00:59	Subject Test Refused
Breath Volume		0.000	00:59	
Air Blank		0.000	00:59	
Control Test		0.080	00:59	
Air Blank		0.000	01:00	
Diagnostics Check		OK	01:00	

---

Data Download Date/Time: 07/31/2017 08:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRACKETT,MARK,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: CLAYTON,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: SCHARROO

DATE/TIME OF TEST: 07/28/2017 01:19:15

---

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.157	01:14
Breath Volume		2.707	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.149	01:17
Breath Volume		2.472	01:17
Air Blank		0.000	01:18
Control Test		0.080	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

---

Data Download Date/Time: 07/31/2017 08:19

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HESSION,CHRISTOPHER,S  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: HESSION,CHRISTOPHER  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 02:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: MCGOWAN JR

DATE/TIME OF TEST: 07/29/2017 03:45:25

---

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.162	03:40
Breath Volume	1.460	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.160	03:43
Breath Volume	2.039	03:43
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

---

Data Download Date/Time: 07/31/2017 08:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CARLSON,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Melbourne PD  
ARRESTING OFFICER: IERVASI,MARIO  
ARRESTING OFFICER AGENCY: Melbourne PD  
LAST AGENCY INSPECTION: 06/22/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 712031  
DRY GAS EXPIRY DATE: 02/21/2020  
VIOLATION CODE: DUI  
ARREST TIME: 03:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ISBELL

DATE/TIME OF TEST: 07/29/2017 04:09:00

---

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.058	04:04
Breath Volume	2.074	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.054	04:07
Breath Volume	2.132	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:09

---

Data Download Date/Time: 07/31/2017 08:19

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GUTTEBO,KENT,N  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BERNARD,MICHAEL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/15/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 03417080A1  
DRY GAS EXPIRY DATE: 03/05/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:46  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: OBRIEN-ABBONIZIO

DATE/TIME OF TEST: 07/22/2017 02:39:05

---

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.080	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.234	02:33
Breath Volume	1.593	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.236	02:37
Breath Volume	1.707	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

---

Data Download Date/Time: 08/25/2017 23:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: JONES,NICHOLAS,W  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: LEWIS,B  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HARVILLE-SMITH

DATE/TIME OF TEST: 07/02/2017 02:29:15

---

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.080	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.133	02:24
Breath Volume	1.710	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.129	02:27
Breath Volume	2.085	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

---

Data Download Date/Time: 07/26/2017 19:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANTON,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GRAVES,BRANDON  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:58  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: SEVERINO

DATE/TIME OF TEST: 07/02/2017 18:09:38

---

Diagnostics Check	OK	18:03
Air Blank	0.000	18:03
Control Test	0.080	18:04
Air Blank	0.000	18:04
Subject Sample # 1	0.186	18:05
Breath Volume	3.535	18:05
Air Blank	0.000	18:05
Air Blank	0.000	18:07
Subject Sample # 2	0.185	18:08
Breath Volume	2.445	18:08
Air Blank	0.000	18:08
Control Test	0.080	18:09
Air Blank	0.000	18:09
Diagnostics Check	OK	18:09

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANTON,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: STANTON,KEVIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:35  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: VADEN

DATE/TIME OF TEST: 07/03/2017 00:04:35

---

Diagnostics Check	OK	23:58
Air Blank	0.000	23:58
Control Test	0.079	23:59
Air Blank	0.000	23:59
Subject Sample # 1	0.138	00:00
Breath Volume	1.738	00:00
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 2	0.133	00:03
Breath Volume	1.812	00:03
Air Blank	0.000	00:03
Control Test	0.081	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LEWIS,JASON,E  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: ELLER,GRANT  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: BUI  
ARREST TIME: 14:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 14:02

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/04/2017 15:00:07

---

Diagnostics Check	OK	14:53
Air Blank	0.000	14:54
Control Test	0.080	14:54
Air Blank	0.000	14:54
Subject Sample # 1	0.207	14:55
Breath Volume	4.031	14:55
Air Blank	0.000	14:56
Air Blank	0.000	14:58
Subject Sample # 2	0.205	14:58
Breath Volume	3.640	14:58
Air Blank	0.000	14:59
Control Test	0.080	14:59
Air Blank	0.000	15:00
Diagnostics Check	OK	15:00

---

Data Download Date/Time: 07/26/2017 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRATSCH,CASEY,R  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BRATSCH,CASEY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:00  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: REVIS

DATE/TIME OF TEST: 07/08/2017 00:00:22

---

Diagnostics Check	OK	23:53
Air Blank	0.000	23:54
Control Test	0.080	23:54
Air Blank	0.000	23:55
Subject Sample # 1	0.091	23:55
Breath Volume	2.519	23:55
Air Blank	0.000	23:56
Air Blank	0.000	23:58
Subject Sample # 2	0.088	23:58
Breath Volume	1.664	23:58
Air Blank	0.000	23:59
Control Test	0.080	23:59
Air Blank	0.000	00:00
Diagnostics Check	OK	00:00

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOWDING,BETHANY,D  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BAILEY,NALICIA  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MCWILLIAMS

DATE/TIME OF TEST: 07/08/2017 03:04:50

---

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.057	03:00
Breath Volume	1.210	03:00
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.060	03:03
Breath Volume	1.835	03:03
Air Blank	0.000	03:03
Control Test	0.079	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLISON, JOSH, G  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HEBB, PAUL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:39  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 07/11/2017 01:20:30

---

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.081	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:18	RFI Detect
Breath Volume	0.457	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank		01:20	RFI Detect
Air Blank	0.000	01:20	

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ELLISON,JOSH,G  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HEBB,PAUL  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: GUTHRIE

DATE/TIME OF TEST: 07/11/2017 01:39:53

---

Diagnostics Check	OK	01:28
Air Blank	0.000	01:29
Control Test	0.081	01:29
Air Blank	0.000	01:30
Subject Sample # 1	0.215	01:30
Breath Volume	1.156	01:30
Air Blank	0.000	01:31
Air Blank	0.000	01:33
Subject Sample # 2	0.193	01:35
Breath Volume	1.218	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 3	0.202	01:38
Breath Volume	1.496	01:38
Air Blank	0.000	01:38
Control Test	0.080	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLEASON,WILLIAM,L  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GLEASON,WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 07/13/2017 22:05:32

---

Diagnostics Check	OK	21:59
Air Blank	0.000	21:59
Control Test	0.080	21:59
Air Blank	0.000	22:00
Subject Sample # 1	0.098	22:00
Breath Volume	2.296	22:00
Air Blank	0.000	22:01
Air Blank	0.000	22:03
Subject Sample # 2	0.105	22:04
Breath Volume	3.011	22:04
Air Blank	0.000	22:04
Control Test	0.079	22:04
Air Blank	0.000	22:05
Diagnostics Check	OK	22:05

---

Data Download Date/Time: 07/26/2017 19:32



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLEASON, WILLIAM, L  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GLEASON, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 07/14/2017 22:22:28

---

Diagnostics Check	OK	22:15
Air Blank	0.000	22:15
Control Test	0.080	22:16
Air Blank	0.000	22:16
Subject Sample # 1	0.154	22:17
Breath Volume	1.820	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:20
Subject Sample # 2	0.144	22:20
Breath Volume	1.363	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:22
Diagnostics Check	OK	22:22

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BRATSCH,CASEY,R  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: BRATSCH,CASEY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:15  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: SILGUERO

DATE/TIME OF TEST: 07/16/2017 02:52:17

---

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.103	02:47
Breath Volume	2.386	02:47
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.105	02:50
Breath Volume	2.136	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

---

Data Download Date/Time: 07/26/2017 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANTON,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: STANTON,KEVIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 19:54  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: WALL

DATE/TIME OF TEST: 07/16/2017 20:48:19

---

Diagnostics Check	OK	20:41
Air Blank	0.000	20:42
Control Test	0.079	20:42
Air Blank	0.000	20:43
Subject Sample # 1	0.322	20:43
Breath Volume	2.816	20:43
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.323	20:46
Breath Volume	1.984	20:46
Air Blank	0.000	20:47
Control Test	0.079	20:47
Air Blank	0.000	20:48
Diagnostics Check	OK	20:48

---

Data Download Date/Time: 07/26/2017 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOWDING,BETHANY,D  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: FERRIERA,ZACHARY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 02:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 07/22/2017 03:32:16

---

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.088	03:27
Breath Volume	2.289	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.081	03:30
Breath Volume	3.039	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

---

Data Download Date/Time: 07/26/2017 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,KIMBERLY,N  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MOORE,KIMBERLY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 22:31  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: MIRANDA VILLANUEVA

DATE/TIME OF TEST: 07/22/2017 23:20:02

---

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.196	23:15
Breath Volume	3.546	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.194	23:18
Breath Volume	3.199	23:18
Air Blank	0.000	23:19
Control Test	0.078	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:20

---

Data Download Date/Time: 07/26/2017 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: DOWDING,BETHANY,D  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: DAUGIRDAS,CHRISTIAN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:29  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 07/23/2017 04:24:03

---

Diagnostics Check	OK	04:17
Air Blank	0.000	04:17
Control Test	0.080	04:18
Air Blank	0.000	04:18
Subject Sample # 1	0.121	04:19
Breath Volume	1.441	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.117	04:22
Breath Volume	1.390	04:22
Air Blank	0.000	04:23
Control Test	0.079	04:23
Air Blank	0.000	04:23
Diagnostics Check	OK	04:24

---

Data Download Date/Time: 07/26/2017 19:32

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PARSONS,C,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: KING-RILEY,A  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:12  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: TYSON

DATE/TIME OF TEST: 07/23/2017 17:17:24

---

Diagnostics Check	OK	17:02	
Air Blank	0.000	17:03	
Control Test	0.080	17:03	
Air Blank	0.000	17:03	
Subject Sample	# 1 0.237	17:06	
Breath Volume	1.718	17:06	
Air Blank	0.000	17:07	
Air Blank	0.000	17:08	
Subject Sample	# 2 0.217	17:11	Volume Not Met
Breath Volume	0.441	17:11	
Air Blank	0.000	17:12	
Air Blank	0.000	17:14	
Subject Sample	# 3 0.237	17:15	
Breath Volume	1.699	17:15	
Air Blank	0.000	17:16	
Control Test	0.079	17:16	
Air Blank	0.000	17:17	
Diagnostics Check	OK	17:17	

---

Data Download Date/Time: 07/26/2017 19:32

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: STANTON,KEVIN,J  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: STANTON,KEVIN  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: PARKS

DATE/TIME OF TEST: 07/25/2017 22:20:22

---

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.082	22:14
Air Blank	0.000	22:15
Subject Sample # 1	0.000	22:15
Breath Volume	2.722	22:15
Air Blank	0.000	22:16
Air Blank	0.000	22:18
Subject Sample # 2	0.000	22:19
Breath Volume	1.550	22:19
Air Blank	0.000	22:19
Control Test	0.079	22:19
Air Blank	0.000	22:20
Diagnostics Check	OK	22:20

---

Data Download Date/Time: 07/26/2017 19:32



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD, WILLIAM,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HOWARD, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 07/29/2017 01:04:31

---

Diagnostics Check	OK	01:02	
Air Blank	0.000	01:02	
Control Test	0.081	01:02	
Air Blank	0.000	01:03	
Subject Sample	# 1	01:03	RFI Detect
Breath Volume	0.000	01:03	
Air Blank		01:04	RFI Detect
Air Blank	0.000	01:04	

---

Data Download Date/Time: 08/23/2017 19:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: HOWARD, WILLIAM,  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: HOWARD, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 00:17  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HERNDON

DATE/TIME OF TEST: 07/29/2017 01:13:22

---

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.080	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.348	01:08
Breath Volume	2.839	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:11
Subject Sample # 2	0.352	01:11
Breath Volume	3.082	01:11
Air Blank	0.000	01:12
Control Test	0.079	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

---

Data Download Date/Time: 08/23/2017 19:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GLEASON, WILLIAM, L  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: GLEASON, WILLIAM  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 20:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SODA

DATE/TIME OF TEST: 07/29/2017 21:39:30

---

Diagnostics Check	OK	21:33
Air Blank	0.000	21:34
Control Test	0.081	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.000	21:35
Breath Volume	1.964	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.000	21:38
Breath Volume	2.359	21:38
Air Blank	0.000	21:38
Control Test	0.081	21:38
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

---

Data Download Date/Time: 08/23/2017 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: MOORE,KIMBERLY,N  
BREATH TEST OPERATOR AGENCY: Brevard County SO  
ARRESTING OFFICER: MOORE,KIMBERLY  
ARRESTING OFFICER AGENCY: Brevard County SO  
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 12716080A5  
DRY GAS EXPIRY DATE: 06/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:03  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: BICANE

DATE/TIME OF TEST: 07/30/2017 00:14:58

---

Diagnostics Check	OK	00:07
Air Blank	0.000	00:08
Control Test	0.081	00:08
Air Blank	0.000	00:08
Subject Sample # 1	0.166	00:10
Breath Volume	1.648	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.173	00:13
Breath Volume	1.832	00:13
Air Blank	0.000	00:14
Control Test	0.079	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:14

---

Data Download Date/Time: 08/23/2017 19:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 06:16

GENDER: Male

OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/15/2017 06:35:25

---

Diagnostics Check	OK	06:28
Air Blank	0.000	06:28
Control Test	0.080	06:28
Air Blank	0.000	06:29
Subject Sample # 1	0.191	06:29
Breath Volume	1.253	06:29
Air Blank	0.000	06:30
Air Blank	0.000	06:32
Subject Sample # 2	0.184	06:33
Breath Volume	1.308	06:33
Air Blank	0.000	06:34
Control Test	0.079	06:34
Air Blank	0.000	06:35
Diagnostics Check	OK	06:35

---

Data Download Date/Time: 08/17/2017 18:14



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001265

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HAMILTON,J,W

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: HAMILTON,J

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 02:38

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: COON

DATE/TIME OF TEST: 07/22/2017 02:53:34

---

Diagnostics Check	OK	02:47	
Air Blank	0.000	02:47	
Control Test	0.080	02:47	
Air Blank	0.000	02:48	
Subject Sample	# 1 0.278	02:49	
Breath Volume	2.648	02:49	
Air Blank	0.000	02:49	
Air Blank	0.000	02:51	
Subject Sample	# 2 0.266	02:52	
Breath Volume	2.933	02:52	
Air Blank	0.000	02:52	
Control Test	0.070	02:52	Control Outside Tolerance
Air Blank	0.000	02:53	

---

Data Download Date/Time: 08/17/2017 18:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PIDGEON,ANDREW,W  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: SPAULDING,CODY  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 26316080A2  
DRY GAS EXPIRY DATE: 11/05/2018  
VIOLATION CODE: DUI  
ARREST TIME: 16:05  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/01/2017 17:19:01

---

Diagnostics Check	OK	17:13
Air Blank	0.000	17:13
Control Test	0.079	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.114	17:14
Breath Volume	2.015	17:14
Air Blank	0.000	17:15
Air Blank	0.000	17:17
Subject Sample # 2	0.111	17:17
Breath Volume	1.574	17:17
Air Blank	0.000	17:18
Control Test	0.080	17:18
Air Blank	0.000	17:18
Diagnostics Check	OK	17:19

---

Data Download Date/Time: 07/12/2017 14:48



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: NON,ARREST

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 22:41

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MIGHTY

DATE/TIME OF TEST: 07/03/2017 22:52:49

---

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.040	22:48
Breath Volume	1.839	22:48
Air Blank	0.000	22:48
Air Blank	0.000	22:50
Subject Sample # 2	0.042	22:51
Breath Volume	1.617	22:51
Air Blank	0.000	22:51
Control Test	0.080	22:52
Air Blank	0.000	22:52
Diagnostics Check	OK	22:52

---

Data Download Date/Time: 07/12/2017 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 06/23/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 26316080A2

DRY GAS EXPIRY DATE: 11/05/2018

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: GARNER

DATE/TIME OF TEST: 07/06/2017 00:37:00

---

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.081	00:31
Air Blank	0.000	00:32
Subject Sample # 1	0.097	00:32
Breath Volume	1.867	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:35
Subject Sample # 2	0.096	00:35
Breath Volume	2.320	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:37

---

Data Download Date/Time: 07/12/2017 14:48

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24515080A2

DRY GAS EXPIRY DATE: 10/05/2017

VIOLATION CODE: DUI

ARREST TIME: 00:11

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: POOLEY

DATE/TIME OF TEST: 07/22/2017 00:29:34

---

Diagnostics Check	OK	00:23
Air Blank	0.000	00:24
Control Test	0.079	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.167	00:25
Breath Volume	2.484	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.166	00:28
Breath Volume	2.242	00:28
Air Blank	0.000	00:28
Control Test	0.078	00:28
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

---

Data Download Date/Time: 08/17/2017 18:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24515080A2

DRY GAS EXPIRY DATE: 10/05/2017

VIOLATION CODE: DUI

ARREST TIME: 03:33

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 07/22/2017 03:52:01

---

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.078	03:46
Air Blank		0.000	03:47
Subject Sample	# 1	0.209	03:47
Breath Volume		1.621	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:49
Subject Sample	# 2	0.223	03:50
Breath Volume		1.941	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:51
Diagnostics Check		OK	03:52

---

Data Download Date/Time: 08/17/2017 18:13

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24515080A2

DRY GAS EXPIRY DATE: 10/05/2017

VIOLATION CODE: DUI

ARREST TIME: 19:32

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: LEKANDER

DATE/TIME OF TEST: 07/22/2017 19:49:23

---

Diagnostics Check	OK	19:43
Air Blank	0.000	19:43
Control Test	0.078	19:44
Air Blank	0.000	19:44
Subject Sample # 1	0.193	19:45
Breath Volume	1.453	19:45
Air Blank	0.000	19:45
Air Blank	0.000	19:47
Subject Sample # 2	0.190	19:47
Breath Volume	1.355	19:47
Air Blank	0.000	19:48
Control Test	0.078	19:48
Air Blank	0.000	19:49
Diagnostics Check	OK	19:49

---

Data Download Date/Time: 08/17/2017 18:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24515080A2

DRY GAS EXPIRY DATE: 10/05/2017

VIOLATION CODE: DUI

ARREST TIME: 01:59

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/23/2017 02:23:04

---

Diagnostics Check		OK	02:09	
Air Blank		0.000	02:09	
Control Test		0.079	02:09	
Air Blank		0.000	02:10	
Subject Sample	# 1	0.105	02:13	Volume Not Met
Breath Volume		0.039	02:13	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 2	0.127	02:17	
Breath Volume		1.667	02:17	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 3	0.124	02:21	
Breath Volume		1.691	02:21	
Air Blank		0.000	02:22	
Control Test		0.078	02:22	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:23	

---

Data Download Date/Time: 08/17/2017 18:13

**Florida Department of Law Enforcement**

**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: LUTZ,EDWIN,J  
BREATH TEST OPERATOR AGENCY: Palm Bay PD  
ARRESTING OFFICER: LUTZ,EDWIN  
ARRESTING OFFICER AGENCY: Palm Bay PD  
LAST AGENCY INSPECTION: 07/12/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 24515080A2  
DRY GAS EXPIRY DATE: 10/05/2017  
VIOLATION CODE: DUI  
ARREST TIME: 00:18  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BOUCHER

DATE/TIME OF TEST: 07/24/2017 00:35:12

---

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:26	
Control Test	0.079	00:27	
Air Blank	0.000	00:27	
Subject Sample	# 1	00:30	No Sample Provided
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Air Blank	0.000	00:33	
Subject Sample	# 2	00:33	Subject Test Refused
Breath Volume	0.000	00:33	
Air Blank	0.000	00:34	
Control Test	0.079	00:34	
Air Blank	0.000	00:35	
Diagnostics Check	OK	00:35	

---

Data Download Date/Time: 08/17/2017 18:13

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: SHODAN,RAGBIR  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1787697  
DRY GAS EXPIRY DATE: 01/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 03:04  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BAGWELL

DATE/TIME OF TEST: 07/29/2017 04:22:52

---

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.133	04:18
Breath Volume	3.816	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.131	04:21
Breath Volume	2.894	04:21
Air Blank	0.000	04:21
Control Test	0.079	04:22
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

---

Data Download Date/Time: 07/30/2017 22:05



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001271

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: TEEHAN,MATTHEW  
ARRESTING OFFICER AGENCY: Altamonte Springs PD  
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 1787697  
DRY GAS EXPIRY DATE: 01/08/2018  
VIOLATION CODE: DUI  
ARREST TIME: 21:40  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: ARVAL JR

DATE/TIME OF TEST: 07/29/2017 22:58:47

---

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.079	22:52
Air Blank	0.000	22:53
Subject Sample # 1	0.211	22:54
Breath Volume	3.457	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.205	22:57
Breath Volume	2.777	22:57
Air Blank	0.000	22:57
Control Test	0.078	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

---

Data Download Date/Time: 07/30/2017 22:05

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HASTINGS,RAYMOND  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: BUI  
ARREST TIME: 19:30  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: GREGORY

DATE/TIME OF TEST: 07/01/2017 21:23:00

---

Diagnostics Check	OK	21:15
Air Blank	0.000	21:16
Control Test	0.081	21:16
Air Blank	0.000	21:17
Subject Sample # 1	0.141	21:18
Breath Volume	3.300	21:18
Air Blank	0.000	21:18
Air Blank	0.000	21:20
Subject Sample # 2	0.138	21:21
Breath Volume	2.953	21:21
Air Blank	0.000	21:22
Control Test	0.080	21:22
Air Blank	0.000	21:22
Diagnostics Check	OK	21:23

---

Data Download Date/Time: 07/03/2017 22:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: FLANAGAN,BRENT  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: SCHNELLMANN

DATE/TIME OF TEST: 07/02/2017 01:04:34

---

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.225	00:59
Breath Volume		1.968	00:59
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.219	01:03
Breath Volume		1.328	01:03
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

---

Data Download Date/Time: 07/03/2017 22:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: RODRIGUEZ,STEPHANIE  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:11  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: APARICIOLAGRAVE

DATE/TIME OF TEST: 07/02/2017 05:18:52

---

Diagnostics Check	OK	05:12
Air Blank	0.000	05:13
Control Test	0.080	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.212	05:14
Breath Volume	3.242	05:14
Air Blank	0.000	05:15
Air Blank	0.000	05:16
Subject Sample # 2	0.204	05:17
Breath Volume	1.109	05:17
Air Blank	0.000	05:17
Control Test	0.080	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

---

Data Download Date/Time: 07/03/2017 22:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: PONDER,CHRISTOPHER,M  
BREATH TEST OPERATOR AGENCY: FHP Troop D  
ARRESTING OFFICER: FLANDERS,J  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:53  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:49

SUBJECT'S LAST NAME: FLANDERS

DATE/TIME OF TEST: 07/02/2017 19:14:11

---

Diagnostics Check	OK	19:11	
Air Blank	0.000	19:11	
Control Test	0.080	19:12	
Air Blank	0.000	19:12	
Subject Sample	# 1	19:12	Subject Test Refused
Breath Volume	0.000	19:12	
Air Blank	0.000	19:13	
Control Test	0.081	19:13	
Air Blank	0.000	19:14	
Diagnostics Check	OK	19:14	

---

Data Download Date/Time: 07/03/2017 22:06

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HASTINGS,RAYMOND  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 20:53  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 07/02/2017 21:59:21

---

Diagnostics Check	OK	21:53
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.134	21:54
Breath Volume	3.800	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.132	21:57
Breath Volume	3.632	21:57
Air Blank	0.000	21:58
Control Test	0.080	21:58
Air Blank	0.000	21:59
Diagnostics Check	OK	21:59

---

Data Download Date/Time: 07/03/2017 22:06

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HASTINGS, RAYMOND  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: BUI  
ARREST TIME: 21:44  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: MCCLLENATHAN

DATE/TIME OF TEST: 07/03/2017 23:21:57

---

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.081	23:16
Air Blank	0.000	23:17
Subject Sample # 1	0.139	23:17
Breath Volume	3.847	23:17
Air Blank	0.000	23:18
Air Blank	0.000	23:19
Subject Sample # 2	0.130	23:20
Breath Volume	3.289	23:20
Air Blank	0.000	23:21
Control Test	0.080	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

---

Data Download Date/Time: 07/30/2017 22:07

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DOUGLAS, NATHAN  
ARRESTING OFFICER AGENCY: FFWCC (Ocala)  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: BUI  
ARREST TIME: 21:38  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: RAGSDALE

DATE/TIME OF TEST: 07/03/2017 23:45:32

---

Diagnostics Check	OK	23:39
Air Blank	0.000	23:39
Control Test	0.080	23:39
Air Blank	0.000	23:40
Subject Sample # 1	0.132	23:41
Breath Volume	2.812	23:41
Air Blank	0.000	23:41
Air Blank	0.000	23:43
Subject Sample # 2	0.131	23:44
Breath Volume	5.113	23:44
Air Blank	0.000	23:44
Control Test	0.080	23:44
Air Blank	0.000	23:45
Diagnostics Check	OK	23:45

---

Data Download Date/Time: 07/30/2017 22:07



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DAIL, EDD  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: PASCUZZO

DATE/TIME OF TEST: 07/04/2017 03:51:41

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.081	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.058	03:47
Breath Volume	1.722	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.060	03:50
Breath Volume	2.000	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

---

Data Download Date/Time: 07/30/2017 22:07

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANCION,CALIX  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 10:06  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: KNUTSON

DATE/TIME OF TEST: 07/06/2017 11:31:01

---

Diagnostics Check	OK	11:23
Air Blank	0.000	11:24
Control Test	0.081	11:24
Air Blank	0.000	11:25
Subject Sample # 1	0.039	11:25
Breath Volume	1.535	11:25
Air Blank	0.000	11:26
Air Blank	0.000	11:28
Subject Sample # 2	0.041	11:29
Breath Volume	2.496	11:29
Air Blank	0.000	11:30
Control Test	0.081	11:30
Air Blank	0.000	11:30
Diagnostics Check	OK	11:31

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: CRAWFORD,L,P  
BREATH TEST OPERATOR AGENCY: Lake Mary PD  
ARRESTING OFFICER: ROSARIOS,J  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 06:16  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: MCCAUL

DATE/TIME OF TEST: 07/08/2017 07:23:06

---

Diagnostics Check	OK	07:16
Air Blank	0.000	07:17
Control Test	0.081	07:17
Air Blank	0.000	07:17
Subject Sample # 1	0.098	07:18
Breath Volume	3.386	07:18
Air Blank	0.000	07:19
Air Blank	0.000	07:21
Subject Sample # 2	0.097	07:21
Breath Volume	3.312	07:21
Air Blank	0.000	07:22
Control Test	0.081	07:22
Air Blank	0.000	07:23
Diagnostics Check	OK	07:23

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: VANHEUSEN,ROBERT  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: REIFE

DATE/TIME OF TEST: 07/08/2017 22:11:27

---

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:05
Air Blank	0.000	22:05
Subject Sample # 1	0.136	22:06
Breath Volume	1.750	22:06
Air Blank	0.000	22:07
Air Blank	0.000	22:09
Subject Sample # 2	0.139	22:09
Breath Volume	1.660	22:09
Air Blank	0.000	22:10
Control Test	0.080	22:10
Air Blank	0.000	22:11
Diagnostics Check	OK	22:11

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DRAUGHON,BOBBY  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:22  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 07/08/2017 23:37:05

---

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.081	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.000	23:32
Breath Volume	1.324	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.000	23:35
Breath Volume	1.484	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:37

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DRAUGHON,BOBBY  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 07/09/2017 03:31:04

---

Diagnostics Check		OK	03:19	
Air Blank		0.000	03:19	
Control Test		0.081	03:20	
Air Blank		0.000	03:20	
Subject Sample	# 1	0.124	03:23	Volume Not Met
Breath Volume		0.742	03:23	
Air Blank		0.000	03:24	
Air Blank		0.000	03:26	
Subject Sample	# 2	0.118	03:29	Volume Not Met
Breath Volume		0.402	03:29	
Air Blank		0.000	03:30	
Control Test		0.080	03:30	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:31	

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DRAUGHON,BOBBY  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:30  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: SINGER

DATE/TIME OF TEST: 07/09/2017 03:39:46

---

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.158	03:35
Breath Volume	1.957	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.148	03:38
Breath Volume	1.503	03:38
Air Blank	0.000	03:38
Control Test	0.081	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HEIM,EDWARD  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: DEONARINE

DATE/TIME OF TEST: 07/12/2017 22:56:47

---

Diagnostics Check	OK	22:45	
Air Blank	0.000	22:45	
Control Test	0.081	22:46	
Air Blank	0.000	22:46	
Subject Sample	# 1 0.033	22:49	Volume Not Met
Breath Volume	0.164	22:49	
Air Blank	0.000	22:50	
Air Blank	0.000	22:52	
Subject Sample	# 2 0.036	22:55	Volume Not Met
Breath Volume	0.847	22:55	
Air Blank	0.000	22:55	
Control Test	0.081	22:56	
Air Blank	0.000	22:56	
Diagnostics Check	OK	22:56	

---

Data Download Date/Time: 07/30/2017 22:08



# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HEIM,EDWARD  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 21:44  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: DEONARINE

DATE/TIME OF TEST: 07/12/2017 23:14:23

---

Diagnostics Check	OK	22:58	
Air Blank	0.000	22:58	
Control Test	0.080	22:59	
Air Blank	0.000	22:59	
Subject Sample	# 1 0.037	23:01	
Breath Volume	1.375	23:01	
Air Blank	0.000	23:02	
Air Blank	0.000	23:04	
Subject Sample	# 2 0.035	23:07	Volume Not Met
Breath Volume	0.753	23:07	
Air Blank	0.000	23:07	
Air Blank	0.000	23:09	
Subject Sample	# 3 0.032	23:12	Volume Not Met
Breath Volume	0.675	23:12	
Air Blank	0.000	23:13	
Control Test	0.081	23:13	
Air Blank	0.000	23:14	
Diagnostics Check	OK	23:14	

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLBURN,MEGAN,E  
BREATH TEST OPERATOR AGENCY: Casselberry PD  
ARRESTING OFFICER: COLBURN,MEGAN  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NEGRON PEREZ

DATE/TIME OF TEST: 07/14/2017 04:27:21

---

Diagnostics Check	OK	04:27	
Air Blank		04:27	RFI Detect
Air Blank		04:27	RFI Detect

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: COLBURN,MEGAN,E  
BREATH TEST OPERATOR AGENCY: Casselberry PD  
ARRESTING OFFICER: COLBURN,MEGAN  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NEGRON PEREZ

DATE/TIME OF TEST: 07/14/2017 04:30:04

---

Diagnostics Check	OK	04:29	
Air Blank		04:29	RFI Detect
Air Blank		04:30	RFI Detect

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMENS,ADAM,J  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: PONDER,CHRISTOPHER  
ARRESTING OFFICER AGENCY: FHP Troop D  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 17:32  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 07/15/2017 19:06:45

---

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:04	
Control Test	0.081	19:04	
Air Blank	0.000	19:05	
Subject Sample	# 1	19:05	Subject Test Refused
Breath Volume	0.000	19:05	
Air Blank	0.000	19:05	
Control Test	0.081	19:06	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:06	

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: SIMMENS,ADAM,J  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ESCUDERO,MADELYN  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:21  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: OCAMPO

DATE/TIME OF TEST: 07/15/2017 23:45:07

---

Diagnostics Check		OK	23:33
Air Blank		0.000	23:34
Control Test		0.081	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.251	23:35
Breath Volume		2.437	23:35
Air Blank		0.000	23:36
Air Blank		0.000	23:38
Subject Sample	# 2	0.225	23:40
Breath Volume		2.312	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:42
Subject Sample	# 3	0.228	23:43
Breath Volume		1.398	23:43
Air Blank		0.000	23:44
Control Test		0.077	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: NOFFKE,SEAN  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 22:02  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 22:02

SUBJECT'S LAST NAME: MCCRAY JR

DATE/TIME OF TEST: 07/17/2017 03:35:40

---

Diagnostics Check	OK	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Subject Sample # 1	0.103	03:31
Breath Volume	1.597	03:31
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.100	03:34
Breath Volume	2.304	03:34
Air Blank	0.000	03:34
Control Test	0.082	03:35
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: DAIL,EDD  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 01:48  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 07/18/2017 02:46:59

---

Diagnostics Check	OK	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Subject Sample # 1	0.270	02:42
Breath Volume	1.351	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.270	02:45
Breath Volume	1.406	02:45
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: BLACKWELL, ROBERT  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:22  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: LUCAS ESCOBAR

DATE/TIME OF TEST: 07/20/2017 01:26:49

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.080	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.249	01:22
Breath Volume	2.851	01:22
Air Blank	0.000	01:22
Air Blank	0.000	01:24
Subject Sample # 2	0.248	01:25
Breath Volume	3.343	01:25
Air Blank	0.000	01:25
Control Test	0.080	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

---

Data Download Date/Time: 07/30/2017 22:08



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ESCUDERO, MADELYN  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:43  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: ZURITA

DATE/TIME OF TEST: 07/20/2017 03:53:07

---

Diagnostics Check	OK	03:44
Air Blank	0.000	03:44
Control Test	0.080	03:45
Air Blank	0.000	03:45
Subject Sample # 1	0.151	03:46
Breath Volume	1.433	03:46
Air Blank	0.000	03:46
Air Blank	0.000	03:48
Subject Sample # 2	0.142	03:51
Breath Volume	3.484	03:51
Air Blank	0.000	03:52
Control Test	0.080	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: VANHEUSEN,ROBERT  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MAYSONET

DATE/TIME OF TEST: 07/22/2017 01:13:29

---

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.138	01:06	Volume Not Met
Breath Volume		0.035	01:06	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.120	01:11	Slope Not Level
Breath Volume		1.160	01:11	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: VANHEUSEN,ROBERT  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:07  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MAYSONET

DATE/TIME OF TEST: 07/22/2017 01:29:12

---

Diagnostics Check	OK	01:20
Air Blank	0.000	01:21
Control Test	0.079	01:21
Air Blank	0.000	01:21
Subject Sample # 1	0.133	01:23
Breath Volume	1.496	01:23
Air Blank	0.000	01:24
Air Blank	0.000	01:26
Subject Sample # 2	0.129	01:27
Breath Volume	1.296	01:27
Air Blank	0.000	01:28
Control Test	0.079	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANDREWS,GARY  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ZACHER

DATE/TIME OF TEST: 07/22/2017 02:06:12

---

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.080	01:55	
Air Blank		0.000	01:55	
Subject Sample	# 1	0.145	01:58	Volume Not Met
Breath Volume		0.375	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2	0.151	02:04	Volume Not Met
Breath Volume		0.484	02:04	
Air Blank		0.000	02:05	
Control Test		0.080	02:05	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: ANDREWS,GARY  
ARRESTING OFFICER AGENCY: Maitland PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 00:10  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: ZACHER

DATE/TIME OF TEST: 07/22/2017 02:22:42

---

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.079	02:08	
Air Blank		0.000	02:08	
Subject Sample	# 1	0.149	02:12	Volume Not Met
Breath Volume		0.863	02:12	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 2	0.150	02:15	
Breath Volume		1.558	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 3	0.144	02:21	Volume Not Met
Breath Volume		0.312	02:21	
Air Blank		0.000	02:21	
Control Test		0.081	02:22	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:22	

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: COLBURN,MEGAN  
ARRESTING OFFICER AGENCY: Casselberry PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 02:35  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: CADENA MESA

DATE/TIME OF TEST: 07/22/2017 03:52:15

---

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.080	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.157	03:47
Breath Volume		2.277	03:47
Air Blank		0.000	03:48
Air Blank		0.000	03:50
Subject Sample	# 2	0.146	03:50
Breath Volume		1.707	03:50
Air Blank		0.000	03:51
Control Test		0.080	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: HUTCHINSON,CHRISTOPER  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:37  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: QUINONES

DATE/TIME OF TEST: 07/22/2017 04:34:06

---

Diagnostics Check	OK	04:27
Air Blank	0.000	04:27
Control Test	0.082	04:27
Air Blank	0.000	04:28
Subject Sample # 1	0.100	04:29
Breath Volume	1.550	04:29
Air Blank	0.000	04:30
Air Blank	0.000	04:31
Subject Sample # 2	0.099	04:32
Breath Volume	1.691	04:32
Air Blank	0.000	04:33
Control Test	0.081	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: BETHAM,KEITH,D  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: VANHUSEN,ROBERT  
ARRESTING OFFICER AGENCY: Oviedo PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 05:20  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: L340558960910

DATE/TIME OF TEST: 07/22/2017 06:00:44

---

Diagnostics Check	OK	05:53
Air Blank	0.000	05:53
Control Test	0.082	05:54
Air Blank	0.000	05:54
Subject Sample # 1	0.175	05:55
Breath Volume	2.070	05:55
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.167	05:59
Breath Volume	2.035	05:59
Air Blank	0.000	05:59
Control Test	0.080	06:00
Air Blank	0.000	06:00
Diagnostics Check	OK	06:00

---

Data Download Date/Time: 07/30/2017 22:08



**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANCION,CALIX,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: RIGGINS,JERRY  
ARRESTING OFFICER AGENCY: Seminole County SO  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 09:45  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: GERLACH

DATE/TIME OF TEST: 07/22/2017 10:29:36

---

Diagnostics Check	OK	10:22
Air Blank	0.000	10:23
Control Test	0.080	10:23
Air Blank	0.000	10:23
Subject Sample # 1	0.149	10:24
Breath Volume	2.425	10:24
Air Blank	0.000	10:25
Air Blank	0.000	10:26
Subject Sample # 2	0.156	10:28
Breath Volume	1.691	10:28
Air Blank	0.000	10:28
Control Test	0.083	10:29
Air Blank	0.000	10:29
Diagnostics Check	OK	10:29

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: FLANAGAN, BRENT  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 04:08  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MAISENHOLDER

DATE/TIME OF TEST: 07/23/2017 04:51:36

---

Diagnostics Check	OK	04:45
Air Blank	0.000	04:45
Control Test	0.082	04:46
Air Blank	0.000	04:46
Subject Sample # 1	0.154	04:47
Breath Volume	1.488	04:47
Air Blank	0.000	04:47
Air Blank	0.000	04:49
Subject Sample # 2	0.167	04:50
Breath Volume	2.492	04:50
Air Blank	0.000	04:50
Control Test	0.081	04:51
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

---

Data Download Date/Time: 07/30/2017 22:08

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: GARCIA, RAY,  
BREATH TEST OPERATOR AGENCY: Seminole County SO  
ARRESTING OFFICER: IRIZARRY, CHRISTOPHER  
ARRESTING OFFICER AGENCY: Sanford PD  
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: 282638  
DRY GAS EXPIRY DATE: 02/15/2019  
VIOLATION CODE: DUI  
ARREST TIME: 03:13  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: BRAVO BERNARDO

DATE/TIME OF TEST: 07/23/2017 05:52:15

---

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.081	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.186	05:47
Breath Volume		2.605	05:47
Air Blank		0.000	05:48
Air Blank		0.000	05:50
Subject Sample	# 2	0.187	05:50
Breath Volume		2.906	05:50
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:52
Diagnostics Check		OK	05:52

---

Data Download Date/Time: 07/30/2017 22:08

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAIR,R  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: BUI  
ARREST TIME: 17:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 07/02/2017 17:35:54

---

Diagnostics Check	OK	17:35	
Air Blank		17:35	RFI Detect
Air Blank		17:35	RFI Detect

---

Data Download Date/Time: 07/28/2017 22:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: ANTHONY,M,  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: BLAIR,R  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: BUI  
ARREST TIME: 17:19  
GENDER: Male  
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 07/02/2017 17:47:57

---

Diagnostics Check	OK	17:39
Air Blank	0.000	17:39
Control Test	0.080	17:40
Air Blank	0.000	17:40
Subject Sample # 1	0.150	17:42
Breath Volume	1.757	17:42
Air Blank	0.000	17:43
Air Blank	0.000	17:45
Subject Sample # 2	0.147	17:46
Breath Volume	1.507	17:46
Air Blank	0.000	17:47
Control Test	0.079	17:47
Air Blank	0.000	17:47
Diagnostics Check	OK	17:47

---

Data Download Date/Time: 07/28/2017 22:14

**Florida Department of Law Enforcement**  
**Alcohol Testing Program**

**Subject Test Electronic Data**

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001273

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: KASELAK,STEPHEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 23:42  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 23:44

SUBJECT'S LAST NAME: SHAPIRO

DATE/TIME OF TEST: 07/16/2017 00:11:20

---

Diagnostics Check	OK	00:05
Air Blank	0.000	00:05
Control Test	0.079	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.193	00:06
Breath Volume	1.535	00:06
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.190	00:09
Breath Volume	1.613	00:09
Air Blank	0.000	00:10
Control Test	0.079	00:10
Air Blank	0.000	00:11
Diagnostics Check	OK	00:11

---

Data Download Date/Time: 07/28/2017 22:14

# Florida Department of Law Enforcement

## Alcohol Testing Program

### Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001274

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO  
SOFTWARE VERSION: 8100.27  
BREATH TEST OPERATOR: KASELAK,STEPHEN,G  
BREATH TEST OPERATOR AGENCY: Pinellas County SO  
ARRESTING OFFICER: KASELAK,STEPHEN  
ARRESTING OFFICER AGENCY: Pinellas County SO  
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No  
DRY GAS STANDARD LOT #: AG629802  
DRY GAS EXPIRY DATE: 10/24/2018  
VIOLATION CODE: DUI  
ARREST TIME: 18:25  
GENDER: Female  
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/02/2017 19:09:07

---

Diagnostics Check	OK	19:05	
Air Blank	0.000	19:05	
Control Test	0.079	19:05	
Air Blank	0.000	19:06	
Subject Sample	# 1	19:07	Subject Test Refused
Breath Volume	0.000	19:07	
Air Blank	0.000	19:08	
Control Test	0.078	19:08	
Air Blank	0.000	19:09	
Diagnostics Check	OK	19:09	

---

Data Download Date/Time: 07/28/2017 23:20