Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: CARRILLO LOPEZ

DATE/TIME OF TEST: 07/01/2017 05:25:59

Diagnostics Check		OK	05:17	
Air Blank		0.000	05:17	
Control Test		0.081	05:18	
Air Blank		0.000	05:18	
Subject Sample	# 1		05:21	No Sample Provided
Breath Volume		0.000	05:21	
Air Blank		0.000	05:22	
Air Blank		0.000	05:24	
Subject Sample	# 2		05:24	Subject Test Refused
Breath Volume		0.000	05:24	
Air Blank		0.000	05:25	
Control Test		0.079	05:25	
Air Blank		0.000	05:25	
Diagnostics Check		OK	05:25	

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: INNOCENT,LESLY
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: PEREZ-DOMINGO

DATE/TIME OF TEST: 07/01/2017 21:39:45

Diagnostics Check		OK	21:32
Air Blank		0.000	21:33
Control Test		0.080	21:33
Air Blank		0.000	21:33
Subject Sample	# 1	0.190	21:34
Breath Volume		2.738	21:34
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.178	21:38
Breath Volume		2.238	21:38
Air Blank		0.000	21:38
Control Test		0.079	21:39
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: LEEDEN,B
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/10/2017 19:49:29

Diagnostics Check		OK	19:37	
Air Blank		0.000	19:37	
Control Test		0.079	19:38	
Air Blank		0.000	19:38	
Subject Sample	# 1	0.192	19:41	Volume Not Met
Breath Volume		0.000	19:41	
Air Blank		0.000	19:42	
Air Blank		0.000	19:44	
Subject Sample	# 2	0.274	19:47	Volume Not Met
Breath Volume		0.265	19:47	
Air Blank		0.000	19:48	
Control Test		0.078	19:48	
Air Blank		0.000	19:49	
Diagnostics Check		OK	19:49	

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: LEEDEN,B
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/10/2017 19:56:01

Diagnostics Check		OK	19:52	
Air Blank		0.000	19:52	
Control Test		0.078	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 1		19:54	Subject Test Refused
Breath Volume		0.015	19:54	
Air Blank		0.000	19:55	
Control Test		0.078	19:55	
Air Blank		0.000	19:55	
Diagnostics Check		OK	19:56	

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,JOSE,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CHEVALIER,JEAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: VERDE-HECHAVARRIA

DATE/TIME OF TEST: 07/12/2017 01:54:46

Diagnostics Check		OK	01:46
Air Blank		0.000	01:46
Control Test		0.081	01:47
Air Blank		0.000	01:47
Subject Sample	# 1	0.321	01:49
Breath Volume		2.855	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.323	01:53
Breath Volume		2.878	01:53
Air Blank		0.000	01:53
Control Test		0.080	01:54
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,GEORGE,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: RODRIGUEZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: CETO-GARCIA

DATE/TIME OF TEST: 07/14/2017 22:46:01

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:40
Subject Sample	# 1	0.199	22:41
Breath Volume		3.714	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:43
Subject Sample	# 2	0.194	22:44
Breath Volume		3.648	22:44
Air Blank		0.000	22:45
Control Test		0.080	22:45
Air Blank		0.000	22:45
Diagnostics Check		OK	22:46

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELO,FRANCISCO,J
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: DEJOHN,CHRISTIAN
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: VALADEZ ARIZAGA

DATE/TIME OF TEST: 07/15/2017 20:13:37

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.080	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.211	20:09
Breath Volume		1.816	20:09
Air Blank		0.000	20:10
Air Blank		0.000	20:11
Subject Sample	# 2	0.213	20:11
Breath Volume		1.734	20:11
Air Blank		0.000	20:12
Control Test		0.079	20:13
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/16/2017 04:21:18

Diagnostics Check		OK	04:12
Air Blank		0.000	04:12
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.208	04:15
Breath Volume		1.195	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.201	04:19
Breath Volume		1.175	04:19
Air Blank		0.000	04:20
Control Test		0.078	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: TANNER,A
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: CARRILLO CORIO

DATE/TIME OF TEST: 07/17/2017 23:46:57

Diagnostics Check		OK	23:44	
Air Blank		0.000	23:44	
Control Test		0.079	23:44	
Air Blank		0.000	23:45	
Subject Sample	# 1		23:45	Subject Test Refused
Breath Volume		0.000	23:45	
Air Blank		0.000	23:46	
Control Test		0.080	23:46	
Air Blank		0.000	23:46	
Diagnostics Check		OK	23:46	

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,JOSE,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: LOPEZ HERNANDEZ

DATE/TIME OF TEST: 07/18/2017 00:33:47

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.189	00:28
Breath Volume		2.386	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:30
Subject Sample	# 2	0.189	00:32
Breath Volume		2.968	00:32
Air Blank		0.000	00:32
Control Test		0.079	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,JOSE,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 07/18/2017 04:21:08

Diagnostics Check		OK	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:15
Subject Sample	# 1	0.125	04:16
Breath Volume		2.203	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:18
Subject Sample	# 2	0.135	04:19
Breath Volume		2.757	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GARCIA,NIVIA
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 18:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: NAVARRO NUNEZ

DATE/TIME OF TEST: 07/22/2017 18:53:54

Diagnostics Check		OK	18:46
Air Blank		0.000	18:47
Control Test		0.079	18:47
Air Blank		0.000	18:48
Subject Sample	# 1	0.182	18:48
Breath Volume		1.613	18:48
Air Blank		0.000	18:49
Air Blank		0.000	18:51
Subject Sample	# 2	0.180	18:52
Breath Volume		1.386	18:52
Air Blank		0.000	18:52
Control Test		0.078	18:53
Air Blank		0.000	18:53
Diagnostics Check		OK	18:53

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: LEAL,GABRIELLE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: RUIZ LOPEZ

DATE/TIME OF TEST: 07/23/2017 00:24:46

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:16	
Control Test		0.080	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:20	No Sample Provided
Breath Volume		0.000	00:20	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2		00:23	Subject Test Refused
Breath Volume		0.000	00:23	
Air Blank		0.000	00:23	
Control Test		0.080	00:24	
Air Blank		0.000	00:24	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLARREAL,JOSE,L
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 07/23/2017 02:57:53

Diagnostics Check		OK	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.154	02:53
Breath Volume		3.148	02:53
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.151	02:56
Breath Volume		2.937	02:56
Air Blank		0.000	02:56
Control Test		0.079	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AGOSTO,VICTOR,M
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGOSTO,VICTOR
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 06/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 10:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:41

SUBJECT'S LAST NAME: RITTER

DATE/TIME OF TEST: 07/27/2017 13:49:30

Diagnostics Check		OK	13:41	
Air Blank		0.000	13:41	
Control Test		0.081	13:41	
Air Blank		0.000	13:42	
Subject Sample	# 1		13:45	No Sample Provided
Breath Volume		0.000	13:45	
Air Blank		0.000	13:45	
Air Blank		0.000	13:47	
Subject Sample	# 2		13:48	Subject Test Refused
Breath Volume		0.000	13:48	
Air Blank		0.000	13:48	
Control Test		0.080	13:48	
Air Blank		0.000	13:49	
Diagnostics Check		OK	13:49	

Data Download Date/Time: 07/28/2017 09:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CHEVALIER

DATE/TIME OF TEST: 07/01/2017 23:20:04

Diagnostics Check		OK	23:13
Air Blank		0.000	23:14
Control Test		0.078	23:14
Air Blank		0.000	23:15
Subject Sample	# 1	0.091	23:15
Breath Volume		3.417	23:15
Air Blank		0.000	23:16
Air Blank		0.000	23:18
Subject Sample	# 2	0.089	23:18
Breath Volume		3.753	23:18
Air Blank		0.000	23:19
Control Test		0.077	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:20

Data Download Date/Time: 07/31/2017 22:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: KOROTAYEV

DATE/TIME OF TEST: 07/06/2017 00:40:37

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.076	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.093	00:36
Breath Volume		1.820	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.086	00:39
Breath Volume		1.609	00:39
Air Blank		0.000	00:39
Control Test		0.075	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 07/31/2017 22:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: BHATT

DATE/TIME OF TEST: 07/08/2017 02:48:39

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.224	02:44
Breath Volume		1.578	02:44
Air Blank		0.000	02:44
Air Blank		0.000	02:46
Subject Sample	# 2	0.229	02:47
Breath Volume		2.425	02:47
Air Blank		0.000	02:47
Control Test		0.077	02:48
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 07/31/2017 22:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ROBEY

DATE/TIME OF TEST: 07/14/2017 22:38:26

Diagnostics Check		OK	22:32
Air Blank		0.000	22:32
Control Test		0.077	22:32
Air Blank		0.000	22:33
Subject Sample	# 1	0.203	22:33
Breath Volume		1.953	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.192	22:36
Breath Volume		1.304	22:36
Air Blank		0.000	22:37
Control Test		0.077	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 07/31/2017 22:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/27/2017 02:25:22

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.078	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.041	02:20
Breath Volume		3.996	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.043	02:23
Breath Volume		3.523	02:23
Air Blank		0.000	02:24
Control Test		0.077	02:24
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 07/31/2017 22:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: STARLING

DATE/TIME OF TEST: 07/28/2017 03:14:11

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.078	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.143	03:07
Breath Volume		1.363	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.129	03:12
Breath Volume		1.246	03:12
Air Blank		0.000	03:13
Control Test		0.077	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 07/31/2017 22:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: FRANK

DATE/TIME OF TEST: 07/02/2017 02:39:08

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.079	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.166	02:34
Breath Volume		2.757	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.162	02:37
Breath Volume		2.015	02:37
Air Blank		0.000	02:38
Control Test		0.078	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BURGESS,TYLER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:32

SUBJECT'S LAST NAME: VINSKUS

DATE/TIME OF TEST: 07/02/2017 19:00:10

Diagnostics Check		OK	18:52
Air Blank		0.000	18:53
Control Test		0.078	18:53
Air Blank		0.000	18:54
Subject Sample	# 1	0.212	18:55
Breath Volume		1.656	18:55
Air Blank		0.000	18:56
Air Blank		0.000	18:57
Subject Sample	# 2	0.206	18:58
Breath Volume		1.281	18:58
Air Blank		0.000	18:59
Control Test		0.077	18:59
Air Blank		0.000	19:00
Diagnostics Check		OK	19:00

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 07/08/2017 19:06:06

Diagnostics Check		OK	18:58
Air Blank		0.000	18:59
Control Test		0.078	18:59
Air Blank		0.000	19:00
Subject Sample	# 1	0.176	19:01
Breath Volume		1.824	19:01
Air Blank		0.000	19:02
Air Blank		0.000	19:03
Subject Sample	# 2	0.176	19:04
Breath Volume		1.781	19:04
Air Blank		0.000	19:05
Control Test		0.077	19:05
Air Blank		0.000	19:06
Diagnostics Check		OK	19:06

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: DEVANE

DATE/TIME OF TEST: 07/09/2017 19:24:39

Diagnostics Check		OK	19:18
Air Blank		0.000	19:18
Control Test		0.079	19:19
Air Blank		0.000	19:19
Subject Sample	# 1	0.269	19:19
Breath Volume		2.167	19:19
Air Blank		0.000	19:20
Air Blank		0.000	19:22
Subject Sample	# 2	0.270	19:23
Breath Volume		2.710	19:23
Air Blank		0.000	19:23
Control Test		0.077	19:24
Air Blank		0.000	19:24
Diagnostics Check		OK	19:24

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: RAVN,SUSANNE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BUSKEY

DATE/TIME OF TEST: 07/12/2017 03:13:13

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.189	03:08
Breath Volume		2.253	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:10
Subject Sample	# 2	0.196	03:11
Breath Volume		1.281	03:11
Air Blank		0.000	03:12
Control Test		0.077	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: KISER

DATE/TIME OF TEST: 07/14/2017 00:38:04

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.079	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.173	00:33
Breath Volume		1.816	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.166	00:36
Breath Volume		1.644	00:36
Air Blank		0.000	00:37
Control Test		0.076	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:38

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: EVELYN

DATE/TIME OF TEST: 07/15/2017 02:40:23

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:35
Subject Sample	# 1	0.044	02:35
Breath Volume		1.734	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.041	02:38
Breath Volume		1.746	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROZYCKI,RICHARD,J
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BONEY,BRANDON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: FALSEY

DATE/TIME OF TEST: 07/18/2017 03:28:21

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.079	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.191	03:23
Breath Volume		2.675	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.198	03:26
Breath Volume		2.816	03:26
Air Blank		0.000	03:27
Control Test		0.078	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: DUMEER

DATE/TIME OF TEST: 07/21/2017 00:23:29

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.056	00:18
Breath Volume		1.425	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 2	0.051	00:21
Breath Volume		1.496	00:21
Air Blank		0.000	00:22
Control Test		0.076	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILKIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/21/2017 02:10:04

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.155	02:05
Breath Volume		2.539	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.151	02:08
Breath Volume		1.738	02:08
Air Blank		0.000	02:09
Control Test		0.078	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:10

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILKIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 07/21/2017 02:42:37

Diagnostics Check		OK	02:30	
Air Blank		0.000	02:31	
Control Test		0.078	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.092	02:35	Volume Not Met
Breath Volume		0.261	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 2	0.087	02:41	Volume Not Met
Breath Volume		0.281	02:41	
Air Blank		0.000	02:41	
Control Test		0.077	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILKIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 07/21/2017 02:57:57

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.077	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1	0.085	02:50	Volume Not Met
Breath Volume		0.472	02:50	
Air Blank		0.000	02:51	
Air Blank		0.000	02:53	
Subject Sample	# 2	0.085	02:56	Volume Not Met
Breath Volume		0.503	02:56	
Air Blank		0.000	02:57	
Control Test		0.077	02:57	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WILKIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DOTY

DATE/TIME OF TEST: 07/21/2017 03:10:51

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.078	03:01
Air Blank		0.000	03:01
Subject Sample	# 1	0.096	03:04
Breath Volume		1.519	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:06
Subject Sample	# 2	0.092	03:09
Breath Volume		1.156	03:09
Air Blank		0.000	03:09
Control Test		0.077	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,JOSH,G
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,LAUREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/25/2017 03:10:06

Diagnostics Check		OK	02:58	
Air Blank		0.000	02:58	
Control Test		0.078	02:59	
Air Blank		0.000	02:59	
Subject Sample	# 1	0.100	03:02	Volume Not Met
Breath Volume		0.089	03:02	
Air Blank		0.000	03:03	
Air Blank		0.000	03:05	
Subject Sample	# 2	0.111	03:08	Volume Not Met
Breath Volume		0.308	03:08	
Air Blank		0.000	03:09	
Control Test		0.077	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,JOSH,G
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,LAUREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/25/2017 03:25:16

Diagnostics Check		OK	03:15	
Air Blank		0.000	03:15	
Control Test		0.077	03:15	
Air Blank		0.000	03:16	
Subject Sample	# 1	0.122	03:19	Volume Not Met
Breath Volume		0.832	03:19	
Air Blank		0.000	03:20	
Air Blank		0.000	03:22	
Subject Sample	# 2		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.076	03:24	
Air Blank		0.000	03:25	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 07/26/2017 19:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PAULAUSKAS

DATE/TIME OF TEST: 07/27/2017 03:14:37

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.079	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.150	03:09
Breath Volume		1.863	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.155	03:13
Breath Volume		1.898	03:13
Air Blank		0.000	03:13
Control Test		0.079	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 08/23/2017 19:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,M,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HUDSON,A
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: GHAWALY

DATE/TIME OF TEST: 07/01/2017 22:45:22

Diagnostics Check		OK	22:42	
Air Blank		0.000	22:42	
Control Test		0.082	22:43	
Air Blank		0.000	22:43	
Subject Sample	# 1		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.082	22:44	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:45	

Data Download Date/Time: 07/23/2017 23:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,M,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: PARKS,T
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: TERRAZA RAYMUNDO

DATE/TIME OF TEST: 07/02/2017 03:55:59

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:53	
Control Test		0.082	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:54	Subject Test Refused
Breath Volume		0.000	03:54	
Air Blank		0.000	03:55	
Control Test		0.082	03:55	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 07/23/2017 23:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: URBANOWSKI,TROY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: STUDLEY

DATE/TIME OF TEST: 07/07/2017 21:13:16

Diagnostics Check		OK	21:06	
Air Blank		0.000	21:07	
Control Test		0.081	21:07	
Air Blank		0.000	21:08	
Subject Sample	# 1	0.212	21:08	
Breath Volume		2.843	21:08	
Air Blank		0.000	21:09	
Air Blank		0.000	21:11	
Subject Sample	# 2	0.198	21:11	
Breath Volume		2.550	21:11	
Air Blank		0.000	21:12	
Control Test		0.065	21:12	Control Outside Tolerance
Air Blank		0.000	21:13	

Data Download Date/Time: 07/23/2017 23:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: URBANOWSKI,TROY
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: STUDLEY

DATE/TIME OF TEST: 07/07/2017 21:27:09

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.081	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.218	21:22
Breath Volume		2.656	21:22
Air Blank		0.000	21:23
Air Blank		0.000	21:24
Subject Sample	# 2	0.219	21:25
Breath Volume		2.832	21:25
Air Blank		0.000	21:26
Control Test		0.081	21:26
Air Blank		0.000	21:27
Diagnostics Check		OK	21:27

Data Download Date/Time: 07/23/2017 23:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,M,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: TAYLOR,A
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 06/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: HARTLEY

DATE/TIME OF TEST: 07/14/2017 23:43:33

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.082	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.161	23:38
Breath Volume		2.792	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.160	23:41
Breath Volume		2.984	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 07/23/2017 23:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OLVERA,JESUS,
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: GIVANS,DAVIN
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 13:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:30

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 07/22/2017 14:56:17

Diagnostics Check		OK	14:53	
Air Blank		0.000	14:53	
Control Test		0.081	14:54	
Air Blank		0.000	14:54	
Subject Sample	# 1		14:54	Subject Test Refused
Breath Volume		0.000	14:54	
Air Blank		0.000	14:55	
Control Test		0.081	14:55	
Air Blank		0.000	14:56	
Diagnostics Check		OK	14:56	

Data Download Date/Time: 07/23/2017 23:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: BAYLY,BLAKE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: ZUNIGA

DATE/TIME OF TEST: 07/22/2017 20:05:19

Diagnostics Check		OK	19:56
Air Blank		0.000	19:56
Control Test		0.082	19:56
Air Blank		0.000	19:57
Subject Sample	# 1	0.224	19:59
Breath Volume		1.519	19:59
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.227	20:03
Breath Volume		1.539	20:03
Air Blank		0.000	20:04
Control Test		0.077	20:04
Air Blank		0.000	20:05
Diagnostics Check		OK	20:05

Data Download Date/Time: 07/23/2017 23:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006234

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,M,R
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: HUDSON,A
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/29/2017 01:24:43

Diagnostics Check		OK	01:21	
Air Blank		0.000	01:22	
Control Test		0.082	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.082	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 08/22/2017 21:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: CARWEIN

DATE/TIME OF TEST: 07/01/2017 23:26:13

Diagnostics Check		OK	23:19
Air Blank		0.000	23:19
Control Test		0.079	23:20
Air Blank		0.000	23:20
Subject Sample	# 1	0.141	23:21
Breath Volume		3.253	23:21
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.159	23:24
Breath Volume		2.746	23:24
Air Blank		0.000	23:25
Control Test		0.080	23:25
Air Blank		0.000	23:26
Diagnostics Check		OK	23:26

Data Download Date/Time: 07/23/2017 16:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: SANFORD

DATE/TIME OF TEST: 07/02/2017 04:45:12

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.080	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.128	04:40
Breath Volume		1.617	04:40
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.132	04:43
Breath Volume		1.984	04:43
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:45
Diagnostics Check		OK	04:45

Data Download Date/Time: 07/23/2017 16:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COE,J,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FREELAND,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 07/04/2017 17:56:57

Diagnostics Check		OK	17:49
Air Blank		0.000	17:50
Control Test		0.079	17:50
Air Blank		0.000	17:50
Subject Sample	# 1	0.260	17:51
Breath Volume		1.917	17:51
Air Blank		0.000	17:52
Air Blank		0.000	17:54
Subject Sample	# 2	0.249	17:55
Breath Volume		1.757	17:55
Air Blank		0.000	17:56
Control Test		0.078	17:56
Air Blank		0.000	17:56
Diagnostics Check		OK	17:56

Data Download Date/Time: 07/23/2017 16:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNEDRS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BOATRIGHT,OFFICER
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: SPIKOL

DATE/TIME OF TEST: 07/06/2017 23:19:57

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.080	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.098	23:15
Breath Volume		2.378	23:15
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.096	23:18
Breath Volume		2.289	23:18
Air Blank		0.000	23:18
Control Test		0.079	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 07/23/2017 16:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: FILLINGIM

DATE/TIME OF TEST: 07/07/2017 00:56:48

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:50
Air Blank		0.000	00:51
Subject Sample	# 1	0.185	00:52
Breath Volume		3.109	00:52
Air Blank		0.000	00:52
Air Blank		0.000	00:54
Subject Sample	# 2	0.183	00:55
Breath Volume		2.265	00:55
Air Blank		0.000	00:55
Control Test		0.079	00:56
Air Blank		0.000	00:56
Diagnostics Check		OK	00:56

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FIELDS,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PADGETT

DATE/TIME OF TEST: 07/08/2017 05:14:13

Diagnostics Check		OK	05:07
Air Blank		0.000	05:07
Control Test		0.079	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.168	05:09
Breath Volume		1.542	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:11
Subject Sample	# 2	0.167	05:12
Breath Volume		1.664	05:12
Air Blank		0.000	05:13
Control Test		0.079	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 07/23/2017 16:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STACY,A,L
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BRIGGS,G
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:43

SUBJECT'S LAST NAME: QUADE

DATE/TIME OF TEST: 07/08/2017 06:15:10

Diagnostics Check		OK	06:06	
Air Blank		0.000	06:07	
Control Test		0.079	06:07	
Air Blank		0.000	06:07	
Subject Sample	# 1		06:11	No Sample Provided
Breath Volume		0.000	06:11	
Air Blank		0.000	06:11	
Air Blank		0.000	06:13	
Subject Sample	# 2		06:13	Subject Test Refused
Breath Volume		0.000	06:13	
Air Blank		0.000	06:14	
Control Test		0.080	06:14	
Air Blank		0.000	06:15	
Diagnostics Check		OK	06:15	

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: RAY,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: DIAZ GUZMAN

DATE/TIME OF TEST: 07/09/2017 01:12:30

Diagnostics Check		OK	01:04
Air Blank		0.000	01:05
Control Test		0.079	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.208	01:07
Breath Volume		1.699	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:09
Subject Sample	# 2	0.212	01:10
Breath Volume		1.722	01:10
Air Blank		0.000	01:11
Control Test		0.078	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SENAK,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ICKES

DATE/TIME OF TEST: 07/09/2017 02:26:12

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.079	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:19	No Sample Provided
Breath Volume		0.000	02:19	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2		02:24	No Sample Provided
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Control Test		0.080	02:25	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: NELSON,OFFICER
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: PAZHANISWAMY

DATE/TIME OF TEST: 07/09/2017 05:03:26

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.080	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:56	No Sample Provided
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 2		05:02	No Sample Provided
Breath Volume		0.000	05:02	
Air Blank		0.000	05:02	
Control Test		0.080	05:02	
Air Blank		0.000	05:03	
Diagnostics Check		OK	05:03	

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VASQUEZ,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 07/09/2017 22:11:24

Diagnostics Check		OK	22:04
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.198	22:06
Breath Volume		3.238	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.188	22:09
Breath Volume		2.847	22:09
Air Blank		0.000	22:10
Control Test		0.079	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 07/23/2017 16:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LINEBERRY,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: ELDRIDGE

DATE/TIME OF TEST: 07/12/2017 03:16:51

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.242	03:11
Breath Volume		2.214	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.238	03:15
Breath Volume		1.601	03:15
Air Blank		0.000	03:15
Control Test		0.077	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JOHN,K
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CASTELLI,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: BUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/15/2017 18:00:11

Diagnostics Check		OK	17:51	
Air Blank		0.000	17:52	
Control Test		0.079	17:52	
Air Blank		0.000	17:52	
Subject Sample	# 1		17:56	No Sample Provided
Breath Volume		0.000	17:56	
Air Blank		0.000	17:56	
Air Blank		0.000	17:58	
Subject Sample	# 2		17:58	Subject Test Refused
Breath Volume		0.000	17:58	
Air Blank		0.000	17:59	
Control Test		0.080	17:59	
Air Blank		0.000	18:00	
Diagnostics Check		OK	18:00	

Data Download Date/Time: 07/23/2017 16:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GRACKI,N
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CRABB

DATE/TIME OF TEST: 07/16/2017 05:57:23

Diagnostics Check		OK	05:50
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:51
Subject Sample	# 1	0.162	05:52
Breath Volume		3.585	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:54
Subject Sample	# 2	0.155	05:55
Breath Volume		2.640	05:55
Air Blank		0.000	05:56
Control Test		0.079	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 07/23/2017 16:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: AVILA-IRIAS

DATE/TIME OF TEST: 07/18/2017 18:34:06

Diagnostics Check		OK	18:26
Air Blank		0.000	18:26
Control Test		0.080	18:27
Air Blank		0.000	18:27
Subject Sample	# 1	0.201	18:28
Breath Volume		1.601	18:28
Air Blank		0.000	18:29
Air Blank		0.000	18:31
Subject Sample	# 2	0.200	18:32
Breath Volume		2.046	18:32
Air Blank		0.000	18:33
Control Test		0.079	18:33
Air Blank		0.000	18:34
Diagnostics Check		OK	18:34

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SAUNDERS,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: STAMER

DATE/TIME OF TEST: 07/21/2017 20:14:06

Diagnostics Check		OK	20:02	
Air Blank		0.000	20:03	
Control Test		0.080	20:03	
Air Blank		0.000	20:03	
Subject Sample	# 1		20:07	No Sample Provided
Breath Volume		0.000	20:07	
Air Blank		0.000	20:07	
Air Blank		0.000	20:09	
Subject Sample	# 2		20:12	No Sample Provided
Breath Volume		0.000	20:12	
Air Blank		0.000	20:13	
Control Test		0.080	20:13	
Air Blank		0.000	20:14	
Diagnostics Check		OK	20:14	

Data Download Date/Time: 07/23/2017 16:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINSON,C,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 06/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GRAYCOCHEA

DATE/TIME OF TEST: 07/21/2017 23:28:19

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.113	23:23
Breath Volume		1.882	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.107	23:26
Breath Volume		1.742	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 07/23/2017 16:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SENAK,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 07/27/2017 03:07:12

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.041	03:02
Breath Volume		2.292	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.040	03:05
Breath Volume		2.589	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 08/22/2017 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,MATTHEW,P
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: THORNTON

DATE/TIME OF TEST: 07/29/2017 05:06:07

Diagnostics Check		OK	04:56
Air Blank		0.000	04:56
Control Test		0.080	04:57
Air Blank		0.000	04:57
Subject Sample	# 1	0.170	04:58
Breath Volume		2.882	04:58
Air Blank		0.000	04:59
Air Blank		0.000	05:00
Subject Sample	# 2	0.148	05:01
Breath Volume		1.250	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 3	0.159	05:04
Breath Volume		2.257	05:04
Air Blank		0.000	05:05
Control Test		0.079	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 08/22/2017 15:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: DRISKELL,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: COURTNEY

DATE/TIME OF TEST: 07/31/2017 22:26:39

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.080	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:22	No Sample Provided
Breath Volume		0.000	22:22	
Air Blank		0.000	22:23	
Air Blank		0.000	22:25	
Subject Sample	# 2		22:25	Subject Test Refused
Breath Volume		0.000	22:25	
Air Blank		0.000	22:25	
Control Test		0.080	22:26	
Air Blank		0.000	22:26	
Diagnostics Check		OK	22:26	

Data Download Date/Time: 08/22/2017 15:46

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: ROSSEL

DATE/TIME OF TEST: 07/01/2017 01:13:56

Diagnostics Check		OK	01:05
Air Blank		0.000	01:05
Control Test		0.081	01:05
Air Blank		0.000	01:06
Subject Sample	# 1	0.199	01:07
Breath Volume		1.382	01:07
Air Blank		0.000	01:08
Air Blank		0.000	01:10
Subject Sample	# 2	0.209	01:12
Breath Volume		2.312	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:13
Diagnostics Check		OK	01:13

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 07/01/2017 03:22:31

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.080	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.199	03:17
Breath Volume		2.250	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:20
Subject Sample	# 2	0.201	03:20
Breath Volume		1.878	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KHATAMI,KHASH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 07/01/2017 21:57:52

Diagnostics Check		OK	21:51
Air Blank		0.000	21:52
Control Test		0.081	21:52
Air Blank		0.000	21:53
Subject Sample	# 1	0.133	21:53
Breath Volume		1.457	21:53
Air Blank		0.000	21:54
Air Blank		0.000	21:55
Subject Sample	# 2	0.140	21:56
Breath Volume		2.320	21:56
Air Blank		0.000	21:56
Control Test		0.081	21:57
Air Blank		0.000	21:57
Diagnostics Check		OK	21:57

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MOTEJO

DATE/TIME OF TEST: 07/02/2017 03:30:14

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.205	03:25
Breath Volume		1.820	03:25
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.219	03:28
Breath Volume		2.734	03:28
Air Blank		0.000	03:29
Control Test		0.080	03:29
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GARCIA,BRIAN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CALEL

DATE/TIME OF TEST: 07/02/2017 04:26:24

Diagnostics Check		OK	04:19
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.183	04:21
Breath Volume		2.824	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.175	04:24
Breath Volume		2.328	04:24
Air Blank		0.000	04:25
Control Test		0.080	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 07/02/2017 23:00:02

Diagnostics Check		OK	22:45	
Air Blank		0.000	22:45	
Control Test		0.080	22:46	
Air Blank		0.000	22:46	
Subject Sample	# 1	0.163	22:47	
Breath Volume		3.132	22:47	
Air Blank		0.000	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 2		22:52	No Sample Provided
Breath Volume		0.000	22:52	
Air Blank		0.000	22:53	
Air Blank		0.000	22:55	
Subject Sample	# 3		22:58	No Sample Provided
Breath Volume		0.000	22:58	
Air Blank		0.000	22:59	
Control Test		0.080	22:59	
Air Blank		0.000	22:59	
Diagnostics Check		OK	23:00	

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 07/02/2017 23:22:23

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.080	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.165	23:17
Breath Volume		2.273	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.157	23:20
Breath Volume		2.980	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: BROOKS

DATE/TIME OF TEST: 07/04/2017 02:29:55

Diagnostics Check		OK	02:23
Air Blank		0.000	02:23
Control Test		0.081	02:24
Air Blank		0.000	02:24
Subject Sample	# 1	0.255	02:25
Breath Volume		2.062	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:27
Subject Sample	# 2	0.255	02:28
Breath Volume		2.066	02:28
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:29
Diagnostics Check		OK	02:29

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: LAURISTON

DATE/TIME OF TEST: 07/04/2017 04:51:18

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.081	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.270	04:46
Breath Volume		2.082	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.270	04:49
Breath Volume		1.792	04:49
Air Blank		0.000	04:50
Control Test		0.080	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: CHANCEY

DATE/TIME OF TEST: 07/04/2017 22:08:30

Diagnostics Check		OK	22:01
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:02
Subject Sample	# 1	0.160	22:03
Breath Volume		1.441	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.149	22:06
Breath Volume		1.148	22:06
Air Blank		0.000	22:07
Control Test		0.081	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: FERRARIS

DATE/TIME OF TEST: 07/05/2017 01:22:48

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.053	01:18
Breath Volume		1.746	01:18
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.053	01:21
Breath Volume		2.343	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:22
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: OWEN,MATTHEW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 07/05/2017 16:10:43

Diagnostics Check		OK	16:02
Air Blank		0.000	16:02
Control Test		0.081	16:02
Air Blank		0.000	16:03
Subject Sample	# 1	0.088	16:05
Breath Volume		1.894	16:05
Air Blank		0.000	16:06
Air Blank		0.000	16:08
Subject Sample	# 2	0.082	16:09
Breath Volume		1.746	16:09
Air Blank		0.000	16:09
Control Test		0.079	16:10
Air Blank		0.000	16:10
Diagnostics Check		OK	16:10

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLEARY,GUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PETER

DATE/TIME OF TEST: 07/06/2017 03:17:16

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.080	03:00	
Air Blank		0.000	03:01	
Subject Sample	# 1	0.155	03:04	Slope Not Level
Breath Volume		1.648	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:06	
Subject Sample	# 2	0.157	03:10	
Breath Volume		1.992	03:10	
Air Blank		0.000	03:10	
Air Blank		0.000	03:12	
Subject Sample	# 3	0.166	03:15	
Breath Volume		2.246	03:15	
Air Blank		0.000	03:16	
Control Test		0.079	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CROY,PAUL
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 07/07/2017 00:35:02

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.081	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STUBBLEBINE,LIVIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: MCKINNEY

DATE/TIME OF TEST: 07/07/2017 01:29:06

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.086	01:24
Breath Volume		1.304	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.083	01:27
Breath Volume		1.679	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODRIGUEZ,LENDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GRAYSON

DATE/TIME OF TEST: 07/08/2017 02:26:00

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.081	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.132	02:21
Breath Volume		1.269	02:21
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.148	02:24
Breath Volume		3.929	02:24
Air Blank		0.000	02:25
Control Test		0.081	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:26

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: CHARLOTIN

DATE/TIME OF TEST: 07/08/2017 03:23:36

Diagnostics Check		OK	03:16
Air Blank		0.000	03:16
Control Test		0.081	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.082	03:19
Breath Volume		1.957	03:19
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.080	03:22
Breath Volume		1.835	03:22
Air Blank		0.000	03:22
Control Test		0.080	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 07/21/2017 12:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: DEL-RIO

DATE/TIME OF TEST: 07/08/2017 04:22:07

Diagnostics Check		OK	04:15
Air Blank		0.000	04:16
Control Test		0.080	04:16
Air Blank		0.000	04:17
Subject Sample	# 1	0.103	04:17
Breath Volume		1.585	04:17
Air Blank		0.000	04:18
Air Blank		0.000	04:20
Subject Sample	# 2	0.100	04:20
Breath Volume		1.429	04:20
Air Blank		0.000	04:21
Control Test		0.080	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: HERNANDEZ ALFARO

DATE/TIME OF TEST: 07/08/2017 23:36:59

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.081	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.119	23:32
Breath Volume		2.363	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:34
Subject Sample	# 2	0.116	23:35
Breath Volume		2.425	23:35
Air Blank		0.000	23:36
Control Test		0.081	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BARKER,ZACHARY
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: BARKER

DATE/TIME OF TEST: 07/09/2017 00:59:25

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.185	00:54
Breath Volume		1.121	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.195	00:57
Breath Volume		1.781	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MENDEZ FELIPE

DATE/TIME OF TEST: 07/09/2017 01:20:47

Diagnostics Check		OK	01:14
Air Blank		0.000	01:15
Control Test		0.080	01:15
Air Blank		0.000	01:15
Subject Sample	# 1	0.170	01:16
Breath Volume		1.113	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:18
Subject Sample	# 2	0.165	01:19
Breath Volume		2.667	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: RODGERS,BRANDY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: GUEVARA CARDENAS

DATE/TIME OF TEST: 07/09/2017 04:18:08

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.080	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.269	04:13
Breath Volume		3.609	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:15
Subject Sample	# 2	0.275	04:16
Breath Volume		3.675	04:16
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 07/10/2017 00:33:03

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.081	00:25
Air Blank		0.000	00:25
Subject Sample	# 1	0.175	00:28
Breath Volume		1.269	00:28
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.169	00:31
Breath Volume		2.785	00:31
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: QUINTANILLA

DATE/TIME OF TEST: 07/11/2017 04:34:47

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.187	04:30
Breath Volume		2.867	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:32
Subject Sample	# 2	0.186	04:33
Breath Volume		1.558	04:33
Air Blank		0.000	04:33
Control Test		0.080	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHENIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 07/12/2017 02:15:36

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:09
Air Blank		0.000	02:09
Subject Sample	# 1	0.033	02:10
Breath Volume		1.488	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.035	02:14
Breath Volume		1.652	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 07/12/2017 04:31:39

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.152	04:27
Breath Volume		1.832	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.152	04:30
Breath Volume		2.789	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FEQUIERE,RALPH
ARRESTING OFFICER AGENCY: Palm Springs PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VALENZUELA GANDARILA

DATE/TIME OF TEST: 07/14/2017 00:02:59

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.150	23:57
Breath Volume		2.160	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.154	00:01
Breath Volume		3.460	00:01
Air Blank		0.000	00:02
Control Test		0.081	00:02
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/14/2017 23:00:09

Diagnostics Check		OK	22:53
Air Blank		0.000	22:54
Control Test		0.080	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.216	22:55
Breath Volume		1.683	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.224	22:58
Breath Volume		2.640	22:58
Air Blank		0.000	22:59
Control Test		0.080	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: CABRERA MANSO

DATE/TIME OF TEST: 07/15/2017 00:19:41

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.081	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.236	00:15
Breath Volume		1.687	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.234	00:18
Breath Volume		2.492	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: POMIANOWSKI

DATE/TIME OF TEST: 07/15/2017 03:04:54

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:01	
Control Test		0.081	03:01	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:03	Subject Test Refused
Breath Volume		0.000	03:03	
Air Blank		0.000	03:03	
Control Test		0.080	03:04	
Air Blank		0.000	03:04	
Diagnostics Check		OK	03:04	

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WEBER,RONALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: MARCHESINI

DATE/TIME OF TEST: 07/15/2017 07:30:21

Diagnostics Check		OK	07:23
Air Blank		0.000	07:24
Control Test		0.081	07:24
Air Blank		0.000	07:25
Subject Sample	# 1	0.146	07:25
Breath Volume		3.347	07:25
Air Blank		0.000	07:26
Air Blank		0.000	07:28
Subject Sample	# 2	0.155	07:28
Breath Volume		2.937	07:28
Air Blank		0.000	07:29
Control Test		0.080	07:29
Air Blank		0.000	07:30
Diagnostics Check		OK	07:30

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 07/16/2017 00:00:09

Diagnostics Check		OK	23:53
Air Blank		0.000	23:54
Control Test		0.081	23:54
Air Blank		0.000	23:55
Subject Sample	# 1	0.111	23:55
Breath Volume		3.230	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.108	23:58
Breath Volume		2.257	23:58
Air Blank		0.000	23:59
Control Test		0.080	23:59
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: MATUTE

DATE/TIME OF TEST: 07/16/2017 00:42:24

Diagnostics Check		OK	00:35
Air Blank		0.000	00:36
Control Test		0.081	00:36
Air Blank		0.000	00:37
Subject Sample	# 1	0.131	00:37
Breath Volume		1.226	00:37
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.135	00:40
Breath Volume		1.347	00:40
Air Blank		0.000	00:41
Control Test		0.080	00:41
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PUJOL,JESUS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: GROOMS

DATE/TIME OF TEST: 07/16/2017 01:25:27

Diagnostics Check		OK	01:19
Air Blank		0.000	01:19
Control Test		0.080	01:20
Air Blank		0.000	01:20
Subject Sample	# 1	0.218	01:21
Breath Volume		2.558	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.232	01:23
Breath Volume		2.812	01:23
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 07/16/2017 02:54:19

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.080	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.000	02:50
Breath Volume		2.511	02:50
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.000	02:52
Breath Volume		2.144	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BANKS

DATE/TIME OF TEST: 07/16/2017 03:58:57

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.069	03:54
Breath Volume		2.968	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:56
Subject Sample	# 2	0.066	03:57
Breath Volume		2.367	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: ZAMBRANO

DATE/TIME OF TEST: 07/16/2017 05:58:16

Diagnostics Check		OK	05:45	
Air Blank		0.000	05:46	
Control Test		0.080	05:46	
Air Blank		0.000	05:47	
Subject Sample	# 1	0.180	05:47	
Breath Volume		2.410	05:47	
Air Blank		0.000	05:48	
Air Blank		0.000	05:50	
Subject Sample	# 2	0.156	05:53	Volume Not Met
Breath Volume		0.417	05:53	
Air Blank		0.000	05:54	
Air Blank		0.000	05:56	
Subject Sample	# 3	0.192	05:56	
Breath Volume		2.976	05:56	
Air Blank		0.000	05:57	
Control Test		0.078	05:57	
Air Blank		0.000	05:58	
Diagnostics Check		OK	05:58	

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO
RODRIGUEZ,LENNDEY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: STETTEN

DATE/TIME OF TEST: 07/16/2017 22:29:59

Diagnostics Check		OK	22:21
Air Blank		0.000	22:22
Control Test		0.080	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.100	22:23
Breath Volume		1.140	22:23
Air Blank		0.000	22:24
Air Blank		0.000	22:25
Subject Sample	# 2	0.098	22:28
Breath Volume		2.613	22:28
Air Blank		0.000	22:29
Control Test		0.081	22:29
Air Blank		0.000	22:29
Diagnostics Check		OK	22:29

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BOORAS,DALTON
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/16/2017 22:55:15

Diagnostics Check		OK	22:49
Air Blank		0.000	22:49
Control Test		0.080	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.226	22:50
Breath Volume		2.687	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.214	22:53
Breath Volume		3.246	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: KELLY,MAX
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 07/17/2017 04:09:04

Diagnostics Check		OK	04:03
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.000	04:04
Breath Volume		1.519	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.000	04:07
Breath Volume		1.281	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:09

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GRUBER,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: HYDE

DATE/TIME OF TEST: 07/18/2017 01:50:07

Diagnostics Check		OK	01:42
Air Blank		0.000	01:43
Control Test		0.081	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.000	01:45
Breath Volume		1.785	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.000	01:48
Breath Volume		2.367	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 07/21/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: SMART

DATE/TIME OF TEST: 07/19/2017 00:49:18

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.000	00:44
Breath Volume		1.718	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.000	00:47
Breath Volume		1.832	00:47
Air Blank		0.000	00:48
Control Test		0.081	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 07/21/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VEGA,JAIME
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: SARTORI

DATE/TIME OF TEST: 07/20/2017 22:59:10

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.080	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.115	22:54
Breath Volume		1.113	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:56
Subject Sample	# 2	0.122	22:57
Breath Volume		1.878	22:57
Air Blank		0.000	22:58
Control Test		0.081	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 07/21/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 07/21/2017 06:23:44

Diagnostics Check		OK	06:16
Air Blank		0.000	06:16
Control Test		0.080	06:17
Air Blank		0.000	06:17
Subject Sample	# 1	0.141	06:18
Breath Volume		2.285	06:18
Air Blank		0.000	06:19
Air Blank		0.000	06:20
Subject Sample	# 2	0.132	06:22
Breath Volume		2.035	06:22
Air Blank		0.000	06:22
Control Test		0.081	06:23
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 07/21/2017 12:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: BELTRAN

DATE/TIME OF TEST: 07/21/2017 21:21:42

Diagnostics Check		OK	21:15
Air Blank		0.000	21:15
Control Test		0.081	21:16
Air Blank		0.000	21:16
Subject Sample	# 1	0.262	21:17
Breath Volume		1.980	21:17
Air Blank		0.000	21:17
Air Blank		0.000	21:19
Subject Sample	# 2	0.263	21:20
Breath Volume		1.457	21:20
Air Blank		0.000	21:20
Control Test		0.081	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 08/04/2017 14:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: GUILD

DATE/TIME OF TEST: 07/21/2017 22:04:19

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.080	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.170	21:59
Breath Volume		1.421	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:01
Subject Sample	# 2	0.169	22:02
Breath Volume		2.000	22:02
Air Blank		0.000	22:03
Control Test		0.080	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MURILLO

DATE/TIME OF TEST: 07/21/2017 23:28:00

Diagnostics Check		OK	23:21
Air Blank		0.000	23:21
Control Test		0.080	23:22
Air Blank		0.000	23:22
Subject Sample	# 1	0.000	23:23
Breath Volume		1.769	23:23
Air Blank		0.000	23:24
Air Blank		0.000	23:26
Subject Sample	# 2	0.000	23:26
Breath Volume		2.261	23:26
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:28

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SENTMANAT,ADOLFO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: VARGAS-AMADOR

DATE/TIME OF TEST: 07/22/2017 01:17:15

Diagnostics Check		OK	01:10
Air Blank		0.000	01:10
Control Test		0.081	01:11
Air Blank		0.000	01:11
Subject Sample	# 1	0.171	01:12
Breath Volume		1.468	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.178	01:15
Breath Volume		2.648	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 07/22/2017 02:50:06

Diagnostics Check		OK	02:32	
Air Blank		0.000	02:33	
Control Test		0.080	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.156	02:37	
Breath Volume		1.183	02:37	
Air Blank		0.000	02:37	
Air Blank		0.000	02:39	
Subject Sample	# 2		02:42	No Sample Provided
Breath Volume		0.000	02:42	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.154	02:48	Volume Not Met
Breath Volume		0.207	02:48	
Air Blank		0.000	02:49	
Control Test		0.080	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: HUGGINS

DATE/TIME OF TEST: 07/22/2017 02:59:56

Diagnostics Check		OK	02:51
Air Blank		0.000	02:52
Control Test		0.079	02:52
Air Blank		0.000	02:52
Subject Sample	# 1	0.161	02:54
Breath Volume		2.097	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.162	02:58
Breath Volume		1.921	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 07/22/2017 03:28:40

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:22
Air Blank		0.000	03:22
Subject Sample	# 1	0.183	03:23
Breath Volume		2.070	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.190	03:27
Breath Volume		2.843	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:28
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: GUSCOTT

DATE/TIME OF TEST: 07/22/2017 05:51:59

Diagnostics Check		OK	05:40	
Air Blank		0.000	05:40	
Control Test		0.081	05:41	
Air Blank		0.000	05:41	
Subject Sample	# 1	0.156	05:44	Volume Not Met
Breath Volume		0.222	05:44	
Air Blank		0.000	05:45	
Air Blank		0.000	05:47	
Subject Sample	# 2	0.157	05:50	Volume Not Met
Breath Volume		0.210	05:50	
Air Blank		0.000	05:51	
Control Test		0.079	05:51	
Air Blank		0.000	05:51	
Diagnostics Check		OK	05:51	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: GUSCOTT

DATE/TIME OF TEST: 07/22/2017 06:00:01

Diagnostics Check		OK	05:53
Air Blank		0.000	05:53
Control Test		0.080	05:54
Air Blank		0.000	05:54
Subject Sample	# 1	0.175	05:55
Breath Volume		2.160	05:55
Air Blank		0.000	05:55
Air Blank		0.000	05:57
Subject Sample	# 2	0.167	05:58
Breath Volume		1.835	05:58
Air Blank		0.000	05:59
Control Test		0.080	05:59
Air Blank		0.000	05:59
Diagnostics Check		OK	06:00

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: ROMERO APARICIO

DATE/TIME OF TEST: 07/22/2017 23:02:25

Diagnostics Check		OK	22:50	
Air Blank		0.000	22:50	
Control Test		0.081	22:51	
Air Blank		0.000	22:51	
Subject Sample	# 1	0.239	22:54	Volume Not Met
Breath Volume		0.828	22:54	
Air Blank		0.000	22:55	
Air Blank		0.000	22:57	
Subject Sample	# 2	0.261	22:57	
Breath Volume		1.468	22:57	
Air Blank		0.000	22:58	
Air Blank		0.000	23:00	
Subject Sample	# 3	0.256	23:00	
Breath Volume		1.140	23:00	
Air Blank		0.000	23:01	
Control Test		0.078	23:01	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: DEPHILLIPS

DATE/TIME OF TEST: 07/23/2017 04:13:22

Diagnostics Check		OK	04:09	
Air Blank		0.000	04:09	
Control Test		0.080	04:09	
Air Blank		0.000	04:10	
Subject Sample	# 1		04:12	Subject Test Refused
Breath Volume		0.000	04:12	
Air Blank		0.000	04:12	
Control Test		0.080	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLEARY,GUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/23/2017 05:54:06

Diagnostics Check		OK	05:48
Air Blank		0.000	05:48
Control Test		0.081	05:48
Air Blank		0.000	05:49
Subject Sample	# 1	0.205	05:49
Breath Volume		2.269	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:52
Subject Sample	# 2	0.196	05:52
Breath Volume		2.304	05:52
Air Blank		0.000	05:53
Control Test		0.080	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENDDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: RAYMUNDO-LOPEZ

DATE/TIME OF TEST: 07/24/2017 03:01:29

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.203	02:56
Breath Volume		1.425	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.203	02:59
Breath Volume		1.210	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 08/04/2017 14:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: COISSY

DATE/TIME OF TEST: 07/24/2017 05:37:09

Diagnostics Check		OK	05:25	
Air Blank		0.000	05:25	
Control Test		0.080	05:25	
Air Blank		0.000	05:26	
Subject Sample	# 1		05:29	No Sample Provided
Breath Volume		0.000	05:29	
Air Blank		0.000	05:29	
Air Blank		0.000	05:31	
Subject Sample	# 2	0.128	05:32	
Breath Volume		1.332	05:32	
Air Blank		0.000	05:33	
Air Blank		0.000	05:35	
Subject Sample	# 3	0.123	05:35	
Breath Volume		2.152	05:35	
Air Blank		0.000	05:36	
Control Test		0.080	05:36	
Air Blank		0.000	05:37	
Diagnostics Check		OK	05:37	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANDZIK,BIRGIT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: SIMSON

DATE/TIME OF TEST: 07/26/2017 00:37:59

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:33	
Control Test		0.081	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:36	Subject Test Refused
Breath Volume		0.339	00:36	
Air Blank		0.000	00:37	
Control Test		0.081	00:37	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHURCH,RYAN
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:37

SUBJECT'S LAST NAME: STODDARD

DATE/TIME OF TEST: 07/26/2017 06:10:28

Diagnostics Check		OK	06:02
Air Blank		0.000	06:02
Control Test		0.081	06:03
Air Blank		0.000	06:03
Subject Sample	# 1	0.272	06:04
Breath Volume		2.929	06:04
Air Blank		0.000	06:05
Air Blank		0.000	06:07
Subject Sample	# 2	0.264	06:08
Breath Volume		2.214	06:08
Air Blank		0.000	06:09
Control Test		0.079	06:09
Air Blank		0.000	06:10
Diagnostics Check		OK	06:10

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: RUND

DATE/TIME OF TEST: 07/26/2017 23:40:20

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:29	
Control Test		0.080	23:29	
Air Blank		0.000	23:30	
Subject Sample	# 1	0.144	23:33	Volume Not Met
Breath Volume		0.226	23:33	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 2	0.109	23:38	Volume Not Met
Breath Volume		0.265	23:38	
Air Blank		0.000	23:39	
Control Test		0.079	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: RUND

DATE/TIME OF TEST: 07/26/2017 23:57:38

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.079	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.179	23:46
Breath Volume		2.140	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:48
Subject Sample	# 2	0.153	23:51
Breath Volume		1.171	23:51
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 3	0.168	23:55
Breath Volume		1.484	23:55
Air Blank		0.000	23:56
Control Test		0.077	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOHNSON,GARFIELD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WOOLARD

DATE/TIME OF TEST: 07/27/2017 21:30:01

Diagnostics Check		OK	21:23
Air Blank		0.000	21:24
Control Test		0.080	21:24
Air Blank		0.000	21:25
Subject Sample	# 1	0.000	21:25
Breath Volume		2.457	21:25
Air Blank		0.000	21:26
Air Blank		0.000	21:28
Subject Sample	# 2	0.000	21:28
Breath Volume		2.046	21:28
Air Blank		0.000	21:29
Control Test		0.080	21:29
Air Blank		0.000	21:29
Diagnostics Check		OK	21:30

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TFANK,ALEXANDER
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 07/28/2017 05:14:15

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:08
Subject Sample	# 1	0.108	05:09
Breath Volume		2.183	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.107	05:12
Breath Volume		1.667	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:14
Diagnostics Check		OK	05:14

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SACKIE-MENSAH,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:45

SUBJECT'S LAST NAME: GARCIA GONZALEZ

DATE/TIME OF TEST: 07/28/2017 16:40:32

Diagnostics Check		OK	16:34
Air Blank		0.000	16:34
Control Test		0.081	16:34
Air Blank		0.000	16:35
Subject Sample	# 1	0.218	16:35
Breath Volume		1.464	16:35
Air Blank		0.000	16:36
Air Blank		0.000	16:38
Subject Sample	# 2	0.214	16:38
Breath Volume		3.222	16:38
Air Blank		0.000	16:39
Control Test		0.080	16:39
Air Blank		0.000	16:40
Diagnostics Check		OK	16:40

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PARENT,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: URREA

DATE/TIME OF TEST: 07/28/2017 20:23:19

Diagnostics Check		OK	20:14
Air Blank		0.000	20:14
Control Test		0.080	20:14
Air Blank		0.000	20:15
Subject Sample	# 1	0.258	20:18
Breath Volume		1.636	20:18
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.262	20:21
Breath Volume		1.105	20:21
Air Blank		0.000	20:22
Control Test		0.078	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JOHNSON,GARFIELD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: GONZALEZ VELASQUEZ

DATE/TIME OF TEST: 07/28/2017 22:12:21

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.197	22:07
Breath Volume		2.152	22:07
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.187	22:10
Breath Volume		1.621	22:10
Air Blank		0.000	22:11
Control Test		0.080	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: SERGIL

DATE/TIME OF TEST: 07/29/2017 00:54:31

Diagnostics Check		OK	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Subject Sample	# 1	0.211	00:49
Breath Volume		2.007	00:49
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.211	00:52
Breath Volume		4.023	00:52
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LYNCH,GREGORY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: BERDUO-MAURICIO

DATE/TIME OF TEST: 07/29/2017 01:26:48

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.080	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.175	01:22
Breath Volume		2.593	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.177	01:25
Breath Volume		2.160	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ST HILAIRE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: KLOSINSKI

DATE/TIME OF TEST: 07/29/2017 06:13:46

Diagnostics Check		OK	06:06
Air Blank		0.000	06:07
Control Test		0.080	06:07
Air Blank		0.000	06:08
Subject Sample	# 1	0.176	06:08
Breath Volume		1.980	06:08
Air Blank		0.000	06:09
Air Blank		0.000	06:11
Subject Sample	# 2	0.172	06:12
Breath Volume		2.562	06:12
Air Blank		0.000	06:12
Control Test		0.081	06:13
Air Blank		0.000	06:13
Diagnostics Check		OK	06:13

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 07/29/2017 20:32:00

Diagnostics Check		OK	20:24
Air Blank		0.000	20:25
Control Test		0.080	20:25
Air Blank		0.000	20:25
Subject Sample	# 1	0.304	20:26
Breath Volume		1.207	20:26
Air Blank		0.000	20:27
Air Blank		0.000	20:29
Subject Sample	# 2	0.287	20:30
Breath Volume		1.121	20:30
Air Blank		0.000	20:31
Control Test		0.078	20:31
Air Blank		0.000	20:31
Diagnostics Check		OK	20:32

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STUBBLEBINE,LIVIA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/30/2017 00:35:23

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.080	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:34	Subject Test Refused
Breath Volume		0.000	00:34	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PARENT,AUSTIN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: VAROQUA

DATE/TIME OF TEST: 07/30/2017 19:12:04

Diagnostics Check		OK	19:04
Air Blank		0.000	19:04
Control Test		0.080	19:04
Air Blank		0.000	19:05
Subject Sample	# 1	0.000	19:06
Breath Volume		2.410	19:06
Air Blank		0.000	19:07
Air Blank		0.000	19:09
Subject Sample	# 2	0.000	19:10
Breath Volume		1.566	19:10
Air Blank		0.000	19:11
Control Test		0.079	19:11
Air Blank		0.000	19:11
Diagnostics Check		OK	19:12

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: AIME,DIANA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GONZALEZ JUAN

DATE/TIME OF TEST: 07/30/2017 20:15:17

Diagnostics Check		OK	20:07
Air Blank		0.000	20:08
Control Test		0.080	20:08
Air Blank		0.000	20:08
Subject Sample	# 1	0.166	20:10
Breath Volume		1.437	20:10
Air Blank		0.000	20:11
Air Blank		0.000	20:13
Subject Sample	# 2	0.157	20:13
Breath Volume		1.171	20:13
Air Blank		0.000	20:14
Control Test		0.078	20:14
Air Blank		0.000	20:15
Diagnostics Check		OK	20:15

Data Download Date/Time: 08/04/2017 14:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LOPEZ GOMEZ

DATE/TIME OF TEST: 07/01/2017 00:33:49

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.082	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.158	00:28
Breath Volume		2.050	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.154	00:32
Breath Volume		2.265	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:33
Air Blank		0.000	00:33
Diagnostics Check		OK	00:33

Data Download Date/Time: 07/21/2017 16:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: VAN DYKE

DATE/TIME OF TEST: 07/01/2017 02:03:25

Diagnostics Check		OK	01:54	
Air Blank		0.000	01:55	
Control Test		0.082	01:55	
Air Blank		0.000	01:56	
Subject Sample	# 1	0.000	01:59	Volume Not Met
Breath Volume		0.410	01:59	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:02	
Control Test		0.082	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 07/21/2017 16:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: SMITH,II

DATE/TIME OF TEST: 07/01/2017 23:11:43

Diagnostics Check		OK	23:05
Air Blank		0.000	23:05
Control Test		0.081	23:06
Air Blank		0.000	23:06
Subject Sample	# 1	0.196	23:07
Breath Volume		2.203	23:07
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.202	23:10
Breath Volume		2.988	23:10
Air Blank		0.000	23:10
Control Test		0.080	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 07/21/2017 16:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOWERS,GARY,W
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: HANKERSON

DATE/TIME OF TEST: 07/02/2017 03:46:54

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.111	03:42
Breath Volume		2.035	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:44
Subject Sample	# 2	0.111	03:45
Breath Volume		1.925	03:45
Air Blank		0.000	03:45
Control Test		0.081	03:46
Air Blank		0.000	03:46
Diagnostics Check		OK	03:46

Data Download Date/Time: 07/21/2017 16:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WEBER,RONALD
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:44

SUBJECT'S LAST NAME: SCHADE,III

DATE/TIME OF TEST: 07/02/2017 16:29:19

Diagnostics Check		OK	16:19
Air Blank		0.000	16:20
Control Test		0.081	16:20
Air Blank		0.000	16:21
Subject Sample	# 1	0.160	16:23
Breath Volume		1.351	16:23
Air Blank		0.000	16:23
Air Blank		0.000	16:25
Subject Sample	# 2	0.167	16:27
Breath Volume		1.351	16:27
Air Blank		0.000	16:28
Control Test		0.079	16:28
Air Blank		0.000	16:29
Diagnostics Check		OK	16:29

Data Download Date/Time: 07/21/2017 16:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DILLARD,DANIEL
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: NOLASCO LOPEZ

DATE/TIME OF TEST: 07/02/2017 19:01:59

Diagnostics Check		OK	18:52
Air Blank		0.000	18:53
Control Test		0.081	18:53
Air Blank		0.000	18:53
Subject Sample	# 1	0.199	18:56
Breath Volume		1.218	18:56
Air Blank		0.000	18:57
Air Blank		0.000	18:59
Subject Sample	# 2	0.199	19:00
Breath Volume		1.105	19:00
Air Blank		0.000	19:01
Control Test		0.078	19:01
Air Blank		0.000	19:01
Diagnostics Check		OK	19:01

Data Download Date/Time: 07/21/2017 16:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006240

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ORDONEZ

DATE/TIME OF TEST: 07/02/2017 23:11:13

Diagnostics Check		OK	22:57	
Air Blank		0.000	22:57	
Control Test		0.080	22:58	
Air Blank		0.000	22:58	
Subject Sample	# 1	0.111	22:59	
Breath Volume		2.441	22:59	
Air Blank		0.000	23:00	
Air Blank		0.000	23:02	
Subject Sample	# 2	0.103	23:05	Volume Not Met
Breath Volume		0.019	23:05	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 3		23:09	Subject Test Refused
Breath Volume		0.019	23:09	
Air Blank		0.000	23:10	
Control Test		0.078	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 07/21/2017 16:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 06/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: CRAIG

DATE/TIME OF TEST: 07/06/2017 02:19:18

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.078	02:13
Air Blank		0.000	02:13
Subject Sample	# 1	0.200	02:14
Breath Volume		2.218	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.210	02:17
Breath Volume		2.562	02:17
Air Blank		0.000	02:18
Control Test		0.078	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 07/06/2017 20:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/08/2017 23:32:32

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.078	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.138	23:27
Breath Volume		1.960	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.149	23:30
Breath Volume		1.589	23:30
Air Blank		0.000	23:31
Control Test		0.078	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 08/12/2017 21:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: RASMUSSEN

DATE/TIME OF TEST: 07/13/2017 03:18:33

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.132	03:13
Breath Volume		1.648	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.133	03:16
Breath Volume		1.511	03:16
Air Blank		0.000	03:17
Control Test		0.079	03:17
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 08/12/2017 21:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: AKINS

DATE/TIME OF TEST: 07/14/2017 02:48:10

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.143	02:43
Breath Volume		2.472	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.139	02:46
Breath Volume		2.449	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:48
Diagnostics Check		OK	02:48

Data Download Date/Time: 08/12/2017 21:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 07/15/2017 03:29:44

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:20	
Control Test		0.078	03:20	
Air Blank		0.000	03:21	
Subject Sample	# 1		03:24	No Sample Provided
Breath Volume		0.000	03:24	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 2		03:28	Subject Test Refused
Breath Volume		0.000	03:28	
Air Blank		0.000	03:28	
Control Test		0.079	03:29	
Air Blank		0.000	03:29	
Diagnostics Check		OK	03:29	

Data Download Date/Time: 08/12/2017 21:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 07/13/2017 00:47:04

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:41
Subject Sample	# 1	0.162	00:42
Breath Volume		1.925	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:44
Subject Sample	# 2	0.159	00:45
Breath Volume		1.234	00:45
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:46
Diagnostics Check		OK	00:47

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 07/14/2017 02:40:06

Diagnostics Check		OK	02:32
Air Blank		0.000	02:33
Control Test		0.081	02:33
Air Blank		0.000	02:34
Subject Sample	# 1	0.168	02:34
Breath Volume		1.253	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.175	02:38
Breath Volume		1.292	02:38
Air Blank		0.000	02:39
Control Test		0.080	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: STARCHENKO

DATE/TIME OF TEST: 07/15/2017 02:33:33

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:21	
Control Test		0.080	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.139	02:25	Volume Not Met
Breath Volume		0.742	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:28	
Subject Sample	# 2	0.196	02:28	
Breath Volume		1.628	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 3	0.216	02:31	
Breath Volume		2.226	02:31	
Air Blank		0.000	02:32	
Control Test		0.079	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 08/14/2017 19:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 07/15/2017 04:16:37

Diagnostics Check		OK	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:10
Air Blank		0.000	04:10
Subject Sample	# 1	0.100	04:11
Breath Volume		2.093	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.100	04:14
Breath Volume		1.691	04:14
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: OLSEN,T
ARRESTING OFFICER AGENCY: Tampa Airport PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SHISGAL

DATE/TIME OF TEST: 07/16/2017 01:07:16

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.080	00:59
Air Blank		0.000	00:59
Subject Sample	# 1	0.134	01:01
Breath Volume		1.468	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.136	01:05
Breath Volume		1.648	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: PATISHTAN

DATE/TIME OF TEST: 07/16/2017 04:10:36

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.191	04:06
Breath Volume		1.285	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.184	04:08
Breath Volume		1.351	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 07/16/2017 21:26:42

Diagnostics Check		OK	21:20
Air Blank		0.000	21:20
Control Test		0.080	21:21
Air Blank		0.000	21:21
Subject Sample	# 1	0.157	21:22
Breath Volume		2.527	21:22
Air Blank		0.000	21:22
Air Blank		0.000	21:24
Subject Sample	# 2	0.161	21:25
Breath Volume		2.062	21:25
Air Blank		0.000	21:25
Control Test		0.080	21:26
Air Blank		0.000	21:26
Diagnostics Check		OK	21:26

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: GRANDE BASILIO

DATE/TIME OF TEST: 07/17/2017 01:30:39

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.080	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.266	01:26
Breath Volume		2.890	01:26
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.263	01:28
Breath Volume		3.203	01:28
Air Blank		0.000	01:29
Control Test		0.079	01:30
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: CORNIEL-LANTIGUA

DATE/TIME OF TEST: 07/17/2017 04:00:01

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.174	03:55
Breath Volume		3.582	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.172	03:58
Breath Volume		4.066	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	04:00

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: LYON,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: SMITHSON

DATE/TIME OF TEST: 07/17/2017 05:51:58

Diagnostics Check		OK	05:45
Air Blank		0.000	05:46
Control Test		0.080	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.188	05:47
Breath Volume		2.718	05:47
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.193	05:50
Breath Volume		1.351	05:50
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 08/14/2017 19:54

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CAYCE,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:27

SUBJECT'S LAST NAME: ETZLER

DATE/TIME OF TEST: 07/17/2017 14:54:28

Diagnostics Check		OK	14:47
Air Blank		0.000	14:48
Control Test		0.080	14:48
Air Blank		0.000	14:49
Subject Sample	# 1	0.189	14:49
Breath Volume		1.539	14:49
Air Blank		0.000	14:50
Air Blank		0.000	14:52
Subject Sample	# 2	0.192	14:52
Breath Volume		1.992	14:52
Air Blank		0.000	14:53
Control Test		0.079	14:53
Air Blank		0.000	14:54
Diagnostics Check		OK	14:54

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: PEEK

DATE/TIME OF TEST: 07/19/2017 04:51:08

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.080	04:45
Air Blank		0.000	04:45
Subject Sample	# 1	0.161	04:46
Breath Volume		1.437	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:48
Subject Sample	# 2	0.170	04:49
Breath Volume		1.523	04:49
Air Blank		0.000	04:50
Control Test		0.079	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: CORDERO PEREZ

DATE/TIME OF TEST: 07/21/2017 04:55:39

Diagnostics Check		OK	04:48
Air Blank		0.000	04:48
Control Test		0.080	04:49
Air Blank		0.000	04:49
Subject Sample	# 1	0.160	04:50
Breath Volume		2.304	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:53
Subject Sample	# 2	0.161	04:54
Breath Volume		1.652	04:54
Air Blank		0.000	04:54
Control Test		0.079	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAVE,R
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/22/2017 03:18:58

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.081	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.142	03:13
Breath Volume		1.460	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.145	03:17
Breath Volume		1.718	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/22/2017 05:17:09

Diagnostics Check		OK	05:11
Air Blank		0.000	05:11
Control Test		0.080	05:11
Air Blank		0.000	05:12
Subject Sample	# 1	0.086	05:12
Breath Volume		2.218	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:15
Subject Sample	# 2	0.088	05:15
Breath Volume		2.433	05:15
Air Blank		0.000	05:16
Control Test		0.080	05:16
Air Blank		0.000	05:17
Diagnostics Check		OK	05:17

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: TERINO

DATE/TIME OF TEST: 07/22/2017 23:29:48

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:24	
Control Test		0.081	23:25	
Air Blank		0.000	23:25	
Subject Sample	# 1		23:28	Interferent Detect
Breath Volume		3.609	23:28	
Air Blank		0.000	23:28	
Control Test		0.080	23:29	
Air Blank		0.000	23:29	
Diagnostics Check		OK	23:29	

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: TERINO

DATE/TIME OF TEST: 07/22/2017 23:34:20

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.080	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:32	Interferent Detect
Breath Volume		2.789	23:32	
Air Blank		0.000	23:33	
Control Test		0.079	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DAWSY,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: STEUERNAGEL

DATE/TIME OF TEST: 07/23/2017 02:26:09

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Subject Sample	# 1	0.136	02:21
Breath Volume		2.863	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:23
Subject Sample	# 2	0.130	02:24
Breath Volume		2.750	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/23/2017 04:04:19

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.000	03:59
Breath Volume		2.894	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.000	04:02
Breath Volume		2.546	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:44

SUBJECT'S LAST NAME: MIRDALA

DATE/TIME OF TEST: 07/23/2017 06:12:08

Diagnostics Check		OK	06:05
Air Blank		0.000	06:05
Control Test		0.080	06:06
Air Blank		0.000	06:06
Subject Sample	# 1	0.121	06:07
Breath Volume		4.738	06:07
Air Blank		0.000	06:08
Air Blank		0.000	06:09
Subject Sample	# 2	0.125	06:10
Breath Volume		4.113	06:10
Air Blank		0.000	06:11
Control Test		0.079	06:11
Air Blank		0.000	06:12
Diagnostics Check		OK	06:12

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: WALTERS

DATE/TIME OF TEST: 07/24/2017 00:46:12

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.189	00:41
Breath Volume		3.566	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:43
Subject Sample	# 2	0.184	00:44
Breath Volume		3.359	00:44
Air Blank		0.000	00:45
Control Test		0.080	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRONSON,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CORDERO ALVAREZ

DATE/TIME OF TEST: 07/24/2017 05:09:15

Diagnostics Check		OK	05:02
Air Blank		0.000	05:02
Control Test		0.080	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.152	05:04
Breath Volume		2.515	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:06
Subject Sample	# 2	0.154	05:07
Breath Volume		1.847	05:07
Air Blank		0.000	05:08
Control Test		0.080	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PERRY,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: DANIELSEN

DATE/TIME OF TEST: 07/25/2017 16:50:50

Diagnostics Check		OK	16:39	
Air Blank		0.000	16:39	
Control Test		0.081	16:40	
Air Blank		0.000	16:40	
Subject Sample	# 1		16:43	No Sample Provided
Breath Volume		0.000	16:43	
Air Blank		0.000	16:44	
Air Blank		0.000	16:46	
Subject Sample	# 2		16:49	No Sample Provided
Breath Volume		0.000	16:49	
Air Blank		0.000	16:49	
Control Test		0.080	16:50	
Air Blank		0.000	16:50	
Diagnostics Check		OK	16:50	

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RAMIREZ VIDAL

DATE/TIME OF TEST: 07/26/2017 02:22:04

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.009	02:17
Breath Volume		3.105	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.009	02:20
Breath Volume		2.863	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: CRANSTON

DATE/TIME OF TEST: 07/26/2017 04:47:47

Diagnostics Check		OK	04:36	
Air Blank		0.000	04:36	
Control Test		0.080	04:37	
Air Blank		0.000	04:37	
Subject Sample	# 1		04:40	No Sample Provided
Breath Volume		0.000	04:40	
Air Blank		0.000	04:41	
Air Blank		0.000	04:43	
Subject Sample	# 2		04:46	No Sample Provided
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.079	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: NORTH

DATE/TIME OF TEST: 07/27/2017 00:49:03

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.081	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.126	00:43
Breath Volume		1.761	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.137	00:47
Breath Volume		2.542	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:49

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NONE,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: ZAMBORI

DATE/TIME OF TEST: 07/27/2017 04:37:32

Diagnostics Check		OK	04:33
Air Blank		0.000	04:33
Control Test		0.080	04:33
Air Blank		0.000	04:34
Subject Sample	# 1	0.257	04:35
Breath Volume		1.406	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:37

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: BADAY

DATE/TIME OF TEST: 07/28/2017 02:18:08

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Subject Sample	# 1	0.090	02:13
Breath Volume		1.812	02:13
Air Blank		0.000	02:14
Air Blank		0.000	02:15
Subject Sample	# 2	0.096	02:16
Breath Volume		1.484	02:16
Air Blank		0.000	02:17
Control Test		0.080	02:17
Air Blank		0.000	02:18
Diagnostics Check		OK	02:18

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/28/2017 05:10:19

Diagnostics Check		OK	04:58	
Air Blank		0.000	04:59	
Control Test		0.080	04:59	
Air Blank		0.000	05:00	
Subject Sample	# 1		05:03	No Sample Provided
Breath Volume		0.000	05:03	
Air Blank		0.000	05:03	
Air Blank		0.000	05:05	
Subject Sample	# 2		05:08	No Sample Provided
Breath Volume		0.000	05:08	
Air Blank		0.000	05:09	
Control Test		0.081	05:09	
Air Blank		0.000	05:10	
Diagnostics Check		OK	05:10	

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ANDERSON,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: SCHIFFER

DATE/TIME OF TEST: 07/29/2017 00:51:25

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:46
Subject Sample	# 1	0.231	00:46
Breath Volume		3.957	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.221	00:49
Breath Volume		3.789	00:49
Air Blank		0.000	00:50
Control Test		0.079	00:50
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: MALDONADO

DATE/TIME OF TEST: 07/29/2017 03:03:36

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.080	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.138	02:58
Breath Volume		1.242	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.139	03:01
Breath Volume		1.628	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 07/29/2017 04:09:55

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.038	04:04
Breath Volume		2.250	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.037	04:08
Breath Volume		2.308	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: WILFREDO-RODRIGUEZ

DATE/TIME OF TEST: 07/30/2017 02:37:35

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.202	02:33
Breath Volume		1.656	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.195	02:35
Breath Volume		1.910	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: MARQUEZ

DATE/TIME OF TEST: 07/30/2017 05:32:39

Diagnostics Check		OK	05:29	
Air Blank		0.000	05:29	
Control Test		0.080	05:29	
Air Blank		0.000	05:30	
Subject Sample	# 1		05:31	Subject Test Refused
Breath Volume		0.000	05:31	
Air Blank		0.000	05:31	
Control Test		0.080	05:32	
Air Blank		0.000	05:32	
Diagnostics Check		OK	05:32	

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 07/31/2017 23:34:06

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:25	
Control Test		0.080	23:25	
Air Blank		0.000	23:25	
Subject Sample	# 1	0.180	23:27	
Breath Volume		3.199	23:27	
Air Blank		0.000	23:28	
Air Blank		0.000	23:30	
Subject Sample	# 2		23:32	Subject Test Refused
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Control Test		0.077	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 08/14/2017 19:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAIRE,TRENTON,M
BREATH TEST OPERATOR AGENCY: OTHER Fed
ARRESTING OFFICER: HAIRE,TRENTON
ARRESTING OFFICER AGENCY: OTHER Fed
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: ALTHOUSE

DATE/TIME OF TEST: 07/02/2017 00:05:38

Diagnostics Check		OK	23:59
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Subject Sample	# 1	0.096	00:01
Breath Volume		3.757	00:01
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.086	00:03
Breath Volume		1.683	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:05
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 07/14/2017 08:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,S
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: HUFFMAN

DATE/TIME OF TEST: 07/05/2017 00:24:29

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.081	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.000	00:20
Breath Volume		1.476	00:20
Air Blank		0.000	00:20
Air Blank		0.000	00:22
Subject Sample	# 2	0.000	00:23
Breath Volume		1.496	00:23
Air Blank		0.000	00:23
Control Test		0.079	00:23
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 07/14/2017 08:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,JOHN,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: HUTCHINSON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 06/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 07/08/2017 21:22:23

Diagnostics Check		OK	21:16
Air Blank		0.000	21:16
Control Test		0.081	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.117	21:17
Breath Volume		1.652	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.125	21:20
Breath Volume		1.199	21:20
Air Blank		0.000	21:21
Control Test		0.081	21:21
Air Blank		0.000	21:22
Diagnostics Check		OK	21:22

Data Download Date/Time: 07/14/2017 08:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GORDON,GREGORY,S
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DONAHUE,BEN
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:56

SUBJECT'S LAST NAME: CALDERON

DATE/TIME OF TEST: 07/18/2017 13:24:11

Diagnostics Check		OK	13:21	
Air Blank		0.000	13:21	
Control Test		0.081	13:22	
Air Blank		0.000	13:22	
Subject Sample	# 1		13:22	Subject Test Refused
Breath Volume		0.000	13:22	
Air Blank		0.000	13:23	
Control Test		0.081	13:23	
Air Blank		0.000	13:24	
Diagnostics Check		OK	13:24	

Data Download Date/Time: 08/09/2017 08:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: ROBINS

DATE/TIME OF TEST: 07/20/2017 23:00:08

Diagnostics Check		OK	22:53
Air Blank		0.000	22:53
Control Test		0.081	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.165	22:55
Breath Volume		2.265	22:55
Air Blank		0.000	22:56
Air Blank		0.000	22:57
Subject Sample	# 2	0.160	22:58
Breath Volume		2.687	22:58
Air Blank		0.000	22:59
Control Test		0.078	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 08/09/2017 08:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: GRAVES,TIMOTHY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 07/22/2017 21:18:35

Diagnostics Check		OK	21:12
Air Blank		0.000	21:12
Control Test		0.082	21:12
Air Blank		0.000	21:13
Subject Sample	# 1	0.022	21:13
Breath Volume		4.296	21:13
Air Blank		0.000	21:14
Air Blank		0.000	21:16
Subject Sample	# 2	0.020	21:17
Breath Volume		2.843	21:17
Air Blank		0.000	21:17
Control Test		0.081	21:17
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 08/09/2017 08:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: NOLEN

DATE/TIME OF TEST: 07/23/2017 22:51:05

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.081	22:45
Air Blank		0.000	22:45
Subject Sample	# 1	0.082	22:46
Breath Volume		1.421	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.078	22:49
Breath Volume		2.269	22:49
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 08/09/2017 08:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006435

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: POWERS,BRITTANY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: SCHILLER

DATE/TIME OF TEST: 07/28/2017 18:52:48

Diagnostics Check		OK	18:46
Air Blank		0.000	18:46
Control Test		0.081	18:47
Air Blank		0.000	18:47
Subject Sample	# 1	0.199	18:48
Breath Volume		1.910	18:48
Air Blank		0.000	18:48
Air Blank		0.000	18:50
Subject Sample	# 2	0.183	18:51
Breath Volume		1.691	18:51
Air Blank		0.000	18:51
Control Test		0.080	18:52
Air Blank		0.000	18:52
Diagnostics Check		OK	18:52

Data Download Date/Time: 08/09/2017 08:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CENAT,F,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CANGE-TALTON,N
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:16

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 07/30/2017 07:42:53

Diagnostics Check		OK	07:40	
Air Blank		0.000	07:40	
Control Test		0.080	07:40	
Air Blank		0.000	07:41	
Subject Sample	# 1		07:41	Subject Test Refused
Breath Volume		0.000	07:41	
Air Blank		0.000	07:41	
Control Test		0.080	07:42	
Air Blank		0.000	07:42	
Diagnostics Check		OK	07:42	

Data Download Date/Time: 09/05/2017 10:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: SORONDO

DATE/TIME OF TEST: 07/31/2017 03:23:01

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:21	Subject Test Refused
Breath Volume		0.000	03:21	
Air Blank		0.000	03:22	
Control Test		0.080	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 09/05/2017 10:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLORES,EMILIO,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PEREZ,CRISTINO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: ROMERO

DATE/TIME OF TEST: 07/31/2017 21:18:32

Diagnostics Check		OK	21:15	
Air Blank		0.000	21:15	
Control Test		0.079	21:16	
Air Blank		0.000	21:16	
Subject Sample	# 1		21:17	Subject Test Refused
Breath Volume		0.000	21:17	
Air Blank		0.000	21:17	
Control Test		0.079	21:17	
Air Blank		0.000	21:18	
Diagnostics Check		OK	21:18	

Data Download Date/Time: 09/05/2017 10:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,CAROLINA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: CARRULLO

DATE/TIME OF TEST: 07/01/2017 01:08:59

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:04	
Control Test		0.080	01:04	
Air Blank		0.000	01:05	
Subject Sample	# 1		01:07	Subject Test Refused
Breath Volume		0.289	01:07	
Air Blank		0.000	01:08	
Control Test		0.079	01:08	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORVA,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: HONIG,ERIK
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: RODRIGUEZ SANCHEZ

DATE/TIME OF TEST: 07/01/2017 02:27:58

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.196	02:23
Breath Volume		1.390	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.197	02:26
Breath Volume		1.390	02:26
Air Blank		0.000	02:27
Control Test		0.077	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: FRANCISCO

DATE/TIME OF TEST: 07/02/2017 03:06:10

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.194	03:01
Breath Volume		1.734	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:03
Subject Sample	# 2	0.191	03:04
Breath Volume		1.730	03:04
Air Blank		0.000	03:05
Control Test		0.077	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,A,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BERNAL,F
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 07/02/2017 04:40:37

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.079	04:35
Air Blank		0.000	04:35
Subject Sample	# 1	0.150	04:36
Breath Volume		2.070	04:36
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.148	04:38
Breath Volume		2.003	04:38
Air Blank		0.000	04:39
Control Test		0.077	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,CAROLINA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ACOSTA,CAROLINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: APARICIO

DATE/TIME OF TEST: 07/02/2017 05:19:13

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:06	
Control Test		0.079	05:06	
Air Blank		0.000	05:06	
Subject Sample	# 1	0.173	05:10	Volume Not Met
Breath Volume		0.585	05:10	
Air Blank		0.000	05:10	
Air Blank		0.000	05:12	
Subject Sample	# 2	0.204	05:13	
Breath Volume		1.218	05:13	
Air Blank		0.000	05:13	
Air Blank		0.000	05:15	
Subject Sample	# 3	0.209	05:17	
Breath Volume		1.476	05:17	
Air Blank		0.000	05:18	
Control Test		0.079	05:18	
Air Blank		0.000	05:19	
Diagnostics Check		OK	05:19	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIEIGA,FRANK,
BREATH TEST OPERATOR AGENCY: North Bay Village PD
ARRESTING OFFICER: PIEIGA,FRANK
ARRESTING OFFICER AGENCY: North Bay Village PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ST. FURCY

DATE/TIME OF TEST: 07/03/2017 00:39:32

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:33
Air Blank		0.000	00:33
Subject Sample	# 1	0.135	00:34
Breath Volume		1.937	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.129	00:37
Breath Volume		2.328	00:37
Air Blank		0.000	00:38
Control Test		0.077	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,DIYANGO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: FERNANDEZ,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/03/2017 03:29:15

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:21
Subject Sample	# 1	0.197	03:22
Breath Volume		1.253	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.191	03:27
Breath Volume		1.886	03:27
Air Blank		0.000	03:28
Control Test		0.076	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,E,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MESA,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/03/2017 23:43:20

Diagnostics Check		OK	23:39	
Air Blank		0.000	23:40	
Control Test		0.080	23:40	
Air Blank		0.000	23:40	
Subject Sample	# 1		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.079	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,E,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MESA,E
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:29

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 07/04/2017 00:19:08

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.080	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.144	00:14
Breath Volume		1.277	00:14
Air Blank		0.000	00:14
Air Blank		0.000	00:16
Subject Sample	# 2	0.150	00:17
Breath Volume		1.796	00:17
Air Blank		0.000	00:18
Control Test		0.079	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORVA,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: SAAVEDRA

DATE/TIME OF TEST: 07/04/2017 03:41:40

Diagnostics Check		OK	03:38	
Air Blank		0.000	03:39	
Control Test		0.079	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 1		03:40	Subject Test Refused
Breath Volume		0.000	03:40	
Air Blank		0.000	03:40	
Control Test		0.079	03:41	
Air Blank		0.000	03:41	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAZQUEZ,JONATHAN,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VAZQUEZ,JONATHAN
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/06/2017 03:09:32

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.080	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.032	03:05
Breath Volume		3.445	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.030	03:08
Breath Volume		2.621	03:08
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,VANESSA,G
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUTIERREZ,JESSICA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 07/06/2017 18:10:33

Diagnostics Check		OK	18:04
Air Blank		0.000	18:04
Control Test		0.079	18:05
Air Blank		0.000	18:05
Subject Sample	# 1	0.214	18:06
Breath Volume		4.218	18:06
Air Blank		0.000	18:06
Air Blank		0.000	18:08
Subject Sample	# 2	0.215	18:09
Breath Volume		4.613	18:09
Air Blank		0.000	18:09
Control Test		0.079	18:09
Air Blank		0.000	18:10
Diagnostics Check		OK	18:10

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WASHINGTON,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 07/07/2017 05:37:13

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.080	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.133	05:32
Breath Volume		2.597	05:32
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.127	05:35
Breath Volume		2.722	05:35
Air Blank		0.000	05:36
Control Test		0.079	05:36
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,KATHERINE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUEVARA,CHARLIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/07/2017 17:53:24

Diagnostics Check		OK	17:50	
Air Blank		0.000	17:51	
Control Test		0.077	17:51	
Air Blank		0.000	17:51	
Subject Sample	# 1		17:52	Subject Test Refused
Breath Volume		0.000	17:52	
Air Blank		0.000	17:52	
Control Test		0.080	17:52	
Air Blank		0.000	17:53	
Diagnostics Check		OK	17:53	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUEVARA,CHARLIE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GUEVARA,CHARLIE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 07/07/2017 17:59:51

Diagnostics Check		OK	17:57	
Air Blank		0.000	17:57	
Control Test		0.079	17:57	
Air Blank		0.000	17:58	
Subject Sample	# 1		17:58	Subject Test Refused
Breath Volume		0.000	17:58	
Air Blank		0.000	17:58	
Control Test		0.080	17:59	
Air Blank		0.000	17:59	
Diagnostics Check		OK	17:59	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASIANO,GIANCARLO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CASIANO,GIANCARLO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 09:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:10

SUBJECT'S LAST NAME: RODRIGUEZ HERNANDEZ

DATE/TIME OF TEST: 07/08/2017 10:38:20

Diagnostics Check		OK	10:31
Air Blank		0.000	10:32
Control Test		0.080	10:32
Air Blank		0.000	10:33
Subject Sample	# 1	0.248	10:33
Breath Volume		1.800	10:33
Air Blank		0.000	10:34
Air Blank		0.000	10:36
Subject Sample	# 2	0.236	10:36
Breath Volume		2.042	10:36
Air Blank		0.000	10:37
Control Test		0.078	10:37
Air Blank		0.000	10:38
Diagnostics Check		OK	10:38

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,ARMANDO,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VALDES,ARMANDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SLAVCHEV

DATE/TIME OF TEST: 07/09/2017 02:41:09

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:40	
Control Test		0.080	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/09/2017 03:06:11

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.079	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,M,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LOPEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 07/09/2017 03:22:25

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.224	03:14
Breath Volume		2.195	03:14
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.246	03:17
Breath Volume		1.757	03:17
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 3	0.241	03:20
Breath Volume		2.406	03:20
Air Blank		0.000	03:21
Control Test		0.078	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,MARCOS,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MARRERO,MARCOS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/10/2017 00:09:03

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.080	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.209	00:04
Breath Volume		1.832	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.209	00:07
Breath Volume		2.273	00:07
Air Blank		0.000	00:08
Control Test		0.078	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:09

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,ARMANDO,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VALDES,ARMANDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/10/2017 02:33:09

Diagnostics Check	OK	02:32	
Air Blank		02:32	RFI Detect
Air Blank	0.000	02:33	

Data Download Date/Time: 07/27/2017 14:13

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,ARMANDO,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VALDES,ARMANDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 07/10/2017 02:49:20

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:46	
Control Test		0.080	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Control Test		0.080	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,CAROLINA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ACOSTA,CAROLINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/13/2017 01:49:25

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.079	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.079	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,CRISTINO,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: PEREZ,CRISTINO
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: LEGANOA PENAS

DATE/TIME OF TEST: 07/14/2017 02:19:58

Diagnostics Check		OK	02:13
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.182	02:15
Breath Volume		2.457	02:15
Air Blank		0.000	02:15
Air Blank		0.000	02:17
Subject Sample	# 2	0.164	02:18
Breath Volume		2.085	02:18
Air Blank		0.000	02:19
Control Test		0.077	02:19
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,M,S
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LOPEZ,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: JUAREZ REYES

DATE/TIME OF TEST: 07/15/2017 01:53:41

Diagnostics Check		OK	01:45
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.207	01:48
Breath Volume		2.406	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:51
Subject Sample	# 2	0.192	01:51
Breath Volume		1.988	01:51
Air Blank		0.000	01:52
Control Test		0.077	01:53
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMSON,ROBERT,H
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: THOMSON,ROBERT
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 13:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:34

SUBJECT'S LAST NAME: MANZENBERGER

DATE/TIME OF TEST: 07/15/2017 14:32:44

Diagnostics Check		OK	14:25
Air Blank		0.000	14:26
Control Test		0.079	14:26
Air Blank		0.000	14:27
Subject Sample	# 1	0.146	14:27
Breath Volume		2.410	14:27
Air Blank		0.000	14:28
Air Blank		0.000	14:30
Subject Sample	# 2	0.140	14:31
Breath Volume		2.625	14:31
Air Blank		0.000	14:31
Control Test		0.078	14:32
Air Blank		0.000	14:32
Diagnostics Check		OK	14:32

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MOISES,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: CAMACHO,CHRISTOPHER
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GUILLEN

DATE/TIME OF TEST: 07/16/2017 02:45:47

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.233	02:41
Breath Volume		2.199	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.231	02:44
Breath Volume		1.863	02:44
Air Blank		0.000	02:44
Control Test		0.079	02:45
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,M,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WASHINGTON,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: POYASTRO

DATE/TIME OF TEST: 07/17/2017 03:45:01

Diagnostics Check		OK	03:38
Air Blank		0.000	03:39
Control Test		0.080	03:39
Air Blank		0.000	03:40
Subject Sample	# 1	0.000	03:40
Breath Volume		4.062	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 2	0.000	03:43
Breath Volume		4.105	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:44
Diagnostics Check		OK	03:45

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JASIEL,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: SOMOZA

DATE/TIME OF TEST: 07/19/2017 04:27:11

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:24	
Control Test		0.079	04:25	
Air Blank		0.000	04:25	
Subject Sample	# 1		04:25	Subject Test Refused
Breath Volume		0.000	04:25	
Air Blank		0.000	04:26	
Control Test		0.079	04:26	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACOSTA,CAROLINA,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ACOSTA,CAROLINA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MILES

DATE/TIME OF TEST: 07/20/2017 02:06:37

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.080	02:00
Air Blank		0.000	02:00
Subject Sample	# 1	0.192	02:02
Breath Volume		2.347	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.189	02:05
Breath Volume		2.007	02:05
Air Blank		0.000	02:05
Control Test		0.080	02:06
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: WASHINGTON,R
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: VAN HOEF

DATE/TIME OF TEST: 07/21/2017 04:18:43

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:13	
Control Test		0.079	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.079	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHAMONDE,JORDANY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BAHAMONDE,JORDANY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: ROJAS

DATE/TIME OF TEST: 07/21/2017 20:36:18

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.079	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.165	20:31
Breath Volume		2.015	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.163	20:34
Breath Volume		1.566	20:34
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMIREZ,MOISES,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LARA,RENE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: VIDAL SANCHEZ

DATE/TIME OF TEST: 07/22/2017 03:10:49

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.190	03:06
Breath Volume		1.265	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.196	03:09
Breath Volume		2.718	03:09
Air Blank		0.000	03:09
Control Test		0.078	03:10
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAHAMONDE,JORDANY,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BROWN,ANDREA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 06/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: HERRING

DATE/TIME OF TEST: 07/23/2017 21:21:24

Diagnostics Check		OK	21:13
Air Blank		0.000	21:13
Control Test		0.080	21:14
Air Blank		0.000	21:14
Subject Sample	# 1	0.168	21:15
Breath Volume		2.132	21:15
Air Blank		0.000	21:16
Air Blank		0.000	21:18
Subject Sample	# 2	0.172	21:19
Breath Volume		2.585	21:19
Air Blank		0.000	21:20
Control Test		0.079	21:20
Air Blank		0.000	21:21
Diagnostics Check		OK	21:21

Data Download Date/Time: 07/27/2017 14:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORTA,RAFAEL,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: TARGIA,FRANK
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: WEISS

DATE/TIME OF TEST: 07/28/2017 00:17:02

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.080	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.000	00:12
Breath Volume		1.199	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.000	00:15
Breath Volume		1.476	00:15
Air Blank		0.000	00:16
Control Test		0.078	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:17

Data Download Date/Time: 09/05/2017 10:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVARRO,IZARY,C
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: AGUILERA,M
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: TORREZ

DATE/TIME OF TEST: 07/28/2017 04:16:56

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.079	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.199	04:12
Breath Volume		2.660	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.201	04:15
Breath Volume		1.476	04:15
Air Blank		0.000	04:16
Control Test		0.077	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 09/05/2017 10:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,ALEXIS,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: JACKSON,MARCEL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:05

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 07/29/2017 14:29:33

Diagnostics Check		OK	14:26	
Air Blank		0.000	14:26	
Control Test		0.078	14:27	
Air Blank		0.000	14:27	
Subject Sample	# 1		14:28	Subject Test Refused
Breath Volume		0.000	14:28	
Air Blank		0.000	14:28	
Control Test		0.078	14:28	
Air Blank		0.000	14:29	
Diagnostics Check		OK	14:29	

Data Download Date/Time: 09/05/2017 10:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,R,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ALVAREZ,J
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:48

SUBJECT'S LAST NAME: DELGADO

DATE/TIME OF TEST: 07/31/2017 00:35:38

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.079	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.179	00:30
Breath Volume		1.421	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:32
Subject Sample	# 2	0.182	00:34
Breath Volume		1.582	00:34
Air Blank		0.000	00:34
Control Test		0.077	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 09/05/2017 10:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILLEGAS,ELIO,S
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VILLEGAS,ELIO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: CASIPLE

DATE/TIME OF TEST: 07/31/2017 05:41:45

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.080	05:35
Air Blank		0.000	05:35
Subject Sample	# 1	0.192	05:36
Breath Volume		1.136	05:36
Air Blank		0.000	05:37
Air Blank		0.000	05:39
Subject Sample	# 2	0.199	05:40
Breath Volume		1.140	05:40
Air Blank		0.000	05:40
Control Test		0.079	05:41
Air Blank		0.000	05:41
Diagnostics Check		OK	05:41

Data Download Date/Time: 09/05/2017 10:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:38

SUBJECT'S LAST NAME: ZERQUERA RIVERO

DATE/TIME OF TEST: 07/06/2017 19:06:57

Diagnostics Check		OK	19:03	
Air Blank		0.000	19:04	
Control Test		0.077	19:04	
Air Blank		0.000	19:05	
Subject Sample	# 1		19:05	Subject Test Refused
Breath Volume		0.000	19:05	
Air Blank		0.000	19:06	
Control Test		0.077	19:06	
Air Blank		0.000	19:06	
Diagnostics Check		OK	19:06	

Data Download Date/Time: 08/15/2017 10:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: NELOMS JR

DATE/TIME OF TEST: 07/01/2017 01:34:41

Diagnostics Check		OK	01:22	
Air Blank		0.000	01:22	
Control Test		0.077	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 1	0.255	01:23	
Breath Volume		2.152	01:23	
Air Blank		0.000	01:24	
Air Blank		0.000	01:26	
Subject Sample	# 2	0.231	01:28	
Breath Volume		2.117	01:28	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 3	0.227	01:33	
Breath Volume		2.527	01:33	
Air Blank		0.000	01:33	
Control Test		0.074	01:34	Control Outside Tolerance
Air Blank		0.000	01:34	

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: NELOMS JR

DATE/TIME OF TEST: 07/01/2017 01:44:22

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.075	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.228	01:39
Breath Volume		2.179	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.225	01:42
Breath Volume		2.109	01:42
Air Blank		0.000	01:43
Control Test		0.075	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: TEICHER

DATE/TIME OF TEST: 07/01/2017 23:07:11

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.076	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.094	23:02
Breath Volume		1.554	23:02
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.092	23:05
Breath Volume		2.031	23:05
Air Blank		0.000	23:06
Control Test		0.075	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DEDEJ,MEDJON
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: DONAGHY

DATE/TIME OF TEST: 07/03/2017 02:26:34

Diagnostics Check		OK	02:19
Air Blank		0.000	02:19
Control Test		0.077	02:20
Air Blank		0.000	02:20
Subject Sample	# 1	0.248	02:21
Breath Volume		2.304	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.233	02:24
Breath Volume		1.679	02:24
Air Blank		0.000	02:25
Control Test		0.077	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CORRO

DATE/TIME OF TEST: 07/03/2017 23:29:12

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.077	23:23
Air Blank		0.000	23:24
Subject Sample	# 1	0.000	23:24
Breath Volume		2.730	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.000	23:27
Breath Volume		2.128	23:27
Air Blank		0.000	23:28
Control Test		0.078	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: MORISSET

DATE/TIME OF TEST: 07/05/2017 03:22:08

Diagnostics Check		OK	03:15
Air Blank		0.000	03:15
Control Test		0.076	03:16
Air Blank		0.000	03:16
Subject Sample	# 1	0.141	03:17
Breath Volume		1.484	03:17
Air Blank		0.000	03:18
Air Blank		0.000	03:19
Subject Sample	# 2	0.134	03:20
Breath Volume		1.722	03:20
Air Blank		0.000	03:21
Control Test		0.077	03:21
Air Blank		0.000	03:22
Diagnostics Check		OK	03:22

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 07/06/2017 02:00:35

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.077	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.287	01:56
Breath Volume		1.183	01:56
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.307	01:58
Breath Volume		1.515	01:58
Air Blank		0.000	01:59
Control Test		0.075	01:59
Air Blank		0.000	02:00
Diagnostics Check		OK	02:00

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 07/07/2017 02:06:37

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.077	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.293	02:01
Breath Volume		2.171	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.278	02:04
Breath Volume		2.000	02:04
Air Blank		0.000	02:05
Control Test		0.076	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: DOUSE

DATE/TIME OF TEST: 07/09/2017 01:40:47

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:38	
Control Test		0.077	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:39	Subject Test Refused
Breath Volume		0.000	01:39	
Air Blank		0.000	01:39	
Control Test		0.077	01:40	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIMAIO,THOMAS
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:27

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 07/09/2017 10:30:30

Diagnostics Check		OK	10:23
Air Blank		0.000	10:23
Control Test		0.078	10:23
Air Blank		0.000	10:24
Subject Sample	# 1	0.139	10:25
Breath Volume		1.117	10:25
Air Blank		0.000	10:25
Air Blank		0.000	10:27
Subject Sample	# 2	0.144	10:28
Breath Volume		1.917	10:28
Air Blank		0.000	10:29
Control Test		0.078	10:29
Air Blank		0.000	10:30
Diagnostics Check		OK	10:30

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON,YVETTE,D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BEDFORD,KEVIN
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:14

SUBJECT'S LAST NAME: WORTHINGTON

DATE/TIME OF TEST: 07/12/2017 14:46:04

Diagnostics Check		OK	14:36
Air Blank		0.000	14:36
Control Test		0.076	14:37
Air Blank		0.000	14:37
Subject Sample	# 1	0.248	14:38
Breath Volume		3.269	14:38
Air Blank		0.000	14:39
Air Blank		0.000	14:40
Subject Sample	# 2	0.224	14:41
Breath Volume		2.230	14:41
Air Blank		0.000	14:42
Air Blank		0.000	14:43
Subject Sample	# 3	0.242	14:44
Breath Volume		1.621	14:44
Air Blank		0.000	14:45
Control Test		0.076	14:45
Air Blank		0.000	14:45
Diagnostics Check		OK	14:46

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: MARINO

DATE/TIME OF TEST: 07/13/2017 03:54:45

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.077	03:49
Air Blank		0.000	03:49
Subject Sample	# 1	0.169	03:50
Breath Volume		1.824	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.163	03:53
Breath Volume		1.285	03:53
Air Blank		0.000	03:53
Control Test		0.077	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:37

SUBJECT'S LAST NAME: SOSA

DATE/TIME OF TEST: 07/14/2017 07:03:49

Diagnostics Check		OK	06:57
Air Blank		0.000	06:58
Control Test		0.078	06:58
Air Blank		0.000	06:58
Subject Sample	# 1	0.174	06:59
Breath Volume		3.957	06:59
Air Blank		0.000	06:59
Air Blank		0.000	07:01
Subject Sample	# 2	0.177	07:02
Breath Volume		3.511	07:02
Air Blank		0.000	07:02
Control Test		0.077	07:03
Air Blank		0.000	07:03
Diagnostics Check		OK	07:03

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ARGENTI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 06/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 07/15/2017 03:00:44

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.077	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.165	02:56
Breath Volume		1.890	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.161	02:59
Breath Volume		1.707	02:59
Air Blank		0.000	02:59
Control Test		0.077	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:00

Data Download Date/Time: 07/21/2017 09:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: YOO

DATE/TIME OF TEST: 07/22/2017 02:37:03

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.077	02:26	
Air Blank		0.000	02:26	
Subject Sample	# 1	0.127	02:29	Volume Not Met
Breath Volume		0.152	02:29	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2	0.135	02:35	Volume Not Met
Breath Volume		0.031	02:35	
Air Blank		0.000	02:36	
Control Test		0.076	02:36	
Air Blank		0.000	02:36	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: YOO

DATE/TIME OF TEST: 07/22/2017 02:47:04

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.076	02:40
Air Blank		0.000	02:40
Subject Sample	# 1	0.138	02:41
Breath Volume		2.093	02:41
Air Blank		0.000	02:42
Air Blank		0.000	02:44
Subject Sample	# 2	0.131	02:45
Breath Volume		1.304	02:45
Air Blank		0.000	02:46
Control Test		0.077	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:47

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: BAZARD

DATE/TIME OF TEST: 07/22/2017 03:23:23

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.075	03:17
Air Blank		0.000	03:18
Subject Sample	# 1	0.061	03:18
Breath Volume		2.597	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.061	03:21
Breath Volume		3.132	03:21
Air Blank		0.000	03:22
Control Test		0.076	03:22
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 07/22/2017 05:45:32

Diagnostics Check		OK	05:40	
Air Blank		0.000	05:40	
Control Test		0.078	05:41	
Air Blank		0.000	05:41	
Subject Sample	# 1		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:44	
Control Test		0.078	05:44	
Air Blank		0.000	05:45	
Diagnostics Check		OK	05:45	

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARCIA,JUSTINE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: PERAGINE

DATE/TIME OF TEST: 07/23/2017 03:29:13

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.077	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.194	03:24
Breath Volume		2.777	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.203	03:27
Breath Volume		2.496	03:27
Air Blank		0.000	03:28
Control Test		0.077	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MEJIA-DIAZ,GIANCARLO
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 07/23/2017 05:23:09

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.077	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1	0.152	05:15	Volume Not Met
Breath Volume		1.097	05:15	
Air Blank		0.000	05:15	
Air Blank		0.000	05:17	
Subject Sample	# 2	0.172	05:18	
Breath Volume		1.191	05:18	
Air Blank		0.000	05:18	
Air Blank		0.000	05:20	
Subject Sample	# 3	0.172	05:21	
Breath Volume		1.863	05:21	
Air Blank		0.000	05:22	
Control Test		0.077	05:22	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ESCOBAR,DIEGO
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SOTELO GOMEZ

DATE/TIME OF TEST: 07/24/2017 04:22:11

Diagnostics Check		OK	04:11
Air Blank		0.000	04:11
Control Test		0.077	04:12
Air Blank		0.000	04:12
Subject Sample	# 1	0.176	04:15
Breath Volume		1.156	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.166	04:20
Breath Volume		1.273	04:20
Air Blank		0.000	04:21
Control Test		0.077	04:21
Air Blank		0.000	04:22
Diagnostics Check		OK	04:22

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PRECIADO

DATE/TIME OF TEST: 07/27/2017 02:31:46

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.077	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.203	02:27
Breath Volume		2.824	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.200	02:30
Breath Volume		2.257	02:30
Air Blank		0.000	02:30
Control Test		0.077	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: IDARRAGA,CARLOS
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: RYGIEL

DATE/TIME OF TEST: 07/28/2017 23:04:24

Diagnostics Check		OK	22:57
Air Blank		0.000	22:58
Control Test		0.077	22:58
Air Blank		0.000	22:59
Subject Sample	# 1	0.151	22:59
Breath Volume		2.214	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.155	23:02
Breath Volume		2.539	23:02
Air Blank		0.000	23:03
Control Test		0.077	23:03
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: VADEBONCOEUR

DATE/TIME OF TEST: 07/28/2017 23:29:32

Diagnostics Check		OK	23:23
Air Blank		0.000	23:23
Control Test		0.077	23:24
Air Blank		0.000	23:24
Subject Sample	# 1	0.019	23:25
Breath Volume		2.460	23:25
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.017	23:28
Breath Volume		1.753	23:28
Air Blank		0.000	23:28
Control Test		0.077	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GARCIA,HEBERTO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: MARTINEZ XOLO

DATE/TIME OF TEST: 07/30/2017 11:40:13

Diagnostics Check		OK	11:33
Air Blank		0.000	11:33
Control Test		0.078	11:34
Air Blank		0.000	11:34
Subject Sample	# 1	0.139	11:35
Breath Volume		1.878	11:35
Air Blank		0.000	11:36
Air Blank		0.000	11:37
Subject Sample	# 2	0.130	11:38
Breath Volume		1.460	11:38
Air Blank		0.000	11:39
Control Test		0.077	11:39
Air Blank		0.000	11:40
Diagnostics Check		OK	11:40

Data Download Date/Time: 08/28/2017 15:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/22/2017 01:02:18

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.082	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.187	00:57
Breath Volume		2.246	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.200	01:00
Breath Volume		2.148	01:00
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 08/04/2017 12:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: VITALINO

DATE/TIME OF TEST: 07/22/2017 02:33:19

Diagnostics Check		OK	02:29	
Air Blank		0.000	02:29	
Control Test		0.081	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 1		02:31	Subject Test Refused
Breath Volume		0.000	02:31	
Air Blank		0.000	02:32	
Control Test		0.080	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 07/22/2017 23:38:37

Diagnostics Check		OK	23:31
Air Blank		0.000	23:32
Control Test		0.081	23:32
Air Blank		0.000	23:33
Subject Sample	# 1	0.133	23:33
Breath Volume		3.210	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.133	23:36
Breath Volume		3.054	23:36
Air Blank		0.000	23:37
Control Test		0.080	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HANDZIK,BIRGIT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: PENA PAZ

DATE/TIME OF TEST: 07/23/2017 00:57:48

Diagnostics Check		OK	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.148	00:52
Breath Volume		1.285	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.137	00:55
Breath Volume		1.332	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHAEFER,JOHN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: CHAPPE VALLADARES

DATE/TIME OF TEST: 07/23/2017 03:54:31

Diagnostics Check		OK	03:47
Air Blank		0.000	03:47
Control Test		0.080	03:47
Air Blank		0.000	03:48
Subject Sample	# 1	0.139	03:49
Breath Volume		1.398	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:51
Subject Sample	# 2	0.144	03:52
Breath Volume		3.089	03:52
Air Blank		0.000	03:53
Control Test		0.080	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FERNANDEZ,RYAN
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: RAMOS CORRALES

DATE/TIME OF TEST: 07/23/2017 04:38:27

Diagnostics Check		OK	04:30
Air Blank		0.000	04:30
Control Test		0.080	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.199	04:33
Breath Volume		2.070	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.196	04:36
Breath Volume		1.519	04:36
Air Blank		0.000	04:37
Control Test		0.076	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 08/04/2017 12:04

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 07/23/2017 05:26:14

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.079	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.169	05:21
Breath Volume		3.054	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.165	05:24
Breath Volume		2.343	05:24
Air Blank		0.000	05:25
Control Test		0.079	05:25
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONEY,MICHAEL
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KEY

DATE/TIME OF TEST: 07/24/2017 00:39:16

Diagnostics Check		OK	00:27	
Air Blank		0.000	00:27	
Control Test		0.081	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 1	0.139	00:31	Volume Not Met
Breath Volume		0.105	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 2	0.153	00:37	Volume Not Met
Breath Volume		0.164	00:37	
Air Blank		0.000	00:38	
Control Test		0.075	00:38	
Air Blank		0.000	00:39	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SAYIH,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MONTIEL

DATE/TIME OF TEST: 07/24/2017 05:28:33

Diagnostics Check		OK	05:21
Air Blank		0.000	05:22
Control Test		0.080	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.192	05:23
Breath Volume		1.898	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.183	05:26
Breath Volume		2.304	05:26
Air Blank		0.000	05:27
Control Test		0.079	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HAYASHI,TATSUAKI
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 07/27/2017 04:29:51

Diagnostics Check		OK	04:22
Air Blank		0.000	04:22
Control Test		0.081	04:22
Air Blank		0.000	04:23
Subject Sample	# 1	0.136	04:24
Breath Volume		1.660	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:26
Subject Sample	# 2	0.148	04:27
Breath Volume		2.835	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:29
Air Blank		0.000	04:29
Diagnostics Check		OK	04:29

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: SARACENO

DATE/TIME OF TEST: 07/28/2017 02:22:48

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.082	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.105	02:17
Breath Volume		2.269	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.112	02:20
Breath Volume		2.980	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LEVEY,SCOTT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 07/29/2017 00:07:47

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.083	00:02
Air Blank		0.000	00:02
Subject Sample	# 1	0.000	00:03
Breath Volume		2.667	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.000	00:06
Breath Volume		2.296	00:06
Air Blank		0.000	00:06
Control Test		0.081	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male

OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: VALLE MURILLO

DATE/TIME OF TEST: 07/29/2017 00:53:40

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.081	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.192	00:48
Breath Volume		1.531	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:50
Subject Sample	# 2	0.187	00:51
Breath Volume		1.371	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: SEAMAN

DATE/TIME OF TEST: 07/29/2017 01:37:03

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.080	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.173	01:32
Breath Volume		2.214	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:34
Subject Sample	# 2	0.169	01:35
Breath Volume		2.898	01:35
Air Blank		0.000	01:36
Control Test		0.079	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:37

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BATISTA,DANNY
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:24

SUBJECT'S LAST NAME: GILMER

DATE/TIME OF TEST: 07/29/2017 06:56:00

Diagnostics Check		OK	06:52	
Air Blank		0.000	06:52	
Control Test		0.081	06:53	
Air Blank		0.000	06:53	
Subject Sample	# 1		06:54	Subject Test Refused
Breath Volume		0.000	06:54	
Air Blank		0.000	06:55	
Control Test		0.080	06:55	
Air Blank		0.000	06:55	
Diagnostics Check		OK	06:56	

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 07/30/2017 02:30:29

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.081	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.124	02:25
Breath Volume		1.429	02:25
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.118	02:28
Breath Volume		2.648	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TAVARES,LUCAS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 05:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CALDERON HIDALGO

DATE/TIME OF TEST: 07/30/2017 06:26:08

Diagnostics Check		OK	06:15	
Air Blank		0.000	06:16	
Control Test		0.081	06:16	
Air Blank		0.000	06:16	
Subject Sample	# 1	0.177	06:20	Volume Not Met
Breath Volume		0.109	06:20	
Air Blank		0.000	06:21	
Air Blank		0.000	06:22	
Subject Sample	# 2		06:24	Subject Test Refused
Breath Volume		0.070	06:24	
Air Blank		0.000	06:25	
Control Test		0.076	06:25	
Air Blank		0.000	06:26	
Diagnostics Check		OK	06:26	

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006476

INSTRUMENT REGISTERED TO: PALM BEACH SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ROYAL

DATE/TIME OF TEST: 07/31/2017 04:52:28

Diagnostics Check		OK	04:44
Air Blank		0.000	04:45
Control Test		0.082	04:45
Air Blank		0.000	04:46
Subject Sample	# 1	0.213	04:46
Breath Volume		2.085	04:46
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.221	04:50
Breath Volume		1.109	04:50
Air Blank		0.000	04:51
Control Test		0.080	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 08/04/2017 12:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,BARBARA,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,BARBARA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VINCI

DATE/TIME OF TEST: 07/01/2017 04:11:55

Diagnostics Check		OK	04:04
Air Blank		0.000	04:04
Control Test		0.081	04:05
Air Blank		0.000	04:05
Subject Sample	# 1	0.164	04:07
Breath Volume		2.750	04:07
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.155	04:10
Breath Volume		1.636	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 07/24/2017 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY,JEFFREY,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: STANLEY,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:45

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 07/02/2017 14:21:15

Diagnostics Check		OK	14:15
Air Blank		0.000	14:15
Control Test		0.081	14:15
Air Blank		0.000	14:16
Subject Sample	# 1	0.000	14:16
Breath Volume		1.289	14:16
Air Blank		0.000	14:17
Air Blank		0.000	14:19
Subject Sample	# 2	0.000	14:19
Breath Volume		1.378	14:19
Air Blank		0.000	14:20
Control Test		0.081	14:20
Air Blank		0.000	14:21
Diagnostics Check		OK	14:21

Data Download Date/Time: 07/24/2017 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITZPATRICK,BRETT,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FITZPATRICK,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BANCROFT

DATE/TIME OF TEST: 07/03/2017 00:03:45

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:02	
Control Test		0.081	00:02	
Air Blank		0.000	00:03	
Subject Sample	# 1		00:03	Sequence Aborted
Breath Volume		0.000	00:03	
Air Blank		0.000	00:03	

Data Download Date/Time: 07/24/2017 10:43

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDGWAY,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: DALTON,JASON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 07/03/2017 19:37:54

Diagnostics Check		OK	19:31
Air Blank		0.000	19:31
Control Test		0.080	19:32
Air Blank		0.000	19:32
Subject Sample	# 1	0.127	19:33
Breath Volume		4.773	19:33
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.126	19:36
Breath Volume		4.480	19:36
Air Blank		0.000	19:36
Control Test		0.080	19:37
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 07/24/2017 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: RIVAS

DATE/TIME OF TEST: 07/04/2017 02:40:19

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.105	02:33
Breath Volume		1.906	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.110	02:38
Breath Volume		2.886	02:38
Air Blank		0.000	02:39
Control Test		0.078	02:39
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 07/24/2017 10:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIDGWAY,NICHOLAS,M
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: HERTEL,NICHOLAS
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: BUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: RICHELIEU

DATE/TIME OF TEST: 07/04/2017 19:48:22

Diagnostics Check		OK	19:40
Air Blank		0.000	19:40
Control Test		0.081	19:40
Air Blank		0.000	19:41
Subject Sample	# 1	0.150	19:43
Breath Volume		2.335	19:43
Air Blank		0.000	19:44
Air Blank		0.000	19:45
Subject Sample	# 2	0.147	19:46
Breath Volume		1.945	19:46
Air Blank		0.000	19:47
Control Test		0.079	19:47
Air Blank		0.000	19:48
Diagnostics Check		OK	19:48

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRUITT,NICHOLAS,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PRUITT,NICHOLAS
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 10:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:27

SUBJECT'S LAST NAME: HEGARTY

DATE/TIME OF TEST: 07/05/2017 12:45:18

Diagnostics Check		OK	12:42	
Air Blank		0.000	12:42	
Control Test		0.081	12:43	
Air Blank		0.000	12:43	
Subject Sample	# 1		12:43	Subject Test Refused
Breath Volume		0.000	12:43	
Air Blank		0.000	12:44	
Control Test		0.081	12:44	
Air Blank		0.000	12:45	
Diagnostics Check		OK	12:45	

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,KELLIE,C
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FERNANDEZ,KELLIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 14:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:13

SUBJECT'S LAST NAME: HAGA

DATE/TIME OF TEST: 07/06/2017 15:43:50

Diagnostics Check		OK	15:34
Air Blank		0.000	15:34
Control Test		0.081	15:35
Air Blank		0.000	15:35
Subject Sample	# 1	0.245	15:36
Breath Volume		3.132	15:36
Air Blank		0.000	15:37
Air Blank		0.000	15:38
Subject Sample	# 2	0.222	15:39
Breath Volume		1.968	15:39
Air Blank		0.000	15:40
Air Blank		0.000	15:41
Subject Sample	# 3	0.223	15:42
Breath Volume		1.953	15:42
Air Blank		0.000	15:42
Control Test		0.079	15:43
Air Blank		0.000	15:43
Diagnostics Check		OK	15:43

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 07/08/2017 04:13:45

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.081	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.217	04:08
Breath Volume		1.707	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.210	04:11
Breath Volume		2.160	04:11
Air Blank		0.000	04:12
Control Test		0.078	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 07/24/2017 10:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EHRHART,BARBARA,B
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: EHRHART,BARBARA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: FULLEN

DATE/TIME OF TEST: 07/08/2017 08:27:27

Diagnostics Check		OK	08:20
Air Blank		0.000	08:20
Control Test		0.080	08:21
Air Blank		0.000	08:21
Subject Sample	# 1	0.168	08:22
Breath Volume		3.191	08:22
Air Blank		0.000	08:23
Air Blank		0.000	08:24
Subject Sample	# 2	0.168	08:25
Breath Volume		1.687	08:25
Air Blank		0.000	08:26
Control Test		0.080	08:26
Air Blank		0.000	08:27
Diagnostics Check		OK	08:27

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITZPATRICK,BRETT,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FITZPATRICK,BRETT
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 07/09/2017 03:41:20

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1		03:35	No Sample Provided
Breath Volume		0.000	03:35	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2		03:40	No Sample Provided
Breath Volume		0.000	03:40	
Air Blank			03:40	Sequence Aborted
Air Blank		0.000	03:41	

Data Download Date/Time: 07/24/2017 10:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 07/09/2017 03:53:32

Diagnostics Check		OK	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.178	03:48
Breath Volume		3.117	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.177	03:51
Breath Volume		3.222	03:51
Air Blank		0.000	03:52
Control Test		0.078	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 07/24/2017 10:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: SANTAMARIA ZAVALA

DATE/TIME OF TEST: 07/10/2017 03:42:59

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.365	03:38
Breath Volume		1.589	03:38
Air Blank		0.000	03:39
Air Blank		0.000	03:40
Subject Sample	# 2	0.354	03:41
Breath Volume		1.554	03:41
Air Blank		0.000	03:42
Control Test		0.077	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:42

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY,JEFFREY,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: STANLEY,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: ROSE

DATE/TIME OF TEST: 07/12/2017 01:42:45

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.192	01:38
Breath Volume		2.085	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.185	01:41
Breath Volume		1.445	01:41
Air Blank		0.000	01:41
Control Test		0.080	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: RAPALO ORELLANA

DATE/TIME OF TEST: 07/13/2017 03:15:41

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:09
Subject Sample	# 1	0.176	03:10
Breath Volume		2.828	03:10
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.165	03:13
Breath Volume		1.417	03:13
Air Blank		0.000	03:14
Control Test		0.078	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: SOLOMON

DATE/TIME OF TEST: 07/14/2017 00:29:29

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.153	00:23
Breath Volume		1.578	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:26
Subject Sample	# 2	0.142	00:27
Breath Volume		1.363	00:27
Air Blank		0.000	00:28
Control Test		0.079	00:28
Air Blank		0.000	00:29
Diagnostics Check		OK	00:29

Data Download Date/Time: 07/24/2017 10:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: AMBROSE

DATE/TIME OF TEST: 07/15/2017 01:44:14

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.227	01:39
Breath Volume		1.871	01:39
Air Blank		0.000	01:40
Air Blank		0.000	01:41
Subject Sample	# 2	0.242	01:42
Breath Volume		1.621	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Diagnostics Check		OK	01:44

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: RODRIGUEZ LABOY

DATE/TIME OF TEST: 07/16/2017 23:07:51

Diagnostics Check		OK	22:59
Air Blank		0.000	22:59
Control Test		0.081	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.220	23:03
Breath Volume		1.625	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:05
Subject Sample	# 2	0.222	23:06
Breath Volume		1.871	23:06
Air Blank		0.000	23:06
Control Test		0.078	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:56

SUBJECT'S LAST NAME: MARIA

DATE/TIME OF TEST: 07/16/2017 23:24:56

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:18
Subject Sample	# 1	0.146	23:19
Breath Volume		2.417	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.150	23:23
Breath Volume		1.378	23:23
Air Blank		0.000	23:23
Control Test		0.079	23:24
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,AARON,R
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: BOWEN,AARON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: KUEBLER

DATE/TIME OF TEST: 07/21/2017 01:45:46

Diagnostics Check		OK	01:39
Air Blank		0.000	01:39
Control Test		0.081	01:39
Air Blank		0.000	01:40
Subject Sample	# 1	0.111	01:41
Breath Volume		2.644	01:41
Air Blank		0.000	01:41
Air Blank		0.000	01:43
Subject Sample	# 2	0.104	01:44
Breath Volume		1.441	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:45
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,JOSEPH,G
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: HIND,TIM
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BRENNAN

DATE/TIME OF TEST: 07/21/2017 01:58:25

Diagnostics Check		OK	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:56	
Air Blank		0.000	01:56	
Subject Sample	# 1		01:57	Subject Test Refused
Breath Volume		0.000	01:57	
Air Blank		0.000	01:57	
Control Test		0.081	01:57	
Air Blank		0.000	01:58	
Diagnostics Check		OK	01:58	

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: DOOM

DATE/TIME OF TEST: 07/22/2017 06:44:05

Diagnostics Check		OK	06:36
Air Blank		0.000	06:37
Control Test		0.081	06:37
Air Blank		0.000	06:38
Subject Sample	# 1	0.148	06:39
Breath Volume		1.593	06:39
Air Blank		0.000	06:40
Air Blank		0.000	06:41
Subject Sample	# 2	0.161	06:42
Breath Volume		1.660	06:42
Air Blank		0.000	06:43
Control Test		0.078	06:43
Air Blank		0.000	06:44
Diagnostics Check		OK	06:44

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: MERAZ

DATE/TIME OF TEST: 07/23/2017 04:29:00

Diagnostics Check		OK	04:20
Air Blank		0.000	04:20
Control Test		0.081	04:21
Air Blank		0.000	04:21
Subject Sample	# 1	0.094	04:22
Breath Volume		1.332	04:22
Air Blank		0.000	04:23
Air Blank		0.000	04:24
Subject Sample	# 2	0.101	04:27
Breath Volume		1.175	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:29

Data Download Date/Time: 07/24/2017 10:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SOLLEY,KATRYNA
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 07/24/2017 22:42:06

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:38	
Control Test		0.080	22:39	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:40	Subject Test Refused
Breath Volume		0.000	22:40	
Air Blank		0.000	22:41	
Control Test		0.080	22:41	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 08/21/2017 11:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,STEFANIE,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PETERSON,STEFANIE
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: HENDRICKS

DATE/TIME OF TEST: 07/25/2017 05:48:09

Diagnostics Check		OK	05:40
Air Blank		0.000	05:40
Control Test		0.081	05:41
Air Blank		0.000	05:41
Subject Sample	# 1	0.208	05:42
Breath Volume		2.207	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.189	05:46
Breath Volume		1.902	05:46
Air Blank		0.000	05:47
Control Test		0.080	05:47
Air Blank		0.000	05:48
Diagnostics Check		OK	05:48

Data Download Date/Time: 08/21/2017 11:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: BALAS

DATE/TIME OF TEST: 07/25/2017 21:12:17

Diagnostics Check		OK	21:05
Air Blank		0.000	21:05
Control Test		0.081	21:06
Air Blank		0.000	21:06
Subject Sample	# 1	0.336	21:07
Breath Volume		2.117	21:07
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.353	21:10
Breath Volume		2.265	21:10
Air Blank		0.000	21:11
Control Test		0.079	21:11
Air Blank		0.000	21:12
Diagnostics Check		OK	21:12

Data Download Date/Time: 08/21/2017 11:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006565

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY,JEFFREY,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: STANLEY,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: AUSTIN

DATE/TIME OF TEST: 07/28/2017 23:34:10

Diagnostics Check		OK	23:27
Air Blank		0.000	23:27
Control Test		0.081	23:27
Air Blank		0.000	23:28
Subject Sample	# 1	0.000	23:29
Breath Volume		1.972	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.000	23:32
Breath Volume		1.566	23:32
Air Blank		0.000	23:33
Control Test		0.080	23:33
Air Blank		0.000	23:34
Diagnostics Check		OK	23:34

Data Download Date/Time: 08/21/2017 11:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: SCHAFFNER

DATE/TIME OF TEST: 07/01/2017 02:32:50

Diagnostics Check		OK	02:25
Air Blank		0.000	02:26
Control Test		0.081	02:26
Air Blank		0.000	02:27
Subject Sample	# 1	0.156	02:27
Breath Volume		2.031	02:27
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.151	02:31
Breath Volume		1.171	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 07/01/2017 05:30:18

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:22
Subject Sample	# 1	0.176	05:24
Breath Volume		1.414	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:27
Subject Sample	# 2	0.172	05:28
Breath Volume		1.175	05:28
Air Blank		0.000	05:29
Control Test		0.078	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: RODRIGUEZ RAMIREZ

DATE/TIME OF TEST: 07/03/2017 01:30:20

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.081	01:24
Air Blank		0.000	01:24
Subject Sample	# 1	0.200	01:25
Breath Volume		3.019	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:28
Subject Sample	# 2	0.205	01:28
Breath Volume		2.972	01:28
Air Blank		0.000	01:29
Control Test		0.080	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: ZULEWSKI

DATE/TIME OF TEST: 07/03/2017 02:02:53

Diagnostics Check		OK	01:56
Air Blank		0.000	01:57
Control Test		0.081	01:57
Air Blank		0.000	01:57
Subject Sample	# 1	0.183	01:58
Breath Volume		1.734	01:58
Air Blank		0.000	01:59
Air Blank		0.000	02:00
Subject Sample	# 2	0.184	02:01
Breath Volume		1.718	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:02
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GOULET

DATE/TIME OF TEST: 07/05/2017 00:37:54

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.200	00:33
Breath Volume		2.617	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:35
Subject Sample	# 2	0.199	00:36
Breath Volume		2.402	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:37
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: TEDESCO

DATE/TIME OF TEST: 07/05/2017 03:40:07

Diagnostics Check		OK	03:37	
Air Blank		0.000	03:37	
Control Test		0.081	03:38	
Air Blank		0.000	03:38	
Subject Sample	# 1		03:38	Subject Test Refused
Breath Volume		0.000	03:38	
Air Blank		0.000	03:39	
Control Test		0.080	03:39	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:40	

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: MASTRILLI

DATE/TIME OF TEST: 07/07/2017 04:14:23

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.098	04:09
Breath Volume		1.972	04:09
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.096	04:12
Breath Volume		1.683	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRK,CASEY,S
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 07/08/2017 02:47:17

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.000	02:42
Breath Volume		3.222	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:45
Breath Volume		3.222	02:45
Air Blank		0.000	02:46
Control Test		0.080	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 07/08/2017 03:13:27

Diagnostics Check		OK	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.140	03:07
Breath Volume		1.281	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.144	03:11
Breath Volume		1.523	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 07/08/2017 05:29:07

Diagnostics Check		OK	05:22
Air Blank		0.000	05:22
Control Test		0.080	05:23
Air Blank		0.000	05:23
Subject Sample	# 1	0.163	05:24
Breath Volume		1.902	05:24
Air Blank		0.000	05:25
Air Blank		0.000	05:26
Subject Sample	# 2	0.161	05:27
Breath Volume		2.632	05:27
Air Blank		0.000	05:28
Control Test		0.080	05:28
Air Blank		0.000	05:29
Diagnostics Check		OK	05:29

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRIMBLE,TERESA,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ST JUSTE

DATE/TIME OF TEST: 07/09/2017 02:19:14

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.081	02:07	
Air Blank		0.000	02:07	
Subject Sample	# 1	0.190	02:08	
Breath Volume		1.371	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.172	02:14	Volume Not Met
Breath Volume		0.128	02:14	
Air Blank		0.000	02:15	
Air Blank		0.000	02:16	
Subject Sample	# 3	0.181	02:17	
Breath Volume		1.757	02:17	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 06/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 07/09/2017 04:08:05

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.080	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.113	04:03
Breath Volume		2.914	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.110	04:06
Breath Volume		3.128	04:06
Air Blank		0.000	04:07
Control Test		0.080	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 07/11/2017 19:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RAMRAJ

DATE/TIME OF TEST: 07/02/2017 04:04:00

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:53	
Control Test		0.078	03:53	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:57	No Sample Provided
Breath Volume		0.726	03:57	
Air Blank		0.000	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 2		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:04	

Data Download Date/Time: 07/21/2017 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: BANYAS

DATE/TIME OF TEST: 07/04/2017 03:51:25

Diagnostics Check		OK	03:48	
Air Blank		0.000	03:49	
Control Test		0.077	03:49	
Air Blank		0.000	03:49	
Subject Sample	# 1		03:50	Subject Test Refused
Breath Volume		0.000	03:50	
Air Blank		0.000	03:50	
Control Test		0.079	03:50	
Air Blank		0.000	03:51	
Diagnostics Check		OK	03:51	

Data Download Date/Time: 07/21/2017 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: COON,REBECCA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HASBUN

DATE/TIME OF TEST: 07/08/2017 03:28:07

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.078	03:23	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:26	Slope Not Met
Breath Volume		0.605	03:26	
Air Blank		0.000	03:27	
Control Test		0.080	03:27	
Air Blank		0.000	03:28	
Diagnostics Check		OK	03:28	

Data Download Date/Time: 07/21/2017 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: COON,REBECCA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HASBUN

DATE/TIME OF TEST: 07/08/2017 03:36:48

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.077	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.156	03:31
Breath Volume		1.480	03:31
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.155	03:35
Breath Volume		1.824	03:35
Air Blank		0.000	03:35
Control Test		0.078	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 07/21/2017 09:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DANIEL,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GRAHAM,D
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ENRIGHT

DATE/TIME OF TEST: 07/08/2017 04:57:57

Diagnostics Check		OK	04:55	
Air Blank		0.000	04:55	
Control Test		0.078	04:55	
Air Blank		0.000	04:56	
Subject Sample	# 1		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:57	
Control Test		0.079	04:57	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 07/21/2017 09:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: COON,REBECCA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 07/08/2017 19:46:33

Diagnostics Check		OK	19:43	
Air Blank		0.000	19:44	
Control Test		0.078	19:44	
Air Blank		0.000	19:44	
Subject Sample	# 1		19:45	Subject Test Refused
Breath Volume		0.000	19:45	
Air Blank		0.000	19:45	
Control Test		0.078	19:45	
Air Blank		0.000	19:46	
Diagnostics Check		OK	19:46	

Data Download Date/Time: 07/21/2017 09:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SCIRE

DATE/TIME OF TEST: 07/09/2017 23:08:55

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:06	
Control Test		0.078	23:06	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:07	Subject Test Refused
Breath Volume		0.000	23:07	
Air Blank		0.000	23:07	
Control Test		0.079	23:08	
Air Blank		0.000	23:08	
Diagnostics Check		OK	23:08	

Data Download Date/Time: 07/21/2017 09:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,BRITTANIE,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: LIMA,ALFREDO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: GORE

DATE/TIME OF TEST: 07/14/2017 02:08:50

Diagnostics Check		OK	02:02
Air Blank		0.000	02:02
Control Test		0.078	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.197	02:04
Breath Volume		2.613	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.195	02:07
Breath Volume		2.648	02:07
Air Blank		0.000	02:07
Control Test		0.078	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 07/21/2017 09:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: LIMA,ALFREDO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 06/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:54

SUBJECT'S LAST NAME: MACDONALD

DATE/TIME OF TEST: 07/15/2017 19:17:56

Diagnostics Check		OK	19:15	
Air Blank		0.000	19:15	
Control Test		0.077	19:15	
Air Blank		0.000	19:16	
Subject Sample	# 1		19:16	Subject Test Refused
Breath Volume		0.000	19:16	
Air Blank		0.000	19:17	
Control Test		0.078	19:17	
Air Blank		0.000	19:17	
Diagnostics Check		OK	19:17	

Data Download Date/Time: 07/21/2017 09:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WYNNE

DATE/TIME OF TEST: 07/22/2017 01:52:41

Diagnostics Check		OK	01:49	
Air Blank		0.000	01:50	
Control Test		0.078	01:50	
Air Blank		0.000	01:51	
Subject Sample	# 1		01:51	Subject Test Refused
Breath Volume		0.000	01:51	
Air Blank		0.000	01:51	
Control Test		0.079	01:52	
Air Blank		0.000	01:52	
Diagnostics Check		OK	01:52	

Data Download Date/Time: 08/22/2017 12:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: LIMA,ALFREDO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: ADAMS SMITH

DATE/TIME OF TEST: 07/23/2017 17:37:18

Diagnostics Check		OK	17:25	
Air Blank		0.000	17:25	
Control Test		0.078	17:26	
Air Blank		0.000	17:26	
Subject Sample	# 1	0.154	17:29	Volume Not Met
Breath Volume		0.019	17:29	
Air Blank		0.000	17:30	
Air Blank		0.000	17:32	
Subject Sample	# 2	0.136	17:35	Volume Not Met
Breath Volume		0.082	17:35	
Air Blank		0.000	17:36	
Control Test		0.078	17:36	
Air Blank		0.000	17:37	
Diagnostics Check		OK	17:37	

Data Download Date/Time: 08/22/2017 12:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: PERMUT

DATE/TIME OF TEST: 07/30/2017 00:11:40

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:09	
Control Test		0.078	00:09	
Air Blank		0.000	00:10	
Subject Sample	# 1		00:10	Subject Test Refused
Breath Volume		0.000	00:10	
Air Blank		0.000	00:10	
Control Test		0.079	00:11	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 08/22/2017 12:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRENZ,JONATHAN,R
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:02

SUBJECT'S LAST NAME: OBERG

DATE/TIME OF TEST: 07/30/2017 18:29:12

Diagnostics Check		OK	18:22
Air Blank		0.000	18:23
Control Test		0.079	18:23
Air Blank		0.000	18:24
Subject Sample	# 1	0.000	18:24
Breath Volume		3.148	18:24
Air Blank		0.000	18:25
Air Blank		0.000	18:27
Subject Sample	# 2	0.000	18:27
Breath Volume		3.257	18:27
Air Blank		0.000	18:28
Control Test		0.075	18:28
Air Blank		0.000	18:29
Diagnostics Check		OK	18:29

Data Download Date/Time: 08/22/2017 12:30

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 07/06/2017 01:56:48

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.079	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Control Test		0.078	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 07/07/2017 00:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTILLO,JOSUET,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: CASTILLO,JOSUET
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: OSCEOLA

DATE/TIME OF TEST: 07/25/2017 01:27:11

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.078	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.078	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 08/25/2017 08:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: CEDENO

DATE/TIME OF TEST: 07/29/2017 21:36:10

Diagnostics Check		OK	21:27
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.285	21:31
Breath Volume		2.144	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:33
Subject Sample	# 2	0.291	21:34
Breath Volume		1.332	21:34
Air Blank		0.000	21:35
Control Test		0.078	21:35
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 08/25/2017 08:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ALARCON

DATE/TIME OF TEST: 07/30/2017 04:05:52

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:02	
Control Test		0.079	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	
Control Test		0.078	04:05	
Air Blank		0.000	04:05	
Diagnostics Check		OK	04:05	

Data Download Date/Time: 08/25/2017 08:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006626

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGLYNN,JOHN,L
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MCGLYNN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: CARBONELL CALAS

DATE/TIME OF TEST: 07/31/2017 01:12:10

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.078	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.158	01:07
Breath Volume		2.101	01:07
Air Blank		0.000	01:07
Air Blank		0.000	01:09
Subject Sample	# 2	0.172	01:10
Breath Volume		2.148	01:10
Air Blank		0.000	01:11
Control Test		0.077	01:11
Air Blank		0.000	01:12
Diagnostics Check		OK	01:12

Data Download Date/Time: 08/25/2017 08:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: GEORGES

DATE/TIME OF TEST: 07/04/2017 02:41:19

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.078	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:40	
Control Test		0.078	02:40	
Air Blank		0.000	02:41	
Diagnostics Check		OK	02:41	

Data Download Date/Time: 07/07/2017 20:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: JIRON

DATE/TIME OF TEST: 07/04/2017 19:16:30

Diagnostics Check		OK	19:14	
Air Blank		0.000	19:14	
Control Test		0.080	19:15	
Air Blank		0.000	19:15	
Subject Sample	# 1		19:15	Subject Test Refused
Breath Volume		0.000	19:15	
Air Blank			19:16	RFI Detect
Air Blank		0.000	19:16	

Data Download Date/Time: 07/07/2017 20:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: JIRON

DATE/TIME OF TEST: 07/04/2017 19:20:52

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:18	
Control Test		0.079	19:18	
Air Blank		0.000	19:19	
Subject Sample	# 1		19:19	Subject Test Refused
Breath Volume		0.000	19:19	
Air Blank		0.000	19:19	
Control Test		0.079	19:20	
Air Blank		0.000	19:20	
Diagnostics Check		OK	19:20	

Data Download Date/Time: 07/07/2017 20:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 06/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: MONTOYA

DATE/TIME OF TEST: 07/06/2017 02:55:41

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.079	02:50
Air Blank		0.000	02:50
Subject Sample	# 1	0.130	02:51
Breath Volume		2.757	02:51
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.138	02:54
Breath Volume		2.843	02:54
Air Blank		0.000	02:54
Control Test		0.078	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 07/07/2017 20:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: BARTOLON

DATE/TIME OF TEST: 07/07/2017 23:48:04

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:46	Subject Test Refused
Breath Volume		0.000	23:46	
Air Blank		0.000	23:47	
Control Test		0.078	23:47	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ JR,RAUL,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ JR,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: PADILLA PUELLO

DATE/TIME OF TEST: 07/08/2017 20:46:37

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.079	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.238	20:41
Breath Volume		1.746	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.243	20:45
Breath Volume		2.121	20:45
Air Blank		0.000	20:45
Control Test		0.079	20:46
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 08/25/2017 08:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: IZNAGA

DATE/TIME OF TEST: 07/09/2017 02:29:01

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:26	
Control Test		0.080	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:27	Subject Test Refused
Breath Volume		0.000	02:27	
Air Blank		0.000	02:28	
Control Test		0.079	02:28	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: YORK

DATE/TIME OF TEST: 07/11/2017 02:09:33

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.079	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.133	02:05
Breath Volume		3.261	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.136	02:08
Breath Volume		3.558	02:08
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: HALTER

DATE/TIME OF TEST: 07/12/2017 03:17:30

Diagnostics Check		OK	03:14	
Air Blank		0.000	03:15	
Control Test		0.079	03:15	
Air Blank		0.000	03:15	
Subject Sample	# 1		03:16	Subject Test Refused
Breath Volume		0.000	03:16	
Air Blank		0.000	03:16	
Control Test		0.078	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 07/15/2017 03:10:22

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.079	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: TUMUS

DATE/TIME OF TEST: 07/17/2017 21:25:07

Diagnostics Check		OK	21:20	
Air Blank		0.000	21:20	
Control Test		0.080	21:20	
Air Blank		0.000	21:21	
Subject Sample	# 1	0.134	21:21	
Breath Volume		2.347	21:21	
Air Blank		0.000	21:22	
Air Blank		0.000	21:24	
Subject Sample	# 2	0.134	21:24	
Breath Volume		2.441	21:24	
Air Blank			21:24	RFI Detect
Air Blank			21:25	RFI Detect

Data Download Date/Time: 08/25/2017 08:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: TUMUS

DATE/TIME OF TEST: 07/17/2017 21:36:40

Diagnostics Check		OK	21:30
Air Blank		0.000	21:30
Control Test		0.079	21:31
Air Blank		0.000	21:31
Subject Sample	# 1	0.136	21:32
Breath Volume		3.417	21:32
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.128	21:35
Breath Volume		1.445	21:35
Air Blank		0.000	21:35
Control Test		0.078	21:36
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: LOPEZ JIMENEZ

DATE/TIME OF TEST: 07/19/2017 03:56:43

Diagnostics Check		OK	03:53	
Air Blank		0.000	03:54	
Control Test		0.080	03:54	
Air Blank		0.000	03:54	
Subject Sample	# 1		03:55	Subject Test Refused
Breath Volume		0.000	03:55	
Air Blank		0.000	03:55	
Control Test		0.078	03:56	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CALVO

DATE/TIME OF TEST: 07/19/2017 23:18:43

Diagnostics Check	OK	23:18	
Air Blank		23:18	RFI Detect
Air Blank	0.000	23:18	

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: CALVO

DATE/TIME OF TEST: 07/19/2017 23:27:08

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.132	23:22
Breath Volume		1.175	23:22
Air Blank		0.000	23:22
Air Blank		0.000	23:24
Subject Sample	# 2	0.124	23:25
Breath Volume		1.558	23:25
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: CASTILLO TRUJILLO

DATE/TIME OF TEST: 07/20/2017 02:20:58

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:18	
Control Test		0.079	02:18	
Air Blank		0.000	02:19	
Subject Sample	# 1		02:19	Subject Test Refused
Breath Volume		0.000	02:19	
Air Blank		0.000	02:20	
Control Test		0.078	02:20	
Air Blank		0.000	02:20	
Diagnostics Check		OK	02:20	

Data Download Date/Time: 08/25/2017 08:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: FROMETA

DATE/TIME OF TEST: 07/20/2017 21:49:32

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:47	
Control Test		0.080	21:47	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:48	Subject Test Refused
Breath Volume		0.000	21:48	
Air Blank		0.000	21:48	
Control Test		0.079	21:48	
Air Blank		0.000	21:49	
Diagnostics Check		OK	21:49	

Data Download Date/Time: 08/25/2017 08:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: VILCHEZ

DATE/TIME OF TEST: 07/22/2017 03:13:54

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.094	03:09
Breath Volume		3.351	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.094	03:12
Breath Volume		3.246	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 08/25/2017 08:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DELEON ANGEL

DATE/TIME OF TEST: 07/26/2017 03:07:09

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:57
Subject Sample	# 1	0.107	03:00
Breath Volume		1.539	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.116	03:05
Breath Volume		2.382	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 08/25/2017 08:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: FRASE

DATE/TIME OF TEST: 07/27/2017 01:34:55

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.079	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.080	01:34	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 08/25/2017 08:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: ALMON

DATE/TIME OF TEST: 07/27/2017 02:42:55

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.079	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.079	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 08/25/2017 08:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 07/29/2017 00:55:51

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.079	00:53	
Air Blank		0.000	00:54	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:54	
Control Test		0.079	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 08/25/2017 08:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: BAUTISTA HERNANDEZ

DATE/TIME OF TEST: 07/01/2017 02:07:33

Diagnostics Check		OK	01:59
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.204	02:02
Breath Volume		1.445	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.194	02:05
Breath Volume		2.398	02:05
Air Blank		0.000	02:06
Control Test		0.078	02:06
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: KNOECKEL

DATE/TIME OF TEST: 07/02/2017 01:07:18

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:01
Air Blank		0.000	01:02
Subject Sample	# 1	0.092	01:02
Breath Volume		2.503	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.091	01:05
Breath Volume		2.476	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 07/04/2017 19:43:28

Diagnostics Check		OK	19:36
Air Blank		0.000	19:37
Control Test		0.079	19:37
Air Blank		0.000	19:38
Subject Sample	# 1	0.180	19:38
Breath Volume		2.839	19:38
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.175	19:41
Breath Volume		3.582	19:41
Air Blank		0.000	19:42
Control Test		0.079	19:42
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/06/2017 02:48:16

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:45	
Control Test		0.079	02:45	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:46	Subject Test Refused
Breath Volume		0.000	02:46	
Air Blank		0.000	02:47	
Control Test		0.079	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: URUCHIMA,ANGEL,R
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: URUCHIMA,ANGEL
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: HATFIELD

DATE/TIME OF TEST: 07/15/2017 04:31:04

Diagnostics Check		OK	04:24
Air Blank		0.000	04:24
Control Test		0.078	04:25
Air Blank		0.000	04:25
Subject Sample	# 1	0.018	04:26
Breath Volume		3.875	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:28
Subject Sample	# 2	0.017	04:29
Breath Volume		3.503	04:29
Air Blank		0.000	04:30
Control Test		0.079	04:30
Air Blank		0.000	04:30
Diagnostics Check		OK	04:31

Data Download Date/Time: 07/24/2017 10:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ISOM

DATE/TIME OF TEST: 07/16/2017 01:28:31

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.080	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.079	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVAGLIO,JAMES,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SCHLAMBACH,BENJAMIN
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:25

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/16/2017 08:54:36

Diagnostics Check		OK	08:46
Air Blank		0.000	08:46
Control Test		0.079	08:47
Air Blank		0.000	08:47
Subject Sample	# 1	0.020	08:48
Breath Volume		1.390	08:48
Air Blank		0.000	08:49
Air Blank		0.000	08:51
Subject Sample	# 2	0.017	08:53
Breath Volume		1.667	08:53
Air Blank		0.000	08:53
Control Test		0.078	08:54
Air Blank		0.000	08:54
Diagnostics Check		OK	08:54

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: GUZMAN-PEREZ

DATE/TIME OF TEST: 07/19/2017 05:57:37

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.079	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.178	05:52
Breath Volume		1.593	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.180	05:55
Breath Volume		1.742	05:55
Air Blank		0.000	05:56
Control Test		0.080	05:57
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: RIOS

DATE/TIME OF TEST: 07/19/2017 21:01:11

Diagnostics Check		OK	20:58	
Air Blank		0.000	20:58	
Control Test		0.079	20:59	
Air Blank		0.000	20:59	
Subject Sample	# 1		20:59	Subject Test Refused
Breath Volume		0.000	20:59	
Air Blank		0.000	21:00	
Control Test		0.079	21:00	
Air Blank		0.000	21:01	
Diagnostics Check		OK	21:01	

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: POTTS

DATE/TIME OF TEST: 07/19/2017 22:59:06

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.079	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.000	22:54
Breath Volume		3.457	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.000	22:57
Breath Volume		2.378	22:57
Air Blank		0.000	22:58
Control Test		0.077	22:58
Air Blank		0.000	22:59
Diagnostics Check		OK	22:59

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERRERO,FRAICOR,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TERRERO,FRAICOR
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 07/23/2017 06:52:36

Diagnostics Check		OK	06:46
Air Blank		0.000	06:46
Control Test		0.079	06:46
Air Blank		0.000	06:47
Subject Sample	# 1	0.165	06:48
Breath Volume		3.296	06:48
Air Blank		0.000	06:48
Air Blank		0.000	06:50
Subject Sample	# 2	0.155	06:50
Breath Volume		3.996	06:50
Air Blank		0.000	06:51
Control Test		0.079	06:51
Air Blank		0.000	06:52
Diagnostics Check		OK	06:52

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: MALDONADO RUIZ

DATE/TIME OF TEST: 07/24/2017 01:08:09

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.080	01:00
Air Blank		0.000	01:01
Subject Sample	# 1	0.043	01:02
Breath Volume		1.347	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.043	01:06
Breath Volume		2.140	01:06
Air Blank		0.000	01:07
Control Test		0.078	01:07
Air Blank		0.000	01:08
Diagnostics Check		OK	01:08

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HUTCHISON

DATE/TIME OF TEST: 07/24/2017 01:20:30

Diagnostics Check		OK	01:10
Air Blank		0.000	01:11
Control Test		0.079	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.212	01:14
Breath Volume		1.339	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.222	01:18
Breath Volume		1.519	01:18
Air Blank		0.000	01:19
Control Test		0.076	01:19
Air Blank		0.000	01:20
Diagnostics Check		OK	01:20

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: MILLA DIAZ

DATE/TIME OF TEST: 07/24/2017 04:48:29

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.154	04:43
Breath Volume		1.644	04:43
Air Blank		0.000	04:44
Air Blank		0.000	04:46
Subject Sample	# 2	0.147	04:46
Breath Volume		1.218	04:46
Air Blank		0.000	04:47
Control Test		0.077	04:47
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 07/24/2017 10:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: BALAS

DATE/TIME OF TEST: 07/25/2017 20:59:35

Diagnostics Check		OK	20:50	
Air Blank		0.000	20:50	
Control Test		0.080	20:51	
Air Blank		0.000	20:51	
Subject Sample	# 1	0.343	20:52	
Breath Volume		1.753	20:52	
Air Blank		0.000	20:52	
Air Blank		0.000	20:54	
Subject Sample	# 2	0.372	20:55	
Breath Volume		2.238	20:55	
Air Blank		0.000	20:55	
Air Blank		0.000	20:57	
Subject Sample	# 3		20:57	Slope Not Met
Breath Volume		2.007	20:57	
Air Blank		0.000	20:58	
Control Test		0.078	20:58	
Air Blank		0.000	20:59	
Diagnostics Check		OK	20:59	

Data Download Date/Time: 08/21/2017 11:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SISK

DATE/TIME OF TEST: 07/26/2017 02:37:32

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.079	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.208	02:33
Breath Volume		1.453	02:33
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.213	02:35
Breath Volume		1.718	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 08/21/2017 11:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: OLAZO CORTES

DATE/TIME OF TEST: 07/27/2017 00:47:57

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.079	00:42
Air Blank		0.000	00:42
Subject Sample	# 1	0.193	00:43
Breath Volume		3.156	00:43
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.195	00:46
Breath Volume		2.250	00:46
Air Blank		0.000	00:47
Control Test		0.078	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 08/21/2017 11:23

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENYAN,M,A
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: KENYAN,MA
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: KOVALSKI

DATE/TIME OF TEST: 07/27/2017 21:33:34

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.078	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.250	21:29
Breath Volume		2.210	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.231	21:31
Breath Volume		1.906	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 08/21/2017 11:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 07/28/2017 23:43:48

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.206	23:39
Breath Volume		2.644	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.200	23:42
Breath Volume		2.621	23:42
Air Blank		0.000	23:42
Control Test		0.078	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 08/21/2017 11:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: 6C08,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 07/29/2017 03:22:18

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:17	
Control Test		0.080	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.523	03:20	
Air Blank		0.000	03:21	
Control Test		0.080	03:21	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 08/21/2017 11:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAKAI,MICHAEL,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MAKAI,MICHAEL
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SALAZAR

DATE/TIME OF TEST: 07/30/2017 00:49:25

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.108	00:44
Breath Volume		2.046	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.113	00:47
Breath Volume		2.781	00:47
Air Blank		0.000	00:48
Control Test		0.080	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 08/21/2017 11:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: TALBOT

DATE/TIME OF TEST: 07/05/2017 21:10:47

Diagnostics Check		OK	21:07	
Air Blank		0.000	21:08	
Control Test		0.079	21:08	
Air Blank		0.000	21:09	
Subject Sample	# 1		21:09	Subject Test Refused
Breath Volume		0.000	21:09	
Air Blank		0.000	21:09	
Control Test		0.077	21:10	
Air Blank		0.000	21:10	
Diagnostics Check		OK	21:10	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: MCCOOL

DATE/TIME OF TEST: 07/05/2017 21:27:54

Diagnostics Check	OK	21:26	
Air Blank	0.000	21:27	
Control Test	0.082	21:27	Interferent Detect
Air Blank	0.000	21:27	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: MCCOOL

DATE/TIME OF TEST: 07/05/2017 21:37:29

Diagnostics Check		OK	21:31
Air Blank		0.000	21:31
Control Test		0.079	21:31
Air Blank		0.000	21:32
Subject Sample	# 1	0.131	21:32
Breath Volume		2.574	21:32
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 2	0.125	21:35
Breath Volume		2.425	21:35
Air Blank		0.000	21:36
Control Test		0.075	21:36
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: TALBOT

DATE/TIME OF TEST: 07/05/2017 22:24:14

Diagnostics Check		OK	22:18
Air Blank		0.000	22:18
Control Test		0.080	22:18
Air Blank		0.000	22:19
Subject Sample	# 1	0.092	22:19
Breath Volume		3.019	22:19
Air Blank		0.000	22:20
Air Blank		0.000	22:22
Subject Sample	# 2	0.091	22:22
Breath Volume		2.890	22:22
Air Blank		0.000	22:23
Control Test		0.077	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 07/31/2017 18:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/06/2017 01:55:43

Diagnostics Check		OK	01:48	
Air Blank		0.000	01:48	
Control Test		0.079	01:49	
Air Blank		0.000	01:49	
Subject Sample	# 1	0.093	01:50	
Breath Volume		1.878	01:50	
Air Blank		0.000	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 2		01:54	Interferent Detect
Breath Volume		2.082	01:54	
Air Blank		0.000	01:54	
Control Test		0.076	01:55	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/06/2017 02:03:44

Diagnostics Check		OK	01:57	
Air Blank		0.000	01:57	
Control Test		0.079	01:57	
Air Blank		0.000	01:58	
Subject Sample	# 1	0.091	01:58	
Breath Volume		1.898	01:58	
Air Blank		0.000	01:59	
Air Blank		0.000	02:01	
Subject Sample	# 2		02:02	Interferent Detect
Breath Volume		2.050	02:02	
Air Blank		0.000	02:02	
Control Test		0.082	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/06/2017 02:09:01

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:07	Interferent Detect
Breath Volume		1.968	02:07	
Air Blank		0.000	02:08	
Control Test		0.080	02:08	
Air Blank		0.000	02:08	
Diagnostics Check		OK	02:09	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 07/06/2017 02:17:15

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.093	02:12
Breath Volume		2.089	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:14
Subject Sample	# 2	0.089	02:15
Breath Volume		2.031	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/07/2017 01:35:27

Diagnostics Check	OK	01:34	
Air Blank		01:34	Ambient Fail
Air Blank	0.000	01:35	

Data Download Date/Time: 07/31/2017 18:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 07/07/2017 01:40:20

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:39	
Control Test		0.080	01:39	
Air Blank		0.000	01:40	
Diagnostics Check		OK	01:40	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: NORDMAN

DATE/TIME OF TEST: 07/07/2017 01:46:00

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.077	01:45	Interferent Detect
Air Blank		0.000	01:46	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 07/07/2017 20:38:02

Diagnostics Check		OK	20:31	
Air Blank		0.000	20:32	
Control Test		0.078	20:32	
Air Blank		0.000	20:33	
Subject Sample	# 1	0.207	20:33	
Breath Volume		2.960	20:33	
Air Blank		0.000	20:34	
Air Blank		0.000	20:36	
Subject Sample	# 2	0.211	20:36	
Breath Volume		2.492	20:36	
Air Blank		0.000	20:37	
Control Test		0.074	20:37	Control Outside Tolerance
Air Blank		0.000	20:38	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: GLEASON-DOVEL

DATE/TIME OF TEST: 07/08/2017 01:40:10

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.078	01:34
Air Blank		0.000	01:34
Subject Sample	# 1	0.113	01:35
Breath Volume		1.878	01:35
Air Blank		0.000	01:36
Air Blank		0.000	01:37
Subject Sample	# 2	0.118	01:38
Breath Volume		2.335	01:38
Air Blank		0.000	01:39
Control Test		0.079	01:39
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 07/31/2017 18:51

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 07/08/2017 03:49:35

Diagnostics Check		OK	03:41
Air Blank		0.000	03:41
Control Test		0.080	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.140	03:43
Breath Volume		2.152	03:43
Air Blank		0.000	03:44
Air Blank		0.000	03:46
Subject Sample	# 2	0.131	03:47
Breath Volume		1.632	03:47
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: WHITTEN

DATE/TIME OF TEST: 07/09/2017 00:01:24

Diagnostics Check		OK	23:54
Air Blank		0.000	23:54
Control Test		0.079	23:55
Air Blank		0.000	23:55
Subject Sample	# 1	0.126	23:56
Breath Volume		3.851	23:56
Air Blank		0.000	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.115	23:59
Breath Volume		4.234	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/09/2017 03:58:28

Diagnostics Check	OK	03:57	
Air Blank	0.000	03:57	
Control Test	0.082	03:57	Interferent Detect
Air Blank	0.000	03:58	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/09/2017 04:03:49

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:00	
Control Test		0.079	04:00	
Air Blank		0.000	04:01	
Subject Sample	# 1		04:02	Interferent Detect
Breath Volume		3.539	04:02	
Air Blank		0.000	04:02	
Control Test		0.078	04:03	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 07/09/2017 04:12:16

Diagnostics Check		OK	04:05
Air Blank		0.000	04:05
Control Test		0.077	04:06
Air Blank		0.000	04:06
Subject Sample	# 1	0.081	04:07
Breath Volume		3.781	04:07
Air Blank		0.000	04:08
Air Blank		0.000	04:09
Subject Sample	# 2	0.080	04:10
Breath Volume		3.636	04:10
Air Blank		0.000	04:11
Control Test		0.078	04:11
Air Blank		0.000	04:12
Diagnostics Check		OK	04:12

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 07/12/2017 00:28:09

Diagnostics Check		OK	00:21
Air Blank		0.000	00:22
Control Test		0.080	00:22
Air Blank		0.000	00:22
Subject Sample	# 1	0.170	00:23
Breath Volume		1.679	00:23
Air Blank		0.000	00:24
Air Blank		0.000	00:25
Subject Sample	# 2	0.162	00:26
Breath Volume		1.390	00:26
Air Blank		0.000	00:27
Control Test		0.080	00:27
Air Blank		0.000	00:28
Diagnostics Check		OK	00:28

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PEREZ MORALES

DATE/TIME OF TEST: 07/13/2017 03:53:23

Diagnostics Check		OK	03:46
Air Blank		0.000	03:47
Control Test		0.078	03:47
Air Blank		0.000	03:47
Subject Sample	# 1	0.091	03:48
Breath Volume		2.742	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:51
Subject Sample	# 2	0.086	03:51
Breath Volume		1.808	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: GARIS

DATE/TIME OF TEST: 07/14/2017 04:16:34

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.079	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.218	04:11
Breath Volume		3.082	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.220	04:14
Breath Volume		2.515	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: WALSH

DATE/TIME OF TEST: 07/14/2017 19:07:58

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.079	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.224	19:03
Breath Volume		3.488	19:03
Air Blank		0.000	19:04
Air Blank		0.000	19:05
Subject Sample	# 2	0.230	19:06
Breath Volume		3.761	19:06
Air Blank		0.000	19:07
Control Test		0.080	19:07
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 07/22/2017 06:08:18

Diagnostics Check		OK	06:05	
Air Blank		0.000	06:05	
Control Test		0.079	06:06	
Air Blank		0.000	06:06	
Subject Sample	# 1		06:06	Subject Test Refused
Breath Volume		0.000	06:06	
Air Blank		0.000	06:07	
Control Test		0.079	06:07	
Air Blank		0.000	06:08	
Diagnostics Check		OK	06:08	

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 06/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:45

SUBJECT'S LAST NAME: GOULD

DATE/TIME OF TEST: 07/22/2017 15:17:58

Diagnostics Check		OK	15:11
Air Blank		0.000	15:12
Control Test		0.077	15:12
Air Blank		0.000	15:13
Subject Sample	# 1	0.000	15:13
Breath Volume		1.339	15:13
Air Blank		0.000	15:14
Air Blank		0.000	15:16
Subject Sample	# 2	0.000	15:16
Breath Volume		1.851	15:16
Air Blank		0.000	15:17
Control Test		0.076	15:17
Air Blank		0.000	15:17
Diagnostics Check		OK	15:17

Data Download Date/Time: 07/31/2017 18:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRASER,THOMAS,J
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: FRASER,THOMAS
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 06/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: CENTENO-CASIQUE

DATE/TIME OF TEST: 07/14/2017 19:18:49

Diagnostics Check		OK	19:12
Air Blank		0.000	19:12
Control Test		0.079	19:12
Air Blank		0.000	19:13
Subject Sample	# 1	0.249	19:14
Breath Volume		1.851	19:14
Air Blank		0.000	19:14
Air Blank		0.000	19:16
Subject Sample	# 2	0.231	19:17
Breath Volume		2.148	19:17
Air Blank		0.000	19:17
Control Test		0.079	19:18
Air Blank		0.000	19:18
Diagnostics Check		OK	19:18

Data Download Date/Time: 07/24/2017 09:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:38

SUBJECT'S LAST NAME: GERLACH

DATE/TIME OF TEST: 07/02/2017 06:05:23

Diagnostics Check		OK	05:59
Air Blank		0.000	05:59
Control Test		0.080	05:59
Air Blank		0.000	06:00
Subject Sample	# 1	0.240	06:00
Breath Volume		2.609	06:00
Air Blank		0.000	06:01
Air Blank		0.000	06:03
Subject Sample	# 2	0.242	06:03
Breath Volume		2.066	06:03
Air Blank		0.000	06:04
Control Test		0.080	06:04
Air Blank		0.000	06:05
Diagnostics Check		OK	06:05

Data Download Date/Time: 07/20/2017 22:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: BOOTHE

DATE/TIME OF TEST: 07/18/2017 00:04:13

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.082	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.271	23:59
Breath Volume		2.738	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.282	00:02
Breath Volume		3.070	00:02
Air Blank		0.000	00:03
Control Test		0.076	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 07/20/2017 22:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: ROEBER,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 07/02/2017 01:01:10

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.079	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.198	00:56
Breath Volume		1.511	00:56
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.200	00:59
Breath Volume		2.347	00:59
Air Blank		0.000	01:00
Control Test		0.078	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 07/28/2017 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 07/21/2017 04:57:52

Diagnostics Check		OK	04:51
Air Blank		0.000	04:52
Control Test		0.079	04:52
Air Blank		0.000	04:52
Subject Sample	# 1	0.000	04:53
Breath Volume		2.847	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.000	04:56
Breath Volume		2.820	04:56
Air Blank		0.000	04:56
Control Test		0.079	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 07/28/2017 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 06/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 07/27/2017 23:47:17

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.091	23:42
Breath Volume		2.597	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.086	23:45
Breath Volume		2.476	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 07/28/2017 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006768

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: ZUCK

DATE/TIME OF TEST: 07/29/2017 23:47:47

Diagnostics Check		OK	23:36	
Air Blank		0.000	23:36	
Control Test		0.081	23:36	
Air Blank		0.000	23:37	
Subject Sample	# 1	0.134	23:40	Volume Not Met
Breath Volume		0.300	23:40	
Air Blank		0.000	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 2	0.239	23:46	Volume Not Met
Breath Volume		0.519	23:46	
Air Blank		0.000	23:46	
Control Test		0.075	23:47	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 08/29/2017 03:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.