

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PONDER, CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BRASWELL JR

DATE/TIME OF TEST: 08/02/2017 04:26:55

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.078	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.189	04:19
Breath Volume	2.222	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.168	04:22
Breath Volume	1.726	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 3	0.176	04:25
Breath Volume	2.949	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEADOWS, SHAWN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: GRYZBOSKI

DATE/TIME OF TEST: 08/02/2017 05:02:32

Diagnostics Check	OK	04:55
Air Blank	0.000	04:56
Control Test	0.079	04:56
Air Blank	0.000	04:56
Subject Sample # 1	0.171	04:57
Breath Volume	1.324	04:57
Air Blank	0.000	04:58
Air Blank	0.000	05:00
Subject Sample # 2	0.167	05:01
Breath Volume	1.402	05:01
Air Blank	0.000	05:01
Control Test	0.078	05:01
Air Blank	0.000	05:02
Diagnostics Check	OK	05:02

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DRAUGHON,BOBBY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:39

SUBJECT'S LAST NAME: MORALES-RODRIGUEZ

DATE/TIME OF TEST: 08/04/2017 22:17:13

Diagnostics Check	OK	22:06
Air Blank	0.000	22:07
Control Test	0.079	22:07
Air Blank	0.000	22:08
Subject Sample # 1	0.240	22:09
Breath Volume	1.343	22:09
Air Blank	0.000	22:09
Air Blank	0.000	22:11
Subject Sample # 2	0.279	22:12
Breath Volume	1.367	22:12
Air Blank	0.000	22:13
Air Blank	0.000	22:15
Subject Sample # 3	0.263	22:15
Breath Volume	1.945	22:15
Air Blank	0.000	22:16
Control Test	0.079	22:16
Air Blank	0.000	22:17
Diagnostics Check	OK	22:17

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RINCON,RENZO
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:52

SUBJECT'S LAST NAME: VALDES OQUENDO

DATE/TIME OF TEST: 08/05/2017 05:18:35

Diagnostics Check	OK	05:12
Air Blank	0.000	05:12
Control Test	0.080	05:13
Air Blank	0.000	05:13
Subject Sample # 1	0.117	05:14
Breath Volume	2.554	05:14
Air Blank	0.000	05:14
Air Blank	0.000	05:16
Subject Sample # 2	0.117	05:17
Breath Volume	2.261	05:17
Air Blank	0.000	05:17
Control Test	0.079	05:18
Air Blank	0.000	05:18
Diagnostics Check	OK	05:18

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RIVERA, CARLOS
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: NOON

DATE/TIME OF TEST: 08/08/2017 03:34:54

Diagnostics Check		OK	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:30
Subject Sample	# 1	0.278	03:30
Breath Volume		1.527	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.273	03:33
Breath Volume		1.128	03:33
Air Blank		0.000	03:34
Control Test		0.078	03:34
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CONDREY,SCOTT
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: SPARKS

DATE/TIME OF TEST: 08/08/2017 17:24:45

Diagnostics Check	OK	17:18
Air Blank	0.000	17:18
Control Test	0.079	17:18
Air Blank	0.000	17:19
Subject Sample # 1	0.000	17:20
Breath Volume	2.457	17:20
Air Blank	0.000	17:20
Air Blank	0.000	17:22
Subject Sample # 2	0.000	17:23
Breath Volume	1.695	17:23
Air Blank	0.000	17:23
Control Test	0.079	17:24
Air Blank	0.000	17:24
Diagnostics Check	OK	17:24

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,PAUL,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CORRALES,ANGEL
ARRESTING OFFICER AGENCY: FDOT MCC
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 10:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:05

SUBJECT'S LAST NAME: SCHLAGEL

DATE/TIME OF TEST: 08/09/2017 11:29:27

Diagnostics Check	OK	11:26	
Air Blank	0.000	11:27	
Control Test	0.079	11:27	
Air Blank	0.000	11:27	
Subject Sample	# 1	11:28	Subject Test Refused
Breath Volume	0.000	11:28	
Air Blank	0.000	11:28	
Control Test	0.080	11:28	
Air Blank	0.000	11:29	
Diagnostics Check	OK	11:29	

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,SHAWN,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: WHEELER,DOUGLAS
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: JAMESON

DATE/TIME OF TEST: 08/10/2017 20:47:30

Diagnostics Check	OK	20:41
Air Blank	0.000	20:41
Control Test	0.079	20:41
Air Blank	0.000	20:42
Subject Sample # 1	0.195	20:43
Breath Volume	2.863	20:43
Air Blank	0.000	20:43
Air Blank	0.000	20:45
Subject Sample # 2	0.179	20:46
Breath Volume	1.121	20:46
Air Blank	0.000	20:46
Control Test	0.080	20:46
Air Blank	0.000	20:47
Diagnostics Check	OK	20:47

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ANDREWS,GARY
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: DIAZ RAMOS

DATE/TIME OF TEST: 08/12/2017 02:33:08

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.143	02:28
Breath Volume	1.363	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.147	02:31
Breath Volume	1.128	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TRANSUE,ZACHERY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: DONALDSON

DATE/TIME OF TEST: 08/12/2017 03:01:00

Diagnostics Check	OK	02:54
Air Blank	0.000	02:55
Control Test	0.079	02:55
Air Blank	0.000	02:55
Subject Sample # 1	0.119	02:56
Breath Volume	2.046	02:56
Air Blank	0.000	02:57
Air Blank	0.000	02:58
Subject Sample # 2	0.117	02:59
Breath Volume	2.195	02:59
Air Blank	0.000	03:00
Control Test	0.078	03:00
Air Blank	0.000	03:00
Diagnostics Check	OK	03:01

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BREWSTER,ADAM
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: HELM

DATE/TIME OF TEST: 08/12/2017 05:03:56

Diagnostics Check	OK	04:57
Air Blank	0.000	04:57
Control Test	0.079	04:58
Air Blank	0.000	04:58
Subject Sample # 1	0.114	04:59
Breath Volume	3.007	04:59
Air Blank	0.000	05:00
Air Blank	0.000	05:01
Subject Sample # 2	0.108	05:02
Breath Volume	1.335	05:02
Air Blank	0.000	05:03
Control Test	0.079	05:03
Air Blank	0.000	05:03
Diagnostics Check	OK	05:03

Data Download Date/Time: 08/13/2017 20:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PEREZ,E
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 15:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:33

SUBJECT'S LAST NAME: GARRITANI

DATE/TIME OF TEST: 08/17/2017 17:01:56

Diagnostics Check	OK	16:54
Air Blank	0.000	16:55
Control Test	0.078	16:55
Air Blank	0.000	16:56
Subject Sample # 1	0.138	16:57
Breath Volume	2.953	16:57
Air Blank	0.000	16:57
Air Blank	0.000	16:59
Subject Sample # 2	0.148	17:00
Breath Volume	1.886	17:00
Air Blank	0.000	17:01
Control Test	0.079	17:01
Air Blank	0.000	17:01
Diagnostics Check	OK	17:01

Data Download Date/Time: 10/03/2017 02:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DELLOVA, JOSEPH
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: ROUCHARD

DATE/TIME OF TEST: 08/25/2017 02:17:36

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.078	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.168	02:12
Breath Volume	1.292	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.163	02:16
Breath Volume	1.199	02:16
Air Blank	0.000	02:16
Control Test	0.078	02:17
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 10/03/2017 02:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEADOWS,SHAWN,M
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: CORNELL,JEREMEY
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: SCHROTH

DATE/TIME OF TEST: 08/28/2017 03:31:23

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:29	
Control Test	0.079	03:29	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.000	03:30	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:31	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 10/03/2017 02:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,PAUL,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TITTLE,JAMES
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 08/28/2017 06:35:21

Diagnostics Check	OK	06:33	
Air Blank	0.000	06:33	
Control Test	0.078	06:34	
Air Blank	0.000	06:34	
Subject Sample	# 1	06:34	Sequence Aborted
Breath Volume	0.000	06:34	
Air Blank	0.000	06:35	

Data Download Date/Time: 10/03/2017 02:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,PAUL,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TITTLE,JAMES
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:00

SUBJECT'S LAST NAME: DENTON

DATE/TIME OF TEST: 08/28/2017 06:40:49

Diagnostics Check	OK	06:37	
Air Blank	0.000	06:38	
Control Test	0.079	06:38	
Air Blank	0.000	06:39	
Subject Sample	# 1	06:39	Subject Test Refused
Breath Volume	0.000	06:39	
Air Blank	0.000	06:39	
Control Test	0.079	06:40	
Air Blank	0.000	06:40	
Diagnostics Check	OK	06:40	

Data Download Date/Time: 10/03/2017 02:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000999

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULSE,ROXANNE,S
BREATH TEST OPERATOR AGENCY: Sanford PD
ARRESTING OFFICER: MCDONNELL,B
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 15:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 08/28/2017 17:11:30

Diagnostics Check	OK	17:05
Air Blank	0.000	17:06
Control Test	0.079	17:06
Air Blank	0.000	17:06
Subject Sample # 1	0.000	17:07
Breath Volume	1.390	17:07
Air Blank	0.000	17:07
Air Blank	0.000	17:09
Subject Sample # 2	0.000	17:10
Breath Volume	1.390	17:10
Air Blank	0.000	17:10
Control Test	0.079	17:10
Air Blank	0.000	17:11
Diagnostics Check	OK	17:11

Data Download Date/Time: 10/03/2017 02:28

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001001

INSTRUMENT REGISTERED TO: SATELLITE BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWENS,STEVE,E
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: OWENS,STEVE
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 646645
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: KLEHN

DATE/TIME OF TEST: 08/06/2017 18:59:28

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:53
Subject Sample # 1	0.250	18:54
Breath Volume	1.941	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:56
Subject Sample # 2	0.234	18:57
Breath Volume	1.863	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 08/21/2017 07:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WARD,SHANE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: URESTE

DATE/TIME OF TEST: 08/06/2017 01:32:11

Diagnostics Check	OK	01:23	
Air Blank	0.000	01:24	
Control Test	0.079	01:24	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:28	No Sample Provided
Breath Volume	0.000	01:28	
Air Blank	0.000	01:28	
Air Blank	0.000	01:30	
Subject Sample	# 2	01:30	Subject Test Refused
Breath Volume	0.000	01:30	
Air Blank	0.000	01:31	
Control Test	0.079	01:31	
Air Blank	0.000	01:32	
Diagnostics Check	OK	01:32	

Data Download Date/Time: 08/17/2017 23:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: BRANDEBERRY,JEREMY
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 08/21/2017 03:43:31

Diagnostics Check	OK	03:36
Air Blank	0.000	03:36
Control Test	0.079	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.055	03:39
Breath Volume	1.199	03:39
Air Blank	0.000	03:39
Air Blank	0.000	03:41
Subject Sample # 2	0.059	03:42
Breath Volume	3.070	03:42
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 09/29/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARAUJO,NESTOR,J
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: EVERSOLE,DONNIE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: LADWIG

DATE/TIME OF TEST: 08/27/2017 20:51:53

Diagnostics Check	OK	20:49	
Air Blank	0.000	20:49	
Control Test	0.078	20:49	
Air Blank	0.000	20:50	
Subject Sample	# 1	20:50	Subject Test Refused
Breath Volume	0.000	20:50	
Air Blank	0.000	20:50	
Control Test	0.079	20:51	
Air Blank	0.000	20:51	
Diagnostics Check	OK	20:51	

Data Download Date/Time: 09/29/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ALVARADO

DATE/TIME OF TEST: 08/03/2017 00:24:08

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.081	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Interferent Detect
Breath Volume	1.609	00:22	
Air Blank	0.000	00:23	
Control Test	0.081	00:23	
Air Blank	0.000	00:24	
Diagnostics Check	OK	00:24	

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: JANEZIC

DATE/TIME OF TEST: 08/03/2017 02:30:10

Diagnostics Check	OK	02:26	
Air Blank	0.000	02:26	
Control Test	0.082	02:26	
Air Blank	0.000	02:27	
Subject Sample	# 1	02:28	Subject Test Refused
Breath Volume	0.000	02:28	
Air Blank	0.000	02:29	
Control Test	0.082	02:29	
Air Blank	0.000	02:30	
Diagnostics Check	OK	02:30	

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 08/03/2017 03:17:41

Diagnostics Check	OK	03:14	
Air Blank	0.000	03:14	
Control Test	0.081	03:15	
Air Blank	0.000	03:15	
Subject Sample	# 1	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:16	
Control Test	0.082	03:17	
Air Blank	0.000	03:17	
Diagnostics Check	OK	03:17	

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: DANTICO

DATE/TIME OF TEST: 08/03/2017 22:17:03

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:05	
Control Test		0.080	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1	0.228	22:09	Volume Not Met
Breath Volume		0.816	22:09	
Air Blank		0.000	22:10	
Air Blank		0.000	22:12	
Subject Sample	# 2	0.215	22:15	Volume Not Met
Breath Volume		0.906	22:15	
Air Blank		0.000	22:16	
Control Test		0.080	22:16	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:17	

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: DECOU

DATE/TIME OF TEST: 08/04/2017 03:29:25

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.081	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.000	03:24
Breath Volume	1.234	03:24
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.000	03:28
Breath Volume	1.359	03:28
Air Blank	0.000	03:28
Control Test	0.081	03:28
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BRUSCA

DATE/TIME OF TEST: 08/04/2017 03:47:27

Diagnostics Check	OK	03:40
Air Blank	0.000	03:41
Control Test	0.081	03:41
Air Blank	0.000	03:41
Subject Sample # 1	0.119	03:43
Breath Volume	1.437	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.116	03:45
Breath Volume	1.410	03:45
Air Blank	0.000	03:46
Control Test	0.081	03:46
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PRINDLE

DATE/TIME OF TEST: 08/05/2017 00:03:09

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 2	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 1	00:01	No Sample Provided
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Control Test	0.082	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: PRINDLE

DATE/TIME OF TEST: 08/05/2017 00:12:29

Diagnostics Check	OK	00:04
Air Blank	0.000	00:05
Control Test	0.081	00:05
Air Blank	0.000	00:06
Subject Sample # 1	0.087	00:07
Breath Volume	1.386	00:07
Air Blank	0.000	00:07
Air Blank	0.000	00:09
Subject Sample # 2	0.085	00:10
Breath Volume	1.238	00:10
Air Blank	0.000	00:11
Control Test	0.081	00:11
Air Blank	0.000	00:12
Diagnostics Check	OK	00:12

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CLOWES

DATE/TIME OF TEST: 08/05/2017 00:46:13

Diagnostics Check	OK	00:42	
Air Blank	0.000	00:43	
Control Test	0.081	00:43	
Air Blank	0.000	00:44	
Subject Sample	# 1	00:44	Subject Test Refused
Breath Volume	0.000	00:44	
Air Blank	0.000	00:45	
Control Test	0.081	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, WILLIAM, T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: DETELICH

DATE/TIME OF TEST: 08/05/2017 02:39:40

Diagnostics Check	OK	02:32
Air Blank	0.000	02:33
Control Test	0.082	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.238	02:35
Breath Volume	1.539	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.243	02:38
Breath Volume	1.386	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 08/10/2017 21:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: CUPERY

DATE/TIME OF TEST: 08/05/2017 04:03:07

Diagnostics Check	OK	03:56
Air Blank	0.000	03:57
Control Test	0.081	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.086	03:58
Breath Volume	2.574	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.090	04:01
Breath Volume	2.707	04:01
Air Blank	0.000	04:02
Control Test	0.081	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 08/10/2017 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REFUSE, RONY,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: REFUSE, RONY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 08/05/2017 06:11:37

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.080	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.046	06:07
Breath Volume	1.597	06:07
Air Blank	0.000	06:07
Air Blank	0.000	06:09
Subject Sample # 2	0.046	06:10
Breath Volume	1.406	06:10
Air Blank	0.000	06:10
Control Test	0.081	06:11
Air Blank	0.000	06:11
Diagnostics Check	OK	06:11

Data Download Date/Time: 08/10/2017 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:46

SUBJECT'S LAST NAME: SHEAFFER

DATE/TIME OF TEST: 08/05/2017 20:17:39

Diagnostics Check	OK	20:06	
Air Blank	0.000	20:07	
Control Test	0.080	20:07	
Air Blank	0.000	20:08	
Subject Sample # 1	0.143	20:11	Volume Not Met
Breath Volume	0.453	20:11	
Air Blank	0.000	20:11	
Air Blank	0.000	20:13	
Subject Sample # 2		20:16	Subject Test Refused
Breath Volume	1.011	20:16	
Air Blank	0.000	20:16	
Control Test	0.079	20:17	
Air Blank	0.000	20:17	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 08/10/2017 21:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, JONATHAN, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA, JANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: KOVERMAN

DATE/TIME OF TEST: 08/07/2017 00:00:34

Diagnostics Check	OK	23:51	
Air Blank	0.000	23:52	
Control Test	0.081	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:56	No Sample Provided
Breath Volume	0.000	23:56	
Air Blank	0.000	23:56	
Air Blank	0.000	23:58	
Subject Sample	# 2	23:59	Subject Test Refused
Breath Volume	0.000	23:59	
Air Blank	0.000	23:59	
Control Test	0.081	23:59	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:00	

Data Download Date/Time: 08/10/2017 21:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: AMATRUDA,CHRIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: CHRISTIAN

DATE/TIME OF TEST: 08/10/2017 17:14:39

Diagnostics Check	OK	17:11	
Air Blank	0.000	17:12	
Control Test	0.081	17:12	
Air Blank	0.000	17:13	
Subject Sample	# 1	17:13	Subject Test Refused
Breath Volume	0.000	17:13	
Air Blank	0.000	17:13	
Control Test	0.081	17:14	
Air Blank	0.000	17:14	
Diagnostics Check	OK	17:14	

Data Download Date/Time: 08/10/2017 21:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: KARAPHILLIS

DATE/TIME OF TEST: 08/11/2017 18:16:19

Diagnostics Check	OK	18:10
Air Blank	0.000	18:10
Control Test	0.081	18:11
Air Blank	0.000	18:11
Subject Sample # 1	0.000	18:12
Breath Volume	2.039	18:12
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.000	18:14
Breath Volume	1.832	18:14
Air Blank	0.000	18:15
Control Test	0.081	18:15
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: REICHNER

DATE/TIME OF TEST: 08/13/2017 02:17:42

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.188	02:09	Volume Not Met
Breath Volume		0.289	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.181	02:12	
Breath Volume		1.753	02:12	
Air Blank		0.000	02:13	
Air Blank		0.000	02:15	
Subject Sample	# 3	0.188	02:16	
Breath Volume		1.183	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORTHINGTON,JUSTIN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WORTHINGTON,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BATISTAIBANEZ

DATE/TIME OF TEST: 08/13/2017 04:05:51

Diagnostics Check	OK	04:01	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:04	Subject Test Refused
Breath Volume	0.000	04:04	
Air Blank	0.000	04:04	
Control Test	0.080	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FORD,ERIC
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: FAGAN

DATE/TIME OF TEST: 08/13/2017 05:33:33

Diagnostics Check	OK	05:30	
Air Blank	0.000	05:30	
Control Test	0.081	05:31	
Air Blank	0.000	05:31	
Subject Sample	# 1	05:32	Subject Test Refused
Breath Volume	0.000	05:32	
Air Blank	0.000	05:32	
Control Test	0.081	05:32	
Air Blank	0.000	05:33	
Diagnostics Check	OK	05:33	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIELDEN,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: CARDINALE

DATE/TIME OF TEST: 08/14/2017 04:53:29

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:42	
Control Test	0.081	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:46	No Sample Provided
Breath Volume	0.000	04:46	
Air Blank	0.000	04:46	
Air Blank	0.000	04:48	
Subject Sample	# 2	04:51	Volume Not Met
Breath Volume	0.257	04:51	
Air Blank	0.000	04:52	
Control Test	0.081	04:52	
Air Blank	0.000	04:53	
Diagnostics Check	OK	04:53	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIELDEN,E
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: CARDINALE

DATE/TIME OF TEST: 08/14/2017 04:59:56

Diagnostics Check	OK	04:55	
Air Blank	0.000	04:55	
Control Test	0.080	04:55	
Air Blank	0.000	04:56	
Subject Sample	# 1	04:58	Subject Test Refused
Breath Volume	0.472	04:58	
Air Blank	0.000	04:59	
Control Test	0.080	04:59	
Air Blank	0.000	04:59	
Diagnostics Check	OK	04:59	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE, JONATHAN, S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER, ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:51

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 08/15/2017 21:20:00

Diagnostics Check	OK	21:12
Air Blank	0.000	21:12
Control Test	0.081	21:12
Air Blank	0.000	21:13
Subject Sample # 1	0.019	21:15
Breath Volume	1.628	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.015	21:18
Breath Volume	2.164	21:18
Air Blank	0.000	21:19
Control Test	0.080	21:19
Air Blank	0.000	21:19
Diagnostics Check	OK	21:20

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/16/2017 23:14:42

Diagnostics Check		OK	23:08
Air Blank		0.000	23:08
Control Test		0.080	23:09
Air Blank		0.000	23:09
Subject Sample	# 1	0.281	23:10
Breath Volume		2.085	23:10
Air Blank		0.000	23:10
Air Blank		0.000	23:12
Subject Sample	# 2	0.282	23:13
Breath Volume		1.605	23:13
Air Blank		0.000	23:13
Control Test		0.078	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HARR,JONATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: SHIVGOLAM

DATE/TIME OF TEST: 08/18/2017 23:58:10

Diagnostics Check	OK	23:55	
Air Blank	0.000	23:55	
Control Test	0.081	23:55	
Air Blank	0.000	23:56	
Subject Sample	# 1	23:56	Subject Test Refused
Breath Volume	0.000	23:56	
Air Blank	0.000	23:57	
Control Test	0.081	23:57	
Air Blank	0.000	23:58	
Diagnostics Check	OK	23:58	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAKE,JONATHAN,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAKE,JONATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MARCH

DATE/TIME OF TEST: 08/19/2017 02:04:11

Diagnostics Check	OK	01:55	
Air Blank	0.000	01:56	
Control Test	0.080	01:56	
Air Blank	0.000	01:56	
Subject Sample	# 1	02:00	No Sample Provided
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Air Blank	0.000	02:02	
Subject Sample	# 2	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank	0.000	02:03	
Control Test	0.081	02:03	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIVECOATS,CURTIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 08:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:46

SUBJECT'S LAST NAME: HATZIPAVLOU

DATE/TIME OF TEST: 08/19/2017 10:22:02

Diagnostics Check	OK	10:15
Air Blank	0.000	10:15
Control Test	0.081	10:16
Air Blank	0.000	10:16
Subject Sample # 1	0.176	10:17
Breath Volume	1.929	10:17
Air Blank	0.000	10:18
Air Blank	0.000	10:19
Subject Sample # 2	0.175	10:20
Breath Volume	1.675	10:20
Air Blank	0.000	10:21
Control Test	0.080	10:21
Air Blank	0.000	10:21
Diagnostics Check	OK	10:22

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,BRYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 11:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: CAULKING

DATE/TIME OF TEST: 08/19/2017 12:41:24

Diagnostics Check	OK	12:34
Air Blank	0.000	12:34
Control Test	0.081	12:35
Air Blank	0.000	12:35
Subject Sample # 1	0.000	12:36
Breath Volume	2.179	12:36
Air Blank	0.000	12:37
Air Blank	0.000	12:39
Subject Sample # 2	0.000	12:40
Breath Volume	1.343	12:40
Air Blank	0.000	12:40
Control Test	0.081	12:40
Air Blank	0.000	12:41
Diagnostics Check	OK	12:41

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: NAGASAWA

DATE/TIME OF TEST: 08/21/2017 02:34:53

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.081	02:32	
Air Blank	0.000	02:32	
Subject Sample	# 1	02:33	Subject Test Refused
Breath Volume	0.000	02:33	
Air Blank	0.000	02:33	
Control Test	0.080	02:34	
Air Blank	0.000	02:34	
Diagnostics Check	OK	02:34	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEELER,BRYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ORRIS

DATE/TIME OF TEST: 08/27/2017 03:25:18

Diagnostics Check	OK	03:21	
Air Blank	0.000	03:21	
Control Test	0.081	03:22	
Air Blank	0.000	03:22	
Subject Sample	# 1	03:23	Subject Test Refused
Breath Volume	0.000	03:23	
Air Blank	0.000	03:24	
Control Test	0.081	03:24	
Air Blank	0.000	03:25	
Diagnostics Check	OK	03:25	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORRAO, MICHAEL, W
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CORRAO, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: CASIMER

DATE/TIME OF TEST: 08/27/2017 19:13:17

Diagnostics Check	OK	19:05
Air Blank	0.000	19:05
Control Test	0.080	19:06
Air Blank	0.000	19:06
Subject Sample # 1	0.205	19:08
Breath Volume	1.894	19:08
Air Blank	0.000	19:08
Air Blank	0.000	19:10
Subject Sample # 2	0.201	19:11
Breath Volume	2.300	19:11
Air Blank	0.000	19:12
Control Test	0.078	19:12
Air Blank	0.000	19:13
Diagnostics Check	OK	19:13

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MARDUKHAYEV

DATE/TIME OF TEST: 08/27/2017 22:46:02

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:34	
Control Test		0.081	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1	0.159	22:38	Volume Not Met
Breath Volume		0.183	22:38	
Air Blank		0.000	22:39	
Air Blank		0.000	22:41	
Subject Sample	# 2	0.166	22:44	Slope Not Level
Breath Volume		1.125	22:44	
Air Blank		0.000	22:45	
Control Test		0.080	22:45	
Air Blank		0.000	22:45	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRENN,DAVID,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,JANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: TOMKINS

DATE/TIME OF TEST: 08/29/2017 04:47:54

Diagnostics Check	OK	04:44	
Air Blank	0.000	04:44	
Control Test	0.081	04:44	
Air Blank	0.000	04:45	
Subject Sample	# 1	04:46	Subject Test Refused
Breath Volume	0.000	04:46	
Air Blank	0.000	04:47	
Control Test	0.082	04:47	
Air Blank	0.000	04:47	
Diagnostics Check	OK	04:47	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SOSA,ROBERTO
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 06:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:54

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 08/29/2017 08:18:25

Diagnostics Check	OK	08:15	
Air Blank	0.000	08:16	
Control Test	0.080	08:16	
Air Blank	0.000	08:16	
Subject Sample	# 1	08:17	Subject Test Refused
Breath Volume	0.000	08:17	
Air Blank	0.000	08:17	
Control Test	0.080	08:17	
Air Blank	0.000	08:18	
Diagnostics Check	OK	08:18	

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PERSAUD,COMPTON,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PERSAUD,COMPTON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BUCHER

DATE/TIME OF TEST: 08/31/2017 03:03:02

Diagnostics Check	OK	02:53
Air Blank	0.000	02:53
Control Test	0.081	02:54
Air Blank	0.000	02:54
Subject Sample # 1	0.184	02:57
Breath Volume	1.445	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.179	03:01
Breath Volume	1.437	03:01
Air Blank	0.000	03:02
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:03

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 08/31/2017 06:18:18

Diagnostics Check	OK	06:10
Air Blank	0.000	06:11
Control Test	0.081	06:11
Air Blank	0.000	06:12
Subject Sample # 1	0.173	06:13
Breath Volume	1.425	06:13
Air Blank	0.000	06:14
Air Blank	0.000	06:15
Subject Sample # 2	0.169	06:16
Breath Volume	1.570	06:16
Air Blank	0.000	06:17
Control Test	0.081	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 09/06/2017 22:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR, RONALD, J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BLAIR, RONALD
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: FAIN

DATE/TIME OF TEST: 08/05/2017 00:46:32

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.079	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:45	
Control Test	0.079	00:45	
Air Blank	0.000	00:46	
Diagnostics Check	OK	00:46	

Data Download Date/Time: 08/10/2017 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: PETIKA

DATE/TIME OF TEST: 08/05/2017 02:58:15

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:57	
Control Test	0.080	02:57	
Air Blank	0.000	02:58	
Diagnostics Check	OK	02:58	

Data Download Date/Time: 08/10/2017 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:19

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 08/05/2017 04:56:03

Diagnostics Check	OK	04:49
Air Blank	0.000	04:50
Control Test	0.079	04:50
Air Blank	0.000	04:51
Subject Sample # 1	0.138	04:51
Breath Volume	2.921	04:51
Air Blank	0.000	04:52
Air Blank	0.000	04:54
Subject Sample # 2	0.136	04:54
Breath Volume	3.039	04:54
Air Blank	0.000	04:55
Control Test	0.079	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:56

Data Download Date/Time: 08/10/2017 20:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRUNDMANN,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/18/2017 01:59:48

Diagnostics Check	OK	01:53
Air Blank	0.000	01:53
Control Test	0.079	01:54
Air Blank	0.000	01:54
Subject Sample # 1	0.174	01:55
Breath Volume	2.460	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.175	01:58
Breath Volume	2.671	01:58
Air Blank	0.000	01:58
Control Test	0.079	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 09/06/2017 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 08/19/2017 00:11:31

Diagnostics Check	OK	00:08	
Air Blank	0.000	00:08	
Control Test	0.079	00:08	
Air Blank	0.000	00:09	
Subject Sample	# 1	00:10	Subject Test Refused
Breath Volume	0.000	00:10	
Air Blank	0.000	00:10	
Control Test	0.079	00:10	
Air Blank	0.000	00:11	
Diagnostics Check	OK	00:11	

Data Download Date/Time: 09/06/2017 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEIB,MELANIE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SESTER

DATE/TIME OF TEST: 08/20/2017 02:15:06

Diagnostics Check	OK	02:08
Air Blank	0.000	02:08
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.165	02:10
Breath Volume	2.250	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.165	02:13
Breath Volume	2.058	02:13
Air Blank	0.000	02:14
Control Test	0.079	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 09/06/2017 20:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,C
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: KIRLEY

DATE/TIME OF TEST: 08/21/2017 22:00:04

Diagnostics Check	OK	21:52
Air Blank	0.000	21:53
Control Test	0.080	21:53
Air Blank	0.000	21:54
Subject Sample # 1	0.260	21:54
Breath Volume	1.722	21:54
Air Blank	0.000	21:55
Air Blank	0.000	21:57
Subject Sample # 2	0.259	21:58
Breath Volume	2.226	21:58
Air Blank	0.000	21:59
Control Test	0.079	21:59
Air Blank	0.000	21:59
Diagnostics Check	OK	22:00

Data Download Date/Time: 09/06/2017 20:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SKALKO,RYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HARDY

DATE/TIME OF TEST: 08/28/2017 02:43:12

Diagnostics Check	OK	02:36
Air Blank	0.000	02:37
Control Test	0.079	02:37
Air Blank	0.000	02:38
Subject Sample # 1	0.204	02:38
Breath Volume	1.535	02:38
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.211	02:41
Breath Volume	1.843	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 09/06/2017 20:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCKENZIE,DEPUTY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MITCHELL,DEPUTY B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 07:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: DOERGE

DATE/TIME OF TEST: 08/28/2017 08:22:03

Diagnostics Check		OK	08:15
Air Blank		0.000	08:16
Control Test		0.079	08:16
Air Blank		0.000	08:17
Subject Sample	# 1	0.305	08:17
Breath Volume		1.496	08:17
Air Blank		0.000	08:18
Air Blank		0.000	08:19
Subject Sample	# 2	0.304	08:20
Breath Volume		1.640	08:20
Air Blank		0.000	08:21
Control Test		0.079	08:21
Air Blank		0.000	08:21
Diagnostics Check		OK	08:22

Data Download Date/Time: 09/06/2017 20:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILHELM,JAMES,D
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: LOPER

DATE/TIME OF TEST: 08/30/2017 01:00:41

Diagnostics Check	OK	00:54
Air Blank	0.000	00:55
Control Test	0.079	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.122	00:56
Breath Volume	1.992	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.122	00:59
Breath Volume	1.601	00:59
Air Blank	0.000	00:59
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 09/06/2017 20:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLSTINE,SCOTT,
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: VILA,ALEXANDER
ARRESTING OFFICER AGENCY: Melbourne Beach PD
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: TREVISANI

DATE/TIME OF TEST: 08/06/2017 19:47:53

Diagnostics Check	OK	19:41
Air Blank	0.000	19:41
Control Test	0.078	19:41
Air Blank	0.000	19:42
Subject Sample # 1	0.068	19:42
Breath Volume	1.414	19:42
Air Blank	0.000	19:43
Air Blank	0.000	19:45
Subject Sample # 2	0.064	19:46
Breath Volume	1.570	19:46
Air Blank	0.000	19:46
Control Test	0.078	19:47
Air Blank	0.000	19:47
Diagnostics Check	OK	19:47

Data Download Date/Time: 08/28/2017 06:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VERDEROSA, TIFFANY, M
BREATH TEST OPERATOR AGENCY: Indialantic PD
ARRESTING OFFICER: VERDEROSA, TIFFANY
ARRESTING OFFICER AGENCY: Indialantic PD
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:05

SUBJECT'S LAST NAME: PACITTO

DATE/TIME OF TEST: 08/23/2017 08:51:17

Diagnostics Check	OK	08:44
Air Blank	0.000	08:44
Control Test	0.078	08:44
Air Blank	0.000	08:45
Subject Sample # 1	0.000	08:46
Breath Volume	4.316	08:46
Air Blank	0.000	08:47
Air Blank	0.000	08:49
Subject Sample # 2	0.000	08:49
Breath Volume	4.082	08:49
Air Blank	0.000	08:50
Control Test	0.077	08:50
Air Blank	0.000	08:51
Diagnostics Check	OK	08:51

Data Download Date/Time: 08/28/2017 06:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTTEBO,KENT,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GUTTEBO,KENT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ELLINGSON

DATE/TIME OF TEST: 08/27/2017 02:30:48

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:25
Subject Sample	# 1	0.211	02:26
Breath Volume		4.292	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.214	02:29
Breath Volume		3.566	02:29
Air Blank		0.000	02:29
Control Test		0.077	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 08/28/2017 06:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001006

INSTRUMENT REGISTERED TO: INDIALANTIC P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTTEBO,KENT,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GUTTEBO,KENT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: BLEAU

DATE/TIME OF TEST: 08/31/2017 19:36:50

Diagnostics Check	OK	19:30
Air Blank	0.000	19:31
Control Test	0.078	19:31
Air Blank	0.000	19:31
Subject Sample # 1	0.258	19:32
Breath Volume	3.632	19:32
Air Blank	0.000	19:33
Air Blank	0.000	19:34
Subject Sample # 2	0.254	19:35
Breath Volume	2.460	19:35
Air Blank	0.000	19:35
Control Test	0.077	19:36
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 09/27/2017 05:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: BEARY,JOHN
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 08/12/2017 05:12:46

Diagnostics Check	OK	05:06
Air Blank	0.000	05:07
Control Test	0.078	05:07
Air Blank	0.000	05:08
Subject Sample # 1	0.152	05:08
Breath Volume	2.136	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:10
Subject Sample # 2	0.151	05:11
Breath Volume	2.054	05:11
Air Blank	0.000	05:11
Control Test	0.078	05:12
Air Blank	0.000	05:12
Diagnostics Check	OK	05:12

Data Download Date/Time: 08/30/2017 20:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: LACEY,JOHN
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: ELAM

DATE/TIME OF TEST: 08/12/2017 22:44:09

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.079	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.273	22:33
Breath Volume	1.929	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.232	22:38
Breath Volume	1.554	22:38
Air Blank	0.000	22:39
Air Blank	0.000	22:40
Subject Sample # 3	0.248	22:42
Breath Volume	1.476	22:42
Air Blank	0.000	22:43
Control Test	0.075	22:43
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 08/30/2017 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUTTEBO,KENT,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GUTTEBO,KENT
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: EMBREE

DATE/TIME OF TEST: 08/26/2017 19:59:33

Diagnostics Check		OK	19:53
Air Blank		0.000	19:53
Control Test		0.079	19:54
Air Blank		0.000	19:54
Subject Sample	# 1	0.245	19:55
Breath Volume		2.253	19:55
Air Blank		0.000	19:55
Air Blank		0.000	19:57
Subject Sample	# 2	0.231	19:58
Breath Volume		1.558	19:58
Air Blank		0.000	19:58
Control Test		0.078	19:58
Air Blank		0.000	19:59
Diagnostics Check		OK	19:59

Data Download Date/Time: 08/30/2017 20:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: MCKINNEY,RHIANNON
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 08/18/2017 03:09:40

Diagnostics Check	OK	03:02
Air Blank	0.000	03:02
Control Test	0.078	03:02
Air Blank	0.000	03:03
Subject Sample # 1	0.081	03:04
Breath Volume	1.683	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 2	0.074	03:08
Breath Volume	1.589	03:08
Air Blank	0.000	03:08
Control Test	0.077	03:09
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 08/18/2017 12:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOOS,JAMES,P
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: SILFIN,JEAN
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MCGREW

DATE/TIME OF TEST: 08/27/2017 02:15:59

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.079	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.000	02:11
Breath Volume	3.796	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.000	02:14
Breath Volume	3.628	02:14
Air Blank	0.000	02:15
Control Test	0.078	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 09/22/2017 14:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOB,MORGANA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: FELICIANO,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: CUNNINGHAM

DATE/TIME OF TEST: 08/03/2017 17:29:25

Diagnostics Check	OK	17:26	
Air Blank	0.000	17:26	
Control Test	0.079	17:27	
Air Blank	0.000	17:27	
Subject Sample	# 1	17:28	Subject Test Refused
Breath Volume	0.000	17:28	
Air Blank	0.000	17:28	
Control Test	0.078	17:28	
Air Blank	0.000	17:29	
Diagnostics Check	OK	17:29	

Data Download Date/Time: 08/09/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROLLESTON,JACOB,M
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ROLLESTON,JACOB
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: MORRIS

DATE/TIME OF TEST: 08/04/2017 22:00:10

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.079	21:58	
Air Blank	0.000	21:58	
Subject Sample	# 1	21:58	Subject Test Refused
Breath Volume	0.000	21:58	
Air Blank	0.000	21:59	
Control Test	0.078	21:59	
Air Blank	0.000	22:00	
Diagnostics Check	OK	22:00	

Data Download Date/Time: 08/09/2017 10:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: JASIUKIEWICZ,MICHAEL
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 07/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: MALONE

DATE/TIME OF TEST: 08/05/2017 04:28:47

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.079	04:23
Air Blank		0.000	04:23
Subject Sample	# 1	0.242	04:24
Breath Volume		2.683	04:24
Air Blank		0.000	04:24
Air Blank		0.000	04:26
Subject Sample	# 2	0.245	04:27
Breath Volume		2.664	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 08/09/2017 10:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUFF,ZACHARY,J
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 08/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: AYOUB

DATE/TIME OF TEST: 08/10/2017 11:21:26

Diagnostics Check	OK	11:14
Air Blank	0.000	11:15
Control Test	0.079	11:15
Air Blank	0.000	11:16
Subject Sample # 1	0.000	11:16
Breath Volume	2.863	11:16
Air Blank	0.000	11:17
Air Blank	0.000	11:19
Subject Sample # 2	0.000	11:20
Breath Volume	2.644	11:20
Air Blank	0.000	11:20
Control Test	0.080	11:20
Air Blank	0.000	11:21
Diagnostics Check	OK	11:21

Data Download Date/Time: 08/16/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOB,MORGANA,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: GUY,JOSHUA
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 08/12/2017 04:23:26

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.078	04:18
Air Blank		0.000	04:18
Subject Sample	# 1	0.154	04:19
Breath Volume		2.218	04:19
Air Blank		0.000	04:19
Air Blank		0.000	04:21
Subject Sample	# 2	0.147	04:21
Breath Volume		2.066	04:21
Air Blank		0.000	04:22
Control Test		0.079	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 08/16/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUCHENEY,WESLEY,A
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: POLETZ,RYAN
ARRESTING OFFICER AGENCY: Pinellas Park PD
LAST AGENCY INSPECTION: 08/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: HELMINTOLLER

DATE/TIME OF TEST: 08/14/2017 22:29:59

Diagnostics Check	OK	22:23
Air Blank	0.000	22:23
Control Test	0.078	22:23
Air Blank	0.000	22:24
Subject Sample # 1	0.000	22:25
Breath Volume	4.914	22:25
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.000	22:28
Breath Volume	4.316	22:28
Air Blank	0.000	22:29
Control Test	0.080	22:29
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 08/16/2017 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ANAMUAH-MENSAH,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 08/15/2017 21:25:47

Diagnostics Check	OK	21:16	
Air Blank	0.000	21:17	
Control Test	0.079	21:17	
Air Blank	0.000	21:18	
Subject Sample # 1	0.310	21:20	
Breath Volume	1.277	21:20	
Air Blank	0.000	21:21	
Air Blank	0.000	21:22	
Subject Sample # 2	0.314	21:23	
Breath Volume	2.480	21:23	
Air Blank	0.000	21:24	
Control Test	0.078	21:24	
Air Blank	0.000	21:25	
Diagnostics Check	Fail	21:25	Temperature Regulation Test
Air Blank	0.000	21:25	

Data Download Date/Time: 08/16/2017 10:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001042

INSTRUMENT REGISTERED TO: PINELLAS PARK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JASIUKIEWICZ,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Pinellas Park PD
ARRESTING OFFICER: ANAMUAH-MENSAH,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 08/15/2017 21:35:42

Diagnostics Check	OK	21:28	
Air Blank	0.000	21:29	
Control Test	0.078	21:29	
Air Blank	0.000	21:29	
Subject Sample	# 1 0.294	21:30	
Breath Volume	1.441	21:30	
Air Blank	0.000	21:30	
Air Blank	0.000	21:32	
Subject Sample	# 2 0.294	21:33	
Breath Volume	1.695	21:33	
Air Blank	0.000	21:34	
Control Test	0.079	21:34	
Air Blank	0.000	21:35	
Diagnostics Check	Fail	21:35	Temperature Regulation Test
Air Blank	0.000	21:35	

Data Download Date/Time: 08/16/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DETTMAN,AMANDA
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: RIGGLES

DATE/TIME OF TEST: 08/06/2017 02:22:11

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.077	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:20	Subject Test Refused
Breath Volume	0.000	02:20	
Air Blank	0.000	02:21	
Control Test	0.078	02:21	
Air Blank	0.000	02:22	
Diagnostics Check	OK	02:22	

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,JOHN,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOTT,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: ADMIN
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 08/13/2017 15:28:03

Diagnostics Check	OK	15:22
Air Blank	0.000	15:22
Control Test	0.077	15:22
Air Blank	0.000	15:23
Subject Sample # 1	0.000	15:23
Breath Volume	1.445	15:23
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.000	15:26
Breath Volume	1.937	15:26
Air Blank	0.000	15:27
Control Test	0.078	15:27
Air Blank	0.000	15:27
Diagnostics Check	OK	15:28

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,I
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,KEVIN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 15:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:32

SUBJECT'S LAST NAME: VISCOMI

DATE/TIME OF TEST: 08/15/2017 16:03:28

Diagnostics Check	OK	15:57
Air Blank	0.000	15:58
Control Test	0.077	15:58
Air Blank	0.000	15:58
Subject Sample # 1	0.000	15:59
Breath Volume	3.585	15:59
Air Blank	0.000	15:59
Air Blank	0.000	16:01
Subject Sample # 2	0.000	16:02
Breath Volume	3.761	16:02
Air Blank	0.000	16:02
Control Test	0.077	16:02
Air Blank	0.000	16:03
Diagnostics Check	OK	16:03

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,I
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,KEVIN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 08/15/2017 16:28:01

Diagnostics Check	OK	16:22
Air Blank	0.000	16:22
Control Test	0.077	16:22
Air Blank	0.000	16:23
Subject Sample # 1	0.000	16:23
Breath Volume	3.359	16:23
Air Blank	0.000	16:24
Air Blank	0.000	16:26
Subject Sample # 2	0.000	16:26
Breath Volume	3.734	16:26
Air Blank	0.000	16:27
Control Test	0.077	16:27
Air Blank	0.000	16:27
Diagnostics Check	OK	16:28

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,V
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROBLES,J
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:52

SUBJECT'S LAST NAME: FLANAGAN

DATE/TIME OF TEST: 08/15/2017 17:18:50

Diagnostics Check	OK	17:12
Air Blank	0.000	17:13
Control Test	0.078	17:13
Air Blank	0.000	17:14
Subject Sample # 1	0.000	17:14
Breath Volume	3.417	17:14
Air Blank	0.000	17:14
Air Blank	0.000	17:16
Subject Sample # 2	0.000	17:17
Breath Volume	3.472	17:17
Air Blank	0.000	17:17
Control Test	0.077	17:18
Air Blank	0.000	17:18
Diagnostics Check	OK	17:18

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,NAKEEM,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BROWN,NAKEEM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:09

SUBJECT'S LAST NAME: CRON

DATE/TIME OF TEST: 08/18/2017 09:43:31

Diagnostics Check	OK	09:42	
Air Blank	0.000	09:42	
Control Test	0.074	09:42	Control Outside Tolerance
Air Blank	0.000	09:43	

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,NAKEEM,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BROWN,NAKEEM
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 09:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:09

SUBJECT'S LAST NAME: CRON

DATE/TIME OF TEST: 08/18/2017 09:51:53

Diagnostics Check	OK	09:46
Air Blank	0.000	09:46
Control Test	0.076	09:46
Air Blank	0.000	09:47
Subject Sample # 1	0.000	09:47
Breath Volume	2.675	09:47
Air Blank	0.000	09:48
Air Blank	0.000	09:50
Subject Sample # 2	0.000	09:50
Breath Volume	2.855	09:50
Air Blank	0.000	09:51
Control Test	0.076	09:51
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,SHANE,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SPENCER,SHANE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: MEGLIO

DATE/TIME OF TEST: 08/18/2017 17:31:59

Diagnostics Check	OK	17:26
Air Blank	0.000	17:26
Control Test	0.078	17:26
Air Blank	0.000	17:27
Subject Sample # 1	0.000	17:27
Breath Volume	1.265	17:27
Air Blank	0.000	17:28
Air Blank	0.000	17:30
Subject Sample # 2	0.000	17:30
Breath Volume	1.191	17:30
Air Blank	0.000	17:31
Control Test	0.077	17:31
Air Blank	0.000	17:31
Diagnostics Check	OK	17:31

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURPHY,J,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RHODEN,D
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: SALVADOR

DATE/TIME OF TEST: 08/19/2017 05:05:45

Diagnostics Check	OK	05:02	
Air Blank	0.000	05:03	
Control Test	0.077	05:03	
Air Blank	0.000	05:03	
Subject Sample	# 1	05:04	Subject Test Refused
Breath Volume	0.000	05:04	
Air Blank	0.000	05:04	
Control Test	0.077	05:05	
Air Blank	0.000	05:05	
Diagnostics Check	OK	05:05	

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: BARTSCH

DATE/TIME OF TEST: 08/23/2017 15:32:23

Diagnostics Check	OK	15:26
Air Blank	0.000	15:27
Control Test	0.077	15:27
Air Blank	0.000	15:27
Subject Sample # 1	0.000	15:28
Breath Volume	2.687	15:28
Air Blank	0.000	15:28
Air Blank	0.000	15:30
Subject Sample # 2	0.000	15:31
Breath Volume	2.324	15:31
Air Blank	0.000	15:31
Control Test	0.077	15:31
Air Blank	0.000	15:32
Diagnostics Check	OK	15:32

Data Download Date/Time: 08/25/2017 12:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/28/2017 10:02:41

Diagnostics Check	OK	09:56
Air Blank	0.000	09:57
Control Test	0.078	09:57
Air Blank	0.000	09:58
Subject Sample # 1	0.000	09:58
Breath Volume	2.699	09:58
Air Blank	0.000	09:58
Air Blank	0.000	10:00
Subject Sample # 2	0.000	10:01
Breath Volume	3.097	10:01
Air Blank	0.000	10:01
Control Test	0.078	10:02
Air Blank	0.000	10:02
Diagnostics Check	OK	10:02

Data Download Date/Time: 09/20/2017 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ROE,ROBERTY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 08/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A3
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: COURT ORD
ARREST TIME: 12:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:20

SUBJECT'S LAST NAME: WARD

DATE/TIME OF TEST: 08/28/2017 12:59:51

Diagnostics Check	OK	12:53
Air Blank	0.000	12:53
Control Test	0.077	12:54
Air Blank	0.000	12:54
Subject Sample # 1	0.000	12:55
Breath Volume	3.785	12:55
Air Blank	0.000	12:55
Air Blank	0.000	12:57
Subject Sample # 2	0.000	12:58
Breath Volume	2.406	12:58
Air Blank	0.000	12:58
Control Test	0.077	12:59
Air Blank	0.000	12:59
Diagnostics Check	OK	12:59

Data Download Date/Time: 09/20/2017 14:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: WEBB

DATE/TIME OF TEST: 08/04/2017 02:26:13

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:20
Air Blank	0.000	02:21
Subject Sample # 1	0.157	02:21
Breath Volume	2.292	02:21
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.153	02:24
Breath Volume	2.113	02:24
Air Blank	0.000	02:25
Control Test	0.080	02:25
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 08/30/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASSILAKIS,GEORGE,E
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ABRAM,JUSTIN
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: WITTE

DATE/TIME OF TEST: 08/04/2017 03:06:50

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.080	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.164	03:02
Breath Volume		2.007	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.150	03:05
Breath Volume		1.179	03:05
Air Blank		0.000	03:05
Control Test		0.078	03:06
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 08/30/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: PATERNO

DATE/TIME OF TEST: 08/09/2017 20:17:53

Diagnostics Check	OK	20:03	
Air Blank	0.000	20:03	
Control Test	0.080	20:04	
Air Blank	0.000	20:04	
Subject Sample	# 1 0.234	20:06	
Breath Volume	1.250	20:06	
Air Blank	0.000	20:07	
Air Blank	0.000	20:09	
Subject Sample	# 2 0.203	20:12	Volume Not Met
Breath Volume	0.031	20:12	
Air Blank	0.000	20:13	
Air Blank	0.000	20:14	
Subject Sample	# 3 0.230	20:16	
Breath Volume	1.386	20:16	
Air Blank	0.000	20:16	
Control Test	0.077	20:17	
Air Blank	0.000	20:17	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 08/30/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 08/13/2017 02:45:26

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.080	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.110	02:41
Breath Volume	2.464	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.111	02:43
Breath Volume	2.378	02:43
Air Blank	0.000	02:44
Control Test	0.079	02:44
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 08/30/2017 11:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: CROWE

DATE/TIME OF TEST: 08/03/2017 03:37:53

Diagnostics Check	OK	03:34	
Air Blank	0.000	03:35	
Control Test	0.079	03:35	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:36	Subject Test Refused
Breath Volume	0.000	03:36	
Air Blank	0.000	03:36	
Control Test	0.079	03:37	
Air Blank	0.000	03:37	
Diagnostics Check	OK	03:37	

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:58

SUBJECT'S LAST NAME: TISSUE

DATE/TIME OF TEST: 08/03/2017 06:25:27

Diagnostics Check		OK	06:19
Air Blank		0.000	06:19
Control Test		0.079	06:19
Air Blank		0.000	06:20
Subject Sample	# 1	0.034	06:20
Breath Volume		2.601	06:20
Air Blank		0.000	06:21
Air Blank		0.000	06:23
Subject Sample	# 2	0.034	06:23
Breath Volume		2.667	06:23
Air Blank		0.000	06:24
Control Test		0.079	06:24
Air Blank		0.000	06:25
Diagnostics Check		OK	06:25

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON, BRADEN, J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF, J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: SOULBY

DATE/TIME OF TEST: 08/07/2017 02:56:24

Diagnostics Check	OK	02:53	
Air Blank	0.000	02:54	
Control Test	0.079	02:54	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:55	Subject Test Refused
Breath Volume	0.000	02:55	
Air Blank	0.000	02:55	
Control Test	0.079	02:55	
Air Blank	0.000	02:56	
Diagnostics Check	OK	02:56	

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPICER,LORETHA,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DUNLAP,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:40

SUBJECT'S LAST NAME: AGUIRRE

DATE/TIME OF TEST: 08/10/2017 09:07:18

Diagnostics Check		OK	09:01
Air Blank		0.000	09:01
Control Test		0.078	09:01
Air Blank		0.000	09:02
Subject Sample	# 1	0.083	09:02
Breath Volume		2.019	09:02
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.081	09:05
Breath Volume		1.519	09:05
Air Blank		0.000	09:06
Control Test		0.078	09:06
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/11/2017 00:47:09

Diagnostics Check	OK	00:44	
Air Blank	0.000	00:44	
Control Test	0.079	00:45	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.078	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MORROBEL CRUZ

DATE/TIME OF TEST: 08/11/2017 03:14:44

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.079	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.170	03:10
Breath Volume	1.585	03:10
Air Blank	0.000	03:10
Air Blank	0.000	03:12
Subject Sample # 2	0.173	03:13
Breath Volume	1.285	03:13
Air Blank	0.000	03:13
Control Test	0.077	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: KHUJA

DATE/TIME OF TEST: 08/11/2017 04:08:48

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:06	
Control Test	0.079	04:06	
Air Blank	0.000	04:07	
Subject Sample	# 1	04:07	Subject Test Refused
Breath Volume	0.000	04:07	
Air Blank	0.000	04:07	
Control Test	0.079	04:08	
Air Blank	0.000	04:08	
Diagnostics Check	OK	04:08	

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J.M.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: CONSTANTINO

DATE/TIME OF TEST: 08/14/2017 01:13:18

Diagnostics Check	OK	01:06
Air Blank	0.000	01:07
Control Test	0.079	01:07
Air Blank	0.000	01:07
Subject Sample # 1	0.192	01:08
Breath Volume	2.777	01:08
Air Blank	0.000	01:09
Air Blank	0.000	01:10
Subject Sample # 2	0.193	01:11
Breath Volume	2.382	01:11
Air Blank	0.000	01:12
Control Test	0.078	01:12
Air Blank	0.000	01:13
Diagnostics Check	OK	01:13

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TOMAS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: SUNDEEN

DATE/TIME OF TEST: 08/14/2017 03:19:33

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.079	03:14
Air Blank	0.000	03:14
Subject Sample # 1	0.075	03:14
Breath Volume	2.671	03:14
Air Blank	0.000	03:15
Air Blank	0.000	03:17
Subject Sample # 2	0.074	03:17
Breath Volume	2.773	03:17
Air Blank	0.000	03:18
Control Test	0.079	03:18
Air Blank	0.000	03:19
Diagnostics Check	OK	03:19

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T.L.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/14/2017 04:12:09

Diagnostics Check	OK	04:05
Air Blank	0.000	04:06
Control Test	0.079	04:06
Air Blank	0.000	04:07
Subject Sample # 1	0.000	04:07
Breath Volume	3.429	04:07
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.000	04:10
Breath Volume	1.773	04:10
Air Blank	0.000	04:11
Control Test	0.079	04:11
Air Blank	0.000	04:12
Diagnostics Check	OK	04:12

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RANGER,R.O.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:02

SUBJECT'S LAST NAME: ALQANAS

DATE/TIME OF TEST: 08/14/2017 06:29:06

Diagnostics Check	OK	06:22
Air Blank	0.000	06:23
Control Test	0.079	06:23
Air Blank	0.000	06:23
Subject Sample # 1	0.157	06:24
Breath Volume	2.835	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:26
Subject Sample # 2	0.149	06:27
Breath Volume	1.734	06:27
Air Blank	0.000	06:28
Control Test	0.079	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T.L.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: HAYES

DATE/TIME OF TEST: 08/15/2017 03:39:40

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.079	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.119	03:35
Breath Volume	1.457	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.118	03:38
Breath Volume	1.304	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 08/16/2017 12:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: STRAND

DATE/TIME OF TEST: 08/18/2017 01:48:41

Diagnostics Check	OK	01:45	
Air Blank	0.000	01:46	
Control Test	0.078	01:46	
Air Blank	0.000	01:47	
Subject Sample	# 1	01:47	Subject Test Refused
Breath Volume	0.000	01:47	
Air Blank	0.000	01:47	
Control Test	0.078	01:48	
Air Blank	0.000	01:48	
Diagnostics Check	OK	01:48	

Data Download Date/Time: 08/29/2017 12:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 08/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: DETTMER

DATE/TIME OF TEST: 08/21/2017 03:06:33

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:01
Subject Sample # 1	0.228	03:01
Breath Volume	3.089	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.227	03:04
Breath Volume	4.437	03:04
Air Blank	0.000	03:05
Control Test	0.078	03:05
Air Blank	0.000	03:06
Diagnostics Check	OK	03:06

Data Download Date/Time: 08/29/2017 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KEEN,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 08/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: FOLLAIN

DATE/TIME OF TEST: 08/28/2017 02:26:10

Diagnostics Check	OK	02:23	
Air Blank	0.000	02:23	
Control Test	0.079	02:24	
Air Blank	0.000	02:24	
Subject Sample	# 1	02:24	Subject Test Refused
Breath Volume	0.000	02:24	
Air Blank	0.000	02:25	
Control Test	0.078	02:25	
Air Blank	0.000	02:26	
Diagnostics Check	OK	02:26	

Data Download Date/Time: 08/29/2017 12:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PINZON,J
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MCGEE

DATE/TIME OF TEST: 08/01/2017 00:15:44

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.078	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.100	00:11
Breath Volume	2.503	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.101	00:14
Breath Volume	2.953	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PRESUME,JEHUD
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SALESMAN

DATE/TIME OF TEST: 08/04/2017 03:33:29

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.079	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.135	03:27
Breath Volume	1.164	03:27
Air Blank	0.000	03:28
Air Blank	0.000	03:29
Subject Sample # 2	0.134	03:32
Breath Volume	1.121	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: RAND

DATE/TIME OF TEST: 08/04/2017 19:22:38

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.078	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.239	19:18
Breath Volume	1.488	19:18
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.250	19:21
Breath Volume	2.167	19:21
Air Blank	0.000	19:21
Control Test	0.078	19:22
Air Blank	0.000	19:22
Diagnostics Check	OK	19:22

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/05/2017 00:27:13

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.078	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:23	No Sample Provided
Breath Volume	0.000	00:23	
Air Blank	0.000	00:23	
Air Blank	0.000	00:25	
Subject Sample	# 2	00:25	Subject Test Refused
Breath Volume	0.000	00:25	
Air Blank	0.000	00:26	
Control Test	0.078	00:26	
Air Blank	0.000	00:27	
Diagnostics Check	OK	00:27	

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 08/05/2017 02:56:16

Diagnostics Check	OK	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Subject Sample # 1	0.266	02:50
Breath Volume	1.730	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.261	02:54
Breath Volume	1.695	02:54
Air Blank	0.000	02:55
Control Test	0.078	02:55
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARIELLO,JOSEPH
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/06/2017 04:20:51

Diagnostics Check	OK	04:05	
Air Blank	0.000	04:05	
Control Test	0.078	04:05	
Air Blank	0.000	04:06	
Subject Sample	# 1	04:09	No Sample Provided
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 2	0.146	04:13
Breath Volume	1.117	04:13	
Air Blank	0.000	04:14	
Air Blank	0.000	04:16	
Subject Sample	# 3	0.149	04:19
Breath Volume	0.136	04:19	Volume Not Met
Air Blank	0.000	04:19	
Control Test	0.077	04:20	
Air Blank	0.000	04:20	
Diagnostics Check	OK	04:20	

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FARIELLO,JOSEPH
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 08/06/2017 04:30:15

Diagnostics Check	OK	04:23
Air Blank	0.000	04:24
Control Test	0.078	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.155	04:25
Breath Volume	1.464	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.155	04:28
Breath Volume	1.457	04:28
Air Blank	0.000	04:29
Control Test	0.078	04:29
Air Blank	0.000	04:30
Diagnostics Check	OK	04:30

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KAY,KYLE
ARRESTING OFFICER AGENCY: Lighthouse Point PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:57

SUBJECT'S LAST NAME: MORROW

DATE/TIME OF TEST: 08/06/2017 22:23:58

Diagnostics Check	OK	22:18
Air Blank	0.000	22:18
Control Test	0.078	22:18
Air Blank	0.000	22:19
Subject Sample # 1	0.298	22:19
Breath Volume	1.507	22:19
Air Blank	0.000	22:20
Air Blank	0.000	22:22
Subject Sample # 2	0.297	22:22
Breath Volume	1.210	22:22
Air Blank	0.000	22:23
Control Test	0.079	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BERMAN,MARC
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: PONCE DE LEON

DATE/TIME OF TEST: 08/06/2017 23:19:12

Diagnostics Check	OK	23:10
Air Blank	0.000	23:10
Control Test	0.079	23:10
Air Blank	0.000	23:11
Subject Sample # 1	0.009	23:13
Breath Volume	1.699	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.007	23:17
Breath Volume	1.203	23:17
Air Blank	0.000	23:18
Control Test	0.078	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: ALMEIDA

DATE/TIME OF TEST: 08/10/2017 21:45:19

Diagnostics Check	OK	21:39
Air Blank	0.000	21:39
Control Test	0.080	21:40
Air Blank	0.000	21:40
Subject Sample # 1	0.000	21:41
Breath Volume	1.570	21:41
Air Blank	0.000	21:41
Air Blank	0.000	21:43
Subject Sample # 2	0.000	21:43
Breath Volume	1.542	21:43
Air Blank	0.000	21:44
Control Test	0.079	21:44
Air Blank	0.000	21:45
Diagnostics Check	OK	21:45

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WILLIAMS,SEAN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/12/2017 00:41:54

Diagnostics Check	OK	00:33
Air Blank	0.000	00:33
Control Test	0.079	00:34
Air Blank	0.000	00:34
Subject Sample # 1	0.208	00:36
Breath Volume	2.292	00:36
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 2	0.196	00:40
Breath Volume	2.156	00:40
Air Blank	0.000	00:41
Control Test	0.079	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PLANTE,MILTON
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: VELOZ

DATE/TIME OF TEST: 08/13/2017 02:33:15

Diagnostics Check	OK	02:26
Air Blank	0.000	02:27
Control Test	0.079	02:27
Air Blank	0.000	02:28
Subject Sample # 1	0.163	02:28
Breath Volume	3.195	02:28
Air Blank	0.000	02:29
Air Blank	0.000	02:31
Subject Sample # 2	0.164	02:31
Breath Volume	1.953	02:31
Air Blank	0.000	02:32
Control Test	0.079	02:32
Air Blank	0.000	02:33
Diagnostics Check	OK	02:33

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:27

SUBJECT'S LAST NAME: CASTRO HERNANDEZ

DATE/TIME OF TEST: 08/13/2017 04:55:53

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.079	04:48
Air Blank	0.000	04:48
Subject Sample # 1	0.211	04:49
Breath Volume	1.679	04:49
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.205	04:54
Breath Volume	1.402	04:54
Air Blank	0.000	04:54
Control Test	0.080	04:55
Air Blank	0.000	04:55
Diagnostics Check	OK	04:55

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HENNESSY,JONATHAN
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:15

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 08/13/2017 07:46:23

Diagnostics Check	OK	07:38
Air Blank	0.000	07:39
Control Test	0.079	07:39
Air Blank	0.000	07:39
Subject Sample # 1	0.115	07:42
Breath Volume	1.222	07:42
Air Blank	0.000	07:42
Air Blank	0.000	07:44
Subject Sample # 2	0.115	07:44
Breath Volume	1.488	07:44
Air Blank	0.000	07:45
Control Test	0.078	07:45
Air Blank	0.000	07:46
Diagnostics Check	OK	07:46

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MATTHEWS,T
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: LARATRO

DATE/TIME OF TEST: 08/14/2017 01:04:42

Diagnostics Check		OK	00:58
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.174	01:00
Breath Volume		2.468	01:00
Air Blank		0.000	01:00
Air Blank		0.000	01:02
Subject Sample	# 2	0.169	01:03
Breath Volume		2.414	01:03
Air Blank		0.000	01:03
Control Test		0.078	01:04
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MALAVE,MICHELLE
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:34

SUBJECT'S LAST NAME: MELO

DATE/TIME OF TEST: 08/14/2017 02:09:26

Diagnostics Check	OK	02:01
Air Blank	0.000	02:01
Control Test	0.079	02:02
Air Blank	0.000	02:02
Subject Sample # 1	0.118	02:05
Breath Volume	3.683	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 2	0.103	02:07
Breath Volume	1.382	02:07
Air Blank	0.000	02:08
Control Test	0.079	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/18/2017 01:13:31

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Subject Sample	# 1	0.136	01:06	Volume Not Met
Breath Volume		0.804	01:06	
Air Blank		0.000	01:07	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.133	01:12	Volume Not Met
Breath Volume		0.968	01:12	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/18/2017 01:27:42

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.078	01:17	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:20	No Sample Provided
Breath Volume	0.000	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample	# 2	01:26	Volume Not Met
Breath Volume	0.328	01:26	
Air Blank	0.000	01:26	
Control Test	0.079	01:27	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALARCON,MARIO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: DEVITO

DATE/TIME OF TEST: 08/18/2017 01:57:05

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.078	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1	0.178	01:49	No .020 agreement
Breath Volume		1.160	01:49	
Air Blank		0.000	01:49	
Air Blank		0.000	01:51	
Subject Sample	# 2	0.203	01:52	No .020 agreement
Breath Volume		3.558	01:52	
Air Blank		0.000	01:53	
Air Blank		0.000	01:55	
Subject Sample	# 3	0.156	01:55	No .020 agreement
Breath Volume		1.785	01:55	
Air Blank		0.000	01:56	
Control Test		0.078	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 08/28/2017 15:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ALARCON,MARIO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: DEVITO

DATE/TIME OF TEST: 08/18/2017 02:24:53

Diagnostics Check	OK	02:17
Air Blank	0.000	02:18
Control Test	0.078	02:18
Air Blank	0.000	02:18
Subject Sample # 1	0.173	02:20
Breath Volume	3.226	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.173	02:23
Breath Volume	3.539	02:23
Air Blank	0.000	02:23
Control Test	0.077	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAMPTON,CHRIS
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: IRIAS

DATE/TIME OF TEST: 08/19/2017 04:59:22

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:48	
Control Test		0.080	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 1	0.173	04:52	Volume Not Met
Breath Volume		0.250	04:52	
Air Blank		0.000	04:52	
Air Blank		0.000	04:54	
Subject Sample	# 2	0.197	04:57	Volume Not Met
Breath Volume		0.281	04:57	
Air Blank		0.000	04:58	
Control Test		0.079	04:58	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HAMPTON,CHRIS
ARRESTING OFFICER AGENCY: Seminole Dept of LE
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:21

SUBJECT'S LAST NAME: IRIAS

DATE/TIME OF TEST: 08/19/2017 05:13:08

Diagnostics Check		OK	05:01	
Air Blank		0.000	05:02	
Control Test		0.078	05:02	
Air Blank		0.000	05:02	
Subject Sample	# 1	0.171	05:06	Volume Not Met
Breath Volume		0.679	05:06	
Air Blank		0.000	05:06	
Air Blank		0.000	05:08	
Subject Sample	# 2	0.190	05:11	Volume Not Met
Breath Volume		0.222	05:11	
Air Blank		0.000	05:12	
Control Test		0.078	05:12	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:08

SUBJECT'S LAST NAME: CORTEZ

DATE/TIME OF TEST: 08/20/2017 04:37:59

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.079	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.205	04:33
Breath Volume	1.542	04:33
Air Blank	0.000	04:33
Air Blank	0.000	04:35
Subject Sample # 2	0.193	04:36
Breath Volume	1.695	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 08/21/2017 00:23:26

Diagnostics Check	OK	00:17
Air Blank	0.000	00:17
Control Test	0.079	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.227	00:18
Breath Volume	1.851	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.210	00:21
Breath Volume	1.148	00:21
Air Blank	0.000	00:22
Control Test	0.079	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: DIAZ,DANNY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: SHERMAN

DATE/TIME OF TEST: 08/22/2017 03:05:45

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.079	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.190	03:01
Breath Volume	2.855	03:01
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.196	03:04
Breath Volume	2.886	03:04
Air Blank	0.000	03:04
Control Test	0.079	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 08/24/2017 01:57:30

Diagnostics Check	OK	01:50
Air Blank	0.000	01:51
Control Test	0.079	01:51
Air Blank	0.000	01:52
Subject Sample # 1	0.113	01:53
Breath Volume	2.089	01:53
Air Blank	0.000	01:53
Air Blank	0.000	01:55
Subject Sample # 2	0.111	01:56
Breath Volume	1.523	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:56
Air Blank	0.000	01:57
Diagnostics Check	OK	01:57

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: FRY-ELLIOTT

DATE/TIME OF TEST: 08/25/2017 04:04:51

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.181	04:00
Breath Volume	1.804	04:00
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.187	04:03
Breath Volume	1.250	04:03
Air Blank	0.000	04:03
Control Test	0.079	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, YVETTE, D
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: WOOLLEY, KYLE
ARRESTING OFFICER AGENCY: Margate PD
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:11

SUBJECT'S LAST NAME: CURRY

DATE/TIME OF TEST: 08/25/2017 14:38:35

Diagnostics Check	OK	14:32
Air Blank	0.000	14:32
Control Test	0.079	14:32
Air Blank	0.000	14:33
Subject Sample # 1	0.252	14:33
Breath Volume	2.875	14:33
Air Blank	0.000	14:34
Air Blank	0.000	14:36
Subject Sample # 2	0.241	14:37
Breath Volume	2.394	14:37
Air Blank	0.000	14:37
Control Test	0.079	14:38
Air Blank	0.000	14:38
Diagnostics Check	OK	14:38

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MEGGIOLARO

DATE/TIME OF TEST: 08/26/2017 01:49:27

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:44
Subject Sample	# 1	0.240	01:44
Breath Volume		1.925	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.247	01:47
Breath Volume		1.914	01:47
Air Blank		0.000	01:48
Control Test		0.078	01:48
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001047

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 07/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: ORELLANA REYES

DATE/TIME OF TEST: 08/27/2017 00:21:16

Diagnostics Check	OK	00:14
Air Blank	0.000	00:14
Control Test	0.079	00:15
Air Blank	0.000	00:15
Subject Sample # 1	0.151	00:16
Breath Volume	2.472	00:16
Air Blank	0.000	00:17
Air Blank	0.000	00:19
Subject Sample # 2	0.145	00:19
Breath Volume	1.832	00:19
Air Blank	0.000	00:20
Control Test	0.078	00:20
Air Blank	0.000	00:21
Diagnostics Check	OK	00:21

Data Download Date/Time: 08/28/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BURGE

DATE/TIME OF TEST: 08/01/2017 01:45:08

Diagnostics Check	OK	01:41	
Air Blank	0.000	01:41	
Control Test	0.081	01:42	
Air Blank	0.000	01:42	
Subject Sample	# 1	01:43	Subject Test Refused
Breath Volume	0.000	01:43	
Air Blank	0.000	01:44	
Control Test	0.080	01:44	
Air Blank	0.000	01:45	
Diagnostics Check	OK	01:45	

Data Download Date/Time: 08/29/2017 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:09

SUBJECT'S LAST NAME: BURGE

DATE/TIME OF TEST: 08/01/2017 01:50:32

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.080	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 08/29/2017 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/05/2017 06:29:05

Diagnostics Check	OK	06:22
Air Blank	0.000	06:23
Control Test	0.081	06:23
Air Blank	0.000	06:24
Subject Sample # 1	0.053	06:24
Breath Volume	3.011	06:24
Air Blank	0.000	06:25
Air Blank	0.000	06:27
Subject Sample # 2	0.054	06:27
Breath Volume	3.386	06:27
Air Blank	0.000	06:28
Control Test	0.081	06:28
Air Blank	0.000	06:29
Diagnostics Check	OK	06:29

Data Download Date/Time: 08/29/2017 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WAYE,MAKINA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 08/06/2017 05:22:17

Diagnostics Check	OK	05:16
Air Blank	0.000	05:16
Control Test	0.080	05:16
Air Blank	0.000	05:17
Subject Sample # 1	0.076	05:17
Breath Volume	3.285	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.075	05:20
Breath Volume	4.296	05:20
Air Blank	0.000	05:21
Control Test	0.081	05:21
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 08/29/2017 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: GIL PETUMA

DATE/TIME OF TEST: 08/06/2017 07:09:48

Diagnostics Check		OK	07:03
Air Blank		0.000	07:04
Control Test		0.081	07:04
Air Blank		0.000	07:04
Subject Sample	# 1	0.158	07:05
Breath Volume		1.808	07:05
Air Blank		0.000	07:06
Air Blank		0.000	07:07
Subject Sample	# 2	0.154	07:08
Breath Volume		1.453	07:08
Air Blank		0.000	07:08
Control Test		0.081	07:09
Air Blank		0.000	07:09
Diagnostics Check		OK	07:09

Data Download Date/Time: 08/29/2017 14:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHISNANT,JASON,C
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: NIEVES,ELLIOTT
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:01

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 08/15/2017 23:41:21

Diagnostics Check	OK	23:28	
Air Blank	0.000	23:29	
Control Test	0.081	23:29	
Air Blank	0.000	23:29	
Subject Sample	# 1	23:32	No Sample Provided
Breath Volume	0.000	23:32	
Air Blank	0.000	23:33	
Air Blank	0.000	23:35	
Subject Sample	# 2	0.161	23:36
Breath Volume	1.757	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 3	0.171	23:39
Breath Volume	3.667	23:39	
Air Blank	0.000	23:40	
Control Test	0.079	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 08/29/2017 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROADHURST,RYAN,E
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BULMAN,CAMERON
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 07/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BALAIS

DATE/TIME OF TEST: 08/18/2017 01:09:44

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.200	01:05
Breath Volume		1.695	01:05
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.193	01:08
Breath Volume		2.890	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 08/29/2017 14:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: YEAGER

DATE/TIME OF TEST: 08/03/2017 03:38:49

Diagnostics Check	OK	03:33
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.000	03:34
Breath Volume	2.769	03:34
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.000	03:37
Breath Volume	2.785	03:37
Air Blank	0.000	03:37
Control Test	0.079	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 08/05/2017 17:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: CAPHART

DATE/TIME OF TEST: 08/12/2017 01:46:05

Diagnostics Check		OK	01:37	
Air Blank		0.000	01:37	
Control Test		0.079	01:38	
Air Blank		0.000	01:38	
Subject Sample	# 1	0.186	01:41	Volume Not Met
Breath Volume		0.222	01:41	
Air Blank		0.000	01:42	
Air Blank		0.000	01:44	
Subject Sample	# 2		01:44	Subject Test Refused
Breath Volume		0.000	01:44	
Air Blank		0.000	01:45	
Control Test		0.077	01:45	
Air Blank		0.000	01:46	
Diagnostics Check		OK	01:46	

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: STALVEY

DATE/TIME OF TEST: 08/12/2017 02:45:08

Diagnostics Check	OK	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Subject Sample	# 1	02:43	Subject Test Refused
Breath Volume	0.000	02:43	
Air Blank	0.000	02:44	
Control Test	0.080	02:44	
Air Blank	0.000	02:45	
Diagnostics Check	OK	02:45	

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/13/2017 19:02:01

Diagnostics Check	OK	18:55
Air Blank	0.000	18:56
Control Test	0.081	18:56
Air Blank	0.000	18:57
Subject Sample # 1	0.000	18:57
Breath Volume	3.253	18:57
Air Blank	0.000	18:58
Air Blank	0.000	19:00
Subject Sample # 2	0.000	19:00
Breath Volume	2.855	19:00
Air Blank	0.000	19:01
Control Test	0.080	19:01
Air Blank	0.000	19:01
Diagnostics Check	OK	19:02

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KRONZ,BRIAN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:34

SUBJECT'S LAST NAME: CURTIS

DATE/TIME OF TEST: 08/13/2017 20:01:07

Diagnostics Check	OK	19:54
Air Blank	0.000	19:54
Control Test	0.079	19:54
Air Blank	0.000	19:55
Subject Sample # 1	0.269	19:56
Breath Volume	1.609	19:56
Air Blank	0.000	19:57
Air Blank	0.000	19:59
Subject Sample # 2	0.253	19:59
Breath Volume	1.121	19:59
Air Blank	0.000	20:00
Control Test	0.078	20:00
Air Blank	0.000	20:01
Diagnostics Check	OK	20:01

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ANAMUAH-MENSAH,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 08/15/2017 22:19:30

Diagnostics Check	OK	22:15	
Air Blank	0.000	22:16	
Control Test	0.079	22:16	
Air Blank	0.000	22:17	
Subject Sample	# 1	22:18	Subject Test Refused
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Control Test	0.079	22:18	
Air Blank	0.000	22:19	
Diagnostics Check	OK	22:19	

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ANAMUAH-MENSAH,A
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BERG

DATE/TIME OF TEST: 08/15/2017 22:27:55

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.080	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.296	22:23
Breath Volume	2.339	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:25
Subject Sample # 2	0.283	22:26
Breath Volume	1.441	22:26
Air Blank	0.000	22:27
Control Test	0.078	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: ETHERIDGE

DATE/TIME OF TEST: 08/18/2017 01:51:50

Diagnostics Check	OK	01:48	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:50	Subject Test Refused
Breath Volume	0.000	01:50	
Air Blank	0.000	01:50	
Control Test	0.079	01:51	
Air Blank	0.000	01:51	
Diagnostics Check	OK	01:51	

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA, VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: SLATER

DATE/TIME OF TEST: 08/18/2017 22:46:03

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.079	22:43	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:44	Subject Test Refused
Breath Volume	0.000	22:44	
Air Blank	0.000	22:45	
Control Test	0.079	22:45	
Air Blank	0.000	22:45	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAUSAM,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RAIMEY

DATE/TIME OF TEST: 08/19/2017 00:43:14

Diagnostics Check	OK	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.119	00:37
Breath Volume	1.210	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:40
Subject Sample # 2	0.120	00:41
Breath Volume	1.171	00:41
Air Blank	0.000	00:42
Control Test	0.079	00:42
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 08/19/2017 02:10:51

Diagnostics Check	OK	02:07	
Air Blank	0.000	02:08	
Control Test	0.079	02:08	
Air Blank	0.000	02:09	
Subject Sample	# 1	02:09	Subject Test Refused
Breath Volume	0.000	02:09	
Air Blank	0.000	02:09	
Control Test	0.078	02:10	
Air Blank	0.000	02:10	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAGER,DANIEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: GWOZDZ

DATE/TIME OF TEST: 08/27/2017 23:27:48

Diagnostics Check	OK	23:21
Air Blank	0.000	23:21
Control Test	0.079	23:22
Air Blank	0.000	23:22
Subject Sample # 1	0.183	23:23
Breath Volume	1.355	23:23
Air Blank	0.000	23:23
Air Blank	0.000	23:25
Subject Sample # 2	0.182	23:26
Breath Volume	1.695	23:26
Air Blank	0.000	23:26
Control Test	0.079	23:27
Air Blank	0.000	23:27
Diagnostics Check	OK	23:27

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES, MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: BAIETTO

DATE/TIME OF TEST: 08/30/2017 02:12:42

Diagnostics Check	OK	02:05
Air Blank	0.000	02:06
Control Test	0.080	02:06
Air Blank	0.000	02:07
Subject Sample # 1	0.195	02:08
Breath Volume	3.386	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.192	02:11
Breath Volume	3.238	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: LAMOUR,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: WOJCIECH

DATE/TIME OF TEST: 08/30/2017 19:06:05

Diagnostics Check	OK	19:03	
Air Blank	0.000	19:03	
Control Test	0.077	19:03	
Air Blank	0.000	19:04	
Subject Sample	# 1	19:04	Subject Test Refused
Breath Volume	0.000	19:04	
Air Blank	0.000	19:05	
Control Test	0.077	19:05	
Air Blank	0.000	19:06	
Diagnostics Check	OK	19:06	

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 08/31/2017 00:57:09

Diagnostics Check		OK	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:51
Subject Sample	# 1	0.206	00:52
Breath Volume		2.480	00:52
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.211	00:55
Breath Volume		2.468	00:55
Air Blank		0.000	00:56
Control Test		0.078	00:56
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 08/31/2017 01:06:16

Diagnostics Check	OK	00:59
Air Blank	0.000	01:00
Control Test	0.078	01:00
Air Blank	0.000	01:01
Subject Sample # 1	0.211	01:01
Breath Volume	2.812	01:01
Air Blank	0.000	01:02
Air Blank	0.000	01:04
Subject Sample # 2	0.206	01:04
Breath Volume	2.484	01:04
Air Blank	0.000	01:05
Control Test	0.079	01:05
Air Blank	0.000	01:06
Diagnostics Check	OK	01:06

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001051

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 08/31/2017 03:24:24

Diagnostics Check	OK	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.062	03:19
Breath Volume	1.500	03:19
Air Blank	0.000	03:19
Air Blank	0.000	03:21
Subject Sample # 2	0.063	03:22
Breath Volume	1.554	03:22
Air Blank	0.000	03:23
Control Test	0.078	03:23
Air Blank	0.000	03:24
Diagnostics Check	OK	03:24

Data Download Date/Time: 09/06/2017 18:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,R
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 08/01/2017 22:23:29

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.079	22:18
Air Blank	0.000	22:18
Subject Sample # 1	0.119	22:19
Breath Volume	1.585	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.122	22:22
Breath Volume	1.449	22:22
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 08/07/2017 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 18015080A2
DRY GAS EXPIRY DATE: 08/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 08/03/2017 02:21:21

Diagnostics Check	OK	02:09	
Air Blank	0.000	02:10	
Control Test	0.079	02:10	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:14	No Sample Provided
Breath Volume	0.000	02:14	
Air Blank	0.000	02:14	
Air Blank	0.000	02:16	
Subject Sample	# 2	02:19	No Sample Provided
Breath Volume	0.000	02:19	
Air Blank	0.000	02:20	
Control Test	0.079	02:20	
Air Blank	0.000	02:21	
Diagnostics Check	OK	02:21	

Data Download Date/Time: 08/07/2017 14:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JESSUP,BRIAN,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: JESSUP,BRIAN
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG530001
DRY GAS EXPIRY DATE: 10/27/2017
VIOLATION CODE: DUI
ARREST TIME: 08:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: QUINTERO

DATE/TIME OF TEST: 08/16/2017 09:08:06

Diagnostics Check		OK	09:02
Air Blank		0.000	09:02
Control Test		0.080	09:02
Air Blank		0.000	09:03
Subject Sample	# 1	0.218	09:03
Breath Volume		2.292	09:03
Air Blank		0.000	09:04
Air Blank		0.000	09:06
Subject Sample	# 2	0.227	09:06
Breath Volume		3.144	09:06
Air Blank		0.000	09:07
Control Test		0.080	09:07
Air Blank		0.000	09:08
Diagnostics Check		OK	09:08

Data Download Date/Time: 09/30/2017 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUEHRENBERG,ERIK,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: MARSECO,VINNY
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG530001
DRY GAS EXPIRY DATE: 10/27/2017
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 08/19/2017 18:40:33

Diagnostics Check	OK	18:31
Air Blank	0.000	18:31
Control Test	0.080	18:32
Air Blank	0.000	18:32
Subject Sample # 1	0.222	18:33
Breath Volume	2.628	18:33
Air Blank	0.000	18:33
Air Blank	0.000	18:35
Subject Sample # 2	0.245	18:36
Breath Volume	2.886	18:36
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 3	0.245	18:38
Breath Volume	2.082	18:38
Air Blank	0.000	18:39
Control Test	0.078	18:39
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 09/30/2017 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAINES,DAVID,K
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: STRUNK,JOHN
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG530001
DRY GAS EXPIRY DATE: 10/27/2017
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: TINOCO ACOSTA

DATE/TIME OF TEST: 08/20/2017 04:28:22

Diagnostics Check	OK	04:20
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Subject Sample # 1	0.133	04:23
Breath Volume	1.550	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.146	04:26
Breath Volume	2.117	04:26
Air Blank	0.000	04:27
Control Test	0.080	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 09/30/2017 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001056

INSTRUMENT REGISTERED TO: DAVIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STUEHRENBERG,ERIK,C
BREATH TEST OPERATOR AGENCY: Davie PD
ARRESTING OFFICER: HAINES,DAVID
ARRESTING OFFICER AGENCY: Davie PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG530001
DRY GAS EXPIRY DATE: 10/27/2017
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 08/21/2017 07:05:23

Diagnostics Check	OK	06:57
Air Blank	0.000	06:58
Control Test	0.080	06:58
Air Blank	0.000	06:59
Subject Sample # 1	0.287	07:00
Breath Volume	2.074	07:00
Air Blank	0.000	07:00
Air Blank	0.000	07:02
Subject Sample # 2	0.283	07:03
Breath Volume	1.621	07:03
Air Blank	0.000	07:04
Control Test	0.079	07:04
Air Blank	0.000	07:05
Diagnostics Check	OK	07:05

Data Download Date/Time: 09/30/2017 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: RUZHIN

DATE/TIME OF TEST: 08/05/2017 01:53:26

Diagnostics Check	OK	01:45
Air Blank	0.000	01:46
Control Test	0.077	01:46
Air Blank	0.000	01:46
Subject Sample # 1	0.046	01:48
Breath Volume	1.761	01:48
Air Blank	0.000	01:48
Air Blank	0.000	01:50
Subject Sample # 2	0.048	01:51
Breath Volume	1.191	01:51
Air Blank	0.000	01:52
Control Test	0.076	01:52
Air Blank	0.000	01:53
Diagnostics Check	OK	01:53

Data Download Date/Time: 08/29/2017 10:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ROJAS-ARRIAGA

DATE/TIME OF TEST: 08/18/2017 22:25:04

Diagnostics Check	OK	22:18	
Air Blank	0.000	22:19	
Control Test	0.078	22:19	
Air Blank	0.000	22:19	
Subject Sample	# 1 0.184	22:20	
Breath Volume	1.261	22:20	
Air Blank	0.000	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 2 0.177	22:23	
Breath Volume	1.667	22:23	
Air Blank	0.000	22:23	
Control Test	0.077	22:24	
Air Blank		22:24	RFI Detect
Air Blank	0.000	22:25	

Data Download Date/Time: 08/29/2017 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JASON,E
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: PAETSCH,JOSHUA
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: MILVID

DATE/TIME OF TEST: 08/04/2017 17:36:09

Diagnostics Check	OK	17:33	
Air Blank	0.000	17:33	
Control Test	0.080	17:33	
Air Blank	0.000	17:34	
Subject Sample	# 1	17:34	Subject Test Refused
Breath Volume	0.000	17:34	
Air Blank	0.000	17:35	
Control Test	0.079	17:35	
Air Blank	0.000	17:36	
Diagnostics Check	OK	17:36	

Data Download Date/Time: 08/18/2017 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,JUSTIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: SHEFFIELD,DEVIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/13/2017 04:01:54

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Subject Sample # 1	0.240	03:57
Breath Volume	2.160	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.228	04:00
Breath Volume	1.914	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 08/18/2017 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JORGE,L
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: HERNANDEZ,JORGE
ARRESTING OFFICER AGENCY: Fernandina Beach PD
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: LEHMAN

DATE/TIME OF TEST: 08/14/2017 00:55:37

Diagnostics Check	OK	00:49
Air Blank	0.000	00:50
Control Test	0.080	00:50
Air Blank	0.000	00:50
Subject Sample # 1	0.212	00:51
Breath Volume	3.269	00:51
Air Blank	0.000	00:51
Air Blank	0.000	00:53
Subject Sample # 2	0.204	00:54
Breath Volume	2.472	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Diagnostics Check	OK	00:55

Data Download Date/Time: 08/18/2017 19:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THOLL,K
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 07/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: RADER

DATE/TIME OF TEST: 08/14/2017 02:46:55

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.080	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.120	02:42
Breath Volume	3.968	02:42
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.120	02:45
Breath Volume	4.453	02:45
Air Blank	0.000	02:45
Control Test	0.081	02:46
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 08/18/2017 19:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULKEARNS, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: BALLARD, J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 08/27/2017 04:37:21

Diagnostics Check	OK	04:30
Air Blank	0.000	04:30
Control Test	0.081	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.120	04:32
Breath Volume	1.519	04:32
Air Blank	0.000	04:32
Air Blank	0.000	04:34
Subject Sample # 2	0.114	04:35
Breath Volume	1.121	04:35
Air Blank	0.000	04:36
Control Test	0.081	04:36
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 10/31/2017 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001059

INSTRUMENT REGISTERED TO: FERNANDINA BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULKEARNS, MICHAEL, S
BREATH TEST OPERATOR AGENCY: Fernandina Beach PD
ARRESTING OFFICER: THOLL, K
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 08/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 08/27/2017 05:12:17

Diagnostics Check	OK	05:09	
Air Blank	0.000	05:09	
Control Test	0.080	05:10	
Air Blank	0.000	05:10	
Subject Sample	# 1	05:10	Subject Test Refused
Breath Volume	0.000	05:10	
Air Blank	0.000	05:11	
Control Test	0.080	05:11	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 10/31/2017 13:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 07/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:53

SUBJECT'S LAST NAME: OSTHOFF

DATE/TIME OF TEST: 08/01/2017 20:17:15

Diagnostics Check	OK	20:14	
Air Blank	0.000	20:14	
Control Test	0.081	20:15	
Air Blank	0.000	20:15	
Subject Sample	# 1	20:15	Subject Test Refused
Breath Volume	0.000	20:15	
Air Blank	0.000	20:16	
Control Test	0.082	20:16	
Air Blank	0.000	20:17	
Diagnostics Check	OK	20:17	

Data Download Date/Time: 08/29/2017 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 07/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: GRIFFIS

DATE/TIME OF TEST: 08/06/2017 03:55:11

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.082	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.180	03:50
Breath Volume	1.593	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.180	03:53
Breath Volume	1.960	03:53
Air Blank	0.000	03:54
Control Test	0.082	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 08/29/2017 09:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONSALL,ROBERT,E
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 07/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/06/2017 22:35:25

Diagnostics Check	OK	22:32	
Air Blank	0.000	22:32	
Control Test	0.081	22:33	
Air Blank	0.000	22:33	
Subject Sample	# 1	22:33	Subject Test Refused
Breath Volume	0.000	22:33	
Air Blank	0.000	22:34	
Control Test	0.081	22:34	
Air Blank	0.000	22:35	
Diagnostics Check	OK	22:35	

Data Download Date/Time: 08/29/2017 09:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ZACHARY,B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: NIEVES,ARION
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 08/03/2017 00:43:38

Diagnostics Check	OK	00:36
Air Blank	0.000	00:36
Control Test	0.080	00:36
Air Blank	0.000	00:37
Subject Sample # 1	0.122	00:38
Breath Volume	1.195	00:38
Air Blank	0.000	00:39
Air Blank	0.000	00:41
Subject Sample # 2	0.119	00:42
Breath Volume	1.644	00:42
Air Blank	0.000	00:42
Control Test	0.077	00:43
Air Blank	0.000	00:43
Diagnostics Check	OK	00:43

Data Download Date/Time: 08/07/2017 16:32

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ZACHARY,B
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: MILLS,ZACHARY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A3
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: SARRA

DATE/TIME OF TEST: 08/06/2017 02:42:17

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:35	
Control Test	0.080	02:35	
Air Blank	0.000	02:36	
Subject Sample	# 1 0.102	02:38	
Breath Volume	1.578	02:38	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2	02:40	Subject Test Refused
Breath Volume	0.000	02:40	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 08/07/2017 16:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,KIMBERLY,L
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: MCDONALD,KIMBERLY
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 07/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PURINTON

DATE/TIME OF TEST: 08/20/2017 23:35:34

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.207	23:30
Breath Volume	3.835	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.196	23:34
Breath Volume	2.148	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 08/25/2017 11:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,HECTOR,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: HOLM,MATHEW
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 08/28/2017 03:19:59

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.080	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:19	
Control Test	0.079	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 09/20/2017 15:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DERIET,RAFAEL,
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: DERIET,RAFAEL
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: PEREZ ALVAREZ

DATE/TIME OF TEST: 08/12/2017 04:51:32

Diagnostics Check		OK	04:45
Air Blank		0.000	04:45
Control Test		0.077	04:46
Air Blank		0.000	04:46
Subject Sample	# 1	0.200	04:47
Breath Volume		2.429	04:47
Air Blank		0.000	04:47
Air Blank		0.000	04:49
Subject Sample	# 2	0.193	04:50
Breath Volume		2.117	04:50
Air Blank		0.000	04:50
Control Test		0.076	04:50
Air Blank		0.000	04:51
Diagnostics Check		OK	04:51

Data Download Date/Time: 08/27/2017 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001063

INSTRUMENT REGISTERED TO: HOLLYWOOD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,MELANIE,A
BREATH TEST OPERATOR AGENCY: Hollywood PD
ARRESTING OFFICER: GARCIA,MELANIE
ARRESTING OFFICER AGENCY: Hollywood PD
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 08/29/2017 23:44:52

Diagnostics Check	OK	23:35	
Air Blank	0.000	23:36	
Control Test	0.077	23:36	
Air Blank	0.000	23:37	
Subject Sample	# 1	23:40	No Sample Provided
Breath Volume	0.000	23:40	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 2	23:43	Subject Test Refused
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Control Test	0.076	23:44	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 09/28/2017 11:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGEL,JASON,S
BREATH TEST OPERATOR AGENCY: New Port Richey PD
ARRESTING OFFICER: ENGEL,CASEY
ARRESTING OFFICER AGENCY: New Port Richey PD
LAST AGENCY INSPECTION: 07/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:44

SUBJECT'S LAST NAME: GASKINS

DATE/TIME OF TEST: 08/27/2017 04:18:13

Diagnostics Check	OK	04:06	
Air Blank	0.000	04:07	
Control Test	0.078	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2	04:16	No Sample Provided
Breath Volume	0.000	04:16	
Air Blank	0.000	04:17	
Control Test	0.079	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 08/30/2017 22:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROQUE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: ROQUE,MICHAEL
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: NEWKIRK

DATE/TIME OF TEST: 08/14/2017 17:01:05

Diagnostics Check	OK	16:58	
Air Blank	0.000	16:58	
Control Test	0.080	16:59	
Air Blank	0.000	16:59	
Subject Sample	# 1	16:59	Subject Test Refused
Breath Volume	0.000	16:59	
Air Blank	0.000	17:00	
Control Test	0.080	17:00	
Air Blank	0.000	17:01	
Diagnostics Check	OK	17:01	

Data Download Date/Time: 08/28/2017 13:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RONDOS,ROBERT,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: RONDOS,ROBERT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 08/31/2017 11:06:49

Diagnostics Check	OK	11:00
Air Blank	0.000	11:00
Control Test	0.080	11:01
Air Blank	0.000	11:01
Subject Sample # 1	0.000	11:02
Breath Volume	1.804	11:02
Air Blank	0.000	11:03
Air Blank	0.000	11:05
Subject Sample # 2	0.000	11:05
Breath Volume	1.578	11:05
Air Blank	0.000	11:05
Control Test	0.080	11:06
Air Blank	0.000	11:06
Diagnostics Check	OK	11:06

Data Download Date/Time: 09/29/2017 16:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 07/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: SEIBERT

DATE/TIME OF TEST: 08/05/2017 04:04:44

Diagnostics Check	OK	03:58
Air Blank	0.000	03:59
Control Test	0.080	03:59
Air Blank	0.000	04:00
Subject Sample # 1	0.000	04:00
Breath Volume	1.996	04:00
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.000	04:03
Breath Volume	1.746	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 08/10/2017 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEANE,M,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEFFINGWELL,J
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 08/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG515306
DRY GAS EXPIRY DATE: 06/02/2017
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: HENDRICKSON

DATE/TIME OF TEST: 08/27/2017 02:57:25

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:55	
Control Test	0.081	02:55	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:56	Subject Test Refused
Breath Volume	0.000	02:56	
Air Blank	0.000	02:56	
Control Test	0.080	02:56	
Air Blank	0.000	02:57	
Diagnostics Check	OK	02:57	

Data Download Date/Time: 09/06/2017 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,D,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,D
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 08/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: RUKAVINA

DATE/TIME OF TEST: 08/31/2017 21:20:12

Diagnostics Check	OK	21:13
Air Blank	0.000	21:14
Control Test	0.080	21:14
Air Blank	0.000	21:14
Subject Sample # 1	0.000	21:15
Breath Volume	2.457	21:15
Air Blank	0.000	21:16
Air Blank	0.000	21:18
Subject Sample # 2	0.000	21:18
Breath Volume	2.113	21:18
Air Blank	0.000	21:19
Control Test	0.079	21:19
Air Blank	0.000	21:20
Diagnostics Check	OK	21:20

Data Download Date/Time: 09/25/2017 01:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: WILLIAMSON

DATE/TIME OF TEST: 08/03/2017 05:01:32

Diagnostics Check	OK	04:53
Air Blank	0.000	04:54
Control Test	0.080	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.157	04:56
Breath Volume	1.648	04:56
Air Blank	0.000	04:57
Air Blank	0.000	04:58
Subject Sample # 2	0.155	05:00
Breath Volume	1.757	05:00
Air Blank	0.000	05:00
Control Test	0.079	05:00
Air Blank	0.000	05:01
Diagnostics Check	OK	05:01

Data Download Date/Time: 08/22/2017 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 08/17/2017 00:09:14

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.079	00:03
Air Blank		0.000	00:03
Subject Sample	# 1	0.098	00:04
Breath Volume		2.406	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.097	00:07
Breath Volume		2.175	00:07
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 08/22/2017 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TORRES,P
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: HERNANDEZ-LOPEZ

DATE/TIME OF TEST: 08/18/2017 00:24:17

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.079	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.176	00:19
Breath Volume	2.507	00:19
Air Blank	0.000	00:20
Air Blank	0.000	00:22
Subject Sample # 2	0.182	00:22
Breath Volume	2.945	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:24
Diagnostics Check	OK	00:24

Data Download Date/Time: 08/22/2017 17:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 07/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: KNEPPER

DATE/TIME OF TEST: 08/21/2017 23:35:40

Diagnostics Check	OK	23:24
Air Blank	0.000	23:25
Control Test	0.079	23:25
Air Blank	0.000	23:26
Subject Sample # 1	0.249	23:27
Breath Volume	2.214	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:29
Subject Sample # 2	0.228	23:30
Breath Volume	2.328	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 3	0.230	23:34
Breath Volume	2.144	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 08/22/2017 17:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: GOLDSTEIN

DATE/TIME OF TEST: 08/23/2017 01:59:05

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.165	01:54
Breath Volume	2.167	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.167	01:57
Breath Volume	2.960	01:57
Air Blank	0.000	01:58
Control Test	0.080	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 09/24/2017 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TORRES,P
ARRESTING OFFICER AGENCY: Green Cove Springs
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BECKMANN

DATE/TIME OF TEST: 08/23/2017 02:47:09

Diagnostics Check	OK	02:35	
Air Blank	0.000	02:36	
Control Test	0.079	02:36	
Air Blank	0.000	02:37	
Subject Sample	# 1	02:40	No Sample Provided
Breath Volume	0.000	02:40	
Air Blank	0.000	02:40	
Air Blank	0.000	02:42	
Subject Sample	# 2	02:45	No Sample Provided
Breath Volume	0.000	02:45	
Air Blank	0.000	02:46	
Control Test	0.080	02:46	
Air Blank	0.000	02:47	
Diagnostics Check	OK	02:47	

Data Download Date/Time: 09/24/2017 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SWIM,G
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: POULIOT

DATE/TIME OF TEST: 08/31/2017 01:23:15

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.079	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.272	01:18
Breath Volume		2.476	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.266	01:21
Breath Volume		1.964	01:21
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 09/24/2017 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CHESTER

DATE/TIME OF TEST: 08/31/2017 02:11:40

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.080	02:05
Air Blank	0.000	02:06
Subject Sample # 1	0.078	02:07
Breath Volume	1.460	02:07
Air Blank	0.000	02:07
Air Blank	0.000	02:09
Subject Sample # 2	0.074	02:10
Breath Volume	1.781	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 09/24/2017 13:32

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: SCHRIDER JARA

DATE/TIME OF TEST: 08/05/2017 23:20:25

Diagnostics Check	OK	23:16	
Air Blank	0.000	23:17	
Control Test	0.078	23:17	
Air Blank	0.000	23:17	
Subject Sample	# 1	23:19	Subject Test Refused
Breath Volume	0.000	23:19	
Air Blank	0.000	23:19	
Control Test	0.079	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 08/23/2017 13:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN, WILLIAM, E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: OATES

DATE/TIME OF TEST: 08/06/2017 03:49:43

Diagnostics Check	OK	03:41	
Air Blank	0.000	03:41	
Control Test	0.078	03:41	
Air Blank	0.000	03:42	
Subject Sample	# 1	03:45	No Sample Provided
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Air Blank	0.000	03:47	
Subject Sample	# 2	03:48	Subject Test Refused
Breath Volume	0.000	03:48	
Air Blank	0.000	03:48	
Control Test	0.077	03:49	
Air Blank	0.000	03:49	
Diagnostics Check	OK	03:49	

Data Download Date/Time: 08/23/2017 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:01

SUBJECT'S LAST NAME: KOGGAN

DATE/TIME OF TEST: 08/12/2017 06:29:03

Diagnostics Check		OK	06:21
Air Blank		0.000	06:22
Control Test		0.077	06:22
Air Blank		0.000	06:23
Subject Sample	# 1	0.164	06:24
Breath Volume		4.210	06:24
Air Blank		0.000	06:24
Air Blank		0.000	06:26
Subject Sample	# 2	0.153	06:27
Breath Volume		3.652	06:27
Air Blank		0.000	06:28
Control Test		0.076	06:28
Air Blank		0.000	06:28
Diagnostics Check		OK	06:29

Data Download Date/Time: 08/23/2017 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEQUEIRA,NATHAN,C
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SEQUEIRA,NATHAN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 08:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BECK

DATE/TIME OF TEST: 08/12/2017 21:49:35

Diagnostics Check	OK	21:46	
Air Blank	0.000	21:47	
Control Test	0.078	21:47	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:48	
Control Test	0.078	21:48	
Air Blank	0.000	21:49	
Diagnostics Check	OK	21:49	

Data Download Date/Time: 08/23/2017 13:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,LANE,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: LLORENS

DATE/TIME OF TEST: 08/16/2017 02:24:56

Diagnostics Check	OK	02:18
Air Blank	0.000	02:19
Control Test	0.078	02:19
Air Blank	0.000	02:20
Subject Sample # 1	0.000	02:20
Breath Volume	2.875	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.000	02:23
Breath Volume	3.050	02:23
Air Blank	0.000	02:24
Control Test	0.078	02:24
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 08/23/2017 13:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIVERO,JOSE,I

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: RIVERO,JOSE

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 23:38

GENDER: Female

OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VALDES

DATE/TIME OF TEST: 08/05/2017 01:21:32

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:19	
Control Test	0.080	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:20	Subject Test Refused
Breath Volume	0.000	01:20	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:21	
Diagnostics Check	OK	01:21	

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ANGEL,G
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HERNANDEZ,ANGEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 07:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:34

SUBJECT'S LAST NAME: PEREZSANCHEZ

DATE/TIME OF TEST: 08/06/2017 08:22:42

Diagnostics Check	OK	08:19	
Air Blank	0.000	08:20	
Control Test	0.079	08:20	
Air Blank	0.000	08:21	
Subject Sample	# 1	08:21	Subject Test Refused
Breath Volume	0.000	08:21	
Air Blank	0.000	08:21	
Control Test	0.080	08:22	
Air Blank	0.000	08:22	
Diagnostics Check	OK	08:22	

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIVERO,JOSE,I

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: RIVERO,JOSE

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 02:21

GENDER: Male

OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: TAVERAS

DATE/TIME OF TEST: 08/09/2017 03:34:46

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:32	
Control Test	0.079	03:32	
Air Blank	0.000	03:33	
Subject Sample	# 1	03:33	Subject Test Refused
Breath Volume	0.000	03:33	
Air Blank	0.000	03:33	
Control Test	0.080	03:34	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: ORBETA,JORGE,S

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: ORBETA,JORGE

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 06:54

GENDER: Male

OBSERVATION PERIOD BEGAN: 07:28

SUBJECT'S LAST NAME: SERRATO

DATE/TIME OF TEST: 08/09/2017 07:56:59

Diagnostics Check		OK	07:49
Air Blank		0.000	07:50
Control Test		0.080	07:50
Air Blank		0.000	07:51
Subject Sample	# 1	0.224	07:52
Breath Volume		1.332	07:52
Air Blank		0.000	07:53
Air Blank		0.000	07:54
Subject Sample	# 2	0.227	07:55
Breath Volume		1.820	07:55
Air Blank		0.000	07:56
Control Test		0.080	07:56
Air Blank		0.000	07:56
Diagnostics Check		OK	07:56

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ BORGES,JORGE,Y
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GRIEGO,CHARLENE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:40

SUBJECT'S LAST NAME: LOPEZ BORGES

DATE/TIME OF TEST: 08/11/2017 07:58:29

Diagnostics Check	OK	07:52
Air Blank	0.000	07:52
Control Test	0.080	07:52
Air Blank	0.000	07:53
Subject Sample # 1	0.197	07:54
Breath Volume	2.203	07:54
Air Blank	0.000	07:54
Air Blank	0.000	07:56
Subject Sample # 2	0.198	07:56
Breath Volume	1.843	07:56
Air Blank	0.000	07:57
Control Test	0.080	07:57
Air Blank	0.000	07:58
Diagnostics Check	OK	07:58

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIVERO,JOSE,I

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: RIVERO,JOSE

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 22:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: RAYMUNDO

DATE/TIME OF TEST: 08/11/2017 23:49:45

Diagnostics Check	OK	23:43
Air Blank	0.000	23:43
Control Test	0.079	23:44
Air Blank	0.000	23:44
Subject Sample # 1	0.213	23:45
Breath Volume	1.757	23:45
Air Blank	0.000	23:46
Air Blank	0.000	23:47
Subject Sample # 2	0.207	23:48
Breath Volume	1.644	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:49
Air Blank	0.000	23:49
Diagnostics Check	OK	23:49

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,TIMOTHY,J

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: BROWN,TIMOTHY

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 04:05

GENDER: Male

OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: VARGAS

DATE/TIME OF TEST: 08/18/2017 05:29:45

Diagnostics Check	OK	05:21
Air Blank	0.000	05:21
Control Test	0.077	05:22
Air Blank	0.000	05:22
Subject Sample # 1	0.127	05:24
Breath Volume	1.511	05:24
Air Blank	0.000	05:24
Air Blank	0.000	05:26
Subject Sample # 2	0.133	05:28
Breath Volume	1.484	05:28
Air Blank	0.000	05:28
Control Test	0.080	05:29
Air Blank	0.000	05:29
Diagnostics Check	OK	05:29

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HARO,EMILIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 07:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:27

SUBJECT'S LAST NAME: TANOS

DATE/TIME OF TEST: 08/19/2017 07:53:49

Diagnostics Check	OK	07:47
Air Blank	0.000	07:47
Control Test	0.080	07:48
Air Blank	0.000	07:48
Subject Sample # 1	0.170	07:49
Breath Volume	2.914	07:49
Air Blank	0.000	07:49
Air Blank	0.000	07:51
Subject Sample # 2	0.170	07:52
Breath Volume	2.609	07:52
Air Blank	0.000	07:52
Control Test	0.080	07:53
Air Blank	0.000	07:53
Diagnostics Check	OK	07:53

Data Download Date/Time: 08/25/2017 08:46

**Florida Department of Law Enforcement
Alcohol Testing Program**

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: CHAVEZ
DATE/TIME OF TEST: 08/19/2017 21:48:18

Diagnostics Check	OK	21:39	
Air Blank	0.000	21:40	
Control Test	0.080	21:40	
Air Blank	0.000	21:41	
Subject Sample	# 1	21:44	No Sample Provided
Breath Volume	0.000	21:44	
Air Blank	0.000	21:44	
Air Blank	0.000	21:46	
Subject Sample	# 2	21:46	Subject Test Refused
Breath Volume	0.000	21:46	
Air Blank	0.000	21:47	
Control Test	0.080	21:47	
Air Blank	0.000	21:48	
Diagnostics Check	OK	21:48	

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SAINTILIE,BOB,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: SAINTILIE,BOB

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 03:34

GENDER: Female

OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 08/20/2017 04:18:27

Diagnostics Check	OK	04:07	
Air Blank	0.000	04:07	
Control Test	0.080	04:07	
Air Blank	0.000	04:08	
Subject Sample	# 1	04:11	No Sample Provided
Breath Volume	0.000	04:11	
Air Blank	0.000	04:11	
Air Blank	0.000	04:13	
Subject Sample	# 2	04:17	No Sample Provided
Breath Volume	0.000	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:18	
Diagnostics Check	OK	04:18	

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: RIVERO,JOSE,I

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: RIVERO,JOSE

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 01:08

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BATISTA ALMAGUER

DATE/TIME OF TEST: 08/21/2017 02:07:32

Diagnostics Check	OK	02:04	
Air Blank	0.000	02:05	
Control Test	0.080	02:05	
Air Blank	0.000	02:05	
Subject Sample	# 1	02:06	Subject Test Refused
Breath Volume	0.000	02:06	
Air Blank	0.000	02:06	
Control Test	0.079	02:06	
Air Blank	0.000	02:07	
Diagnostics Check	OK	02:07	

Data Download Date/Time: 08/25/2017 08:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: SAINTILIEEN,BOB,

BREATH TEST OPERATOR AGENCY: FHP Troop E

ARRESTING OFFICER: SAINTILIEEN,BOB

ARRESTING OFFICER AGENCY: FHP Troop E

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 1824329

DRY GAS EXPIRY DATE: 03/09/2018

VIOLATION CODE: DUI

ARREST TIME: 23:37

GENDER: Female

OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 08/24/2017 01:05:14

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.000	01:03	
Air Blank	0.000	01:04	
Control Test	0.080	01:04	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 08/25/2017 08:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001075

INSTRUMENT REGISTERED TO: FHP TROOP E
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAVINON,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SAVINON,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MOISE

DATE/TIME OF TEST: 08/26/2017 02:16:57

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.079	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.164	02:12
Breath Volume	2.656	02:12
Air Blank	0.000	02:12
Air Blank	0.000	02:14
Subject Sample # 2	0.155	02:15
Breath Volume	1.851	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 09/29/2017 12:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,NATHAN,R
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: BUTERA,JOSEPH
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: SOUTHWICK

DATE/TIME OF TEST: 08/08/2017 23:30:00

Diagnostics Check	OK	23:23
Air Blank	0.000	23:24
Control Test	0.079	23:24
Air Blank	0.000	23:24
Subject Sample # 1	0.242	23:25
Breath Volume	2.617	23:25
Air Blank	0.000	23:26
Air Blank	0.000	23:27
Subject Sample # 2	0.251	23:28
Breath Volume	2.843	23:28
Air Blank	0.000	23:29
Control Test	0.078	23:29
Air Blank	0.000	23:29
Diagnostics Check	OK	23:30

Data Download Date/Time: 08/31/2017 03:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON, JOSHUA, A
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: GRIFFIN, RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 08/27/2017 04:48:48

Diagnostics Check	OK	04:45	
Air Blank	0.000	04:46	
Control Test	0.079	04:46	
Air Blank	0.000	04:47	
Subject Sample	# 1	04:47	Subject Test Refused
Breath Volume	0.000	04:47	
Air Blank	0.000	04:47	
Control Test	0.079	04:48	
Air Blank	0.000	04:48	
Diagnostics Check	OK	04:48	

Data Download Date/Time: 08/31/2017 03:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,NATHAN,R
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: FELIX,JOSE
ARRESTING OFFICER AGENCY: Dade City PD
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:38

SUBJECT'S LAST NAME: LOCASTO

DATE/TIME OF TEST: 08/30/2017 18:10:35

Diagnostics Check	OK	18:04
Air Blank	0.000	18:04
Control Test	0.078	18:05
Air Blank	0.000	18:05
Subject Sample # 1	0.164	18:06
Breath Volume	1.441	18:06
Air Blank	0.000	18:06
Air Blank	0.000	18:08
Subject Sample # 2	0.163	18:09
Breath Volume	1.449	18:09
Air Blank	0.000	18:09
Control Test	0.077	18:10
Air Blank	0.000	18:10
Diagnostics Check	OK	18:10

Data Download Date/Time: 08/31/2017 03:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 07/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 08/05/2017 02:21:54

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.079	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.179	02:11
Breath Volume		1.414	02:11
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 2	0.210	02:15
Breath Volume		2.707	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 3	0.184	02:20
Breath Volume		1.343	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 08/05/2017 17:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI, ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: GOODWIN

DATE/TIME OF TEST: 08/05/2017 21:31:55

Diagnostics Check	OK	21:25
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.000	21:27
Breath Volume	1.121	21:27
Air Blank	0.000	21:27
Air Blank	0.000	21:29
Subject Sample # 2	0.000	21:30
Breath Volume	2.183	21:30
Air Blank	0.000	21:30
Control Test	0.079	21:31
Air Blank	0.000	21:31
Diagnostics Check	OK	21:31

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: HEINKEL

DATE/TIME OF TEST: 08/19/2017 01:54:02

Diagnostics Check	OK	01:47
Air Blank	0.000	01:47
Control Test	0.080	01:48
Air Blank	0.000	01:48
Subject Sample # 1	0.175	01:49
Breath Volume	3.925	01:49
Air Blank	0.000	01:49
Air Blank	0.000	01:51
Subject Sample # 2	0.174	01:52
Breath Volume	3.906	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:53
Diagnostics Check	OK	01:54

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ELLIS,SKYLER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/19/2017 02:39:30

Diagnostics Check	OK	02:33
Air Blank	0.000	02:33
Control Test	0.079	02:34
Air Blank	0.000	02:34
Subject Sample # 1	0.184	02:35
Breath Volume	1.980	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.178	02:38
Breath Volume	1.750	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:38
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COPE,ALFRED,L
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 08/20/2017 00:51:01

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:39	
Control Test		0.075	00:39	
Air Blank		0.000	00:40	
Subject Sample	# 1	0.205	00:43	Volume Not Met
Breath Volume		0.015	00:43	
Air Blank		0.000	00:44	
Air Blank		0.000	00:46	
Subject Sample	# 2	0.177	00:49	Volume Not Met
Breath Volume		0.257	00:49	
Air Blank		0.000	00:50	
Control Test		0.077	00:50	
Air Blank		0.000	00:50	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 08/20/2017 01:08:49

Diagnostics Check		OK	00:57	
Air Blank		0.000	00:57	
Control Test		0.078	00:58	
Air Blank		0.000	00:58	
Subject Sample	# 1	0.216	01:01	Volume Not Met
Breath Volume		0.074	01:01	
Air Blank		0.000	01:02	
Air Blank		0.000	01:04	
Subject Sample	# 2	0.204	01:07	Volume Not Met
Breath Volume		0.160	01:07	
Air Blank		0.000	01:07	
Control Test		0.078	01:08	
Air Blank		0.000	01:08	
Diagnostics Check		OK	01:08	

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WEISKOPF,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: MCALLISTER

DATE/TIME OF TEST: 08/20/2017 01:20:02

Diagnostics Check		OK	01:11	
Air Blank		0.000	01:12	
Control Test		0.077	01:12	
Air Blank		0.000	01:12	
Subject Sample	# 1	0.144	01:16	Volume Not Met
Breath Volume		0.132	01:16	
Air Blank		0.000	01:16	
Air Blank		0.000	01:18	
Subject Sample	# 2		01:18	Subject Test Refused
Breath Volume		0.000	01:18	
Air Blank		0.000	01:19	
Control Test		0.079	01:19	
Air Blank		0.000	01:19	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 08/20/2017 03:20:16

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.080	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.101	03:15
Breath Volume	2.847	03:15
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.100	03:18
Breath Volume	2.929	03:18
Air Blank	0.000	03:19
Control Test	0.080	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/20/2017 21:09:25

Diagnostics Check	OK	21:08	
Air Blank		21:08	RFI Detect
Air Blank	0.000	21:09	

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTEN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 08/20/2017 21:35:06

Diagnostics Check	OK	21:28
Air Blank	0.000	21:29
Control Test	0.080	21:29
Air Blank	0.000	21:30
Subject Sample # 1	0.170	21:30
Breath Volume	1.589	21:30
Air Blank	0.000	21:31
Air Blank	0.000	21:33
Subject Sample # 2	0.174	21:33
Breath Volume	1.242	21:33
Air Blank	0.000	21:34
Control Test	0.080	21:34
Air Blank	0.000	21:35
Diagnostics Check	OK	21:35

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: HAUSAM,M
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:48

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 08/22/2017 23:18:25

Diagnostics Check	OK	23:08
Air Blank	0.000	23:08
Control Test	0.080	23:09
Air Blank	0.000	23:09
Subject Sample # 1	0.192	23:10
Breath Volume	1.714	23:10
Air Blank	0.000	23:11
Air Blank	0.000	23:13
Subject Sample # 2	0.219	23:14
Breath Volume	3.613	23:14
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 3	0.212	23:16
Breath Volume	2.562	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:18
Diagnostics Check	OK	23:18

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA, VLADAMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: BILSKY

DATE/TIME OF TEST: 08/24/2017 01:50:27

Diagnostics Check	OK	01:47	
Air Blank	0.000	01:48	
Control Test	0.080	01:48	
Air Blank	0.000	01:48	
Subject Sample	# 1	01:49	Subject Test Refused
Breath Volume	0.000	01:49	
Air Blank	0.000	01:49	
Control Test	0.079	01:49	
Air Blank	0.000	01:50	
Diagnostics Check	OK	01:50	

Data Download Date/Time: 09/06/2017 18:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: DOYLE

DATE/TIME OF TEST: 08/26/2017 22:08:43

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:06	
Control Test	0.080	22:06	
Air Blank	0.000	22:07	
Subject Sample	# 1	22:07	Subject Test Refused
Breath Volume	0.000	22:07	
Air Blank	0.000	22:07	
Control Test	0.079	22:08	
Air Blank	0.000	22:08	
Diagnostics Check	OK	22:08	

Data Download Date/Time: 09/06/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 08/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MIFFIN

DATE/TIME OF TEST: 08/29/2017 19:38:12

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:35	
Control Test	0.080	19:36	
Air Blank	0.000	19:36	
Subject Sample	# 1	19:36	Subject Test Refused
Breath Volume	0.000	19:36	
Air Blank	0.000	19:37	
Control Test	0.079	19:37	
Air Blank	0.000	19:38	
Diagnostics Check	OK	19:38	

Data Download Date/Time: 09/06/2017 18:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEASE, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: PEASE, MATTHEW
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: HARTMAN

DATE/TIME OF TEST: 08/24/2017 03:56:12

Diagnostics Check	OK	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.194	03:51
Breath Volume	2.593	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:53
Subject Sample # 2	0.191	03:54
Breath Volume	1.679	03:54
Air Blank	0.000	03:55
Control Test	0.077	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 09/27/2017 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI,ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:53

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/25/2017 16:29:40

Diagnostics Check	OK	16:23
Air Blank	0.000	16:23
Control Test	0.079	16:23
Air Blank	0.000	16:24
Subject Sample # 1	0.000	16:25
Breath Volume	2.175	16:25
Air Blank	0.000	16:25
Air Blank	0.000	16:27
Subject Sample # 2	0.000	16:28
Breath Volume	2.804	16:28
Air Blank	0.000	16:28
Control Test	0.079	16:29
Air Blank	0.000	16:29
Diagnostics Check	OK	16:29

Data Download Date/Time: 09/27/2017 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: RIAD

DATE/TIME OF TEST: 08/27/2017 01:07:42

Diagnostics Check	OK	01:01
Air Blank	0.000	01:02
Control Test	0.079	01:02
Air Blank	0.000	01:03
Subject Sample # 1	0.000	01:03
Breath Volume	1.824	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.000	01:06
Breath Volume	1.207	01:06
Air Blank	0.000	01:06
Control Test	0.080	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 09/27/2017 13:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: DILLON

DATE/TIME OF TEST: 08/28/2017 21:22:16

Diagnostics Check	OK	21:12
Air Blank	0.000	21:13
Control Test	0.079	21:13
Air Blank	0.000	21:14
Subject Sample # 1	0.077	21:15
Breath Volume	1.457	21:15
Air Blank	0.000	21:15
Air Blank	0.000	21:17
Subject Sample # 2	0.074	21:20
Breath Volume	1.652	21:20
Air Blank	0.000	21:21
Control Test	0.079	21:21
Air Blank	0.000	21:22
Diagnostics Check	OK	21:22

Data Download Date/Time: 09/27/2017 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEASE, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FAULK, TAYLOR
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: LANNING

DATE/TIME OF TEST: 08/29/2017 01:21:18

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.078	01:14
Air Blank	0.000	01:14
Subject Sample # 1	0.132	01:16
Breath Volume	2.328	01:16
Air Blank	0.000	01:17
Air Blank	0.000	01:19
Subject Sample # 2	0.133	01:19
Breath Volume	2.773	01:19
Air Blank	0.000	01:20
Control Test	0.078	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 09/27/2017 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEASE, MATTHEW, W
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: FITSCHEN, BRYAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 08/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: KRANE

DATE/TIME OF TEST: 08/30/2017 04:01:50

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.149	03:57
Breath Volume	3.312	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.152	04:00
Breath Volume	3.117	04:00
Air Blank	0.000	04:00
Control Test	0.079	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 09/27/2017 13:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'BRIEN, WILLIAM, P
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD, JEFFREY
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 07/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:22

SUBJECT'S LAST NAME: WILCOX

DATE/TIME OF TEST: 08/10/2017 17:56:03

Diagnostics Check	OK	17:43	
Air Blank	0.000	17:43	
Control Test	0.078	17:44	
Air Blank	0.000	17:44	
Subject Sample	# 1	17:47	No Sample Provided
Breath Volume	0.000	17:47	
Air Blank	0.000	17:48	
Air Blank	0.000	17:50	
Subject Sample	# 2	0.259	17:51
Breath Volume	1.570	17:51	
Air Blank	0.000	17:51	
Air Blank	0.000	17:53	
Subject Sample	# 3	0.267	17:54
Breath Volume	1.523	17:54	
Air Blank	0.000	17:55	
Control Test	0.077	17:55	
Air Blank	0.000	17:55	
Diagnostics Check	OK	17:56	

Data Download Date/Time: 08/11/2017 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSKE, OFC J,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: GUE, OFC R
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 08/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: HAAS

DATE/TIME OF TEST: 08/27/2017 03:33:27

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.077	03:24
Air Blank	0.000	03:24
Subject Sample # 1	0.203	03:26
Breath Volume	1.167	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.199	03:31
Breath Volume	1.140	03:31
Air Blank	0.000	03:32
Control Test	0.075	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 09/27/2017 15:36

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MERO ONA

DATE/TIME OF TEST: 08/12/2017 03:41:48

Diagnostics Check	OK	03:34
Air Blank	0.000	03:35
Control Test	0.077	03:35
Air Blank	0.000	03:36
Subject Sample # 1	0.102	03:36
Breath Volume	1.828	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.103	03:40
Breath Volume	1.320	03:40
Air Blank	0.000	03:40
Control Test	0.077	03:41
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 08/24/2017 03:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/12/2017 23:04:55

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.078	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.176	23:00
Breath Volume	1.824	23:00
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.167	23:03
Breath Volume	1.531	23:03
Air Blank	0.000	23:03
Control Test	0.076	23:04
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 08/24/2017 03:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001109

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 08/26/2017 00:15:38

Diagnostics Check	OK	00:09
Air Blank	0.000	00:09
Control Test	0.078	00:10
Air Blank	0.000	00:10
Subject Sample # 1	0.089	00:11
Breath Volume	3.152	00:11
Air Blank	0.000	00:11
Air Blank	0.000	00:13
Subject Sample # 2	0.087	00:14
Breath Volume	2.523	00:14
Air Blank	0.000	00:14
Control Test	0.078	00:15
Air Blank	0.000	00:15
Diagnostics Check	OK	00:15

Data Download Date/Time: 09/07/2017 22:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERO,JOSE,I
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: RIVERO,JOSE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: MORALES AVILA

DATE/TIME OF TEST: 08/27/2017 03:48:47

Diagnostics Check	OK	03:42
Air Blank	0.000	03:42
Control Test	0.082	03:42
Air Blank	0.000	03:43
Subject Sample # 1	0.148	03:44
Breath Volume	1.265	03:44
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.144	03:47
Breath Volume	1.269	03:47
Air Blank	0.000	03:47
Control Test	0.079	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 09/29/2017 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINTILIE,BOB,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: SAINTILIE,BOB
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 08/27/2017 05:15:29

Diagnostics Check	OK	05:08
Air Blank	0.000	05:08
Control Test	0.079	05:08
Air Blank	0.000	05:09
Subject Sample # 1	0.111	05:11
Breath Volume	1.617	05:11
Air Blank	0.000	05:11
Air Blank	0.000	05:13
Subject Sample # 2	0.121	05:13
Breath Volume	1.859	05:13
Air Blank	0.000	05:14
Control Test	0.082	05:14
Air Blank	0.000	05:15
Diagnostics Check	OK	05:15

Data Download Date/Time: 09/29/2017 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: LOPEZMACHADO

DATE/TIME OF TEST: 08/28/2017 01:09:41

Diagnostics Check		OK	01:02
Air Blank		0.000	01:02
Control Test		0.081	01:02
Air Blank		0.000	01:03
Subject Sample	# 1	0.136	01:04
Breath Volume		2.445	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.134	01:08
Breath Volume		2.390	01:08
Air Blank		0.000	01:08
Control Test		0.082	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 09/29/2017 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS, MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 08/28/2017 01:55:30

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:52	
Control Test	0.082	01:52	
Air Blank	0.000	01:52	
Subject Sample	# 1	01:54	Subject Test Refused
Breath Volume	0.000	01:54	
Air Blank	0.000	01:54	
Control Test	0.081	01:54	
Air Blank	0.000	01:55	
Diagnostics Check	OK	01:55	

Data Download Date/Time: 09/29/2017 11:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001110

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEL CASTILLO,RAUL,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DELCASTILLO,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: DUERO3

DATE/TIME OF TEST: 08/31/2017 09:42:22

Diagnostics Check	OK	09:39	
Air Blank	0.000	09:39	
Control Test	0.081	09:40	
Air Blank	0.000	09:40	
Subject Sample	# 1	09:41	Subject Test Refused
Breath Volume	0.000	09:41	
Air Blank	0.000	09:41	
Control Test	0.081	09:41	
Air Blank	0.000	09:42	
Diagnostics Check	OK	09:42	

Data Download Date/Time: 09/29/2017 11:35

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JONATHAN, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 07/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: KANUCH

DATE/TIME OF TEST: 08/03/2017 19:50:30

Diagnostics Check	OK	19:43
Air Blank	0.000	19:44
Control Test	0.081	19:44
Air Blank	0.000	19:45
Subject Sample # 1	0.133	19:45
Breath Volume	2.703	19:45
Air Blank	0.000	19:46
Air Blank	0.000	19:48
Subject Sample # 2	0.140	19:49
Breath Volume	3.214	19:49
Air Blank	0.000	19:49
Control Test	0.082	19:49
Air Blank	0.000	19:50
Diagnostics Check	OK	19:50

Data Download Date/Time: 08/12/2017 14:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: RHODES

DATE/TIME OF TEST: 08/20/2017 21:51:48

Diagnostics Check	OK	21:44
Air Blank	0.000	21:45
Control Test	0.081	21:45
Air Blank	0.000	21:45
Subject Sample # 1	0.148	21:46
Breath Volume	3.027	21:46
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.133	21:50
Breath Volume	1.695	21:50
Air Blank	0.000	21:50
Control Test	0.081	21:51
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 09/04/2017 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 12:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:15

SUBJECT'S LAST NAME: BYARS

DATE/TIME OF TEST: 08/23/2017 13:45:04

Diagnostics Check	OK	13:37	
Air Blank	0.000	13:37	
Control Test	0.081	13:37	
Air Blank	0.000	13:38	
Subject Sample # 1	0.000	13:39	
Breath Volume	1.195	13:39	
Air Blank	0.000	13:39	
Air Blank	0.000	13:41	
Subject Sample # 2		13:43	Subject Test Refused
Breath Volume	0.832	13:43	
Air Blank	0.000	13:44	
Control Test	0.082	13:44	
Air Blank	0.000	13:44	
Diagnostics Check	OK	13:45	

Data Download Date/Time: 09/04/2017 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JONATHAN, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 08/26/2017 19:55:54

Diagnostics Check		OK	19:44	
Air Blank		0.000	19:44	
Control Test		0.081	19:45	
Air Blank		0.000	19:45	
Subject Sample	# 1	0.161	19:48	Volume Not Met
Breath Volume		0.074	19:48	
Air Blank		0.000	19:49	
Air Blank		0.000	19:51	
Subject Sample	# 2	0.214	19:54	Volume Not Met
Breath Volume		0.257	19:54	
Air Blank		0.000	19:54	
Control Test		0.082	19:55	
Air Blank		0.000	19:55	
Diagnostics Check		OK	19:55	

Data Download Date/Time: 09/04/2017 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JONATHAN, P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 08/26/2017 20:10:35

Diagnostics Check		OK	19:58	
Air Blank		0.000	19:59	
Control Test		0.080	19:59	
Air Blank		0.000	20:00	
Subject Sample	# 1	0.189	20:03	Volume Not Met
Breath Volume		0.234	20:03	
Air Blank		0.000	20:04	
Air Blank		0.000	20:05	
Subject Sample	# 2	0.188	20:09	Volume Not Met
Breath Volume		0.890	20:09	
Air Blank		0.000	20:09	
Control Test		0.080	20:09	
Air Blank		0.000	20:10	
Diagnostics Check		OK	20:10	

Data Download Date/Time: 09/04/2017 22:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, CHRISTOPHER, M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY, JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 08/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 08/26/2017 20:20:15

Diagnostics Check	OK	20:12
Air Blank	0.000	20:13
Control Test	0.080	20:13
Air Blank	0.000	20:13
Subject Sample # 1	0.240	20:14
Breath Volume	1.585	20:14
Air Blank	0.000	20:15
Air Blank	0.000	20:17
Subject Sample # 2	0.233	20:18
Breath Volume	1.484	20:18
Air Blank	0.000	20:19
Control Test	0.081	20:19
Air Blank	0.000	20:20
Diagnostics Check	OK	20:20

Data Download Date/Time: 09/04/2017 22:34

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: WOLFF, ROSE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: BUBES

DATE/TIME OF TEST: 08/12/2017 05:17:08

Diagnostics Check	OK	05:11
Air Blank	0.000	05:11
Control Test	0.082	05:11
Air Blank	0.000	05:12
Subject Sample # 1	0.000	05:12
Breath Volume	1.667	05:12
Air Blank	0.000	05:13
Air Blank	0.000	05:15
Subject Sample # 2	0.000	05:15
Breath Volume	1.398	05:15
Air Blank	0.000	05:16
Control Test	0.082	05:16
Air Blank	0.000	05:17
Diagnostics Check	OK	05:17

Data Download Date/Time: 08/17/2017 11:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK, WILLIAM, D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: BLACKMON

DATE/TIME OF TEST: 08/19/2017 03:17:37

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.081	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.000	03:13
Breath Volume	2.378	03:13
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.000	03:16
Breath Volume	2.070	03:16
Air Blank	0.000	03:16
Control Test	0.082	03:17
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 09/10/2017 02:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: BOZIAN

DATE/TIME OF TEST: 08/03/2017 04:25:38

Diagnostics Check		OK	04:18
Air Blank		0.000	04:18
Control Test		0.082	04:19
Air Blank		0.000	04:19
Subject Sample	# 1	0.058	04:20
Breath Volume		2.285	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:22
Subject Sample	# 2	0.059	04:24
Breath Volume		2.156	04:24
Air Blank		0.000	04:24
Control Test		0.082	04:25
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 08/24/2017 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: CHOWDHURY

DATE/TIME OF TEST: 08/04/2017 03:44:54

Diagnostics Check	OK	03:38
Air Blank	0.000	03:39
Control Test	0.081	03:39
Air Blank	0.000	03:40
Subject Sample # 1	0.096	03:40
Breath Volume	2.253	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:43
Subject Sample # 2	0.095	03:43
Breath Volume	2.093	03:43
Air Blank	0.000	03:43
Control Test	0.081	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 08/24/2017 03:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 08/08/2017 23:41:07

Diagnostics Check	OK	23:35
Air Blank	0.000	23:35
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.000	23:36
Breath Volume	2.410	23:36
Air Blank	0.000	23:37
Air Blank	0.000	23:39
Subject Sample # 2	0.000	23:39
Breath Volume	1.882	23:39
Air Blank	0.000	23:40
Control Test	0.081	23:40
Air Blank	0.000	23:41
Diagnostics Check	OK	23:41

Data Download Date/Time: 08/24/2017 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: CHADWICK

DATE/TIME OF TEST: 08/17/2017 01:44:57

Diagnostics Check	OK	01:38
Air Blank	0.000	01:39
Control Test	0.082	01:39
Air Blank	0.000	01:40
Subject Sample # 1	0.219	01:40
Breath Volume	1.886	01:40
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.219	01:43
Breath Volume	1.980	01:43
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:44
Diagnostics Check	OK	01:44

Data Download Date/Time: 08/24/2017 03:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 07/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MARTINEZ MARTINEZ

DATE/TIME OF TEST: 08/22/2017 02:20:50

Diagnostics Check	OK	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.136	02:16
Breath Volume	1.250	02:16
Air Blank	0.000	02:16
Air Blank	0.000	02:18
Subject Sample # 2	0.152	02:19
Breath Volume	1.578	02:19
Air Blank	0.000	02:19
Control Test	0.081	02:20
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 08/24/2017 03:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: GARCIA ABRAHANTES

DATE/TIME OF TEST: 08/26/2017 02:40:16

Diagnostics Check	OK	02:34
Air Blank	0.000	02:34
Control Test	0.080	02:34
Air Blank	0.000	02:35
Subject Sample # 1	0.160	02:35
Breath Volume	1.394	02:35
Air Blank	0.000	02:36
Air Blank	0.000	02:38
Subject Sample # 2	0.155	02:38
Breath Volume	1.679	02:38
Air Blank	0.000	02:39
Control Test	0.080	02:39
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 09/07/2017 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 01:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: ALCANTARA RIVERA

DATE/TIME OF TEST: 08/27/2017 02:12:58

Diagnostics Check	OK	02:05
Air Blank	0.000	02:05
Control Test	0.081	02:06
Air Blank	0.000	02:06
Subject Sample # 1	0.189	02:08
Breath Volume	3.140	02:08
Air Blank	0.000	02:09
Air Blank	0.000	02:10
Subject Sample # 2	0.203	02:11
Breath Volume	2.312	02:11
Air Blank	0.000	02:12
Control Test	0.081	02:12
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 09/07/2017 22:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001113

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNSBERRY,J,N
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: THORNSBERRY,J
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 08/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 796576
DRY GAS EXPIRY DATE: 06/14/2020
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: KELLNER

DATE/TIME OF TEST: 08/27/2017 04:46:37

Diagnostics Check	OK	04:40
Air Blank	0.000	04:41
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.187	04:42
Breath Volume	3.191	04:42
Air Blank	0.000	04:42
Air Blank	0.000	04:44
Subject Sample # 2	0.189	04:45
Breath Volume	2.894	04:45
Air Blank	0.000	04:45
Control Test	0.080	04:46
Air Blank	0.000	04:46
Diagnostics Check	OK	04:46

Data Download Date/Time: 09/07/2017 22:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: KRYSHCHENKO,ILLYA
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: ABRAHAMSON

DATE/TIME OF TEST: 08/17/2017 21:56:16

Diagnostics Check	OK	21:53	
Air Blank	0.000	21:53	
Control Test	0.077	21:54	
Air Blank	0.000	21:54	
Subject Sample	# 1	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Control Test	0.076	21:55	
Air Blank	0.000	21:56	
Diagnostics Check	OK	21:56	

Data Download Date/Time: 08/29/2017 14:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BICE,BRIAN
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MCINNIS

DATE/TIME OF TEST: 08/22/2017 00:21:47

Diagnostics Check	OK	00:18	
Air Blank	0.000	00:19	
Control Test	0.076	00:19	
Air Blank	0.000	00:20	
Subject Sample	# 1	00:20	Subject Test Refused
Breath Volume	0.000	00:20	
Air Blank	0.000	00:20	
Control Test	0.077	00:21	
Air Blank	0.000	00:21	
Diagnostics Check	OK	00:21	

Data Download Date/Time: 08/29/2017 14:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMIDT,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 08/01/2017 02:55:53

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.079	02:46
Air Blank	0.000	02:47
Subject Sample # 1	0.219	02:50
Breath Volume	1.156	02:50
Air Blank	0.000	02:51
Air Blank	0.000	02:53
Subject Sample # 2	0.220	02:54
Breath Volume	2.066	02:54
Air Blank	0.000	02:54
Control Test	0.077	02:55
Air Blank	0.000	02:55
Diagnostics Check	OK	02:55

Data Download Date/Time: 08/07/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HURTADO

DATE/TIME OF TEST: 08/02/2017 01:58:11

Diagnostics Check		OK	01:51
Air Blank		0.000	01:52
Control Test		0.077	01:52
Air Blank		0.000	01:53
Subject Sample	# 1	0.234	01:53
Breath Volume		1.878	01:53
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.233	01:56
Breath Volume		2.187	01:56
Air Blank		0.000	01:57
Control Test		0.076	01:57
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 08/07/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEIGNER,BRIAN,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SPEIGNER,BRIAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 08/03/2017 04:02:42

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.077	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.192	03:55
Breath Volume	3.089	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.165	03:58
Breath Volume	1.710	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 3	0.193	04:01
Breath Volume	3.515	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 08/07/2017 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 08/04/2017 03:14:55

Diagnostics Check	OK	03:08
Air Blank	0.000	03:09
Control Test	0.079	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.044	03:10
Breath Volume	3.628	03:10
Air Blank	0.000	03:11
Air Blank	0.000	03:13
Subject Sample # 2	0.043	03:13
Breath Volume	3.777	03:13
Air Blank	0.000	03:14
Control Test	0.078	03:14
Air Blank	0.000	03:14
Diagnostics Check	OK	03:14

Data Download Date/Time: 08/07/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 08/05/2017 04:16:30

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.079	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.175	04:12
Breath Volume	1.347	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.155	04:14
Breath Volume	1.500	04:14
Air Blank	0.000	04:15
Control Test	0.078	04:15
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 08/07/2017 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: SCHAEFER

DATE/TIME OF TEST: 08/06/2017 03:43:02

Diagnostics Check		OK	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:37
Subject Sample	# 1	0.177	03:38
Breath Volume		1.195	03:38
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.165	03:41
Breath Volume		1.523	03:41
Air Blank		0.000	03:42
Control Test		0.078	03:42
Air Blank		0.000	03:42
Diagnostics Check		OK	03:43

Data Download Date/Time: 08/07/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROUGHTON,J,
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: SEARS,M
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 07/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ROSIER-TERRY

DATE/TIME OF TEST: 08/06/2017 22:56:17

Diagnostics Check	OK	22:50
Air Blank	0.000	22:50
Control Test	0.078	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.099	22:51
Breath Volume	1.199	22:51
Air Blank	0.000	22:52
Air Blank	0.000	22:54
Subject Sample # 2	0.118	22:54
Breath Volume	2.683	22:54
Air Blank	0.000	22:55
Control Test	0.078	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 08/07/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINDSAY,RYAN,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: LINDSAY,RYAN
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 08/09/2017 22:32:33

Diagnostics Check	OK	22:26
Air Blank	0.000	22:26
Control Test	0.078	22:26
Air Blank	0.000	22:27
Subject Sample # 1	0.073	22:27
Breath Volume	2.722	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.072	22:31
Breath Volume	3.039	22:31
Air Blank	0.000	22:31
Control Test	0.077	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: MINCY

DATE/TIME OF TEST: 08/12/2017 02:45:42

Diagnostics Check	OK	02:39
Air Blank	0.000	02:40
Control Test	0.078	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.128	02:41
Breath Volume	1.332	02:41
Air Blank	0.000	02:41
Air Blank	0.000	02:43
Subject Sample # 2	0.127	02:44
Breath Volume	2.250	02:44
Air Blank	0.000	02:44
Control Test	0.078	02:45
Air Blank	0.000	02:45
Diagnostics Check	OK	02:45

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVERETT, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: EVERETT, MATTHEW
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:59

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 08/14/2017 18:29:49

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.077	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.245	18:25
Breath Volume	1.863	18:25
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.247	18:28
Breath Volume	1.292	18:28
Air Blank	0.000	18:28
Control Test	0.077	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FOOTE, BERNARD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 08/14/2017 19:29:52

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.078	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.176	19:25
Breath Volume	1.843	19:25
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.188	19:28
Breath Volume	1.519	19:28
Air Blank	0.000	19:28
Control Test	0.077	19:29
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAN,DONALD,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: DEAN,DONALD
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 08/14/2017 22:36:31

Diagnostics Check	OK	22:30
Air Blank	0.000	22:30
Control Test	0.078	22:31
Air Blank	0.000	22:31
Subject Sample # 1	0.025	22:32
Breath Volume	2.140	22:32
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.024	22:35
Breath Volume	2.476	22:35
Air Blank	0.000	22:35
Control Test	0.078	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILSON, BRANDON, J
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: WILSON, BRANDON
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: RUSSELL

DATE/TIME OF TEST: 08/15/2017 23:37:26

Diagnostics Check	OK	23:30
Air Blank	0.000	23:31
Control Test	0.078	23:31
Air Blank	0.000	23:32
Subject Sample # 1	0.220	23:32
Breath Volume	1.582	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.220	23:35
Breath Volume	1.886	23:35
Air Blank	0.000	23:36
Control Test	0.078	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VANN,T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 08/17/2017 00:37:28

Diagnostics Check		OK	00:31
Air Blank		0.000	00:31
Control Test		0.078	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.207	00:33
Breath Volume		1.718	00:33
Air Blank		0.000	00:33
Air Blank		0.000	00:35
Subject Sample	# 2	0.211	00:35
Breath Volume		1.839	00:35
Air Blank		0.000	00:36
Control Test		0.076	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: VARNER,TYLER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 08/17/2017 02:46:11

Diagnostics Check	OK	02:39
Air Blank	0.000	02:39
Control Test	0.079	02:40
Air Blank	0.000	02:40
Subject Sample # 1	0.107	02:41
Breath Volume	2.242	02:41
Air Blank	0.000	02:42
Air Blank	0.000	02:44
Subject Sample # 2	0.109	02:44
Breath Volume	1.863	02:44
Air Blank	0.000	02:45
Control Test	0.077	02:45
Air Blank	0.000	02:46
Diagnostics Check	OK	02:46

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BUIS,J
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: SHACKELFORD

DATE/TIME OF TEST: 08/19/2017 07:07:25

Diagnostics Check	OK	07:01
Air Blank	0.000	07:01
Control Test	0.078	07:01
Air Blank	0.000	07:02
Subject Sample # 1	0.122	07:02
Breath Volume	1.523	07:02
Air Blank	0.000	07:03
Air Blank	0.000	07:05
Subject Sample # 2	0.119	07:05
Breath Volume	1.906	07:05
Air Blank	0.000	07:06
Control Test	0.077	07:06
Air Blank	0.000	07:07
Diagnostics Check	OK	07:07

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMIDT,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 08/20/2017 02:56:51

Diagnostics Check	OK	02:47
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Subject Sample # 1	0.157	02:49
Breath Volume	1.882	02:49
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.132	02:52
Breath Volume	2.539	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 3	0.154	02:55
Breath Volume	3.289	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LECKINGER, MICHAEL, E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LECKINGER, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ELKINS

DATE/TIME OF TEST: 08/23/2017 00:21:25

Diagnostics Check		OK	00:14
Air Blank		0.000	00:14
Control Test		0.078	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.219	00:16
Breath Volume		1.566	00:16
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 2	0.228	00:19
Breath Volume		1.546	00:19
Air Blank		0.000	00:20
Control Test		0.077	00:20
Air Blank		0.000	00:21
Diagnostics Check		OK	00:21

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: FORSYTHE

DATE/TIME OF TEST: 08/24/2017 01:56:27

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.077	01:51
Air Blank	0.000	01:51
Subject Sample # 1	0.186	01:52
Breath Volume	1.882	01:52
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.192	01:54
Breath Volume	2.316	01:54
Air Blank	0.000	01:55
Control Test	0.078	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,PARISE,L
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: ADAMS,PARISE
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: CREW

DATE/TIME OF TEST: 08/25/2017 00:50:55

Diagnostics Check	OK	00:48	
Air Blank	0.000	00:48	
Control Test	0.078	00:48	
Air Blank	0.000	00:49	
Subject Sample	# 1	00:49	Subject Test Refused
Breath Volume	0.000	00:49	
Air Blank	0.000	00:49	
Control Test	0.079	00:50	
Air Blank	0.000	00:50	
Diagnostics Check	OK	00:50	

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BULLOCK,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BULLOCK,JOSHUA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HONES

DATE/TIME OF TEST: 08/25/2017 03:44:00

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.078	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.166	03:33
Breath Volume	1.765	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.145	03:38
Breath Volume	1.308	03:38
Air Blank	0.000	03:39
Air Blank	0.000	03:40
Subject Sample # 3	0.155	03:42
Breath Volume	1.445	03:42
Air Blank	0.000	03:43
Control Test	0.077	03:43
Air Blank	0.000	03:43
Diagnostics Check	OK	03:44

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEGRAW,JACK,D
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: VARNER,TYLER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: SAWYER

DATE/TIME OF TEST: 08/26/2017 02:04:59

Diagnostics Check	OK	01:52	
Air Blank	0.000	01:52	
Control Test	0.078	01:53	
Air Blank	0.000	01:53	
Subject Sample	# 1 0.167	01:56	Volume Not Met
Breath Volume	0.093	01:56	
Air Blank	0.000	01:57	
Air Blank	0.000	01:59	
Subject Sample	# 2 0.166	02:00	
Breath Volume	1.781	02:00	
Air Blank	0.000	02:01	
Air Blank	0.000	02:02	
Subject Sample	# 3 0.160	02:03	
Breath Volume	1.621	02:03	
Air Blank	0.000	02:04	
Control Test	0.076	02:04	
Air Blank	0.000	02:04	
Diagnostics Check	OK	02:04	

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 08/26/2017 04:39:24

Diagnostics Check	OK	04:28	
Air Blank	0.000	04:28	
Control Test	0.078	04:28	
Air Blank	0.000	04:29	
Subject Sample	# 1	04:32	No Sample Provided
Breath Volume	0.000	04:32	
Air Blank	0.000	04:32	
Air Blank	0.000	04:34	
Subject Sample	# 2	04:38	No Sample Provided
Breath Volume	0.000	04:38	
Air Blank	0.000	04:38	
Control Test	0.079	04:38	
Air Blank	0.000	04:39	
Diagnostics Check	OK	04:39	

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,JEFFREY,A
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: THOMPSON,JEFFREY
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 08/26/2017 04:46:02

Diagnostics Check	OK	04:41	
Air Blank	0.000	04:42	
Control Test	0.078	04:42	
Air Blank	0.000	04:43	
Subject Sample	# 1	04:44	Subject Test Refused
Breath Volume	0.050	04:44	
Air Blank	0.000	04:45	
Control Test	0.078	04:45	
Air Blank	0.000	04:45	
Diagnostics Check	OK	04:46	

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: VANN,T
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 08/27/2017 02:29:14

Diagnostics Check	OK	02:23
Air Blank	0.000	02:23
Control Test	0.078	02:23
Air Blank	0.000	02:24
Subject Sample # 1	0.159	02:24
Breath Volume	1.375	02:24
Air Blank	0.000	02:25
Air Blank	0.000	02:27
Subject Sample # 2	0.163	02:27
Breath Volume	1.894	02:27
Air Blank	0.000	02:28
Control Test	0.077	02:28
Air Blank	0.000	02:29
Diagnostics Check	OK	02:29

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SODERHOLM,CHELSEA,C
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SODERHOLM,CHELSEA
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: SHEFFIELD

DATE/TIME OF TEST: 08/30/2017 03:12:09

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Subject Sample	# 1	0.184	03:07
Breath Volume		1.687	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.184	03:10
Breath Volume		1.675	03:10
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 08/31/2017 01:11:13

Diagnostics Check	OK	01:04
Air Blank	0.000	01:05
Control Test	0.078	01:05
Air Blank	0.000	01:06
Subject Sample # 1	0.223	01:06
Breath Volume	2.199	01:06
Air Blank	0.000	01:07
Air Blank	0.000	01:08
Subject Sample # 2	0.222	01:09
Breath Volume	1.585	01:09
Air Blank	0.000	01:10
Control Test	0.077	01:10
Air Blank	0.000	01:11
Diagnostics Check	OK	01:11

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001115

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAXTON,D,L
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: BRAXTON,D
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: DEROSIER

DATE/TIME OF TEST: 08/31/2017 02:56:38

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.079	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.130	02:51
Breath Volume	1.574	02:51
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.133	02:55
Breath Volume	2.140	02:55
Air Blank	0.000	02:55
Control Test	0.078	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 09/25/2017 10:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINERA,ORESTE,I
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROBINSON,IVAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 07/10/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 19:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: CABRAL

DATE/TIME OF TEST: 08/03/2017 20:59:19

Diagnostics Check		OK	20:52
Air Blank		0.000	20:53
Control Test		0.081	20:53
Air Blank		0.000	20:54
Subject Sample	# 1	0.201	20:54
Breath Volume		2.496	20:54
Air Blank		0.000	20:55
Air Blank		0.000	20:57
Subject Sample	# 2	0.201	20:57
Breath Volume		2.355	20:57
Air Blank		0.000	20:58
Control Test		0.079	20:58
Air Blank		0.000	20:59
Diagnostics Check		OK	20:59

Data Download Date/Time: 08/29/2017 17:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001119

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEMAN, EDWARD, A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: NUNEZ, ANDREW
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 437738
DRY GAS EXPIRY DATE: 06/08/2019
VIOLATION CODE: DUI
ARREST TIME: 19:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: ABELL

DATE/TIME OF TEST: 08/31/2017 20:08:53

Diagnostics Check	OK	20:00
Air Blank	0.000	20:01
Control Test	0.080	20:01
Air Blank	0.000	20:01
Subject Sample # 1	0.130	20:02
Breath Volume	3.375	20:02
Air Blank	0.000	20:03
Air Blank	0.000	20:05
Subject Sample # 2	0.122	20:07
Breath Volume	1.851	20:07
Air Blank	0.000	20:07
Control Test	0.078	20:08
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 09/26/2017 18:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001120

INSTRUMENT REGISTERED TO: FHP TROOP K

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: VARON, MARCOS, H

BREATH TEST OPERATOR AGENCY: FHP Troop K

ARRESTING OFFICER: VARON, MARCOS

ARRESTING OFFICER AGENCY: FHP Troop K

LAST AGENCY INSPECTION: 07/07/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 366849

DRY GAS EXPIRY DATE: 03/30/2019

VIOLATION CODE: DUI

ARREST TIME: 19:17

GENDER: Male

OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CRUZ GONZALEZ

DATE/TIME OF TEST: 08/21/2017 21:55:25

Diagnostics Check	OK	21:52	
Air Blank	0.000	21:52	
Control Test	0.080	21:53	
Air Blank	0.000	21:53	
Subject Sample	# 1	21:54	Subject Test Refused
Breath Volume	0.000	21:54	
Air Blank	0.000	21:54	
Control Test	0.080	21:54	
Air Blank	0.000	21:55	
Diagnostics Check	OK	21:55	

Data Download Date/Time: 08/25/2017 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001129

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS, TIFFANY, R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS, TIFFANY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 08/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 08/31/2017 23:23:01

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.079	23:17
Air Blank	0.000	23:18
Subject Sample # 1	0.078	23:18
Breath Volume	2.285	23:18
Air Blank	0.000	23:19
Air Blank	0.000	23:21
Subject Sample # 2	0.077	23:21
Breath Volume	2.421	23:21
Air Blank	0.000	23:22
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:23

Data Download Date/Time: 10/02/2017 02:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:09

SUBJECT'S LAST NAME: HARMON

DATE/TIME OF TEST: 08/06/2017 20:43:09

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.078	20:37
Air Blank	0.000	20:37
Subject Sample # 1	0.170	20:38
Breath Volume	2.167	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:40
Subject Sample # 2	0.170	20:41
Breath Volume	2.128	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 08/29/2017 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,JOSEPH,W
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: MYERS,MATTHEW
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:18

SUBJECT'S LAST NAME: MICHEL

DATE/TIME OF TEST: 08/08/2017 23:51:16

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.078	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.213	23:43
Breath Volume	1.386	23:43
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.234	23:46
Breath Volume	1.714	23:46
Air Blank	0.000	23:47
Air Blank	0.000	23:48
Subject Sample # 3	0.234	23:49
Breath Volume	1.574	23:49
Air Blank	0.000	23:50
Control Test	0.077	23:50
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 08/29/2017 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WILSON, JEREMIAH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: COATES

DATE/TIME OF TEST: 08/16/2017 23:21:43

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.078	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.206	23:13
Breath Volume	1.781	23:13
Air Blank	0.000	23:14
Air Blank	0.000	23:16
Subject Sample # 2	0.183	23:17
Breath Volume	1.253	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 3	0.203	23:20
Breath Volume	1.988	23:20
Air Blank	0.000	23:20
Control Test	0.078	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 08/29/2017 14:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001130

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: ROLLY, CAVIDA
ARRESTING OFFICER AGENCY: Orange City PD
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: WELLER

DATE/TIME OF TEST: 08/27/2017 22:41:15

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.078	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.163	22:36
Breath Volume	1.281	22:36
Air Blank	0.000	22:36
Air Blank	0.000	22:38
Subject Sample # 2	0.163	22:39
Breath Volume	1.746	22:39
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 08/29/2017 14:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: DORION

DATE/TIME OF TEST: 08/02/2017 00:28:30

Diagnostics Check	OK	00:25	
Air Blank	0.000	00:26	
Control Test	0.080	00:26	
Air Blank	0.000	00:26	
Subject Sample	# 1	00:27	Subject Test Refused
Breath Volume	0.000	00:27	
Air Blank	0.000	00:27	
Control Test	0.080	00:27	
Air Blank	0.000	00:28	
Diagnostics Check	OK	00:28	

Data Download Date/Time: 08/29/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,KEITH,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: BOISSAT

DATE/TIME OF TEST: 08/07/2017 04:20:00

Diagnostics Check	OK	04:17	
Air Blank	0.000	04:17	
Control Test	0.080	04:17	
Air Blank	0.000	04:18	
Subject Sample	# 1	04:18	Subject Test Refused
Breath Volume	0.000	04:18	
Air Blank	0.000	04:19	
Control Test	0.080	04:19	
Air Blank		04:19	RFI Detect
Air Blank	0.000	04:20	

Data Download Date/Time: 08/29/2017 14:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: MILLIKEN

DATE/TIME OF TEST: 08/13/2017 22:38:34

Diagnostics Check	OK	22:31
Air Blank	0.000	22:32
Control Test	0.079	22:32
Air Blank	0.000	22:33
Subject Sample # 1	0.143	22:33
Breath Volume	3.230	22:33
Air Blank	0.000	22:34
Air Blank	0.000	22:36
Subject Sample # 2	0.146	22:37
Breath Volume	3.421	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Diagnostics Check	OK	22:38

Data Download Date/Time: 08/29/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: TAVAREZ

DATE/TIME OF TEST: 08/27/2017 20:34:43

Diagnostics Check	OK	20:27
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:28
Subject Sample # 1	0.134	20:29
Breath Volume	3.613	20:29
Air Blank	0.000	20:29
Air Blank	0.000	20:31
Subject Sample # 2	0.128	20:33
Breath Volume	2.339	20:33
Air Blank	0.000	20:33
Control Test	0.080	20:34
Air Blank	0.000	20:34
Diagnostics Check	OK	20:34

Data Download Date/Time: 08/29/2017 14:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001131

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTENOUR,COREY,T
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RITTENOUR,COREY
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 07/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 08/28/2017 03:27:53

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:17	
Air Blank		0.000	03:17	
Subject Sample	# 1	0.188	03:20	Volume Not Met
Breath Volume		0.035	03:20	
Air Blank		0.000	03:21	
Air Blank		0.000	03:23	
Subject Sample	# 2		03:26	No Sample Provided
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Control Test		0.079	03:27	
Air Blank		0.000	03:27	
Diagnostics Check		OK	03:27	

Data Download Date/Time: 08/29/2017 14:20