

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LONES,SHAUN,G
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: LONES,SHAUN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: SPENCER

DATE/TIME OF TEST: 09/03/2017 19:12:12

Diagnostics Check	OK	19:06	
Air Blank	0.000	19:07	
Control Test	0.083	19:07	
Air Blank	0.000	19:08	
Subject Sample	# 1	19:10	Subject Test Refused
Breath Volume	0.000	19:10	
Air Blank	0.000	19:11	
Control Test	0.082	19:11	
Air Blank	0.000	19:12	
Diagnostics Check	OK	19:12	

Data Download Date/Time: 09/27/2017 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOPP,SETH,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: HOPP,SETH
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 09/22/2017 20:48:14

Diagnostics Check	OK	20:45	
Air Blank	0.000	20:45	
Control Test	0.080	20:46	
Air Blank	0.000	20:46	
Subject Sample	# 1	20:46	Subject Test Refused
Breath Volume	0.000	20:46	
Air Blank	0.000	20:47	
Control Test	0.080	20:47	
Air Blank	0.000	20:48	
Diagnostics Check	OK	20:48	

Data Download Date/Time: 09/27/2017 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEDETTI,MARIO,E
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KILMURRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ANDRICAIN

DATE/TIME OF TEST: 09/23/2017 02:52:55

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.081	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.132	02:48
Breath Volume	1.722	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:50
Subject Sample # 2	0.128	02:51
Breath Volume	2.277	02:51
Air Blank	0.000	02:52
Control Test	0.080	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 09/27/2017 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: EISINGER

DATE/TIME OF TEST: 09/23/2017 17:14:11

Diagnostics Check		OK	17:02	
Air Blank		0.000	17:03	
Control Test		0.081	17:03	
Air Blank		0.000	17:03	
Subject Sample	# 1	0.071	17:07	Volume Not Met
Breath Volume		0.746	17:07	
Air Blank		0.000	17:07	
Air Blank		0.000	17:09	
Subject Sample	# 2	0.077	17:12	Volume Not Met
Breath Volume		0.347	17:12	
Air Blank		0.000	17:13	
Control Test		0.080	17:13	
Air Blank		0.000	17:14	
Diagnostics Check		OK	17:14	

Data Download Date/Time: 09/27/2017 13:26

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000868

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 08/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 15:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:38

SUBJECT'S LAST NAME: EISINGER

DATE/TIME OF TEST: 09/23/2017 17:25:37

Diagnostics Check	OK	17:15	
Air Blank	0.000	17:16	
Control Test	0.079	17:16	
Air Blank	0.000	17:16	
Subject Sample # 1	0.061	17:18	
Breath Volume	1.566	17:18	
Air Blank	0.000	17:19	
Air Blank	0.000	17:21	
Subject Sample # 2		17:24	Subject Test Refused
Breath Volume	0.234	17:24	
Air Blank	0.000	17:24	
Control Test	0.080	17:25	
Air Blank	0.000	17:25	
Diagnostics Check	OK	17:25	

Data Download Date/Time: 09/27/2017 13:26

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:03

SUBJECT'S LAST NAME: SEETHALER

DATE/TIME OF TEST: 09/01/2017 22:26:56

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.082	22:24	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.082	22:26	
Air Blank	0.000	22:26	
Diagnostics Check	OK	22:26	

Data Download Date/Time: 09/26/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 09/06/2017 22:02:59

Diagnostics Check	OK	21:57	
Air Blank	0.000	21:57	
Control Test	0.081	21:57	
Air Blank	0.000	21:58	
Subject Sample	# 1 0.197	22:01	
Breath Volume	1.781	22:01	
Air Blank		22:02	Purge Fail
Air Blank		22:02	Purge Fail

Data Download Date/Time: 09/26/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 08/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: LITTLE

DATE/TIME OF TEST: 09/06/2017 22:12:33

Diagnostics Check	OK	22:06
Air Blank	0.000	22:06
Control Test	0.080	22:06
Air Blank	0.000	22:07
Subject Sample # 1	0.213	22:08
Breath Volume	1.593	22:08
Air Blank	0.000	22:08
Air Blank	0.000	22:10
Subject Sample # 2	0.204	22:11
Breath Volume	1.242	22:11
Air Blank	0.000	22:11
Control Test	0.081	22:11
Air Blank	0.000	22:12
Diagnostics Check	OK	22:12

Data Download Date/Time: 09/26/2017 14:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:48

SUBJECT'S LAST NAME: FULPER

DATE/TIME OF TEST: 09/26/2017 21:14:22

Diagnostics Check	OK	21:10	
Air Blank	0.000	21:10	
Control Test	0.081	21:11	
Air Blank	0.000	21:11	
Subject Sample	# 1	21:13	Subject Test Refused
Breath Volume	0.000	21:13	
Air Blank	0.000	21:13	
Control Test	0.081	21:13	
Air Blank	0.000	21:14	
Diagnostics Check	OK	21:14	

Data Download Date/Time: 10/01/2017 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: VILLANUEVA,MAURICIO
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:43

SUBJECT'S LAST NAME: DUBON SANTOS

DATE/TIME OF TEST: 09/29/2017 22:10:17

Diagnostics Check	OK	22:04
Air Blank	0.000	22:04
Control Test	0.080	22:04
Air Blank	0.000	22:05
Subject Sample # 1	0.308	22:05
Breath Volume	2.367	22:05
Air Blank	0.000	22:06
Air Blank	0.000	22:08
Subject Sample # 2	0.306	22:08
Breath Volume	3.769	22:08
Air Blank	0.000	22:09
Control Test	0.082	22:09
Air Blank	0.000	22:10
Diagnostics Check	OK	22:10

Data Download Date/Time: 10/01/2017 18:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: CHRISTLE-IKRAMOV

DATE/TIME OF TEST: 09/04/2017 05:55:55

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.082	05:49
Air Blank	0.000	05:50
Subject Sample # 1	0.226	05:51
Breath Volume	1.890	05:51
Air Blank	0.000	05:51
Air Blank	0.000	05:53
Subject Sample # 2	0.219	05:54
Breath Volume	1.449	05:54
Air Blank	0.000	05:55
Control Test	0.080	05:55
Air Blank	0.000	05:55
Diagnostics Check	OK	05:55

Data Download Date/Time: 09/27/2017 18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000870

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 08/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: TUREK

DATE/TIME OF TEST: 09/05/2017 01:17:58

Diagnostics Check	OK	01:15	
Air Blank	0.000	01:15	
Control Test	0.082	01:15	
Air Blank	0.000	01:16	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.082	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 09/27/2017 18:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000875

INSTRUMENT REGISTERED TO: CALHOUN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLIDAY,DAISY,J
BREATH TEST OPERATOR AGENCY: Blountstown PD
ARRESTING OFFICER: DICKEY,WILLIAM
ARRESTING OFFICER AGENCY: Calhoun County SO
LAST AGENCY INSPECTION: 08/08/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: JAMES

DATE/TIME OF TEST: 09/02/2017 00:32:51

Diagnostics Check	OK	00:29	
Air Blank	0.000	00:30	
Control Test	0.079	00:30	
Air Blank	0.000	00:31	
Subject Sample	# 1	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.079	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 09/18/2017 09:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CRUZ,JORGE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: GODINEZ MARTIN

DATE/TIME OF TEST: 09/02/2017 00:02:08

Diagnostics Check	OK	23:54
Air Blank	0.000	23:55
Control Test	0.082	23:55
Air Blank	0.000	23:55
Subject Sample # 2	0.190	23:57
Breath Volume	1.136	23:57
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.204	00:00
Breath Volume	2.390	00:00
Air Blank	0.000	00:01
Control Test	0.081	00:01
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 09/26/2017 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: FREMOND

DATE/TIME OF TEST: 09/02/2017 00:48:03

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:42
Subject Sample # 1	0.248	00:43
Breath Volume	1.542	00:43
Air Blank	0.000	00:44
Air Blank	0.000	00:45
Subject Sample # 2	0.257	00:46
Breath Volume	1.492	00:46
Air Blank	0.000	00:47
Control Test	0.080	00:47
Air Blank	0.000	00:47
Diagnostics Check	OK	00:48

Data Download Date/Time: 09/26/2017 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,JESUS,O
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: CRUZ,JORGE
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GUERRERO CHAVEZ

DATE/TIME OF TEST: 09/03/2017 00:03:13

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.082	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:02	
Control Test	0.082	00:02	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:03	

Data Download Date/Time: 09/26/2017 08:21

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: XINIC PACHECO

DATE/TIME OF TEST: 09/03/2017 22:23:52

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.082	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.020	22:19
Breath Volume	2.324	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.022	22:22
Breath Volume	1.835	22:22
Air Blank	0.000	22:22
Control Test	0.083	22:23
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 09/26/2017 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCORE, MICHAEL, A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: WILSON, CHRISTOPHER
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 08/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: FIGLIUOLO

DATE/TIME OF TEST: 09/04/2017 22:29:36

Diagnostics Check	OK	22:26	
Air Blank	0.000	22:27	
Control Test	0.082	22:27	
Air Blank	0.000	22:27	
Subject Sample	# 1	22:28	Subject Test Refused
Breath Volume	0.000	22:28	
Air Blank	0.000	22:28	
Control Test	0.081	22:28	
Air Blank	0.000	22:29	
Diagnostics Check	OK	22:29	

Data Download Date/Time: 09/26/2017 08:21

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELGADO,MIGUEL,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DELGADO,MIGUEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PEARL

DATE/TIME OF TEST: 09/19/2017 23:30:24

Diagnostics Check	OK	23:27	
Air Blank	0.000	23:28	
Control Test	0.077	23:28	
Air Blank	0.000	23:28	
Subject Sample	# 1	23:29	Subject Test Refused
Breath Volume	0.000	23:29	
Air Blank	0.000	23:29	
Control Test	0.076	23:29	
Air Blank	0.000	23:30	
Diagnostics Check	OK	23:30	

Data Download Date/Time: 10/25/2017 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGHESE,JIBY,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: PAVAO,JOSHUA
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: MARIS

DATE/TIME OF TEST: 09/24/2017 23:09:24

Diagnostics Check	OK	23:06	
Air Blank	0.000	23:07	
Control Test	0.077	23:07	
Air Blank	0.000	23:07	
Subject Sample	# 1	23:08	Subject Test Refused
Breath Volume	0.000	23:08	
Air Blank	0.000	23:08	
Control Test	0.078	23:08	
Air Blank	0.000	23:09	
Diagnostics Check	OK	23:09	

Data Download Date/Time: 10/25/2017 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANKOFF,BRIAN,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: HANKOFF,BRIAN
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: MENDES

DATE/TIME OF TEST: 09/30/2017 06:56:59

Diagnostics Check	OK	06:54	
Air Blank	0.000	06:54	
Control Test	0.079	06:54	
Air Blank	0.000	06:55	
Subject Sample	# 1	06:55	Subject Test Refused
Breath Volume	0.000	06:55	
Air Blank	0.000	06:56	
Control Test	0.077	06:56	
Air Blank	0.000	06:56	
Diagnostics Check	OK	06:56	

Data Download Date/Time: 10/25/2017 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RAMIREZ,DANIL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: LIMA

DATE/TIME OF TEST: 09/16/2017 23:06:59

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.082	23:01
Air Blank	0.000	23:01
Subject Sample # 1	0.132	23:02
Breath Volume	2.914	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:04
Subject Sample # 2	0.134	23:05
Breath Volume	3.507	23:05
Air Blank	0.000	23:06
Control Test	0.081	23:06
Air Blank	0.000	23:06
Diagnostics Check	OK	23:06

Data Download Date/Time: 10/25/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIVERA,ALBERTO,J
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: GEBBIA,SALVATORE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: KIRKLAND

DATE/TIME OF TEST: 09/30/2017 02:34:35

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.196	02:29
Breath Volume	1.261	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.198	02:32
Breath Volume	2.066	02:32
Air Blank	0.000	02:33
Control Test	0.080	02:33
Air Blank	0.000	02:34
Diagnostics Check	OK	02:34

Data Download Date/Time: 10/25/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: LOZANO

DATE/TIME OF TEST: 09/01/2017 02:37:02

Diagnostics Check	OK	02:28
Air Blank	0.000	02:28
Control Test	0.081	02:29
Air Blank	0.000	02:29
Subject Sample # 1	0.245	02:32
Breath Volume	3.121	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.248	02:35
Breath Volume	3.218	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:37

Data Download Date/Time: 09/04/2017 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: ROSALES

DATE/TIME OF TEST: 09/02/2017 04:45:52

Diagnostics Check	OK	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Subject Sample # 1	0.130	04:41
Breath Volume	1.929	04:41
Air Blank	0.000	04:41
Air Blank	0.000	04:43
Subject Sample # 2	0.120	04:44
Breath Volume	1.406	04:44
Air Blank	0.000	04:44
Control Test	0.079	04:45
Air Blank	0.000	04:45
Diagnostics Check	OK	04:45

Data Download Date/Time: 09/04/2017 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KENSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BOLUFE

DATE/TIME OF TEST: 09/03/2017 03:32:02

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.081	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.135	03:27
Breath Volume		2.566	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.126	03:30
Breath Volume		2.714	03:30
Air Blank		0.000	03:31
Control Test		0.078	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:32

Data Download Date/Time: 09/04/2017 10:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/03/2017 04:58:47

Diagnostics Check	OK	04:53	
Air Blank	0.000	04:54	
Control Test	0.081	04:54	
Air Blank	0.000	04:54	
Subject Sample	# 1	04:57	Subject Test Refused
Breath Volume	0.000	04:57	
Air Blank	0.000	04:57	
Control Test	0.080	04:58	
Air Blank	0.000	04:58	
Diagnostics Check	OK	04:58	

Data Download Date/Time: 09/04/2017 10:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: ALTAMIRANOJR

DATE/TIME OF TEST: 09/03/2017 06:37:00

Diagnostics Check	OK	06:30
Air Blank	0.000	06:31
Control Test	0.080	06:31
Air Blank	0.000	06:31
Subject Sample # 1	0.140	06:32
Breath Volume	3.027	06:32
Air Blank	0.000	06:33
Air Blank	0.000	06:34
Subject Sample # 2	0.139	06:35
Breath Volume	2.242	06:35
Air Blank	0.000	06:36
Control Test	0.080	06:36
Air Blank	0.000	06:36
Diagnostics Check	OK	06:37

Data Download Date/Time: 09/04/2017 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/03/2017 07:39:36

Diagnostics Check	OK	07:35	
Air Blank	0.000	07:36	
Control Test	0.081	07:36	
Air Blank	0.000	07:37	
Subject Sample	# 1	07:38	Subject Test Refused
Breath Volume	0.000	07:38	
Air Blank	0.000	07:38	
Control Test	0.080	07:38	
Air Blank	0.000	07:39	
Diagnostics Check	OK	07:39	

Data Download Date/Time: 09/04/2017 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/04/2017 00:32:09

Diagnostics Check	OK	00:28	
Air Blank	0.000	00:29	
Control Test	0.082	00:29	
Air Blank	0.000	00:29	
Subject Sample	# 1	00:30	Subject Test Refused
Breath Volume	0.000	00:30	
Air Blank	0.000	00:31	
Control Test	0.080	00:31	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 09/04/2017 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: CONTRERAS

DATE/TIME OF TEST: 09/05/2017 01:59:57

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Subject Sample	# 1	0.155	01:52
Breath Volume		1.308	01:52
Air Blank		0.000	01:53
Air Blank		0.000	01:55
Subject Sample	# 2	0.155	01:58
Breath Volume		1.558	01:58
Air Blank		0.000	01:59
Control Test		0.079	01:59
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 09/12/2017 01:20:58

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.082	01:18	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:19	Subject Test Refused
Breath Volume	0.000	01:19	
Air Blank	0.000	01:20	
Control Test	0.080	01:20	
Air Blank	0.000	01:20	
Diagnostics Check	OK	01:20	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: ALEMAN

DATE/TIME OF TEST: 09/12/2017 01:35:36

Diagnostics Check	OK	01:32	
Air Blank	0.000	01:32	
Control Test	0.081	01:32	
Air Blank	0.000	01:33	
Subject Sample	# 1	01:34	Subject Test Refused
Breath Volume	0.000	01:34	
Air Blank	0.000	01:34	
Control Test	0.080	01:35	
Air Blank	0.000	01:35	
Diagnostics Check	OK	01:35	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:50

SUBJECT'S LAST NAME: RICO

DATE/TIME OF TEST: 09/13/2017 20:17:18

Diagnostics Check	OK	20:10
Air Blank	0.000	20:10
Control Test	0.082	20:11
Air Blank	0.000	20:11
Subject Sample # 1	0.159	20:12
Breath Volume	3.109	20:12
Air Blank	0.000	20:13
Air Blank	0.000	20:14
Subject Sample # 2	0.156	20:15
Breath Volume	3.144	20:15
Air Blank	0.000	20:16
Control Test	0.081	20:16
Air Blank	0.000	20:17
Diagnostics Check	OK	20:17

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: PINEDA ESCOBAR

DATE/TIME OF TEST: 09/18/2017 03:37:21

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.130	03:32
Breath Volume	2.628	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:34
Subject Sample # 2	0.119	03:35
Breath Volume	1.195	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: CARDONA

DATE/TIME OF TEST: 09/22/2017 03:47:43

Diagnostics Check	OK	03:39	
Air Blank	0.000	03:39	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Subject Sample	# 1	03:43	No Sample Provided
Breath Volume	0.000	03:43	
Air Blank	0.000	03:44	
Air Blank	0.000	03:46	
Subject Sample	# 2	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SERRANO ESCOBAR

DATE/TIME OF TEST: 09/22/2017 04:14:32

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.081	04:02
Air Blank	0.000	04:02
Subject Sample # 1	0.227	04:04
Breath Volume	1.226	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:06
Subject Sample # 2	0.251	04:08
Breath Volume	1.988	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 3	0.241	04:12
Breath Volume	1.757	04:12
Air Blank	0.000	04:13
Control Test	0.077	04:13
Air Blank	0.000	04:14
Diagnostics Check	OK	04:14

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: LOPEZ RANGEL

DATE/TIME OF TEST: 09/23/2017 03:05:20

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.080	02:59
Air Blank	0.000	03:00
Subject Sample # 1	0.145	03:00
Breath Volume	1.203	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.147	03:03
Breath Volume	1.144	03:03
Air Blank	0.000	03:04
Control Test	0.080	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: MCELLIOTT

DATE/TIME OF TEST: 09/23/2017 23:43:16

Diagnostics Check	OK	23:35
Air Blank	0.000	23:36
Control Test	0.081	23:36
Air Blank	0.000	23:36
Subject Sample # 1	0.193	23:37
Breath Volume	1.328	23:37
Air Blank	0.000	23:38
Air Blank	0.000	23:40
Subject Sample # 2	0.198	23:41
Breath Volume	1.757	23:41
Air Blank	0.000	23:42
Control Test	0.080	23:42
Air Blank	0.000	23:43
Diagnostics Check	OK	23:43

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: SAINT JEAN

DATE/TIME OF TEST: 09/24/2017 04:18:19

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:05	
Control Test		0.081	04:05	
Air Blank		0.000	04:06	
Subject Sample	# 1	0.096	04:09	Volume Not Met
Breath Volume		0.761	04:09	
Air Blank		0.000	04:10	
Air Blank		0.000	04:11	
Subject Sample	# 2	0.116	04:13	
Breath Volume		1.867	04:13	
Air Blank		0.000	04:14	
Air Blank		0.000	04:16	
Subject Sample	# 3	0.116	04:16	
Breath Volume		1.699	04:16	
Air Blank		0.000	04:17	
Control Test		0.078	04:17	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/25/2017 05:19:24

Diagnostics Check		OK	05:09
Air Blank		0.000	05:10
Control Test		0.081	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.143	05:13
Breath Volume		1.472	05:13
Air Blank		0.000	05:14
Air Blank		0.000	05:15
Subject Sample	# 2	0.140	05:17
Breath Volume		1.621	05:17
Air Blank		0.000	05:18
Control Test		0.080	05:18
Air Blank		0.000	05:19
Diagnostics Check		OK	05:19

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAHA,K
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 09/28/2017 00:49:00

Diagnostics Check	OK	00:41
Air Blank	0.000	00:42
Control Test	0.081	00:42
Air Blank	0.000	00:43
Subject Sample # 1	0.000	00:44
Breath Volume	2.335	00:44
Air Blank	0.000	00:45
Air Blank	0.000	00:47
Subject Sample # 2	0.000	00:47
Breath Volume	1.792	00:47
Air Blank	0.000	00:48
Control Test	0.080	00:48
Air Blank	0.000	00:48
Diagnostics Check	OK	00:49

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 09/28/2017 04:42:10

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.081	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 1	0.000	04:34	Volume Not Met
Breath Volume		0.375	04:34	
Air Blank		0.000	04:34	
Air Blank		0.000	04:36	
Subject Sample	# 2	0.000	04:37	
Breath Volume		1.683	04:37	
Air Blank		0.000	04:37	
Air Blank		0.000	04:39	
Subject Sample	# 3	0.000	04:40	
Breath Volume		1.187	04:40	
Air Blank		0.000	04:41	
Control Test		0.080	04:41	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,G
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 09/28/2017 04:54:01

Diagnostics Check		OK	04:47
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:49
Subject Sample	# 1	0.115	04:49
Breath Volume		3.050	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:51
Subject Sample	# 2	0.112	04:52
Breath Volume		3.242	04:52
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:53
Diagnostics Check		OK	04:54

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 09/29/2017 03:59:03

Diagnostics Check	OK	03:56	
Air Blank	0.000	03:56	
Control Test	0.080	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:58	
Control Test	0.078	03:58	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:59	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO, YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/30/2017 20:38:28

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.196	20:33
Breath Volume	2.125	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:35
Subject Sample # 2	0.183	20:36
Breath Volume	1.832	20:36
Air Blank	0.000	20:37
Control Test	0.080	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO, YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 09/30/2017 20:53:01

Diagnostics Check	OK	20:42
Air Blank	0.000	20:43
Control Test	0.080	20:43
Air Blank	0.000	20:44
Subject Sample # 1	0.165	20:45
Breath Volume	1.339	20:45
Air Blank	0.000	20:45
Air Blank	0.000	20:47
Subject Sample # 2	0.188	20:48
Breath Volume	1.234	20:48
Air Blank	0.000	20:48
Air Blank	0.000	20:50
Subject Sample # 3	0.189	20:51
Breath Volume	2.312	20:51
Air Blank	0.000	20:52
Control Test	0.080	20:52
Air Blank	0.000	20:52
Diagnostics Check	OK	20:53

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: MATE

DATE/TIME OF TEST: 09/30/2017 22:12:16

Diagnostics Check		OK	22:05
Air Blank		0.000	22:06
Control Test		0.080	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.281	22:07
Breath Volume		1.566	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:09
Subject Sample	# 2	0.283	22:10
Breath Volume		1.664	22:10
Air Blank		0.000	22:11
Control Test		0.080	22:11
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RODRIGUEZ,N
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 09/02/2017 20:04:01

Diagnostics Check	OK	20:03	
Air Blank		20:03	RFI Detect
Air Blank	0.000	20:04	

Data Download Date/Time: 09/15/2017 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: RODRIGUEZ,N
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:25

SUBJECT'S LAST NAME: FRIAS

DATE/TIME OF TEST: 09/02/2017 20:12:56

Diagnostics Check	OK	20:05
Air Blank	0.000	20:05
Control Test	0.079	20:06
Air Blank	0.000	20:06
Subject Sample # 1	0.342	20:07
Breath Volume	1.683	20:07
Air Blank	0.000	20:08
Air Blank	0.000	20:10
Subject Sample # 2	0.347	20:11
Breath Volume	1.945	20:11
Air Blank	0.000	20:12
Control Test	0.079	20:12
Air Blank	0.000	20:12
Diagnostics Check	OK	20:12

Data Download Date/Time: 09/15/2017 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: JUAN CRISTOBAL

DATE/TIME OF TEST: 09/03/2017 01:47:48

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.152	01:43
Breath Volume	1.824	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.152	01:46
Breath Volume	2.375	01:46
Air Blank	0.000	01:46
Control Test	0.080	01:47
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 09/15/2017 11:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILPOT,C,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: RODRIGUEZ PEREZ

DATE/TIME OF TEST: 09/03/2017 12:41:46

Diagnostics Check	OK	12:35
Air Blank	0.000	12:36
Control Test	0.079	12:36
Air Blank	0.000	12:36
Subject Sample # 1	0.162	12:37
Breath Volume	1.847	12:37
Air Blank	0.000	12:37
Air Blank	0.000	12:39
Subject Sample # 2	0.164	12:40
Breath Volume	1.296	12:40
Air Blank	0.000	12:40
Control Test	0.079	12:41
Air Blank	0.000	12:41
Diagnostics Check	OK	12:41

Data Download Date/Time: 09/15/2017 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILPOT,C,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/04/2017 22:11:05

Diagnostics Check	OK	22:05	
Air Blank	0.000	22:05	
Control Test	0.079	22:06	
Air Blank	0.000	22:06	
Subject Sample # 1	0.203	22:07	
Breath Volume	1.378	22:07	
Air Blank	0.000	22:07	
Air Blank		22:09	Improper Sample
Air Blank	0.000	22:10	
Control Test	0.080	22:10	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 09/15/2017 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PHILPOT,C,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PHILPOT,C
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 09/04/2017 22:21:44

Diagnostics Check	OK	22:14
Air Blank	0.000	22:14
Control Test	0.078	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.185	22:17
Breath Volume	1.351	22:17
Air Blank	0.000	22:18
Air Blank	0.000	22:19
Subject Sample # 2	0.199	22:20
Breath Volume	1.789	22:20
Air Blank	0.000	22:20
Control Test	0.078	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 09/15/2017 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAPANES,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/06/2017 01:53:11

Diagnostics Check		OK	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Subject Sample	# 1	0.244	01:47
Breath Volume		2.101	01:47
Air Blank		0.000	01:48
Air Blank		0.000	01:50
Subject Sample	# 2	0.230	01:51
Breath Volume		1.925	01:51
Air Blank		0.000	01:52
Control Test		0.080	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 09/15/2017 11:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: EKMEIRO

DATE/TIME OF TEST: 09/03/2017 02:03:44

Diagnostics Check	OK	02:00	
Air Blank	0.000	02:00	
Control Test	0.078	02:00	
Air Blank	0.000	02:01	
Subject Sample	# 1	02:02	Subject Test Refused
Breath Volume	0.000	02:02	
Air Blank		02:03	RFI Detect
Air Blank	0.000	02:03	

Data Download Date/Time: 09/15/2017 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLIMAK,M,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: SLIMAK,M
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: EKMEIRO

DATE/TIME OF TEST: 09/03/2017 02:09:34

Diagnostics Check	OK	02:06	
Air Blank	0.000	02:07	
Control Test	0.079	02:07	
Air Blank	0.000	02:08	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.000	02:08	
Air Blank	0.000	02:08	
Control Test	0.078	02:08	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:09	

Data Download Date/Time: 09/15/2017 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWSON, GLADSTON,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LAWSON, GLADSTON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: AGUILAR

DATE/TIME OF TEST: 09/04/2017 03:42:21

Diagnostics Check	OK	03:35
Air Blank	0.000	03:35
Control Test	0.079	03:36
Air Blank	0.000	03:36
Subject Sample # 1	0.049	03:37
Breath Volume	2.230	03:37
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.047	03:40
Breath Volume	2.250	03:40
Air Blank	0.000	03:41
Control Test	0.078	03:41
Air Blank	0.000	03:42
Diagnostics Check	OK	03:42

Data Download Date/Time: 09/15/2017 11:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORRELL,BRIAN,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 09/04/2017 22:31:06

Diagnostics Check	OK	22:27	
Air Blank	0.000	22:28	
Control Test	0.078	22:28	
Air Blank	0.000	22:29	
Subject Sample	# 1	22:30	RFI Detect
Breath Volume	0.132	22:30	
Air Blank		22:30	RFI Detect
Air Blank	0.000	22:31	

Data Download Date/Time: 09/15/2017 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SORRELL,BRIAN,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HERNANDEZ,LUIS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22315080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: RODRIGUEZ JR

DATE/TIME OF TEST: 09/04/2017 22:44:30

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:33	
Control Test		0.078	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 1	0.149	22:37	Volume Not Met
Breath Volume		0.070	22:37	
Air Blank		0.000	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 2	0.150	22:42	Volume Not Met
Breath Volume		0.332	22:42	
Air Blank		0.000	22:43	
Control Test		0.079	22:43	
Air Blank		0.000	22:44	
Diagnostics Check		OK	22:44	

Data Download Date/Time: 09/15/2017 11:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: BILLU

DATE/TIME OF TEST: 09/02/2017 02:02:36

Diagnostics Check		OK	01:50	
Air Blank		0.000	01:51	
Control Test		0.079	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1	0.103	01:55	Volume Not Met
Breath Volume		0.175	01:55	
Air Blank		0.000	01:56	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.106	02:00	Volume Not Met
Breath Volume		0.574	02:00	
Air Blank		0.000	02:01	
Control Test		0.078	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 09/30/2017 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MIKE
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19615080A2
DRY GAS EXPIRY DATE: 09/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: MARINHO

DATE/TIME OF TEST: 09/02/2017 06:22:23

Diagnostics Check	OK	06:10	
Air Blank	0.000	06:11	
Control Test	0.078	06:11	
Air Blank	0.000	06:12	
Subject Sample	# 1	06:15	No Sample Provided
Breath Volume	0.000	06:15	
Air Blank	0.000	06:15	
Air Blank	0.000	06:17	
Subject Sample	# 2	06:21	No Sample Provided
Breath Volume	0.000	06:21	
Air Blank	0.000	06:21	
Control Test	0.079	06:21	
Air Blank	0.000	06:22	
Diagnostics Check	OK	06:22	

Data Download Date/Time: 09/30/2017 20:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVADO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MANINT

DATE/TIME OF TEST: 09/15/2017 22:26:08

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:20	
Control Test	0.080	22:21	
Air Blank	0.000	22:21	
Subject Sample	# 1 0.295	22:24	Volume Not Met
Breath Volume	0.179	22:24	
Air Blank		22:25	Purge Fail
Air Blank	0.000	22:26	

Data Download Date/Time: 09/30/2017 20:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVADO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MANINT

DATE/TIME OF TEST: 09/15/2017 22:44:04

Diagnostics Check		OK	22:28	
Air Blank		0.000	22:29	
Control Test		0.077	22:29	
Air Blank		0.000	22:29	
Subject Sample	# 1	0.258	22:33	Volume Not Met
Breath Volume		0.730	22:33	
Air Blank		0.000	22:33	
Air Blank		0.000	22:35	
Subject Sample	# 2	0.299	22:37	
Breath Volume		1.128	22:37	
Air Blank		0.000	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 3	0.294	22:42	
Breath Volume		1.410	22:42	
Air Blank		0.000	22:43	
Control Test		0.076	22:43	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:44	

Data Download Date/Time: 09/30/2017 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AZEVEDO,MIKE,C
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DOSTALY,S
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 08/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 09/24/2017 06:26:32

Diagnostics Check	OK	06:19
Air Blank	0.000	06:20
Control Test	0.081	06:20
Air Blank	0.000	06:21
Subject Sample # 1	0.066	06:21
Breath Volume	1.406	06:21
Air Blank	0.000	06:22
Air Blank	0.000	06:24
Subject Sample # 2	0.066	06:24
Breath Volume	1.894	06:24
Air Blank	0.000	06:25
Control Test	0.080	06:25
Air Blank	0.000	06:26
Diagnostics Check	OK	06:26

Data Download Date/Time: 09/30/2017 20:55

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,REID,B
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KLEIS,KEVIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 08/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1568991
DRY GAS EXPIRY DATE: 10/01/2016
VIOLATION CODE: BUI
ARREST TIME: 18:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: ERNST

DATE/TIME OF TEST: 09/02/2017 19:41:33

Diagnostics Check	OK	19:34
Air Blank	0.000	19:35
Control Test	0.080	19:35
Air Blank	0.000	19:36
Subject Sample # 1	0.073	19:37
Breath Volume	2.027	19:37
Air Blank	0.000	19:37
Air Blank	0.000	19:39
Subject Sample # 2	0.063	19:40
Breath Volume	2.261	19:40
Air Blank	0.000	19:40
Control Test	0.079	19:40
Air Blank	0.000	19:41
Diagnostics Check	OK	19:41

Data Download Date/Time: 09/29/2017 11:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: GAILLARD

DATE/TIME OF TEST: 09/07/2017 00:08:18

Diagnostics Check	OK	00:04	
Air Blank	0.000	00:05	
Control Test	0.078	00:05	
Air Blank	0.000	00:06	
Subject Sample	# 1	00:06	Subject Test Refused
Breath Volume	0.000	00:06	
Air Blank	0.000	00:07	
Control Test	0.078	00:07	
Air Blank	0.000	00:08	
Diagnostics Check	OK	00:08	

Data Download Date/Time: 09/18/2017 10:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000907

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 08/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: BARATH

DATE/TIME OF TEST: 09/07/2017 23:34:28

Diagnostics Check	OK	23:25	
Air Blank	0.000	23:26	
Control Test	0.078	23:26	
Air Blank	0.000	23:27	
Subject Sample	# 1	23:30	No Sample Provided
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Air Blank	0.000	23:32	
Subject Sample	# 2	23:33	Subject Test Refused
Breath Volume	0.000	23:33	
Air Blank	0.000	23:33	
Control Test	0.078	23:33	
Air Blank	0.000	23:34	
Diagnostics Check	OK	23:34	

Data Download Date/Time: 09/18/2017 10:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/21/2017 03:40:22

Diagnostics Check	OK	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:34
Subject Sample # 1	0.109	03:35
Breath Volume	2.121	03:35
Air Blank	0.000	03:35
Air Blank	0.000	03:37
Subject Sample # 2	0.106	03:38
Breath Volume	2.601	03:38
Air Blank	0.000	03:39
Control Test	0.078	03:39
Air Blank	0.000	03:40
Diagnostics Check	OK	03:40

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARGO,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: BECHTOL

DATE/TIME OF TEST: 09/24/2017 00:41:07

Diagnostics Check	OK	00:29
Air Blank	0.000	00:30
Control Test	0.078	00:30
Air Blank	0.000	00:31
Subject Sample # 1	0.199	00:32
Breath Volume	1.324	00:32
Air Blank	0.000	00:32
Air Blank	0.000	00:34
Subject Sample # 2	0.231	00:35
Breath Volume	2.781	00:35
Air Blank	0.000	00:36
Air Blank	0.000	00:38
Subject Sample # 3	0.223	00:39
Breath Volume	1.863	00:39
Air Blank	0.000	00:40
Control Test	0.077	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: CARRILLO,LUIS
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GRONDIN

DATE/TIME OF TEST: 09/25/2017 22:36:13

Diagnostics Check	OK	22:29
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:31
Subject Sample # 1	0.243	22:31
Breath Volume	2.046	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:34
Subject Sample # 2	0.242	22:34
Breath Volume	2.253	22:34
Air Blank	0.000	22:35
Control Test	0.079	22:35
Air Blank	0.000	22:36
Diagnostics Check	OK	22:36

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 09/27/2017 22:07:00

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.079	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.234	22:02
Breath Volume		1.644	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 2	0.239	22:05
Breath Volume		1.757	22:05
Air Blank		0.000	22:06
Control Test		0.079	22:06
Air Blank		0.000	22:06
Diagnostics Check		OK	22:07

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARGO,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: MARKOWSKI

DATE/TIME OF TEST: 09/28/2017 00:06:09

Diagnostics Check	OK	23:59
Air Blank	0.000	23:59
Control Test	0.080	00:00
Air Blank	0.000	00:00
Subject Sample # 1	0.000	00:01
Breath Volume	3.363	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.000	00:04
Breath Volume	3.406	00:04
Air Blank	0.000	00:05
Control Test	0.080	00:05
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM,JOHN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: MADRID BARAHONA

DATE/TIME OF TEST: 09/30/2017 05:41:19

Diagnostics Check	OK	05:34
Air Blank	0.000	05:35
Control Test	0.079	05:35
Air Blank	0.000	05:35
Subject Sample # 1	0.124	05:36
Breath Volume	2.121	05:36
Air Blank	0.000	05:37
Air Blank	0.000	05:39
Subject Sample # 2	0.113	05:39
Breath Volume	1.964	05:39
Air Blank	0.000	05:40
Control Test	0.075	05:40
Air Blank	0.000	05:41
Diagnostics Check	OK	05:41

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000937

INSTRUMENT REGISTERED TO: SANIBEL ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHWEWOOD,JOHN,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: SHERWOOD,JOHN
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 08/29/2000

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 05514080A1
DRY GAS EXPIRY DATE: 12/05/2016
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: LOZO

DATE/TIME OF TEST: 09/01/2017 18:32:38

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.080	18:26
Air Blank	0.000	18:27
Subject Sample # 1	0.198	18:27
Breath Volume	2.148	18:27
Air Blank	0.000	18:28
Air Blank	0.000	18:30
Subject Sample # 2	0.197	18:31
Breath Volume	1.820	18:31
Air Blank	0.000	18:31
Control Test	0.082	18:32
Air Blank	0.000	18:32
Diagnostics Check	OK	18:32

Data Download Date/Time: 09/02/2016 02:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: OUELLETTE,SARAH
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 09/27/2017 03:17:20

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.078	03:09
Air Blank	0.000	03:10
Subject Sample # 1	0.152	03:12
Breath Volume	1.421	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.156	03:15
Breath Volume	2.628	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:17
Diagnostics Check	OK	03:17

Data Download Date/Time: 10/25/2017 16:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,LAUREN,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: LAKEMAN,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: ABELSON

DATE/TIME OF TEST: 09/28/2017 23:36:36

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.078	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.199	23:32
Breath Volume	1.527	23:32
Air Blank	0.000	23:32
Air Blank	0.000	23:34
Subject Sample # 2	0.201	23:35
Breath Volume	1.667	23:35
Air Blank	0.000	23:35
Control Test	0.077	23:36
Air Blank	0.000	23:36
Diagnostics Check	OK	23:36

Data Download Date/Time: 10/25/2017 16:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,KEVIN,G
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CATANI,JOHN
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: RAMIREZ-RAMOS

DATE/TIME OF TEST: 09/14/2017 05:03:32

Diagnostics Check		OK	04:57
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:58
Subject Sample	# 1	0.145	04:59
Breath Volume		2.839	04:59
Air Blank		0.000	04:59
Air Blank		0.000	05:01
Subject Sample	# 2	0.154	05:02
Breath Volume		2.589	05:02
Air Blank		0.000	05:02
Control Test		0.079	05:02
Air Blank		0.000	05:03
Diagnostics Check		OK	05:03

Data Download Date/Time: 10/03/2017 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,KEVIN,G
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MORALES,GUALBERTO
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A2
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: PETITFRERE

DATE/TIME OF TEST: 09/23/2017 23:09:49

Diagnostics Check	OK	23:03
Air Blank	0.000	23:03
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.128	23:05
Breath Volume	1.976	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:07
Subject Sample # 2	0.132	23:08
Breath Volume	2.949	23:08
Air Blank	0.000	23:08
Control Test	0.078	23:09
Air Blank	0.000	23:09
Diagnostics Check	OK	23:09

Data Download Date/Time: 10/03/2017 11:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROHRAFF,HARRY,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: STRUTHERS

DATE/TIME OF TEST: 09/01/2017 21:32:13

Diagnostics Check	OK	21:25
Air Blank	0.000	21:25
Control Test	0.079	21:26
Air Blank	0.000	21:26
Subject Sample # 1	0.000	21:27
Breath Volume	3.519	21:27
Air Blank	0.000	21:28
Air Blank	0.000	21:30
Subject Sample # 2	0.000	21:30
Breath Volume	2.859	21:30
Air Blank	0.000	21:31
Control Test	0.079	21:31
Air Blank	0.000	21:32
Diagnostics Check	OK	21:32

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: SUREDA-MACHADO

DATE/TIME OF TEST: 09/02/2017 04:44:22

Diagnostics Check	OK	04:35	
Air Blank	0.000	04:36	
Control Test	0.079	04:36	
Air Blank	0.000	04:37	
Subject Sample	# 1	04:40	No Sample Provided
Breath Volume	0.000	04:40	
Air Blank	0.000	04:40	
Air Blank	0.000	04:42	
Subject Sample	# 2	04:43	Subject Test Refused
Breath Volume	0.000	04:43	
Air Blank	0.000	04:43	
Control Test	0.079	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDUFFIE,B.,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BAXTER,J.
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: ALTEEN

DATE/TIME OF TEST: 09/02/2017 12:58:49

Diagnostics Check	OK	12:52
Air Blank	0.000	12:53
Control Test	0.079	12:53
Air Blank	0.000	12:54
Subject Sample # 1	0.000	12:54
Breath Volume	1.976	12:54
Air Blank	0.000	12:54
Air Blank	0.000	12:56
Subject Sample # 2	0.000	12:57
Breath Volume	2.375	12:57
Air Blank	0.000	12:57
Control Test	0.079	12:58
Air Blank	0.000	12:58
Diagnostics Check	OK	12:58

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: PLUSSA,KYLE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: BUI
ARREST TIME: 19:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:03

SUBJECT'S LAST NAME: LASTINGER

DATE/TIME OF TEST: 09/02/2017 20:29:53

Diagnostics Check	OK	20:23
Air Blank	0.000	20:23
Control Test	0.080	20:24
Air Blank	0.000	20:24
Subject Sample # 1	0.067	20:25
Breath Volume	3.464	20:25
Air Blank	0.000	20:25
Air Blank	0.000	20:27
Subject Sample # 2	0.066	20:28
Breath Volume	3.546	20:28
Air Blank	0.000	20:28
Control Test	0.079	20:29
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 09/03/2017 00:36:58

Diagnostics Check	OK	00:30
Air Blank	0.000	00:31
Control Test	0.079	00:31
Air Blank	0.000	00:31
Subject Sample # 1	0.271	00:32
Breath Volume	1.957	00:32
Air Blank	0.000	00:33
Air Blank	0.000	00:34
Subject Sample # 2	0.271	00:35
Breath Volume	1.738	00:35
Air Blank	0.000	00:36
Control Test	0.079	00:36
Air Blank	0.000	00:36
Diagnostics Check	OK	00:36

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/03/2017 22:46:19

Diagnostics Check		OK	22:31	
Air Blank		0.000	22:31	
Control Test		0.079	22:32	
Air Blank		0.000	22:32	
Subject Sample	# 1	0.247	22:35	Volume Not Met
Breath Volume		0.894	22:35	
Air Blank		0.000	22:36	
Air Blank		0.000	22:38	
Subject Sample	# 2	0.294	22:41	
Breath Volume		1.976	22:41	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 3	0.280	22:44	
Breath Volume		2.167	22:44	
Air Blank		0.000	22:45	
Control Test		0.077	22:45	
Air Blank		0.000	22:46	
Diagnostics Check		OK	22:46	

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JESSE,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ESTRADA

DATE/TIME OF TEST: 09/03/2017 23:19:18

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.079	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.000	23:14
Breath Volume	4.167	23:14
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.000	23:17
Breath Volume	4.078	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,COLTON,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCGILVREY,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SHAMOON

DATE/TIME OF TEST: 09/04/2017 03:59:49

Diagnostics Check		OK	03:53
Air Blank		0.000	03:54
Control Test		0.079	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.128	03:55
Breath Volume		1.894	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:57
Subject Sample	# 2	0.122	03:58
Breath Volume		1.933	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000944

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CHAPMAN,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 08/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: GALLEN

DATE/TIME OF TEST: 09/04/2017 05:36:41

Diagnostics Check	OK	05:27
Air Blank	0.000	05:28
Control Test	0.079	05:28
Air Blank	0.000	05:29
Subject Sample # 1	0.154	05:30
Breath Volume	1.261	05:30
Air Blank	0.000	05:31
Air Blank	0.000	05:32
Subject Sample # 2	0.149	05:34
Breath Volume	1.398	05:34
Air Blank	0.000	05:35
Control Test	0.080	05:36
Air Blank	0.000	05:36
Diagnostics Check	OK	05:36

Data Download Date/Time: 09/05/2017 10:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: KNAPP,E
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: RS

DATE/TIME OF TEST: 09/21/2017 18:35:14

Diagnostics Check	OK	18:26
Air Blank	0.000	18:27
Control Test	0.082	18:27
Air Blank	0.000	18:28
Subject Sample # 1	0.213	18:30
Breath Volume	1.914	18:30
Air Blank	0.000	18:31
Air Blank	0.000	18:33
Subject Sample # 2	0.212	18:33
Breath Volume	1.628	18:33
Air Blank	0.000	18:34
Control Test	0.080	18:34
Air Blank	0.000	18:35
Diagnostics Check	OK	18:35

Data Download Date/Time: 09/22/2017 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SWARTZENTRUBER,B
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 09/21/2017 22:01:41

Diagnostics Check	OK	21:50	
Air Blank	0.000	21:50	
Control Test	0.081	21:51	
Air Blank	0.000	21:51	
Subject Sample	# 1	21:54	No Sample Provided
Breath Volume	0.000	21:54	
Air Blank	0.000	21:55	
Air Blank	0.000	21:57	
Subject Sample	# 2	22:00	No Sample Provided
Breath Volume	0.000	22:00	
Air Blank	0.000	22:00	
Control Test	0.081	22:01	
Air Blank	0.000	22:01	
Diagnostics Check	OK	22:01	

Data Download Date/Time: 09/22/2017 08:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROGERS,JUSTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: HOUSE

DATE/TIME OF TEST: 09/22/2017 23:00:09

Diagnostics Check	OK	22:53
Air Blank	0.000	22:53
Control Test	0.081	22:53
Air Blank	0.000	22:54
Subject Sample # 1	0.232	22:55
Breath Volume	1.582	22:55
Air Blank	0.000	22:55
Air Blank	0.000	22:57
Subject Sample # 2	0.239	22:58
Breath Volume	2.457	22:58
Air Blank	0.000	22:59
Control Test	0.081	22:59
Air Blank	0.000	23:00
Diagnostics Check	OK	23:00

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: GRAVEL,CPL
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 09/24/2017 22:32:17

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.081	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.000	22:27
Breath Volume	1.273	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.000	22:30
Breath Volume	1.109	22:30
Air Blank	0.000	22:31
Control Test	0.082	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: LONESKY,M
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: SERNA

DATE/TIME OF TEST: 09/26/2017 01:02:34

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.082	00:56
Air Blank	0.000	00:57
Subject Sample # 1	0.138	00:58
Breath Volume	1.574	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.137	01:01
Breath Volume	1.214	01:01
Air Blank	0.000	01:01
Control Test	0.081	01:01
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: CIOLINO,L
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: MCKLVEEN

DATE/TIME OF TEST: 09/27/2017 01:34:26

Diagnostics Check	OK	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Subject Sample # 1	0.295	01:29
Breath Volume	2.054	01:29
Air Blank	0.000	01:30
Air Blank	0.000	01:32
Subject Sample # 2	0.275	01:32
Breath Volume	2.109	01:32
Air Blank	0.000	01:33
Control Test	0.081	01:33
Air Blank	0.000	01:34
Diagnostics Check	OK	01:34

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,ANGIE,S
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROGERS,JUSTIN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:50

SUBJECT'S LAST NAME: ROESSLER

DATE/TIME OF TEST: 09/27/2017 11:16:57

Diagnostics Check	OK	11:10
Air Blank	0.000	11:11
Control Test	0.082	11:11
Air Blank	0.000	11:12
Subject Sample # 1	0.356	11:12
Breath Volume	2.445	11:12
Air Blank	0.000	11:13
Air Blank	0.000	11:15
Subject Sample # 2	0.352	11:15
Breath Volume	2.402	11:15
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:16
Diagnostics Check	OK	11:16

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGAN,MELISSA,
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: TRUFAN,J
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: GREEK

DATE/TIME OF TEST: 09/30/2017 04:54:28

Diagnostics Check	OK	04:47
Air Blank	0.000	04:48
Control Test	0.081	04:48
Air Blank	0.000	04:49
Subject Sample # 1	0.150	04:50
Breath Volume	1.859	04:50
Air Blank	0.000	04:50
Air Blank	0.000	04:52
Subject Sample # 2	0.149	04:53
Breath Volume	2.257	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:53
Air Blank	0.000	04:54
Diagnostics Check	OK	04:54

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI,R
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: MERINO

DATE/TIME OF TEST: 09/01/2017 23:16:19

Diagnostics Check	OK	23:07
Air Blank	0.000	23:07
Control Test	0.078	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.133	23:11
Breath Volume	1.960	23:11
Air Blank	0.000	23:12
Air Blank	0.000	23:13
Subject Sample # 2	0.126	23:14
Breath Volume	1.554	23:14
Air Blank	0.000	23:15
Control Test	0.075	23:15
Air Blank	0.000	23:16
Diagnostics Check	OK	23:16

Data Download Date/Time: 09/29/2017 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALGADO,T
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 09/11/2017 23:44:35

Diagnostics Check	OK	23:32	
Air Blank	0.000	23:32	
Control Test	0.077	23:33	
Air Blank	0.000	23:33	
Subject Sample	# 1	23:36	No Sample Provided
Breath Volume	0.000	23:36	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 2	0.203	23:39
Breath Volume	3.582	23:39	
Air Blank	0.000	23:40	
Air Blank	0.000	23:42	
Subject Sample	# 3	0.199	23:42
Breath Volume	3.832	23:42	
Air Blank	0.000	23:43	
Control Test	0.077	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 09/29/2017 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,JOEL,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: SALGADO,WILMER
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: BONILLA

DATE/TIME OF TEST: 09/12/2017 07:31:36

Diagnostics Check	OK	07:25
Air Blank	0.000	07:26
Control Test	0.078	07:26
Air Blank	0.000	07:26
Subject Sample # 1	0.072	07:27
Breath Volume	1.753	07:27
Air Blank	0.000	07:27
Air Blank	0.000	07:29
Subject Sample # 2	0.075	07:29
Breath Volume	2.519	07:29
Air Blank	0.000	07:30
Control Test	0.076	07:30
Air Blank	0.000	07:31
Diagnostics Check	OK	07:31

Data Download Date/Time: 09/29/2017 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,J,D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING,JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 09/23/2017 17:33:25

Diagnostics Check	OK	17:27
Air Blank	0.000	17:27
Control Test	0.077	17:28
Air Blank	0.000	17:28
Subject Sample # 1	0.169	17:28
Breath Volume	1.746	17:28
Air Blank	0.000	17:29
Air Blank	0.000	17:31
Subject Sample # 2	0.177	17:31
Breath Volume	2.535	17:31
Air Blank	0.000	17:32
Control Test	0.076	17:32
Air Blank	0.000	17:33
Diagnostics Check	OK	17:33

Data Download Date/Time: 09/29/2017 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BONILLA, JOSUE, D
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDING, JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 16:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WEAVER

DATE/TIME OF TEST: 09/24/2017 00:46:29

Diagnostics Check	OK	00:40
Air Blank	0.000	00:40
Control Test	0.077	00:41
Air Blank	0.000	00:41
Subject Sample # 1	0.027	00:42
Breath Volume	2.855	00:42
Air Blank	0.000	00:42
Air Blank	0.000	00:44
Subject Sample # 2	0.025	00:44
Breath Volume	3.007	00:44
Air Blank	0.000	00:45
Control Test	0.077	00:45
Air Blank	0.000	00:46
Diagnostics Check	OK	00:46

Data Download Date/Time: 09/29/2017 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER, DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: CUTCLIFFE, BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 09/01/2017 00:22:36

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.051	00:18
Breath Volume	1.308	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.049	00:21
Breath Volume	1.312	00:21
Air Blank	0.000	00:21
Control Test	0.082	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 10/02/2017 06:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000949

INSTRUMENT REGISTERED TO: TALLAHASSEE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KELLER,DAVID,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: KELLER,DAVID
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/31/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34416090A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:04

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 09/21/2017 06:34:42

Diagnostics Check		OK	06:28
Air Blank		0.000	06:28
Control Test		0.081	06:28
Air Blank		0.000	06:29
Subject Sample	# 1	0.201	06:30
Breath Volume		2.449	06:30
Air Blank		0.000	06:30
Air Blank		0.000	06:32
Subject Sample	# 2	0.205	06:33
Breath Volume		2.648	06:33
Air Blank		0.000	06:33
Control Test		0.081	06:34
Air Blank		0.000	06:34
Diagnostics Check		OK	06:34

Data Download Date/Time: 10/02/2017 06:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: PFEIFER

DATE/TIME OF TEST: 09/30/2017 01:23:43

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:22	
Control Test	0.082	01:22	
Air Blank		01:23	RFI Detect
Air Blank	0.000	01:23	

Data Download Date/Time: 10/24/2017 08:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000952

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAUVERSBURK,BRENNA,M
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HAUVERSBURK,BRENNA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: PFEIFER

DATE/TIME OF TEST: 09/30/2017 01:32:01

Diagnostics Check	OK	01:26
Air Blank	0.000	01:26
Control Test	0.083	01:26
Air Blank	0.000	01:27
Subject Sample # 1	0.000	01:27
Breath Volume	1.671	01:27
Air Blank	0.000	01:28
Air Blank	0.000	01:30
Subject Sample # 2	0.000	01:30
Breath Volume	1.949	01:30
Air Blank	0.000	01:31
Control Test	0.081	01:31
Air Blank	0.000	01:31
Diagnostics Check	OK	01:32

Data Download Date/Time: 10/24/2017 08:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT, JOSHUA, N
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: MATTHEWS, DALTON
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:28

SUBJECT'S LAST NAME: COX

DATE/TIME OF TEST: 09/04/2017 17:39:32

Diagnostics Check	OK	17:28	
Air Blank	0.000	17:28	
Control Test	0.079	17:28	
Air Blank	0.000	17:29	
Subject Sample	# 1	17:32	No Sample Provided
Breath Volume	0.000	17:32	
Air Blank	0.000	17:33	
Air Blank	0.000	17:35	
Subject Sample	# 2	17:38	No Sample Provided
Breath Volume	0.000	17:38	
Air Blank	0.000	17:38	
Control Test	0.079	17:38	
Air Blank	0.000	17:39	
Diagnostics Check	OK	17:39	

Data Download Date/Time: 09/20/2017 15:46

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AH SAM,GREGORY,S
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: AH SAM,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/22/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: DAROZA

DATE/TIME OF TEST: 09/08/2017 22:14:59

Diagnostics Check	OK	22:11	
Air Blank	0.000	22:11	
Control Test	0.078	22:12	
Air Blank	0.000	22:12	
Subject Sample	# 1	22:13	Subject Test Refused
Breath Volume	0.000	22:13	
Air Blank	0.000	22:14	
Control Test	0.078	22:14	
Air Blank	0.000	22:14	
Diagnostics Check	OK	22:14	

Data Download Date/Time: 09/20/2017 15:46

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCNEIL,DARRIN,J
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: MOLINA,ALFREDO
ARRESTING OFFICER AGENCY: Clewiston PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 11:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:19

SUBJECT'S LAST NAME: SALINAS HERNANDEZ

DATE/TIME OF TEST: 09/24/2017 13:02:42

Diagnostics Check	OK	12:56
Air Blank	0.000	12:56
Control Test	0.081	12:57
Air Blank	0.000	12:57
Subject Sample # 1	0.375	12:58
Breath Volume	2.671	12:58
Air Blank	0.000	12:58
Air Blank	0.000	13:00
Subject Sample # 2	0.374	13:01
Breath Volume	3.457	13:01
Air Blank	0.000	13:01
Control Test	0.079	13:02
Air Blank	0.000	13:02
Diagnostics Check	OK	13:02

Data Download Date/Time: 10/05/2017 14:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: WILLIAMS,N,

BREATH TEST OPERATOR AGENCY: Clewiston PD

ARRESTING OFFICER: WILLIAMS,N

ARRESTING OFFICER AGENCY: Clewiston PD

LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 644476

DRY GAS EXPIRY DATE: 12/01/2019

VIOLATION CODE: DUI

ARREST TIME: 19:37

GENDER: Male

OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: UDONI

DATE/TIME OF TEST: 09/30/2017 20:22:49

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.081	20:16
Air Blank		0.000	20:16
Subject Sample	# 1	0.242	20:18
Breath Volume		1.675	20:18
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.251	20:21
Breath Volume		1.613	20:21
Air Blank		0.000	20:21
Control Test		0.080	20:22
Air Blank		0.000	20:22
Diagnostics Check		OK	20:22

Data Download Date/Time: 10/05/2017 14:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEAS, BRANDYN, M
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: DEAS, BRANDYN
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: COGGINS

DATE/TIME OF TEST: 09/03/2017 03:51:54

Diagnostics Check	OK	03:44
Air Blank	0.000	03:45
Control Test	0.080	03:45
Air Blank	0.000	03:46
Subject Sample # 1	0.138	03:46
Breath Volume	4.550	03:46
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.131	03:50
Breath Volume	3.128	03:50
Air Blank	0.000	03:50
Control Test	0.080	03:51
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 09/25/2017 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMIDT,M,
BREATH TEST OPERATOR AGENCY: Tallahassee PD
ARRESTING OFFICER: SMIDT,M
ARRESTING OFFICER AGENCY: Tallahassee PD
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: PASSALACQUA

DATE/TIME OF TEST: 09/03/2017 04:13:29

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:07
Subject Sample # 1	0.242	04:09
Breath Volume	2.218	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.237	04:11
Breath Volume	3.066	04:11
Air Blank	0.000	04:12
Control Test	0.082	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 09/25/2017 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,MARY,P
BREATH TEST OPERATOR AGENCY: Leon County SO
ARRESTING OFFICER: TURNER,JOSHUA
ARRESTING OFFICER AGENCY: Leon County SO
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: CROSBY

DATE/TIME OF TEST: 09/05/2017 03:58:34

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.081	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1	03:57	Subject Test Refused
Breath Volume	0.000	03:57	
Air Blank	0.000	03:57	
Control Test	0.081	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 09/25/2017 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: MIMS,CHARLES
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 08/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:33

SUBJECT'S LAST NAME: FESHCHENKO

DATE/TIME OF TEST: 09/24/2017 02:02:02

Diagnostics Check		OK	01:53	
Air Blank		0.000	01:53	
Control Test		0.080	01:54	
Air Blank		0.000	01:54	
Subject Sample	# 1	0.195	01:57	Volume Not Met
Breath Volume		0.714	01:57	
Air Blank		0.000	01:58	
Air Blank		0.000	02:00	
Subject Sample	# 2		02:00	Subject Test Refused
Breath Volume		0.000	02:00	
Air Blank		0.000	02:01	
Control Test		0.077	02:01	
Air Blank		0.000	02:01	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 09/25/2017 10:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAINT LOUIS, JUNIOR,
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: VARNER, TYLER
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/26/2017 04:03:25

Diagnostics Check	OK	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:01	
Subject Sample	# 1	04:02	Subject Test Refused
Breath Volume	0.000	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Diagnostics Check	OK	04:03	

Data Download Date/Time: 10/25/2017 11:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000961

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENING,FREDRICK,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: HENING,FREDRICK
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 08/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: KIRTON

DATE/TIME OF TEST: 09/13/2017 21:10:20

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.000	21:05
Breath Volume	1.183	21:05
Air Blank	0.000	21:05
Air Blank	0.000	21:07
Subject Sample # 2	0.000	21:08
Breath Volume	2.031	21:08
Air Blank	0.000	21:09
Control Test	0.080	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 09/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: STONE, THOMAS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/25/2017 00:04:02

Diagnostics Check	OK	00:01	
Air Blank	0.000	00:01	
Control Test	0.080	00:01	
Air Blank	0.000	00:02	
Subject Sample	# 1	00:02	RFI Detect
Breath Volume	0.000	00:02	
Air Blank	0.000	00:03	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Diagnostics Check	OK	00:04	

Data Download Date/Time: 10/09/2017 11:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELMS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: STONE, THOMAS
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 09/25/2017 00:09:36

Diagnostics Check	OK	00:06	
Air Blank	0.000	00:07	
Control Test	0.080	00:07	
Air Blank	0.000	00:07	
Subject Sample	# 1	00:08	Subject Test Refused
Breath Volume	0.000	00:08	
Air Blank	0.000	00:08	
Control Test	0.080	00:09	
Air Blank	0.000	00:09	
Diagnostics Check	OK	00:09	

Data Download Date/Time: 10/09/2017 11:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JAMES, SETH
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: MALDONADO NUNEZ

DATE/TIME OF TEST: 09/01/2017 23:37:19

Diagnostics Check	OK	23:30
Air Blank	0.000	23:30
Control Test	0.080	23:31
Air Blank	0.000	23:31
Subject Sample # 1	0.106	23:32
Breath Volume	2.289	23:32
Air Blank	0.000	23:33
Air Blank	0.000	23:35
Subject Sample # 2	0.101	23:35
Breath Volume	1.710	23:35
Air Blank	0.000	23:36
Control Test	0.080	23:36
Air Blank	0.000	23:37
Diagnostics Check	OK	23:37

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROJAS,JEAN CARLO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: LUNA

DATE/TIME OF TEST: 09/02/2017 01:42:19

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.080	01:37
Air Blank	0.000	01:37
Subject Sample # 1	0.150	01:38
Breath Volume	1.386	01:38
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.144	01:40
Breath Volume	1.261	01:40
Air Blank	0.000	01:41
Control Test	0.080	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: CHA

DATE/TIME OF TEST: 09/02/2017 05:30:24

Diagnostics Check	OK	05:24
Air Blank	0.000	05:24
Control Test	0.081	05:24
Air Blank	0.000	05:25
Subject Sample # 1	0.060	05:26
Breath Volume	1.820	05:26
Air Blank	0.000	05:26
Air Blank	0.000	05:28
Subject Sample # 2	0.057	05:28
Breath Volume	2.511	05:28
Air Blank	0.000	05:29
Control Test	0.080	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RABLE, JOSEPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: DOLAN

DATE/TIME OF TEST: 09/02/2017 21:41:52

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.081	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.000	21:37
Breath Volume	1.695	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.000	21:40
Breath Volume	3.046	21:40
Air Blank	0.000	21:40
Control Test	0.080	21:41
Air Blank	0.000	21:41
Diagnostics Check	OK	21:41

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS, BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 09/03/2017 00:31:50

Diagnostics Check	OK	00:25
Air Blank	0.000	00:25
Control Test	0.081	00:25
Air Blank	0.000	00:26
Subject Sample # 1	0.088	00:26
Breath Volume	1.691	00:26
Air Blank	0.000	00:27
Air Blank	0.000	00:29
Subject Sample # 2	0.091	00:30
Breath Volume	2.375	00:30
Air Blank	0.000	00:30
Control Test	0.080	00:31
Air Blank	0.000	00:31
Diagnostics Check	OK	00:31

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE,PETER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: RANDALL

DATE/TIME OF TEST: 09/03/2017 01:36:07

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:25	
Control Test		0.080	01:25	
Air Blank		0.000	01:25	
Subject Sample	# 1	0.121	01:29	Volume Not Met
Breath Volume		0.035	01:29	
Air Blank		0.000	01:29	
Air Blank		0.000	01:31	
Subject Sample	# 2	0.089	01:34	Volume Not Met
Breath Volume		0.015	01:34	
Air Blank		0.000	01:35	
Control Test		0.079	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MATHISEN, LEIGH
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:19

SUBJECT'S LAST NAME: PARTRIDGE

DATE/TIME OF TEST: 09/03/2017 02:47:48

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.000	02:43
Breath Volume	2.644	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.000	02:46
Breath Volume	3.218	02:46
Air Blank	0.000	02:46
Control Test	0.080	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRWIN,CHIP
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: LOPRESTO

DATE/TIME OF TEST: 09/03/2017 03:43:25

Diagnostics Check	OK	03:36
Air Blank	0.000	03:37
Control Test	0.080	03:37
Air Blank	0.000	03:37
Subject Sample # 1	0.181	03:38
Breath Volume	1.273	03:38
Air Blank	0.000	03:38
Air Blank	0.000	03:40
Subject Sample # 2	0.168	03:41
Breath Volume	1.375	03:41
Air Blank	0.000	03:42
Control Test	0.079	03:42
Air Blank	0.000	03:43
Diagnostics Check	OK	03:43

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:22

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 09/03/2017 05:53:33

Diagnostics Check	OK	05:46
Air Blank	0.000	05:47
Control Test	0.080	05:47
Air Blank	0.000	05:48
Subject Sample # 1	0.089	05:48
Breath Volume	2.000	05:48
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.088	05:52
Breath Volume	1.453	05:52
Air Blank	0.000	05:52
Control Test	0.080	05:52
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RABLE, JOSEPH
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: COLLAZO

DATE/TIME OF TEST: 09/03/2017 22:40:44

Diagnostics Check	OK	22:34
Air Blank	0.000	22:35
Control Test	0.080	22:35
Air Blank	0.000	22:35
Subject Sample # 1	0.133	22:36
Breath Volume	1.738	22:36
Air Blank	0.000	22:37
Air Blank	0.000	22:38
Subject Sample # 2	0.131	22:39
Breath Volume	1.609	22:39
Air Blank	0.000	22:39
Control Test	0.080	22:40
Air Blank	0.000	22:40
Diagnostics Check	OK	22:40

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ERSHOWSKY,ERICA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: ESPINOZA-MARTINEZ

DATE/TIME OF TEST: 09/03/2017 23:35:52

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.079	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.197	23:30
Breath Volume	1.484	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.202	23:34
Breath Volume	1.609	23:34
Air Blank	0.000	23:34
Control Test	0.079	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WASMUND, ALLEN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: LUCERO

DATE/TIME OF TEST: 09/04/2017 21:08:25

Diagnostics Check		OK	21:02
Air Blank		0.000	21:02
Control Test		0.081	21:03
Air Blank		0.000	21:03
Subject Sample	# 1	0.156	21:04
Breath Volume		2.621	21:04
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.155	21:06
Breath Volume		2.382	21:06
Air Blank		0.000	21:07
Control Test		0.081	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONGSHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:24

SUBJECT'S LAST NAME: HORTON

DATE/TIME OF TEST: 09/06/2017 11:54:53

Diagnostics Check	OK	11:48
Air Blank	0.000	11:48
Control Test	0.082	11:48
Air Blank	0.000	11:49
Subject Sample # 1	0.132	11:50
Breath Volume	4.855	11:50
Air Blank	0.000	11:50
Air Blank	0.000	11:52
Subject Sample # 2	0.123	11:53
Breath Volume	2.085	11:53
Air Blank	0.000	11:53
Control Test	0.081	11:54
Air Blank	0.000	11:54
Diagnostics Check	OK	11:54

Data Download Date/Time: 09/07/2017 16:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: BONNER

DATE/TIME OF TEST: 09/08/2017 05:19:04

Diagnostics Check	OK	05:12	
Air Blank	0.000	05:13	
Control Test	0.081	05:13	
Air Blank	0.000	05:14	
Subject Sample	# 1 0.090	05:14	
Breath Volume	1.421	05:14	
Air Blank	0.000	05:15	
Air Blank	0.000	05:17	
Subject Sample	# 2	05:17	Subject Test Refused
Breath Volume	0.000	05:17	
Air Blank	0.000	05:18	
Control Test	0.082	05:18	
Air Blank	0.000	05:18	
Diagnostics Check	OK	05:19	

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: THOMAS, RICK
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: DRESCH

DATE/TIME OF TEST: 09/08/2017 23:47:23

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.082	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.217	23:42
Breath Volume	4.210	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.197	23:45
Breath Volume	2.652	23:45
Air Blank	0.000	23:46
Control Test	0.080	23:46
Air Blank	0.000	23:47
Diagnostics Check	OK	23:47

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KNIGHT,RYAN
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 09/09/2017 02:15:44

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.082	02:10
Air Blank	0.000	02:10
Subject Sample # 1	0.144	02:11
Breath Volume	4.046	02:11
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.143	02:14
Breath Volume	3.796	02:14
Air Blank	0.000	02:14
Control Test	0.081	02:15
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROZEFORT,MICHELLE
ARRESTING OFFICER AGENCY: Eatonville PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: LALANE ANGLADA

DATE/TIME OF TEST: 09/10/2017 02:27:20

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.082	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.189	02:22
Breath Volume	3.503	02:22
Air Blank	0.000	02:23
Air Blank	0.000	02:25
Subject Sample # 2	0.173	02:25
Breath Volume	1.605	02:25
Air Blank	0.000	02:26
Control Test	0.080	02:26
Air Blank	0.000	02:27
Diagnostics Check	OK	02:27

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: GUTIERREZ RODRIGUEZ

DATE/TIME OF TEST: 09/12/2017 00:11:27

Diagnostics Check		OK	23:59	
Air Blank		0.000	23:59	
Control Test		0.081	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1	0.168	00:03	Volume Not Met
Breath Volume		0.019	00:03	
Air Blank		0.000	00:04	
Air Blank		0.000	00:05	
Subject Sample	# 2	0.160	00:06	
Breath Volume		2.324	00:06	
Air Blank		0.000	00:07	
Air Blank		0.000	00:09	
Subject Sample	# 3	0.169	00:09	
Breath Volume		2.652	00:09	
Air Blank		0.000	00:10	
Control Test		0.079	00:10	
Air Blank		0.000	00:11	
Diagnostics Check		OK	00:11	

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA, ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: MEDOR

DATE/TIME OF TEST: 09/12/2017 02:08:52

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.081	02:03
Air Blank	0.000	02:03
Subject Sample # 1	0.229	02:04
Breath Volume	1.933	02:04
Air Blank	0.000	02:04
Air Blank	0.000	02:06
Subject Sample # 2	0.229	02:07
Breath Volume	1.125	02:07
Air Blank	0.000	02:07
Control Test	0.081	02:08
Air Blank	0.000	02:08
Diagnostics Check	OK	02:08

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, S
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 09/12/2017 04:26:07

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.081	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.139	04:21
Breath Volume	1.187	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.126	04:24
Breath Volume	1.257	04:24
Air Blank	0.000	04:25
Control Test	0.080	04:25
Air Blank	0.000	04:26
Diagnostics Check	OK	04:26

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: URIBE, SERGIO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: GRADY

DATE/TIME OF TEST: 09/13/2017 01:00:51

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.081	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.169	00:56
Breath Volume	2.503	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.165	00:59
Breath Volume	1.691	00:59
Air Blank	0.000	00:59
Control Test	0.081	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,B
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 09/13/2017 23:32:10

Diagnostics Check	OK	23:25
Air Blank	0.000	23:26
Control Test	0.082	23:26
Air Blank	0.000	23:27
Subject Sample # 1	0.128	23:27
Breath Volume	1.578	23:27
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.132	23:30
Breath Volume	1.972	23:30
Air Blank	0.000	23:31
Control Test	0.080	23:31
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/14/2017 12:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: DAUGHERTY

DATE/TIME OF TEST: 09/15/2017 05:43:41

Diagnostics Check	OK	05:34
Air Blank	0.000	05:34
Control Test	0.082	05:34
Air Blank	0.000	05:35
Subject Sample # 1	0.205	05:35
Breath Volume	3.453	05:35
Air Blank	0.000	05:36
Air Blank	0.000	05:38
Subject Sample # 2	0.182	05:38
Breath Volume	2.003	05:38
Air Blank	0.000	05:39
Air Blank	0.000	05:41
Subject Sample # 3	0.199	05:42
Breath Volume	4.062	05:42
Air Blank	0.000	05:42
Control Test	0.080	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:43

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE, PETER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: PADILLA

DATE/TIME OF TEST: 09/15/2017 23:17:19

Diagnostics Check	OK	23:07
Air Blank	0.000	23:08
Control Test	0.082	23:08
Air Blank	0.000	23:08
Subject Sample # 1	0.156	23:10
Breath Volume	2.617	23:10
Air Blank	0.000	23:10
Air Blank	0.000	23:12
Subject Sample # 2	0.133	23:12
Breath Volume	1.191	23:12
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 3	0.142	23:15
Breath Volume	1.851	23:15
Air Blank	0.000	23:16
Control Test	0.080	23:16
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,MARK
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: MERCEDES SANCHEZ

DATE/TIME OF TEST: 09/16/2017 00:29:44

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.081	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.183	00:25
Breath Volume	1.781	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.185	00:28
Breath Volume	3.242	00:28
Air Blank	0.000	00:28
Control Test	0.080	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:29

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WATTS, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: DIEMEL

DATE/TIME OF TEST: 09/16/2017 03:20:31

Diagnostics Check	OK	03:14
Air Blank	0.000	03:14
Control Test	0.082	03:15
Air Blank	0.000	03:15
Subject Sample # 1	0.192	03:16
Breath Volume	1.445	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.207	03:19
Breath Volume	2.105	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:19
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: ASHE

DATE/TIME OF TEST: 09/16/2017 05:39:32

Diagnostics Check	OK	05:28	
Air Blank	0.000	05:28	
Control Test	0.081	05:28	
Air Blank	0.000	05:29	
Subject Sample	# 1	05:32	No Sample Provided
Breath Volume	0.000	05:32	
Air Blank	0.000	05:33	
Air Blank	0.000	05:35	
Subject Sample	# 2	05:38	No Sample Provided
Breath Volume	0.000	05:38	
Air Blank	0.000	05:38	
Control Test	0.080	05:38	
Air Blank	0.000	05:39	
Diagnostics Check	OK	05:39	

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WOOD, CHRISTOPHER
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MIRANDA-LOPEZ

DATE/TIME OF TEST: 09/16/2017 22:02:27

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.081	21:57
Air Blank		0.000	21:57
Subject Sample	# 1	0.230	21:58
Breath Volume		2.222	21:58
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.242	22:00
Breath Volume		2.679	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: LAVENDER

DATE/TIME OF TEST: 09/17/2017 04:50:04

Diagnostics Check	OK	04:43
Air Blank	0.000	04:43
Control Test	0.082	04:43
Air Blank	0.000	04:44
Subject Sample # 1	0.202	04:45
Breath Volume	2.542	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:48
Subject Sample # 2	0.193	04:48
Breath Volume	2.675	04:48
Air Blank	0.000	04:49
Control Test	0.077	04:49
Air Blank	0.000	04:49
Diagnostics Check	OK	04:50

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBERTSON, DANIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:14

SUBJECT'S LAST NAME: PRESLEY

DATE/TIME OF TEST: 09/17/2017 19:42:04

Diagnostics Check		OK	19:35
Air Blank		0.000	19:36
Control Test		0.081	19:36
Air Blank		0.000	19:36
Subject Sample	# 1	0.245	19:37
Breath Volume		1.468	19:37
Air Blank		0.000	19:38
Air Blank		0.000	19:40
Subject Sample	# 2	0.254	19:40
Breath Volume		1.519	19:40
Air Blank		0.000	19:41
Control Test		0.081	19:41
Air Blank		0.000	19:41
Diagnostics Check		OK	19:42

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GAUTAM,ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: DAVIS JR

DATE/TIME OF TEST: 09/17/2017 22:06:36

Diagnostics Check	OK	22:00
Air Blank	0.000	22:00
Control Test	0.083	22:01
Air Blank	0.000	22:01
Subject Sample # 1	0.183	22:02
Breath Volume	1.218	22:02
Air Blank	0.000	22:02
Air Blank	0.000	22:04
Subject Sample # 2	0.175	22:05
Breath Volume	1.148	22:05
Air Blank	0.000	22:05
Control Test	0.081	22:05
Air Blank	0.000	22:06
Diagnostics Check	OK	22:06

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SORRELLS,BRIAN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: MICOURT

DATE/TIME OF TEST: 09/17/2017 22:48:43

Diagnostics Check	OK	22:41
Air Blank	0.000	22:42
Control Test	0.081	22:42
Air Blank	0.000	22:43
Subject Sample # 1	0.243	22:43
Breath Volume	2.531	22:43
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.245	22:47
Breath Volume	1.582	22:47
Air Blank	0.000	22:47
Control Test	0.081	22:48
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:04

SUBJECT'S LAST NAME: POLEN

DATE/TIME OF TEST: 09/18/2017 00:30:26

Diagnostics Check	OK	00:24
Air Blank	0.000	00:24
Control Test	0.081	00:25
Air Blank	0.000	00:25
Subject Sample # 1	0.170	00:26
Breath Volume	1.199	00:26
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.170	00:28
Breath Volume	1.410	00:28
Air Blank	0.000	00:29
Control Test	0.080	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN, E
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 09/18/2017 04:00:35

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:49	
Control Test	0.081	03:49	
Air Blank	0.000	03:50	
Subject Sample	# 1	03:53	No Sample Provided
Breath Volume	0.000	03:53	
Air Blank	0.000	03:54	
Air Blank	0.000	03:56	
Subject Sample	# 2	03:59	No Sample Provided
Breath Volume	0.000	03:59	
Air Blank	0.000	03:59	
Control Test	0.081	03:59	
Air Blank	0.000	04:00	
Diagnostics Check	OK	04:00	

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,JUAN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 09/18/2017 05:22:47

Diagnostics Check	OK	05:14
Air Blank	0.000	05:14
Control Test	0.082	05:14
Air Blank	0.000	05:15
Subject Sample # 1	0.129	05:17
Breath Volume	1.593	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.134	05:21
Breath Volume	1.972	05:21
Air Blank	0.000	05:21
Control Test	0.078	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALYL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DELUCA,ALEX
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: ROLFE

DATE/TIME OF TEST: 09/19/2017 04:19:10

Diagnostics Check		OK	04:07	
Air Blank		0.000	04:08	
Control Test		0.082	04:08	
Air Blank		0.000	04:08	
Subject Sample	# 1	0.152	04:12	Volume Not Met
Breath Volume		0.210	04:12	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 2	0.151	04:17	Volume Not Met
Breath Volume		0.347	04:17	
Air Blank		0.000	04:18	
Control Test		0.078	04:18	
Air Blank		0.000	04:19	
Diagnostics Check		OK	04:19	

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TEW, JEREMY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: ESTIEN RIVERA

DATE/TIME OF TEST: 09/20/2017 06:38:03

Diagnostics Check		OK	06:30
Air Blank		0.000	06:30
Control Test		0.082	06:31
Air Blank		0.000	06:31
Subject Sample	# 1	0.156	06:33
Breath Volume		2.699	06:33
Air Blank		0.000	06:33
Air Blank		0.000	06:35
Subject Sample	# 2	0.159	06:36
Breath Volume		1.421	06:36
Air Blank		0.000	06:37
Control Test		0.081	06:37
Air Blank		0.000	06:37
Diagnostics Check		OK	06:38

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: PENA

DATE/TIME OF TEST: 09/21/2017 01:14:12

Diagnostics Check		OK	01:05	
Air Blank		0.000	01:05	
Control Test		0.081	01:06	
Air Blank		0.000	01:06	
Subject Sample	# 1	0.199	01:09	Volume Not Met
Breath Volume		0.000	01:09	
Air Blank		0.000	01:10	
Air Blank		0.000	01:12	
Subject Sample	# 2		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:13	
Control Test		0.076	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 09/21/2017 03:53:33

Diagnostics Check	OK	03:46
Air Blank	0.000	03:47
Control Test	0.082	03:47
Air Blank	0.000	03:48
Subject Sample # 1	0.128	03:48
Breath Volume	1.175	03:48
Air Blank	0.000	03:49
Air Blank	0.000	03:51
Subject Sample # 2	0.124	03:52
Breath Volume	1.808	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:52
Air Blank	0.000	03:53
Diagnostics Check	OK	03:53

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:59

SUBJECT'S LAST NAME: CULBERSON

DATE/TIME OF TEST: 09/21/2017 05:38:12

Diagnostics Check	OK	05:26	
Air Blank	0.000	05:27	
Control Test	0.081	05:27	
Air Blank	0.000	05:28	
Subject Sample	# 1	05:31	No Sample Provided
Breath Volume	0.000	05:31	
Air Blank	0.000	05:31	
Air Blank	0.000	05:33	
Subject Sample	# 2	05:36	No Sample Provided
Breath Volume	0.000	05:36	
Air Blank	0.000	05:37	
Control Test	0.082	05:37	
Air Blank	0.000	05:38	
Diagnostics Check	OK	05:38	

Data Download Date/Time: 09/21/2017 12:38

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNS,M
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 09/21/2017 23:39:38

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.081	23:33
Air Blank	0.000	23:34
Subject Sample # 1	0.175	23:35
Breath Volume	2.718	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.170	23:38
Breath Volume	2.398	23:38
Air Blank	0.000	23:38
Control Test	0.081	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBIJ
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: NEDABYLEK

DATE/TIME OF TEST: 09/22/2017 01:18:04

Diagnostics Check	OK	01:09	
Air Blank	0.000	01:10	
Control Test	0.082	01:10	
Air Blank	0.000	01:10	
Subject Sample	# 1	01:14	No Sample Provided
Breath Volume	0.000	01:14	
Air Blank	0.000	01:14	
Air Blank	0.000	01:16	
Subject Sample	# 2	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:17	
Control Test	0.081	01:17	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:18	

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROJAS,JEANCARLO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: OMALLEY

DATE/TIME OF TEST: 09/22/2017 14:50:18

Diagnostics Check	OK	14:47	
Air Blank	0.000	14:47	
Control Test	0.081	14:47	
Air Blank	0.000	14:48	
Subject Sample	# 1	14:48	Interferent Detect
Breath Volume	3.101	14:48	
Air Blank	0.000	14:49	
Control Test	0.082	14:49	
Air Blank	0.000	14:50	
Diagnostics Check	OK	14:50	

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROJAS,JEANCARLO
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: OMALLEY

DATE/TIME OF TEST: 09/22/2017 14:55:25

Diagnostics Check	OK	14:52	
Air Blank	0.000	14:52	
Control Test	0.081	14:53	
Air Blank	0.000	14:53	
Subject Sample	# 1	14:53	Interferent Detect
Breath Volume	2.984	14:53	
Air Blank	0.000	14:54	
Control Test	0.080	14:54	
Air Blank	0.000	14:55	
Diagnostics Check	OK	14:55	

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 09/23/2017 03:36:34

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.082	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.118	03:31
Breath Volume	1.343	03:31
Air Blank	0.000	03:32
Air Blank	0.000	03:33
Subject Sample # 2	0.109	03:35
Breath Volume	2.210	03:35
Air Blank	0.000	03:35
Control Test	0.080	03:35
Air Blank	0.000	03:36
Diagnostics Check	OK	03:36

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VERDUZCO RIOS

DATE/TIME OF TEST: 09/23/2017 04:28:49

Diagnostics Check		OK	04:22
Air Blank		0.000	04:23
Control Test		0.081	04:23
Air Blank		0.000	04:24
Subject Sample	# 1	0.215	04:24
Breath Volume		2.285	04:24
Air Blank		0.000	04:25
Air Blank		0.000	04:27
Subject Sample	# 2	0.200	04:27
Breath Volume		1.367	04:27
Air Blank		0.000	04:27
Control Test		0.080	04:28
Air Blank		0.000	04:28
Diagnostics Check		OK	04:28

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:11

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 09/24/2017 04:40:23

Diagnostics Check	OK	04:34
Air Blank	0.000	04:34
Control Test	0.081	04:34
Air Blank	0.000	04:35
Subject Sample # 1	0.137	04:35
Breath Volume	3.816	04:35
Air Blank	0.000	04:36
Air Blank	0.000	04:38
Subject Sample # 2	0.135	04:38
Breath Volume	2.511	04:38
Air Blank	0.000	04:39
Control Test	0.081	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AGOSTO,ALBA
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: AGUSTINHIPOLITO

DATE/TIME OF TEST: 09/24/2017 18:01:52

Diagnostics Check	OK	17:55
Air Blank	0.000	17:55
Control Test	0.081	17:56
Air Blank	0.000	17:56
Subject Sample # 1	0.154	17:57
Breath Volume	2.984	17:57
Air Blank	0.000	17:58
Air Blank	0.000	17:59
Subject Sample # 2	0.149	18:00
Breath Volume	2.378	18:00
Air Blank	0.000	18:00
Control Test	0.081	18:01
Air Blank	0.000	18:01
Diagnostics Check	OK	18:01

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARIANO,S
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: RAMRATTAN

DATE/TIME OF TEST: 09/25/2017 00:22:38

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.081	00:17
Air Blank		0.000	00:17
Subject Sample	# 1	0.220	00:18
Breath Volume		2.062	00:18
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.231	00:21
Breath Volume		1.160	00:21
Air Blank		0.000	00:21
Control Test		0.080	00:22
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLLAND,D
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 09/25/2017 04:11:12

Diagnostics Check	OK	04:08	
Air Blank	0.000	04:08	
Control Test	0.081	04:09	
Air Blank	0.000	04:09	
Subject Sample	# 1	04:09	Subject Test Refused
Breath Volume	0.000	04:09	
Air Blank	0.000	04:10	
Control Test	0.081	04:10	
Air Blank	0.000	04:11	
Diagnostics Check	OK	04:11	

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RANDALL,SHAWN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: FRANZESE

DATE/TIME OF TEST: 09/25/2017 23:32:51

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.083	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.000	23:28
Breath Volume	2.253	23:28
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.000	23:31
Breath Volume	2.398	23:31
Air Blank	0.000	23:31
Control Test	0.081	23:32
Air Blank	0.000	23:32
Diagnostics Check	OK	23:32

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OJEDA, MARIA, C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN, ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: OJEDA

DATE/TIME OF TEST: 09/28/2017 01:42:27

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:31	
Control Test	0.082	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:35	No Sample Provided
Breath Volume	0.000	01:35	
Air Blank	0.000	01:35	
Air Blank	0.000	01:37	
Subject Sample	# 2	01:41	No Sample Provided
Breath Volume	0.000	01:41	
Air Blank	0.000	01:41	
Control Test	0.081	01:41	
Air Blank	0.000	01:42	
Diagnostics Check	OK	01:42	

Data Download Date/Time: 09/28/2017 13:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: DAWSON

DATE/TIME OF TEST: 09/29/2017 03:48:51

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.082	03:40
Air Blank	0.000	03:40
Subject Sample # 1	0.223	03:43
Breath Volume	1.722	03:43
Air Blank	0.000	03:43
Air Blank	0.000	03:45
Subject Sample # 2	0.215	03:47
Breath Volume	1.425	03:47
Air Blank	0.000	03:47
Control Test	0.080	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CSIZMADIA,ANDY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 09/30/2017 03:44:51

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.082	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.051	03:40
Breath Volume	2.031	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.050	03:43
Breath Volume	2.289	03:43
Air Blank	0.000	03:43
Control Test	0.082	03:44
Air Blank	0.000	03:44
Diagnostics Check	OK	03:44

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER, TAYLOR
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: DEIDA CRUZ

DATE/TIME OF TEST: 09/30/2017 04:01:38

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.141	03:57
Breath Volume	1.472	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.135	04:00
Breath Volume	1.890	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MAJESKE

DATE/TIME OF TEST: 09/30/2017 04:29:52

Diagnostics Check		OK	04:18	
Air Blank		0.000	04:18	
Control Test		0.080	04:19	
Air Blank		0.000	04:19	
Subject Sample	# 1	0.132	04:22	Volume Not Met
Breath Volume		0.308	04:22	
Air Blank		0.000	04:23	
Air Blank		0.000	04:25	
Subject Sample	# 2	0.155	04:28	Volume Not Met
Breath Volume		0.117	04:28	
Air Blank		0.000	04:28	
Control Test		0.081	04:29	
Air Blank		0.000	04:29	
Diagnostics Check		OK	04:29	

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 09/22/2017 23:28:26

Diagnostics Check	OK	23:22
Air Blank	0.000	23:22
Control Test	0.078	23:23
Air Blank	0.000	23:23
Subject Sample # 1	0.338	23:23
Breath Volume	1.816	23:23
Air Blank	0.000	23:24
Air Blank	0.000	23:26
Subject Sample # 2	0.337	23:26
Breath Volume	1.910	23:26
Air Blank	0.000	23:27
Control Test	0.078	23:27
Air Blank	0.000	23:28
Diagnostics Check	OK	23:28

Data Download Date/Time: 10/04/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: MUNOZ TEJADA

DATE/TIME OF TEST: 09/24/2017 03:35:01

Diagnostics Check	OK	03:28
Air Blank	0.000	03:29
Control Test	0.077	03:29
Air Blank	0.000	03:30
Subject Sample # 1	0.190	03:30
Breath Volume	2.242	03:30
Air Blank	0.000	03:31
Air Blank	0.000	03:33
Subject Sample # 2	0.197	03:33
Breath Volume	2.140	03:33
Air Blank	0.000	03:34
Control Test	0.078	03:34
Air Blank	0.000	03:34
Diagnostics Check	OK	03:35

Data Download Date/Time: 10/04/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JACOB
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: ALLGARY

DATE/TIME OF TEST: 09/01/2017 04:02:59

Diagnostics Check	OK	03:52
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.202	03:54
Breath Volume	4.449	03:54
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.236	03:58
Breath Volume	3.839	03:58
Air Blank	0.000	03:58
Air Blank	0.000	04:00
Subject Sample # 3	0.212	04:01
Breath Volume	3.070	04:01
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:02
Diagnostics Check	OK	04:02

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MONDELLO, ANTHONY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 09/02/2017 23:50:06

Diagnostics Check		OK	23:38	
Air Blank		0.000	23:38	
Control Test		0.080	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 1	0.114	23:42	Volume Not Met
Breath Volume		0.593	23:42	
Air Blank		0.000	23:43	
Air Blank		0.000	23:45	
Subject Sample	# 2	0.125	23:48	Volume Not Met
Breath Volume		0.519	23:48	
Air Blank		0.000	23:49	
Control Test		0.079	23:49	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 09/03/2017 04:30:40

Diagnostics Check	OK	04:24	
Air Blank	0.000	04:24	
Control Test	0.079	04:24	
Air Blank	0.000	04:25	
Subject Sample # 1	0.176	04:26	
Breath Volume	1.851	04:26	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample # 2		04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:29	
Control Test	0.079	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:30	

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MANGAN,WILLIAM
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 15:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: VARIAN

DATE/TIME OF TEST: 09/03/2017 15:26:55

Diagnostics Check	OK	15:20
Air Blank	0.000	15:21
Control Test	0.080	15:21
Air Blank	0.000	15:22
Subject Sample # 1	0.284	15:22
Breath Volume	3.160	15:22
Air Blank	0.000	15:23
Air Blank	0.000	15:24
Subject Sample # 2	0.279	15:25
Breath Volume	2.859	15:25
Air Blank	0.000	15:26
Control Test	0.078	15:26
Air Blank	0.000	15:26
Diagnostics Check	OK	15:26

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,NEFTALIF
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: BINDER,JARED
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: GALLARDO

DATE/TIME OF TEST: 09/04/2017 02:18:06

Diagnostics Check	OK	02:10
Air Blank	0.000	02:10
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.197	02:12
Breath Volume	1.179	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.186	02:16
Breath Volume	1.296	02:16
Air Blank	0.000	02:17
Control Test	0.079	02:17
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: COLLEY,JOHN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 14:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: CARTER-MURCH

DATE/TIME OF TEST: 09/07/2017 15:24:22

Diagnostics Check	OK	15:16
Air Blank	0.000	15:17
Control Test	0.080	15:17
Air Blank	0.000	15:18
Subject Sample # 1	0.211	15:19
Breath Volume	1.238	15:19
Air Blank	0.000	15:20
Air Blank	0.000	15:22
Subject Sample # 2	0.206	15:22
Breath Volume	1.250	15:22
Air Blank	0.000	15:23
Control Test	0.078	15:23
Air Blank	0.000	15:24
Diagnostics Check	OK	15:24

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: PYE,CHRISTOPHER
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: SERAPHIN

DATE/TIME OF TEST: 09/08/2017 01:15:48

Diagnostics Check	OK	01:08
Air Blank	0.000	01:08
Control Test	0.079	01:09
Air Blank	0.000	01:09
Subject Sample # 1	0.285	01:11
Breath Volume	2.449	01:11
Air Blank	0.000	01:12
Air Blank	0.000	01:13
Subject Sample # 2	0.281	01:14
Breath Volume	2.558	01:14
Air Blank	0.000	01:14
Control Test	0.077	01:15
Air Blank	0.000	01:15
Diagnostics Check	OK	01:15

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,C,B
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: ANTHONY,C
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 06:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: OCANA

DATE/TIME OF TEST: 09/10/2017 07:36:12

Diagnostics Check	OK	07:29
Air Blank	0.000	07:30
Control Test	0.079	07:30
Air Blank	0.000	07:31
Subject Sample # 1	0.165	07:31
Breath Volume	1.789	07:31
Air Blank	0.000	07:32
Air Blank	0.000	07:34
Subject Sample # 2	0.170	07:34
Breath Volume	1.511	07:34
Air Blank	0.000	07:35
Control Test	0.078	07:35
Air Blank	0.000	07:36
Diagnostics Check	OK	07:36

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA, EMILIANO, R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA, EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 09/15/2017 04:02:16

Diagnostics Check	OK	03:50	
Air Blank	0.000	03:51	
Control Test	0.080	03:51	
Air Blank	0.000	03:52	
Subject Sample	# 1	03:55	No Sample Provided
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Air Blank	0.000	03:57	
Subject Sample	# 2	04:00	No Sample Provided
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.080	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: JASON,DURAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: VALERA

DATE/TIME OF TEST: 09/17/2017 06:18:33

Diagnostics Check	OK	06:12
Air Blank	0.000	06:12
Control Test	0.080	06:13
Air Blank	0.000	06:13
Subject Sample # 1	0.213	06:14
Breath Volume	2.746	06:14
Air Blank	0.000	06:14
Air Blank	0.000	06:16
Subject Sample # 2	0.213	06:16
Breath Volume	2.316	06:16
Air Blank	0.000	06:17
Control Test	0.080	06:17
Air Blank	0.000	06:18
Diagnostics Check	OK	06:18

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 09/18/2017 22:49:54

Diagnostics Check	OK	22:46	
Air Blank	0.000	22:47	
Control Test	0.081	22:47	
Air Blank	0.000	22:47	
Subject Sample	# 1	22:48	Subject Test Refused
Breath Volume	0.000	22:48	
Air Blank	0.000	22:49	
Control Test	0.080	22:49	
Air Blank	0.000	22:49	
Diagnostics Check	OK	22:49	

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:42

SUBJECT'S LAST NAME: MANNING

DATE/TIME OF TEST: 09/18/2017 23:12:03

Diagnostics Check	OK	23:05
Air Blank	0.000	23:06
Control Test	0.080	23:06
Air Blank	0.000	23:06
Subject Sample # 1	0.000	23:07
Breath Volume	1.570	23:07
Air Blank	0.000	23:08
Air Blank	0.000	23:10
Subject Sample # 2	0.000	23:10
Breath Volume	1.542	23:10
Air Blank	0.000	23:11
Control Test	0.081	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:12

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: HERERRA CASTILLO

DATE/TIME OF TEST: 09/20/2017 23:03:46

Diagnostics Check	OK	22:56
Air Blank	0.000	22:57
Control Test	0.080	22:57
Air Blank	0.000	22:58
Subject Sample # 1	0.134	22:58
Breath Volume	2.378	22:58
Air Blank	0.000	22:59
Air Blank	0.000	23:01
Subject Sample # 2	0.143	23:02
Breath Volume	1.605	23:02
Air Blank	0.000	23:02
Control Test	0.080	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BADMAN-PATON,LIN,M
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: SNOW,BRANDON
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: SLATER

DATE/TIME OF TEST: 09/21/2017 20:28:33

Diagnostics Check	OK	20:21
Air Blank	0.000	20:21
Control Test	0.080	20:22
Air Blank	0.000	20:22
Subject Sample # 1	0.169	20:23
Breath Volume	2.902	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:25
Subject Sample # 2	0.169	20:26
Breath Volume	2.757	20:26
Air Blank	0.000	20:27
Control Test	0.081	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: DUNN, COLLIN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: NEMITZ

DATE/TIME OF TEST: 09/23/2017 03:35:23

Diagnostics Check	OK	03:27
Air Blank	0.000	03:28
Control Test	0.080	03:28
Air Blank	0.000	03:29
Subject Sample # 1	0.125	03:29
Breath Volume	2.246	03:29
Air Blank	0.000	03:30
Air Blank	0.000	03:32
Subject Sample # 2	0.130	03:33
Breath Volume	2.023	03:33
Air Blank	0.000	03:34
Control Test	0.080	03:34
Air Blank	0.000	03:35
Diagnostics Check	OK	03:35

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MANNING, RYAN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/24/2017 02:20:15

Diagnostics Check	OK	02:14	
Air Blank	0.000	02:14	
Control Test	0.081	02:15	
Air Blank	0.000	02:15	
Subject Sample	# 1 0.206	02:18	Volume Not Met
Breath Volume	0.585	02:18	
Air Blank		02:19	Purge Fail
Air Blank	0.000	02:20	

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MANNING, RYAN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/24/2017 02:27:28

Diagnostics Check	OK	02:22	
Air Blank	0.000	02:22	
Control Test	0.079	02:22	
Air Blank	0.000	02:23	
Subject Sample # 1	0.261	02:26	
Breath Volume	1.593	02:26	
Air Blank		02:26	Purge Fail
Air Blank	0.000	02:27	

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONDELLO, ANTHONY, J
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: MANNING, RYAN
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 08/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: COOPER

DATE/TIME OF TEST: 09/24/2017 02:40:58

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.077	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.247	02:31
Breath Volume	1.105	02:31
Air Blank	0.000	02:32
Air Blank	0.000	02:33
Subject Sample # 2	0.277	02:34
Breath Volume	2.007	02:34
Air Blank	0.000	02:35
Air Blank	0.000	02:36
Subject Sample # 3	0.262	02:39
Breath Volume	1.199	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:40
Diagnostics Check	OK	02:40

Data Download Date/Time: 09/28/2017 14:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WAGER, WILLIAM, W
BREATH TEST OPERATOR AGENCY: St. Cloud PD
ARRESTING OFFICER: WAGER, WILLIAM
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: NICHOLSON

DATE/TIME OF TEST: 09/29/2017 04:00:48

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.080	03:54
Air Blank	0.000	03:54
Subject Sample # 1	0.231	03:55
Breath Volume	1.617	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.217	03:59
Breath Volume	1.273	03:59
Air Blank	0.000	03:59
Control Test	0.079	04:00
Air Blank	0.000	04:00
Diagnostics Check	OK	04:00

Data Download Date/Time: 10/18/2017 20:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000995

INSTRUMENT REGISTERED TO: COCOA BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GUINN,DAVID,J
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: PAYNE,TAYLOR
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 08/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: BAGGETT

DATE/TIME OF TEST: 09/01/2017 20:32:30

Diagnostics Check	OK	20:25
Air Blank	0.000	20:25
Control Test	0.079	20:25
Air Blank	0.000	20:26
Subject Sample # 1	0.299	20:26
Breath Volume	1.695	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.284	20:30
Breath Volume	1.492	20:30
Air Blank	0.000	20:31
Control Test	0.078	20:31
Air Blank	0.000	20:32
Diagnostics Check	OK	20:32

Data Download Date/Time: 02/09/2018 16:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 09/01/2017 22:47:18

Diagnostics Check	OK	22:41
Air Blank	0.000	22:41
Control Test	0.079	22:41
Air Blank	0.000	22:42
Subject Sample # 1	0.183	22:42
Breath Volume	1.562	22:42
Air Blank	0.000	22:43
Air Blank	0.000	22:45
Subject Sample # 2	0.175	22:45
Breath Volume	1.539	22:45
Air Blank	0.000	22:46
Control Test	0.079	22:46
Air Blank	0.000	22:47
Diagnostics Check	OK	22:47

Data Download Date/Time: 09/16/2017 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/14/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: GREEK

DATE/TIME OF TEST: 09/02/2017 22:26:39

Diagnostics Check	OK	22:20
Air Blank	0.000	22:21
Control Test	0.079	22:21
Air Blank	0.000	22:21
Subject Sample # 1	0.000	22:22
Breath Volume	1.550	22:22
Air Blank	0.000	22:22
Air Blank	0.000	22:24
Subject Sample # 2	0.000	22:25
Breath Volume	1.792	22:25
Air Blank	0.000	22:25
Control Test	0.080	22:26
Air Blank	0.000	22:26
Diagnostics Check	OK	22:26

Data Download Date/Time: 09/16/2017 10:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JASON,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SANCHEZ,JASON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: STOECKEL

DATE/TIME OF TEST: 09/04/2017 01:59:19

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:53
Subject Sample # 1	0.186	01:54
Breath Volume	1.164	01:54
Air Blank	0.000	01:55
Air Blank	0.000	01:56
Subject Sample # 2	0.177	01:57
Breath Volume	1.390	01:57
Air Blank	0.000	01:58
Control Test	0.079	01:58
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 09/16/2017 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SANCHEZ,JASON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 08/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: ZOLLER

DATE/TIME OF TEST: 09/04/2017 22:32:05

Diagnostics Check	OK	22:24
Air Blank	0.000	22:25
Control Test	0.079	22:25
Air Blank	0.000	22:26
Subject Sample # 1	0.155	22:27
Breath Volume	1.320	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:29
Subject Sample # 2	0.161	22:30
Breath Volume	1.574	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 09/16/2017 10:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: CABRANES,JORGE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: LOYD

DATE/TIME OF TEST: 09/22/2017 04:19:10

Diagnostics Check	OK	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Subject Sample # 1	0.000	04:14
Breath Volume	2.046	04:14
Air Blank	0.000	04:15
Air Blank	0.000	04:17
Subject Sample # 2	0.000	04:17
Breath Volume	1.941	04:17
Air Blank	0.000	04:18
Control Test	0.079	04:18
Air Blank	0.000	04:19
Diagnostics Check	OK	04:19

Data Download Date/Time: 10/23/2017 11:57