

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PECK,KEITH
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:36

SUBJECT'S LAST NAME: MAVRAKIS

DATE/TIME OF TEST: 10/21/2017 05:03:03

Diagnostics Check	OK	04:56
Air Blank	0.000	04:57
Control Test	0.080	04:57
Air Blank	0.000	04:58
Subject Sample # 1	0.122	04:58
Breath Volume	1.335	04:58
Air Blank	0.000	04:59
Air Blank	0.000	05:00
Subject Sample # 2	0.113	05:01
Breath Volume	1.242	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:02
Diagnostics Check	OK	05:03

Data Download Date/Time: 11/06/2017 13:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001132

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAMOS,JOSE,L
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: RAMOS,JOSE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:28

SUBJECT'S LAST NAME: CRADY

DATE/TIME OF TEST: 10/21/2017 15:57:14

Diagnostics Check	OK	15:49
Air Blank	0.000	15:49
Control Test	0.080	15:50
Air Blank	0.000	15:50
Subject Sample # 1	0.000	15:52
Breath Volume	3.757	15:52
Air Blank	0.000	15:53
Air Blank	0.000	15:55
Subject Sample # 2	0.000	15:55
Breath Volume	1.867	15:55
Air Blank	0.000	15:56
Control Test	0.081	15:56
Air Blank	0.000	15:57
Diagnostics Check	OK	15:57

Data Download Date/Time: 11/06/2017 13:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELOIT,BRIAN,W
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KELLY,EZEKIEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 10/01/2017 03:09:44

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.081	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.173	03:05
Breath Volume		2.867	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.170	03:08
Breath Volume		2.890	03:08
Air Blank		0.000	03:08
Control Test		0.079	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHIRSHIKOVA,ALINA,A
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BALLOUN,SAMUEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 10/01/2017 04:47:11

Diagnostics Check	OK	04:40
Air Blank	0.000	04:40
Control Test	0.080	04:41
Air Blank	0.000	04:41
Subject Sample # 1	0.170	04:42
Breath Volume	3.296	04:42
Air Blank	0.000	04:43
Air Blank	0.000	04:44
Subject Sample # 2	0.166	04:45
Breath Volume	2.257	04:45
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Diagnostics Check	OK	04:47

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 22:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 10/03/2017 23:08:28

Diagnostics Check	OK	23:01
Air Blank	0.000	23:02
Control Test	0.080	23:02
Air Blank	0.000	23:03
Subject Sample # 1	0.170	23:04
Breath Volume	3.074	23:04
Air Blank	0.000	23:04
Air Blank	0.000	23:06
Subject Sample # 2	0.164	23:06
Breath Volume	3.519	23:06
Air Blank	0.000	23:07
Control Test	0.079	23:07
Air Blank	0.000	23:08
Diagnostics Check	OK	23:08

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: DONATY

DATE/TIME OF TEST: 10/04/2017 02:02:07

Diagnostics Check		OK	01:55
Air Blank		0.000	01:55
Control Test		0.080	01:56
Air Blank		0.000	01:56
Subject Sample	# 1	0.214	01:57
Breath Volume		2.496	01:57
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.223	02:00
Breath Volume		1.917	02:00
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: QUILLEN

DATE/TIME OF TEST: 10/04/2017 04:06:37

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.080	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.153	04:02
Breath Volume	2.765	04:02
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.152	04:05
Breath Volume	2.828	04:05
Air Blank	0.000	04:05
Control Test	0.080	04:06
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICKEY,M,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: AIPLE,C
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:43

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/06/2017 15:09:27

Diagnostics Check	OK	15:03
Air Blank	0.000	15:03
Control Test	0.081	15:03
Air Blank	0.000	15:04
Subject Sample # 1	0.214	15:04
Breath Volume	2.746	15:04
Air Blank	0.000	15:05
Air Blank	0.000	15:07
Subject Sample # 2	0.199	15:07
Breath Volume	1.898	15:07
Air Blank	0.000	15:08
Control Test	0.080	15:08
Air Blank	0.000	15:09
Diagnostics Check	OK	15:09

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: YINGLING

DATE/TIME OF TEST: 10/06/2017 22:41:59

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.080	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.000	22:37
Breath Volume	2.480	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.000	22:40
Breath Volume	2.863	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:41
Diagnostics Check	OK	22:41

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: LAY,STEVEN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KINLAW

DATE/TIME OF TEST: 10/07/2017 03:13:50

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.202	03:09
Breath Volume		1.917	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:11
Subject Sample	# 2	0.208	03:12
Breath Volume		3.710	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,THOMAS,H
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: EVANS,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 10/07/2017 03:30:33

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.080	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.153	03:26
Breath Volume	2.703	03:26
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.154	03:29
Breath Volume	2.535	03:29
Air Blank	0.000	03:29
Control Test	0.079	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SMITH,EREK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: PARSONS II

DATE/TIME OF TEST: 10/07/2017 03:45:33

Diagnostics Check	OK	03:38
Air Blank	0.000	03:38
Control Test	0.079	03:39
Air Blank	0.000	03:39
Subject Sample # 1	0.137	03:40
Breath Volume	3.175	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.129	03:44
Breath Volume	1.582	03:44
Air Blank	0.000	03:44
Control Test	0.079	03:44
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 10/07/2017 04:16:42

Diagnostics Check	OK	04:09
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:10
Subject Sample # 1	0.207	04:12
Breath Volume	2.089	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.201	04:15
Breath Volume	2.152	04:15
Air Blank	0.000	04:15
Control Test	0.079	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: SKEEN

DATE/TIME OF TEST: 10/07/2017 04:38:17

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:32
Subject Sample # 1	0.074	04:33
Breath Volume	1.480	04:33
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.073	04:36
Breath Volume	1.394	04:36
Air Blank	0.000	04:37
Control Test	0.079	04:37
Air Blank	0.000	04:38
Diagnostics Check	OK	04:38

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GORDON,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 12:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: VAN KIRK

DATE/TIME OF TEST: 10/08/2017 01:56:30

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.179	01:52
Breath Volume		2.195	01:52
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.183	01:54
Breath Volume		1.710	01:54
Air Blank		0.000	01:55
Control Test		0.079	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MONIE,JARED
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: POPP

DATE/TIME OF TEST: 10/08/2017 03:16:13

Diagnostics Check	OK	03:09
Air Blank	0.000	03:09
Control Test	0.080	03:10
Air Blank	0.000	03:10
Subject Sample # 1	0.023	03:11
Breath Volume	2.023	03:11
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.022	03:14
Breath Volume	1.871	03:14
Air Blank	0.000	03:15
Control Test	0.080	03:15
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON, TYLER, B
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: THOMPSON, TYLER
ARRESTING OFFICER AGENCY: St. Augustine Beach
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: FONT

DATE/TIME OF TEST: 10/09/2017 17:48:13

Diagnostics Check	OK	17:45	
Air Blank	0.000	17:45	
Control Test	0.080	17:46	
Air Blank	0.000	17:46	
Subject Sample	# 1	17:46	Subject Test Refused
Breath Volume	0.000	17:46	
Air Blank	0.000	17:47	
Control Test	0.080	17:47	
Air Blank	0.000	17:48	
Diagnostics Check	OK	17:48	

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANNING,SEAN,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: GORDON,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 10/17/2017 01:25:13

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Subject Sample # 1	0.234	01:15	
Breath Volume	3.339	01:15	
Air Blank	0.000	01:15	
Air Blank	0.000	01:17	
Subject Sample # 2	0.208	01:20	Volume Not Met
Breath Volume	0.347	01:20	
Air Blank	0.000	01:21	
Air Blank	0.000	01:23	
Subject Sample # 3	0.220	01:23	
Breath Volume	2.878	01:23	
Air Blank	0.000	01:24	
Control Test	0.076	01:24	
Air Blank	0.000	01:25	
Diagnostics Check	OK	01:25	

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINSKY,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: LINSKY,MICHAEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 10/20/2017 01:00:57

Diagnostics Check	OK	00:54
Air Blank	0.000	00:54
Control Test	0.080	00:55
Air Blank	0.000	00:55
Subject Sample # 1	0.178	00:56
Breath Volume	3.007	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.176	00:59
Breath Volume	2.609	00:59
Air Blank	0.000	01:00
Control Test	0.080	01:00
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CREPPEL,SEAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: BACH

DATE/TIME OF TEST: 10/21/2017 04:04:19

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.011	03:59
Breath Volume	1.230	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.012	04:02
Breath Volume	1.625	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,TIMOTHY,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: FISCHER

DATE/TIME OF TEST: 10/22/2017 01:19:04

Diagnostics Check	OK	01:16	
Air Blank	0.000	01:16	
Control Test	0.080	01:16	
Air Blank	0.000	01:17	
Subject Sample	# 1	01:17	Subject Test Refused
Breath Volume	0.000	01:17	
Air Blank	0.000	01:18	
Control Test	0.079	01:18	
Air Blank	0.000	01:18	
Diagnostics Check	OK	01:19	

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SMITH,EREK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: SILVERDAHL

DATE/TIME OF TEST: 10/22/2017 02:20:27

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.080	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.154	02:14
Breath Volume	2.074	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.138	02:18
Breath Volume	1.554	02:18
Air Blank	0.000	02:19
Control Test	0.080	02:19
Air Blank	0.000	02:20
Diagnostics Check	OK	02:20

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,ALAN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: LEGGE

DATE/TIME OF TEST: 10/23/2017 00:59:32

Diagnostics Check	OK	00:53	
Air Blank	0.000	00:53	
Control Test	0.079	00:53	
Air Blank	0.000	00:54	
Subject Sample	# 1 0.189	00:55	
Breath Volume	2.625	00:55	
Air Blank	0.000	00:55	
Air Blank	0.000	00:57	
Subject Sample	# 2 0.180	00:58	
Breath Volume	1.925	00:58	
Air Blank	0.000	00:58	
Control Test	0.079	00:58	
Air Blank		00:59	RFI Detect
Air Blank	0.000	00:59	

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABEL,PAUL,J
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: COWARD,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 10/24/2017 21:03:15

Diagnostics Check	OK	21:00	
Air Blank	0.000	21:01	
Control Test	0.080	21:01	
Air Blank	0.000	21:02	
Subject Sample	# 1	21:02	RFI Detect
Breath Volume	0.000	21:02	
Air Blank		21:02	RFI Detect
Air Blank	0.000	21:03	

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABEL,PAUL,J
BREATH TEST OPERATOR AGENCY: St. Augustine Beach
ARRESTING OFFICER: COWARD,THOMAS
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: KENDRICK

DATE/TIME OF TEST: 10/24/2017 21:12:22

Diagnostics Check	OK	21:06	
Air Blank	0.000	21:07	
Control Test	0.079	21:07	
Air Blank	0.000	21:07	
Subject Sample	# 1	21:10	Subject Test Refused
Breath Volume	0.000	21:10	
Air Blank	0.000	21:11	
Control Test	0.079	21:11	
Air Blank		21:11	RFI Detect
Air Blank	0.000	21:12	

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001133

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINSKY,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: LINSKY,MICHAEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 09/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/24/2017 21:52:59

Diagnostics Check	OK	21:46
Air Blank	0.000	21:47
Control Test	0.076	21:47
Air Blank	0.000	21:47
Subject Sample # 1	0.204	21:48
Breath Volume	3.375	21:48
Air Blank	0.000	21:49
Air Blank	0.000	21:50
Subject Sample # 2	0.213	21:51
Breath Volume	2.761	21:51
Air Blank	0.000	21:52
Control Test	0.077	21:52
Air Blank	0.000	21:52
Diagnostics Check	OK	21:52

Data Download Date/Time: 10/25/2017 11:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,DAVID,M
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: MITCHELL,DAVID
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 18:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: KUPER

DATE/TIME OF TEST: 10/25/2017 20:25:58

Diagnostics Check	OK	20:23	
Air Blank	0.000	20:23	
Control Test	0.079	20:23	
Air Blank	0.000	20:24	
Subject Sample	# 1	20:24	Subject Test Refused
Breath Volume	0.000	20:24	
Air Blank	0.000	20:25	
Control Test	0.079	20:25	
Air Blank	0.000	20:25	
Diagnostics Check	OK	20:25	

Data Download Date/Time: 11/28/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:02

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/26/2017 04:31:04

Diagnostics Check	OK	04:22	
Air Blank	0.000	04:23	
Control Test	0.078	04:23	
Air Blank	0.000	04:23	
Subject Sample	# 1	04:27	No Sample Provided
Breath Volume	0.000	04:27	
Air Blank	0.000	04:27	
Air Blank	0.000	04:29	
Subject Sample	# 2	04:29	Subject Test Refused
Breath Volume	0.000	04:29	
Air Blank	0.000	04:30	
Control Test	0.078	04:30	
Air Blank	0.000	04:30	
Diagnostics Check	OK	04:31	

Data Download Date/Time: 11/28/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WALLACE,NATHAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 17:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:10

SUBJECT'S LAST NAME: MC AFEE

DATE/TIME OF TEST: 10/26/2017 18:39:57

Diagnostics Check	OK	18:30	
Air Blank	0.000	18:31	
Control Test	0.078	18:31	
Air Blank	0.000	18:31	
Subject Sample	# 1	18:35	No Sample Provided
Breath Volume	0.000	18:35	
Air Blank	0.000	18:35	
Air Blank	0.000	18:37	
Subject Sample	# 2	18:38	Subject Test Refused
Breath Volume	0.000	18:38	
Air Blank	0.000	18:39	
Control Test	0.078	18:39	
Air Blank	0.000	18:39	
Diagnostics Check	OK	18:39	

Data Download Date/Time: 11/28/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YARBOROUGH,ERIC,L
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: SMITH,EREK
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 10/27/2017 03:03:49

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.078	02:59	
Air Blank	0.000	02:59	
Subject Sample	# 1	03:02	Subject Test Refused
Breath Volume	0.000	03:02	
Air Blank	0.000	03:02	
Control Test	0.080	03:03	
Air Blank	0.000	03:03	
Diagnostics Check	OK	03:03	

Data Download Date/Time: 11/28/2017 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LINSKY,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BALLOUN,SAMUEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: KELLY

DATE/TIME OF TEST: 10/29/2017 00:30:30

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.078	00:24
Air Blank	0.000	00:24
Subject Sample # 1	0.137	00:25
Breath Volume	1.761	00:25
Air Blank	0.000	00:26
Air Blank	0.000	00:28
Subject Sample # 2	0.131	00:29
Breath Volume	1.691	00:29
Air Blank	0.000	00:29
Control Test	0.078	00:29
Air Blank	0.000	00:30
Diagnostics Check	OK	00:30

Data Download Date/Time: 11/28/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: OESTERLE

DATE/TIME OF TEST: 10/31/2017 03:37:43

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.084	03:32
Breath Volume	2.523	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.085	03:36
Breath Volume	2.554	03:36
Air Blank	0.000	03:36
Control Test	0.080	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 11/28/2017 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001134

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: GORDON,JOSHUA
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 366849
DRY GAS EXPIRY DATE: 03/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: BRESSETTE

DATE/TIME OF TEST: 10/31/2017 21:30:38

Diagnostics Check	OK	21:24
Air Blank	0.000	21:24
Control Test	0.078	21:25
Air Blank	0.000	21:25
Subject Sample # 1	0.060	21:26
Breath Volume	2.039	21:26
Air Blank	0.000	21:26
Air Blank	0.000	21:28
Subject Sample # 2	0.059	21:29
Breath Volume	1.453	21:29
Air Blank	0.000	21:29
Control Test	0.079	21:30
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 11/28/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KEATING,KAYLA
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 17:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: BARNES-SPENCER

DATE/TIME OF TEST: 10/02/2017 18:34:26

Diagnostics Check	OK	18:28
Air Blank	0.000	18:29
Control Test	0.080	18:29
Air Blank	0.000	18:29
Subject Sample # 1	0.000	18:30
Breath Volume	1.558	18:30
Air Blank	0.000	18:30
Air Blank	0.000	18:32
Subject Sample # 2	0.000	18:33
Breath Volume	1.406	18:33
Air Blank	0.000	18:33
Control Test	0.080	18:33
Air Blank	0.000	18:34
Diagnostics Check	OK	18:34

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CAZI
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MENDOZA ROQUE

DATE/TIME OF TEST: 10/08/2017 02:38:31

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.081	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.221	02:34
Breath Volume		2.230	02:34
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.223	02:37
Breath Volume		2.171	02:37
Air Blank		0.000	02:37
Control Test		0.080	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 10/08/2017 21:11:29

Diagnostics Check	OK	21:05
Air Blank	0.000	21:05
Control Test	0.080	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.179	21:06
Breath Volume	1.988	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.185	21:10
Breath Volume	3.113	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:10
Air Blank	0.000	21:11
Diagnostics Check	OK	21:11

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: MCCUTCHEON

DATE/TIME OF TEST: 10/10/2017 02:39:06

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:27	
Control Test		0.080	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1	0.155	02:31	Volume Not Met
Breath Volume		0.507	02:31	
Air Blank		0.000	02:32	
Air Blank		0.000	02:34	
Subject Sample	# 2	0.166	02:34	
Breath Volume		1.328	02:34	
Air Blank		0.000	02:35	
Air Blank		0.000	02:37	
Subject Sample	# 3	0.165	02:37	
Breath Volume		1.125	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SUN,TONY
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 08:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 10/11/2017 10:52:23

Diagnostics Check	OK	10:42
Air Blank	0.000	10:42
Control Test	0.081	10:43
Air Blank	0.000	10:43
Subject Sample # 1	0.194	10:46
Breath Volume	1.246	10:46
Air Blank	0.000	10:46
Air Blank	0.000	10:48
Subject Sample # 2	0.186	10:50
Breath Volume	1.152	10:50
Air Blank	0.000	10:51
Control Test	0.081	10:51
Air Blank	0.000	10:52
Diagnostics Check	OK	10:52

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CABRERA,JOSEPH
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 18:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ITURRIA

DATE/TIME OF TEST: 10/14/2017 19:48:12

Diagnostics Check	OK	19:42
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:43
Subject Sample # 1	0.149	19:43
Breath Volume	2.312	19:43
Air Blank	0.000	19:44
Air Blank	0.000	19:46
Subject Sample # 2	0.139	19:46
Breath Volume	1.281	19:46
Air Blank	0.000	19:47
Control Test	0.082	19:47
Air Blank	0.000	19:48
Diagnostics Check	OK	19:48

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: JIMENEZ,JOSE
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 16:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: STAFFORD

DATE/TIME OF TEST: 10/18/2017 18:39:32

Diagnostics Check	OK	18:31
Air Blank	0.000	18:32
Control Test	0.081	18:32
Air Blank	0.000	18:33
Subject Sample # 1	0.027	18:35
Breath Volume	1.511	18:35
Air Blank	0.000	18:35
Air Blank	0.000	18:37
Subject Sample # 2	0.026	18:38
Breath Volume	1.652	18:38
Air Blank	0.000	18:38
Control Test	0.082	18:38
Air Blank	0.000	18:39
Diagnostics Check	OK	18:39

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: SHINGU

DATE/TIME OF TEST: 10/20/2017 22:30:46

Diagnostics Check	OK	22:24
Air Blank	0.000	22:24
Control Test	0.081	22:25
Air Blank	0.000	22:25
Subject Sample # 1	0.213	22:26
Breath Volume	3.636	22:26
Air Blank	0.000	22:26
Air Blank	0.000	22:28
Subject Sample # 2	0.214	22:29
Breath Volume	3.339	22:29
Air Blank	0.000	22:29
Control Test	0.081	22:30
Air Blank	0.000	22:30
Diagnostics Check	OK	22:30

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FERNANDEZ,NICHOLAS
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: CAPERS

DATE/TIME OF TEST: 10/21/2017 04:12:57

Diagnostics Check	OK	04:10	
Air Blank	0.000	04:10	
Control Test	0.081	04:10	
Air Blank	0.000	04:11	
Subject Sample	# 1	04:11	Subject Test Refused
Breath Volume	0.000	04:11	
Air Blank	0.000	04:12	
Control Test	0.081	04:12	
Air Blank	0.000	04:12	
Diagnostics Check	OK	04:12	

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FITHIAN,J
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: SLOCUM

DATE/TIME OF TEST: 10/22/2017 01:37:53

Diagnostics Check	OK	01:31
Air Blank	0.000	01:32
Control Test	0.081	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.253	01:33
Breath Volume	3.414	01:33
Air Blank	0.000	01:33
Air Blank	0.000	01:35
Subject Sample # 2	0.250	01:36
Breath Volume	2.906	01:36
Air Blank	0.000	01:36
Control Test	0.081	01:37
Air Blank	0.000	01:37
Diagnostics Check	OK	01:37

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: QUIST

DATE/TIME OF TEST: 10/22/2017 04:13:25

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:06
Air Blank	0.000	04:06
Subject Sample # 1	0.175	04:08
Breath Volume	1.492	04:08
Air Blank	0.000	04:08
Air Blank	0.000	04:10
Subject Sample # 2	0.185	04:11
Breath Volume	1.570	04:11
Air Blank	0.000	04:12
Control Test	0.080	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SEMPERE,DIANA
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: RIOS NERI

DATE/TIME OF TEST: 10/23/2017 21:08:08

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.081	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.252	21:03
Breath Volume		2.953	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.263	21:06
Breath Volume		3.003	21:06
Air Blank		0.000	21:07
Control Test		0.080	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MCCOOL

DATE/TIME OF TEST: 10/24/2017 02:22:02

Diagnostics Check	OK	02:16
Air Blank	0.000	02:16
Control Test	0.081	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.000	02:17
Breath Volume	2.292	02:17
Air Blank	0.000	02:18
Air Blank	0.000	02:20
Subject Sample # 2	0.000	02:20
Breath Volume	1.882	02:20
Air Blank	0.000	02:21
Control Test	0.081	02:21
Air Blank	0.000	02:21
Diagnostics Check	OK	02:22

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEDESMA,OLIVER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GULLETT JR

DATE/TIME OF TEST: 10/24/2017 03:22:49

Diagnostics Check	OK	03:13
Air Blank	0.000	03:13
Control Test	0.082	03:13
Air Blank	0.000	03:14
Subject Sample # 1	0.249	03:17
Breath Volume	2.171	03:17
Air Blank	0.000	03:17
Air Blank	0.000	03:19
Subject Sample # 2	0.248	03:21
Breath Volume	1.617	03:21
Air Blank	0.000	03:21
Control Test	0.081	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: BASTIEN

DATE/TIME OF TEST: 10/29/2017 03:38:58

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:32
Subject Sample # 1	0.165	03:33
Breath Volume	1.582	03:33
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.179	03:37
Breath Volume	1.914	03:37
Air Blank	0.000	03:38
Control Test	0.081	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SABAN,SHAHEED
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: RAHMING

DATE/TIME OF TEST: 10/30/2017 04:11:27

Diagnostics Check	OK	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.102	04:06
Breath Volume	1.523	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.108	04:10
Breath Volume	1.523	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:10
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001139

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OLEA,ELIEZER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 562838
DRY GAS EXPIRY DATE: 09/06/2019
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: LEUNG

DATE/TIME OF TEST: 10/31/2017 20:59:16

Diagnostics Check	OK	20:52
Air Blank	0.000	20:53
Control Test	0.082	20:53
Air Blank	0.000	20:53
Subject Sample # 1	0.000	20:54
Breath Volume	3.125	20:54
Air Blank	0.000	20:55
Air Blank	0.000	20:57
Subject Sample # 2	0.000	20:57
Breath Volume	1.808	20:57
Air Blank	0.000	20:58
Control Test	0.081	20:58
Air Blank	0.000	20:59
Diagnostics Check	OK	20:59

Data Download Date/Time: 11/06/2017 12:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/01/2017 04:01:37

Diagnostics Check	OK	03:55
Air Blank	0.000	03:55
Control Test	0.082	03:55
Air Blank	0.000	03:56
Subject Sample # 1	0.162	03:57
Breath Volume	1.105	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.171	04:00
Breath Volume	1.519	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 10/22/2017 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER, ADAM, H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER, ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: LARSON

DATE/TIME OF TEST: 10/07/2017 02:28:29

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.082	02:17	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:21	No Sample Provided
Breath Volume	0.000	02:21	
Air Blank	0.000	02:21	
Air Blank	0.000	02:23	
Subject Sample	# 2	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:27	
Control Test	0.082	02:27	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 10/22/2017 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,LATORIAN,T
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: TURNER,LATORIAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: FEAGLE

DATE/TIME OF TEST: 10/13/2017 21:40:29

Diagnostics Check	OK	21:34
Air Blank	0.000	21:34
Control Test	0.082	21:35
Air Blank	0.000	21:35
Subject Sample # 1	0.192	21:36
Breath Volume	2.156	21:36
Air Blank	0.000	21:36
Air Blank	0.000	21:38
Subject Sample # 2	0.188	21:38
Breath Volume	1.781	21:38
Air Blank	0.000	21:39
Control Test	0.080	21:39
Air Blank	0.000	21:40
Diagnostics Check	OK	21:40

Data Download Date/Time: 10/22/2017 10:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001151

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 09/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BUCKMAN

DATE/TIME OF TEST: 10/14/2017 02:56:56

Diagnostics Check	OK	02:50
Air Blank	0.000	02:50
Control Test	0.080	02:50
Air Blank	0.000	02:51
Subject Sample # 1	0.177	02:52
Breath Volume	1.121	02:52
Air Blank	0.000	02:52
Air Blank	0.000	02:54
Subject Sample # 2	0.184	02:55
Breath Volume	1.601	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Diagnostics Check	OK	02:56

Data Download Date/Time: 10/22/2017 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001154

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANSARD,JOSHUA,D
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HANSARD,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 10/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 05:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: OSANI

DATE/TIME OF TEST: 10/28/2017 06:06:36

Diagnostics Check		OK	06:00
Air Blank		0.000	06:00
Control Test		0.081	06:01
Air Blank		0.000	06:01
Subject Sample	# 1	0.230	06:02
Breath Volume		2.375	06:02
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.244	06:05
Breath Volume		2.484	06:05
Air Blank		0.000	06:05
Control Test		0.082	06:05
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 11/06/2017 13:35

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: JORDAN, DANIEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: CIRAFISI

DATE/TIME OF TEST: 10/08/2017 08:31:29

Diagnostics Check	OK	08:28	
Air Blank	0.000	08:28	
Control Test	0.080	08:29	
Air Blank	0.000	08:29	
Subject Sample	# 1	08:30	Subject Test Refused
Breath Volume	0.000	08:30	
Air Blank	0.000	08:30	
Control Test	0.080	08:30	
Air Blank	0.000	08:31	
Diagnostics Check	OK	08:31	

Data Download Date/Time: 10/25/2017 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,ROBERT,B
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: CARBONE,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 10:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:24

SUBJECT'S LAST NAME: POZZUOLI

DATE/TIME OF TEST: 10/13/2017 11:52:55

Diagnostics Check		OK	11:46
Air Blank		0.000	11:47
Control Test		0.080	11:47
Air Blank		0.000	11:47
Subject Sample	# 1	0.053	11:48
Breath Volume		2.937	11:48
Air Blank		0.000	11:49
Air Blank		0.000	11:51
Subject Sample	# 2	0.044	11:51
Breath Volume		1.250	11:51
Air Blank		0.000	11:52
Control Test		0.078	11:52
Air Blank		0.000	11:52
Diagnostics Check		OK	11:52

Data Download Date/Time: 10/25/2017 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEARSON,ROBERT,B
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: CARBONE,ROBERT
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 08:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:27

SUBJECT'S LAST NAME: RYAN

DATE/TIME OF TEST: 10/18/2017 09:56:41

Diagnostics Check		OK	09:49
Air Blank		0.000	09:50
Control Test		0.080	09:50
Air Blank		0.000	09:50
Subject Sample	# 1	0.295	09:51
Breath Volume		2.984	09:51
Air Blank		0.000	09:51
Air Blank		0.000	09:53
Subject Sample	# 2	0.310	09:55
Breath Volume		2.792	09:55
Air Blank		0.000	09:55
Control Test		0.081	09:56
Air Blank		0.000	09:56
Diagnostics Check		OK	09:56

Data Download Date/Time: 10/25/2017 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER, MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 10/20/2017 02:02:55

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:01	Subject Test Refused
Breath Volume	0.000	02:01	
Air Blank	0.000	02:02	
Control Test	0.082	02:02	
Air Blank	0.000	02:02	
Diagnostics Check	OK	02:02	

Data Download Date/Time: 10/25/2017 10:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBIN,RICHARD,D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: WARMINGTON,W
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: DUBOSE

DATE/TIME OF TEST: 10/21/2017 00:57:08

Diagnostics Check	OK	00:54	
Air Blank	0.000	00:54	
Control Test	0.081	00:55	
Air Blank	0.000	00:55	
Subject Sample	# 1	00:55	Subject Test Refused
Breath Volume	0.000	00:55	
Air Blank	0.000	00:56	
Control Test	0.081	00:56	
Air Blank	0.000	00:57	
Diagnostics Check	OK	00:57	

Data Download Date/Time: 10/25/2017 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001156

INSTRUMENT REGISTERED TO: ORMOND BEACH P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARNER, MICHAEL, D
BREATH TEST OPERATOR AGENCY: Ormond Beach PD
ARRESTING OFFICER: GARNER, MICHAEL
ARRESTING OFFICER AGENCY: Ormond Beach PD
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: TRACY

DATE/TIME OF TEST: 10/21/2017 04:34:19

Diagnostics Check	OK	04:25
Air Blank	0.000	04:26
Control Test	0.081	04:26
Air Blank	0.000	04:26
Subject Sample # 1	0.231	04:29
Breath Volume	1.726	04:29
Air Blank	0.000	04:29
Air Blank	0.000	04:31
Subject Sample # 2	0.213	04:32
Breath Volume	2.027	04:32
Air Blank	0.000	04:33
Control Test	0.078	04:33
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 10/25/2017 10:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

INSTRUMENT REGISTERED TO: PONCE INLET PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINES,BRANDON,W
BREATH TEST OPERATOR AGENCY: Ponce Inlet PD
ARRESTING OFFICER: OGDEN,ALTON
ARRESTING OFFICER AGENCY: Daytona Beach Shores
LAST AGENCY INSPECTION: 09/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 10/13/2017 04:51:10

Diagnostics Check	OK	04:43
Air Blank	0.000	04:44
Control Test	0.081	04:44
Air Blank	0.000	04:44
Subject Sample # 1	0.185	04:45
Breath Volume	1.183	04:45
Air Blank	0.000	04:46
Air Blank	0.000	04:47
Subject Sample # 2	0.184	04:49
Breath Volume	2.199	04:49
Air Blank	0.000	04:50
Control Test	0.080	04:50
Air Blank	0.000	04:51
Diagnostics Check	OK	04:51

Data Download Date/Time: 10/23/2017 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001158

INSTRUMENT REGISTERED TO: PONCE INLET PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BENEZETTE,BENJAMIN,J
BREATH TEST OPERATOR AGENCY: Port Orange PD
ARRESTING OFFICER: BENEZETTE,BENJAMIN
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 10/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1787697
DRY GAS EXPIRY DATE: 01/08/2018
VIOLATION CODE: DUI
ARREST TIME: 17:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: DONOHUE

DATE/TIME OF TEST: 10/31/2017 19:24:46

Diagnostics Check	OK	19:18
Air Blank	0.000	19:19
Control Test	0.081	19:19
Air Blank	0.000	19:19
Subject Sample # 1	0.193	19:20
Breath Volume	1.210	19:20
Air Blank	0.000	19:20
Air Blank	0.000	19:22
Subject Sample # 2	0.208	19:23
Breath Volume	1.277	19:23
Air Blank	0.000	19:23
Control Test	0.081	19:24
Air Blank	0.000	19:24
Diagnostics Check	OK	19:24

Data Download Date/Time: 11/16/2017 07:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO,JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/24/2019
VIOLATION CODE: BUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:39

SUBJECT'S LAST NAME: LEON PAZMINO

DATE/TIME OF TEST: 10/01/2017 22:12:05

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.080	22:09	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:10	Subject Test Refused
Breath Volume	0.000	22:10	
Air Blank	0.000	22:11	
Control Test	0.079	22:11	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:12	

Data Download Date/Time: 10/25/2017 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WASHINGTON, RONALD, M
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: KLEIS, KEVIN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/24/2019
VIOLATION CODE: BUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:43

SUBJECT'S LAST NAME: ANDRADE

DATE/TIME OF TEST: 10/07/2017 20:36:48

Diagnostics Check	OK	20:29
Air Blank	0.000	20:29
Control Test	0.079	20:30
Air Blank	0.000	20:30
Subject Sample # 1	0.070	20:31
Breath Volume	3.152	20:31
Air Blank	0.000	20:31
Air Blank	0.000	20:33
Subject Sample # 2	0.071	20:35
Breath Volume	2.542	20:35
Air Blank	0.000	20:35
Control Test	0.078	20:36
Air Blank	0.000	20:36
Diagnostics Check	OK	20:36

Data Download Date/Time: 10/25/2017 12:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ADRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/24/2019
VIOLATION CODE: BUI
ARREST TIME: 21:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: GAZON

DATE/TIME OF TEST: 10/07/2017 21:51:32

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.078	21:45
Air Blank	0.000	21:46
Subject Sample # 1	0.067	21:47
Breath Volume	2.121	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.064	21:50
Breath Volume	2.906	21:50
Air Blank	0.000	21:50
Control Test	0.078	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 10/25/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO,JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/24/2019
VIOLATION CODE: BUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ARENAS

DATE/TIME OF TEST: 10/07/2017 22:17:46

Diagnostics Check	OK	22:14	
Air Blank	0.000	22:14	
Control Test	0.078	22:14	
Air Blank	0.000	22:15	
Subject Sample	# 1	22:16	Subject Test Refused
Breath Volume	0.000	22:16	
Air Blank	0.000	22:16	
Control Test	0.078	22:17	
Air Blank	0.000	22:17	
Diagnostics Check	OK	22:17	

Data Download Date/Time: 10/25/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: PEREZ,ADRIAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/24/2019
VIOLATION CODE: BUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:06

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 10/08/2017 18:33:55

Diagnostics Check	OK	18:29	
Air Blank	0.000	18:30	
Control Test	0.079	18:30	
Air Blank	0.000	18:30	
Subject Sample	# 1	18:32	Subject Test Refused
Breath Volume	0.000	18:32	
Air Blank	0.000	18:32	
Control Test	0.078	18:33	
Air Blank	0.000	18:33	
Diagnostics Check	OK	18:33	

Data Download Date/Time: 10/25/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001179

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ADRIAN,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO,JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/24/2019
VIOLATION CODE: BUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 10/08/2017 22:11:47

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:09	
Control Test	0.078	22:09	
Air Blank	0.000	22:10	
Subject Sample	# 1	22:10	Subject Test Refused
Breath Volume	0.000	22:10	
Air Blank	0.000	22:10	
Control Test	0.078	22:11	
Air Blank	0.000	22:11	
Diagnostics Check	OK	22:11	

Data Download Date/Time: 10/25/2017 12:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: BUNDRICK,MATTHEW
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DEFALCO

DATE/TIME OF TEST: 10/06/2017 02:55:23

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:43	
Control Test		0.079	02:44	
Air Blank		0.000	02:44	
Subject Sample	# 1	0.000	02:47	Volume Not Met
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Air Blank		0.000	02:50	
Subject Sample	# 2	0.219	02:50	
Breath Volume		2.351	02:50	
Air Blank		0.000	02:51	
Air Blank		0.000	02:53	
Subject Sample	# 3	0.216	02:53	
Breath Volume		2.601	02:53	
Air Blank		0.000	02:54	
Control Test		0.077	02:54	
Air Blank		0.000	02:55	
Diagnostics Check		OK	02:55	

Data Download Date/Time: 10/23/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001190

INSTRUMENT REGISTERED TO: BOYNTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASTRO,DENNIS,
BREATH TEST OPERATOR AGENCY: Boynton Beach PD
ARRESTING OFFICER: BUNDRICK,MATTHEW
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:44

SUBJECT'S LAST NAME: MCGRUFF

DATE/TIME OF TEST: 10/08/2017 01:10:14

Diagnostics Check	OK	01:07	
Air Blank	0.000	01:07	
Control Test	0.080	01:08	
Air Blank	0.000	01:08	
Subject Sample	# 1	01:08	Subject Test Refused
Breath Volume	0.000	01:08	
Air Blank	0.000	01:09	
Control Test	0.079	01:09	
Air Blank	0.000	01:10	
Diagnostics Check	OK	01:10	

Data Download Date/Time: 10/23/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHERI,OFC,M
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: CHERI,M
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/01/2017 22:58:12

Diagnostics Check	OK	22:53	
Air Blank	0.000	22:53	
Control Test	0.080	22:53	
Air Blank	0.000	22:54	
Subject Sample	# 1	22:56	Subject Test Refused
Breath Volume	0.000	22:56	
Air Blank	0.000	22:57	
Control Test	0.080	22:57	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 10/26/2017 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: THOMAS,STEPHANIE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 10/02/2017 03:41:17

Diagnostics Check	OK	03:32
Air Blank	0.000	03:33
Control Test	0.080	03:33
Air Blank	0.000	03:34
Subject Sample # 1	0.261	03:36
Breath Volume	2.386	03:36
Air Blank	0.000	03:37
Air Blank	0.000	03:39
Subject Sample # 2	0.248	03:39
Breath Volume	2.101	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Diagnostics Check	OK	03:41

Data Download Date/Time: 10/26/2017 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001192

INSTRUMENT REGISTERED TO: PANAMA CITY BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,STEPHANIE,N
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: THOMAS,STEPHANIE
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/02/2017 03:58:45

Diagnostics Check		OK	03:51
Air Blank		0.000	03:52
Control Test		0.080	03:52
Air Blank		0.000	03:53
Subject Sample	# 1	0.252	03:53
Breath Volume		2.992	03:53
Air Blank		0.000	03:54
Air Blank		0.000	03:56
Subject Sample	# 2	0.253	03:57
Breath Volume		3.078	03:57
Air Blank		0.000	03:57
Control Test		0.082	03:58
Air Blank		0.000	03:58
Diagnostics Check		OK	03:58

Data Download Date/Time: 10/26/2017 11:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: RUBINO

DATE/TIME OF TEST: 10/01/2017 04:21:07

Diagnostics Check	OK	04:14
Air Blank	0.000	04:15
Control Test	0.080	04:15
Air Blank	0.000	04:16
Subject Sample # 1	0.056	04:16
Breath Volume	1.539	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.058	04:19
Breath Volume	3.351	04:19
Air Blank	0.000	04:20
Control Test	0.080	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 10/01/2017 15:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANFIELD,BRANDON,N
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: SWENSON,NATHANIEL
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:02

SUBJECT'S LAST NAME: BELLAMY

DATE/TIME OF TEST: 10/02/2017 19:29:05

Diagnostics Check	OK	19:23
Air Blank	0.000	19:23
Control Test	0.081	19:23
Air Blank	0.000	19:24
Subject Sample # 1	0.000	19:24
Breath Volume	1.500	19:24
Air Blank	0.000	19:25
Air Blank	0.000	19:27
Subject Sample # 2	0.000	19:27
Breath Volume	3.066	19:27
Air Blank	0.000	19:28
Control Test	0.080	19:28
Air Blank	0.000	19:29
Diagnostics Check	OK	19:29

Data Download Date/Time: 11/30/2017 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORGAN,AQUILA,A
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: TAYLOR,CHARLES
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: GLADDING

DATE/TIME OF TEST: 10/12/2017 20:05:13

Diagnostics Check	OK	19:58
Air Blank	0.000	19:59
Control Test	0.080	19:59
Air Blank	0.000	19:59
Subject Sample # 1	0.104	20:00
Breath Volume	2.562	20:00
Air Blank	0.000	20:01
Air Blank	0.000	20:03
Subject Sample # 2	0.089	20:03
Breath Volume	1.152	20:03
Air Blank	0.000	20:04
Control Test	0.079	20:04
Air Blank	0.000	20:05
Diagnostics Check	OK	20:05

Data Download Date/Time: 11/30/2017 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUMFIELD, TYLER, G
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: BRUMFIELD, TYLER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 10/14/2017 08:07:00

Diagnostics Check		OK	08:00
Air Blank		0.000	08:01
Control Test		0.080	08:01
Air Blank		0.000	08:01
Subject Sample	# 1	0.189	08:02
Breath Volume		1.566	08:02
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.184	08:05
Breath Volume		1.574	08:05
Air Blank		0.000	08:06
Control Test		0.080	08:06
Air Blank		0.000	08:06
Diagnostics Check		OK	08:07

Data Download Date/Time: 11/30/2017 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JEFFREY,J
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: WEBER,JEFFREY
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: BRYK

DATE/TIME OF TEST: 10/15/2017 01:41:34

Diagnostics Check	OK	01:38	
Air Blank	0.000	01:39	
Control Test	0.079	01:39	
Air Blank	0.000	01:40	
Subject Sample	# 1	01:40	Subject Test Refused
Breath Volume	0.000	01:40	
Air Blank	0.000	01:40	
Control Test	0.080	01:40	
Air Blank	0.000	01:41	
Diagnostics Check	OK	01:41	

Data Download Date/Time: 11/30/2017 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLACE,NICHOLAS,T
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: FREDETTE,CAMERON
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: IZQUIERDO

DATE/TIME OF TEST: 10/21/2017 03:22:52

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1	0.075	03:12	Volume Not Met
Breath Volume		0.578	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 2	0.091	03:16	
Breath Volume		1.132	03:16	
Air Blank		0.000	03:17	
Air Blank		0.000	03:19	
Subject Sample	# 3	0.088	03:21	
Breath Volume		1.753	03:21	
Air Blank		0.000	03:21	
Control Test		0.080	03:22	
Air Blank		0.000	03:22	
Diagnostics Check		OK	03:22	

Data Download Date/Time: 11/30/2017 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001230

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: Tyndall AFB PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Tyndall AFB PD
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: DIXON

DATE/TIME OF TEST: 10/28/2017 01:50:37

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.080	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.196	01:46
Breath Volume	4.101	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.194	01:49
Breath Volume	4.410	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 11/30/2017 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001231

INSTRUMENT REGISTERED TO: AUBURNDALE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Auburndale PD
ARRESTING OFFICER: LEAL, AUSTIN
ARRESTING OFFICER AGENCY: Auburndale PD
LAST AGENCY INSPECTION: 09/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: MATAI

DATE/TIME OF TEST: 10/05/2017 02:48:01

Diagnostics Check	OK	02:40
Air Blank	0.000	02:40
Control Test	0.077	02:40
Air Blank	0.000	02:41
Subject Sample # 1	0.170	02:43
Breath Volume	2.183	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:45
Subject Sample # 2	0.165	02:46
Breath Volume	1.453	02:46
Air Blank	0.000	02:47
Control Test	0.076	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:48

Data Download Date/Time: 10/30/2017 10:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,D
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: WHITMARSH

DATE/TIME OF TEST: 10/01/2017 15:47:42

Diagnostics Check	OK	15:36	
Air Blank	0.000	15:36	
Control Test	0.080	15:37	
Air Blank	0.000	15:37	
Subject Sample	# 1	15:40	No Sample Provided
Breath Volume	0.000	15:40	
Air Blank	0.000	15:41	
Air Blank	0.000	15:43	
Subject Sample	# 2	15:46	No Sample Provided
Breath Volume	0.000	15:46	
Air Blank	0.000	15:46	
Control Test	0.080	15:47	
Air Blank	0.000	15:47	
Diagnostics Check	OK	15:47	

Data Download Date/Time: 11/30/2017 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,D
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: WHITMASH

DATE/TIME OF TEST: 10/01/2017 16:24:44

Diagnostics Check	OK	16:11	
Air Blank	0.000	16:12	
Control Test	0.080	16:12	
Air Blank	0.000	16:12	
Subject Sample	# 1	16:16	No Sample Provided
Breath Volume	0.000	16:16	
Air Blank	0.000	16:16	
Air Blank	0.000	16:18	
Subject Sample	# 2	0.181	16:19
Breath Volume	1.296	16:19	
Air Blank	0.000	16:20	
Air Blank	0.000	16:22	
Subject Sample	# 3	0.181	16:23
Breath Volume	1.777	16:23	
Air Blank	0.000	16:23	
Control Test	0.079	16:24	
Air Blank	0.000	16:24	
Diagnostics Check	OK	16:24	

Data Download Date/Time: 11/30/2017 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DILLARD,D
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: WHITMARSH

DATE/TIME OF TEST: 10/01/2017 16:38:28

Diagnostics Check	OK	16:29
Air Blank	0.000	16:30
Control Test	0.080	16:30
Air Blank	0.000	16:31
Subject Sample # 1	0.170	16:32
Breath Volume	2.808	16:32
Air Blank	0.000	16:32
Air Blank	0.000	16:34
Subject Sample # 2	0.166	16:36
Breath Volume	1.226	16:36
Air Blank	0.000	16:37
Control Test	0.080	16:37
Air Blank	0.000	16:38
Diagnostics Check	OK	16:38

Data Download Date/Time: 11/30/2017 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANNESE,MORRIS,J
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: ANNESE,MORRIS
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: SEBASTIAN REANDA

DATE/TIME OF TEST: 10/06/2017 21:34:16

Diagnostics Check	OK	21:29	
Air Blank	0.000	21:30	
Control Test	0.080	21:30	
Air Blank	0.000	21:30	
Subject Sample	# 1	21:32	Subject Test Refused
Breath Volume	0.000	21:32	
Air Blank	0.000	21:33	
Control Test	0.080	21:33	
Air Blank	0.000	21:34	
Diagnostics Check	OK	21:34	

Data Download Date/Time: 11/30/2017 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: SHEALY

DATE/TIME OF TEST: 10/12/2017 00:18:15

Diagnostics Check	OK	00:14	
Air Blank	0.000	00:14	
Control Test	0.081	00:15	
Air Blank	0.000	00:15	
Subject Sample	# 1	00:16	Subject Test Refused
Breath Volume	0.000	00:16	
Air Blank	0.000	00:17	
Control Test	0.081	00:17	
Air Blank	0.000	00:18	
Diagnostics Check	OK	00:18	

Data Download Date/Time: 11/30/2017 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: DONDE,J
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PERSAUD

DATE/TIME OF TEST: 10/25/2017 05:02:59

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:51	
Control Test	0.080	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1 0.182	04:55	Volume Not Met
Breath Volume	0.101	04:55	
Air Blank	0.000	04:56	
Air Blank	0.000	04:58	
Subject Sample	# 2 0.183	05:01	Volume Not Met
Breath Volume	0.078	05:01	
Air Blank	0.000	05:02	
Control Test	0.078	05:02	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 11/30/2017 08:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001235

INSTRUMENT REGISTERED TO: WEST PALM BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DONDE,J,
BREATH TEST OPERATOR AGENCY: West Palm Beach PD
ARRESTING OFFICER: 0133,DONDE
ARRESTING OFFICER AGENCY: West Palm Beach PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 152169
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: HEATH

DATE/TIME OF TEST: 10/26/2017 03:40:55

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:35	
Control Test	0.081	03:36	
Air Blank	0.000	03:36	
Subject Sample	# 1	03:39	Subject Test Refused
Breath Volume	0.000	03:39	
Air Blank	0.000	03:40	
Control Test	0.081	03:40	
Air Blank	0.000	03:40	
Diagnostics Check	OK	03:40	

Data Download Date/Time: 11/30/2017 08:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,ANTHONY,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BINZ,ANTHONY
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 18:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:58

SUBJECT'S LAST NAME: KELEHAN

DATE/TIME OF TEST: 10/14/2017 19:25:16

Diagnostics Check	OK	19:19
Air Blank	0.000	19:19
Control Test	0.079	19:19
Air Blank	0.000	19:20
Subject Sample # 1	0.276	19:20
Breath Volume	1.535	19:20
Air Blank	0.000	19:21
Air Blank	0.000	19:23
Subject Sample # 2	0.269	19:23
Breath Volume	1.398	19:23
Air Blank	0.000	19:24
Control Test	0.078	19:24
Air Blank	0.000	19:25
Diagnostics Check	OK	19:25

Data Download Date/Time: 10/15/2017 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001243

INSTRUMENT REGISTERED TO: EDGEWATER P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINZ,ANTHONY,E
BREATH TEST OPERATOR AGENCY: Edgewater PD
ARRESTING OFFICER: BINZ,ANTHONY
ARRESTING OFFICER AGENCY: Edgewater PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 765462
DRY GAS EXPIRY DATE: 04/13/2020
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: JUAREZ

DATE/TIME OF TEST: 10/15/2017 02:10:00

Diagnostics Check	OK	02:05	
Air Blank	0.000	02:05	
Control Test	0.080	02:06	
Air Blank	0.000	02:06	
Subject Sample	# 1	02:08	Subject Test Refused
Breath Volume	0.312	02:08	
Air Blank	0.000	02:09	
Control Test	0.080	02:09	
Air Blank	0.000	02:09	
Diagnostics Check	OK	02:10	

Data Download Date/Time: 10/15/2017 23:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001244

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,KELVIN,
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ROUNTREE,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:53

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/15/2017 22:29:11

Diagnostics Check	OK	22:21
Air Blank	0.000	22:21
Control Test	0.078	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.213	22:24
Breath Volume	1.531	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.213	22:27
Breath Volume	1.429	22:27
Air Blank	0.000	22:28
Control Test	0.079	22:28
Air Blank	0.000	22:29
Diagnostics Check	OK	22:29

Data Download Date/Time: 10/30/2017 17:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: PATSCH,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: WAHBA

DATE/TIME OF TEST: 10/05/2017 03:28:22

Diagnostics Check	OK	03:22
Air Blank	0.000	03:22
Control Test	0.078	03:22
Air Blank	0.000	03:23
Subject Sample # 1	0.198	03:23
Breath Volume	2.714	03:23
Air Blank	0.000	03:24
Air Blank	0.000	03:26
Subject Sample # 2	0.193	03:26
Breath Volume	1.992	03:26
Air Blank	0.000	03:27
Control Test	0.078	03:27
Air Blank	0.000	03:28
Diagnostics Check	OK	03:28

Data Download Date/Time: 10/18/2017 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: HIGHAM,KELSIE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 10/14/2017 02:39:45

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.079	02:32
Air Blank	0.000	02:32
Subject Sample # 1	0.168	02:35
Breath Volume	3.367	02:35
Air Blank	0.000	02:35
Air Blank	0.000	02:37
Subject Sample # 2	0.173	02:38
Breath Volume	2.195	02:38
Air Blank	0.000	02:38
Control Test	0.079	02:39
Air Blank	0.000	02:39
Diagnostics Check	OK	02:39

Data Download Date/Time: 10/18/2017 19:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELODY,FRANCIS,D
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: HIGHAM,KELSIE
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/15/2017 01:04:12

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.079	00:56	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:59	No Sample Provided
Breath Volume	0.000	00:59	
Air Blank	0.000	01:00	
Air Blank	0.000	01:02	
Subject Sample	# 2	01:02	Subject Test Refused
Breath Volume	0.000	01:02	
Air Blank	0.000	01:03	
Control Test	0.080	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 10/18/2017 19:20

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001246

INSTRUMENT REGISTERED TO: SOUTH DAYTONA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANK,GABRIELLE,A
BREATH TEST OPERATOR AGENCY: South Daytona PD
ARRESTING OFFICER: CAMARA,CRYSTA
ARRESTING OFFICER AGENCY: Port Orange PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: RUE

DATE/TIME OF TEST: 10/18/2017 21:16:26

Diagnostics Check	OK	21:10
Air Blank	0.000	21:10
Control Test	0.080	21:11
Air Blank	0.000	21:11
Subject Sample # 1	0.320	21:11
Breath Volume	1.964	21:11
Air Blank	0.000	21:12
Air Blank	0.000	21:14
Subject Sample # 2	0.323	21:14
Breath Volume	2.121	21:14
Air Blank	0.000	21:15
Control Test	0.080	21:15
Air Blank	0.000	21:16
Diagnostics Check	OK	21:16

Data Download Date/Time: 10/26/2017 10:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: BASSETT

DATE/TIME OF TEST: 10/01/2017 23:51:39

Diagnostics Check	OK	23:45
Air Blank	0.000	23:45
Control Test	0.077	23:45
Air Blank	0.000	23:46
Subject Sample # 1	0.214	23:47
Breath Volume	1.851	23:47
Air Blank	0.000	23:47
Air Blank	0.000	23:49
Subject Sample # 2	0.218	23:50
Breath Volume	1.769	23:50
Air Blank	0.000	23:50
Control Test	0.077	23:51
Air Blank	0.000	23:51
Diagnostics Check	OK	23:51

Data Download Date/Time: 10/02/2017 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 09/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 10/02/2017 02:37:38

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.076	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.168	02:33
Breath Volume	1.980	02:33
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.165	02:36
Breath Volume	1.593	02:36
Air Blank	0.000	02:36
Control Test	0.076	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/02/2017 14:29

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: KOTT,DOUGLAS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: CULVERWELL

DATE/TIME OF TEST: 10/04/2017 22:04:29

Diagnostics Check		OK	21:57
Air Blank		0.000	21:58
Control Test		0.077	21:58
Air Blank		0.000	21:59
Subject Sample	# 1	0.208	21:59
Breath Volume		1.992	21:59
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.210	22:02
Breath Volume		1.753	22:02
Air Blank		0.000	22:03
Control Test		0.077	22:03
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: GALLE

DATE/TIME OF TEST: 10/06/2017 04:13:49

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.074	04:09
Breath Volume	2.410	04:09
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.072	04:12
Breath Volume	2.359	04:12
Air Blank	0.000	04:12
Control Test	0.077	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PERDUE,AARON
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:20

SUBJECT'S LAST NAME: HALLORAN

DATE/TIME OF TEST: 10/06/2017 10:56:56

Diagnostics Check	OK	10:50
Air Blank	0.000	10:50
Control Test	0.078	10:50
Air Blank	0.000	10:51
Subject Sample # 1	0.220	10:52
Breath Volume	1.703	10:52
Air Blank	0.000	10:53
Air Blank	0.000	10:54
Subject Sample # 2	0.209	10:55
Breath Volume	1.386	10:55
Air Blank	0.000	10:56
Control Test	0.077	10:56
Air Blank	0.000	10:56
Diagnostics Check	OK	10:56

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: DEVLIN

DATE/TIME OF TEST: 10/07/2017 03:13:03

Diagnostics Check	OK	03:06
Air Blank	0.000	03:07
Control Test	0.077	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.159	03:08
Breath Volume	2.335	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.159	03:11
Breath Volume	1.726	03:11
Air Blank	0.000	03:12
Control Test	0.077	03:12
Air Blank	0.000	03:12
Diagnostics Check	OK	03:13

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SACKMANN,ANDREW
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:37

SUBJECT'S LAST NAME: MCMICKELL

DATE/TIME OF TEST: 10/08/2017 21:04:56

Diagnostics Check	OK	20:57
Air Blank	0.000	20:58
Control Test	0.077	20:58
Air Blank	0.000	20:59
Subject Sample # 1	0.049	20:59
Breath Volume	2.007	20:59
Air Blank	0.000	21:00
Air Blank	0.000	21:02
Subject Sample # 2	0.049	21:03
Breath Volume	1.257	21:03
Air Blank	0.000	21:04
Control Test	0.078	21:04
Air Blank	0.000	21:04
Diagnostics Check	OK	21:04

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LUNA,KRISTINA
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: STUBBS

DATE/TIME OF TEST: 10/10/2017 06:12:06

Diagnostics Check	OK	06:05
Air Blank	0.000	06:05
Control Test	0.078	06:06
Air Blank	0.000	06:06
Subject Sample # 1	0.241	06:07
Breath Volume	1.597	06:07
Air Blank	0.000	06:08
Air Blank	0.000	06:09
Subject Sample # 2	0.238	06:10
Breath Volume	1.281	06:10
Air Blank	0.000	06:11
Control Test	0.077	06:11
Air Blank	0.000	06:12
Diagnostics Check	OK	06:12

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: PACK,JEFFERY
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: ALEMAN

DATE/TIME OF TEST: 10/10/2017 17:03:03

Diagnostics Check	OK	16:56
Air Blank	0.000	16:56
Control Test	0.078	16:57
Air Blank	0.000	16:57
Subject Sample # 1	0.251	16:58
Breath Volume	1.886	16:58
Air Blank	0.000	16:58
Air Blank	0.000	17:00
Subject Sample # 2	0.238	17:01
Breath Volume	1.167	17:01
Air Blank	0.000	17:02
Control Test	0.078	17:02
Air Blank	0.000	17:02
Diagnostics Check	OK	17:03

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: STOVALL

DATE/TIME OF TEST: 10/10/2017 22:49:37

Diagnostics Check	OK	22:42
Air Blank	0.000	22:42
Control Test	0.078	22:43
Air Blank	0.000	22:43
Subject Sample # 1	0.157	22:44
Breath Volume	2.226	22:44
Air Blank	0.000	22:45
Air Blank	0.000	22:47
Subject Sample # 2	0.160	22:48
Breath Volume	2.128	22:48
Air Blank	0.000	22:48
Control Test	0.078	22:49
Air Blank	0.000	22:49
Diagnostics Check	OK	22:49

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: ARROYO

DATE/TIME OF TEST: 10/12/2017 04:22:34

Diagnostics Check	OK	04:16
Air Blank	0.000	04:16
Control Test	0.079	04:17
Air Blank	0.000	04:17
Subject Sample # 1	0.000	04:18
Breath Volume	2.464	04:18
Air Blank	0.000	04:18
Air Blank	0.000	04:20
Subject Sample # 2	0.000	04:21
Breath Volume	2.023	04:21
Air Blank	0.000	04:21
Control Test	0.078	04:21
Air Blank	0.000	04:22
Diagnostics Check	OK	04:22

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: OLEA,ELIEZER
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: DRISCOLL

DATE/TIME OF TEST: 10/12/2017 23:53:21

Diagnostics Check	OK	23:47
Air Blank	0.000	23:47
Control Test	0.079	23:47
Air Blank	0.000	23:48
Subject Sample # 1	0.128	23:48
Breath Volume	2.800	23:48
Air Blank	0.000	23:49
Air Blank	0.000	23:51
Subject Sample # 2	0.130	23:51
Breath Volume	3.324	23:51
Air Blank	0.000	23:52
Control Test	0.078	23:52
Air Blank	0.000	23:53
Diagnostics Check	OK	23:53

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FACCINI,DAVID
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 10/13/2017 03:56:44

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.078	03:51
Air Blank		0.000	03:51
Subject Sample	# 1	0.208	03:52
Breath Volume		2.347	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.204	03:55
Breath Volume		2.257	03:55
Air Blank		0.000	03:55
Control Test		0.078	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: NAVARRO,CASIMIRO
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RAMCHARAN

DATE/TIME OF TEST: 10/14/2017 03:58:47

Diagnostics Check	OK	03:51
Air Blank	0.000	03:51
Control Test	0.078	03:52
Air Blank	0.000	03:52
Subject Sample # 1	0.263	03:53
Breath Volume	1.261	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.250	03:57
Breath Volume	1.269	03:57
Air Blank	0.000	03:57
Control Test	0.077	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: POLANCO BARRERA

DATE/TIME OF TEST: 10/16/2017 02:42:31

Diagnostics Check	OK	02:33	
Air Blank	0.000	02:34	
Control Test	0.079	02:34	
Air Blank	0.000	02:34	
Subject Sample	# 1	02:37	No Sample Provided
Breath Volume	0.000	02:37	
Air Blank	0.000	02:38	
Air Blank	0.000	02:40	
Subject Sample	# 2	02:41	Subject Test Refused
Breath Volume	0.000	02:41	
Air Blank	0.000	02:41	
Control Test	0.079	02:41	
Air Blank	0.000	02:42	
Diagnostics Check	OK	02:42	

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: CARTER,JAMES
ARRESTING OFFICER AGENCY: Fort Lauderdale PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RUZ

DATE/TIME OF TEST: 10/18/2017 02:12:35

Diagnostics Check	OK	02:06
Air Blank	0.000	02:06
Control Test	0.078	02:07
Air Blank	0.000	02:07
Subject Sample # 1	0.000	02:08
Breath Volume	1.496	02:08
Air Blank	0.000	02:08
Air Blank	0.000	02:10
Subject Sample # 2	0.000	02:11
Breath Volume	2.652	02:11
Air Blank	0.000	02:11
Control Test	0.078	02:11
Air Blank	0.000	02:12
Diagnostics Check	OK	02:12

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GONZALEZ,HECTOR
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: CHOW

DATE/TIME OF TEST: 10/21/2017 04:06:41

Diagnostics Check		OK	03:55	
Air Blank		0.000	03:55	
Control Test		0.078	03:56	
Air Blank		0.000	03:56	
Subject Sample	# 1	0.198	03:59	Volume Not Met
Breath Volume		0.101	03:59	
Air Blank		0.000	04:00	
Air Blank		0.000	04:02	
Subject Sample	# 2		04:05	No Sample Provided
Breath Volume		0.000	04:05	
Air Blank		0.000	04:05	
Control Test		0.076	04:06	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROBERTS,CURTIS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: NAVARRETE FLEMING

DATE/TIME OF TEST: 10/22/2017 06:05:34

Diagnostics Check	OK	06:02	
Air Blank	0.000	06:02	
Control Test	0.078	06:02	
Air Blank	0.000	06:03	
Subject Sample	# 1	06:04	Subject Test Refused
Breath Volume	0.000	06:04	
Air Blank	0.000	06:04	
Control Test	0.078	06:04	
Air Blank	0.000	06:05	
Diagnostics Check	OK	06:05	

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLMAN,SHEILA,L
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROBERTS,CURTIS
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/22/2017 07:30:21

Diagnostics Check	OK	07:23
Air Blank	0.000	07:24
Control Test	0.078	07:24
Air Blank	0.000	07:25
Subject Sample # 1	0.143	07:25
Breath Volume	1.472	07:25
Air Blank	0.000	07:26
Air Blank	0.000	07:28
Subject Sample # 2	0.143	07:28
Breath Volume	1.714	07:28
Air Blank	0.000	07:29
Control Test	0.078	07:29
Air Blank	0.000	07:30
Diagnostics Check	OK	07:30

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: FLINK,ANDREW
ARRESTING OFFICER AGENCY: Miramar PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: AUGUSTIN FELIZ

DATE/TIME OF TEST: 10/22/2017 23:05:11

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:59
Subject Sample # 1	0.158	23:00
Breath Volume	2.343	23:00
Air Blank	0.000	23:01
Air Blank	0.000	23:03
Subject Sample # 2	0.150	23:03
Breath Volume	1.636	23:03
Air Blank	0.000	23:04
Control Test	0.077	23:04
Air Blank	0.000	23:05
Diagnostics Check	OK	23:05

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEDINA,MARMILLY,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: JEANCHARLES

DATE/TIME OF TEST: 10/24/2017 00:32:49

Diagnostics Check	OK	00:24	
Air Blank	0.000	00:24	
Control Test	0.077	00:25	
Air Blank	0.000	00:25	
Subject Sample	# 1	00:28	No Sample Provided
Breath Volume	0.000	00:28	
Air Blank	0.000	00:29	
Air Blank	0.000	00:31	
Subject Sample	# 2	00:31	Subject Test Refused
Breath Volume	0.000	00:31	
Air Blank	0.000	00:31	
Control Test	0.078	00:32	
Air Blank	0.000	00:32	
Diagnostics Check	OK	00:32	

Data Download Date/Time: 11/06/2017 12:52

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: BUDISH,HARRISON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: RETAMOZO

DATE/TIME OF TEST: 10/25/2017 05:22:39

Diagnostics Check	OK	05:15
Air Blank	0.000	05:16
Control Test	0.078	05:16
Air Blank	0.000	05:16
Subject Sample # 1	0.111	05:17
Breath Volume	1.156	05:17
Air Blank	0.000	05:18
Air Blank	0.000	05:20
Subject Sample # 2	0.120	05:21
Breath Volume	1.675	05:21
Air Blank	0.000	05:21
Control Test	0.078	05:22
Air Blank	0.000	05:22
Diagnostics Check	OK	05:22

Data Download Date/Time: 11/06/2017 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: ROJAS,RANDY
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: MARTINSON

DATE/TIME OF TEST: 10/26/2017 06:08:02

Diagnostics Check	OK	06:00
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:01
Subject Sample # 1	0.080	06:03
Breath Volume	1.808	06:03
Air Blank	0.000	06:03
Air Blank	0.000	06:05
Subject Sample # 2	0.085	06:06
Breath Volume	1.937	06:06
Air Blank	0.000	06:07
Control Test	0.079	06:07
Air Blank	0.000	06:07
Diagnostics Check	OK	06:08

Data Download Date/Time: 11/06/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRAY,SANDRA,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: MCCLELLAND,ANDREW
ARRESTING OFFICER AGENCY: Hallandale Beach PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: LOCKHART

DATE/TIME OF TEST: 10/28/2017 02:55:13

Diagnostics Check		OK	02:48
Air Blank		0.000	02:48
Control Test		0.079	02:49
Air Blank		0.000	02:49
Subject Sample	# 1	0.200	02:50
Breath Volume		2.328	02:50
Air Blank		0.000	02:51
Air Blank		0.000	02:53
Subject Sample	# 2	0.205	02:53
Breath Volume		2.367	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Diagnostics Check		OK	02:55

Data Download Date/Time: 11/06/2017 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SEVERICHE

DATE/TIME OF TEST: 10/28/2017 04:30:28

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.078	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1	0.193	04:19	No .020 agreement
Breath Volume		1.832	04:19	
Air Blank		0.000	04:19	
Air Blank		0.000	04:21	
Subject Sample	# 2	0.137	04:24	Volume Not Met
Breath Volume		0.437	04:24	
Air Blank		0.000	04:25	
Air Blank		0.000	04:27	
Subject Sample	# 3	0.165	04:28	No .020 agreement
Breath Volume		1.472	04:28	
Air Blank		0.000	04:29	
Control Test		0.076	04:29	
Air Blank		0.000	04:30	
Diagnostics Check		OK	04:30	

Data Download Date/Time: 11/06/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: HARRIS,MILTON
ARRESTING OFFICER AGENCY: Pembroke Pines PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: SEVERICHE

DATE/TIME OF TEST: 10/28/2017 04:40:09

Diagnostics Check	OK	04:32
Air Blank	0.000	04:32
Control Test	0.076	04:33
Air Blank	0.000	04:33
Subject Sample # 1	0.169	04:34
Breath Volume	1.562	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.160	04:38
Breath Volume	1.324	04:38
Air Blank	0.000	04:39
Control Test	0.078	04:39
Air Blank	0.000	04:40
Diagnostics Check	OK	04:40

Data Download Date/Time: 11/06/2017 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRAZIER,ANAYA,S
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: GUTSHALL,CHARLES
ARRESTING OFFICER AGENCY: Sunrise PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: SHAW

DATE/TIME OF TEST: 10/30/2017 03:33:46

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.079	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.165	03:29
Breath Volume	2.207	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.161	03:32
Breath Volume	2.566	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 11/06/2017 12:53

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001249

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARNANDAN,INDROWTIE,
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: LEE,TIMMOTHY
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A3
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BLANDI

DATE/TIME OF TEST: 10/31/2017 22:02:05

Diagnostics Check	OK	21:55
Air Blank	0.000	21:56
Control Test	0.078	21:56
Air Blank	0.000	21:57
Subject Sample # 1	0.167	21:57
Breath Volume	2.316	21:57
Air Blank	0.000	21:58
Air Blank	0.000	21:59
Subject Sample # 2	0.165	22:00
Breath Volume	1.824	22:00
Air Blank	0.000	22:01
Control Test	0.078	22:01
Air Blank	0.000	22:01
Diagnostics Check	OK	22:02

Data Download Date/Time: 11/06/2017 12:53

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001254

INSTRUMENT REGISTERED TO: MONROE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEMAN,NICHOLIS,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: WHITEMAN,NICHOLIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 09/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: KARI

DATE/TIME OF TEST: 10/02/2017 22:05:07

Diagnostics Check	OK	22:02	
Air Blank	0.000	22:02	
Control Test	0.080	22:03	
Air Blank	0.000	22:03	
Subject Sample	# 1	22:03	Subject Test Refused
Breath Volume	0.000	22:03	
Air Blank	0.000	22:04	
Control Test	0.079	22:04	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:05	

Data Download Date/Time: 10/04/2017 16:36

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: CLaar

DATE/TIME OF TEST: 10/12/2017 00:05:36

Diagnostics Check	OK	23:58
Air Blank	0.000	23:59
Control Test	0.079	23:59
Air Blank	0.000	00:00
Subject Sample # 1	0.229	00:00
Breath Volume	1.425	00:00
Air Blank	0.000	00:01
Air Blank	0.000	00:03
Subject Sample # 2	0.241	00:04
Breath Volume	2.148	00:04
Air Blank	0.000	00:04
Control Test	0.078	00:04
Air Blank	0.000	00:05
Diagnostics Check	OK	00:05

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: MORAVEK

DATE/TIME OF TEST: 10/12/2017 01:50:38

Diagnostics Check	OK	01:44
Air Blank	0.000	01:44
Control Test	0.079	01:45
Air Blank	0.000	01:45
Subject Sample # 1	0.000	01:46
Breath Volume	2.761	01:46
Air Blank	0.000	01:46
Air Blank	0.000	01:48
Subject Sample # 2	0.000	01:49
Breath Volume	2.429	01:49
Air Blank	0.000	01:49
Control Test	0.078	01:50
Air Blank	0.000	01:50
Diagnostics Check	OK	01:50

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: AMATUCCIO

DATE/TIME OF TEST: 10/12/2017 04:02:02

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:56	
Control Test	0.078	03:56	
Air Blank	0.000	03:57	
Subject Sample	# 1 0.155	03:57	
Breath Volume	1.140	03:57	
Air Blank	0.000	03:58	
Air Blank	0.000	04:00	
Subject Sample	# 2 0.149	04:00	
Breath Volume	1.777	04:00	
Air Blank	0.000	04:01	
Control Test	0.078	04:01	
Air Blank		04:01	RFI Detect
Air Blank		04:02	RFI Detect

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES, NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: AMATUCCIO

DATE/TIME OF TEST: 10/12/2017 04:13:14

Diagnostics Check	OK	04:07
Air Blank	0.000	04:07
Control Test	0.078	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.146	04:08
Breath Volume	1.808	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.140	04:11
Breath Volume	1.359	04:11
Air Blank	0.000	04:12
Control Test	0.078	04:12
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: BOX

DATE/TIME OF TEST: 10/13/2017 03:45:40

Diagnostics Check	OK	03:37
Air Blank	0.000	03:38
Control Test	0.079	03:38
Air Blank	0.000	03:39
Subject Sample # 1	0.170	03:40
Breath Volume	1.589	03:40
Air Blank	0.000	03:41
Air Blank	0.000	03:42
Subject Sample # 2	0.167	03:44
Breath Volume	1.933	03:44
Air Blank	0.000	03:44
Control Test	0.078	03:45
Air Blank	0.000	03:45
Diagnostics Check	OK	03:45

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:14

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 10/13/2017 07:45:23

Diagnostics Check	OK	07:38
Air Blank	0.000	07:38
Control Test	0.079	07:39
Air Blank	0.000	07:39
Subject Sample # 1	0.113	07:40
Breath Volume	1.472	07:40
Air Blank	0.000	07:40
Air Blank	0.000	07:42
Subject Sample # 2	0.120	07:43
Breath Volume	1.367	07:43
Air Blank	0.000	07:44
Control Test	0.079	07:44
Air Blank	0.000	07:45
Diagnostics Check	OK	07:45

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRIAN,HARRIS
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: RODRIGUEZ ARGUETA

DATE/TIME OF TEST: 10/14/2017 03:07:36

Diagnostics Check	OK	02:57	
Air Blank	0.000	02:57	
Control Test	0.078	02:57	
Air Blank	0.000	02:58	
Subject Sample	# 1	03:01	No Sample Provided
Breath Volume	0.000	03:01	
Air Blank	0.000	03:01	
Air Blank	0.000	03:03	
Subject Sample	# 2	03:06	Subject Test Refused
Breath Volume	0.000	03:06	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO, FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 10/14/2017 05:38:48

Diagnostics Check	OK	05:32
Air Blank	0.000	05:32
Control Test	0.079	05:33
Air Blank	0.000	05:33
Subject Sample # 1	0.089	05:34
Breath Volume	2.031	05:34
Air Blank	0.000	05:34
Air Blank	0.000	05:36
Subject Sample # 2	0.092	05:37
Breath Volume	1.800	05:37
Air Blank	0.000	05:37
Control Test	0.080	05:38
Air Blank	0.000	05:38
Diagnostics Check	OK	05:38

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: SASKO

DATE/TIME OF TEST: 10/15/2017 03:02:46

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.078	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.167	02:58
Breath Volume	2.656	02:58
Air Blank	0.000	02:58
Air Blank	0.000	03:00
Subject Sample # 2	0.163	03:01
Breath Volume	1.328	03:01
Air Blank	0.000	03:01
Control Test	0.079	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 10/19/2017 09:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: OCAMPO,OSCAR
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:57

SUBJECT'S LAST NAME: NERON

DATE/TIME OF TEST: 10/16/2017 16:26:16

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.079	16:20
Air Blank	0.000	16:21
Subject Sample # 1	0.000	16:21
Breath Volume	1.589	16:21
Air Blank	0.000	16:22
Air Blank	0.000	16:24
Subject Sample # 2	0.000	16:24
Breath Volume	2.859	16:24
Air Blank	0.000	16:25
Control Test	0.079	16:25
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 10/19/2017 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ORTIZ, ROBERTO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/18/2017 03:13:26

Diagnostics Check	OK	03:07
Air Blank	0.000	03:07
Control Test	0.080	03:07
Air Blank	0.000	03:08
Subject Sample # 1	0.126	03:08
Breath Volume	1.445	03:08
Air Blank	0.000	03:09
Air Blank	0.000	03:11
Subject Sample # 2	0.126	03:11
Breath Volume	1.890	03:11
Air Blank	0.000	03:12
Control Test	0.079	03:12
Air Blank	0.000	03:13
Diagnostics Check	OK	03:13

Data Download Date/Time: 10/19/2017 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BURKHART,CHRISTOPHER
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: GUNESCH

DATE/TIME OF TEST: 10/19/2017 00:41:35

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.079	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.146	00:36
Breath Volume	2.343	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.145	00:40
Breath Volume	2.250	00:40
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 10/19/2017 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBIJ
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,ETHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 10/19/2017 03:33:36

Diagnostics Check	OK	03:24
Air Blank	0.000	03:25
Control Test	0.079	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.194	03:26
Breath Volume	3.066	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.181	03:32
Breath Volume	2.097	03:32
Air Blank	0.000	03:32
Control Test	0.078	03:32
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 10/19/2017 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALMEYDA,ERIC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 10/19/2017 04:13:51

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.079	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.128	04:09
Breath Volume	1.894	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.130	04:12
Breath Volume	1.398	04:12
Air Blank	0.000	04:12
Control Test	0.079	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/19/2017 09:28

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KANE, ALEXANDER
ARRESTING OFFICER AGENCY: Edgewood PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 10/20/2017 00:25:54

Diagnostics Check	OK	00:19
Air Blank	0.000	00:19
Control Test	0.079	00:20
Air Blank	0.000	00:20
Subject Sample # 1	0.127	00:21
Breath Volume	1.695	00:21
Air Blank	0.000	00:22
Air Blank	0.000	00:23
Subject Sample # 2	0.128	00:24
Breath Volume	2.414	00:24
Air Blank	0.000	00:24
Control Test	0.079	00:25
Air Blank	0.000	00:25
Diagnostics Check	OK	00:25

Data Download Date/Time: 10/26/2017 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GOMEZ SANTIAGO

DATE/TIME OF TEST: 10/20/2017 01:31:16

Diagnostics Check	OK	01:24
Air Blank	0.000	01:25
Control Test	0.079	01:25
Air Blank	0.000	01:25
Subject Sample # 1	0.150	01:26
Breath Volume	2.359	01:26
Air Blank	0.000	01:27
Air Blank	0.000	01:29
Subject Sample # 2	0.146	01:29
Breath Volume	1.878	01:29
Air Blank	0.000	01:30
Control Test	0.078	01:30
Air Blank	0.000	01:31
Diagnostics Check	OK	01:31

Data Download Date/Time: 10/26/2017 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: BLOCKER

DATE/TIME OF TEST: 10/21/2017 03:32:05

Diagnostics Check	OK	03:25
Air Blank	0.000	03:25
Control Test	0.080	03:26
Air Blank	0.000	03:26
Subject Sample # 1	0.164	03:27
Breath Volume	2.214	03:27
Air Blank	0.000	03:27
Air Blank	0.000	03:29
Subject Sample # 2	0.162	03:30
Breath Volume	1.527	03:30
Air Blank	0.000	03:31
Control Test	0.079	03:31
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 10/26/2017 17:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: SASSER

DATE/TIME OF TEST: 10/21/2017 23:22:46

Diagnostics Check	OK	23:16
Air Blank	0.000	23:17
Control Test	0.078	23:17
Air Blank	0.000	23:17
Subject Sample # 1	0.154	23:18
Breath Volume	2.812	23:18
Air Blank	0.000	23:18
Air Blank	0.000	23:20
Subject Sample # 2	0.161	23:21
Breath Volume	4.480	23:21
Air Blank	0.000	23:21
Control Test	0.078	23:22
Air Blank	0.000	23:22
Diagnostics Check	OK	23:22

Data Download Date/Time: 10/26/2017 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001257

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADRON,MARCIE,L
BREATH TEST OPERATOR AGENCY: Orlando PD
ARRESTING OFFICER: WONGSHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: LOVE

DATE/TIME OF TEST: 10/23/2017 07:34:20

Diagnostics Check	OK	07:24
Air Blank	0.000	07:24
Control Test	0.079	07:24
Air Blank	0.000	07:25
Subject Sample # 1	0.139	07:28
Breath Volume	1.523	07:28
Air Blank	0.000	07:28
Air Blank	0.000	07:30
Subject Sample # 2	0.140	07:32
Breath Volume	1.171	07:32
Air Blank	0.000	07:33
Control Test	0.077	07:33
Air Blank	0.000	07:34
Diagnostics Check	OK	07:34

Data Download Date/Time: 10/26/2017 17:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO,ELDALY,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KASSELMAN,S
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: RIVERA SANTIAGO

DATE/TIME OF TEST: 10/01/2017 04:26:02

Diagnostics Check	OK	04:18
Air Blank	0.000	04:18
Control Test	0.080	04:19
Air Blank	0.000	04:19
Subject Sample # 1	0.105	04:21
Breath Volume	1.972	04:21
Air Blank	0.000	04:22
Air Blank	0.000	04:23
Subject Sample # 2	0.102	04:24
Breath Volume	1.300	04:24
Air Blank	0.000	04:25
Control Test	0.078	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:26

Data Download Date/Time: 10/05/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:46

SUBJECT'S LAST NAME: KELMENTS

DATE/TIME OF TEST: 10/01/2017 05:22:38

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:10	
Control Test		0.078	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1	0.066	05:14	Volume Not Met
Breath Volume		0.621	05:14	
Air Blank		0.000	05:15	
Air Blank		0.000	05:17	
Subject Sample	# 2	0.067	05:17	
Breath Volume		2.484	05:17	
Air Blank		0.000	05:18	
Air Blank		0.000	05:20	
Subject Sample	# 3	0.061	05:21	
Breath Volume		1.511	05:21	
Air Blank		0.000	05:21	
Control Test		0.078	05:22	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 10/05/2017 15:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCWHERTER,PETER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:25

SUBJECT'S LAST NAME: MOLASKEY

DATE/TIME OF TEST: 10/01/2017 18:58:54

Diagnostics Check	OK	18:52
Air Blank	0.000	18:53
Control Test	0.079	18:53
Air Blank	0.000	18:54
Subject Sample # 1	0.000	18:54
Breath Volume	1.324	18:54
Air Blank	0.000	18:55
Air Blank	0.000	18:57
Subject Sample # 2	0.000	18:57
Breath Volume	4.054	18:57
Air Blank	0.000	18:58
Control Test	0.079	18:58
Air Blank	0.000	18:58
Diagnostics Check	OK	18:58

Data Download Date/Time: 10/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER, OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER, DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: MIRANDA-FUENTES

DATE/TIME OF TEST: 10/03/2017 22:44:39

Diagnostics Check	OK	22:37
Air Blank	0.000	22:38
Control Test	0.080	22:38
Air Blank	0.000	22:39
Subject Sample # 1	0.115	22:40
Breath Volume	3.066	22:40
Air Blank	0.000	22:40
Air Blank	0.000	22:42
Subject Sample # 2	0.113	22:43
Breath Volume	3.324	22:43
Air Blank	0.000	22:43
Control Test	0.079	22:44
Air Blank	0.000	22:44
Diagnostics Check	OK	22:44

Data Download Date/Time: 10/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROLAND, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: MCRAE

DATE/TIME OF TEST: 10/05/2017 01:55:18

Diagnostics Check	OK	01:49
Air Blank	0.000	01:49
Control Test	0.079	01:49
Air Blank	0.000	01:50
Subject Sample # 1	0.256	01:50
Breath Volume	1.351	01:50
Air Blank	0.000	01:51
Air Blank	0.000	01:53
Subject Sample # 2	0.267	01:53
Breath Volume	1.179	01:53
Air Blank	0.000	01:54
Control Test	0.078	01:54
Air Blank	0.000	01:55
Diagnostics Check	OK	01:55

Data Download Date/Time: 10/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE, ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:25

SUBJECT'S LAST NAME: RADINGER

DATE/TIME OF TEST: 10/05/2017 07:55:10

Diagnostics Check	OK	07:49
Air Blank	0.000	07:49
Control Test	0.079	07:49
Air Blank	0.000	07:50
Subject Sample # 1	0.000	07:50
Breath Volume	1.691	07:50
Air Blank	0.000	07:51
Air Blank	0.000	07:53
Subject Sample # 2	0.000	07:53
Breath Volume	3.390	07:53
Air Blank	0.000	07:54
Control Test	0.079	07:54
Air Blank	0.000	07:55
Diagnostics Check	OK	07:55

Data Download Date/Time: 10/05/2017 15:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:54

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/06/2017 22:25:08

Diagnostics Check	OK	22:18
Air Blank	0.000	22:19
Control Test	0.080	22:19
Air Blank	0.000	22:19
Subject Sample # 1	0.118	22:20
Breath Volume	1.558	22:20
Air Blank	0.000	22:21
Air Blank	0.000	22:23
Subject Sample # 2	0.120	22:23
Breath Volume	1.312	22:23
Air Blank	0.000	22:24
Control Test	0.079	22:24
Air Blank	0.000	22:25
Diagnostics Check	OK	22:25

Data Download Date/Time: 10/11/2017 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PROCTOR, RHETT
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: SALDANA VIRGUEZ

DATE/TIME OF TEST: 10/07/2017 04:21:34

Diagnostics Check	OK	04:15
Air Blank	0.000	04:15
Control Test	0.080	04:16
Air Blank	0.000	04:16
Subject Sample # 1	0.049	04:17
Breath Volume	2.023	04:17
Air Blank	0.000	04:17
Air Blank	0.000	04:19
Subject Sample # 2	0.046	04:20
Breath Volume	1.121	04:20
Air Blank	0.000	04:20
Control Test	0.079	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 10/11/2017 11:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH, STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:17

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 10/07/2017 06:52:12

Diagnostics Check		OK	06:39
Air Blank		0.000	06:40
Control Test		0.080	06:40
Air Blank		0.000	06:40
Subject Sample	# 1	0.238	06:42
Breath Volume		1.531	06:42
Air Blank		0.000	06:42
Air Blank		0.000	06:44
Subject Sample	# 2	0.268	06:45
Breath Volume		2.699	06:45
Air Blank		0.000	06:46
Air Blank		0.000	06:48
Subject Sample	# 3	0.277	06:50
Breath Volume		1.691	06:50
Air Blank		0.000	06:51
Control Test		0.078	06:51
Air Blank		0.000	06:52
Diagnostics Check		OK	06:52

Data Download Date/Time: 10/11/2017 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: SYLVIA

DATE/TIME OF TEST: 10/08/2017 02:05:45

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.080	02:00
Air Blank	0.000	02:00
Subject Sample # 1	0.097	02:01
Breath Volume	2.496	02:01
Air Blank	0.000	02:01
Air Blank	0.000	02:03
Subject Sample # 2	0.096	02:04
Breath Volume	2.496	02:04
Air Blank	0.000	02:04
Control Test	0.077	02:05
Air Blank	0.000	02:05
Diagnostics Check	OK	02:05

Data Download Date/Time: 10/11/2017 11:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MAYORGA,OSCAR
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: BRADFORD

DATE/TIME OF TEST: 10/11/2017 23:21:43

Diagnostics Check	OK	23:15
Air Blank	0.000	23:16
Control Test	0.079	23:16
Air Blank	0.000	23:16
Subject Sample # 1	0.116	23:17
Breath Volume	2.390	23:17
Air Blank	0.000	23:17
Air Blank	0.000	23:19
Subject Sample # 2	0.117	23:20
Breath Volume	2.472	23:20
Air Blank	0.000	23:20
Control Test	0.079	23:21
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: DITROIA

DATE/TIME OF TEST: 10/12/2017 01:49:31

Diagnostics Check	OK	01:41
Air Blank	0.000	01:41
Control Test	0.079	01:42
Air Blank	0.000	01:42
Subject Sample # 1	0.194	01:43
Breath Volume	1.105	01:43
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.196	01:47
Breath Volume	2.050	01:47
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Diagnostics Check	OK	01:49

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LEROY,SANTIA
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 10/12/2017 04:48:15

Diagnostics Check	OK	04:41
Air Blank	0.000	04:42
Control Test	0.079	04:42
Air Blank	0.000	04:42
Subject Sample # 1	0.230	04:43
Breath Volume	4.203	04:43
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.234	04:46
Breath Volume	3.902	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:47
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CRUZ,AYLER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: PATINO

DATE/TIME OF TEST: 10/12/2017 23:38:43

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:34
Breath Volume	1.382	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.000	23:37
Breath Volume	2.339	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COMPTON,A
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: GARRETT

DATE/TIME OF TEST: 10/13/2017 01:41:29

Diagnostics Check		OK	01:29	
Air Blank		0.000	01:29	
Control Test		0.080	01:29	
Air Blank		0.000	01:30	
Subject Sample	# 1	0.177	01:33	Volume Not Met
Breath Volume		0.140	01:33	
Air Blank		0.000	01:34	
Air Blank		0.000	01:35	
Subject Sample	# 2	0.171	01:36	
Breath Volume		1.882	01:36	
Air Blank		0.000	01:37	
Air Blank		0.000	01:39	
Subject Sample	# 3	0.179	01:39	
Breath Volume		1.687	01:39	
Air Blank		0.000	01:40	
Control Test		0.077	01:40	
Air Blank		0.000	01:41	
Diagnostics Check		OK	01:41	

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SHEIMAN, RICHARD
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: DENIS

DATE/TIME OF TEST: 10/13/2017 12:34:18

Diagnostics Check	OK	12:25	
Air Blank	0.000	12:26	
Control Test	0.079	12:26	
Air Blank	0.000	12:27	
Subject Sample	# 1	12:30	No Sample Provided
Breath Volume	0.000	12:30	
Air Blank	0.000	12:30	
Air Blank	0.000	12:32	
Subject Sample	# 2	12:32	Subject Test Refused
Breath Volume	0.000	12:32	
Air Blank	0.000	12:33	
Control Test	0.079	12:33	
Air Blank	0.000	12:34	
Diagnostics Check	OK	12:34	

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ,ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: AGUERO MORCIEGO

DATE/TIME OF TEST: 10/14/2017 01:38:51

Diagnostics Check	OK	01:31
Air Blank	0.000	01:31
Control Test	0.079	01:32
Air Blank	0.000	01:32
Subject Sample # 1	0.150	01:33
Breath Volume	2.074	01:33
Air Blank	0.000	01:34
Air Blank	0.000	01:36
Subject Sample # 2	0.147	01:37
Breath Volume	1.539	01:37
Air Blank	0.000	01:37
Control Test	0.079	01:38
Air Blank	0.000	01:38
Diagnostics Check	OK	01:38

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JOHNSON,LATESHA
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/15/2017 03:39:34

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.318	03:35
Breath Volume		2.480	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.323	03:37
Breath Volume		2.753	03:37
Air Blank		0.000	03:38
Control Test		0.077	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HERNANDEZ, EDUARDO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 10/15/2017 09:51:17

Diagnostics Check	OK	09:44
Air Blank	0.000	09:44
Control Test	0.079	09:45
Air Blank	0.000	09:45
Subject Sample # 1	0.040	09:46
Breath Volume	2.132	09:46
Air Blank	0.000	09:47
Air Blank	0.000	09:49
Subject Sample # 2	0.040	09:49
Breath Volume	1.871	09:49
Air Blank	0.000	09:50
Control Test	0.078	09:50
Air Blank	0.000	09:51
Diagnostics Check	OK	09:51

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: HUERTA

DATE/TIME OF TEST: 10/16/2017 05:28:53

Diagnostics Check	OK	05:22
Air Blank	0.000	05:23
Control Test	0.079	05:23
Air Blank	0.000	05:23
Subject Sample # 1	0.209	05:24
Breath Volume	3.238	05:24
Air Blank	0.000	05:25
Air Blank	0.000	05:26
Subject Sample # 2	0.206	05:27
Breath Volume	3.335	05:27
Air Blank	0.000	05:27
Control Test	0.079	05:28
Air Blank	0.000	05:28
Diagnostics Check	OK	05:28

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PELAEZ

DATE/TIME OF TEST: 10/17/2017 03:59:29

Diagnostics Check	OK	03:53
Air Blank	0.000	03:53
Control Test	0.079	03:53
Air Blank	0.000	03:54
Subject Sample # 1	0.112	03:55
Breath Volume	1.652	03:55
Air Blank	0.000	03:55
Air Blank	0.000	03:57
Subject Sample # 2	0.111	03:57
Breath Volume	1.511	03:57
Air Blank	0.000	03:58
Control Test	0.078	03:58
Air Blank	0.000	03:59
Diagnostics Check	OK	03:59

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COX,JONATHAN
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: HASAN

DATE/TIME OF TEST: 10/18/2017 03:31:03

Diagnostics Check	OK	03:28	
Air Blank	0.000	03:28	
Control Test	0.080	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:29	Subject Test Refused
Breath Volume	0.000	03:29	
Air Blank	0.000	03:30	
Control Test	0.079	03:30	
Air Blank	0.000	03:30	
Diagnostics Check	OK	03:31	

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GLADHILL,KAITLYN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: REID

DATE/TIME OF TEST: 10/19/2017 00:24:03

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.079	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.173	00:18
Breath Volume	1.179	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:20
Subject Sample # 2	0.166	00:22
Breath Volume	1.289	00:22
Air Blank	0.000	00:23
Control Test	0.078	00:23
Air Blank	0.000	00:23
Diagnostics Check	OK	00:24

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:37

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/19/2017 05:05:24

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.137	05:01
Breath Volume	2.468	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.128	05:03
Breath Volume	1.445	05:03
Air Blank	0.000	05:04
Control Test	0.080	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 10/19/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SANTIAGO,WILLIAM
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:07

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 10/19/2017 23:39:11

Diagnostics Check	OK	23:29
Air Blank	0.000	23:30
Control Test	0.079	23:30
Air Blank	0.000	23:31
Subject Sample # 1	0.160	23:33
Breath Volume	1.265	23:33
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.164	23:37
Breath Volume	1.312	23:37
Air Blank	0.000	23:38
Control Test	0.077	23:38
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 10/26/2017 17:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 10/26/2017 05:30:20

Diagnostics Check	OK	05:23
Air Blank	0.000	05:23
Control Test	0.079	05:23
Air Blank	0.000	05:24
Subject Sample # 1	0.239	05:25
Breath Volume	2.828	05:25
Air Blank	0.000	05:26
Air Blank	0.000	05:27
Subject Sample # 2	0.251	05:28
Breath Volume	2.824	05:28
Air Blank	0.000	05:29
Control Test	0.078	05:29
Air Blank	0.000	05:30
Diagnostics Check	OK	05:30

Data Download Date/Time: 10/26/2017 17:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BORRAS,TAMARA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/27/2017 00:19:37

Diagnostics Check	OK	00:12
Air Blank	0.000	00:13
Control Test	0.080	00:13
Air Blank	0.000	00:14
Subject Sample # 1	0.149	00:14
Breath Volume	2.558	00:14
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.136	00:18
Breath Volume	1.773	00:18
Air Blank	0.000	00:18
Control Test	0.079	00:19
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 11/02/2017 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: CARTAGENA

DATE/TIME OF TEST: 10/28/2017 05:13:12

Diagnostics Check	OK	05:05
Air Blank	0.000	05:06
Control Test	0.079	05:06
Air Blank	0.000	05:07
Subject Sample # 1	0.126	05:08
Breath Volume	1.179	05:08
Air Blank	0.000	05:09
Air Blank	0.000	05:11
Subject Sample # 2	0.129	05:11
Breath Volume	1.300	05:11
Air Blank	0.000	05:12
Control Test	0.079	05:12
Air Blank	0.000	05:13
Diagnostics Check	OK	05:13

Data Download Date/Time: 11/02/2017 15:00

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001258

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GAUTAM,ROBERT
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:36

SUBJECT'S LAST NAME: BARRALAGO

DATE/TIME OF TEST: 10/29/2017 02:06:11

Diagnostics Check	OK	01:59
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Subject Sample # 1	0.205	02:01
Breath Volume	2.808	02:01
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.201	02:04
Breath Volume	2.976	02:04
Air Blank	0.000	02:05
Control Test	0.077	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 11/02/2017 15:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PICHECO,JOHN,A
BREATH TEST OPERATOR AGENCY: Cocoa PD
ARRESTING OFFICER: TORELLO,AMANDA
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:46

SUBJECT'S LAST NAME: VORACHACK

DATE/TIME OF TEST: 10/01/2017 18:16:39

Diagnostics Check	OK	18:09
Air Blank	0.000	18:10
Control Test	0.080	18:10
Air Blank	0.000	18:10
Subject Sample # 1	0.133	18:11
Breath Volume	1.250	18:11
Air Blank	0.000	18:12
Air Blank	0.000	18:14
Subject Sample # 2	0.127	18:15
Breath Volume	1.468	18:15
Air Blank	0.000	18:15
Control Test	0.078	18:16
Air Blank	0.000	18:16
Diagnostics Check	OK	18:16

Data Download Date/Time: 10/28/2017 20:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001260

INSTRUMENT REGISTERED TO: COCOA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK,NICHOLAS,B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: HERNANDEZ,L
ARRESTING OFFICER AGENCY: Cocoa PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 10/21/2017 03:58:17

Diagnostics Check	OK	03:55	
Air Blank	0.000	03:55	
Control Test	0.081	03:56	
Air Blank	0.000	03:56	
Subject Sample	# 1	03:56	Subject Test Refused
Breath Volume	0.000	03:56	
Air Blank	0.000	03:57	
Control Test	0.080	03:57	
Air Blank	0.000	03:58	
Diagnostics Check	OK	03:58	

Data Download Date/Time: 10/28/2017 20:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: KENNEDY,PATRICK
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: BOEHM

DATE/TIME OF TEST: 10/04/2017 04:18:47

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Subject Sample	# 1	0.260	04:14
Breath Volume		2.691	04:14
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.276	04:17
Breath Volume		2.464	04:17
Air Blank		0.000	04:17
Control Test		0.079	04:18
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 10/26/2017 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,ROBERT,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LYNCH,ROBERT
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 10/21/2017 03:34:01

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.079	03:23	
Air Blank		0.000	03:23	
Subject Sample	# 1	0.194	03:26	Volume Not Met
Breath Volume		0.187	03:26	
Air Blank		0.000	03:27	
Air Blank		0.000	03:29	
Subject Sample	# 2	0.184	03:32	Volume Not Met
Breath Volume		0.152	03:32	
Air Blank		0.000	03:33	
Control Test		0.078	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:34	

Data Download Date/Time: 10/26/2017 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYNCH,ROBERT,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: LYNCH,ROBERT
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 10/21/2017 03:56:15

Diagnostics Check		OK	03:39	
Air Blank		0.000	03:39	
Control Test		0.078	03:39	
Air Blank		0.000	03:40	
Subject Sample	# 1	0.185	03:43	
Breath Volume		1.253	03:43	
Air Blank		0.000	03:44	
Air Blank		0.000	03:45	
Subject Sample	# 2	0.168	03:49	Volume Not Met
Breath Volume		0.703	03:49	
Air Blank		0.000	03:49	
Air Blank		0.000	03:51	
Subject Sample	# 3	0.153	03:54	Volume Not Met
Breath Volume		0.363	03:54	
Air Blank		0.000	03:55	
Control Test		0.077	03:55	
Air Blank		0.000	03:56	
Diagnostics Check		OK	03:56	

Data Download Date/Time: 10/26/2017 02:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,PATRICK,W
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: KENNEDY,PATRICK
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 23:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:26

SUBJECT'S LAST NAME: COOKSHOTT

DATE/TIME OF TEST: 10/21/2017 23:58:19

Diagnostics Check	OK	23:51
Air Blank	0.000	23:52
Control Test	0.080	23:52
Air Blank	0.000	23:52
Subject Sample # 1	0.000	23:53
Breath Volume	4.503	23:53
Air Blank	0.000	23:54
Air Blank	0.000	23:56
Subject Sample # 2	0.000	23:56
Breath Volume	3.507	23:56
Air Blank	0.000	23:57
Control Test	0.078	23:57
Air Blank	0.000	23:58
Diagnostics Check	OK	23:58

Data Download Date/Time: 10/26/2017 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASEY,COREY,M
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: SIMMONS,BRIAN
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 09/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: SESCO

DATE/TIME OF TEST: 10/25/2017 02:36:57

Diagnostics Check	OK	02:34	
Air Blank	0.000	02:34	
Control Test	0.080	02:34	
Air Blank	0.000	02:35	
Subject Sample	# 1	02:35	Subject Test Refused
Breath Volume	0.000	02:35	
Air Blank	0.000	02:36	
Control Test	0.080	02:36	
Air Blank	0.000	02:36	
Diagnostics Check	OK	02:36	

Data Download Date/Time: 10/26/2017 02:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001261

INSTRUMENT REGISTERED TO: ROCKLEDGE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBROOK, NICHOLAS, B
BREATH TEST OPERATOR AGENCY: Rockledge PD
ARRESTING OFFICER: HERBENER, TIMOTHY
ARRESTING OFFICER AGENCY: Rockledge PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/13/2019
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: POUNCEY

DATE/TIME OF TEST: 10/27/2017 01:59:45

Diagnostics Check	OK	01:52
Air Blank	0.000	01:53
Control Test	0.079	01:53
Air Blank	0.000	01:54
Subject Sample # 1	0.231	01:55
Breath Volume	1.203	01:55
Air Blank	0.000	01:55
Air Blank	0.000	01:57
Subject Sample # 2	0.227	01:58
Breath Volume	1.503	01:58
Air Blank	0.000	01:58
Control Test	0.078	01:59
Air Blank	0.000	01:59
Diagnostics Check	OK	01:59

Data Download Date/Time: 11/28/2017 11:29

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANQUI,DANIEL,C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: FRANQUI,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 10/22/2017 03:42:47

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.080	03:31	
Air Blank		0.000	03:32	
Subject Sample	# 1	0.118	03:35	Volume Not Met
Breath Volume		0.101	03:35	
Air Blank		0.000	03:36	
Air Blank		0.000	03:37	
Subject Sample	# 2	0.101	03:40	Volume Not Met
Breath Volume		0.019	03:40	
Air Blank		0.000	03:41	
Control Test		0.080	03:42	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 11/29/2017 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANQUI,DANIEL,C
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: FRANQUI,DANIEL
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 10/22/2017 03:57:14

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:46	
Control Test		0.078	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.105	03:50	Volume Not Met
Breath Volume		0.089	03:50	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 2	0.081	03:55	Volume Not Met
Breath Volume		0.238	03:55	
Air Blank		0.000	03:56	
Control Test		0.079	03:56	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 11/29/2017 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,JEFFERY,L
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: BLAIR,JEFFERY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 09:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:19

SUBJECT'S LAST NAME: DEFRANCESCO

DATE/TIME OF TEST: 10/25/2017 10:51:07

Diagnostics Check	OK	10:44
Air Blank	0.000	10:44
Control Test	0.081	10:44
Air Blank	0.000	10:45
Subject Sample # 1	0.000	10:46
Breath Volume	2.371	10:46
Air Blank	0.000	10:46
Air Blank	0.000	10:48
Subject Sample # 2	0.000	10:49
Breath Volume	2.480	10:49
Air Blank	0.000	10:50
Control Test	0.081	10:50
Air Blank	0.000	10:51
Diagnostics Check	OK	10:51

Data Download Date/Time: 11/29/2017 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER, MATTHEW, G
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DUMMER, LUKE
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MONRREAL RODRIGUEZ

DATE/TIME OF TEST: 10/29/2017 02:49:42

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:43
Air Blank	0.000	02:44
Subject Sample # 1	0.193	02:44
Breath Volume	3.828	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.192	02:48
Breath Volume	3.828	02:48
Air Blank	0.000	02:48
Control Test	0.078	02:49
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 11/29/2017 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLES,GABRIEL,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DAVIS,ANTHONY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DITTIGER

DATE/TIME OF TEST: 10/29/2017 03:54:57

Diagnostics Check	OK	03:54	
Air Blank		03:54	RFI Detect
Air Blank	0.000	03:54	

Data Download Date/Time: 11/29/2017 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOLES,GABRIEL,D
BREATH TEST OPERATOR AGENCY: Melbourne PD
ARRESTING OFFICER: DAVIS,ANTHONY
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: DITTIGER

DATE/TIME OF TEST: 10/29/2017 04:04:47

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.259	03:59
Breath Volume	1.484	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.268	04:03
Breath Volume	1.425	04:03
Air Blank	0.000	04:03
Control Test	0.081	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 11/29/2017 12:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001262

INSTRUMENT REGISTERED TO: MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LACEY,JOHN,J
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: WHITE,TAYLOR
ARRESTING OFFICER AGENCY: Melbourne PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 04:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:07

SUBJECT'S LAST NAME: HARDEN

DATE/TIME OF TEST: 10/30/2017 04:57:13

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.074	04:52
Breath Volume	1.718	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:54
Subject Sample # 2	0.069	04:55
Breath Volume	1.914	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 11/29/2017 12:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAMLETT,TIMOTHY,J
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: BRAMLETT,T
ARRESTING OFFICER AGENCY: Indian Harbour Beach
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: STREB

DATE/TIME OF TEST: 10/06/2017 02:30:31

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.080	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.109	02:25
Breath Volume	2.191	02:25
Air Blank	0.000	02:26
Air Blank	0.000	02:28
Subject Sample # 2	0.116	02:28
Breath Volume	1.480	02:28
Air Blank	0.000	02:29
Control Test	0.081	02:29
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 11/19/2017 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: MCCORD

DATE/TIME OF TEST: 10/20/2017 00:59:42

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.081	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:58	Subject Test Refused
Breath Volume	0.000	00:58	
Air Blank	0.000	00:58	
Control Test	0.081	00:59	
Air Blank	0.000	00:59	
Diagnostics Check	OK	00:59	

Data Download Date/Time: 11/19/2017 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MIDDLEBROOKS,JAMES,E
BREATH TEST OPERATOR AGENCY: Indian Harbour Beach
ARRESTING OFFICER: MARTINEZ,PHILIP
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:21

SUBJECT'S LAST NAME: MIDDLEBROOKS

DATE/TIME OF TEST: 10/26/2017 15:51:21

Diagnostics Check	OK	15:45
Air Blank	0.000	15:45
Control Test	0.081	15:45
Air Blank	0.000	15:46
Subject Sample # 1	0.163	15:46
Breath Volume	1.156	15:46
Air Blank	0.000	15:47
Air Blank	0.000	15:49
Subject Sample # 2	0.159	15:49
Breath Volume	1.597	15:49
Air Blank	0.000	15:50
Control Test	0.081	15:50
Air Blank	0.000	15:51
Diagnostics Check	OK	15:51

Data Download Date/Time: 11/19/2017 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAY, AMBERLY, L
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: CLAY, AMBERLY
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: PEABODY

DATE/TIME OF TEST: 10/27/2017 04:08:36

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:03
Subject Sample # 1	0.119	04:04
Breath Volume	1.257	04:04
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.123	04:07
Breath Volume	1.566	04:07
Air Blank	0.000	04:07
Control Test	0.079	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 11/19/2017 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PAULUS

DATE/TIME OF TEST: 10/29/2017 02:52:06

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:36	
Control Test		0.081	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1	0.134	02:40	Volume Not Met
Breath Volume		0.675	02:40	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.146	02:44	
Breath Volume		1.281	02:44	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 3	0.156	02:50	Volume Not Met
Breath Volume		0.234	02:50	
Air Blank		0.000	02:51	
Control Test		0.078	02:51	
Air Blank		0.000	02:52	
Diagnostics Check		OK	02:52	

Data Download Date/Time: 11/19/2017 09:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,R
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: PAULUS

DATE/TIME OF TEST: 10/29/2017 03:02:53

Diagnostics Check	OK	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Subject Sample # 1	0.149	02:57
Breath Volume	1.214	02:57
Air Blank	0.000	02:58
Air Blank	0.000	02:59
Subject Sample # 2	0.146	03:01
Breath Volume	1.550	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 11/19/2017 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001263

INSTRUMENT REGISTERED TO: INDIAN HARBOUR PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOUSMAN,JOHN,D
BREATH TEST OPERATOR AGENCY: Satellite Beach PD
ARRESTING OFFICER: HOUSMAN,JOHN
ARRESTING OFFICER AGENCY: Satellite Beach PD
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:12

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/29/2017 06:42:48

Diagnostics Check	OK	06:35
Air Blank	0.000	06:36
Control Test	0.080	06:36
Air Blank	0.000	06:36
Subject Sample # 1	0.148	06:38
Breath Volume	1.652	06:38
Air Blank	0.000	06:38
Air Blank	0.000	06:40
Subject Sample # 2	0.154	06:41
Breath Volume	1.191	06:41
Air Blank	0.000	06:41
Control Test	0.080	06:42
Air Blank	0.000	06:42
Diagnostics Check	OK	06:42

Data Download Date/Time: 11/19/2017 09:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BULA

DATE/TIME OF TEST: 10/01/2017 01:18:25

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:12	
Control Test	0.080	01:13	
Air Blank	0.000	01:13	
Subject Sample	# 1 0.156	01:16	Volume Not Met
Breath Volume	0.121	01:16	
Air Blank		01:17	Purge Fail
Air Blank		01:18	Purge Fail

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BULA

DATE/TIME OF TEST: 10/01/2017 01:21:29

Diagnostics Check	OK	01:20	
Air Blank	0.000	01:20	
Control Test	0.067	01:20	Control Outside Tolerance
Air Blank	0.000	01:21	

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BULA

DATE/TIME OF TEST: 10/01/2017 01:24:16

Diagnostics Check	OK	01:22	
Air Blank	0.000	01:23	
Control Test	0.073	01:23	Control Outside Tolerance
Air Blank	0.000	01:24	

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BULA

DATE/TIME OF TEST: 10/01/2017 01:32:16

Diagnostics Check	OK	01:26	
Air Blank	0.000	01:27	
Control Test	0.076	01:27	
Air Blank	0.000	01:27	
Subject Sample	# 1 0.158	01:30	
Breath Volume	1.730	01:30	
Air Blank		01:31	Purge Fail
Air Blank	0.000	01:32	

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BRATSCH,CASEY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: BULA

DATE/TIME OF TEST: 10/01/2017 01:35:36

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:34	
Control Test	0.074	01:35	Control Outside Tolerance
Air Blank	0.000	01:35	

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HEBB,PAUL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: TICE

DATE/TIME OF TEST: 10/03/2017 01:17:25

Diagnostics Check	OK	01:13	
Air Blank	0.000	01:14	
Control Test	0.081	01:14	
Air Blank	0.000	01:15	
Subject Sample	# 1	01:16	Subject Test Refused
Breath Volume	0.000	01:16	
Air Blank	0.000	01:16	
Control Test	0.081	01:16	
Air Blank	0.000	01:17	
Diagnostics Check	OK	01:17	

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/07/2017 02:58:19

Diagnostics Check	OK	02:52
Air Blank	0.000	02:52
Control Test	0.080	02:53
Air Blank	0.000	02:53
Subject Sample # 1	0.082	02:53
Breath Volume	2.562	02:53
Air Blank	0.000	02:54
Air Blank	0.000	02:56
Subject Sample # 2	0.083	02:56
Breath Volume	2.406	02:56
Air Blank	0.000	02:57
Control Test	0.080	02:57
Air Blank	0.000	02:58
Diagnostics Check	OK	02:58

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MORETTO,RACHEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: HAWTHORNE

DATE/TIME OF TEST: 10/07/2017 22:53:28

Diagnostics Check	OK	22:47
Air Blank	0.000	22:47
Control Test	0.079	22:48
Air Blank	0.000	22:48
Subject Sample # 1	0.088	22:49
Breath Volume	2.171	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.089	22:51
Breath Volume	2.101	22:51
Air Blank	0.000	22:52
Control Test	0.080	22:52
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,NALICIA,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DONALDSON,LAUREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: TOPPE

DATE/TIME OF TEST: 10/18/2017 18:59:50

Diagnostics Check	OK	18:53
Air Blank	0.000	18:54
Control Test	0.081	18:54
Air Blank	0.000	18:55
Subject Sample # 1	0.000	18:55
Breath Volume	3.230	18:55
Air Blank	0.000	18:56
Air Blank	0.000	18:58
Subject Sample # 2	0.000	18:58
Breath Volume	3.207	18:58
Air Blank	0.000	18:58
Control Test	0.082	18:59
Air Blank	0.000	18:59
Diagnostics Check	OK	18:59

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEVELAND, WILLIAM, S
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DRULARD, BRIAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: LINCOLN

DATE/TIME OF TEST: 10/22/2017 22:27:21

Diagnostics Check	OK	22:23	
Air Blank	0.000	22:24	
Control Test	0.080	22:24	
Air Blank	0.000	22:25	
Subject Sample	# 1	22:25	Subject Test Refused
Breath Volume	0.000	22:25	
Air Blank	0.000	22:26	
Control Test	0.080	22:26	
Air Blank	0.000	22:27	
Diagnostics Check	OK	22:27	

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOWDING,BETHANY,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: DOWDING,BETHANY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: TOLBERT

DATE/TIME OF TEST: 10/25/2017 04:40:32

Diagnostics Check		OK	04:30
Air Blank		0.000	04:31
Control Test		0.079	04:31
Air Blank		0.000	04:31
Subject Sample	# 1	0.142	04:33
Breath Volume		1.644	04:33
Air Blank		0.000	04:34
Air Blank		0.000	04:35
Subject Sample	# 2	0.139	04:39
Breath Volume		1.789	04:39
Air Blank		0.000	04:39
Control Test		0.079	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 10/25/2017 16:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,NALICIA,K
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: WINSTEAD,JUSTIN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: JUAN PABLO

DATE/TIME OF TEST: 10/28/2017 03:46:59

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.081	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.054	03:42
Breath Volume	3.195	03:42
Air Blank	0.000	03:42
Air Blank	0.000	03:44
Subject Sample # 2	0.055	03:45
Breath Volume	2.773	03:45
Air Blank	0.000	03:46
Control Test	0.080	03:46
Air Blank	0.000	03:46
Diagnostics Check	OK	03:46

Data Download Date/Time: 11/16/2017 17:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001264

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHEINER,JAMES,R
BREATH TEST OPERATOR AGENCY: Cocoa Beach PD
ARRESTING OFFICER: PENDER,DEREK
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: IRWIN

DATE/TIME OF TEST: 10/31/2017 04:19:59

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.136	04:15
Breath Volume		4.007	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:17
Subject Sample	# 2	0.139	04:18
Breath Volume		4.042	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 11/16/2017 17:56

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 24515080A2
DRY GAS EXPIRY DATE: 10/05/2017
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: CASELLA

DATE/TIME OF TEST: 10/02/2017 23:03:41

Diagnostics Check	OK	22:57
Air Blank	0.000	22:58
Control Test	0.079	22:58
Air Blank	0.000	22:58
Subject Sample # 1	0.104	22:59
Breath Volume	3.320	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:01
Subject Sample # 2	0.108	23:02
Breath Volume	3.191	23:02
Air Blank	0.000	23:02
Control Test	0.079	23:03
Air Blank	0.000	23:03
Diagnostics Check	OK	23:03

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24515080A2

DRY GAS EXPIRY DATE: 10/05/2017

VIOLATION CODE: DUI

ARREST TIME: 02:06

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: MCMAHAN

DATE/TIME OF TEST: 10/03/2017 02:21:05

Diagnostics Check	OK	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:15
Air Blank	0.000	02:16
Subject Sample # 1	0.150	02:16
Breath Volume	1.843	02:16
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.150	02:19
Breath Volume	1.324	02:19
Air Blank	0.000	02:20
Control Test	0.080	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 24515080A2

DRY GAS EXPIRY DATE: 10/05/2017

VIOLATION CODE: DUI

ARREST TIME: 01:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: GEORGE

DATE/TIME OF TEST: 10/04/2017 01:25:52

Diagnostics Check		OK	01:19
Air Blank		0.000	01:20
Control Test		0.080	01:20
Air Blank		0.000	01:21
Subject Sample	# 1	0.072	01:21
Breath Volume		1.921	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.069	01:24
Breath Volume		2.023	01:24
Air Blank		0.000	01:24
Control Test		0.079	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:14

GENDER: Female

OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: SEYFORD

DATE/TIME OF TEST: 10/10/2017 01:36:16

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.079	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.300	01:27
Breath Volume		1.226	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.279	01:30
Breath Volume		1.289	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 3	0.282	01:34
Breath Volume		1.152	01:34
Air Blank		0.000	01:35
Control Test		0.077	01:35
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 00:26

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/18/2017 01:02:48

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:48	
Control Test		0.078	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1	0.125	00:50	
Breath Volume		2.210	00:50	
Air Blank		0.000	00:50	
Air Blank		0.000	00:52	
Subject Sample	# 2	0.086	00:55	Volume Not Met
Breath Volume		0.601	00:55	
Air Blank		0.000	00:56	
Air Blank		0.000	00:58	
Subject Sample	# 3	0.090	01:01	Volume Not Met
Breath Volume		0.816	01:01	
Air Blank		0.000	01:01	
Control Test		0.078	01:02	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 10/18/2017 01:12:13

Diagnostics Check	OK	01:04	
Air Blank	0.000	01:04	
Control Test	0.078	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1 0.109	01:06	
Breath Volume	2.187	01:06	
Air Blank	0.000	01:07	
Air Blank	0.000	01:09	
Subject Sample	# 2	01:10	Subject Test Refused
Breath Volume	0.000	01:10	
Air Blank	0.000	01:11	
Control Test	0.078	01:11	
Air Blank	0.000	01:12	
Diagnostics Check	OK	01:12	

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: LUTZ,EDWIN,J

BREATH TEST OPERATOR AGENCY: Palm Bay PD

ARRESTING OFFICER: LUTZ,EDWIN

ARRESTING OFFICER AGENCY: Palm Bay PD

LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 22817080A4

DRY GAS EXPIRY DATE: 10/05/2019

VIOLATION CODE: DUI

ARREST TIME: 00:27

GENDER: Male

OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: CABOT

DATE/TIME OF TEST: 10/19/2017 00:47:11

Diagnostics Check	OK	00:43	
Air Blank	0.000	00:44	
Control Test	0.079	00:44	
Air Blank	0.000	00:45	
Subject Sample	# 1	00:45	Subject Test Refused
Breath Volume	0.000	00:45	
Air Blank	0.000	00:46	
Control Test	0.079	00:46	
Air Blank	0.000	00:47	
Diagnostics Check	OK	00:47	

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001266

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUTZ,EDWIN,J
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: LUTZ,EDWIN
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 09/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SCHRODER

DATE/TIME OF TEST: 10/19/2017 21:16:31

Diagnostics Check	OK	21:05		
Air Blank	0.000	21:05		
Control Test	0.079	21:05		
Air Blank	0.000	21:06		
Subject Sample	# 1	21:09	No Sample Provided	
Breath Volume	0.000	21:09		
Air Blank	0.000	21:09		
Air Blank	0.000	21:11		
Subject Sample	# 2	0.182	21:15	Volume Not Met
Breath Volume	0.363	21:15		
Air Blank	0.000	21:15		
Control Test	0.078	21:15		
Air Blank	0.000	21:16		
Diagnostics Check	OK	21:16		

Data Download Date/Time: 10/27/2017 21:59

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: ARGENTINE, BRITTANY
ARRESTING OFFICER AGENCY: Oviedo PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WHITEN

DATE/TIME OF TEST: 10/09/2017 04:44:27

Diagnostics Check	OK	04:33	
Air Blank	0.000	04:33	
Control Test	0.080	04:33	
Air Blank	0.000	04:34	
Subject Sample	# 1 0.178	04:37	Volume Not Met
Breath Volume	0.257	04:37	
Air Blank	0.000	04:38	
Air Blank	0.000	04:40	
Subject Sample	# 2	04:42	Subject Test Refused
Breath Volume	0.039	04:42	
Air Blank	0.000	04:43	
Control Test	0.080	04:43	
Air Blank	0.000	04:44	
Diagnostics Check	OK	04:44	

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMENS,ADAM,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BUHLER,A
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 15:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: BURANOSKY

DATE/TIME OF TEST: 10/10/2017 16:48:09

Diagnostics Check	OK	16:41
Air Blank	0.000	16:41
Control Test	0.080	16:42
Air Blank	0.000	16:42
Subject Sample # 1	0.166	16:43
Breath Volume	5.093	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:45
Subject Sample # 2	0.152	16:46
Breath Volume	3.929	16:46
Air Blank	0.000	16:47
Control Test	0.078	16:47
Air Blank	0.000	16:48
Diagnostics Check	OK	16:48

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERGUSON,RICHARD,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: MOREL,ANTHONY
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:59

SUBJECT'S LAST NAME: DAVIERO

DATE/TIME OF TEST: 10/10/2017 19:43:01

Diagnostics Check	OK	19:35
Air Blank	0.000	19:36
Control Test	0.081	19:36
Air Blank	0.000	19:37
Subject Sample # 1	0.000	19:38
Breath Volume	3.250	19:38
Air Blank	0.000	19:38
Air Blank	0.000	19:40
Subject Sample # 2	0.000	19:41
Breath Volume	3.207	19:41
Air Blank	0.000	19:42
Control Test	0.081	19:42
Air Blank	0.000	19:42
Diagnostics Check	OK	19:43

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: DEL CASTILLO, BRET
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HENNESEY

DATE/TIME OF TEST: 10/11/2017 01:59:28

Diagnostics Check	OK	01:51	
Air Blank	0.000	01:51	
Control Test	0.081	01:51	
Air Blank	0.000	01:52	
Subject Sample	# 1 0.211	01:55	Volume Not Met
Breath Volume	0.695	01:55	
Air Blank	0.000	01:56	
Air Blank	0.000	01:57	
Subject Sample	# 2	01:58	Subject Test Refused
Breath Volume	0.000	01:58	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Diagnostics Check	OK	01:59	

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RICE, MATHEW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CABALLERO

DATE/TIME OF TEST: 10/11/2017 04:29:10

Diagnostics Check	OK	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.191	04:24
Breath Volume	2.558	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.193	04:27
Breath Volume	2.507	04:27
Air Blank	0.000	04:28
Control Test	0.080	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: MEDUGNO, EMILIO
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MORRICAL

DATE/TIME OF TEST: 10/12/2017 02:09:34

Diagnostics Check	OK	02:00
Air Blank	0.000	02:01
Control Test	0.081	02:01
Air Blank	0.000	02:01
Subject Sample # 1	0.122	02:02
Breath Volume	1.921	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.145	02:05
Breath Volume	4.570	02:05
Air Blank	0.000	02:05
Air Blank	0.000	02:07
Subject Sample # 3	0.132	02:08
Breath Volume	3.410	02:08
Air Blank	0.000	02:08
Control Test	0.081	02:08
Air Blank	0.000	02:09
Diagnostics Check	OK	02:09

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HALL, ANDREW
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CHRISTIANO

DATE/TIME OF TEST: 10/12/2017 03:26:42

Diagnostics Check	OK	03:20
Air Blank	0.000	03:21
Control Test	0.081	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.193	03:22
Breath Volume	3.859	03:22
Air Blank	0.000	03:22
Air Blank	0.000	03:24
Subject Sample # 2	0.193	03:25
Breath Volume	4.160	03:25
Air Blank	0.000	03:25
Control Test	0.081	03:26
Air Blank	0.000	03:26
Diagnostics Check	OK	03:26

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HUNT,R
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/12/2017 11:20:09

Diagnostics Check	OK	11:14
Air Blank	0.000	11:14
Control Test	0.081	11:15
Air Blank	0.000	11:15
Subject Sample # 1	0.000	11:15
Breath Volume	3.644	11:15
Air Blank	0.000	11:16
Air Blank	0.000	11:18
Subject Sample # 2	0.000	11:18
Breath Volume	3.242	11:18
Air Blank	0.000	11:19
Control Test	0.081	11:19
Air Blank	0.000	11:20
Diagnostics Check	OK	11:20

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: UMANA,CLAUDIA
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: KOLDENHOVEN

DATE/TIME OF TEST: 10/13/2017 00:31:49

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.080	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.250	00:26
Breath Volume		2.656	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.250	00:30
Breath Volume		2.753	00:30
Air Blank		0.000	00:30
Control Test		0.081	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:31

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BLACKWELL,ROBERT
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: HINNERS

DATE/TIME OF TEST: 10/13/2017 04:41:42

Diagnostics Check	OK	04:34
Air Blank	0.000	04:35
Control Test	0.081	04:35
Air Blank	0.000	04:36
Subject Sample # 1	0.232	04:36
Breath Volume	2.046	04:36
Air Blank	0.000	04:37
Air Blank	0.000	04:39
Subject Sample # 2	0.222	04:40
Breath Volume	1.718	04:40
Air Blank	0.000	04:40
Control Test	0.081	04:41
Air Blank	0.000	04:41
Diagnostics Check	OK	04:41

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RICH,LINDSEY
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: TROMBLY

DATE/TIME OF TEST: 10/13/2017 23:33:33

Diagnostics Check	OK	23:27
Air Blank	0.000	23:27
Control Test	0.080	23:27
Air Blank	0.000	23:28
Subject Sample # 1	0.226	23:29
Breath Volume	3.082	23:29
Air Blank	0.000	23:29
Air Blank	0.000	23:31
Subject Sample # 2	0.225	23:32
Breath Volume	2.640	23:32
Air Blank	0.000	23:32
Control Test	0.080	23:32
Air Blank	0.000	23:33
Diagnostics Check	OK	23:33

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: TORRES CRUZ,KENNY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: DANNER

DATE/TIME OF TEST: 10/14/2017 03:20:51

Diagnostics Check	OK	03:13
Air Blank	0.000	03:14
Control Test	0.081	03:14
Air Blank	0.000	03:15
Subject Sample # 1	0.217	03:16
Breath Volume	1.121	03:16
Air Blank	0.000	03:16
Air Blank	0.000	03:18
Subject Sample # 2	0.227	03:19
Breath Volume	2.191	03:19
Air Blank	0.000	03:19
Control Test	0.081	03:20
Air Blank	0.000	03:20
Diagnostics Check	OK	03:20

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: PRICE,ZACHARY
ARRESTING OFFICER AGENCY: Casselberry PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: BORREGO

DATE/TIME OF TEST: 10/14/2017 03:55:02

Diagnostics Check		OK	03:40	
Air Blank		0.000	03:41	
Control Test		0.081	03:41	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.131	03:45	Volume Not Met
Breath Volume		0.269	03:45	
Air Blank		0.000	03:45	
Air Blank		0.000	03:47	
Subject Sample	# 2	0.124	03:50	
Breath Volume		1.625	03:50	
Air Blank		0.000	03:50	
Air Blank		0.000	03:52	
Subject Sample	# 3	0.132	03:53	
Breath Volume		2.039	03:53	
Air Blank		0.000	03:54	
Control Test		0.080	03:54	
Air Blank		0.000	03:54	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HILAIRE,FRED
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: UPPER

DATE/TIME OF TEST: 10/14/2017 05:44:04

Diagnostics Check	OK	05:37
Air Blank	0.000	05:38
Control Test	0.080	05:38
Air Blank	0.000	05:38
Subject Sample # 1	0.064	05:39
Breath Volume	4.613	05:39
Air Blank	0.000	05:40
Air Blank	0.000	05:42
Subject Sample # 2	0.059	05:42
Breath Volume	3.406	05:42
Air Blank	0.000	05:43
Control Test	0.079	05:43
Air Blank	0.000	05:43
Diagnostics Check	OK	05:44

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERGUSON,RICHARD,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: MANSO,C
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 19:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: PETIT

DATE/TIME OF TEST: 10/14/2017 20:57:24

Diagnostics Check	OK	20:50
Air Blank	0.000	20:51
Control Test	0.081	20:51
Air Blank	0.000	20:51
Subject Sample # 1	0.174	20:52
Breath Volume	3.042	20:52
Air Blank	0.000	20:53
Air Blank	0.000	20:55
Subject Sample # 2	0.177	20:55
Breath Volume	3.910	20:55
Air Blank	0.000	20:56
Control Test	0.081	20:56
Air Blank	0.000	20:57
Diagnostics Check	OK	20:57

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LUXON,M
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 05:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:20

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/15/2017 09:49:24

Diagnostics Check	OK	09:41
Air Blank	0.000	09:42
Control Test	0.080	09:42
Air Blank	0.000	09:43
Subject Sample # 1	0.100	09:43
Breath Volume	1.367	09:43
Air Blank	0.000	09:44
Air Blank	0.000	09:46
Subject Sample # 2	0.104	09:47
Breath Volume	1.105	09:47
Air Blank	0.000	09:48
Control Test	0.079	09:48
Air Blank	0.000	09:49
Diagnostics Check	OK	09:49

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERGUSON,RICHARD,A
BREATH TEST OPERATOR AGENCY: Lake Mary PD
ARRESTING OFFICER: GOMEZ,KRISTOPHER
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 19:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: BOCK III

DATE/TIME OF TEST: 10/15/2017 20:31:58

Diagnostics Check	OK	20:24
Air Blank	0.000	20:25
Control Test	0.081	20:25
Air Blank	0.000	20:25
Subject Sample # 1	0.348	20:26
Breath Volume	2.332	20:26
Air Blank	0.000	20:27
Air Blank	0.000	20:29
Subject Sample # 2	0.347	20:30
Breath Volume	2.285	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:31
Diagnostics Check	OK	20:31

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HENRY, CAITLIN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: STREETER

DATE/TIME OF TEST: 10/17/2017 00:45:40

Diagnostics Check	OK	00:39
Air Blank	0.000	00:39
Control Test	0.081	00:40
Air Blank	0.000	00:40
Subject Sample # 1	0.284	00:41
Breath Volume	1.824	00:41
Air Blank	0.000	00:41
Air Blank	0.000	00:43
Subject Sample # 2	0.268	00:44
Breath Volume	1.292	00:44
Air Blank	0.000	00:44
Control Test	0.081	00:45
Air Blank	0.000	00:45
Diagnostics Check	OK	00:45

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: GONZALEZ, JUAN
ARRESTING OFFICER AGENCY: Longwood PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PENTECOST

DATE/TIME OF TEST: 10/17/2017 02:18:44

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.081	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.127	02:14
Breath Volume	4.140	02:14
Air Blank	0.000	02:14
Air Blank	0.000	02:16
Subject Sample # 2	0.125	02:17
Breath Volume	4.343	02:17
Air Blank	0.000	02:17
Control Test	0.081	02:18
Air Blank	0.000	02:18
Diagnostics Check	OK	02:18

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: REHDER, PATRICK
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: LAWSON

DATE/TIME OF TEST: 10/22/2017 03:56:52

Diagnostics Check	OK	03:50
Air Blank	0.000	03:51
Control Test	0.080	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.140	03:52
Breath Volume	1.535	03:52
Air Blank	0.000	03:53
Air Blank	0.000	03:55
Subject Sample # 2	0.129	03:55
Breath Volume	1.312	03:55
Air Blank	0.000	03:55
Control Test	0.080	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANCION,CALIX,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RIGGS,B
ARRESTING OFFICER AGENCY: Altamonte Springs PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 15:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:48

SUBJECT'S LAST NAME: BOJARA

DATE/TIME OF TEST: 10/22/2017 16:18:30

Diagnostics Check	OK	16:12
Air Blank	0.000	16:12
Control Test	0.080	16:12
Air Blank	0.000	16:13
Subject Sample # 1	0.136	16:14
Breath Volume	1.402	16:14
Air Blank	0.000	16:14
Air Blank	0.000	16:16
Subject Sample # 2	0.128	16:16
Breath Volume	1.296	16:16
Air Blank	0.000	16:17
Control Test	0.081	16:17
Air Blank	0.000	16:18
Diagnostics Check	OK	16:18

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,PAUL,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: BUHLER,ALLISON
ARRESTING OFFICER AGENCY: Winter Springs PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 13:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:12

SUBJECT'S LAST NAME: ROSARIO MAISONET

DATE/TIME OF TEST: 10/23/2017 14:39:36

Diagnostics Check	OK	14:36	
Air Blank	0.000	14:37	
Control Test	0.081	14:37	
Air Blank	0.000	14:37	
Subject Sample	# 1	14:38	Subject Test Refused
Breath Volume	0.000	14:38	
Air Blank	0.000	14:38	
Control Test	0.081	14:39	
Air Blank	0.000	14:39	
Diagnostics Check	OK	14:39	

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,PAUL,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: LANGWORTHY,HAROLD
ARRESTING OFFICER AGENCY: Lake Mary PD
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 14:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: SCHOND

DATE/TIME OF TEST: 10/23/2017 15:27:59

Diagnostics Check	OK	15:22
Air Blank	0.000	15:22
Control Test	0.081	15:22
Air Blank	0.000	15:23
Subject Sample # 1	0.000	15:23
Breath Volume	2.226	15:23
Air Blank	0.000	15:24
Air Blank	0.000	15:26
Subject Sample # 2	0.000	15:26
Breath Volume	1.308	15:26
Air Blank	0.000	15:27
Control Test	0.082	15:27
Air Blank	0.000	15:27
Diagnostics Check	OK	15:27

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA, RAY,
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: STEWART, NICHOLAS
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: RIDGEWELL-DODD

DATE/TIME OF TEST: 10/23/2017 22:51:19

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.199	22:46
Breath Volume	1.808	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.193	22:49
Breath Volume	2.746	22:49
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VILA,GEORGE,N
BREATH TEST OPERATOR AGENCY: Altamonte Springs PD
ARRESTING OFFICER: GROH,STEPHEN
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 12:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:25

SUBJECT'S LAST NAME: GRAY

DATE/TIME OF TEST: 10/26/2017 14:04:41

Diagnostics Check	OK	13:56
Air Blank	0.000	13:57
Control Test	0.082	13:57
Air Blank	0.000	13:57
Subject Sample # 1	0.303	13:59
Breath Volume	2.515	13:59
Air Blank	0.000	14:00
Air Blank	0.000	14:02
Subject Sample # 2	0.305	14:03
Breath Volume	1.558	14:03
Air Blank	0.000	14:03
Control Test	0.081	14:04
Air Blank	0.000	14:04
Diagnostics Check	OK	14:04

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRIOS,STEPHANIE,A
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HALL,JONATHAN
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 10/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: IRIARTE

DATE/TIME OF TEST: 10/28/2017 01:29:55

Diagnostics Check	OK	01:18	
Air Blank	0.000	01:18	
Control Test	0.081	01:19	
Air Blank	0.000	01:19	
Subject Sample	# 1	01:22	No Sample Provided
Breath Volume	0.000	01:22	
Air Blank	0.000	01:23	
Air Blank	0.000	01:25	
Subject Sample	# 2	01:28	No Sample Provided
Breath Volume	0.000	01:28	
Air Blank	0.000	01:29	
Control Test	0.081	01:29	
Air Blank	0.000	01:29	
Diagnostics Check	OK	01:29	

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BETHAM,KEITH,D
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: HOOPER,JUSTINE
ARRESTING OFFICER AGENCY: Sanford PD
LAST AGENCY INSPECTION: 10/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:09

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 10/28/2017 04:37:52

Diagnostics Check	OK	04:30
Air Blank	0.000	04:31
Control Test	0.079	04:31
Air Blank	0.000	04:31
Subject Sample # 1	0.096	04:32
Breath Volume	2.050	04:32
Air Blank	0.000	04:33
Air Blank	0.000	04:34
Subject Sample # 2	0.088	04:36
Breath Volume	1.675	04:36
Air Blank	0.000	04:36
Control Test	0.080	04:37
Air Blank	0.000	04:37
Diagnostics Check	OK	04:37

Data Download Date/Time: 10/30/2017 23:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001272

INSTRUMENT REGISTERED TO: SEMINOLE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RIGGINS,PAUL,J
BREATH TEST OPERATOR AGENCY: Seminole County SO
ARRESTING OFFICER: RIZZO,ALEXANDRA
ARRESTING OFFICER AGENCY: Seminole County SO
LAST AGENCY INSPECTION: 10/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 282638
DRY GAS EXPIRY DATE: 02/15/2019
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:25

SUBJECT'S LAST NAME: LAURENT

DATE/TIME OF TEST: 10/29/2017 06:53:21

Diagnostics Check	OK	06:45
Air Blank	0.000	06:46
Control Test	0.081	06:46
Air Blank	0.000	06:47
Subject Sample # 1	0.104	06:48
Breath Volume	1.667	06:48
Air Blank	0.000	06:49
Air Blank	0.000	06:51
Subject Sample # 2	0.099	06:51
Breath Volume	1.554	06:51
Air Blank	0.000	06:52
Control Test	0.080	06:52
Air Blank	0.000	06:53
Diagnostics Check	OK	06:53

Data Download Date/Time: 10/30/2017 23:08