

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: BERNAL

DATE/TIME OF TEST: 10/02/2017 21:22:54

Diagnostics Check	OK	21:18	
Air Blank	0.000	21:19	
Control Test	0.081	21:19	
Air Blank	0.000	21:20	
Subject Sample	# 1	21:21	Subject Test Refused
Breath Volume	0.000	21:21	
Air Blank	0.000	21:22	
Control Test	0.082	21:22	
Air Blank	0.000	21:22	
Diagnostics Check	OK	21:22	

Data Download Date/Time: 11/01/2017 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/18/2017 22:14:23

Diagnostics Check	OK	22:08
Air Blank	0.000	22:08
Control Test	0.082	22:08
Air Blank	0.000	22:09
Subject Sample # 1	0.257	22:09
Breath Volume	4.339	22:09
Air Blank	0.000	22:10
Air Blank	0.000	22:12
Subject Sample # 2	0.255	22:12
Breath Volume	4.761	22:12
Air Blank	0.000	22:13
Control Test	0.082	22:13
Air Blank	0.000	22:14
Diagnostics Check	OK	22:14

Data Download Date/Time: 11/01/2017 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HODGES,TITUS,R
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: THALER,NICOHLAS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: MACHADO

DATE/TIME OF TEST: 10/20/2017 05:34:37

Diagnostics Check	OK	05:28
Air Blank	0.000	05:28
Control Test	0.082	05:29
Air Blank	0.000	05:29
Subject Sample # 1	0.188	05:30
Breath Volume	3.402	05:30
Air Blank	0.000	05:30
Air Blank	0.000	05:32
Subject Sample # 2	0.176	05:33
Breath Volume	3.671	05:33
Air Blank	0.000	05:33
Control Test	0.083	05:34
Air Blank	0.000	05:34
Diagnostics Check	OK	05:34

Data Download Date/Time: 11/01/2017 13:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLORZANO,DYLAN,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: KILMURRAY,CHRIS
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 279870
DRY GAS EXPIRY DATE: 02/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 10/23/2017 01:15:55

Diagnostics Check	OK	01:12	
Air Blank	0.000	01:13	
Control Test	0.081	01:13	
Air Blank	0.000	01:14	
Subject Sample	# 1	01:14	Subject Test Refused
Breath Volume	0.000	01:14	
Air Blank	0.000	01:15	
Control Test	0.082	01:15	
Air Blank	0.000	01:15	
Diagnostics Check	OK	01:15	

Data Download Date/Time: 11/01/2017 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 15:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: SEDA

DATE/TIME OF TEST: 10/26/2017 16:54:38

Diagnostics Check	OK	16:48
Air Blank	0.000	16:48
Control Test	0.081	16:49
Air Blank	0.000	16:49
Subject Sample # 1	0.157	16:50
Breath Volume	1.203	16:50
Air Blank	0.000	16:50
Air Blank	0.000	16:52
Subject Sample # 2	0.175	16:53
Breath Volume	1.628	16:53
Air Blank	0.000	16:53
Control Test	0.082	16:54
Air Blank	0.000	16:54
Diagnostics Check	OK	16:54

Data Download Date/Time: 11/01/2017 13:42

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KILMURRAY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: SCHEID

DATE/TIME OF TEST: 10/31/2017 21:30:08

Diagnostics Check	OK	21:23
Air Blank	0.000	21:23
Control Test	0.082	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.191	21:25
Breath Volume	1.445	21:25
Air Blank	0.000	21:26
Air Blank	0.000	21:27
Subject Sample # 2	0.195	21:28
Breath Volume	2.382	21:28
Air Blank	0.000	21:29
Control Test	0.082	21:29
Air Blank	0.000	21:30
Diagnostics Check	OK	21:30

Data Download Date/Time: 11/01/2017 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KERNS,CODY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 10/31/2017 23:30:05

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
Control Test	0.081	23:21	
Air Blank	0.000	23:22	
Subject Sample # 1	0.181	23:23	
Breath Volume	1.304	23:23	
Air Blank	0.000	23:23	
Air Blank	0.000	23:25	
Subject Sample # 2	0.179	23:28	Volume Not Met
Breath Volume	0.187	23:28	
Air Blank		23:29	Purge Fail
Air Blank	0.000	23:30	

Data Download Date/Time: 11/01/2017 13:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KERNS,CODY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: PENNINGTON

DATE/TIME OF TEST: 10/31/2017 23:39:55

Diagnostics Check	OK	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Subject Sample # 1	0.158	23:35
Breath Volume	1.496	23:35
Air Blank	0.000	23:35
Air Blank	0.000	23:37
Subject Sample # 2	0.174	23:38
Breath Volume	1.222	23:38
Air Blank	0.000	23:39
Control Test	0.082	23:39
Air Blank	0.000	23:39
Diagnostics Check	OK	23:39

Data Download Date/Time: 11/01/2017 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000869

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: KERNS,CODY
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 10/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/31/2017 23:51:21

Diagnostics Check		OK	23:44
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.167	23:46
Breath Volume		2.207	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.158	23:49
Breath Volume		2.128	23:49
Air Blank		0.000	23:50
Control Test		0.082	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 11/01/2017 13:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000874

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROIG,DAWN,M
BREATH TEST OPERATOR AGENCY: Miami Shores PD
ARRESTING OFFICER: ROIG,DAWN
ARRESTING OFFICER AGENCY: Miami Shores PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:16

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 10/31/2017 20:47:03

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:39	
Control Test	0.080	20:39	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:43	No Sample Provided
Breath Volume	0.000	20:43	
Air Blank	0.000	20:43	
Air Blank	0.000	20:45	
Subject Sample	# 2	20:45	Subject Test Refused
Breath Volume	0.000	20:45	
Air Blank	0.000	20:46	
Control Test	0.079	20:46	
Air Blank	0.000	20:46	
Diagnostics Check	OK	20:47	

Data Download Date/Time: 11/30/2017 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:11

SUBJECT'S LAST NAME: MEJIA COC

DATE/TIME OF TEST: 10/08/2017 20:50:24

Diagnostics Check		OK	20:38	
Air Blank		0.000	20:38	
Control Test		0.082	20:38	
Air Blank		0.000	20:39	
Subject Sample	# 1	0.183	20:42	Volume Not Met
Breath Volume		0.449	20:42	
Air Blank		0.000	20:43	
Air Blank		0.000	20:45	
Subject Sample	# 2	0.176	20:45	
Breath Volume		1.359	20:45	
Air Blank		0.000	20:46	
Air Blank		0.000	20:48	
Subject Sample	# 3	0.179	20:48	
Breath Volume		1.332	20:48	
Air Blank		0.000	20:49	
Control Test		0.083	20:49	
Air Blank		0.000	20:50	
Diagnostics Check		OK	20:50	

Data Download Date/Time: 10/28/2017 01:13

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 09/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MILIAN

DATE/TIME OF TEST: 10/26/2017 00:39:31

Diagnostics Check	OK	00:36	
Air Blank	0.000	00:36	
Control Test	0.081	00:36	
Air Blank	0.000	00:37	
Subject Sample	# 1	00:38	Subject Test Refused
Breath Volume	0.000	00:38	
Air Blank	0.000	00:38	
Control Test	0.082	00:38	
Air Blank	0.000	00:39	
Diagnostics Check	OK	00:39	

Data Download Date/Time: 10/28/2017 01:13

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: MARTINEZ HERNANDEZ

DATE/TIME OF TEST: 10/28/2017 22:22:16

Diagnostics Check	OK	22:13	
Air Blank	0.000	22:14	
Control Test	0.081	22:14	
Air Blank	0.000	22:14	
Subject Sample	# 1	22:18	No Sample Provided
Breath Volume	0.000	22:18	
Air Blank	0.000	22:18	
Air Blank	0.000	22:20	
Subject Sample	# 2	22:20	Subject Test Refused
Breath Volume	0.000	22:20	
Air Blank	0.000	22:21	
Control Test	0.082	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 11/29/2017 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000878

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,A
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: KISSNER,FREDERICK
ARRESTING OFFICER AGENCY: Biscayne Nat. Park
LAST AGENCY INSPECTION: 10/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1870563
DRY GAS EXPIRY DATE: 06/17/2018
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 10/29/2017 18:13:26

Diagnostics Check	OK	18:06
Air Blank	0.000	18:07
Control Test	0.081	18:07
Air Blank	0.000	18:08
Subject Sample # 1	0.195	18:08
Breath Volume	1.324	18:08
Air Blank	0.000	18:09
Air Blank	0.000	18:11
Subject Sample # 2	0.205	18:11
Breath Volume	1.625	18:11
Air Blank	0.000	18:12
Control Test	0.082	18:12
Air Blank	0.000	18:13
Diagnostics Check	OK	18:13

Data Download Date/Time: 11/29/2017 19:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAVAS,DAVID,E
BREATH TEST OPERATOR AGENCY: Bal Harbour Vil PD
ARRESTING OFFICER: NAVAS,DAVID
ARRESTING OFFICER AGENCY: Bal Harbour Vil PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: ZAKHARCHENKO

DATE/TIME OF TEST: 10/14/2017 02:19:20

Diagnostics Check	OK	02:15	
Air Blank	0.000	02:16	
Control Test	0.076	02:16	
Air Blank	0.000	02:17	
Subject Sample	# 1	02:17	Subject Test Refused
Breath Volume	0.000	02:17	
Air Blank	0.000	02:18	
Control Test	0.078	02:18	
Air Blank	0.000	02:19	
Diagnostics Check	OK	02:19	

Data Download Date/Time: 10/25/2017 11:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000880

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAESTRE,HANS,
BREATH TEST OPERATOR AGENCY: Aventura PD
ARRESTING OFFICER: MARTIN,JAMES
ARRESTING OFFICER AGENCY: Aventura PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: WIEGANDT

DATE/TIME OF TEST: 10/18/2017 22:32:17

Diagnostics Check	OK	22:23
Air Blank	0.000	22:24
Control Test	0.080	22:24
Air Blank	0.000	22:25
Subject Sample # 1	0.187	22:27
Breath Volume	1.738	22:27
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 2	0.184	22:30
Breath Volume	1.792	22:30
Air Blank	0.000	22:31
Control Test	0.075	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 10/25/2017 11:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000881

INSTRUMENT REGISTERED TO: MIAMI DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ROMAY

DATE/TIME OF TEST: 10/14/2017 21:06:55

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.082	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.250	21:02
Breath Volume		1.699	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:04
Subject Sample	# 2	0.254	21:05
Breath Volume		2.054	21:05
Air Blank		0.000	21:06
Control Test		0.079	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 10/25/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: MAJORS

DATE/TIME OF TEST: 10/04/2017 01:09:16

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.081	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:04	No Sample Provided
Breath Volume	0.000	01:04	
Air Blank	0.000	01:05	
Air Blank	0.000	01:07	
Subject Sample	# 2	01:07	Subject Test Refused
Breath Volume	0.000	01:07	
Air Blank	0.000	01:08	
Control Test	0.081	01:08	
Air Blank	0.000	01:09	
Diagnostics Check	OK	01:09	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO, YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: DRAGO BARRUETO

DATE/TIME OF TEST: 10/05/2017 16:26:24

Diagnostics Check	OK	16:19
Air Blank	0.000	16:20
Control Test	0.081	16:20
Air Blank	0.000	16:21
Subject Sample # 1	0.117	16:21
Breath Volume	1.558	16:21
Air Blank	0.000	16:22
Air Blank	0.000	16:23
Subject Sample # 2	0.121	16:24
Breath Volume	1.378	16:24
Air Blank	0.000	16:25
Control Test	0.080	16:25
Air Blank	0.000	16:26
Diagnostics Check	OK	16:26

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/06/2017 04:59:41

Diagnostics Check	OK	04:53
Air Blank	0.000	04:53
Control Test	0.081	04:54
Air Blank	0.000	04:54
Subject Sample # 1	0.157	04:55
Breath Volume	1.937	04:55
Air Blank	0.000	04:55
Air Blank	0.000	04:57
Subject Sample # 2	0.152	04:58
Breath Volume	1.156	04:58
Air Blank	0.000	04:58
Control Test	0.081	04:59
Air Blank	0.000	04:59
Diagnostics Check	OK	04:59

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: GARZA

DATE/TIME OF TEST: 10/06/2017 23:54:04

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.083	23:50	
Air Blank	0.000	23:51	
Subject Sample	# 1	23:52	Subject Test Refused
Breath Volume	0.000	23:52	
Air Blank	0.000	23:53	
Control Test	0.081	23:53	
Air Blank	0.000	23:53	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORVA,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: LEONTARITIS

DATE/TIME OF TEST: 10/07/2017 05:09:12

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.081	05:02
Air Blank	0.000	05:02
Subject Sample # 1	0.221	05:04
Breath Volume	1.738	05:04
Air Blank	0.000	05:05
Air Blank	0.000	05:06
Subject Sample # 2	0.229	05:07
Breath Volume	1.988	05:07
Air Blank	0.000	05:08
Control Test	0.078	05:08
Air Blank	0.000	05:09
Diagnostics Check	OK	05:09

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PADILLA FLORES

DATE/TIME OF TEST: 10/07/2017 20:28:28

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.082	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.229	20:23
Breath Volume	1.101	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.227	20:26
Breath Volume	1.507	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: VEGA

DATE/TIME OF TEST: 10/07/2017 23:55:15

Diagnostics Check	OK	23:48
Air Blank	0.000	23:49
Control Test	0.080	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.145	23:50
Breath Volume	1.523	23:50
Air Blank	0.000	23:51
Air Blank	0.000	23:52
Subject Sample # 2	0.142	23:53
Breath Volume	1.582	23:53
Air Blank	0.000	23:54
Control Test	0.078	23:54
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 10/08/2017 01:04:08

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.122	00:59
Breath Volume	3.128	00:59
Air Blank	0.000	00:59
Air Blank	0.000	01:01
Subject Sample # 2	0.115	01:02
Breath Volume	2.355	01:02
Air Blank	0.000	01:03
Control Test	0.080	01:03
Air Blank	0.000	01:04
Diagnostics Check	OK	01:04

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: MELARA PINEDA

DATE/TIME OF TEST: 10/08/2017 01:14:55

Diagnostics Check		OK	01:08
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.279	01:10
Breath Volume		1.796	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.270	01:13
Breath Volume		1.250	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PACHECO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PACHECO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: VELAZQUEZ

DATE/TIME OF TEST: 10/08/2017 03:09:15

Diagnostics Check		OK	02:57	
Air Blank		0.000	02:57	
Control Test		0.082	02:57	
Air Blank		0.000	02:58	
Subject Sample	# 1	0.129	03:01	Volume Not Met
Breath Volume		0.675	03:01	
Air Blank		0.000	03:02	
Air Blank		0.000	03:03	
Subject Sample	# 2	0.162	03:04	
Breath Volume		2.910	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 3	0.149	03:07	
Breath Volume		1.335	03:07	
Air Blank		0.000	03:08	
Control Test		0.079	03:08	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO, YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MOSCOSO

DATE/TIME OF TEST: 10/08/2017 22:33:56

Diagnostics Check		OK	22:21	
Air Blank		0.000	22:22	
Control Test		0.081	22:22	
Air Blank		0.000	22:23	
Subject Sample	# 1	0.172	22:26	Volume Not Met
Breath Volume		0.039	22:26	
Air Blank		0.000	22:27	
Air Blank		0.000	22:28	
Subject Sample	# 2	0.166	22:31	Volume Not Met
Breath Volume		0.023	22:31	
Air Blank		0.000	22:32	
Control Test		0.080	22:33	
Air Blank		0.000	22:33	
Diagnostics Check		OK	22:33	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO, YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO, YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: MOSCOSO

DATE/TIME OF TEST: 10/08/2017 22:44:55

Diagnostics Check		OK	22:36	
Air Blank		0.000	22:36	
Control Test		0.078	22:37	
Air Blank		0.000	22:37	
Subject Sample	# 1	0.219	22:40	Volume Not Met
Breath Volume		0.121	22:40	
Air Blank		0.000	22:41	
Air Blank		0.000	22:43	
Subject Sample	# 2		22:43	Subject Test Refused
Breath Volume		0.000	22:43	
Air Blank		0.000	22:44	
Control Test		0.077	22:44	
Air Blank		0.000	22:44	
Diagnostics Check		OK	22:44	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/09/2017 04:14:00

Diagnostics Check	OK	04:11	
Air Blank	0.000	04:11	
Control Test	0.081	04:11	
Air Blank	0.000	04:12	
Subject Sample	# 1	04:12	Subject Test Refused
Breath Volume	0.000	04:12	
Air Blank	0.000	04:13	
Control Test	0.080	04:13	
Air Blank	0.000	04:13	
Diagnostics Check	OK	04:14	

Data Download Date/Time: 10/26/2017 11:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MEILAN

DATE/TIME OF TEST: 10/12/2017 02:57:52

Diagnostics Check	OK	02:55	
Air Blank	0.000	02:56	
Control Test	0.081	02:56	
Air Blank		02:57	Purge Fail
Air Blank	0.000	02:57	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MEILAN

DATE/TIME OF TEST: 10/12/2017 03:05:55

Diagnostics Check	OK	02:59
Air Blank	0.000	03:00
Control Test	0.081	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.000	03:01
Breath Volume	2.105	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:04
Subject Sample # 2	0.000	03:04
Breath Volume	2.398	03:04
Air Blank	0.000	03:05
Control Test	0.080	03:05
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: ST FORT

DATE/TIME OF TEST: 10/12/2017 22:22:24

Diagnostics Check	OK	22:19	
Air Blank	0.000	22:19	
Control Test	0.082	22:19	
Air Blank	0.000	22:20	
Subject Sample	# 1	22:21	Subject Test Refused
Breath Volume	0.000	22:21	
Air Blank	0.000	22:21	
Control Test	0.081	22:21	
Air Blank	0.000	22:22	
Diagnostics Check	OK	22:22	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BARDELAS,MAYLEE
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: PADRON

DATE/TIME OF TEST: 10/13/2017 04:34:39

Diagnostics Check	OK	04:30	
Air Blank	0.000	04:30	
Control Test	0.081	04:30	
Air Blank	0.000	04:31	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: PEREZ SANTAMARIA

DATE/TIME OF TEST: 10/14/2017 02:01:43

Diagnostics Check	OK	01:58	
Air Blank	0.000	01:58	
Control Test	0.081	01:59	
Air Blank	0.000	01:59	
Subject Sample	# 1	02:00	Subject Test Refused
Breath Volume	0.000	02:00	
Air Blank	0.000	02:00	
Control Test	0.079	02:01	
Air Blank	0.000	02:01	
Diagnostics Check	OK	02:01	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:33

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/14/2017 05:00:49

Diagnostics Check	OK	04:54
Air Blank	0.000	04:54
Control Test	0.082	04:54
Air Blank	0.000	04:55
Subject Sample # 1	0.150	04:56
Breath Volume	2.960	04:56
Air Blank	0.000	04:56
Air Blank	0.000	04:58
Subject Sample # 2	0.143	04:59
Breath Volume	2.113	04:59
Air Blank	0.000	04:59
Control Test	0.079	05:00
Air Blank	0.000	05:00
Diagnostics Check	OK	05:00

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,DALE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: BORREGO

DATE/TIME OF TEST: 10/15/2017 03:18:21

Diagnostics Check	OK	03:08	
Air Blank	0.000	03:09	
Control Test	0.080	03:09	
Air Blank	0.000	03:10	
Subject Sample	# 1	03:13	No Sample Provided
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Air Blank	0.000	03:15	
Subject Sample	# 2	03:16	Subject Test Refused
Breath Volume	0.000	03:16	
Air Blank	0.000	03:17	
Control Test	0.081	03:17	
Air Blank	0.000	03:18	
Diagnostics Check	OK	03:18	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARMINES, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: MARTIN, NATALIE
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: CARMINES

DATE/TIME OF TEST: 10/16/2017 03:09:29

Diagnostics Check	OK	02:59
Air Blank	0.000	02:59
Control Test	0.080	03:00
Air Blank	0.000	03:00
Subject Sample # 1	0.215	03:01
Breath Volume	2.828	03:01
Air Blank	0.000	03:02
Air Blank	0.000	03:03
Subject Sample # 2	0.173	03:04
Breath Volume	3.625	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:07
Subject Sample # 3	0.210	03:07
Breath Volume	4.500	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/17/2017 00:49:27

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.082	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:48	Subject Test Refused
Breath Volume	0.000	00:48	
Air Blank	0.000	00:48	
Control Test	0.080	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: PEREZ,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: PUPO

DATE/TIME OF TEST: 10/17/2017 01:53:18

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.081	01:50	
Air Blank	0.000	01:50	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:52	
Control Test	0.080	01:52	
Air Blank	0.000	01:53	
Diagnostics Check	OK	01:53	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,R,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MIRABAL,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/17/2017 03:02:54

Diagnostics Check	OK	02:56
Air Blank	0.000	02:57
Control Test	0.082	02:57
Air Blank	0.000	02:57
Subject Sample # 1	0.000	02:58
Breath Volume	1.445	02:58
Air Blank	0.000	02:59
Air Blank	0.000	03:01
Subject Sample # 2	0.000	03:01
Breath Volume	1.246	03:01
Air Blank	0.000	03:01
Control Test	0.080	03:02
Air Blank	0.000	03:02
Diagnostics Check	OK	03:02

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: KINCHENS

DATE/TIME OF TEST: 10/19/2017 20:16:21

Diagnostics Check	OK	20:09
Air Blank	0.000	20:10
Control Test	0.081	20:10
Air Blank	0.000	20:10
Subject Sample # 1	0.174	20:11
Breath Volume	1.496	20:11
Air Blank	0.000	20:12
Air Blank	0.000	20:14
Subject Sample # 2	0.173	20:14
Breath Volume	2.261	20:14
Air Blank	0.000	20:15
Control Test	0.079	20:15
Air Blank	0.000	20:16
Diagnostics Check	OK	20:16

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: HERNANDEZ BOFFILL

DATE/TIME OF TEST: 10/20/2017 04:23:18

Diagnostics Check	OK	04:16
Air Blank	0.000	04:17
Control Test	0.081	04:17
Air Blank	0.000	04:18
Subject Sample # 1	0.131	04:18
Breath Volume	1.867	04:18
Air Blank	0.000	04:19
Air Blank	0.000	04:21
Subject Sample # 2	0.125	04:21
Breath Volume	1.570	04:21
Air Blank	0.000	04:22
Control Test	0.075	04:22
Air Blank	0.000	04:23
Diagnostics Check	OK	04:23

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZORRILLA,J,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ZORRILLA,J
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 10/20/2017 13:04:27

Diagnostics Check	OK	12:53	
Air Blank	0.000	12:53	
Control Test	0.081	12:53	
Air Blank	0.000	12:54	
Subject Sample	# 1	12:57	No Sample Provided
Breath Volume	0.000	12:57	
Air Blank	0.000	12:57	
Air Blank	0.000	12:59	
Subject Sample	# 2	13:03	No Sample Provided
Breath Volume	0.000	13:03	
Air Blank	0.000	13:03	
Control Test	0.081	13:03	
Air Blank	0.000	13:04	
Diagnostics Check	OK	13:04	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/20/2017 19:30:30

Diagnostics Check	OK	19:24
Air Blank	0.000	19:24
Control Test	0.080	19:24
Air Blank	0.000	19:25
Subject Sample # 1	0.236	19:25
Breath Volume	1.589	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 2	0.246	19:28
Breath Volume	1.699	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:46

SUBJECT'S LAST NAME: OZORES

DATE/TIME OF TEST: 10/21/2017 01:14:36

Diagnostics Check	OK	01:11	
Air Blank	0.000	01:12	
Control Test	0.080	01:12	
Air Blank	0.000	01:13	
Subject Sample	# 1	01:13	Subject Test Refused
Breath Volume	0.000	01:13	
Air Blank	0.000	01:13	
Control Test	0.081	01:14	
Air Blank	0.000	01:14	
Diagnostics Check	OK	01:14	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ABDEL-AZIZ

DATE/TIME OF TEST: 10/21/2017 02:45:44

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:34	
Control Test		0.081	02:34	
Air Blank		0.000	02:35	
Subject Sample	# 1	0.123	02:38	Volume Not Met
Breath Volume		0.714	02:38	
Air Blank		0.000	02:39	
Air Blank		0.000	02:40	
Subject Sample	# 2	0.097	02:43	Volume Not Met
Breath Volume		0.335	02:43	
Air Blank		0.000	02:44	
Control Test		0.078	02:45	
Air Blank		0.000	02:45	
Diagnostics Check		OK	02:45	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ABDEL-AZIZ

DATE/TIME OF TEST: 10/21/2017 02:52:28

Diagnostics Check	OK	02:48	
Air Blank	0.000	02:48	
Control Test	0.076	02:48	
Air Blank	0.000	02:49	
Subject Sample	# 1	02:50	Subject Test Refused
Breath Volume	0.085	02:50	
Air Blank	0.000	02:51	
Control Test	0.078	02:51	
Air Blank	0.000	02:52	
Diagnostics Check	OK	02:52	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: GALLEGO SANCHEZ

DATE/TIME OF TEST: 10/21/2017 21:39:40

Diagnostics Check	OK	21:33
Air Blank	0.000	21:33
Control Test	0.081	21:34
Air Blank	0.000	21:34
Subject Sample # 1	0.153	21:35
Breath Volume	1.308	21:35
Air Blank	0.000	21:35
Air Blank	0.000	21:37
Subject Sample # 2	0.153	21:38
Breath Volume	1.136	21:38
Air Blank	0.000	21:38
Control Test	0.080	21:39
Air Blank	0.000	21:39
Diagnostics Check	OK	21:39

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUJILLO,GEORGE,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TRUJILLO,GEORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: GASTESI

DATE/TIME OF TEST: 10/21/2017 23:20:04

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.080	23:14
Air Blank	0.000	23:14
Subject Sample # 1	0.198	23:15
Breath Volume	1.371	23:15
Air Blank	0.000	23:16
Air Blank	0.000	23:17
Subject Sample # 2	0.201	23:18
Breath Volume	1.597	23:18
Air Blank	0.000	23:19
Control Test	0.078	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:20

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,F
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: KRISSEL

DATE/TIME OF TEST: 10/22/2017 01:06:40

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.080	01:06	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGOS,O,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MIRABAL,R
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: AUGUSTE

DATE/TIME OF TEST: 10/22/2017 02:47:29

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.081	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.259	02:42
Breath Volume		2.011	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.243	02:45
Breath Volume		1.800	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENNIS,RHETT,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ENNIS

DATE/TIME OF TEST: 10/22/2017 03:48:03

Diagnostics Check	OK	03:44	
Air Blank	0.000	03:44	
Control Test	0.080	03:45	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:46	Subject Test Refused
Breath Volume	0.000	03:46	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZORRILLA,J,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ZORRILLA,J
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:26

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/25/2017 12:54:53

Diagnostics Check	OK	12:47	
Air Blank	0.000	12:47	
Control Test	0.081	12:48	
Air Blank	0.000	12:48	
Subject Sample	# 1 0.203	12:48	
Breath Volume	2.402	12:48	
Air Blank	0.000	12:49	
Air Blank	0.000	12:51	
Subject Sample	# 2 0.236	12:51	
Breath Volume	2.140	12:51	
Air Blank	0.000	12:52	
Air Blank	0.000	12:54	
Subject Sample	# 3 0.212	12:54	
Breath Volume	1.929	12:54	
Air Blank		12:54	RFI Detect
Air Blank		12:54	RFI Detect

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: DIAZ,JAVIER
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 09/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: POVEDA

DATE/TIME OF TEST: 10/26/2017 01:33:47

Diagnostics Check	OK	01:30	
Air Blank	0.000	01:31	
Control Test	0.081	01:31	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:32	Subject Test Refused
Breath Volume	0.000	01:32	
Air Blank	0.000	01:32	
Control Test	0.081	01:33	
Air Blank	0.000	01:33	
Diagnostics Check	OK	01:33	

Data Download Date/Time: 10/26/2017 11:18

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAHA,KHALED,
BREATH TEST OPERATOR AGENCY: Coral Gables PD
ARRESTING OFFICER: NAYA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HARRIMAN

DATE/TIME OF TEST: 10/28/2017 01:28:36

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:26	
Control Test	0.082	01:26	
Air Blank	0.000	01:26	
Subject Sample	# 1	01:27	Subject Test Refused
Breath Volume	0.000	01:27	
Air Blank	0.000	01:27	
Control Test	0.081	01:27	
Air Blank	0.000	01:28	
Diagnostics Check	OK	01:28	

Data Download Date/Time: 11/10/2017 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORGES,SEAN,P
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BORGES,SEAN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: FIMIA MARTIN

DATE/TIME OF TEST: 10/28/2017 09:16:43

Diagnostics Check		OK	09:10
Air Blank		0.000	09:10
Control Test		0.081	09:11
Air Blank		0.000	09:11
Subject Sample	# 1	0.228	09:11
Breath Volume		2.003	09:11
Air Blank		0.000	09:12
Air Blank		0.000	09:14
Subject Sample	# 2	0.235	09:14
Breath Volume		2.910	09:14
Air Blank		0.000	09:15
Control Test		0.079	09:16
Air Blank		0.000	09:16
Diagnostics Check		OK	09:16

Data Download Date/Time: 11/10/2017 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: TAHA,KHALED
ARRESTING OFFICER AGENCY: Coral Gables PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: HERRERA

DATE/TIME OF TEST: 10/29/2017 03:37:11

Diagnostics Check	OK	03:30
Air Blank	0.000	03:31
Control Test	0.081	03:31
Air Blank	0.000	03:31
Subject Sample # 1	0.115	03:32
Breath Volume	1.390	03:32
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.123	03:35
Breath Volume	1.562	03:35
Air Blank	0.000	03:36
Control Test	0.079	03:36
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 11/10/2017 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: PIERI

DATE/TIME OF TEST: 10/29/2017 03:52:37

Diagnostics Check	OK	03:49	
Air Blank	0.000	03:50	
Control Test	0.080	03:50	
Air Blank	0.000	03:51	
Subject Sample	# 1	03:51	Subject Test Refused
Breath Volume	0.000	03:51	
Air Blank	0.000	03:51	
Control Test	0.080	03:52	
Air Blank	0.000	03:52	
Diagnostics Check	OK	03:52	

Data Download Date/Time: 11/10/2017 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:09

SUBJECT'S LAST NAME: ZAPATA

DATE/TIME OF TEST: 10/29/2017 07:37:52

Diagnostics Check	OK	07:29
Air Blank	0.000	07:29
Control Test	0.082	07:30
Air Blank	0.000	07:30
Subject Sample # 1	0.138	07:32
Breath Volume	2.074	07:32
Air Blank	0.000	07:33
Air Blank	0.000	07:34
Subject Sample # 2	0.131	07:36
Breath Volume	2.031	07:36
Air Blank	0.000	07:36
Control Test	0.081	07:37
Air Blank	0.000	07:37
Diagnostics Check	OK	07:37

Data Download Date/Time: 11/10/2017 10:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000883

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MONDRAGON

DATE/TIME OF TEST: 10/30/2017 05:47:26

Diagnostics Check		OK	05:40
Air Blank		0.000	05:41
Control Test		0.081	05:41
Air Blank		0.000	05:42
Subject Sample	# 1	0.224	05:42
Breath Volume		1.265	05:42
Air Blank		0.000	05:43
Air Blank		0.000	05:45
Subject Sample	# 2	0.219	05:45
Breath Volume		1.210	05:45
Air Blank		0.000	05:46
Control Test		0.081	05:46
Air Blank		0.000	05:47
Diagnostics Check		OK	05:47

Data Download Date/Time: 11/10/2017 10:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTOS,J,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LAWSON,G
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:29

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 10/18/2017 06:22:08

Diagnostics Check	OK	06:19	
Air Blank	0.000	06:19	
Control Test	0.080	06:20	
Air Blank	0.000	06:20	
Subject Sample	# 1	06:20	Subject Test Refused
Breath Volume	0.000	06:20	
Air Blank	0.000	06:21	
Control Test	0.080	06:21	
Air Blank	0.000	06:22	
Diagnostics Check	OK	06:22	

Data Download Date/Time: 11/21/2017 07:37

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000884

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VARGAS,D,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: VARGAS,D
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: FLORES

DATE/TIME OF TEST: 10/31/2017 02:16:53

Diagnostics Check	OK	02:10
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:11
Subject Sample # 1	0.155	02:12
Breath Volume	3.218	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:14
Subject Sample # 2	0.151	02:15
Breath Volume	3.300	02:15
Air Blank	0.000	02:15
Control Test	0.080	02:16
Air Blank	0.000	02:16
Diagnostics Check	OK	02:16

Data Download Date/Time: 11/21/2017 07:37

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MESA,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MESA,JONATHAN
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: CALONGE

DATE/TIME OF TEST: 10/13/2017 01:38:50

Diagnostics Check	OK	01:35	
Air Blank	0.000	01:36	
Control Test	0.079	01:36	
Air Blank	0.000	01:37	
Subject Sample	# 1	01:37	Subject Test Refused
Breath Volume	0.000	01:37	
Air Blank	0.000	01:37	
Control Test	0.079	01:38	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 11/21/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARDAN,SAVANNA,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: REEVE,Y
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: ARDAN

DATE/TIME OF TEST: 10/14/2017 09:41:50

Diagnostics Check	OK	09:37	
Air Blank	0.000	09:37	
Control Test	0.079	09:37	
Air Blank	0.000	09:38	
Subject Sample	# 1	09:40	Subject Test Refused
Breath Volume	0.000	09:40	
Air Blank	0.000	09:40	
Control Test	0.080	09:41	
Air Blank	0.000	09:41	
Diagnostics Check	OK	09:41	

Data Download Date/Time: 11/21/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVE, YAARA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: REEVE, YAARA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:10

SUBJECT'S LAST NAME: ARDAN

DATE/TIME OF TEST: 10/14/2017 09:57:41

Diagnostics Check	OK	09:45	
Air Blank	0.000	09:45	
Control Test	0.080	09:45	
Air Blank	0.000	09:46	
Subject Sample	# 1	09:49	No Sample Provided
Breath Volume	0.000	09:49	
Air Blank	0.000	09:49	
Air Blank	0.000	09:51	
Subject Sample	# 2	0.272	09:52
Breath Volume	1.437	09:52	
Air Blank	0.000	09:53	
Air Blank	0.000	09:55	
Subject Sample	# 3	0.278	09:55
Breath Volume	2.492	09:55	
Air Blank	0.000	09:56	
Control Test	0.079	09:57	
Air Blank	0.000	09:57	
Diagnostics Check	OK	09:57	

Data Download Date/Time: 11/21/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,CHRISTOPHER,E
BREATH TEST OPERATOR AGENCY: South Miami PD
ARRESTING OFFICER: JOHNSON,CHRISTOPHER
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: CISNEROS SICLE

DATE/TIME OF TEST: 10/25/2017 23:31:22

Diagnostics Check	OK	23:21	
Air Blank	0.000	23:22	
Control Test	0.081	23:22	
Air Blank	0.000	23:23	
Subject Sample	# 1	23:26	No Sample Provided
Breath Volume	0.000	23:26	
Air Blank	0.000	23:26	
Air Blank	0.000	23:28	
Subject Sample	# 2	23:30	Subject Test Refused
Breath Volume	0.000	23:30	
Air Blank	0.000	23:30	
Control Test	0.080	23:30	
Air Blank	0.000	23:31	
Diagnostics Check	OK	23:31	

Data Download Date/Time: 11/21/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTOS,J,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: LAWSON,G
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: LAFUENTE

DATE/TIME OF TEST: 10/27/2017 01:05:43

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.079	01:01	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:03	Subject Test Refused
Breath Volume	0.765	01:03	
Air Blank	0.000	01:04	
Control Test	0.081	01:05	
Air Blank	0.000	01:05	
Diagnostics Check	OK	01:05	

Data Download Date/Time: 11/21/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000885

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REEVE, YAARA,
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: REEVE, YAARA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 10/30/2017 11:09:25

Diagnostics Check	OK	11:06	
Air Blank	0.000	11:06	
Control Test	0.079	11:07	
Air Blank	0.000	11:07	
Subject Sample	# 1	11:08	Subject Test Refused
Breath Volume	0.000	11:08	
Air Blank	0.000	11:08	
Control Test	0.079	11:08	
Air Blank	0.000	11:09	
Diagnostics Check	OK	11:09	

Data Download Date/Time: 11/21/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BONDARENKO

DATE/TIME OF TEST: 10/12/2017 00:02:55

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.080	23:56
Air Blank	0.000	23:57
Subject Sample # 2	0.202	23:58
Breath Volume	3.261	23:58
Air Blank	0.000	23:58
Air Blank	0.000	00:00
Subject Sample # 1	0.205	00:01
Breath Volume	3.562	00:01
Air Blank	0.000	00:02
Control Test	0.077	00:02
Air Blank	0.000	00:02
Diagnostics Check	OK	00:02

Data Download Date/Time: 10/30/2017 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA, INGRID,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO, MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:23

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 10/14/2017 22:53:38

Diagnostics Check	OK	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:45
Air Blank	0.000	22:46
Subject Sample # 1	0.152	22:49
Breath Volume	2.085	22:49
Air Blank	0.000	22:49
Air Blank	0.000	22:51
Subject Sample # 2	0.159	22:52
Breath Volume	2.289	22:52
Air Blank	0.000	22:52
Control Test	0.077	22:53
Air Blank	0.000	22:53
Diagnostics Check	OK	22:53

Data Download Date/Time: 10/30/2017 13:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TANIS,MIKE,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KIRKUS,JOSHUA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: RUSIN

DATE/TIME OF TEST: 10/15/2017 06:22:00

Diagnostics Check		OK	06:15
Air Blank		0.000	06:15
Control Test		0.080	06:16
Air Blank		0.000	06:16
Subject Sample	# 1	0.204	06:17
Breath Volume		1.234	06:17
Air Blank		0.000	06:18
Air Blank		0.000	06:19
Subject Sample	# 2	0.200	06:20
Breath Volume		1.257	06:20
Air Blank		0.000	06:21
Control Test		0.080	06:21
Air Blank		0.000	06:21
Diagnostics Check		OK	06:22

Data Download Date/Time: 10/30/2017 13:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: AZEVEDO,MICHAEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: VASQUEZ EDUARDO

DATE/TIME OF TEST: 10/15/2017 19:30:08

Diagnostics Check	OK	19:21	
Air Blank	0.000	19:22	
Control Test	0.080	19:22	
Air Blank	0.000	19:22	
Subject Sample	# 1	19:25	No Sample Provided
Breath Volume	0.000	19:25	
Air Blank	0.000	19:26	
Air Blank	0.000	19:28	
Subject Sample	# 2	19:28	Subject Test Refused
Breath Volume	0.000	19:28	
Air Blank	0.000	19:29	
Control Test	0.080	19:29	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 10/30/2017 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AJO,MICHAEL,D
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: OLIVARES,ANDRES
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: COFFMAN

DATE/TIME OF TEST: 10/18/2017 03:14:49

Diagnostics Check	OK	03:11	
Air Blank	0.000	03:12	
Control Test	0.081	03:12	
Air Blank	0.000	03:13	
Subject Sample	# 1	03:13	Subject Test Refused
Breath Volume	0.000	03:13	
Air Blank	0.000	03:13	
Control Test	0.080	03:14	
Air Blank	0.000	03:14	
Diagnostics Check	OK	03:14	

Data Download Date/Time: 10/30/2017 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DOSTALY,STEVEN
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 10/22/2017 01:22:59

Diagnostics Check		OK	01:09	
Air Blank		0.000	01:09	
Control Test		0.080	01:10	
Air Blank		0.000	01:10	
Subject Sample	# 1	0.300	01:13	
Breath Volume		1.363	01:13	
Air Blank		0.000	01:13	
Air Blank		0.000	01:15	
Subject Sample	# 2	0.284	01:18	Volume Not Met
Breath Volume		0.898	01:18	
Air Blank		0.000	01:19	
Air Blank		0.000	01:21	
Subject Sample	# 3		01:21	Subject Test Refused
Breath Volume		0.000	01:21	
Air Blank		0.000	01:22	
Control Test		0.077	01:22	
Air Blank		0.000	01:22	
Diagnostics Check		OK	01:22	

Data Download Date/Time: 10/30/2017 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KIRKUS,JOSHUA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: PHANOR

DATE/TIME OF TEST: 10/22/2017 04:39:06

Diagnostics Check	OK	04:31
Air Blank	0.000	04:31
Control Test	0.080	04:32
Air Blank	0.000	04:32
Subject Sample # 1	0.180	04:34
Breath Volume	1.578	04:34
Air Blank	0.000	04:34
Air Blank	0.000	04:36
Subject Sample # 2	0.172	04:37
Breath Volume	1.257	04:37
Air Blank	0.000	04:38
Control Test	0.078	04:38
Air Blank	0.000	04:39
Diagnostics Check	OK	04:39

Data Download Date/Time: 10/30/2017 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CABRERA, INGRID,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: ACEVEDO, MIGUEL
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 10/22/2017 23:07:34

Diagnostics Check	OK	23:02	
Air Blank	0.000	23:03	
Control Test	0.080	23:03	
Air Blank	0.000	23:04	
Subject Sample	# 1	23:06	Subject Test Refused
Breath Volume	0.000	23:06	
Air Blank	0.000	23:06	
Control Test	0.079	23:06	
Air Blank	0.000	23:07	
Diagnostics Check	OK	23:07	

Data Download Date/Time: 10/30/2017 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIRKUS,JOSHUA,B
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: KIRKUS,JOSHUA
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:47

SUBJECT'S LAST NAME: BURRIS

DATE/TIME OF TEST: 10/27/2017 17:15:53

Diagnostics Check	OK	17:07	
Air Blank	0.000	17:07	
Control Test	0.078	17:08	
Air Blank	0.000	17:08	
Subject Sample	# 1	17:11	No Sample Provided
Breath Volume	0.000	17:11	
Air Blank	0.000	17:12	
Air Blank	0.000	17:14	
Subject Sample	# 2	17:14	Subject Test Refused
Breath Volume	0.000	17:14	
Air Blank	0.000	17:14	
Control Test	0.080	17:15	
Air Blank	0.000	17:15	
Diagnostics Check	OK	17:15	

Data Download Date/Time: 10/30/2017 13:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000886

INSTRUMENT REGISTERED TO: NORTH MIAMI BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOSTALY,STEVENS,
BREATH TEST OPERATOR AGENCY: North Miami Beach PD
ARRESTING OFFICER: DOSTALY,STEVENS
ARRESTING OFFICER AGENCY: North Miami Beach PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 10/29/2017 05:21:44

Diagnostics Check	OK	05:14
Air Blank	0.000	05:15
Control Test	0.080	05:15
Air Blank	0.000	05:15
Subject Sample # 1	0.173	05:16
Breath Volume	1.632	05:16
Air Blank	0.000	05:17
Air Blank	0.000	05:19
Subject Sample # 2	0.171	05:20
Breath Volume	2.273	05:20
Air Blank	0.000	05:20
Control Test	0.079	05:21
Air Blank	0.000	05:21
Diagnostics Check	OK	05:21

Data Download Date/Time: 10/30/2017 13:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: JACKSON,KARL,

BREATH TEST OPERATOR AGENCY: Bay County SO

ARRESTING OFFICER: JAMES,WILLIAM

ARRESTING OFFICER AGENCY: Naval Lab Pana City

LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 477627

DRY GAS EXPIRY DATE: 06/29/2019

VIOLATION CODE: DUI

ARREST TIME: 17:15

GENDER: Male

OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/22/2017 18:16:59

Diagnostics Check	OK	18:16	
Air Blank		18:16	RFI Detect
Air Blank	0.000	18:16	

Data Download Date/Time: 11/19/2017 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,KARL,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JAMES,WILLIAM
ARRESTING OFFICER AGENCY: Naval Lab Pana City
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/22/2017 18:21:00

Diagnostics Check	OK	18:20	
Air Blank		18:20	Improper Sample
Air Blank	0.000	18:21	

Data Download Date/Time: 11/19/2017 12:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,KARL,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JAMES,WILLIAM
ARRESTING OFFICER AGENCY: Naval Lab Pana City
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 17:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:43

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 10/22/2017 18:29:10

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.079	18:23
Air Blank	0.000	18:24
Subject Sample # 1	0.120	18:24
Breath Volume	2.750	18:24
Air Blank	0.000	18:25
Air Blank	0.000	18:27
Subject Sample # 2	0.119	18:27
Breath Volume	5.289	18:27
Air Blank	0.000	18:28
Control Test	0.078	18:28
Air Blank	0.000	18:29
Diagnostics Check	OK	18:29

Data Download Date/Time: 11/19/2017 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000892

INSTRUMENT REGISTERED TO: FWC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACKSON,KARL,
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: JAMES,WILLIAM
ARRESTING OFFICER AGENCY: Naval Lab Pana City
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 17:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/22/2017 18:40:06

Diagnostics Check	OK	18:34
Air Blank	0.000	18:34
Control Test	0.079	18:34
Air Blank	0.000	18:35
Subject Sample # 1	0.208	18:35
Breath Volume	2.867	18:35
Air Blank	0.000	18:36
Air Blank	0.000	18:38
Subject Sample # 2	0.211	18:38
Breath Volume	2.812	18:38
Air Blank	0.000	18:39
Control Test	0.079	18:39
Air Blank	0.000	18:40
Diagnostics Check	OK	18:40

Data Download Date/Time: 11/19/2017 12:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000906

INSTRUMENT REGISTERED TO: FISH & WILDLIFE CC
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:23

SUBJECT'S LAST NAME: LUIZAGA

DATE/TIME OF TEST: 10/29/2017 20:56:02

Diagnostics Check	OK	20:53	
Air Blank	0.000	20:53	
Control Test	0.081	20:53	
Air Blank	0.000	20:54	
Subject Sample	# 1	20:54	Subject Test Refused
Breath Volume	0.000	20:54	
Air Blank	0.000	20:55	
Control Test	0.082	20:55	
Air Blank	0.000	20:55	
Diagnostics Check	OK	20:56	

Data Download Date/Time: 11/14/2017 11:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARGO,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MIESSAU

DATE/TIME OF TEST: 10/03/2017 02:49:25

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.079	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.106	02:44
Breath Volume	2.058	02:44
Air Blank	0.000	02:45
Air Blank	0.000	02:47
Subject Sample # 2	0.122	02:47
Breath Volume	1.820	02:47
Air Blank	0.000	02:48
Control Test	0.078	02:48
Air Blank	0.000	02:49
Diagnostics Check	OK	02:49

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,A,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOTELLO,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: BRYL

DATE/TIME OF TEST: 10/04/2017 02:14:43

Diagnostics Check	OK	02:08
Air Blank	0.000	02:09
Control Test	0.079	02:09
Air Blank	0.000	02:09
Subject Sample # 1	0.000	02:10
Breath Volume	2.695	02:10
Air Blank	0.000	02:10
Air Blank	0.000	02:12
Subject Sample # 2	0.000	02:13
Breath Volume	2.179	02:13
Air Blank	0.000	02:13
Control Test	0.079	02:14
Air Blank	0.000	02:14
Diagnostics Check	OK	02:14

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,CARL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: BUCKNER

DATE/TIME OF TEST: 10/05/2017 15:02:48

Diagnostics Check	OK	14:56	
Air Blank	0.000	14:56	
Control Test	0.080	14:57	
Air Blank	0.000	14:57	
Subject Sample	# 1 0.365	14:58	
Breath Volume	1.109	14:58	
Air Blank	0.000	14:59	
Air Blank	0.000	15:00	
Subject Sample	# 2	15:01	Slope Not Met
Breath Volume	1.234	15:01	
Air Blank	0.000	15:01	
Control Test	0.079	15:02	
Air Blank	0.000	15:02	
Diagnostics Check	OK	15:02	

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,CARL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: BUCKNER

DATE/TIME OF TEST: 10/05/2017 15:11:53

Diagnostics Check	OK	15:05	
Air Blank	0.000	15:06	
Control Test	0.078	15:06	
Air Blank	0.000	15:06	
Subject Sample	# 1 0.387	15:07	
Breath Volume	1.972	15:07	
Air Blank	0.000	15:08	
Air Blank	0.000	15:09	
Subject Sample	# 2	15:10	Slope Not Met
Breath Volume	0.839	15:10	
Air Blank	0.000	15:10	
Control Test	0.078	15:11	
Air Blank	0.000	15:11	
Diagnostics Check	OK	15:11	

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARGO,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 10/06/2017 15:44:36

Diagnostics Check	OK	15:36
Air Blank	0.000	15:36
Control Test	0.079	15:36
Air Blank	0.000	15:37
Subject Sample # 1	0.043	15:39
Breath Volume	1.898	15:39
Air Blank	0.000	15:39
Air Blank	0.000	15:41
Subject Sample # 2	0.045	15:43
Breath Volume	1.804	15:43
Air Blank	0.000	15:43
Control Test	0.080	15:44
Air Blank	0.000	15:44
Diagnostics Check	OK	15:44

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,A,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOTELLO,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LOZADA-FLORES

DATE/TIME OF TEST: 10/07/2017 02:24:29

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.078	02:16
Air Blank	0.000	02:17
Subject Sample # 1	0.173	02:20
Breath Volume	2.843	02:20
Air Blank	0.000	02:20
Air Blank	0.000	02:22
Subject Sample # 2	0.177	02:22
Breath Volume	3.070	02:22
Air Blank	0.000	02:23
Control Test	0.076	02:23
Air Blank	0.000	02:24
Diagnostics Check	OK	02:24

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,A,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOTELLO,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/07/2017 07:10:26

Diagnostics Check	OK	07:04
Air Blank	0.000	07:04
Control Test	0.079	07:05
Air Blank	0.000	07:05
Subject Sample # 1	0.083	07:05
Breath Volume	2.054	07:05
Air Blank	0.000	07:06
Air Blank	0.000	07:08
Subject Sample # 2	0.081	07:08
Breath Volume	3.085	07:08
Air Blank	0.000	07:09
Control Test	0.078	07:09
Air Blank	0.000	07:10
Diagnostics Check	OK	07:10

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: JASLANEK

DATE/TIME OF TEST: 10/08/2017 21:10:14

Diagnostics Check	OK	21:04
Air Blank	0.000	21:04
Control Test	0.079	21:04
Air Blank	0.000	21:05
Subject Sample # 1	0.156	21:05
Breath Volume	1.367	21:05
Air Blank	0.000	21:06
Air Blank	0.000	21:08
Subject Sample # 2	0.153	21:08
Breath Volume	1.199	21:08
Air Blank	0.000	21:09
Control Test	0.078	21:09
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,BILLY,S
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: SCHILKE,CARL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: MULHOLLAND

DATE/TIME OF TEST: 10/10/2017 20:28:07

Diagnostics Check	OK	20:21
Air Blank	0.000	20:22
Control Test	0.080	20:22
Air Blank	0.000	20:23
Subject Sample # 1	0.152	20:23
Breath Volume	2.816	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.160	20:26
Breath Volume	2.957	20:26
Air Blank	0.000	20:27
Control Test	0.079	20:27
Air Blank	0.000	20:28
Diagnostics Check	OK	20:28

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HALL,CHRISTINE,K
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: HALL,CHRISTINE
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:58

SUBJECT'S LAST NAME: MORAN

DATE/TIME OF TEST: 10/12/2017 18:30:05

Diagnostics Check	OK	18:23
Air Blank	0.000	18:23
Control Test	0.079	18:24
Air Blank	0.000	18:24
Subject Sample # 1	0.202	18:25
Breath Volume	1.675	18:25
Air Blank	0.000	18:26
Air Blank	0.000	18:27
Subject Sample # 2	0.202	18:28
Breath Volume	2.417	18:28
Air Blank	0.000	18:29
Control Test	0.079	18:29
Air Blank	0.000	18:29
Diagnostics Check	OK	18:30

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MALDONADO,RYAN,A
BREATH TEST OPERATOR AGENCY: Fort Myers PD
ARRESTING OFFICER: FEARS,KAREEM
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: ANDREWS

DATE/TIME OF TEST: 10/13/2017 22:32:21

Diagnostics Check	OK	22:25
Air Blank	0.000	22:26
Control Test	0.079	22:26
Air Blank	0.000	22:26
Subject Sample # 1	0.215	22:27
Breath Volume	1.722	22:27
Air Blank	0.000	22:28
Air Blank	0.000	22:30
Subject Sample # 2	0.217	22:30
Breath Volume	1.164	22:30
Air Blank	0.000	22:31
Control Test	0.079	22:31
Air Blank	0.000	22:32
Diagnostics Check	OK	22:32

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLEURANT,RODSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BROWN,EARTHEN
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: PETTIT

DATE/TIME OF TEST: 10/15/2017 02:53:18

Diagnostics Check	OK	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.109	02:48
Breath Volume	3.824	02:48
Air Blank	0.000	02:49
Air Blank	0.000	02:51
Subject Sample # 2	0.104	02:51
Breath Volume	2.585	02:51
Air Blank	0.000	02:52
Control Test	0.079	02:52
Air Blank	0.000	02:53
Diagnostics Check	OK	02:53

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLEURANT,RODSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HAMILTON-REYES,CORTNEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 10/15/2017 03:29:43

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.078	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.167	03:22
Breath Volume	1.765	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.159	03:28
Breath Volume	2.390	03:28
Air Blank	0.000	03:28
Control Test	0.079	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLEURANT,RODSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: PERRY,MIKE
ARRESTING OFFICER AGENCY: Fort Myers PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: ALMEIDA LLERENA

DATE/TIME OF TEST: 10/15/2017 04:56:11

Diagnostics Check	OK	04:51	
Air Blank	0.000	04:52	
Control Test	0.079	04:52	
Air Blank	0.000	04:52	
Subject Sample	# 1	04:55	Sequence Aborted
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	

Data Download Date/Time: 10/18/2017 11:01

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN, JEREMY, L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: GARRETT, R
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 10/15/2017 07:41:01

Diagnostics Check	OK	07:34
Air Blank	0.000	07:35
Control Test	0.079	07:35
Air Blank	0.000	07:36
Subject Sample # 1	0.122	07:36
Breath Volume	1.296	07:36
Air Blank	0.000	07:37
Air Blank	0.000	07:38
Subject Sample # 2	0.111	07:39
Breath Volume	1.265	07:39
Air Blank	0.000	07:40
Control Test	0.078	07:40
Air Blank	0.000	07:40
Diagnostics Check	OK	07:41

Data Download Date/Time: 10/18/2017 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHERWOOD,JOHN,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: SHERWOOD,JOHN
ARRESTING OFFICER AGENCY: Sanibel PD
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ORNELAS HERNANDEZ

DATE/TIME OF TEST: 10/16/2017 01:47:18

Diagnostics Check	OK	01:40
Air Blank	0.000	01:41
Control Test	0.079	01:41
Air Blank	0.000	01:42
Subject Sample # 1	0.095	01:42
Breath Volume	1.898	01:42
Air Blank	0.000	01:43
Air Blank	0.000	01:45
Subject Sample # 2	0.089	01:45
Breath Volume	2.277	01:45
Air Blank	0.000	01:46
Control Test	0.079	01:46
Air Blank	0.000	01:47
Diagnostics Check	OK	01:47

Data Download Date/Time: 10/18/2017 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHERWOOD,JOHN,M
BREATH TEST OPERATOR AGENCY: Sanibel PD
ARRESTING OFFICER: HUAPILLA,MIGUEL
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: SERGROVIES

DATE/TIME OF TEST: 10/16/2017 02:21:34

Diagnostics Check	OK	02:15
Air Blank	0.000	02:16
Control Test	0.079	02:16
Air Blank	0.000	02:16
Subject Sample # 1	0.000	02:17
Breath Volume	3.828	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.000	02:20
Breath Volume	1.371	02:20
Air Blank	0.000	02:20
Control Test	0.079	02:20
Air Blank	0.000	02:21
Diagnostics Check	OK	02:21

Data Download Date/Time: 10/18/2017 11:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: EUBANKS,AARON
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: NOWALK

DATE/TIME OF TEST: 10/19/2017 11:46:11

Diagnostics Check	OK	11:40
Air Blank	0.000	11:40
Control Test	0.080	11:40
Air Blank	0.000	11:41
Subject Sample # 1	0.000	11:41
Breath Volume	1.175	11:41
Air Blank	0.000	11:42
Air Blank	0.000	11:44
Subject Sample # 2	0.000	11:44
Breath Volume	1.445	11:44
Air Blank	0.000	11:45
Control Test	0.079	11:45
Air Blank	0.000	11:46
Diagnostics Check	OK	11:46

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: LINDHOLM, J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: GOSKI

DATE/TIME OF TEST: 10/28/2017 02:52:04

Diagnostics Check	OK	02:45
Air Blank	0.000	02:45
Control Test	0.080	02:45
Air Blank	0.000	02:46
Subject Sample # 1	0.132	02:47
Breath Volume	1.730	02:47
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.143	02:50
Breath Volume	1.890	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:52

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR, J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 10/28/2017 03:57:04

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.079	03:51
Air Blank	0.000	03:51
Subject Sample # 1	0.147	03:52
Breath Volume	1.953	03:52
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.147	03:55
Breath Volume	1.949	03:55
Air Blank	0.000	03:56
Control Test	0.079	03:56
Air Blank	0.000	03:56
Diagnostics Check	OK	03:57

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLEURANT,RODSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLEURANT,RODSON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:51

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 10/28/2017 08:34:13

Diagnostics Check	OK	08:28
Air Blank	0.000	08:28
Control Test	0.079	08:28
Air Blank	0.000	08:29
Subject Sample # 1	0.049	08:29
Breath Volume	2.316	08:29
Air Blank	0.000	08:30
Air Blank	0.000	08:32
Subject Sample # 2	0.049	08:32
Breath Volume	2.242	08:32
Air Blank	0.000	08:33
Control Test	0.080	08:33
Air Blank	0.000	08:34
Diagnostics Check	OK	08:34

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JOSIAH,R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BROWN,E
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: SLINGERLAND

DATE/TIME OF TEST: 10/28/2017 23:30:42

Diagnostics Check	OK	23:24
Air Blank	0.000	23:24
Control Test	0.079	23:25
Air Blank	0.000	23:25
Subject Sample # 1	0.165	23:26
Breath Volume	2.347	23:26
Air Blank	0.000	23:26
Air Blank	0.000	23:28
Subject Sample # 2	0.170	23:29
Breath Volume	2.718	23:29
Air Blank	0.000	23:29
Control Test	0.079	23:30
Air Blank	0.000	23:30
Diagnostics Check	OK	23:30

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: BAIR, J
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: EARNSHAW JR

DATE/TIME OF TEST: 10/29/2017 03:04:36

Diagnostics Check	OK	02:58
Air Blank	0.000	02:58
Control Test	0.079	02:58
Air Blank	0.000	02:59
Subject Sample # 1	0.168	02:59
Breath Volume	2.109	02:59
Air Blank	0.000	03:00
Air Blank	0.000	03:02
Subject Sample # 2	0.158	03:03
Breath Volume	2.429	03:03
Air Blank	0.000	03:03
Control Test	0.078	03:04
Air Blank	0.000	03:04
Diagnostics Check	OK	03:04

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDIN,JEREMY,L
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: MANN,KOBY
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:14

SUBJECT'S LAST NAME: MYERS

DATE/TIME OF TEST: 10/29/2017 08:58:10

Diagnostics Check	OK	08:52
Air Blank	0.000	08:52
Control Test	0.080	08:53
Air Blank	0.000	08:53
Subject Sample # 1	0.073	08:53
Breath Volume	3.242	08:53
Air Blank	0.000	08:54
Air Blank	0.000	08:56
Subject Sample # 2	0.070	08:56
Breath Volume	2.917	08:56
Air Blank	0.000	08:57
Control Test	0.079	08:57
Air Blank	0.000	08:58
Diagnostics Check	OK	08:58

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH, JOSIAH, R
BREATH TEST OPERATOR AGENCY: Lee County SO
ARRESTING OFFICER: WALDROUP, K
ARRESTING OFFICER AGENCY: Lee County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: KIRSH

DATE/TIME OF TEST: 10/29/2017 23:46:45

Diagnostics Check	OK	23:36
Air Blank	0.000	23:37
Control Test	0.080	23:37
Air Blank	0.000	23:38
Subject Sample # 1	0.221	23:38
Breath Volume	2.460	23:38
Air Blank	0.000	23:39
Air Blank	0.000	23:41
Subject Sample # 2	0.199	23:41
Breath Volume	1.312	23:41
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 3	0.206	23:45
Breath Volume	2.007	23:45
Air Blank	0.000	23:45
Control Test	0.079	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000914

INSTRUMENT REGISTERED TO: FORT MYERS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARGO,TIMOTHY,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WARGO,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:57

SUBJECT'S LAST NAME: POWERS

DATE/TIME OF TEST: 10/31/2017 12:31:18

Diagnostics Check	OK	12:25
Air Blank	0.000	12:25
Control Test	0.080	12:25
Air Blank	0.000	12:26
Subject Sample # 1	0.000	12:26
Breath Volume	2.652	12:26
Air Blank	0.000	12:27
Air Blank	0.000	12:29
Subject Sample # 2	0.000	12:29
Breath Volume	2.511	12:29
Air Blank	0.000	12:30
Control Test	0.080	12:30
Air Blank	0.000	12:31
Diagnostics Check	OK	12:31

Data Download Date/Time: 11/08/2017 12:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON,LAUREN,F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: HANNON,LAUREN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 10/08/2017 02:37:08

Diagnostics Check	OK	02:31
Air Blank	0.000	02:31
Control Test	0.078	02:31
Air Blank	0.000	02:32
Subject Sample # 1	0.173	02:32
Breath Volume	1.312	02:32
Air Blank	0.000	02:33
Air Blank	0.000	02:35
Subject Sample # 2	0.164	02:35
Breath Volume	2.628	02:35
Air Blank	0.000	02:36
Control Test	0.078	02:36
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 10/25/2017 16:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARRITY, THOMAS, A
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: GARRITY, THOMAS
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: COFER

DATE/TIME OF TEST: 10/08/2017 03:17:04

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.079	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.093	03:12
Breath Volume	3.628	03:12
Air Blank	0.000	03:13
Air Blank	0.000	03:15
Subject Sample # 2	0.093	03:15
Breath Volume	2.863	03:15
Air Blank	0.000	03:16
Control Test	0.079	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:17

Data Download Date/Time: 10/25/2017 16:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HANNON, LAUREN, F
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: PAYNE, TAYLOR
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 09/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: INGOLD

DATE/TIME OF TEST: 10/13/2017 02:47:57

Diagnostics Check	OK	02:41
Air Blank	0.000	02:41
Control Test	0.079	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.153	02:43
Breath Volume	1.472	02:43
Air Blank	0.000	02:43
Air Blank	0.000	02:45
Subject Sample # 2	0.154	02:46
Breath Volume	1.777	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:47
Diagnostics Check	OK	02:47

Data Download Date/Time: 10/25/2017 16:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000940

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRATSCH,CASEY,R
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: REITES,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 09/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:44

SUBJECT'S LAST NAME: HARRELL

DATE/TIME OF TEST: 10/13/2017 23:11:06

Diagnostics Check	OK	23:04
Air Blank	0.000	23:04
Control Test	0.078	23:05
Air Blank	0.000	23:05
Subject Sample # 1	0.114	23:06
Breath Volume	1.437	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.107	23:09
Breath Volume	2.289	23:09
Air Blank	0.000	23:10
Control Test	0.077	23:10
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 10/25/2017 16:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000942

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRADLEY,KELLY,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,KURT
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/03/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 10/22/2017 00:04:29

Diagnostics Check	OK	23:55
Air Blank	0.000	23:56
Control Test	0.078	23:56
Air Blank	0.000	23:56
Subject Sample # 2	0.074	23:59
Breath Volume	1.460	23:59
Air Blank	0.000	00:00
Air Blank	0.000	00:02
Subject Sample # 1	0.074	00:03
Breath Volume	2.292	00:03
Air Blank	0.000	00:03
Control Test	0.077	00:03
Air Blank	0.000	00:04
Diagnostics Check	OK	00:04

Data Download Date/Time: 11/01/2017 12:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:28

SUBJECT'S LAST NAME: MALLOY

DATE/TIME OF TEST: 10/03/2017 19:57:51

Diagnostics Check	OK	19:51
Air Blank	0.000	19:52
Control Test	0.079	19:52
Air Blank	0.000	19:53
Subject Sample # 1	0.000	19:53
Breath Volume	1.437	19:53
Air Blank	0.000	19:54
Air Blank	0.000	19:56
Subject Sample # 2	0.000	19:56
Breath Volume	1.863	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:57

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MERTILUS,FREDELYNE,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:56

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 10/03/2017 20:27:09

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.078	20:19
Air Blank	0.000	20:19
Subject Sample # 1	0.177	20:22
Breath Volume	3.054	20:22
Air Blank	0.000	20:23
Air Blank	0.000	20:24
Subject Sample # 2	0.178	20:25
Breath Volume	2.699	20:25
Air Blank	0.000	20:26
Control Test	0.076	20:26
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: FORD,L
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: RAINONE

DATE/TIME OF TEST: 10/05/2017 01:55:55

Diagnostics Check		OK	01:42	
Air Blank		0.000	01:43	
Control Test		0.080	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1	0.102	01:47	Volume Not Met
Breath Volume		0.429	01:47	
Air Blank		0.000	01:47	
Air Blank		0.000	01:49	
Subject Sample	# 2	0.116	01:50	
Breath Volume		1.652	01:50	
Air Blank		0.000	01:50	
Air Blank		0.000	01:52	
Subject Sample	# 3	0.116	01:54	
Breath Volume		1.484	01:54	
Air Blank		0.000	01:55	
Control Test		0.079	01:55	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLINE,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SMITH,L
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 10/05/2017 02:06:29

Diagnostics Check	OK	01:59
Air Blank	0.000	02:00
Control Test	0.079	02:00
Air Blank	0.000	02:01
Subject Sample # 1	0.105	02:02
Breath Volume	1.613	02:02
Air Blank	0.000	02:02
Air Blank	0.000	02:04
Subject Sample # 2	0.110	02:04
Breath Volume	1.832	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:06
Diagnostics Check	OK	02:06

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,COLTON,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COMBS,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: TABARES

DATE/TIME OF TEST: 10/05/2017 02:52:52

Diagnostics Check	OK	02:46
Air Blank	0.000	02:46
Control Test	0.079	02:47
Air Blank	0.000	02:47
Subject Sample # 1	0.000	02:48
Breath Volume	1.320	02:48
Air Blank	0.000	02:48
Air Blank	0.000	02:50
Subject Sample # 2	0.000	02:51
Breath Volume	1.707	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:52
Air Blank	0.000	02:52
Diagnostics Check	OK	02:52

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORNELIUS,TIM,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORNELIUS,TIM
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 10/07/2017 02:47:59

Diagnostics Check		OK	02:41
Air Blank		0.000	02:41
Control Test		0.079	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.184	02:43
Breath Volume		4.011	02:43
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.186	02:46
Breath Volume		4.394	02:46
Air Blank		0.000	02:47
Control Test		0.079	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELIAS,D,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCORMICK,P
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: BRANDTS

DATE/TIME OF TEST: 10/07/2017 04:00:03

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.079	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.255	03:51
Breath Volume		2.015	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.230	03:55
Breath Volume		1.414	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 3	0.222	03:58
Breath Volume		2.132	03:58
Air Blank		0.000	03:59
Control Test		0.077	03:59
Air Blank		0.000	03:59
Diagnostics Check		OK	04:00

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CURTSINGER,CODY,A
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DIAZ,R
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: LECCIMA

DATE/TIME OF TEST: 10/08/2017 21:51:23

Diagnostics Check	OK	21:45
Air Blank	0.000	21:45
Control Test	0.079	21:46
Air Blank	0.000	21:46
Subject Sample # 1	0.000	21:47
Breath Volume	2.847	21:47
Air Blank	0.000	21:47
Air Blank	0.000	21:49
Subject Sample # 2	0.000	21:50
Breath Volume	2.503	21:50
Air Blank	0.000	21:50
Control Test	0.080	21:50
Air Blank	0.000	21:51
Diagnostics Check	OK	21:51

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JESSE,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCGILVREY,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MOSCATO

DATE/TIME OF TEST: 10/10/2017 04:07:19

Diagnostics Check	OK	04:01
Air Blank	0.000	04:01
Control Test	0.079	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.036	04:02
Breath Volume	1.957	04:02
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.035	04:05
Breath Volume	2.117	04:05
Air Blank	0.000	04:06
Control Test	0.080	04:06
Air Blank	0.000	04:07
Diagnostics Check	OK	04:07

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRILLO,XAVIER,R
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCORMICK,P
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NAGY

DATE/TIME OF TEST: 10/12/2017 04:27:02

Diagnostics Check	OK	04:20
Air Blank	0.000	04:21
Control Test	0.080	04:21
Air Blank	0.000	04:22
Subject Sample # 1	0.231	04:22
Breath Volume	2.691	04:22
Air Blank	0.000	04:23
Air Blank	0.000	04:25
Subject Sample # 2	0.223	04:25
Breath Volume	2.519	04:25
Air Blank	0.000	04:26
Control Test	0.079	04:26
Air Blank	0.000	04:26
Diagnostics Check	OK	04:27

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIERMAN,BRENT,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PULIZZOTTO,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:33

SUBJECT'S LAST NAME: COWAN

DATE/TIME OF TEST: 10/12/2017 20:03:57

Diagnostics Check		OK	19:57
Air Blank		0.000	19:57
Control Test		0.079	19:58
Air Blank		0.000	19:58
Subject Sample	# 1	0.260	19:59
Breath Volume		2.562	19:59
Air Blank		0.000	19:59
Air Blank		0.000	20:01
Subject Sample	# 2	0.263	20:02
Breath Volume		2.621	20:02
Air Blank		0.000	20:03
Control Test		0.079	20:03
Air Blank		0.000	20:03
Diagnostics Check		OK	20:03

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKS,D,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: CORWIN,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KABASHI

DATE/TIME OF TEST: 10/13/2017 02:36:44

Diagnostics Check	OK	02:29
Air Blank	0.000	02:29
Control Test	0.080	02:30
Air Blank	0.000	02:30
Subject Sample # 1	0.144	02:32
Breath Volume	3.015	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.129	02:35
Breath Volume	2.566	02:35
Air Blank	0.000	02:35
Control Test	0.077	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DICARLO,J,T
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: LUKASZ,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 10/14/2017 02:57:28

Diagnostics Check	OK	02:50
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.201	02:52
Breath Volume	3.152	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.185	02:55
Breath Volume	1.847	02:55
Air Blank	0.000	02:56
Control Test	0.079	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLOPS,C,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: COSTELLO,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/14/2017 03:17:17

Diagnostics Check		OK	03:04	
Air Blank		0.000	03:04	
Control Test		0.079	03:05	
Air Blank		0.000	03:05	
Subject Sample	# 1	0.147	03:08	Volume Not Met
Breath Volume		0.265	03:08	
Air Blank		0.000	03:09	
Air Blank		0.000	03:11	
Subject Sample	# 2	0.136	03:12	
Breath Volume		1.312	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:14	
Subject Sample	# 3	0.152	03:15	
Breath Volume		2.398	03:15	
Air Blank		0.000	03:16	
Control Test		0.077	03:16	
Air Blank		0.000	03:17	
Diagnostics Check		OK	03:17	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCOY,D
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:18

SUBJECT'S LAST NAME: HAINES

DATE/TIME OF TEST: 10/14/2017 03:48:57

Diagnostics Check	OK	03:39
Air Blank	0.000	03:40
Control Test	0.079	03:40
Air Blank	0.000	03:41
Subject Sample # 1	0.202	03:43
Breath Volume	1.136	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:46
Subject Sample # 2	0.189	03:47
Breath Volume	1.593	03:47
Air Blank	0.000	03:48
Control Test	0.076	03:48
Air Blank	0.000	03:48
Diagnostics Check	OK	03:48

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BATSON, MITCHELL, J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VONSOOSTEN, S
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:23

SUBJECT'S LAST NAME: SAINTILAIRE

DATE/TIME OF TEST: 10/14/2017 04:59:41

Diagnostics Check		OK	04:46	
Air Blank		0.000	04:47	
Control Test		0.079	04:47	
Air Blank		0.000	04:47	
Subject Sample	# 1	0.111	04:51	Volume Not Met
Breath Volume		0.718	04:51	
Air Blank		0.000	04:51	
Air Blank		0.000	04:53	
Subject Sample	# 2	0.128	04:54	
Breath Volume		1.914	04:54	
Air Blank		0.000	04:55	
Air Blank		0.000	04:57	
Subject Sample	# 3	0.131	04:58	
Breath Volume		1.437	04:58	
Air Blank		0.000	04:58	
Control Test		0.078	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,Y,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: GROSSKLAS,.
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:41

SUBJECT'S LAST NAME: SANKEY

DATE/TIME OF TEST: 10/15/2017 00:07:03

Diagnostics Check	OK	00:02	
Air Blank	0.000	00:02	
Control Test	0.079	00:03	
Air Blank	0.000	00:03	
Subject Sample	# 1	00:05	Subject Test Refused
Breath Volume	0.000	00:05	
Air Blank	0.000	00:06	
Control Test	0.079	00:06	
Air Blank	0.000	00:06	
Diagnostics Check	OK	00:07	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MAHOLTZ,S
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: FORREST

DATE/TIME OF TEST: 10/15/2017 01:22:57

Diagnostics Check	OK	01:16
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:17
Subject Sample # 1	0.000	01:18
Breath Volume	2.128	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.000	01:21
Breath Volume	1.300	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:22
Diagnostics Check	OK	01:22

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS, MATTHEW, L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER, K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 10/17/2017 04:29:32

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.080	04:23
Air Blank	0.000	04:24
Subject Sample # 1	0.145	04:24
Breath Volume	2.238	04:24
Air Blank	0.000	04:25
Air Blank	0.000	04:27
Subject Sample # 2	0.154	04:27
Breath Volume	2.316	04:27
Air Blank	0.000	04:28
Control Test	0.079	04:28
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: VIGLIANO

DATE/TIME OF TEST: 10/17/2017 22:27:49

Diagnostics Check	OK	22:21
Air Blank	0.000	22:22
Control Test	0.079	22:22
Air Blank	0.000	22:22
Subject Sample # 1	0.172	22:23
Breath Volume	2.207	22:23
Air Blank	0.000	22:23
Air Blank	0.000	22:25
Subject Sample # 2	0.163	22:26
Breath Volume	1.488	22:26
Air Blank	0.000	22:26
Control Test	0.079	22:27
Air Blank	0.000	22:27
Diagnostics Check	OK	22:27

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DRISCOLL,JAMES,P
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DRISCOLL,JAMES
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:13

SUBJECT'S LAST NAME: SAVOV

DATE/TIME OF TEST: 10/18/2017 23:40:08

Diagnostics Check	OK	23:34
Air Blank	0.000	23:34
Control Test	0.078	23:34
Air Blank	0.000	23:35
Subject Sample # 1	0.187	23:35
Breath Volume	3.140	23:35
Air Blank	0.000	23:36
Air Blank	0.000	23:38
Subject Sample # 2	0.196	23:38
Breath Volume	2.554	23:38
Air Blank	0.000	23:39
Control Test	0.079	23:39
Air Blank	0.000	23:40
Diagnostics Check	OK	23:40

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUSSA,KYLE,A
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: YAXLEY,BEATRICE
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 10/20/2017 21:27:30

Diagnostics Check	OK	21:21
Air Blank	0.000	21:21
Control Test	0.079	21:22
Air Blank	0.000	21:22
Subject Sample # 1	0.000	21:23
Breath Volume	2.242	21:23
Air Blank	0.000	21:23
Air Blank	0.000	21:25
Subject Sample # 2	0.000	21:26
Breath Volume	1.742	21:26
Air Blank	0.000	21:26
Control Test	0.079	21:26
Air Blank	0.000	21:27
Diagnostics Check	OK	21:27

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRATT,NICHOLAS,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MILLER,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 10/21/2017 01:06:57

Diagnostics Check		OK	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.220	01:02
Breath Volume		2.218	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.212	01:05
Breath Volume		3.070	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELMORE,D,A
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: ELMORE,D
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: LLAMAS

DATE/TIME OF TEST: 10/21/2017 01:39:58

Diagnostics Check	OK	01:33
Air Blank	0.000	01:33
Control Test	0.079	01:33
Air Blank	0.000	01:34
Subject Sample # 1	0.209	01:35
Breath Volume	3.777	01:35
Air Blank	0.000	01:35
Air Blank	0.000	01:37
Subject Sample # 2	0.210	01:38
Breath Volume	3.691	01:38
Air Blank	0.000	01:39
Control Test	0.078	01:39
Air Blank	0.000	01:39
Diagnostics Check	OK	01:39

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUONPANE,V,D
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCCORMICK,B
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: SUKER

DATE/TIME OF TEST: 10/21/2017 02:51:36

Diagnostics Check	OK	02:44
Air Blank	0.000	02:45
Control Test	0.079	02:45
Air Blank	0.000	02:45
Subject Sample # 1	0.165	02:46
Breath Volume	1.566	02:46
Air Blank	0.000	02:47
Air Blank	0.000	02:49
Subject Sample # 2	0.159	02:50
Breath Volume	1.429	02:50
Air Blank	0.000	02:50
Control Test	0.078	02:51
Air Blank	0.000	02:51
Diagnostics Check	OK	02:51

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ONDREA,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ALONSO,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CAMPBELL

DATE/TIME OF TEST: 10/21/2017 22:15:33

Diagnostics Check	OK	22:08
Air Blank	0.000	22:09
Control Test	0.080	22:09
Air Blank	0.000	22:10
Subject Sample # 1	0.071	22:10
Breath Volume	1.281	22:10
Air Blank	0.000	22:11
Air Blank	0.000	22:13
Subject Sample # 2	0.081	22:13
Breath Volume	2.503	22:13
Air Blank	0.000	22:14
Control Test	0.080	22:14
Air Blank	0.000	22:15
Diagnostics Check	OK	22:15

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,MARTIN,C
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: STUDZINSKI,K
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: FARIAS

DATE/TIME OF TEST: 10/22/2017 00:06:58

Diagnostics Check	OK	00:00
Air Blank	0.000	00:00
Control Test	0.079	00:00
Air Blank	0.000	00:01
Subject Sample # 1	0.141	00:01
Breath Volume	2.386	00:01
Air Blank	0.000	00:02
Air Blank	0.000	00:04
Subject Sample # 2	0.140	00:05
Breath Volume	2.609	00:05
Air Blank	0.000	00:06
Control Test	0.079	00:06
Air Blank	0.000	00:06
Diagnostics Check	OK	00:06

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO, JONATHAN, B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SMITH, L
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VILLELA

DATE/TIME OF TEST: 10/23/2017 23:45:46

Diagnostics Check	OK	23:43	
Air Blank	0.000	23:43	
Control Test	0.080	23:44	
Air Blank	0.000	23:44	
Subject Sample	# 1	23:44	Slope Not Met
Breath Volume	1.683	23:44	
Air Blank		23:45	Improper Sample
Air Blank	0.000	23:45	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: SMITH,L
ARRESTING OFFICER AGENCY: Marco Island PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VILLELA

DATE/TIME OF TEST: 10/23/2017 23:54:24

Diagnostics Check	OK	23:48
Air Blank	0.000	23:48
Control Test	0.079	23:49
Air Blank	0.000	23:49
Subject Sample # 1	0.342	23:50
Breath Volume	2.910	23:50
Air Blank	0.000	23:50
Air Blank	0.000	23:52
Subject Sample # 2	0.324	23:52
Breath Volume	3.539	23:52
Air Blank	0.000	23:53
Control Test	0.078	23:53
Air Blank	0.000	23:54
Diagnostics Check	OK	23:54

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOEL, YVES,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: VONSOOSTEN, S
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: CHERRY

DATE/TIME OF TEST: 10/25/2017 04:29:59

Diagnostics Check	OK	04:23
Air Blank	0.000	04:23
Control Test	0.079	04:24
Air Blank	0.000	04:24
Subject Sample # 1	0.096	04:25
Breath Volume	1.777	04:25
Air Blank	0.000	04:26
Air Blank	0.000	04:27
Subject Sample # 2	0.096	04:28
Breath Volume	2.019	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:29
Diagnostics Check	OK	04:29

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: Naples PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 10/25/2017 12:56:46

Diagnostics Check	OK	12:51	
Air Blank	0.000	12:51	
Control Test	0.080	12:52	
Air Blank	0.000	12:52	
Subject Sample	# 1 0.171	12:53	
Breath Volume	1.843	12:53	
Air Blank	0.000	12:53	
Air Blank		12:55	RFI Detect
Air Blank	0.000	12:55	
Control Test	0.081	12:56	
Air Blank	0.000	12:56	
Diagnostics Check	OK	12:56	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARP,RYAN,M
BREATH TEST OPERATOR AGENCY: Naples PD
ARRESTING OFFICER: HARP,RYAN
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: CHARLES

DATE/TIME OF TEST: 10/25/2017 13:04:36

Diagnostics Check		OK	12:58
Air Blank		0.000	12:58
Control Test		0.080	12:59
Air Blank		0.000	12:59
Subject Sample	# 1	0.176	13:00
Breath Volume		2.207	13:00
Air Blank		0.000	13:00
Air Blank		0.000	13:02
Subject Sample	# 2	0.179	13:03
Breath Volume		2.386	13:03
Air Blank		0.000	13:03
Control Test		0.080	13:04
Air Blank		0.000	13:04
Diagnostics Check		OK	13:04

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAGE,L,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: BENOIT,J
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:56

SUBJECT'S LAST NAME: KNOPP

DATE/TIME OF TEST: 10/26/2017 16:48:08

Diagnostics Check	OK	16:41
Air Blank	0.000	16:42
Control Test	0.081	16:42
Air Blank	0.000	16:43
Subject Sample # 1	0.178	16:43
Breath Volume	2.082	16:43
Air Blank	0.000	16:44
Air Blank	0.000	16:46
Subject Sample # 2	0.181	16:46
Breath Volume	2.824	16:46
Air Blank	0.000	16:47
Control Test	0.080	16:47
Air Blank	0.000	16:48
Diagnostics Check	OK	16:48

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHLIN,C,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ROWLEY,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:27

SUBJECT'S LAST NAME: WILD

DATE/TIME OF TEST: 10/27/2017 19:58:03

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.080	19:52
Air Blank	0.000	19:52
Subject Sample # 1	0.206	19:53
Breath Volume	1.781	19:53
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.218	19:56
Breath Volume	2.628	19:56
Air Blank	0.000	19:57
Control Test	0.079	19:57
Air Blank	0.000	19:57
Diagnostics Check	OK	19:58

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JESSE,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: DRUM

DATE/TIME OF TEST: 10/28/2017 00:28:09

Diagnostics Check		OK	00:16	
Air Blank		0.000	00:16	
Control Test		0.079	00:17	
Air Blank		0.000	00:17	
Subject Sample	# 1	0.201	00:20	Volume Not Met
Breath Volume		0.265	00:20	
Air Blank		0.000	00:21	
Air Blank		0.000	00:23	
Subject Sample	# 2	0.192	00:26	Volume Not Met
Breath Volume		0.144	00:26	
Air Blank		0.000	00:27	
Control Test		0.079	00:27	
Air Blank		0.000	00:28	
Diagnostics Check		OK	00:28	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TURNER,JESSE,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: NOVA,M
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: DRUM

DATE/TIME OF TEST: 10/28/2017 00:44:18

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:33	
Control Test		0.079	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 1	0.212	00:37	Volume Not Met
Breath Volume		0.343	00:37	
Air Blank		0.000	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 2		00:42	Subject Test Refused
Breath Volume		0.179	00:42	
Air Blank		0.000	00:43	
Control Test		0.078	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 11/02/2017 10:04

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEORGE,SHANNON,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MCGILVREY,J
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BIALECK

DATE/TIME OF TEST: 10/28/2017 02:19:07

Diagnostics Check	OK	02:12
Air Blank	0.000	02:12
Control Test	0.079	02:13
Air Blank	0.000	02:13
Subject Sample # 1	0.092	02:14
Breath Volume	1.328	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:16
Subject Sample # 2	0.098	02:17
Breath Volume	1.171	02:17
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 11/02/2017 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROHRAFF,HARRY,E
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: ZERPA,E
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: CLEMONS

DATE/TIME OF TEST: 10/28/2017 23:10:28

Diagnostics Check	OK	23:03
Air Blank	0.000	23:04
Control Test	0.079	23:04
Air Blank	0.000	23:04
Subject Sample # 1	0.230	23:05
Breath Volume	1.875	23:05
Air Blank	0.000	23:06
Air Blank	0.000	23:08
Subject Sample # 2	0.249	23:08
Breath Volume	1.414	23:08
Air Blank	0.000	23:09
Control Test	0.079	23:09
Air Blank	0.000	23:10
Diagnostics Check	OK	23:10

Data Download Date/Time: 11/02/2017 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NATSON,ANDRAE,M
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MARTINEZ,E
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: CANIZALES

DATE/TIME OF TEST: 10/29/2017 03:18:36

Diagnostics Check	OK	03:11
Air Blank	0.000	03:12
Control Test	0.080	03:12
Air Blank	0.000	03:13
Subject Sample # 1	0.152	03:13
Breath Volume	1.308	03:13
Air Blank	0.000	03:14
Air Blank	0.000	03:16
Subject Sample # 2	0.159	03:16
Breath Volume	1.417	03:16
Air Blank	0.000	03:17
Control Test	0.079	03:18
Air Blank	0.000	03:18
Diagnostics Check	OK	03:18

Data Download Date/Time: 11/02/2017 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAETJENS,MELISSA,N
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: MULRY,THOMAS
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/29/2017 07:19:59

Diagnostics Check	OK	07:13
Air Blank	0.000	07:14
Control Test	0.080	07:14
Air Blank	0.000	07:14
Subject Sample # 1	0.181	07:15
Breath Volume	3.664	07:15
Air Blank	0.000	07:16
Air Blank	0.000	07:18
Subject Sample # 2	0.163	07:18
Breath Volume	2.398	07:18
Air Blank	0.000	07:19
Control Test	0.078	07:19
Air Blank	0.000	07:19
Diagnostics Check	OK	07:19

Data Download Date/Time: 11/02/2017 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRUBIN,L,
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: DEPASQUALE,C
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:25

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/29/2017 16:53:30

Diagnostics Check	OK	16:46
Air Blank	0.000	16:47
Control Test	0.080	16:47
Air Blank	0.000	16:48
Subject Sample # 1	0.273	16:48
Breath Volume	1.867	16:48
Air Blank	0.000	16:49
Air Blank	0.000	16:51
Subject Sample # 2	0.273	16:51
Breath Volume	1.750	16:51
Air Blank	0.000	16:52
Control Test	0.078	16:52
Air Blank	0.000	16:53
Diagnostics Check	OK	16:53

Data Download Date/Time: 11/02/2017 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,KEVIN,J
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: PEREZ,M
ARRESTING OFFICER AGENCY: Naples PD
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 10/30/2017 03:56:12

Diagnostics Check	OK	03:49
Air Blank	0.000	03:50
Control Test	0.080	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.189	03:51
Breath Volume	2.507	03:51
Air Blank	0.000	03:52
Air Blank	0.000	03:54
Subject Sample # 2	0.174	03:54
Breath Volume	1.542	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:56
Diagnostics Check	OK	03:56

Data Download Date/Time: 11/02/2017 10:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000943

INSTRUMENT REGISTERED TO: COLLIER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KENNEDY,DANIEL,L
BREATH TEST OPERATOR AGENCY: Collier County SO
ARRESTING OFFICER: KENNEDY,DANIEL
ARRESTING OFFICER AGENCY: Collier County SO
LAST AGENCY INSPECTION: 10/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A3
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: PITONZO

DATE/TIME OF TEST: 10/31/2017 00:31:02

Diagnostics Check		OK	00:24
Air Blank		0.000	00:25
Control Test		0.079	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.205	00:26
Breath Volume		2.390	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.216	00:29
Breath Volume		2.187	00:29
Air Blank		0.000	00:30
Control Test		0.079	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:31

Data Download Date/Time: 11/02/2017 10:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: KITTS

DATE/TIME OF TEST: 10/02/2017 21:33:51

Diagnostics Check	OK	21:27
Air Blank	0.000	21:28
Control Test	0.082	21:28
Air Blank	0.000	21:29
Subject Sample # 1	0.000	21:29
Breath Volume	2.218	21:29
Air Blank	0.000	21:30
Air Blank	0.000	21:32
Subject Sample # 2	0.000	21:32
Breath Volume	1.902	21:32
Air Blank	0.000	21:32
Control Test	0.082	21:33
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WALKUSKI,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: KANE

DATE/TIME OF TEST: 10/05/2017 20:49:11

Diagnostics Check	OK	20:40
Air Blank	0.000	20:40
Control Test	0.082	20:41
Air Blank	0.000	20:41
Subject Sample # 1	0.107	20:44
Breath Volume	1.273	20:44
Air Blank	0.000	20:44
Air Blank	0.000	20:46
Subject Sample # 2	0.102	20:47
Breath Volume	1.390	20:47
Air Blank	0.000	20:48
Control Test	0.081	20:48
Air Blank	0.000	20:49
Diagnostics Check	OK	20:49

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KLOS,LINDSAY,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PELFREY,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: LAUREANO

DATE/TIME OF TEST: 10/07/2017 02:38:38

Diagnostics Check	OK	02:32
Air Blank	0.000	02:32
Control Test	0.080	02:33
Air Blank	0.000	02:33
Subject Sample # 1	0.223	02:34
Breath Volume	1.324	02:34
Air Blank	0.000	02:34
Air Blank	0.000	02:36
Subject Sample # 2	0.232	02:37
Breath Volume	1.683	02:37
Air Blank	0.000	02:37
Control Test	0.081	02:38
Air Blank	0.000	02:38
Diagnostics Check	OK	02:38

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: WOLKE,OFFICER
ARRESTING OFFICER AGENCY: Punta Gorda PD
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 10/09/2017 01:45:29

Diagnostics Check	OK	01:39
Air Blank	0.000	01:39
Control Test	0.080	01:40
Air Blank	0.000	01:40
Subject Sample # 1	0.246	01:41
Breath Volume	3.113	01:41
Air Blank	0.000	01:41
Air Blank	0.000	01:43
Subject Sample # 2	0.237	01:44
Breath Volume	3.074	01:44
Air Blank	0.000	01:44
Control Test	0.081	01:44
Air Blank	0.000	01:45
Diagnostics Check	OK	01:45

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FISHER,SCOTT,R
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: ROOT,DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: HEMMELGARN

DATE/TIME OF TEST: 10/12/2017 01:58:48

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.081	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.205	01:54
Breath Volume		1.671	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.203	01:57
Breath Volume		1.582	01:57
Air Blank		0.000	01:57
Control Test		0.081	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:58

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLATZ,KEVIN,L
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: PATTEN,JACOB
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: LUTEROTTY

DATE/TIME OF TEST: 10/13/2017 23:56:15

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.081	23:49
Air Blank	0.000	23:50
Subject Sample # 1	0.151	23:51
Breath Volume	2.890	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.147	23:54
Breath Volume	2.894	23:54
Air Blank	0.000	23:55
Control Test	0.082	23:55
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,VANESSA,N
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: SOLTIS,TONY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:53

SUBJECT'S LAST NAME: FARRELL

DATE/TIME OF TEST: 10/14/2017 01:20:11

Diagnostics Check	OK	01:13
Air Blank	0.000	01:14
Control Test	0.082	01:14
Air Blank	0.000	01:15
Subject Sample # 1	0.052	01:15
Breath Volume	1.707	01:15
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.050	01:18
Breath Volume	1.597	01:18
Air Blank	0.000	01:19
Control Test	0.082	01:19
Air Blank	0.000	01:20
Diagnostics Check	OK	01:20

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCLAIN,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: MYRIE,K
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: DICKERSON

DATE/TIME OF TEST: 10/15/2017 23:19:38

Diagnostics Check	OK	23:11
Air Blank	0.000	23:11
Control Test	0.082	23:11
Air Blank	0.000	23:12
Subject Sample # 1	0.185	23:15
Breath Volume	1.312	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.200	23:18
Breath Volume	1.800	23:18
Air Blank	0.000	23:18
Control Test	0.080	23:19
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLEURANT,RODSON,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: FLEURANT,RODSON
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:59

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 10/16/2017 06:58:28

Diagnostics Check	OK	06:52
Air Blank	0.000	06:52
Control Test	0.081	06:52
Air Blank	0.000	06:53
Subject Sample # 1	0.100	06:54
Breath Volume	2.949	06:54
Air Blank	0.000	06:54
Air Blank	0.000	06:56
Subject Sample # 2	0.086	06:56
Breath Volume	1.578	06:56
Air Blank	0.000	06:57
Control Test	0.081	06:57
Air Blank	0.000	06:58
Diagnostics Check	OK	06:58

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000945

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER, JEREMY, A
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: JONES, DEPUTY
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 09/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 10/17/2017 01:49:49

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:37	
Control Test	0.081	01:38	
Air Blank	0.000	01:38	
Subject Sample # 1	0.131	01:39	
Breath Volume	1.347	01:39	
Air Blank	0.000	01:39	
Air Blank	0.000	01:41	
Subject Sample # 2	0.139	01:44	Volume Not Met
Breath Volume	0.699	01:44	
Air Blank	0.000	01:45	
Air Blank	0.000	01:47	
Subject Sample # 3	0.135	01:48	
Breath Volume	2.195	01:48	
Air Blank	0.000	01:48	
Control Test	0.080	01:49	
Air Blank	0.000	01:49	
Diagnostics Check	OK	01:49	

Data Download Date/Time: 10/18/2017 09:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ, GLORIA, C
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: CLENDENING, JAMES
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:35

SUBJECT'S LAST NAME: BIEN AIME

DATE/TIME OF TEST: 10/07/2017 10:11:52

Diagnostics Check	OK	10:05
Air Blank	0.000	10:05
Control Test	0.076	10:06
Air Blank	0.000	10:06
Subject Sample # 1	0.197	10:07
Breath Volume	3.144	10:07
Air Blank	0.000	10:08
Air Blank	0.000	10:09
Subject Sample # 2	0.195	10:10
Breath Volume	3.402	10:10
Air Blank	0.000	10:10
Control Test	0.075	10:11
Air Blank	0.000	10:11
Diagnostics Check	OK	10:11

Data Download Date/Time: 10/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: SAMMONS

DATE/TIME OF TEST: 10/13/2017 21:24:03

Diagnostics Check	OK	21:17
Air Blank	0.000	21:18
Control Test	0.077	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.000	21:19
Breath Volume	1.800	21:19
Air Blank	0.000	21:20
Air Blank	0.000	21:22
Subject Sample # 2	0.000	21:22
Breath Volume	2.332	21:22
Air Blank	0.000	21:23
Control Test	0.077	21:23
Air Blank	0.000	21:23
Diagnostics Check	OK	21:24

Data Download Date/Time: 10/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: OKUPSKI,RYAN
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:56

SUBJECT'S LAST NAME: CARVER

DATE/TIME OF TEST: 10/13/2017 21:42:58

Diagnostics Check	OK	21:35
Air Blank	0.000	21:36
Control Test	0.077	21:36
Air Blank	0.000	21:37
Subject Sample # 1	0.087	21:37
Breath Volume	1.437	21:37
Air Blank	0.000	21:38
Air Blank	0.000	21:40
Subject Sample # 2	0.089	21:41
Breath Volume	1.867	21:41
Air Blank	0.000	21:42
Control Test	0.077	21:42
Air Blank	0.000	21:42
Diagnostics Check	OK	21:42

Data Download Date/Time: 10/31/2017 11:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000947

INSTRUMENT REGISTERED TO: GLADES COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOMS,DONNA,J
BREATH TEST OPERATOR AGENCY: Glades County SO
ARRESTING OFFICER: HERNANDEZ,DANIEL
ARRESTING OFFICER AGENCY: Glades County SO
LAST AGENCY INSPECTION: 09/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 10/15/2017 03:01:34

Diagnostics Check	OK	02:58	
Air Blank	0.000	02:59	
Control Test	0.078	02:59	
Air Blank	0.000	03:00	
Subject Sample	# 1	03:00	Subject Test Refused
Breath Volume	0.000	03:00	
Air Blank	0.000	03:00	
Control Test	0.078	03:00	
Air Blank	0.000	03:01	
Diagnostics Check	OK	03:01	

Data Download Date/Time: 10/31/2017 11:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000950

INSTRUMENT REGISTERED TO: HAVANA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: MEADOWS

DATE/TIME OF TEST: 10/22/2017 05:53:07

Diagnostics Check		OK	05:46
Air Blank		0.000	05:47
Control Test		0.077	05:47
Air Blank		0.000	05:48
Subject Sample	# 1	0.201	05:48
Breath Volume		2.707	05:48
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 2	0.207	05:51
Breath Volume		2.355	05:51
Air Blank		0.000	05:52
Control Test		0.076	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 11/14/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,HENRY,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAMM,JAMES
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/28/2017 00:06:07

Diagnostics Check	OK	00:05	
Air Blank		00:05	RFI Detect
Air Blank	0.000	00:06	

Data Download Date/Time: 11/21/2017 11:22

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000953

INSTRUMENT REGISTERED TO: FRANKLIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,HENRY,E
BREATH TEST OPERATOR AGENCY: Franklin County SO
ARRESTING OFFICER: HAMM,JAMES
ARRESTING OFFICER AGENCY: Franklin County SO
LAST AGENCY INSPECTION: 10/24/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 10/28/2017 00:15:02

Diagnostics Check	OK	00:08
Air Blank	0.000	00:09
Control Test	0.079	00:09
Air Blank	0.000	00:09
Subject Sample # 1	0.114	00:10
Breath Volume	3.031	00:10
Air Blank	0.000	00:11
Air Blank	0.000	00:12
Subject Sample # 2	0.112	00:13
Breath Volume	2.816	00:13
Air Blank	0.000	00:14
Control Test	0.077	00:14
Air Blank	0.000	00:14
Diagnostics Check	OK	00:15

Data Download Date/Time: 11/21/2017 11:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG, WILLIAM, B
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: YOUNG, WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: ROBERTS

DATE/TIME OF TEST: 10/07/2017 03:51:35

Diagnostics Check	OK	03:45
Air Blank	0.000	03:45
Control Test	0.078	03:46
Air Blank	0.000	03:46
Subject Sample # 1	0.152	03:47
Breath Volume	2.347	03:47
Air Blank	0.000	03:47
Air Blank	0.000	03:49
Subject Sample # 2	0.152	03:50
Breath Volume	2.593	03:50
Air Blank	0.000	03:50
Control Test	0.078	03:50
Air Blank	0.000	03:51
Diagnostics Check	OK	03:51

Data Download Date/Time: 10/11/2017 11:27

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLEIGHER,JOHN,C
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: PITTS,JACK
ARRESTING OFFICER AGENCY: Monticello PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 10/15/2017 01:02:57

Diagnostics Check	OK	01:00	
Air Blank	0.000	01:00	
Control Test	0.079	01:00	
Air Blank	0.000	01:01	
Subject Sample	# 1	01:01	Subject Test Refused
Breath Volume	0.000	01:01	
Air Blank	0.000	01:02	
Control Test	0.079	01:02	
Air Blank	0.000	01:02	
Diagnostics Check	OK	01:02	

Data Download Date/Time: 11/15/2017 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000954

INSTRUMENT REGISTERED TO: JEFFERSON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLEMENT,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: CAMPER,ALAN
ARRESTING OFFICER AGENCY: Jefferson County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: PHILLIPS

DATE/TIME OF TEST: 10/22/2017 02:22:12

Diagnostics Check	OK	02:14
Air Blank	0.000	02:15
Control Test	0.079	02:15
Air Blank	0.000	02:15
Subject Sample # 1	0.198	02:17
Breath Volume	2.031	02:17
Air Blank	0.000	02:17
Air Blank	0.000	02:19
Subject Sample # 2	0.186	02:20
Breath Volume	1.433	02:20
Air Blank	0.000	02:21
Control Test	0.078	02:21
Air Blank	0.000	02:22
Diagnostics Check	OK	02:22

Data Download Date/Time: 11/15/2017 10:47

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000955

INSTRUMENT REGISTERED TO: CLEWISTON P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAMBERLAIN,STEVEN,M
BREATH TEST OPERATOR AGENCY: Clewiston PD
ARRESTING OFFICER: WOODS,JOSHUA
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 10/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: GLORIA

DATE/TIME OF TEST: 10/07/2017 03:56:38

Diagnostics Check	OK	03:53	
Air Blank	0.000	03:54	
Control Test	0.080	03:54	
Air Blank	0.000	03:55	
Subject Sample	# 1	03:55	Subject Test Refused
Breath Volume	0.000	03:55	
Air Blank	0.000	03:55	
Control Test	0.081	03:56	
Air Blank	0.000	03:56	
Diagnostics Check	OK	03:56	

Data Download Date/Time: 11/28/2017 21:00

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,ERIC,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ANDERSON,CASEY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 10/25/2017 23:41:05

Diagnostics Check	OK	23:34	
Air Blank	0.000	23:35	
Control Test	0.081	23:35	
Air Blank	0.000	23:36	
Subject Sample # 1	0.190	23:37	
Breath Volume	1.207	23:37	
Air Blank	0.000	23:38	
Air Blank		23:39	Improper Sample
Air Blank	0.000	23:40	
Control Test	0.082	23:40	
Air Blank	0.000	23:41	
Diagnostics Check	OK	23:41	

Data Download Date/Time: 11/10/2017 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLETCHER,ERIC,D
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: ANDERSON,CASEY
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: THIGPEN

DATE/TIME OF TEST: 10/26/2017 00:03:06

Diagnostics Check	OK	23:56
Air Blank	0.000	23:57
Control Test	0.080	23:57
Air Blank	0.000	23:57
Subject Sample # 2	0.187	23:58
Breath Volume	1.652	23:58
Air Blank	0.000	23:59
Air Blank	0.000	00:01
Subject Sample # 1	0.194	00:01
Breath Volume	2.332	00:01
Air Blank	0.000	00:02
Control Test	0.081	00:02
Air Blank	0.000	00:03
Diagnostics Check	OK	00:03

Data Download Date/Time: 11/10/2017 10:09

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000957

INSTRUMENT REGISTERED TO: LEON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTCLIFFE,BRIAN,R
BREATH TEST OPERATOR AGENCY: FL State University
ARRESTING OFFICER: CUTCLIFFE,BRIAN
ARRESTING OFFICER AGENCY: FL State University
LAST AGENCY INSPECTION: 10/25/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: PAULSEN

DATE/TIME OF TEST: 10/27/2017 02:44:28

Diagnostics Check	OK	02:38
Air Blank	0.000	02:38
Control Test	0.081	02:39
Air Blank	0.000	02:39
Subject Sample # 1	0.142	02:40
Breath Volume	2.457	02:40
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.131	02:42
Breath Volume	1.871	02:42
Air Blank	0.000	02:43
Control Test	0.081	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 11/10/2017 10:09

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHEPARD,JOSHUA,M
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: SHEPARD,JOSHUA
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: REED

DATE/TIME OF TEST: 10/13/2017 17:44:17

Diagnostics Check	OK	17:41	
Air Blank	0.000	17:41	
Control Test	0.081	17:41	
Air Blank	0.000	17:42	
Subject Sample	# 1	17:42	Subject Test Refused
Breath Volume	0.000	17:42	
Air Blank	0.000	17:43	
Control Test	0.080	17:43	
Air Blank	0.000	17:44	
Diagnostics Check	OK	17:44	

Data Download Date/Time: 10/23/2017 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000960

INSTRUMENT REGISTERED TO: LIBERTY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,ADAM,O
BREATH TEST OPERATOR AGENCY: Liberty County SO
ARRESTING OFFICER: WALKER,ADAM
ARRESTING OFFICER AGENCY: Liberty County SO
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: DURDEN

DATE/TIME OF TEST: 10/15/2017 00:30:06

Diagnostics Check	OK	00:23
Air Blank	0.000	00:23
Control Test	0.082	00:23
Air Blank	0.000	00:24
Subject Sample # 1	0.268	00:25
Breath Volume	1.355	00:25
Air Blank	0.000	00:25
Air Blank	0.000	00:27
Subject Sample # 2	0.259	00:28
Breath Volume	1.984	00:28
Air Blank	0.000	00:29
Control Test	0.081	00:29
Air Blank	0.000	00:29
Diagnostics Check	OK	00:30

Data Download Date/Time: 10/23/2017 10:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,VICKI,N
BREATH TEST OPERATOR AGENCY: Wakulla County SO
ARRESTING OFFICER: SHEFFIELD,BRANDON
ARRESTING OFFICER AGENCY: Wakulla County SO
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: HICKS

DATE/TIME OF TEST: 10/13/2017 11:46:06

Diagnostics Check	OK	11:34	
Air Blank	0.000	11:34	
Control Test	0.079	11:34	
Air Blank	0.000	11:35	
Subject Sample	# 1	11:38	No Sample Provided
Breath Volume	0.000	11:38	
Air Blank	0.000	11:39	
Air Blank	0.000	11:41	
Subject Sample	# 2	0.000	11:41
Breath Volume	3.535	11:41	
Air Blank	0.000	11:42	
Air Blank	0.000	11:44	
Subject Sample	# 3	0.000	11:44
Breath Volume	2.500	11:44	
Air Blank	0.000	11:45	
Control Test	0.079	11:45	
Air Blank	0.000	11:46	
Diagnostics Check	OK	11:46	

Data Download Date/Time: 11/09/2017 08:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000962

INSTRUMENT REGISTERED TO: WAKULLA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TALLMAN,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: TALLMAN,JOHN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: MOLDER

DATE/TIME OF TEST: 10/18/2017 17:06:51

Diagnostics Check	OK	17:00
Air Blank	0.000	17:01
Control Test	0.080	17:01
Air Blank	0.000	17:01
Subject Sample # 1	0.253	17:02
Breath Volume	2.566	17:02
Air Blank	0.000	17:03
Air Blank	0.000	17:04
Subject Sample # 2	0.252	17:05
Breath Volume	3.230	17:05
Air Blank	0.000	17:05
Control Test	0.078	17:06
Air Blank	0.000	17:06
Diagnostics Check	OK	17:06

Data Download Date/Time: 11/09/2017 08:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MCFERREN,JASON
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: JACKMAN

DATE/TIME OF TEST: 10/01/2017 00:54:38

Diagnostics Check		OK	00:48
Air Blank		0.000	00:49
Control Test		0.082	00:49
Air Blank		0.000	00:49
Subject Sample	# 1	0.178	00:50
Breath Volume		2.562	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.186	00:53
Breath Volume		2.406	00:53
Air Blank		0.000	00:53
Control Test		0.082	00:54
Air Blank		0.000	00:54
Diagnostics Check		OK	00:54

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: FRU RODNEY

DATE/TIME OF TEST: 10/01/2017 03:26:37

Diagnostics Check	OK	03:15	
Air Blank	0.000	03:15	
Control Test	0.082	03:16	
Air Blank	0.000	03:16	
Subject Sample	# 1	03:19	No Sample Provided
Breath Volume	0.000	03:19	
Air Blank	0.000	03:20	
Air Blank	0.000	03:22	
Subject Sample	# 2	03:25	No Sample Provided
Breath Volume	0.000	03:25	
Air Blank	0.000	03:25	
Control Test	0.082	03:26	
Air Blank	0.000	03:26	
Diagnostics Check	OK	03:26	

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KERN,E
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 10/01/2017 04:24:09

Diagnostics Check	OK	04:17
Air Blank	0.000	04:18
Control Test	0.082	04:18
Air Blank	0.000	04:19
Subject Sample # 1	0.188	04:19
Breath Volume	2.730	04:19
Air Blank	0.000	04:20
Air Blank	0.000	04:22
Subject Sample # 2	0.184	04:22
Breath Volume	2.875	04:22
Air Blank	0.000	04:23
Control Test	0.081	04:23
Air Blank	0.000	04:24
Diagnostics Check	OK	04:24

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,JORGE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: DURAN ESPINAL

DATE/TIME OF TEST: 10/01/2017 04:44:00

Diagnostics Check	OK	04:37
Air Blank	0.000	04:37
Control Test	0.082	04:38
Air Blank	0.000	04:38
Subject Sample # 1	0.108	04:39
Breath Volume	2.117	04:39
Air Blank	0.000	04:40
Air Blank	0.000	04:42
Subject Sample # 2	0.111	04:42
Breath Volume	1.808	04:42
Air Blank	0.000	04:43
Control Test	0.080	04:43
Air Blank	0.000	04:43
Diagnostics Check	OK	04:44

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,D
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 10/01/2017 06:23:03

Diagnostics Check	OK	06:17
Air Blank	0.000	06:17
Control Test	0.081	06:17
Air Blank	0.000	06:18
Subject Sample # 1	0.179	06:18
Breath Volume	1.898	06:18
Air Blank	0.000	06:19
Air Blank	0.000	06:21
Subject Sample # 2	0.177	06:21
Breath Volume	1.980	06:21
Air Blank	0.000	06:22
Control Test	0.081	06:22
Air Blank	0.000	06:22
Diagnostics Check	OK	06:23

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 10/02/2017 23:04:13

Diagnostics Check	OK	22:58
Air Blank	0.000	22:58
Control Test	0.081	22:59
Air Blank	0.000	22:59
Subject Sample # 1	0.021	22:59
Breath Volume	2.257	22:59
Air Blank	0.000	23:00
Air Blank	0.000	23:02
Subject Sample # 2	0.022	23:02
Breath Volume	3.878	23:02
Air Blank	0.000	23:03
Control Test	0.082	23:03
Air Blank	0.000	23:04
Diagnostics Check	OK	23:04

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LUFKIN,WILLIAM
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: FORTIER

DATE/TIME OF TEST: 10/05/2017 02:15:07

Diagnostics Check	OK	02:09
Air Blank	0.000	02:09
Control Test	0.083	02:09
Air Blank	0.000	02:10
Subject Sample # 1	0.126	02:10
Breath Volume	2.398	02:10
Air Blank	0.000	02:11
Air Blank	0.000	02:13
Subject Sample # 2	0.128	02:13
Breath Volume	2.394	02:13
Air Blank	0.000	02:14
Control Test	0.082	02:14
Air Blank	0.000	02:15
Diagnostics Check	OK	02:15

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: MONTES

DATE/TIME OF TEST: 10/05/2017 05:37:21

Diagnostics Check	OK	05:29
Air Blank	0.000	05:30
Control Test	0.082	05:30
Air Blank	0.000	05:30
Subject Sample # 1	0.167	05:32
Breath Volume	2.464	05:32
Air Blank	0.000	05:32
Air Blank	0.000	05:34
Subject Sample # 2	0.169	05:35
Breath Volume	2.062	05:35
Air Blank	0.000	05:36
Control Test	0.082	05:36
Air Blank	0.000	05:37
Diagnostics Check	OK	05:37

Data Download Date/Time: 10/05/2017 15:54

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KEANE,PETER
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: RIVASALONZO

DATE/TIME OF TEST: 10/06/2017 02:41:39

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.082	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.245	02:37
Breath Volume	1.714	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.233	02:40
Breath Volume	2.753	02:40
Air Blank	0.000	02:40
Control Test	0.081	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: COMPTON,ALAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: GIRON

DATE/TIME OF TEST: 10/06/2017 04:42:54

Diagnostics Check	OK	04:36
Air Blank	0.000	04:37
Control Test	0.081	04:37
Air Blank	0.000	04:38
Subject Sample # 1	0.074	04:38
Breath Volume	1.871	04:38
Air Blank	0.000	04:39
Air Blank	0.000	04:41
Subject Sample # 2	0.075	04:41
Breath Volume	1.460	04:41
Air Blank	0.000	04:42
Control Test	0.082	04:42
Air Blank	0.000	04:42
Diagnostics Check	OK	04:42

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAIGHT,MONTAGUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:18

SUBJECT'S LAST NAME: BURFORD

DATE/TIME OF TEST: 10/06/2017 07:47:15

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.082	07:41
Air Blank	0.000	07:42
Subject Sample # 1	0.000	07:42
Breath Volume	2.089	07:42
Air Blank	0.000	07:43
Air Blank	0.000	07:45
Subject Sample # 2	0.000	07:45
Breath Volume	3.066	07:45
Air Blank	0.000	07:46
Control Test	0.082	07:46
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: THURSTON

DATE/TIME OF TEST: 10/07/2017 03:30:47

Diagnostics Check	OK	03:24
Air Blank	0.000	03:24
Control Test	0.081	03:25
Air Blank	0.000	03:25
Subject Sample # 1	0.185	03:26
Breath Volume	3.128	03:26
Air Blank	0.000	03:27
Air Blank	0.000	03:28
Subject Sample # 2	0.185	03:29
Breath Volume	2.019	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:30
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALLEN, MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:30

SUBJECT'S LAST NAME: OAKLEY

DATE/TIME OF TEST: 10/07/2017 13:06:43

Diagnostics Check	OK	13:00
Air Blank	0.000	13:00
Control Test	0.082	13:01
Air Blank	0.000	13:01
Subject Sample # 1	0.000	13:01
Breath Volume	1.210	13:01
Air Blank	0.000	13:02
Air Blank	0.000	13:04
Subject Sample # 2	0.000	13:05
Breath Volume	2.398	13:05
Air Blank	0.000	13:05
Control Test	0.082	13:06
Air Blank	0.000	13:06
Diagnostics Check	OK	13:06

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLLEY,WILLIAM
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:16

SUBJECT'S LAST NAME: MANION

DATE/TIME OF TEST: 10/07/2017 18:50:12

Diagnostics Check	OK	18:42
Air Blank	0.000	18:43
Control Test	0.082	18:43
Air Blank	0.000	18:43
Subject Sample # 1	0.131	18:44
Breath Volume	1.218	18:44
Air Blank	0.000	18:45
Air Blank	0.000	18:46
Subject Sample # 2	0.119	18:48
Breath Volume	1.640	18:48
Air Blank	0.000	18:49
Control Test	0.081	18:49
Air Blank	0.000	18:50
Diagnostics Check	OK	18:50

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: LANDIN,CARLOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SANTIESTEBAN

DATE/TIME OF TEST: 10/08/2017 02:28:55

Diagnostics Check	OK	02:17	
Air Blank	0.000	02:17	
Control Test	0.081	02:18	
Air Blank	0.000	02:18	
Subject Sample	# 1	02:21	No Sample Provided
Breath Volume	0.000	02:21	
Air Blank	0.000	02:22	
Air Blank	0.000	02:24	
Subject Sample	# 2	02:27	No Sample Provided
Breath Volume	0.000	02:27	
Air Blank	0.000	02:28	
Control Test	0.080	02:28	
Air Blank	0.000	02:28	
Diagnostics Check	OK	02:28	

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BORRAS,TAMARA
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: CHASTAIN

DATE/TIME OF TEST: 10/08/2017 04:16:41

Diagnostics Check	OK	04:10
Air Blank	0.000	04:10
Control Test	0.081	04:11
Air Blank	0.000	04:11
Subject Sample # 1	0.203	04:12
Breath Volume	3.359	04:12
Air Blank	0.000	04:12
Air Blank	0.000	04:14
Subject Sample # 2	0.202	04:15
Breath Volume	3.246	04:15
Air Blank	0.000	04:15
Control Test	0.081	04:16
Air Blank	0.000	04:16
Diagnostics Check	OK	04:16

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,D
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: DENAVARD

DATE/TIME OF TEST: 10/08/2017 06:01:09

Diagnostics Check	OK	05:54
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:55
Subject Sample # 1	0.240	05:56
Breath Volume	1.238	05:56
Air Blank	0.000	05:56
Air Blank	0.000	05:58
Subject Sample # 2	0.245	05:59
Breath Volume	1.398	05:59
Air Blank	0.000	06:00
Control Test	0.081	06:00
Air Blank	0.000	06:01
Diagnostics Check	OK	06:01

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,GEAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: SHERWOOD

DATE/TIME OF TEST: 10/08/2017 19:59:37

Diagnostics Check		OK	19:47	
Air Blank		0.000	19:48	
Control Test		0.082	19:48	
Air Blank		0.000	19:49	
Subject Sample	# 1	0.266	19:52	Volume Not Met
Breath Volume		1.078	19:52	
Air Blank		0.000	19:52	
Air Blank		0.000	19:54	
Subject Sample	# 2	0.277	19:54	
Breath Volume		1.363	19:54	
Air Blank		0.000	19:55	
Air Blank		0.000	19:57	
Subject Sample	# 3	0.284	19:58	
Breath Volume		1.457	19:58	
Air Blank		0.000	19:58	
Control Test		0.080	19:59	
Air Blank		0.000	19:59	
Diagnostics Check		OK	19:59	

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VINER,DENNIS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: PRUITT

DATE/TIME OF TEST: 10/09/2017 02:47:20

Diagnostics Check		OK	02:35	
Air Blank		0.000	02:35	
Control Test		0.082	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1	0.010	02:39	Volume Not Met
Breath Volume		0.937	02:39	
Air Blank		0.000	02:40	
Air Blank		0.000	02:42	
Subject Sample	# 2	0.011	02:43	
Breath Volume		1.406	02:43	
Air Blank		0.000	02:43	
Air Blank		0.000	02:45	
Subject Sample	# 3	0.009	02:45	
Breath Volume		1.683	02:45	
Air Blank		0.000	02:46	
Control Test		0.081	02:46	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PARRISH,KAYLA
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/10/2017 03:47:22

Diagnostics Check	OK	03:35	
Air Blank	0.000	03:36	
Control Test	0.082	03:36	
Air Blank	0.000	03:37	
Subject Sample	# 1	03:40	No Sample Provided
Breath Volume	0.000	03:40	
Air Blank	0.000	03:40	
Air Blank	0.000	03:42	
Subject Sample	# 2	03:46	No Sample Provided
Breath Volume	0.000	03:46	
Air Blank	0.000	03:46	
Control Test	0.082	03:46	
Air Blank	0.000	03:47	
Diagnostics Check	OK	03:47	

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BRITT,JOHN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: ORTIZFREYTES

DATE/TIME OF TEST: 10/10/2017 05:04:54

Diagnostics Check	OK	04:58
Air Blank	0.000	04:59
Control Test	0.083	04:59
Air Blank	0.000	05:00
Subject Sample # 1	0.120	05:00
Breath Volume	2.914	05:00
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.120	05:03
Breath Volume	3.042	05:03
Air Blank	0.000	05:03
Control Test	0.081	05:04
Air Blank	0.000	05:04
Diagnostics Check	OK	05:04

Data Download Date/Time: 10/11/2017 09:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WING,RYAN
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: MATOS

DATE/TIME OF TEST: 10/12/2017 02:03:27

Diagnostics Check	OK	01:56
Air Blank	0.000	01:57
Control Test	0.082	01:57
Air Blank	0.000	01:58
Subject Sample # 1	0.196	01:58
Breath Volume	3.078	01:58
Air Blank	0.000	01:59
Air Blank	0.000	02:01
Subject Sample # 2	0.192	02:01
Breath Volume	3.296	02:01
Air Blank	0.000	02:02
Control Test	0.081	02:02
Air Blank	0.000	02:03
Diagnostics Check	OK	02:03

Data Download Date/Time: 10/19/2017 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: CABREJA

DATE/TIME OF TEST: 10/13/2017 00:22:27

Diagnostics Check	OK	00:15
Air Blank	0.000	00:16
Control Test	0.081	00:16
Air Blank	0.000	00:17
Subject Sample # 1	0.213	00:17
Breath Volume	3.554	00:17
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.216	00:20
Breath Volume	1.929	00:20
Air Blank	0.000	00:21
Control Test	0.081	00:21
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 10/19/2017 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: IRWIN,CHIP
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: HARPER

DATE/TIME OF TEST: 10/14/2017 02:26:37

Diagnostics Check	OK	02:20
Air Blank	0.000	02:20
Control Test	0.081	02:21
Air Blank	0.000	02:21
Subject Sample # 1	0.138	02:22
Breath Volume	1.886	02:22
Air Blank	0.000	02:22
Air Blank	0.000	02:24
Subject Sample # 2	0.135	02:25
Breath Volume	1.554	02:25
Air Blank	0.000	02:25
Control Test	0.081	02:26
Air Blank	0.000	02:26
Diagnostics Check	OK	02:26

Data Download Date/Time: 10/19/2017 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MONTGOMERY,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 10/15/2017 03:19:11

Diagnostics Check		OK	03:07	
Air Blank		0.000	03:08	
Control Test		0.081	03:08	
Air Blank		0.000	03:08	
Subject Sample	# 1	0.187	03:12	Volume Not Met
Breath Volume		0.000	03:12	
Air Blank		0.000	03:12	
Air Blank		0.000	03:14	
Subject Sample	# 2	0.191	03:17	Volume Not Met
Breath Volume		0.210	03:17	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 10/19/2017 09:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,AUGUSTIN
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:07

SUBJECT'S LAST NAME: ALYAMMAHI

DATE/TIME OF TEST: 10/17/2017 03:37:37

Diagnostics Check	OK	03:31
Air Blank	0.000	03:31
Control Test	0.082	03:32
Air Blank	0.000	03:32
Subject Sample # 1	0.139	03:33
Breath Volume	2.746	03:33
Air Blank	0.000	03:33
Air Blank	0.000	03:35
Subject Sample # 2	0.134	03:36
Breath Volume	1.238	03:36
Air Blank	0.000	03:36
Control Test	0.082	03:37
Air Blank	0.000	03:37
Diagnostics Check	OK	03:37

Data Download Date/Time: 10/19/2017 09:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RAMOSRODRIGUEZ

DATE/TIME OF TEST: 10/21/2017 03:45:52

Diagnostics Check	OK	03:37	
Air Blank	0.000	03:37	
Control Test	0.081	03:38	
Air Blank	0.000	03:38	
Subject Sample	# 1	03:41	No Sample Provided
Breath Volume	0.000	03:41	
Air Blank	0.000	03:42	
Air Blank	0.000	03:44	
Subject Sample	# 2	03:44	Subject Test Refused
Breath Volume	0.000	03:44	
Air Blank	0.000	03:44	
Control Test	0.081	03:45	
Air Blank	0.000	03:45	
Diagnostics Check	OK	03:45	

Data Download Date/Time: 10/26/2017 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOYE,JAMES
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: RAMOSRODRIGUEZ

DATE/TIME OF TEST: 10/21/2017 03:58:36

Diagnostics Check	OK	03:50
Air Blank	0.000	03:50
Control Test	0.082	03:50
Air Blank	0.000	03:51
Subject Sample # 1	0.161	03:54
Breath Volume	1.269	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.166	03:57
Breath Volume	1.566	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/26/2017 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KELEMEN,JANOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: KNOWLES

DATE/TIME OF TEST: 10/21/2017 04:25:47

Diagnostics Check	OK	04:19
Air Blank	0.000	04:19
Control Test	0.081	04:20
Air Blank	0.000	04:20
Subject Sample # 1	0.239	04:21
Breath Volume	3.968	04:21
Air Blank	0.000	04:21
Air Blank	0.000	04:23
Subject Sample # 2	0.236	04:24
Breath Volume	3.730	04:24
Air Blank	0.000	04:24
Control Test	0.080	04:25
Air Blank	0.000	04:25
Diagnostics Check	OK	04:25

Data Download Date/Time: 10/26/2017 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: TENORIONORENO

DATE/TIME OF TEST: 10/22/2017 01:31:51

Diagnostics Check		OK	01:25
Air Blank		0.000	01:26
Control Test		0.082	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.156	01:27
Breath Volume		2.660	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.158	01:30
Breath Volume		2.472	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 10/26/2017 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TORRES,NICOLE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: GASPAR

DATE/TIME OF TEST: 10/22/2017 04:06:12

Diagnostics Check	OK	04:00
Air Blank	0.000	04:00
Control Test	0.082	04:01
Air Blank	0.000	04:01
Subject Sample # 1	0.150	04:01
Breath Volume	1.406	04:01
Air Blank	0.000	04:02
Air Blank	0.000	04:04
Subject Sample # 2	0.149	04:04
Breath Volume	1.453	04:04
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Diagnostics Check	OK	04:06

Data Download Date/Time: 10/26/2017 17:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GIARRUSSO,AMY
ARRESTING OFFICER AGENCY: Winter Park PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 10/25/2017 01:37:44

Diagnostics Check		OK	01:26
Air Blank		0.000	01:26
Control Test		0.081	01:26
Air Blank		0.000	01:27
Subject Sample	# 1	0.231	01:27
Breath Volume		1.351	01:27
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.252	01:30
Breath Volume		2.441	01:30
Air Blank		0.000	01:31
Air Blank		0.000	01:33
Subject Sample	# 3	0.245	01:36
Breath Volume		2.738	01:36
Air Blank		0.000	01:36
Control Test		0.081	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 10/26/2017 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 10/26/2017 01:04:10

Diagnostics Check	OK	00:52	
Air Blank	0.000	00:53	
Control Test	0.082	00:53	
Air Blank	0.000	00:53	
Subject Sample	# 1	00:57	No Sample Provided
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Air Blank	0.000	00:59	
Subject Sample	# 2	01:02	Volume Not Met
Breath Volume	0.781	01:02	
Air Blank	0.000	01:03	
Control Test	0.082	01:03	
Air Blank	0.000	01:04	
Diagnostics Check	OK	01:04	

Data Download Date/Time: 10/26/2017 17:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MOORE,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: SCHWARTZ

DATE/TIME OF TEST: 10/26/2017 19:28:27

Diagnostics Check	OK	19:22
Air Blank	0.000	19:22
Control Test	0.083	19:22
Air Blank	0.000	19:23
Subject Sample # 1	0.179	19:23
Breath Volume	1.589	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:26
Subject Sample # 2	0.166	19:26
Breath Volume	2.179	19:26
Air Blank	0.000	19:27
Control Test	0.081	19:27
Air Blank	0.000	19:28
Diagnostics Check	OK	19:28

Data Download Date/Time: 11/02/2017 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SABAT GARCIA

DATE/TIME OF TEST: 10/28/2017 03:55:14

Diagnostics Check	OK	03:48
Air Blank	0.000	03:49
Control Test	0.082	03:49
Air Blank	0.000	03:50
Subject Sample # 1	0.119	03:50
Breath Volume	1.820	03:50
Air Blank	0.000	03:51
Air Blank	0.000	03:53
Subject Sample # 2	0.118	03:53
Breath Volume	2.257	03:53
Air Blank	0.000	03:54
Control Test	0.081	03:54
Air Blank	0.000	03:55
Diagnostics Check	OK	03:55

Data Download Date/Time: 11/02/2017 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:26

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 10/28/2017 22:58:11

Diagnostics Check	OK	22:51
Air Blank	0.000	22:51
Control Test	0.081	22:52
Air Blank	0.000	22:52
Subject Sample # 1	0.000	22:53
Breath Volume	1.691	22:53
Air Blank	0.000	22:53
Air Blank	0.000	22:55
Subject Sample # 2	0.000	22:56
Breath Volume	3.230	22:56
Air Blank	0.000	22:57
Control Test	0.082	22:57
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 11/02/2017 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: REARDON, SARAH
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:01

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 10/29/2017 04:37:40

Diagnostics Check	OK	04:26	
Air Blank	0.000	04:26	
Control Test	0.082	04:27	
Air Blank	0.000	04:27	
Subject Sample	# 1	04:30	No Sample Provided
Breath Volume	0.000	04:30	
Air Blank	0.000	04:31	
Air Blank	0.000	04:33	
Subject Sample	# 2	04:36	No Sample Provided
Breath Volume	0.000	04:36	
Air Blank	0.000	04:36	
Control Test	0.082	04:37	
Air Blank	0.000	04:37	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 11/02/2017 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSARIO SANTIAGO, ELDALY, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: TEMPLES, JUSTIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 10/30/2017 06:12:04

Diagnostics Check	OK	06:00	
Air Blank	0.000	06:01	
Control Test	0.083	06:01	
Air Blank	0.000	06:01	
Subject Sample	# 1	06:05	No Sample Provided
Breath Volume	0.000	06:05	
Air Blank	0.000	06:05	
Air Blank	0.000	06:07	
Subject Sample	# 2	06:10	No Sample Provided
Breath Volume	0.000	06:10	
Air Blank	0.000	06:11	
Control Test	0.082	06:11	
Air Blank	0.000	06:11	
Diagnostics Check	OK	06:12	

Data Download Date/Time: 11/02/2017 15:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000963

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLIDAY,LUCUS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: BIZJAK

DATE/TIME OF TEST: 10/31/2017 21:50:17

Diagnostics Check	OK	21:47	
Air Blank	0.000	21:47	
Control Test	0.081	21:48	
Air Blank	0.000	21:48	
Subject Sample	# 1	21:48	Subject Test Refused
Breath Volume	0.000	21:48	
Air Blank	0.000	21:49	
Control Test	0.082	21:49	
Air Blank	0.000	21:50	
Diagnostics Check	OK	21:50	

Data Download Date/Time: 11/02/2017 15:02

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: PARRA

DATE/TIME OF TEST: 10/20/2017 02:34:04

Diagnostics Check	OK	02:27
Air Blank	0.000	02:28
Control Test	0.079	02:28
Air Blank	0.000	02:29
Subject Sample # 1	0.174	02:29
Breath Volume	2.023	02:29
Air Blank	0.000	02:30
Air Blank	0.000	02:32
Subject Sample # 2	0.173	02:32
Breath Volume	1.710	02:32
Air Blank	0.000	02:33
Control Test	0.078	02:33
Air Blank	0.000	02:33
Diagnostics Check	OK	02:34

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: NUCUP

DATE/TIME OF TEST: 10/21/2017 04:13:00

Diagnostics Check	OK	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:08
Subject Sample # 1	0.175	04:08
Breath Volume	3.121	04:08
Air Blank	0.000	04:09
Air Blank	0.000	04:11
Subject Sample # 2	0.179	04:11
Breath Volume	3.070	04:11
Air Blank	0.000	04:12
Control Test	0.079	04:12
Air Blank	0.000	04:12
Diagnostics Check	OK	04:13

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 10/21/2017 06:02:30

Diagnostics Check	OK	05:56
Air Blank	0.000	05:56
Control Test	0.079	05:57
Air Blank	0.000	05:57
Subject Sample # 1	0.000	05:58
Breath Volume	2.429	05:58
Air Blank	0.000	05:58
Air Blank	0.000	06:00
Subject Sample # 2	0.000	06:01
Breath Volume	2.800	06:01
Air Blank	0.000	06:01
Control Test	0.079	06:01
Air Blank	0.000	06:02
Diagnostics Check	OK	06:02

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN, JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:34

SUBJECT'S LAST NAME: HUERTAS-BAEZ

DATE/TIME OF TEST: 10/21/2017 07:01:23

Diagnostics Check	OK	06:55
Air Blank	0.000	06:56
Control Test	0.079	06:56
Air Blank	0.000	06:56
Subject Sample # 1	0.125	06:57
Breath Volume	1.316	06:57
Air Blank	0.000	06:57
Air Blank	0.000	06:59
Subject Sample # 2	0.129	06:59
Breath Volume	1.234	06:59
Air Blank	0.000	07:00
Control Test	0.080	07:00
Air Blank	0.000	07:01
Diagnostics Check	OK	07:01

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL, CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: BARAJASBARRAGAN

DATE/TIME OF TEST: 10/21/2017 08:01:39

Diagnostics Check	OK	07:50	
Air Blank	0.000	07:50	
Control Test	0.080	07:51	
Air Blank	0.000	07:51	
Subject Sample	# 1	07:54	No Sample Provided
Breath Volume	0.000	07:54	
Air Blank	0.000	07:55	
Air Blank	0.000	07:57	
Subject Sample	# 2	08:00	No Sample Provided
Breath Volume	0.000	08:00	
Air Blank	0.000	08:00	
Control Test	0.080	08:01	
Air Blank	0.000	08:01	
Diagnostics Check	OK	08:01	

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CHAPLIN, CHRIS
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:05

SUBJECT'S LAST NAME: SILVA

DATE/TIME OF TEST: 10/21/2017 16:52:46

Diagnostics Check	OK	16:46
Air Blank	0.000	16:47
Control Test	0.079	16:47
Air Blank	0.000	16:47
Subject Sample # 1	0.233	16:48
Breath Volume	1.878	16:48
Air Blank	0.000	16:49
Air Blank	0.000	16:50
Subject Sample # 2	0.246	16:51
Breath Volume	1.703	16:51
Air Blank	0.000	16:51
Control Test	0.080	16:52
Air Blank	0.000	16:52
Diagnostics Check	OK	16:52

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: QUESENBERRY

DATE/TIME OF TEST: 10/22/2017 02:00:19

Diagnostics Check	OK	01:54
Air Blank	0.000	01:54
Control Test	0.079	01:54
Air Blank	0.000	01:55
Subject Sample # 1	0.210	01:55
Breath Volume	1.468	01:55
Air Blank	0.000	01:56
Air Blank	0.000	01:58
Subject Sample # 2	0.204	01:58
Breath Volume	2.316	01:58
Air Blank	0.000	01:59
Control Test	0.079	01:59
Air Blank	0.000	02:00
Diagnostics Check	OK	02:00

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SPEER,ANDREW
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SWANSON

DATE/TIME OF TEST: 10/22/2017 03:58:00

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.079	03:52
Air Blank	0.000	03:53
Subject Sample # 1	0.172	03:53
Breath Volume	2.214	03:53
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.182	03:56
Breath Volume	2.261	03:56
Air Blank	0.000	03:57
Control Test	0.079	03:57
Air Blank	0.000	03:57
Diagnostics Check	OK	03:58

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HOLIDAY,LUCAS
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:50

SUBJECT'S LAST NAME: VALENCIA-CAUPER

DATE/TIME OF TEST: 10/22/2017 07:24:42

Diagnostics Check		OK	07:13	
Air Blank		0.000	07:13	
Control Test		0.080	07:14	
Air Blank		0.000	07:14	
Subject Sample	# 1	0.107	07:17	Volume Not Met
Breath Volume		0.050	07:17	
Air Blank		0.000	07:18	
Air Blank		0.000	07:20	
Subject Sample	# 2	0.111	07:23	Volume Not Met
Breath Volume		0.015	07:23	
Air Blank		0.000	07:23	
Control Test		0.080	07:24	
Air Blank		0.000	07:24	
Diagnostics Check		OK	07:24	

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MICHAEL, J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CLARK, ANDREW
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:37

SUBJECT'S LAST NAME: BARRERA

DATE/TIME OF TEST: 10/22/2017 08:11:32

Diagnostics Check	OK	08:05
Air Blank	0.000	08:05
Control Test	0.079	08:06
Air Blank	0.000	08:06
Subject Sample # 1	0.075	08:07
Breath Volume	2.328	08:07
Air Blank	0.000	08:07
Air Blank	0.000	08:09
Subject Sample # 2	0.076	08:10
Breath Volume	2.187	08:10
Air Blank	0.000	08:10
Control Test	0.079	08:10
Air Blank	0.000	08:11
Diagnostics Check	OK	08:11

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,JOSE
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:44

SUBJECT'S LAST NAME: PECHACEK

DATE/TIME OF TEST: 10/22/2017 08:35:40

Diagnostics Check	OK	08:29
Air Blank	0.000	08:29
Control Test	0.080	08:30
Air Blank	0.000	08:30
Subject Sample # 1	0.079	08:31
Breath Volume	1.187	08:31
Air Blank	0.000	08:31
Air Blank	0.000	08:33
Subject Sample # 2	0.086	08:34
Breath Volume	2.171	08:34
Air Blank	0.000	08:34
Control Test	0.080	08:35
Air Blank	0.000	08:35
Diagnostics Check	OK	08:35

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:41

SUBJECT'S LAST NAME: DE VLUGT

DATE/TIME OF TEST: 10/22/2017 21:10:48

Diagnostics Check	OK	21:04
Air Blank	0.000	21:05
Control Test	0.078	21:05
Air Blank	0.000	21:06
Subject Sample # 1	0.195	21:06
Breath Volume	1.406	21:06
Air Blank	0.000	21:07
Air Blank	0.000	21:09
Subject Sample # 2	0.196	21:09
Breath Volume	1.371	21:09
Air Blank	0.000	21:09
Control Test	0.079	21:10
Air Blank	0.000	21:10
Diagnostics Check	OK	21:10

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,ANDREW
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: KINZEL

DATE/TIME OF TEST: 10/24/2017 03:16:41

Diagnostics Check	OK	03:10
Air Blank	0.000	03:11
Control Test	0.080	03:11
Air Blank	0.000	03:11
Subject Sample # 1	0.000	03:12
Breath Volume	1.242	03:12
Air Blank	0.000	03:12
Air Blank	0.000	03:14
Subject Sample # 2	0.000	03:15
Breath Volume	2.195	03:15
Air Blank	0.000	03:15
Control Test	0.080	03:16
Air Blank	0.000	03:16
Diagnostics Check	OK	03:16

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WADEMAN,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:23

SUBJECT'S LAST NAME: POLEC

DATE/TIME OF TEST: 10/26/2017 14:55:45

Diagnostics Check		OK	14:49
Air Blank		0.000	14:50
Control Test		0.080	14:50
Air Blank		0.000	14:50
Subject Sample	# 1	0.333	14:51
Breath Volume		1.917	14:51
Air Blank		0.000	14:52
Air Blank		0.000	14:53
Subject Sample	# 2	0.316	14:54
Breath Volume		1.312	14:54
Air Blank		0.000	14:54
Control Test		0.080	14:55
Air Blank		0.000	14:55
Diagnostics Check		OK	14:55

Data Download Date/Time: 10/26/2017 17:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VENTURA,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: WELCH

DATE/TIME OF TEST: 10/27/2017 03:27:25

Diagnostics Check	OK	03:20
Air Blank	0.000	03:20
Control Test	0.079	03:21
Air Blank	0.000	03:21
Subject Sample # 1	0.211	03:22
Breath Volume	2.035	03:22
Air Blank	0.000	03:23
Air Blank	0.000	03:25
Subject Sample # 2	0.216	03:25
Breath Volume	3.121	03:25
Air Blank	0.000	03:26
Control Test	0.080	03:26
Air Blank	0.000	03:27
Diagnostics Check	OK	03:27

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS, DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: OSBORNE

DATE/TIME OF TEST: 10/27/2017 04:34:48

Diagnostics Check	OK	04:28
Air Blank	0.000	04:29
Control Test	0.079	04:29
Air Blank	0.000	04:29
Subject Sample # 1	0.226	04:30
Breath Volume	2.140	04:30
Air Blank	0.000	04:30
Air Blank	0.000	04:32
Subject Sample # 2	0.228	04:33
Breath Volume	2.996	04:33
Air Blank	0.000	04:33
Control Test	0.079	04:34
Air Blank	0.000	04:34
Diagnostics Check	OK	04:34

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HUDSON,CODY
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:33

SUBJECT'S LAST NAME: WALAS

DATE/TIME OF TEST: 10/27/2017 23:07:30

Diagnostics Check	OK	23:00
Air Blank	0.000	23:00
Control Test	0.080	23:00
Air Blank	0.000	23:01
Subject Sample # 1	0.173	23:02
Breath Volume	1.613	23:02
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.165	23:06
Breath Volume	1.699	23:06
Air Blank	0.000	23:06
Control Test	0.079	23:06
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BELARDES,DAVID
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: GERHARDT

DATE/TIME OF TEST: 10/28/2017 02:57:30

Diagnostics Check	OK	02:51
Air Blank	0.000	02:51
Control Test	0.079	02:51
Air Blank	0.000	02:52
Subject Sample # 1	0.184	02:52
Breath Volume	3.570	02:52
Air Blank	0.000	02:53
Air Blank	0.000	02:55
Subject Sample # 2	0.180	02:56
Breath Volume	3.761	02:56
Air Blank	0.000	02:56
Control Test	0.080	02:56
Air Blank	0.000	02:57
Diagnostics Check	OK	02:57

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON, TRAVIS, W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARTINEZ, ERNESTO
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 10/28/2017 04:05:23

Diagnostics Check	OK	03:59
Air Blank	0.000	03:59
Control Test	0.081	04:00
Air Blank	0.000	04:00
Subject Sample # 1	0.100	04:01
Breath Volume	2.960	04:01
Air Blank	0.000	04:01
Air Blank	0.000	04:03
Subject Sample # 2	0.102	04:03
Breath Volume	3.152	04:03
Air Blank	0.000	04:04
Control Test	0.080	04:04
Air Blank	0.000	04:05
Diagnostics Check	OK	04:05

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: ECHEVERRI

DATE/TIME OF TEST: 10/29/2017 05:08:08

Diagnostics Check	OK	05:01
Air Blank	0.000	05:02
Control Test	0.080	05:02
Air Blank	0.000	05:03
Subject Sample # 1	0.194	05:03
Breath Volume	2.019	05:03
Air Blank	0.000	05:04
Air Blank	0.000	05:06
Subject Sample # 2	0.194	05:06
Breath Volume	2.187	05:06
Air Blank	0.000	05:07
Control Test	0.079	05:07
Air Blank	0.000	05:08
Diagnostics Check	OK	05:08

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: ENIOLA

DATE/TIME OF TEST: 10/29/2017 07:10:35

Diagnostics Check	OK	07:01
Air Blank	0.000	07:02
Control Test	0.080	07:02
Air Blank	0.000	07:02
Subject Sample # 1	0.165	07:05
Breath Volume	1.136	07:05
Air Blank	0.000	07:06
Air Blank	0.000	07:08
Subject Sample # 2	0.163	07:09
Breath Volume	1.148	07:09
Air Blank	0.000	07:09
Control Test	0.080	07:10
Air Blank	0.000	07:10
Diagnostics Check	OK	07:10

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORTON,CODY
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: BUDD

DATE/TIME OF TEST: 10/31/2017 01:01:38

Diagnostics Check	OK	00:55
Air Blank	0.000	00:56
Control Test	0.079	00:56
Air Blank	0.000	00:56
Subject Sample # 1	0.142	00:57
Breath Volume	4.257	00:57
Air Blank	0.000	00:57
Air Blank	0.000	00:59
Subject Sample # 2	0.136	01:00
Breath Volume	4.281	01:00
Air Blank	0.000	01:00
Control Test	0.080	01:01
Air Blank	0.000	01:01
Diagnostics Check	OK	01:01

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MACFAWN,BRITTANY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: KIMPLING

DATE/TIME OF TEST: 10/31/2017 06:19:26

Diagnostics Check	OK	06:13
Air Blank	0.000	06:13
Control Test	0.080	06:13
Air Blank	0.000	06:14
Subject Sample # 1	0.171	06:14
Breath Volume	3.128	06:14
Air Blank	0.000	06:15
Air Blank	0.000	06:17
Subject Sample # 2	0.169	06:17
Breath Volume	3.480	06:17
Air Blank	0.000	06:18
Control Test	0.080	06:18
Air Blank	0.000	06:19
Diagnostics Check	OK	06:19

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000964

INSTRUMENT REGISTERED TO: ORANGE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELVILLE,KELLY,M
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GOMES,JERIEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 10/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:12

SUBJECT'S LAST NAME: AQUINOAPONTE

DATE/TIME OF TEST: 10/31/2017 11:49:11

Diagnostics Check	OK	11:42
Air Blank	0.000	11:43
Control Test	0.080	11:43
Air Blank	0.000	11:43
Subject Sample # 1	0.000	11:44
Breath Volume	1.164	11:44
Air Blank	0.000	11:45
Air Blank	0.000	11:47
Subject Sample # 2	0.000	11:47
Breath Volume	1.843	11:47
Air Blank	0.000	11:48
Control Test	0.080	11:48
Air Blank	0.000	11:49
Diagnostics Check	OK	11:49

Data Download Date/Time: 11/02/2017 15:05

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: FINUFF,DANIEL,W

BREATH TEST OPERATOR AGENCY: FHP Troop H

ARRESTING OFFICER: FINUFF,DANIEL

ARRESTING OFFICER AGENCY: FHP Troop H

LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 34015080A2

DRY GAS EXPIRY DATE: 01/05/2018

VIOLATION CODE: DUI

ARREST TIME: 01:45

GENDER: Male

OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: VELASQUEZ HERNANDEZ

DATE/TIME OF TEST: 10/02/2017 02:19:32

Diagnostics Check	OK	02:12
Air Blank	0.000	02:13
Control Test	0.077	02:13
Air Blank	0.000	02:14
Subject Sample # 1	0.161	02:14
Breath Volume	1.781	02:14
Air Blank	0.000	02:15
Air Blank	0.000	02:17
Subject Sample # 2	0.163	02:18
Breath Volume	2.636	02:18
Air Blank	0.000	02:18
Control Test	0.079	02:18
Air Blank	0.000	02:19
Diagnostics Check	OK	02:19

Data Download Date/Time: 10/04/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 09/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: GALANTE

DATE/TIME OF TEST: 10/02/2017 23:19:32

Diagnostics Check	OK	23:13
Air Blank	0.000	23:13
Control Test	0.077	23:13
Air Blank	0.000	23:14
Subject Sample # 1	0.034	23:15
Breath Volume	2.015	23:15
Air Blank	0.000	23:15
Air Blank	0.000	23:17
Subject Sample # 2	0.035	23:18
Breath Volume	2.835	23:18
Air Blank	0.000	23:18
Control Test	0.078	23:18
Air Blank	0.000	23:19
Diagnostics Check	OK	23:19

Data Download Date/Time: 10/04/2017 11:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: FOLEY

DATE/TIME OF TEST: 10/05/2017 21:44:39

Diagnostics Check	OK	21:38
Air Blank	0.000	21:38
Control Test	0.077	21:39
Air Blank	0.000	21:39
Subject Sample # 1	0.135	21:40
Breath Volume	2.671	21:40
Air Blank	0.000	21:40
Air Blank	0.000	21:42
Subject Sample # 2	0.142	21:43
Breath Volume	2.660	21:43
Air Blank	0.000	21:43
Control Test	0.078	21:44
Air Blank	0.000	21:44
Diagnostics Check	OK	21:44

Data Download Date/Time: 11/27/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LAZZARI,RYAN
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 10/06/2017 20:20:36

Diagnostics Check	OK	20:17	
Air Blank	0.000	20:17	
Control Test	0.078	20:18	
Air Blank	0.000	20:18	
Subject Sample	# 1	20:19	Slope Not Met
Breath Volume	0.968	20:19	
Air Blank	0.000	20:19	
Control Test	0.077	20:20	
Air Blank	0.000	20:20	
Diagnostics Check	OK	20:20	

Data Download Date/Time: 11/27/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: LAZZARI,RYAN
ARRESTING OFFICER AGENCY: Quincy PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 10/06/2017 20:29:02

Diagnostics Check		OK	20:22
Air Blank		0.000	20:22
Control Test		0.077	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.206	20:24
Breath Volume		1.527	20:24
Air Blank		0.000	20:25
Air Blank		0.000	20:26
Subject Sample	# 2	0.210	20:27
Breath Volume		1.515	20:27
Air Blank		0.000	20:28
Control Test		0.077	20:28
Air Blank		0.000	20:28
Diagnostics Check		OK	20:29

Data Download Date/Time: 11/27/2017 10:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: MCKNIGHT

DATE/TIME OF TEST: 10/06/2017 21:17:13

Diagnostics Check	OK	21:10
Air Blank	0.000	21:11
Control Test	0.077	21:11
Air Blank	0.000	21:12
Subject Sample # 1	0.179	21:12
Breath Volume	1.730	21:12
Air Blank	0.000	21:13
Air Blank	0.000	21:15
Subject Sample # 2	0.184	21:15
Breath Volume	2.218	21:15
Air Blank	0.000	21:16
Control Test	0.077	21:16
Air Blank	0.000	21:17
Diagnostics Check	OK	21:17

Data Download Date/Time: 11/27/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FINUFF,DANIEL,W
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: FINUFF,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 10/07/2017 21:23:31

Diagnostics Check	OK	21:17
Air Blank	0.000	21:17
Control Test	0.077	21:18
Air Blank	0.000	21:18
Subject Sample # 1	0.188	21:19
Breath Volume	1.824	21:19
Air Blank	0.000	21:19
Air Blank	0.000	21:21
Subject Sample # 2	0.178	21:22
Breath Volume	2.640	21:22
Air Blank	0.000	21:22
Control Test	0.078	21:22
Air Blank	0.000	21:23
Diagnostics Check	OK	21:23

Data Download Date/Time: 11/27/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,JAMES,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: ALLEN,JAMES
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34015080A2
DRY GAS EXPIRY DATE: 01/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:58

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 10/20/2017 20:31:08

Diagnostics Check		OK	20:18	
Air Blank		0.000	20:19	
Control Test		0.077	20:19	
Air Blank		0.000	20:20	
Subject Sample	# 1		20:23	No Sample Provided
Breath Volume		0.000	20:23	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 2	0.218	20:26	No .020 agreement
Breath Volume		2.082	20:26	
Air Blank		0.000	20:27	
Air Blank		0.000	20:28	
Subject Sample	# 3	0.195	20:29	No .020 agreement
Breath Volume		2.320	20:29	
Air Blank		0.000	20:30	
Control Test		0.077	20:30	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 11/27/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,CHARLES,E

BREATH TEST OPERATOR AGENCY: Gretna PD

ARRESTING OFFICER: GONZALES,IVAN

ARRESTING OFFICER AGENCY: Gretna PD

LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: Yes

DRY GAS STANDARD LOT #: 14816080A3

DRY GAS EXPIRY DATE: 07/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:54

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: DILWORTH

DATE/TIME OF TEST: 10/26/2017 21:38:05

Diagnostics Check	OK	21:32	
Air Blank	0.000	21:32	
Control Test	0.081	21:32	
Air Blank	0.000	21:33	
Subject Sample	# 1 0.219	21:36	Volume Not Met
Breath Volume	0.312	21:36	
Air Blank		21:37	Purge Fail
Air Blank		21:38	Purge Fail

Data Download Date/Time: 11/27/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: BROWN,CHARLES,E

BREATH TEST OPERATOR AGENCY: Gretna PD

ARRESTING OFFICER: GONZALES,IVAN

ARRESTING OFFICER AGENCY: Gretna PD

LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 14816080A3

DRY GAS EXPIRY DATE: 07/05/2018

VIOLATION CODE: DUI

ARREST TIME: 20:54

GENDER: Female

OBSERVATION PERIOD BEGAN: 21:03

SUBJECT'S LAST NAME: DILWORTH

DATE/TIME OF TEST: 10/26/2017 21:43:38

Diagnostics Check	OK	21:40	
Air Blank	0.000	21:41	
Control Test	0.079	21:41	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:42	Subject Test Refused
Breath Volume	0.000	21:42	
Air Blank	0.000	21:42	
Control Test	0.080	21:43	
Air Blank	0.000	21:43	
Diagnostics Check	OK	21:43	

Data Download Date/Time: 11/27/2017 10:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SGT R,N
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: WILLIAMS,SGT. R
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 10/27/2017 23:38:45

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.081	23:33	
Air Blank	0.000	23:34	
Subject Sample # 1	0.045	23:35	
Breath Volume	1.757	23:35	
Air Blank	0.000	23:35	
Air Blank		23:37	Improper Sample
Air Blank	0.000	23:37	
Control Test	0.082	23:38	
Air Blank	0.000	23:38	
Diagnostics Check	OK	23:38	

Data Download Date/Time: 11/27/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000965

INSTRUMENT REGISTERED TO: QUINCY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,SGT. R,N
BREATH TEST OPERATOR AGENCY: Midway PD
ARRESTING OFFICER: WILLIAMS,SGT. R
ARRESTING OFFICER AGENCY: Midway PD
LAST AGENCY INSPECTION: 10/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 10/27/2017 23:46:58

Diagnostics Check	OK	23:40
Air Blank	0.000	23:41
Control Test	0.080	23:41
Air Blank	0.000	23:42
Subject Sample # 1	0.043	23:42
Breath Volume	2.199	23:42
Air Blank	0.000	23:43
Air Blank	0.000	23:45
Subject Sample # 2	0.040	23:45
Breath Volume	2.554	23:45
Air Blank	0.000	23:46
Control Test	0.081	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 11/27/2017 10:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:57

SUBJECT'S LAST NAME: ORREA

DATE/TIME OF TEST: 10/01/2017 05:39:47

Diagnostics Check	OK	05:33
Air Blank	0.000	05:33
Control Test	0.080	05:34
Air Blank	0.000	05:34
Subject Sample # 1	0.234	05:35
Breath Volume	1.617	05:35
Air Blank	0.000	05:35
Air Blank	0.000	05:37
Subject Sample # 2	0.216	05:38
Breath Volume	1.484	05:38
Air Blank	0.000	05:38
Control Test	0.078	05:39
Air Blank	0.000	05:39
Diagnostics Check	OK	05:39

Data Download Date/Time: 10/18/2017 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA,EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: REAVIS

DATE/TIME OF TEST: 10/02/2017 01:34:02

Diagnostics Check	OK	01:31	
Air Blank	0.000	01:32	
Control Test	0.080	01:32	
Air Blank	0.000	01:32	
Subject Sample	# 1	01:33	Slope Not Met
Breath Volume	1.054	01:33	
Air Blank		01:33	Improper Sample
Air Blank	0.000	01:34	

Data Download Date/Time: 10/18/2017 20:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVA,EMILIANO,R
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SILVA,EMILIANO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: REAVIS

DATE/TIME OF TEST: 10/02/2017 01:42:18

Diagnostics Check	OK	01:36
Air Blank	0.000	01:36
Control Test	0.078	01:36
Air Blank	0.000	01:37
Subject Sample # 1	0.408	01:37
Breath Volume	3.433	01:37
Air Blank	0.000	01:38
Air Blank	0.000	01:40
Subject Sample # 2	0.426	01:40
Breath Volume	3.417	01:40
Air Blank	0.000	01:41
Control Test	0.078	01:41
Air Blank	0.000	01:42
Diagnostics Check	OK	01:42

Data Download Date/Time: 10/18/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYLE,OWEN,D
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LAWSON,CASEY
ARRESTING OFFICER AGENCY: St. Cloud PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:50

SUBJECT'S LAST NAME: BONILLA MATIAS

DATE/TIME OF TEST: 10/05/2017 06:47:23

Diagnostics Check	OK	06:40
Air Blank	0.000	06:40
Control Test	0.080	06:40
Air Blank	0.000	06:41
Subject Sample # 1	0.159	06:42
Breath Volume	1.468	06:42
Air Blank	0.000	06:43
Air Blank	0.000	06:44
Subject Sample # 2	0.152	06:45
Breath Volume	1.316	06:45
Air Blank	0.000	06:46
Control Test	0.079	06:46
Air Blank	0.000	06:47
Diagnostics Check	OK	06:47

Data Download Date/Time: 10/18/2017 20:39

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTONE,LACEY,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: KING,SHADRACH
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SIERRA AREVALO

DATE/TIME OF TEST: 10/06/2017 23:07:36

Diagnostics Check	OK	23:00
Air Blank	0.000	23:01
Control Test	0.081	23:01
Air Blank	0.000	23:02
Subject Sample # 1	0.108	23:03
Breath Volume	1.746	23:03
Air Blank	0.000	23:03
Air Blank	0.000	23:05
Subject Sample # 2	0.097	23:06
Breath Volume	2.078	23:06
Air Blank	0.000	23:06
Control Test	0.079	23:07
Air Blank	0.000	23:07
Diagnostics Check	OK	23:07

Data Download Date/Time: 10/18/2017 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHARLES,F
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: THOMAS,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: RODRIGUEZ PADILLA

DATE/TIME OF TEST: 10/07/2017 12:45:09

Diagnostics Check	OK	12:39
Air Blank	0.000	12:39
Control Test	0.081	12:40
Air Blank	0.000	12:40
Subject Sample # 1	0.000	12:40
Breath Volume	2.183	12:40
Air Blank	0.000	12:41
Air Blank	0.000	12:43
Subject Sample # 2	0.000	12:43
Breath Volume	1.402	12:43
Air Blank	0.000	12:44
Control Test	0.079	12:44
Air Blank	0.000	12:45
Diagnostics Check	OK	12:45

Data Download Date/Time: 10/18/2017 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JACOB,C
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: REYES-SOTO,RICHARD
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 09/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 10/11/2017 20:29:20

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.080	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.234	20:23
Breath Volume	1.871	20:23
Air Blank	0.000	20:24
Air Blank	0.000	20:26
Subject Sample # 2	0.249	20:27
Breath Volume	2.300	20:27
Air Blank	0.000	20:28
Control Test	0.079	20:28
Air Blank	0.000	20:29
Diagnostics Check	OK	20:29

Data Download Date/Time: 10/18/2017 20:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,BRIAN,F
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DECKER,DAVID
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 07:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:22

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 10/20/2017 08:49:19

Diagnostics Check	OK	08:42
Air Blank	0.000	08:43
Control Test	0.079	08:43
Air Blank	0.000	08:44
Subject Sample # 1	0.097	08:44
Breath Volume	1.921	08:44
Air Blank	0.000	08:45
Air Blank	0.000	08:46
Subject Sample # 2	0.096	08:47
Breath Volume	1.535	08:47
Air Blank	0.000	08:48
Control Test	0.079	08:48
Air Blank	0.000	08:49
Diagnostics Check	OK	08:49

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTONE,LACEY,M
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: PHILLIPS,ADAM
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 04:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: PACHECO SANTIAGO

DATE/TIME OF TEST: 10/21/2017 04:54:35

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:43	
Control Test		0.079	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.177	04:47	Volume Not Met
Breath Volume		0.605	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 2	0.177	04:53	Volume Not Met
Breath Volume		0.460	04:53	
Air Blank		0.000	04:53	
Control Test		0.077	04:53	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: REYES,I
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:21

SUBJECT'S LAST NAME: CAPUTO

DATE/TIME OF TEST: 10/23/2017 21:49:50

Diagnostics Check	OK	21:42
Air Blank	0.000	21:42
Control Test	0.079	21:43
Air Blank	0.000	21:43
Subject Sample # 1	0.229	21:45
Breath Volume	2.078	21:45
Air Blank	0.000	21:46
Air Blank	0.000	21:47
Subject Sample # 2	0.217	21:48
Breath Volume	1.738	21:48
Air Blank	0.000	21:48
Control Test	0.079	21:49
Air Blank	0.000	21:49
Diagnostics Check	OK	21:49

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:12

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 10/24/2017 21:46:39

Diagnostics Check	OK	21:35	
Air Blank	0.000	21:35	
Control Test	0.080	21:36	
Air Blank	0.000	21:36	
Subject Sample	# 1 0.250	21:39	Volume Not Met
Breath Volume	0.148	21:39	
Air Blank	0.000	21:40	
Air Blank	0.000	21:42	
Subject Sample	# 2	21:45	No Sample Provided
Breath Volume	0.000	21:45	
Air Blank	0.000	21:45	
Control Test	0.076	21:46	
Air Blank	0.000	21:46	
Diagnostics Check	OK	21:46	

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 10/24/2017 22:19:12

Diagnostics Check		OK	22:04	
Air Blank		0.000	22:04	
Control Test		0.080	22:05	
Air Blank		0.000	22:05	
Subject Sample	# 1	0.272	22:06	
Breath Volume		2.410	22:06	
Air Blank		0.000	22:06	
Air Blank		0.000	22:08	
Subject Sample	# 2	0.246	22:11	Volume Not Met
Breath Volume		0.160	22:11	
Air Blank		0.000	22:12	
Air Blank		0.000	22:14	
Subject Sample	# 3	0.194	22:17	Volume Not Met
Breath Volume		0.398	22:17	
Air Blank		0.000	22:18	
Control Test		0.076	22:18	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,D
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: GOODMAN

DATE/TIME OF TEST: 10/24/2017 22:31:09

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.250	22:23
Breath Volume	1.363	22:23
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.240	22:29
Breath Volume	1.593	22:29
Air Blank	0.000	22:30
Control Test	0.078	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,D,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: OCHOA,Y
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:46

SUBJECT'S LAST NAME: GALVEZ

DATE/TIME OF TEST: 10/25/2017 00:19:35

Diagnostics Check	OK	00:13
Air Blank	0.000	00:13
Control Test	0.079	00:14
Air Blank	0.000	00:14
Subject Sample # 1	0.175	00:15
Breath Volume	2.511	00:15
Air Blank	0.000	00:15
Air Blank	0.000	00:17
Subject Sample # 2	0.173	00:17
Breath Volume	2.437	00:17
Air Blank	0.000	00:18
Control Test	0.078	00:18
Air Blank	0.000	00:19
Diagnostics Check	OK	00:19

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEEKMAN,JOSHUA,J
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: BEEKMAN,JOSHUA
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 07:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 10/29/2017 08:28:26

Diagnostics Check	OK	08:21
Air Blank	0.000	08:22
Control Test	0.080	08:22
Air Blank	0.000	08:22
Subject Sample # 1	0.030	08:23
Breath Volume	1.953	08:23
Air Blank	0.000	08:24
Air Blank	0.000	08:26
Subject Sample # 2	0.026	08:26
Breath Volume	1.476	08:26
Air Blank	0.000	08:27
Control Test	0.079	08:27
Air Blank	0.000	08:28
Diagnostics Check	OK	08:28

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORUJO,PEDRO,L
BREATH TEST OPERATOR AGENCY: Kissimmee PD
ARRESTING OFFICER: CORUJO,PEDRO
ARRESTING OFFICER AGENCY: Kissimmee PD
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: LUPIAN

DATE/TIME OF TEST: 10/30/2017 00:48:56

Diagnostics Check	OK	00:45	
Air Blank	0.000	00:46	
Control Test	0.081	00:46	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:48	
Diagnostics Check	OK	00:48	

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000966

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: GANDULLA FIGUEROA

DATE/TIME OF TEST: 10/30/2017 03:54:17

Diagnostics Check	OK	03:47
Air Blank	0.000	03:48
Control Test	0.082	03:48
Air Blank	0.000	03:49
Subject Sample # 1	0.051	03:49
Breath Volume	2.914	03:49
Air Blank	0.000	03:50
Air Blank	0.000	03:52
Subject Sample # 2	0.051	03:52
Breath Volume	1.808	03:52
Air Blank	0.000	03:53
Control Test	0.083	03:53
Air Blank	0.000	03:54
Diagnostics Check	OK	03:54

Data Download Date/Time: 11/07/2017 18:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000968

INSTRUMENT REGISTERED TO: OSCEOLA CORRECTIONS
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PYE,CHRISTOPHER,T
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MARMOL,EUNICE
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 10/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 10/20/2017 02:30:51

Diagnostics Check	OK	02:23
Air Blank	0.000	02:24
Control Test	0.081	02:24
Air Blank	0.000	02:25
Subject Sample # 1	0.160	02:26
Breath Volume	4.066	02:26
Air Blank	0.000	02:27
Air Blank	0.000	02:28
Subject Sample # 2	0.167	02:29
Breath Volume	3.917	02:29
Air Blank	0.000	02:29
Control Test	0.079	02:30
Air Blank	0.000	02:30
Diagnostics Check	OK	02:30

Data Download Date/Time: 11/07/2017 18:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: LENGFELD,JESSE
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MACULEWICH

DATE/TIME OF TEST: 10/02/2017 21:47:57

Diagnostics Check	OK	21:41
Air Blank	0.000	21:41
Control Test	0.080	21:42
Air Blank	0.000	21:42
Subject Sample # 1	0.189	21:43
Breath Volume	3.484	21:43
Air Blank	0.000	21:44
Air Blank	0.000	21:45
Subject Sample # 2	0.180	21:46
Breath Volume	2.843	21:46
Air Blank	0.000	21:47
Control Test	0.079	21:47
Air Blank	0.000	21:47
Diagnostics Check	OK	21:47

Data Download Date/Time: 10/23/2017 11:57

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,JASON,M
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: SANCHEZ,JASON
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:22

SUBJECT'S LAST NAME: FEENEY

DATE/TIME OF TEST: 10/07/2017 14:50:57

Diagnostics Check	OK	14:44
Air Blank	0.000	14:45
Control Test	0.079	14:45
Air Blank	0.000	14:45
Subject Sample # 1	0.000	14:46
Breath Volume	1.671	14:46
Air Blank	0.000	14:47
Air Blank	0.000	14:49
Subject Sample # 2	0.000	14:49
Breath Volume	1.394	14:49
Air Blank	0.000	14:50
Control Test	0.079	14:50
Air Blank	0.000	14:50
Diagnostics Check	OK	14:50

Data Download Date/Time: 10/23/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDOLE,GEOFFERY,E
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCDOLE,GEOFFERY
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: TURBERVILLE

DATE/TIME OF TEST: 10/11/2017 20:18:06

Diagnostics Check	OK	20:11
Air Blank	0.000	20:12
Control Test	0.079	20:12
Air Blank	0.000	20:13
Subject Sample # 1	0.148	20:13
Breath Volume	2.824	20:13
Air Blank	0.000	20:14
Air Blank	0.000	20:16
Subject Sample # 2	0.143	20:16
Breath Volume	2.804	20:16
Air Blank	0.000	20:17
Control Test	0.079	20:17
Air Blank	0.000	20:18
Diagnostics Check	OK	20:18

Data Download Date/Time: 10/23/2017 11:57

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000996

INSTRUMENT REGISTERED TO: TITUSVILLE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRORY,DIARMUID,J
BREATH TEST OPERATOR AGENCY: Titusville PD
ARRESTING OFFICER: MCGRORY,DIARMUID
ARRESTING OFFICER AGENCY: Titusville PD
LAST AGENCY INSPECTION: 09/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 10/15/2017 03:07:14

Diagnostics Check		OK	03:00
Air Blank		0.000	03:01
Control Test		0.079	03:01
Air Blank		0.000	03:02
Subject Sample	# 1	0.204	03:02
Breath Volume		1.667	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.204	03:05
Breath Volume		1.304	03:05
Air Blank		0.000	03:06
Control Test		0.079	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 10/23/2017 11:57