

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: DEAN

DATE/TIME OF TEST: 12/02/2017 04:37:02

Diagnostics Check	OK	04:34	
Air Blank	0.000	04:34	
Control Test	0.079	04:34	
Air Blank	0.000	04:35	
Subject Sample	# 1	04:35	Subject Test Refused
Breath Volume	0.000	04:35	
Air Blank	0.000	04:36	
Control Test	0.080	04:36	
Air Blank	0.000	04:36	
Diagnostics Check	OK	04:37	

Data Download Date/Time: 12/06/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:41

SUBJECT'S LAST NAME: SALUP-RAMOS

DATE/TIME OF TEST: 12/02/2017 05:11:03

Diagnostics Check	OK	05:04
Air Blank	0.000	05:05
Control Test	0.079	05:05
Air Blank	0.000	05:06
Subject Sample # 1	0.150	05:06
Breath Volume	2.421	05:06
Air Blank	0.000	05:07
Air Blank	0.000	05:08
Subject Sample # 2	0.150	05:09
Breath Volume	3.191	05:09
Air Blank	0.000	05:10
Control Test	0.079	05:10
Air Blank	0.000	05:10
Diagnostics Check	OK	05:11

Data Download Date/Time: 12/06/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: QUINTANA,JORGE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: DUARTE

DATE/TIME OF TEST: 12/02/2017 05:53:44

Diagnostics Check	OK	05:47
Air Blank	0.000	05:47
Control Test	0.080	05:48
Air Blank	0.000	05:48
Subject Sample # 1	0.100	05:49
Breath Volume	2.351	05:49
Air Blank	0.000	05:49
Air Blank	0.000	05:51
Subject Sample # 2	0.105	05:52
Breath Volume	2.648	05:52
Air Blank	0.000	05:52
Control Test	0.079	05:53
Air Blank	0.000	05:53
Diagnostics Check	OK	05:53

Data Download Date/Time: 12/06/2017 09:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: GONZALEZ LLANO

DATE/TIME OF TEST: 12/03/2017 05:35:02

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:29
Subject Sample	# 1	0.050	05:30
Breath Volume		2.203	05:30
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.050	05:33
Breath Volume		2.671	05:33
Air Blank		0.000	05:34
Control Test		0.079	05:34
Air Blank		0.000	05:34
Diagnostics Check		OK	05:35

Data Download Date/Time: 12/06/2017 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: KUPPERSCHMID

DATE/TIME OF TEST: 12/06/2017 03:09:55

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.176	03:05
Breath Volume		1.742	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.180	03:08
Breath Volume		1.285	03:08
Air Blank		0.000	03:08
Control Test		0.078	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 12/06/2017 09:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DURAN,JARVYS
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: AUGUSTE

DATE/TIME OF TEST: 12/07/2017 02:57:39

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.171	02:52
Breath Volume		1.210	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.165	02:55
Breath Volume		1.601	02:55
Air Blank		0.000	02:56
Control Test		0.079	02:57
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: REESE

DATE/TIME OF TEST: 12/09/2017 04:48:39

Diagnostics Check	OK	04:42
Air Blank	0.000	04:42
Control Test	0.079	04:43
Air Blank	0.000	04:43
Subject Sample # 1	0.072	04:44
Breath Volume	2.730	04:44
Air Blank	0.000	04:44
Air Blank	0.000	04:46
Subject Sample # 2	0.074	04:46
Breath Volume	3.359	04:46
Air Blank	0.000	04:47
Control Test	0.078	04:48
Air Blank	0.000	04:48
Diagnostics Check	OK	04:48

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/09/2017 05:50:51

Diagnostics Check	OK	05:44
Air Blank	0.000	05:45
Control Test	0.079	05:45
Air Blank	0.000	05:45
Subject Sample # 1	0.114	05:46
Breath Volume	1.957	05:46
Air Blank	0.000	05:47
Air Blank	0.000	05:48
Subject Sample # 2	0.108	05:49
Breath Volume	1.527	05:49
Air Blank	0.000	05:49
Control Test	0.079	05:50
Air Blank	0.000	05:50
Diagnostics Check	OK	05:50

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: BENCOSME

DATE/TIME OF TEST: 12/10/2017 03:09:10

Diagnostics Check	OK	03:02
Air Blank	0.000	03:03
Control Test	0.081	03:03
Air Blank	0.000	03:04
Subject Sample # 1	0.201	03:04
Breath Volume	3.847	03:04
Air Blank	0.000	03:05
Air Blank	0.000	03:06
Subject Sample # 2	0.190	03:07
Breath Volume	4.136	03:07
Air Blank	0.000	03:08
Control Test	0.080	03:08
Air Blank	0.000	03:09
Diagnostics Check	OK	03:09

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: LAVIN

DATE/TIME OF TEST: 12/15/2017 04:09:04

Diagnostics Check	OK	04:02
Air Blank	0.000	04:03
Control Test	0.080	04:03
Air Blank	0.000	04:04
Subject Sample # 1	0.128	04:04
Breath Volume	1.695	04:04
Air Blank	0.000	04:05
Air Blank	0.000	04:07
Subject Sample # 2	0.121	04:07
Breath Volume	1.609	04:07
Air Blank	0.000	04:08
Control Test	0.080	04:08
Air Blank	0.000	04:08
Diagnostics Check	OK	04:09

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: DOSIO

DATE/TIME OF TEST: 12/23/2017 04:04:53

Diagnostics Check	OK	03:58
Air Blank	0.000	03:58
Control Test	0.080	03:58
Air Blank	0.000	03:59
Subject Sample # 1	0.067	03:59
Breath Volume	2.628	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:02
Subject Sample # 2	0.067	04:03
Breath Volume	5.296	04:03
Air Blank	0.000	04:03
Control Test	0.080	04:04
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: DOSIO

DATE/TIME OF TEST: 12/23/2017 05:05:24

Diagnostics Check	OK	04:59
Air Blank	0.000	04:59
Control Test	0.080	05:00
Air Blank	0.000	05:00
Subject Sample # 1	0.046	05:01
Breath Volume	2.617	05:01
Air Blank	0.000	05:01
Air Blank	0.000	05:03
Subject Sample # 2	0.042	05:03
Breath Volume	1.843	05:03
Air Blank	0.000	05:04
Control Test	0.079	05:04
Air Blank	0.000	05:05
Diagnostics Check	OK	05:05

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 12/24/2017 02:37:59

Diagnostics Check	OK	02:29
Air Blank	0.000	02:30
Control Test	0.079	02:30
Air Blank	0.000	02:31
Subject Sample # 1	0.073	02:33
Breath Volume	1.753	02:33
Air Blank	0.000	02:34
Air Blank	0.000	02:35
Subject Sample # 2	0.077	02:36
Breath Volume	1.722	02:36
Air Blank	0.000	02:37
Control Test	0.078	02:37
Air Blank	0.000	02:37
Diagnostics Check	OK	02:37

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: URBINA

DATE/TIME OF TEST: 12/28/2017 01:27:31

Diagnostics Check	OK	01:24	
Air Blank	0.000	01:25	
Control Test	0.080	01:25	
Air Blank	0.000	01:25	
Subject Sample	# 1	01:26	Subject Test Refused
Breath Volume	0.000	01:26	
Air Blank	0.000	01:26	
Control Test	0.080	01:26	
Air Blank	0.000	01:27	
Diagnostics Check	OK	01:27	

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:56

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/29/2017 05:36:06

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.080	05:33	
Air Blank	0.000	05:34	
Subject Sample	# 1	05:34	Subject Test Refused
Breath Volume	0.000	05:34	
Air Blank	0.000	05:35	
Control Test	0.081	05:35	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: RAGIN

DATE/TIME OF TEST: 12/31/2017 02:14:20

Diagnostics Check		OK	02:08
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.245	02:09
Breath Volume		2.210	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.258	02:12
Breath Volume		2.238	02:12
Air Blank		0.000	02:13
Control Test		0.079	02:13
Air Blank		0.000	02:14
Diagnostics Check		OK	02:14

Data Download Date/Time: 01/04/2018 12:12

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J.R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/01/2017 19:27:52

Diagnostics Check	OK	19:20
Air Blank	0.000	19:21
Control Test	0.079	19:21
Air Blank	0.000	19:22
Subject Sample # 1	0.226	19:23
Breath Volume	1.179	19:23
Air Blank	0.000	19:24
Air Blank	0.000	19:25
Subject Sample # 2	0.236	19:26
Breath Volume	1.265	19:26
Air Blank	0.000	19:26
Control Test	0.080	19:27
Air Blank	0.000	19:27
Diagnostics Check	OK	19:27

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JORDAN,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 12/01/2017 22:19:41

Diagnostics Check	OK	22:13
Air Blank	0.000	22:13
Control Test	0.080	22:13
Air Blank	0.000	22:14
Subject Sample # 1	0.147	22:15
Breath Volume	1.871	22:15
Air Blank	0.000	22:15
Air Blank	0.000	22:17
Subject Sample # 2	0.140	22:18
Breath Volume	1.859	22:18
Air Blank	0.000	22:18
Control Test	0.079	22:19
Air Blank	0.000	22:19
Diagnostics Check	OK	22:19

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRATT,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:22

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/03/2017 16:49:43

Diagnostics Check	OK	16:42
Air Blank	0.000	16:42
Control Test	0.080	16:43
Air Blank	0.000	16:43
Subject Sample # 1	0.240	16:44
Breath Volume	1.253	16:44
Air Blank	0.000	16:45
Air Blank	0.000	16:47
Subject Sample # 2	0.237	16:48
Breath Volume	1.214	16:48
Air Blank	0.000	16:48
Control Test	0.080	16:49
Air Blank	0.000	16:49
Diagnostics Check	OK	16:49

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: KATHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:19

SUBJECT'S LAST NAME: CASON

DATE/TIME OF TEST: 12/03/2017 17:46:08

Diagnostics Check	OK	17:39
Air Blank	0.000	17:39
Control Test	0.079	17:40
Air Blank	0.000	17:40
Subject Sample # 1	0.029	17:41
Breath Volume	1.488	17:41
Air Blank	0.000	17:42
Air Blank	0.000	17:43
Subject Sample # 2	0.030	17:44
Breath Volume	1.171	17:44
Air Blank	0.000	17:45
Control Test	0.080	17:45
Air Blank	0.000	17:46
Diagnostics Check	OK	17:46

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:27

SUBJECT'S LAST NAME: VASQUEZ

DATE/TIME OF TEST: 12/05/2017 18:54:00

Diagnostics Check	OK	18:47
Air Blank	0.000	18:47
Control Test	0.080	18:48
Air Blank	0.000	18:48
Subject Sample # 1	0.317	18:49
Breath Volume	1.371	18:49
Air Blank	0.000	18:50
Air Blank	0.000	18:51
Subject Sample # 2	0.304	18:52
Breath Volume	1.761	18:52
Air Blank	0.000	18:53
Control Test	0.080	18:53
Air Blank	0.000	18:53
Diagnostics Check	OK	18:54

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: N/A,N/A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:00

SUBJECT'S LAST NAME: LEGGETT

DATE/TIME OF TEST: 12/07/2017 19:28:26

Diagnostics Check	OK	19:25	
Air Blank	0.000	19:26	
Control Test	0.080	19:26	
Air Blank	0.000	19:26	
Subject Sample	# 1	19:27	Subject Test Refused
Breath Volume	0.000	19:27	
Air Blank	0.000	19:27	
Control Test	0.080	19:27	
Air Blank	0.000	19:28	
Diagnostics Check	OK	19:28	

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COLEMAN,F
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:19

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 12/08/2017 16:43:57

Diagnostics Check	OK	16:41	
Air Blank	0.000	16:41	
Control Test	0.080	16:41	
Air Blank	0.000	16:42	
Subject Sample	# 1	16:42	Subject Test Refused
Breath Volume	0.000	16:42	
Air Blank	0.000	16:43	
Control Test	0.080	16:43	
Air Blank	0.000	16:43	
Diagnostics Check	OK	16:43	

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THIGPEN,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: GRUBB

DATE/TIME OF TEST: 12/08/2017 22:06:00

Diagnostics Check	OK	22:03	
Air Blank	0.000	22:03	
Control Test	0.080	22:03	
Air Blank	0.000	22:04	
Subject Sample	# 1	22:04	Subject Test Refused
Breath Volume	0.000	22:04	
Air Blank	0.000	22:05	
Control Test	0.081	22:05	
Air Blank	0.000	22:05	
Diagnostics Check	OK	22:06	

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T.A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:24

SUBJECT'S LAST NAME: MUNFORD

DATE/TIME OF TEST: 12/08/2017 22:50:44

Diagnostics Check	OK	22:44
Air Blank	0.000	22:44
Control Test	0.081	22:45
Air Blank	0.000	22:45
Subject Sample # 1	0.151	22:46
Breath Volume	4.070	22:46
Air Blank	0.000	22:46
Air Blank	0.000	22:48
Subject Sample # 2	0.147	22:49
Breath Volume	4.113	22:49
Air Blank	0.000	22:49
Control Test	0.081	22:50
Air Blank	0.000	22:50
Diagnostics Check	OK	22:50

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,A,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:04

SUBJECT'S LAST NAME: LAURO

DATE/TIME OF TEST: 12/10/2017 21:30:06

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:27	
Control Test	0.080	21:27	
Air Blank	0.000	21:28	
Subject Sample	# 1	21:28	Subject Test Refused
Breath Volume	0.000	21:28	
Air Blank	0.000	21:29	
Control Test	0.081	21:29	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:30	

Data Download Date/Time: 12/11/2017 16:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANTONINI,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/13/2017 16:13:33

Diagnostics Check	OK	16:06
Air Blank	0.000	16:06
Control Test	0.080	16:06
Air Blank	0.000	16:07
Subject Sample # 1	0.000	16:07
Breath Volume	1.863	16:07
Air Blank	0.000	16:08
Air Blank	0.000	16:10
Subject Sample # 2	0.000	16:12
Breath Volume	1.984	16:12
Air Blank	0.000	16:12
Control Test	0.081	16:12
Air Blank	0.000	16:13
Diagnostics Check	OK	16:13

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:03

SUBJECT'S LAST NAME: DEMERY

DATE/TIME OF TEST: 12/15/2017 18:36:53

Diagnostics Check		OK	18:23	
Air Blank		0.000	18:23	
Control Test		0.080	18:23	
Air Blank		0.000	18:24	
Subject Sample	# 1	0.099	18:27	Volume Not Met
Breath Volume		0.632	18:27	
Air Blank		0.000	18:28	
Air Blank		0.000	18:30	
Subject Sample	# 2	0.117	18:30	
Breath Volume		2.246	18:30	
Air Blank		0.000	18:31	
Air Blank		0.000	18:33	
Subject Sample	# 3	0.122	18:35	
Breath Volume		1.906	18:35	
Air Blank		0.000	18:35	
Control Test		0.083	18:36	
Air Blank		0.000	18:36	
Diagnostics Check		OK	18:36	

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WALLACE,B
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:48

SUBJECT'S LAST NAME: STEPHENS

DATE/TIME OF TEST: 12/15/2017 20:15:21

Diagnostics Check	OK	20:08
Air Blank	0.000	20:09
Control Test	0.083	20:09
Air Blank	0.000	20:10
Subject Sample # 1	0.000	20:11
Breath Volume	1.339	20:11
Air Blank	0.000	20:11
Air Blank	0.000	20:13
Subject Sample # 2	0.000	20:13
Breath Volume	1.621	20:13
Air Blank	0.000	20:14
Control Test	0.076	20:14
Air Blank	0.000	20:15
Diagnostics Check	OK	20:15

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: PEARCE

DATE/TIME OF TEST: 12/15/2017 22:58:51

Diagnostics Check	OK	22:52
Air Blank	0.000	22:52
Control Test	0.081	22:53
Air Blank	0.000	22:53
Subject Sample # 1	0.000	22:54
Breath Volume	3.285	22:54
Air Blank	0.000	22:54
Air Blank	0.000	22:56
Subject Sample # 2	0.000	22:57
Breath Volume	3.613	22:57
Air Blank	0.000	22:57
Control Test	0.080	22:58
Air Blank	0.000	22:58
Diagnostics Check	OK	22:58

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:29

SUBJECT'S LAST NAME: FELICE

DATE/TIME OF TEST: 12/16/2017 20:56:24

Diagnostics Check	OK	20:49
Air Blank	0.000	20:49
Control Test	0.080	20:50
Air Blank	0.000	20:50
Subject Sample # 1	0.021	20:51
Breath Volume	2.109	20:51
Air Blank	0.000	20:52
Air Blank	0.000	20:54
Subject Sample # 2	0.018	20:54
Breath Volume	1.917	20:54
Air Blank	0.000	20:55
Control Test	0.081	20:55
Air Blank	0.000	20:56
Diagnostics Check	OK	20:56

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 12/16/2017 22:50:33

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.081	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.204	22:45
Breath Volume		2.457	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.213	22:49
Breath Volume		2.703	22:49
Air Blank		0.000	22:49
Control Test		0.080	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: ESPLING

DATE/TIME OF TEST: 12/17/2017 19:30:17

Diagnostics Check	OK	19:27	
Air Blank	0.000	19:27	
Control Test	0.080	19:28	
Air Blank	0.000	19:28	
Subject Sample	# 1	19:28	Subject Test Refused
Breath Volume	0.000	19:28	
Air Blank	0.000	19:29	
Control Test	0.080	19:29	
Air Blank	0.000	19:30	
Diagnostics Check	OK	19:30	

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXIS,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: GRAVLEE

DATE/TIME OF TEST: 12/20/2017 19:25:53

Diagnostics Check	OK	19:23	
Air Blank	0.000	19:24	
Control Test	0.080	19:24	
Air Blank		19:25	Purge Fail
Air Blank	0.000	19:25	

Data Download Date/Time: 12/29/2017 14:23

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,ALEXIS,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: GRAVLEE

DATE/TIME OF TEST: 12/20/2017 19:33:58

Diagnostics Check	OK	19:27
Air Blank	0.000	19:27
Control Test	0.078	19:28
Air Blank	0.000	19:28
Subject Sample # 1	0.200	19:29
Breath Volume	3.718	19:29
Air Blank	0.000	19:30
Air Blank	0.000	19:31
Subject Sample # 2	0.215	19:32
Breath Volume	3.277	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, TYLER, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ADMIN, ADMIN
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ADMIN

DATE/TIME OF TEST: 12/22/2017 17:41:12

Diagnostics Check	OK	17:35
Air Blank	0.000	17:35
Control Test	0.080	17:36
Air Blank	0.000	17:36
Subject Sample # 1	0.085	17:36
Breath Volume	2.539	17:36
Air Blank	0.000	17:37
Air Blank	0.000	17:39
Subject Sample # 2	0.083	17:39
Breath Volume	2.195	17:39
Air Blank	0.000	17:40
Control Test	0.080	17:40
Air Blank	0.000	17:41
Diagnostics Check	OK	17:41

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER, TYLER, R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: N/A, N/A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: ADMINISTRATIVE TEST

DATE/TIME OF TEST: 12/22/2017 18:04:42

Diagnostics Check	OK	17:58
Air Blank	0.000	17:58
Control Test	0.080	17:59
Air Blank	0.000	17:59
Subject Sample # 1	0.072	18:00
Breath Volume	2.359	18:00
Air Blank	0.000	18:00
Air Blank	0.000	18:02
Subject Sample # 2	0.075	18:03
Breath Volume	2.320	18:03
Air Blank	0.000	18:03
Control Test	0.080	18:04
Air Blank	0.000	18:04
Diagnostics Check	OK	18:04

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 12/22/2017 20:38:22

Diagnostics Check	OK	20:30
Air Blank	0.000	20:31
Control Test	0.080	20:31
Air Blank	0.000	20:32
Subject Sample # 1	0.151	20:32
Breath Volume	1.468	20:32
Air Blank	0.000	20:33
Air Blank	0.000	20:35
Subject Sample # 2	0.150	20:36
Breath Volume	1.457	20:36
Air Blank	0.000	20:37
Control Test	0.081	20:37
Air Blank	0.000	20:38
Diagnostics Check	OK	20:38

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 12/22/2017 22:39:40

Diagnostics Check	OK	22:36	
Air Blank	0.000	22:37	
Control Test	0.080	22:37	
Air Blank	0.000	22:38	
Subject Sample	# 1	22:38	Subject Test Refused
Breath Volume	0.000	22:38	
Air Blank	0.000	22:38	
Control Test	0.081	22:39	
Air Blank	0.000	22:39	
Diagnostics Check	OK	22:39	

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: PUTMAN

DATE/TIME OF TEST: 12/23/2017 18:20:50

Diagnostics Check	OK	18:14
Air Blank	0.000	18:14
Control Test	0.080	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.000	18:16
Breath Volume	1.859	18:16
Air Blank	0.000	18:16
Air Blank	0.000	18:18
Subject Sample # 2	0.000	18:19
Breath Volume	1.621	18:19
Air Blank	0.000	18:19
Control Test	0.081	18:20
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: ROSALES CARDENAS

DATE/TIME OF TEST: 12/23/2017 20:27:55

Diagnostics Check	OK	20:20
Air Blank	0.000	20:21
Control Test	0.079	20:21
Air Blank	0.000	20:22
Subject Sample # 1	0.179	20:23
Breath Volume	1.933	20:23
Air Blank	0.000	20:23
Air Blank	0.000	20:25
Subject Sample # 2	0.173	20:26
Breath Volume	1.523	20:26
Air Blank	0.000	20:27
Control Test	0.080	20:27
Air Blank	0.000	20:27
Diagnostics Check	OK	20:27

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:59

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 12/24/2017 21:26:50

Diagnostics Check	OK	21:20
Air Blank	0.000	21:20
Control Test	0.080	21:20
Air Blank	0.000	21:21
Subject Sample # 1	0.227	21:22
Breath Volume	2.195	21:22
Air Blank	0.000	21:22
Air Blank	0.000	21:24
Subject Sample # 2	0.215	21:25
Breath Volume	2.312	21:25
Air Blank	0.000	21:25
Control Test	0.080	21:26
Air Blank	0.000	21:26
Diagnostics Check	OK	21:26

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 12/26/2017 21:05:34

Diagnostics Check	OK	20:59
Air Blank	0.000	20:59
Control Test	0.080	20:59
Air Blank	0.000	21:00
Subject Sample # 1	0.145	21:01
Breath Volume	2.050	21:01
Air Blank	0.000	21:01
Air Blank	0.000	21:03
Subject Sample # 2	0.151	21:04
Breath Volume	1.507	21:04
Air Blank	0.000	21:04
Control Test	0.080	21:04
Air Blank	0.000	21:05
Diagnostics Check	OK	21:05

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: DEMARS

DATE/TIME OF TEST: 12/28/2017 22:03:38

Diagnostics Check	OK	21:57
Air Blank	0.000	21:57
Control Test	0.080	21:58
Air Blank	0.000	21:58
Subject Sample # 1	0.270	21:59
Breath Volume	2.484	21:59
Air Blank	0.000	21:59
Air Blank	0.000	22:01
Subject Sample # 2	0.276	22:02
Breath Volume	1.957	22:02
Air Blank	0.000	22:02
Control Test	0.080	22:03
Air Blank	0.000	22:03
Diagnostics Check	OK	22:03

Data Download Date/Time: 12/29/2017 14:24

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,A,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EASTERLY,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: GRUBB

DATE/TIME OF TEST: 12/30/2017 18:13:41

Diagnostics Check	OK	18:10	
Air Blank	0.000	18:11	
Control Test	0.081	18:11	
Air Blank	0.000	18:11	
Subject Sample	# 1	18:12	Subject Test Refused
Breath Volume	0.000	18:12	
Air Blank	0.000	18:12	
Control Test	0.081	18:13	
Air Blank	0.000	18:13	
Diagnostics Check	OK	18:13	

Data Download Date/Time: 01/19/2018 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,A,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:43

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/31/2017 21:08:02

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:07	
		21:07	RFI Detect
Air Blank	0.000	21:08	

Data Download Date/Time: 01/19/2018 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,A,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/31/2017 21:31:05

Diagnostics Check	OK	21:27	
Air Blank	0.000	21:28	
Control Test	0.080	21:28	
Air Blank	0.000	21:29	
Subject Sample	# 1	21:29	Subject Test Refused
Breath Volume	0.000	21:29	
Air Blank	0.000	21:30	
Control Test	0.081	21:30	
Air Blank	0.000	21:30	
Diagnostics Check	OK	21:31	

Data Download Date/Time: 01/19/2018 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BENTON

DATE/TIME OF TEST: 12/01/2017 02:13:16

Diagnostics Check	OK	02:10	
Air Blank	0.000	02:10	
Control Test	0.080	02:11	
Air Blank	0.000	02:11	
Subject Sample	# 1	02:11	Subject Test Refused
Breath Volume	0.000	02:11	
Air Blank	0.000	02:12	
Control Test	0.080	02:12	
Air Blank	0.000	02:13	
Diagnostics Check	OK	02:13	

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J.T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/01/2017 22:42:12

Diagnostics Check	OK	22:35
Air Blank	0.000	22:36
Control Test	0.079	22:36
Air Blank	0.000	22:36
Subject Sample # 1	0.013	22:37
Breath Volume	2.234	22:37
Air Blank	0.000	22:38
Air Blank	0.000	22:40
Subject Sample # 2	0.012	22:40
Breath Volume	2.417	22:40
Air Blank	0.000	22:41
Control Test	0.080	22:41
Air Blank	0.000	22:42
Diagnostics Check	OK	22:42

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:06

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 12/02/2017 07:35:04

Diagnostics Check	OK	07:32	
Air Blank	0.000	07:32	
Control Test	0.079	07:32	
Air Blank	0.000	07:33	
Subject Sample	# 1	07:33	Subject Test Refused
Breath Volume	0.000	07:33	
Air Blank	0.000	07:34	
Control Test	0.080	07:34	
Air Blank	0.000	07:34	
Diagnostics Check	OK	07:35	

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J.K.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: RAMOSSANCHEZ

DATE/TIME OF TEST: 12/03/2017 00:35:34

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.079	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.114	00:30
Breath Volume		2.312	00:30
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.110	00:34
Breath Volume		1.414	00:34
Air Blank		0.000	00:34
Control Test		0.078	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J.M.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: VALDEZ PAULINA

DATE/TIME OF TEST: 12/03/2017 01:24:29

Diagnostics Check	OK	01:17
Air Blank	0.000	01:18
Control Test	0.079	01:18
Air Blank	0.000	01:19
Subject Sample # 1	0.229	01:19
Breath Volume	2.179	01:19
Air Blank	0.000	01:20
Air Blank	0.000	01:22
Subject Sample # 2	0.226	01:22
Breath Volume	2.179	01:22
Air Blank	0.000	01:23
Control Test	0.079	01:23
Air Blank	0.000	01:24
Diagnostics Check	OK	01:24

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RODRIGUEZ,M.R.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: NEALY

DATE/TIME OF TEST: 12/03/2017 03:13:29

Diagnostics Check	OK	03:10	
Air Blank	0.000	03:10	
Control Test	0.079	03:11	
Air Blank	0.000	03:11	
Subject Sample	# 1	03:12	Subject Test Refused
Breath Volume	0.000	03:12	
Air Blank	0.000	03:12	
Control Test	0.079	03:12	
Air Blank	0.000	03:13	
Diagnostics Check	OK	03:13	

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T.E.
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:21

SUBJECT'S LAST NAME: CALLE MURILLO

DATE/TIME OF TEST: 12/03/2017 03:47:56

Diagnostics Check	OK	03:41
Air Blank	0.000	03:41
Control Test	0.079	03:42
Air Blank	0.000	03:42
Subject Sample # 1	0.194	03:43
Breath Volume	1.929	03:43
Air Blank	0.000	03:44
Air Blank	0.000	03:45
Subject Sample # 2	0.185	03:46
Breath Volume	2.113	03:46
Air Blank	0.000	03:47
Control Test	0.079	03:47
Air Blank	0.000	03:47
Diagnostics Check	OK	03:47

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBER,JOSEPH,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: LANE,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/03/2017 05:45:14

Diagnostics Check	OK	05:39
Air Blank	0.000	05:39
Control Test	0.079	05:39
Air Blank	0.000	05:40
Subject Sample # 1	0.101	05:40
Breath Volume	2.441	05:40
Air Blank	0.000	05:41
Air Blank	0.000	05:43
Subject Sample # 2	0.096	05:43
Breath Volume	1.539	05:43
Air Blank	0.000	05:44
Control Test	0.078	05:44
Air Blank	0.000	05:45
Diagnostics Check	OK	05:45

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:21

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/03/2017 07:47:47

Diagnostics Check	OK	07:41
Air Blank	0.000	07:41
Control Test	0.080	07:42
Air Blank	0.000	07:42
Subject Sample # 1	0.128	07:43
Breath Volume	3.082	07:43
Air Blank	0.000	07:44
Air Blank	0.000	07:45
Subject Sample # 2	0.128	07:46
Breath Volume	2.761	07:46
Air Blank	0.000	07:46
Control Test	0.079	07:47
Air Blank	0.000	07:47
Diagnostics Check	OK	07:47

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:29

SUBJECT'S LAST NAME: EMERSON

DATE/TIME OF TEST: 12/06/2017 23:52:33

Diagnostics Check	OK	23:49	
Air Blank	0.000	23:50	
Control Test	0.080	23:50	
Air Blank	0.000	23:50	
Subject Sample	# 1	23:51	Subject Test Refused
Breath Volume	0.000	23:51	
Air Blank	0.000	23:51	
Control Test	0.080	23:51	
Air Blank	0.000	23:52	
Diagnostics Check	OK	23:52	

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: PIPPINS

DATE/TIME OF TEST: 12/08/2017 01:41:43

Diagnostics Check	OK	01:35
Air Blank	0.000	01:35
Control Test	0.080	01:36
Air Blank	0.000	01:36
Subject Sample # 1	0.167	01:37
Breath Volume	1.511	01:37
Air Blank	0.000	01:37
Air Blank	0.000	01:39
Subject Sample # 2	0.163	01:40
Breath Volume	2.160	01:40
Air Blank	0.000	01:40
Control Test	0.080	01:41
Air Blank	0.000	01:41
Diagnostics Check	OK	01:41

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: DUNNIE

DATE/TIME OF TEST: 12/10/2017 02:17:10

Diagnostics Check	OK	02:11
Air Blank	0.000	02:11
Control Test	0.080	02:11
Air Blank	0.000	02:12
Subject Sample # 1	0.100	02:12
Breath Volume	2.257	02:12
Air Blank	0.000	02:13
Air Blank	0.000	02:15
Subject Sample # 2	0.088	02:15
Breath Volume	1.289	02:15
Air Blank	0.000	02:16
Control Test	0.080	02:16
Air Blank	0.000	02:17
Diagnostics Check	OK	02:17

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: RODRIGUEZ, M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:26

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 12/10/2017 03:52:19

Diagnostics Check	OK	03:46
Air Blank	0.000	03:46
Control Test	0.081	03:47
Air Blank	0.000	03:47
Subject Sample # 1	0.000	03:48
Breath Volume	2.574	03:48
Air Blank	0.000	03:48
Air Blank	0.000	03:50
Subject Sample # 2	0.000	03:50
Breath Volume	1.308	03:50
Air Blank	0.000	03:51
Control Test	0.081	03:51
Air Blank	0.000	03:52
Diagnostics Check	OK	03:52

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: BURNETT

DATE/TIME OF TEST: 12/10/2017 05:09:16

Diagnostics Check		OK	05:02
Air Blank		0.000	05:03
Control Test		0.081	05:03
Air Blank		0.000	05:04
Subject Sample	# 1	0.220	05:04
Breath Volume		1.652	05:04
Air Blank		0.000	05:05
Air Blank		0.000	05:07
Subject Sample	# 2	0.212	05:07
Breath Volume		1.531	05:07
Air Blank		0.000	05:08
Control Test		0.081	05:08
Air Blank		0.000	05:09
Diagnostics Check		OK	05:09

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER, R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/11/2017 00:59:25

Diagnostics Check	OK	00:53
Air Blank	0.000	00:53
Control Test	0.080	00:53
Air Blank	0.000	00:54
Subject Sample # 1	0.212	00:54
Breath Volume	3.292	00:54
Air Blank	0.000	00:55
Air Blank	0.000	00:57
Subject Sample # 2	0.199	00:57
Breath Volume	1.316	00:57
Air Blank	0.000	00:58
Control Test	0.079	00:58
Air Blank	0.000	00:59
Diagnostics Check	OK	00:59

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:28

SUBJECT'S LAST NAME: VALENTIN

DATE/TIME OF TEST: 12/11/2017 05:55:04

Diagnostics Check	OK	05:48
Air Blank	0.000	05:49
Control Test	0.080	05:49
Air Blank	0.000	05:49
Subject Sample # 1	0.177	05:50
Breath Volume	1.281	05:50
Air Blank	0.000	05:51
Air Blank	0.000	05:52
Subject Sample # 2	0.176	05:53
Breath Volume	2.210	05:53
Air Blank	0.000	05:54
Control Test	0.081	05:54
Air Blank	0.000	05:54
Diagnostics Check	OK	05:55

Data Download Date/Time: 12/11/2017 16:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: BOLING

DATE/TIME OF TEST: 12/12/2017 00:22:43

Diagnostics Check	OK	00:16
Air Blank	0.000	00:16
Control Test	0.080	00:17
Air Blank	0.000	00:17
Subject Sample # 1	0.007	00:18
Breath Volume	2.039	00:18
Air Blank	0.000	00:18
Air Blank	0.000	00:20
Subject Sample # 2	0.000	00:21
Breath Volume	2.546	00:21
Air Blank	0.000	00:21
Control Test	0.079	00:22
Air Blank	0.000	00:22
Diagnostics Check	OK	00:22

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE, J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SMALL

DATE/TIME OF TEST: 12/12/2017 03:27:22

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.103	03:22
Breath Volume		1.402	03:22
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.105	03:25
Breath Volume		2.011	03:25
Air Blank		0.000	03:26
Control Test		0.080	03:26
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:57

SUBJECT'S LAST NAME: TRAUGOTT

DATE/TIME OF TEST: 12/12/2017 23:24:07

Diagnostics Check	OK	23:17
Air Blank	0.000	23:18
Control Test	0.079	23:18
Air Blank	0.000	23:19
Subject Sample # 1	0.222	23:19
Breath Volume	1.941	23:19
Air Blank	0.000	23:20
Air Blank	0.000	23:22
Subject Sample # 2	0.215	23:22
Breath Volume	1.589	23:22
Air Blank	0.000	23:23
Control Test	0.079	23:23
Air Blank	0.000	23:24
Diagnostics Check	OK	23:24

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:43

SUBJECT'S LAST NAME: EVANS

DATE/TIME OF TEST: 12/13/2017 01:06:33

Diagnostics Check	OK	01:03	
Air Blank	0.000	01:04	
Control Test	0.081	01:04	
Air Blank	0.000	01:05	
Subject Sample	# 1	01:05	Subject Test Refused
Breath Volume	0.000	01:05	
Air Blank	0.000	01:05	
Control Test	0.079	01:05	
Air Blank	0.000	01:06	
Diagnostics Check	OK	01:06	

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: CLARKE

DATE/TIME OF TEST: 12/13/2017 02:32:00

Diagnostics Check	OK	02:29	
Air Blank	0.000	02:29	
Control Test	0.080	02:29	
Air Blank	0.000	02:30	
Subject Sample	# 1	02:30	Subject Test Refused
Breath Volume	0.000	02:30	
Air Blank	0.000	02:31	
Control Test	0.079	02:31	
Air Blank	0.000	02:31	
Diagnostics Check	OK	02:32	

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/13/2017 06:40:15

Diagnostics Check		OK	06:30
Air Blank		0.000	06:30
Control Test		0.080	06:31
Air Blank		0.000	06:31
Subject Sample	# 1	0.243	06:34
Breath Volume		1.746	06:34
Air Blank		0.000	06:34
Air Blank		0.000	06:36
Subject Sample	# 2	0.250	06:38
Breath Volume		2.437	06:38
Air Blank		0.000	06:39
Control Test		0.078	06:39
Air Blank		0.000	06:40
Diagnostics Check		OK	06:40

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MAYS,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: HILDRETH

DATE/TIME OF TEST: 12/13/2017 15:15:16

Diagnostics Check	OK	15:08
Air Blank	0.000	15:09
Control Test	0.080	15:09
Air Blank	0.000	15:09
Subject Sample # 1	0.007	15:10
Breath Volume	2.132	15:10
Air Blank	0.000	15:11
Air Blank	0.000	15:13
Subject Sample # 2	0.000	15:13
Breath Volume	1.347	15:13
Air Blank	0.000	15:14
Control Test	0.080	15:14
Air Blank	0.000	15:15
Diagnostics Check	OK	15:15

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,DONALD,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ANTONINI,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:31

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/13/2017 15:59:18

Diagnostics Check	OK	15:52
Air Blank	0.000	15:52
Control Test	0.080	15:53
Air Blank	0.000	15:53
Subject Sample # 1	0.007	15:54
Breath Volume	2.750	15:54
Air Blank	0.000	15:55
Air Blank	0.000	15:57
Subject Sample # 2	0.000	15:57
Breath Volume	2.074	15:57
Air Blank	0.000	15:58
Control Test	0.080	15:58
Air Blank	0.000	15:59
Diagnostics Check	OK	15:59

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTON,D,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:42

SUBJECT'S LAST NAME: CAMERON

DATE/TIME OF TEST: 12/14/2017 22:09:58

Diagnostics Check	OK	22:02
Air Blank	0.000	22:02
Control Test	0.080	22:03
Air Blank	0.000	22:03
Subject Sample # 1	0.209	22:04
Breath Volume	1.660	22:04
Air Blank	0.000	22:05
Air Blank	0.000	22:07
Subject Sample # 2	0.209	22:08
Breath Volume	2.156	22:08
Air Blank	0.000	22:09
Control Test	0.078	22:09
Air Blank	0.000	22:09
Diagnostics Check	OK	22:09

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAYES,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/15/2017 01:01:19

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:55
Subject Sample	# 1	0.211	00:56
Breath Volume		2.203	00:56
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.217	00:59
Breath Volume		2.230	00:59
Air Blank		0.000	01:00
Control Test		0.078	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 12/15/2017 04:01:18

Diagnostics Check	OK	03:54
Air Blank	0.000	03:55
Control Test	0.079	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.218	03:56
Breath Volume	2.386	03:56
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.220	03:59
Breath Volume	1.136	03:59
Air Blank	0.000	04:00
Control Test	0.078	04:00
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: STATKUS

DATE/TIME OF TEST: 12/15/2017 23:35:06

Diagnostics Check	OK	23:29
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:30
Subject Sample # 1	0.170	23:30
Breath Volume	2.109	23:30
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 2	0.168	23:33
Breath Volume	1.468	23:33
Air Blank	0.000	23:34
Control Test	0.080	23:34
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/16/2017 03:48:37

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.080	03:46	
Air Blank	0.000	03:47	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.081	03:48	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 12/17/2017 01:40:37

Diagnostics Check	OK	01:37	
Air Blank	0.000	01:38	
Control Test	0.081	01:38	
Air Blank	0.000	01:39	
Subject Sample	# 1	01:39	Subject Test Refused
Breath Volume	0.000	01:39	
Air Blank	0.000	01:39	
Control Test	0.079	01:40	
Air Blank	0.000	01:40	
Diagnostics Check	OK	01:40	

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 12/17/2017 04:31:12

Diagnostics Check	OK	04:24
Air Blank	0.000	04:25
Control Test	0.081	04:25
Air Blank	0.000	04:25
Subject Sample # 1	0.154	04:26
Breath Volume	2.621	04:26
Air Blank	0.000	04:26
Air Blank	0.000	04:28
Subject Sample # 2	0.152	04:29
Breath Volume	2.667	04:29
Air Blank	0.000	04:30
Control Test	0.078	04:30
Air Blank	0.000	04:31
Diagnostics Check	OK	04:31

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: VANHOF,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ELLEDGE

DATE/TIME OF TEST: 12/17/2017 23:11:37

Diagnostics Check	OK	23:05
Air Blank	0.000	23:05
Control Test	0.080	23:05
Air Blank	0.000	23:06
Subject Sample # 1	0.181	23:06
Breath Volume	1.687	23:06
Air Blank	0.000	23:07
Air Blank	0.000	23:09
Subject Sample # 2	0.170	23:09
Breath Volume	1.699	23:09
Air Blank	0.000	23:10
Control Test	0.079	23:11
Air Blank	0.000	23:11
Diagnostics Check	OK	23:11

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/18/2017 02:48:16

Diagnostics Check	OK	02:41
Air Blank	0.000	02:42
Control Test	0.080	02:42
Air Blank	0.000	02:42
Subject Sample # 1	0.176	02:43
Breath Volume	2.265	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.179	02:46
Breath Volume	2.589	02:46
Air Blank	0.000	02:47
Control Test	0.079	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: CLAYTON

DATE/TIME OF TEST: 12/18/2017 23:54:53

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.080	23:52	
Air Blank	0.000	23:53	
Subject Sample	# 1	23:53	Subject Test Refused
Breath Volume	0.000	23:53	
Air Blank	0.000	23:53	
Control Test	0.080	23:54	
Air Blank	0.000	23:54	
Diagnostics Check	OK	23:54	

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/19/2017 01:03:12

Diagnostics Check		OK	00:57
Air Blank		0.000	00:57
Control Test		0.079	00:57
Air Blank		0.000	00:58
Subject Sample	# 1	0.314	00:58
Breath Volume		2.476	00:58
Air Blank		0.000	00:59
Air Blank		0.000	01:01
Subject Sample	# 2	0.309	01:01
Breath Volume		2.699	01:01
Air Blank		0.000	01:02
Control Test		0.079	01:02
Air Blank		0.000	01:03
Diagnostics Check		OK	01:03

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ALKHMOOR

DATE/TIME OF TEST: 12/19/2017 03:34:19

Diagnostics Check	OK	03:31	
Air Blank	0.000	03:31	
Control Test	0.080	03:32	
Air Blank	0.000	03:32	
Subject Sample	# 1	03:32	Subject Test Refused
Breath Volume	0.000	03:32	
Air Blank	0.000	03:33	
Control Test	0.083	03:33	
Air Blank	0.000	03:34	
Diagnostics Check	OK	03:34	

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY, T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/20/2017 02:36:57

Diagnostics Check	OK	02:30
Air Blank	0.000	02:31
Control Test	0.079	02:31
Air Blank	0.000	02:31
Subject Sample # 1	0.000	02:32
Breath Volume	2.234	02:32
Air Blank	0.000	02:32
Air Blank	0.000	02:34
Subject Sample # 2	0.007	02:35
Breath Volume	3.238	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:36
Diagnostics Check	OK	02:36

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001280

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERRY, SHANON, E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WELLS, D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 12/20/2017 04:03:06

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.000	03:58
Breath Volume	2.109	03:58
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.000	04:01
Breath Volume	2.015	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 12/21/2017 14:03

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEWART,DILLON,J
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BEGUIRISTAIN,VICTOR
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 12/01/2017 03:46:44

Diagnostics Check	OK	03:43	
Air Blank	0.000	03:44	
Control Test	0.076	03:44	
Air Blank	0.000	03:45	
Subject Sample	# 1	03:45	Subject Test Refused
Breath Volume	0.000	03:45	
Air Blank	0.000	03:45	
Control Test	0.076	03:46	
Air Blank	0.000	03:46	
Diagnostics Check	OK	03:46	

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: PERRYMAN

DATE/TIME OF TEST: 12/02/2017 03:33:36

Diagnostics Check	OK	03:27
Air Blank	0.000	03:27
Control Test	0.077	03:28
Air Blank	0.000	03:28
Subject Sample # 1	0.100	03:29
Breath Volume	2.730	03:29
Air Blank	0.000	03:29
Air Blank	0.000	03:31
Subject Sample # 2	0.098	03:32
Breath Volume	2.347	03:32
Air Blank	0.000	03:32
Control Test	0.077	03:33
Air Blank	0.000	03:33
Diagnostics Check	OK	03:33

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: STEWART,DILLON
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: PATEL

DATE/TIME OF TEST: 12/06/2017 03:22:49

Diagnostics Check	OK	03:16
Air Blank	0.000	03:16
Control Test	0.077	03:17
Air Blank	0.000	03:17
Subject Sample # 1	0.101	03:18
Breath Volume	3.460	03:18
Air Blank	0.000	03:19
Air Blank	0.000	03:20
Subject Sample # 2	0.092	03:21
Breath Volume	1.347	03:21
Air Blank	0.000	03:21
Control Test	0.080	03:22
Air Blank	0.000	03:22
Diagnostics Check	OK	03:22

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BECKHAM,MARCUS,W
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: GONZALEZ PACHECO

DATE/TIME OF TEST: 12/08/2017 03:05:29

Diagnostics Check	OK	02:58
Air Blank	0.000	02:59
Control Test	0.077	02:59
Air Blank	0.000	02:59
Subject Sample # 1	0.151	03:00
Breath Volume	3.359	03:00
Air Blank	0.000	03:01
Air Blank	0.000	03:03
Subject Sample # 2	0.146	03:03
Breath Volume	2.941	03:03
Air Blank	0.000	03:04
Control Test	0.077	03:04
Air Blank	0.000	03:05
Diagnostics Check	OK	03:05

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: ZARATE,JESSICA
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: WASSERFALL

DATE/TIME OF TEST: 12/10/2017 03:32:05

Diagnostics Check	OK	03:27	
Air Blank	0.000	03:28	
Control Test	0.077	03:28	
Air Blank	0.000	03:29	
Subject Sample	# 1	03:30	Subject Test Refused
Breath Volume	0.152	03:30	
Air Blank	0.000	03:31	
Control Test	0.079	03:31	
Air Blank	0.000	03:32	
Diagnostics Check	OK	03:32	

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DIAZ VAZQUEZ

DATE/TIME OF TEST: 12/16/2017 02:57:27

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.077	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1	0.000	02:50	Volume Not Met
Breath Volume		0.109	02:50	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.097	02:55	Volume Not Met
Breath Volume		0.410	02:55	
Air Blank		0.000	02:56	
Control Test		0.079	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DIAZ VAZQUEZ

DATE/TIME OF TEST: 12/16/2017 03:11:53

Diagnostics Check		OK	03:00	
Air Blank		0.000	03:00	
Control Test		0.076	03:01	
Air Blank		0.000	03:01	
Subject Sample	# 1	0.087	03:04	Volume Not Met
Breath Volume		0.113	03:04	
Air Blank		0.000	03:05	
Air Blank		0.000	03:07	
Subject Sample	# 2	0.094	03:10	Volume Not Met
Breath Volume		0.074	03:10	
Air Blank		0.000	03:10	
Control Test		0.078	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAIN II,KENNETH,A
BREATH TEST OPERATOR AGENCY: U of FI PD
ARRESTING OFFICER: BAIN II,KENNETH
ARRESTING OFFICER AGENCY: U of FI PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: DIAZ VAZQUEZ

DATE/TIME OF TEST: 12/16/2017 03:28:05

Diagnostics Check	OK	03:13	
Air Blank	0.000	03:13	
Control Test	0.076	03:13	
Air Blank	0.000	03:14	
Subject Sample	# 1 0.090	03:16	
Breath Volume	1.429	03:16	
Air Blank	0.000	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 2 0.090	03:22	Volume Not Met
Breath Volume	1.000	03:22	
Air Blank	0.000	03:22	
Air Blank	0.000	03:24	
Subject Sample	# 3 0.082	03:26	
Breath Volume	1.277	03:26	
Air Blank	0.000	03:27	
Control Test	0.076	03:27	
Air Blank	0.000	03:28	
Diagnostics Check	OK	03:28	

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: RAY

DATE/TIME OF TEST: 12/20/2017 23:27:00

Diagnostics Check		OK	23:14	
Air Blank		0.000	23:15	
Control Test		0.075	23:15	
Air Blank		0.000	23:16	
Subject Sample	# 1	0.098	23:19	Volume Not Met
Breath Volume		0.476	23:19	
Air Blank		0.000	23:19	
Air Blank		0.000	23:21	
Subject Sample	# 2	0.098	23:22	
Breath Volume		1.390	23:22	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 3	0.101	23:25	
Breath Volume		1.593	23:25	
Air Blank		0.000	23:26	
Control Test		0.076	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 12/22/2017 10:44

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,LYNNE,S
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GOULD,PATRICIA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: MAST

DATE/TIME OF TEST: 12/31/2017 05:15:54

Diagnostics Check	OK	05:04	
Air Blank	0.000	05:04	
Control Test	0.077	05:05	
Air Blank	0.000	05:05	
Subject Sample	# 1	05:08	No Sample Provided
Breath Volume	0.000	05:08	
Air Blank	0.000	05:09	
Air Blank	0.000	05:11	
Subject Sample	# 2	05:14	No Sample Provided
Breath Volume	0.000	05:14	
Air Blank	0.000	05:14	
Control Test	0.078	05:15	
Air Blank	0.000	05:15	
Diagnostics Check	OK	05:15	

Data Download Date/Time: 01/30/2018 15:20

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANGFORD,BRIAN,A
BREATH TEST OPERATOR AGENCY: Gilchrist County SO
ARRESTING OFFICER: RALEY,B
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 11/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 18:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ELDER

DATE/TIME OF TEST: 12/20/2017 20:43:07

Diagnostics Check	OK	20:37
Air Blank	0.000	20:37
Control Test	0.079	20:37
Air Blank	0.000	20:38
Subject Sample # 1	0.158	20:38
Breath Volume	2.769	20:38
Air Blank	0.000	20:39
Air Blank	0.000	20:41
Subject Sample # 2	0.165	20:41
Breath Volume	2.414	20:41
Air Blank	0.000	20:42
Control Test	0.078	20:42
Air Blank	0.000	20:43
Diagnostics Check	OK	20:43

Data Download Date/Time: 12/31/2017 18:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CHARLES,C
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BRANNON,AARON
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 12/17/2017 00:49:06

Diagnostics Check	OK	00:46	
Air Blank	0.000	00:46	
Control Test	0.080	00:47	
Air Blank	0.000	00:47	
Subject Sample	# 1	00:47	Subject Test Refused
Breath Volume	0.000	00:47	
Air Blank	0.000	00:48	
Control Test	0.081	00:48	
Air Blank	0.000	00:49	
Diagnostics Check	OK	00:49	

Data Download Date/Time: 01/31/2018 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001289

INSTRUMENT REGISTERED TO: UNION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1824329
DRY GAS EXPIRY DATE: 03/09/2018
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:38

SUBJECT'S LAST NAME: FLICK

DATE/TIME OF TEST: 12/17/2017 20:23:55

Diagnostics Check	OK	20:17
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.206	20:19
Breath Volume	4.527	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.200	20:22
Breath Volume	3.824	20:22
Air Blank	0.000	20:23
Control Test	0.079	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 01/31/2018 15:11

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,CHARLES,C
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WILLIAMS,CHARLES
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: LOUDERMILK

DATE/TIME OF TEST: 12/06/2017 00:23:55

Diagnostics Check	OK	00:20	
Air Blank	0.000	00:21	
Control Test	0.079	00:21	
Air Blank	0.000	00:22	
Subject Sample	# 1	00:22	Subject Test Refused
Breath Volume	0.000	00:22	
Air Blank	0.000	00:23	
Control Test	0.078	00:23	
Air Blank	0.000	00:23	
Diagnostics Check	OK	00:23	

Data Download Date/Time: 12/26/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FROST,KEVIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: POWELL,JAMES
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: GIVENS

DATE/TIME OF TEST: 12/08/2017 23:44:02

Diagnostics Check	OK	23:37
Air Blank	0.000	23:37
Control Test	0.080	23:38
Air Blank	0.000	23:38
Subject Sample # 1	0.200	23:39
Breath Volume	1.562	23:39
Air Blank	0.000	23:40
Air Blank	0.000	23:41
Subject Sample # 2	0.196	23:42
Breath Volume	1.710	23:42
Air Blank	0.000	23:43
Control Test	0.078	23:43
Air Blank	0.000	23:43
Diagnostics Check	OK	23:44

Data Download Date/Time: 12/26/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WILLIAMS,TERRELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MOODY

DATE/TIME OF TEST: 12/15/2017 22:58:52

Diagnostics Check	OK	22:50	
Air Blank	0.000	22:50	
Control Test	0.079	22:51	
Air Blank	0.000	22:51	
Subject Sample	# 1	22:54	No Sample Provided
Breath Volume	0.000	22:54	
Air Blank	0.000	22:55	
Air Blank	0.000	22:57	
Subject Sample	# 2	22:57	Subject Test Refused
Breath Volume	0.000	22:57	
Air Blank	0.000	22:57	
Control Test	0.079	22:58	
Air Blank	0.000	22:58	
Diagnostics Check	OK	22:58	

Data Download Date/Time: 12/26/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBINSON,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: ROBINSON,KYLE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 07:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:29

SUBJECT'S LAST NAME: BOLES

DATE/TIME OF TEST: 12/20/2017 07:57:13

Diagnostics Check	OK	07:54	
Air Blank	0.000	07:54	
Control Test	0.080	07:54	
Air Blank	0.000	07:55	
Subject Sample	# 1	07:55	Subject Test Refused
Breath Volume	0.000	07:55	
Air Blank	0.000	07:56	
Control Test	0.080	07:56	
Air Blank	0.000	07:57	
Diagnostics Check	OK	07:57	

Data Download Date/Time: 12/26/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WILLIAMS,TERRELL
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/21/2017 03:02:58

Diagnostics Check	OK	02:52	
Air Blank	0.000	02:53	
Control Test	0.079	02:53	
Air Blank	0.000	02:54	
Subject Sample	# 1	02:57	No Sample Provided
Breath Volume	0.000	02:57	
Air Blank	0.000	02:57	
Air Blank	0.000	02:59	
Subject Sample	# 2	03:01	Subject Test Refused
Breath Volume	0.000	03:01	
Air Blank	0.000	03:02	
Control Test	0.080	03:02	
Air Blank	0.000	03:02	
Diagnostics Check	OK	03:02	

Data Download Date/Time: 12/26/2017 12:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HULLENDER,ROBERT,S
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: OTTOMEYER,JAMES
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: TUSA

DATE/TIME OF TEST: 12/23/2017 03:29:56

Diagnostics Check	OK	03:23
Air Blank	0.000	03:23
Control Test	0.079	03:23
Air Blank	0.000	03:24
Subject Sample # 1	0.122	03:25
Breath Volume	1.562	03:25
Air Blank	0.000	03:25
Air Blank	0.000	03:27
Subject Sample # 2	0.108	03:28
Breath Volume	1.234	03:28
Air Blank	0.000	03:29
Control Test	0.078	03:29
Air Blank	0.000	03:29
Diagnostics Check	OK	03:29

Data Download Date/Time: 12/26/2017 12:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001302

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KAHALLEY,JAMEY,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: KAHALLEY,JAMEY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 12/31/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 12/31/2017 22:46:32

Diagnostics Check	OK	22:42	
Air Blank	0.000	22:42	
Control Test	0.080	22:42	
Air Blank	0.000	22:43	
Subject Sample	# 1	22:45	Subject Test Refused
Breath Volume	0.000	22:45	
Air Blank	0.000	22:45	
Control Test	0.080	22:45	
Air Blank	0.000	22:46	
Diagnostics Check	OK	22:46	

Data Download Date/Time: 02/01/2018 00:06

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNDON,DAVID,E

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: SIMMONS,CHARLES

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A2

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 18:10

GENDER: Male

OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/01/2017 19:01:24

Diagnostics Check	OK	18:58	
Air Blank	0.000	18:58	
Control Test	0.078	18:59	
Air Blank	0.000	18:59	
Subject Sample	# 1	19:00	Subject Test Refused
Breath Volume	0.000	19:00	
Air Blank	0.000	19:00	
Control Test	0.079	19:00	
Air Blank	0.000	19:01	
Diagnostics Check	OK	19:01	

Data Download Date/Time: 02/09/2018 11:08

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD

SOFTWARE VERSION: 8100.27

BREATH TEST OPERATOR: HERNDON,DAVID,E

BREATH TEST OPERATOR AGENCY: Marianna PD

ARRESTING OFFICER: HERNDON,DAVID

ARRESTING OFFICER AGENCY: Marianna PD

LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No

DRY GAS STANDARD LOT #: 32316080A2

DRY GAS EXPIRY DATE: 01/05/2019

VIOLATION CODE: DUI

ARREST TIME: 01:42

GENDER: Male

OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: QAMAR

DATE/TIME OF TEST: 12/12/2017 03:07:42

Diagnostics Check	OK	02:54	
Air Blank	0.000	02:54	
Control Test	0.080	02:54	
Air Blank	0.000	02:55	
Subject Sample	# 1	02:58	No Sample Provided
Breath Volume	0.589	02:58	
Air Blank	0.000	02:59	
Air Blank	0.000	03:01	
Subject Sample	# 2	03:02	
Breath Volume	1.527	03:02	
Air Blank	0.000	03:03	
Air Blank	0.000	03:05	
Subject Sample	# 3	03:06	
Breath Volume	1.121	03:06	
Air Blank	0.000	03:06	
Control Test	0.077	03:07	
Air Blank	0.000	03:07	
Diagnostics Check	OK	03:07	

Data Download Date/Time: 02/09/2018 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001305

INSTRUMENT REGISTERED TO: MARIANNA PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COWAN,RICHARD,T
BREATH TEST OPERATOR AGENCY: Jackson County SO
ARRESTING OFFICER: COWAN,RICHARD
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: DURAN-ROSAS

DATE/TIME OF TEST: 12/24/2017 02:43:41

Diagnostics Check	OK	02:37
Air Blank	0.000	02:38
Control Test	0.080	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.107	02:39
Breath Volume	2.523	02:39
Air Blank	0.000	02:39
Air Blank	0.000	02:41
Subject Sample # 2	0.101	02:42
Breath Volume	2.312	02:42
Air Blank	0.000	02:42
Control Test	0.078	02:43
Air Blank	0.000	02:43
Diagnostics Check	OK	02:43

Data Download Date/Time: 02/09/2018 11:08

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/03/2017 04:04:08

Diagnostics Check	OK	03:57
Air Blank	0.000	03:58
Control Test	0.079	03:58
Air Blank	0.000	03:58
Subject Sample # 1	0.180	03:59
Breath Volume	1.617	03:59
Air Blank	0.000	04:00
Air Blank	0.000	04:01
Subject Sample # 2	0.171	04:02
Breath Volume	2.417	04:02
Air Blank	0.000	04:03
Control Test	0.079	04:03
Air Blank	0.000	04:04
Diagnostics Check	OK	04:04

Data Download Date/Time: 12/22/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/03/2017 22:43:09

Diagnostics Check	OK	22:40	
Air Blank	0.000	22:40	
Control Test	0.080	22:41	
Air Blank	0.000	22:41	
Subject Sample	# 1	22:41	Subject Test Refused
Breath Volume	0.000	22:41	
Air Blank	0.000	22:42	
Control Test	0.079	22:42	
Air Blank	0.000	22:43	
Diagnostics Check	OK	22:43	

Data Download Date/Time: 12/22/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,JOHN,J
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BOHRER,MARK
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:23

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 12/05/2017 05:53:49

Diagnostics Check		OK	05:44
Air Blank		0.000	05:44
Control Test		0.080	05:45
Air Blank		0.000	05:45
Subject Sample	# 1	0.130	05:46
Breath Volume		2.332	05:46
Air Blank		0.000	05:46
Air Blank		0.000	05:48
Subject Sample	# 2	0.108	05:49
Breath Volume		1.191	05:49
Air Blank		0.000	05:49
Air Blank		0.000	05:51
Subject Sample	# 3	0.131	05:52
Breath Volume		1.875	05:52
Air Blank		0.000	05:52
Control Test		0.079	05:53
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 12/22/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOYN,MATTHEW,S
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: MERRITT,DAKOTA
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: TUCKER

DATE/TIME OF TEST: 12/07/2017 16:44:26

Diagnostics Check	OK	16:37
Air Blank	0.000	16:38
Control Test	0.080	16:38
Air Blank	0.000	16:39
Subject Sample # 1	0.012	16:39
Breath Volume	1.500	16:39
Air Blank	0.000	16:40
Air Blank	0.000	16:42
Subject Sample # 2	0.011	16:43
Breath Volume	1.832	16:43
Air Blank	0.000	16:43
Control Test	0.080	16:43
Air Blank	0.000	16:44
Diagnostics Check	OK	16:44

Data Download Date/Time: 12/22/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:28

SUBJECT'S LAST NAME: LAMBERT

DATE/TIME OF TEST: 12/09/2017 06:51:42

Diagnostics Check	OK	06:48	
Air Blank	0.000	06:48	
Control Test	0.080	06:49	
Air Blank	0.000	06:49	
Subject Sample	# 1	06:50	Subject Test Refused
Breath Volume	0.000	06:50	
Air Blank	0.000	06:50	
Control Test	0.080	06:51	
Air Blank	0.000	06:51	
Diagnostics Check	OK	06:51	

Data Download Date/Time: 12/22/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOBER,JOSHUA,P
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: GOBER,JOSHUA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 12/15/2017 03:48:25

Diagnostics Check	OK	03:45	
Air Blank	0.000	03:46	
Control Test	0.078	03:46	
Air Blank	0.000	03:46	
Subject Sample	# 1	03:47	Subject Test Refused
Breath Volume	0.000	03:47	
Air Blank	0.000	03:47	
Control Test	0.080	03:47	
Air Blank	0.000	03:48	
Diagnostics Check	OK	03:48	

Data Download Date/Time: 12/22/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,CARLON,E
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PRICE,CARLON
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: BOWEN

DATE/TIME OF TEST: 12/16/2017 02:48:18

Diagnostics Check	OK	02:42
Air Blank	0.000	02:42
Control Test	0.081	02:42
Air Blank	0.000	02:43
Subject Sample # 1	0.087	02:43
Breath Volume	4.101	02:43
Air Blank	0.000	02:44
Air Blank	0.000	02:46
Subject Sample # 2	0.083	02:46
Breath Volume	3.972	02:46
Air Blank	0.000	02:47
Control Test	0.078	02:47
Air Blank	0.000	02:48
Diagnostics Check	OK	02:48

Data Download Date/Time: 12/22/2017 12:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: YOUNG,A
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:19

SUBJECT'S LAST NAME: HARVELL

DATE/TIME OF TEST: 12/17/2017 00:42:30

Diagnostics Check	OK	00:39	
Air Blank	0.000	00:40	
Control Test	0.080	00:40	
Air Blank	0.000	00:40	
Subject Sample	# 1	00:41	Subject Test Refused
Breath Volume	0.000	00:41	
Air Blank	0.000	00:41	
Control Test	0.081	00:41	
Air Blank	0.000	00:42	
Diagnostics Check	OK	00:42	

Data Download Date/Time: 12/22/2017 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSIUS,STANLEY,W
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: PIERCE,B
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: CONWAY

DATE/TIME OF TEST: 12/17/2017 06:14:30

Diagnostics Check		OK	06:08
Air Blank		0.000	06:08
Control Test		0.080	06:08
Air Blank		0.000	06:09
Subject Sample	# 1	0.115	06:10
Breath Volume		1.582	06:10
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.111	06:12
Breath Volume		1.519	06:12
Air Blank		0.000	06:13
Control Test		0.079	06:13
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 12/22/2017 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HELLETT,KARL,R
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: LETCHER,CROWMELL
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 11:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:40

SUBJECT'S LAST NAME: PURVIANCE

DATE/TIME OF TEST: 12/19/2017 13:13:12

Diagnostics Check	OK	13:06
Air Blank	0.000	13:06
Control Test	0.080	13:07
Air Blank	0.000	13:07
Subject Sample # 1	0.220	13:08
Breath Volume	2.300	13:08
Air Blank	0.000	13:09
Air Blank	0.000	13:11
Subject Sample # 2	0.201	13:11
Breath Volume	1.894	13:11
Air Blank	0.000	13:12
Control Test	0.079	13:12
Air Blank	0.000	13:13
Diagnostics Check	OK	13:13

Data Download Date/Time: 12/22/2017 12:17

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 12/21/2017 09:16:51

Diagnostics Check		OK	09:10
Air Blank		0.000	09:10
Control Test		0.078	09:11
Air Blank		0.000	09:11
Subject Sample	# 1	0.201	09:12
Breath Volume		2.593	09:12
Air Blank		0.000	09:13
Air Blank		0.000	09:14
Subject Sample	# 2	0.205	09:15
Breath Volume		2.339	09:15
Air Blank		0.000	09:15
Control Test		0.078	09:16
Air Blank		0.000	09:16
Diagnostics Check		OK	09:16

Data Download Date/Time: 12/22/2017 12:17

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,PATRICK,T
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: PIERCE,BRAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: SHAWEN

DATE/TIME OF TEST: 12/23/2017 04:04:53

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.080	04:02	
Air Blank	0.000	04:03	
Subject Sample	# 1	04:03	Subject Test Refused
Breath Volume	0.000	04:03	
Air Blank	0.000	04:03	
Control Test	0.080	04:04	
Air Blank	0.000	04:04	
Diagnostics Check	OK	04:04	

Data Download Date/Time: 02/16/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: SLAY

DATE/TIME OF TEST: 12/30/2017 01:48:33

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.033	01:44
Breath Volume		1.246	01:44
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.033	01:47
Breath Volume		1.207	01:47
Air Blank		0.000	01:47
Control Test		0.081	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 02/16/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRELL,CHAD,A
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: HARRELL,CHAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: NORMAN

DATE/TIME OF TEST: 12/30/2017 04:57:15

Diagnostics Check	OK	04:52	
Air Blank	0.000	04:53	
Control Test	0.080	04:53	
Air Blank	0.000	04:53	
Subject Sample	# 1	04:55	Subject Test Refused
Breath Volume	0.000	04:55	
Air Blank	0.000	04:56	
Control Test	0.081	04:56	
Air Blank	0.000	04:57	
Diagnostics Check	OK	04:57	

Data Download Date/Time: 02/16/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:01

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 12/30/2017 22:31:23

Diagnostics Check	OK	22:22
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Subject Sample # 1	0.267	22:24
Breath Volume	1.136	22:24
Air Blank	0.000	22:24
Air Blank	0.000	22:26
Subject Sample # 2	0.288	22:26
Breath Volume	2.125	22:26
Air Blank	0.000	22:27
Air Blank	0.000	22:29
Subject Sample # 3	0.288	22:29
Breath Volume	2.476	22:29
Air Blank	0.000	22:30
Control Test	0.079	22:30
Air Blank	0.000	22:31
Diagnostics Check	OK	22:31

Data Download Date/Time: 02/16/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001308

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 06:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:21

SUBJECT'S LAST NAME: ARNSPIGER

DATE/TIME OF TEST: 12/31/2017 08:50:45

Diagnostics Check	OK	08:41
Air Blank	0.000	08:42
Control Test	0.079	08:42
Air Blank	0.000	08:42
Subject Sample # 1	0.108	08:43
Breath Volume	1.121	08:43
Air Blank	0.000	08:44
Air Blank	0.000	08:46
Subject Sample # 2	0.115	08:49
Breath Volume	1.765	08:49
Air Blank	0.000	08:49
Control Test	0.079	08:50
Air Blank	0.000	08:50
Diagnostics Check	OK	08:50

Data Download Date/Time: 02/16/2018 11:55

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENNING,LUKE,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: DENNING,LUKE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 10:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:43

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 12/19/2017 11:30:14

Diagnostics Check		OK	11:18	
Air Blank		0.000	11:19	
Control Test		0.080	11:19	
Air Blank		0.000	11:20	
Subject Sample	# 1	0.259	11:23	Volume Not Met
Breath Volume		0.097	11:23	
Air Blank		0.000	11:24	
Air Blank		0.000	11:25	
Subject Sample	# 2		11:28	Subject Test Refused
Breath Volume		0.437	11:28	
Air Blank		0.000	11:29	
Control Test		0.074	11:29	Control Outside Tolerance
Air Blank		0.000	11:30	

Data Download Date/Time: 01/18/2018 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOLNAR,JARROD,T
BREATH TEST OPERATOR AGENCY: FFWCC (Panama City)
ARRESTING OFFICER: MOLNAR,JARROD
ARRESTING OFFICER AGENCY: FFWCC (Panama City)
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: WAGNON

DATE/TIME OF TEST: 12/21/2017 05:42:54

Diagnostics Check		OK	05:34	
Air Blank		0.000	05:34	
Control Test		0.080	05:35	
Air Blank		0.000	05:35	
Subject Sample	# 1	0.115	05:38	Volume Not Met
Breath Volume		0.035	05:38	
Air Blank		0.000	05:39	
Air Blank		0.000	05:41	
Subject Sample	# 2		05:41	Subject Test Refused
Breath Volume		0.000	05:41	
Air Blank		0.000	05:41	
Control Test		0.079	05:42	
Air Blank		0.000	05:42	
Diagnostics Check		OK	05:42	

Data Download Date/Time: 01/18/2018 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUTTO,KYLE,K
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: YOUNGHANSE,KYLE
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: DIECKMANN

DATE/TIME OF TEST: 12/21/2017 20:08:16

Diagnostics Check	OK	20:01
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:02
Subject Sample # 1	0.142	20:03
Breath Volume	3.687	20:03
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.143	20:06
Breath Volume	4.054	20:06
Air Blank	0.000	20:07
Control Test	0.079	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 01/18/2018 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GUNN,CONNOR
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: BATRES

DATE/TIME OF TEST: 12/22/2017 01:26:49

Diagnostics Check	OK	01:19
Air Blank	0.000	01:20
Control Test	0.081	01:20
Air Blank	0.000	01:21
Subject Sample # 1	0.142	01:22
Breath Volume	1.867	01:22
Air Blank	0.000	01:23
Air Blank	0.000	01:24
Subject Sample # 2	0.146	01:25
Breath Volume	1.628	01:25
Air Blank	0.000	01:25
Control Test	0.081	01:26
Air Blank	0.000	01:26
Diagnostics Check	OK	01:26

Data Download Date/Time: 01/18/2018 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,STEVEN,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: YOUNGHANSE,KYLE
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: KRECKER

DATE/TIME OF TEST: 12/31/2017 23:48:59

Diagnostics Check	OK	23:42
Air Blank	0.000	23:43
Control Test	0.081	23:43
Air Blank	0.000	23:43
Subject Sample # 1	0.000	23:44
Breath Volume	1.589	23:44
Air Blank	0.000	23:44
Air Blank	0.000	23:46
Subject Sample # 2	0.007	23:47
Breath Volume	1.398	23:47
Air Blank	0.000	23:48
Control Test	0.081	23:48
Air Blank	0.000	23:48
Diagnostics Check	OK	23:48

Data Download Date/Time: 01/18/2018 11:47

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORD, DAVID,
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: PONCEY, KATHLEEN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:05

SUBJECT'S LAST NAME: LORD

DATE/TIME OF TEST: 12/04/2017 12:33:24

Diagnostics Check	OK	12:27
Air Blank	0.000	12:27
Control Test	0.078	12:28
Air Blank	0.000	12:28
Subject Sample # 1	0.036	12:28
Breath Volume	1.199	12:28
Air Blank	0.000	12:29
Air Blank	0.000	12:31
Subject Sample # 2	0.036	12:31
Breath Volume	1.257	12:31
Air Blank	0.000	12:32
Control Test	0.077	12:32
Air Blank	0.000	12:33
Diagnostics Check	OK	12:33

Data Download Date/Time: 12/15/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLISON,STEVEN,L
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COLLINS,FARRON
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:19

SUBJECT'S LAST NAME: DOOLY

DATE/TIME OF TEST: 12/07/2017 22:46:43

Diagnostics Check	OK	22:40
Air Blank	0.000	22:40
Control Test	0.079	22:40
Air Blank	0.000	22:41
Subject Sample # 1	0.143	22:42
Breath Volume	1.835	22:42
Air Blank	0.000	22:42
Air Blank	0.000	22:44
Subject Sample # 2	0.148	22:45
Breath Volume	1.789	22:45
Air Blank	0.000	22:45
Control Test	0.079	22:46
Air Blank	0.000	22:46
Diagnostics Check	OK	22:46

Data Download Date/Time: 12/15/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,JEFFREY,A
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: BOBLITT,RICK
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/09/2017 21:44:27

Diagnostics Check	OK	21:41	
Air Blank	0.000	21:41	
Control Test	0.079	21:42	
Air Blank	0.000	21:42	
Subject Sample	# 1	21:43	Subject Test Refused
Breath Volume	0.000	21:43	
Air Blank	0.000	21:43	
Control Test	0.079	21:43	
Air Blank	0.000	21:44	
Diagnostics Check	OK	21:44	

Data Download Date/Time: 12/15/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,JEFFREY,A
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: BOBLITT,RICK
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/09/2017 22:10:52

Diagnostics Check	OK	22:08	
Air Blank	0.000	22:08	
Control Test	0.079	22:08	
Air Blank	0.000	22:09	
Subject Sample	# 1	22:10	Sequence Aborted
Breath Volume	0.000	22:10	
Air Blank	0.000	22:10	

Data Download Date/Time: 12/15/2017 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITEHEAD,JEFFREY,A
BREATH TEST OPERATOR AGENCY: DeFuniak Springs PD
ARRESTING OFFICER: BOBLITT,RICK
ARRESTING OFFICER AGENCY: DeFuniak Springs PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 12/09/2017 22:21:17

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:15
Subject Sample # 1	0.232	22:16
Breath Volume	1.464	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.238	22:19
Breath Volume	2.230	22:19
Air Blank	0.000	22:20
Control Test	0.078	22:20
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 12/15/2017 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: SIMPSON,TREY
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: HONEYCUTT

DATE/TIME OF TEST: 12/12/2017 19:23:31

Diagnostics Check	OK	19:15
Air Blank	0.000	19:16
Control Test	0.079	19:16
Air Blank	0.000	19:16
Subject Sample # 1	0.192	19:17
Breath Volume	3.269	19:17
Air Blank	0.000	19:18
Air Blank	0.000	19:20
Subject Sample # 2	0.195	19:21
Breath Volume	1.617	19:21
Air Blank	0.000	19:22
Control Test	0.080	19:22
Air Blank	0.000	19:23
Diagnostics Check	OK	19:23

Data Download Date/Time: 12/15/2017 16:33

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE, MITCHELL, W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MCMILLIAN, RUSSELL
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: PETERNELL

DATE/TIME OF TEST: 12/13/2017 20:03:17

Diagnostics Check	OK	19:56
Air Blank	0.000	19:56
Control Test	0.079	19:56
Air Blank	0.000	19:57
Subject Sample # 1	0.200	19:58
Breath Volume	1.367	19:58
Air Blank	0.000	19:58
Air Blank	0.000	20:00
Subject Sample # 2	0.204	20:01
Breath Volume	1.921	20:01
Air Blank	0.000	20:02
Control Test	0.080	20:02
Air Blank	0.000	20:03
Diagnostics Check	OK	20:03

Data Download Date/Time: 12/15/2017 16:33

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,JONATHAN,B
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MILLER,BLAZE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:50

SUBJECT'S LAST NAME: TENORIO

DATE/TIME OF TEST: 12/17/2017 18:31:49

Diagnostics Check	OK	18:24
Air Blank	0.000	18:25
Control Test	0.078	18:25
Air Blank	0.000	18:26
Subject Sample # 1	0.243	18:27
Breath Volume	1.933	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.242	18:30
Breath Volume	2.199	18:30
Air Blank	0.000	18:30
Control Test	0.079	18:31
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 01/18/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGAN,JAMIE,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: MOUNT,DEBORAH
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: PERRIN

DATE/TIME OF TEST: 12/24/2017 20:37:58

Diagnostics Check	OK	20:31
Air Blank	0.000	20:31
Control Test	0.078	20:32
Air Blank	0.000	20:32
Subject Sample # 1	0.291	20:33
Breath Volume	2.972	20:33
Air Blank	0.000	20:34
Air Blank	0.000	20:35
Subject Sample # 2	0.288	20:36
Breath Volume	2.855	20:36
Air Blank	0.000	20:37
Control Test	0.077	20:37
Air Blank	0.000	20:37
Diagnostics Check	OK	20:37

Data Download Date/Time: 01/18/2018 11:42

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROGERS,DAVID,E
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: OTT,TYLER
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: SUMMERFORD

DATE/TIME OF TEST: 12/11/2017 23:08:55

Diagnostics Check	OK	23:04	
Air Blank	0.000	23:04	
Control Test	0.078	23:04	
Air Blank	0.000	23:05	
Subject Sample	# 1	23:07	Subject Test Refused
Breath Volume	0.000	23:07	
Air Blank	0.000	23:08	
Control Test	0.081	23:08	
Air Blank	0.000	23:08	
Diagnostics Check	OK	23:08	

Data Download Date/Time: 01/11/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:01

SUBJECT'S LAST NAME: MENDOZA YANEZ

DATE/TIME OF TEST: 12/17/2017 12:32:01

Diagnostics Check	OK	12:22
Air Blank	0.000	12:22
Control Test	0.080	12:22
Air Blank	0.000	12:23
Subject Sample # 1	0.314	12:24
Breath Volume	3.292	12:24
Air Blank	0.000	12:25
Air Blank	0.000	12:27
Subject Sample # 2	0.291	12:27
Breath Volume	1.757	12:27
Air Blank	0.000	12:28
Air Blank	0.000	12:30
Subject Sample # 3	0.318	12:30
Breath Volume	3.304	12:30
Air Blank	0.000	12:31
Control Test	0.079	12:31
Air Blank	0.000	12:31
Diagnostics Check	OK	12:32

Data Download Date/Time: 01/11/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAJERE,ARIC,
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: SATALLANTE,DEVON
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: CHURCH

DATE/TIME OF TEST: 12/28/2017 23:34:42

Diagnostics Check	OK	23:28
Air Blank	0.000	23:29
Control Test	0.080	23:29
Air Blank	0.000	23:29
Subject Sample # 1	0.136	23:30
Breath Volume	3.082	23:30
Air Blank	0.000	23:30
Air Blank	0.000	23:32
Subject Sample # 2	0.134	23:33
Breath Volume	2.500	23:33
Air Blank	0.000	23:33
Control Test	0.080	23:34
Air Blank	0.000	23:34
Diagnostics Check	OK	23:34

Data Download Date/Time: 01/11/2018 10:25

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 16:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: MOYES

DATE/TIME OF TEST: 12/29/2017 18:20:49

Diagnostics Check	OK	18:14
Air Blank	0.000	18:15
Control Test	0.080	18:15
Air Blank	0.000	18:15
Subject Sample # 1	0.412	18:16
Breath Volume	2.125	18:16
Air Blank	0.000	18:17
Air Blank	0.000	18:18
Subject Sample # 2	0.415	18:19
Breath Volume	2.199	18:19
Air Blank	0.000	18:19
Control Test	0.080	18:20
Air Blank	0.000	18:20
Diagnostics Check	OK	18:20

Data Download Date/Time: 01/11/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 18:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:01

SUBJECT'S LAST NAME: PEREZ GALICIA

DATE/TIME OF TEST: 12/30/2017 19:30:22

Diagnostics Check	OK	19:21
Air Blank	0.000	19:21
Control Test	0.079	19:22
Air Blank	0.000	19:22
Subject Sample # 1	0.381	19:23
Breath Volume	3.230	19:23
Air Blank	0.000	19:23
Air Blank	0.000	19:25
Subject Sample # 2	0.354	19:25
Breath Volume	1.257	19:25
Air Blank	0.000	19:26
Air Blank	0.000	19:28
Subject Sample # 3	0.364	19:28
Breath Volume	3.343	19:28
Air Blank	0.000	19:29
Control Test	0.079	19:29
Air Blank	0.000	19:30
Diagnostics Check	OK	19:30

Data Download Date/Time: 01/11/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 12/04/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: FLEENOR GROOMS

DATE/TIME OF TEST: 12/31/2017 01:33:26

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.216	01:28
Breath Volume		1.972	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.199	01:31
Breath Volume		1.773	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 01/11/2018 10:25

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARGERUM,KELLEY,J
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: DIXON,DEMARCUS
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MARSHALL

DATE/TIME OF TEST: 12/23/2017 05:35:58

Diagnostics Check		OK	05:24	
Air Blank		0.000	05:24	
Control Test		0.079	05:24	
Air Blank		0.000	05:25	
Subject Sample	# 1	0.154	05:28	Volume Not Met
Breath Volume		0.015	05:28	
Air Blank		0.000	05:29	
Air Blank		0.000	05:31	
Subject Sample	# 2	0.181	05:34	Volume Not Met
Breath Volume		0.164	05:34	
Air Blank		0.000	05:35	
Control Test		0.075	05:35	
Air Blank		0.000	05:35	
Diagnostics Check		OK	05:35	

Data Download Date/Time: 12/31/2017 13:07

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: THOMPSON,BRIAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: GILCES-TUMBACO

DATE/TIME OF TEST: 12/15/2017 22:43:42

Diagnostics Check	OK	22:37
Air Blank	0.000	22:37
Control Test	0.080	22:37
Air Blank	0.000	22:38
Subject Sample # 1	0.223	22:39
Breath Volume	1.476	22:39
Air Blank	0.000	22:39
Air Blank	0.000	22:41
Subject Sample # 2	0.223	22:42
Breath Volume	1.300	22:42
Air Blank	0.000	22:42
Control Test	0.079	22:43
Air Blank	0.000	22:43
Diagnostics Check	OK	22:43

Data Download Date/Time: 12/31/2017 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TREVISOL,JOSEPH,D
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TREVISOL,JOSEPH
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/16/2017 21:08:43

Diagnostics Check	OK	21:02
Air Blank	0.000	21:02
Control Test	0.079	21:03
Air Blank	0.000	21:03
Subject Sample # 1	0.083	21:04
Breath Volume	2.839	21:04
Air Blank	0.000	21:04
Air Blank	0.000	21:06
Subject Sample # 2	0.078	21:07
Breath Volume	1.902	21:07
Air Blank	0.000	21:07
Control Test	0.079	21:08
Air Blank	0.000	21:08
Diagnostics Check	OK	21:08

Data Download Date/Time: 12/31/2017 12:51

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,BRIAN,J
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: REYNOLDS,MATTHEW
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: BETANCOURT

DATE/TIME OF TEST: 12/23/2017 01:07:51

Diagnostics Check	OK	01:01
Air Blank	0.000	01:01
Control Test	0.080	01:01
Air Blank	0.000	01:02
Subject Sample # 1	0.179	01:03
Breath Volume	2.105	01:03
Air Blank	0.000	01:03
Air Blank	0.000	01:05
Subject Sample # 2	0.175	01:06
Breath Volume	1.441	01:06
Air Blank	0.000	01:06
Control Test	0.079	01:07
Air Blank	0.000	01:07
Diagnostics Check	OK	01:07

Data Download Date/Time: 12/31/2017 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001323

INSTRUMENT REGISTERED TO: PORT ST LUCIE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRUDEAU,RYAN,S
BREATH TEST OPERATOR AGENCY: Port St. Lucie PD
ARRESTING OFFICER: TRUDEAU,RYAN
ARRESTING OFFICER AGENCY: Port St. Lucie PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: FINK

DATE/TIME OF TEST: 12/24/2017 19:37:49

Diagnostics Check	OK	19:31
Air Blank	0.000	19:32
Control Test	0.080	19:32
Air Blank	0.000	19:33
Subject Sample # 1	0.288	19:33
Breath Volume	3.226	19:33
Air Blank	0.000	19:34
Air Blank	0.000	19:36
Subject Sample # 2	0.297	19:36
Breath Volume	2.875	19:36
Air Blank	0.000	19:36
Control Test	0.078	19:37
Air Blank	0.000	19:37
Diagnostics Check	OK	19:37

Data Download Date/Time: 12/31/2017 12:51

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN, R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: TEITELBAUM, D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: LARROUDE

DATE/TIME OF TEST: 12/01/2017 00:02:28

Diagnostics Check	OK	23:59	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Subject Sample	# 1	00:01	Subject Test Refused
Breath Volume	0.000	00:01	
Air Blank	0.000	00:01	
Control Test	0.079	00:01	
Air Blank	0.000	00:02	
Diagnostics Check	OK	00:02	

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, MATTHEW, T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ERVIN

DATE/TIME OF TEST: 12/02/2017 03:32:55

Diagnostics Check	OK	03:25
Air Blank	0.000	03:26
Control Test	0.081	03:26
Air Blank	0.000	03:27
Subject Sample # 1	0.173	03:28
Breath Volume	1.605	03:28
Air Blank	0.000	03:28
Air Blank	0.000	03:30
Subject Sample # 2	0.156	03:31
Breath Volume	1.960	03:31
Air Blank	0.000	03:32
Control Test	0.080	03:32
Air Blank	0.000	03:32
Diagnostics Check	OK	03:32

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, MATTHEW, T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 12/03/2017 03:30:47

Diagnostics Check	OK	03:30	
Air Blank		03:30	RFI Detect
Air Blank		03:30	RFI Detect

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, MATTHEW, T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 12/03/2017 03:33:54

Diagnostics Check	OK	03:33	
Air Blank		03:33	RFI Detect
Air Blank		03:33	RFI Detect

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GALLAGHER, MATTHEW, T
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GALLAGHER, MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: MCLEOD

DATE/TIME OF TEST: 12/03/2017 03:44:12

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.080	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.175	03:39
Breath Volume		2.289	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:42
Subject Sample	# 2	0.169	03:42
Breath Volume		2.148	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Diagnostics Check		OK	03:44

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EVANS,Z
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: HUDMON

DATE/TIME OF TEST: 12/05/2017 23:20:19

Diagnostics Check	OK	23:17	
Air Blank	0.000	23:17	
Control Test	0.081	23:18	
Air Blank	0.000	23:18	
Subject Sample	# 1	23:18	Subject Test Refused
Breath Volume	0.000	23:18	
Air Blank	0.000	23:19	
Control Test	0.080	23:19	
Air Blank	0.000	23:20	
Diagnostics Check	OK	23:20	

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATFIELD,CHAYSE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: KOTECKI

DATE/TIME OF TEST: 12/06/2017 02:43:25

Diagnostics Check	OK	02:40	
Air Blank	0.000	02:41	
Control Test	0.080	02:41	
Air Blank	0.000	02:41	
Subject Sample	# 1	02:42	Subject Test Refused
Breath Volume	0.000	02:42	
Air Blank	0.000	02:42	
Control Test	0.080	02:42	
Air Blank	0.000	02:43	
Diagnostics Check	OK	02:43	

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD, THERESA, K
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: ROBINSON, D
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 08:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: LEWARK

DATE/TIME OF TEST: 12/07/2017 09:09:25

Diagnostics Check	OK	09:03
Air Blank	0.000	09:03
Control Test	0.080	09:04
Air Blank	0.000	09:04
Subject Sample # 1	0.000	09:05
Breath Volume	1.562	09:05
Air Blank	0.000	09:05
Air Blank	0.000	09:07
Subject Sample # 2	0.000	09:08
Breath Volume	1.597	09:08
Air Blank	0.000	09:08
Control Test	0.080	09:08
Air Blank	0.000	09:09
Diagnostics Check	OK	09:09

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY, JOSHUA, J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GENTRY, JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: KEYSER

DATE/TIME OF TEST: 12/09/2017 04:11:44

Diagnostics Check	OK	04:05
Air Blank	0.000	04:05
Control Test	0.081	04:05
Air Blank	0.000	04:06
Subject Sample # 1	0.134	04:06
Breath Volume	2.507	04:06
Air Blank	0.000	04:07
Air Blank	0.000	04:09
Subject Sample # 2	0.129	04:10
Breath Volume	1.652	04:10
Air Blank	0.000	04:10
Control Test	0.080	04:11
Air Blank	0.000	04:11
Diagnostics Check	OK	04:11

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK, JIMMY, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: WILLIAMS, ROBERT
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: LETA

DATE/TIME OF TEST: 12/11/2017 00:23:25

Diagnostics Check	OK	00:16
Air Blank	0.000	00:17
Control Test	0.082	00:17
Air Blank	0.000	00:18
Subject Sample # 1	0.146	00:18
Breath Volume	1.835	00:18
Air Blank	0.000	00:19
Air Blank	0.000	00:21
Subject Sample # 2	0.152	00:21
Breath Volume	1.417	00:21
Air Blank	0.000	00:22
Control Test	0.082	00:22
Air Blank	0.000	00:23
Diagnostics Check	OK	00:23

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ, OSCAR,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BRASHEARS, ERIC
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: NIYOMKUN

DATE/TIME OF TEST: 12/11/2017 01:38:11

Diagnostics Check	OK	01:34	
Air Blank	0.000	01:35	
Control Test	0.081	01:35	
Air Blank	0.000	01:36	
Subject Sample	# 1	01:36	Subject Test Refused
Breath Volume	0.000	01:36	
Air Blank	0.000	01:37	
Control Test	0.082	01:37	
Air Blank	0.000	01:38	
Diagnostics Check	OK	01:38	

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIMAVERA,S,M
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: DALEY,S
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: KRAUSE

DATE/TIME OF TEST: 12/11/2017 04:28:06

Diagnostics Check	OK	04:21
Air Blank	0.000	04:21
Control Test	0.082	04:22
Air Blank	0.000	04:22
Subject Sample # 1	0.096	04:23
Breath Volume	1.777	04:23
Air Blank	0.000	04:24
Air Blank	0.000	04:26
Subject Sample # 2	0.094	04:26
Breath Volume	1.328	04:26
Air Blank	0.000	04:27
Control Test	0.082	04:27
Air Blank	0.000	04:28
Diagnostics Check	OK	04:28

Data Download Date/Time: 12/22/2017 05:30

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,QUINTON,T
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:40

SUBJECT'S LAST NAME: CALDWELL

DATE/TIME OF TEST: 12/11/2017 18:11:18

Diagnostics Check	OK	18:08	
Air Blank	0.000	18:08	
Control Test	0.081	18:09	
Air Blank	0.000	18:09	
Subject Sample	# 1	18:09	Subject Test Refused
Breath Volume	0.000	18:09	
Air Blank	0.000	18:10	
Control Test	0.082	18:10	
Air Blank	0.000	18:11	
Diagnostics Check	OK	18:11	

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JANNETTY,QUINTON
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: BURCH

DATE/TIME OF TEST: 12/13/2017 01:29:16

Diagnostics Check	OK	01:22
Air Blank	0.000	01:22
Control Test	0.083	01:23
Air Blank	0.000	01:23
Subject Sample # 1	0.199	01:24
Breath Volume	1.832	01:24
Air Blank	0.000	01:25
Air Blank	0.000	01:26
Subject Sample # 2	0.208	01:27
Breath Volume	2.679	01:27
Air Blank	0.000	01:28
Control Test	0.082	01:28
Air Blank	0.000	01:29
Diagnostics Check	OK	01:29

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOD,T,K
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MCKENZIE,DS
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:11

SUBJECT'S LAST NAME: DITTUS

DATE/TIME OF TEST: 12/15/2017 11:39:07

Diagnostics Check	OK	11:33
Air Blank	0.000	11:33
Control Test	0.081	11:33
Air Blank	0.000	11:34
Subject Sample # 1	0.000	11:34
Breath Volume	3.402	11:34
Air Blank	0.000	11:35
Air Blank	0.000	11:37
Subject Sample # 2	0.000	11:37
Breath Volume	3.160	11:37
Air Blank	0.000	11:38
Control Test	0.080	11:38
Air Blank	0.000	11:39
Diagnostics Check	OK	11:39

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: NIEVES,L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: IRLE

DATE/TIME OF TEST: 12/15/2017 22:23:38

Diagnostics Check	OK	22:20	
Air Blank	0.000	22:21	
Control Test	0.081	22:21	
Air Blank	0.000	22:22	
Subject Sample	# 1	22:22	Subject Test Refused
Breath Volume	0.000	22:22	
Air Blank	0.000	22:22	
Control Test	0.080	22:23	
Air Blank	0.000	22:23	
Diagnostics Check	OK	22:23	

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SMITH,COLBY
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: MARMOLEJO

DATE/TIME OF TEST: 12/17/2017 04:34:48

Diagnostics Check	OK	04:31	
Air Blank	0.000	04:32	
Control Test	0.081	04:32	
Air Blank	0.000	04:33	
Subject Sample	# 1	04:33	Subject Test Refused
Breath Volume	0.000	04:33	
Air Blank	0.000	04:33	
Control Test	0.080	04:34	
Air Blank	0.000	04:34	
Diagnostics Check	OK	04:34	

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YORK, JIMMY, A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MELIUS, BRYAN
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: WESTON OWENS

DATE/TIME OF TEST: 12/19/2017 01:52:38

Diagnostics Check	OK	01:49	
Air Blank	0.000	01:50	
Control Test	0.080	01:50	
Air Blank	0.000	01:51	
Subject Sample	# 1	01:51	Subject Test Refused
Breath Volume	0.000	01:51	
Air Blank	0.000	01:51	
Control Test	0.080	01:52	
Air Blank	0.000	01:52	
Diagnostics Check	OK	01:52	

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHEELER,MEGAN,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BRASHEARS,ERIC
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 12/20/2017 03:38:51

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.080	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.126	03:34
Breath Volume	2.687	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.117	03:37
Breath Volume	1.984	03:37
Air Blank	0.000	03:37
Control Test	0.080	03:38
Air Blank	0.000	03:38
Diagnostics Check	OK	03:38

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STOVALL,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: MONNEY,T
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 16:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: HOFMAN

DATE/TIME OF TEST: 12/21/2017 17:56:24

Diagnostics Check		OK	17:44
Air Blank		0.000	17:45
Control Test		0.081	17:45
Air Blank		0.000	17:46
Subject Sample	# 1	0.288	17:47
Breath Volume		1.742	17:47
Air Blank		0.000	17:47
Air Blank		0.000	17:49
Subject Sample	# 2	0.313	17:51
Breath Volume		2.824	17:51
Air Blank		0.000	17:51
Air Blank		0.000	17:53
Subject Sample	# 3	0.297	17:54
Breath Volume		1.453	17:54
Air Blank		0.000	17:55
Control Test		0.078	17:55
Air Blank		0.000	17:56
Diagnostics Check		OK	17:56

Data Download Date/Time: 12/22/2017 05:31

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EDER,K
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: BISCHOFF

DATE/TIME OF TEST: 12/22/2017 22:18:04

Diagnostics Check	OK	22:12
Air Blank	0.000	22:12
Control Test	0.081	22:12
Air Blank	0.000	22:13
Subject Sample # 1	0.000	22:13
Breath Volume	3.328	22:13
Air Blank	0.000	22:14
Air Blank	0.000	22:16
Subject Sample # 2	0.000	22:16
Breath Volume	2.695	22:16
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:17
Diagnostics Check	OK	22:18

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,DEVIN,C
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: CHOI,HWAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/23/2017 01:21:10

Diagnostics Check	OK	01:13
Air Blank	0.000	01:13
Control Test	0.080	01:13
Air Blank	0.000	01:14
Subject Sample # 1	0.156	01:16
Breath Volume	2.136	01:16
Air Blank	0.000	01:16
Air Blank	0.000	01:18
Subject Sample # 2	0.154	01:19
Breath Volume	1.785	01:19
Air Blank	0.000	01:20
Control Test	0.080	01:20
Air Blank	0.000	01:21
Diagnostics Check	OK	01:21

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ, MAEGAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER, S
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: BIRKS

DATE/TIME OF TEST: 12/23/2017 19:45:42

Diagnostics Check	OK	19:42	
Air Blank	0.000	19:43	
Control Test	0.081	19:43	
Air Blank	0.000	19:44	
Subject Sample	# 1	19:44	Subject Test Refused
Breath Volume	0.000	19:44	
Air Blank	0.000	19:44	
Control Test	0.081	19:45	
Air Blank	0.000	19:45	
Diagnostics Check	OK	19:45	

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOMINGUEZ,OSCAR,
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BRASHEARS,ERIC
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: SPENCE-THOMAS

DATE/TIME OF TEST: 12/23/2017 22:51:59

Diagnostics Check	OK	22:45
Air Blank	0.000	22:46
Control Test	0.080	22:46
Air Blank	0.000	22:47
Subject Sample # 1	0.237	22:47
Breath Volume	2.160	22:47
Air Blank	0.000	22:48
Air Blank	0.000	22:49
Subject Sample # 2	0.245	22:50
Breath Volume	2.343	22:50
Air Blank	0.000	22:51
Control Test	0.080	22:51
Air Blank	0.000	22:51
Diagnostics Check	OK	22:51

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:38

SUBJECT'S LAST NAME: LUPI

DATE/TIME OF TEST: 12/24/2017 02:16:05

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:01	
Control Test		0.080	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1	0.205	02:04	
Breath Volume		1.812	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 2	0.134	02:10	Volume Not Met
Breath Volume		0.835	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 3	0.191	02:14	
Breath Volume		1.515	02:14	
Air Blank		0.000	02:15	
Control Test		0.078	02:15	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: SHARPETA,C
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/28/2017 21:03:40

Diagnostics Check	OK	20:56
Air Blank	0.000	20:56
Control Test	0.080	20:56
Air Blank	0.000	20:57
Subject Sample # 1	0.181	20:59
Breath Volume	1.410	20:59
Air Blank	0.000	20:59
Air Blank	0.000	21:01
Subject Sample # 2	0.187	21:02
Breath Volume	1.230	21:02
Air Blank	0.000	21:02
Control Test	0.080	21:03
Air Blank	0.000	21:03
Diagnostics Check	OK	21:03

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROWLEY,SEAN,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANI
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/30/2017 04:05:36

Diagnostics Check	OK	04:02	
Air Blank	0.000	04:02	
Control Test	0.081	04:03	
Air Blank	0.000	04:03	
Subject Sample	# 1 0.166	04:03	
Breath Volume	1.570	04:03	
Air Blank		04:04	RFI Detect
Air Blank	0.000	04:04	
Control Test	0.081	04:05	
Air Blank	0.000	04:05	
Diagnostics Check	OK	04:05	

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROWLEY,SEAN,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANI
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/30/2017 04:13:53

Diagnostics Check	OK	04:07
Air Blank	0.000	04:08
Control Test	0.081	04:08
Air Blank	0.000	04:08
Subject Sample # 1	0.166	04:09
Breath Volume	1.304	04:09
Air Blank	0.000	04:10
Air Blank	0.000	04:11
Subject Sample # 2	0.158	04:12
Breath Volume	1.406	04:12
Air Blank	0.000	04:12
Control Test	0.080	04:13
Air Blank	0.000	04:13
Diagnostics Check	OK	04:13

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FERGUSON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 12/30/2017 20:34:13

Diagnostics Check	OK	20:30	
Air Blank	0.000	20:31	
Control Test	0.081	20:31	
Air Blank	0.000	20:31	
Subject Sample	# 1 0.192	20:33	
Breath Volume	1.210	20:33	
Air Blank		20:33	RFI Detect
Air Blank		20:33	RFI Detect
Air Blank	0.000	20:34	

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FERGUSON,M
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: ENGLISH

DATE/TIME OF TEST: 12/30/2017 20:42:36

Diagnostics Check	OK	20:36
Air Blank	0.000	20:36
Control Test	0.080	20:36
Air Blank	0.000	20:37
Subject Sample # 1	0.187	20:38
Breath Volume	1.492	20:38
Air Blank	0.000	20:38
Air Blank	0.000	20:40
Subject Sample # 2	0.191	20:41
Breath Volume	1.722	20:41
Air Blank	0.000	20:41
Control Test	0.080	20:42
Air Blank	0.000	20:42
Diagnostics Check	OK	20:42

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MAEGAN,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 12/31/2017 21:33:06

Diagnostics Check	OK	21:27
Air Blank	0.000	21:27
Control Test	0.081	21:27
Air Blank	0.000	21:28
Subject Sample # 1	0.000	21:28
Breath Volume	1.789	21:28
Air Blank	0.000	21:29
Air Blank	0.000	21:31
Subject Sample # 2	0.000	21:31
Breath Volume	1.453	21:31
Air Blank	0.000	21:32
Control Test	0.081	21:32
Air Blank	0.000	21:33
Diagnostics Check	OK	21:33

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/31/2017 23:53:16

Diagnostics Check	OK	23:52	
Air Blank	0.000	23:52	
Control Test	0.000	23:53	RFI Detect
Air Blank		23:53	RFI Detect

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/31/2017 23:55:18

Diagnostics Check	OK	23:55	
Air Blank		23:55	RFI Detect
Air Blank		23:55	RFI Detect

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 12/31/2017 23:57:17

Diagnostics Check	OK	23:57	
Air Blank		23:57	RFI Detect
Air Blank		23:57	RFI Detect

Data Download Date/Time: 01/30/2018 07:50

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL, MATTHEW, A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL, MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 12/07/2017 20:23:55

Diagnostics Check	OK	20:18
Air Blank	0.000	20:18
Control Test	0.080	20:18
Air Blank	0.000	20:19
Subject Sample # 1	0.000	20:19
Breath Volume	2.152	20:19
Air Blank	0.000	20:20
Air Blank	0.000	20:22
Subject Sample # 2	0.000	20:22
Breath Volume	2.144	20:22
Air Blank	0.000	20:23
Control Test	0.081	20:23
Air Blank	0.000	20:23
Diagnostics Check	OK	20:23

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: VALARIO-POPOCA

DATE/TIME OF TEST: 12/08/2017 04:57:17

Diagnostics Check	OK	04:50
Air Blank	0.000	04:51
Control Test	0.081	04:51
Air Blank	0.000	04:51
Subject Sample # 1	0.266	04:52
Breath Volume	3.652	04:52
Air Blank	0.000	04:53
Air Blank	0.000	04:55
Subject Sample # 2	0.267	04:55
Breath Volume	3.503	04:55
Air Blank	0.000	04:56
Control Test	0.080	04:56
Air Blank	0.000	04:57
Diagnostics Check	OK	04:57

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:39

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/14/2017 04:14:01

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.218	04:09
Breath Volume		2.992	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:11
Subject Sample	# 2	0.214	04:12
Breath Volume		4.097	04:12
Air Blank		0.000	04:13
Control Test		0.080	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:14

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MORALES,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:04

SUBJECT'S LAST NAME: NAGLE

DATE/TIME OF TEST: 12/15/2017 01:54:22

Diagnostics Check	OK	01:48
Air Blank	0.000	01:48
Control Test	0.079	01:48
Air Blank	0.000	01:49
Subject Sample # 1	0.165	01:49
Breath Volume	2.957	01:49
Air Blank	0.000	01:50
Air Blank	0.000	01:52
Subject Sample # 2	0.165	01:52
Breath Volume	2.312	01:52
Air Blank	0.000	01:53
Control Test	0.078	01:53
Air Blank	0.000	01:54
Diagnostics Check	OK	01:54

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MITCHELL,A
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: REID III

DATE/TIME OF TEST: 12/15/2017 04:02:10

Diagnostics Check	OK	03:58	
Air Blank	0.000	03:59	
Control Test	0.079	03:59	
Air Blank	0.000	04:00	
Subject Sample	# 1	04:00	Subject Test Refused
Breath Volume	0.000	04:00	
Air Blank	0.000	04:01	
Control Test	0.081	04:01	
Air Blank	0.000	04:02	
Diagnostics Check	OK	04:02	

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SHOPE,K
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: SOMERS

DATE/TIME OF TEST: 12/16/2017 04:08:10

Diagnostics Check	OK	04:01
Air Blank	0.000	04:02
Control Test	0.080	04:02
Air Blank	0.000	04:03
Subject Sample # 1	0.057	04:03
Breath Volume	4.121	04:03
Air Blank	0.000	04:04
Air Blank	0.000	04:06
Subject Sample # 2	0.055	04:06
Breath Volume	3.500	04:06
Air Blank	0.000	04:07
Control Test	0.081	04:07
Air Blank	0.000	04:08
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NICK,A,
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: NICK,A
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 12/19/2017 21:29:32

Diagnostics Check	OK	21:23
Air Blank	0.000	21:24
Control Test	0.080	21:24
Air Blank	0.000	21:24
Subject Sample # 1	0.000	21:25
Breath Volume	5.539	21:25
Air Blank	0.000	21:25
Air Blank	0.000	21:27
Subject Sample # 2	0.000	21:28
Breath Volume	4.738	21:28
Air Blank	0.000	21:28
Control Test	0.080	21:28
Air Blank	0.000	21:29
Diagnostics Check	OK	21:29

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANIK,JONATHAN,F
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: COLLINS,RIDGE
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:32

SUBJECT'S LAST NAME: DI MASCIO

DATE/TIME OF TEST: 12/21/2017 21:46:45

Diagnostics Check	OK	21:40
Air Blank	0.000	21:41
Control Test	0.080	21:41
Air Blank	0.000	21:42
Subject Sample # 1	0.161	21:42
Breath Volume	1.421	21:42
Air Blank	0.000	21:43
Air Blank	0.000	21:44
Subject Sample # 2	0.156	21:45
Breath Volume	1.484	21:45
Air Blank	0.000	21:45
Control Test	0.079	21:46
Air Blank	0.000	21:46
Diagnostics Check	OK	21:46

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOBAS,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARKEY,L
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/22/2017 19:39:19

Diagnostics Check	OK	19:32
Air Blank	0.000	19:33
Control Test	0.080	19:33
Air Blank	0.000	19:34
Subject Sample # 1	0.198	19:34
Breath Volume	3.210	19:34
Air Blank	0.000	19:35
Air Blank	0.000	19:37
Subject Sample # 2	0.187	19:37
Breath Volume	3.082	19:37
Air Blank	0.000	19:38
Control Test	0.080	19:38
Air Blank	0.000	19:39
Diagnostics Check	OK	19:39

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: VAN HORN

DATE/TIME OF TEST: 12/23/2017 02:02:51

Diagnostics Check	OK	01:56
Air Blank	0.000	01:56
Control Test	0.079	01:57
Air Blank	0.000	01:57
Subject Sample # 1	0.102	01:58
Breath Volume	3.179	01:58
Air Blank	0.000	01:58
Air Blank	0.000	02:00
Subject Sample # 2	0.099	02:01
Breath Volume	3.378	02:01
Air Blank	0.000	02:01
Control Test	0.080	02:02
Air Blank	0.000	02:02
Diagnostics Check	OK	02:02

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLINT,SERGEANT,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FLINT,SERGEANT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: CLINGER

DATE/TIME OF TEST: 12/23/2017 04:08:04

Diagnostics Check	OK	04:00
Air Blank	0.000	04:01
Control Test	0.080	04:01
Air Blank	0.000	04:02
Subject Sample # 1	0.142	04:03
Breath Volume	3.078	04:03
Air Blank	0.000	04:03
Air Blank	0.000	04:05
Subject Sample # 2	0.138	04:06
Breath Volume	2.359	04:06
Air Blank	0.000	04:07
Control Test	0.079	04:07
Air Blank	0.000	04:07
Diagnostics Check	OK	04:08

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CHRISTIAN,D
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 12/25/2017 01:53:15

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.079	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.301	01:45
Breath Volume		1.164	01:45
Air Blank		0.000	01:46
Air Blank		0.000	01:47
Subject Sample	# 2	0.274	01:48
Breath Volume		1.625	01:48
Air Blank		0.000	01:49
Air Blank		0.000	01:50
Subject Sample	# 3	0.298	01:51
Breath Volume		2.351	01:51
Air Blank		0.000	01:52
Control Test		0.079	01:52
Air Blank		0.000	01:53
Diagnostics Check		OK	01:53

Data Download Date/Time: 12/29/2017 07:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERR,CALEB,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: KERR,CALEB
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 12/31/2017 06:31:34

Diagnostics Check		OK	06:18	
Air Blank		0.000	06:18	
Control Test		0.079	06:19	
Air Blank		0.000	06:19	
Subject Sample	# 1	0.208	06:20	No .020 agreement
Breath Volume		1.824	06:20	
Air Blank		0.000	06:21	
Air Blank		0.000	06:23	
Subject Sample	# 2	0.174	06:26	Volume Not Met
Breath Volume		0.035	06:26	
Air Blank		0.000	06:27	
Air Blank		0.000	06:28	
Subject Sample	# 3	0.184	06:29	No .020 agreement
Breath Volume		2.019	06:29	
Air Blank		0.000	06:30	
Control Test		0.077	06:30	
Air Blank		0.000	06:31	
Diagnostics Check		OK	06:31	

Data Download Date/Time: 01/31/2018 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS, MICHAEL, K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TUCK, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: THAMER

DATE/TIME OF TEST: 12/31/2017 16:22:06

Diagnostics Check	OK	16:10	
Air Blank	0.000	16:10	
Control Test	0.080	16:10	
Air Blank	0.000	16:11	
Subject Sample	# 1	16:14	No Sample Provided
Breath Volume	0.000	16:14	
Air Blank	0.000	16:14	
Air Blank	0.000	16:16	
Subject Sample	# 2 0.212	16:20	Volume Not Met
Breath Volume	0.152	16:20	
Air Blank	0.000	16:21	
Control Test	0.079	16:21	
Air Blank	0.000	16:22	
Diagnostics Check	OK	16:22	

Data Download Date/Time: 01/31/2018 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS, MICHAEL, K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TUCK, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: THAMER

DATE/TIME OF TEST: 12/31/2017 16:41:13

Diagnostics Check	OK	16:23	
Air Blank	0.000	16:24	
Control Test	0.079	16:24	
Air Blank	0.000	16:24	
Subject Sample	# 1 0.210	16:28	
Breath Volume	1.601	16:28	
Air Blank	0.000	16:29	
Air Blank	0.000	16:30	
Subject Sample	# 2 0.195	16:33	Volume Not Met
Breath Volume	0.101	16:33	
Air Blank	0.000	16:34	
Air Blank	0.000	16:36	
Subject Sample	# 3 0.180	16:39	Volume Not Met
Breath Volume	0.742	16:39	
Air Blank	0.000	16:40	
Control Test	0.077	16:40	
Air Blank	0.000	16:41	
Diagnostics Check	OK	16:41	

Data Download Date/Time: 01/31/2018 07:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSS, MICHAEL, K
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TUCK, DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:42

SUBJECT'S LAST NAME: THAMER

DATE/TIME OF TEST: 12/31/2017 16:51:19

Diagnostics Check	OK	16:42	
Air Blank	0.000	16:43	
Control Test	0.078	16:43	
Air Blank	0.000	16:44	
Subject Sample	# 1 0.195	16:47	Volume Not Met
Breath Volume	0.105	16:47	
Air Blank	0.000	16:48	
Air Blank	0.000	16:49	
Subject Sample	# 2	16:49	Subject Test Refused
Breath Volume	0.000	16:49	
Air Blank	0.000	16:50	
Control Test	0.077	16:50	
Air Blank	0.000	16:51	
Diagnostics Check	OK	16:51	

Data Download Date/Time: 01/31/2018 07:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES, RONNIE,
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: FIEDLER, CHRISTOPHER
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 11/20/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 10:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:54

SUBJECT'S LAST NAME: ESTUDILLO AGUILAR

DATE/TIME OF TEST: 12/10/2017 11:22:48

Diagnostics Check	OK	11:16
Air Blank	0.000	11:16
Control Test	0.080	11:16
Air Blank	0.000	11:17
Subject Sample # 1	0.161	11:17
Breath Volume	2.003	11:17
Air Blank	0.000	11:18
Air Blank	0.000	11:20
Subject Sample # 2	0.166	11:21
Breath Volume	3.863	11:21
Air Blank	0.000	11:21
Control Test	0.080	11:22
Air Blank	0.000	11:22
Diagnostics Check	OK	11:22

Data Download Date/Time: 12/22/2017 08:40

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001345

INSTRUMENT REGISTERED TO: EUSTIS P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HACKETT,SEAN,J
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: HACKETT,SEAN
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 11/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: ROACH

DATE/TIME OF TEST: 12/20/2017 05:02:17

Diagnostics Check	OK	04:59	
Air Blank	0.000	04:59	
Control Test	0.080	05:00	
Air Blank	0.000	05:00	
Subject Sample	# 1	05:00	Subject Test Refused
Breath Volume	0.000	05:00	
Air Blank	0.000	05:01	
Control Test	0.079	05:01	
Air Blank	0.000	05:02	
Diagnostics Check	OK	05:02	

Data Download Date/Time: 12/22/2017 08:40

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 12/02/2017 04:04:02

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.118	03:59
Breath Volume		2.105	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.109	04:02
Breath Volume		1.343	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:04

Data Download Date/Time: 12/29/2017 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/12/2017 23:46:42

Diagnostics Check	OK	23:40
Air Blank	0.000	23:40
Control Test	0.080	23:41
Air Blank	0.000	23:41
Subject Sample # 1	0.085	23:42
Breath Volume	3.562	23:42
Air Blank	0.000	23:42
Air Blank	0.000	23:44
Subject Sample # 2	0.082	23:45
Breath Volume	3.914	23:45
Air Blank	0.000	23:45
Control Test	0.080	23:46
Air Blank	0.000	23:46
Diagnostics Check	OK	23:46

Data Download Date/Time: 12/29/2017 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/21/2017 02:11:14

Diagnostics Check	OK	02:04
Air Blank	0.000	02:05
Control Test	0.079	02:05
Air Blank	0.000	02:05
Subject Sample # 1	0.195	02:06
Breath Volume	3.550	02:06
Air Blank	0.000	02:07
Air Blank	0.000	02:08
Subject Sample # 2	0.190	02:09
Breath Volume	3.417	02:09
Air Blank	0.000	02:10
Control Test	0.078	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 12/29/2017 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: RADEFELD

DATE/TIME OF TEST: 12/23/2017 01:00:31

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.079	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.250	00:56
Breath Volume	3.156	00:56
Air Blank	0.000	00:56
Air Blank	0.000	00:58
Subject Sample # 2	0.244	00:59
Breath Volume	1.929	00:59
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 12/29/2017 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: KOVALSKY JR

DATE/TIME OF TEST: 12/31/2017 23:38:37

Diagnostics Check	OK	23:32
Air Blank	0.000	23:32
Control Test	0.079	23:33
Air Blank	0.000	23:33
Subject Sample # 1	0.000	23:34
Breath Volume	3.636	23:34
Air Blank	0.000	23:34
Air Blank	0.000	23:36
Subject Sample # 2	0.000	23:37
Breath Volume	3.113	23:37
Air Blank	0.000	23:37
Control Test	0.079	23:38
Air Blank	0.000	23:38
Diagnostics Check	OK	23:38

Data Download Date/Time: 01/31/2018 08:39

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MESZAROS

DATE/TIME OF TEST: 12/07/2017 02:30:34

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.078	02:17	
Air Blank		0.000	02:17	
Subject Sample	# 1	0.218	02:18	No .020 agreement
Breath Volume		2.015	02:18	
Air Blank		0.000	02:19	
Air Blank		0.000	02:21	
Subject Sample	# 2		02:24	No Sample Provided
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Air Blank		0.000	02:26	
Subject Sample	# 3	0.191	02:29	No .020 agreement
Breath Volume		1.703	02:29	
Air Blank		0.000	02:29	
Control Test		0.079	02:29	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 12/29/2017 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: MESZAROS

DATE/TIME OF TEST: 12/07/2017 02:44:12

Diagnostics Check	OK	02:37
Air Blank	0.000	02:37
Control Test	0.078	02:38
Air Blank	0.000	02:38
Subject Sample # 1	0.183	02:39
Breath Volume	1.292	02:39
Air Blank	0.000	02:40
Air Blank	0.000	02:42
Subject Sample # 2	0.202	02:42
Breath Volume	1.738	02:42
Air Blank	0.000	02:43
Control Test	0.078	02:43
Air Blank	0.000	02:44
Diagnostics Check	OK	02:44

Data Download Date/Time: 12/29/2017 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: STEPHENSON

DATE/TIME OF TEST: 12/08/2017 22:23:07

Diagnostics Check	OK	22:16
Air Blank	0.000	22:17
Control Test	0.078	22:17
Air Blank	0.000	22:17
Subject Sample # 1	0.111	22:18
Breath Volume	3.890	22:18
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.120	22:21
Breath Volume	4.261	22:21
Air Blank	0.000	22:22
Control Test	0.078	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 12/29/2017 08:58

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: DRESSENDORFER

DATE/TIME OF TEST: 12/17/2017 02:41:59

Diagnostics Check	OK	02:35
Air Blank	0.000	02:36
Control Test	0.079	02:36
Air Blank	0.000	02:37
Subject Sample # 1	0.214	02:37
Breath Volume	3.082	02:37
Air Blank	0.000	02:38
Air Blank	0.000	02:39
Subject Sample # 2	0.209	02:40
Breath Volume	3.125	02:40
Air Blank	0.000	02:41
Control Test	0.079	02:41
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 12/29/2017 08:58

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001348

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATT,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATT
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:35

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 12/30/2017 02:16:53

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.079	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.158	02:09	Volume Not Met
Breath Volume		1.031	02:09	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.191	02:12	
Breath Volume		2.031	02:12	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 3	0.204	02:15	
Breath Volume		1.425	02:15	
Air Blank		0.000	02:15	
Control Test		0.078	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 01/31/2018 07:43

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001363

INSTRUMENT REGISTERED TO: CHARLOTTE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERN,MICHAEL,P
BREATH TEST OPERATOR AGENCY: Charlotte County SO
ARRESTING OFFICER: DOE,JOHN
ARRESTING OFFICER AGENCY: Charlotte County SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:10

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 12/27/2017 22:48:20

Diagnostics Check	OK	22:42
Air Blank	0.000	22:43
Control Test	0.079	22:43
Air Blank	0.000	22:44
Subject Sample # 1	0.000	22:44
Breath Volume	1.925	22:44
Air Blank	0.000	22:44
Air Blank	0.000	22:46
Subject Sample # 2	0.000	22:46
Breath Volume	1.898	22:46
Air Blank	0.000	22:47
Control Test	0.079	22:47
Air Blank	0.000	22:48
Diagnostics Check	OK	22:48

Data Download Date/Time: 01/26/2018 07:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,DEBRA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BRICE,DAVID
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/01/2017 15:18:41

Diagnostics Check	OK	15:17	
Air Blank	0.000	15:17	
Control Test	0.078	15:18	Interferent Detect
Air Blank	0.000	15:18	

Data Download Date/Time: 12/06/2017 20:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PAGE,DEBRA,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BRICE,DAVID
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/01/2017 15:33:22

Diagnostics Check	OK	15:32	
Air Blank	0.000	15:32	
Control Test	0.079	15:32	Interferent Detect
Air Blank	0.000	15:33	

Data Download Date/Time: 12/06/2017 20:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,ROBERT,P
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WHITE,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: MINGOIA

DATE/TIME OF TEST: 12/05/2017 13:07:49

Diagnostics Check	OK	13:06	
Air Blank		13:07	Ambient Fail
Air Blank		13:07	Purge Fail

Data Download Date/Time: 12/06/2017 20:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,ROBERT,P
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WHITE,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: MINGOIA

DATE/TIME OF TEST: 12/05/2017 13:11:56

Diagnostics Check	OK	13:10	
Air Blank	0.000	13:11	
Control Test	0.079	13:11	Interferent Detect
Air Blank	0.000	13:11	

Data Download Date/Time: 12/06/2017 20:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,ROBERT,P
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: WHITE,ROBERT
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:44

SUBJECT'S LAST NAME: MINGOIA

DATE/TIME OF TEST: 12/05/2017 13:15:57

Diagnostics Check	OK	13:14	
Air Blank	0.000	13:15	
Control Test	0.078	13:15	Interferent Detect
Air Blank	0.000	13:15	

Data Download Date/Time: 12/06/2017 20:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001364

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: SYS CHECK
ARREST TIME: 17:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: DOE

DATE/TIME OF TEST: 12/06/2017 18:31:15

Diagnostics Check	OK	18:25
Air Blank	0.000	18:26
Control Test	0.080	18:26
Air Blank	0.000	18:26
Subject Sample # 1	0.000	18:27
Breath Volume	1.855	18:27
Air Blank	0.000	18:27
Air Blank	0.000	18:29
Subject Sample # 2	0.000	18:29
Breath Volume	1.867	18:29
Air Blank	0.000	18:30
Control Test	0.081	18:30
Air Blank	0.000	18:31
Diagnostics Check	OK	18:31

Data Download Date/Time: 12/06/2017 20:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINKOFER,ZACHARY,R
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: ROE,J
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:01

SUBJECT'S LAST NAME: TREHARNE

DATE/TIME OF TEST: 12/05/2017 18:37:11

Diagnostics Check	OK	18:25	
Air Blank	0.000	18:26	
Control Test	0.081	18:26	
Air Blank	0.000	18:27	
Subject Sample	# 1 0.000	18:30	Volume Not Met
Breath Volume	0.910	18:30	
Air Blank	0.000	18:30	
Air Blank	0.000	18:32	
Subject Sample	# 2 0.000	18:35	Volume Not Met
Breath Volume	1.039	18:35	
Air Blank	0.000	18:36	
Control Test	0.080	18:36	
Air Blank	0.000	18:37	
Diagnostics Check	OK	18:37	

Data Download Date/Time: 12/26/2017 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,RICK,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BROWN,RICK
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: BRAY

DATE/TIME OF TEST: 12/11/2017 04:01:42

Diagnostics Check	OK	03:54
Air Blank	0.000	03:54
Control Test	0.081	03:55
Air Blank	0.000	03:55
Subject Sample # 1	0.151	03:57
Breath Volume	3.398	03:57
Air Blank	0.000	03:57
Air Blank	0.000	03:59
Subject Sample # 2	0.150	04:00
Breath Volume	1.687	04:00
Air Blank	0.000	04:00
Control Test	0.081	04:01
Air Blank	0.000	04:01
Diagnostics Check	OK	04:01

Data Download Date/Time: 12/26/2017 15:41

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHNER, TIFFANY, A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: DAVIS, KEVIN
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 11/02/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: FULLER

DATE/TIME OF TEST: 12/18/2017 00:41:10

Diagnostics Check	OK	00:34
Air Blank	0.000	00:34
Control Test	0.081	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.176	00:36
Breath Volume	1.457	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:38
Subject Sample # 2	0.185	00:39
Breath Volume	2.406	00:39
Air Blank	0.000	00:40
Control Test	0.078	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 12/26/2017 15:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: PERKINS,MARCUS
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: EMERY

DATE/TIME OF TEST: 12/03/2017 03:25:53

Diagnostics Check	OK	03:19
Air Blank	0.000	03:19
Control Test	0.079	03:20
Air Blank	0.000	03:20
Subject Sample # 1	0.105	03:21
Breath Volume	1.128	03:21
Air Blank	0.000	03:21
Air Blank	0.000	03:23
Subject Sample # 2	0.107	03:24
Breath Volume	2.453	03:24
Air Blank	0.000	03:24
Control Test	0.079	03:25
Air Blank	0.000	03:25
Diagnostics Check	OK	03:25

Data Download Date/Time: 12/28/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: CLAUDIO

DATE/TIME OF TEST: 12/16/2017 03:08:27

Diagnostics Check		OK	03:01
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Subject Sample	# 1	0.109	03:04
Breath Volume		2.277	03:04
Air Blank		0.000	03:04
Air Blank		0.000	03:06
Subject Sample	# 2	0.100	03:07
Breath Volume		1.480	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Diagnostics Check		OK	03:08

Data Download Date/Time: 12/28/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: POLLOCK,DEREK
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 12/17/2017 03:39:19

Diagnostics Check	OK	03:32
Air Blank	0.000	03:32
Control Test	0.079	03:33
Air Blank	0.000	03:33
Subject Sample # 1	0.190	03:34
Breath Volume	1.453	03:34
Air Blank	0.000	03:34
Air Blank	0.000	03:36
Subject Sample # 2	0.189	03:37
Breath Volume	1.605	03:37
Air Blank	0.000	03:38
Control Test	0.080	03:38
Air Blank	0.000	03:39
Diagnostics Check	OK	03:39

Data Download Date/Time: 12/28/2017 12:34

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAHNSEN, A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/06/2017 22:43:04

Diagnostics Check	OK	22:33
Air Blank	0.000	22:33
Control Test	0.081	22:34
Air Blank	0.000	22:34
Subject Sample # 1	0.076	22:37
Breath Volume	1.554	22:37
Air Blank	0.000	22:37
Air Blank	0.000	22:39
Subject Sample # 2	0.071	22:41
Breath Volume	1.386	22:41
Air Blank	0.000	22:42
Control Test	0.080	22:42
Air Blank	0.000	22:42
Diagnostics Check	OK	22:43

Data Download Date/Time: 12/23/2017 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CORDERIO,E
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 12/13/2017 00:10:23

Diagnostics Check	OK	00:04
Air Blank	0.000	00:04
Control Test	0.081	00:05
Air Blank	0.000	00:05
Subject Sample # 1	0.183	00:06
Breath Volume	1.925	00:06
Air Blank	0.000	00:06
Air Blank	0.000	00:08
Subject Sample # 2	0.168	00:08
Breath Volume	1.359	00:08
Air Blank	0.000	00:09
Control Test	0.081	00:09
Air Blank	0.000	00:10
Diagnostics Check	OK	00:10

Data Download Date/Time: 12/23/2017 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WARREN,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: RICHMAN

DATE/TIME OF TEST: 12/13/2017 03:30:27

Diagnostics Check	OK	03:23
Air Blank	0.000	03:24
Control Test	0.081	03:24
Air Blank	0.000	03:25
Subject Sample # 1	0.123	03:25
Breath Volume	1.265	03:25
Air Blank	0.000	03:26
Air Blank	0.000	03:28
Subject Sample # 2	0.119	03:29
Breath Volume	1.726	03:29
Air Blank	0.000	03:29
Control Test	0.081	03:29
Air Blank	0.000	03:30
Diagnostics Check	OK	03:30

Data Download Date/Time: 12/23/2017 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FREHULFER, S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: CUNLIFFE

DATE/TIME OF TEST: 12/16/2017 23:35:42

Diagnostics Check	OK	23:26
Air Blank	0.000	23:26
Control Test	0.081	23:27
Air Blank	0.000	23:27
Subject Sample # 1	0.331	23:28
Breath Volume	1.957	23:28
Air Blank	0.000	23:28
Air Blank	0.000	23:30
Subject Sample # 2	0.293	23:31
Breath Volume	2.164	23:31
Air Blank	0.000	23:31
Air Blank	0.000	23:33
Subject Sample # 3	0.295	23:34
Breath Volume	2.902	23:34
Air Blank	0.000	23:34
Control Test	0.080	23:35
Air Blank	0.000	23:35
Diagnostics Check	OK	23:35

Data Download Date/Time: 12/23/2017 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FORD, A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: MCNUTT

DATE/TIME OF TEST: 12/18/2017 01:03:22

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.081	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1	0.186	00:52	
Breath Volume		1.292	00:52	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 2	0.126	00:57	Volume Not Met
Breath Volume		0.085	00:57	
Air Blank		0.000	00:58	
Air Blank		0.000	01:00	
Subject Sample	# 3	0.185	01:01	
Breath Volume		1.406	01:01	
Air Blank		0.000	01:02	
Control Test		0.079	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 12/23/2017 16:15

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BURSE, E
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:12

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 12/18/2017 04:38:24

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.080	04:32
Air Blank		0.000	04:33
Subject Sample	# 1	0.104	04:34
Breath Volume		1.296	04:34
Air Blank		0.000	04:34
Air Blank		0.000	04:36
Subject Sample	# 2	0.104	04:36
Breath Volume		1.382	04:36
Air Blank		0.000	04:37
Control Test		0.080	04:37
Air Blank		0.000	04:38
Diagnostics Check		OK	04:38

Data Download Date/Time: 12/23/2017 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: MCREE, M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: ELWER

DATE/TIME OF TEST: 12/22/2017 01:56:16

Diagnostics Check	OK	01:50
Air Blank	0.000	01:50
Control Test	0.081	01:50
Air Blank	0.000	01:51
Subject Sample # 1	0.136	01:51
Breath Volume	1.730	01:51
Air Blank	0.000	01:52
Air Blank	0.000	01:54
Subject Sample # 2	0.136	01:54
Breath Volume	1.718	01:54
Air Blank	0.000	01:55
Control Test	0.080	01:55
Air Blank	0.000	01:56
Diagnostics Check	OK	01:56

Data Download Date/Time: 12/23/2017 16:15

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL, DAKOTA, W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FREHULFER, S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/26/2017 23:21:04

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.000	23:16
Breath Volume	1.222	23:16
Air Blank	0.000	23:16
Air Blank	0.000	23:18
Subject Sample # 2	0.000	23:19
Breath Volume	1.507	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:20
Diagnostics Check	OK	23:21

Data Download Date/Time: 01/14/2018 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 12/28/2017 00:28:34

Diagnostics Check	OK	00:26	
Air Blank	0.000	00:27	
Control Test	0.081	00:27	
Air Blank		00:28	RFI Detect
Air Blank	0.000	00:28	

Data Download Date/Time: 01/14/2018 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: COHEN

DATE/TIME OF TEST: 12/30/2017 23:44:25

Diagnostics Check	OK	23:33	
Air Blank	0.000	23:33	
Control Test	0.081	23:33	
Air Blank	0.000	23:34	
Subject Sample	# 1	23:37	No Sample Provided
Breath Volume	0.000	23:37	
Air Blank	0.000	23:37	
Air Blank	0.000	23:39	
Subject Sample	# 2	23:43	No Sample Provided
Breath Volume	0.000	23:43	
Air Blank	0.000	23:43	
Control Test	0.082	23:43	
Air Blank	0.000	23:44	
Diagnostics Check	OK	23:44	

Data Download Date/Time: 01/14/2018 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FREHULFER,S
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:54

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 12/31/2017 04:21:16

Diagnostics Check	OK	04:14
Air Blank	0.000	04:14
Control Test	0.081	04:15
Air Blank	0.000	04:15
Subject Sample # 1	0.089	04:16
Breath Volume	1.167	04:16
Air Blank	0.000	04:17
Air Blank	0.000	04:18
Subject Sample # 2	0.085	04:19
Breath Volume	1.617	04:19
Air Blank	0.000	04:20
Control Test	0.081	04:20
Air Blank	0.000	04:21
Diagnostics Check	OK	04:21

Data Download Date/Time: 01/14/2018 15:19

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:21

SUBJECT'S LAST NAME: GUTIERREZ

DATE/TIME OF TEST: 12/31/2017 22:56:10

Diagnostics Check	OK	22:49
Air Blank	0.000	22:50
Control Test	0.081	22:50
Air Blank	0.000	22:51
Subject Sample # 1	0.228	22:51
Breath Volume	1.382	22:51
Air Blank	0.000	22:51
Air Blank	0.000	22:53
Subject Sample # 2	0.220	22:54
Breath Volume	1.679	22:54
Air Blank	0.000	22:55
Control Test	0.081	22:55
Air Blank	0.000	22:56
Diagnostics Check	OK	22:56

Data Download Date/Time: 01/14/2018 15:19

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRABER,KIMBERLY,L
BREATH TEST OPERATOR AGENCY: Putnam County SO
ARRESTING OFFICER: TRABER,KIMBERLY
ARRESTING OFFICER AGENCY: Putnam County SO
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 06/01/2018
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 12/04/2017 22:41:47

Diagnostics Check	OK	22:30	
Air Blank	0.000	22:30	
Control Test	0.078	22:31	
Air Blank	0.000	22:31	
Subject Sample	# 1	22:34	No Sample Provided
Breath Volume	0.000	22:34	
Air Blank	0.000	22:35	
Air Blank	0.000	22:37	
Subject Sample	# 2	22:40	No Sample Provided
Breath Volume	0.000	22:40	
Air Blank	0.000	22:40	
Control Test	0.079	22:41	
Air Blank	0.000	22:41	
Diagnostics Check	OK	22:41	

Data Download Date/Time: 01/01/2018 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,W,
BREATH TEST OPERATOR AGENCY: Crescent City PD
ARRESTING OFFICER: MELTON,W
ARRESTING OFFICER AGENCY: Crescent City PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MORALES RAMIREZ

DATE/TIME OF TEST: 12/07/2017 04:52:57

Diagnostics Check	OK	04:46
Air Blank	0.000	04:46
Control Test	0.078	04:46
Air Blank	0.000	04:47
Subject Sample # 1	0.104	04:48
Breath Volume	2.332	04:48
Air Blank	0.000	04:48
Air Blank	0.000	04:50
Subject Sample # 2	0.101	04:51
Breath Volume	2.273	04:51
Air Blank	0.000	04:52
Control Test	0.078	04:52
Air Blank	0.000	04:52
Diagnostics Check	OK	04:52

Data Download Date/Time: 01/01/2018 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,W,
BREATH TEST OPERATOR AGENCY: Crescent City PD
ARRESTING OFFICER: MELTON,W
ARRESTING OFFICER AGENCY: Crescent City PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: GILL

DATE/TIME OF TEST: 12/16/2017 02:00:41

Diagnostics Check	OK	01:57	
Air Blank	0.000	01:58	
Control Test	0.079	01:58	
Air Blank	0.000	01:59	
Subject Sample	# 1	01:59	Subject Test Refused
Breath Volume	0.000	01:59	
Air Blank	0.000	01:59	
Control Test	0.079	02:00	
Air Blank	0.000	02:00	
Diagnostics Check	OK	02:00	

Data Download Date/Time: 01/01/2018 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001648

INSTRUMENT REGISTERED TO: PUTNAM COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTON,W,
BREATH TEST OPERATOR AGENCY: Crescent City PD
ARRESTING OFFICER: MELTON,W
ARRESTING OFFICER AGENCY: Crescent City PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: OXENDINE

DATE/TIME OF TEST: 12/16/2017 23:40:37

Diagnostics Check	OK	23:37	
Air Blank	0.000	23:37	
Control Test	0.079	23:37	
Air Blank	0.000	23:38	
Subject Sample	# 1	23:39	Subject Test Refused
Breath Volume	0.000	23:39	
Air Blank	0.000	23:39	
Control Test	0.079	23:40	
Air Blank	0.000	23:40	
Diagnostics Check	OK	23:40	

Data Download Date/Time: 01/01/2018 21:41

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/17/2017 06:03:17

Diagnostics Check	OK	05:57
Air Blank	0.000	05:57
Control Test	0.080	05:57
Air Blank	0.000	05:58
Subject Sample # 1	0.000	05:58
Breath Volume	1.609	05:58
Air Blank	0.000	05:59
Air Blank	0.000	06:01
Subject Sample # 2	0.000	06:01
Breath Volume	2.121	06:01
Air Blank	0.000	06:02
Control Test	0.080	06:02
Air Blank	0.000	06:03
Diagnostics Check	OK	06:03

Data Download Date/Time: 12/29/2017 14:43

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: O'BANNON

DATE/TIME OF TEST: 12/31/2017 02:39:48

Diagnostics Check	OK	02:32	
Air Blank	0.000	02:32	
Control Test	0.080	02:33	
Air Blank	0.000	02:33	
Subject Sample	# 1 0.164	02:34	
Breath Volume	1.921	02:34	
Air Blank	0.000	02:35	
Air Blank	0.000	02:37	
Subject Sample	# 2 0.174	02:37	
Breath Volume	1.210	02:37	
Air Blank	0.000	02:38	
Control Test	0.079	02:38	
Air Blank		02:39	RFI Detect
Air Blank	0.000	02:39	

Data Download Date/Time: 01/30/2018 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001649

INSTRUMENT REGISTERED TO: KENNETH CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUNER,SEAN,R
BREATH TEST OPERATOR AGENCY: Kenneth City PD
ARRESTING OFFICER: BRUNER,SEAN
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A7
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: O BANNON

DATE/TIME OF TEST: 12/31/2017 02:50:34

Diagnostics Check	OK	02:43
Air Blank	0.000	02:43
Control Test	0.080	02:44
Air Blank	0.000	02:44
Subject Sample # 1	0.176	02:45
Breath Volume	1.609	02:45
Air Blank	0.000	02:46
Air Blank	0.000	02:47
Subject Sample # 2	0.160	02:48
Breath Volume	1.726	02:48
Air Blank	0.000	02:49
Control Test	0.078	02:49
Air Blank	0.000	02:50
Diagnostics Check	OK	02:50

Data Download Date/Time: 01/30/2018 17:02

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:07

SUBJECT'S LAST NAME: HASSAN

DATE/TIME OF TEST: 12/01/2017 19:36:09

Diagnostics Check	OK	19:29
Air Blank	0.000	19:29
Control Test	0.079	19:30
Air Blank	0.000	19:30
Subject Sample # 1	0.182	19:31
Breath Volume	2.062	19:31
Air Blank	0.000	19:32
Air Blank	0.000	19:33
Subject Sample # 2	0.184	19:34
Breath Volume	2.082	19:34
Air Blank	0.000	19:35
Control Test	0.079	19:35
Air Blank	0.000	19:36
Diagnostics Check	OK	19:36

Data Download Date/Time: 12/06/2017 10:16

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: WISE JR

DATE/TIME OF TEST: 12/07/2017 00:58:22

Diagnostics Check	OK	00:55	
Air Blank	0.000	00:55	
Control Test	0.079	00:55	
Air Blank	0.000	00:56	
Subject Sample	# 1	00:57	Subject Test Refused
Breath Volume	0.000	00:57	
Air Blank	0.000	00:57	
Control Test	0.079	00:57	
Air Blank	0.000	00:58	
Diagnostics Check	OK	00:58	

Data Download Date/Time: 01/17/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BEDARD JR

DATE/TIME OF TEST: 12/08/2017 20:08:28

Diagnostics Check	OK	20:02
Air Blank	0.000	20:02
Control Test	0.078	20:02
Air Blank	0.000	20:03
Subject Sample # 1	0.048	20:04
Breath Volume	1.539	20:04
Air Blank	0.000	20:04
Air Blank	0.000	20:06
Subject Sample # 2	0.051	20:07
Breath Volume	3.066	20:07
Air Blank	0.000	20:07
Control Test	0.078	20:07
Air Blank	0.000	20:08
Diagnostics Check	OK	20:08

Data Download Date/Time: 01/17/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HORNEMAN

DATE/TIME OF TEST: 12/09/2017 03:09:54

Diagnostics Check	OK	03:05	
Air Blank	0.000	03:06	
Control Test	0.078	03:06	
Air Blank	0.000	03:07	
Subject Sample	# 1	03:08	Subject Test Refused
Breath Volume	0.000	03:08	
Air Blank	0.000	03:09	
Control Test	0.078	03:09	
Air Blank	0.000	03:09	
Diagnostics Check	OK	03:09	

Data Download Date/Time: 01/17/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA, VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DANIEL

DATE/TIME OF TEST: 12/10/2017 00:08:03

Diagnostics Check	OK	00:01
Air Blank	0.000	00:01
Control Test	0.080	00:02
Air Blank	0.000	00:02
Subject Sample # 1	0.075	00:03
Breath Volume	1.835	00:03
Air Blank	0.000	00:04
Air Blank	0.000	00:05
Subject Sample # 2	0.084	00:06
Breath Volume	2.433	00:06
Air Blank	0.000	00:07
Control Test	0.080	00:07
Air Blank	0.000	00:07
Diagnostics Check	OK	00:08

Data Download Date/Time: 01/17/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI, ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO, AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: CAPOBIANCO

DATE/TIME OF TEST: 12/10/2017 02:11:10

Diagnostics Check	OK	02:02
Air Blank	0.000	02:02
Control Test	0.080	02:02
Air Blank	0.000	02:03
Subject Sample # 1	0.184	02:05
Breath Volume	2.378	02:05
Air Blank	0.000	02:06
Air Blank	0.000	02:08
Subject Sample # 2	0.178	02:09
Breath Volume	2.281	02:09
Air Blank	0.000	02:10
Control Test	0.079	02:10
Air Blank	0.000	02:11
Diagnostics Check	OK	02:11

Data Download Date/Time: 01/17/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HODGE

DATE/TIME OF TEST: 12/10/2017 04:03:29

Diagnostics Check	OK	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:57
Air Blank	0.000	03:58
Subject Sample # 1	0.116	03:59
Breath Volume	2.539	03:59
Air Blank	0.000	03:59
Air Blank	0.000	04:01
Subject Sample # 2	0.112	04:02
Breath Volume	2.152	04:02
Air Blank	0.000	04:02
Control Test	0.079	04:02
Air Blank	0.000	04:03
Diagnostics Check	OK	04:03

Data Download Date/Time: 01/17/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,MATTHEW,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CARTER,MATTHEW
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: GOODWYN

DATE/TIME OF TEST: 12/10/2017 05:12:43

Diagnostics Check	OK	05:08	
Air Blank	0.000	05:09	
Control Test	0.079	05:09	
Air Blank	0.000	05:10	
Subject Sample	# 1	05:11	Subject Test Refused
Breath Volume	0.000	05:11	
Air Blank	0.000	05:11	
Control Test	0.079	05:12	
Air Blank	0.000	05:12	
Diagnostics Check	OK	05:12	

Data Download Date/Time: 01/17/2018 08:48

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARK,BRENT,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CLARK,BRENT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:56

SUBJECT'S LAST NAME: STAMPER

DATE/TIME OF TEST: 12/10/2017 18:23:53

Diagnostics Check	OK	18:17
Air Blank	0.000	18:17
Control Test	0.079	18:17
Air Blank	0.000	18:18
Subject Sample # 1	0.232	18:19
Breath Volume	1.578	18:19
Air Blank	0.000	18:19
Air Blank	0.000	18:21
Subject Sample # 2	0.230	18:22
Breath Volume	1.953	18:22
Air Blank	0.000	18:22
Control Test	0.079	18:23
Air Blank	0.000	18:23
Diagnostics Check	OK	18:23

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKINNER,DAVID,D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,Z
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/11/2017 19:57:01

Diagnostics Check	OK	19:51
Air Blank	0.000	19:51
Control Test	0.080	19:51
Air Blank	0.000	19:52
Subject Sample # 1	0.196	19:52
Breath Volume	1.878	19:52
Air Blank	0.000	19:53
Air Blank	0.000	19:55
Subject Sample # 2	0.186	19:55
Breath Volume	2.253	19:55
Air Blank	0.000	19:56
Control Test	0.081	19:56
Air Blank	0.000	19:56
Diagnostics Check	OK	19:57

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEISKOPF, MICHAEL, D
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BUCHANAN, CHRISTOPHER
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:07

SUBJECT'S LAST NAME: BUNAISKY

DATE/TIME OF TEST: 12/13/2017 10:33:53

Diagnostics Check	OK	10:28
Air Blank	0.000	10:28
Control Test	0.080	10:28
Air Blank	0.000	10:29
Subject Sample # 1	0.000	10:29
Breath Volume	1.750	10:29
Air Blank	0.000	10:30
Air Blank	0.000	10:32
Subject Sample # 2	0.000	10:32
Breath Volume	1.117	10:32
Air Blank	0.000	10:33
Control Test	0.081	10:33
Air Blank	0.000	10:33
Diagnostics Check	OK	10:33

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: ADAME

DATE/TIME OF TEST: 12/13/2017 23:59:57

Diagnostics Check	OK	23:49
Air Blank	0.000	23:49
Control Test	0.080	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.130	23:51
Breath Volume	1.492	23:51
Air Blank	0.000	23:52
Air Blank	0.000	23:54
Subject Sample # 2	0.154	23:55
Breath Volume	2.164	23:55
Air Blank	0.000	23:55
Air Blank	0.000	23:57
Subject Sample # 3	0.132	23:58
Breath Volume	1.148	23:58
Air Blank	0.000	23:59
Control Test	0.078	23:59
Air Blank	0.000	23:59
Diagnostics Check	OK	23:59

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEFFEK, MAURICIO,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BUCHANAN, CHRIS
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:48

SUBJECT'S LAST NAME: ALBERT

DATE/TIME OF TEST: 12/14/2017 14:18:20

Diagnostics Check	OK	14:11
Air Blank	0.000	14:12
Control Test	0.079	14:12
Air Blank	0.000	14:13
Subject Sample # 1	0.316	14:13
Breath Volume	3.179	14:13
Air Blank	0.000	14:14
Air Blank	0.000	14:16
Subject Sample # 2	0.308	14:16
Breath Volume	3.195	14:16
Air Blank	0.000	14:17
Control Test	0.078	14:17
Air Blank	0.000	14:18
Diagnostics Check	OK	14:18

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SPETRINI

DATE/TIME OF TEST: 12/15/2017 00:48:06

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.182	00:43
Breath Volume		1.363	00:43
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.185	00:46
Breath Volume		1.394	00:46
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: STRUNK

DATE/TIME OF TEST: 12/15/2017 02:41:15

Diagnostics Check	OK	02:35
Air Blank	0.000	02:35
Control Test	0.079	02:36
Air Blank	0.000	02:36
Subject Sample # 1	0.000	02:37
Breath Volume	2.402	02:37
Air Blank	0.000	02:37
Air Blank	0.000	02:39
Subject Sample # 2	0.000	02:39
Breath Volume	2.433	02:39
Air Blank	0.000	02:40
Control Test	0.079	02:40
Air Blank	0.000	02:41
Diagnostics Check	OK	02:41

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:57

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 12/16/2017 01:23:23

Diagnostics Check	OK	01:17
Air Blank	0.000	01:17
Control Test	0.079	01:17
Air Blank	0.000	01:18
Subject Sample # 1	0.123	01:18
Breath Volume	3.890	01:18
Air Blank	0.000	01:19
Air Blank	0.000	01:21
Subject Sample # 2	0.121	01:21
Breath Volume	3.632	01:21
Air Blank	0.000	01:22
Control Test	0.079	01:22
Air Blank	0.000	01:23
Diagnostics Check	OK	01:23

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KRONZ,BRIAN,M
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: BIBLER,RANDALL
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: BUI
ARREST TIME: 22:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: LEAVITT

DATE/TIME OF TEST: 12/16/2017 23:56:59

Diagnostics Check	OK	23:51
Air Blank	0.000	23:51
Control Test	0.080	23:51
Air Blank	0.000	23:52
Subject Sample # 1	0.115	23:52
Breath Volume	2.960	23:52
Air Blank	0.000	23:53
Air Blank	0.000	23:55
Subject Sample # 2	0.112	23:55
Breath Volume	2.906	23:55
Air Blank	0.000	23:56
Control Test	0.079	23:56
Air Blank	0.000	23:56
Diagnostics Check	OK	23:56

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: THOMAS,KRISTIN
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: WYSMAN

DATE/TIME OF TEST: 12/18/2017 23:55:45

Diagnostics Check	OK	23:49
Air Blank	0.000	23:50
Control Test	0.079	23:50
Air Blank	0.000	23:50
Subject Sample # 1	0.136	23:51
Breath Volume	2.445	23:51
Air Blank	0.000	23:51
Air Blank	0.000	23:53
Subject Sample # 2	0.144	23:54
Breath Volume	1.843	23:54
Air Blank	0.000	23:54
Control Test	0.080	23:55
Air Blank	0.000	23:55
Diagnostics Check	OK	23:55

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CISSELL,ZACHARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: GEER

DATE/TIME OF TEST: 12/19/2017 01:46:33

Diagnostics Check	OK	01:43	
Air Blank	0.000	01:44	
Control Test	0.079	01:44	
Air Blank	0.000	01:45	
Subject Sample	# 1	01:45	Subject Test Refused
Breath Volume	0.000	01:45	
Air Blank	0.000	01:45	
Control Test	0.079	01:45	
Air Blank	0.000	01:46	
Diagnostics Check	OK	01:46	

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 12/22/2017 00:35:34

Diagnostics Check	OK	00:29
Air Blank	0.000	00:29
Control Test	0.079	00:29
Air Blank	0.000	00:30
Subject Sample # 1	0.190	00:30
Breath Volume	3.437	00:30
Air Blank	0.000	00:31
Air Blank	0.000	00:33
Subject Sample # 2	0.179	00:34
Breath Volume	2.472	00:34
Air Blank	0.000	00:34
Control Test	0.078	00:34
Air Blank	0.000	00:35
Diagnostics Check	OK	00:35

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PIERCE JR,SCOTT,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: PIERCE JR,SCOTT
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:18

SUBJECT'S LAST NAME: TIMMONS

DATE/TIME OF TEST: 12/22/2017 10:44:44

Diagnostics Check	OK	10:38
Air Blank	0.000	10:38
Control Test	0.078	10:39
Air Blank	0.000	10:39
Subject Sample # 1	0.304	10:40
Breath Volume	2.398	10:40
Air Blank	0.000	10:40
Air Blank	0.000	10:42
Subject Sample # 2	0.290	10:43
Breath Volume	2.718	10:43
Air Blank	0.000	10:43
Control Test	0.079	10:44
Air Blank	0.000	10:44
Diagnostics Check	OK	10:44

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MATTSON

DATE/TIME OF TEST: 12/24/2017 00:01:04

Diagnostics Check	OK	23:58	
Air Blank	0.000	23:58	
Control Test	0.079	23:58	
Air Blank	0.000	23:59	
Subject Sample	# 1	23:59	Subject Test Refused
Breath Volume	0.000	23:59	
Air Blank	0.000	00:00	
Control Test	0.080	00:00	
Air Blank	0.000	00:00	
Diagnostics Check	OK	00:01	

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/24/2017 03:19:42

Diagnostics Check	OK	03:16	
Air Blank	0.000	03:17	
Control Test	0.079	03:17	
Air Blank	0.000	03:18	
Subject Sample	# 1	03:18	Subject Test Refused
Breath Volume	0.000	03:18	
Air Blank	0.000	03:18	
Control Test	0.078	03:19	
Air Blank	0.000	03:19	
Diagnostics Check	OK	03:19	

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: IOAKIMEDES

DATE/TIME OF TEST: 12/29/2017 22:21:41

Diagnostics Check	OK	22:14
Air Blank	0.000	22:15
Control Test	0.079	22:15
Air Blank	0.000	22:16
Subject Sample # 1	0.147	22:16
Breath Volume	1.781	22:16
Air Blank	0.000	22:17
Air Blank	0.000	22:19
Subject Sample # 2	0.146	22:20
Breath Volume	1.800	22:20
Air Blank	0.000	22:20
Control Test	0.079	22:21
Air Blank	0.000	22:21
Diagnostics Check	OK	22:21

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WAITERS

DATE/TIME OF TEST: 12/31/2017 03:01:48

Diagnostics Check	OK	02:55
Air Blank	0.000	02:55
Control Test	0.080	02:56
Air Blank	0.000	02:56
Subject Sample # 1	0.149	02:57
Breath Volume	1.328	02:57
Air Blank	0.000	02:57
Air Blank	0.000	02:59
Subject Sample # 2	0.139	03:00
Breath Volume	1.289	03:00
Air Blank	0.000	03:00
Control Test	0.078	03:01
Air Blank	0.000	03:01
Diagnostics Check	OK	03:01

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISTI,JORDAN,A
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: MORGAN,QUENTIN-LEE
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 12/31/2017 06:21:39

Diagnostics Check		OK	06:15
Air Blank		0.000	06:16
Control Test		0.079	06:16
Air Blank		0.000	06:16
Subject Sample	# 1	0.224	06:17
Breath Volume		2.656	06:17
Air Blank		0.000	06:17
Air Blank		0.000	06:19
Subject Sample	# 2	0.230	06:20
Breath Volume		1.851	06:20
Air Blank		0.000	06:20
Control Test		0.079	06:21
Air Blank		0.000	06:21
Diagnostics Check		OK	06:21

Data Download Date/Time: 01/17/2018 08:49

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,CHASE,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WALLACE,NATHAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:44

SUBJECT'S LAST NAME: BANKO

DATE/TIME OF TEST: 12/01/2017 18:11:35

Diagnostics Check	OK	18:05
Air Blank	0.000	18:05
Control Test	0.079	18:05
Air Blank	0.000	18:06
Subject Sample # 1	0.309	18:07
Breath Volume	1.667	18:07
Air Blank	0.000	18:07
Air Blank	0.000	18:09
Subject Sample # 2	0.314	18:10
Breath Volume	2.003	18:10
Air Blank	0.000	18:10
Control Test	0.079	18:11
Air Blank	0.000	18:11
Diagnostics Check	OK	18:11

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,ERIN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SMALL,RYAN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/07/2017 00:51:48

Diagnostics Check	OK	00:45
Air Blank	0.000	00:46
Control Test	0.081	00:46
Air Blank	0.000	00:47
Subject Sample # 1	0.221	00:47
Breath Volume	1.859	00:47
Air Blank	0.000	00:48
Air Blank	0.000	00:49
Subject Sample # 2	0.212	00:50
Breath Volume	2.066	00:50
Air Blank	0.000	00:50
Control Test	0.080	00:51
Air Blank	0.000	00:51
Diagnostics Check	OK	00:51

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ACKERMAN,JUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: HOLSOMBACK

DATE/TIME OF TEST: 12/08/2017 23:21:21

Diagnostics Check	OK	23:14
Air Blank	0.000	23:14
Control Test	0.081	23:15
Air Blank	0.000	23:15
Subject Sample # 1	0.162	23:16
Breath Volume	1.503	23:16
Air Blank	0.000	23:17
Air Blank	0.000	23:18
Subject Sample # 2	0.175	23:19
Breath Volume	2.281	23:19
Air Blank	0.000	23:20
Control Test	0.080	23:20
Air Blank	0.000	23:21
Diagnostics Check	OK	23:21

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: HALL,SAMUEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SPITTEL

DATE/TIME OF TEST: 12/09/2017 00:41:23

Diagnostics Check	OK	00:33
Air Blank	0.000	00:34
Control Test	0.080	00:34
Air Blank	0.000	00:35
Subject Sample # 1	0.165	00:36
Breath Volume	1.523	00:36
Air Blank	0.000	00:37
Air Blank	0.000	00:39
Subject Sample # 2	0.167	00:39
Breath Volume	1.382	00:39
Air Blank	0.000	00:40
Control Test	0.079	00:40
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPROUSE,JOHN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SPROUSE,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: DROPPS

DATE/TIME OF TEST: 12/09/2017 01:00:10

Diagnostics Check	OK	00:53
Air Blank	0.000	00:54
Control Test	0.080	00:54
Air Blank	0.000	00:55
Subject Sample # 1	0.274	00:55
Breath Volume	1.984	00:55
Air Blank	0.000	00:56
Air Blank	0.000	00:57
Subject Sample # 2	0.268	00:58
Breath Volume	1.664	00:58
Air Blank	0.000	00:59
Control Test	0.079	00:59
Air Blank	0.000	01:00
Diagnostics Check	OK	01:00

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: ANDERSON,JUSTIN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 12/13/2017 22:23:26

Diagnostics Check	OK	22:17
Air Blank	0.000	22:17
Control Test	0.081	22:17
Air Blank	0.000	22:18
Subject Sample # 1	0.190	22:19
Breath Volume	2.089	22:19
Air Blank	0.000	22:19
Air Blank	0.000	22:21
Subject Sample # 2	0.188	22:21
Breath Volume	2.402	22:21
Air Blank	0.000	22:22
Control Test	0.080	22:22
Air Blank	0.000	22:23
Diagnostics Check	OK	22:23

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBENEDETTO,JUSTIN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WILSON,MARK
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 12:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DIBENEDETTO

DATE/TIME OF TEST: 12/17/2017 01:37:10

Diagnostics Check	OK	01:33	
Air Blank	0.000	01:33	
Control Test	0.083	01:33	
Air Blank	0.000	01:34	
Subject Sample	# 1	01:36	RFI Detect
Breath Volume	0.246	01:36	
Air Blank		01:36	RFI Detect
Air Blank	0.000	01:37	

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIBENEDETO,JUSTIN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WILSON,MARK
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 12:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: DIBENEDETTO

DATE/TIME OF TEST: 12/17/2017 01:51:04

Diagnostics Check		OK	01:39	
Air Blank		0.000	01:40	
Control Test		0.080	01:40	
Air Blank		0.000	01:40	
Subject Sample	# 1	0.120	01:44	Volume Not Met
Breath Volume		0.101	01:44	
Air Blank		0.000	01:44	
Air Blank		0.000	01:46	
Subject Sample	# 2		01:49	No Sample Provided
Breath Volume		0.000	01:49	
Air Blank		0.000	01:50	
Control Test		0.079	01:50	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:51	

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPROUSE,JOHN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:04

SUBJECT'S LAST NAME: KILPATRICK

DATE/TIME OF TEST: 12/17/2017 20:42:43

Diagnostics Check	OK	20:38	
Air Blank	0.000	20:38	
Control Test	0.081	20:38	
Air Blank	0.000	20:39	
Subject Sample	# 1	20:41	Subject Test Refused
Breath Volume	0.000	20:41	
Air Blank	0.000	20:41	
Control Test	0.081	20:42	
Air Blank	0.000	20:42	
Diagnostics Check	OK	20:42	

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPROUSE,JOHN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SPROUSE,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 12:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:36

SUBJECT'S LAST NAME: O NEAL

DATE/TIME OF TEST: 12/20/2017 01:02:56

Diagnostics Check	OK	00:56
Air Blank	0.000	00:56
Control Test	0.081	00:57
Air Blank	0.000	00:57
Subject Sample # 1	0.240	00:58
Breath Volume	3.664	00:58
Air Blank	0.000	00:58
Air Blank	0.000	01:00
Subject Sample # 2	0.229	01:01
Breath Volume	3.570	01:01
Air Blank	0.000	01:02
Control Test	0.081	01:02
Air Blank	0.000	01:02
Diagnostics Check	OK	01:02

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KING,CHASE,E
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: KING,CHASE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 18:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: MEREDITH

DATE/TIME OF TEST: 12/20/2017 19:33:30

Diagnostics Check	OK	19:26
Air Blank	0.000	19:26
Control Test	0.081	19:26
Air Blank	0.000	19:27
Subject Sample # 1	0.138	19:28
Breath Volume	1.468	19:28
Air Blank	0.000	19:28
Air Blank	0.000	19:30
Subject Sample # 2	0.134	19:32
Breath Volume	1.488	19:32
Air Blank	0.000	19:32
Control Test	0.081	19:32
Air Blank	0.000	19:33
Diagnostics Check	OK	19:33

Data Download Date/Time: 12/21/2017 16:10

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WILSON,JOSH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: COFFIN

DATE/TIME OF TEST: 12/24/2017 22:00:42

Diagnostics Check	OK	21:54
Air Blank	0.000	21:54
Control Test	0.081	21:55
Air Blank	0.000	21:55
Subject Sample # 1	0.298	21:56
Breath Volume	1.109	21:56
Air Blank	0.000	21:56
Air Blank	0.000	21:58
Subject Sample # 2	0.297	21:59
Breath Volume	2.082	21:59
Air Blank	0.000	21:59
Control Test	0.081	22:00
Air Blank	0.000	22:00
Diagnostics Check	OK	22:00

Data Download Date/Time: 02/01/2018 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 12/28/2017 01:27:03

Diagnostics Check	OK	01:25	
Air Blank	0.000	01:25	
Control Test	0.082	01:26	
Air Blank		01:26	RFI Detect
Air Blank	0.000	01:27	

Data Download Date/Time: 02/01/2018 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ACKERMAN,JUSTIN,T
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: BRIGGS,JASON
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: DENNIS

DATE/TIME OF TEST: 12/28/2017 01:31:01

Diagnostics Check	OK	01:28	
Air Blank	0.000	01:28	
Control Test	0.082	01:28	
Air Blank	0.000	01:29	
Subject Sample	# 1	01:29	Subject Test Refused
Breath Volume	0.000	01:29	
Air Blank	0.000	01:30	
Control Test	0.082	01:30	
Air Blank	0.000	01:30	
Diagnostics Check	OK	01:31	

Data Download Date/Time: 02/01/2018 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRICKEY,MICHAEL,D
BREATH TEST OPERATOR AGENCY: St. Augustine PD
ARRESTING OFFICER: BRICKEY,MICHAEL
ARRESTING OFFICER AGENCY: St. Augustine PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 21:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:09

SUBJECT'S LAST NAME: BUKSAR

DATE/TIME OF TEST: 12/29/2017 22:35:48

Diagnostics Check	OK	22:29
Air Blank	0.000	22:29
Control Test	0.082	22:30
Air Blank	0.000	22:30
Subject Sample # 1	0.290	22:31
Breath Volume	1.988	22:31
Air Blank	0.000	22:32
Air Blank	0.000	22:33
Subject Sample # 2	0.302	22:34
Breath Volume	1.964	22:34
Air Blank	0.000	22:34
Control Test	0.082	22:35
Air Blank	0.000	22:35
Diagnostics Check	OK	22:35

Data Download Date/Time: 02/01/2018 04:50

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,ERIN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: NELSON,MICHAEL
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/30/2017 03:07:21

Diagnostics Check	OK	03:00
Air Blank	0.000	03:01
Control Test	0.082	03:01
Air Blank	0.000	03:02
Subject Sample # 1	0.191	03:02
Breath Volume	2.332	03:02
Air Blank	0.000	03:03
Air Blank	0.000	03:05
Subject Sample # 2	0.183	03:05
Breath Volume	2.753	03:05
Air Blank	0.000	03:06
Control Test	0.080	03:06
Air Blank	0.000	03:07
Diagnostics Check	OK	03:07

Data Download Date/Time: 02/01/2018 04:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGINNIS,JOSEPH,D
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MCGINNIS,JOSEPH
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: GOOGE

DATE/TIME OF TEST: 12/30/2017 03:58:37

Diagnostics Check	OK	03:52
Air Blank	0.000	03:52
Control Test	0.082	03:53
Air Blank	0.000	03:53
Subject Sample # 1	0.177	03:54
Breath Volume	1.894	03:54
Air Blank	0.000	03:54
Air Blank	0.000	03:56
Subject Sample # 2	0.169	03:57
Breath Volume	1.988	03:57
Air Blank	0.000	03:57
Control Test	0.080	03:58
Air Blank	0.000	03:58
Diagnostics Check	OK	03:58

Data Download Date/Time: 02/01/2018 04:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PATE,MICHAEL,J
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MONIE,JARED
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 10:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:22

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/31/2017 11:52:35

Diagnostics Check	OK	11:44	
Air Blank	0.000	11:44	
Control Test	0.082	11:44	
Air Blank	0.000	11:45	
Subject Sample	# 1	11:48	No Sample Provided
Breath Volume	0.000	11:48	
Air Blank	0.000	11:49	
Air Blank	0.000	11:51	
Subject Sample	# 2	11:51	Subject Test Refused
Breath Volume	0.000	11:51	
Air Blank	0.000	11:51	
Control Test	0.082	11:52	
Air Blank	0.000	11:52	
Diagnostics Check	OK	11:52	

Data Download Date/Time: 02/01/2018 04:56

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEANITE,DUHAMEL,
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:15

SUBJECT'S LAST NAME: LANSING

DATE/TIME OF TEST: 12/09/2017 11:38:55

Diagnostics Check	OK	11:35	
Air Blank	0.000	11:36	
Control Test	0.081	11:36	
Air Blank	0.000	11:37	
Subject Sample	# 1	11:37	Subject Test Refused
Breath Volume	0.000	11:37	
Air Blank	0.000	11:37	
Control Test	0.081	11:38	
Air Blank	0.000	11:38	
Diagnostics Check	OK	11:38	

Data Download Date/Time: 12/21/2017 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,KEVIN,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: LOPEZ,KEVIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:06

SUBJECT'S LAST NAME: BACA

DATE/TIME OF TEST: 12/19/2017 05:36:23

Diagnostics Check	OK	05:33	
Air Blank	0.000	05:33	
Control Test	0.082	05:34	
Air Blank	0.000	05:34	
Subject Sample	# 1	05:34	Subject Test Refused
Breath Volume	0.000	05:34	
Air Blank	0.000	05:35	
Control Test	0.081	05:35	
Air Blank	0.000	05:36	
Diagnostics Check	OK	05:36	

Data Download Date/Time: 12/21/2017 10:22

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CARO

DATE/TIME OF TEST: 12/01/2017 05:59:51

Diagnostics Check	OK	05:56	
Air Blank	0.000	05:57	
Control Test	0.079	05:57	
Air Blank	0.000	05:58	
Subject Sample	# 1	05:58	RFI Detect
Breath Volume	0.000	05:58	
Air Blank	0.000	05:58	
Control Test	0.080	05:59	
Air Blank	0.000	05:59	
Diagnostics Check	OK	05:59	

Data Download Date/Time: 12/01/2017 18:14

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 11/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 05:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: CARO

DATE/TIME OF TEST: 12/01/2017 06:09:28

Diagnostics Check	OK	06:01
Air Blank	0.000	06:01
Control Test	0.080	06:01
Air Blank	0.000	06:02
Subject Sample # 1	0.231	06:03
Breath Volume	1.746	06:03
Air Blank	0.000	06:04
Air Blank	0.000	06:05
Subject Sample # 2	0.220	06:07
Breath Volume	2.019	06:07
Air Blank	0.000	06:08
Control Test	0.077	06:08
Air Blank	0.000	06:09
Diagnostics Check	OK	06:09

Data Download Date/Time: 12/01/2017 18:14

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CARHART

DATE/TIME OF TEST: 12/01/2017 23:09:21

Diagnostics Check	OK	23:08	
Air Blank	0.000	23:08	
Control Test	0.080	23:08	
Air Blank		23:09	RFI Detect
Air Blank		23:09	RFI Detect

Data Download Date/Time: 01/08/2018 17:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:47

SUBJECT'S LAST NAME: CARHART

DATE/TIME OF TEST: 12/01/2017 23:17:43

Diagnostics Check	OK	23:11
Air Blank	0.000	23:12
Control Test	0.079	23:12
Air Blank	0.000	23:12
Subject Sample # 1	0.148	23:13
Breath Volume	2.574	23:13
Air Blank	0.000	23:13
Air Blank	0.000	23:15
Subject Sample # 2	0.142	23:16
Breath Volume	2.917	23:16
Air Blank	0.000	23:16
Control Test	0.079	23:17
Air Blank	0.000	23:17
Diagnostics Check	OK	23:17

Data Download Date/Time: 01/08/2018 17:27

Florida Department of Law Enforcement
Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: LYONS,DANIEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 00:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:14

SUBJECT'S LAST NAME: MEADE

DATE/TIME OF TEST: 12/24/2017 00:41:48

Diagnostics Check	OK	00:35
Air Blank	0.000	00:35
Control Test	0.080	00:35
Air Blank	0.000	00:36
Subject Sample # 1	0.328	00:37
Breath Volume	2.789	00:37
Air Blank	0.000	00:38
Air Blank	0.000	00:39
Subject Sample # 2	0.328	00:40
Breath Volume	2.710	00:40
Air Blank	0.000	00:40
Control Test	0.078	00:41
Air Blank	0.000	00:41
Diagnostics Check	OK	00:41

Data Download Date/Time: 01/08/2018 17:27