Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ALVAREZ MARRERO

DATE/TIME OF TEST: 12/02/2017 02:33:39

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1		02:29	No Sample Provided
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Air Blank		0.000	02:32	
Subject Sample	# 2		02:32	Subject Test Refused
Breath Volume		0.000	02:32	
Air Blank		0.000	02:32	
Control Test		0.080	02:33	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/04/2017 01:11:21

Diagnostics Check		OK	01:04
Air Blank		0.000	01:04
Control Test		0.081	01:05
Air Blank		0.000	01:05
Subject Sample	# 1	0.264	01:06
Breath Volume		3.324	01:06
Air Blank		0.000	01:07
Air Blank		0.000	01:08
Subject Sample	# 2	0.278	01:09
Breath Volume		2.054	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MURGUIDO,ALEX
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 13:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:25

SUBJECT'S LAST NAME: DE DAGHER

DATE/TIME OF TEST: 12/07/2017 14:58:40

Diagnostics Check		OK	14:50
Air Blank		0.000	14:50
Control Test		0.080	14:51
Air Blank		0.000	14:51
Subject Sample	# 1	0.151	14:53
Breath Volume		1.199	14:53
Air Blank		0.000	14:53
Air Blank		0.000	14:55
Subject Sample	# 2	0.150	14:56
Breath Volume		1.199	14:56
Air Blank		0.000	14:57
Control Test		0.078	14:58
Air Blank		0.000	14:58
Diagnostics Check		OK	14:58

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MAYSONET

DATE/TIME OF TEST: 12/09/2017 01:17:44

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.081	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.135	01:13
Breath Volume		2.972	01:13
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.141	01:16
Breath Volume		3.046	01:16
Air Blank		0.000	01:16
Control Test		0.080	01:17
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: KAVANAUGH,J
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: VALDEZ VIEYRA

DATE/TIME OF TEST: 12/15/2017 04:03:37

Diagnostics Check		OK	03:56
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.207	03:58
Breath Volume		3.160	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.205	04:01
Breath Volume		2.765	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUIRRE,RICARDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:30

SUBJECT'S LAST NAME: ARNULFO DE LA CRUZ

DATE/TIME OF TEST: 12/16/2017 18:06:03

Diagnostics Check		OK	17:58	
Air Blank		0.000	17:58	
Control Test		0.080	17:58	
Air Blank		0.000	17:59	
Subject Sample	# 1	0.232	17:59	
Breath Volume		1.500	17:59	
Air Blank		0.000	18:00	
Air Blank		0.000	18:02	
Subject Sample	# 2	0.241	18:04	
Breath Volume		2.234	18:04	
Air Blank			18:05	Purge Fail
Air Blank		0.000	18:06	

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUIRRE,RICARDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:05

SUBJECT'S LAST NAME: ARNULFO DE LA CRUZ

DATE/TIME OF TEST: 12/16/2017 18:33:04

Diagnostics Check		OK	18:28	
Air Blank		0.000	18:29	
Control Test		0.076	18:29	
Air Blank		0.000	18:29	
Subject Sample	# 1	0.231	18:31	
Breath Volume		1.945	18:31	
Air Blank			18:32	Purge Fail
Air Blank		0.000	18:33	

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEX,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: AGUIRRE,RICARDO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:35

SUBJECT'S LAST NAME: ARNULFO DE LA CRUZ

DATE/TIME OF TEST: 12/16/2017 19:05:40

Diagnostics Check		OK	18:57	
Air Blank		0.000	18:57	
Control Test		0.078	18:57	
Air Blank		0.000	18:58	
Subject Sample	# 1		19:01	No Sample Provided
Breath Volume		0.000	19:01	
Air Blank		0.000	19:01	
Air Blank		0.000	19:03	
Subject Sample	# 2		19:04	Subject Test Refused
Breath Volume		0.000	19:04	
Air Blank		0.000	19:04	
Control Test		0.080	19:05	
Air Blank		0.000	19:05	
Diagnostics Check		OK	19:05	

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: FREER

DATE/TIME OF TEST: 12/17/2017 03:47:57

Diagnostics Check		OK	03:33	
Air Blank		0.000	03:33	
Control Test		0.081	03:34	
Air Blank		0.000	03:34	
Subject Sample	# 1	0.181	03:37	Volume Not Met
Breath Volume		0.085	03:37	
Air Blank		0.000	03:38	
Air Blank		0.000	03:40	
Subject Sample	# 2	0.174	03:41	
Breath Volume		1.765	03:41	
Air Blank		0.000	03:41	
Air Blank		0.000	03:43	
Subject Sample	# 3		03:46	Subject Test Refused
Breath Volume		0.148	03:46	
Air Blank		0.000	03:47	
Control Test		0.079	03:47	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 12/18/2017 03:02:00

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:53	
Control Test		0.081	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:57	No Sample Provided
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Air Blank		0.000	03:00	
Subject Sample	# 2		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:01	
Control Test		0.081	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:02	

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: GUERRERO,L
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/18/2017 04:08:46

Diagnostics Check		OK	04:05	
Air Blank		0.000	04:06	
Control Test		0.080	04:06	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:07	Subject Test Refused
Breath Volume		0.000	04:07	
Air Blank		0.000	04:07	
Control Test		0.080	04:08	
Air Blank		0.000	04:08	
Diagnostics Check		OK	04:08	

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: DIAZ GARCIA

DATE/TIME OF TEST: 12/19/2017 02:02:28

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:54
Subject Sample	# 1	0.256	01:56
Breath Volume		1.613	01:56
Air Blank		0.000	01:57
Air Blank		0.000	01:59
Subject Sample	# 2	0.238	02:00
Breath Volume		1.589	02:00
Air Blank		0.000	02:01
Control Test		0.077	02:01
Air Blank		0.000	02:02
Diagnostics Check		OK	02:02

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYS,DARRYL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MAYS,DARRYL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 18:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/22/2017 18:54:26

Diagnostics Check		OK	18:47
Air Blank		0.000	18:47
Control Test		0.081	18:47
Air Blank		0.000	18:48
Subject Sample	# 1	0.149	18:49
Breath Volume		1.605	18:49
Air Blank		0.000	18:50
Air Blank		0.000	18:51
Subject Sample	# 2	0.163	18:52
Breath Volume		2.085	18:52
Air Blank		0.000	18:53
Control Test		0.080	18:53
Air Blank		0.000	18:54
Diagnostics Check		OK	18:54

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 12/23/2017 00:55:56

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.081	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.173	00:50
Breath Volume		1.605	00:50
Air Blank		0.000	00:50
Air Blank		0.000	00:52
Subject Sample	# 2	0.163	00:54
Breath Volume		1.828	00:54
Air Blank		0.000	00:55
Control Test		0.079	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ORTIZ

DATE/TIME OF TEST: 12/23/2017 02:34:00

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.173	02:29
Breath Volume		1.144	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:31
Subject Sample	# 2	0.156	02:32
Breath Volume		1.140	02:32
Air Blank		0.000	02:33
Control Test		0.078	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 12/27/2017 17:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURGUIDO,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: MURGUIDO,AALEJANDRO
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:50

SUBJECT'S LAST NAME: MADRID

DATE/TIME OF TEST: 12/29/2017 16:26:06

Diagnostics Check		OK	16:18
Air Blank		0.000	16:18
Control Test		0.080	16:19
Air Blank		0.000	16:19
Subject Sample	# 1	0.327	16:21
Breath Volume		1.382	16:21
Air Blank		0.000	16:22
Air Blank		0.000	16:23
Subject Sample	# 2	0.328	16:24
Breath Volume		1.652	16:24
Air Blank		0.000	16:25
Control Test		0.077	16:25
Air Blank		0.000	16:25
Diagnostics Check		OK	16:26

Data Download Date/Time: 01/29/2018 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 06:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: CARRERAS

DATE/TIME OF TEST: 12/30/2017 06:46:21

Diagnostics Check		OK	06:35
Air Blank		0.000	06:35
Control Test		0.080	06:36
Air Blank		0.000	06:36
Subject Sample	# 1	0.365	06:37
Breath Volume		2.019	06:37
Air Blank		0.000	06:38
Air Blank		0.000	06:40
Subject Sample	# 2	0.393	06:41
Breath Volume		2.222	06:41
Air Blank		0.000	06:42
Air Blank		0.000	06:43
Subject Sample	# 3	0.371	06:44
Breath Volume		1.144	06:44
Air Blank		0.000	06:45
Control Test		0.078	06:45
Air Blank		0.000	06:46
Diagnostics Check		OK	06:46

Data Download Date/Time: 01/29/2018 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006199

INSTRUMENT REGISTERED TO: HOMESTEAD PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,RAUL,
BREATH TEST OPERATOR AGENCY: Homestead PD
ARRESTING OFFICER: PEREZ,RAUL
ARRESTING OFFICER AGENCY: Homestead PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 477627
DRY GAS EXPIRY DATE: 06/29/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: CHAVEZ

DATE/TIME OF TEST: 12/31/2017 02:52:40

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.172	02:47
Breath Volume		3.570	02:47
Air Blank		0.000	02:48
Air Blank		0.000	02:50
Subject Sample	# 2	0.169	02:51
Breath Volume		3.097	02:51
Air Blank		0.000	02:51
Control Test		0.080	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:52

Data Download Date/Time: 01/29/2018 15:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006214

INSTRUMENT REGISTERED TO: OCALA POLICE DEPT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOEDICKE,WILLIAM,R
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: JOEDICKE,WILLIAM
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/01/2017 21:59:33

Diagnostics Check		OK	21:52
Air Blank		0.000	21:53
Control Test		0.078	21:53
Air Blank		0.000	21:53
Subject Sample	# 1	0.215	21:54
Breath Volume		1.289	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.222	21:58
Breath Volume		1.640	21:58
Air Blank		0.000	21:58
Control Test		0.078	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 12/03/2017 20:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BARRETT,TAYLOR
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/02/2017 03:15:25

Diagnostics Check		OK	03:06
Air Blank		0.000	03:07
Control Test		0.078	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.182	03:09
Breath Volume		1.476	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.186	03:13
Breath Volume		1.738	03:13
Air Blank		0.000	03:14
Control Test		0.077	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 01/02/2018 20:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILKIN,SEAN,D
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: JACOBS,JEFFRIE
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: RODRIGUEZ SALGADO

DATE/TIME OF TEST: 12/03/2017 00:36:56

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:31
Subject Sample	# 1	0.201	00:32
Breath Volume		4.312	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.203	00:35
Breath Volume		4.511	00:35
Air Blank		0.000	00:36
Control Test		0.078	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: QUINTAL,VIRGEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 12/06/2017 21:18:47

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:10	
Control Test		0.077	21:10	
Air Blank		0.000	21:11	
Subject Sample	# 1	0.246	21:12	
Breath Volume		1.207	21:12	
Air Blank		0.000	21:13	
Air Blank		0.000	21:15	
Subject Sample	# 2	0.229	21:17	
Breath Volume		1.218	21:17	
Air Blank		0.000	21:17	
Control Test		0.074	21:18	Control Outside Tolerance
Air Blank		0.000	21:18	

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: QUINTAL,VIRGEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 12/06/2017 21:34:30

Diagnostics Check		OK	21:21	
Air Blank		0.000	21:21	
Control Test		0.075	21:22	
Air Blank		0.000	21:22	
Subject Sample	# 1	0.230	21:23	No .020 agreement
Breath Volume		1.750	21:23	
Air Blank		0.000	21:24	
Air Blank		0.000	21:26	
Subject Sample	# 2	0.255	21:26	No .020 agreement
Breath Volume		2.261	21:26	
Air Blank		0.000	21:27	
Air Blank		0.000	21:29	
Subject Sample	# 3	0.233	21:32	Volume Not Met
Breath Volume		0.085	21:32	
Air Blank		0.000	21:33	
Control Test		0.075	21:33	
Air Blank		0.000	21:34	
Diagnostics Check		OK	21:34	

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: QUINTAL,VIRGEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 12/06/2017 21:39:06

Diagnostics Check	OK	21:37	
Air Blank	0.000	21:38	
Control Test	0.073	21:38	Control Outside Tolerance
Air Blank	0.000	21:39	

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAVN,SUSANNE,M
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: QUINTAL,VIRGEL
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:45

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 12/06/2017 21:49:20

Diagnostics Check		OK	21:40	
Air Blank		0.000	21:40	
Control Test		0.075	21:41	
Air Blank		0.000	21:41	
Subject Sample	# 1	0.246	21:43	
Breath Volume		1.679	21:43	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 2	0.231	21:47	
Breath Volume		1.449	21:47	
Air Blank		0.000	21:48	
Control Test		0.074	21:48	Control Outside Tolerance
Air Blank		0.000	21:49	

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: CRONIN,SEAN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: TATE

DATE/TIME OF TEST: 12/15/2017 03:19:07

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.078	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.197	03:13
Breath Volume		2.355	03:13
Air Blank		0.000	03:14
Air Blank		0.000	03:16
Subject Sample	# 2	0.180	03:17
Breath Volume		1.183	03:17
Air Blank		0.000	03:18
Control Test		0.076	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEAHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: PIEDRA

DATE/TIME OF TEST: 12/19/2017 23:44:31

Diagnostics Check		OK	23:38
Air Blank		0.000	23:38
Control Test		0.078	23:38
Air Blank		0.000	23:39
Subject Sample	# 1	0.053	23:39
Breath Volume		2.484	23:39
Air Blank		0.000	23:40
Air Blank		0.000	23:42
Subject Sample	# 2	0.051	23:42
Breath Volume		2.363	23:42
Air Blank		0.000	23:43
Control Test		0.078	23:43
Air Blank		0.000	23:44
Diagnostics Check		OK	23:44

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: VON HOLLEN

DATE/TIME OF TEST: 12/22/2017 02:22:08

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.064	02:17
Breath Volume		2.195	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.067	02:20
Breath Volume		2.187	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MANGRUM,JULIANN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: COHAN

DATE/TIME OF TEST: 12/24/2017 00:26:55

Diagnostics Check		OK	00:15	
Air Blank		0.000	00:15	
Control Test		0.079	00:16	
Air Blank		0.000	00:16	
Subject Sample	# 1		00:19	No Sample Provided
Breath Volume		0.000	00:19	
Air Blank		0.000	00:20	
Air Blank		0.000	00:22	
Subject Sample	# 2		00:25	No Sample Provided
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.079	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MANGRUM,JULIANN
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: COHAN

DATE/TIME OF TEST: 12/24/2017 00:49:39

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.079	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.171	00:44
Breath Volume		2.210	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.162	00:48
Breath Volume		1.703	00:48
Air Blank		0.000	00:48
Control Test		0.078	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIVINS,HEATHER,C
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: BIVINS,HEATHER
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/24/2017 03:21:38

Diagnostics Check		OK	03:14
Air Blank		0.000	03:15
Control Test		0.078	03:15
Air Blank		0.000	03:15
Subject Sample	# 1	0.122	03:16
Breath Volume		1.421	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:19
Subject Sample	# 2	0.119	03:20
Breath Volume		1.480	03:20
Air Blank		0.000	03:20
Control Test		0.078	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:12

SUBJECT'S LAST NAME: CZYZEWSKI

DATE/TIME OF TEST: 12/28/2017 00:39:36

Diagnostics Check		OK	00:32
Air Blank		0.000	00:33
Control Test		0.077	00:33
Air Blank		0.000	00:34
Subject Sample	# 1	0.168	00:34
Breath Volume		1.125	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.177	00:38
Breath Volume		1.554	00:38
Air Blank		0.000	00:38
Control Test		0.077	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:34

SUBJECT'S LAST NAME: QUATTROCCHI

DATE/TIME OF TEST: 12/29/2017 23:57:23

Diagnostics Check		OK	23:54	
Air Blank		0.000	23:54	
Control Test		0.078	23:55	
Air Blank		0.000	23:55	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:56	
Control Test		0.078	23:56	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006233

INSTRUMENT REGISTERED TO: BREVARD COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOORE,KIMBERLY,N
BREATH TEST OPERATOR AGENCY: Brevard County SO
ARRESTING OFFICER: MOORE,KIMBERLY
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: LOVEDAY

DATE/TIME OF TEST: 12/30/2017 02:41:58

Diagnostics Check		OK	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:36
Subject Sample	# 1	0.087	02:37
Breath Volume		2.707	02:37
Air Blank		0.000	02:38
Air Blank		0.000	02:39
Subject Sample	# 2	0.079	02:40
Breath Volume		1.781	02:40
Air Blank		0.000	02:41
Control Test		0.077	02:41
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 01/02/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: MENDEZ

DATE/TIME OF TEST: 12/10/2017 18:47:08

Diagnostics Check		OK	18:40
Air Blank		0.000	18:41
Control Test		0.083	18:41
Air Blank		0.000	18:41
Subject Sample	# 1	0.128	18:42
Breath Volume		1.605	18:42
Air Blank		0.000	18:43
Air Blank		0.000	18:44
Subject Sample	# 2	0.135	18:45
Breath Volume		1.707	18:45
Air Blank		0.000	18:46
Control Test		0.083	18:46
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 12/18/2017 11:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006236

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: STEWARD

DATE/TIME OF TEST: 12/13/2017 21:14:15

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.083	21:08
Air Blank		0.000	21:08
Subject Sample	# 1	0.213	21:09
Breath Volume		1.597	21:09
Air Blank		0.000	21:10
Air Blank		0.000	21:11
Subject Sample	# 2	0.218	21:12
Breath Volume		2.464	21:12
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 12/18/2017 11:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:37

SUBJECT'S LAST NAME: MURPHY

DATE/TIME OF TEST: 12/03/2017 09:07:31

Diagnostics Check		OK	08:59
Air Blank		0.000	09:00
Control Test		0.080	09:00
Air Blank		0.000	09:01
Subject Sample	# 1	0.154	09:02
Breath Volume		1.617	09:02
Air Blank		0.000	09:03
Air Blank		0.000	09:05
Subject Sample	# 2	0.154	09:05
Breath Volume		2.582	09:05
Air Blank		0.000	09:06
Control Test		0.079	09:06
Air Blank		0.000	09:07
Diagnostics Check		OK	09:07

Data Download Date/Time: 12/23/2017 16:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: CAMPAU

DATE/TIME OF TEST: 12/04/2017 22:52:10

Diagnostics Check		OK	22:41	
Air Blank		0.000	22:41	
Control Test		0.080	22:41	
Air Blank		0.000	22:42	
Subject Sample	# 1	0.181	22:42	
Breath Volume		2.125	22:42	
Air Blank		0.000	22:43	
Air Blank		0.000	22:45	
Subject Sample	# 2	0.156	22:47	
Breath Volume		1.433	22:47	
Air Blank		0.000	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 3	0.163	22:50	
Breath Volume		2.921	22:50	
Air Blank		0.000	22:51	
Control Test		0.074	22:51	Control Outside Tolerance
Air Blank		0.000	22:52	

Data Download Date/Time: 12/23/2017 16:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAUNDERS,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SYDOROWICZ,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 12/23/2017 16:36:58

Diagnostics Check		OK	16:28
Air Blank		0.000	16:29
Control Test		0.080	16:29
Air Blank		0.000	16:30
Subject Sample	# 1	0.267	16:31
Breath Volume		1.128	16:31
Air Blank		0.000	16:32
Air Blank		0.000	16:33
Subject Sample	# 2	0.254	16:35
Breath Volume		1.195	16:35
Air Blank		0.000	16:36
Control Test		0.079	16:36
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 01/14/2018 14:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006237

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 12/28/2017 00:53:12

Diagnostics Check		OK	00:46
Air Blank		0.000	00:47
Control Test		0.080	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.164	00:48
Breath Volume		1.851	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.162	00:51
Breath Volume		1.679	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 01/14/2018 14:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LESKOW,JORDAN
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: GANDIA

DATE/TIME OF TEST: 12/15/2017 23:43:15

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Subject Sample	# 1	0.113	23:38
Breath Volume		2.011	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.110	23:41
Breath Volume		2.140	23:41
Air Blank		0.000	23:42
Control Test		0.080	23:42
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO
RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:59
GENDER: Male

OBSERVATION PERIOD BEGAN: 02:25

SUBJECT'S LAST NAME: BAEZ

DATE/TIME OF TEST: 12/16/2017 02:56:51

Diagnostics Check		OK	02:49
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.174	02:51
Breath Volume		2.023	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.168	02:55
Breath Volume		1.851	02:55
Air Blank		0.000	02:55
Control Test		0.080	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:17

SUBJECT'S LAST NAME: GOSNELL

DATE/TIME OF TEST: 12/16/2017 05:46:14

Diagnostics Check		OK	05:41	
Air Blank		0.000	05:41	
Control Test		0.080	05:41	
Air Blank		0.000	05:42	
Subject Sample	# 1		05:44	Subject Test Refused
Breath Volume		0.000	05:44	
Air Blank		0.000	05:45	
Control Test		0.080	05:45	
Air Blank		0.000	05:46	
Diagnostics Check		OK	05:46	

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DEVAUL,ERIC
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: JOLLY

DATE/TIME OF TEST: 12/17/2017 02:34:21

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.163	02:29
Breath Volume		3.328	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.161	02:32
Breath Volume		3.429	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SALVEMINI,MICHAEL
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: POPP

DATE/TIME OF TEST: 12/17/2017 03:31:59

Diagnostics Check		OK	03:20	
Air Blank		0.000	03:20	
Control Test		0.081	03:21	
Air Blank		0.000	03:21	
Subject Sample	# 1	0.145	03:24	Volume Not Met
Breath Volume		0.781	03:24	
Air Blank		0.000	03:25	
Air Blank		0.000	03:27	
Subject Sample	# 2	0.119	03:30	Volume Not Met
Breath Volume		0.355	03:30	
Air Blank		0.000	03:31	
Control Test		0.078	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PETTI,JENNA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RUNGE

DATE/TIME OF TEST: 12/18/2017 03:55:11

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:50	
Control Test		0.080	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1	0.101	03:51	
Breath Volume		1.734	03:51	
Air Blank		0.000	03:52	
Air Blank			03:53	RFI Detect
Air Blank		0.000	03:54	
Control Test		0.081	03:54	
Air Blank		0.000	03:55	
Diagnostics Check		OK	03:55	

Data Download Date/Time: 01/19/2018 12:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PETTI,JENNA
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: RUNGE

DATE/TIME OF TEST: 12/18/2017 04:09:31

Diagnostics Check		OK	03:56	
Air Blank		0.000	03:56	
Control Test		0.080	03:57	
Air Blank		0.000	03:57	
Subject Sample	# 1	0.103	03:58	No .020 agreement
Breath Volume		2.296	03:58	
Air Blank		0.000	03:58	
Air Blank		0.000	04:00	
Subject Sample	# 2	0.076	04:03	Volume Not Met
Breath Volume		0.050	04:03	
Air Blank		0.000	04:04	
Air Blank		0.000	04:06	
Subject Sample	# 3	0.082	04:07	No .020 agreement
Breath Volume		1.519	04:07	
Air Blank		0.000	04:08	
Control Test		0.078	04:08	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 01/19/2018 12:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTILLO,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: CORVIN

DATE/TIME OF TEST: 12/19/2017 02:17:23

Diagnostics Check		OK	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:11
Air Blank		0.000	02:12
Subject Sample	# 1	0.210	02:12
Breath Volume		1.375	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.210	02:15
Breath Volume		1.789	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Diagnostics Check		OK	02:17

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PHILIPPE,BENDJY
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 12/19/2017 02:34:01

Diagnostics Check		OK	02:27
Air Blank		0.000	02:28
Control Test		0.080	02:28
Air Blank		0.000	02:29
Subject Sample	# 1	0.195	02:29
Breath Volume		2.632	02:29
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.193	02:32
Breath Volume		2.519	02:32
Air Blank		0.000	02:33
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:34

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,KEITH
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: WUNDER

DATE/TIME OF TEST: 12/20/2017 23:16:34

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:04	
Control Test		0.079	23:04	
Air Blank		0.000	23:05	
Subject Sample	# 1		23:08	No Sample Provided
Breath Volume		0.000	23:08	
Air Blank		0.000	23:08	
Air Blank		0.000	23:10	
Subject Sample	# 2	0.172	23:12	
Breath Volume		1.835	23:12	
Air Blank		0.000	23:12	
Air Blank		0.000	23:14	
Subject Sample	# 3	0.178	23:14	
Breath Volume		1.554	23:14	
Air Blank		0.000	23:15	
Control Test		0.077	23:15	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 01/19/2018 12:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SORIANO,REYNALDO
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ELIAS

DATE/TIME OF TEST: 12/21/2017 00:19:16

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.080	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.293	00:14
Breath Volume		1.574	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.300	00:17
Breath Volume		1.917	00:17
Air Blank		0.000	00:18
Control Test		0.079	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 01/19/2018 12:26

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONEY,MICHAEL
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:37

SUBJECT'S LAST NAME: GAGLIARDI

DATE/TIME OF TEST: 12/21/2017 03:06:11

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.020	03:01
Breath Volume		2.699	03:01
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.017	03:04
Breath Volume		1.187	03:04
Air Blank		0.000	03:05
Control Test		0.079	03:05
Air Blank		0.000	03:06
Diagnostics Check		OK	03:06

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BALDASSARI

DATE/TIME OF TEST: 12/21/2017 04:54:55

Diagnostics Check		OK	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:50
Subject Sample	# 1	0.216	04:50
Breath Volume		1.343	04:50
Air Blank		0.000	04:51
Air Blank		0.000	04:52
Subject Sample	# 2	0.208	04:53
Breath Volume		1.457	04:53
Air Blank		0.000	04:54
Control Test		0.079	04:54
Air Blank		0.000	04:54
Diagnostics Check		OK	04:54

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: SIMS,JR.

DATE/TIME OF TEST: 12/21/2017 22:26:32

Diagnostics Check		OK	22:19
Air Blank		0.000	22:19
Control Test		0.080	22:20
Air Blank		0.000	22:20
Subject Sample	# 1	0.000	22:21
Breath Volume		2.281	22:21
Air Blank		0.000	22:22
Air Blank		0.000	22:24
Subject Sample	# 2	0.000	22:25
Breath Volume		1.742	22:25
Air Blank		0.000	22:25
Control Test		0.080	22:25
Air Blank		0.000	22:26
Diagnostics Check		OK	22:26

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MYERSCOUGH,DONALD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: VILLATORO

DATE/TIME OF TEST: 12/23/2017 03:01:08

Diagnostics Check		OK	02:54
Air Blank		0.000	02:54
Control Test		0.080	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.115	02:56
Breath Volume		1.628	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.113	02:59
Breath Volume		1.324	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ROBINSON,ROBERT
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: NORWOOD IV

DATE/TIME OF TEST: 12/23/2017 03:31:53

Diagnostics Check		OK	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:23
Subject Sample	# 1	0.154	03:27
Breath Volume		1.273	03:27
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.160	03:30
Breath Volume		1.523	03:30
Air Blank		0.000	03:30
Control Test		0.077	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZUCCARELLI,MICHAEL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 12/23/2017 04:47:45

Diagnostics Check		OK	04:42	
Air Blank		0.000	04:42	
Control Test		0.079	04:43	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:46	Subject Test Refused
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.081	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ZUCCARELLI,MICHAEL
ARRESTING OFFICER AGENCY: North Palm Beach
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 12/23/2017 04:57:30

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.080	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1		04:56	Subject Test Refused
Breath Volume		0.000	04:56	
Air Blank		0.000	04:56	
Control Test		0.081	04:56	
Air Blank		0.000	04:57	
Diagnostics Check		OK	04:57	

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: ESCARAN,LESTER
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:49

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/23/2017 07:25:16

Diagnostics Check		OK	07:13	
Air Blank		0.000	07:13	
Control Test		0.080	07:13	
Air Blank		0.000	07:14	
Subject Sample	# 1	0.100	07:17	Volume Not Met
Breath Volume		0.742	07:17	
Air Blank		0.000	07:18	
Air Blank		0.000	07:20	
Subject Sample	# 2	0.105	07:20	
Breath Volume		1.687	07:20	
Air Blank		0.000	07:21	
Air Blank		0.000	07:23	
Subject Sample	# 3	0.107	07:23	
Breath Volume		1.351	07:23	
Air Blank		0.000	07:24	
Control Test		0.079	07:24	
Air Blank		0.000	07:25	
Diagnostics Check		OK	07:25	

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HENNESSY,DONALD
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: TZUNUN-PAR

DATE/TIME OF TEST: 12/24/2017 01:59:29

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.357	01:54
Breath Volume		1.531	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.354	01:57
Breath Volume		2.031	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PUJOL,JESUS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: KARVELA

DATE/TIME OF TEST: 12/24/2017 02:54:16

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.080	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.184	02:49
Breath Volume		1.328	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.182	02:52
Breath Volume		1.312	02:52
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: FRANCISCO-MIGUEL

DATE/TIME OF TEST: 12/24/2017 07:15:12

Diagnostics Check		OK	07:08
Air Blank		0.000	07:08
Control Test		0.080	07:08
Air Blank		0.000	07:09
Subject Sample	# 1	0.167	07:10
Breath Volume		2.355	07:10
Air Blank		0.000	07:11
Air Blank		0.000	07:13
Subject Sample	# 2	0.171	07:13
Breath Volume		2.503	07:13
Air Blank		0.000	07:14
Control Test		0.080	07:14
Air Blank		0.000	07:15
Diagnostics Check		OK	07:15

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: VALERIO,MICHAEL
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 12:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:43

SUBJECT'S LAST NAME: KASICA

DATE/TIME OF TEST: 12/24/2017 14:14:29

Diagnostics Check		OK	14:07
Air Blank		0.000	14:08
Control Test		0.080	14:08
Air Blank		0.000	14:08
Subject Sample	# 1	0.000	14:10
Breath Volume		1.585	14:10
Air Blank		0.000	14:10
Air Blank		0.000	14:12
Subject Sample	# 2	0.000	14:13
Breath Volume		1.250	14:13
Air Blank		0.000	14:13
Control Test		0.080	14:13
Air Blank		0.000	14:14
Diagnostics Check		OK	14:14

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WARREN,SAM
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: PANTUSO

DATE/TIME OF TEST: 12/25/2017 01:37:49

Diagnostics Check		OK	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:32
Subject Sample	# 1	0.189	01:33
Breath Volume		2.343	01:33
Air Blank		0.000	01:34
Air Blank		0.000	01:35
Subject Sample	# 2	0.194	01:36
Breath Volume		2.019	01:36
Air Blank		0.000	01:36
Control Test		0.079	01:37
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male

OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: LOPEZ REYES

DATE/TIME OF TEST: 12/25/2017 03:37:39

Diagnostics Check		OK	03:31
Air Blank		0.000	03:31
Control Test		0.080	03:32
Air Blank		0.000	03:32
Subject Sample	# 1	0.219	03:33
Breath Volume		1.722	03:33
Air Blank		0.000	03:33
Air Blank		0.000	03:35
Subject Sample	# 2	0.226	03:36
Breath Volume		1.437	03:36
Air Blank		0.000	03:36
Control Test		0.079	03:37
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/19/2018 12:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMNATHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CAVER,CAMERON
ARRESTING OFFICER AGENCY: Palm Beach Gardens
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: GONZALEZ CHAJ

DATE/TIME OF TEST: 12/25/2017 05:46:44

Diagnostics Check		OK	05:39
Air Blank		0.000	05:39
Control Test		0.080	05:39
Air Blank		0.000	05:40
Subject Sample	# 1	0.233	05:41
Breath Volume		2.117	05:41
Air Blank		0.000	05:42
Air Blank		0.000	05:44
Subject Sample	# 2	0.243	05:45
Breath Volume		2.589	05:45
Air Blank		0.000	05:45
Control Test		0.079	05:46
Air Blank		0.000	05:46
Diagnostics Check		OK	05:46

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BEESLEY,MARCUS
ARRESTING OFFICER AGENCY: Lantana PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:59

SUBJECT'S LAST NAME: HARBIN

DATE/TIME OF TEST: 12/25/2017 17:03:20

Diagnostics Check		OK	16:48	
Air Blank		0.000	16:49	
Control Test		0.079	16:49	
Air Blank		0.000	16:50	
Subject Sample	# 1	0.000	16:50	
Breath Volume		1.843	16:50	
Air Blank		0.000	16:51	
Air Blank		0.000	16:53	
Subject Sample	# 2		16:56	No Sample Provided
Breath Volume		0.000	16:56	
Air Blank		0.000	16:57	
Air Blank		0.000	16:59	
Subject Sample	# 3	0.000	17:01	
Breath Volume		1.460	17:01	
Air Blank		0.000	17:02	
Control Test		0.080	17:02	
Air Blank		0.000	17:03	
Diagnostics Check		OK	17:03	

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 12/26/2017 04:31:50

Diagnostics Check		OK	04:28	
Air Blank		0.000	04:29	
Control Test		0.080	04:29	
Air Blank		0.000	04:29	
Subject Sample	# 1		04:30	Slope Not Met
Breath Volume		0.910	04:30	
Air Blank		0.000	04:30	
Control Test		0.080	04:31	
Air Blank		0.000	04:31	
Diagnostics Check		OK	04:31	

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BIGGS,JAMES,G
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,DAVID
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 12/26/2017 04:58:36

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.080	04:52
Air Blank		0.000	04:53
Subject Sample	# 1	0.237	04:54
Breath Volume		1.996	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.256	04:56
Breath Volume		2.835	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: AFANADOR

DATE/TIME OF TEST: 12/26/2017 22:40:59

Diagnostics Check		OK	22:30
Air Blank		0.000	22:30
Control Test		0.079	22:31
Air Blank		0.000	22:31
Subject Sample	# 1	0.248	22:32
Breath Volume		1.484	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 2	0.224	22:35
Breath Volume		1.132	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 3	0.233	22:39
Breath Volume		1.839	22:39
Air Blank		0.000	22:40
Control Test		0.078	22:40
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 01/19/2018 12:27

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: RIZVEE

DATE/TIME OF TEST: 12/28/2017 04:21:16

Diagnostics Check		OK	04:13
Air Blank		0.000	04:14
Control Test		0.080	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.122	04:16
Breath Volume		1.632	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:18
Subject Sample	# 2	0.120	04:19
Breath Volume		1.417	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: DENIS,ISMARY
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: BURGER

DATE/TIME OF TEST: 12/28/2017 21:07:04

Diagnostics Check		OK	21:01
Air Blank		0.000	21:01
Control Test		0.081	21:01
Air Blank		0.000	21:02
Subject Sample	# 1	0.000	21:02
Breath Volume		3.683	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.000	21:05
Breath Volume		3.363	21:05
Air Blank		0.000	21:06
Control Test		0.081	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:07

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCHNEIDER,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: QUINTERO NARVAEZ

DATE/TIME OF TEST: 12/29/2017 03:16:57

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.081	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.122	03:12
Breath Volume		2.839	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.119	03:15
Breath Volume		2.367	03:15
Air Blank		0.000	03:16
Control Test		0.079	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PETRICONE,MICHELLE
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: TELLO ORTIZ

DATE/TIME OF TEST: 12/30/2017 00:07:56

Diagnostics Check		OK	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Subject Sample	# 1	0.134	00:02
Breath Volume		1.339	00:02
Air Blank		0.000	00:02
Air Blank		0.000	00:04
Subject Sample	# 2	0.126	00:06
Breath Volume		2.105	00:06
Air Blank		0.000	00:07
Control Test		0.079	00:07
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FICARRA,CHRISTOPHER
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 12/30/2017 02:59:48

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.080	02:54
Air Blank		0.000	02:54
Subject Sample	# 1	0.145	02:55
Breath Volume		1.417	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:57
Subject Sample	# 2	0.139	02:58
Breath Volume		1.464	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006239

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MERCIER,DANIEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: KEUHNART

DATE/TIME OF TEST: 12/30/2017 05:40:57

Diagnostics Check		OK	05:34
Air Blank		0.000	05:34
Control Test		0.081	05:35
Air Blank		0.000	05:35
Subject Sample	# 1	0.141	05:36
Breath Volume		2.160	05:36
Air Blank		0.000	05:36
Air Blank		0.000	05:38
Subject Sample	# 2	0.141	05:39
Breath Volume		1.640	05:39
Air Blank		0.000	05:40
Control Test		0.080	05:40
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 01/19/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: TILLEY

DATE/TIME OF TEST: 12/06/2017 21:22:05

Diagnostics Check		OK	21:15
Air Blank		0.000	21:16
Control Test		0.081	21:16
Air Blank		0.000	21:17
Subject Sample	# 1	0.068	21:17
Breath Volume		2.183	21:17
Air Blank		0.000	21:18
Air Blank		0.000	21:20
Subject Sample	# 2	0.067	21:20
Breath Volume		1.996	21:20
Air Blank		0.000	21:21
Control Test		0.081	21:21
Air Blank		0.000	21:21
Diagnostics Check		OK	21:22

Data Download Date/Time: 12/28/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PATSCH,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:34

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 12/12/2017 21:01:11

Diagnostics Check		OK	20:54	
Air Blank		0.000	20:54	
Control Test		0.080	20:55	
Air Blank		0.000	20:55	
Subject Sample	# 1	0.287	20:56	
Breath Volume		3.027	20:56	
Air Blank		0.000	20:56	
Air Blank		0.000	20:58	
Subject Sample	# 2		21:00	Sequence Aborted
Breath Volume		0.000	21:00	
Air Blank		0.000	21:01	

Data Download Date/Time: 12/28/2017 11:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: PATSCH,JOSHUA
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: STACK

DATE/TIME OF TEST: 12/12/2017 21:28:26

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.080	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.273	21:23
Breath Volume		2.914	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.276	21:26
Breath Volume		2.906	21:26
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 12/28/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006262

INSTRUMENT REGISTERED TO: VOLUSIA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CALKINS,BRIAN,M
BREATH TEST OPERATOR AGENCY: Volusia County SO
ARRESTING OFFICER: HELMS,WILLIAM
ARRESTING OFFICER AGENCY: New Smyrna Beach PD
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: CAMBRIDGE

DATE/TIME OF TEST: 12/17/2017 19:43:57

Diagnostics Check		OK	19:37
Air Blank		0.000	19:37
Control Test		0.079	19:37
Air Blank		0.000	19:38
Subject Sample	# 1	0.262	19:39
Breath Volume		2.468	19:39
Air Blank		0.000	19:39
Air Blank		0.000	19:41
Subject Sample	# 2	0.269	19:42
Breath Volume		1.343	19:42
Air Blank		0.000	19:43
Control Test		0.078	19:43
Air Blank		0.000	19:43
Diagnostics Check		OK	19:43

Data Download Date/Time: 12/28/2017 11:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: HOLTZ

DATE/TIME OF TEST: 12/02/2017 02:20:12

Diagnostics Check		OK	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:14
Subject Sample	# 1	0.133	02:15
Breath Volume		2.953	02:15
Air Blank		0.000	02:16
Air Blank		0.000	02:18
Subject Sample	# 2	0.126	02:18
Breath Volume		2.839	02:18
Air Blank		0.000	02:19
Control Test		0.080	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 12/06/2017 20:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 11/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SHULTZ

DATE/TIME OF TEST: 12/03/2017 03:29:27

Diagnostics Check		OK	03:22
Air Blank		0.000	03:22
Control Test		0.082	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.136	03:24
Breath Volume		2.003	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.127	03:27
Breath Volume		1.367	03:27
Air Blank		0.000	03:28
Control Test		0.079	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 12/06/2017 20:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:19

SUBJECT'S LAST NAME: LOLLIS

DATE/TIME OF TEST: 12/07/2017 21:55:47

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.079	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.196	21:50
Breath Volume		1.621	21:50
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.194	21:54
Breath Volume		1.500	21:54
Air Blank		0.000	21:54
Control Test		0.079	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:55

Data Download Date/Time: 01/11/2018 20:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: HIGGINBOTHAM

DATE/TIME OF TEST: 12/08/2017 03:18:50

Diagnostics Check		OK	03:09	
Air Blank		0.000	03:09	
Control Test		0.081	03:10	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:13	No Sample Provided
Breath Volume		0.000	03:13	
Air Blank		0.000	03:14	
Air Blank		0.000	03:16	
Subject Sample	# 2		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:17	
Control Test		0.081	03:18	
Air Blank		0.000	03:18	
Diagnostics Check		OK	03:18	

Data Download Date/Time: 01/11/2018 20:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:27

SUBJECT'S LAST NAME: LEON

DATE/TIME OF TEST: 12/10/2017 01:01:23

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.081	00:52	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:56	No Sample Provided
Breath Volume		0.000	00:56	
Air Blank		0.000	00:57	
Air Blank		0.000	00:59	
Subject Sample	# 2		00:59	Subject Test Refused
Breath Volume		0.000	00:59	
Air Blank		0.000	01:00	
Control Test		0.081	01:00	
Air Blank		0.000	01:01	
Diagnostics Check		OK	01:01	

Data Download Date/Time: 01/11/2018 20:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: VENCIL

DATE/TIME OF TEST: 12/13/2017 20:55:05

Diagnostics Check		OK	20:48
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Subject Sample	# 1	0.125	20:50
Breath Volume		2.046	20:50
Air Blank		0.000	20:50
Air Blank		0.000	20:52
Subject Sample	# 2	0.112	20:53
Breath Volume		1.597	20:53
Air Blank		0.000	20:54
Control Test		0.081	20:54
Air Blank		0.000	20:54
Diagnostics Check		OK	20:55

Data Download Date/Time: 01/11/2018 20:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: BEJERANO

DATE/TIME OF TEST: 12/22/2017 04:44:35

Diagnostics Check		OK	04:35	
Air Blank		0.000	04:35	
Control Test		0.080	04:35	
Air Blank		0.000	04:36	
Subject Sample	# 1		04:39	No Sample Provided
Breath Volume		0.000	04:39	
Air Blank		0.000	04:40	
Air Blank		0.000	04:42	
Subject Sample	# 2		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:43	
Control Test		0.080	04:43	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 01/11/2018 20:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: KRANZLER

DATE/TIME OF TEST: 12/23/2017 03:33:48

Diagnostics Check		OK	03:27
Air Blank		0.000	03:27
Control Test		0.079	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.200	03:29
Breath Volume		3.808	03:29
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.195	03:32
Breath Volume		3.855	03:32
Air Blank		0.000	03:32
Control Test		0.079	03:33
Air Blank		0.000	03:33
Diagnostics Check		OK	03:33

Data Download Date/Time: 01/11/2018 20:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:16

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 12/29/2017 02:42:58

Diagnostics Check		OK	02:36
Air Blank		0.000	02:36
Control Test		0.081	02:37
Air Blank		0.000	02:37
Subject Sample	# 1	0.082	02:38
Breath Volume		4.097	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:40
Subject Sample	# 2	0.082	02:41
Breath Volume		2.863	02:41
Air Blank		0.000	02:42
Control Test		0.081	02:42
Air Blank		0.000	02:42
Diagnostics Check		OK	02:42

Data Download Date/Time: 01/11/2018 20:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006384

INSTRUMENT REGISTERED TO: GAINESVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STEMEN,AARON,M
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: STEMEN,AARON
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: MCMANUS

DATE/TIME OF TEST: 12/30/2017 04:06:15

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.128	04:01
Breath Volume		3.570	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.132	04:04
Breath Volume		4.015	04:04
Air Blank		0.000	04:05
Control Test		0.081	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 01/11/2018 20:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: MCKITTRICK

DATE/TIME OF TEST: 12/19/2017 22:58:11

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:52
Air Blank		0.000	22:52
Subject Sample	# 1	0.136	22:53
Breath Volume		3.597	22:53
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.142	22:56
Breath Volume		4.207	22:56
Air Blank		0.000	22:57
Control Test		0.080	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: ORENDAIN

DATE/TIME OF TEST: 12/20/2017 02:06:19

Diagnostics Check		OK	01:59
Air Blank		0.000	02:00
Control Test		0.081	02:00
Air Blank		0.000	02:01
Subject Sample	# 1	0.158	02:01
Breath Volume		2.843	02:01
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.162	02:04
Breath Volume		3.718	02:04
Air Blank		0.000	02:05
Control Test		0.081	02:05
Air Blank		0.000	02:06
Diagnostics Check		OK	02:06

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: WESSLING

DATE/TIME OF TEST: 12/20/2017 22:35:01

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.080	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.150	22:30
Breath Volume		3.359	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:32
Subject Sample	# 2	0.130	22:33
Breath Volume		1.722	22:33
Air Blank		0.000	22:34
Control Test		0.079	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: DOSHIER

DATE/TIME OF TEST: 12/21/2017 01:07:35

Diagnostics Check		OK	00:57
Air Blank		0.000	00:58
Control Test		0.080	00:58
Air Blank		0.000	00:59
Subject Sample	# 1	0.228	01:01
Breath Volume		1.207	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:04
Subject Sample	# 2	0.225	01:05
Breath Volume		1.199	01:05
Air Blank		0.000	01:06
Control Test		0.078	01:06
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BARATY

DATE/TIME OF TEST: 12/21/2017 02:17:04

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.173	02:09	Volume Not Met
Breath Volume		0.621	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.149	02:15	Volume Not Met
Breath Volume		0.105	02:15	
Air Blank		0.000	02:16	
Control Test		0.078	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:13

SUBJECT'S LAST NAME: MOELLER

DATE/TIME OF TEST: 12/22/2017 04:40:49

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.272	04:35
Breath Volume		2.476	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.261	04:39
Breath Volume		2.351	04:39
Air Blank		0.000	04:39
Control Test		0.081	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R 48501
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:43

SUBJECT'S LAST NAME: CASTANEDA

DATE/TIME OF TEST: 12/22/2017 07:12:09

Diagnostics Check		OK	07:03	
Air Blank		0.000	07:04	
Control Test		0.081	07:04	
Air Blank		0.000	07:04	
Subject Sample	# 1	0.164	07:08	Volume Not Met
Breath Volume		0.187	07:08	
Air Blank		0.000	07:08	
Air Blank		0.000	07:10	
Subject Sample	# 2		07:10	Subject Test Refused
Breath Volume		0.000	07:10	
Air Blank		0.000	07:11	
Control Test		0.079	07:11	
Air Blank		0.000	07:12	
Diagnostics Check		OK	07:12	

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MONTOUTE,S 38228
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:12

SUBJECT'S LAST NAME: FLEMING

DATE/TIME OF TEST: 12/22/2017 08:40:09

Diagnostics Check		OK	08:32
Air Blank		0.000	08:33
Control Test		0.080	08:33
Air Blank		0.000	08:34
Subject Sample	# 1	0.133	08:35
Breath Volume		2.167	08:35
Air Blank		0.000	08:35
Air Blank		0.000	08:37
Subject Sample	# 2	0.128	08:38
Breath Volume		1.449	08:38
Air Blank		0.000	08:39
Control Test		0.080	08:39
Air Blank		0.000	08:40
Diagnostics Check		OK	08:40

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNIDER,P
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HOPE

DATE/TIME OF TEST: 12/23/2017 04:34:48

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.080	04:26	
Air Blank		0.000	04:27	
Subject Sample	# 1	0.153	04:30	Volume Not Met
Breath Volume		0.363	04:30	
Air Blank		0.000	04:31	
Air Blank		0.000	04:32	
Subject Sample	# 2		04:33	Subject Test Refused
Breath Volume		0.000	04:33	
Air Blank		0.000	04:33	
Control Test		0.079	04:34	
Air Blank		0.000	04:34	
Diagnostics Check		OK	04:34	

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MACIAS,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:26

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/23/2017 05:00:32

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:48	
Control Test		0.080	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 1	0.190	04:52	Slope Not Level
Breath Volume		1.429	04:52	
Air Blank		0.000	04:53	
Air Blank		0.000	04:54	
Subject Sample	# 2	0.193	04:55	
Breath Volume		1.582	04:55	
Air Blank		0.000	04:56	
Air Blank		0.000	04:58	
Subject Sample	# 3	0.190	04:58	
Breath Volume		1.328	04:58	
Air Blank		0.000	04:59	
Control Test		0.079	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARR,BARBARA,H
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HILL,B
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: CORDERO SALINA

DATE/TIME OF TEST: 12/23/2017 12:42:02

Diagnostics Check		OK	12:36
Air Blank		0.000	12:36
Control Test		0.080	12:36
Air Blank		0.000	12:37
Subject Sample	# 1	0.000	12:37
Breath Volume		2.859	12:37
Air Blank		0.000	12:38
Air Blank		0.000	12:40
Subject Sample	# 2	0.000	12:40
Breath Volume		3.144	12:40
Air Blank		0.000	12:41
Control Test		0.081	12:41
Air Blank		0.000	12:41
Diagnostics Check		OK	12:42

Data Download Date/Time: 01/03/2018 19:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: PELILLO

DATE/TIME OF TEST: 12/24/2017 02:04:00

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.194	01:59
Breath Volume		2.343	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:01
Subject Sample	# 2	0.188	02:02
Breath Volume		2.886	02:02
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:04

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: GOMEZ DIAZ

DATE/TIME OF TEST: 12/24/2017 20:41:33

Diagnostics Check		OK	20:34
Air Blank		0.000	20:35
Control Test		0.081	20:35
Air Blank		0.000	20:35
Subject Sample	# 1	0.130	20:36
Breath Volume		1.941	20:36
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.137	20:39
Breath Volume		2.761	20:39
Air Blank		0.000	20:40
Control Test		0.080	20:40
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: ROOSA

DATE/TIME OF TEST: 12/25/2017 23:58:59

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:53
Subject Sample	# 1	0.144	23:54
Breath Volume		2.828	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:56
Subject Sample	# 2	0.132	23:57
Breath Volume		1.273	23:57
Air Blank		0.000	23:58
Control Test		0.080	23:58
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: PETERSON JR.

DATE/TIME OF TEST: 12/27/2017 02:20:16

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.329	02:14
Breath Volume		1.265	02:14
Air Blank		0.000	02:15
Air Blank		0.000	02:16
Subject Sample	# 2	0.327	02:18
Breath Volume		1.617	02:18
Air Blank		0.000	02:19
Control Test		0.079	02:19
Air Blank		0.000	02:20
Diagnostics Check		OK	02:20

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:29

SUBJECT'S LAST NAME: KARUP

DATE/TIME OF TEST: 12/28/2017 03:56:37

Diagnostics Check		OK	03:49
Air Blank		0.000	03:49
Control Test		0.081	03:50
Air Blank		0.000	03:50
Subject Sample	# 1	0.196	03:51
Breath Volume		3.167	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.182	03:54
Breath Volume		3.003	03:54
Air Blank		0.000	03:55
Control Test		0.080	03:56
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 12/28/2017 23:35:54

Diagnostics Check		OK	23:29
Air Blank		0.000	23:29
Control Test		0.081	23:30
Air Blank		0.000	23:30
Subject Sample	# 1	0.000	23:31
Breath Volume		1.457	23:31
Air Blank		0.000	23:31
Air Blank		0.000	23:33
Subject Sample	# 2	0.000	23:34
Breath Volume		1.671	23:34
Air Blank		0.000	23:34
Control Test		0.082	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:35

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: BLESSING

DATE/TIME OF TEST: 12/29/2017 00:10:05

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.081	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.159	00:05
Breath Volume		2.675	00:05
Air Blank		0.000	00:06
Air Blank		0.000	00:08
Subject Sample	# 2	0.155	00:08
Breath Volume		1.730	00:08
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Diagnostics Check		OK	00:10

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: VALLEJO

DATE/TIME OF TEST: 12/29/2017 04:32:42

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.081	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.177	04:27
Breath Volume		1.257	04:27
Air Blank		0.000	04:28
Air Blank		0.000	04:30
Subject Sample	# 2	0.168	04:31
Breath Volume		1.113	04:31
Air Blank		0.000	04:31
Control Test		0.080	04:32
Air Blank		0.000	04:32
Diagnostics Check		OK	04:32

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DE JESUS,V
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: SINHOLD

DATE/TIME OF TEST: 12/29/2017 06:24:10

Diagnostics Check		OK	06:17
Air Blank		0.000	06:18
Control Test		0.081	06:18
Air Blank		0.000	06:19
Subject Sample	# 1	0.203	06:19
Breath Volume		2.246	06:19
Air Blank		0.000	06:20
Air Blank		0.000	06:22
Subject Sample	# 2	0.190	06:22
Breath Volume		2.484	06:22
Air Blank		0.000	06:23
Control Test		0.081	06:23
Air Blank		0.000	06:24
Diagnostics Check		OK	06:24

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PETRIE

DATE/TIME OF TEST: 12/30/2017 02:17:05

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.082	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.170	02:09	Volume Not Met
Breath Volume		0.050	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2	0.132	02:15	Volume Not Met
Breath Volume		1.015	02:15	
Air Blank		0.000	02:16	
Control Test		0.078	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: PETRIE

DATE/TIME OF TEST: 12/30/2017 02:33:24

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:21	
Control Test		0.079	02:21	
Air Blank		0.000	02:22	
Subject Sample	# 1	0.119	02:25	Volume Not Met
Breath Volume		0.296	02:25	
Air Blank		0.000	02:26	
Air Blank		0.000	02:27	
Subject Sample	# 2	0.151	02:28	
Breath Volume		1.550	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 3	0.159	02:31	
Breath Volume		1.273	02:31	
Air Blank		0.000	02:32	
Control Test		0.080	02:32	
Air Blank		0.000	02:33	
Diagnostics Check		OK	02:33	

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: HERNANDEZ-LOPEZ

DATE/TIME OF TEST: 12/30/2017 21:42:13

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.082	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.196	21:37
Breath Volume		2.664	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:39
Subject Sample	# 2	0.189	21:40
Breath Volume		2.210	21:40
Air Blank		0.000	21:41
Control Test		0.081	21:41
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006385

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: PEREZ-OSORIO

DATE/TIME OF TEST: 12/31/2017 22:05:54

Diagnostics Check		OK	21:58
Air Blank		0.000	21:59
Control Test		0.081	21:59
Air Blank		0.000	22:00
Subject Sample	# 1	0.253	22:01
Breath Volume		2.792	22:01
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.257	22:04
Breath Volume		3.156	22:04
Air Blank		0.000	22:04
Control Test		0.081	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 01/03/2018 19:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAIN,KYLE,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MCLAIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 21:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: PADRON

DATE/TIME OF TEST: 12/14/2017 22:40:46

Diagnostics Check		OK	22:37	
Air Blank		0.000	22:38	
Control Test		0.080	22:38	
Air Blank		0.000	22:39	
Subject Sample	# 1		22:39	Subject Test Refused
Breath Volume		0.000	22:39	
Air Blank		0.000	22:39	
Control Test		0.081	22:40	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 12/26/2017 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMERON,DAVID,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: CORONILLA,DANIEL
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:28

SUBJECT'S LAST NAME: FORD

DATE/TIME OF TEST: 12/19/2017 00:05:03

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Subject Sample	# 1	0.107	00:00
Breath Volume		3.187	00:00
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 2	0.100	00:03
Breath Volume		2.128	00:03
Air Blank		0.000	00:04
Control Test		0.079	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:05

Data Download Date/Time: 12/26/2017 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006386

INSTRUMENT REGISTERED TO: HENDRY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCLAIN,KYLE,J
BREATH TEST OPERATOR AGENCY: Hendry County SO
ARRESTING OFFICER: MCLAIN,KYLE
ARRESTING OFFICER AGENCY: Hendry County SO
LAST AGENCY INSPECTION: 11/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 05115080A5
DRY GAS EXPIRY DATE: 03/05/2017
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: WELLS

DATE/TIME OF TEST: 12/26/2017 00:04:38

Diagnostics Check		OK	23:57
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.187	23:59
Breath Volume		2.476	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.202	00:02
Breath Volume		1.984	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:04
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 12/26/2017 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILLOW,JEREMY,T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PILLOW,JEREMY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 12/02/2017 20:23:02

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:16	
Control Test		0.078	20:16	
Air Blank		0.000	20:16	
Subject Sample	# 1	0.161	20:18	
Breath Volume		2.316	20:18	
Air Blank		0.000	20:19	
Air Blank		0.000	20:20	
Subject Sample	# 2	0.155	20:21	
Breath Volume		1.796	20:21	
Air Blank		0.000	20:22	
Control Test		0.074	20:22	Control Outside Tolerance
Air Blank		0.000	20:23	

Data Download Date/Time: 12/12/2017 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PILLOW,JEREMY,T
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: PILLOW,JEREMY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: TERRELL

DATE/TIME OF TEST: 12/02/2017 20:32:10

Diagnostics Check		OK	20:25
Air Blank		0.000	20:26
Control Test		0.078	20:26
Air Blank		0.000	20:27
Subject Sample	# 1	0.166	20:27
Breath Volume		2.375	20:27
Air Blank		0.000	20:28
Air Blank		0.000	20:30
Subject Sample	# 2	0.160	20:30
Breath Volume		2.050	20:30
Air Blank		0.000	20:31
Control Test		0.078	20:31
Air Blank		0.000	20:32
Diagnostics Check		OK	20:32

Data Download Date/Time: 12/12/2017 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNOWLES,CONSTANCE,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KNOWLES,CONSTANCE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:04

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/06/2017 15:31:59

Diagnostics Check		OK	15:25
Air Blank		0.000	15:26
Control Test		0.079	15:26
Air Blank		0.000	15:26
Subject Sample	# 1	0.171	15:27
Breath Volume		1.683	15:27
Air Blank		0.000	15:28
Air Blank		0.000	15:29
Subject Sample	# 2	0.170	15:30
Breath Volume		1.406	15:30
Air Blank		0.000	15:31
Control Test		0.078	15:31
Air Blank		0.000	15:31
Diagnostics Check		OK	15:31

Data Download Date/Time: 12/12/2017 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARACLE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BUZBEE,HEATHER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:10

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/09/2017 06:37:31

Diagnostics Check		OK	06:30
Air Blank		0.000	06:30
Control Test		0.079	06:30
Air Blank		0.000	06:31
Subject Sample	# 1	0.222	06:32
Breath Volume		3.640	06:32
Air Blank		0.000	06:33
Air Blank		0.000	06:34
Subject Sample	# 2	0.215	06:35
Breath Volume		3.484	06:35
Air Blank		0.000	06:36
Control Test		0.079	06:36
Air Blank		0.000	06:37
Diagnostics Check		OK	06:37

Data Download Date/Time: 12/12/2017 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINH,TAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DINH,TAM
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: MOLTER

DATE/TIME OF TEST: 12/09/2017 23:34:00

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.080	23:28
Air Blank		0.000	23:28
Subject Sample	# 1	0.188	23:29
Breath Volume		3.796	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:31
Subject Sample	# 2	0.195	23:32
Breath Volume		3.570	23:32
Air Blank		0.000	23:33
Control Test		0.079	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:34

Data Download Date/Time: 12/12/2017 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,BRANDT,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KESSINGER,B
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/11/2017 04:00:48

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:58	
Control Test		0.079	03:58	
Air Blank		0.000	03:59	
Subject Sample	# 1		03:59	RFI Detect
Breath Volume		0.000	03:59	
Air Blank		0.000	03:59	
Control Test		0.080	04:00	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 12/12/2017 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,BRANDT,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KESSINGER,B
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/11/2017 04:04:52

Diagnostics Check		OK	04:02	
Air Blank		0.000	04:03	
Control Test		0.079	04:03	
Air Blank		0.000	04:03	
Subject Sample	# 1		04:04	Sequence Aborted
Breath Volume		0.000	04:04	
Air Blank		0.000	04:04	

Data Download Date/Time: 12/12/2017 09:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KESSINGER,BRANDT,J
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: KESSINGER,B
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 12/11/2017 04:13:44

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.151	04:08
Breath Volume		2.183	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.154	04:12
Breath Volume		3.445	04:12
Air Blank		0.000	04:12
Control Test		0.075	04:13
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 12/12/2017 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VO,SEBASTIAN,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: VO,SEBASTIAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:57

SUBJECT'S LAST NAME: RICKMAN

DATE/TIME OF TEST: 12/12/2017 06:42:04

Diagnostics Check		OK	06:30	
Air Blank		0.000	06:31	
Control Test		0.080	06:31	
Air Blank		0.000	06:31	
Subject Sample	# 1		06:35	No Sample Provided
Breath Volume		0.000	06:35	
Air Blank		0.000	06:35	
Air Blank		0.000	06:37	
Subject Sample	# 2		06:40	No Sample Provided
Breath Volume		0.000	06:40	
Air Blank		0.000	06:41	
Control Test		0.079	06:41	
Air Blank		0.000	06:41	
Diagnostics Check		OK	06:42	

Data Download Date/Time: 12/12/2017 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: COWART

DATE/TIME OF TEST: 12/13/2017 22:35:06

Diagnostics Check		OK	22:28
Air Blank		0.000	22:28
Control Test		0.079	22:29
Air Blank		0.000	22:29
Subject Sample	# 1	0.124	22:30
Breath Volume		3.789	22:30
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.119	22:33
Breath Volume		3.769	22:33
Air Blank		0.000	22:34
Control Test		0.079	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHORETTE,BRIAN,E
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SHORETTE,BRIAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: OLIVER

DATE/TIME OF TEST: 12/14/2017 00:48:47

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.078	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.193	00:44
Breath Volume		2.257	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:46
Subject Sample	# 2	0.180	00:47
Breath Volume		1.933	00:47
Air Blank		0.000	00:47
Control Test		0.075	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COTE,JACOB,M
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: COTE,JACOB
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: BOCCIA

DATE/TIME OF TEST: 12/14/2017 02:47:48

Diagnostics Check		OK	02:44	
Air Blank		0.000	02:45	
Control Test		0.079	02:45	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:46	Subject Test Refused
Breath Volume		0.000	02:46	
Air Blank		0.000	02:46	
Control Test		0.079	02:47	
Air Blank		0.000	02:47	
Diagnostics Check		OK	02:47	

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: CALEY,TYLER
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: HOBBS

DATE/TIME OF TEST: 12/16/2017 03:05:46

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.080	03:00
Air Blank		0.000	03:00
Subject Sample	# 1	0.228	03:01
Breath Volume		4.246	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.218	03:04
Breath Volume		3.687	03:04
Air Blank		0.000	03:04
Control Test		0.080	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HAMPTON,LOGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/16/2017 22:42:16

Diagnostics Check		OK	22:32	
Air Blank		0.000	22:32	
Control Test		0.080	22:33	
Air Blank		0.000	22:33	
Subject Sample	# 1	0.179	22:36	Volume Not Met
Breath Volume		0.585	22:36	
Air Blank		0.000	22:37	
Air Blank		0.000	22:39	
Subject Sample	# 2		22:40	Subject Test Refused
Breath Volume		0.484	22:40	
Air Blank		0.000	22:41	
Control Test		0.078	22:41	
Air Blank		0.000	22:42	
Diagnostics Check		OK	22:42	

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: SMITH,MORGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: LIVINGSTON

DATE/TIME OF TEST: 12/17/2017 03:11:47

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:09	
Control Test		0.080	03:09	
Air Blank		0.000	03:10	
Subject Sample	# 1		03:10	Subject Test Refused
Breath Volume		0.000	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:11	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENLEY,WILLIAM,C
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HENLEY,WILLIAM
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:29

SUBJECT'S LAST NAME: LABERGE

DATE/TIME OF TEST: 12/18/2017 21:56:11

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.079	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.231	21:51
Breath Volume		3.945	21:51
Air Blank		0.000	21:52
Air Blank		0.000	21:53
Subject Sample	# 2	0.231	21:54
Breath Volume		3.269	21:54
Air Blank		0.000	21:55
Control Test		0.078	21:55
Air Blank		0.000	21:56
Diagnostics Check		OK	21:56

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SHELLEY,SAMUEL,L
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: HAMPTON,LOGAN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 12/23/2017 22:55:30

Diagnostics Check		OK	22:48
Air Blank		0.000	22:48
Control Test		0.078	22:49
Air Blank		0.000	22:49
Subject Sample	# 1	0.202	22:50
Breath Volume		4.078	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.201	22:53
Breath Volume		4.355	22:53
Air Blank		0.000	22:54
Control Test		0.078	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: RICHARDSON,LARRY
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:28

SUBJECT'S LAST NAME: STALLWORTH

DATE/TIME OF TEST: 12/24/2017 21:57:13

Diagnostics Check		OK	21:54	
Air Blank		0.000	21:54	
Control Test		0.078	21:54	
Air Blank		0.000	21:55	
Subject Sample	# 1		21:55	Subject Test Refused
Breath Volume		0.000	21:55	
Air Blank		0.000	21:56	
Control Test		0.079	21:56	
Air Blank		0.000	21:57	
Diagnostics Check		OK	21:57	

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEITE,STEVEN,G
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: LEITE,STEVEN
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:58

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/25/2017 00:26:35

Diagnostics Check		OK	00:18
Air Blank		0.000	00:19
Control Test		0.079	00:19
Air Blank		0.000	00:20
Subject Sample	# 1	0.177	00:21
Breath Volume		1.796	00:21
Air Blank		0.000	00:22
Air Blank		0.000	00:23
Subject Sample	# 2	0.169	00:24
Breath Volume		1.230	00:24
Air Blank		0.000	00:25
Control Test		0.075	00:25
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: HARRIS,ZACHARY
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 12/27/2017 02:47:48

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.079	02:41
Air Blank		0.000	02:41
Subject Sample	# 1	0.109	02:42
Breath Volume		1.304	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.120	02:46
Breath Volume		1.300	02:46
Air Blank		0.000	02:46
Control Test		0.079	02:47
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006436

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,AMANDA,L
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: DAVIS,AMANDA
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: ROMERO MARQUEZ

DATE/TIME OF TEST: 12/31/2017 05:13:26

Diagnostics Check		OK	05:06
Air Blank		0.000	05:06
Control Test		0.080	05:07
Air Blank		0.000	05:07
Subject Sample	# 1	0.153	05:08
Breath Volume		3.000	05:08
Air Blank		0.000	05:09
Air Blank		0.000	05:11
Subject Sample	# 2	0.158	05:11
Breath Volume		3.308	05:11
Air Blank		0.000	05:12
Control Test		0.077	05:12
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 01/31/2018 11:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEYER,THOMAS,S
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MEYER,THOMAS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/01/2017 21:19:54

Diagnostics Check		OK	21:16	
Air Blank		0.000	21:17	
Control Test		0.079	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 1		21:18	Subject Test Refused
Breath Volume		0.000	21:18	
Air Blank		0.000	21:18	
Control Test		0.080	21:19	
Air Blank		0.000	21:19	
Diagnostics Check		OK	21:19	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REYES,A,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GOMEZ,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:02

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/03/2017 03:40:16

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.079	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.210	03:35
Breath Volume		1.644	03:35
Air Blank		0.000	03:36
Air Blank		0.000	03:37
Subject Sample	# 2	0.207	03:38
Breath Volume		1.960	03:38
Air Blank		0.000	03:39
Control Test		0.078	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALBERTO,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: VEGA,KAREENA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:47

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 12/03/2017 16:20:51

Diagnostics Check		OK	16:07	
Air Blank		0.000	16:07	
Control Test		0.079	16:08	
Air Blank		0.000	16:08	
Subject Sample	# 1	0.244	16:11	Volume Not Met
Breath Volume		1.042	16:11	
Air Blank		0.000	16:12	
Air Blank		0.000	16:14	
Subject Sample	# 2	0.254	16:15	
Breath Volume		1.519	16:15	
Air Blank		0.000	16:16	
Air Blank		0.000	16:18	
Subject Sample	# 3	0.253	16:19	
Breath Volume		1.546	16:19	
Air Blank		0.000	16:19	
Control Test		0.075	16:20	
Air Blank		0.000	16:20	
Diagnostics Check		OK	16:20	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMINGTON,EARL,J
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: SIMINGTON,EARL
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: BROWN PHILLIPS

DATE/TIME OF TEST: 12/05/2017 03:47:41

Diagnostics Check		OK	03:43	
Air Blank		0.000	03:43	
Control Test		0.080	03:44	
Air Blank		0.000	03:44	
Subject Sample	# 1		03:46	Subject Test Refused
Breath Volume		0.000	03:46	
Air Blank		0.000	03:46	
Control Test		0.080	03:47	
Air Blank		0.000	03:47	
Diagnostics Check		OK	03:47	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLANCO,JUAN,S
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: BLANCO,JUAN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 22:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LEON BOLIVAR

DATE/TIME OF TEST: 12/05/2017 23:27:22

Diagnostics Check		OK	23:24	
Air Blank		0.000	23:24	
Control Test		0.079	23:25	
Air Blank		0.000	23:25	
Subject Sample	# 1		23:25	Subject Test Refused
Breath Volume		0.000	23:25	
Air Blank		0.000	23:26	
Control Test		0.079	23:26	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:27	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ESTEVEZ,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:15

SUBJECT'S LAST NAME: NAVARRO

DATE/TIME OF TEST: 12/07/2017 04:13:08

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:07	No Sample Provided
Breath Volume		0.000	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2		04:11	Subject Test Refused
Breath Volume		0.000	04:11	
Air Blank		0.000	04:12	
Control Test		0.078	04:12	
Air Blank		0.000	04:13	
Diagnostics Check		OK	04:13	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLINGER,TI'ANDRE,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BELLINGER,TI'ANDRE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: DE JESUS

DATE/TIME OF TEST: 12/08/2017 00:55:09

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:52	
Control Test		0.079	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:53	Subject Test Refused
Breath Volume		0.000	00:53	
Air Blank		0.000	00:54	
Control Test		0.079	00:54	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,GERARDO,L
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: QUINONES,GERARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 06:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:02

SUBJECT'S LAST NAME: PERALTA AYALA

DATE/TIME OF TEST: 12/08/2017 07:34:56

Diagnostics Check		OK	07:26
Air Blank		0.000	07:27
Control Test		0.079	07:27
Air Blank		0.000	07:28
Subject Sample	# 1	0.235	07:30
Breath Volume		1.808	07:30
Air Blank		0.000	07:31
Air Blank		0.000	07:32
Subject Sample	# 2	0.227	07:33
Breath Volume		1.707	07:33
Air Blank		0.000	07:34
Control Test		0.076	07:34
Air Blank		0.000	07:34
Diagnostics Check		OK	07:34

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LLOVET,GERARDO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: LLOVET,GERARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: ROSSI

DATE/TIME OF TEST: 12/08/2017 23:06:22

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:03	
Control Test		0.080	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1		23:05	Subject Test Refused
Breath Volume		0.000	23:05	
Air Blank		0.000	23:05	
Control Test		0.079	23:05	
Air Blank		0.000	23:06	
Diagnostics Check		OK	23:06	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,SANDRA,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GARCIA,SANDRA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: GANDOLFI

DATE/TIME OF TEST: 12/09/2017 03:13:36

Diagnostics Check		OK	03:03
Air Blank		0.000	03:04
Control Test		0.079	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.237	03:07
Breath Volume		1.507	03:07
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.230	03:11
Breath Volume		1.464	03:11
Air Blank		0.000	03:12
Control Test		0.078	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARRERO,MARCOS,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: MARRERO,MARCOS
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:15

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/09/2017 10:01:42

Diagnostics Check		OK	09:56	
Air Blank		0.000	09:57	
Control Test		0.080	09:57	
Air Blank		0.000	09:58	
Subject Sample	# 1		10:00	Subject Test Refused
Breath Volume		0.000	10:00	
Air Blank		0.000	10:00	
Control Test		0.080	10:01	
Air Blank		0.000	10:01	
Diagnostics Check		OK	10:01	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOPEZ,M,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: BETANCOURT,Y
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: BERMUDEZ

DATE/TIME OF TEST: 12/10/2017 03:39:20

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.080	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.194	03:34
Breath Volume		1.281	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.197	03:37
Breath Volume		1.898	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GOMEZ,JOSEPH
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:45

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 12/10/2017 08:23:00

Diagnostics Check		OK	08:10	
Air Blank		0.000	08:11	
Control Test		0.080	08:11	
Air Blank		0.000	08:12	
Subject Sample	# 1	0.200	08:15	Volume Not Met
Breath Volume		0.574	08:15	
Air Blank		0.000	08:15	
Air Blank		0.000	08:17	
Subject Sample	# 2	0.193	08:18	
Breath Volume		2.675	08:18	
Air Blank		0.000	08:19	
Air Blank		0.000	08:20	
Subject Sample	# 3	0.197	08:21	
Breath Volume		1.273	08:21	
Air Blank		0.000	08:22	
Control Test		0.079	08:22	
Air Blank		0.000	08:22	
Diagnostics Check		OK	08:23	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ROBERTO,
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: MCKINLEY,STEVEN
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: BUI
ARREST TIME: 22:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: MORANTES

DATE/TIME OF TEST: 12/10/2017 23:26:51

Diagnostics Check		OK	23:23	
Air Blank		0.000	23:23	
Control Test		0.080	23:24	
Air Blank		0.000	23:24	
Subject Sample	# 1		23:25	Subject Test Refused
Breath Volume		0.000	23:25	
Air Blank		0.000	23:25	
Control Test		0.080	23:26	
Air Blank		0.000	23:26	
Diagnostics Check		OK	23:26	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENA,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HARMON,FABIANA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: GONZALEZ LOPEZ

DATE/TIME OF TEST: 12/11/2017 03:41:52

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.078	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.190	03:37
Breath Volume		1.972	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.189	03:40
Breath Volume		3.644	03:40
Air Blank		0.000	03:40
Control Test		0.080	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,K,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GONZALEZ,K
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:53

SUBJECT'S LAST NAME: JIMOH

DATE/TIME OF TEST: 12/11/2017 08:18:57

Diagnostics Check		OK	08:14	
Air Blank		0.000	08:14	
Control Test		0.080	08:15	
Air Blank		0.000	08:15	
Subject Sample	# 1		08:17	Subject Test Refused
Breath Volume		0.000	08:17	
Air Blank		0.000	08:18	
Control Test		0.080	08:18	
Air Blank		0.000	08:18	
Diagnostics Check		OK	08:18	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GOMEZ,JOSEPH,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: GOMEZ,JOSEPH
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: LEIVS

DATE/TIME OF TEST: 12/16/2017 07:37:07

Diagnostics Check		OK	07:29
Air Blank		0.000	07:29
Control Test		0.080	07:29
Air Blank		0.000	07:30
Subject Sample	# 1	0.194	07:32
Breath Volume		3.007	07:32
Air Blank		0.000	07:32
Air Blank		0.000	07:34
Subject Sample	# 2	0.189	07:35
Breath Volume		2.128	07:35
Air Blank		0.000	07:36
Control Test		0.079	07:36
Air Blank		0.000	07:37
Diagnostics Check		OK	07:37

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANCHEZ,AQUILES,M
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: HERRERA,ENRIQUE
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: BLANDON

DATE/TIME OF TEST: 12/17/2017 04:33:58

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.080	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.235	04:29
Breath Volume		2.535	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.218	04:32
Breath Volume		1.183	04:32
Air Blank		0.000	04:33
Control Test		0.078	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLANAS,A,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PLANAS,A
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/18/2017 00:30:42

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:22	
Control Test		0.080	00:22	
Air Blank		0.000	00:23	
Subject Sample	# 1		00:26	No Sample Provided
Breath Volume		0.000	00:26	
Air Blank		0.000	00:26	
Air Blank		0.000	00:28	
Subject Sample	# 2		00:29	Subject Test Refused
Breath Volume		0.000	00:29	
Air Blank		0.000	00:29	
Control Test		0.080	00:30	
Air Blank		0.000	00:30	
Diagnostics Check		OK	00:30	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ESTEVEZ,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: PIERRE

DATE/TIME OF TEST: 12/20/2017 04:06:19

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:04	
Air Blank		0.000	04:04	
Subject Sample	# 1		04:04	Subject Test Refused
Breath Volume		0.000	04:04	
Air Blank		0.000	04:05	
Control Test		0.080	04:05	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,DALIA,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ECHEVARRIA,DALIA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 10:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:45

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/23/2017 11:16:13

Diagnostics Check		OK	11:10	
Air Blank		0.000	11:11	
Control Test		0.081	11:11	
Air Blank		0.000	11:12	
Subject Sample	# 1		11:14	Subject Test Refused
Breath Volume		0.000	11:14	
Air Blank		0.000	11:15	
Control Test		0.081	11:15	
Air Blank		0.000	11:16	
Diagnostics Check		OK	11:16	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARKER,DEONTAY,T
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: PARKER,DEONTAY
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 12/24/2017 02:48:06

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:45	
Control Test		0.080	02:45	
Air Blank		0.000	02:46	
Subject Sample	# 1		02:46	Subject Test Refused
Breath Volume		0.000	02:46	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCANTARA,PATRICIA,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OLMEDO,REBECCA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:01

SUBJECT'S LAST NAME: ALCANTARA

DATE/TIME OF TEST: 12/26/2017 08:23:19

Diagnostics Check	OK	08:21	
Air Blank	0.000	08:22	
Control Test	0.081	08:22	
Air Blank		08:22	Improper Sample
Air Blank	0.000	08:23	

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALCANTARA,PATRICIA,E
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: OLMEDO,REBECCA
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 07:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: ALCANTARA

DATE/TIME OF TEST: 12/26/2017 08:32:54

Diagnostics Check		OK	08:26
Air Blank		0.000	08:27
Control Test		0.080	08:27
Air Blank		0.000	08:27
Subject Sample	# 1	0.141	08:28
Breath Volume		1.128	08:28
Air Blank		0.000	08:29
Air Blank		0.000	08:31
Subject Sample	# 2	0.155	08:31
Breath Volume		2.636	08:31
Air Blank		0.000	08:31
Control Test		0.080	08:32
Air Blank		0.000	08:32
Diagnostics Check		OK	08:32

Data Download Date/Time: 12/28/2017 15:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DIAZ,LEONARDO,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: DIAZ,LEONARDO
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: PAGES

DATE/TIME OF TEST: 12/26/2017 21:22:24

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.080	21:12	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:15	No Sample Provided
Breath Volume		0.000	21:15	
Air Blank		0.000	21:16	
Air Blank		0.000	21:18	
Subject Sample	# 2		21:20	Subject Test Refused
Breath Volume		0.000	21:20	
Air Blank		0.000	21:21	
Control Test		0.080	21:21	
Air Blank			21:21	RFI Detect
Air Blank		0.000	21:22	

Data Download Date/Time: 12/28/2017 15:52

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006456

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VASQUEZ,JASON,A
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: RIOS,RICHARD
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: BAYONA

DATE/TIME OF TEST: 12/28/2017 20:18:27

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:16	
Control Test		0.080	20:16	
Air Blank		0.000	20:16	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:17	
Control Test		0.081	20:17	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:18	

Data Download Date/Time: 01/30/2018 13:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006457

INSTRUMENT REGISTERED TO: MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESTEVEZ,D,
BREATH TEST OPERATOR AGENCY: Miami PD
ARRESTING OFFICER: ESTEVEZ,D
ARRESTING OFFICER AGENCY: Miami PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 05:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:35

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/01/2017 06:03:27

Diagnostics Check		OK	05:55	
Air Blank		0.000	05:56	
Control Test		0.078	05:56	
Air Blank		0.000	05:56	
Subject Sample	# 1	0.080	05:58	
Breath Volume		1.855	05:58	
Air Blank		0.000	05:59	
Air Blank		0.000	06:01	
Subject Sample	# 2	0.070	06:01	
Breath Volume		1.859	06:01	
Air Blank		0.000	06:02	
Control Test		0.000	06:02	Control Outside Tolerance
Air Blank		0.000	06:03	

Data Download Date/Time: 12/11/2017 10:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DURHAM,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: DURHAM,JAMES
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 12/15/2017 00:03:33

Diagnostics Check		OK	23:59	
Air Blank		0.000	23:59	
Control Test		0.080	23:59	
Air Blank		0.000	00:00	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:02	
Control Test		0.081	00:02	
Air Blank		0.000	00:03	
Diagnostics Check		OK	00:03	

Data Download Date/Time: 01/31/2018 14:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYERS,BARBI,J
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: MYERS,BARBI
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 17:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:03

SUBJECT'S LAST NAME: COOKE

DATE/TIME OF TEST: 12/19/2017 19:35:09

Diagnostics Check		OK	19:25
Air Blank		0.000	19:25
Control Test		0.080	19:26
Air Blank		0.000	19:26
Subject Sample	# 1	0.269	19:27
Breath Volume		1.773	19:27
Air Blank		0.000	19:28
Air Blank		0.000	19:29
Subject Sample	# 2	0.299	19:30
Breath Volume		2.195	19:30
Air Blank		0.000	19:31
Air Blank		0.000	19:32
Subject Sample	# 3	0.275	19:33
Breath Volume		2.273	19:33
Air Blank		0.000	19:34
Control Test		0.078	19:34
Air Blank		0.000	19:35
Diagnostics Check		OK	19:35

Data Download Date/Time: 01/31/2018 14:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006472

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,TIMOTHY,S
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BROWN,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 390314
DRY GAS EXPIRY DATE: 04/19/2019
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: STEEN

DATE/TIME OF TEST: 12/25/2017 01:04:15

Diagnostics Check		OK	00:55
Air Blank		0.000	00:55
Control Test		0.081	00:56
Air Blank		0.000	00:56
Subject Sample	# 1	0.169	00:57
Breath Volume		1.171	00:57
Air Blank		0.000	00:58
Air Blank		0.000	00:59
Subject Sample	# 2	0.174	01:02
Breath Volume		1.437	01:02
Air Blank		0.000	01:03
Control Test		0.079	01:03
Air Blank		0.000	01:04
Diagnostics Check		OK	01:04

Data Download Date/Time: 01/31/2018 14:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRY,HEATHER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: SCHRY,HEATHER
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06717080A6
DRY GAS EXPIRY DATE: 05/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: STOUPAKIS

DATE/TIME OF TEST: 12/19/2017 22:39:35

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:36	
Control Test		0.078	22:36	
Air Blank		0.000	22:37	
Subject Sample	# 1		22:38	Subject Test Refused
Breath Volume		0.000	22:38	
Air Blank		0.000	22:38	
Control Test		0.079	22:38	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 12/21/2017 10:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRY,HEATHER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: GUERRA,JULIAN
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: JACOBS

DATE/TIME OF TEST: 12/26/2017 19:54:54

Diagnostics Check		OK	19:51	
Air Blank		0.000	19:52	
Control Test		0.078	19:52	
Air Blank		0.000	19:53	
Subject Sample	# 1		19:53	Subject Test Refused
Breath Volume		0.000	19:53	
Air Blank		0.000	19:53	
Control Test		0.078	19:54	
Air Blank		0.000	19:54	
Diagnostics Check		OK	19:54	

Data Download Date/Time: 01/24/2018 10:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRY,HEATHER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/28/2017 20:09:55

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:01	
Control Test		0.078	20:01	
Air Blank		0.000	20:02	
Subject Sample	# 1	0.247	20:04	
Breath Volume		1.812	20:04	
Air Blank		0.000	20:05	
Air Blank		0.000	20:07	
Subject Sample	# 2		20:08	RFI Detect
Breath Volume		0.089	20:08	
Air Blank		0.000	20:08	
Control Test		0.077	20:09	
Air Blank		0.000	20:09	
Diagnostics Check		OK	20:09	

Data Download Date/Time: 01/24/2018 10:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRY,HEATHER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/28/2017 20:15:08

Diagnostics Check		OK	20:11	
Air Blank		0.000	20:11	
Control Test		0.077	20:11	
Air Blank		0.000	20:12	
Subject Sample	# 1		20:13	RFI Detect
Breath Volume		0.035	20:13	
Air Blank		0.000	20:14	
Control Test		0.077	20:14	
Air Blank		0.000	20:15	
Diagnostics Check		OK	20:15	

Data Download Date/Time: 01/24/2018 10:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHRY,HEATHER,
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: LANDA,CHELSY
ARRESTING OFFICER AGENCY: South Miami PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/28/2017 20:29:33

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.078	20:17	
Air Blank		0.000	20:17	
Subject Sample	# 1	0.251	20:18	
Breath Volume		1.234	20:18	
Air Blank		0.000	20:19	
Air Blank		0.000	20:20	
Subject Sample	# 2	0.242	20:24	Volume Not Met
Breath Volume		0.625	20:24	
Air Blank		0.000	20:25	
Air Blank		0.000	20:26	
Subject Sample	# 3	0.249	20:28	
Breath Volume		1.578	20:28	
Air Blank			20:28	Purge Fail
Air Blank		0.000	20:29	

Data Download Date/Time: 01/24/2018 10:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006473

INSTRUMENT REGISTERED TO: PINECREST PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,EDISON,B
BREATH TEST OPERATOR AGENCY: Village of Pinecrest
ARRESTING OFFICER: CRUZ,EDISON
ARRESTING OFFICER AGENCY: Village of Pinecrest
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TABB

DATE/TIME OF TEST: 12/30/2017 00:23:13

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.078	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1		00:21	Subject Test Refused
Breath Volume		0.000	00:21	
Air Blank		0.000	00:22	
Control Test		0.078	00:22	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:23	

Data Download Date/Time: 01/24/2018 10:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: FLICKSTEIN

DATE/TIME OF TEST: 12/03/2017 23:48:49

Diagnostics Check		OK	23:42
Air Blank		0.000	23:42
Control Test		0.078	23:43
Air Blank		0.000	23:43
Subject Sample	# 1	0.112	23:44
Breath Volume		2.074	23:44
Air Blank		0.000	23:44
Air Blank		0.000	23:46
Subject Sample	# 2	0.119	23:47
Breath Volume		2.253	23:47
Air Blank		0.000	23:47
Control Test		0.079	23:48
Air Blank		0.000	23:48
Diagnostics Check		OK	23:48

Data Download Date/Time: 12/13/2017 17:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 11/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:07

SUBJECT'S LAST NAME: PROUTY

DATE/TIME OF TEST: 12/06/2017 01:37:24

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.078	01:31
Air Blank		0.000	01:31
Subject Sample	# 1	0.000	01:32
Breath Volume		1.187	01:32
Air Blank		0.000	01:33
Air Blank		0.000	01:35
Subject Sample	# 2	0.000	01:36
Breath Volume		1.152	01:36
Air Blank		0.000	01:36
Control Test		0.078	01:36
Air Blank		0.000	01:37
Diagnostics Check		OK	01:37

Data Download Date/Time: 12/13/2017 17:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: VAIL Y VAIL

DATE/TIME OF TEST: 12/18/2017 23:54:36

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.077	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.199	23:43
Breath Volume		1.941	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:45
Subject Sample	# 2	0.171	23:47
Breath Volume		1.449	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 3	0.173	23:52
Breath Volume		1.953	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 01/10/2018 12:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: HAMILTON

DATE/TIME OF TEST: 12/19/2017 04:23:34

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.077	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.159	04:17
Breath Volume		1.753	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.140	04:21
Breath Volume		1.695	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 01/10/2018 12:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/24/2017 23:29:14

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.077	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.000	23:24
Breath Volume		2.214	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:27
Subject Sample	# 2	0.000	23:27
Breath Volume		3.097	23:27
Air Blank		0.000	23:28
Control Test		0.077	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 01/10/2018 12:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERPHOS,STEVEN,A
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SERPHOS,STEVEN
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:53

SUBJECT'S LAST NAME: IVOVIC

DATE/TIME OF TEST: 12/27/2017 03:21:39

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.184	03:16
Breath Volume		1.531	03:16
Air Blank		0.000	03:16
Air Blank		0.000	03:18
Subject Sample	# 2	0.203	03:20
Breath Volume		1.296	03:20
Air Blank		0.000	03:20
Control Test		0.077	03:21
Air Blank		0.000	03:21
Diagnostics Check		OK	03:21

Data Download Date/Time: 01/10/2018 12:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006475

INSTRUMENT REGISTERED TO: BROWARD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SAPP,JOSHUA,R
BREATH TEST OPERATOR AGENCY: Broward County SO
ARRESTING OFFICER: SAPP,JOSHUA
ARRESTING OFFICER AGENCY: Broward County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 19817080A1
DRY GAS EXPIRY DATE: 09/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 12/31/2017 02:50:01

Diagnostics Check		OK	02:41
Air Blank		0.000	02:42
Control Test		0.077	02:42
Air Blank		0.000	02:42
Subject Sample	# 1	0.145	02:45
Breath Volume		1.476	02:45
Air Blank		0.000	02:46
Air Blank		0.000	02:47
Subject Sample	# 2	0.163	02:48
Breath Volume		1.777	02:48
Air Blank		0.000	02:49
Control Test		0.076	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:50

Data Download Date/Time: 01/10/2018 12:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male

OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/01/2017 02:04:20

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.080	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:56	No Sample Provided
Breath Volume		0.000	01:56	
Air Blank		0.000	01:57	
Air Blank		0.000	01:59	
Subject Sample	# 2	0.218	01:59	
Breath Volume		2.730	01:59	
Air Blank		0.000	02:00	
Air Blank		0.000	02:02	
Subject Sample	# 3	0.230	02:02	
Breath Volume		2.890	02:02	
Air Blank		0.000	02:03	
Control Test		0.079	02:03	
Air Blank		0.000	02:04	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CLEARY,GUS
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: CLEMENTE

DATE/TIME OF TEST: 12/01/2017 21:41:25

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.079	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.277	21:36
Breath Volume		3.304	21:36
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.272	21:39
Breath Volume		3.542	21:39
Air Blank		0.000	21:40
Control Test		0.079	21:40
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JIMESON,JANAE
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: MELVIN

DATE/TIME OF TEST: 12/02/2017 04:19:11

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.080	04:11
Air Blank		0.000	04:11
Subject Sample	# 1	0.198	04:14
Breath Volume		1.335	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.196	04:17
Breath Volume		1.148	04:17
Air Blank		0.000	04:18
Control Test		0.077	04:18
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:42

SUBJECT'S LAST NAME: CHACON MALDONADO

DATE/TIME OF TEST: 12/02/2017 05:11:50

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.079	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.154	05:07
Breath Volume		1.441	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:09
Subject Sample	# 2	0.143	05:10
Breath Volume		3.117	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FREY,JACOB
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:33

SUBJECT'S LAST NAME: GONZALEZ JR

DATE/TIME OF TEST: 12/02/2017 07:08:48

Diagnostics Check		OK	07:00
Air Blank		0.000	07:01
Control Test		0.080	07:01
Air Blank		0.000	07:01
Subject Sample	# 1	0.124	07:03
Breath Volume		1.882	07:03
Air Blank		0.000	07:04
Air Blank		0.000	07:06
Subject Sample	# 2	0.109	07:07
Breath Volume		1.347	07:07
Air Blank		0.000	07:07
Control Test		0.079	07:08
Air Blank		0.000	07:08
Diagnostics Check		OK	07:08

Data Download Date/Time: 12/15/2017 15:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CHIQUITO-RODRIGUEZ,LENNDY
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:52
GENDER: Female

OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: SORKIN

DATE/TIME OF TEST: 12/02/2017 20:56:12

Diagnostics Check		OK	20:49
Air Blank		0.000	20:49
Control Test		0.080	20:50
Air Blank		0.000	20:50
Subject Sample	# 1	0.012	20:51
Breath Volume		1.597	20:51
Air Blank		0.000	20:51
Air Blank		0.000	20:53
Subject Sample	# 2	0.011	20:54
Breath Volume		2.234	20:54
Air Blank		0.000	20:55
Control Test		0.080	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: JARRELL,THOMAS
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:21

SUBJECT'S LAST NAME: SANCHEZ CANE

DATE/TIME OF TEST: 12/02/2017 23:55:53

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.080	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.136	23:50
Breath Volume		1.937	23:50
Air Blank		0.000	23:50
Air Blank		0.000	23:52
Subject Sample	# 2	0.133	23:54
Breath Volume		1.726	23:54
Air Blank		0.000	23:54
Control Test		0.080	23:55
Air Blank		0.000	23:55
Diagnostics Check		OK	23:55

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: GUBISH,MATTHEW
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: GATTO

DATE/TIME OF TEST: 12/03/2017 01:55:15

Diagnostics Check		OK	01:43	
Air Blank		0.000	01:43	
Control Test		0.079	01:43	
Air Blank		0.000	01:44	
Subject Sample	# 1	0.218	01:47	Volume Not Met
Breath Volume		0.101	01:47	
Air Blank		0.000	01:48	
Air Blank		0.000	01:49	
Subject Sample	# 2	0.289	01:50	No .020 agreement
Breath Volume		1.566	01:50	
Air Blank		0.000	01:51	
Air Blank		0.000	01:53	
Subject Sample	# 3	0.256	01:53	No .020 agreement
Breath Volume		1.308	01:53	
Air Blank		0.000	01:54	
Control Test		0.078	01:54	
Air Blank		0.000	01:55	
Diagnostics Check		OK	01:55	

Data Download Date/Time: 12/15/2017 15:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: LONEY,MICHAEL
ARRESTING OFFICER AGENCY: Tequesta PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 12/03/2017 02:54:28

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.207	02:49
Breath Volume		2.621	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.207	02:52
Breath Volume		2.664	02:52
Air Blank		0.000	02:53
Control Test		0.079	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OWEN,SUE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MACCARONE,DAVID
ARRESTING OFFICER AGENCY: Palm Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: LACORTE

DATE/TIME OF TEST: 12/03/2017 03:27:51

Diagnostics Check		OK	03:21
Air Blank		0.000	03:21
Control Test		0.079	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.087	03:23
Breath Volume		2.207	03:23
Air Blank		0.000	03:23
Air Blank		0.000	03:25
Subject Sample	# 2	0.081	03:26
Breath Volume		1.863	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:27
Air Blank		0.000	03:27
Diagnostics Check		OK	03:27

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: PANIAGUA,MIGUEL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: CERNA

DATE/TIME OF TEST: 12/03/2017 15:28:51

Diagnostics Check		OK	15:20	
Air Blank		0.000	15:20	
Control Test		0.081	15:20	
Air Blank		0.000	15:21	
Subject Sample	# 1	0.127	15:24	Volume Not Met
Breath Volume		0.015	15:24	
Air Blank		0.000	15:25	
Air Blank		0.000	15:26	
Subject Sample	# 2		15:27	Subject Test Refused
Breath Volume		0.000	15:27	
Air Blank		0.000	15:27	
Control Test		0.078	15:28	
Air Blank		0.000	15:28	
Diagnostics Check		OK	15:28	

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: MIRANDA,EDWIN
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: ROBLERO ZACARIAS

DATE/TIME OF TEST: 12/03/2017 19:29:53

Diagnostics Check		OK	19:22
Air Blank		0.000	19:23
Control Test		0.080	19:23
Air Blank		0.000	19:24
Subject Sample	# 1	0.232	19:25
Breath Volume		2.726	19:25
Air Blank		0.000	19:25
Air Blank		0.000	19:27
Subject Sample	# 2	0.227	19:28
Breath Volume		2.582	19:28
Air Blank		0.000	19:28
Control Test		0.078	19:29
Air Blank		0.000	19:29
Diagnostics Check		OK	19:29

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CARHART,JOSHUA
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: HURST

DATE/TIME OF TEST: 12/03/2017 23:12:11

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.164	23:07
Breath Volume		2.996	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.159	23:10
Breath Volume		2.683	23:10
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: TODD,ZION
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: DOMINGO- JUAN

DATE/TIME OF TEST: 12/04/2017 02:28:34

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:15	
Control Test		0.080	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1	0.108	02:19	Volume Not Met
Breath Volume		1.027	02:19	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2	0.115	02:22	
Breath Volume		1.156	02:22	
Air Blank		0.000	02:23	
Air Blank		0.000	02:25	
Subject Sample	# 3	0.108	02:26	
Breath Volume		1.101	02:26	
Air Blank		0.000	02:27	
Control Test		0.078	02:27	
Air Blank		0.000	02:28	
Diagnostics Check		OK	02:28	

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HUYETT,MARK
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:41

SUBJECT'S LAST NAME: CAPLAN

DATE/TIME OF TEST: 12/04/2017 16:09:59

Diagnostics Check		OK	15:56	
Air Blank		0.000	15:56	
Control Test		0.080	15:57	
Air Blank		0.000	15:57	
Subject Sample	# 1	0.179	16:00	Volume Not Met
Breath Volume		0.253	16:00	
Air Blank		0.000	16:01	
Air Blank		0.000	16:03	
Subject Sample	# 2	0.203	16:04	
Breath Volume		1.773	16:04	
Air Blank		0.000	16:05	
Air Blank		0.000	16:07	
Subject Sample	# 3	0.209	16:08	
Breath Volume		1.320	16:08	
Air Blank		0.000	16:09	
Control Test		0.077	16:09	
Air Blank		0.000	16:09	
Diagnostics Check		OK	16:09	

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:42

SUBJECT'S LAST NAME: PEREZ HERNANDEZ

DATE/TIME OF TEST: 12/04/2017 21:14:05

Diagnostics Check		OK	21:06
Air Blank		0.000	21:06
Control Test		0.080	21:07
Air Blank		0.000	21:07
Subject Sample	# 1	0.156	21:09
Breath Volume		1.156	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.150	21:12
Breath Volume		1.546	21:12
Air Blank		0.000	21:13
Control Test		0.079	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WINDSOR,NICHOLAS
ARRESTING OFFICER AGENCY: Delray Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 12/07/2017 20:36:50

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.080	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.240	20:31
Breath Volume		1.207	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.227	20:35
Breath Volume		1.570	20:35
Air Blank		0.000	20:35
Control Test		0.080	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'NEAL,SHARI,L
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: HOPE,PAUL
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: BRUCELLA

DATE/TIME OF TEST: 12/08/2017 11:06:31

Diagnostics Check		OK	11:00
Air Blank		0.000	11:00
Control Test		0.080	11:01
Air Blank		0.000	11:01
Subject Sample	# 1	0.000	11:02
Breath Volume		2.296	11:02
Air Blank		0.000	11:02
Air Blank		0.000	11:04
Subject Sample	# 2	0.000	11:05
Breath Volume		1.925	11:05
Air Blank		0.000	11:05
Control Test		0.081	11:05
Air Blank		0.000	11:06
Diagnostics Check		OK	11:06

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: CASTRO,DENNIS
ARRESTING OFFICER AGENCY: Boynton Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:21

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 12/08/2017 20:54:06

Diagnostics Check		OK	20:47
Air Blank		0.000	20:47
Control Test		0.080	20:48
Air Blank		0.000	20:48
Subject Sample	# 1	0.000	20:49
Breath Volume		4.937	20:49
Air Blank		0.000	20:49
Air Blank		0.000	20:51
Subject Sample	# 2	0.000	20:52
Breath Volume		4.746	20:52
Air Blank		0.000	20:53
Control Test		0.080	20:53
Air Blank		0.000	20:54
Diagnostics Check		OK	20:54

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WHITE,EDWARD
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: PELLEGRINI JR

DATE/TIME OF TEST: 12/09/2017 01:55:33

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.000	01:51
Breath Volume		1.472	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.000	01:54
Breath Volume		1.476	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:27

SUBJECT'S LAST NAME: PHAN

DATE/TIME OF TEST: 12/09/2017 23:08:25

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.081	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.000	23:03
Breath Volume		2.542	23:03
Air Blank		0.000	23:04
Air Blank		0.000	23:06
Subject Sample	# 2	0.000	23:07
Breath Volume		2.906	23:07
Air Blank		0.000	23:07
Control Test		0.080	23:07
Air Blank		0.000	23:08
Diagnostics Check		OK	23:08

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: BORROWS,ANDREW
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: ROBEY

DATE/TIME OF TEST: 12/10/2017 00:32:39

Diagnostics Check		OK	00:25
Air Blank		0.000	00:25
Control Test		0.081	00:25
Air Blank		0.000	00:26
Subject Sample	# 1	0.229	00:27
Breath Volume		2.968	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.229	00:30
Breath Volume		3.042	00:30
Air Blank		0.000	00:31
Control Test		0.079	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:32

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALMER,SAMANTHA,M
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: FANDREY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Jupiter PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: PATRICK

DATE/TIME OF TEST: 12/10/2017 03:36:10

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.081	03:26
Air Blank		0.000	03:26
Subject Sample	# 1	0.234	03:27
Breath Volume		2.640	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.212	03:30
Breath Volume		1.843	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 3	0.213	03:34
Breath Volume		2.160	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARENT,GARY,J
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: WALTON,THOMAS
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: JIMENEZ LOPEZ

DATE/TIME OF TEST: 12/10/2017 20:36:33

Diagnostics Check		OK	20:30
Air Blank		0.000	20:30
Control Test		0.080	20:30
Air Blank		0.000	20:31
Subject Sample	# 1	0.192	20:31
Breath Volume		1.179	20:31
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.202	20:34
Breath Volume		1.664	20:34
Air Blank		0.000	20:35
Control Test		0.080	20:35
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: STEPHAN,ROBERT
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:18

SUBJECT'S LAST NAME: DIAZ-CRUZ

DATE/TIME OF TEST: 12/11/2017 16:54:39

Diagnostics Check		OK	16:48
Air Blank		0.000	16:48
Control Test		0.081	16:48
Air Blank		0.000	16:49
Subject Sample	# 1	0.136	16:49
Breath Volume		1.265	16:49
Air Blank		0.000	16:50
Air Blank		0.000	16:52
Subject Sample	# 2	0.138	16:53
Breath Volume		1.132	16:53
Air Blank		0.000	16:53
Control Test		0.080	16:54
Air Blank		0.000	16:54
Diagnostics Check		OK	16:54

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006477

INSTRUMENT REGISTERED TO: PALM BEACH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAIN,JEANETTE,
BREATH TEST OPERATOR AGENCY: Palm Beach County SO
ARRESTING OFFICER: SCARTOZZI,PATRICK
ARRESTING OFFICER AGENCY: Palm Beach County SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12317080A3
DRY GAS EXPIRY DATE: 07/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: ROCK

DATE/TIME OF TEST: 12/12/2017 22:55:29

Diagnostics Check		OK	22:48
Air Blank		0.000	22:49
Control Test		0.081	22:49
Air Blank		0.000	22:50
Subject Sample	# 1	0.172	22:50
Breath Volume		1.847	22:50
Air Blank		0.000	22:51
Air Blank		0.000	22:53
Subject Sample	# 2	0.171	22:53
Breath Volume		2.300	22:53
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Diagnostics Check		OK	22:55

Data Download Date/Time: 12/15/2017 15:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: ECKERT

DATE/TIME OF TEST: 12/08/2017 01:47:07

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.079	01:41
Air Blank		0.000	01:41
Subject Sample	# 1	0.073	01:42
Breath Volume		2.800	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:44
Subject Sample	# 2	0.073	01:45
Breath Volume		2.925	01:45
Air Blank		0.000	01:46
Control Test		0.080	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 12/18/2017 09:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLINE,GLENN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: CLINE,GLENN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 12/14/2017 13:53:33

Diagnostics Check	OK	13:52	
Air Blank		13:53	Ambient Fail
Air Blank	0.000	13:53	

Data Download Date/Time: 12/18/2017 09:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DIN,IMRAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 12/15/2017 16:45:21

Diagnostics Check	OK	16:44	
Air Blank		16:44	RFI Detect
Air Blank	0.000	16:45	

Data Download Date/Time: 12/18/2017 09:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DIN,IMRAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 12/15/2017 16:51:13

Diagnostics Check		OK	16:47	
Air Blank		0.000	16:47	
Control Test		0.078	16:48	
Air Blank		0.000	16:48	
Subject Sample	# 1	0.201	16:49	
Breath Volume		2.828	16:49	
Air Blank			16:49	RFI Detect
Air Blank		0.000	16:50	
Control Test		0.079	16:50	
Air Blank		0.000	16:51	
Diagnostics Check		OK	16:51	

Data Download Date/Time: 12/18/2017 09:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006559

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: DIN,IMRAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1841081
DRY GAS EXPIRY DATE: 04/08/2018
VIOLATION CODE: DUI
ARREST TIME: 15:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 12/15/2017 17:01:17

Diagnostics Check		OK	16:55
Air Blank		0.000	16:55
Control Test		0.078	16:55
Air Blank		0.000	16:56
Subject Sample	# 1	0.172	16:56
Breath Volume		2.117	16:56
Air Blank		0.000	16:57
Air Blank		0.000	16:59
Subject Sample	# 2	0.179	16:59
Breath Volume		2.835	16:59
Air Blank		0.000	17:00
Control Test		0.078	17:00
Air Blank		0.000	17:01
Diagnostics Check		OK	17:01

Data Download Date/Time: 12/18/2017 09:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:58

SUBJECT'S LAST NAME: LUGO

DATE/TIME OF TEST: 12/01/2017 04:25:09

Diagnostics Check		OK	04:18
Air Blank		0.000	04:19
Control Test		0.081	04:19
Air Blank		0.000	04:20
Subject Sample	# 1	0.193	04:20
Breath Volume		2.933	04:20
Air Blank		0.000	04:21
Air Blank		0.000	04:23
Subject Sample	# 2	0.195	04:23
Breath Volume		2.304	04:23
Air Blank		0.000	04:24
Control Test		0.080	04:24
Air Blank		0.000	04:25
Diagnostics Check		OK	04:25

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:54

SUBJECT'S LAST NAME: RITCHEL

DATE/TIME OF TEST: 12/01/2017 21:27:31

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:14	
Control Test		0.080	21:15	
Air Blank		0.000	21:15	
Subject Sample	# 1	0.189	21:18	Volume Not Met
Breath Volume		0.953	21:18	
Air Blank		0.000	21:19	
Air Blank		0.000	21:21	
Subject Sample	# 2	0.189	21:22	
Breath Volume		1.246	21:22	
Air Blank		0.000	21:23	
Air Blank		0.000	21:25	
Subject Sample	# 3	0.194	21:25	
Breath Volume		1.335	21:25	
Air Blank		0.000	21:26	
Control Test		0.078	21:26	
Air Blank		0.000	21:27	
Diagnostics Check		OK	21:27	

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,WILLIAM
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: STURGIS

DATE/TIME OF TEST: 12/02/2017 01:50:10

Diagnostics Check		OK	01:43
Air Blank		0.000	01:43
Control Test		0.081	01:44
Air Blank		0.000	01:44
Subject Sample	# 1	0.140	01:45
Breath Volume		1.996	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.133	01:48
Breath Volume		1.300	01:48
Air Blank		0.000	01:49
Control Test		0.080	01:49
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/02/2017 03:03:33

Diagnostics Check		OK	02:57
Air Blank		0.000	02:57
Control Test		0.080	02:58
Air Blank		0.000	02:58
Subject Sample	# 1	0.045	02:59
Breath Volume		3.406	02:59
Air Blank		0.000	02:59
Air Blank		0.000	03:01
Subject Sample	# 2	0.039	03:02
Breath Volume		2.062	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:02
Air Blank		0.000	03:03
Diagnostics Check		OK	03:03

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: BAYNE JR

DATE/TIME OF TEST: 12/02/2017 04:47:21

Diagnostics Check		OK	04:41
Air Blank		0.000	04:41
Control Test		0.080	04:41
Air Blank		0.000	04:42
Subject Sample	# 1	0.176	04:42
Breath Volume		2.878	04:42
Air Blank		0.000	04:43
Air Blank		0.000	04:45
Subject Sample	# 2	0.185	04:45
Breath Volume		2.929	04:45
Air Blank		0.000	04:46
Control Test		0.079	04:46
Air Blank		0.000	04:47
Diagnostics Check		OK	04:47

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: ESTEVEZ

DATE/TIME OF TEST: 12/03/2017 02:12:46

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.080	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.162	02:08
Breath Volume		2.960	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.170	02:11
Breath Volume		3.429	02:11
Air Blank		0.000	02:11
Control Test		0.080	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: SALEM

DATE/TIME OF TEST: 12/03/2017 04:26:08

Diagnostics Check		OK	04:17
Air Blank		0.000	04:17
Control Test		0.080	04:17
Air Blank		0.000	04:18
Subject Sample	# 1	0.105	04:21
Breath Volume		1.429	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:24
Subject Sample	# 2	0.107	04:24
Breath Volume		1.996	04:24
Air Blank		0.000	04:25
Control Test		0.079	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:48

SUBJECT'S LAST NAME: MAHANEY

DATE/TIME OF TEST: 12/03/2017 06:15:48

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.080	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.209	06:11
Breath Volume		1.886	06:11
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.206	06:14
Breath Volume		1.316	06:14
Air Blank		0.000	06:14
Control Test		0.080	06:15
Air Blank		0.000	06:15
Diagnostics Check		OK	06:15

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: MCGOWAN

DATE/TIME OF TEST: 12/04/2017 00:03:27

Diagnostics Check		OK	23:57
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:58
Subject Sample	# 2	0.141	23:58
Breath Volume		3.371	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.142	00:01
Breath Volume		3.656	00:01
Air Blank		0.000	00:02
Control Test		0.079	00:02
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: ULLMAN

DATE/TIME OF TEST: 12/05/2017 00:50:16

Diagnostics Check		OK	00:42
Air Blank		0.000	00:43
Control Test		0.080	00:43
Air Blank		0.000	00:43
Subject Sample	# 1	0.298	00:45
Breath Volume		1.664	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.305	00:48
Breath Volume		2.457	00:48
Air Blank		0.000	00:49
Control Test		0.080	00:49
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: SOTO JR

DATE/TIME OF TEST: 12/05/2017 02:11:40

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.080	02:03
Air Blank		0.000	02:04
Subject Sample	# 1	0.166	02:05
Breath Volume		1.179	02:05
Air Blank		0.000	02:06
Air Blank		0.000	02:07
Subject Sample	# 2	0.177	02:10
Breath Volume		1.906	02:10
Air Blank		0.000	02:10
Control Test		0.079	02:11
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:49

SUBJECT'S LAST NAME: BETZ

DATE/TIME OF TEST: 12/05/2017 18:17:17

Diagnostics Check		OK	18:09
Air Blank		0.000	18:10
Control Test		0.080	18:10
Air Blank		0.000	18:11
Subject Sample	# 1	0.279	18:12
Breath Volume		1.757	18:12
Air Blank		0.000	18:13
Air Blank		0.000	18:14
Subject Sample	# 2	0.293	18:15
Breath Volume		2.492	18:15
Air Blank		0.000	18:16
Control Test		0.079	18:16
Air Blank		0.000	18:17
Diagnostics Check		OK	18:17

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NAYLOR,I
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/06/2017 19:33:08

Diagnostics Check		OK	19:26
Air Blank		0.000	19:27
Control Test		0.079	19:27
Air Blank		0.000	19:28
Subject Sample	# 1	0.193	19:28
Breath Volume		3.535	19:28
Air Blank		0.000	19:29
Air Blank		0.000	19:31
Subject Sample	# 2	0.191	19:31
Breath Volume		3.613	19:31
Air Blank		0.000	19:32
Control Test		0.080	19:32
Air Blank		0.000	19:33
Diagnostics Check		OK	19:33

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: NIEMELAE

DATE/TIME OF TEST: 12/07/2017 02:17:34

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1		02:10	No Sample Provided
Breath Volume		0.000	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:13	
Subject Sample	# 2		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 12/19/2017 19:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: GALLOWAY

DATE/TIME OF TEST: 12/08/2017 00:23:56

Diagnostics Check		OK	00:09
Air Blank		0.000	00:09
Control Test		0.080	00:09
Air Blank		0.000	00:10
Subject Sample	# 1	0.201	00:13
Breath Volume		1.191	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.223	00:17
Breath Volume		1.371	00:17
Air Blank		0.000	00:17
Air Blank		0.000	00:19
Subject Sample	# 3	0.227	00:22
Breath Volume		1.191	00:22
Air Blank		0.000	00:22
Control Test		0.079	00:23
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 12/19/2017 19:32

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PIGOTT,C
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:38

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/08/2017 07:08:03

Diagnostics Check		OK	07:01
Air Blank		0.000	07:01
Control Test		0.080	07:02
Air Blank		0.000	07:02
Subject Sample	# 1	0.129	07:03
Breath Volume		1.824	07:03
Air Blank		0.000	07:04
Air Blank		0.000	07:06
Subject Sample	# 2	0.128	07:06
Breath Volume		1.804	07:06
Air Blank		0.000	07:07
Control Test		0.080	07:07
Air Blank		0.000	07:07
Diagnostics Check		OK	07:08

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: SHEEHAN

DATE/TIME OF TEST: 12/09/2017 01:17:37

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.256	01:11
Breath Volume		1.523	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:13
Subject Sample	# 2	0.243	01:15
Breath Volume		1.500	01:15
Air Blank		0.000	01:16
Control Test		0.078	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:08

SUBJECT'S LAST NAME: DILULLO

DATE/TIME OF TEST: 12/09/2017 05:38:20

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.080	05:30
Air Blank		0.000	05:30
Subject Sample	# 1	0.136	05:33
Breath Volume		1.738	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.149	05:36
Breath Volume		3.488	05:36
Air Blank		0.000	05:37
Control Test		0.079	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRENNAN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 12/09/2017 21:29:12

Diagnostics Check		OK	21:22
Air Blank		0.000	21:22
Control Test		0.081	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.280	21:24
Breath Volume		2.300	21:24
Air Blank		0.000	21:25
Air Blank		0.000	21:26
Subject Sample	# 2	0.271	21:27
Breath Volume		1.988	21:27
Air Blank		0.000	21:28
Control Test		0.080	21:28
Air Blank		0.000	21:29
Diagnostics Check		OK	21:29

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: O'CONNOR

DATE/TIME OF TEST: 12/10/2017 02:12:48

Diagnostics Check		OK	02:05
Air Blank		0.000	02:06
Control Test		0.080	02:06
Air Blank		0.000	02:07
Subject Sample	# 1	0.105	02:07
Breath Volume		1.523	02:07
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.119	02:11
Breath Volume		3.382	02:11
Air Blank		0.000	02:11
Control Test		0.079	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMMONS,ANGELA,V
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:12

SUBJECT'S LAST NAME: PIZARRO

DATE/TIME OF TEST: 12/10/2017 07:39:32

Diagnostics Check		OK	07:33
Air Blank		0.000	07:33
Control Test		0.080	07:33
Air Blank		0.000	07:34
Subject Sample	# 1	0.134	07:35
Breath Volume		3.265	07:35
Air Blank		0.000	07:35
Air Blank		0.000	07:37
Subject Sample	# 2	0.133	07:37
Breath Volume		2.886	07:37
Air Blank		0.000	07:38
Control Test		0.080	07:38
Air Blank		0.000	07:39
Diagnostics Check		OK	07:39

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: BRANDT

DATE/TIME OF TEST: 12/10/2017 22:05:45

Diagnostics Check		OK	21:59
Air Blank		0.000	21:59
Control Test		0.080	22:00
Air Blank		0.000	22:00
Subject Sample	# 1	0.146	22:01
Breath Volume		3.031	22:01
Air Blank		0.000	22:01
Air Blank		0.000	22:03
Subject Sample	# 2	0.153	22:04
Breath Volume		3.121	22:04
Air Blank		0.000	22:04
Control Test		0.081	22:05
Air Blank		0.000	22:05
Diagnostics Check		OK	22:05

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: MILLICAN

DATE/TIME OF TEST: 12/11/2017 05:43:27

Diagnostics Check		OK	05:36
Air Blank		0.000	05:36
Control Test		0.081	05:37
Air Blank		0.000	05:37
Subject Sample	# 1	0.042	05:38
Breath Volume		2.042	05:38
Air Blank		0.000	05:39
Air Blank		0.000	05:41
Subject Sample	# 2	0.038	05:41
Breath Volume		1.320	05:41
Air Blank		0.000	05:42
Control Test		0.080	05:42
Air Blank		0.000	05:43
Diagnostics Check		OK	05:43

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:51

SUBJECT'S LAST NAME: BURNHAM

DATE/TIME OF TEST: 12/12/2017 00:18:02

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.081	00:12
Air Blank		0.000	00:12
Subject Sample	# 1	0.155	00:13
Breath Volume		2.585	00:13
Air Blank		0.000	00:14
Air Blank		0.000	00:15
Subject Sample	# 2	0.152	00:16
Breath Volume		1.882	00:16
Air Blank		0.000	00:17
Control Test		0.081	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:18

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 12/12/2017 04:09:13

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.082	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.121	04:04
Breath Volume		1.828	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.127	04:07
Breath Volume		1.796	04:07
Air Blank		0.000	04:08
Control Test		0.080	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/14/2017 04:06:12

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.080	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.134	04:01
Breath Volume		1.250	04:01
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.132	04:04
Breath Volume		1.679	04:04
Air Blank		0.000	04:05
Control Test		0.079	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:18

SUBJECT'S LAST NAME: LIGAS

DATE/TIME OF TEST: 12/15/2017 00:44:45

Diagnostics Check		OK	00:38
Air Blank		0.000	00:39
Control Test		0.080	00:39
Air Blank		0.000	00:39
Subject Sample	# 1	0.000	00:40
Breath Volume		4.718	00:40
Air Blank		0.000	00:40
Air Blank		0.000	00:42
Subject Sample	# 2	0.000	00:43
Breath Volume		4.781	00:43
Air Blank		0.000	00:43
Control Test		0.080	00:44
Air Blank		0.000	00:44
Diagnostics Check		OK	00:44

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: RODRIGUEZ FLORES

DATE/TIME OF TEST: 12/16/2017 01:24:46

Diagnostics Check		OK	01:19	
Air Blank		0.000	01:19	
Control Test		0.081	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:23	Subject Test Refused
Breath Volume		0.000	01:23	
Air Blank		0.000	01:23	
Control Test		0.081	01:24	
Air Blank		0.000	01:24	
Diagnostics Check		OK	01:24	

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 12/16/2017 03:04:46

Diagnostics Check		OK	02:58
Air Blank		0.000	02:58
Control Test		0.080	02:59
Air Blank		0.000	02:59
Subject Sample	# 1	0.152	03:00
Breath Volume		2.351	03:00
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.149	03:03
Breath Volume		2.414	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/16/2017 05:26:59

Diagnostics Check		OK	05:18
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.165	05:21
Breath Volume		2.031	05:21
Air Blank		0.000	05:22
Air Blank		0.000	05:24
Subject Sample	# 2	0.159	05:25
Breath Volume		1.722	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:26
Diagnostics Check		OK	05:26

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/17/2017 01:50:51

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.191	01:46
Breath Volume		2.660	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.189	01:49
Breath Volume		1.972	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: CARRILLO

DATE/TIME OF TEST: 12/17/2017 05:57:34

Diagnostics Check		OK	05:51
Air Blank		0.000	05:51
Control Test		0.080	05:51
Air Blank		0.000	05:52
Subject Sample	# 1	0.192	05:52
Breath Volume		3.281	05:52
Air Blank		0.000	05:53
Air Blank		0.000	05:55
Subject Sample	# 2	0.188	05:55
Breath Volume		1.484	05:55
Air Blank		0.000	05:56
Control Test		0.080	05:56
Air Blank		0.000	05:57
Diagnostics Check		OK	05:57

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:13

SUBJECT'S LAST NAME: MOLINA

DATE/TIME OF TEST: 12/17/2017 20:40:06

Diagnostics Check		OK	20:33
Air Blank		0.000	20:34
Control Test		0.080	20:34
Air Blank		0.000	20:35
Subject Sample	# 1	0.148	20:35
Breath Volume		1.347	20:35
Air Blank		0.000	20:36
Air Blank		0.000	20:38
Subject Sample	# 2	0.155	20:38
Breath Volume		1.355	20:38
Air Blank		0.000	20:39
Control Test		0.079	20:39
Air Blank		0.000	20:40
Diagnostics Check		OK	20:40

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J.A.
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/18/2017 02:48:33

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.081	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:46	Slope Not Met
Breath Volume		0.597	02:46	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: MILLER,J.A.
ARRESTING OFFICER AGENCY: Temple Terrace PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:23

SUBJECT'S LAST NAME: BRYANT

DATE/TIME OF TEST: 12/18/2017 02:58:10

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.154	02:52
Breath Volume		1.332	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.149	02:56
Breath Volume		1.660	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SINANI,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: COLL

DATE/TIME OF TEST: 12/18/2017 21:10:51

Diagnostics Check		OK	21:04
Air Blank		0.000	21:04
Control Test		0.080	21:05
Air Blank		0.000	21:05
Subject Sample	# 1	0.162	21:06
Breath Volume		2.996	21:06
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 2	0.153	21:09
Breath Volume		4.000	21:09
Air Blank		0.000	21:09
Control Test		0.080	21:10
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: LLOYD

DATE/TIME OF TEST: 12/18/2017 22:58:20

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.080	22:52
Air Blank		0.000	22:53
Subject Sample	# 1	0.000	22:53
Breath Volume		2.031	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.000	22:56
Breath Volume		2.636	22:56
Air Blank		0.000	22:57
Control Test		0.079	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006567

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: KENNEDY,D 129956
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/19/2017 15:15:44

Diagnostics Check		OK	15:08
Air Blank		0.000	15:08
Control Test		0.080	15:09
Air Blank		0.000	15:09
Subject Sample	# 1	0.016	15:10
Breath Volume		2.125	15:10
Air Blank		0.000	15:11
Air Blank		0.000	15:13
Subject Sample	# 2	0.016	15:14
Breath Volume		1.562	15:14
Air Blank		0.000	15:14
Control Test		0.080	15:15
Air Blank		0.000	15:15
Diagnostics Check		OK	15:15

Data Download Date/Time: 12/19/2017 19:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,CHRISTOPHER,B
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WALL,BRAXTON
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:33

SUBJECT'S LAST NAME: FREDRICKSON

DATE/TIME OF TEST: 12/01/2017 14:01:37

Diagnostics Check		OK	13:53
Air Blank		0.000	13:53
Control Test		0.080	13:54
Air Blank		0.000	13:54
Subject Sample	# 1	0.173	13:57
Breath Volume		2.410	13:57
Air Blank		0.000	13:57
Air Blank		0.000	13:59
Subject Sample	# 2	0.177	14:00
Breath Volume		1.835	14:00
Air Blank		0.000	14:00
Control Test		0.078	14:01
Air Blank		0.000	14:01
Diagnostics Check		OK	14:01

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: TOWNS,R
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/15/2017 02:44:00

Diagnostics Check		OK	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.191	02:39
Breath Volume		2.191	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:41
Subject Sample	# 2	0.189	02:42
Breath Volume		1.980	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:44

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEHAN,SHANE,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: MEEHAN,SHANE
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: RESNICK

DATE/TIME OF TEST: 12/16/2017 22:44:14

Diagnostics Check		OK	22:37
Air Blank		0.000	22:38
Control Test		0.081	22:38
Air Blank		0.000	22:38
Subject Sample	# 1	0.000	22:39
Breath Volume		1.765	22:39
Air Blank		0.000	22:40
Air Blank		0.000	22:42
Subject Sample	# 2	0.000	22:42
Breath Volume		1.667	22:42
Air Blank		0.000	22:43
Control Test		0.080	22:43
Air Blank		0.000	22:44
Diagnostics Check		OK	22:44

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,D,
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: GADDIE,K
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:06

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/17/2017 19:47:31

Diagnostics Check		OK	19:44	
Air Blank		0.000	19:44	
Control Test		0.080	19:45	
Air Blank		0.000	19:45	
Subject Sample	# 1		19:46	Subject Test Refused
Breath Volume		0.000	19:46	
Air Blank		0.000	19:46	
Control Test		0.080	19:46	
Air Blank		0.000	19:47	
Diagnostics Check		OK	19:47	

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: PETRUZZELLA,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: STONE

DATE/TIME OF TEST: 12/19/2017 04:55:38

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:44	
Control Test		0.080	04:44	
Air Blank		0.000	04:45	
Subject Sample	# 1	0.149	04:48	Volume Not Met
Breath Volume		0.136	04:48	
Air Blank		0.000	04:49	
Air Blank		0.000	04:50	
Subject Sample	# 2	0.148	04:53	Volume Not Met
Breath Volume		0.292	04:53	
Air Blank		0.000	04:54	
Control Test		0.076	04:55	
Air Blank		0.000	04:55	
Diagnostics Check		OK	04:55	

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,F,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: 2C12,D
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 12/21/2017 20:53:09

Diagnostics Check		OK	20:45
Air Blank		0.000	20:46
Control Test		0.080	20:46
Air Blank		0.000	20:47
Subject Sample	# 1	0.141	20:48
Breath Volume		1.886	20:48
Air Blank		0.000	20:48
Air Blank		0.000	20:50
Subject Sample	# 2	0.129	20:51
Breath Volume		1.800	20:51
Air Blank		0.000	20:52
Control Test		0.077	20:52
Air Blank		0.000	20:53
Diagnostics Check		OK	20:53

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GADSON,LAUREN,J
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: LIMA,J
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SWEENEY

DATE/TIME OF TEST: 12/24/2017 21:39:18

Diagnostics Check		OK	21:35	
Air Blank		0.000	21:36	
Control Test		0.080	21:36	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:37	Subject Test Refused
Breath Volume		0.000	21:37	
Air Blank		0.000	21:38	
Control Test		0.079	21:38	
Air Blank		0.000	21:39	
Diagnostics Check		OK	21:39	

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,J,G
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: ROSSI,CRAIG
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:53

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/26/2017 18:37:56

Diagnostics Check		OK	18:30
Air Blank		0.000	18:30
Control Test		0.081	18:31
Air Blank		0.000	18:31
Subject Sample	# 1	0.261	18:32
Breath Volume		1.359	18:32
Air Blank		0.000	18:33
Air Blank		0.000	18:34
Subject Sample	# 2	0.270	18:35
Breath Volume		1.535	18:35
Air Blank		0.000	18:36
Control Test		0.078	18:37
Air Blank		0.000	18:37
Diagnostics Check		OK	18:37

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEMP,F,R
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OLIVA,J
ARRESTING OFFICER AGENCY: Flagler Beach PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: HATZL

DATE/TIME OF TEST: 12/27/2017 00:34:55

Diagnostics Check		OK	00:32	
Air Blank		0.000	00:32	
Control Test		0.080	00:32	
Air Blank		0.000	00:33	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.079	00:34	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEHAN,SHANE,P
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: WOODCOCK,M
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: FESMIRE

DATE/TIME OF TEST: 12/28/2017 00:09:50

Diagnostics Check		OK	00:03
Air Blank		0.000	00:04
Control Test		0.081	00:04
Air Blank		0.000	00:05
Subject Sample	# 1	0.000	00:05
Breath Volume		1.496	00:05
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.000	00:08
Breath Volume		1.207	00:08
Air Blank		0.000	00:08
Control Test		0.080	00:09
Air Blank		0.000	00:09
Diagnostics Check		OK	00:09

Data Download Date/Time: 12/29/2017 16:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,J,G
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OCAMPO,OMAR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: DECARLO

DATE/TIME OF TEST: 12/30/2017 23:48:59

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:45	
Control Test		0.081	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:48	Sequence Aborted
Breath Volume		0.000	23:48	
Air Blank		0.000	23:48	

Data Download Date/Time: 01/31/2018 14:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,J,G
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: OCAMPO,OMAR
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: DECARLO

DATE/TIME OF TEST: 12/30/2017 23:57:54

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.081	23:55	
Air Blank		0.000	23:56	
Subject Sample	# 1		23:56	Subject Test Refused
Breath Volume		0.000	23:56	
Air Blank		0.000	23:56	
Control Test		0.080	23:57	
Air Blank		0.000	23:57	
Diagnostics Check		OK	23:57	

Data Download Date/Time: 01/31/2018 14:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006616

INSTRUMENT REGISTERED TO: FLAGLER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHEWNING,AUSTIN,M
BREATH TEST OPERATOR AGENCY: Flagler County SO
ARRESTING OFFICER: CHEWNING,AUSTIN
ARRESTING OFFICER AGENCY: Flagler County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A1
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/31/2017 04:08:05

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.081	04:00
Air Blank		0.000	04:01
Subject Sample	# 1	0.000	04:01
Breath Volume		1.589	04:01
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.000	04:06
Breath Volume		1.410	04:06
Air Blank		0.000	04:07
Control Test		0.079	04:07
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 01/31/2018 14:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FONG,KENNETH,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ZIADIE,ANDREW
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: FURLONG

DATE/TIME OF TEST: 12/01/2017 17:15:16

Diagnostics Check		OK	17:12	
Air Blank		0.000	17:12	
Control Test		0.077	17:12	
Air Blank		0.000	17:13	
Subject Sample	# 1		17:13	Subject Test Refused
Breath Volume		0.000	17:13	
Air Blank		0.000	17:14	
Control Test		0.077	17:14	
Air Blank		0.000	17:15	
Diagnostics Check		OK	17:15	

Data Download Date/Time: 12/20/2017 11:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DANIEL,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: DAVIES

DATE/TIME OF TEST: 12/04/2017 00:08:38

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:06	
Control Test		0.078	00:06	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:07	
Control Test		0.078	00:08	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRAWFORD,A,
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: CASTILLO,J
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:03

SUBJECT'S LAST NAME: FAHME

DATE/TIME OF TEST: 12/05/2017 07:35:22

Diagnostics Check		OK	07:28
Air Blank		0.000	07:28
Control Test		0.078	07:29
Air Blank		0.000	07:29
Subject Sample	# 1	0.000	07:30
Breath Volume		2.281	07:30
Air Blank		0.000	07:31
Air Blank		0.000	07:33
Subject Sample	# 2	0.000	07:33
Breath Volume		1.281	07:33
Air Blank		0.000	07:34
Control Test		0.079	07:34
Air Blank		0.000	07:35
Diagnostics Check		OK	07:35

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALVAREZ,AMANDA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: TATUM

DATE/TIME OF TEST: 12/07/2017 04:20:17

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.077	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1		04:16	No Sample Provided
Breath Volume		0.000	04:16	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2		04:18	Subject Test Refused
Breath Volume		0.000	04:18	
Air Blank		0.000	04:19	
Control Test		0.078	04:19	
Air Blank		0.000	04:20	
Diagnostics Check		OK	04:20	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GANNON,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: BARR

DATE/TIME OF TEST: 12/08/2017 02:51:08

Diagnostics Check		OK	02:48	
Air Blank		0.000	02:48	
Control Test		0.078	02:49	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:50	
Control Test		0.078	02:50	
Air Blank		0.000	02:51	
Diagnostics Check		OK	02:51	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,A
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 12/09/2017 03:19:32

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.078	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.202	03:14
Breath Volume		1.718	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.205	03:17
Breath Volume		1.984	03:17
Air Blank		0.000	03:18
Control Test		0.077	03:18
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: MERRITT

DATE/TIME OF TEST: 12/11/2017 00:57:30

Diagnostics Check		OK	00:54	
Air Blank		0.000	00:55	
Control Test		0.079	00:55	
Air Blank		0.000	00:55	
Subject Sample	# 1		00:56	Subject Test Refused
Breath Volume		0.000	00:56	
Air Blank		0.000	00:56	
Control Test		0.080	00:56	
Air Blank		0.000	00:57	
Diagnostics Check		OK	00:57	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: AUER

DATE/TIME OF TEST: 12/13/2017 02:43:57

Diagnostics Check		OK	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Subject Sample	# 1	0.204	02:38
Breath Volume		2.023	02:38
Air Blank		0.000	02:39
Air Blank		0.000	02:41
Subject Sample	# 2	0.206	02:42
Breath Volume		1.457	02:42
Air Blank		0.000	02:43
Control Test		0.078	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:43

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: BUNCH

DATE/TIME OF TEST: 12/15/2017 22:11:22

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:08	
Control Test		0.078	22:09	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:10	
Control Test		0.078	22:10	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,STEPHEN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 12/15/2017 23:20:29

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.078	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.148	23:14
Breath Volume		2.113	23:14
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.136	23:18
Breath Volume		1.195	23:18
Air Blank		0.000	23:19
Control Test		0.075	23:19
Air Blank		0.000	23:20
Diagnostics Check		OK	23:20

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: REISSI,DANIEL
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: TASSIS

DATE/TIME OF TEST: 12/16/2017 00:16:36

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:14	
Control Test		0.077	00:14	
Air Blank		0.000	00:15	
Subject Sample	# 1		00:15	Subject Test Refused
Breath Volume		0.000	00:15	
Air Blank		0.000	00:15	
Control Test		0.078	00:16	
Air Blank		0.000	00:16	
Diagnostics Check		OK	00:16	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: CAVOUR

DATE/TIME OF TEST: 12/16/2017 01:02:19

Diagnostics Check		OK	00:59	
Air Blank		0.000	00:59	
Control Test		0.078	01:00	
Air Blank		0.000	01:00	
Subject Sample	# 1		01:00	Subject Test Refused
Breath Volume		0.000	01:00	
Air Blank		0.000	01:01	
Control Test		0.078	01:01	
Air Blank		0.000	01:02	
Diagnostics Check		OK	01:02	

Data Download Date/Time: 12/20/2017 11:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: ALVAREZ,AMANDA
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: LANGEVIN

DATE/TIME OF TEST: 12/16/2017 02:29:04

Diagnostics Check		OK	02:22
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.137	02:24
Breath Volume		2.734	02:24
Air Blank		0.000	02:24
Air Blank		0.000	02:26
Subject Sample	# 2	0.134	02:27
Breath Volume		3.324	02:27
Air Blank		0.000	02:28
Control Test		0.077	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:29

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHITE,RYAN,P
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: VANCAMP,JEFFERY
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: LETSCHERT

DATE/TIME OF TEST: 12/18/2017 00:44:11

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:41	
Control Test		0.078	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:42	Subject Test Refused
Breath Volume		0.000	00:42	
Air Blank		0.000	00:43	
Control Test		0.079	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VAN CAMP,JEFFERY,A
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: GENDEN,ERIC
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: FUCHSBERG

DATE/TIME OF TEST: 12/18/2017 22:51:38

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.078	22:46
Air Blank		0.000	22:46
Subject Sample	# 1	0.000	22:47
Breath Volume		1.664	22:47
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.000	22:50
Breath Volume		1.503	22:50
Air Blank		0.000	22:50
Control Test		0.075	22:51
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PARE,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 07:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:55

SUBJECT'S LAST NAME: LOZANO

DATE/TIME OF TEST: 12/19/2017 08:25:05

Diagnostics Check		OK	08:19	
Air Blank		0.000	08:19	
Control Test		0.079	08:20	
Air Blank		0.000	08:20	
Subject Sample	# 1	0.000	08:21	
Breath Volume		2.812	08:21	
Air Blank		0.000	08:21	
Air Blank		0.000	08:23	
Subject Sample	# 2		08:23	Subject Test Refused
Breath Volume		0.000	08:23	
Air Blank		0.000	08:24	
Control Test		0.078	08:24	
Air Blank		0.000	08:25	
Diagnostics Check		OK	08:25	

Data Download Date/Time: 12/20/2017 11:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PARE,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 08:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 12/19/2017 08:38:12

Diagnostics Check		OK	08:32	
Air Blank		0.000	08:32	
Control Test		0.079	08:33	
Air Blank		0.000	08:33	
Subject Sample	# 1	0.000	08:34	
Breath Volume		3.394	08:34	
Air Blank		0.000	08:34	
Air Blank		0.000	08:36	
Subject Sample	# 2		08:36	Subject Test Refused
Breath Volume		0.000	08:36	
Air Blank		0.000	08:37	
Control Test		0.078	08:37	
Air Blank		0.000	08:38	
Diagnostics Check		OK	08:38	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PARE,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: DAVENPORT

DATE/TIME OF TEST: 12/19/2017 09:24:49

Diagnostics Check		OK	09:18	
Air Blank		0.000	09:19	
Control Test		0.078	09:19	
Air Blank		0.000	09:20	
Subject Sample	# 1	0.000	09:20	
Breath Volume		1.496	09:20	
Air Blank		0.000	09:21	
Air Blank		0.000	09:23	
Subject Sample	# 2		09:23	Subject Test Refused
Breath Volume		0.000	09:23	
Air Blank		0.000	09:23	
Control Test		0.078	09:24	
Air Blank		0.000	09:24	
Diagnostics Check		OK	09:24	

Data Download Date/Time: 12/20/2017 11:34

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PARE,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 09:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:00

SUBJECT'S LAST NAME: SAINT FORT

DATE/TIME OF TEST: 12/19/2017 09:36:32

Diagnostics Check		OK	09:30	
Air Blank		0.000	09:31	
Control Test		0.078	09:31	
Air Blank		0.000	09:32	
Subject Sample	# 1	0.000	09:32	
Breath Volume		2.222	09:32	
Air Blank		0.000	09:32	
Air Blank		0.000	09:34	
Subject Sample	# 2		09:35	Subject Test Refused
Breath Volume		0.000	09:35	
Air Blank		0.000	09:35	
Control Test		0.078	09:35	
Air Blank		0.000	09:36	
Diagnostics Check		OK	09:36	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PARE,BRIAN,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: PARE,BRIAN
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 09:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: FRANKENSTEIN

DATE/TIME OF TEST: 12/20/2017 10:04:02

Diagnostics Check		OK	09:58	
Air Blank		0.000	09:58	
Control Test		0.078	09:59	
Air Blank		0.000	09:59	
Subject Sample	# 1	0.000	10:00	
Breath Volume		1.667	10:00	
Air Blank		0.000	10:00	
Air Blank		0.000	10:02	
Subject Sample	# 2		10:02	Subject Test Refused
Breath Volume		0.000	10:02	
Air Blank		0.000	10:03	
Control Test		0.078	10:03	
Air Blank		0.000	10:03	
Diagnostics Check		OK	10:04	

Data Download Date/Time: 12/20/2017 11:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: SAAVEDRA,ALONSO
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: JUAN PEDRO

DATE/TIME OF TEST: 12/22/2017 22:32:42

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:30	
Control Test		0.078	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 1		22:31	Subject Test Refused
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Control Test		0.078	22:32	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 01/30/2018 14:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFALKO,TRAVIS,K
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: DANIEL,REISSI
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: INGLE

DATE/TIME OF TEST: 12/28/2017 20:36:53

Diagnostics Check		OK	20:34	
Air Blank		0.000	20:34	
Control Test		0.078	20:34	
Air Blank		0.000	20:35	
Subject Sample	# 1		20:35	Subject Test Refused
Breath Volume		0.000	20:35	
Air Blank		0.000	20:35	
Control Test		0.079	20:36	
Air Blank		0.000	20:36	
Diagnostics Check		OK	20:36	

Data Download Date/Time: 01/30/2018 14:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DEEN,ALYIA,S
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: FONG,KENNETH
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/29/2017 01:03:50

Diagnostics Check		OK	01:00	
Air Blank		0.000	01:01	
Control Test		0.078	01:01	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:02	Subject Test Refused
Breath Volume		0.000	01:02	
Air Blank		0.000	01:02	
Control Test		0.078	01:03	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 01/30/2018 14:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006622

INSTRUMENT REGISTERED TO: BOCA RATON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REISSI,DAN,J
BREATH TEST OPERATOR AGENCY: Boca Raton PD
ARRESTING OFFICER: BISSOON,S
ARRESTING OFFICER AGENCY: Boca Raton PD
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A1
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/29/2017 23:50:33

Diagnostics Check		OK	23:47	
Air Blank		0.000	23:48	
Control Test		0.078	23:48	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:49	Subject Test Refused
Breath Volume		0.000	23:49	
Air Blank		0.000	23:49	
Control Test		0.078	23:49	
Air Blank		0.000	23:50	
Diagnostics Check		OK	23:50	

Data Download Date/Time: 01/30/2018 14:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 21:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:18

SUBJECT'S LAST NAME: EVERETT

DATE/TIME OF TEST: 12/25/2017 21:45:29

Diagnostics Check		OK	21:39
Air Blank		0.000	21:39
Control Test		0.081	21:40
Air Blank		0.000	21:40
Subject Sample	# 1	0.346	21:41
Breath Volume		1.355	21:41
Air Blank		0.000	21:41
Air Blank		0.000	21:43
Subject Sample	# 2	0.337	21:43
Breath Volume		1.257	21:43
Air Blank		0.000	21:44
Control Test		0.081	21:44
Air Blank		0.000	21:45
Diagnostics Check		OK	21:45

Data Download Date/Time: 12/30/2017 07:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006630

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ-CORTINA,RAUL,H
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: GONZALEZ-CORTINA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: JIMENEZ

DATE/TIME OF TEST: 12/31/2017 01:23:16

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.082	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.149	01:16
Breath Volume		1.382	01:16
Air Blank		0.000	01:17
Air Blank		0.000	01:19
Subject Sample	# 2	0.149	01:21
Breath Volume		1.187	01:21
Air Blank		0.000	01:22
Control Test		0.081	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 01/26/2018 10:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:11

SUBJECT'S LAST NAME: MARTINEZ RODRIGUEZ

DATE/TIME OF TEST: 12/23/2017 00:49:07

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.079	00:35	
Air Blank		0.000	00:36	
Subject Sample	# 1	0.148	00:39	Volume Not Met
Breath Volume		0.554	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2	0.170	00:43	
Breath Volume		1.148	00:43	
Air Blank		0.000	00:43	
Air Blank		0.000	00:45	
Subject Sample	# 3	0.161	00:47	
Breath Volume		1.105	00:47	
Air Blank		0.000	00:48	
Control Test		0.078	00:48	
Air Blank		0.000	00:49	
Diagnostics Check		OK	00:49	

Data Download Date/Time: 01/22/2018 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006631

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWE,BRANDON,D
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: ROWE,BRANDON
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: RODRIGUEZ CARRIZALES

DATE/TIME OF TEST: 12/23/2017 03:50:51

Diagnostics Check		OK	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Subject Sample	# 1	0.119	03:46
Breath Volume		1.570	03:46
Air Blank		0.000	03:46
Air Blank		0.000	03:48
Subject Sample	# 2	0.118	03:49
Breath Volume		1.421	03:49
Air Blank		0.000	03:49
Control Test		0.079	03:50
Air Blank		0.000	03:50
Diagnostics Check		OK	03:50

Data Download Date/Time: 01/22/2018 09:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: FODEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:20

SUBJECT'S LAST NAME: JUHLER-ROBBINS

DATE/TIME OF TEST: 12/01/2017 04:44:46

Diagnostics Check		OK	04:41	
Air Blank		0.000	04:42	
Control Test		0.080	04:42	
Air Blank		0.000	04:43	
Subject Sample	# 1		04:43	Subject Test Refused
Breath Volume		0.000	04:43	
Air Blank		0.000	04:43	
Control Test		0.079	04:44	
Air Blank		0.000	04:44	
Diagnostics Check		OK	04:44	

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: BURTON

DATE/TIME OF TEST: 12/09/2017 02:38:08

Diagnostics Check		OK	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Subject Sample	# 1	0.100	02:33
Breath Volume		1.710	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:35
Subject Sample	# 2	0.103	02:36
Breath Volume		2.093	02:36
Air Blank		0.000	02:37
Control Test		0.081	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 12/10/2017 00:00:42

Diagnostics Check		OK	23:53
Air Blank		0.000	23:53
Control Test		0.080	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.154	23:55
Breath Volume		1.207	23:55
Air Blank		0.000	23:56
Air Blank		0.000	23:58
Subject Sample	# 2	0.157	23:58
Breath Volume		1.773	23:58
Air Blank		0.000	23:59
Control Test		0.080	00:00
Air Blank		0.000	00:00
Diagnostics Check		OK	00:00

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GOMEZ SERRANO

DATE/TIME OF TEST: 12/10/2017 02:57:43

Diagnostics Check	OK	02:56	
Air Blank	0.000	02:56	
Control Test	0.084	02:57	Interferent Detect
Air Blank	0.000	02:57	

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GOMEZ SERRANO

DATE/TIME OF TEST: 12/10/2017 03:05:41

Diagnostics Check		OK	02:59
Air Blank		0.000	02:59
Control Test		0.082	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.100	03:01
Breath Volume		3.480	03:01
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.097	03:04
Breath Volume		3.527	03:04
Air Blank		0.000	03:04
Control Test		0.079	03:05
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/22/2017 03:42:15

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:30	
Control Test		0.081	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1	0.197	03:34	Volume Not Met
Breath Volume		0.144	03:34	
Air Blank		0.000	03:35	
Air Blank		0.000	03:37	
Subject Sample	# 2	0.185	03:40	Volume Not Met
Breath Volume		0.031	03:40	
Air Blank		0.000	03:41	
Control Test		0.075	03:41	
Air Blank		0.000	03:42	
Diagnostics Check		OK	03:42	

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: GRANTHAM

DATE/TIME OF TEST: 12/22/2017 20:12:10

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.080	20:06
Air Blank		0.000	20:06
Subject Sample	# 1	0.196	20:07
Breath Volume		2.382	20:07
Air Blank		0.000	20:08
Air Blank		0.000	20:09
Subject Sample	# 2	0.188	20:10
Breath Volume		2.253	20:10
Air Blank		0.000	20:11
Control Test		0.077	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: MILANO

DATE/TIME OF TEST: 12/22/2017 23:00:37

Diagnostics Check		OK	22:52
Air Blank		0.000	22:53
Control Test		0.078	22:53
Air Blank		0.000	22:54
Subject Sample	# 1	0.203	22:54
Breath Volume		2.167	22:54
Air Blank		0.000	22:55
Air Blank		0.000	22:57
Subject Sample	# 2	0.202	22:58
Breath Volume		1.976	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	23:00
Diagnostics Check		OK	23:00

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:30

SUBJECT'S LAST NAME: RODRIGUEZ PONS

DATE/TIME OF TEST: 12/24/2017 19:04:30

Diagnostics Check		OK	18:58
Air Blank		0.000	18:58
Control Test		0.080	18:59
Air Blank		0.000	18:59
Subject Sample	# 1	0.147	18:59
Breath Volume		2.136	18:59
Air Blank		0.000	19:00
Air Blank		0.000	19:02
Subject Sample	# 2	0.143	19:02
Breath Volume		2.250	19:02
Air Blank		0.000	19:03
Control Test		0.076	19:03
Air Blank		0.000	19:04
Diagnostics Check		OK	19:04

Data Download Date/Time: 12/28/2017 12:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 12/29/2017 00:08:09

Diagnostics Check		OK	00:05	
Air Blank		0.000	00:05	
Control Test		0.083	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1	0.141	00:06	
Breath Volume		2.347	00:06	
Air Blank			00:07	Purge Fail
Air Blank		0.000	00:08	

Data Download Date/Time: 01/30/2018 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006633

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANGELICCHI,JOSEPH,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: ANGELICCHI,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 12/29/2017 00:16:26

Diagnostics Check		OK	00:09
Air Blank		0.000	00:10
Control Test		0.080	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.139	00:11
Breath Volume		2.027	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.141	00:14
Breath Volume		2.527	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 01/30/2018 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: PENA-GARCIA

DATE/TIME OF TEST: 12/05/2017 22:50:30

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.080	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.178	22:45
Breath Volume		1.609	22:45
Air Blank		0.000	22:46
Air Blank		0.000	22:48
Subject Sample	# 2	0.177	22:48
Breath Volume		1.789	22:48
Air Blank		0.000	22:49
Control Test		0.079	22:49
Air Blank		0.000	22:50
Diagnostics Check		OK	22:50

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: GONZALES,AARON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:34

SUBJECT'S LAST NAME: GANT

DATE/TIME OF TEST: 12/06/2017 22:10:40

Diagnostics Check		OK	22:03
Air Blank		0.000	22:03
Control Test		0.079	22:03
Air Blank		0.000	22:04
Subject Sample	# 1	0.176	22:06
Breath Volume		1.660	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.186	22:09
Breath Volume		1.464	22:09
Air Blank		0.000	22:09
Control Test		0.078	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,J
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 12/09/2017 04:35:15

Diagnostics Check	OK	04:34	
Air Blank		04:34	RFI Detect
Air Blank	0.000	04:35	

Data Download Date/Time: 12/18/2017 11:05

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,GWODZ,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: HOUSTON

DATE/TIME OF TEST: 12/09/2017 04:43:25

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.192	04:39
Breath Volume		2.656	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.189	04:41
Breath Volume		2.761	04:41
Air Blank		0.000	04:42
Control Test		0.078	04:42
Air Blank		0.000	04:43
Diagnostics Check		OK	04:43

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: BERGERON

DATE/TIME OF TEST: 12/09/2017 20:12:50

Diagnostics Check		OK	20:05
Air Blank		0.000	20:05
Control Test		0.080	20:06
Air Blank		0.000	20:06
Subject Sample	# 1	0.187	20:07
Breath Volume		1.628	20:07
Air Blank		0.000	20:08
Air Blank		0.000	20:10
Subject Sample	# 2	0.191	20:11
Breath Volume		2.136	20:11
Air Blank		0.000	20:11
Control Test		0.079	20:12
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LEON-PAULINO

DATE/TIME OF TEST: 12/10/2017 23:29:05

Diagnostics Check		OK	23:22
Air Blank		0.000	23:22
Control Test		0.081	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.240	23:24
Breath Volume		3.417	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.232	23:27
Breath Volume		2.437	23:27
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:29
Diagnostics Check		OK	23:29

Data Download Date/Time: 12/18/2017 11:05

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: LOPEZ LOPEZ

DATE/TIME OF TEST: 12/10/2017 23:41:45

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.080	23:34
Air Blank		0.000	23:35
Subject Sample	# 1	0.176	23:37
Breath Volume		1.585	23:37
Air Blank		0.000	23:37
Air Blank		0.000	23:39
Subject Sample	# 2	0.167	23:40
Breath Volume		1.984	23:40
Air Blank		0.000	23:40
Control Test		0.076	23:41
Air Blank		0.000	23:41
Diagnostics Check		OK	23:41

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: O'STEEN

DATE/TIME OF TEST: 12/14/2017 01:43:56

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.080	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.222	01:39
Breath Volume		2.558	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.217	01:42
Breath Volume		2.789	01:42
Air Blank		0.000	01:43
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,RONALD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: PETERSON,RONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: KLUSON

DATE/TIME OF TEST: 12/16/2017 21:33:56

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.000	21:29
Breath Volume		2.894	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.000	21:32
Breath Volume		2.785	21:32
Air Blank		0.000	21:33
Control Test		0.080	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,RONALD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: PETERSON,RONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: KLUSON

DATE/TIME OF TEST: 12/16/2017 21:48:07

Diagnostics Check		OK	21:41
Air Blank		0.000	21:42
Control Test		0.080	21:42
Air Blank		0.000	21:43
Subject Sample	# 1	0.000	21:43
Breath Volume		2.828	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.000	21:46
Breath Volume		3.089	21:46
Air Blank		0.000	21:47
Control Test		0.079	21:47
Air Blank		0.000	21:48
Diagnostics Check		OK	21:48

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 12/16/2017 23:22:57

Diagnostics Check		OK	23:11	
Air Blank		0.000	23:11	
Control Test		0.079	23:11	
Air Blank		0.000	23:12	
Subject Sample	# 1	0.238	23:15	Volume Not Met
Breath Volume		0.738	23:15	
Air Blank		0.000	23:16	
Air Blank		0.000	23:18	
Subject Sample	# 2	0.182	23:21	Volume Not Met
Breath Volume		0.558	23:21	
Air Blank		0.000	23:21	
Control Test		0.075	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 12/18/2017 11:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALES,AARON,K
BREATH TEST OPERATOR AGENCY: FFWCC (Lakeland)
ARRESTING OFFICER: GONZALES,AARON
ARRESTING OFFICER AGENCY: FFWCC (Lakeland)
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:06

SUBJECT'S LAST NAME: CHARBONEAU

DATE/TIME OF TEST: 12/22/2017 22:40:12

Diagnostics Check		OK	22:33
Air Blank		0.000	22:34
Control Test		0.080	22:34
Air Blank		0.000	22:35
Subject Sample	# 1	0.176	22:35
Breath Volume		1.695	22:35
Air Blank		0.000	22:36
Air Blank		0.000	22:38
Subject Sample	# 2	0.168	22:38
Breath Volume		1.417	22:38
Air Blank		0.000	22:39
Control Test		0.079	22:39
Air Blank		0.000	22:40
Diagnostics Check		OK	22:40

Data Download Date/Time: 01/22/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: SANDS,B
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 572920
DRY GAS EXPIRY DATE: 09/23/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: CHILICAS

DATE/TIME OF TEST: 12/23/2017 04:24:05

Diagnostics Check		OK	04:19	
Air Blank		0.000	04:20	
Control Test		0.080	04:20	
Air Blank		0.000	04:20	
Subject Sample	# 1		04:22	Subject Test Refused
Breath Volume		0.000	04:22	
Air Blank		0.000	04:23	
Control Test		0.081	04:23	
Air Blank		0.000	04:24	
Diagnostics Check		OK	04:24	

Data Download Date/Time: 01/22/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: SAVCHENKO

DATE/TIME OF TEST: 12/24/2017 21:10:34

Diagnostics Check		OK	21:05	
Air Blank		0.000	21:05	
Control Test		0.080	21:06	
Air Blank		0.000	21:06	
Subject Sample	# 1		21:08	Subject Test Refused
Breath Volume		0.273	21:08	
Air Blank		0.000	21:09	
Control Test		0.080	21:09	
Air Blank		0.000	21:10	
Diagnostics Check		OK	21:10	

Data Download Date/Time: 01/22/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 12/25/2017 02:59:01

Diagnostics Check		OK	02:56	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Subject Sample	# 1		02:57	Subject Test Refused
Breath Volume		0.000	02:57	
Air Blank		0.000	02:58	
Control Test		0.080	02:58	
Air Blank		0.000	02:58	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 01/22/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006634

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPURLOCK,ROBERT,V
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SPURLOCK,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: OTHER
ARREST TIME: 15:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 12/31/2017 15:53:07

Diagnostics Check		OK	15:47
Air Blank		0.000	15:47
Control Test		0.080	15:48
Air Blank		0.000	15:48
Subject Sample	# 1	0.000	15:48
Breath Volume		2.656	15:48
Air Blank		0.000	15:49
Air Blank		0.000	15:51
Subject Sample	# 2	0.000	15:51
Breath Volume		2.996	15:51
Air Blank		0.000	15:52
Control Test		0.078	15:52
Air Blank		0.000	15:53
Diagnostics Check		OK	15:53

Data Download Date/Time: 01/22/2018 12:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006636

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A3
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: HARRISON

DATE/TIME OF TEST: 12/28/2017 23:07:40

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:02
Air Blank		0.000	23:02
Subject Sample	# 1	0.100	23:03
Breath Volume		2.332	23:03
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.098	23:06
Breath Volume		2.523	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 12/31/2017 05:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: CARLSON

DATE/TIME OF TEST: 12/01/2017 21:48:03

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.077	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.000	21:43
Breath Volume		1.609	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.000	21:46
Breath Volume		1.527	21:46
Air Blank		0.000	21:47
Control Test		0.078	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:48

Data Download Date/Time: 12/28/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:35

SUBJECT'S LAST NAME: MCCULLOUGH

DATE/TIME OF TEST: 12/02/2017 05:01:33

Diagnostics Check		OK	04:55
Air Blank		0.000	04:55
Control Test		0.079	04:56
Air Blank		0.000	04:56
Subject Sample	# 1	0.088	04:57
Breath Volume		1.445	04:57
Air Blank		0.000	04:57
Air Blank		0.000	04:59
Subject Sample	# 2	0.085	04:59
Breath Volume		1.605	04:59
Air Blank		0.000	05:00
Control Test		0.077	05:00
Air Blank		0.000	05:01
Diagnostics Check		OK	05:01

Data Download Date/Time: 12/28/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MACNAUGHTON

DATE/TIME OF TEST: 12/15/2017 21:47:54

Diagnostics Check		OK	21:36	
Air Blank		0.000	21:36	
Control Test		0.079	21:36	
Air Blank		0.000	21:37	
Subject Sample	# 1		21:40	No Sample Provided
Breath Volume		0.000	21:40	
Air Blank		0.000	21:41	
Air Blank		0.000	21:43	
Subject Sample	# 2	0.135	21:46	Volume Not Met
Breath Volume		0.125	21:46	
Air Blank		0.000	21:46	
Control Test		0.076	21:47	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 12/28/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: MACNAUGHTON

DATE/TIME OF TEST: 12/15/2017 21:53:35

Diagnostics Check		OK	21:48	
Air Blank		0.000	21:49	
Control Test		0.078	21:49	
Air Blank		0.000	21:50	
Subject Sample	# 1		21:51	Subject Test Refused
Breath Volume		0.097	21:51	
Air Blank		0.000	21:52	
Control Test		0.078	21:52	
Air Blank		0.000	21:53	
Diagnostics Check		OK	21:53	

Data Download Date/Time: 12/28/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALFY,RACHEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: LASHLEY

DATE/TIME OF TEST: 12/17/2017 02:39:04

Diagnostics Check		OK	02:31
Air Blank		0.000	02:32
Control Test		0.080	02:32
Air Blank		0.000	02:33
Subject Sample	# 1	0.209	02:33
Breath Volume		1.195	02:33
Air Blank		0.000	02:34
Air Blank		0.000	02:36
Subject Sample	# 2	0.226	02:37
Breath Volume		1.382	02:37
Air Blank		0.000	02:38
Control Test		0.078	02:38
Air Blank		0.000	02:38
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/28/2017 12:32

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARRIS,SKYLAAR,D
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: PALFY,RACHEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/30/2017 20:34:45

Diagnostics Check		OK	20:28
Air Blank		0.000	20:28
Control Test		0.079	20:28
Air Blank		0.000	20:29
Subject Sample	# 1	0.184	20:30
Breath Volume		1.902	20:30
Air Blank		0.000	20:30
Air Blank		0.000	20:32
Subject Sample	# 2	0.192	20:33
Breath Volume		1.343	20:33
Air Blank		0.000	20:33
Control Test		0.079	20:34
Air Blank		0.000	20:34
Diagnostics Check		OK	20:34

Data Download Date/Time: 01/30/2018 13:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006638

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:03

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/30/2017 23:42:57

Diagnostics Check		OK	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:36
Subject Sample	# 1	0.126	23:38
Breath Volume		1.417	23:38
Air Blank		0.000	23:39
Air Blank		0.000	23:40
Subject Sample	# 2	0.115	23:41
Breath Volume		1.503	23:41
Air Blank		0.000	23:42
Control Test		0.079	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 01/30/2018 13:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BEHLENDORF

DATE/TIME OF TEST: 12/06/2017 00:31:00

Diagnostics Check		OK	00:24
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:25
Subject Sample	# 1	0.193	00:26
Breath Volume		2.457	00:26
Air Blank		0.000	00:27
Air Blank		0.000	00:29
Subject Sample	# 2	0.199	00:29
Breath Volume		2.632	00:29
Air Blank		0.000	00:30
Control Test		0.078	00:30
Air Blank		0.000	00:30
Diagnostics Check		OK	00:31

Data Download Date/Time: 12/18/2017 09:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BORES,ALAN,P
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: MEDICO III

DATE/TIME OF TEST: 12/23/2017 05:29:49

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:19	
Control Test		0.078	05:20	
Air Blank		0.000	05:20	
Subject Sample	# 1		05:23	No Sample Provided
Breath Volume		0.000	05:23	
Air Blank		0.000	05:24	
Air Blank		0.000	05:26	
Subject Sample	# 2		05:28	Subject Test Refused
Breath Volume		0.000	05:28	
Air Blank		0.000	05:28	
Control Test		0.079	05:29	
Air Blank		0.000	05:29	
Diagnostics Check		OK	05:29	

Data Download Date/Time: 01/22/2018 11:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006692

INSTRUMENT REGISTERED TO: HOLMES BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,MIKE,J
BREATH TEST OPERATOR AGENCY: Holmes Beach PD
ARRESTING OFFICER: BORES,ALAN
ARRESTING OFFICER AGENCY: Holmes Beach PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:20

SUBJECT'S LAST NAME: GUNNING

DATE/TIME OF TEST: 12/23/2017 05:51:55

Diagnostics Check		OK	05:45
Air Blank		0.000	05:45
Control Test		0.078	05:46
Air Blank		0.000	05:46
Subject Sample	# 1	0.094	05:47
Breath Volume		1.902	05:47
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.093	05:50
Breath Volume		2.207	05:50
Air Blank		0.000	05:51
Control Test		0.077	05:51
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 01/22/2018 11:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: STEINMANN

DATE/TIME OF TEST: 12/02/2017 23:58:52

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:56	
Control Test		0.079	23:56	
Air Blank		0.000	23:57	
Subject Sample	# 1		23:57	Subject Test Refused
Breath Volume		0.000	23:57	
Air Blank		0.000	23:57	
Control Test		0.079	23:58	
Air Blank		0.000	23:58	
Diagnostics Check		OK	23:58	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:55

SUBJECT'S LAST NAME: UHLAND

DATE/TIME OF TEST: 12/03/2017 05:20:57

Diagnostics Check		OK	05:18	
Air Blank		0.000	05:18	
Control Test		0.078	05:18	
Air Blank		0.000	05:19	
Subject Sample	# 1		05:19	Subject Test Refused
Breath Volume		0.000	05:19	
Air Blank		0.000	05:20	
Control Test		0.079	05:20	
Air Blank		0.000	05:20	
Diagnostics Check		OK	05:20	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 17:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:41

SUBJECT'S LAST NAME: TOLER

DATE/TIME OF TEST: 12/06/2017 18:07:04

Diagnostics Check		OK	18:04	
Air Blank		0.000	18:04	
Control Test		0.078	18:04	
Air Blank		0.000	18:05	
Subject Sample	# 1		18:05	Subject Test Refused
Breath Volume		0.000	18:05	
Air Blank		0.000	18:06	
Control Test		0.079	18:06	
Air Blank		0.000	18:06	
Diagnostics Check		OK	18:07	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JORDAN,MICHAEL,R
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: JORDAN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 12/10/2017 00:16:33

Diagnostics Check		OK	00:08	
Air Blank		0.000	00:08	
Control Test		0.080	00:09	
Air Blank		0.000	00:09	
Subject Sample	# 1		00:12	No Sample Provided
Breath Volume		0.000	00:12	
Air Blank		0.000	00:13	
Air Blank		0.000	00:15	
Subject Sample	# 2		00:16	Sequence Aborted
Breath Volume		0.000	00:16	
Air Blank		0.000	00:16	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SEAGO,SHEILA,Y
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: TORRES,JOEL
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 12/10/2017 01:06:01

Diagnostics Check		OK	00:59
Air Blank		0.000	00:59
Control Test		0.078	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.138	01:01
Breath Volume		2.410	01:01
Air Blank		0.000	01:02
Air Blank		0.000	01:03
Subject Sample	# 2	0.129	01:04
Breath Volume		1.609	01:04
Air Blank		0.000	01:05
Control Test		0.076	01:05
Air Blank		0.000	01:05
Diagnostics Check		OK	01:06

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 12/14/2017 00:34:29

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:32	
Control Test		0.079	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:33	Subject Test Refused
Breath Volume		0.000	00:33	
Air Blank		0.000	00:33	
Control Test		0.080	00:33	
Air Blank		0.000	00:34	
Diagnostics Check		OK	00:34	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FAJARDO,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: REINER

DATE/TIME OF TEST: 12/16/2017 04:59:58

Diagnostics Check		OK	04:56	
Air Blank		0.000	04:57	
Control Test		0.078	04:57	
Air Blank		0.000	04:58	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:59	
Control Test		0.080	04:59	
Air Blank		0.000	04:59	
Diagnostics Check		OK	04:59	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KERNS,CODY,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: FAJARDO,MARIO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 11/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: MEDRANO FLORES

DATE/TIME OF TEST: 12/17/2017 23:15:54

Diagnostics Check		OK	23:12	
Air Blank		0.000	23:13	
Control Test		0.079	23:13	
Air Blank		0.000	23:14	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:14	
Control Test		0.079	23:15	
Air Blank		0.000	23:15	
Diagnostics Check		OK	23:15	

Data Download Date/Time: 12/21/2017 15:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: O'KEEFE,PATRICIA,M
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRANCO,BRYAN
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 18:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: GOOCH

DATE/TIME OF TEST: 12/24/2017 19:16:14

Diagnostics Check		OK	19:13	
Air Blank		0.000	19:13	
Control Test		0.079	19:14	
Air Blank		0.000	19:14	
Subject Sample	# 1		19:14	Subject Test Refused
Breath Volume		0.000	19:14	
Air Blank		0.000	19:15	
Control Test		0.079	19:15	
Air Blank		0.000	19:16	
Diagnostics Check		OK	19:16	

Data Download Date/Time: 01/18/2018 12:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALVAREZ,ORLANDO,
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: ALVAREZ,ORLANDO
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: RAMIREZ ARAUZ

DATE/TIME OF TEST: 12/30/2017 00:53:39

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:50	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:52	Subject Test Refused
Breath Volume		0.000	00:52	
Air Blank		0.000	00:52	
Control Test		0.080	00:53	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 01/18/2018 12:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006693

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,W
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: MACHADO LEON

DATE/TIME OF TEST: 12/30/2017 19:12:24

Diagnostics Check		OK	19:09	
Air Blank		0.000	19:09	
Control Test		0.081	19:10	
Air Blank		0.000	19:10	
Subject Sample	# 1		19:11	Subject Test Refused
Breath Volume		0.000	19:11	
Air Blank		0.000	19:11	
Control Test		0.080	19:11	
Air Blank		0.000	19:12	
Diagnostics Check		OK	19:12	

Data Download Date/Time: 01/18/2018 12:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 16:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:43

SUBJECT'S LAST NAME: GOKEY III

DATE/TIME OF TEST: 12/02/2017 17:12:47

Diagnostics Check		OK	17:06
Air Blank		0.000	17:06
Control Test		0.080	17:07
Air Blank		0.000	17:07
Subject Sample	# 1	0.078	17:08
Breath Volume		4.308	17:08
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 2	0.075	17:11
Breath Volume		4.285	17:11
Air Blank		0.000	17:11
Control Test		0.080	17:12
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 18:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:13

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/02/2017 22:14:48

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:12	
Control Test		0.081	22:12	
Air Blank		0.000	22:13	
Subject Sample	# 1		22:13	Subject Test Refused
Breath Volume		0.000	22:13	
Air Blank		0.000	22:13	
Control Test		0.081	22:14	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: WOOSLEY

DATE/TIME OF TEST: 12/02/2017 23:45:16

Diagnostics Check		OK	23:38
Air Blank		0.000	23:39
Control Test		0.081	23:39
Air Blank		0.000	23:40
Subject Sample	# 1	0.145	23:40
Breath Volume		3.390	23:40
Air Blank		0.000	23:41
Air Blank		0.000	23:43
Subject Sample	# 2	0.150	23:43
Breath Volume		2.792	23:43
Air Blank		0.000	23:44
Control Test		0.081	23:44
Air Blank		0.000	23:45
Diagnostics Check		OK	23:45

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: WARRICK

DATE/TIME OF TEST: 12/03/2017 22:18:27

Diagnostics Check		OK	22:11
Air Blank		0.000	22:11
Control Test		0.080	22:12
Air Blank		0.000	22:12
Subject Sample	# 1	0.266	22:13
Breath Volume		1.523	22:13
Air Blank		0.000	22:14
Air Blank		0.000	22:15
Subject Sample	# 2	0.250	22:16
Breath Volume		2.125	22:16
Air Blank		0.000	22:17
Control Test		0.079	22:17
Air Blank		0.000	22:18
Diagnostics Check		OK	22:18

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GUTIERREZ-RODRIGUEZ

DATE/TIME OF TEST: 12/04/2017 00:55:45

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.080	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:51	No Sample Provided
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 2		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:54	
Control Test		0.080	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 12/18/2017 12:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: GOMEZ

DATE/TIME OF TEST: 12/04/2017 01:29:05

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.081	01:23
Air Blank		0.000	01:23
Subject Sample	# 1	0.114	01:24
Breath Volume		1.644	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.107	01:27
Breath Volume		1.769	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:02

SUBJECT'S LAST NAME: ANDRIES

DATE/TIME OF TEST: 12/06/2017 17:45:14

Diagnostics Check		OK	17:38
Air Blank		0.000	17:39
Control Test		0.080	17:39
Air Blank		0.000	17:40
Subject Sample	# 1	0.000	17:40
Breath Volume		1.898	17:40
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 2	0.000	17:43
Breath Volume		2.230	17:43
Air Blank		0.000	17:44
Control Test		0.080	17:44
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:52

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/07/2017 01:19:53

Diagnostics Check		OK	01:12
Air Blank		0.000	01:12
Control Test		0.080	01:13
Air Blank		0.000	01:13
Subject Sample	# 1	0.221	01:14
Breath Volume		2.414	01:14
Air Blank		0.000	01:14
Air Blank		0.000	01:16
Subject Sample	# 2	0.236	01:18
Breath Volume		3.339	01:18
Air Blank		0.000	01:18
Control Test		0.077	01:19
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:08

SUBJECT'S LAST NAME: BRINSON

DATE/TIME OF TEST: 12/08/2017 03:38:50

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.123	03:33
Breath Volume		1.441	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.136	03:37
Breath Volume		1.917	03:37
Air Blank		0.000	03:37
Control Test		0.079	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILDT,ROBERT,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SMITH,BRENT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:45

SUBJECT'S LAST NAME: HAMIDI

DATE/TIME OF TEST: 12/10/2017 09:42:40

Diagnostics Check		OK	09:32
Air Blank		0.000	09:32
Control Test		0.080	09:33
Air Blank		0.000	09:33
Subject Sample	# 1	0.094	09:36
Breath Volume		1.253	09:36
Air Blank		0.000	09:36
Air Blank		0.000	09:38
Subject Sample	# 2	0.085	09:41
Breath Volume		1.417	09:41
Air Blank		0.000	09:41
Control Test		0.080	09:42
Air Blank		0.000	09:42
Diagnostics Check		OK	09:42

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: GRAVES

DATE/TIME OF TEST: 12/10/2017 23:56:47

Diagnostics Check		OK	23:50
Air Blank		0.000	23:51
Control Test		0.081	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.164	23:52
Breath Volume		2.238	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.166	23:55
Breath Volume		2.449	23:55
Air Blank		0.000	23:55
Control Test		0.080	23:56
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATE,ROBERT,J
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: TATE,ROBERT
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 10:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:14

SUBJECT'S LAST NAME: CHRISTMAN

DATE/TIME OF TEST: 12/13/2017 11:41:07

Diagnostics Check		OK	11:34
Air Blank		0.000	11:35
Control Test		0.082	11:35
Air Blank		0.000	11:36
Subject Sample	# 1	0.000	11:36
Breath Volume		2.988	11:36
Air Blank		0.000	11:37
Air Blank		0.000	11:39
Subject Sample	# 2	0.000	11:39
Breath Volume		2.015	11:39
Air Blank		0.000	11:40
Control Test		0.081	11:40
Air Blank		0.000	11:41
Diagnostics Check		OK	11:41

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIMPSON,STEVEN,M
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: SIMPSON,STEVEN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:38

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 12/14/2017 00:05:28

Diagnostics Check		OK	23:58
Air Blank		0.000	23:59
Control Test		0.081	23:59
Air Blank		0.000	00:00
Subject Sample	# 1	0.108	00:00
Breath Volume		1.468	00:00
Air Blank		0.000	00:01
Air Blank		0.000	00:03
Subject Sample	# 2	0.119	00:03
Breath Volume		3.125	00:03
Air Blank		0.000	00:04
Control Test		0.081	00:04
Air Blank		0.000	00:05
Diagnostics Check		OK	00:05

Data Download Date/Time: 12/18/2017 12:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATSON,DANIEL,L
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: WATSON,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 13:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:05

SUBJECT'S LAST NAME: ZANDERS

DATE/TIME OF TEST: 12/14/2017 16:36:17

Diagnostics Check		OK	16:30
Air Blank		0.000	16:30
Control Test		0.080	16:31
Air Blank		0.000	16:31
Subject Sample	# 1	0.187	16:31
Breath Volume		2.050	16:31
Air Blank		0.000	16:32
Air Blank		0.000	16:34
Subject Sample	# 2	0.179	16:34
Breath Volume		1.589	16:34
Air Blank		0.000	16:35
Control Test		0.080	16:35
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 12/18/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: MAYHUGH

DATE/TIME OF TEST: 12/15/2017 01:10:08

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:09	
Control Test		0.080	01:09	
Air Blank		0.000	01:10	
Diagnostics Check		OK	01:10	

Data Download Date/Time: 12/18/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: VENTURA-SANDAVOL

DATE/TIME OF TEST: 12/15/2017 23:37:23

Diagnostics Check		OK	23:31
Air Blank		0.000	23:31
Control Test		0.079	23:31
Air Blank		0.000	23:32
Subject Sample	# 1	0.160	23:32
Breath Volume		2.007	23:32
Air Blank		0.000	23:33
Air Blank		0.000	23:35
Subject Sample	# 2	0.161	23:35
Breath Volume		2.613	23:35
Air Blank		0.000	23:36
Control Test		0.080	23:36
Air Blank		0.000	23:37
Diagnostics Check		OK	23:37

Data Download Date/Time: 12/18/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PADOVA,BRIAN,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: PADOVA,BRIAN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: CUMMINGS

DATE/TIME OF TEST: 12/16/2017 00:44:26

Diagnostics Check		OK	00:41	
Air Blank		0.000	00:42	
Control Test		0.081	00:42	
Air Blank		0.000	00:42	
Subject Sample	# 1		00:43	Subject Test Refused
Breath Volume		0.000	00:43	
Air Blank		0.000	00:43	
Control Test		0.081	00:43	
Air Blank		0.000	00:44	
Diagnostics Check		OK	00:44	

Data Download Date/Time: 12/18/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:17

SUBJECT'S LAST NAME: GAZIL

DATE/TIME OF TEST: 12/16/2017 18:51:22

Diagnostics Check		OK	18:44
Air Blank		0.000	18:45
Control Test		0.081	18:45
Air Blank		0.000	18:45
Subject Sample	# 1	0.000	18:46
Breath Volume		1.152	18:46
Air Blank		0.000	18:47
Air Blank		0.000	18:49
Subject Sample	# 2	0.000	18:49
Breath Volume		1.203	18:49
Air Blank		0.000	18:50
Control Test		0.079	18:50
Air Blank		0.000	18:51
Diagnostics Check		OK	18:51

Data Download Date/Time: 12/18/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PETERSON,RONALD,E
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: PETERSON,RONALD
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:01

SUBJECT'S LAST NAME: KLUSON

DATE/TIME OF TEST: 12/16/2017 22:01:21

Diagnostics Check		OK	21:55	
Air Blank		0.000	21:55	
Control Test		0.081	21:56	
Air Blank		0.000	21:56	
Subject Sample	# 1	0.000	21:57	
Breath Volume		3.214	21:57	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2		21:59	Subject Test Refused
Breath Volume		0.000	21:59	
Air Blank		0.000	22:00	
Control Test		0.081	22:00	
Air Blank		0.000	22:01	
Diagnostics Check		OK	22:01	

Data Download Date/Time: 12/18/2017 12:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,B
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: HAMPTON

DATE/TIME OF TEST: 12/17/2017 01:42:40

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.080	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1	0.192	01:32	
Breath Volume		1.660	01:32	
Air Blank		0.000	01:32	
Air Blank		0.000	01:34	
Subject Sample	# 2	0.138	01:37	Volume Not Met
Breath Volume		0.156	01:37	
Air Blank		0.000	01:38	
Air Blank		0.000	01:40	
Subject Sample	# 3	0.187	01:41	
Breath Volume		1.332	01:41	
Air Blank		0.000	01:41	
Control Test		0.079	01:42	
Air Blank		0.000	01:42	
Diagnostics Check		OK	01:42	

Data Download Date/Time: 12/18/2017 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 12/20/2017 22:16:20

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.080	22:14	
Air Blank		0.000	22:14	
Subject Sample	# 1		22:14	Subject Test Refused
Breath Volume		0.000	22:14	
Air Blank		0.000	22:15	
Control Test		0.080	22:15	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/21/2017 22:07:05

Diagnostics Check		OK	22:00
Air Blank		0.000	22:01
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.091	22:02
Breath Volume		2.675	22:02
Air Blank		0.000	22:03
Air Blank		0.000	22:04
Subject Sample	# 2	0.090	22:05
Breath Volume		2.824	22:05
Air Blank		0.000	22:06
Control Test		0.080	22:06
Air Blank		0.000	22:07
Diagnostics Check		OK	22:07

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: ACOSTA

DATE/TIME OF TEST: 12/22/2017 03:25:48

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.081	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.136	03:20
Breath Volume		2.343	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.138	03:24
Breath Volume		2.308	03:24
Air Blank		0.000	03:24
Control Test		0.080	03:25
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: CANNON

DATE/TIME OF TEST: 12/23/2017 23:57:09

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.081	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.174	23:52
Breath Volume		2.777	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.171	23:55
Breath Volume		2.679	23:55
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 16:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:20

SUBJECT'S LAST NAME: MISKOWSKY

DATE/TIME OF TEST: 12/24/2017 20:07:16

Diagnostics Check		OK	20:04	
Air Blank		0.000	20:04	
Control Test		0.081	20:05	
Air Blank		0.000	20:05	
Subject Sample	# 1		20:05	Subject Test Refused
Breath Volume		0.000	20:05	
Air Blank		0.000	20:06	
Control Test		0.080	20:06	
Air Blank		0.000	20:07	
Diagnostics Check		OK	20:07	

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA-PAZ,SAMUEL,D
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: GARCIA-PAZ,SAMUEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: QUINTANILLA APARICIO

DATE/TIME OF TEST: 12/25/2017 00:33:00

Diagnostics Check		OK	00:26
Air Blank		0.000	00:26
Control Test		0.080	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.211	00:28
Breath Volume		3.164	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:30
Subject Sample	# 2	0.204	00:31
Breath Volume		3.621	00:31
Air Blank		0.000	00:32
Control Test		0.078	00:32
Air Blank		0.000	00:32
Diagnostics Check		OK	00:33

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: HERNANDEZ CALDERON

DATE/TIME OF TEST: 12/25/2017 01:33:29

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.081	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.229	01:28
Breath Volume		3.250	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.221	01:31
Breath Volume		3.511	01:31
Air Blank		0.000	01:32
Control Test		0.080	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GWODZ,OFC,R
BREATH TEST OPERATOR AGENCY: Bradenton PD
ARRESTING OFFICER: GWODZ,OFC
ARRESTING OFFICER AGENCY: Bradenton PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 03:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:56

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 12/25/2017 04:25:51

Diagnostics Check		OK	04:22	
Air Blank		0.000	04:23	
Control Test		0.080	04:23	
Air Blank		0.000	04:24	
Subject Sample	# 1		04:24	Subject Test Refused
Breath Volume		0.000	04:24	
Air Blank		0.000	04:24	
Control Test		0.080	04:25	
Air Blank		0.000	04:25	
Diagnostics Check		OK	04:25	

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,GERRY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,GERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:25

SUBJECT'S LAST NAME: JUPIN

DATE/TIME OF TEST: 12/26/2017 22:27:06

Diagnostics Check		OK	22:24	
Air Blank		0.000	22:24	
Control Test		0.081	22:25	
Air Blank		0.000	22:25	
Subject Sample	# 1		22:25	Subject Test Refused
Breath Volume		0.000	22:25	
Air Blank		0.000	22:26	
Control Test		0.081	22:26	
Air Blank		0.000	22:27	
Diagnostics Check		OK	22:27	

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,GERY,W
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,GERRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 19:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: JUPIN

DATE/TIME OF TEST: 12/26/2017 22:52:40

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.082	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.046	22:47
Breath Volume		2.304	22:47
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.044	22:51
Breath Volume		2.074	22:51
Air Blank		0.000	22:51
Control Test		0.081	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTEMAGNO,CARLO,C
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: MONTEMAGNO,CARLO
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: MCCLELLAN

DATE/TIME OF TEST: 12/27/2017 02:32:34

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.081	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.109	02:28
Breath Volume		2.589	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.108	02:31
Breath Volume		2.960	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CATHERWOOD

DATE/TIME OF TEST: 12/28/2017 01:56:20

Diagnostics Check		OK	01:40	
Air Blank		0.000	01:40	
Control Test		0.081	01:40	
Air Blank		0.000	01:41	
Subject Sample	# 1	0.140	01:43	
Breath Volume		1.429	01:43	
Air Blank		0.000	01:44	
Air Blank		0.000	01:45	
Subject Sample	# 2	0.119	01:49	Volume Not Met
Breath Volume		0.468	01:49	
Air Blank		0.000	01:49	
Air Blank		0.000	01:51	
Subject Sample	# 3	0.132	01:54	Volume Not Met
Breath Volume		0.199	01:54	
Air Blank		0.000	01:55	
Control Test		0.080	01:55	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: HERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: CATHERWOOD

DATE/TIME OF TEST: 12/28/2017 02:07:59

Diagnostics Check		OK	01:58
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.150	02:02
Breath Volume		1.847	02:02
Air Blank		0.000	02:02
Air Blank		0.000	02:04
Subject Sample	# 2	0.134	02:06
Breath Volume		1.636	02:06
Air Blank		0.000	02:07
Control Test		0.079	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: PENALOZA ANTUNEZ

DATE/TIME OF TEST: 12/28/2017 21:19:38

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:13
Subject Sample	# 1	0.184	21:14
Breath Volume		3.507	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.185	21:18
Breath Volume		2.433	21:18
Air Blank		0.000	21:18
Control Test		0.078	21:19
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAEZ,H,
BREATH TEST OPERATOR AGENCY: Palmetto PD
ARRESTING OFFICER: GONZALEZ,K
ARRESTING OFFICER AGENCY: Palmetto PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 12/30/2017 03:23:51

Diagnostics Check		OK	03:21	
Air Blank		0.000	03:21	
Control Test		0.081	03:21	
Air Blank		0.000	03:22	
Subject Sample	# 1		03:22	Subject Test Refused
Breath Volume		0.000	03:22	
Air Blank		0.000	03:22	
Control Test		0.081	03:23	
Air Blank		0.000	03:23	
Diagnostics Check		OK	03:23	

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUICK,BENJAMIN,
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: QUICK,BENJAMIN
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:13

SUBJECT'S LAST NAME: TRYGG

DATE/TIME OF TEST: 12/30/2017 03:40:01

Diagnostics Check		OK	03:33
Air Blank		0.000	03:33
Control Test		0.081	03:34
Air Blank		0.000	03:34
Subject Sample	# 1	0.156	03:35
Breath Volume		3.960	03:35
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.152	03:38
Breath Volume		3.792	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:39
Diagnostics Check		OK	03:40

Data Download Date/Time: 01/22/2018 13:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSTERMEYER,DAMIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OSTERMEYER,DAMIAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 05:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: ALBRITTON

DATE/TIME OF TEST: 12/30/2017 06:23:13

Diagnostics Check		OK	06:16
Air Blank		0.000	06:16
Control Test		0.081	06:16
Air Blank		0.000	06:17
Subject Sample	# 1	0.104	06:18
Breath Volume		1.820	06:18
Air Blank		0.000	06:19
Air Blank		0.000	06:20
Subject Sample	# 2	0.087	06:21
Breath Volume		1.488	06:21
Air Blank		0.000	06:22
Control Test		0.079	06:22
Air Blank		0.000	06:23
Diagnostics Check		OK	06:23

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OSTERMEYER,DAMIAN,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: OSTERMEYER,DAMIAN
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 06:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: MORALES CARMEN

DATE/TIME OF TEST: 12/31/2017 07:31:07

Diagnostics Check		OK	07:24
Air Blank		0.000	07:25
Control Test		0.081	07:25
Air Blank		0.000	07:26
Subject Sample	# 1	0.196	07:26
Breath Volume		2.003	07:26
Air Blank		0.000	07:27
Air Blank		0.000	07:29
Subject Sample	# 2	0.183	07:29
Breath Volume		1.914	07:29
Air Blank		0.000	07:30
Control Test		0.078	07:30
Air Blank		0.000	07:31
Diagnostics Check		OK	07:31

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006694

INSTRUMENT REGISTERED TO: MANATEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VANOVER,ANDREW,P
BREATH TEST OPERATOR AGENCY: Manatee County SO
ARRESTING OFFICER: VANOVER,ANDREW
ARRESTING OFFICER AGENCY: Manatee County SO
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 256835
DRY GAS EXPIRY DATE: 01/11/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: WENZEL

DATE/TIME OF TEST: 12/31/2017 23:59:06

Diagnostics Check		OK	23:52
Air Blank		0.000	23:53
Control Test		0.081	23:53
Air Blank		0.000	23:54
Subject Sample	# 1	0.031	23:54
Breath Volume		2.746	23:54
Air Blank		0.000	23:55
Air Blank		0.000	23:57
Subject Sample	# 2	0.030	23:57
Breath Volume		3.019	23:57
Air Blank		0.000	23:58
Control Test		0.080	23:58
Air Blank		0.000	23:59
Diagnostics Check		OK	23:59

Data Download Date/Time: 01/22/2018 13:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: PEREZ ROBLERO

DATE/TIME OF TEST: 12/09/2017 03:38:57

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.170	03:34
Breath Volume		1.957	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:36
Subject Sample	# 2	0.173	03:37
Breath Volume		3.441	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 12/18/2017 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006758

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGUSKY,DANIEL,J
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: BOGUSKY,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 123456789
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 12/22/2017 22:20:45

Diagnostics Check		OK	22:17	
Air Blank		0.000	22:17	
Control Test		0.082	22:18	
Air Blank		0.000	22:18	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank		0.000	22:19	
Control Test		0.082	22:20	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 01/22/2018 09:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: CHAPELL

DATE/TIME OF TEST: 12/17/2017 00:17:45

Diagnostics Check		OK	00:11
Air Blank		0.000	00:11
Control Test		0.083	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.159	00:12
Breath Volume		4.398	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.164	00:15
Breath Volume		4.289	00:15
Air Blank		0.000	00:16
Control Test		0.082	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 12/28/2017 08:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006760

INSTRUMENT REGISTERED TO: FHP
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWMAN,JOANNA,N
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BOWMAN,JOANNA
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: WEBER

DATE/TIME OF TEST: 12/22/2017 01:47:31

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:45	
Control Test		0.080	01:45	
Air Blank		0.000	01:45	
Subject Sample	# 1		01:46	Subject Test Refused
Breath Volume		0.000	01:46	
Air Blank		0.000	01:46	
Control Test		0.081	01:46	
Air Blank		0.000	01:47	
Diagnostics Check		OK	01:47	

Data Download Date/Time: 12/28/2017 08:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: GAMBINO

DATE/TIME OF TEST: 12/17/2017 01:17:24

Diagnostics Check		OK	01:11
Air Blank		0.000	01:11
Control Test		0.080	01:11
Air Blank		0.000	01:12
Subject Sample	# 1	0.152	01:12
Breath Volume		1.878	01:12
Air Blank		0.000	01:13
Air Blank		0.000	01:15
Subject Sample	# 2	0.147	01:15
Breath Volume		1.582	01:15
Air Blank		0.000	01:16
Control Test		0.080	01:16
Air Blank		0.000	01:17
Diagnostics Check		OK	01:17

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 12/21/2017 20:36:46

Diagnostics Check		OK	20:30	
Air Blank		0.000	20:31	
Control Test		0.079	20:31	
Air Blank		0.000	20:32	
Subject Sample	# 1	0.155	20:32	
Breath Volume		3.000	20:32	
Air Blank		0.000	20:33	
Air Blank		0.000	20:35	
Subject Sample	# 2	0.155	20:35	
Breath Volume		2.207	20:35	
Air Blank			20:36	RFI Detect
Air Blank		0.000	20:36	

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: WALTON

DATE/TIME OF TEST: 12/21/2017 20:45:32

Diagnostics Check		OK	20:38
Air Blank		0.000	20:39
Control Test		0.081	20:39
Air Blank		0.000	20:40
Subject Sample	# 1	0.140	20:40
Breath Volume		2.265	20:40
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.147	20:43
Breath Volume		2.812	20:43
Air Blank		0.000	20:44
Control Test		0.079	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PONKEY

DATE/TIME OF TEST: 12/22/2017 00:37:03

Diagnostics Check		OK	00:30
Air Blank		0.000	00:31
Control Test		0.080	00:31
Air Blank		0.000	00:32
Subject Sample	# 1	0.181	00:32
Breath Volume		2.335	00:32
Air Blank		0.000	00:33
Air Blank		0.000	00:34
Subject Sample	# 2	0.180	00:35
Breath Volume		3.007	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:37

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELSEY,RICHARD,D
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: GELSEY,RICHARD
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 21:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 12/25/2017 22:34:25

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.000	22:29
Breath Volume		2.039	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.000	22:33
Breath Volume		2.039	22:33
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 12/27/2017 03:33:48

Diagnostics Check		OK	03:30	
Air Blank		0.000	03:31	
Control Test		0.081	03:31	
Air Blank		0.000	03:31	
Subject Sample	# 1	0.000	03:32	
Breath Volume		2.503	03:32	
Air Blank			03:32	Improper Sample
Air Blank		0.000	03:32	
Control Test		0.079	03:33	
Air Blank		0.000	03:33	
Diagnostics Check		OK	03:33	

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: NAPIER

DATE/TIME OF TEST: 12/27/2017 03:41:19

Diagnostics Check		OK	03:35
Air Blank		0.000	03:35
Control Test		0.080	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.000	03:37
Breath Volume		3.371	03:37
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.000	03:39
Breath Volume		3.343	03:39
Air Blank		0.000	03:40
Control Test		0.078	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 12/30/2017 01:14:27

Diagnostics Check		OK	01:08	
Air Blank		0.000	01:09	
Control Test		0.082	01:09	
Air Blank		0.000	01:10	
Subject Sample	# 1		01:12	RFI Detect
Breath Volume		0.019	01:12	
Air Blank		0.000	01:13	
Control Test		0.078	01:13	
Air Blank		0.000	01:14	
Diagnostics Check		OK	01:14	

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:47

SUBJECT'S LAST NAME: MCGREGOR

DATE/TIME OF TEST: 12/30/2017 01:26:46

Diagnostics Check		OK	01:17	
Air Blank		0.000	01:17	
Control Test		0.079	01:17	
Air Blank		0.000	01:18	
Subject Sample	# 1	0.117	01:21	Volume Not Met
Breath Volume		0.296	01:21	
Air Blank		0.000	01:22	
Air Blank		0.000	01:23	
Subject Sample	# 2	0.122	01:25	
Breath Volume		1.406	01:25	
Air Blank			01:26	Purge Fail
Air Blank		0.000	01:26	

Data Download Date/Time: 01/12/2018 23:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006764

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONTGOMERY,KENNETH,R
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: MONTGOMERY,KENNETH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/16/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 840875
DRY GAS EXPIRY DATE: 09/19/2020
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: SONO

DATE/TIME OF TEST: 12/31/2017 03:54:46

Diagnostics Check		OK	03:48
Air Blank		0.000	03:48
Control Test		0.081	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.082	03:50
Breath Volume		2.339	03:50
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.088	03:53
Breath Volume		3.281	03:53
Air Blank		0.000	03:53
Control Test		0.081	03:54
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 01/12/2018 23:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:22

SUBJECT'S LAST NAME: BOWERS

DATE/TIME OF TEST: 12/01/2017 22:48:26

Diagnostics Check		OK	22:42
Air Blank		0.000	22:42
Control Test		0.081	22:43
Air Blank		0.000	22:43
Subject Sample	# 1	0.131	22:43
Breath Volume		2.218	22:43
Air Blank		0.000	22:44
Air Blank		0.000	22:46
Subject Sample	# 2	0.138	22:46
Breath Volume		2.984	22:46
Air Blank		0.000	22:47
Control Test		0.081	22:47
Air Blank		0.000	22:48
Diagnostics Check		OK	22:48

Data Download Date/Time: 12/15/2017 21:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: TORRES-MANZO

DATE/TIME OF TEST: 12/02/2017 01:39:49

Diagnostics Check		OK	01:32
Air Blank		0.000	01:33
Control Test		0.080	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.162	01:34
Breath Volume		1.828	01:34
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.179	01:38
Breath Volume		3.761	01:38
Air Blank		0.000	01:38
Control Test		0.080	01:39
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 12/15/2017 21:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 12/02/2017 22:54:21

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:51	
Control Test		0.077	22:51	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:52	Subject Test Refused
Breath Volume		0.000	22:52	
Air Blank		0.000	22:53	
Control Test		0.080	22:53	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 12/15/2017 21:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: WOODARD

DATE/TIME OF TEST: 12/09/2017 21:50:17

Diagnostics Check		OK	21:44
Air Blank		0.000	21:44
Control Test		0.082	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.000	21:46
Breath Volume		1.402	21:46
Air Blank		0.000	21:46
Air Blank		0.000	21:48
Subject Sample	# 2	0.000	21:48
Breath Volume		1.316	21:48
Air Blank		0.000	21:49
Control Test		0.080	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 12/15/2017 21:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: SARVEY

DATE/TIME OF TEST: 12/10/2017 01:36:42

Diagnostics Check		OK	01:30
Air Blank		0.000	01:30
Control Test		0.081	01:30
Air Blank		0.000	01:31
Subject Sample	# 1	0.190	01:32
Breath Volume		3.355	01:32
Air Blank		0.000	01:32
Air Blank		0.000	01:34
Subject Sample	# 2	0.186	01:35
Breath Volume		3.257	01:35
Air Blank		0.000	01:35
Control Test		0.081	01:36
Air Blank		0.000	01:36
Diagnostics Check		OK	01:36

Data Download Date/Time: 12/15/2017 21:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HEALY,GREGORY,B
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HEALY,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: CLARDY

DATE/TIME OF TEST: 12/16/2017 00:41:34

Diagnostics Check		OK	00:35
Air Blank		0.000	00:35
Control Test		0.081	00:36
Air Blank		0.000	00:36
Subject Sample	# 1	0.121	00:37
Breath Volume		2.843	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.113	00:40
Breath Volume		1.871	00:40
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 01/12/2018 21:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006765

INSTRUMENT REGISTERED TO: FHP TROOP G
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: KNERR

DATE/TIME OF TEST: 12/20/2017 22:04:41

Diagnostics Check		OK	21:58
Air Blank		0.000	21:58
Control Test		0.081	21:59
Air Blank		0.000	21:59
Subject Sample	# 1	0.144	22:00
Breath Volume		3.027	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.144	22:03
Breath Volume		2.960	22:03
Air Blank		0.000	22:03
Control Test		0.081	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 01/12/2018 21:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006770

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTELLO,A,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: SMITH,A
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12345678
DRY GAS EXPIRY DATE: 11/11/2018
VIOLATION CODE: DUI
ARREST TIME: 00:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:07

SUBJECT'S LAST NAME: FELIPE

DATE/TIME OF TEST: 12/17/2017 00:49:28

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.081	00:43
Air Blank		0.000	00:44
Subject Sample	# 1	0.203	00:44
Breath Volume		3.160	00:44
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.195	00:47
Breath Volume		3.019	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:48
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 01/10/2018 12:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,GABRIEL,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: ROBERSON,GABRIEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:22

SUBJECT'S LAST NAME: HATCHER

DATE/TIME OF TEST: 12/02/2017 03:54:14

Diagnostics Check		OK	03:47
Air Blank		0.000	03:48
Control Test		0.078	03:48
Air Blank		0.000	03:49
Subject Sample	# 1	0.223	03:49
Breath Volume		2.000	03:49
Air Blank		0.000	03:50
Air Blank		0.000	03:52
Subject Sample	# 2	0.232	03:52
Breath Volume		2.226	03:52
Air Blank		0.000	03:53
Control Test		0.077	03:53
Air Blank		0.000	03:54
Diagnostics Check		OK	03:54

Data Download Date/Time: 01/02/2018 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUELLO,JAVIER,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: PUELLO,JAVIER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CRIMMINS

DATE/TIME OF TEST: 12/12/2017 22:11:15

Diagnostics Check		OK	22:08	
Air Blank		0.000	22:08	
Control Test		0.080	22:09	
Air Blank		0.000	22:09	
Subject Sample	# 1		22:09	Subject Test Refused
Breath Volume		0.000	22:09	
Air Blank		0.000	22:10	
Control Test		0.080	22:10	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:11	

Data Download Date/Time: 01/02/2018 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: SYKES

DATE/TIME OF TEST: 12/20/2017 23:34:27

Diagnostics Check		OK	23:30	
Air Blank		0.000	23:31	
Control Test		0.078	23:31	
Air Blank		0.000	23:32	
Subject Sample	# 1		23:33	Subject Test Refused
Breath Volume		0.000	23:33	
Air Blank		0.000	23:33	
Control Test		0.078	23:33	
Air Blank		0.000	23:34	
Diagnostics Check		OK	23:34	

Data Download Date/Time: 01/02/2018 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLL,MICHAEL,
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: WOLL,MICHAEL
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/21/2017 23:02:38

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.079	23:00	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:01	
Control Test		0.078	23:02	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 01/02/2018 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SIRMANS,ABNER,K
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: SIRMANS,ABNER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: HISER

DATE/TIME OF TEST: 12/23/2017 02:41:33

Diagnostics Check		OK	02:34
Air Blank		0.000	02:35
Control Test		0.076	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.149	02:36
Breath Volume		1.812	02:36
Air Blank		0.000	02:37
Air Blank		0.000	02:39
Subject Sample	# 2	0.150	02:39
Breath Volume		3.226	02:39
Air Blank		0.000	02:40
Control Test		0.075	02:40
Air Blank		0.000	02:41
Diagnostics Check		OK	02:41

Data Download Date/Time: 01/02/2018 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-006931

INSTRUMENT REGISTERED TO: FT WALTON BEACH PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUELLO,JAVIER,A
BREATH TEST OPERATOR AGENCY: Ft. Walton Beach PD
ARRESTING OFFICER: PUELLO,JAVIER
ARRESTING OFFICER AGENCY: Ft. Walton Beach PD
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: OLIVA-MUNOZ

DATE/TIME OF TEST: 12/31/2017 04:07:35

Diagnostics Check		OK	03:59	
Air Blank		0.000	03:59	
Control Test		0.079	04:00	
Air Blank		0.000	04:00	
Subject Sample	# 1	0.212	04:02	
Breath Volume		1.238	04:02	
Air Blank		0.000	04:03	
Air Blank		0.000	04:05	
Subject Sample	# 2	0.208	04:05	
Breath Volume		2.128	04:05	
Air Blank		0.000	04:06	
Control Test		0.074	04:06	Control Outside Tolerance
Air Blank		0.000	04:07	

Data Download Date/Time: 01/02/2018 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.