Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTWRIGHT JR,GARY,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: ABBOTT,RYAN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: STEGER

DATE/TIME OF TEST: 12/09/2017 02:57:24

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.079	02:51
Air Blank		0.000	02:51
Subject Sample	# 1	0.092	02:52
Breath Volume		1.511	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.089	02:55
Breath Volume		1.179	02:55
Air Blank		0.000	02:56
Control Test		0.078	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 12/27/2017 16:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSICK,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: MCLEOD,JOHN
ARRESTING OFFICER AGENCY: Wauchula PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: PORTILLO

DATE/TIME OF TEST: 12/11/2017 23:49:13

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.080	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.194	23:43
Breath Volume		2.429	23:43
Air Blank		0.000	23:44
Air Blank		0.000	23:45
Subject Sample	# 2	0.194	23:47
Breath Volume		2.597	23:47
Air Blank		0.000	23:48
Control Test		0.077	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 12/27/2017 16:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSICK,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: WHITE,KEVIN
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:28

SUBJECT'S LAST NAME: FUENTES

DATE/TIME OF TEST: 12/16/2017 03:58:04

Diagnostics Check		OK	03:48
Air Blank		0.000	03:49
Control Test		0.079	03:49
Air Blank		0.000	03:50
Subject Sample	# 1	0.236	03:52
Breath Volume		1.238	03:52
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.220	03:56
Breath Volume		2.000	03:56
Air Blank		0.000	03:57
Control Test		0.076	03:57
Air Blank		0.000	03:57
Diagnostics Check		OK	03:58

Data Download Date/Time: 12/27/2017 16:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANDA,ROBERT,W
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: HART,LYLE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 12/29/2017 00:12:34

Diagnostics Check		OK	00:09	
Air Blank		0.000	00:10	
Control Test		0.079	00:10	
Air Blank		0.000	00:11	
Subject Sample	# 1		00:11	Subject Test Refused
Breath Volume		0.000	00:11	
Air Blank		0.000	00:11	
Control Test		0.079	00:11	
Air Blank		0.000	00:12	
Diagnostics Check		OK	00:12	

Data Download Date/Time: 01/29/2018 20:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001002

INSTRUMENT REGISTERED TO: HARDEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CASSICK,PATRICIA,A
BREATH TEST OPERATOR AGENCY: Hardee County SO
ARRESTING OFFICER: HART,LYLE
ARRESTING OFFICER AGENCY: Hardee County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: MINTZ

DATE/TIME OF TEST: 12/29/2017 20:56:19

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:51	
Control Test		0.078	20:52	
Air Blank		0.000	20:52	
Subject Sample	# 1		20:54	Subject Test Refused
Breath Volume		0.000	20:54	
Air Blank		0.000	20:55	
Control Test		0.078	20:55	
Air Blank		0.000	20:56	
Diagnostics Check		OK	20:56	

Data Download Date/Time: 01/29/2018 20:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: ARMITAGE

DATE/TIME OF TEST: 12/01/2017 00:08:57

Diagnostics Check		OK	23:57	
Air Blank		0.000	23:58	
Control Test		0.080	23:58	
Air Blank		0.000	23:58	
Subject Sample	# 1	0.000	00:02	Volume Not Met
Breath Volume		0.871	00:02	
Air Blank		0.000	00:02	
Air Blank		0.000	00:04	
Subject Sample	# 2		00:07	No Sample Provided
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Control Test		0.080	00:08	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:23

SUBJECT'S LAST NAME: DENNISTON

DATE/TIME OF TEST: 12/02/2017 03:53:07

Diagnostics Check		OK	03:45
Air Blank		0.000	03:46
Control Test		0.080	03:46
Air Blank		0.000	03:46
Subject Sample	# 1	0.177	03:48
Breath Volume		1.320	03:48
Air Blank		0.000	03:49
Air Blank		0.000	03:50
Subject Sample	# 2	0.180	03:51
Breath Volume		1.195	03:51
Air Blank		0.000	03:52
Control Test		0.079	03:52
Air Blank		0.000	03:53
Diagnostics Check		OK	03:53

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEELER,BRYAN,C
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEELER,BRYAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: CALLEJA

DATE/TIME OF TEST: 12/03/2017 05:17:03

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.078	05:06
Air Blank		0.000	05:07
Subject Sample	# 1	0.202	05:10
Breath Volume		1.550	05:10
Air Blank		0.000	05:10
Air Blank		0.000	05:12
Subject Sample	# 2	0.205	05:15
Breath Volume		1.589	05:15
Air Blank		0.000	05:16
Control Test		0.078	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:17

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: MIKELL

DATE/TIME OF TEST: 12/04/2017 04:06:29

Diagnostics Check		OK	03:59
Air Blank		0.000	04:00
Control Test		0.080	04:00
Air Blank		0.000	04:00
Subject Sample	# 1	0.126	04:02
Breath Volume		1.503	04:02
Air Blank		0.000	04:02
Air Blank		0.000	04:04
Subject Sample	# 2	0.129	04:05
Breath Volume		1.367	04:05
Air Blank		0.000	04:05
Control Test		0.080	04:05
Air Blank		0.000	04:06
Diagnostics Check		OK	04:06

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WELSHANE,DEPUTY D.
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: KOCH

DATE/TIME OF TEST: 12/04/2017 05:30:10

Diagnostics Check		OK	05:20
Air Blank		0.000	05:20
Control Test		0.080	05:21
Air Blank		0.000	05:21
Subject Sample	# 1	0.149	05:23
Breath Volume		1.359	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:25
Subject Sample	# 2	0.147	05:28
Breath Volume		1.128	05:28
Air Blank		0.000	05:29
Control Test		0.081	05:29
Air Blank		0.000	05:30
Diagnostics Check		OK	05:30

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: DORSEY

DATE/TIME OF TEST: 12/06/2017 23:09:54

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:05	No Sample Provided
Breath Volume		0.000	23:05	
Air Blank		0.000	23:05	
Air Blank		0.000	23:07	
Subject Sample	# 2		23:08	Subject Test Refused
Breath Volume		0.000	23:08	
Air Blank		0.000	23:09	
Control Test		0.080	23:09	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HOSKINS

DATE/TIME OF TEST: 12/06/2017 23:43:24

Diagnostics Check		OK	23:33	
Air Blank		0.000	23:33	
Control Test		0.079	23:33	
Air Blank		0.000	23:34	
Subject Sample	# 1		23:37	No Sample Provided
Breath Volume		0.000	23:37	
Air Blank		0.000	23:37	
Air Blank		0.000	23:39	
Subject Sample	# 2		23:41	Subject Test Refused
Breath Volume		0.000	23:41	
Air Blank		0.000	23:42	
Control Test		0.081	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BARSELOU

DATE/TIME OF TEST: 12/08/2017 01:11:24

Diagnostics Check		OK	01:01
Air Blank		0.000	01:02
Control Test		0.080	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.105	01:05
Breath Volume		1.125	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:07
Subject Sample	# 2	0.103	01:09
Breath Volume		1.292	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Diagnostics Check		OK	01:11

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/08/2017 03:09:45

Diagnostics Check		OK	03:06	
Air Blank		0.000	03:07	
Control Test		0.080	03:07	
Air Blank		0.000	03:08	
Subject Sample	# 1		03:08	Subject Test Refused
Breath Volume		0.000	03:08	
Air Blank		0.000	03:08	
Control Test		0.079	03:09	
Air Blank		0.000	03:09	
Diagnostics Check		OK	03:09	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: MINGLEDORFF

DATE/TIME OF TEST: 12/08/2017 04:40:09

Diagnostics Check		OK	04:33
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.207	04:35
Breath Volume		1.148	04:35
Air Blank		0.000	04:36
Air Blank		0.000	04:37
Subject Sample	# 2	0.203	04:38
Breath Volume		1.308	04:38
Air Blank		0.000	04:39
Control Test		0.080	04:39
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: VOZDA

DATE/TIME OF TEST: 12/09/2017 02:03:16

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.080	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:01	Subject Test Refused
Breath Volume		0.000	02:01	
Air Blank		0.000	02:02	
Control Test		0.081	02:02	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:03	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: RODRIGUEZ RODRIGUEZ

DATE/TIME OF TEST: 12/09/2017 03:07:18

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:04	
Control Test		0.080	03:04	
Air Blank		0.000	03:05	
Subject Sample	# 1		03:05	Subject Test Refused
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Control Test		0.080	03:06	
Air Blank		0.000	03:07	
Diagnostics Check		OK	03:07	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DENHARD,JUSTIN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: SWARTS

DATE/TIME OF TEST: 12/09/2017 18:47:39

Diagnostics Check		OK	18:41
Air Blank		0.000	18:41
Control Test		0.081	18:42
Air Blank		0.000	18:42
Subject Sample	# 1	0.036	18:43
Breath Volume		2.699	18:43
Air Blank		0.000	18:43
Air Blank		0.000	18:45
Subject Sample	# 2	0.035	18:46
Breath Volume		1.878	18:46
Air Blank		0.000	18:46
Control Test		0.081	18:47
Air Blank		0.000	18:47
Diagnostics Check		OK	18:47

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/10/2017 04:05:55

Diagnostics Check	OK	04:04	
Air Blank	0.000	04:05	
Control Test	0.080	04:05	
Air Blank		04:05	RFI Detect
Air Blank	0.000	04:05	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGER,CHARLES,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGER,CHARLES
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/10/2017 04:16:03

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Subject Sample	# 1	0.139	04:10
Breath Volume		1.515	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:12
Subject Sample	# 2	0.157	04:14
Breath Volume		2.695	04:14
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:15
Diagnostics Check		OK	04:16

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 12/15/2017 03:16:11

Diagnostics Check		OK	03:08
Air Blank		0.000	03:09
Control Test		0.080	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.126	03:11
Breath Volume		1.257	03:11
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.142	03:14
Breath Volume		2.093	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:21

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/15/2017 19:53:43

Diagnostics Check		OK	19:42	
Air Blank		0.000	19:42	
Control Test		0.080	19:43	
Air Blank		0.000	19:43	
Subject Sample	# 1		19:46	No Sample Provided
Breath Volume		0.000	19:46	
Air Blank		0.000	19:47	
Air Blank		0.000	19:49	
Subject Sample	# 2		19:52	No Sample Provided
Breath Volume		0.000	19:52	
Air Blank		0.000	19:52	
Control Test		0.080	19:53	
Air Blank		0.000	19:53	
Diagnostics Check		OK	19:53	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: OBEROI

DATE/TIME OF TEST: 12/16/2017 00:29:31

Diagnostics Check		OK	00:26	
Air Blank		0.000	00:27	
Control Test		0.080	00:27	
Air Blank		0.000	00:27	
Subject Sample	# 1		00:28	Subject Test Refused
Breath Volume		0.000	00:28	
Air Blank		0.000	00:28	
Control Test		0.081	00:28	
Air Blank		0.000	00:29	
Diagnostics Check		OK	00:29	

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:24

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 12/16/2017 03:51:15

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.122	03:46
Breath Volume		1.859	03:46
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.125	03:49
Breath Volume		2.625	03:49
Air Blank		0.000	03:50
Control Test		0.080	03:50
Air Blank		0.000	03:51
Diagnostics Check		OK	03:51

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: LYON

DATE/TIME OF TEST: 12/17/2017 00:16:07

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:10
Air Blank		0.000	00:11
Subject Sample	# 1	0.150	00:11
Breath Volume		1.433	00:11
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.148	00:14
Breath Volume		1.570	00:14
Air Blank		0.000	00:15
Control Test		0.079	00:15
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 12/27/2017 21:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:01

SUBJECT'S LAST NAME: FILLIE

DATE/TIME OF TEST: 12/17/2017 02:24:39

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.080	02:22	
Air Blank		0.000	02:22	
Subject Sample	# 1		02:23	Subject Test Refused
Breath Volume		0.000	02:23	
Air Blank		0.000	02:23	
Control Test		0.079	02:24	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRIS,BRIDGETT,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HELLER,DARREN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 17:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:20

SUBJECT'S LAST NAME: KROUZEK

DATE/TIME OF TEST: 12/17/2017 18:48:42

Diagnostics Check		OK	18:40
Air Blank		0.000	18:41
Control Test		0.079	18:41
Air Blank		0.000	18:41
Subject Sample	# 1	0.239	18:43
Breath Volume		1.351	18:43
Air Blank		0.000	18:44
Air Blank		0.000	18:45
Subject Sample	# 2	0.237	18:47
Breath Volume		1.546	18:47
Air Blank		0.000	18:47
Control Test		0.079	18:48
Air Blank		0.000	18:48
Diagnostics Check		OK	18:48

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: BACKMAN

DATE/TIME OF TEST: 12/18/2017 00:52:20

Diagnostics Check		OK	00:49	
Air Blank		0.000	00:49	
Control Test		0.079	00:49	
Air Blank		0.000	00:50	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Control Test		0.080	00:51	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LEHR,GEORGE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 06:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:05

SUBJECT'S LAST NAME: MANDATO

DATE/TIME OF TEST: 12/19/2017 07:31:30

Diagnostics Check		OK	07:25
Air Blank		0.000	07:25
Control Test		0.079	07:26
Air Blank		0.000	07:26
Subject Sample	# 1	0.231	07:27
Breath Volume		2.851	07:27
Air Blank		0.000	07:27
Air Blank		0.000	07:29
Subject Sample	# 2	0.236	07:29
Breath Volume		2.277	07:29
Air Blank		0.000	07:30
Control Test		0.079	07:30
Air Blank		0.000	07:31
Diagnostics Check		OK	07:31

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/20/2017 05:24:24

Diagnostics Check		OK	05:18
Air Blank		0.000	05:18
Control Test		0.080	05:19
Air Blank		0.000	05:19
Subject Sample	# 1	0.325	05:19
Breath Volume		3.761	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.343	05:22
Breath Volume		3.851	05:22
Air Blank		0.000	05:23
Control Test		0.079	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:58

SUBJECT'S LAST NAME: SECOSH

DATE/TIME OF TEST: 12/20/2017 23:24:23

Diagnostics Check		OK	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.202	23:19
Breath Volume		1.730	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.199	23:22
Breath Volume		2.023	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:24
Diagnostics Check		OK	23:24

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BURKE

DATE/TIME OF TEST: 12/21/2017 01:49:47

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.081	01:43
Air Blank		0.000	01:43
Subject Sample	# 1	0.149	01:45
Breath Volume		1.804	01:45
Air Blank		0.000	01:45
Air Blank		0.000	01:47
Subject Sample	# 2	0.140	01:48
Breath Volume		1.687	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:49
Air Blank		0.000	01:49
Diagnostics Check		OK	01:49

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ELSER

DATE/TIME OF TEST: 12/21/2017 02:30:42

Diagnostics Check		OK	02:26	
Air Blank		0.000	02:26	
Control Test		0.081	02:27	
Air Blank		0.000	02:27	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:29	
Control Test		0.080	02:30	
Air Blank		0.000	02:30	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: UNRUH

DATE/TIME OF TEST: 12/21/2017 20:27:22

Diagnostics Check		OK	20:21
Air Blank		0.000	20:21
Control Test		0.080	20:22
Air Blank		0.000	20:22
Subject Sample	# 1	0.179	20:22
Breath Volume		2.031	20:22
Air Blank		0.000	20:23
Air Blank		0.000	20:25
Subject Sample	# 2	0.167	20:25
Breath Volume		1.277	20:25
Air Blank		0.000	20:26
Control Test		0.080	20:26
Air Blank		0.000	20:27
Diagnostics Check		OK	20:27

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: QUINN

DATE/TIME OF TEST: 12/22/2017 21:32:38

Diagnostics Check		OK	21:26
Air Blank		0.000	21:27
Control Test		0.079	21:27
Air Blank		0.000	21:27
Subject Sample	# 1	0.000	21:28
Breath Volume		2.191	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.000	21:31
Breath Volume		2.136	21:31
Air Blank		0.000	21:31
Control Test		0.079	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BRUNER,S
ARRESTING OFFICER AGENCY: Kenneth City PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:25

SUBJECT'S LAST NAME: BARNES JR

DATE/TIME OF TEST: 12/23/2017 05:13:56

Diagnostics Check		OK	05:07
Air Blank		0.000	05:08
Control Test		0.079	05:08
Air Blank		0.000	05:09
Subject Sample	# 1	0.204	05:09
Breath Volume		1.308	05:09
Air Blank		0.000	05:10
Air Blank		0.000	05:11
Subject Sample	# 2	0.209	05:12
Breath Volume		1.921	05:12
Air Blank		0.000	05:13
Control Test		0.080	05:13
Air Blank		0.000	05:13
Diagnostics Check		OK	05:13

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUDBRINK,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SUDBRINK,BRIAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: DUFF

DATE/TIME OF TEST: 12/24/2017 05:42:33

Diagnostics Check		OK	05:35
Air Blank		0.000	05:36
Control Test		0.080	05:36
Air Blank		0.000	05:37
Subject Sample	# 1	0.155	05:37
Breath Volume		1.179	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:40
Subject Sample	# 2	0.153	05:41
Breath Volume		1.257	05:41
Air Blank		0.000	05:41
Control Test		0.081	05:41
Air Blank		0.000	05:42
Diagnostics Check		OK	05:42

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AMATRUDA,CHRIS,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SAPANARA,MATT
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 06:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:10

SUBJECT'S LAST NAME: FLEMMING

DATE/TIME OF TEST: 12/24/2017 07:34:46

Diagnostics Check		OK	07:31	
Air Blank		0.000	07:32	
Control Test		0.081	07:32	
Air Blank		0.000	07:33	
Subject Sample	# 1		07:33	Subject Test Refused
Breath Volume		0.000	07:33	
Air Blank		0.000	07:33	
Control Test		0.080	07:34	
Air Blank		0.000	07:34	
Diagnostics Check		OK	07:34	

Data Download Date/Time: 12/27/2017 21:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:14

SUBJECT'S LAST NAME: SWINEHART

DATE/TIME OF TEST: 12/28/2017 04:41:33

Diagnostics Check		OK	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:34
Air Blank		0.000	04:35
Subject Sample	# 1	0.163	04:36
Breath Volume		1.371	04:36
Air Blank		0.000	04:37
Air Blank		0.000	04:39
Subject Sample	# 2	0.169	04:40
Breath Volume		1.960	04:40
Air Blank		0.000	04:40
Control Test		0.080	04:40
Air Blank		0.000	04:41
Diagnostics Check		OK	04:41

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: NICHOLS

DATE/TIME OF TEST: 12/28/2017 22:14:56

Diagnostics Check		OK	22:03	
Air Blank		0.000	22:04	
Control Test		0.080	22:04	
Air Blank		0.000	22:04	
Subject Sample	# 1		22:07	No Sample Provided
Breath Volume		0.000	22:07	
Air Blank		0.000	22:08	
Air Blank		0.000	22:10	
Subject Sample	# 2		22:13	No Sample Provided
Breath Volume		0.000	22:13	
Air Blank		0.000	22:14	
Control Test		0.080	22:14	
Air Blank		0.000	22:14	
Diagnostics Check		OK	22:14	

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: MIKLOS

DATE/TIME OF TEST: 12/29/2017 04:17:47

Diagnostics Check		OK	04:14	
Air Blank		0.000	04:15	
Control Test		0.080	04:15	
Air Blank		0.000	04:16	
Subject Sample	# 1		04:16	Subject Test Refused
Breath Volume		0.000	04:16	
Air Blank		0.000	04:16	
Control Test		0.080	04:17	
Air Blank		0.000	04:17	
Diagnostics Check		OK	04:17	

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILTSE,DEPUTY WILLIAM,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WILTSE,DEPUTY WILLIAM
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:53

SUBJECT'S LAST NAME: BURT

DATE/TIME OF TEST: 12/29/2017 19:18:46

Diagnostics Check		OK	19:13	
Air Blank		0.000	19:13	
Control Test		0.079	19:14	
Air Blank		0.000	19:14	
Subject Sample	# 1		19:17	Subject Test Refused
Breath Volume		0.000	19:17	
Air Blank		0.000	19:17	
Control Test		0.080	19:18	
Air Blank		0.000	19:18	
Diagnostics Check		OK	19:18	

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 12/29/2017 20:57:07

Diagnostics Check		OK	20:48	
Air Blank		0.000	20:48	
Control Test		0.080	20:49	
Air Blank		0.000	20:49	
Subject Sample	# 1	0.074	20:52	Volume Not Met
Breath Volume		0.183	20:52	
Air Blank		0.000	20:53	
Air Blank		0.000	20:55	
Subject Sample	# 2		20:55	Subject Test Refused
Breath Volume		0.000	20:55	
Air Blank		0.000	20:56	
Control Test		0.079	20:56	
Air Blank		0.000	20:57	
Diagnostics Check		OK	20:57	

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: BYCZKOWSKI

DATE/TIME OF TEST: 12/30/2017 00:01:21

Diagnostics Check		OK	23:55
Air Blank		0.000	23:55
Control Test		0.079	23:56
Air Blank		0.000	23:56
Subject Sample	# 1	0.147	23:57
Breath Volume		2.058	23:57
Air Blank		0.000	23:57
Air Blank		0.000	23:59
Subject Sample	# 2	0.128	23:59
Breath Volume		1.121	23:59
Air Blank		0.000	00:00
Control Test		0.080	00:00
Air Blank		0.000	00:01
Diagnostics Check		OK	00:01

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,JUSTIN,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/30/2017 04:07:39

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.079	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.111	04:03
Breath Volume		3.683	04:03
Air Blank		0.000	04:03
Air Blank		0.000	04:05
Subject Sample	# 2	0.105	04:06
Breath Volume		1.886	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Diagnostics Check		OK	04:07

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: BAICH

DATE/TIME OF TEST: 12/30/2017 20:31:46

Diagnostics Check		OK	20:28	
Air Blank		0.000	20:29	
Control Test		0.080	20:29	
Air Blank		0.000	20:29	
Subject Sample	# 1		20:30	Subject Test Refused
Breath Volume		0.000	20:30	
Air Blank		0.000	20:30	
Control Test		0.081	20:31	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:31	

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:28

SUBJECT'S LAST NAME: WRIGHT

DATE/TIME OF TEST: 12/31/2017 00:52:14

Diagnostics Check		OK	00:48	
Air Blank		0.000	00:48	
Control Test		0.081	00:49	
Air Blank		0.000	00:49	
Subject Sample	# 1		00:50	Subject Test Refused
Breath Volume		0.000	00:50	
Air Blank		0.000	00:51	
Control Test		0.081	00:51	
Air Blank		0.000	00:52	
Diagnostics Check		OK	00:52	

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: ALVAREZ GUERRA

DATE/TIME OF TEST: 12/31/2017 04:22:47

Diagnostics Check		OK	04:13	
Air Blank		0.000	04:13	
Control Test		0.081	04:14	
Air Blank		0.000	04:14	
Subject Sample	# 1	0.161	04:17	
Breath Volume		1.582	04:17	
Air Blank		0.000	04:18	
Air Blank		0.000	04:20	
Subject Sample	# 2		04:21	Subject Test Refused
Breath Volume		0.121	04:21	
Air Blank		0.000	04:21	
Control Test		0.079	04:22	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:22	

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANEY,DAMON,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: LANEY,DAMON
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: SIMON

DATE/TIME OF TEST: 12/31/2017 04:48:55

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Subject Sample	# 1	0.201	04:44
Breath Volume		1.773	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 2	0.200	04:47
Breath Volume		1.792	04:47
Air Blank		0.000	04:48
Control Test		0.080	04:48
Air Blank		0.000	04:48
Diagnostics Check		OK	04:48

Data Download Date/Time: 01/25/2018 22:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001003

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:54

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/31/2017 23:21:06

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.153	23:16
Breath Volume		2.726	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.153	23:19
Breath Volume		2.550	23:19
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 01/25/2018 22:25

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JOHNSON,BRUCE,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JOHNSON,BRUCE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:31

SUBJECT'S LAST NAME: STEC

DATE/TIME OF TEST: 12/02/2017 19:57:22

Diagnostics Check		OK	19:54	
Air Blank		0.000	19:54	
Control Test		0.079	19:55	
Air Blank		0.000	19:55	
Subject Sample	# 1		19:56	Subject Test Refused
Breath Volume		0.000	19:56	
Air Blank		0.000	19:56	
Control Test		0.080	19:56	
Air Blank		0.000	19:57	
Diagnostics Check		OK	19:57	

Data Download Date/Time: 12/27/2017 20:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BERGSTROM,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BERGSTROM,S
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: LYNCH

DATE/TIME OF TEST: 12/04/2017 00:03:45

Diagnostics Check		OK	23:56
Air Blank		0.000	23:56
Control Test		0.080	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.183	23:58
Breath Volume		1.199	23:58
Air Blank		0.000	23:59
Air Blank		0.000	00:01
Subject Sample	# 1	0.163	00:02
Breath Volume		1.988	00:02
Air Blank		0.000	00:02
Control Test		0.079	00:03
Air Blank		0.000	00:03
Diagnostics Check		OK	00:03

Data Download Date/Time: 12/27/2017 20:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SZYMANSKI,JACEK,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SZYMANSKI,JACEK
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: BRETT

DATE/TIME OF TEST: 12/20/2017 03:14:28

Diagnostics Check		OK	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.191	03:10
Breath Volume		3.164	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.184	03:12
Breath Volume		1.250	03:12
Air Blank		0.000	03:13
Control Test		0.080	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 12/27/2017 20:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: REYES

DATE/TIME OF TEST: 12/21/2017 04:00:14

Diagnostics Check		OK	03:57	
Air Blank		0.000	03:57	
Control Test		0.080	03:57	
Air Blank		0.000	03:58	
Subject Sample	# 1		03:58	Subject Test Refused
Breath Volume		0.000	03:58	
Air Blank		0.000	03:59	
Control Test		0.080	03:59	
Air Blank		0.000	04:00	
Diagnostics Check		OK	04:00	

Data Download Date/Time: 12/27/2017 20:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLAIRE,J,L
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: SCHUTZ,T
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 12/25/2017 19:31:07

Diagnostics Check		OK	19:28	
Air Blank		0.000	19:28	
Control Test		0.080	19:29	
Air Blank		0.000	19:29	
Subject Sample	# 1		19:29	Subject Test Refused
Breath Volume		0.000	19:29	
Air Blank		0.000	19:30	
Control Test		0.080	19:30	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:31	

Data Download Date/Time: 12/27/2017 20:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001005

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BREGON,JAMES,S
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BREGON,JAMES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 07:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:35

SUBJECT'S LAST NAME: HUFF

DATE/TIME OF TEST: 12/27/2017 08:18:30

Diagnostics Check		OK	08:04	
Air Blank		0.000	08:05	
Control Test		0.081	08:05	
Air Blank		0.000	08:06	
Subject Sample	# 1	0.111	08:09	Volume Not Met
Breath Volume		1.074	08:09	
Air Blank		0.000	08:10	
Air Blank		0.000	08:11	
Subject Sample	# 2	0.112	08:13	
Breath Volume		1.589	08:13	
Air Blank		0.000	08:14	
Air Blank		0.000	08:16	
Subject Sample	# 3	0.125	08:16	
Breath Volume		2.707	08:16	
Air Blank		0.000	08:17	
Control Test		0.080	08:17	
Air Blank		0.000	08:18	
Diagnostics Check		OK	08:18	

Data Download Date/Time: 12/27/2017 20:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELTZER,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MELTZER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 12:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:58

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 12/20/2017 13:24:56

Diagnostics Check		OK	13:18
Air Blank		0.000	13:19
Control Test		0.082	13:19
Air Blank		0.000	13:19
Subject Sample	# 1	0.293	13:20
Breath Volume		1.816	13:20
Air Blank		0.000	13:21
Air Blank		0.000	13:22
Subject Sample	# 2	0.281	13:23
Breath Volume		1.273	13:23
Air Blank		0.000	13:24
Control Test		0.081	13:24
Air Blank		0.000	13:24
Diagnostics Check		OK	13:24

Data Download Date/Time: 12/26/2017 07:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001007

INSTRUMENT REGISTERED TO: HAINES CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGER,KEVIN,M
BREATH TEST OPERATOR AGENCY: Haines City PD
ARRESTING OFFICER: ARMBRUSTER,HANAH
ARRESTING OFFICER AGENCY: Haines City PD
LAST AGENCY INSPECTION: 11/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:15

SUBJECT'S LAST NAME: JOLY

DATE/TIME OF TEST: 12/23/2017 06:41:42

Diagnostics Check		OK	06:35
Air Blank		0.000	06:35
Control Test		0.082	06:36
Air Blank		0.000	06:36
Subject Sample	# 1	0.124	06:37
Breath Volume		2.464	06:37
Air Blank		0.000	06:38
Air Blank		0.000	06:39
Subject Sample	# 2	0.122	06:40
Breath Volume		1.140	06:40
Air Blank		0.000	06:40
Control Test		0.080	06:41
Air Blank		0.000	06:41
Diagnostics Check		OK	06:41

Data Download Date/Time: 12/26/2017 07:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABU SOUD,AHMAD,A
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RUTHERFORD,CHRIS
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:10

SUBJECT'S LAST NAME: ADAMS

DATE/TIME OF TEST: 12/17/2017 03:37:25

Diagnostics Check		OK	03:30
Air Blank		0.000	03:30
Control Test		0.080	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.247	03:32
Breath Volume		1.363	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 2	0.238	03:35
Breath Volume		1.835	03:35
Air Blank		0.000	03:36
Control Test		0.079	03:36
Air Blank		0.000	03:37
Diagnostics Check		OK	03:37

Data Download Date/Time: 12/25/2017 12:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001008

INSTRUMENT REGISTERED TO: WEST MELBOURNE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,MARK,E
BREATH TEST OPERATOR AGENCY: West Melbourne PD
ARRESTING OFFICER: RUTHERFORD,OFC
ARRESTING OFFICER AGENCY: West Melbourne PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:23

SUBJECT'S LAST NAME: ARTHUR

DATE/TIME OF TEST: 12/20/2017 18:02:52

Diagnostics Check		OK	17:56
Air Blank		0.000	17:57
Control Test		0.079	17:57
Air Blank		0.000	17:57
Subject Sample	# 1	0.000	17:58
Breath Volume		1.687	17:58
Air Blank		0.000	17:58
Air Blank		0.000	18:00
Subject Sample	# 2	0.000	18:01
Breath Volume		4.375	18:01
Air Blank		0.000	18:01
Control Test		0.078	18:02
Air Blank		0.000	18:02
Diagnostics Check		OK	18:02

Data Download Date/Time: 12/25/2017 12:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: PARTEE

DATE/TIME OF TEST: 12/19/2017 02:40:35

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:29	
Control Test		0.079	02:29	
Air Blank		0.000	02:30	
Subject Sample	# 1	0.066	02:33	Volume Not Met
Breath Volume		0.574	02:33	
Air Blank		0.000	02:33	
Air Blank		0.000	02:35	
Subject Sample	# 2	0.074	02:38	Volume Not Met
Breath Volume		0.925	02:38	
Air Blank		0.000	02:39	
Control Test		0.078	02:39	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 12/26/2017 18:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: PARTEE

DATE/TIME OF TEST: 12/19/2017 02:49:38

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.079	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.081	02:44
Breath Volume		2.183	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.083	02:48
Breath Volume		2.480	02:48
Air Blank		0.000	02:48
Control Test		0.078	02:49
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 12/26/2017 18:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 12/19/2017 05:26:47

Diagnostics Check		OK	05:10	
Air Blank		0.000	05:11	
Control Test		0.080	05:11	
Air Blank		0.000	05:11	
Subject Sample	# 1	0.157	05:13	
Breath Volume		1.113	05:13	
Air Blank		0.000	05:14	
Air Blank		0.000	05:16	
Subject Sample	# 2	0.142	05:19	Volume Not Met
Breath Volume		0.433	05:19	
Air Blank		0.000	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 3	0.167	05:25	Volume Not Met
Breath Volume		0.390	05:25	
Air Blank		0.000	05:25	
Control Test		0.078	05:26	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 12/26/2017 18:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: GILMORE

DATE/TIME OF TEST: 12/19/2017 05:36:27

Diagnostics Check		OK	05:28
Air Blank		0.000	05:29
Control Test		0.079	05:29
Air Blank		0.000	05:30
Subject Sample	# 1	0.165	05:31
Breath Volume		2.062	05:31
Air Blank		0.000	05:32
Air Blank		0.000	05:33
Subject Sample	# 2	0.159	05:34
Breath Volume		1.730	05:34
Air Blank		0.000	05:35
Control Test		0.078	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 12/26/2017 18:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MCNAMARA

DATE/TIME OF TEST: 12/24/2017 22:04:58

Diagnostics Check		OK	21:57
Air Blank		0.000	21:57
Control Test		0.080	21:58
Air Blank		0.000	21:58
Subject Sample	# 1	0.252	22:00
Breath Volume		1.488	22:00
Air Blank		0.000	22:00
Air Blank		0.000	22:02
Subject Sample	# 2	0.255	22:03
Breath Volume		2.105	22:03
Air Blank		0.000	22:04
Control Test		0.080	22:04
Air Blank		0.000	22:04
Diagnostics Check		OK	22:04

Data Download Date/Time: 12/26/2017 18:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001009

INSTRUMENT REGISTERED TO: PALM BAY P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,J,W
BREATH TEST OPERATOR AGENCY: Palm Bay PD
ARRESTING OFFICER: HAMILTON,J
ARRESTING OFFICER AGENCY: Palm Bay PD
LAST AGENCY INSPECTION: 12/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:03

SUBJECT'S LAST NAME: LANGSTON

DATE/TIME OF TEST: 12/30/2017 02:30:36

Diagnostics Check		OK	02:23
Air Blank		0.000	02:24
Control Test		0.080	02:24
Air Blank		0.000	02:25
Subject Sample	# 1	0.139	02:26
Breath Volume		2.328	02:26
Air Blank		0.000	02:26
Air Blank		0.000	02:28
Subject Sample	# 2	0.125	02:29
Breath Volume		1.441	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 01/30/2018 14:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MURPHY,J
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: DION

DATE/TIME OF TEST: 12/01/2017 03:01:14

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.079	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.163	02:56
Breath Volume		1.550	02:56
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.164	02:59
Breath Volume		2.546	02:59
Air Blank		0.000	03:00
Control Test		0.077	03:00
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FARLEY,J
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:41

SUBJECT'S LAST NAME: BONDARENKO

DATE/TIME OF TEST: 12/01/2017 04:10:34

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.078	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.109	04:05
Breath Volume		1.664	04:05
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.105	04:08
Breath Volume		1.308	04:08
Air Blank		0.000	04:09
Control Test		0.078	04:09
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/01/2017 05:25:27

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.079	05:20
Air Blank		0.000	05:20
Subject Sample	# 1	0.000	05:21
Breath Volume		2.652	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.000	05:24
Breath Volume		2.121	05:24
Air Blank		0.000	05:24
Control Test		0.078	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:21

SUBJECT'S LAST NAME: ABDELL

DATE/TIME OF TEST: 12/02/2017 02:44:43

Diagnostics Check		OK	02:41	
Air Blank		0.000	02:42	
Control Test		0.079	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 1		02:43	Subject Test Refused
Breath Volume		0.000	02:43	
Air Blank		0.000	02:43	
Control Test		0.078	02:44	
Air Blank		0.000	02:44	
Diagnostics Check		OK	02:44	

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: OMERBASHI

DATE/TIME OF TEST: 12/02/2017 03:53:49

Diagnostics Check		OK	03:50	
Air Blank		0.000	03:51	
Control Test		0.079	03:51	
Air Blank		0.000	03:52	
Subject Sample	# 1		03:52	Subject Test Refused
Breath Volume		0.000	03:52	
Air Blank		0.000	03:52	
Control Test		0.078	03:53	
Air Blank		0.000	03:53	
Diagnostics Check		OK	03:53	

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: WALLACE

DATE/TIME OF TEST: 12/02/2017 04:57:18

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.078	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.197	04:52
Breath Volume		2.453	04:52
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.199	04:55
Breath Volume		2.046	04:55
Air Blank		0.000	04:56
Control Test		0.078	04:56
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: KIMBERLY

DATE/TIME OF TEST: 12/02/2017 23:33:08

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.061	23:28
Breath Volume		3.367	23:28
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.059	23:31
Breath Volume		2.527	23:31
Air Blank		0.000	23:32
Control Test		0.078	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: GORDON

DATE/TIME OF TEST: 12/03/2017 00:50:04

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.212	00:45
Breath Volume		2.855	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.211	00:48
Breath Volume		2.265	00:48
Air Blank		0.000	00:49
Control Test		0.077	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:50

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: SCHWEGMAN

DATE/TIME OF TEST: 12/03/2017 02:11:24

Diagnostics Check		OK	02:04
Air Blank		0.000	02:04
Control Test		0.079	02:05
Air Blank		0.000	02:05
Subject Sample	# 1	0.140	02:06
Breath Volume		1.664	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.138	02:09
Breath Volume		1.492	02:09
Air Blank		0.000	02:10
Control Test		0.078	02:10
Air Blank		0.000	02:11
Diagnostics Check		OK	02:11

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DEMPSEY

DATE/TIME OF TEST: 12/03/2017 03:30:43

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.124	03:26
Breath Volume		2.410	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.114	03:29
Breath Volume		1.335	03:29
Air Blank		0.000	03:29
Control Test		0.077	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: DINKINS

DATE/TIME OF TEST: 12/03/2017 04:18:38

Diagnostics Check		OK	04:15	
Air Blank		0.000	04:16	
Control Test		0.078	04:16	
Air Blank		0.000	04:17	
Subject Sample	# 1		04:17	Subject Test Refused
Breath Volume		0.000	04:17	
Air Blank		0.000	04:17	
Control Test		0.078	04:18	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 12/11/2017 15:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 03:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: DINKINS

DATE/TIME OF TEST: 12/03/2017 05:07:19

Diagnostics Check		OK	05:00
Air Blank		0.000	05:00
Control Test		0.079	05:01
Air Blank		0.000	05:01
Subject Sample	# 1	0.055	05:02
Breath Volume		1.667	05:02
Air Blank		0.000	05:03
Air Blank		0.000	05:05
Subject Sample	# 2	0.054	05:05
Breath Volume		1.664	05:05
Air Blank		0.000	05:06
Control Test		0.079	05:06
Air Blank		0.000	05:07
Diagnostics Check		OK	05:07

Data Download Date/Time: 12/11/2017 15:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:16

SUBJECT'S LAST NAME: PURGASON

DATE/TIME OF TEST: 12/03/2017 05:42:20

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:39	
Control Test		0.079	05:40	
Air Blank		0.000	05:40	
Subject Sample	# 1		05:40	Subject Test Refused
Breath Volume		0.000	05:40	
Air Blank		0.000	05:41	
Control Test		0.078	05:41	
Air Blank		0.000	05:42	
Diagnostics Check		OK	05:42	

Data Download Date/Time: 12/11/2017 15:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: MATZKE

DATE/TIME OF TEST: 12/03/2017 23:16:20

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:13	
Control Test		0.079	23:14	
Air Blank		0.000	23:14	
Subject Sample	# 1		23:14	Subject Test Refused
Breath Volume		0.000	23:14	
Air Blank		0.000	23:15	
Control Test		0.079	23:15	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 12/11/2017 15:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: MORALES ROLDAN

DATE/TIME OF TEST: 12/04/2017 01:13:40

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.080	01:11	
Air Blank		0.000	01:12	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Control Test		0.079	01:13	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:36

SUBJECT'S LAST NAME: CARIAS

DATE/TIME OF TEST: 12/04/2017 03:04:50

Diagnostics Check		OK	02:56
Air Blank		0.000	02:57
Control Test		0.079	02:57
Air Blank		0.000	02:58
Subject Sample	# 1	0.144	02:59
Breath Volume		2.378	02:59
Air Blank		0.000	03:00
Air Blank		0.000	03:02
Subject Sample	# 2	0.141	03:03
Breath Volume		2.855	03:03
Air Blank		0.000	03:03
Control Test		0.078	03:04
Air Blank		0.000	03:04
Diagnostics Check		OK	03:04

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: MEMOALI

DATE/TIME OF TEST: 12/04/2017 05:26:05

Diagnostics Check		OK	05:23	
Air Blank		0.000	05:23	
Control Test		0.079	05:23	
Air Blank		0.000	05:24	
Subject Sample	# 1		05:24	Subject Test Refused
Breath Volume		0.000	05:24	
Air Blank		0.000	05:25	
Control Test		0.079	05:25	
Air Blank		0.000	05:26	
Diagnostics Check		OK	05:26	

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: DEHNER

DATE/TIME OF TEST: 12/05/2017 01:30:10

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.087	01:25
Breath Volume		1.414	01:25
Air Blank		0.000	01:26
Air Blank		0.000	01:27
Subject Sample	# 2	0.085	01:28
Breath Volume		2.289	01:28
Air Blank		0.000	01:29
Control Test		0.078	01:29
Air Blank		0.000	01:30
Diagnostics Check		OK	01:30

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIGHTSEY,ROBERT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DAVENPORT,R
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: HUNTER

DATE/TIME OF TEST: 12/05/2017 02:46:50

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.080	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Control Test		0.078	02:46	
Air Blank		0.000	02:46	
Diagnostics Check		OK	02:46	

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 12/08/2017 02:36:40

Diagnostics Check		OK	02:29
Air Blank		0.000	02:29
Control Test		0.080	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.185	02:31
Breath Volume		1.421	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.187	02:34
Breath Volume		1.378	02:34
Air Blank		0.000	02:35
Control Test		0.077	02:36
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BUDNIK

DATE/TIME OF TEST: 12/09/2017 22:53:34

Diagnostics Check	OK	22:52	
Air Blank		22:53	Ambient Fail
Air Blank	0.000	22:53	

Data Download Date/Time: 12/11/2017 15:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: BUDNIK

DATE/TIME OF TEST: 12/09/2017 23:03:58

Diagnostics Check		OK	22:57
Air Blank		0.000	22:57
Control Test		0.079	22:58
Air Blank		0.000	22:58
Subject Sample	# 1	0.138	22:59
Breath Volume		3.203	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:01
Subject Sample	# 2	0.134	23:02
Breath Volume		2.667	23:02
Air Blank		0.000	23:03
Control Test		0.078	23:03
Air Blank		0.000	23:03
Diagnostics Check		OK	23:03

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: ADERMAN

DATE/TIME OF TEST: 12/10/2017 02:23:54

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.078	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.137	02:18
Breath Volume		2.593	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.131	02:22
Breath Volume		2.027	02:22
Air Blank		0.000	02:22
Control Test		0.078	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:31

SUBJECT'S LAST NAME: RIVERA MORALES

DATE/TIME OF TEST: 12/10/2017 04:09:44

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:08	
Subject Sample	# 1		04:08	Subject Test Refused
Breath Volume		0.000	04:08	
Air Blank		0.000	04:08	
Control Test		0.079	04:09	
Air Blank		0.000	04:09	
Diagnostics Check		OK	04:09	

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:18

SUBJECT'S LAST NAME: BATTLES

DATE/TIME OF TEST: 12/10/2017 05:44:38

Diagnostics Check		OK	05:38
Air Blank		0.000	05:38
Control Test		0.080	05:39
Air Blank		0.000	05:39
Subject Sample	# 1	0.073	05:40
Breath Volume		2.351	05:40
Air Blank		0.000	05:40
Air Blank		0.000	05:42
Subject Sample	# 2	0.067	05:43
Breath Volume		2.398	05:43
Air Blank		0.000	05:43
Control Test		0.080	05:43
Air Blank		0.000	05:44
Diagnostics Check		OK	05:44

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WOLFORD,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:30

SUBJECT'S LAST NAME: FAUSKEE

DATE/TIME OF TEST: 12/10/2017 09:57:05

Diagnostics Check		OK	09:50
Air Blank		0.000	09:50
Control Test		0.080	09:51
Air Blank		0.000	09:51
Subject Sample	# 1	0.292	09:52
Breath Volume		2.480	09:52
Air Blank		0.000	09:53
Air Blank		0.000	09:54
Subject Sample	# 2	0.287	09:55
Breath Volume		3.109	09:55
Air Blank		0.000	09:56
Control Test		0.078	09:56
Air Blank		0.000	09:56
Diagnostics Check		OK	09:57

Data Download Date/Time: 12/11/2017 15:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 12/14/2017 00:22:20

Diagnostics Check		OK	00:16
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Subject Sample	# 1	0.187	00:17
Breath Volume		2.191	00:17
Air Blank		0.000	00:18
Air Blank		0.000	00:20
Subject Sample	# 2	0.168	00:20
Breath Volume		1.796	00:20
Air Blank		0.000	00:21
Control Test		0.079	00:21
Air Blank		0.000	00:22
Diagnostics Check		OK	00:22

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRADDAM,H
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/14/2017 02:36:02

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:30
Subject Sample	# 1	0.214	02:31
Breath Volume		2.371	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:33
Subject Sample	# 2	0.208	02:34
Breath Volume		2.425	02:34
Air Blank		0.000	02:35
Control Test		0.078	02:35
Air Blank		0.000	02:35
Diagnostics Check		OK	02:36

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:22

SUBJECT'S LAST NAME: HODGES

DATE/TIME OF TEST: 12/20/2017 00:49:40

Diagnostics Check		OK	00:43
Air Blank		0.000	00:43
Control Test		0.079	00:44
Air Blank		0.000	00:44
Subject Sample	# 1	0.242	00:45
Breath Volume		2.503	00:45
Air Blank		0.000	00:45
Air Blank		0.000	00:47
Subject Sample	# 2	0.254	00:47
Breath Volume		3.113	00:47
Air Blank		0.000	00:48
Control Test		0.078	00:49
Air Blank		0.000	00:49
Diagnostics Check		OK	00:49

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: BURLS

DATE/TIME OF TEST: 12/21/2017 01:16:45

Diagnostics Check		OK	01:09
Air Blank		0.000	01:10
Control Test		0.079	01:10
Air Blank		0.000	01:11
Subject Sample	# 1	0.107	01:11
Breath Volume		2.085	01:11
Air Blank		0.000	01:12
Air Blank		0.000	01:14
Subject Sample	# 2	0.107	01:15
Breath Volume		2.882	01:15
Air Blank		0.000	01:15
Control Test		0.078	01:16
Air Blank		0.000	01:16
Diagnostics Check		OK	01:16

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: THOMPSON,M
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: COBB

DATE/TIME OF TEST: 12/21/2017 03:07:16

Diagnostics Check		OK	02:58
Air Blank		0.000	02:59
Control Test		0.079	02:59
Air Blank		0.000	03:00
Subject Sample	# 1	0.121	03:02
Breath Volume		2.007	03:02
Air Blank		0.000	03:02
Air Blank		0.000	03:04
Subject Sample	# 2	0.123	03:05
Breath Volume		2.199	03:05
Air Blank		0.000	03:06
Control Test		0.077	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:59

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 12/21/2017 04:26:33

Diagnostics Check		OK	04:19
Air Blank		0.000	04:20
Control Test		0.079	04:20
Air Blank		0.000	04:20
Subject Sample	# 1	0.141	04:21
Breath Volume		1.375	04:21
Air Blank		0.000	04:22
Air Blank		0.000	04:23
Subject Sample	# 2	0.142	04:24
Breath Volume		2.253	04:24
Air Blank		0.000	04:25
Control Test		0.078	04:25
Air Blank		0.000	04:26
Diagnostics Check		OK	04:26

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLARIDA,K,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 10:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:00

SUBJECT'S LAST NAME: MARION

DATE/TIME OF TEST: 12/21/2017 13:27:37

Diagnostics Check		OK	13:20
Air Blank		0.000	13:21
Control Test		0.078	13:21
Air Blank		0.000	13:22
Subject Sample	# 1	0.302	13:22
Breath Volume		1.753	13:22
Air Blank		0.000	13:23
Air Blank		0.000	13:25
Subject Sample	# 2	0.310	13:25
Breath Volume		2.152	13:25
Air Blank		0.000	13:26
Control Test		0.078	13:26
Air Blank		0.000	13:27
Diagnostics Check		OK	13:27

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: GORDEN JR

DATE/TIME OF TEST: 12/21/2017 23:31:52

Diagnostics Check		OK	23:25
Air Blank		0.000	23:25
Control Test		0.079	23:25
Air Blank		0.000	23:26
Subject Sample	# 1	0.213	23:27
Breath Volume		2.964	23:27
Air Blank		0.000	23:27
Air Blank		0.000	23:29
Subject Sample	# 2	0.202	23:30
Breath Volume		2.531	23:30
Air Blank		0.000	23:30
Control Test		0.078	23:31
Air Blank		0.000	23:31
Diagnostics Check		OK	23:31

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SIRMONS

DATE/TIME OF TEST: 12/22/2017 03:14:38

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:09
Subject Sample	# 1	0.108	03:09
Breath Volume		1.492	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.110	03:12
Breath Volume		1.816	03:12
Air Blank		0.000	03:13
Control Test		0.078	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: COHILL,D
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:13

SUBJECT'S LAST NAME: CLEARY

DATE/TIME OF TEST: 12/22/2017 06:40:14

Diagnostics Check		OK	06:33
Air Blank		0.000	06:34
Control Test		0.079	06:34
Air Blank		0.000	06:35
Subject Sample	# 1	0.131	06:35
Breath Volume		3.515	06:35
Air Blank		0.000	06:36
Air Blank		0.000	06:38
Subject Sample	# 2	0.129	06:38
Breath Volume		3.238	06:38
Air Blank		0.000	06:39
Control Test		0.078	06:39
Air Blank		0.000	06:40
Diagnostics Check		OK	06:40

Data Download Date/Time: 12/29/2017 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAKER,TYLER,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 12/29/2017 21:58:32

Diagnostics Check		OK	21:50
Air Blank		0.000	21:51
Control Test		0.079	21:51
Air Blank		0.000	21:51
Subject Sample	# 1	0.176	21:52
Breath Volume		3.851	21:52
Air Blank		0.000	21:53
Air Blank		0.000	21:55
Subject Sample	# 2	0.179	21:56
Breath Volume		3.269	21:56
Air Blank		0.000	21:57
Control Test		0.079	21:57
Air Blank		0.000	21:58
Diagnostics Check		OK	21:58

Data Download Date/Time: 01/19/2018 14:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: LATHAM

DATE/TIME OF TEST: 12/31/2017 02:45:22

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.079	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.090	02:40
Breath Volume		2.289	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.091	02:43
Breath Volume		2.632	02:43
Air Blank		0.000	02:44
Control Test		0.079	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 01/19/2018 14:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001018

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: CERRTO

DATE/TIME OF TEST: 12/31/2017 05:00:20

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:57	
Control Test		0.080	04:58	
Air Blank		0.000	04:58	
Subject Sample	# 1		04:58	Subject Test Refused
Breath Volume		0.000	04:58	
Air Blank		0.000	04:59	
Control Test		0.079	04:59	
Air Blank		0.000	05:00	
Diagnostics Check		OK	05:00	

Data Download Date/Time: 01/19/2018 14:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 15:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 12/08/2017 15:51:03

Diagnostics Check		OK	15:45
Air Blank		0.000	15:45
Control Test		0.078	15:45
Air Blank		0.000	15:46
Subject Sample	# 1	0.000	15:46
Breath Volume		1.707	15:46
Air Blank		0.000	15:47
Air Blank		0.000	15:49
Subject Sample	# 2	0.000	15:49
Breath Volume		1.617	15:49
Air Blank		0.000	15:50
Control Test		0.078	15:50
Air Blank		0.000	15:50
Diagnostics Check		OK	15:51

Data Download Date/Time: 12/30/2017 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: ABELN,SHON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 15:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 12/08/2017 16:01:50

Diagnostics Check		OK	15:55
Air Blank		0.000	15:56
Control Test		0.078	15:56
Air Blank		0.000	15:57
Subject Sample	# 1	0.000	15:57
Breath Volume		2.835	15:57
Air Blank		0.000	15:58
Air Blank		0.000	16:00
Subject Sample	# 2	0.000	16:00
Breath Volume		2.726	16:00
Air Blank		0.000	16:00
Control Test		0.078	16:01
Air Blank		0.000	16:01
Diagnostics Check		OK	16:01

Data Download Date/Time: 12/30/2017 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SANDERS

DATE/TIME OF TEST: 12/22/2017 02:44:03

Diagnostics Check		OK	02:37
Air Blank		0.000	02:37
Control Test		0.077	02:38
Air Blank		0.000	02:38
Subject Sample	# 1	0.145	02:39
Breath Volume		1.769	02:39
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.141	02:42
Breath Volume		2.289	02:42
Air Blank		0.000	02:43
Control Test		0.077	02:43
Air Blank		0.000	02:43
Diagnostics Check		OK	02:44

Data Download Date/Time: 12/30/2017 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MCLELLAN,LARRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 12/31/2017 21:42:36

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:39	
Control Test		0.078	21:39	
Air Blank		0.000	21:40	
Subject Sample	# 1	0.000	21:41	
Breath Volume		4.335	21:41	
Air Blank			21:41	RFI Detect
Air Blank		0.000	21:41	
Control Test		0.078	21:42	
Air Blank		0.000	21:42	
Diagnostics Check		OK	21:42	

Data Download Date/Time: 01/27/2018 04:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001043

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MCLELLAN,LARRY
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: DUNCAN

DATE/TIME OF TEST: 12/31/2017 21:54:45

Diagnostics Check		OK	21:48
Air Blank		0.000	21:49
Control Test		0.078	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.000	21:50
Breath Volume		4.382	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.000	21:53
Breath Volume		3.488	21:53
Air Blank		0.000	21:53
Control Test		0.077	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 01/27/2018 04:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: MENENDEZ

DATE/TIME OF TEST: 12/03/2017 23:43:36

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.079	23:37
Air Blank		0.000	23:38
Subject Sample	# 1	0.169	23:39
Breath Volume		2.167	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.162	23:42
Breath Volume		2.101	23:42
Air Blank		0.000	23:42
Control Test		0.079	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 12/27/2017 15:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: DRAKE

DATE/TIME OF TEST: 12/17/2017 02:19:21

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.124	02:12
Breath Volume		1.281	02:12
Air Blank		0.000	02:13
Air Blank		0.000	02:15
Subject Sample	# 2	0.114	02:17
Breath Volume		1.171	02:17
Air Blank		0.000	02:18
Control Test		0.080	02:18
Air Blank		0.000	02:19
Diagnostics Check		OK	02:19

Data Download Date/Time: 12/27/2017 15:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: LORENZ

DATE/TIME OF TEST: 12/18/2017 02:21:13

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:19	RFI Detect
Breath Volume		0.152	02:19	
Air Blank		0.000	02:20	
Control Test		0.081	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 12/27/2017 15:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: LORENZ

DATE/TIME OF TEST: 12/18/2017 02:35:46

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:23	
Control Test		0.079	02:23	
Air Blank		0.000	02:24	
Subject Sample	# 1	0.089	02:27	Volume Not Met
Breath Volume		0.125	02:27	
Air Blank		0.000	02:28	
Air Blank		0.000	02:29	
Subject Sample	# 2	0.098	02:31	
Breath Volume		1.171	02:31	
Air Blank		0.000	02:31	
Air Blank		0.000	02:33	
Subject Sample	# 3	0.098	02:34	
Breath Volume		1.617	02:34	
Air Blank		0.000	02:34	
Control Test		0.078	02:35	
Air Blank		0.000	02:35	
Diagnostics Check		OK	02:35	

Data Download Date/Time: 12/27/2017 15:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GIOVINCO

DATE/TIME OF TEST: 12/19/2017 02:26:52

Diagnostics Check		OK	02:16	
Air Blank		0.000	02:17	
Control Test		0.079	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 1		02:21	No Sample Provided
Breath Volume		0.000	02:21	
Air Blank		0.000	02:21	
Air Blank		0.000	02:23	
Subject Sample	# 2		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:25	
Control Test		0.080	02:26	
Air Blank		0.000	02:26	
Diagnostics Check		OK	02:26	

Data Download Date/Time: 12/27/2017 15:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001044

INSTRUMENT REGISTERED TO: COCONUT CREEK PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESCOBAR,ALEJANDRO,
BREATH TEST OPERATOR AGENCY: Coconut Creek Pub Sf
ARRESTING OFFICER: ESCOBAR,ALEJANDRO
ARRESTING OFFICER AGENCY: Coconut Creek Pub Sf
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: GIOVINCO

DATE/TIME OF TEST: 12/19/2017 02:40:46

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.080	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.071	02:35
Breath Volume		1.738	02:35
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.073	02:39
Breath Volume		3.136	02:39
Air Blank		0.000	02:39
Control Test		0.080	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 12/27/2017 15:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SCOTT,A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A3
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: STRICKLAND

DATE/TIME OF TEST: 12/21/2017 20:56:45

Diagnostics Check		OK	20:50
Air Blank		0.000	20:50
Control Test		0.080	20:51
Air Blank		0.000	20:51
Subject Sample	# 1	0.144	20:52
Breath Volume		2.375	20:52
Air Blank		0.000	20:52
Air Blank		0.000	20:54
Subject Sample	# 2	0.141	20:55
Breath Volume		2.058	20:55
Air Blank		0.000	20:55
Control Test		0.079	20:56
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A3
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/21/2017 22:16:57

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.079	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.172	22:12
Breath Volume		2.648	22:12
Air Blank		0.000	22:13
Air Blank		0.000	22:14
Subject Sample	# 2	0.170	22:15
Breath Volume		2.179	22:15
Air Blank		0.000	22:16
Control Test		0.079	22:16
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: CALLAO

DATE/TIME OF TEST: 12/22/2017 02:50:46

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.095	02:46
Breath Volume		1.882	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:48
Subject Sample	# 2	0.094	02:49
Breath Volume		2.597	02:49
Air Blank		0.000	02:49
Control Test		0.079	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: BUTLER

DATE/TIME OF TEST: 12/22/2017 04:15:00

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.079	04:09
Air Blank		0.000	04:10
Subject Sample	# 1	0.078	04:10
Breath Volume		3.867	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.076	04:13
Breath Volume		3.941	04:13
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:15

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DIAZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:18

SUBJECT'S LAST NAME: HOLMES

DATE/TIME OF TEST: 12/22/2017 06:45:07

Diagnostics Check		OK	06:38
Air Blank		0.000	06:39
Control Test		0.079	06:39
Air Blank		0.000	06:40
Subject Sample	# 1	0.129	06:40
Breath Volume		1.546	06:40
Air Blank		0.000	06:41
Air Blank		0.000	06:42
Subject Sample	# 2	0.123	06:43
Breath Volume		1.664	06:43
Air Blank		0.000	06:44
Control Test		0.079	06:44
Air Blank		0.000	06:45
Diagnostics Check		OK	06:45

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: BARRIE

DATE/TIME OF TEST: 12/23/2017 00:40:37

Diagnostics Check		OK	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.334	00:36
Breath Volume		3.550	00:36
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.329	00:38
Breath Volume		2.738	00:38
Air Blank		0.000	00:39
Control Test		0.078	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/23/2017 03:16:01

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.079	03:10
Air Blank		0.000	03:10
Subject Sample	# 1	0.166	03:11
Breath Volume		2.503	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:13
Subject Sample	# 2	0.164	03:14
Breath Volume		1.886	03:14
Air Blank		0.000	03:15
Control Test		0.079	03:15
Air Blank		0.000	03:15
Diagnostics Check		OK	03:16

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/23/2017 04:03:33

Diagnostics Check		OK	04:00	
Air Blank		0.000	04:01	
Control Test		0.079	04:01	
Air Blank		0.000	04:02	
Subject Sample	# 1		04:02	Subject Test Refused
Breath Volume		0.000	04:02	
Air Blank		0.000	04:02	
Control Test		0.079	04:02	
Air Blank		0.000	04:03	
Diagnostics Check		OK	04:03	

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: MORALESPEREZ

DATE/TIME OF TEST: 12/24/2017 02:03:58

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.207	01:59
Breath Volume		2.300	01:59
Air Blank		0.000	02:00
Air Blank		0.000	02:01
Subject Sample	# 2	0.202	02:02
Breath Volume		2.148	02:02
Air Blank		0.000	02:03
Control Test		0.079	02:03
Air Blank		0.000	02:03
Diagnostics Check		OK	02:03

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MORGAN,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:02

SUBJECT'S LAST NAME: AKPABIO

DATE/TIME OF TEST: 12/24/2017 05:28:25

Diagnostics Check		OK	05:22
Air Blank		0.000	05:22
Control Test		0.079	05:22
Air Blank		0.000	05:23
Subject Sample	# 1	0.137	05:23
Breath Volume		1.679	05:23
Air Blank		0.000	05:24
Air Blank		0.000	05:26
Subject Sample	# 2	0.136	05:26
Breath Volume		1.312	05:26
Air Blank		0.000	05:27
Control Test		0.079	05:27
Air Blank		0.000	05:28
Diagnostics Check		OK	05:28

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: WHEELER

DATE/TIME OF TEST: 12/25/2017 23:12:05

Diagnostics Check		OK	23:05
Air Blank		0.000	23:06
Control Test		0.081	23:06
Air Blank		0.000	23:07
Subject Sample	# 1	0.114	23:07
Breath Volume		2.085	23:07
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.110	23:10
Breath Volume		3.695	23:10
Air Blank		0.000	23:11
Control Test		0.079	23:11
Air Blank		0.000	23:11
Diagnostics Check		OK	23:12

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,M
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: SAULSBERRY

DATE/TIME OF TEST: 12/26/2017 02:37:06

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:32
Subject Sample	# 1	0.132	02:32
Breath Volume		2.757	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.128	02:35
Breath Volume		2.996	02:35
Air Blank		0.000	02:36
Control Test		0.079	02:36
Air Blank		0.000	02:37
Diagnostics Check		OK	02:37

Data Download Date/Time: 12/29/2017 14:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:32

SUBJECT'S LAST NAME: WINGATE

DATE/TIME OF TEST: 12/26/2017 04:59:30

Diagnostics Check		OK	04:52
Air Blank		0.000	04:53
Control Test		0.080	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.182	04:54
Breath Volume		2.164	04:54
Air Blank		0.000	04:55
Air Blank		0.000	04:56
Subject Sample	# 2	0.173	04:57
Breath Volume		1.707	04:57
Air Blank		0.000	04:58
Control Test		0.079	04:58
Air Blank		0.000	04:59
Diagnostics Check		OK	04:59

Data Download Date/Time: 12/29/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 12/27/2017 01:55:47

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.152	01:51
Breath Volume		1.410	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.162	01:54
Breath Volume		1.593	01:54
Air Blank		0.000	01:54
Control Test		0.079	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 12/29/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 12/27/2017 03:40:38

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.177	03:36
Breath Volume		1.156	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.162	03:39
Breath Volume		2.250	03:39
Air Blank		0.000	03:39
Control Test		0.080	03:40
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 12/29/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/28/2017 01:43:38

Diagnostics Check		OK	01:37
Air Blank		0.000	01:37
Control Test		0.079	01:38
Air Blank		0.000	01:38
Subject Sample	# 1	0.115	01:39
Breath Volume		1.175	01:39
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.123	01:42
Breath Volume		1.285	01:42
Air Blank		0.000	01:42
Control Test		0.079	01:43
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 12/29/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001045

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: TRIPPETT

DATE/TIME OF TEST: 12/28/2017 03:13:42

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.079	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.132	03:09
Breath Volume		1.105	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.129	03:12
Breath Volume		1.156	03:12
Air Blank		0.000	03:12
Control Test		0.080	03:13
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 12/29/2017 14:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: LAFOND

DATE/TIME OF TEST: 12/09/2017 02:25:46

Diagnostics Check		OK	02:18
Air Blank		0.000	02:18
Control Test		0.080	02:19
Air Blank		0.000	02:19
Subject Sample	# 1	0.173	02:20
Breath Volume		4.609	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:23
Subject Sample	# 2	0.178	02:24
Breath Volume		4.597	02:24
Air Blank		0.000	02:24
Control Test		0.080	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 12/20/2017 10:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROJAS,BRIENNA,R
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: ROJAS,BRIENNA
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:56

SUBJECT'S LAST NAME: KAPLAN

DATE/TIME OF TEST: 12/14/2017 15:27:30

Diagnostics Check		OK	15:17
Air Blank		0.000	15:18
Control Test		0.080	15:18
Air Blank		0.000	15:19
Subject Sample	# 1	0.245	15:21
Breath Volume		1.671	15:21
Air Blank		0.000	15:22
Air Blank		0.000	15:23
Subject Sample	# 2	0.245	15:25
Breath Volume		1.394	15:25
Air Blank		0.000	15:26
Control Test		0.078	15:26
Air Blank		0.000	15:27
Diagnostics Check		OK	15:27

Data Download Date/Time: 12/20/2017 10:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTOM,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: WETHERINGTON,STEPHEN
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:37

SUBJECT'S LAST NAME: OMALLEY

DATE/TIME OF TEST: 12/27/2017 22:08:17

Diagnostics Check		OK	22:02
Air Blank		0.000	22:02
Control Test		0.080	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.196	22:03
Breath Volume		2.289	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.182	22:06
Breath Volume		1.699	22:06
Air Blank		0.000	22:07
Control Test		0.080	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 01/30/2018 13:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001050

INSTRUMENT REGISTERED TO: CORAL SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WETHERINGTON,STEPHEN,L
BREATH TEST OPERATOR AGENCY: Coral Springs PD
ARRESTING OFFICER: BEST,RICH
ARRESTING OFFICER AGENCY: Coral Springs PD
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: DUREMY

DATE/TIME OF TEST: 12/28/2017 03:28:07

Diagnostics Check		OK	03:20
Air Blank		0.000	03:21
Control Test		0.080	03:21
Air Blank		0.000	03:22
Subject Sample	# 1	0.199	03:23
Breath Volume		1.968	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:25
Subject Sample	# 2	0.186	03:26
Breath Volume		2.167	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 01/30/2018 13:51

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: BORLAND

DATE/TIME OF TEST: 12/03/2017 02:17:03

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.191	02:12
Breath Volume		3.082	02:12
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.171	02:15
Breath Volume		1.285	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:16
Diagnostics Check		OK	02:17

Data Download Date/Time: 12/14/2017 07:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,D,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 12/03/2017 23:27:36

Diagnostics Check		OK	23:19
Air Blank		0.000	23:20
Control Test		0.081	23:20
Air Blank		0.000	23:21
Subject Sample	# 1	0.195	23:23
Breath Volume		2.300	23:23
Air Blank		0.000	23:23
Air Blank		0.000	23:25
Subject Sample	# 2	0.186	23:26
Breath Volume		2.023	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Diagnostics Check		OK	23:27

Data Download Date/Time: 12/14/2017 07:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FEDDERLY,GARRETT,A
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DOUGLAS,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:43

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/09/2017 13:18:05

Diagnostics Check		OK	13:11
Air Blank		0.000	13:12
Control Test		0.081	13:12
Air Blank		0.000	13:13
Subject Sample	# 1	0.182	13:13
Breath Volume		3.132	13:13
Air Blank		0.000	13:14
Air Blank		0.000	13:16
Subject Sample	# 2	0.182	13:16
Breath Volume		2.683	13:16
Air Blank		0.000	13:17
Control Test		0.080	13:17
Air Blank		0.000	13:17
Diagnostics Check		OK	13:18

Data Download Date/Time: 12/14/2017 07:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNTON,N,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: DELEON,L
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:58

SUBJECT'S LAST NAME: BEGAN

DATE/TIME OF TEST: 12/10/2017 22:32:51

Diagnostics Check		OK	22:30	
Air Blank		0.000	22:30	
Control Test		0.081	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 1		22:31	Subject Test Refused
Breath Volume		0.000	22:31	
Air Blank		0.000	22:31	
Control Test		0.081	22:32	
Air Blank		0.000	22:32	
Diagnostics Check		OK	22:32	

Data Download Date/Time: 12/14/2017 07:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,SHANNON,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: GORDAN,G
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/11/2017 18:06:43

Diagnostics Check		OK	17:58
Air Blank		0.000	17:59
Control Test		0.080	17:59
Air Blank		0.000	17:59
Subject Sample	# 1	0.225	18:00
Breath Volume		1.464	18:00
Air Blank		0.000	18:00
Air Blank		0.000	18:02
Subject Sample	# 2	0.228	18:04
Breath Volume		1.199	18:04
Air Blank		0.000	18:05
Control Test		0.081	18:06
Air Blank		0.000	18:06
Diagnostics Check		OK	18:06

Data Download Date/Time: 12/14/2017 07:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,ROBERT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 15:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: HUDDLESTUN

DATE/TIME OF TEST: 12/12/2017 17:33:48

Diagnostics Check		OK	17:27
Air Blank		0.000	17:27
Control Test		0.081	17:27
Air Blank		0.000	17:28
Subject Sample	# 1	0.213	17:29
Breath Volume		3.222	17:29
Air Blank		0.000	17:30
Air Blank		0.000	17:31
Subject Sample	# 2	0.213	17:32
Breath Volume		3.265	17:32
Air Blank		0.000	17:32
Control Test		0.080	17:33
Air Blank		0.000	17:33
Diagnostics Check		OK	17:33

Data Download Date/Time: 12/14/2017 07:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: PALECEK,D
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: MOWERY

DATE/TIME OF TEST: 12/15/2017 23:57:59

Diagnostics Check		OK	23:51
Air Blank		0.000	23:52
Control Test		0.082	23:52
Air Blank		0.000	23:53
Subject Sample	# 1	0.143	23:53
Breath Volume		1.871	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.135	23:56
Breath Volume		1.367	23:56
Air Blank		0.000	23:57
Control Test		0.082	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 01/21/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,C,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: JUAREZ LIEVANO

DATE/TIME OF TEST: 12/17/2017 05:22:45

Diagnostics Check		OK	05:19	
Air Blank		0.000	05:20	
Control Test		0.080	05:20	
Air Blank		0.000	05:21	
Subject Sample	# 1		05:21	Subject Test Refused
Breath Volume		0.000	05:21	
Air Blank		0.000	05:21	
Control Test		0.082	05:22	
Air Blank		0.000	05:22	
Diagnostics Check		OK	05:22	

Data Download Date/Time: 01/21/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILL,T,V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ENNIS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 12/19/2017 21:05:37

Diagnostics Check		OK	21:02	
Air Blank		0.000	21:03	
Control Test		0.081	21:03	
Air Blank		0.000	21:04	
Subject Sample	# 1		21:04	Subject Test Refused
Breath Volume		0.000	21:04	
Air Blank		0.000	21:04	
Control Test		0.081	21:05	
Air Blank		0.000	21:05	
Diagnostics Check		OK	21:05	

Data Download Date/Time: 01/21/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:51

SUBJECT'S LAST NAME: COYLE

DATE/TIME OF TEST: 12/20/2017 04:21:09

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:12	
Control Test		0.080	04:12	
Air Blank		0.000	04:13	
Subject Sample	# 1	0.160	04:16	Volume Not Met
Breath Volume		0.121	04:16	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2		04:19	Subject Test Refused
Breath Volume		0.527	04:19	
Air Blank		0.000	04:20	
Control Test		0.079	04:20	
Air Blank		0.000	04:21	
Diagnostics Check		OK	04:21	

Data Download Date/Time: 01/21/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNTON,N,
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: THORNTON,N
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: ADMIN
ARREST TIME: 17:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: GARRIGUS

DATE/TIME OF TEST: 12/22/2017 19:07:22

Diagnostics Check		OK	19:01
Air Blank		0.000	19:01
Control Test		0.080	19:02
Air Blank		0.000	19:02
Subject Sample	# 1	0.242	19:03
Breath Volume		1.636	19:03
Air Blank		0.000	19:03
Air Blank		0.000	19:05
Subject Sample	# 2	0.242	19:05
Breath Volume		1.925	19:05
Air Blank		0.000	19:06
Control Test		0.080	19:06
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 01/21/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GILL,T,V
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: CREEL,S
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: GOODSON

DATE/TIME OF TEST: 12/22/2017 21:06:46

Diagnostics Check		OK	20:57
Air Blank		0.000	20:58
Control Test		0.080	20:58
Air Blank		0.000	20:59
Subject Sample	# 1	0.206	21:02
Breath Volume		1.824	21:02
Air Blank		0.000	21:02
Air Blank		0.000	21:04
Subject Sample	# 2	0.205	21:05
Breath Volume		1.175	21:05
Air Blank		0.000	21:05
Control Test		0.079	21:06
Air Blank		0.000	21:06
Diagnostics Check		OK	21:06

Data Download Date/Time: 01/21/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HIRES,J,L
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: ENNIS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: ROBBINS

DATE/TIME OF TEST: 12/22/2017 23:22:28

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:20	
Control Test		0.080	23:20	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:21	
Control Test		0.081	23:21	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 01/21/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THORNTON,N,E
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: HUNTER,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:56

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 12/23/2017 01:21:45

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.082	01:19	
Air Blank		0.000	01:20	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:20	
Control Test		0.081	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 01/21/2018 08:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOOLARD,RAYMOND,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: JACKSON,ROBERT
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:12

SUBJECT'S LAST NAME: WHEELDON

DATE/TIME OF TEST: 12/30/2017 16:56:57

Diagnostics Check		OK	16:50
Air Blank		0.000	16:51
Control Test		0.081	16:51
Air Blank		0.000	16:51
Subject Sample	# 1	0.103	16:52
Breath Volume		2.109	16:52
Air Blank		0.000	16:53
Air Blank		0.000	16:55
Subject Sample	# 2	0.106	16:55
Breath Volume		2.144	16:55
Air Blank		0.000	16:56
Control Test		0.080	16:56
Air Blank		0.000	16:56
Diagnostics Check		OK	16:56

Data Download Date/Time: 01/21/2018 08:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOWEN,JUSTIN,W
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: BOWEN,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:13

SUBJECT'S LAST NAME: KUKLINSKI

DATE/TIME OF TEST: 12/30/2017 17:37:13

Diagnostics Check		OK	17:33	
Air Blank		0.000	17:33	
Control Test		0.081	17:34	
Air Blank		0.000	17:34	
Subject Sample	# 1		17:35	Subject Test Refused
Breath Volume		0.000	17:35	
Air Blank		0.000	17:36	
Control Test		0.081	17:36	
Air Blank		0.000	17:37	
Diagnostics Check		OK	17:37	

Data Download Date/Time: 01/21/2018 08:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARDONA,K,C
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: REYNOLDS,M
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:15

SUBJECT'S LAST NAME: WIGGINS

DATE/TIME OF TEST: 12/31/2017 01:43:01

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.080	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.191	01:38
Breath Volume		3.019	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:40
Subject Sample	# 2	0.173	01:41
Breath Volume		2.187	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:43

Data Download Date/Time: 01/21/2018 08:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEBB,C,D
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MOYERS,J
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 12/31/2017 04:52:07

Diagnostics Check		OK	04:46
Air Blank		0.000	04:46
Control Test		0.081	04:46
Air Blank		0.000	04:47
Subject Sample	# 1	0.187	04:47
Breath Volume		2.343	04:47
Air Blank		0.000	04:48
Air Blank		0.000	04:50
Subject Sample	# 2	0.184	04:50
Breath Volume		2.265	04:50
Air Blank		0.000	04:51
Control Test		0.081	04:51
Air Blank		0.000	04:52
Diagnostics Check		OK	04:52

Data Download Date/Time: 01/21/2018 08:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001053

INSTRUMENT REGISTERED TO: NASSAU CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRICE,SHANNON,R
BREATH TEST OPERATOR AGENCY: Nassau County SO
ARRESTING OFFICER: MORGAN,W
ARRESTING OFFICER AGENCY: Nassau County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:47

SUBJECT'S LAST NAME: LAWLER

DATE/TIME OF TEST: 12/31/2017 07:23:12

Diagnostics Check		OK	07:16
Air Blank		0.000	07:16
Control Test		0.081	07:17
Air Blank		0.000	07:17
Subject Sample	# 1	0.120	07:18
Breath Volume		2.105	07:18
Air Blank		0.000	07:19
Air Blank		0.000	07:21
Subject Sample	# 2	0.113	07:21
Breath Volume		2.125	07:21
Air Blank		0.000	07:22
Control Test		0.081	07:22
Air Blank		0.000	07:23
Diagnostics Check		OK	07:23

Data Download Date/Time: 01/21/2018 08:43

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: DOUBERLEY,CHAD
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:48

SUBJECT'S LAST NAME: BENIER

DATE/TIME OF TEST: 12/02/2017 03:11:05

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.079	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.080	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 12/30/2017 08:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KEEFE,JAYME,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: KEEFE,JAYME
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 15:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 12/02/2017 15:51:59

Diagnostics Check		OK	15:46
Air Blank		0.000	15:46
Control Test		0.079	15:46
Air Blank		0.000	15:47
Subject Sample	# 1	0.000	15:47
Breath Volume		1.730	15:47
Air Blank		0.000	15:48
Air Blank		0.000	15:50
Subject Sample	# 2	0.000	15:50
Breath Volume		1.679	15:50
Air Blank		0.000	15:51
Control Test		0.080	15:51
Air Blank		0.000	15:51
Diagnostics Check		OK	15:51

Data Download Date/Time: 12/30/2017 08:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: CLOUD,M
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: VANDE HEI

DATE/TIME OF TEST: 12/04/2017 01:21:53

Diagnostics Check		OK	01:18	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:19	
Subject Sample	# 1		01:20	Subject Test Refused
Breath Volume		0.000	01:20	
Air Blank		0.000	01:20	
Control Test		0.080	01:21	
Air Blank		0.000	01:21	
Diagnostics Check		OK	01:21	

Data Download Date/Time: 12/30/2017 08:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: ZERO TOL
ARREST TIME: 15:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:30

SUBJECT'S LAST NAME: STEINER

DATE/TIME OF TEST: 12/06/2017 15:57:22

Diagnostics Check		OK	15:50
Air Blank		0.000	15:51
Control Test		0.080	15:51
Air Blank		0.000	15:52
Subject Sample	# 1	0.000	15:52
Breath Volume		3.171	15:52
Air Blank		0.000	15:53
Air Blank		0.000	15:55
Subject Sample	# 2	0.000	15:56
Breath Volume		3.300	15:56
Air Blank		0.000	15:56
Control Test		0.081	15:56
Air Blank		0.000	15:57
Diagnostics Check		OK	15:57

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLMES,K,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: SWAN

DATE/TIME OF TEST: 12/07/2017 03:09:36

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.080	03:04
Air Blank		0.000	03:04
Subject Sample	# 1	0.135	03:05
Breath Volume		2.402	03:05
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.127	03:08
Breath Volume		2.125	03:08
Air Blank		0.000	03:08
Control Test		0.080	03:09
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARIN,ADRIAN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:17

SUBJECT'S LAST NAME: KENWORTHY

DATE/TIME OF TEST: 12/09/2017 08:45:13

Diagnostics Check		OK	08:39
Air Blank		0.000	08:39
Control Test		0.081	08:39
Air Blank		0.000	08:40
Subject Sample	# 1	0.000	08:40
Breath Volume		2.128	08:40
Air Blank		0.000	08:41
Air Blank		0.000	08:43
Subject Sample	# 2	0.000	08:43
Breath Volume		1.988	08:43
Air Blank		0.000	08:44
Control Test		0.079	08:44
Air Blank		0.000	08:45
Diagnostics Check		OK	08:45

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: WELLS,BRANDON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:42

SUBJECT'S LAST NAME: PRICE

DATE/TIME OF TEST: 12/09/2017 20:06:12

Diagnostics Check		OK	20:02	
Air Blank		0.000	20:03	
Control Test		0.081	20:03	
Air Blank		0.000	20:04	
Subject Sample	# 1		20:04	Subject Test Refused
Breath Volume		0.000	20:04	
Air Blank		0.000	20:05	
Control Test		0.081	20:05	
Air Blank		0.000	20:06	
Diagnostics Check		OK	20:06	

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRACKETT,MARK,O
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: BRACKETT,MARK
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: ZERO TOL
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 12/11/2017 11:36:40

Diagnostics Check		OK	11:30
Air Blank		0.000	11:31
Control Test		0.080	11:31
Air Blank		0.000	11:31
Subject Sample	# 1	0.000	11:32
Breath Volume		2.664	11:32
Air Blank		0.000	11:32
Air Blank		0.000	11:34
Subject Sample	# 2	0.000	11:35
Breath Volume		2.753	11:35
Air Blank		0.000	11:35
Control Test		0.082	11:36
Air Blank		0.000	11:36
Diagnostics Check		OK	11:36

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,COREY,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SPENCER,S
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: PLUMMER

DATE/TIME OF TEST: 12/12/2017 02:22:19

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.081	02:20	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:21	
Control Test		0.081	02:21	
Air Blank		0.000	02:22	
Diagnostics Check		OK	02:22	

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:02

SUBJECT'S LAST NAME: PADBERG

DATE/TIME OF TEST: 12/12/2017 03:29:07

Diagnostics Check		OK	03:23
Air Blank		0.000	03:23
Control Test		0.081	03:23
Air Blank		0.000	03:24
Subject Sample	# 1	0.125	03:24
Breath Volume		1.921	03:24
Air Blank		0.000	03:25
Air Blank		0.000	03:27
Subject Sample	# 2	0.116	03:27
Breath Volume		2.003	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: MARLOW,ROYCE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: OTHER
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: BOYD

DATE/TIME OF TEST: 12/15/2017 15:00:00

Diagnostics Check		OK	14:54
Air Blank		0.000	14:54
Control Test		0.080	14:54
Air Blank		0.000	14:55
Subject Sample	# 1	0.000	14:55
Breath Volume		1.621	14:55
Air Blank		0.000	14:56
Air Blank		0.000	14:58
Subject Sample	# 2	0.000	14:58
Breath Volume		1.519	14:58
Air Blank		0.000	14:59
Control Test		0.081	14:59
Air Blank		0.000	14:59
Diagnostics Check		OK	15:00

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,E
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:49

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 12/16/2017 01:13:31

Diagnostics Check		OK	01:10	
Air Blank		0.000	01:11	
Control Test		0.080	01:11	
Air Blank		0.000	01:11	
Subject Sample	# 1		01:12	Subject Test Refused
Breath Volume		0.000	01:12	
Air Blank		0.000	01:12	
Control Test		0.079	01:12	
Air Blank		0.000	01:13	
Diagnostics Check		OK	01:13	

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/16/2017 06:24:00

Diagnostics Check		OK	06:16
Air Blank		0.000	06:17
Control Test		0.081	06:17
Air Blank		0.000	06:18
Subject Sample	# 1	0.145	06:19
Breath Volume		1.687	06:19
Air Blank		0.000	06:19
Air Blank		0.000	06:21
Subject Sample	# 2	0.143	06:22
Breath Volume		1.300	06:22
Air Blank		0.000	06:23
Control Test		0.080	06:23
Air Blank		0.000	06:23
Diagnostics Check		OK	06:24

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINISH,BARRY,A
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: HINISH,BARRY
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/16/2017 19:22:16

Diagnostics Check		OK	19:16
Air Blank		0.000	19:16
Control Test		0.079	19:16
Air Blank		0.000	19:17
Subject Sample	# 1	0.000	19:18
Breath Volume		2.996	19:18
Air Blank		0.000	19:18
Air Blank		0.000	19:20
Subject Sample	# 2	0.000	19:20
Breath Volume		2.414	19:20
Air Blank		0.000	19:21
Control Test		0.080	19:21
Air Blank		0.000	19:22
Diagnostics Check		OK	19:22

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMILTON,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: GENTRY,W
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: BRYAN

DATE/TIME OF TEST: 12/17/2017 22:54:17

Diagnostics Check		OK	22:51	
Air Blank		0.000	22:51	
Control Test		0.080	22:52	
Air Blank		0.000	22:52	
Subject Sample	# 1		22:52	Subject Test Refused
Breath Volume		0.000	22:52	
Air Blank		0.000	22:53	
Control Test		0.080	22:53	
Air Blank		0.000	22:54	
Diagnostics Check		OK	22:54	

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ABELN,SHON,W
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: GAMMAGE

DATE/TIME OF TEST: 12/18/2017 07:28:10

Diagnostics Check		OK	07:21
Air Blank		0.000	07:22
Control Test		0.080	07:22
Air Blank		0.000	07:23
Subject Sample	# 1	0.000	07:23
Breath Volume		1.218	07:23
Air Blank		0.000	07:24
Air Blank		0.000	07:26
Subject Sample	# 2	0.000	07:26
Breath Volume		1.203	07:26
Air Blank		0.000	07:27
Control Test		0.080	07:27
Air Blank		0.000	07:28
Diagnostics Check		OK	07:28

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:22

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/18/2017 13:52:33

Diagnostics Check		OK	13:46
Air Blank		0.000	13:47
Control Test		0.079	13:47
Air Blank		0.000	13:47
Subject Sample	# 1	0.000	13:48
Breath Volume		2.125	13:48
Air Blank		0.000	13:48
Air Blank		0.000	13:50
Subject Sample	# 2	0.000	13:51
Breath Volume		2.445	13:51
Air Blank		0.000	13:51
Control Test		0.081	13:51
Air Blank		0.000	13:52
Diagnostics Check		OK	13:52

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBLES,J,
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: NONE,NONE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:30

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 12/18/2017 14:59:06

Diagnostics Check		OK	14:52
Air Blank		0.000	14:53
Control Test		0.080	14:53
Air Blank		0.000	14:54
Subject Sample	# 1	0.000	14:54
Breath Volume		2.859	14:54
Air Blank		0.000	14:55
Air Blank		0.000	14:57
Subject Sample	# 2	0.000	14:57
Breath Volume		2.851	14:57
Air Blank		0.000	14:58
Control Test		0.081	14:58
Air Blank		0.000	14:59
Diagnostics Check		OK	14:59

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RHODEN,DAVID
ARRESTING OFFICER AGENCY: Lake Placid PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: FIALHO

DATE/TIME OF TEST: 12/19/2017 02:31:03

Diagnostics Check		OK	02:21
Air Blank		0.000	02:22
Control Test		0.079	02:22
Air Blank		0.000	02:23
Subject Sample	# 1	0.128	02:26
Breath Volume		1.781	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:28
Subject Sample	# 2	0.116	02:29
Breath Volume		1.453	02:29
Air Blank		0.000	02:30
Control Test		0.078	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:31

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNELL,C,J
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SNELL,C
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 12:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/21/2017 13:02:43

Diagnostics Check		OK	12:56
Air Blank		0.000	12:57
Control Test		0.080	12:57
Air Blank		0.000	12:57
Subject Sample	# 1	0.000	12:58
Breath Volume		2.273	12:58
Air Blank		0.000	12:58
Air Blank		0.000	13:00
Subject Sample	# 2	0.000	13:01
Breath Volume		1.339	13:01
Air Blank		0.000	13:01
Control Test		0.080	13:02
Air Blank		0.000	13:02
Diagnostics Check		OK	13:02

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROE,ROBERT,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SPENCER,SHANE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: OLIVAS

DATE/TIME OF TEST: 12/22/2017 04:13:21

Diagnostics Check		OK	04:06
Air Blank		0.000	04:06
Control Test		0.080	04:07
Air Blank		0.000	04:07
Subject Sample	# 1	0.127	04:08
Breath Volume		2.449	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:10
Subject Sample	# 2	0.130	04:11
Breath Volume		2.226	04:11
Air Blank		0.000	04:12
Control Test		0.080	04:12
Air Blank		0.000	04:13
Diagnostics Check		OK	04:13

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PECK,JOHN,R
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:30

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/24/2017 17:12:19

Diagnostics Check		OK	17:05
Air Blank		0.000	17:06
Control Test		0.080	17:06
Air Blank		0.000	17:06
Subject Sample	# 1	0.000	17:07
Breath Volume		2.195	17:07
Air Blank		0.000	17:08
Air Blank		0.000	17:10
Subject Sample	# 2	0.000	17:10
Breath Volume		2.414	17:10
Air Blank		0.000	17:11
Control Test		0.081	17:11
Air Blank		0.000	17:12
Diagnostics Check		OK	17:12

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,I
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SMITH,KEVIN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: PROB/PAROL
ARREST TIME: 17:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:45

SUBJECT'S LAST NAME: GRIFFIN

DATE/TIME OF TEST: 12/24/2017 18:13:30

Diagnostics Check		OK	18:07
Air Blank		0.000	18:08
Control Test		0.081	18:08
Air Blank		0.000	18:08
Subject Sample	# 1	0.000	18:09
Breath Volume		3.160	18:09
Air Blank		0.000	18:09
Air Blank		0.000	18:11
Subject Sample	# 2	0.000	18:12
Breath Volume		3.875	18:12
Air Blank		0.000	18:12
Control Test		0.080	18:12
Air Blank		0.000	18:13
Diagnostics Check		OK	18:13

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SAVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: VAN DER LIKE

DATE/TIME OF TEST: 12/25/2017 21:08:13

Diagnostics Check		OK	21:01
Air Blank		0.000	21:02
Control Test		0.080	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.216	21:03
Breath Volume		1.500	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:05
Subject Sample	# 2	0.210	21:06
Breath Volume		1.941	21:06
Air Blank		0.000	21:07
Control Test		0.081	21:07
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALTON,LATOYA,D
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PECK,CHARLIE
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: CRAWFORD

DATE/TIME OF TEST: 12/27/2017 20:12:22

Diagnostics Check		OK	20:05
Air Blank		0.000	20:06
Control Test		0.079	20:06
Air Blank		0.000	20:07
Subject Sample	# 1	0.256	20:07
Breath Volume		1.945	20:07
Air Blank		0.000	20:08
Air Blank		0.000	20:10
Subject Sample	# 2	0.252	20:10
Breath Volume		1.984	20:10
Air Blank		0.000	20:11
Control Test		0.080	20:11
Air Blank		0.000	20:12
Diagnostics Check		OK	20:12

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,I
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: PANTOJA,YOSMANY
ARRESTING OFFICER AGENCY: Sebring PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: COUSO

DATE/TIME OF TEST: 12/28/2017 02:24:20

Diagnostics Check		OK	02:20	
Air Blank		0.000	02:20	
Control Test		0.079	02:21	
Air Blank		0.000	02:21	
Subject Sample	# 1		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:23	
Control Test		0.079	02:23	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RAGOODIAL,TRISTON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 08:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:50

SUBJECT'S LAST NAME: FISHER

DATE/TIME OF TEST: 12/28/2017 09:26:17

Diagnostics Check		OK	09:19
Air Blank		0.000	09:20
Control Test		0.079	09:20
Air Blank		0.000	09:21
Subject Sample	# 1	0.000	09:21
Breath Volume		2.695	09:21
Air Blank		0.000	09:22
Air Blank		0.000	09:24
Subject Sample	# 2	0.000	09:24
Breath Volume		2.515	09:24
Air Blank		0.000	09:25
Control Test		0.079	09:25
Air Blank		0.000	09:26
Diagnostics Check		OK	09:26

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGOODIAL,TRISTON,T
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: RAGOODIAL,TRISTON
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: COURT ORD
ARREST TIME: 11:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:10

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/28/2017 11:50:49

Diagnostics Check		OK	11:44
Air Blank		0.000	11:45
Control Test		0.080	11:45
Air Blank		0.000	11:45
Subject Sample	# 1	0.000	11:46
Breath Volume		3.570	11:46
Air Blank		0.000	11:46
Air Blank		0.000	11:48
Subject Sample	# 2	0.000	11:49
Breath Volume		2.964	11:49
Air Blank		0.000	11:49
Control Test		0.080	11:50
Air Blank		0.000	11:50
Diagnostics Check		OK	11:50

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARLOW,ROYCE,B
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: SALVADOR,ELDER
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: OSBORN

DATE/TIME OF TEST: 12/29/2017 23:51:45

Diagnostics Check		OK	23:48	
Air Blank		0.000	23:49	
Control Test		0.081	23:49	
Air Blank		0.000	23:49	
Subject Sample	# 1		23:50	Subject Test Refused
Breath Volume		0.000	23:50	
Air Blank		0.000	23:50	
Control Test		0.080	23:51	
Air Blank		0.000	23:51	
Diagnostics Check		OK	23:51	

Data Download Date/Time: 12/30/2017 08:11

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001054

INSTRUMENT REGISTERED TO: HIGHLANDS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LOTT,JOHN,E
BREATH TEST OPERATOR AGENCY: Highlands County SO
ARRESTING OFFICER: LOTT,JOHN
ARRESTING OFFICER AGENCY: Highlands County SO
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 15:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 12/30/2017 15:51:47

Diagnostics Check		OK	15:45
Air Blank		0.000	15:46
Control Test		0.080	15:46
Air Blank		0.000	15:46
Subject Sample	# 1	0.000	15:47
Breath Volume		2.542	15:47
Air Blank		0.000	15:47
Air Blank		0.000	15:49
Subject Sample	# 2	0.000	15:50
Breath Volume		2.222	15:50
Air Blank		0.000	15:50
Control Test		0.081	15:51
Air Blank		0.000	15:51
Diagnostics Check		OK	15:51

Data Download Date/Time: 01/27/2018 03:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINCLAIR,CAMERON,W
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SINCLAIR,CAMERON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:09

SUBJECT'S LAST NAME: LANDOLT

DATE/TIME OF TEST: 12/01/2017 18:41:10

Diagnostics Check		OK	18:33
Air Blank		0.000	18:34
Control Test		0.080	18:34
Air Blank		0.000	18:34
Subject Sample	# 1	0.213	18:36
Breath Volume		1.515	18:36
Air Blank		0.000	18:37
Air Blank		0.000	18:38
Subject Sample	# 2	0.212	18:39
Breath Volume		1.378	18:39
Air Blank		0.000	18:40
Control Test		0.079	18:40
Air Blank		0.000	18:41
Diagnostics Check		OK	18:41

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LAWES,ERIC,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LAWES,ERIC
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:13

SUBJECT'S LAST NAME: VARHOLICK

DATE/TIME OF TEST: 12/03/2017 18:54:53

Diagnostics Check		OK	18:43	
Air Blank		0.000	18:43	
Control Test		0.080	18:44	
Air Blank		0.000	18:44	
Subject Sample	# 1		18:47	No Sample Provided
Breath Volume		0.000	18:47	
Air Blank		0.000	18:48	
Air Blank		0.000	18:50	
Subject Sample	# 2		18:53	No Sample Provided
Breath Volume		0.000	18:53	
Air Blank		0.000	18:53	
Control Test		0.080	18:54	
Air Blank		0.000	18:54	
Diagnostics Check		OK	18:54	

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:53

SUBJECT'S LAST NAME: WADLEY

DATE/TIME OF TEST: 12/03/2017 23:22:07

Diagnostics Check		OK	23:15
Air Blank		0.000	23:16
Control Test		0.080	23:16
Air Blank		0.000	23:17
Subject Sample	# 1	0.220	23:17
Breath Volume		3.199	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:20
Subject Sample	# 2	0.215	23:20
Breath Volume		2.097	23:20
Air Blank		0.000	23:21
Control Test		0.079	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOLLEY,KATRYNA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: VALE

DATE/TIME OF TEST: 12/05/2017 22:24:05

Diagnostics Check		OK	22:12	
Air Blank		0.000	22:12	
Control Test		0.080	22:12	
Air Blank		0.000	22:13	
Subject Sample	# 1	0.116	22:16	Volume Not Met
Breath Volume		0.390	22:16	
Air Blank		0.000	22:17	
Air Blank		0.000	22:19	
Subject Sample	# 2	0.140	22:19	
Breath Volume		1.433	22:19	
Air Blank		0.000	22:20	
Air Blank		0.000	22:22	
Subject Sample	# 3	0.135	22:22	
Breath Volume		2.707	22:22	
Air Blank		0.000	22:23	
Control Test		0.080	22:23	
Air Blank		0.000	22:24	
Diagnostics Check		OK	22:24	

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:49

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 12/06/2017 21:25:50

Diagnostics Check		OK	21:12	
Air Blank		0.000	21:12	
Control Test		0.078	21:13	
Air Blank		0.000	21:13	
Subject Sample	# 1	0.227	21:14	
Breath Volume		1.460	21:14	
Air Blank		0.000	21:15	
Air Blank		0.000	21:17	
Subject Sample	# 2	0.201	21:20	Volume Not Met
Breath Volume		0.742	21:20	
Air Blank		0.000	21:21	
Air Blank		0.000	21:23	
Subject Sample	# 3	0.226	21:24	
Breath Volume		1.191	21:24	
Air Blank		0.000	21:24	
Control Test		0.078	21:25	
Air Blank		0.000	21:25	
Diagnostics Check		OK	21:25	

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JAMES,MICHAEL
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/09/2017 21:22:42

Diagnostics Check	OK	21:21	
Air Blank	0.000	21:21	
Control Test	0.081	21:22	
Air Blank		21:22	RFI Detect
Air Blank	0.000	21:22	

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: JAMES,MICHAEL
ARRESTING OFFICER AGENCY: Davenport PD
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:53

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/09/2017 21:34:06

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.081	21:27
Air Blank		0.000	21:27
Subject Sample	# 1	0.123	21:29
Breath Volume		1.949	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.116	21:32
Breath Volume		1.796	21:32
Air Blank		0.000	21:33
Control Test		0.080	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:47

SUBJECT'S LAST NAME: SHAW-SLEDGE

DATE/TIME OF TEST: 12/10/2017 19:23:00

Diagnostics Check		OK	19:16
Air Blank		0.000	19:17
Control Test		0.081	19:17
Air Blank		0.000	19:18
Subject Sample	# 1	0.182	19:18
Breath Volume		1.710	19:18
Air Blank		0.000	19:19
Air Blank		0.000	19:21
Subject Sample	# 2	0.183	19:21
Breath Volume		1.644	19:21
Air Blank		0.000	19:22
Control Test		0.080	19:22
Air Blank		0.000	19:22
Diagnostics Check		OK	19:23

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: JIGUAN

DATE/TIME OF TEST: 12/10/2017 19:51:34

Diagnostics Check		OK	19:43
Air Blank		0.000	19:44
Control Test		0.081	19:44
Air Blank		0.000	19:45
Subject Sample	# 1	0.294	19:46
Breath Volume		1.761	19:46
Air Blank		0.000	19:47
Air Blank		0.000	19:49
Subject Sample	# 2	0.281	19:50
Breath Volume		1.156	19:50
Air Blank		0.000	19:50
Control Test		0.080	19:50
Air Blank		0.000	19:51
Diagnostics Check		OK	19:51

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOLLEY,KATRYNA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: RODRIGUEZ PENA

DATE/TIME OF TEST: 12/11/2017 00:27:11

Diagnostics Check		OK	00:20
Air Blank		0.000	00:21
Control Test		0.081	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.144	00:22
Breath Volume		1.542	00:22
Air Blank		0.000	00:23
Air Blank		0.000	00:25
Subject Sample	# 2	0.140	00:25
Breath Volume		1.875	00:25
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:27
Diagnostics Check		OK	00:27

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUCAS,NATHANIEL,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LUCAS,NATHANIEL
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: MASON

DATE/TIME OF TEST: 12/14/2017 21:24:37

Diagnostics Check		OK	21:18
Air Blank		0.000	21:18
Control Test		0.080	21:19
Air Blank		0.000	21:19
Subject Sample	# 1	0.145	21:19
Breath Volume		1.328	21:19
Air Blank		0.000	21:20
Air Blank		0.000	21:22
Subject Sample	# 2	0.138	21:23
Breath Volume		1.277	21:23
Air Blank		0.000	21:23
Control Test		0.080	21:24
Air Blank		0.000	21:24
Diagnostics Check		OK	21:24

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SYKES,STEVEN,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: SYKES,STEVEN
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/14/2017 21:34:13

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.080	21:29
Air Blank		0.000	21:29
Subject Sample	# 1	0.125	21:29
Breath Volume		1.562	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.123	21:32
Breath Volume		1.640	21:32
Air Blank		0.000	21:33
Control Test		0.079	21:33
Air Blank		0.000	21:34
Diagnostics Check		OK	21:34

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHRISTELLO,CLIFFORD,A
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: CHRISTELLO,CLIFFORD
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: GUTIERREZ-ZEFERINO

DATE/TIME OF TEST: 12/15/2017 00:34:21

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.081	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.273	00:29
Breath Volume		1.765	00:29
Air Blank		0.000	00:30
Air Blank		0.000	00:32
Subject Sample	# 2	0.267	00:32
Breath Volume		1.152	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,ADAM,H
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: POTTER,ADAM
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: WELTY

DATE/TIME OF TEST: 12/17/2017 03:07:11

Diagnostics Check		OK	02:59
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.192	03:02
Breath Volume		1.351	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.202	03:05
Breath Volume		2.000	03:05
Air Blank		0.000	03:06
Control Test		0.080	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WARREN,BILLY
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: FRADY

DATE/TIME OF TEST: 12/17/2017 04:42:54

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:32	
Control Test		0.081	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:35	No Sample Provided
Breath Volume		0.000	04:35	
Air Blank		0.000	04:36	
Air Blank		0.000	04:38	
Subject Sample	# 2		04:41	No Sample Provided
Breath Volume		0.000	04:41	
Air Blank		0.000	04:42	
Control Test		0.081	04:42	
Air Blank		0.000	04:42	
Diagnostics Check		OK	04:42	

Data Download Date/Time: 12/23/2017 19:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WARREN,BILLY,J
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: YOUNG,JEVON
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:57

SUBJECT'S LAST NAME: MARSON

DATE/TIME OF TEST: 12/17/2017 07:27:35

Diagnostics Check		OK	07:24	
Air Blank		0.000	07:25	
Control Test		0.081	07:25	
Air Blank		0.000	07:26	
Subject Sample	# 1		07:26	Subject Test Refused
Breath Volume		0.000	07:26	
Air Blank		0.000	07:26	
Control Test		0.081	07:27	
Air Blank		0.000	07:27	
Diagnostics Check		OK	07:27	

Data Download Date/Time: 12/23/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDORFF,RUSSELL,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: WALDORFF,RUSSELL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: HUNT

DATE/TIME OF TEST: 12/21/2017 02:11:38

Diagnostics Check		OK	02:00	
Air Blank		0.000	02:00	
Control Test		0.081	02:01	
Air Blank		0.000	02:01	
Subject Sample	# 1		02:04	No Sample Provided
Breath Volume		0.000	02:04	
Air Blank		0.000	02:05	
Air Blank		0.000	02:07	
Subject Sample	# 2		02:10	No Sample Provided
Breath Volume		0.000	02:10	
Air Blank		0.000	02:10	
Control Test		0.081	02:11	
Air Blank		0.000	02:11	
Diagnostics Check		OK	02:11	

Data Download Date/Time: 12/23/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAMMONDS,TEAQULIA,L
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: LOVETT,DOUGLAS
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: WOLFE

DATE/TIME OF TEST: 12/21/2017 23:51:10

Diagnostics Check		OK	23:45
Air Blank		0.000	23:45
Control Test		0.081	23:45
Air Blank		0.000	23:46
Subject Sample	# 1	0.150	23:46
Breath Volume		2.683	23:46
Air Blank		0.000	23:47
Air Blank		0.000	23:49
Subject Sample	# 2	0.161	23:49
Breath Volume		2.394	23:49
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:51
Diagnostics Check		OK	23:51

Data Download Date/Time: 12/23/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLEY,KATRYNA,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SOLLEY,KATRYNA
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: MUSICO JR

DATE/TIME OF TEST: 12/22/2017 22:10:52

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.079	22:05
Air Blank		0.000	22:05
Subject Sample	# 1	0.153	22:06
Breath Volume		1.386	22:06
Air Blank		0.000	22:06
Air Blank		0.000	22:08
Subject Sample	# 2	0.162	22:09
Breath Volume		1.769	22:09
Air Blank		0.000	22:09
Control Test		0.075	22:10
Air Blank		0.000	22:10
Diagnostics Check		OK	22:10

Data Download Date/Time: 12/23/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: WOFFORD,CLINT
ARRESTING OFFICER AGENCY: Polk County SO
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 12/23/2017 01:48:14

Diagnostics Check		OK	01:42
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.000	01:43
Breath Volume		2.050	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.000	01:46
Breath Volume		2.500	01:46
Air Blank		0.000	01:47
Control Test		0.080	01:47
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 12/23/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001055

INSTRUMENT REGISTERED TO: POLK COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JEUDY,RINALDO,
BREATH TEST OPERATOR AGENCY: Polk County SO
ARRESTING OFFICER: MCKEE,JOHN
ARRESTING OFFICER AGENCY: Bartow PD
LAST AGENCY INSPECTION: 11/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080AI
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:22

SUBJECT'S LAST NAME: HERNANDEZ-FRIAS

DATE/TIME OF TEST: 12/23/2017 02:57:06

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.081	02:54	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:56	
Control Test		0.080	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 12/23/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:49

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 12/17/2017 03:20:27

Diagnostics Check		OK	03:13
Air Blank		0.000	03:14
Control Test		0.079	03:14
Air Blank		0.000	03:15
Subject Sample	# 1	0.116	03:15
Breath Volume		1.207	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.112	03:18
Breath Volume		1.171	03:18
Air Blank		0.000	03:19
Control Test		0.079	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 12/27/2017 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/22/2017 03:00:45

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:47	
Control Test		0.078	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1	0.150	02:49	
Breath Volume		1.187	02:49	
Air Blank		0.000	02:50	
Air Blank		0.000	02:52	
Subject Sample	# 2	0.126	02:55	Volume Not Met
Breath Volume		0.460	02:55	
Air Blank		0.000	02:56	
Air Blank		0.000	02:58	
Subject Sample	# 3	0.152	02:59	
Breath Volume		1.496	02:59	
Air Blank		0.000	02:59	
Control Test		0.079	03:00	
Air Blank		0.000	03:00	
Diagnostics Check		OK	03:00	

Data Download Date/Time: 12/27/2017 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001057

INSTRUMENT REGISTERED TO: LAKELAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTLE,CORY,A
BREATH TEST OPERATOR AGENCY: Lakeland PD
ARRESTING OFFICER: SUTTLE,CORY
ARRESTING OFFICER AGENCY: Lakeland PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: HIGDON

DATE/TIME OF TEST: 12/24/2017 01:14:57

Diagnostics Check		OK	01:05
Air Blank		0.000	01:06
Control Test		0.080	01:06
Air Blank		0.000	01:07
Subject Sample	# 1	0.097	01:09
Breath Volume		1.292	01:09
Air Blank		0.000	01:09
Air Blank		0.000	01:11
Subject Sample	# 2	0.090	01:13
Breath Volume		1.375	01:13
Air Blank		0.000	01:14
Control Test		0.079	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 12/27/2017 09:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: LEWIS,RYAN
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:13

SUBJECT'S LAST NAME: DICKSON

DATE/TIME OF TEST: 12/01/2017 22:37:33

Diagnostics Check		OK	22:34	
Air Blank		0.000	22:35	
Control Test		0.081	22:35	
Air Blank		0.000	22:35	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:36	
Control Test		0.082	22:36	
Air Blank		0.000	22:37	
Diagnostics Check		OK	22:37	

Data Download Date/Time: 12/12/2017 12:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,CODY,W
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: GRANTHAM,CODY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: RILEY

DATE/TIME OF TEST: 12/05/2017 01:25:40

Diagnostics Check		OK	01:18
Air Blank		0.000	01:19
Control Test		0.081	01:19
Air Blank		0.000	01:20
Subject Sample	# 1	0.199	01:21
Breath Volume		2.062	01:21
Air Blank		0.000	01:21
Air Blank		0.000	01:23
Subject Sample	# 2	0.207	01:24
Breath Volume		3.078	01:24
Air Blank		0.000	01:24
Control Test		0.082	01:25
Air Blank		0.000	01:25
Diagnostics Check		OK	01:25

Data Download Date/Time: 12/12/2017 12:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,BRANDON,S
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: BOONE,BRANDON
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:49

SUBJECT'S LAST NAME: SAN MINGUEL

DATE/TIME OF TEST: 12/14/2017 12:15:08

Diagnostics Check		OK	12:11	
Air Blank		0.000	12:12	
Control Test		0.082	12:12	
Air Blank		0.000	12:13	
Subject Sample	# 1		12:13	Subject Test Refused
Breath Volume		0.000	12:13	
Air Blank		0.000	12:14	
Control Test		0.083	12:14	
Air Blank		0.000	12:15	
Diagnostics Check		OK	12:15	

Data Download Date/Time: 01/16/2018 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: ORPHEE,MICHAEL
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 12/15/2017 22:58:27

Diagnostics Check		OK	22:50
Air Blank		0.000	22:51
Control Test		0.081	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.132	22:53
Breath Volume		2.621	22:53
Air Blank		0.000	22:54
Air Blank		0.000	22:55
Subject Sample	# 2	0.129	22:56
Breath Volume		2.710	22:56
Air Blank		0.000	22:57
Control Test		0.083	22:57
Air Blank		0.000	22:58
Diagnostics Check		OK	22:58

Data Download Date/Time: 01/16/2018 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ORPHEE,MICHAEL,
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: ORPHEE,MICHAEL
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: SWEET

DATE/TIME OF TEST: 12/17/2017 02:21:39

Diagnostics Check		OK	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.103	02:16
Breath Volume		2.125	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.103	02:19
Breath Volume		2.152	02:19
Air Blank		0.000	02:20
Control Test		0.082	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 01/16/2018 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,CODY,W
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: GRANTHAM,CODY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MONTEZ

DATE/TIME OF TEST: 12/18/2017 22:30:10

Diagnostics Check		OK	22:22
Air Blank		0.000	22:22
Control Test		0.081	22:23
Air Blank		0.000	22:23
Subject Sample	# 1	0.144	22:24
Breath Volume		1.105	22:24
Air Blank		0.000	22:24
Air Blank		0.000	22:26
Subject Sample	# 2	0.144	22:28
Breath Volume		2.722	22:28
Air Blank		0.000	22:29
Control Test		0.084	22:29
Air Blank		0.000	22:30
Diagnostics Check		OK	22:30

Data Download Date/Time: 01/16/2018 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,CODY,W
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: GRANTHAM,CODY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 12/20/2017 04:58:09

Diagnostics Check		OK	04:52	
Air Blank		0.000	04:52	
Control Test		0.081	04:52	
Air Blank		0.000	04:53	
Subject Sample	# 1	0.161	04:56	Volume Not Met
Breath Volume		0.035	04:56	
Air Blank			04:57	Purge Fail
Air Blank			04:58	Purge Fail

Data Download Date/Time: 01/16/2018 16:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001060

INSTRUMENT REGISTERED TO: WINTER HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,CODY,W
BREATH TEST OPERATOR AGENCY: Winter Haven PD
ARRESTING OFFICER: GRANTHAM,CODY
ARRESTING OFFICER AGENCY: Winter Haven PD
LAST AGENCY INSPECTION: 12/12/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: MCDANIEL

DATE/TIME OF TEST: 12/20/2017 05:07:02

Diagnostics Check		OK	04:59
Air Blank		0.000	05:00
Control Test		0.080	05:00
Air Blank		0.000	05:01
Subject Sample	# 1	0.167	05:02
Breath Volume		3.367	05:02
Air Blank		0.000	05:02
Air Blank		0.000	05:04
Subject Sample	# 2	0.158	05:05
Breath Volume		2.800	05:05
Air Blank		0.000	05:06
Control Test		0.081	05:06
Air Blank		0.000	05:06
Diagnostics Check		OK	05:07

Data Download Date/Time: 01/16/2018 16:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:16

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 12/05/2017 23:40:24

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.080	23:38	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:39	Subject Test Refused
Breath Volume		0.000	23:39	
Air Blank		0.000	23:39	
Control Test		0.080	23:39	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 12/31/2017 15:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: NIEVES,ARION
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 12/28/2017 20:33:18

Diagnostics Check		OK	20:26
Air Blank		0.000	20:27
Control Test		0.081	20:27
Air Blank		0.000	20:27
Subject Sample	# 1	0.000	20:28
Breath Volume		1.722	20:28
Air Blank		0.000	20:29
Air Blank		0.000	20:31
Subject Sample	# 2	0.000	20:31
Breath Volume		2.492	20:31
Air Blank		0.000	20:32
Control Test		0.080	20:32
Air Blank		0.000	20:33
Diagnostics Check		OK	20:33

Data Download Date/Time: 12/31/2017 15:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001061

INSTRUMENT REGISTERED TO: GULFPORT PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARTER,TIMOTHY,R
BREATH TEST OPERATOR AGENCY: Gulfport PD
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:51

SUBJECT'S LAST NAME: ELMIGER

DATE/TIME OF TEST: 12/30/2017 03:20:05

Diagnostics Check		OK	03:12
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.203	03:14
Breath Volume		1.519	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.202	03:18
Breath Volume		1.679	03:18
Air Blank		0.000	03:19
Control Test		0.080	03:19
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 12/31/2017 15:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRITTEN,STEVEN,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: BRITTEN,STEVEN
ARRESTING OFFICER AGENCY: Treasure Island PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:36

SUBJECT'S LAST NAME: DEMKO

DATE/TIME OF TEST: 12/05/2017 06:06:09

Diagnostics Check		OK	05:59
Air Blank		0.000	06:00
Control Test		0.080	06:00
Air Blank		0.000	06:00
Subject Sample	# 1	0.156	06:01
Breath Volume		1.628	06:01
Air Blank		0.000	06:02
Air Blank		0.000	06:04
Subject Sample	# 2	0.163	06:04
Breath Volume		1.832	06:04
Air Blank		0.000	06:05
Control Test		0.080	06:05
Air Blank		0.000	06:06
Diagnostics Check		OK	06:06

Data Download Date/Time: 12/21/2017 09:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SWETNICH,JARED,A
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: JANIK,STEPHEN
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 12/13/2017 01:18:54

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:16	
Control Test		0.080	01:16	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:18	
Control Test		0.080	01:18	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 12/21/2017 09:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001062

INSTRUMENT REGISTERED TO: TREASURE ISLAND PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRITTEN,STEVEN,
BREATH TEST OPERATOR AGENCY: Treasure Island PD
ARRESTING OFFICER: JANIK,STEPHEN
ARRESTING OFFICER AGENCY: Indian Shores PD
LAST AGENCY INSPECTION: 11/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A6
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:45

SUBJECT'S LAST NAME: MCFADDEN

DATE/TIME OF TEST: 12/17/2017 06:16:52

Diagnostics Check		OK	06:05	
Air Blank		0.000	06:05	
Control Test		0.080	06:06	
Air Blank		0.000	06:06	
Subject Sample	# 1		06:09	No Sample Provided
Breath Volume		0.000	06:09	
Air Blank		0.000	06:10	
Air Blank		0.000	06:12	
Subject Sample	# 2		06:15	No Sample Provided
Breath Volume		0.000	06:15	
Air Blank		0.000	06:15	
Control Test		0.081	06:16	
Air Blank		0.000	06:16	
Diagnostics Check		OK	06:16	

Data Download Date/Time: 12/21/2017 09:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001064

INSTRUMENT REGISTERED TO: NEW PORT RICHEY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A1
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:34

SUBJECT'S LAST NAME: GANLEY

DATE/TIME OF TEST: 12/04/2017 23:01:06

Diagnostics Check		OK	22:54
Air Blank		0.000	22:54
Control Test		0.079	22:54
Air Blank		0.000	22:55
Subject Sample	# 1	0.138	22:56
Breath Volume		1.386	22:56
Air Blank		0.000	22:56
Air Blank		0.000	22:58
Subject Sample	# 2	0.135	22:59
Breath Volume		1.367	22:59
Air Blank		0.000	23:00
Control Test		0.079	23:00
Air Blank		0.000	23:01
Diagnostics Check		OK	23:01

Data Download Date/Time: 12/30/2017 22:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: WARWICK,ANDREW
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: MASTERSON

DATE/TIME OF TEST: 12/01/2017 20:18:46

Diagnostics Check		OK	20:15	
Air Blank		0.000	20:16	
Control Test		0.080	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:17	
Control Test		0.080	20:18	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:18	

Data Download Date/Time: 12/18/2017 12:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOONE,ANTHONY,L
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: ROQUE,MIKE
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:53

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/06/2017 15:20:51

Diagnostics Check		OK	15:13
Air Blank		0.000	15:14
Control Test		0.079	15:14
Air Blank		0.000	15:15
Subject Sample	# 1	0.232	15:16
Breath Volume		2.843	15:16
Air Blank		0.000	15:16
Air Blank		0.000	15:18
Subject Sample	# 2	0.247	15:19
Breath Volume		1.937	15:19
Air Blank		0.000	15:19
Control Test		0.078	15:20
Air Blank		0.000	15:20
Diagnostics Check		OK	15:20

Data Download Date/Time: 12/18/2017 12:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: WARWICK,ANDREW
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 12/11/2017 22:28:20

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:25	
Control Test		0.079	22:26	
Air Blank		0.000	22:26	
Subject Sample	# 1		22:26	Subject Test Refused
Breath Volume		0.000	22:26	
Air Blank		0.000	22:27	
Control Test		0.079	22:27	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 12/18/2017 12:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: SPATZ,JOHN
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 11/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: HEADLEY

DATE/TIME OF TEST: 12/16/2017 04:06:59

Diagnostics Check		OK	04:04	
Air Blank		0.000	04:04	
Control Test		0.081	04:04	
Air Blank		0.000	04:05	
Subject Sample	# 1		04:05	Subject Test Refused
Breath Volume		0.000	04:05	
Air Blank		0.000	04:06	
Control Test		0.079	04:06	
Air Blank		0.000	04:06	
Diagnostics Check		OK	04:06	

Data Download Date/Time: 12/18/2017 12:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TRINIDAD,MICHAEL,E
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: MACISAAC,SCOTT
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: BURCH

DATE/TIME OF TEST: 12/20/2017 22:40:41

Diagnostics Check		OK	22:29	
Air Blank		0.000	22:30	
Control Test		0.078	22:30	
Air Blank		0.000	22:31	
Subject Sample	# 1		22:34	No Sample Provided
Breath Volume		0.000	22:34	
Air Blank		0.000	22:34	
Air Blank		0.000	22:36	
Subject Sample	# 2		22:39	Subject Test Refused
Breath Volume		0.000	22:39	
Air Blank		0.000	22:39	
Control Test		0.080	22:40	
Air Blank		0.000	22:40	
Diagnostics Check		OK	22:40	

Data Download Date/Time: 01/29/2018 13:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPATZ,JOHN,M
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: WARWICK,ANDREW
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: KOWALSKI

DATE/TIME OF TEST: 12/22/2017 00:52:06

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.079	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.113	00:47
Breath Volume		3.488	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.114	00:50
Breath Volume		3.460	00:50
Air Blank		0.000	00:51
Control Test		0.078	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 01/29/2018 13:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GELLATLY,ROBERT,A
BREATH TEST OPERATOR AGENCY: Tarpon Springs PD
ARRESTING OFFICER: MILLER,JACOB
ARRESTING OFFICER AGENCY: Tarpon Springs PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:25

SUBJECT'S LAST NAME: GENTRY

DATE/TIME OF TEST: 12/23/2017 21:00:51

Diagnostics Check		OK	20:54
Air Blank		0.000	20:54
Control Test		0.079	20:55
Air Blank		0.000	20:55
Subject Sample	# 1	0.229	20:56
Breath Volume		2.722	20:56
Air Blank		0.000	20:56
Air Blank		0.000	20:58
Subject Sample	# 2	0.216	20:59
Breath Volume		3.371	20:59
Air Blank		0.000	20:59
Control Test		0.078	21:00
Air Blank		0.000	21:00
Diagnostics Check		OK	21:00

Data Download Date/Time: 01/29/2018 13:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001065

INSTRUMENT REGISTERED TO: TARPON SPRINGS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,WILLIAM,T
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: SMITH,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ADAY

DATE/TIME OF TEST: 12/30/2017 04:04:05

Diagnostics Check		OK	03:57
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.149	03:59
Breath Volume		2.359	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.144	04:02
Breath Volume		2.640	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 01/29/2018 13:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/22/2017 23:28:56

Diagnostics Check	OK	23:28	
Air Blank		23:28	RFI Detect
Air Blank	0.000	23:28	

Data Download Date/Time: 12/27/2017 20:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 12/22/2017 23:37:01

Diagnostics Check		OK	23:30
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:31
Subject Sample	# 1	0.000	23:32
Breath Volume		1.988	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.000	23:35
Breath Volume		2.964	23:35
Air Blank		0.000	23:36
Control Test		0.079	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:37

Data Download Date/Time: 12/27/2017 20:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001066

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: CARTER,TIMOTHY
ARRESTING OFFICER AGENCY: Gulfport PD
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: CHROMCZAK

DATE/TIME OF TEST: 12/23/2017 05:13:48

Diagnostics Check		OK	05:11	
Air Blank		0.000	05:11	
Control Test		0.079	05:11	
Air Blank		0.000	05:12	
Subject Sample	# 1		05:12	Subject Test Refused
Breath Volume		0.000	05:12	
Air Blank		0.000	05:12	
Control Test		0.079	05:13	
Air Blank		0.000	05:13	
Diagnostics Check		OK	05:13	

Data Download Date/Time: 12/27/2017 20:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: LICHTENBERG

DATE/TIME OF TEST: 12/01/2017 00:42:50

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.165	00:38
Breath Volume		1.171	00:38
Air Blank		0.000	00:38
Air Blank		0.000	00:40
Subject Sample	# 2	0.168	00:41
Breath Volume		2.281	00:41
Air Blank		0.000	00:41
Control Test		0.079	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:42

Data Download Date/Time: 12/21/2017 20:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: CHANDA

DATE/TIME OF TEST: 12/02/2017 01:37:23

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:24	
Control Test		0.079	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:28	No Sample Provided
Breath Volume		0.000	01:28	
Air Blank		0.000	01:28	
Air Blank		0.000	01:30	
Subject Sample	# 2	0.205	01:32	
Breath Volume		2.691	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 3	0.195	01:35	
Breath Volume		2.464	01:35	
Air Blank		0.000	01:36	
Control Test		0.076	01:36	
Air Blank		0.000	01:37	
Diagnostics Check		OK	01:37	

Data Download Date/Time: 12/21/2017 20:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: PENA SILIS

DATE/TIME OF TEST: 12/03/2017 00:32:45

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:30	
Control Test		0.079	00:30	
Air Blank		0.000	00:31	
Subject Sample	# 1	0.238	00:31	
Breath Volume		3.843	00:31	
Air Blank			00:32	Purge Fail
Air Blank		0.000	00:32	

Data Download Date/Time: 12/21/2017 20:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: PENA SILIS

DATE/TIME OF TEST: 12/03/2017 00:35:58

Diagnostics Check	OK	00:34	
Air Blank	0.000	00:34	
Control Test	0.074	00:35	Control Outside Tolerance
Air Blank	0.000	00:35	

Data Download Date/Time: 12/21/2017 20:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NUGENT,DAVID
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:08

SUBJECT'S LAST NAME: PENA SILIS

DATE/TIME OF TEST: 12/03/2017 00:43:51

Diagnostics Check		OK	00:37	
Air Blank		0.000	00:38	
Control Test		0.077	00:38	
Air Blank		0.000	00:39	
Subject Sample	# 1	0.242	00:39	
Breath Volume		4.472	00:39	
Air Blank		0.000	00:40	
Air Blank		0.000	00:42	
Subject Sample	# 2	0.232	00:42	
Breath Volume		4.339	00:42	
Air Blank			00:43	Purge Fail
Air Blank		0.000	00:43	

Data Download Date/Time: 12/21/2017 20:40

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: REED,CHAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: WOOD

DATE/TIME OF TEST: 12/16/2017 01:51:14

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:44
Air Blank		0.000	01:45
Subject Sample	# 1	0.157	01:46
Breath Volume		2.250	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:48
Subject Sample	# 2	0.152	01:49
Breath Volume		2.812	01:49
Air Blank		0.000	01:50
Control Test		0.079	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 12/21/2017 20:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 11/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SOSA FLORES

DATE/TIME OF TEST: 12/16/2017 03:02:52

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.126	02:58
Breath Volume		1.511	02:58
Air Blank		0.000	02:59
Air Blank		0.000	03:00
Subject Sample	# 2	0.126	03:01
Breath Volume		1.636	03:01
Air Blank		0.000	03:01
Control Test		0.079	03:02
Air Blank		0.000	03:02
Diagnostics Check		OK	03:02

Data Download Date/Time: 12/21/2017 20:40

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,CHAD,R
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: NAJJIH MOLL

DATE/TIME OF TEST: 12/22/2017 03:34:13

Diagnostics Check		OK	03:26
Air Blank		0.000	03:27
Control Test		0.080	03:27
Air Blank		0.000	03:28
Subject Sample	# 1	0.167	03:29
Breath Volume		2.574	03:29
Air Blank		0.000	03:30
Air Blank		0.000	03:31
Subject Sample	# 2	0.178	03:32
Breath Volume		2.000	03:32
Air Blank		0.000	03:33
Control Test		0.077	03:33
Air Blank		0.000	03:34
Diagnostics Check		OK	03:34

Data Download Date/Time: 01/02/2018 18:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLEN,BRAD,A
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ALLEN,BRAD
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: PHOMTHISENE

DATE/TIME OF TEST: 12/22/2017 05:00:36

Diagnostics Check		OK	04:53
Air Blank		0.000	04:53
Control Test		0.079	04:54
Air Blank		0.000	04:54
Subject Sample	# 1	0.186	04:55
Breath Volume		2.285	04:55
Air Blank		0.000	04:56
Air Blank		0.000	04:58
Subject Sample	# 2	0.185	04:58
Breath Volume		2.898	04:58
Air Blank		0.000	04:59
Control Test		0.077	04:59
Air Blank		0.000	05:00
Diagnostics Check		OK	05:00

Data Download Date/Time: 01/02/2018 18:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:11

SUBJECT'S LAST NAME: MCNAMEE

DATE/TIME OF TEST: 12/23/2017 23:47:32

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.079	23:41
Air Blank		0.000	23:41
Subject Sample	# 1	0.190	23:42
Breath Volume		1.953	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.182	23:45
Breath Volume		2.312	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/02/2018 18:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: HANKINSON

DATE/TIME OF TEST: 12/24/2017 02:31:45

Diagnostics Check		OK	02:24
Air Blank		0.000	02:25
Control Test		0.080	02:25
Air Blank		0.000	02:26
Subject Sample	# 1	0.172	02:26
Breath Volume		2.613	02:26
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.169	02:30
Breath Volume		2.351	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:31
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 01/02/2018 18:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ENGLAND,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: ROBERTS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:15

SUBJECT'S LAST NAME: GENNINGER

DATE/TIME OF TEST: 12/26/2017 17:45:38

Diagnostics Check		OK	17:39
Air Blank		0.000	17:39
Control Test		0.080	17:39
Air Blank		0.000	17:40
Subject Sample	# 1	0.000	17:41
Breath Volume		2.988	17:41
Air Blank		0.000	17:41
Air Blank		0.000	17:43
Subject Sample	# 2	0.000	17:44
Breath Volume		2.640	17:44
Air Blank		0.000	17:44
Control Test		0.079	17:45
Air Blank		0.000	17:45
Diagnostics Check		OK	17:45

Data Download Date/Time: 01/02/2018 18:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: NEGERSMITH,DANIEL
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 12/27/2017 23:54:47

Diagnostics Check		OK	23:46
Air Blank		0.000	23:47
Control Test		0.081	23:47
Air Blank		0.000	23:48
Subject Sample	# 1	0.124	23:49
Breath Volume		2.074	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:51
Subject Sample	# 2	0.121	23:53
Breath Volume		1.890	23:53
Air Blank		0.000	23:53
Control Test		0.078	23:54
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 01/02/2018 18:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001068

INSTRUMENT REGISTERED TO: CLEARWATER PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NEGERSMITH,DANIEL,M
BREATH TEST OPERATOR AGENCY: Clearwater PD
ARRESTING OFFICER: HURT,JONATHAN
ARRESTING OFFICER AGENCY: Clearwater PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: COSTA

DATE/TIME OF TEST: 12/28/2017 01:31:37

Diagnostics Check		OK	01:24
Air Blank		0.000	01:24
Control Test		0.079	01:24
Air Blank		0.000	01:25
Subject Sample	# 1	0.121	01:26
Breath Volume		3.410	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:28
Subject Sample	# 2	0.121	01:29
Breath Volume		3.417	01:29
Air Blank		0.000	01:30
Control Test		0.078	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 01/02/2018 18:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARLSON,P
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: STERLING

DATE/TIME OF TEST: 12/01/2017 00:06:18

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:03	
Control Test		0.081	00:04	
Air Blank		0.000	00:04	
Subject Sample	# 1		00:04	Subject Test Refused
Breath Volume		0.000	00:04	
Air Blank		0.000	00:05	
Control Test		0.081	00:05	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 12/11/2017 15:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: SOLIMAN

DATE/TIME OF TEST: 12/01/2017 03:25:30

Diagnostics Check		OK	03:19
Air Blank		0.000	03:19
Control Test		0.081	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.070	03:21
Breath Volume		1.261	03:21
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.064	03:23
Breath Volume		1.664	03:23
Air Blank		0.000	03:24
Control Test		0.081	03:24
Air Blank		0.000	03:25
Diagnostics Check		OK	03:25

Data Download Date/Time: 12/11/2017 15:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: WORTH

DATE/TIME OF TEST: 12/01/2017 05:06:14

Diagnostics Check		OK	04:59
Air Blank		0.000	04:59
Control Test		0.081	05:00
Air Blank		0.000	05:00
Subject Sample	# 1	0.221	05:01
Breath Volume		2.304	05:01
Air Blank		0.000	05:02
Air Blank		0.000	05:03
Subject Sample	# 2	0.215	05:04
Breath Volume		2.339	05:04
Air Blank		0.000	05:05
Control Test		0.080	05:05
Air Blank		0.000	05:06
Diagnostics Check		OK	05:06

Data Download Date/Time: 12/11/2017 15:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:33

SUBJECT'S LAST NAME: CAVEY

DATE/TIME OF TEST: 12/01/2017 05:56:41

Diagnostics Check		OK	05:53	
Air Blank		0.000	05:54	
Control Test		0.081	05:54	
Air Blank		0.000	05:55	
Subject Sample	# 1		05:55	Subject Test Refused
Breath Volume		0.000	05:55	
Air Blank		0.000	05:55	
Control Test		0.081	05:56	
Air Blank		0.000	05:56	
Diagnostics Check		OK	05:56	

Data Download Date/Time: 12/11/2017 15:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: HARDESTY

DATE/TIME OF TEST: 12/01/2017 19:07:22

Diagnostics Check		OK	19:04	
Air Blank		0.000	19:04	
Control Test		0.080	19:05	
Air Blank		0.000	19:05	
Subject Sample	# 1		19:05	Subject Test Refused
Breath Volume		0.000	19:05	
Air Blank		0.000	19:06	
Control Test		0.080	19:06	
Air Blank		0.000	19:07	
Diagnostics Check		OK	19:07	

Data Download Date/Time: 12/11/2017 15:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: JEREZ URRUTIA

DATE/TIME OF TEST: 12/02/2017 01:06:02

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:03	
Control Test		0.082	01:03	
Air Blank		0.000	01:04	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:05	
Control Test		0.081	01:05	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:06	

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROLLE

DATE/TIME OF TEST: 12/02/2017 03:35:45

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.082	03:33	
Air Blank		0.000	03:34	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.081	03:35	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MILLER,C
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:52

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 12/02/2017 04:19:40

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.121	04:15
Breath Volume		1.472	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.116	04:18
Breath Volume		1.140	04:18
Air Blank		0.000	04:18
Control Test		0.080	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:21

SUBJECT'S LAST NAME: JOHNS

DATE/TIME OF TEST: 12/02/2017 05:45:02

Diagnostics Check		OK	05:42	
Air Blank		0.000	05:42	
Control Test		0.081	05:42	
Air Blank		0.000	05:43	
Subject Sample	# 1		05:43	Subject Test Refused
Breath Volume		0.000	05:43	
Air Blank		0.000	05:44	
Control Test		0.081	05:44	
Air Blank		0.000	05:44	
Diagnostics Check		OK	05:45	

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROSEPILER,JONATHAN,M
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/05/2017 01:26:43

Diagnostics Check		OK	01:20
Air Blank		0.000	01:20
Control Test		0.082	01:21
Air Blank		0.000	01:21
Subject Sample	# 1	0.123	01:22
Breath Volume		2.667	01:22
Air Blank		0.000	01:22
Air Blank		0.000	01:24
Subject Sample	# 2	0.124	01:25
Breath Volume		3.398	01:25
Air Blank		0.000	01:25
Control Test		0.081	01:26
Air Blank		0.000	01:26
Diagnostics Check		OK	01:26

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 12/06/2017 03:25:03

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.081	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:23	Subject Test Refused
Breath Volume		0.000	03:23	
Air Blank		0.000	03:24	
Control Test		0.082	03:24	
Air Blank		0.000	03:24	
Diagnostics Check		OK	03:25	

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRIER,A,R
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: BUI

DATE/TIME OF TEST: 12/08/2017 02:57:32

Diagnostics Check		OK	02:54	
Air Blank		0.000	02:54	
Control Test		0.083	02:55	
Air Blank		0.000	02:55	
Subject Sample	# 1		02:56	Subject Test Refused
Breath Volume		0.000	02:56	
Air Blank		0.000	02:56	
Control Test		0.083	02:56	
Air Blank		0.000	02:57	
Diagnostics Check		OK	02:57	

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUTTON,BRADEN,J
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: DOUGHFMAN,D
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: FLORES CUIN

DATE/TIME OF TEST: 12/09/2017 23:11:09

Diagnostics Check		OK	23:04
Air Blank		0.000	23:04
Control Test		0.082	23:05
Air Blank		0.000	23:05
Subject Sample	# 1	0.128	23:06
Breath Volume		2.503	23:06
Air Blank		0.000	23:07
Air Blank		0.000	23:09
Subject Sample	# 2	0.113	23:09
Breath Volume		1.101	23:09
Air Blank		0.000	23:10
Control Test		0.083	23:10
Air Blank		0.000	23:11
Diagnostics Check		OK	23:11

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRETT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:37

SUBJECT'S LAST NAME: HADDOCK

DATE/TIME OF TEST: 12/10/2017 04:04:02

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.082	03:58
Air Blank		0.000	03:58
Subject Sample	# 1	0.233	03:59
Breath Volume		1.218	03:59
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.235	04:02
Breath Volume		4.226	04:02
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:03
Diagnostics Check		OK	04:04

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: FERNANDEZ

DATE/TIME OF TEST: 12/10/2017 21:13:38

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.081	21:11	
Air Blank		0.000	21:11	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Control Test		0.082	21:13	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 12/11/2017 16:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MACK,C,S
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GARCIA,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:01

SUBJECT'S LAST NAME: ROYER

DATE/TIME OF TEST: 12/12/2017 20:28:31

Diagnostics Check		OK	20:21
Air Blank		0.000	20:22
Control Test		0.082	20:22
Air Blank		0.000	20:23
Subject Sample	# 1	0.141	20:23
Breath Volume		2.613	20:23
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.138	20:26
Breath Volume		2.128	20:26
Air Blank		0.000	20:27
Control Test		0.082	20:27
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: NETTLES

DATE/TIME OF TEST: 12/13/2017 00:04:35

Diagnostics Check		OK	23:58
Air Blank		0.000	23:58
Control Test		0.081	23:58
Air Blank		0.000	23:59
Subject Sample	# 2	0.243	23:59
Breath Volume		2.605	23:59
Air Blank		0.000	00:00
Air Blank		0.000	00:02
Subject Sample	# 1	0.227	00:02
Breath Volume		1.488	00:02
Air Blank		0.000	00:03
Control Test		0.081	00:03
Air Blank		0.000	00:04
Diagnostics Check		OK	00:04

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: PENNY

DATE/TIME OF TEST: 12/13/2017 02:09:39

Diagnostics Check		OK	02:03
Air Blank		0.000	02:03
Control Test		0.081	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.197	02:05
Breath Volume		1.496	02:05
Air Blank		0.000	02:05
Air Blank		0.000	02:07
Subject Sample	# 2	0.183	02:08
Breath Volume		1.546	02:08
Air Blank		0.000	02:08
Control Test		0.081	02:09
Air Blank		0.000	02:09
Diagnostics Check		OK	02:09

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: GREEN,J
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 12/13/2017 04:28:08

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:25	
Control Test		0.081	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:26	Subject Test Refused
Breath Volume		0.000	04:26	
Air Blank		0.000	04:27	
Control Test		0.081	04:27	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JONES,M
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: CUSHMAN

DATE/TIME OF TEST: 12/14/2017 02:22:54

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.081	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.000	02:18
Breath Volume		2.683	02:18
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.000	02:21
Breath Volume		2.367	02:21
Air Blank		0.000	02:21
Control Test		0.082	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:58

SUBJECT'S LAST NAME: SAMAYOAPEREZ

DATE/TIME OF TEST: 12/17/2017 02:21:49

Diagnostics Check		OK	02:19	
Air Blank		0.000	02:19	
Control Test		0.081	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	Subject Test Refused
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Control Test		0.081	02:21	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CARTER,C
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: ATKINSRICKETTS

DATE/TIME OF TEST: 12/17/2017 05:58:38

Diagnostics Check		OK	05:52
Air Blank		0.000	05:52
Control Test		0.081	05:53
Air Blank		0.000	05:53
Subject Sample	# 1	0.245	05:54
Breath Volume		2.359	05:54
Air Blank		0.000	05:54
Air Blank		0.000	05:56
Subject Sample	# 2	0.242	05:57
Breath Volume		2.546	05:57
Air Blank		0.000	05:57
Control Test		0.082	05:58
Air Blank		0.000	05:58
Diagnostics Check		OK	05:58

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:52

SUBJECT'S LAST NAME: FRICKE

DATE/TIME OF TEST: 12/17/2017 20:23:23

Diagnostics Check		OK	20:12	
Air Blank		0.000	20:13	
Control Test		0.081	20:13	
Air Blank		0.000	20:14	
Subject Sample	# 1	0.137	20:17	Volume Not Met
Breath Volume		0.292	20:17	
Air Blank		0.000	20:18	
Air Blank		0.000	20:19	
Subject Sample	# 2		20:21	Subject Test Refused
Breath Volume		0.226	20:21	
Air Blank		0.000	20:22	
Control Test		0.080	20:22	
Air Blank		0.000	20:23	
Diagnostics Check		OK	20:23	

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: WEST

DATE/TIME OF TEST: 12/17/2017 23:36:03

Diagnostics Check		OK	23:29
Air Blank		0.000	23:30
Control Test		0.081	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.150	23:31
Breath Volume		2.582	23:31
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.145	23:34
Breath Volume		2.546	23:34
Air Blank		0.000	23:35
Control Test		0.082	23:35
Air Blank		0.000	23:35
Diagnostics Check		OK	23:36

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001069

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POWELL,EDWARD,A
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: TUBBS

DATE/TIME OF TEST: 12/19/2017 00:32:04

Diagnostics Check		OK	00:25
Air Blank		0.000	00:26
Control Test		0.082	00:26
Air Blank		0.000	00:27
Subject Sample	# 1	0.095	00:27
Breath Volume		2.417	00:27
Air Blank		0.000	00:28
Air Blank		0.000	00:30
Subject Sample	# 2	0.095	00:30
Breath Volume		3.562	00:30
Air Blank		0.000	00:31
Control Test		0.081	00:31
Air Blank		0.000	00:31
Diagnostics Check		OK	00:32

Data Download Date/Time: 12/21/2017 15:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001070

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:26

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 12/31/2017 21:52:22

Diagnostics Check		OK	21:49	
Air Blank		0.000	21:49	
Control Test		0.078	21:50	
Air Blank		0.000	21:50	
Subject Sample	# 1		21:50	Subject Test Refused
Breath Volume		0.000	21:50	
Air Blank		0.000	21:51	
Control Test		0.077	21:51	
Air Blank		0.000	21:52	
Diagnostics Check		OK	21:52	

Data Download Date/Time: 01/17/2018 17:12

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:37

SUBJECT'S LAST NAME: BREWER

DATE/TIME OF TEST: 12/01/2017 17:09:54

Diagnostics Check		OK	17:03
Air Blank		0.000	17:03
Control Test		0.080	17:04
Air Blank		0.000	17:04
Subject Sample	# 1	0.000	17:05
Breath Volume		1.898	17:05
Air Blank		0.000	17:05
Air Blank		0.000	17:07
Subject Sample	# 2	0.000	17:08
Breath Volume		1.753	17:08
Air Blank		0.000	17:09
Control Test		0.080	17:09
Air Blank		0.000	17:09
Diagnostics Check		OK	17:09

Data Download Date/Time: 12/23/2017 16:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTIAGO,DEPUTY,
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: GADT,DEPUTY
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:50

SUBJECT'S LAST NAME: HILLIARD

DATE/TIME OF TEST: 12/06/2017 15:29:35

Diagnostics Check		OK	15:20
Air Blank		0.000	15:20
Control Test		0.079	15:21
Air Blank		0.000	15:21
Subject Sample	# 1	0.084	15:23
Breath Volume		1.761	15:23
Air Blank		0.000	15:24
Air Blank		0.000	15:26
Subject Sample	# 2	0.088	15:28
Breath Volume		1.835	15:28
Air Blank		0.000	15:28
Control Test		0.079	15:29
Air Blank		0.000	15:29
Diagnostics Check		OK	15:29

Data Download Date/Time: 12/23/2017 16:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,M,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHEELER,A
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: DANIELS

DATE/TIME OF TEST: 12/10/2017 01:59:31

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.081	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.174	01:54
Breath Volume		3.144	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.175	01:58
Breath Volume		3.007	01:58
Air Blank		0.000	01:58
Control Test		0.080	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 12/23/2017 16:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:22

SUBJECT'S LAST NAME: VERNON

DATE/TIME OF TEST: 12/12/2017 11:52:09

Diagnostics Check		OK	11:49	
Air Blank		0.000	11:49	
Control Test		0.080	11:50	
Air Blank		0.000	11:50	
Subject Sample	# 1		11:50	RFI Detect
Breath Volume		0.000	11:50	
Air Blank		0.000	11:51	
Control Test		0.080	11:51	
Air Blank		0.000	11:52	
Diagnostics Check		OK	11:52	

Data Download Date/Time: 12/23/2017 16:16

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROOKS,E,T
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PRITCHARD,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 10:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:22

SUBJECT'S LAST NAME: VERNON

DATE/TIME OF TEST: 12/12/2017 12:00:16

Diagnostics Check		OK	11:53
Air Blank		0.000	11:54
Control Test		0.080	11:54
Air Blank		0.000	11:55
Subject Sample	# 1	0.000	11:55
Breath Volume		2.410	11:55
Air Blank		0.000	11:56
Air Blank		0.000	11:58
Subject Sample	# 2	0.000	11:58
Breath Volume		2.378	11:58
Air Blank		0.000	11:59
Control Test		0.080	11:59
Air Blank		0.000	12:00
Diagnostics Check		OK	12:00

Data Download Date/Time: 12/23/2017 16:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 12/14/2017 01:06:54

Diagnostics Check		OK	01:00
Air Blank		0.000	01:01
Control Test		0.080	01:01
Air Blank		0.000	01:01
Subject Sample	# 1	0.186	01:02
Breath Volume		4.015	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:04
Subject Sample	# 2	0.179	01:05
Breath Volume		3.503	01:05
Air Blank		0.000	01:06
Control Test		0.079	01:06
Air Blank		0.000	01:06
Diagnostics Check		OK	01:06

Data Download Date/Time: 12/23/2017 16:17

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JOHNSON,G
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:10

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/14/2017 04:43:55

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:32	
Control Test		0.079	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1	0.213	04:33	
Breath Volume		2.121	04:33	
Air Blank		0.000	04:34	
Air Blank		0.000	04:35	
Subject Sample	# 2		04:39	No Sample Provided
Breath Volume		0.000	04:39	
Air Blank		0.000	04:39	
Air Blank		0.000	04:41	
Subject Sample	# 3	0.208	04:42	
Breath Volume		2.117	04:42	
Air Blank		0.000	04:43	
Control Test		0.080	04:43	
Air Blank		0.000	04:43	
Diagnostics Check		OK	04:43	

Data Download Date/Time: 12/23/2017 16:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: THEUS,J
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: WADSWORTH

DATE/TIME OF TEST: 12/15/2017 02:44:30

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.159	02:40
Breath Volume		2.542	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.155	02:43
Breath Volume		2.406	02:43
Air Blank		0.000	02:43
Control Test		0.079	02:43
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 12/23/2017 16:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TIMPE,J
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:07

SUBJECT'S LAST NAME: LANDRENEAU

DATE/TIME OF TEST: 12/18/2017 21:33:43

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.079	21:28
Air Blank		0.000	21:28
Subject Sample	# 1	0.203	21:29
Breath Volume		2.218	21:29
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.211	21:32
Breath Volume		2.375	21:32
Air Blank		0.000	21:32
Control Test		0.079	21:33
Air Blank		0.000	21:33
Diagnostics Check		OK	21:33

Data Download Date/Time: 12/23/2017 16:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,C,D
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: COX,Z
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:22

SUBJECT'S LAST NAME: PERROTTA

DATE/TIME OF TEST: 12/24/2017 21:50:08

Diagnostics Check		OK	21:43
Air Blank		0.000	21:44
Control Test		0.080	21:44
Air Blank		0.000	21:44
Subject Sample	# 1	0.322	21:45
Breath Volume		1.675	21:45
Air Blank		0.000	21:46
Air Blank		0.000	21:47
Subject Sample	# 2	0.304	21:48
Breath Volume		1.867	21:48
Air Blank		0.000	21:49
Control Test		0.080	21:49
Air Blank		0.000	21:50
Diagnostics Check		OK	21:50

Data Download Date/Time: 01/14/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 12/28/2017 00:37:30

Diagnostics Check		OK	00:31	
Air Blank		0.000	00:31	
Control Test		0.079	00:32	
Air Blank		0.000	00:32	
Subject Sample	# 1	0.164	00:33	
Breath Volume		1.656	00:33	
Air Blank		0.000	00:34	
Air Blank		0.000	00:36	
Subject Sample	# 2	0.160	00:36	
Breath Volume		1.828	00:36	
Air Blank			00:37	RFI Detect
Air Blank		0.000	00:37	

Data Download Date/Time: 01/14/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: TBER,T
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: HINES

DATE/TIME OF TEST: 12/28/2017 00:46:31

Diagnostics Check	OK	00:45	
Air Blank		00:46	RFI Detect
Air Blank	0.000	00:46	

Data Download Date/Time: 01/14/2018 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001071

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANTORO,CAMERON,J
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: JETT,C
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A2
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: SAAVEDRA PEREZ

DATE/TIME OF TEST: 12/28/2017 03:37:05

Diagnostics Check		OK	03:26
Air Blank		0.000	03:26
Control Test		0.079	03:27
Air Blank		0.000	03:27
Subject Sample	# 1	0.118	03:28
Breath Volume		1.417	03:28
Air Blank		0.000	03:29
Air Blank		0.000	03:31
Subject Sample	# 2	0.139	03:32
Breath Volume		2.238	03:32
Air Blank		0.000	03:33
Air Blank		0.000	03:34
Subject Sample	# 3	0.135	03:35
Breath Volume		2.250	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:37

Data Download Date/Time: 01/14/2018 14:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001073

INSTRUMENT REGISTERED TO: DADE CITY PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DOTY,DERRICK,J
BREATH TEST OPERATOR AGENCY: Dade City PD
ARRESTING OFFICER: LARKIN,JESSE
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A3
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: FINKBEINER

DATE/TIME OF TEST: 12/04/2017 05:36:20

Diagnostics Check		OK	05:30
Air Blank		0.000	05:30
Control Test		0.079	05:31
Air Blank		0.000	05:31
Subject Sample	# 1	0.000	05:32
Breath Volume		2.707	05:32
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.000	05:34
Breath Volume		3.320	05:34
Air Blank		0.000	05:35
Control Test		0.078	05:35
Air Blank		0.000	05:36
Diagnostics Check		OK	05:36

Data Download Date/Time: 12/13/2017 18:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI,ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 12/14/2017 02:21:54

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.079	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.138	02:17
Breath Volume		1.796	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:19
Subject Sample	# 2	0.143	02:20
Breath Volume		2.167	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 12/27/2017 16:37

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARQUES,JASON,R
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: STOCKTON,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:12

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/17/2017 10:40:56

Diagnostics Check		OK	10:34
Air Blank		0.000	10:34
Control Test		0.079	10:35
Air Blank		0.000	10:35
Subject Sample	# 1	0.000	10:36
Breath Volume		3.828	10:36
Air Blank		0.000	10:37
Air Blank		0.000	10:39
Subject Sample	# 2	0.000	10:39
Breath Volume		1.972	10:39
Air Blank		0.000	10:40
Control Test		0.079	10:40
Air Blank		0.000	10:40
Diagnostics Check		OK	10:40

Data Download Date/Time: 12/27/2017 16:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: MEIER,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:35

SUBJECT'S LAST NAME: SESSA

DATE/TIME OF TEST: 12/19/2017 07:02:10

Diagnostics Check		OK	06:55
Air Blank		0.000	06:56
Control Test		0.078	06:56
Air Blank		0.000	06:56
Subject Sample	# 1	0.132	06:57
Breath Volume		1.500	06:57
Air Blank		0.000	06:58
Air Blank		0.000	06:59
Subject Sample	# 2	0.135	07:00
Breath Volume		2.687	07:00
Air Blank		0.000	07:01
Control Test		0.079	07:01
Air Blank		0.000	07:02
Diagnostics Check		OK	07:02

Data Download Date/Time: 12/27/2017 16:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001074

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:25

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/19/2017 21:48:55

Diagnostics Check		OK	21:46	
Air Blank		0.000	21:46	
Control Test		0.079	21:46	
Air Blank		0.000	21:47	
Subject Sample	# 1		21:47	Subject Test Refused
Breath Volume		0.000	21:47	
Air Blank		0.000	21:48	
Control Test		0.078	21:48	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 12/27/2017 16:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001076

INSTRUMENT REGISTERED TO: ZEPHYRHILLS PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARDNER,NATHAN,R
BREATH TEST OPERATOR AGENCY: Zephyrhills PD
ARRESTING OFFICER: RICE,CALEB
ARRESTING OFFICER AGENCY: Zephyrhills PD
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 12/31/2017 04:23:12

Diagnostics Check		OK	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Subject Sample	# 1	0.199	04:18
Breath Volume		1.953	04:18
Air Blank		0.000	04:19
Air Blank		0.000	04:20
Subject Sample	# 2	0.201	04:21
Breath Volume		2.570	04:21
Air Blank		0.000	04:22
Control Test		0.078	04:22
Air Blank		0.000	04:23
Diagnostics Check		OK	04:23

Data Download Date/Time: 12/31/2017 11:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: MCMEEKIN

DATE/TIME OF TEST: 12/01/2017 21:51:08

Diagnostics Check		OK	21:45
Air Blank		0.000	21:45
Control Test		0.080	21:45
Air Blank		0.000	21:46
Subject Sample	# 1	0.148	21:46
Breath Volume		1.484	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:49
Subject Sample	# 2	0.149	21:49
Breath Volume		1.554	21:49
Air Blank		0.000	21:50
Control Test		0.080	21:50
Air Blank		0.000	21:51
Diagnostics Check		OK	21:51

Data Download Date/Time: 12/06/2017 10:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BATTLE

DATE/TIME OF TEST: 12/01/2017 23:28:04

Diagnostics Check		OK	23:21	
Air Blank		0.000	23:22	
Control Test		0.080	23:22	
Air Blank		0.000	23:23	
Subject Sample	# 1	0.095	23:23	
Breath Volume		2.316	23:23	
Air Blank		0.000	23:23	
Air Blank		0.000	23:25	
Subject Sample	# 2		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:27	
Control Test		0.080	23:27	
Air Blank		0.000	23:27	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 12/06/2017 10:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/02/2017 01:42:44

Diagnostics Check		OK	01:36
Air Blank		0.000	01:36
Control Test		0.080	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.000	01:38
Breath Volume		1.855	01:38
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.000	01:41
Breath Volume		1.726	01:41
Air Blank		0.000	01:41
Control Test		0.080	01:42
Air Blank		0.000	01:42
Diagnostics Check		OK	01:42

Data Download Date/Time: 12/06/2017 10:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 12/02/2017 02:50:05

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:49	
Control Test		0.081	02:49	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 12/06/2017 10:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ALLI,ANTHONY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:39

SUBJECT'S LAST NAME: FARINA

DATE/TIME OF TEST: 12/02/2017 18:32:36

Diagnostics Check		OK	18:29	
Air Blank		0.000	18:30	
Control Test		0.080	18:30	
Air Blank		0.000	18:31	
Subject Sample	# 1		18:31	Subject Test Refused
Breath Volume		0.000	18:31	
Air Blank		0.000	18:31	
Control Test		0.080	18:32	
Air Blank		0.000	18:32	
Diagnostics Check		OK	18:32	

Data Download Date/Time: 12/06/2017 10:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 11/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SANCHEZ

DATE/TIME OF TEST: 12/02/2017 23:57:55

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.080	23:51
Air Blank		0.000	23:52
Subject Sample	# 1	0.163	23:52
Breath Volume		1.523	23:52
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.147	23:56
Breath Volume		1.324	23:56
Air Blank		0.000	23:57
Control Test		0.080	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 12/06/2017 10:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALLI,ANTHONY,
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:16

SUBJECT'S LAST NAME: PANEK

DATE/TIME OF TEST: 12/09/2017 01:43:13

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.079	01:37
Air Blank		0.000	01:38
Subject Sample	# 1	0.218	01:38
Breath Volume		3.835	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.214	01:41
Breath Volume		3.882	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:45

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 12/13/2017 20:17:16

Diagnostics Check		OK	20:10
Air Blank		0.000	20:11
Control Test		0.081	20:11
Air Blank		0.000	20:12
Subject Sample	# 1	0.000	20:12
Breath Volume		3.445	20:12
Air Blank		0.000	20:13
Air Blank		0.000	20:15
Subject Sample	# 2	0.000	20:15
Breath Volume		3.304	20:15
Air Blank		0.000	20:16
Control Test		0.079	20:16
Air Blank		0.000	20:17
Diagnostics Check		OK	20:17

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: WILBURN

DATE/TIME OF TEST: 12/14/2017 21:31:05

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:28	
Control Test		0.080	21:28	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:29	Subject Test Refused
Breath Volume		0.000	21:29	
Air Blank		0.000	21:30	
Control Test		0.080	21:30	
Air Blank		0.000	21:30	
Diagnostics Check		OK	21:31	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:08

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 12/15/2017 19:31:32

Diagnostics Check		OK	19:28	
Air Blank		0.000	19:29	
Control Test		0.080	19:29	
Air Blank		0.000	19:29	
Subject Sample	# 1		19:30	Subject Test Refused
Breath Volume		0.000	19:30	
Air Blank		0.000	19:30	
Control Test		0.081	19:30	
Air Blank		0.000	19:31	
Diagnostics Check		OK	19:31	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 12/15/2017 22:18:14

Diagnostics Check		OK	22:13	
Air Blank		0.000	22:13	
Control Test		0.080	22:13	
Air Blank		0.000	22:14	
Subject Sample	# 1	0.189	22:14	
Breath Volume		3.449	22:14	
Air Blank		0.000	22:15	
Air Blank		0.000	22:17	
Subject Sample	# 2	0.175	22:17	
Breath Volume		2.562	22:17	
Air Blank			22:18	RFI Detect
Air Blank			22:18	RFI Detect

Data Download Date/Time: 01/17/2018 08:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: DREW

DATE/TIME OF TEST: 12/15/2017 22:27:47

Diagnostics Check		OK	22:21
Air Blank		0.000	22:21
Control Test		0.079	22:22
Air Blank		0.000	22:22
Subject Sample	# 1	0.179	22:23
Breath Volume		2.070	22:23
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.171	22:26
Breath Volume		1.835	22:26
Air Blank		0.000	22:26
Control Test		0.078	22:27
Air Blank		0.000	22:27
Diagnostics Check		OK	22:27

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: BETZ

DATE/TIME OF TEST: 12/15/2017 23:14:52

Diagnostics Check		OK	23:08
Air Blank		0.000	23:09
Control Test		0.080	23:09
Air Blank		0.000	23:10
Subject Sample	# 1	0.140	23:10
Breath Volume		1.750	23:10
Air Blank		0.000	23:11
Air Blank		0.000	23:12
Subject Sample	# 2	0.140	23:13
Breath Volume		1.792	23:13
Air Blank		0.000	23:13
Control Test		0.080	23:14
Air Blank		0.000	23:14
Diagnostics Check		OK	23:14

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:13

SUBJECT'S LAST NAME: RUHUMULIZA

DATE/TIME OF TEST: 12/16/2017 01:39:36

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:37	
Control Test		0.081	01:37	
Air Blank		0.000	01:38	
Subject Sample	# 1		01:38	Subject Test Refused
Breath Volume		0.000	01:38	
Air Blank		0.000	01:38	
Control Test		0.080	01:39	
Air Blank		0.000	01:39	
Diagnostics Check		OK	01:39	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 12/17/2017 02:38:06

Diagnostics Check		OK	02:30
Air Blank		0.000	02:31
Control Test		0.080	02:31
Air Blank		0.000	02:31
Subject Sample	# 1	0.164	02:32
Breath Volume		2.027	02:32
Air Blank		0.000	02:33
Air Blank		0.000	02:35
Subject Sample	# 2	0.176	02:36
Breath Volume		2.707	02:36
Air Blank		0.000	02:37
Control Test		0.078	02:37
Air Blank		0.000	02:38
Diagnostics Check		OK	02:38

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: LOVELACE

DATE/TIME OF TEST: 12/22/2017 22:00:08

Diagnostics Check		OK	21:53
Air Blank		0.000	21:54
Control Test		0.079	21:54
Air Blank		0.000	21:54
Subject Sample	# 1	0.127	21:55
Breath Volume		1.792	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.129	21:58
Breath Volume		2.097	21:58
Air Blank		0.000	21:59
Control Test		0.079	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KELLETT

DATE/TIME OF TEST: 12/23/2017 02:26:33

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.080	02:14	
Air Blank		0.000	02:15	
Subject Sample	# 1	0.254	02:15	
Breath Volume		1.527	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:18	
Subject Sample	# 2	0.197	02:21	Volume Not Met
Breath Volume		0.109	02:21	
Air Blank		0.000	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 3	0.249	02:24	
Breath Volume		1.242	02:24	
Air Blank		0.000	02:25	
Control Test		0.064	02:25	Control Outside Tolerance
Air Blank		0.000	02:26	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KELLETT

DATE/TIME OF TEST: 12/23/2017 02:33:03

Diagnostics Check	OK	02:31	
Air Blank	0.000	02:32	
Control Test	0.074	02:32	Control Outside Tolerance
Air Blank		02:33	RFI Detect

Data Download Date/Time: 01/17/2018 08:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: KELLETT

DATE/TIME OF TEST: 12/23/2017 02:39:41

Diagnostics Check		OK	02:36	
Air Blank		0.000	02:36	
Control Test		0.078	02:36	
Air Blank		0.000	02:37	
Subject Sample	# 1		02:38	Subject Test Refused
Breath Volume		0.000	02:38	
Air Blank		0.000	02:38	
Control Test		0.078	02:39	
Air Blank		0.000	02:39	
Diagnostics Check		OK	02:39	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: STUNZIG

DATE/TIME OF TEST: 12/29/2017 00:26:39

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:24	
Control Test		0.080	00:24	
Air Blank		0.000	00:24	
Subject Sample	# 1		00:25	Subject Test Refused
Breath Volume		0.000	00:25	
Air Blank		0.000	00:25	
Control Test		0.080	00:26	
Air Blank		0.000	00:26	
Diagnostics Check		OK	00:26	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:43

SUBJECT'S LAST NAME: DREHSEN

DATE/TIME OF TEST: 12/29/2017 23:13:36

Diagnostics Check		OK	23:03	
Air Blank		0.000	23:04	
Control Test		0.081	23:04	
Air Blank		0.000	23:04	
Subject Sample	# 1	0.209	23:05	
Breath Volume		2.453	23:05	
Air Blank		0.000	23:06	
Air Blank		0.000	23:07	
Subject Sample	# 2	0.231	23:08	
Breath Volume		1.957	23:08	
Air Blank		0.000	23:09	
Air Blank		0.000	23:11	
Subject Sample	# 3		23:12	RFI Detect
Breath Volume		0.000	23:12	
Air Blank		0.000	23:12	
Control Test		0.078	23:13	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: DREHSEN

DATE/TIME OF TEST: 12/29/2017 23:21:22

Diagnostics Check		OK	23:14
Air Blank		0.000	23:15
Control Test		0.079	23:15
Air Blank		0.000	23:16
Subject Sample	# 1	0.216	23:16
Breath Volume		1.980	23:16
Air Blank		0.000	23:17
Air Blank		0.000	23:19
Subject Sample	# 2	0.227	23:19
Breath Volume		2.492	23:19
Air Blank		0.000	23:20
Control Test		0.079	23:20
Air Blank		0.000	23:21
Diagnostics Check		OK	23:21

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 12/30/2017 02:04:56

Diagnostics Check		OK	01:58
Air Blank		0.000	01:59
Control Test		0.081	01:59
Air Blank		0.000	02:00
Subject Sample	# 1	0.112	02:00
Breath Volume		2.761	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.110	02:03
Breath Volume		3.156	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:04
Diagnostics Check		OK	02:04

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001078

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 12/06/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:15

SUBJECT'S LAST NAME: ALESSANDRO

DATE/TIME OF TEST: 12/30/2017 20:04:04

Diagnostics Check		OK	20:01	
Air Blank		0.000	20:01	
Control Test		0.081	20:02	
Air Blank		0.000	20:02	
Subject Sample	# 1		20:02	Subject Test Refused
Breath Volume		0.000	20:02	
Air Blank		0.000	20:03	
Control Test		0.081	20:03	
Air Blank		0.000	20:03	
Diagnostics Check		OK	20:04	

Data Download Date/Time: 01/17/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:39

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 12/01/2017 01:03:18

Diagnostics Check		OK	00:59	
Air Blank		0.000	01:00	
Control Test		0.080	01:00	
Air Blank		0.000	01:01	
Subject Sample	# 1		01:01	Subject Test Refused
Breath Volume		0.000	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:03	
Diagnostics Check		OK	01:03	

Data Download Date/Time: 12/18/2017 17:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ALFOND

DATE/TIME OF TEST: 12/01/2017 04:28:42

Diagnostics Check		OK	04:25	
Air Blank		0.000	04:26	
Control Test		0.080	04:26	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:27	Subject Test Refused
Breath Volume		0.000	04:27	
Air Blank		0.000	04:27	
Control Test		0.080	04:28	
Air Blank		0.000	04:28	
Diagnostics Check		OK	04:28	

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HUNT,JERRY
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 08:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 09:49

SUBJECT'S LAST NAME: KAFERLEIN

DATE/TIME OF TEST: 12/01/2017 10:20:32

Diagnostics Check		OK	10:13
Air Blank		0.000	10:13
Control Test		0.081	10:13
Air Blank		0.000	10:14
Subject Sample	# 1	0.000	10:16
Breath Volume		1.597	10:16
Air Blank		0.000	10:16
Air Blank		0.000	10:18
Subject Sample	# 2	0.000	10:19
Breath Volume		1.742	10:19
Air Blank		0.000	10:19
Control Test		0.080	10:19
Air Blank		0.000	10:20
Diagnostics Check		OK	10:20

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,JERRY,D
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: ASBURY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:49

SUBJECT'S LAST NAME: KAFERLEIN

DATE/TIME OF TEST: 12/01/2017 10:35:34

Diagnostics Check		OK	10:29
Air Blank		0.000	10:29
Control Test		0.080	10:30
Air Blank		0.000	10:30
Subject Sample	# 1	0.000	10:31
Breath Volume		3.339	10:31
Air Blank		0.000	10:31
Air Blank		0.000	10:33
Subject Sample	# 2	0.000	10:34
Breath Volume		3.019	10:34
Air Blank		0.000	10:34
Control Test		0.080	10:34
Air Blank		0.000	10:35
Diagnostics Check		OK	10:35

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:41

SUBJECT'S LAST NAME: CORWIN

DATE/TIME OF TEST: 12/01/2017 22:08:16

Diagnostics Check		OK	22:02
Air Blank		0.000	22:02
Control Test		0.081	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.117	22:03
Breath Volume		1.808	22:03
Air Blank		0.000	22:04
Air Blank		0.000	22:06
Subject Sample	# 2	0.128	22:06
Breath Volume		1.812	22:06
Air Blank		0.000	22:07
Control Test		0.081	22:07
Air Blank		0.000	22:08
Diagnostics Check		OK	22:08

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: MADGE

DATE/TIME OF TEST: 12/01/2017 23:18:18

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.172	23:13
Breath Volume		3.101	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.163	23:16
Breath Volume		3.230	23:16
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: SCHOENHERR

DATE/TIME OF TEST: 12/03/2017 02:33:46

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.080	02:27
Air Blank		0.000	02:28
Subject Sample	# 1	0.113	02:29
Breath Volume		3.527	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.114	02:32
Breath Volume		2.625	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: JUNG

DATE/TIME OF TEST: 12/03/2017 02:57:07

Diagnostics Check		OK	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Subject Sample	# 1	0.162	02:52
Breath Volume		2.269	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.164	02:55
Breath Volume		1.734	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:57
Diagnostics Check		OK	02:57

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:33

SUBJECT'S LAST NAME: BIRD

DATE/TIME OF TEST: 12/08/2017 01:00:21

Diagnostics Check		OK	00:53
Air Blank		0.000	00:54
Control Test		0.081	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.166	00:55
Breath Volume		1.488	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.162	00:58
Breath Volume		1.632	00:58
Air Blank		0.000	00:59
Control Test		0.081	00:59
Air Blank		0.000	01:00
Diagnostics Check		OK	01:00

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:22

SUBJECT'S LAST NAME: GUENTHER

DATE/TIME OF TEST: 12/09/2017 01:51:21

Diagnostics Check		OK	01:44
Air Blank		0.000	01:45
Control Test		0.081	01:45
Air Blank		0.000	01:46
Subject Sample	# 1	0.186	01:46
Breath Volume		2.128	01:46
Air Blank		0.000	01:47
Air Blank		0.000	01:49
Subject Sample	# 2	0.196	01:49
Breath Volume		2.687	01:49
Air Blank		0.000	01:50
Control Test		0.081	01:50
Air Blank		0.000	01:51
Diagnostics Check		OK	01:51

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: NUNLEY,JARED
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:05

SUBJECT'S LAST NAME: KLEIN

DATE/TIME OF TEST: 12/10/2017 15:32:43

Diagnostics Check		OK	15:26
Air Blank		0.000	15:26
Control Test		0.081	15:27
Air Blank		0.000	15:27
Subject Sample	# 1	0.000	15:28
Breath Volume		1.347	15:28
Air Blank		0.000	15:28
Air Blank		0.000	15:30
Subject Sample	# 2	0.000	15:31
Breath Volume		1.566	15:31
Air Blank		0.000	15:31
Control Test		0.082	15:32
Air Blank		0.000	15:32
Diagnostics Check		OK	15:32

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: RICE,JOHN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: EMANUEL

DATE/TIME OF TEST: 12/12/2017 22:12:00

Diagnostics Check		OK	22:09	
Air Blank		0.000	22:09	
Control Test		0.080	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 1		22:10	Subject Test Refused
Breath Volume		0.000	22:10	
Air Blank		0.000	22:11	
Control Test		0.081	22:11	
Air Blank		0.000	22:11	
Diagnostics Check		OK	22:12	

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,JOHN,B
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: GRASSIA,ANTONIO
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 12/12/2017 22:18:14

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:15	
Control Test		0.080	22:16	
Air Blank		0.000	22:16	
Subject Sample	# 1		22:16	Subject Test Refused
Breath Volume		0.000	22:16	
Air Blank		0.000	22:17	
Control Test		0.080	22:17	
Air Blank		0.000	22:18	
Diagnostics Check		OK	22:18	

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:38

SUBJECT'S LAST NAME: HAHN

DATE/TIME OF TEST: 12/13/2017 23:02:50

Diagnostics Check		OK	22:59	
Air Blank		0.000	23:00	
Control Test		0.081	23:00	
Air Blank		0.000	23:01	
Subject Sample	# 1		23:01	Subject Test Refused
Breath Volume		0.000	23:01	
Air Blank		0.000	23:01	
Control Test		0.081	23:02	
Air Blank		0.000	23:02	
Diagnostics Check		OK	23:02	

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: VAN AUKEN

DATE/TIME OF TEST: 12/14/2017 23:33:15

Diagnostics Check		OK	23:26
Air Blank		0.000	23:26
Control Test		0.080	23:27
Air Blank		0.000	23:27
Subject Sample	# 1	0.224	23:28
Breath Volume		1.816	23:28
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.227	23:31
Breath Volume		2.027	23:31
Air Blank		0.000	23:32
Control Test		0.080	23:32
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: WAGNER

DATE/TIME OF TEST: 12/14/2017 23:46:27

Diagnostics Check		OK	23:40
Air Blank		0.000	23:40
Control Test		0.080	23:40
Air Blank		0.000	23:41
Subject Sample	# 1	0.164	23:41
Breath Volume		2.722	23:41
Air Blank		0.000	23:42
Air Blank		0.000	23:44
Subject Sample	# 2	0.158	23:44
Breath Volume		2.417	23:44
Air Blank		0.000	23:45
Control Test		0.079	23:45
Air Blank		0.000	23:46
Diagnostics Check		OK	23:46

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: MACK

DATE/TIME OF TEST: 12/15/2017 03:19:45

Diagnostics Check		OK	03:11
Air Blank		0.000	03:12
Control Test		0.080	03:12
Air Blank		0.000	03:13
Subject Sample	# 1	0.164	03:15
Breath Volume		3.003	03:15
Air Blank		0.000	03:15
Air Blank		0.000	03:17
Subject Sample	# 2	0.156	03:18
Breath Volume		2.363	03:18
Air Blank		0.000	03:18
Control Test		0.080	03:19
Air Blank		0.000	03:19
Diagnostics Check		OK	03:19

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:07

SUBJECT'S LAST NAME: CONLEY

DATE/TIME OF TEST: 12/16/2017 02:39:37

Diagnostics Check		OK	02:28
Air Blank		0.000	02:28
Control Test		0.081	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.180	02:30
Breath Volume		1.917	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:33
Subject Sample	# 2	0.205	02:34
Breath Volume		2.414	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 3	0.191	02:38
Breath Volume		2.093	02:38
Air Blank		0.000	02:38
Control Test		0.080	02:39
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 12/18/2017 17:09

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: ROWE

DATE/TIME OF TEST: 12/16/2017 02:53:34

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.213	02:49
Breath Volume		1.347	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.233	02:52
Breath Volume		1.500	02:52
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 11/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:41

SUBJECT'S LAST NAME: QUEVEDO FUENTES

DATE/TIME OF TEST: 12/17/2017 03:06:18

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:02	
Control Test		0.081	03:02	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.081	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 12/18/2017 17:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: ROBLERO ROBLERO

DATE/TIME OF TEST: 12/21/2017 01:08:04

Diagnostics Check		OK	00:59
Air Blank		0.000	01:00
Control Test		0.081	01:00
Air Blank		0.000	01:00
Subject Sample	# 1	0.165	01:02
Breath Volume		3.039	01:02
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.164	01:06
Breath Volume		2.988	01:06
Air Blank		0.000	01:07
Control Test		0.080	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:08

Data Download Date/Time: 01/22/2018 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BALDIE,BRYAN,A
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: KATZ,JUSTIN
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: GIBBS

DATE/TIME OF TEST: 12/21/2017 22:16:42

Diagnostics Check		OK	22:11	
Air Blank		0.000	22:11	
Control Test		0.081	22:11	
Air Blank		0.000	22:12	
Subject Sample	# 1		22:15	Subject Test Refused
Breath Volume		0.000	22:15	
Air Blank		0.000	22:15	
Control Test		0.080	22:16	
Air Blank		0.000	22:16	
Diagnostics Check		OK	22:16	

Data Download Date/Time: 01/22/2018 11:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 12/22/2017 04:33:45

Diagnostics Check		OK	04:25
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.215	04:28
Breath Volume		2.507	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.206	04:32
Breath Volume		2.449	04:32
Air Blank		0.000	04:32
Control Test		0.081	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 01/22/2018 11:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:34

SUBJECT'S LAST NAME: DE PINTO

DATE/TIME OF TEST: 12/23/2017 01:01:49

Diagnostics Check		OK	00:54
Air Blank		0.000	00:55
Control Test		0.080	00:55
Air Blank		0.000	00:56
Subject Sample	# 1	0.170	00:57
Breath Volume		2.292	00:57
Air Blank		0.000	00:57
Air Blank		0.000	00:59
Subject Sample	# 2	0.176	01:00
Breath Volume		2.757	01:00
Air Blank		0.000	01:00
Control Test		0.080	01:01
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/22/2018 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 12/23/2017 04:39:10

Diagnostics Check		OK	04:27	
Air Blank		0.000	04:28	
Control Test		0.081	04:28	
Air Blank		0.000	04:28	
Subject Sample	# 1		04:32	No Sample Provided
Breath Volume		0.000	04:32	
Air Blank		0.000	04:32	
Air Blank		0.000	04:34	
Subject Sample	# 2	0.134	04:37	Volume Not Met
Breath Volume		0.050	04:37	
Air Blank		0.000	04:38	
Control Test		0.079	04:38	
Air Blank		0.000	04:39	
Diagnostics Check		OK	04:39	

Data Download Date/Time: 01/22/2018 11:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: GUZMAN

DATE/TIME OF TEST: 12/23/2017 04:54:37

Diagnostics Check		OK	04:43	
Air Blank		0.000	04:43	
Control Test		0.080	04:43	
Air Blank		0.000	04:44	
Subject Sample	# 1	0.131	04:47	Volume Not Met
Breath Volume		0.015	04:47	
Air Blank		0.000	04:48	
Air Blank		0.000	04:49	
Subject Sample	# 2		04:53	No Sample Provided
Breath Volume		0.000	04:53	
Air Blank		0.000	04:53	
Control Test		0.079	04:54	
Air Blank		0.000	04:54	
Diagnostics Check		OK	04:54	

Data Download Date/Time: 01/22/2018 11:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SILVERSTEIN,CHRISTINA,
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: SILVERSTEIN,CHRISTINA
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: SILVERSTEIN

DATE/TIME OF TEST: 12/24/2017 00:39:03

Diagnostics Check		OK	00:35	
Air Blank		0.000	00:35	
Control Test		0.080	00:36	
Air Blank		0.000	00:36	
Subject Sample	# 1		00:37	Subject Test Refused
Breath Volume		0.000	00:37	
Air Blank		0.000	00:38	
Control Test		0.080	00:38	
Air Blank		0.000	00:38	
Diagnostics Check		OK	00:39	

Data Download Date/Time: 01/22/2018 11:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001079

INSTRUMENT REGISTERED TO: LARGO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HASTINGS,KENNETH,S
BREATH TEST OPERATOR AGENCY: Largo PD
ARRESTING OFFICER: HASTINGS,KENNETH
ARRESTING OFFICER AGENCY: Largo PD
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: WATTS

DATE/TIME OF TEST: 12/28/2017 23:43:34

Diagnostics Check		OK	23:40	
Air Blank		0.000	23:41	
Control Test		0.081	23:41	
Air Blank		0.000	23:41	
Subject Sample	# 1		23:42	Subject Test Refused
Breath Volume		0.000	23:42	
Air Blank		0.000	23:42	
Control Test		0.081	23:42	
Air Blank		0.000	23:43	
Diagnostics Check		OK	23:43	

Data Download Date/Time: 01/22/2018 11:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR,ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GOODWALD,DARREN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: FREDRICKSON

DATE/TIME OF TEST: 12/01/2017 00:08:16

Diagnostics Check		OK	23:55	
Air Blank		0.000	23:55	
Control Test		0.079	23:56	
Air Blank		0.000	23:56	
Subject Sample	# 3	0.137	23:59	Volume Not Met
Breath Volume		0.316	23:59	
Air Blank		0.000	00:00	
Air Blank		0.000	00:02	
Subject Sample	# 1	0.157	00:03	No .020 agreement
Breath Volume		1.246	00:03	
Air Blank		0.000	00:03	
Air Blank		0.000	00:05	
Subject Sample	# 2	0.178	00:06	No .020 agreement
Breath Volume		1.121	00:06	
Air Blank		0.000	00:07	
Control Test		0.079	00:07	
Air Blank		0.000	00:08	
Diagnostics Check		OK	00:08	

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MONSEGUR,ROBERT,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: GOODWALD,DARREN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:55

SUBJECT'S LAST NAME: FREDRICKSON

DATE/TIME OF TEST: 12/01/2017 00:25:11

Diagnostics Check		OK	00:16
Air Blank		0.000	00:17
Control Test		0.078	00:17
Air Blank		0.000	00:18
Subject Sample	# 1	0.167	00:20
Breath Volume		2.261	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.165	00:23
Breath Volume		1.457	00:23
Air Blank		0.000	00:24
Control Test		0.078	00:24
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 12/03/2017 00:20:25

Diagnostics Check		OK	00:12
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Subject Sample	# 1	0.196	00:15
Breath Volume		1.128	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.201	00:18
Breath Volume		1.175	00:18
Air Blank		0.000	00:19
Control Test		0.077	00:19
Air Blank		0.000	00:20
Diagnostics Check		OK	00:20

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,WILLIAM,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GREEN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:23

SUBJECT'S LAST NAME: TOMLINSON

DATE/TIME OF TEST: 12/05/2017 19:49:31

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.080	19:43
Air Blank		0.000	19:44
Subject Sample	# 1	0.085	19:44
Breath Volume		1.339	19:44
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.085	19:48
Breath Volume		1.453	19:48
Air Blank		0.000	19:48
Control Test		0.078	19:48
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROWAND,ART,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CRAWFORD,CHRIS
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: SHASHURA

DATE/TIME OF TEST: 12/06/2017 20:09:24

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.080	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.240	20:02
Breath Volume		2.218	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.211	20:04
Breath Volume		1.210	20:04
Air Blank		0.000	20:05
Air Blank		0.000	20:07
Subject Sample	# 3	0.227	20:07
Breath Volume		2.515	20:07
Air Blank		0.000	20:08
Control Test		0.079	20:08
Air Blank		0.000	20:09
Diagnostics Check		OK	20:09

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: ALBRIGHT

DATE/TIME OF TEST: 12/07/2017 21:33:00

Diagnostics Check		OK	21:27
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:28
Subject Sample	# 1	0.000	21:28
Breath Volume		3.695	21:28
Air Blank		0.000	21:29
Air Blank		0.000	21:31
Subject Sample	# 2	0.000	21:31
Breath Volume		4.128	21:31
Air Blank		0.000	21:32
Control Test		0.079	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:33

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,WILLIAM,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:31

SUBJECT'S LAST NAME: NEWKIRK

DATE/TIME OF TEST: 12/08/2017 14:59:20

Diagnostics Check		OK	14:51
Air Blank		0.000	14:52
Control Test		0.080	14:52
Air Blank		0.000	14:52
Subject Sample	# 1	0.149	14:53
Breath Volume		2.148	14:53
Air Blank		0.000	14:54
Air Blank		0.000	14:55
Subject Sample	# 2	0.158	14:57
Breath Volume		2.683	14:57
Air Blank		0.000	14:58
Control Test		0.079	14:58
Air Blank		0.000	14:59
Diagnostics Check		OK	14:59

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: EDWARDS

DATE/TIME OF TEST: 12/09/2017 02:31:26

Diagnostics Check		OK	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Subject Sample	# 1	0.000	02:27
Breath Volume		3.777	02:27
Air Blank		0.000	02:27
Air Blank		0.000	02:29
Subject Sample	# 2	0.000	02:30
Breath Volume		3.117	02:30
Air Blank		0.000	02:30
Control Test		0.079	02:30
Air Blank		0.000	02:31
Diagnostics Check		OK	02:31

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/09/2017 20:57:40

Diagnostics Check		OK	20:51
Air Blank		0.000	20:51
Control Test		0.081	20:52
Air Blank		0.000	20:52
Subject Sample	# 1	0.000	20:53
Breath Volume		1.265	20:53
Air Blank		0.000	20:53
Air Blank		0.000	20:55
Subject Sample	# 2	0.000	20:56
Breath Volume		1.386	20:56
Air Blank		0.000	20:56
Control Test		0.080	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CINELLI,ADAM,D
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CINELLI,ADAM
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: LISLE

DATE/TIME OF TEST: 12/10/2017 03:13:56

Diagnostics Check		OK	03:10	
Air Blank		0.000	03:10	
Control Test		0.080	03:11	
Air Blank		0.000	03:11	
Subject Sample	# 1		03:12	Subject Test Refused
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	
Control Test		0.080	03:13	
Air Blank		0.000	03:13	
Diagnostics Check		OK	03:13	

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: HAPPLE

DATE/TIME OF TEST: 12/11/2017 04:16:36

Diagnostics Check		OK	04:10
Air Blank		0.000	04:10
Control Test		0.080	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.204	04:12
Breath Volume		2.375	04:12
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.196	04:15
Breath Volume		2.640	04:15
Air Blank		0.000	04:15
Control Test		0.078	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IVEY,BRIAN,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: CRAWFORD,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: ZAMORA

DATE/TIME OF TEST: 12/11/2017 23:04:42

Diagnostics Check		OK	22:58
Air Blank		0.000	22:59
Control Test		0.079	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.238	23:00
Breath Volume		2.125	23:00
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.229	23:03
Breath Volume		1.902	23:03
Air Blank		0.000	23:03
Control Test		0.078	23:04
Air Blank		0.000	23:04
Diagnostics Check		OK	23:04

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,DANIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:27

SUBJECT'S LAST NAME: DESHAZO

DATE/TIME OF TEST: 12/12/2017 17:54:22

Diagnostics Check		OK	17:48
Air Blank		0.000	17:48
Control Test		0.080	17:48
Air Blank		0.000	17:49
Subject Sample	# 1	0.106	17:49
Breath Volume		1.601	17:49
Air Blank		0.000	17:50
Air Blank		0.000	17:52
Subject Sample	# 2	0.107	17:52
Breath Volume		1.878	17:52
Air Blank		0.000	17:53
Control Test		0.080	17:53
Air Blank		0.000	17:54
Diagnostics Check		OK	17:54

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: SLADE

DATE/TIME OF TEST: 12/12/2017 22:16:25

Diagnostics Check		OK	22:10
Air Blank		0.000	22:10
Control Test		0.080	22:11
Air Blank		0.000	22:11
Subject Sample	# 1	0.151	22:11
Breath Volume		2.523	22:11
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.149	22:14
Breath Volume		2.589	22:14
Air Blank		0.000	22:15
Control Test		0.080	22:15
Air Blank		0.000	22:16
Diagnostics Check		OK	22:16

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,DANIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: NISI

DATE/TIME OF TEST: 12/14/2017 02:04:04

Diagnostics Check		OK	02:01	
Air Blank		0.000	02:01	
Control Test		0.079	02:01	
Air Blank		0.000	02:02	
Subject Sample	# 1		02:02	Subject Test Refused
Breath Volume		0.000	02:02	
Air Blank		0.000	02:03	
Control Test		0.079	02:03	
Air Blank		0.000	02:03	
Diagnostics Check		OK	02:04	

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: VORIS

DATE/TIME OF TEST: 12/14/2017 04:10:53

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.078	04:04
Air Blank		0.000	04:05
Subject Sample	# 1	0.203	04:06
Breath Volume		2.460	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.194	04:09
Breath Volume		1.574	04:09
Air Blank		0.000	04:09
Control Test		0.079	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:10

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRISCO,RICKY,A
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DOMINGUEZ,JAMES
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:40

SUBJECT'S LAST NAME: LAMPARSKI

DATE/TIME OF TEST: 12/14/2017 08:10:27

Diagnostics Check		OK	08:04
Air Blank		0.000	08:04
Control Test		0.080	08:05
Air Blank		0.000	08:05
Subject Sample	# 1	0.103	08:06
Breath Volume		2.058	08:06
Air Blank		0.000	08:06
Air Blank		0.000	08:08
Subject Sample	# 2	0.102	08:08
Breath Volume		2.066	08:08
Air Blank		0.000	08:09
Control Test		0.080	08:09
Air Blank		0.000	08:10
Diagnostics Check		OK	08:10

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: KLIPPEL

DATE/TIME OF TEST: 12/14/2017 21:32:50

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.081	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.000	21:28
Breath Volume		1.597	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.000	21:31
Breath Volume		1.488	21:31
Air Blank		0.000	21:31
Control Test		0.080	21:32
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,JACQUES,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LANG,JACQUES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: PIZARRO

DATE/TIME OF TEST: 12/15/2017 00:39:30

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.204	00:34
Breath Volume		1.359	00:34
Air Blank		0.000	00:35
Air Blank		0.000	00:37
Subject Sample	# 2	0.202	00:37
Breath Volume		1.617	00:37
Air Blank		0.000	00:38
Control Test		0.079	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: GRIGATO

DATE/TIME OF TEST: 12/15/2017 04:09:46

Diagnostics Check		OK	04:00
Air Blank		0.000	04:01
Control Test		0.079	04:01
Air Blank		0.000	04:02
Subject Sample	# 1	0.165	04:03
Breath Volume		2.394	04:03
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.170	04:08
Breath Volume		1.125	04:08
Air Blank		0.000	04:08
Control Test		0.077	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: SHEPPARD

DATE/TIME OF TEST: 12/15/2017 23:54:15

Diagnostics Check		OK	23:47
Air Blank		0.000	23:48
Control Test		0.079	23:48
Air Blank		0.000	23:48
Subject Sample	# 1	0.234	23:49
Breath Volume		2.382	23:49
Air Blank		0.000	23:50
Air Blank		0.000	23:51
Subject Sample	# 2	0.249	23:52
Breath Volume		3.105	23:52
Air Blank		0.000	23:53
Control Test		0.078	23:53
Air Blank		0.000	23:54
Diagnostics Check		OK	23:54

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: IRWIN,JEFFREY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: FRANCIS

DATE/TIME OF TEST: 12/16/2017 04:08:38

Diagnostics Check		OK	04:02
Air Blank		0.000	04:02
Control Test		0.079	04:03
Air Blank		0.000	04:03
Subject Sample	# 1	0.147	04:04
Breath Volume		1.289	04:04
Air Blank		0.000	04:04
Air Blank		0.000	04:06
Subject Sample	# 2	0.151	04:07
Breath Volume		2.238	04:07
Air Blank		0.000	04:07
Control Test		0.078	04:08
Air Blank		0.000	04:08
Diagnostics Check		OK	04:08

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:44

SUBJECT'S LAST NAME: POSEY

DATE/TIME OF TEST: 12/17/2017 02:15:32

Diagnostics Check		OK	02:05
Air Blank		0.000	02:05
Control Test		0.079	02:05
Air Blank		0.000	02:06
Subject Sample	# 1	0.234	02:07
Breath Volume		3.152	02:07
Air Blank		0.000	02:07
Air Blank		0.000	02:09
Subject Sample	# 2	0.194	02:10
Breath Volume		1.300	02:10
Air Blank		0.000	02:11
Air Blank		0.000	02:13
Subject Sample	# 3	0.213	02:14
Breath Volume		2.101	02:14
Air Blank		0.000	02:14
Control Test		0.079	02:14
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANDO,CHRISTOPHER,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRANDO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 07:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:00

SUBJECT'S LAST NAME: DEWITT

DATE/TIME OF TEST: 12/17/2017 08:07:26

Diagnostics Check		OK	08:00
Air Blank		0.000	08:01
Control Test		0.079	08:01
Air Blank		0.000	08:02
Subject Sample	# 1	0.189	08:03
Breath Volume		1.410	08:03
Air Blank		0.000	08:03
Air Blank		0.000	08:05
Subject Sample	# 2	0.180	08:05
Breath Volume		1.140	08:05
Air Blank		0.000	08:06
Control Test		0.078	08:06
Air Blank		0.000	08:07
Diagnostics Check		OK	08:07

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANG,JACQUES,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: LANG,JACQUES
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: BUZBY

DATE/TIME OF TEST: 12/18/2017 21:59:16

Diagnostics Check		OK	21:52
Air Blank		0.000	21:53
Control Test		0.080	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.139	21:54
Breath Volume		1.507	21:54
Air Blank		0.000	21:55
Air Blank		0.000	21:57
Subject Sample	# 2	0.143	21:57
Breath Volume		1.285	21:57
Air Blank		0.000	21:58
Control Test		0.079	21:58
Air Blank		0.000	21:59
Diagnostics Check		OK	21:59

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: KENYON

DATE/TIME OF TEST: 12/19/2017 21:32:16

Diagnostics Check		OK	21:26
Air Blank		0.000	21:26
Control Test		0.080	21:26
Air Blank		0.000	21:27
Subject Sample	# 1	0.000	21:28
Breath Volume		3.574	21:28
Air Blank		0.000	21:28
Air Blank		0.000	21:30
Subject Sample	# 2	0.000	21:30
Breath Volume		3.449	21:30
Air Blank		0.000	21:31
Control Test		0.080	21:31
Air Blank		0.000	21:32
Diagnostics Check		OK	21:32

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,NELSON,X
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: RODRIGUEZ,NELSON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: GRANGER

DATE/TIME OF TEST: 12/20/2017 03:06:09

Diagnostics Check		OK	03:03	
Air Blank		0.000	03:03	
Control Test		0.079	03:04	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:05	
Control Test		0.076	03:05	
Air Blank		0.000	03:06	
Diagnostics Check		OK	03:06	

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRIOLAT,CASEY,J
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BRIOLAT,CASEY
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 06:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: MCCORRY

DATE/TIME OF TEST: 12/21/2017 07:47:16

Diagnostics Check		OK	07:40
Air Blank		0.000	07:40
Control Test		0.080	07:41
Air Blank		0.000	07:41
Subject Sample	# 1	0.147	07:42
Breath Volume		2.640	07:42
Air Blank		0.000	07:43
Air Blank		0.000	07:45
Subject Sample	# 2	0.147	07:45
Breath Volume		3.910	07:45
Air Blank		0.000	07:46
Control Test		0.079	07:46
Air Blank		0.000	07:47
Diagnostics Check		OK	07:47

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: COTTER

DATE/TIME OF TEST: 12/22/2017 02:47:04

Diagnostics Check		OK	02:40
Air Blank		0.000	02:41
Control Test		0.080	02:41
Air Blank		0.000	02:42
Subject Sample	# 1	0.000	02:42
Breath Volume		2.812	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:45
Subject Sample	# 2	0.000	02:45
Breath Volume		2.453	02:45
Air Blank		0.000	02:46
Control Test		0.079	02:46
Air Blank		0.000	02:46
Diagnostics Check		OK	02:47

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,JACOB,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BLYTH,JACOB
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: BURTT

DATE/TIME OF TEST: 12/23/2017 01:48:46

Diagnostics Check		OK	01:41
Air Blank		0.000	01:42
Control Test		0.080	01:42
Air Blank		0.000	01:43
Subject Sample	# 1	0.157	01:44
Breath Volume		1.925	01:44
Air Blank		0.000	01:45
Air Blank		0.000	01:46
Subject Sample	# 2	0.173	01:47
Breath Volume		1.753	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:48
Air Blank		0.000	01:48
Diagnostics Check		OK	01:48

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: SMALLEY

DATE/TIME OF TEST: 12/23/2017 03:41:13

Diagnostics Check		OK	03:34
Air Blank		0.000	03:35
Control Test		0.080	03:35
Air Blank		0.000	03:36
Subject Sample	# 1	0.054	03:36
Breath Volume		2.496	03:36
Air Blank		0.000	03:37
Air Blank		0.000	03:39
Subject Sample	# 2	0.058	03:39
Breath Volume		3.214	03:39
Air Blank		0.000	03:40
Control Test		0.080	03:40
Air Blank		0.000	03:41
Diagnostics Check		OK	03:41

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROWN,JASON,G
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: BROWN,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: GAMLEN

DATE/TIME OF TEST: 12/24/2017 03:42:04

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.080	03:36
Air Blank		0.000	03:36
Subject Sample	# 1	0.131	03:37
Breath Volume		3.031	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.118	03:40
Breath Volume		1.476	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:41
Diagnostics Check		OK	03:42

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: CAULK,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: SOULIERE

DATE/TIME OF TEST: 12/25/2017 19:21:10

Diagnostics Check		OK	19:12	
Air Blank		0.000	19:13	
Control Test		0.080	19:13	
Air Blank		0.000	19:13	
Subject Sample	# 1		19:17	No Sample Provided
Breath Volume		0.000	19:17	
Air Blank		0.000	19:17	
Air Blank		0.000	19:19	
Subject Sample	# 2		19:19	Subject Test Refused
Breath Volume		0.000	19:19	
Air Blank		0.000	19:20	
Control Test		0.079	19:20	
Air Blank		0.000	19:21	
Diagnostics Check		OK	19:21	

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUGO-FELIX,DANNY,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LUGO-FELIX,DANNY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: SARIC

DATE/TIME OF TEST: 12/25/2017 20:45:23

Diagnostics Check		OK	20:37
Air Blank		0.000	20:37
Control Test		0.080	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.183	20:39
Breath Volume		2.578	20:39
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.183	20:43
Breath Volume		3.039	20:43
Air Blank		0.000	20:44
Control Test		0.080	20:44
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COFFEE,JASON,A
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: COFFEE,JASON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 11/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MCCABE

DATE/TIME OF TEST: 12/27/2017 00:59:04

Diagnostics Check		OK	00:52
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.000	00:54
Breath Volume		1.574	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:57
Subject Sample	# 2	0.000	00:57
Breath Volume		1.484	00:57
Air Blank		0.000	00:58
Control Test		0.079	00:58
Air Blank		0.000	00:58
Diagnostics Check		OK	00:59

Data Download Date/Time: 12/27/2017 16:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NOTTINGHAM,CHRISTOPHER,R
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NOTTINGHAM,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 11:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: LAMANA

DATE/TIME OF TEST: 12/28/2017 00:53:29

Diagnostics Check		OK	00:46
Air Blank		0.000	00:46
Control Test		0.080	00:47
Air Blank		0.000	00:47
Subject Sample	# 1	0.165	00:48
Breath Volume		2.757	00:48
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.159	00:51
Breath Volume		2.933	00:51
Air Blank		0.000	00:52
Control Test		0.080	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HODGSON

DATE/TIME OF TEST: 12/30/2017 01:15:15

Diagnostics Check		OK	01:03	
Air Blank		0.000	01:04	
Control Test		0.080	01:04	
Air Blank		0.000	01:04	
Subject Sample	# 1	0.180	01:08	Volume Not Met
Breath Volume		0.238	01:08	
Air Blank		0.000	01:08	
Air Blank		0.000	01:10	
Subject Sample	# 2	0.186	01:13	Volume Not Met
Breath Volume		0.250	01:13	
Air Blank		0.000	01:14	
Control Test		0.078	01:14	
Air Blank		0.000	01:15	
Diagnostics Check		OK	01:15	

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,RAEFORD,L
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: GRIFFIN,RAEFORD
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: HODGSON

DATE/TIME OF TEST: 12/30/2017 01:27:49

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:25	
Control Test		0.079	01:25	
Air Blank		0.000	01:26	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:26	
Control Test		0.080	01:27	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: WINEMAN

DATE/TIME OF TEST: 12/30/2017 02:34:56

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:30
Subject Sample	# 1	0.129	02:30
Breath Volume		2.210	02:30
Air Blank		0.000	02:31
Air Blank		0.000	02:32
Subject Sample	# 2	0.138	02:33
Breath Volume		3.246	02:33
Air Blank		0.000	02:34
Control Test		0.079	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: MUTZABAUGH

DATE/TIME OF TEST: 12/30/2017 03:12:38

Diagnostics Check		OK	03:06
Air Blank		0.000	03:06
Control Test		0.079	03:07
Air Blank		0.000	03:07
Subject Sample	# 1	0.069	03:08
Breath Volume		3.886	03:08
Air Blank		0.000	03:08
Air Blank		0.000	03:10
Subject Sample	# 2	0.070	03:11
Breath Volume		3.710	03:11
Air Blank		0.000	03:11
Control Test		0.079	03:12
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: POLSON,BRIAN
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: NEELEN

DATE/TIME OF TEST: 12/30/2017 03:36:24

Diagnostics Check		OK	03:25
Air Blank		0.000	03:25
Control Test		0.079	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.199	03:27
Breath Volume		2.019	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.174	03:32
Breath Volume		1.246	03:32
Air Blank		0.000	03:32
Air Blank		0.000	03:34
Subject Sample	# 3	0.168	03:34
Breath Volume		1.464	03:34
Air Blank		0.000	03:35
Control Test		0.077	03:35
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 01/23/2018 13:41

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAUGHERTY,DANIEL,M
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: DAUGHERTY,DANIEL
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 12/31/2017 01:07:39

Diagnostics Check		OK	01:01
Air Blank		0.000	01:01
Control Test		0.079	01:02
Air Blank		0.000	01:02
Subject Sample	# 1	0.198	01:03
Breath Volume		2.800	01:03
Air Blank		0.000	01:03
Air Blank		0.000	01:05
Subject Sample	# 2	0.205	01:06
Breath Volume		3.523	01:06
Air Blank		0.000	01:06
Control Test		0.078	01:07
Air Blank		0.000	01:07
Diagnostics Check		OK	01:07

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KNORR,PETER,J
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: KNORR,PETER
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: GRIFFITH

DATE/TIME OF TEST: 12/31/2017 02:22:47

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.078	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.224	02:18
Breath Volume		3.121	02:18
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.229	02:21
Breath Volume		3.371	02:21
Air Blank		0.000	02:21
Control Test		0.079	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: SQUIRE

DATE/TIME OF TEST: 12/31/2017 03:13:29

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.080	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.074	03:09
Breath Volume		1.906	03:09
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.071	03:12
Breath Volume		1.343	03:12
Air Blank		0.000	03:12
Control Test		0.079	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,MARC,T
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: LANE,MARC
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/31/2017 05:45:14

Diagnostics Check		OK	05:35
Air Blank		0.000	05:35
Control Test		0.080	05:36
Air Blank		0.000	05:36
Subject Sample	# 1	0.148	05:37
Breath Volume		3.062	05:37
Air Blank		0.000	05:38
Air Blank		0.000	05:39
Subject Sample	# 2	0.169	05:40
Breath Volume		3.734	05:40
Air Blank		0.000	05:41
Air Blank		0.000	05:43
Subject Sample	# 3	0.156	05:43
Breath Volume		2.617	05:43
Air Blank		0.000	05:44
Control Test		0.079	05:44
Air Blank		0.000	05:45
Diagnostics Check		OK	05:45

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,ROBERT,R
BREATH TEST OPERATOR AGENCY: Port Richey PD
ARRESTING OFFICER: BABCOCK,EDWARD
ARRESTING OFFICER AGENCY: Port Richey PD
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:17

SUBJECT'S LAST NAME: BULLOCK

DATE/TIME OF TEST: 12/31/2017 20:43:57

Diagnostics Check		OK	20:37
Air Blank		0.000	20:37
Control Test		0.080	20:38
Air Blank		0.000	20:38
Subject Sample	# 1	0.000	20:39
Breath Volume		1.996	20:39
Air Blank		0.000	20:39
Air Blank		0.000	20:41
Subject Sample	# 2	0.000	20:42
Breath Volume		2.226	20:42
Air Blank		0.000	20:43
Control Test		0.080	20:43
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001081

INSTRUMENT REGISTERED TO: PASCO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,BARRY,
BREATH TEST OPERATOR AGENCY: Pasco County SO
ARRESTING OFFICER: NIXON,BARRY
ARRESTING OFFICER AGENCY: Pasco County SO
LAST AGENCY INSPECTION: 12/27/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: PETRO

DATE/TIME OF TEST: 12/31/2017 23:12:29

Diagnostics Check		OK	23:06
Air Blank		0.000	23:06
Control Test		0.080	23:07
Air Blank		0.000	23:07
Subject Sample	# 1	0.000	23:08
Breath Volume		2.175	23:08
Air Blank		0.000	23:08
Air Blank		0.000	23:10
Subject Sample	# 2	0.000	23:11
Breath Volume		2.070	23:11
Air Blank		0.000	23:11
Control Test		0.080	23:11
Air Blank		0.000	23:12
Diagnostics Check		OK	23:12

Data Download Date/Time: 01/23/2018 13:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAMPBELL,STUART,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: VEGA,BRENNAN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: HENRY

DATE/TIME OF TEST: 12/04/2017 23:58:08

Diagnostics Check		OK	23:51
Air Blank		0.000	23:51
Control Test		0.077	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.172	23:53
Breath Volume		2.878	23:53
Air Blank		0.000	23:54
Air Blank		0.000	23:56
Subject Sample	# 2	0.159	23:56
Breath Volume		1.566	23:56
Air Blank		0.000	23:57
Control Test		0.077	23:57
Air Blank		0.000	23:58
Diagnostics Check		OK	23:58

Data Download Date/Time: 12/28/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,ERIC,C
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: VEGA,BRENNAN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:46

SUBJECT'S LAST NAME: LEZAMA

DATE/TIME OF TEST: 12/11/2017 19:15:51

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.078	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.309	19:10
Breath Volume		1.257	19:10
Air Blank		0.000	19:11
Air Blank		0.000	19:13
Subject Sample	# 2	0.307	19:14
Breath Volume		2.660	19:14
Air Blank		0.000	19:14
Control Test		0.076	19:15
Air Blank		0.000	19:15
Diagnostics Check		OK	19:15

Data Download Date/Time: 12/28/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,BRENNAN,R
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:31

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 12/14/2017 21:06:26

Diagnostics Check		OK	20:51	
Air Blank		0.000	20:52	
Control Test		0.078	20:52	
Air Blank		0.000	20:52	
Subject Sample	# 1	0.192	20:56	Volume Not Met
Breath Volume		0.074	20:56	
Air Blank		0.000	20:56	
Air Blank		0.000	20:58	
Subject Sample	# 2	0.216	20:58	
Breath Volume		1.640	20:58	
Air Blank		0.000	20:59	
Air Blank		0.000	21:01	
Subject Sample	# 3	0.173	21:04	Volume Not Met
Breath Volume		1.023	21:04	
Air Blank		0.000	21:05	
Control Test		0.076	21:05	
Air Blank		0.000	21:06	
Diagnostics Check		OK	21:06	

Data Download Date/Time: 12/28/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VEGA,BRENNAN,R
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: FARQUHARSON,KEVIN
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 11/23/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:49

SUBJECT'S LAST NAME: ALEXANDRE

DATE/TIME OF TEST: 12/14/2017 22:24:07

Diagnostics Check		OK	22:13
Air Blank		0.000	22:13
Control Test		0.077	22:13
Air Blank		0.000	22:14
Subject Sample	# 1	0.204	22:15
Breath Volume		1.226	22:15
Air Blank		0.000	22:16
Air Blank		0.000	22:18
Subject Sample	# 2	0.177	22:18
Breath Volume		1.679	22:18
Air Blank		0.000	22:19
Air Blank		0.000	22:21
Subject Sample	# 3	0.190	22:22
Breath Volume		1.996	22:22
Air Blank		0.000	22:23
Control Test		0.077	22:23
Air Blank		0.000	22:24
Diagnostics Check		OK	22:24

Data Download Date/Time: 12/28/2017 19:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001082

INSTRUMENT REGISTERED TO: PLANTATION PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BROSKIE,OFC. J.,
BREATH TEST OPERATOR AGENCY: Plantation PD
ARRESTING OFFICER: BEAUREGARD,OFC. J.
ARRESTING OFFICER AGENCY: Plantation PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A2
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: OSULLIVAN

DATE/TIME OF TEST: 12/28/2017 22:29:00

Diagnostics Check		OK	22:15	
Air Blank		0.000	22:16	
Control Test		0.077	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 1	0.126	22:20	Volume Not Met
Breath Volume		0.699	22:20	
Air Blank		0.000	22:21	
Air Blank		0.000	22:22	
Subject Sample	# 2	0.126	22:24	
Breath Volume		1.707	22:24	
Air Blank		0.000	22:24	
Air Blank		0.000	22:26	
Subject Sample	# 3	0.131	22:27	
Breath Volume		2.265	22:27	
Air Blank		0.000	22:28	
Control Test		0.076	22:28	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:29	

Data Download Date/Time: 01/25/2018 15:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISSETT,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: COVIELLO

DATE/TIME OF TEST: 12/03/2017 23:18:41

Diagnostics Check		OK	23:12
Air Blank		0.000	23:13
Control Test		0.080	23:13
Air Blank		0.000	23:13
Subject Sample	# 1	0.144	23:14
Breath Volume		2.234	23:14
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.150	23:17
Breath Volume		1.992	23:17
Air Blank		0.000	23:17
Control Test		0.080	23:18
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 12/11/2017 15:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:35
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: KNIGHT

DATE/TIME OF TEST: 12/04/2017 03:47:44

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.081	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.108	03:43
Breath Volume		1.222	03:43
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.113	03:46
Breath Volume		1.820	03:46
Air Blank		0.000	03:46
Control Test		0.080	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:47

Data Download Date/Time: 12/11/2017 15:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DINKINS,CHRISTOPHER,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:22

SUBJECT'S LAST NAME: HALIGOWSKI

DATE/TIME OF TEST: 12/06/2017 20:49:07

Diagnostics Check		OK	20:43
Air Blank		0.000	20:43
Control Test		0.080	20:43
Air Blank		0.000	20:44
Subject Sample	# 1	0.000	20:44
Breath Volume		3.480	20:44
Air Blank		0.000	20:45
Air Blank		0.000	20:47
Subject Sample	# 2	0.000	20:47
Breath Volume		3.113	20:47
Air Blank		0.000	20:48
Control Test		0.080	20:48
Air Blank		0.000	20:49
Diagnostics Check		OK	20:49

Data Download Date/Time: 12/11/2017 15:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 12/09/2017 02:16:13

Diagnostics Check		OK	02:13	
Air Blank		0.000	02:13	
Control Test		0.080	02:13	
Air Blank		0.000	02:14	
Subject Sample	# 1		02:14	RFI Detect
Breath Volume		0.000	02:14	
Air Blank		0.000	02:15	
Control Test		0.080	02:15	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 12/11/2017 15:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: EANES,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: DE LA CRUZ

DATE/TIME OF TEST: 12/09/2017 02:24:47

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.081	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.179	02:20
Breath Volume		2.421	02:20
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.195	02:23
Breath Volume		1.226	02:23
Air Blank		0.000	02:23
Control Test		0.080	02:24
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 12/11/2017 15:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:51

SUBJECT'S LAST NAME: PECK

DATE/TIME OF TEST: 12/10/2017 23:18:18

Diagnostics Check		OK	23:11
Air Blank		0.000	23:12
Control Test		0.081	23:12
Air Blank		0.000	23:13
Subject Sample	# 1	0.072	23:13
Breath Volume		2.878	23:13
Air Blank		0.000	23:14
Air Blank		0.000	23:16
Subject Sample	# 2	0.071	23:16
Breath Volume		3.242	23:16
Air Blank		0.000	23:17
Control Test		0.080	23:17
Air Blank		0.000	23:18
Diagnostics Check		OK	23:18

Data Download Date/Time: 12/11/2017 15:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 11/17/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 12/11/2017 04:05:12

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.081	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.099	04:00
Breath Volume		1.179	04:00
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.096	04:03
Breath Volume		2.253	04:03
Air Blank		0.000	04:04
Control Test		0.082	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 12/11/2017 15:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FILBERT,E
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 12/12/2017 01:18:23

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:16	
Control Test		0.081	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:17	Subject Test Refused
Breath Volume		0.000	01:17	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: HART

DATE/TIME OF TEST: 12/15/2017 02:10:51

Diagnostics Check		OK	02:07	
Air Blank		0.000	02:08	
Control Test		0.079	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:09	Subject Test Refused
Breath Volume		0.000	02:09	
Air Blank		0.000	02:09	
Control Test		0.080	02:10	
Air Blank		0.000	02:10	
Diagnostics Check		OK	02:10	

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERTS,CHUCK,E
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:48

SUBJECT'S LAST NAME: ROSI KOKELJ

DATE/TIME OF TEST: 12/15/2017 04:14:44

Diagnostics Check		OK	04:08
Air Blank		0.000	04:09
Control Test		0.080	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.124	04:10
Breath Volume		2.535	04:10
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.124	04:13
Breath Volume		2.386	04:13
Air Blank		0.000	04:13
Control Test		0.079	04:14
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,KEVIN,W
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GARCIA

DATE/TIME OF TEST: 12/16/2017 01:23:29

Diagnostics Check		OK	01:20	
Air Blank		0.000	01:21	
Control Test		0.081	01:21	
Air Blank		0.000	01:21	
Subject Sample	# 1		01:22	Subject Test Refused
Breath Volume		0.000	01:22	
Air Blank		0.000	01:22	
Control Test		0.081	01:22	
Air Blank		0.000	01:23	
Diagnostics Check		OK	01:23	

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CAMACHO,K
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: ZEILER

DATE/TIME OF TEST: 12/20/2017 02:36:25

Diagnostics Check		OK	02:29
Air Blank		0.000	02:30
Control Test		0.080	02:30
Air Blank		0.000	02:31
Subject Sample	# 1	0.102	02:31
Breath Volume		1.664	02:31
Air Blank		0.000	02:32
Air Blank		0.000	02:34
Subject Sample	# 2	0.107	02:34
Breath Volume		2.023	02:34
Air Blank		0.000	02:35
Control Test		0.080	02:35
Air Blank		0.000	02:36
Diagnostics Check		OK	02:36

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: VAN MOURIK

DATE/TIME OF TEST: 12/20/2017 04:32:35

Diagnostics Check		OK	04:29	
Air Blank		0.000	04:30	
Control Test		0.080	04:30	
Air Blank		0.000	04:31	
Subject Sample	# 1		04:31	Subject Test Refused
Breath Volume		0.000	04:31	
Air Blank		0.000	04:31	
Control Test		0.080	04:31	
Air Blank		0.000	04:32	
Diagnostics Check		OK	04:32	

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 12/21/2017 03:09:06

Diagnostics Check		OK	03:02
Air Blank		0.000	03:03
Control Test		0.080	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.139	03:04
Breath Volume		1.437	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.145	03:07
Breath Volume		2.128	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURNS,D,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:57

SUBJECT'S LAST NAME: CARMONA

DATE/TIME OF TEST: 12/21/2017 04:27:39

Diagnostics Check		OK	04:24	
Air Blank		0.000	04:25	
Control Test		0.080	04:25	
Air Blank		0.000	04:26	
Subject Sample	# 1		04:26	Subject Test Refused
Breath Volume		0.000	04:26	
Air Blank		0.000	04:26	
Control Test		0.081	04:27	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: CEPEDA

DATE/TIME OF TEST: 12/23/2017 00:43:39

Diagnostics Check		OK	00:37
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:39
Subject Sample	# 1	0.000	00:39
Breath Volume		1.851	00:39
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.000	00:42
Breath Volume		2.183	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:43
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AUGUSTINE,G,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/11/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:57

SUBJECT'S LAST NAME: GUARD

DATE/TIME OF TEST: 12/23/2017 03:24:13

Diagnostics Check		OK	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:19
Subject Sample	# 1	0.139	03:19
Breath Volume		2.257	03:19
Air Blank		0.000	03:20
Air Blank		0.000	03:22
Subject Sample	# 2	0.134	03:22
Breath Volume		2.609	03:22
Air Blank		0.000	03:23
Control Test		0.080	03:23
Air Blank		0.000	03:24
Diagnostics Check		OK	03:24

Data Download Date/Time: 12/29/2017 14:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HOOVER,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:25

SUBJECT'S LAST NAME: GREEN

DATE/TIME OF TEST: 12/29/2017 23:52:19

Diagnostics Check		OK	23:45
Air Blank		0.000	23:46
Control Test		0.080	23:46
Air Blank		0.000	23:47
Subject Sample	# 1	0.253	23:47
Breath Volume		4.167	23:47
Air Blank		0.000	23:48
Air Blank		0.000	23:50
Subject Sample	# 2	0.252	23:50
Breath Volume		4.296	23:50
Air Blank		0.000	23:51
Control Test		0.079	23:51
Air Blank		0.000	23:52
Diagnostics Check		OK	23:52

Data Download Date/Time: 01/19/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: N,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUSSELL,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: JORRIN JR FLORES

DATE/TIME OF TEST: 12/30/2017 04:20:13

Diagnostics Check		OK	04:13
Air Blank		0.000	04:13
Control Test		0.081	04:14
Air Blank		0.000	04:14
Subject Sample	# 1	0.216	04:15
Breath Volume		1.949	04:15
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.207	04:18
Breath Volume		1.972	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 01/19/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: WHITE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: CHANEY

DATE/TIME OF TEST: 12/30/2017 05:25:30

Diagnostics Check		OK	05:19
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Subject Sample	# 1	0.262	05:21
Breath Volume		3.046	05:21
Air Blank		0.000	05:21
Air Blank		0.000	05:23
Subject Sample	# 2	0.255	05:23
Breath Volume		2.851	05:23
Air Blank		0.000	05:24
Control Test		0.080	05:24
Air Blank		0.000	05:25
Diagnostics Check		OK	05:25

Data Download Date/Time: 01/19/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: ULYSSE,S
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: ALBERT

DATE/TIME OF TEST: 12/31/2017 01:50:50

Diagnostics Check		OK	01:44
Air Blank		0.000	01:44
Control Test		0.080	01:45
Air Blank		0.000	01:45
Subject Sample	# 1	0.131	01:46
Breath Volume		2.054	01:46
Air Blank		0.000	01:46
Air Blank		0.000	01:48
Subject Sample	# 2	0.138	01:49
Breath Volume		2.003	01:49
Air Blank		0.000	01:49
Control Test		0.080	01:50
Air Blank		0.000	01:50
Diagnostics Check		OK	01:50

Data Download Date/Time: 01/19/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001083

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAND,JARRETT,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ESTES

DATE/TIME OF TEST: 12/31/2017 04:44:09

Diagnostics Check		OK	04:36
Air Blank		0.000	04:37
Control Test		0.081	04:37
Air Blank		0.000	04:38
Subject Sample	# 1	0.082	04:39
Breath Volume		1.296	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.081	04:42
Breath Volume		1.179	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 01/19/2018 14:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 15:15

SUBJECT'S LAST NAME: SEAGROVES

DATE/TIME OF TEST: 12/05/2017 15:42:30

Diagnostics Check		OK	15:36
Air Blank		0.000	15:36
Control Test		0.081	15:36
Air Blank		0.000	15:37
Subject Sample	# 1	0.312	15:37
Breath Volume		1.500	15:37
Air Blank		0.000	15:38
Air Blank		0.000	15:40
Subject Sample	# 2	0.302	15:41
Breath Volume		1.320	15:41
Air Blank		0.000	15:41
Control Test		0.082	15:41
Air Blank		0.000	15:42
Diagnostics Check		OK	15:42

Data Download Date/Time: 12/28/2017 07:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PIERCE,BRAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 07:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 09:13

SUBJECT'S LAST NAME: GRINDLE

DATE/TIME OF TEST: 12/06/2017 09:41:01

Diagnostics Check		OK	09:34
Air Blank		0.000	09:35
Control Test		0.080	09:35
Air Blank		0.000	09:35
Subject Sample	# 1	0.000	09:36
Breath Volume		2.148	09:36
Air Blank		0.000	09:37
Air Blank		0.000	09:39
Subject Sample	# 2	0.000	09:39
Breath Volume		2.191	09:39
Air Blank		0.000	09:40
Control Test		0.081	09:40
Air Blank		0.000	09:40
Diagnostics Check		OK	09:41

Data Download Date/Time: 12/28/2017 07:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINN,CAMERON,B
BREATH TEST OPERATOR AGENCY: Panama City PD
ARRESTING OFFICER: BONNER,ALEC
ARRESTING OFFICER AGENCY: Panama City PD
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 09:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:01

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 12/06/2017 11:31:50

Diagnostics Check		OK	11:21	
Air Blank		0.000	11:21	
Control Test		0.081	11:21	
Air Blank		0.000	11:22	
Subject Sample	# 1	0.000	11:25	Volume Not Met
Breath Volume		0.390	11:25	
Air Blank		0.000	11:26	
Air Blank		0.000	11:28	
Subject Sample	# 2		11:30	Subject Test Refused
Breath Volume		0.136	11:30	
Air Blank		0.000	11:30	
Control Test		0.082	11:31	
Air Blank		0.000	11:31	
Diagnostics Check		OK	11:31	

Data Download Date/Time: 12/28/2017 07:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,TIMOTHY,B
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: PIERCE,BRAD
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 05:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:45

SUBJECT'S LAST NAME: WILLIS

DATE/TIME OF TEST: 12/11/2017 07:14:07

Diagnostics Check		OK	07:06
Air Blank		0.000	07:06
Control Test		0.083	07:06
Air Blank		0.000	07:07
Subject Sample	# 1	0.148	07:09
Breath Volume		1.406	07:09
Air Blank		0.000	07:09
Air Blank		0.000	07:11
Subject Sample	# 2	0.151	07:12
Breath Volume		1.511	07:12
Air Blank		0.000	07:13
Control Test		0.082	07:13
Air Blank		0.000	07:14
Diagnostics Check		OK	07:14

Data Download Date/Time: 12/28/2017 07:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,KEVIN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: JOHNSON,JUSTIN
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:07

SUBJECT'S LAST NAME: BUSH JR

DATE/TIME OF TEST: 12/14/2017 20:44:53

Diagnostics Check		OK	20:37	
Air Blank		0.000	20:37	
Control Test		0.081	20:38	
Air Blank		0.000	20:38	
Subject Sample	# 1	0.317	20:39	
Breath Volume		1.203	20:39	
Air Blank		0.000	20:40	
Air Blank		0.000	20:42	
Subject Sample	# 2		20:43	Subject Test Refused
Breath Volume		0.730	20:43	
Air Blank		0.000	20:43	
Control Test		0.082	20:44	
Air Blank		0.000	20:44	
Diagnostics Check		OK	20:44	

Data Download Date/Time: 12/28/2017 07:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOWALSKI,KEVIN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: KOWALSKI,KEVIN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:07

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 12/19/2017 22:34:58

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:29
Subject Sample	# 1	0.205	22:30
Breath Volume		1.316	22:30
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.207	22:33
Breath Volume		1.691	22:33
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:34
Diagnostics Check		OK	22:34

Data Download Date/Time: 12/28/2017 07:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SISK,CARLOS,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: SISK,CARLOS
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 19:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: LAW

DATE/TIME OF TEST: 12/21/2017 20:11:22

Diagnostics Check		OK	20:03
Air Blank		0.000	20:03
Control Test		0.081	20:03
Air Blank		0.000	20:04
Subject Sample	# 1	0.221	20:05
Breath Volume		2.078	20:05
Air Blank		0.000	20:06
Air Blank		0.000	20:08
Subject Sample	# 2	0.218	20:09
Breath Volume		1.878	20:09
Air Blank		0.000	20:10
Control Test		0.082	20:10
Air Blank		0.000	20:11
Diagnostics Check		OK	20:11

Data Download Date/Time: 12/28/2017 07:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JONATHAN,P
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: GENTRY,JONATHAN
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/01/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 10:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 11:17

SUBJECT'S LAST NAME: SIMPSON

DATE/TIME OF TEST: 12/23/2017 11:52:51

Diagnostics Check		OK	11:38	
Air Blank		0.000	11:38	
Control Test		0.081	11:39	
Air Blank		0.000	11:39	
Subject Sample	# 1	0.248	11:40	
Breath Volume		1.242	11:40	
Air Blank		0.000	11:41	
Air Blank		0.000	11:43	
Subject Sample	# 2	0.229	11:46	Volume Not Met
Breath Volume		0.242	11:46	
Air Blank		0.000	11:47	
Air Blank		0.000	11:48	
Subject Sample	# 3	0.241	11:51	
Breath Volume		2.488	11:51	
Air Blank		0.000	11:51	
Control Test		0.080	11:52	
Air Blank		0.000	11:52	
Diagnostics Check		OK	11:52	

Data Download Date/Time: 12/28/2017 07:33

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAILEY,RAYMOND,F
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BAILEY,RAYMOND
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: GRAHAM

DATE/TIME OF TEST: 12/30/2017 00:36:51

Diagnostics Check		OK	00:23	
Air Blank		0.000	00:24	
Control Test		0.082	00:24	
Air Blank		0.000	00:24	
Subject Sample	# 1	0.108	00:27	Volume Not Met
Breath Volume		0.472	00:27	
Air Blank		0.000	00:28	
Air Blank		0.000	00:30	
Subject Sample	# 2	0.101	00:32	
Breath Volume		1.371	00:32	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 3	0.107	00:35	
Breath Volume		1.894	00:35	
Air Blank		0.000	00:35	
Control Test		0.082	00:36	
Air Blank		0.000	00:36	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 01/16/2018 19:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001111

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YASUDA,SHOHEI,
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: YOUNG,ALEX
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 712031
DRY GAS EXPIRY DATE: 02/21/2020
VIOLATION CODE: DUI
ARREST TIME: 22:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:05

SUBJECT'S LAST NAME: RIVERS

DATE/TIME OF TEST: 12/31/2017 23:28:19

Diagnostics Check		OK	23:25	
Air Blank		0.000	23:26	
Control Test		0.082	23:26	
Air Blank		0.000	23:26	
Subject Sample	# 1		23:26	Subject Test Refused
Breath Volume		0.000	23:26	
Air Blank		0.000	23:27	
Control Test		0.082	23:27	
Air Blank		0.000	23:28	
Diagnostics Check		OK	23:28	

Data Download Date/Time: 01/16/2018 19:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: LEDESMA

DATE/TIME OF TEST: 12/03/2017 04:24:11

Diagnostics Check		OK	04:17
Air Blank		0.000	04:18
Control Test		0.082	04:18
Air Blank		0.000	04:19
Subject Sample	# 1	0.111	04:19
Breath Volume		2.300	04:19
Air Blank		0.000	04:20
Air Blank		0.000	04:22
Subject Sample	# 2	0.113	04:22
Breath Volume		3.218	04:22
Air Blank		0.000	04:23
Control Test		0.082	04:23
Air Blank		0.000	04:24
Diagnostics Check		OK	04:24

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MCNEELY,STEVE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:45

SUBJECT'S LAST NAME: RAMSEY

DATE/TIME OF TEST: 12/04/2017 01:23:13

Diagnostics Check		OK	01:17
Air Blank		0.000	01:17
Control Test		0.082	01:17
Air Blank		0.000	01:18
Subject Sample	# 1	0.102	01:18
Breath Volume		2.230	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:21
Subject Sample	# 2	0.103	01:21
Breath Volume		2.402	01:21
Air Blank		0.000	01:22
Control Test		0.082	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WENZEL,THOMAS,J
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: TAYLOR,CHANNING
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 10:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:35

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 12/13/2017 11:36:23

Diagnostics Check		OK	11:30
Air Blank		0.000	11:30
Control Test		0.082	11:31
Air Blank		0.000	11:31
Subject Sample	# 1	0.000	11:32
Breath Volume		2.031	11:32
Air Blank		0.000	11:32
Air Blank		0.000	11:34
Subject Sample	# 2	0.000	11:35
Breath Volume		2.046	11:35
Air Blank		0.000	11:35
Control Test		0.082	11:35
Air Blank		0.000	11:36
Diagnostics Check		OK	11:36

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STANLEY,JEFFREY,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: STANLEY,JEFFREY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: GEANACOPOULOS

DATE/TIME OF TEST: 12/16/2017 05:53:16

Diagnostics Check		OK	05:43
Air Blank		0.000	05:44
Control Test		0.082	05:44
Air Blank		0.000	05:44
Subject Sample	# 1	0.171	05:46
Breath Volume		1.152	05:46
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.170	05:51
Breath Volume		1.250	05:51
Air Blank		0.000	05:52
Control Test		0.082	05:52
Air Blank		0.000	05:53
Diagnostics Check		OK	05:53

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: VALDES,DAVID,S
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: VALDES,DAVID
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:40

SUBJECT'S LAST NAME: GONZALEZ

DATE/TIME OF TEST: 12/16/2017 20:23:08

Diagnostics Check		OK	20:15
Air Blank		0.000	20:16
Control Test		0.082	20:16
Air Blank		0.000	20:17
Subject Sample	# 1	0.152	20:18
Breath Volume		2.265	20:18
Air Blank		0.000	20:18
Air Blank		0.000	20:20
Subject Sample	# 2	0.141	20:21
Breath Volume		1.792	20:21
Air Blank		0.000	20:22
Control Test		0.079	20:22
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: MACLEOD

DATE/TIME OF TEST: 12/18/2017 04:14:14

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.259	04:09
Breath Volume		3.039	04:09
Air Blank		0.000	04:10
Air Blank		0.000	04:12
Subject Sample	# 2	0.263	04:12
Breath Volume		2.457	04:12
Air Blank		0.000	04:13
Control Test		0.081	04:13
Air Blank		0.000	04:14
Diagnostics Check		OK	04:14

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 12/24/2017 04:00:40

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.082	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.157	03:56
Breath Volume		3.113	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.154	03:59
Breath Volume		1.984	03:59
Air Blank		0.000	03:59
Control Test		0.081	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001112

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULLIGAN,WILLIAM,C
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: MULLIGAN,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 11/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:12

SUBJECT'S LAST NAME: MATTESON

DATE/TIME OF TEST: 12/26/2017 02:37:48

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:35	
Control Test		0.082	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:36	Subject Test Refused
Breath Volume		0.000	02:36	
Air Blank		0.000	02:36	
Control Test		0.083	02:37	
Air Blank		0.000	02:37	
Diagnostics Check		OK	02:37	

Data Download Date/Time: 12/31/2017 04:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER,WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: PARRIS

DATE/TIME OF TEST: 12/10/2017 20:44:47

Diagnostics Check	OK	20:44	
Air Blank		20:44	Ambient Fail
Air Blank		20:44	Improper Sample

Data Download Date/Time: 12/30/2017 14:36

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER,WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: PARRIS

DATE/TIME OF TEST: 12/10/2017 20:48:52

Diagnostics Check	OK	20:48	
Air Blank		20:48	Improper Sample
Air Blank	0.000	20:48	

Data Download Date/Time: 12/30/2017 14:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOPER,WILLIAM,J
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: COOPER,WILLIAM
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 19:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: PARRIS

DATE/TIME OF TEST: 12/10/2017 20:57:55

Diagnostics Check		OK	20:51
Air Blank		0.000	20:52
Control Test		0.077	20:52
Air Blank		0.000	20:53
Subject Sample	# 1	0.201	20:53
Breath Volume		2.007	20:53
Air Blank		0.000	20:54
Air Blank		0.000	20:56
Subject Sample	# 2	0.213	20:56
Breath Volume		1.859	20:56
Air Blank		0.000	20:56
Control Test		0.078	20:57
Air Blank		0.000	20:57
Diagnostics Check		OK	20:57

Data Download Date/Time: 12/30/2017 14:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CODY,RICHARD,D
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: CODY,RICHARD
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/11/2017 00:07:39

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.078	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:06	Subject Test Refused
Breath Volume		0.000	00:06	
Air Blank		0.000	00:06	
Control Test		0.079	00:07	
Air Blank		0.000	00:07	
Diagnostics Check		OK	00:07	

Data Download Date/Time: 12/30/2017 14:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: BICE,BRIAN
ARRESTING OFFICER AGENCY: Deland PD
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 21:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: CRICHE

DATE/TIME OF TEST: 12/11/2017 22:23:52

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.078	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1	0.081	22:19	
Breath Volume		4.632	22:19	
Air Blank		0.000	22:20	
Air Blank		0.000	22:22	
Subject Sample	# 2		22:22	Subject Test Refused
Breath Volume		0.000	22:22	
Air Blank		0.000	22:22	
Control Test		0.078	22:23	
Air Blank		0.000	22:23	
Diagnostics Check		OK	22:23	

Data Download Date/Time: 12/30/2017 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001114

INSTRUMENT REGISTERED TO: DELAND P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MULERO,JOHAN,
BREATH TEST OPERATOR AGENCY: Deland PD
ARRESTING OFFICER: JAMES,ROYCE
ARRESTING OFFICER AGENCY: Volusia County SO
LAST AGENCY INSPECTION: 11/26/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG616503
DRY GAS EXPIRY DATE: 06/13/2018
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: FLANAGAN

DATE/TIME OF TEST: 12/30/2017 03:36:47

Diagnostics Check		OK	03:28
Air Blank		0.000	03:28
Control Test		0.077	03:28
Air Blank		0.000	03:29
Subject Sample	# 1	0.164	03:30
Breath Volume		1.687	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.165	03:35
Breath Volume		2.023	03:35
Air Blank		0.000	03:35
Control Test		0.078	03:36
Air Blank		0.000	03:36
Diagnostics Check		OK	03:36

Data Download Date/Time: 12/30/2017 14:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COLE,DENNIS,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: COLE,DENNIS
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:15

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 12/30/2017 18:42:38

Diagnostics Check		OK	18:38	
Air Blank		0.000	18:39	
Control Test		0.080	18:39	
Air Blank		0.000	18:40	
Subject Sample	# 1		18:41	Subject Test Refused
Breath Volume		0.000	18:41	
Air Blank		0.000	18:41	
Control Test		0.080	18:42	
Air Blank		0.000	18:42	
Diagnostics Check		OK	18:42	

Data Download Date/Time: 01/26/2018 10:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001121

INSTRUMENT REGISTERED TO: FHP TROOP E MIAMI
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MANGELS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: IGLESIAS,YIMY
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 620051
DRY GAS EXPIRY DATE: 11/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: VICENTE

DATE/TIME OF TEST: 12/31/2017 01:01:32

Diagnostics Check		OK	00:53
Air Blank		0.000	00:53
Control Test		0.080	00:53
Air Blank		0.000	00:54
Subject Sample	# 1	0.200	00:55
Breath Volume		2.097	00:55
Air Blank		0.000	00:56
Air Blank		0.000	00:58
Subject Sample	# 2	0.192	00:59
Breath Volume		2.011	00:59
Air Blank		0.000	01:00
Control Test		0.079	01:00
Air Blank		0.000	01:01
Diagnostics Check		OK	01:01

Data Download Date/Time: 01/26/2018 10:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.