Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BRUNSON

DATE/TIME OF TEST: 01/01/2018 00:23:29

Diagnostics Check		OK	00:13
Air Blank		0.000	00:14
Control Test		0.083	00:14
Air Blank		0.000	00:15
Subject Sample	# 1	0.277	00:15
Breath Volume		3.316	00:15
Air Blank		0.000	00:16
Air Blank		0.000	00:18
Subject Sample	# 2	0.255	00:18
Breath Volume		2.640	00:18
Air Blank		0.000	00:19
Air Blank		0.000	00:21
Subject Sample	# 3	0.255	00:21
Breath Volume		3.000	00:21
Air Blank		0.000	00:22
Control Test		0.079	00:22
Air Blank		0.000	00:23
Diagnostics Check		OK	00:23

Data Download Date/Time: 01/12/2018 21:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MARTIRONE

DATE/TIME OF TEST: 01/18/2018 21:31:16

Diagnostics Check	OK	21:30	
Air Blank		21:30	Ambient Fail
Air Blank	0.000	21:31	

Data Download Date/Time: 02/15/2018 21:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MARTIRONE

DATE/TIME OF TEST: 01/18/2018 21:33:52

Diagnostics Check	OK	21:32	
Air Blank		21:33	Ambient Fail
Air Blank	0.000	21:33	

Data Download Date/Time: 02/15/2018 21:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: MARTIRONE

DATE/TIME OF TEST: 01/18/2018 21:41:42

Diagnostics Check		OK	21:35
Air Blank		0.000	21:36
Control Test		0.083	21:36
Air Blank		0.000	21:36
Subject Sample	# 1	0.116	21:37
Breath Volume		2.386	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.115	21:40
Breath Volume		2.000	21:40
Air Blank		0.000	21:40
Control Test		0.082	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 02/15/2018 21:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BACKUS

DATE/TIME OF TEST: 01/25/2018 02:31:40

Diagnostics Check		OK	02:25	
Air Blank		0.000	02:25	
Control Test		0.083	02:25	
Air Blank		0.000	02:26	
Subject Sample	# 1	0.188	02:27	
Breath Volume		4.347	02:27	
Air Blank		0.000	02:27	
Air Blank		0.000	02:29	
Subject Sample	# 2	0.182	02:30	
Breath Volume		1.734	02:30	
Air Blank		0.000	02:30	
Control Test		0.088	02:31	Interferent Detect
Air Blank		0.000	02:31	

Data Download Date/Time: 02/15/2018 21:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BACKUS

DATE/TIME OF TEST: 01/25/2018 02:37:52

Diagnostics Check		OK	02:33	
Air Blank		0.000	02:33	
Control Test		0.082	02:33	
Air Blank		0.000	02:34	
Subject Sample	# 1	0.176	02:34	
Breath Volume		4.003	02:34	
Air Blank		0.000	02:35	
Air Blank			02:37	Ambient Fail
Air Blank		0.000	02:37	

Data Download Date/Time: 02/15/2018 21:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: BACKUS

DATE/TIME OF TEST: 01/25/2018 02:45:31

Diagnostics Check		OK	02:39
Air Blank		0.000	02:39
Control Test		0.082	02:39
Air Blank		0.000	02:40
Subject Sample	# 1	0.177	02:41
Breath Volume		4.218	02:41
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.170	02:43
Breath Volume		4.582	02:43
Air Blank		0.000	02:44
Control Test		0.083	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 02/15/2018 21:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PHILLIPPO

DATE/TIME OF TEST: 01/26/2018 23:49:20

Diagnostics Check	OK	23:48	
Air Blank		23:48	Ambient Fail
Air Blank	0.000	23:49	

Data Download Date/Time: 02/15/2018 21:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000227

INSTRUMENT REGISTERED TO: FDLE
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 814181
DRY GAS EXPIRY DATE: 07/24/2020
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:06

SUBJECT'S LAST NAME: PHILLIPPO

DATE/TIME OF TEST: 01/26/2018 23:54:31

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:51	
Control Test		0.081	23:51	
Air Blank		0.000	23:52	
Subject Sample	# 1		23:53	Subject Test Refused
Breath Volume		0.000	23:53	
Air Blank		0.000	23:53	
Control Test		0.082	23:53	
Air Blank		0.000	23:54	
Diagnostics Check		OK	23:54	

Data Download Date/Time: 02/15/2018 21:19

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/01/2018 02:59:07

Diagnostics Check		OK	02:42	
Air Blank		0.000	02:42	
Control Test		0.081	02:42	
Air Blank		0.000	02:43	
Subject Sample	# 1	0.229	02:46	
Breath Volume		1.753	02:46	
Air Blank		0.000	02:47	
Air Blank		0.000	02:48	
Subject Sample	# 2	0.199	02:51	Volume Not Met
Breath Volume		0.441	02:51	
Air Blank		0.000	02:52	
Air Blank		0.000	02:54	
Subject Sample	# 3	0.000	02:57	Volume Not Met
Breath Volume		0.160	02:57	
Air Blank		0.000	02:58	
Control Test		0.076	02:58	
Air Blank		0.000	02:59	
Diagnostics Check		OK	02:59	

Data Download Date/Time: 01/22/2018 10:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUNT,CHARLES,R
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: HUNT,CHARLES
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/09/2018 04:21:34

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.234	04:17
Breath Volume		2.402	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.240	04:20
Breath Volume		2.347	04:20
Air Blank		0.000	04:20
Control Test		0.078	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 01/22/2018 10:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: MINCEY

DATE/TIME OF TEST: 01/09/2018 20:45:37

Diagnostics Check		OK	20:39
Air Blank		0.000	20:40
Control Test		0.079	20:40
Air Blank		0.000	20:40
Subject Sample	# 1	0.140	20:41
Breath Volume		2.773	20:41
Air Blank		0.000	20:41
Air Blank		0.000	20:43
Subject Sample	# 2	0.136	20:44
Breath Volume		2.667	20:44
Air Blank		0.000	20:44
Control Test		0.078	20:45
Air Blank		0.000	20:45
Diagnostics Check		OK	20:45

Data Download Date/Time: 01/22/2018 10:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MARTINEZ,CHRISTOPHER
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:31

SUBJECT'S LAST NAME: BRUNEL

DATE/TIME OF TEST: 01/15/2018 00:02:08

Diagnostics Check		OK	23:55
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:57
Subject Sample	# 2	0.266	23:57
Breath Volume		2.085	23:57
Air Blank		0.000	23:58
Air Blank		0.000	00:00
Subject Sample	# 1	0.264	00:00
Breath Volume		2.472	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:02
Diagnostics Check		OK	00:02

Data Download Date/Time: 01/22/2018 10:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WELCH,CHAD,J
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:05

SUBJECT'S LAST NAME: MCCOY

DATE/TIME OF TEST: 01/16/2018 22:53:46

Diagnostics Check		OK	22:38	
Air Blank		0.000	22:38	
Control Test		0.080	22:39	
Air Blank		0.000	22:39	
Subject Sample	# 1	0.153	22:42	Volume Not Met
Breath Volume		0.449	22:42	
Air Blank		0.000	22:43	
Air Blank		0.000	22:45	
Subject Sample	# 2	0.155	22:47	
Breath Volume		1.792	22:47	
Air Blank		0.000	22:48	
Air Blank		0.000	22:49	
Subject Sample	# 3	0.148	22:52	
Breath Volume		2.113	22:52	
Air Blank		0.000	22:52	
Control Test		0.076	22:53	
Air Blank		0.000	22:53	
Diagnostics Check		OK	22:53	

Data Download Date/Time: 01/22/2018 10:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,DANIEL,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MARTIN,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/21/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:59

SUBJECT'S LAST NAME: WEGHORST

DATE/TIME OF TEST: 01/18/2018 03:21:34

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.083	03:19	
Air Blank		0.000	03:19	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.083	03:20	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 01/22/2018 10:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOWARDS,JACOB,T
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: MILLER,HAYLEY
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:50

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/23/2018 22:39:46

Diagnostics Check		OK	22:26	
Air Blank		0.000	22:27	
Control Test		0.080	22:27	
Air Blank		0.000	22:27	
Subject Sample	# 1		22:30	No Sample Provided
Breath Volume		0.000	22:30	
Air Blank		0.000	22:31	
Air Blank		0.000	22:33	
Subject Sample	# 2	0.201	22:35	
Breath Volume		2.292	22:35	
Air Blank		0.000	22:35	
Air Blank		0.000	22:37	
Subject Sample	# 3	0.183	22:38	
Breath Volume		2.410	22:38	
Air Blank		0.000	22:38	
Control Test		0.079	22:39	
Air Blank		0.000	22:39	
Diagnostics Check		OK	22:39	

Data Download Date/Time: 02/21/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALLETT,LANCE,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PALLETT,LANCE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/24/2018 20:41:11

Diagnostics Check		OK	20:36	
Air Blank		0.000	20:36	
Control Test		0.081	20:36	
Air Blank		0.000	20:37	
Subject Sample	# 1		20:39	Slope Not Met
Breath Volume		1.191	20:39	
Air Blank		0.000	20:40	
Control Test		0.081	20:40	
Air Blank		0.000	20:41	
Diagnostics Check		OK	20:41	

Data Download Date/Time: 02/21/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALLETT,LANCE,P
BREATH TEST OPERATOR AGENCY: Alachua County SO
ARRESTING OFFICER: PALLETT,LANCE
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:40

SUBJECT'S LAST NAME: ALVAREZ

DATE/TIME OF TEST: 01/24/2018 21:10:25

Diagnostics Check		OK	21:00
Air Blank		0.000	21:01
Control Test		0.080	21:01
Air Blank		0.000	21:01
Subject Sample	# 1	0.283	21:02
Breath Volume		2.191	21:02
Air Blank		0.000	21:03
Air Blank		0.000	21:05
Subject Sample	# 2	0.260	21:05
Breath Volume		1.117	21:05
Air Blank		0.000	21:06
Air Blank		0.000	21:08
Subject Sample	# 3	0.271	21:08
Breath Volume		1.445	21:08
Air Blank		0.000	21:09
Control Test		0.078	21:09
Air Blank		0.000	21:10
Diagnostics Check		OK	21:10

Data Download Date/Time: 02/21/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRINSKO,THOMAS,K
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: GUYTON,JOHN
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:35

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/26/2018 00:09:17

Diagnostics Check		OK	00:04	
Air Blank		0.000	00:05	
Control Test		0.081	00:05	
Air Blank		0.000	00:06	
Subject Sample	# 1		00:07	Subject Test Refused
Breath Volume		0.000	00:07	
Air Blank		0.000	00:08	
Control Test		0.082	00:08	
Air Blank		0.000	00:09	
Diagnostics Check		OK	00:09	

Data Download Date/Time: 02/21/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCORMICK,ERIC,R
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: COX,CLEVELAND
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:16

SUBJECT'S LAST NAME: MAYBEN

DATE/TIME OF TEST: 01/28/2018 03:49:40

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.080	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.157	03:44
Breath Volume		1.925	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.151	03:48
Breath Volume		1.636	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 02/21/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,JESSE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 01/31/2018 03:30:38

Diagnostics Check		OK	03:23	
Air Blank		0.000	03:24	
Control Test		0.082	03:24	
Air Blank		0.000	03:25	
Subject Sample	# 1	0.183	03:25	
Breath Volume		2.882	03:25	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 2	0.175	03:28	
Breath Volume		2.785	03:28	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank			03:30	RFI Detect
Air Blank		0.000	03:30	

Data Download Date/Time: 02/21/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000737

INSTRUMENT REGISTERED TO: ALACHUA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOSTICK,JESSE,W
BREATH TEST OPERATOR AGENCY: Gainesville PD
ARRESTING OFFICER: BOSTICK,JESSE
ARRESTING OFFICER AGENCY: Gainesville PD
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:56

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 01/31/2018 03:39:01

Diagnostics Check		OK	03:32
Air Blank		0.000	03:33
Control Test		0.081	03:33
Air Blank		0.000	03:34
Subject Sample	# 1	0.169	03:34
Breath Volume		1.894	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.159	03:37
Breath Volume		1.515	03:37
Air Blank		0.000	03:38
Control Test		0.080	03:38
Air Blank		0.000	03:38
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/21/2018 10:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WIGFIELD

DATE/TIME OF TEST: 01/01/2018 01:35:14

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:32	
Control Test		0.082	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1	0.205	01:33	
Breath Volume		4.363	01:33	
Air Blank			01:34	Purge Fail
Air Blank		0.000	01:35	

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:11

SUBJECT'S LAST NAME: WIGFIELD

DATE/TIME OF TEST: 01/01/2018 01:45:03

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.078	01:38
Air Blank		0.000	01:39
Subject Sample	# 1	0.190	01:40
Breath Volume		1.566	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:42
Subject Sample	# 2	0.189	01:43
Breath Volume		3.164	01:43
Air Blank		0.000	01:44
Control Test		0.079	01:44
Air Blank		0.000	01:44
Diagnostics Check		OK	01:45

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:39

SUBJECT'S LAST NAME: HECHAVARRIA

DATE/TIME OF TEST: 01/01/2018 05:08:47

Diagnostics Check		OK	05:01
Air Blank		0.000	05:02
Control Test		0.082	05:02
Air Blank		0.000	05:03
Subject Sample	# 1	0.173	05:03
Breath Volume		2.542	05:03
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.172	05:07
Breath Volume		3.515	05:07
Air Blank		0.000	05:07
Control Test		0.081	05:08
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:28

SUBJECT'S LAST NAME: PACHECO

DATE/TIME OF TEST: 01/02/2018 03:01:41

Diagnostics Check		OK	02:55
Air Blank		0.000	02:55
Control Test		0.083	02:55
Air Blank		0.000	02:56
Subject Sample	# 1	0.180	02:57
Breath Volume		1.378	02:57
Air Blank		0.000	02:57
Air Blank		0.000	02:59
Subject Sample	# 2	0.174	03:00
Breath Volume		1.433	03:00
Air Blank		0.000	03:00
Control Test		0.082	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 20:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: LAND

DATE/TIME OF TEST: 01/02/2018 21:28:22

Diagnostics Check		OK	21:21
Air Blank		0.000	21:22
Control Test		0.082	21:22
Air Blank		0.000	21:23
Subject Sample	# 1	0.160	21:23
Breath Volume		3.632	21:23
Air Blank		0.000	21:24
Air Blank		0.000	21:26
Subject Sample	# 2	0.160	21:26
Breath Volume		3.542	21:26
Air Blank		0.000	21:27
Control Test		0.080	21:27
Air Blank		0.000	21:28
Diagnostics Check		OK	21:28

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOODARD,TAYLOR,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ECHOLS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/06/2018 02:53:04

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.082	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.122	02:47
Breath Volume		1.488	02:47
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.118	02:51
Breath Volume		3.734	02:51
Air Blank		0.000	02:52
Control Test		0.083	02:52
Air Blank		0.000	02:52
Diagnostics Check		OK	02:53

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILSON,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 01/07/2018 02:37:16

Diagnostics Check	OK	02:36	
Air Blank		02:36	RFI Detect
Air Blank	0.000	02:37	

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WILSON,BRIAN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:08

SUBJECT'S LAST NAME: QUEZADA

DATE/TIME OF TEST: 01/07/2018 02:42:52

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:40	
Control Test		0.083	02:40	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:41	Subject Test Refused
Breath Volume		0.000	02:41	
Air Blank		0.000	02:41	
Control Test		0.083	02:42	
Air Blank		0.000	02:42	
Diagnostics Check		OK	02:42	

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENRICH,KRISTEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KRAMER,MELODY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NOLING

DATE/TIME OF TEST: 01/09/2018 04:25:41

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:21	
Control Test		0.081	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1	0.173	04:22	
Breath Volume		1.582	04:22	
Air Blank		0.000	04:23	
Air Blank			04:25	Ambient Fail
Air Blank		0.000	04:25	

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENRICH,KRISTEN,R
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KRAMER,MELODY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: NOLING

DATE/TIME OF TEST: 01/09/2018 04:34:58

Diagnostics Check		OK	04:28
Air Blank		0.000	04:28
Control Test		0.077	04:28
Air Blank		0.000	04:29
Subject Sample	# 1	0.163	04:29
Breath Volume		2.332	04:29
Air Blank		0.000	04:30
Air Blank		0.000	04:32
Subject Sample	# 2	0.151	04:33
Breath Volume		1.339	04:33
Air Blank		0.000	04:34
Control Test		0.076	04:34
Air Blank		0.000	04:34
Diagnostics Check		OK	04:34

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER, JACOB L,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: MCMILLIN

DATE/TIME OF TEST: 01/09/2018 05:11:31

Diagnostics Check		OK	05:05	
Air Blank		0.000	05:06	
Control Test		0.080	05:06	
Air Blank		0.000	05:07	
Subject Sample	# 1		05:10	Subject Test Refused
Breath Volume		0.000	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 01/10/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HAGGAN,WESLEY,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HAGGAN,WESLEY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/13/2018 02:17:32

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.081	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.082	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FREMBGEN,TIA,A
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: MARTIN,DON
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:31

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 01/13/2018 06:04:39

Diagnostics Check		OK	05:53	
Air Blank		0.000	05:53	
Control Test		0.082	05:54	
Air Blank		0.000	05:54	
Subject Sample	# 1		05:57	No Sample Provided
Breath Volume		0.000	05:57	
Air Blank		0.000	05:58	
Air Blank		0.000	06:00	
Subject Sample	# 2		06:03	No Sample Provided
Breath Volume		0.000	06:03	
Air Blank		0.000	06:03	
Control Test		0.082	06:04	
Air Blank		0.000	06:04	
Diagnostics Check		OK	06:04	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 16:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:58

SUBJECT'S LAST NAME: BRONSON

DATE/TIME OF TEST: 01/13/2018 17:25:37

Diagnostics Check		OK	17:19
Air Blank		0.000	17:19
Control Test		0.082	17:19
Air Blank		0.000	17:20
Subject Sample	# 1	0.029	17:21
Breath Volume		3.812	17:21
Air Blank		0.000	17:21
Air Blank		0.000	17:23
Subject Sample	# 2	0.020	17:24
Breath Volume		3.390	17:24
Air Blank		0.000	17:24
Control Test		0.080	17:25
Air Blank		0.000	17:25
Diagnostics Check		OK	17:25

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: NOBLE

DATE/TIME OF TEST: 01/15/2018 23:23:41

Diagnostics Check		OK	23:17
Air Blank		0.000	23:17
Control Test		0.082	23:18
Air Blank		0.000	23:18
Subject Sample	# 1	0.216	23:19
Breath Volume		1.453	23:19
Air Blank		0.000	23:19
Air Blank		0.000	23:21
Subject Sample	# 2	0.212	23:22
Breath Volume		1.589	23:22
Air Blank		0.000	23:22
Control Test		0.080	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:23

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: BOIKO

DATE/TIME OF TEST: 01/20/2018 03:16:26

Diagnostics Check		OK	03:09
Air Blank		0.000	03:10
Control Test		0.083	03:10
Air Blank		0.000	03:11
Subject Sample	# 1	0.159	03:11
Breath Volume		1.945	03:11
Air Blank		0.000	03:12
Air Blank		0.000	03:14
Subject Sample	# 2	0.156	03:14
Breath Volume		2.265	03:14
Air Blank		0.000	03:15
Control Test		0.080	03:15
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BASBAS

DATE/TIME OF TEST: 01/21/2018 00:22:59

Diagnostics Check		OK	00:13	
Air Blank		0.000	00:13	
Control Test		0.082	00:13	
Air Blank		0.000	00:14	
Subject Sample	# 1	0.149	00:15	
Breath Volume		1.730	00:15	
Air Blank		0.000	00:16	
Air Blank		0.000	00:18	
Subject Sample	# 2	0.133	00:21	
Breath Volume		2.234	00:21	
Air Blank		0.000	00:22	
Control Test		0.074	00:22	Control Outside Tolerance
Air Blank		0.000	00:22	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: BASBAS

DATE/TIME OF TEST: 01/21/2018 00:37:09

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.145	00:31
Breath Volume		2.171	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.144	00:35
Breath Volume		1.878	00:35
Air Blank		0.000	00:36
Control Test		0.078	00:36
Air Blank		0.000	00:37
Diagnostics Check		OK	00:37

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ECHOLS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: CRUZ

DATE/TIME OF TEST: 01/21/2018 02:56:00

Diagnostics Check		OK	02:49
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:51
Subject Sample	# 1	0.147	02:51
Breath Volume		1.261	02:51
Air Blank		0.000	02:52
Air Blank		0.000	02:54
Subject Sample	# 2	0.142	02:54
Breath Volume		1.468	02:54
Air Blank		0.000	02:55
Control Test		0.081	02:55
Air Blank		0.000	02:55
Diagnostics Check		OK	02:56

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 01:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: WHITEHORN

DATE/TIME OF TEST: 01/21/2018 03:11:09

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:08	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:10	
Control Test		0.081	03:10	
Air Blank		0.000	03:11	
Diagnostics Check		OK	03:11	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WATTS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WATTS,MICHAEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: SETTLE

DATE/TIME OF TEST: 01/22/2018 00:17:48

Diagnostics Check		OK	00:06	
Air Blank		0.000	00:06	
Control Test		0.081	00:07	
Air Blank		0.000	00:07	
Subject Sample	# 1		00:10	No Sample Provided
Breath Volume		0.000	00:10	
Air Blank		0.000	00:11	
Air Blank		0.000	00:13	
Subject Sample	# 2		00:16	No Sample Provided
Breath Volume		0.000	00:16	
Air Blank		0.000	00:16	
Control Test		0.082	00:17	
Air Blank		0.000	00:17	
Diagnostics Check		OK	00:17	

Data Download Date/Time: 02/27/2018 10:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPEAKMAN,ROBERT,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: SPEAKMAN,ROBERT
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 01/22/2018 04:33:57

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.080	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.152	04:29
Breath Volume		2.023	04:29
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.157	04:32
Breath Volume		3.824	04:32
Air Blank		0.000	04:33
Control Test		0.079	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LANE,AMANDA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: BOYETT

DATE/TIME OF TEST: 01/22/2018 23:44:35

Diagnostics Check		OK	23:42	
Air Blank		0.000	23:43	
Control Test		0.081	23:43	
Air Blank		0.000	23:43	
Subject Sample	# 1		23:44	Sequence Aborted
Breath Volume		0.000	23:44	
Air Blank		0.000	23:44	

Data Download Date/Time: 02/27/2018 10:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANE,AMANDA,M
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: HERRMANN,JOSEPH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: JUDSON

DATE/TIME OF TEST: 01/23/2018 00:39:14

Diagnostics Check		OK	00:32
Air Blank		0.000	00:32
Control Test		0.080	00:32
Air Blank		0.000	00:33
Subject Sample	# 1	0.206	00:34
Breath Volume		1.917	00:34
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.207	00:37
Breath Volume		1.671	00:37
Air Blank		0.000	00:38
Control Test		0.080	00:38
Air Blank		0.000	00:39
Diagnostics Check		OK	00:39

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KIMBRIL,JOSHUA,B
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KIMBRIL,JOSHUA
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:18

SUBJECT'S LAST NAME: HALSTEAD

DATE/TIME OF TEST: 01/25/2018 01:45:07

Diagnostics Check		OK	01:38
Air Blank		0.000	01:38
Control Test		0.082	01:39
Air Blank		0.000	01:39
Subject Sample	# 1	0.171	01:40
Breath Volume		2.855	01:40
Air Blank		0.000	01:41
Air Blank		0.000	01:42
Subject Sample	# 2	0.168	01:43
Breath Volume		2.648	01:43
Air Blank		0.000	01:44
Control Test		0.081	01:44
Air Blank		0.000	01:45
Diagnostics Check		OK	01:45

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,RAYMOND,S
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: ECHOLS,DANIEL
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:08

SUBJECT'S LAST NAME: ROGERS

DATE/TIME OF TEST: 01/25/2018 21:42:39

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.082	21:35
Air Blank		0.000	21:35
Subject Sample	# 1	0.149	21:37
Breath Volume		2.207	21:37
Air Blank		0.000	21:38
Air Blank		0.000	21:39
Subject Sample	# 2	0.143	21:41
Breath Volume		2.441	21:41
Air Blank		0.000	21:41
Control Test		0.079	21:42
Air Blank		0.000	21:42
Diagnostics Check		OK	21:42

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BURGESS,MASON,W
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BURGESS,MASON
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: VELASQUEZ GOMEZ

DATE/TIME OF TEST: 01/27/2018 04:09:18

Diagnostics Check		OK	04:01
Air Blank		0.000	04:01
Control Test		0.081	04:02
Air Blank		0.000	04:02
Subject Sample	# 1	0.192	04:04
Breath Volume		2.207	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.194	04:07
Breath Volume		2.445	04:07
Air Blank		0.000	04:08
Control Test		0.079	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FITCHETT,ADAM,
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: LYBRAND,CHRIS
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 04:40
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: JONES

DATE/TIME OF TEST: 01/27/2018 05:27:26

Diagnostics Check		OK	05:20
Air Blank		0.000	05:21
Control Test		0.081	05:21
Air Blank		0.000	05:21
Subject Sample	# 1	0.160	05:22
Breath Volume		2.953	05:22
Air Blank		0.000	05:23
Air Blank		0.000	05:25
Subject Sample	# 2	0.158	05:25
Breath Volume		2.746	05:25
Air Blank		0.000	05:26
Control Test		0.079	05:26
Air Blank		0.000	05:27
Diagnostics Check		OK	05:27

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRISTOL,JARRET,W
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: KRAMER,MELODY
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:43

SUBJECT'S LAST NAME: HAGOOD

DATE/TIME OF TEST: 01/28/2018 00:06:36

Diagnostics Check		OK	00:03	
Air Blank		0.000	00:04	
Control Test		0.080	00:04	
Air Blank		0.000	00:05	
Subject Sample	# 1		00:05	Subject Test Refused
Breath Volume		0.000	00:05	
Air Blank		0.000	00:05	
Control Test		0.081	00:06	
Air Blank		0.000	00:06	
Diagnostics Check		OK	00:06	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:55

SUBJECT'S LAST NAME: SAQUIC

DATE/TIME OF TEST: 01/28/2018 06:19:23

Diagnostics Check		OK	06:15	
Air Blank		0.000	06:16	
Control Test		0.081	06:16	
Air Blank		0.000	06:16	
Subject Sample	# 1	0.147	06:17	
Breath Volume		2.898	06:17	
Air Blank			06:18	Purge Fail
Air Blank		0.000	06:19	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: SAQUIC

DATE/TIME OF TEST: 01/28/2018 06:42:12

Diagnostics Check	OK	06:40	
Air Blank	0.000	06:41	
Control Test	0.067	06:41	Control Outside Tolerance
Air Blank	0.000	06:42	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:42

SUBJECT'S LAST NAME: SAQUIC

DATE/TIME OF TEST: 01/28/2018 07:04:17

Diagnostics Check	OK	07:02	
Air Blank	0.000	07:03	
Control Test	0.072	07:03	Control Outside Tolerance
Air Blank	0.000	07:04	

Data Download Date/Time: 02/27/2018 10:06

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000739

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RACER,JACOB,L
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: RACER,JACOB
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:04

SUBJECT'S LAST NAME: SAQUIC

DATE/TIME OF TEST: 01/28/2018 07:25:41

Diagnostics Check	OK	07:24	
Air Blank	0.000	07:24	
Control Test	0.074	07:25	Control Outside Tolerance
Air Blank	0.000	07:25	

Data Download Date/Time: 02/27/2018 10:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,DANNY,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:50

SUBJECT'S LAST NAME: DALY

DATE/TIME OF TEST: 01/06/2018 04:27:06

Diagnostics Check		OK	04:20	
Air Blank		0.000	04:20	
Control Test		0.081	04:21	
Air Blank		0.000	04:21	
Subject Sample	# 1	0.000	04:21	
Breath Volume		1.191	04:21	
Air Blank		0.000	04:22	
Air Blank		0.000	04:24	
Subject Sample	# 2		04:25	Subject Test Refused
Breath Volume		1.078	04:25	
Air Blank		0.000	04:26	
Control Test		0.082	04:26	
Air Blank		0.000	04:27	
Diagnostics Check		OK	04:27	

Data Download Date/Time: 01/29/2018 11:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,DANNY,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:55

SUBJECT'S LAST NAME: PATTERSON

DATE/TIME OF TEST: 01/07/2018 23:24:01

Diagnostics Check		OK	23:17
Air Blank		0.000	23:18
Control Test		0.082	23:18
Air Blank		0.000	23:19
Subject Sample	# 1	0.198	23:19
Breath Volume		3.085	23:19
Air Blank		0.000	23:20
Air Blank		0.000	23:22
Subject Sample	# 2	0.188	23:22
Breath Volume		2.550	23:22
Air Blank		0.000	23:23
Control Test		0.082	23:23
Air Blank		0.000	23:23
Diagnostics Check		OK	23:24

Data Download Date/Time: 01/29/2018 11:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARRICO,KELLY,D
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: CARRICO,KELLY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 13:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:34

SUBJECT'S LAST NAME: CLEMENTINO

DATE/TIME OF TEST: 01/12/2018 15:10:43

Diagnostics Check		OK	15:04
Air Blank		0.000	15:04
Control Test		0.082	15:05
Air Blank		0.000	15:05
Subject Sample	# 1	0.000	15:06
Breath Volume		1.472	15:06
Air Blank		0.000	15:06
Air Blank		0.000	15:08
Subject Sample	# 2	0.000	15:09
Breath Volume		2.046	15:09
Air Blank		0.000	15:09
Control Test		0.081	15:10
Air Blank		0.000	15:10
Diagnostics Check		OK	15:10

Data Download Date/Time: 01/29/2018 11:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000742

INSTRUMENT REGISTERED TO: VALPARAISO PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,DANNY,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 12/30/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: MASK

DATE/TIME OF TEST: 01/20/2018 03:40:17

Diagnostics Check		OK	03:34
Air Blank		0.000	03:34
Control Test		0.081	03:35
Air Blank		0.000	03:35
Subject Sample	# 1	0.000	03:36
Breath Volume		4.449	03:36
Air Blank		0.000	03:36
Air Blank		0.000	03:38
Subject Sample	# 2	0.000	03:38
Breath Volume		4.078	03:38
Air Blank		0.000	03:39
Control Test		0.081	03:39
Air Blank		0.000	03:40
Diagnostics Check		OK	03:40

Data Download Date/Time: 01/29/2018 11:18

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HYDER,CHRIS,M
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: LACKEY,NICHOLAS
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: SNOW

DATE/TIME OF TEST: 01/06/2018 21:18:57

Diagnostics Check		OK	21:13
Air Blank		0.000	21:13
Control Test		0.078	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.251	21:14
Breath Volume		1.820	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.248	21:17
Breath Volume		1.898	21:17
Air Blank		0.000	21:18
Control Test		0.079	21:18
Air Blank		0.000	21:18
Diagnostics Check		OK	21:18

Data Download Date/Time: 01/31/2018 11:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MEEKS,JEREMIAH,S
BREATH TEST OPERATOR AGENCY: Escambia County SO
ARRESTING OFFICER: MEEKS,JEREMIAH
ARRESTING OFFICER AGENCY: Escambia County SO
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/09/2018 01:15:10

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.078	01:09
Air Blank		0.000	01:10
Subject Sample	# 1	0.191	01:10
Breath Volume		3.023	01:10
Air Blank		0.000	01:11
Air Blank		0.000	01:13
Subject Sample	# 2	0.187	01:13
Breath Volume		2.769	01:13
Air Blank		0.000	01:14
Control Test		0.076	01:14
Air Blank		0.000	01:15
Diagnostics Check		OK	01:15

Data Download Date/Time: 01/31/2018 11:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENDRIX,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: NEAL

DATE/TIME OF TEST: 01/21/2018 20:56:17

Diagnostics Check		OK	20:50
Air Blank		0.000	20:50
Control Test		0.078	20:50
Air Blank		0.000	20:51
Subject Sample	# 1	0.209	20:51
Breath Volume		2.515	20:51
Air Blank		0.000	20:52
Air Blank		0.000	20:54
Subject Sample	# 2	0.208	20:54
Breath Volume		1.121	20:54
Air Blank		0.000	20:55
Control Test		0.077	20:55
Air Blank		0.000	20:56
Diagnostics Check		OK	20:56

Data Download Date/Time: 01/31/2018 11:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BYRD,BRANDON,C
BREATH TEST OPERATOR AGENCY: Pensacola PD
ARRESTING OFFICER: BYRD,BRANDON
ARRESTING OFFICER AGENCY: Pensacola PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:57

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 01/22/2018 21:29:46

Diagnostics Check		OK	21:17	
Air Blank		0.000	21:18	
Control Test		0.077	21:18	
Air Blank		0.000	21:18	
Subject Sample	# 1	0.247	21:19	
Breath Volume		1.296	21:19	
Air Blank		0.000	21:20	
Air Blank		0.000	21:22	
Subject Sample	# 2	0.238	21:25	Volume Not Met
Breath Volume		0.148	21:25	
Air Blank		0.000	21:26	
Air Blank		0.000	21:27	
Subject Sample	# 3		21:28	Subject Test Refused
Breath Volume		0.000	21:28	
Air Blank		0.000	21:28	
Control Test		0.075	21:29	
Air Blank		0.000	21:29	
Diagnostics Check		OK	21:29	

Data Download Date/Time: 01/31/2018 11:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000744

INSTRUMENT REGISTERED TO: ESCAMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENDRIX,TIMOTHY,D
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HENDRIX,TIMOTHY
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: ARD

DATE/TIME OF TEST: 01/25/2018 20:18:30

Diagnostics Check		OK	20:12
Air Blank		0.000	20:12
Control Test		0.077	20:12
Air Blank		0.000	20:13
Subject Sample	# 1	0.197	20:13
Breath Volume		2.863	20:13
Air Blank		0.000	20:14
Air Blank		0.000	20:16
Subject Sample	# 2	0.197	20:16
Breath Volume		3.609	20:16
Air Blank		0.000	20:17
Control Test		0.078	20:17
Air Blank		0.000	20:18
Diagnostics Check		OK	20:18

Data Download Date/Time: 01/31/2018 11:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 01:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: CONNELL

DATE/TIME OF TEST: 01/01/2018 01:59:00

Diagnostics Check		OK	01:52
Air Blank		0.000	01:52
Control Test		0.082	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.114	01:54
Breath Volume		1.238	01:54
Air Blank		0.000	01:54
Air Blank		0.000	01:56
Subject Sample	# 2	0.109	01:57
Breath Volume		1.136	01:57
Air Blank		0.000	01:58
Control Test		0.082	01:58
Air Blank		0.000	01:58
Diagnostics Check		OK	01:59

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALESE,ANTHONY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PALESE,A
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 1766638
DRY GAS EXPIRY DATE: 11/05/2017
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: PRINCE

DATE/TIME OF TEST: 01/01/2018 03:57:37

Diagnostics Check		OK	03:45	
Air Blank		0.000	03:45	
Control Test		0.082	03:46	
Air Blank		0.000	03:46	
Subject Sample	# 1	0.108	03:47	
Breath Volume		1.296	03:47	
Air Blank		0.000	03:48	
Air Blank		0.000	03:49	
Subject Sample	# 2	0.097	03:52	Volume Not Met
Breath Volume		0.609	03:52	
Air Blank		0.000	03:53	
Air Blank		0.000	03:55	
Subject Sample	# 3	0.108	03:55	
Breath Volume		1.714	03:55	
Air Blank		0.000	03:56	
Control Test		0.078	03:57	
Air Blank		0.000	03:57	
Diagnostics Check		OK	03:57	

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:24

SUBJECT'S LAST NAME: HAWKINS

DATE/TIME OF TEST: 01/01/2018 04:47:42

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:45	
Control Test		0.081	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:46	Subject Test Refused
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.081	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAFFERTY,CHRISTOPHER,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RAFFERTY,CHRISTOPHER
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:10

SUBJECT'S LAST NAME: CAMACHO

DATE/TIME OF TEST: 01/01/2018 20:36:56

Diagnostics Check		OK	20:30
Air Blank		0.000	20:31
Control Test		0.083	20:31
Air Blank		0.000	20:31
Subject Sample	# 1	0.000	20:32
Breath Volume		3.328	20:32
Air Blank		0.000	20:32
Air Blank		0.000	20:34
Subject Sample	# 2	0.000	20:35
Breath Volume		3.640	20:35
Air Blank		0.000	20:35
Control Test		0.081	20:36
Air Blank		0.000	20:36
Diagnostics Check		OK	20:36

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HECKMAN,SHELBY,L
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: RIVERA,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 19:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: MARIANO

DATE/TIME OF TEST: 01/07/2018 21:26:56

Diagnostics Check		OK	21:14	
Air Blank		0.000	21:14	
Control Test		0.083	21:15	
Air Blank		0.000	21:15	
Subject Sample	# 1	0.097	21:16	
Breath Volume		2.417	21:16	
Air Blank		0.000	21:17	
Air Blank		0.000	21:18	
Subject Sample	# 2	0.068	21:21	Volume Not Met
Breath Volume		0.324	21:21	
Air Blank		0.000	21:22	
Air Blank		0.000	21:24	
Subject Sample	# 3	0.091	21:25	
Breath Volume		1.894	21:25	
Air Blank		0.000	21:25	
Control Test		0.079	21:26	
Air Blank		0.000	21:26	
Diagnostics Check		OK	21:26	

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:30

SUBJECT'S LAST NAME: DELLINGER

DATE/TIME OF TEST: 01/07/2018 23:57:18

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.082	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.161	23:52
Breath Volume		1.695	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.156	23:55
Breath Volume		2.386	23:55
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 03:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:03

SUBJECT'S LAST NAME: FONTES-RODRIGUEZ

DATE/TIME OF TEST: 01/12/2018 04:30:16

Diagnostics Check		OK	04:23
Air Blank		0.000	04:24
Control Test		0.082	04:24
Air Blank		0.000	04:25
Subject Sample	# 1	0.141	04:25
Breath Volume		3.078	04:25
Air Blank		0.000	04:26
Air Blank		0.000	04:28
Subject Sample	# 2	0.135	04:28
Breath Volume		2.652	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:30
Diagnostics Check		OK	04:30

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: BIANCULLI

DATE/TIME OF TEST: 01/12/2018 21:47:44

Diagnostics Check		OK	21:41
Air Blank		0.000	21:41
Control Test		0.081	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.079	21:43
Breath Volume		1.441	21:43
Air Blank		0.000	21:43
Air Blank		0.000	21:45
Subject Sample	# 2	0.078	21:46
Breath Volume		1.953	21:46
Air Blank		0.000	21:46
Control Test		0.081	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:47

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 01/13/2018 01:28:54

Diagnostics Check		OK	01:25	
Air Blank		0.000	01:26	
Control Test		0.082	01:26	
Air Blank		0.000	01:27	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.081	01:28	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:04

SUBJECT'S LAST NAME: ROSARIO

DATE/TIME OF TEST: 01/13/2018 03:32:00

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.081	03:25
Air Blank		0.000	03:26
Subject Sample	# 1	0.209	03:27
Breath Volume		2.781	03:27
Air Blank		0.000	03:28
Air Blank		0.000	03:29
Subject Sample	# 2	0.201	03:30
Breath Volume		1.519	03:30
Air Blank		0.000	03:31
Control Test		0.080	03:31
Air Blank		0.000	03:31
Diagnostics Check		OK	03:32

Data Download Date/Time: 01/16/2018 18:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: BOLTON

DATE/TIME OF TEST: 01/14/2018 00:27:00

Diagnostics Check		OK	00:20
Air Blank		0.000	00:20
Control Test		0.082	00:21
Air Blank		0.000	00:21
Subject Sample	# 1	0.179	00:22
Breath Volume		2.601	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.181	00:25
Breath Volume		1.867	00:25
Air Blank		0.000	00:26
Control Test		0.082	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:27

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALESE,ANTHONY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PALESE,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PADIN RODRIGUEZ

DATE/TIME OF TEST: 01/14/2018 04:10:50

Diagnostics Check		OK	04:08	
Air Blank		0.000	04:08	
Control Test		0.082	04:08	
Air Blank		0.000	04:09	
Subject Sample	# 1	0.133	04:09	
Breath Volume		1.863	04:09	
Air Blank			04:10	RFI Detect
Air Blank			04:10	RFI Detect
Air Blank		0.000	04:10	

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PALESE,ANTHONY,W
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: PALESE,ANTHONY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 02:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: PADIN RODRIGUEZ

DATE/TIME OF TEST: 01/14/2018 04:18:17

Diagnostics Check		OK	04:11
Air Blank		0.000	04:12
Control Test		0.082	04:12
Air Blank		0.000	04:13
Subject Sample	# 1	0.128	04:13
Breath Volume		1.652	04:13
Air Blank		0.000	04:14
Air Blank		0.000	04:16
Subject Sample	# 2	0.114	04:16
Breath Volume		1.339	04:16
Air Blank		0.000	04:17
Control Test		0.081	04:17
Air Blank		0.000	04:18
Diagnostics Check		OK	04:18

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:42

SUBJECT'S LAST NAME: DASGUPTA

DATE/TIME OF TEST: 01/15/2018 01:09:52

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.082	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.203	01:05
Breath Volume		3.757	01:05
Air Blank		0.000	01:06
Air Blank		0.000	01:07
Subject Sample	# 2	0.199	01:08
Breath Volume		4.429	01:08
Air Blank		0.000	01:08
Control Test		0.082	01:09
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DELRUSSO,CHRISTOPHER,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: DELRUSSO,CHRISTOPHER
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: SAIN

DATE/TIME OF TEST: 01/15/2018 06:14:54

Diagnostics Check		OK	06:08
Air Blank		0.000	06:08
Control Test		0.083	06:09
Air Blank		0.000	06:09
Subject Sample	# 1	0.117	06:10
Breath Volume		1.683	06:10
Air Blank		0.000	06:10
Air Blank		0.000	06:12
Subject Sample	# 2	0.113	06:13
Breath Volume		1.273	06:13
Air Blank		0.000	06:13
Control Test		0.081	06:14
Air Blank		0.000	06:14
Diagnostics Check		OK	06:14

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WORFEL,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WORFEL,KYLE
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 21:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: NOEL

DATE/TIME OF TEST: 01/15/2018 23:22:21

Diagnostics Check		OK	23:13
Air Blank		0.000	23:13
Control Test		0.082	23:14
Air Blank		0.000	23:14
Subject Sample	# 1	0.243	23:17
Breath Volume		2.246	23:17
Air Blank		0.000	23:18
Air Blank		0.000	23:19
Subject Sample	# 2	0.256	23:20
Breath Volume		1.871	23:20
Air Blank		0.000	23:21
Control Test		0.080	23:21
Air Blank		0.000	23:22
Diagnostics Check		OK	23:22

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,KELLEY,R
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CARPENTER,DONALD
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 01/16/2018 00:24:29

Diagnostics Check		OK	00:20	
Air Blank		0.000	00:20	
Control Test		0.082	00:20	
Air Blank		0.000	00:21	
Subject Sample	# 1	0.194	00:23	
Breath Volume		1.222	00:23	
Air Blank			00:23	Purge Fail
Air Blank		0.000	00:24	

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000752

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARPENTER,D,
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: CARPENTER,D
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 23:45
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: KELLEY

DATE/TIME OF TEST: 01/16/2018 00:35:40

Diagnostics Check		OK	00:29
Air Blank		0.000	00:29
Control Test		0.080	00:29
Air Blank		0.000	00:30
Subject Sample	# 1	0.203	00:31
Breath Volume		1.527	00:31
Air Blank		0.000	00:32
Air Blank		0.000	00:33
Subject Sample	# 2	0.191	00:34
Breath Volume		1.492	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 01/16/2018 18:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000759

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFITH,DALTON,M
BREATH TEST OPERATOR AGENCY: Marianna PD
ARRESTING OFFICER: SIMMONS,CHARLES
ARRESTING OFFICER AGENCY: Marianna PD
LAST AGENCY INSPECTION: 12/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: DAVIS

DATE/TIME OF TEST: 01/13/2018 04:17:44

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.079	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.179	04:12
Breath Volume		2.234	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:15
Subject Sample	# 2	0.181	04:16
Breath Volume		1.421	04:16
Air Blank		0.000	04:16
Control Test		0.079	04:17
Air Blank		0.000	04:17
Diagnostics Check		OK	04:17

Data Download Date/Time: 01/15/2018 14:27

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ESNEUL,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: WALKER,LANDON
ARRESTING OFFICER AGENCY: Crestview PD
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: HERNANDEZ RIVERA

DATE/TIME OF TEST: 01/14/2018 00:51:38

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.151	00:46
Breath Volume		1.308	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:49
Subject Sample	# 2	0.171	00:50
Breath Volume		2.300	00:50
Air Blank		0.000	00:50
Control Test		0.080	00:51
Air Blank		0.000	00:51
Diagnostics Check		OK	00:51

Data Download Date/Time: 02/06/2018 09:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,KENNETH,D
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: TAYLOR,KENNETH
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: LA POINTE

DATE/TIME OF TEST: 01/26/2018 00:36:03

Diagnostics Check		OK	00:33	
Air Blank		0.000	00:33	
Control Test		0.081	00:33	
Air Blank		0.000	00:34	
Subject Sample	# 1		00:34	Subject Test Refused
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Control Test		0.081	00:35	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:36	

Data Download Date/Time: 02/06/2018 09:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000763

INSTRUMENT REGISTERED TO: OKALOOSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOGGS,JUSTIN,E
BREATH TEST OPERATOR AGENCY: Okaloosa County SO
ARRESTING OFFICER: BOGGS,JUSTIN
ARRESTING OFFICER AGENCY: Okaloosa County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 666491
DRY GAS EXPIRY DATE: 12/30/2019
VIOLATION CODE: DUI
ARREST TIME: 14:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:00

SUBJECT'S LAST NAME: DALE

DATE/TIME OF TEST: 01/29/2018 15:33:30

Diagnostics Check		OK	15:20	
Air Blank		0.000	15:20	
Control Test		0.079	15:21	
Air Blank		0.000	15:21	
Subject Sample	# 1	0.136	15:23	
Breath Volume		1.875	15:23	
Air Blank		0.000	15:24	
Air Blank		0.000	15:25	
Subject Sample	# 2	0.110	15:28	Volume Not Met
Breath Volume		0.398	15:28	
Air Blank		0.000	15:29	
Air Blank		0.000	15:31	
Subject Sample	# 3	0.136	15:31	
Breath Volume		2.429	15:31	
Air Blank		0.000	15:32	
Control Test		0.079	15:32	
Air Blank		0.000	15:33	
Diagnostics Check		OK	15:33	

Data Download Date/Time: 02/06/2018 09:25

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ROBINSON,STUART
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: DOMINGUE

DATE/TIME OF TEST: 01/04/2018 02:40:52

Diagnostics Check		OK	02:33
Air Blank		0.000	02:33
Control Test		0.078	02:34
Air Blank		0.000	02:34
Subject Sample	# 1	0.243	02:35
Breath Volume		1.382	02:35
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.234	02:39
Breath Volume		1.828	02:39
Air Blank		0.000	02:39
Control Test		0.075	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: CREWS,MICHAEL
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: RAMER

DATE/TIME OF TEST: 01/06/2018 23:47:22

Diagnostics Check		OK	23:40
Air Blank		0.000	23:41
Control Test		0.079	23:41
Air Blank		0.000	23:42
Subject Sample	# 1	0.126	23:42
Breath Volume		1.898	23:42
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.138	23:45
Breath Volume		2.937	23:45
Air Blank		0.000	23:46
Control Test		0.079	23:46
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: JACKSON

DATE/TIME OF TEST: 01/07/2018 03:16:48

Diagnostics Check		OK	03:10
Air Blank		0.000	03:11
Control Test		0.079	03:11
Air Blank		0.000	03:11
Subject Sample	# 1	0.153	03:12
Breath Volume		1.746	03:12
Air Blank		0.000	03:13
Air Blank		0.000	03:14
Subject Sample	# 2	0.153	03:15
Breath Volume		1.996	03:15
Air Blank		0.000	03:15
Control Test		0.077	03:16
Air Blank		0.000	03:16
Diagnostics Check		OK	03:16

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLAY,DARREL,E
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BASS,ERIKA
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: FARMER

DATE/TIME OF TEST: 01/07/2018 04:31:49

Diagnostics Check		OK	04:25
Air Blank		0.000	04:25
Control Test		0.079	04:26
Air Blank		0.000	04:26
Subject Sample	# 1	0.240	04:27
Breath Volume		2.761	04:27
Air Blank		0.000	04:27
Air Blank		0.000	04:29
Subject Sample	# 2	0.237	04:30
Breath Volume		2.304	04:30
Air Blank		0.000	04:30
Control Test		0.076	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: ROBINSON,STUART
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:11

SUBJECT'S LAST NAME: JOHNSTON

DATE/TIME OF TEST: 01/08/2018 12:37:27

Diagnostics Check		OK	12:31
Air Blank		0.000	12:31
Control Test		0.077	12:31
Air Blank		0.000	12:32
Subject Sample	# 1	0.298	12:32
Breath Volume		2.949	12:32
Air Blank		0.000	12:33
Air Blank		0.000	12:35
Subject Sample	# 2	0.278	12:35
Breath Volume		1.425	12:35
Air Blank		0.000	12:36
Control Test		0.076	12:36
Air Blank		0.000	12:37
Diagnostics Check		OK	12:37

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: PEREZ,YAYAZ
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:38

SUBJECT'S LAST NAME: DUCKER

DATE/TIME OF TEST: 01/11/2018 22:07:10

Diagnostics Check		OK	21:58	
Air Blank		0.000	21:59	
Control Test		0.079	21:59	
Air Blank		0.000	22:00	
Subject Sample	# 1	0.000	22:02	
Breath Volume		2.585	22:02	
Air Blank		0.000	22:03	
Air Blank		0.000	22:05	
Subject Sample	# 2		22:05	Subject Test Refused
Breath Volume		0.613	22:05	
Air Blank		0.000	22:06	
Control Test		0.078	22:06	
Air Blank		0.000	22:07	
Diagnostics Check		OK	22:07	

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVEY,DAVID,A
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: HARVEY,DAVID
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: HADEN

DATE/TIME OF TEST: 01/20/2018 03:10:33

Diagnostics Check		OK	03:01	
Air Blank		0.000	03:01	
Control Test		0.079	03:02	
Air Blank		0.000	03:02	
Subject Sample	# 1		03:05	No Sample Provided
Breath Volume		0.000	03:05	
Air Blank		0.000	03:06	
Air Blank		0.000	03:08	
Subject Sample	# 2		03:09	Subject Test Refused
Breath Volume		0.000	03:09	
Air Blank		0.000	03:09	
Control Test		0.079	03:09	
Air Blank		0.000	03:10	
Diagnostics Check		OK	03:10	

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDENBAUGH,JUSTIN,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BEDENBAUGH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/28/2017

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:19

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/20/2018 23:56:57

Diagnostics Check		OK	23:51	
Air Blank		0.000	23:52	
Control Test		0.078	23:52	
Air Blank		0.000	23:52	
Subject Sample	# 1		23:55	Subject Test Refused
Breath Volume		0.000	23:55	
Air Blank		0.000	23:56	
Control Test		0.078	23:56	
Air Blank		0.000	23:56	
Diagnostics Check		OK	23:56	

Data Download Date/Time: 01/23/2018 07:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CREWS,MICHAEL,B
BREATH TEST OPERATOR AGENCY: Columbia County SO
ARRESTING OFFICER: BEDENBAUGH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: BEEKMAN

DATE/TIME OF TEST: 01/25/2018 02:27:16

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.079	02:25	
Air Blank		0.000	02:25	
Subject Sample	# 1		02:25	Subject Test Refused
Breath Volume		0.000	02:25	
Air Blank		0.000	02:26	
Control Test		0.079	02:26	
Air Blank		0.000	02:27	
Diagnostics Check		OK	02:27	

Data Download Date/Time: 02/28/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BEDENBAUGH,JUSTIN,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: BEDENBAUGH,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: BEEKMAN

DATE/TIME OF TEST: 01/25/2018 04:11:03

Diagnostics Check		OK	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:04
Subject Sample	# 1	0.146	04:06
Breath Volume		1.183	04:06
Air Blank		0.000	04:06
Air Blank		0.000	04:08
Subject Sample	# 2	0.144	04:09
Breath Volume		1.550	04:09
Air Blank		0.000	04:10
Control Test		0.078	04:10
Air Blank		0.000	04:10
Diagnostics Check		OK	04:11

Data Download Date/Time: 02/28/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: TILLIE,JOHN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: KENNEDY

DATE/TIME OF TEST: 01/27/2018 04:49:18

Diagnostics Check		OK	04:39
Air Blank		0.000	04:40
Control Test		0.078	04:40
Air Blank		0.000	04:41
Subject Sample	# 1	0.120	04:41
Breath Volume		4.292	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.096	04:44
Breath Volume		2.191	04:44
Air Blank		0.000	04:45
Air Blank		0.000	04:47
Subject Sample	# 3	0.111	04:47
Breath Volume		3.929	04:47
Air Blank		0.000	04:48
Control Test		0.077	04:48
Air Blank		0.000	04:49
Diagnostics Check		OK	04:49

Data Download Date/Time: 02/28/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLIGAN,CONNIE,A
BREATH TEST OPERATOR AGENCY: Lake City PD
ARRESTING OFFICER: PETERSON,A
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: MCCORMICK

DATE/TIME OF TEST: 01/27/2018 21:47:00

Diagnostics Check		OK	21:40
Air Blank		0.000	21:40
Control Test		0.079	21:41
Air Blank		0.000	21:41
Subject Sample	# 1	0.178	21:42
Breath Volume		1.871	21:42
Air Blank		0.000	21:42
Air Blank		0.000	21:44
Subject Sample	# 2	0.167	21:45
Breath Volume		1.203	21:45
Air Blank		0.000	21:46
Control Test		0.077	21:46
Air Blank		0.000	21:46
Diagnostics Check		OK	21:47

Data Download Date/Time: 02/28/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TILLIE,JOHN,R
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: AGUILA,STEPHEN
ARRESTING OFFICER AGENCY: Columbia County SO
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: PANTONE

DATE/TIME OF TEST: 01/27/2018 22:13:29

Diagnostics Check		OK	22:05	
Air Blank		0.000	22:06	
Control Test		0.078	22:06	
Air Blank		0.000	22:06	
Subject Sample	# 1	0.201	22:08	
Breath Volume		1.984	22:08	
Air Blank		0.000	22:09	
Air Blank		0.000	22:10	
Subject Sample	# 2	0.191	22:11	
Breath Volume		1.472	22:11	
Air Blank		0.000	22:12	
Control Test		0.073	22:12	Control Outside Tolerance
Air Blank		0.000	22:13	

Data Download Date/Time: 02/28/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000775

INSTRUMENT REGISTERED TO: COLUMBIA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SINGLETARY,MATHEW,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: COLBERT,BRANDON
ARRESTING OFFICER AGENCY: Lake City PD
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:06

SUBJECT'S LAST NAME: LAYTON

DATE/TIME OF TEST: 01/31/2018 02:31:10

Diagnostics Check		OK	02:28	
Air Blank		0.000	02:28	
Control Test		0.078	02:29	
Air Blank		0.000	02:29	
Subject Sample	# 1		02:29	Subject Test Refused
Breath Volume		0.000	02:29	
Air Blank		0.000	02:30	
Control Test		0.080	02:30	
Air Blank		0.000	02:31	
Diagnostics Check		OK	02:31	

Data Download Date/Time: 02/28/2018 20:26

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOEN,MARK,A
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HOEN

DATE/TIME OF TEST: 01/31/2018 21:54:59

Diagnostics Check		OK	21:49
Air Blank		0.000	21:49
Control Test		0.081	21:49
Air Blank		0.000	21:50
Subject Sample	# 1	0.267	21:50
Breath Volume		2.207	21:50
Air Blank		0.000	21:51
Air Blank		0.000	21:53
Subject Sample	# 2	0.263	21:53
Breath Volume		1.601	21:53
Air Blank		0.000	21:54
Control Test		0.080	21:54
Air Blank		0.000	21:54
Diagnostics Check		OK	21:54

Data Download Date/Time: 03/21/2018 10:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000776

INSTRUMENT REGISTERED TO: LAFAYETTE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PUTNEL,JOHN,D
BREATH TEST OPERATOR AGENCY: Lafayette County SO
ARRESTING OFFICER: DEAS,ALFRED
ARRESTING OFFICER AGENCY: Lafayette County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17316080A2
DRY GAS EXPIRY DATE: 08/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: HOEN

DATE/TIME OF TEST: 01/31/2018 22:37:22

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.227	22:29
Breath Volume		1.445	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:32
Subject Sample	# 2	0.252	22:32
Breath Volume		2.152	22:32
Air Blank		0.000	22:33
Air Blank		0.000	22:35
Subject Sample	# 3	0.246	22:35
Breath Volume		2.148	22:35
Air Blank		0.000	22:36
Control Test		0.078	22:36
Air Blank		0.000	22:37
Diagnostics Check		OK	22:37

Data Download Date/Time: 03/21/2018 10:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000779

INSTRUMENT REGISTERED TO: MADISON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUMNER,ETHAN,R
BREATH TEST OPERATOR AGENCY: FHP Troop H
ARRESTING OFFICER: SUMNER,ETHAN
ARRESTING OFFICER AGENCY: FHP Troop H
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A3
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: HAYNES

DATE/TIME OF TEST: 01/28/2018 23:06:11

Diagnostics Check		OK	22:55
Air Blank		0.000	22:55
Control Test		0.081	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.082	22:59
Breath Volume		2.121	22:59
Air Blank		0.000	23:00
Air Blank		0.000	23:02
Subject Sample	# 2	0.078	23:04
Breath Volume		3.058	23:04
Air Blank		0.000	23:05
Control Test		0.081	23:05
Air Blank		0.000	23:06
Diagnostics Check		OK	23:06

Data Download Date/Time: 02/13/2018 10:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HENRY,MICHAEL,S
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: MIARA,JEFF
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/01/2018 00:59:28

Diagnostics Check		OK	00:51
Air Blank		0.000	00:52
Control Test		0.082	00:52
Air Blank		0.000	00:53
Subject Sample	# 1	0.201	00:54
Breath Volume		1.390	00:54
Air Blank		0.000	00:55
Air Blank		0.000	00:56
Subject Sample	# 2	0.196	00:57
Breath Volume		1.691	00:57
Air Blank		0.000	00:58
Control Test		0.081	00:58
Air Blank		0.000	00:59
Diagnostics Check		OK	00:59

Data Download Date/Time: 01/28/2018 22:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLAUGHTER,D,F
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SLAUGHTER,SLAUGHTER
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: TOLEDANO

DATE/TIME OF TEST: 01/07/2018 00:37:20

Diagnostics Check		OK	00:24	
Air Blank		0.000	00:24	
Control Test		0.080	00:25	
Air Blank		0.000	00:25	
Subject Sample	# 1	0.228	00:26	
Breath Volume		1.445	00:26	
Air Blank		0.000	00:27	
Air Blank		0.000	00:28	
Subject Sample	# 2		00:31	No Sample Provided
Breath Volume		0.000	00:31	
Air Blank		0.000	00:32	
Air Blank		0.000	00:34	
Subject Sample	# 3	0.230	00:35	
Breath Volume		1.246	00:35	
Air Blank		0.000	00:36	
Control Test		0.082	00:36	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 01/28/2018 22:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT,DAVID,G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: WATSON,CLAYTON
ARRESTING OFFICER AGENCY: Suwannee County SO
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:28

SUBJECT'S LAST NAME: BYRD

DATE/TIME OF TEST: 01/10/2018 01:57:33

Diagnostics Check		OK	01:52	
Air Blank		0.000	01:52	
Control Test		0.082	01:52	
Air Blank		0.000	01:53	
Subject Sample	# 1		01:55	Subject Test Refused
Breath Volume		0.023	01:55	
Air Blank		0.000	01:56	
Control Test		0.085	01:56	
Air Blank		0.000	01:57	
Diagnostics Check		OK	01:57	

Data Download Date/Time: 01/28/2018 22:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SLAUGHTER,DEREK,F
BREATH TEST OPERATOR AGENCY: Live Oak PD
ARRESTING OFFICER: SLAUGHTER,DEREK
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: GROSS

DATE/TIME OF TEST: 01/11/2018 00:51:45

Diagnostics Check		OK	00:39	
Air Blank		0.000	00:40	
Control Test		0.082	00:40	
Air Blank		0.000	00:41	
Subject Sample	# 1	0.106	00:44	Volume Not Met
Breath Volume		0.230	00:44	
Air Blank		0.000	00:45	
Air Blank		0.000	00:46	
Subject Sample	# 2	0.092	00:49	Volume Not Met
Breath Volume		0.335	00:49	
Air Blank		0.000	00:50	
Control Test		0.079	00:51	
Air Blank		0.000	00:51	
Diagnostics Check		OK	00:51	

Data Download Date/Time: 01/28/2018 22:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUGHES,ALFRED,L
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HUGHES,ALFRED
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: BENNETT

DATE/TIME OF TEST: 01/25/2018 22:51:06

Diagnostics Check		OK	22:44
Air Blank		0.000	22:45
Control Test		0.083	22:45
Air Blank		0.000	22:46
Subject Sample	# 1	0.169	22:46
Breath Volume		1.796	22:46
Air Blank		0.000	22:47
Air Blank		0.000	22:49
Subject Sample	# 2	0.168	22:49
Breath Volume		1.632	22:49
Air Blank		0.000	22:50
Control Test		0.081	22:50
Air Blank		0.000	22:51
Diagnostics Check		OK	22:51

Data Download Date/Time: 01/28/2018 22:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000780

INSTRUMENT REGISTERED TO: SUWANNEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FALGOUT,DAVID,G
BREATH TEST OPERATOR AGENCY: Suwannee County SO
ARRESTING OFFICER: FLEMING,R
ARRESTING OFFICER AGENCY: Live Oak PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 29017080A2
DRY GAS EXPIRY DATE: 12/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:00

SUBJECT'S LAST NAME: VARGO

DATE/TIME OF TEST: 01/28/2018 04:31:41

Diagnostics Check		OK	04:24
Air Blank		0.000	04:25
Control Test		0.082	04:25
Air Blank		0.000	04:26
Subject Sample	# 1	0.187	04:26
Breath Volume		1.472	04:26
Air Blank		0.000	04:27
Air Blank		0.000	04:28
Subject Sample	# 2	0.186	04:29
Breath Volume		1.761	04:29
Air Blank		0.000	04:30
Control Test		0.081	04:31
Air Blank		0.000	04:31
Diagnostics Check		OK	04:31

Data Download Date/Time: 01/28/2018 22:23

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAMMES

DATE/TIME OF TEST: 01/07/2018 02:15:11

Diagnostics Check		OK	02:10	
Air Blank		0.000	02:10	
Control Test		0.081	02:11	
Air Blank		0.000	02:11	
Subject Sample	# 1	0.176	02:13	
Breath Volume		1.601	02:13	
Air Blank			02:14	Purge Fail
Air Blank		0.000	02:15	

Data Download Date/Time: 01/17/2018 17:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BANKS,SEAN,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: BANKS,SEAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: HAMMES

DATE/TIME OF TEST: 01/07/2018 02:23:50

Diagnostics Check		OK	02:17
Air Blank		0.000	02:17
Control Test		0.081	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.196	02:19
Breath Volume		2.625	02:19
Air Blank		0.000	02:19
Air Blank		0.000	02:21
Subject Sample	# 2	0.187	02:22
Breath Volume		1.730	02:22
Air Blank		0.000	02:22
Control Test		0.082	02:23
Air Blank		0.000	02:23
Diagnostics Check		OK	02:23

Data Download Date/Time: 01/17/2018 17:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ROBERSON,THOMAS,C
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: ROBERSON,THOMAS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:24

SUBJECT'S LAST NAME: RICH

DATE/TIME OF TEST: 01/08/2018 16:51:04

Diagnostics Check		OK	16:45
Air Blank		0.000	16:45
Control Test		0.081	16:45
Air Blank		0.000	16:46
Subject Sample	# 1	0.073	16:46
Breath Volume		1.656	16:46
Air Blank		0.000	16:47
Air Blank		0.000	16:49
Subject Sample	# 2	0.073	16:49
Breath Volume		1.339	16:49
Air Blank		0.000	16:50
Control Test		0.080	16:50
Air Blank		0.000	16:50
Diagnostics Check		OK	16:51

Data Download Date/Time: 01/17/2018 17:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: FITZWATER

DATE/TIME OF TEST: 01/14/2018 04:53:56

Diagnostics Check		OK	04:48	
Air Blank		0.000	04:49	
Control Test		0.081	04:49	
Air Blank		0.000	04:49	
Subject Sample	# 1		04:52	Interferent Detect
Breath Volume		2.671	04:52	
Air Blank			04:53	Purge Fail
Air Blank		0.000	04:53	

Data Download Date/Time: 01/17/2018 17:47

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEIERS,BLAINE,T
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: WEIERS,BLAINE
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:28

SUBJECT'S LAST NAME: FITZWATER

DATE/TIME OF TEST: 01/14/2018 05:01:18

Diagnostics Check		OK	04:57	
Air Blank		0.000	04:57	
Control Test		0.081	04:58	
Air Blank		0.000	04:58	
Subject Sample	# 1		04:59	Subject Test Refused
Breath Volume		0.453	04:59	
Air Blank		0.000	05:00	
Control Test		0.083	05:00	
Air Blank		0.000	05:01	
Diagnostics Check		OK	05:01	

Data Download Date/Time: 01/17/2018 17:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNGREN,DENNIS,A
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: YOUNGREN,DENNIS
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:29

SUBJECT'S LAST NAME: RUGGS

DATE/TIME OF TEST: 01/21/2018 12:55:23

Diagnostics Check		OK	12:49
Air Blank		0.000	12:49
Control Test		0.081	12:50
Air Blank		0.000	12:50
Subject Sample	# 1	0.000	12:51
Breath Volume		4.164	12:51
Air Blank		0.000	12:51
Air Blank		0.000	12:53
Subject Sample	# 2	0.000	12:54
Breath Volume		4.171	12:54
Air Blank		0.000	12:54
Control Test		0.082	12:54
Air Blank		0.000	12:55
Diagnostics Check		OK	12:55

Data Download Date/Time: 02/23/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCAPIN,BRANDON,L
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: NEFF,STEPHEN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:04

SUBJECT'S LAST NAME: D'ALEMBERTE

DATE/TIME OF TEST: 01/25/2018 10:30:56

Diagnostics Check		OK	10:24
Air Blank		0.000	10:25
Control Test		0.082	10:25
Air Blank		0.000	10:25
Subject Sample	# 1	0.000	10:26
Breath Volume		2.027	10:26
Air Blank		0.000	10:26
Air Blank		0.000	10:28
Subject Sample	# 2	0.000	10:29
Breath Volume		1.507	10:29
Air Blank		0.000	10:30
Control Test		0.082	10:30
Air Blank		0.000	10:30
Diagnostics Check		OK	10:30

Data Download Date/Time: 02/23/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000787

INSTRUMENT REGISTERED TO: GULF BREEZE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TATRO,KERSTAN,J
BREATH TEST OPERATOR AGENCY: Gulf Breeze PD
ARRESTING OFFICER: TATRO,KERSTAN
ARRESTING OFFICER AGENCY: Gulf Breeze PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A5
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: AUSHERMAN

DATE/TIME OF TEST: 01/27/2018 03:01:51

Diagnostics Check		OK	02:59	
Air Blank		0.000	02:59	
Control Test		0.081	02:59	
Air Blank		0.000	03:00	
Subject Sample	# 1		03:00	Subject Test Refused
Breath Volume		0.000	03:00	
Air Blank		0.000	03:00	
Control Test		0.080	03:01	
Air Blank		0.000	03:01	
Diagnostics Check		OK	03:01	

Data Download Date/Time: 02/23/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: BEHRSIN,BLAKE
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:30

SUBJECT'S LAST NAME: OMEALLY

DATE/TIME OF TEST: 01/18/2018 21:00:24

Diagnostics Check		OK	20:52	
Air Blank		0.000	20:53	
Control Test		0.082	20:53	
Air Blank		0.000	20:54	
Subject Sample	# 1		20:57	No Sample Provided
Breath Volume		0.000	20:57	
Air Blank		0.000	20:57	
Air Blank		0.000	20:59	
Subject Sample	# 2		20:59	Sequence Aborted
Breath Volume		0.000	20:59	
Air Blank		0.000	21:00	

Data Download Date/Time: 02/27/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,TANYA,L
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: STICKNEY,RON
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: HEAD

DATE/TIME OF TEST: 01/19/2018 04:34:34

Diagnostics Check		OK	04:31	
Air Blank		0.000	04:32	
Control Test		0.082	04:32	
Air Blank		0.000	04:32	
Subject Sample	# 1		04:34	Sequence Aborted
Breath Volume		0.000	04:34	
Air Blank		0.000	04:34	

Data Download Date/Time: 02/27/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUES,RUI,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: PARRETT,AUSTIN
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: HIRSCH

DATE/TIME OF TEST: 01/27/2018 03:45:30

Diagnostics Check		OK	03:35
Air Blank		0.000	03:36
Control Test		0.081	03:36
Air Blank		0.000	03:37
Subject Sample	# 1	0.263	03:37
Breath Volume		2.253	03:37
Air Blank		0.000	03:38
Air Blank		0.000	03:40
Subject Sample	# 2	0.288	03:40
Breath Volume		1.972	03:40
Air Blank		0.000	03:41
Air Blank		0.000	03:43
Subject Sample	# 3	0.270	03:43
Breath Volume		2.039	03:43
Air Blank		0.000	03:44
Control Test		0.080	03:44
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/27/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,ERIC,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PORTALES-BONILLA

DATE/TIME OF TEST: 01/28/2018 01:42:02

Diagnostics Check		OK	01:36	
Air Blank		0.000	01:37	
Control Test		0.082	01:37	
Air Blank		0.000	01:37	
Subject Sample	# 1	0.143	01:38	
Breath Volume		1.195	01:38	
Air Blank		0.000	01:39	
Air Blank			01:40	RFI Detect
Air Blank		0.000	01:41	
			01:41	RFI Detect
Air Blank		0.000	01:42	

Data Download Date/Time: 02/27/2018 08:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLBERT,ERIC,R
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SCUOPPO,ANTHONY
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: PORTALES-BONILLA

DATE/TIME OF TEST: 01/28/2018 01:56:55

Diagnostics Check		OK	01:44	
Air Blank		0.000	01:44	
Control Test		0.082	01:45	
Air Blank		0.000	01:45	
Subject Sample	# 1	0.143	01:48	Volume Not Met
Breath Volume		0.574	01:48	
Air Blank		0.000	01:49	
Air Blank		0.000	01:51	
Subject Sample	# 2	0.148	01:51	
Breath Volume		2.328	01:51	
Air Blank		0.000	01:52	
Air Blank		0.000	01:54	
Subject Sample	# 3	0.151	01:55	
Breath Volume		1.238	01:55	
Air Blank		0.000	01:55	
Control Test		0.081	01:56	
Air Blank		0.000	01:56	
Diagnostics Check		OK	01:56	

Data Download Date/Time: 02/27/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000788

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,TYLER,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: WITT,TYLER
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:59

SUBJECT'S LAST NAME: LIQUORI

DATE/TIME OF TEST: 01/31/2018 00:25:40

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.082	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.000	00:21
Breath Volume		4.589	00:21
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.000	00:24
Breath Volume		4.210	00:24
Air Blank		0.000	00:24
Control Test		0.081	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 02/27/2018 08:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,DAVID,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SNOW,DAVID
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 10:11

SUBJECT'S LAST NAME: BECKER

DATE/TIME OF TEST: 01/06/2018 10:38:13

Diagnostics Check		OK	10:31
Air Blank		0.000	10:32
Control Test		0.080	10:32
Air Blank		0.000	10:33
Subject Sample	# 1	0.161	10:33
Breath Volume		2.714	10:33
Air Blank		0.000	10:34
Air Blank		0.000	10:36
Subject Sample	# 2	0.163	10:36
Breath Volume		3.027	10:36
Air Blank		0.000	10:37
Control Test		0.080	10:37
Air Blank		0.000	10:38
Diagnostics Check		OK	10:38

Data Download Date/Time: 01/17/2018 10:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUES,RUI,
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: RODRIGUES,RUI
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: HOSTETLER

DATE/TIME OF TEST: 01/15/2018 03:09:18

Diagnostics Check		OK	03:02
Air Blank		0.000	03:02
Control Test		0.080	03:03
Air Blank		0.000	03:03
Subject Sample	# 1	0.047	03:04
Breath Volume		1.597	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.040	03:07
Breath Volume		1.351	03:07
Air Blank		0.000	03:08
Control Test		0.078	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 01/17/2018 10:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000792

INSTRUMENT REGISTERED TO: ST LUCIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNOW,DAVID,J
BREATH TEST OPERATOR AGENCY: St. Lucie County SO
ARRESTING OFFICER: SNOW,DAVID
ARRESTING OFFICER AGENCY: St. Lucie County SO
LAST AGENCY INSPECTION: 12/05/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/15/2018 18:24:29

Diagnostics Check		OK	18:17
Air Blank		0.000	18:17
Control Test		0.079	18:17
Air Blank		0.000	18:18
Subject Sample	# 1	0.167	18:19
Breath Volume		1.484	18:19
Air Blank		0.000	18:20
Air Blank		0.000	18:22
Subject Sample	# 2	0.173	18:22
Breath Volume		1.363	18:22
Air Blank		0.000	18:23
Control Test		0.078	18:23
Air Blank		0.000	18:24
Diagnostics Check		OK	18:24

Data Download Date/Time: 01/17/2018 10:31

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000795

INSTRUMENT REGISTERED TO: GULF COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BAXLEY,JAMES,P
BREATH TEST OPERATOR AGENCY: Gulf County SO
ARRESTING OFFICER: BAXLEY,JAMES
ARRESTING OFFICER AGENCY: Gulf County SO
LAST AGENCY INSPECTION: 01/12/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A2
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:00

SUBJECT'S LAST NAME: CARROLL

DATE/TIME OF TEST: 01/31/2018 10:34:14

Diagnostics Check		OK	10:28
Air Blank		0.000	10:28
Control Test		0.079	10:28
Air Blank		0.000	10:29
Subject Sample	# 1	0.000	10:29
Breath Volume		4.386	10:29
Air Blank		0.000	10:30
Air Blank		0.000	10:32
Subject Sample	# 2	0.000	10:32
Breath Volume		4.621	10:32
Air Blank		0.000	10:33
Control Test		0.079	10:33
Air Blank		0.000	10:34
Diagnostics Check		OK	10:34

Data Download Date/Time: 02/23/2018 12:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000797

INSTRUMENT REGISTERED TO: HAMILTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALKER,SHELIA,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: WALKER,SHELIA
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 12/18/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 14816080A3
DRY GAS EXPIRY DATE: 07/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 01/04/2018 08:04:07

Diagnostics Check		OK	07:58
Air Blank		0.000	07:58
Control Test		0.080	07:58
Air Blank		0.000	07:59
Subject Sample	# 1	0.000	07:59
Breath Volume		1.160	07:59
Air Blank		0.000	08:00
Air Blank		0.000	08:02
Subject Sample	# 2	0.000	08:02
Breath Volume		1.382	08:02
Air Blank		0.000	08:03
Control Test		0.080	08:03
Air Blank		0.000	08:04
Diagnostics Check		OK	08:04

Data Download Date/Time: 01/31/2018 18:37

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN-TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/01/2018 01:56:17

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.081	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.050	01:51
Breath Volume		1.871	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.051	01:54
Breath Volume		1.871	01:54
Air Blank		0.000	01:55
Control Test		0.081	01:55
Air Blank		0.000	01:56
Diagnostics Check		OK	01:56

Data Download Date/Time: 01/03/2018 19:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMPKINS,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:54

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/01/2018 05:25:01

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.081	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.186	05:20
Breath Volume		1.246	05:20
Air Blank		0.000	05:21
Air Blank		0.000	05:22
Subject Sample	# 2	0.190	05:23
Breath Volume		1.761	05:23
Air Blank		0.000	05:24
Control Test		0.080	05:24
Air Blank		0.000	05:24
Diagnostics Check		OK	05:25

Data Download Date/Time: 01/03/2018 19:24

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:53

SUBJECT'S LAST NAME: ANGEL-LEON

DATE/TIME OF TEST: 01/24/2018 05:20:19

Diagnostics Check		OK	05:13
Air Blank		0.000	05:13
Control Test		0.082	05:14
Air Blank		0.000	05:14
Subject Sample	# 1	0.230	05:15
Breath Volume		2.421	05:15
Air Blank		0.000	05:16
Air Blank		0.000	05:18
Subject Sample	# 2	0.223	05:18
Breath Volume		1.410	05:18
Air Blank		0.000	05:19
Control Test		0.080	05:19
Air Blank		0.000	05:20
Diagnostics Check		OK	05:20

Data Download Date/Time: 02/14/2018 10:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BAILEY,D
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:23

SUBJECT'S LAST NAME: DOWNEY

DATE/TIME OF TEST: 01/24/2018 21:51:02

Diagnostics Check		OK	21:44
Air Blank		0.000	21:45
Control Test		0.082	21:45
Air Blank		0.000	21:45
Subject Sample	# 1	0.158	21:46
Breath Volume		2.328	21:46
Air Blank		0.000	21:47
Air Blank		0.000	21:48
Subject Sample	# 2	0.164	21:49
Breath Volume		1.785	21:49
Air Blank		0.000	21:50
Control Test		0.081	21:50
Air Blank		0.000	21:50
Diagnostics Check		OK	21:51

Data Download Date/Time: 02/14/2018 10:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARLAUG,G
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:47

SUBJECT'S LAST NAME: GREENE

DATE/TIME OF TEST: 01/26/2018 03:14:28

Diagnostics Check		OK	03:07
Air Blank		0.000	03:08
Control Test		0.081	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.143	03:09
Breath Volume		2.335	03:09
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.141	03:12
Breath Volume		1.890	03:12
Air Blank		0.000	03:13
Control Test		0.081	03:13
Air Blank		0.000	03:14
Diagnostics Check		OK	03:14

Data Download Date/Time: 02/14/2018 10:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: DEVETO

DATE/TIME OF TEST: 01/28/2018 00:58:45

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:43	
Control Test		0.081	00:43	
Air Blank		0.000	00:44	
Subject Sample	# 1	0.223	00:46	
Breath Volume		1.511	00:46	
Air Blank		0.000	00:46	
Air Blank		0.000	00:48	
Subject Sample	# 2	0.199	00:51	Volume Not Met
Breath Volume		0.187	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 3		00:57	No Sample Provided
Breath Volume		0.000	00:57	
Air Blank		0.000	00:57	
Control Test		0.079	00:58	
Air Blank		0.000	00:58	
Diagnostics Check		OK	00:58	

Data Download Date/Time: 02/14/2018 10:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:08

SUBJECT'S LAST NAME: RAFUS

DATE/TIME OF TEST: 01/28/2018 01:36:35

Diagnostics Check		OK	01:28	
Air Blank		0.000	01:28	
Control Test		0.080	01:28	
Air Blank		0.000	01:29	
Subject Sample	# 1		01:32	No Sample Provided
Breath Volume		0.000	01:32	
Air Blank		0.000	01:33	
Air Blank		0.000	01:35	
Subject Sample	# 2		01:35	Subject Test Refused
Breath Volume		0.000	01:35	
Air Blank		0.000	01:35	
Control Test		0.081	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 02/14/2018 10:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:42

SUBJECT'S LAST NAME: CORSO JR

DATE/TIME OF TEST: 01/28/2018 04:09:07

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.081	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.160	04:04
Breath Volume		3.031	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:06
Subject Sample	# 2	0.149	04:07
Breath Volume		1.480	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/14/2018 10:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000808

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SIMS,T
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:29

SUBJECT'S LAST NAME: HUAPILLA-MARTINEZ

DATE/TIME OF TEST: 01/28/2018 04:55:52

Diagnostics Check		OK	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:50
Air Blank		0.000	04:50
Subject Sample	# 1	0.142	04:51
Breath Volume		2.578	04:51
Air Blank		0.000	04:52
Air Blank		0.000	04:53
Subject Sample	# 2	0.138	04:54
Breath Volume		1.890	04:54
Air Blank		0.000	04:54
Control Test		0.081	04:55
Air Blank		0.000	04:55
Diagnostics Check		OK	04:55

Data Download Date/Time: 02/14/2018 10:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FULTON,DARIN,M
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FULTON,DARIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:08
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:39

SUBJECT'S LAST NAME: RAMEY

DATE/TIME OF TEST: 01/21/2018 17:06:35

Diagnostics Check		OK	16:59
Air Blank		0.000	17:00
Control Test		0.080	17:00
Air Blank		0.000	17:01
Subject Sample	# 1	0.242	17:02
Breath Volume		1.300	17:02
Air Blank		0.000	17:02
Air Blank		0.000	17:04
Subject Sample	# 2	0.229	17:05
Breath Volume		1.867	17:05
Air Blank		0.000	17:05
Control Test		0.079	17:05
Air Blank		0.000	17:06
Diagnostics Check		OK	17:06

Data Download Date/Time: 02/26/2018 09:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: MOSS

DATE/TIME OF TEST: 01/27/2018 00:40:53

Diagnostics Check		OK	00:36	
Air Blank		0.000	00:37	
Control Test		0.080	00:37	
Air Blank		0.000	00:38	
Subject Sample	# 1		00:39	Subject Test Refused
Breath Volume		0.000	00:39	
Air Blank		0.000	00:39	
Control Test		0.079	00:40	
Air Blank		0.000	00:40	
Diagnostics Check		OK	00:40	

Data Download Date/Time: 02/26/2018 09:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000810

INSTRUMENT REGISTERED TO: DIXIE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCE,KEITH,D
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SPENCE,KEITH
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:30

SUBJECT'S LAST NAME: BASS

DATE/TIME OF TEST: 01/27/2018 20:07:48

Diagnostics Check		OK	20:01
Air Blank		0.000	20:02
Control Test		0.079	20:02
Air Blank		0.000	20:02
Subject Sample	# 1	0.182	20:03
Breath Volume		1.585	20:03
Air Blank		0.000	20:04
Air Blank		0.000	20:05
Subject Sample	# 2	0.182	20:06
Breath Volume		2.429	20:06
Air Blank		0.000	20:06
Control Test		0.079	20:07
Air Blank		0.000	20:07
Diagnostics Check		OK	20:07

Data Download Date/Time: 02/26/2018 09:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: BROWN,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80-000814
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:38

SUBJECT'S LAST NAME: FLORES CONCEPCION

DATE/TIME OF TEST: 01/20/2018 04:05:21

Diagnostics Check		OK	03:58
Air Blank		0.000	03:59
Control Test		0.081	03:59
Air Blank		0.000	03:59
Subject Sample	# 1	0.172	04:00
Breath Volume		2.425	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:03
Subject Sample	# 2	0.171	04:03
Breath Volume		2.402	04:03
Air Blank		0.000	04:04
Control Test		0.080	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: STIMPSON,RACHEL,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HARSANYI,WILLIAM
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 80-000814
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:00

SUBJECT'S LAST NAME: MONSEGUR

DATE/TIME OF TEST: 01/20/2018 05:23:57

Diagnostics Check		OK	05:21	
Air Blank		0.000	05:21	
Control Test		0.082	05:21	
Air Blank		0.000	05:22	
Subject Sample	# 1		05:22	Subject Test Refused
Breath Volume		0.000	05:22	
Air Blank		0.000	05:23	
Control Test		0.081	05:23	
Air Blank		0.000	05:23	
Diagnostics Check		OK	05:23	

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:47

SUBJECT'S LAST NAME: BUCHHOLTZ

DATE/TIME OF TEST: 01/20/2018 21:14:36

Diagnostics Check		OK	21:08
Air Blank		0.000	21:08
Control Test		0.081	21:08
Air Blank		0.000	21:09
Subject Sample	# 1	0.040	21:10
Breath Volume		3.800	21:10
Air Blank		0.000	21:10
Air Blank		0.000	21:12
Subject Sample	# 2	0.032	21:13
Breath Volume		1.417	21:13
Air Blank		0.000	21:13
Control Test		0.081	21:13
Air Blank		0.000	21:14
Diagnostics Check		OK	21:14

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JULIOS,ANTHONY,J
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HARSANYI,J
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:10

SUBJECT'S LAST NAME: MELCHING

DATE/TIME OF TEST: 01/20/2018 23:36:40

Diagnostics Check		OK	23:30
Air Blank		0.000	23:30
Control Test		0.081	23:30
Air Blank		0.000	23:31
Subject Sample	# 1	0.169	23:32
Breath Volume		3.539	23:32
Air Blank		0.000	23:32
Air Blank		0.000	23:34
Subject Sample	# 2	0.160	23:35
Breath Volume		3.308	23:35
Air Blank		0.000	23:35
Control Test		0.081	23:36
Air Blank		0.000	23:36
Diagnostics Check		OK	23:36

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITHERELL,BENJAMIN,T
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: WITHERELL,BENJAMIN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 07:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:26

SUBJECT'S LAST NAME: FRANCO

DATE/TIME OF TEST: 01/24/2018 08:52:49

Diagnostics Check		OK	08:46
Air Blank		0.000	08:46
Control Test		0.082	08:47
Air Blank		0.000	08:47
Subject Sample	# 1	0.000	08:48
Breath Volume		1.558	08:48
Air Blank		0.000	08:48
Air Blank		0.000	08:50
Subject Sample	# 2	0.000	08:51
Breath Volume		2.261	08:51
Air Blank		0.000	08:51
Control Test		0.080	08:52
Air Blank		0.000	08:52
Diagnostics Check		OK	08:52

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: QUINONES,D,
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: HOFFLAND,R
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 01/25/2018 03:56:15

Diagnostics Check		OK	03:50
Air Blank		0.000	03:50
Control Test		0.081	03:50
Air Blank		0.000	03:51
Subject Sample	# 1	0.129	03:51
Breath Volume		2.988	03:51
Air Blank		0.000	03:52
Air Blank		0.000	03:54
Subject Sample	# 2	0.136	03:54
Breath Volume		2.503	03:54
Air Blank		0.000	03:55
Control Test		0.081	03:55
Air Blank		0.000	03:56
Diagnostics Check		OK	03:56

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: COCHRANE,AARON
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: MITCHELL

DATE/TIME OF TEST: 01/26/2018 23:16:25

Diagnostics Check		OK	23:13	
Air Blank		0.000	23:13	
Control Test		0.081	23:14	
Air Blank		0.000	23:14	
Subject Sample	# 1		23:15	Subject Test Refused
Breath Volume		0.000	23:15	
Air Blank		0.000	23:15	
Control Test		0.081	23:15	
Air Blank		0.000	23:16	
Diagnostics Check		OK	23:16	

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEIRCE,HAROLD,F
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: SMITH,BRIAN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:56

SUBJECT'S LAST NAME: SANTIAGO

DATE/TIME OF TEST: 01/27/2018 14:19:58

Diagnostics Check		OK	14:16	
Air Blank		0.000	14:16	
Control Test		0.081	14:16	
Air Blank		0.000	14:17	
Subject Sample	# 1		14:18	Subject Test Refused
Breath Volume		0.000	14:18	
Air Blank		0.000	14:19	
Control Test		0.081	14:19	
Air Blank		0.000	14:19	
Diagnostics Check		OK	14:19	

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COCHRANE,AARON,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: LEWIS,SAMATHA
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:54

SUBJECT'S LAST NAME: BEANE

DATE/TIME OF TEST: 01/28/2018 00:41:00

Diagnostics Check		OK	00:31
Air Blank		0.000	00:32
Control Test		0.081	00:32
Air Blank		0.000	00:32
Subject Sample	# 1	0.153	00:33
Breath Volume		3.117	00:33
Air Blank		0.000	00:34
Air Blank		0.000	00:36
Subject Sample	# 2	0.132	00:36
Breath Volume		1.128	00:36
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 3	0.128	00:39
Breath Volume		1.300	00:39
Air Blank		0.000	00:40
Control Test		0.081	00:40
Air Blank		0.000	00:40
Diagnostics Check		OK	00:41

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000814

INSTRUMENT REGISTERED TO: HERNANDO COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRUZ,CHRISTINA,M
BREATH TEST OPERATOR AGENCY: Hernando County SO
ARRESTING OFFICER: MATHIS,JOHN
ARRESTING OFFICER AGENCY: Hernando County SO
LAST AGENCY INSPECTION: 01/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 12716080A7
DRY GAS EXPIRY DATE: 06/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: BUSTRIN

DATE/TIME OF TEST: 01/31/2018 02:19:24

Diagnostics Check		OK	02:08	
Air Blank		0.000	02:08	
Control Test		0.081	02:08	
Air Blank		0.000	02:09	
Subject Sample	# 1		02:12	No Sample Provided
Breath Volume		0.000	02:12	
Air Blank		0.000	02:12	
Air Blank		0.000	02:14	
Subject Sample	# 2		02:18	No Sample Provided
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.082	02:18	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 02/17/2018 13:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000815

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOUTROS,PHILLIP
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 11:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:15

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 01/31/2018 12:38:22

Diagnostics Check		OK	12:35	
Air Blank		0.000	12:35	
Control Test		0.080	12:36	
Air Blank		0.000	12:36	
Subject Sample	# 1		12:36	Subject Test Refused
Breath Volume		0.000	12:36	
Air Blank		0.000	12:37	
Control Test		0.080	12:37	
Air Blank		0.000	12:38	
Diagnostics Check		OK	12:38	

Data Download Date/Time: 02/25/2018 19:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEWIS,JARROD,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SEARS,TRAVIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: GHIOTO

DATE/TIME OF TEST: 01/01/2018 01:34:37

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:31	
Control Test		0.079	01:32	
Air Blank		0.000	01:32	
Subject Sample	# 1		01:33	Subject Test Refused
Breath Volume		0.000	01:33	
Air Blank		0.000	01:33	
Control Test		0.079	01:33	
Air Blank		0.000	01:34	
Diagnostics Check		OK	01:34	

Data Download Date/Time: 01/29/2018 01:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: COLLINS,MARVIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:55

SUBJECT'S LAST NAME: BROWN III

DATE/TIME OF TEST: 01/02/2018 03:19:06

Diagnostics Check		OK	03:16	
Air Blank		0.000	03:16	
Control Test		0.080	03:16	
Air Blank		0.000	03:17	
Subject Sample	# 1		03:17	Subject Test Refused
Breath Volume		0.000	03:17	
Air Blank		0.000	03:18	
Control Test		0.080	03:18	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 01/29/2018 01:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: DEMO

DATE/TIME OF TEST: 01/10/2018 02:40:52

Diagnostics Check		OK	02:34
Air Blank		0.000	02:34
Control Test		0.077	02:35
Air Blank		0.000	02:35
Subject Sample	# 1	0.170	02:36
Breath Volume		2.343	02:36
Air Blank		0.000	02:36
Air Blank		0.000	02:38
Subject Sample	# 2	0.171	02:39
Breath Volume		1.929	02:39
Air Blank		0.000	02:39
Control Test		0.078	02:40
Air Blank		0.000	02:40
Diagnostics Check		OK	02:40

Data Download Date/Time: 01/29/2018 01:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLOW,VINESSA,
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: CIULLA,J
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 13:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 13:26

SUBJECT'S LAST NAME: HARRIS

DATE/TIME OF TEST: 01/10/2018 13:59:27

Diagnostics Check		OK	13:49
Air Blank		0.000	13:49
Control Test		0.078	13:50
Air Blank		0.000	13:50
Subject Sample	# 1	0.321	13:51
Breath Volume		3.628	13:51
Air Blank		0.000	13:52
Air Blank		0.000	13:54
Subject Sample	# 2	0.291	13:54
Breath Volume		2.660	13:54
Air Blank		0.000	13:55
Air Blank		0.000	13:57
Subject Sample	# 3	0.296	13:57
Breath Volume		1.414	13:57
Air Blank		0.000	13:58
Control Test		0.076	13:58
Air Blank		0.000	13:59
Diagnostics Check		OK	13:59

Data Download Date/Time: 01/29/2018 01:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,JUSTIN,P
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: NAVARRO,BENJAMIN
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:31

SUBJECT'S LAST NAME: STEFANIK

DATE/TIME OF TEST: 01/13/2018 22:59:00

Diagnostics Check		OK	22:52
Air Blank		0.000	22:52
Control Test		0.079	22:53
Air Blank		0.000	22:53
Subject Sample	# 1	0.157	22:54
Breath Volume		2.117	22:54
Air Blank		0.000	22:54
Air Blank		0.000	22:56
Subject Sample	# 2	0.152	22:57
Breath Volume		1.968	22:57
Air Blank		0.000	22:58
Control Test		0.080	22:58
Air Blank		0.000	22:58
Diagnostics Check		OK	22:59

Data Download Date/Time: 01/29/2018 01:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BOUTROS,PHILLIP
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: KNAPP

DATE/TIME OF TEST: 01/18/2018 21:34:19

Diagnostics Check		OK	21:22	
Air Blank		0.000	21:22	
Control Test		0.080	21:23	
Air Blank		0.000	21:23	
Subject Sample	# 1	0.150	21:26	Volume Not Met
Breath Volume		0.441	21:26	
Air Blank		0.000	21:27	
Air Blank		0.000	21:29	
Subject Sample	# 2		21:32	Subject Test Refused
Breath Volume		0.585	21:32	
Air Blank		0.000	21:33	
Control Test		0.080	21:33	
Air Blank		0.000	21:34	
Diagnostics Check		OK	21:34	

Data Download Date/Time: 01/29/2018 01:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GEOGHEGAN,ASHLEY,M
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: BARD,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:08

SUBJECT'S LAST NAME: P200481944700

DATE/TIME OF TEST: 01/18/2018 23:40:47

Diagnostics Check		OK	23:28	
Air Blank		0.000	23:28	
Control Test		0.079	23:28	
Air Blank		0.000	23:29	
Subject Sample	# 1		23:32	No Sample Provided
Breath Volume		0.000	23:32	
Air Blank		0.000	23:33	
Air Blank		0.000	23:35	
Subject Sample	# 2	0.159	23:35	
Breath Volume		2.972	23:35	
Air Blank		0.000	23:36	
Air Blank		0.000	23:38	
Subject Sample	# 3	0.164	23:39	
Breath Volume		2.046	23:39	
Air Blank		0.000	23:39	
Control Test		0.079	23:40	
Air Blank		0.000	23:40	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 01/29/2018 01:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: ALEXANDER,DENZEL
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/20/2018 00:34:37

Diagnostics Check		OK	00:26
Air Blank		0.000	00:27
Control Test		0.079	00:27
Air Blank		0.000	00:27
Subject Sample	# 1	0.135	00:28
Breath Volume		2.015	00:28
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.136	00:32
Breath Volume		3.097	00:32
Air Blank		0.000	00:33
Control Test		0.079	00:33
Air Blank		0.000	00:34
Diagnostics Check		OK	00:34

Data Download Date/Time: 01/29/2018 01:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LIVERS,DAWN,T
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: SEARS,TRAVIS
ARRESTING OFFICER AGENCY: Sumter County SO
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:04

SUBJECT'S LAST NAME: MARQUETTE

DATE/TIME OF TEST: 01/21/2018 19:27:52

Diagnostics Check		OK	19:24	
Air Blank		0.000	19:25	
Control Test		0.077	19:25	
Air Blank		0.000	19:26	
Subject Sample	# 1		19:26	Subject Test Refused
Breath Volume		0.000	19:26	
Air Blank		0.000	19:26	
Control Test		0.077	19:27	
Air Blank		0.000	19:27	
Diagnostics Check		OK	19:27	

Data Download Date/Time: 01/29/2018 01:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000816

INSTRUMENT REGISTERED TO: SUMTER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBSON,JAMES,W
BREATH TEST OPERATOR AGENCY: Sumter County SO
ARRESTING OFFICER: PEREZ,CONRAD
ARRESTING OFFICER AGENCY: Center Hill PD
LAST AGENCY INSPECTION: 12/20/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A3
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: TRINIDAD SANCHEZ

DATE/TIME OF TEST: 01/28/2018 02:43:31

Diagnostics Check		OK	02:31	
Air Blank		0.000	02:31	
Control Test		0.077	02:31	
Air Blank		0.000	02:32	
Subject Sample	# 1	0.159	02:35	Volume Not Met
Breath Volume		0.132	02:35	
Air Blank		0.000	02:36	
Air Blank		0.000	02:38	
Subject Sample	# 2	0.163	02:38	
Breath Volume		1.441	02:38	
Air Blank		0.000	02:39	
Air Blank		0.000	02:41	
Subject Sample	# 3	0.161	02:41	
Breath Volume		1.980	02:41	
Air Blank		0.000	02:42	
Control Test		0.077	02:42	
Air Blank		0.000	02:43	
Diagnostics Check		OK	02:43	

Data Download Date/Time: 01/29/2018 01:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 19:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:06

SUBJECT'S LAST NAME: WATERS

DATE/TIME OF TEST: 01/06/2018 20:37:31

Diagnostics Check		OK	20:27	
Air Blank		0.000	20:27	
Control Test		0.081	20:28	
Air Blank		0.000	20:28	
Subject Sample	# 1	0.182	20:31	Volume Not Met
Breath Volume		0.093	20:31	
Air Blank		0.000	20:32	
Air Blank		0.000	20:34	
Subject Sample	# 2		20:36	Subject Test Refused
Breath Volume		0.000	20:36	
Air Blank		0.000	20:36	
Control Test		0.079	20:36	
Air Blank		0.000	20:37	
Diagnostics Check		OK	20:37	

Data Download Date/Time: 01/24/2018 11:38

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: METZ,KYLE,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TRNSIT,CITRUS
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 11:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:13

SUBJECT'S LAST NAME: CRIDER

DATE/TIME OF TEST: 01/09/2018 11:39:53

Diagnostics Check		OK	11:34
Air Blank		0.000	11:34
Control Test		0.080	11:34
Air Blank		0.000	11:35
Subject Sample	# 1	0.000	11:35
Breath Volume		2.128	11:35
Air Blank		0.000	11:36
Air Blank		0.000	11:38
Subject Sample	# 2	0.000	11:38
Breath Volume		1.644	11:38
Air Blank		0.000	11:38
Control Test		0.080	11:39
Air Blank		0.000	11:39
Diagnostics Check		OK	11:39

Data Download Date/Time: 01/24/2018 11:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ECHEVARRIA,LISA,V
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,ERIC
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:09

SUBJECT'S LAST NAME: BRISTOL

DATE/TIME OF TEST: 01/16/2018 02:51:57

Diagnostics Check		OK	02:44
Air Blank		0.000	02:45
Control Test		0.080	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.092	02:46
Breath Volume		1.148	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.096	02:50
Breath Volume		1.187	02:50
Air Blank		0.000	02:51
Control Test		0.080	02:51
Air Blank		0.000	02:51
Diagnostics Check		OK	02:51

Data Download Date/Time: 01/24/2018 11:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 01:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: COIMBRE

DATE/TIME OF TEST: 01/17/2018 02:45:32

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.072	02:40
Breath Volume		2.855	02:40
Air Blank		0.000	02:41
Air Blank		0.000	02:43
Subject Sample	# 2	0.068	02:43
Breath Volume		1.722	02:43
Air Blank		0.000	02:44
Control Test		0.081	02:44
Air Blank		0.000	02:45
Diagnostics Check		OK	02:45

Data Download Date/Time: 01/24/2018 11:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: FELIX,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 20:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: SUDDUTH

DATE/TIME OF TEST: 01/19/2018 21:13:51

Diagnostics Check		OK	21:10	
Air Blank		0.000	21:11	
Control Test		0.080	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:12	Subject Test Refused
Breath Volume		0.000	21:12	
Air Blank		0.000	21:12	
Control Test		0.080	21:13	
Air Blank		0.000	21:13	
Diagnostics Check		OK	21:13	

Data Download Date/Time: 01/24/2018 11:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHOETTNER,ADRIAN,J
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: KJEKSTAD,E
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 06/05/2017
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:16

SUBJECT'S LAST NAME: CONRAD

DATE/TIME OF TEST: 01/21/2018 19:49:55

Diagnostics Check		OK	19:43
Air Blank		0.000	19:43
Control Test		0.079	19:44
Air Blank		0.000	19:44
Subject Sample	# 1	0.167	19:45
Breath Volume		1.781	19:45
Air Blank		0.000	19:45
Air Blank		0.000	19:47
Subject Sample	# 2	0.175	19:48
Breath Volume		2.800	19:48
Air Blank		0.000	19:48
Control Test		0.079	19:49
Air Blank		0.000	19:49
Diagnostics Check		OK	19:49

Data Download Date/Time: 01/24/2018 11:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: ELMORE,PATRICK
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 12/13/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: SCALF

DATE/TIME OF TEST: 01/22/2018 23:28:24

Diagnostics Check		OK	23:20
Air Blank		0.000	23:20
Control Test		0.080	23:21
Air Blank		0.000	23:21
Subject Sample	# 1	0.170	23:22
Breath Volume		1.183	23:22
Air Blank		0.000	23:23
Air Blank		0.000	23:24
Subject Sample	# 2	0.168	23:26
Breath Volume		1.339	23:26
Air Blank		0.000	23:27
Control Test		0.079	23:27
Air Blank		0.000	23:28
Diagnostics Check		OK	23:28

Data Download Date/Time: 01/24/2018 11:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LATHAN,MICHAEL,S
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: HOWARD,R
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 13115080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:00

SUBJECT'S LAST NAME: TEETER

DATE/TIME OF TEST: 01/27/2018 00:37:07

Diagnostics Check		OK	00:34	
Air Blank		0.000	00:34	
Control Test		0.080	00:34	
Air Blank		0.000	00:35	
Subject Sample	# 1		00:35	Subject Test Refused
Breath Volume		0.000	00:35	
Air Blank		0.000	00:36	
Control Test		0.080	00:36	
Air Blank		0.000	00:37	
Diagnostics Check		OK	00:37	

Data Download Date/Time: 02/27/2018 11:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: DANNER,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: WHITEMAN

DATE/TIME OF TEST: 01/27/2018 22:04:18

Diagnostics Check		OK	21:52	
Air Blank		0.000	21:53	
Control Test		0.079	21:53	
Air Blank		0.000	21:54	
Subject Sample	# 1		21:57	No Sample Provided
Breath Volume		0.000	21:57	
Air Blank		0.000	21:57	
Air Blank		0.000	21:59	
Subject Sample	# 2		22:02	No Sample Provided
Breath Volume		0.000	22:02	
Air Blank		0.000	22:03	
Control Test		0.079	22:03	
Air Blank		0.000	22:04	
Diagnostics Check		OK	22:04	

Data Download Date/Time: 02/27/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,DANIEL,Z
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: TAYLOR,DEP
ARRESTING OFFICER AGENCY: Citrus County SO
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: BEACH

DATE/TIME OF TEST: 01/28/2018 02:49:07

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Subject Sample	# 1		02:42	No Sample Provided
Breath Volume		0.000	02:42	
Air Blank		0.000	02:42	
Air Blank		0.000	02:44	
Subject Sample	# 2		02:47	No Sample Provided
Breath Volume		0.000	02:47	
Air Blank		0.000	02:48	
Control Test		0.081	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 02/27/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: COUNTY WORKE,COUNTY WORKER
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/15/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:00
GENDER: Male

OBSERVATION PERIOD BEGAN: 10:40

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 01/29/2018 11:11:48

Diagnostics Check		OK	11:05
Air Blank		0.000	11:05
Control Test		0.079	11:05
Air Blank		0.000	11:06
Subject Sample	# 1	0.000	11:07
Breath Volume		1.855	11:07
Air Blank		0.000	11:07
Air Blank		0.000	11:09
Subject Sample	# 2	0.000	11:10
Breath Volume		1.925	11:10
Air Blank		0.000	11:10
Control Test		0.080	11:11
Air Blank		0.000	11:11
Diagnostics Check		OK	11:11

Data Download Date/Time: 02/27/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000819

INSTRUMENT REGISTERED TO: CITRUS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WING,JOSHUA,M
BREATH TEST OPERATOR AGENCY: CCA Citrus
ARRESTING OFFICER: COUNTY WORKER,COUNTY WORKER
ARRESTING OFFICER AGENCY: CCA Citrus
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: ZERO TOL
ARREST TIME: 00:00
GENDER: Male

OBSERVATION PERIOD BEGAN: 11:02

SUBJECT'S LAST NAME: PETERSON

DATE/TIME OF TEST: 01/29/2018 11:28:09

Diagnostics Check		OK	11:22
Air Blank		0.000	11:22
Control Test		0.080	11:22
Air Blank		0.000	11:23
Subject Sample	# 1	0.000	11:23
Breath Volume		1.453	11:23
Air Blank		0.000	11:24
Air Blank		0.000	11:26
Subject Sample	# 2	0.000	11:26
Breath Volume		1.660	11:26
Air Blank		0.000	11:27
Control Test		0.079	11:27
Air Blank		0.000	11:28
Diagnostics Check		OK	11:28

Data Download Date/Time: 02/27/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUKENS,SHAWN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 01/01/2018 02:17:51

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:07	
Subject Sample	# 1	0.073	02:10	Volume Not Met
Breath Volume		0.992	02:10	
Air Blank		0.000	02:11	
Air Blank		0.000	02:13	
Subject Sample	# 2	0.067	02:16	Volume Not Met
Breath Volume		0.074	02:16	
Air Blank		0.000	02:16	
Control Test		0.080	02:17	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 01/04/2018 18:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: LUKENS,SHAWN
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 12/09/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: SNYDER

DATE/TIME OF TEST: 01/01/2018 02:30:42

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.078	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.078	02:25
Breath Volume		2.511	02:25
Air Blank		0.000	02:25
Air Blank		0.000	02:27
Subject Sample	# 2	0.069	02:29
Breath Volume		1.804	02:29
Air Blank		0.000	02:29
Control Test		0.079	02:30
Air Blank		0.000	02:30
Diagnostics Check		OK	02:30

Data Download Date/Time: 01/04/2018 18:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAINS,HENRY,E
BREATH TEST OPERATOR AGENCY: Fruitland Park PD
ARRESTING OFFICER: HILBERER,BRIAN
ARRESTING OFFICER AGENCY: Fruitland Park PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: HUGHES

DATE/TIME OF TEST: 01/05/2018 21:02:44

Diagnostics Check		OK	20:56
Air Blank		0.000	20:57
Control Test		0.080	20:57
Air Blank		0.000	20:57
Subject Sample	# 1	0.031	20:58
Breath Volume		2.621	20:58
Air Blank		0.000	20:58
Air Blank		0.000	21:00
Subject Sample	# 2	0.029	21:01
Breath Volume		1.714	21:01
Air Blank		0.000	21:01
Control Test		0.085	21:02
Air Blank		0.000	21:02
Diagnostics Check		OK	21:02

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARRETT,SHANE,M
BREATH TEST OPERATOR AGENCY: Lady Lake PD
ARRESTING OFFICER: BARRETT,SHANE
ARRESTING OFFICER AGENCY: Lady Lake PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:48

SUBJECT'S LAST NAME: BOUDREAU

DATE/TIME OF TEST: 01/08/2018 01:28:25

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.079	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:27	Subject Test Refused
Breath Volume		0.000	01:27	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RITTER,ALEC,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: RITTER,ALEC
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:35

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 01/11/2018 23:02:29

Diagnostics Check		OK	22:55
Air Blank		0.000	22:56
Control Test		0.079	22:56
Air Blank		0.000	22:56
Subject Sample	# 1	0.180	22:57
Breath Volume		1.175	22:57
Air Blank		0.000	22:58
Air Blank		0.000	23:00
Subject Sample	# 2	0.185	23:01
Breath Volume		1.460	23:01
Air Blank		0.000	23:01
Control Test		0.080	23:01
Air Blank		0.000	23:02
Diagnostics Check		OK	23:02

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PITCHER,DARRYL,P
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CHESSHER,SANDRA
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:53

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/12/2018 00:24:52

Diagnostics Check		OK	00:18
Air Blank		0.000	00:18
Control Test		0.080	00:19
Air Blank		0.000	00:19
Subject Sample	# 1	0.136	00:20
Breath Volume		2.027	00:20
Air Blank		0.000	00:21
Air Blank		0.000	00:22
Subject Sample	# 2	0.132	00:23
Breath Volume		1.777	00:23
Air Blank		0.000	00:23
Control Test		0.080	00:24
Air Blank		0.000	00:24
Diagnostics Check		OK	00:24

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: SILVERIO,ALVIN
ARRESTING OFFICER AGENCY: Mascotte PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 01/13/2018 03:18:59

Diagnostics Check		OK	03:11
Air Blank		0.000	03:11
Control Test		0.080	03:12
Air Blank		0.000	03:12
Subject Sample	# 1	0.081	03:14
Breath Volume		1.777	03:14
Air Blank		0.000	03:15
Air Blank		0.000	03:16
Subject Sample	# 2	0.082	03:17
Breath Volume		1.320	03:17
Air Blank		0.000	03:18
Control Test		0.080	03:18
Air Blank		0.000	03:18
Diagnostics Check		OK	03:18

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHESSHER,SANDRA,W
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PITCHER,D
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: FIELDS

DATE/TIME OF TEST: 01/14/2018 02:50:53

Diagnostics Check		OK	02:42
Air Blank		0.000	02:42
Control Test		0.080	02:43
Air Blank		0.000	02:43
Subject Sample	# 1	0.225	02:46
Breath Volume		1.199	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.221	02:49
Breath Volume		1.402	02:49
Air Blank		0.000	02:49
Control Test		0.079	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: REED,GREGORY,A
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: REED,GREGORY
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:30

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/14/2018 06:00:55

Diagnostics Check		OK	05:54
Air Blank		0.000	05:54
Control Test		0.080	05:54
Air Blank		0.000	05:55
Subject Sample	# 1	0.095	05:56
Breath Volume		1.652	05:56
Air Blank		0.000	05:56
Air Blank		0.000	05:58
Subject Sample	# 2	0.091	05:59
Breath Volume		1.738	05:59
Air Blank		0.000	06:00
Control Test		0.081	06:00
Air Blank		0.000	06:00
Diagnostics Check		OK	06:00

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARTANOWICZ,MARTIN,
BREATH TEST OPERATOR AGENCY: FHP Troop K
ARRESTING OFFICER: KARTANOWICZ,MARTIN
ARRESTING OFFICER AGENCY: FHP Troop K
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: INGRAM

DATE/TIME OF TEST: 01/17/2018 02:15:03

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:06	
Control Test		0.080	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.061	02:09	Volume Not Met
Breath Volume		0.019	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2		02:13	Subject Test Refused
Breath Volume		0.050	02:13	
Air Blank		0.000	02:14	
Control Test		0.079	02:14	
Air Blank		0.000	02:14	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,ETHAN,C
BREATH TEST OPERATOR AGENCY: Clermont PD
ARRESTING OFFICER: BELL,ETHAN
ARRESTING OFFICER AGENCY: Clermont PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:20

SUBJECT'S LAST NAME: KING

DATE/TIME OF TEST: 01/17/2018 17:02:22

Diagnostics Check		OK	16:53	
Air Blank		0.000	16:54	
Control Test		0.079	16:54	
Air Blank		0.000	16:55	
Subject Sample	# 1		16:58	No Sample Provided
Breath Volume		0.000	16:58	
Air Blank		0.000	16:58	
Air Blank		0.000	17:00	
Subject Sample	# 2		17:01	Subject Test Refused
Breath Volume		0.000	17:01	
Air Blank		0.000	17:01	
Control Test		0.079	17:01	
Air Blank		0.000	17:02	
Diagnostics Check		OK	17:02	

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DENTON,AMY
ARRESTING OFFICER AGENCY: Astatula PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:40

SUBJECT'S LAST NAME: ULLOA

DATE/TIME OF TEST: 01/18/2018 04:09:50

Diagnostics Check		OK	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Subject Sample	# 1	0.181	04:04
Breath Volume		1.554	04:04
Air Blank		0.000	04:05
Air Blank		0.000	04:07
Subject Sample	# 2	0.183	04:08
Breath Volume		2.843	04:08
Air Blank		0.000	04:08
Control Test		0.082	04:09
Air Blank		0.000	04:09
Diagnostics Check		OK	04:09

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARMON,RICK,A
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: PEREZ,DAVID
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:30

SUBJECT'S LAST NAME: CURLEY

DATE/TIME OF TEST: 01/19/2018 03:23:54

Diagnostics Check		OK	03:16
Air Blank		0.000	03:17
Control Test		0.080	03:17
Air Blank		0.000	03:17
Subject Sample	# 1	0.144	03:18
Breath Volume		1.699	03:18
Air Blank		0.000	03:19
Air Blank		0.000	03:21
Subject Sample	# 2	0.149	03:22
Breath Volume		1.281	03:22
Air Blank		0.000	03:23
Control Test		0.082	03:23
Air Blank		0.000	03:23
Diagnostics Check		OK	03:23

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORALES,RONNIE,
BREATH TEST OPERATOR AGENCY: Eustis PD
ARRESTING OFFICER: PERRY JR,WAYNE
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: STUCKEY

DATE/TIME OF TEST: 01/21/2018 01:29:07

Diagnostics Check		OK	01:23
Air Blank		0.000	01:23
Control Test		0.079	01:23
Air Blank		0.000	01:24
Subject Sample	# 1	0.301	01:24
Breath Volume		2.632	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:27
Subject Sample	# 2	0.286	01:27
Breath Volume		2.660	01:27
Air Blank		0.000	01:28
Control Test		0.080	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: GREEN,SAMUEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:27

SUBJECT'S LAST NAME: PEREZ

DATE/TIME OF TEST: 01/24/2018 01:54:23

Diagnostics Check		OK	01:48
Air Blank		0.000	01:48
Control Test		0.080	01:48
Air Blank		0.000	01:49
Subject Sample	# 1	0.181	01:49
Breath Volume		3.796	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:52
Subject Sample	# 2	0.182	01:52
Breath Volume		3.996	01:52
Air Blank		0.000	01:53
Control Test		0.081	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRENSHAW,SHANNON,L
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: SORAPURU,C
ARRESTING OFFICER AGENCY: Tavares PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:29

SUBJECT'S LAST NAME: ZAHN

DATE/TIME OF TEST: 01/25/2018 10:56:02

Diagnostics Check		OK	10:49
Air Blank		0.000	10:49
Control Test		0.080	10:50
Air Blank		0.000	10:50
Subject Sample	# 1	0.229	10:51
Breath Volume		3.386	10:51
Air Blank		0.000	10:51
Air Blank		0.000	10:53
Subject Sample	# 2	0.222	10:54
Breath Volume		3.078	10:54
Air Blank		0.000	10:55
Control Test		0.081	10:55
Air Blank		0.000	10:55
Diagnostics Check		OK	10:56

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BINDER,BRONSON,M
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: BINDER,BRONSON
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: SPEAR

DATE/TIME OF TEST: 01/26/2018 22:42:11

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.080	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.000	22:37
Breath Volume		2.699	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:40
Subject Sample	# 2	0.000	22:40
Breath Volume		2.484	22:40
Air Blank		0.000	22:41
Control Test		0.080	22:41
Air Blank		0.000	22:42
Diagnostics Check		OK	22:42

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANT,DANIEL,F
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: GRANT,DANIEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: MCGUIRE

DATE/TIME OF TEST: 01/28/2018 01:22:19

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.079	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.188	01:17
Breath Volume		4.472	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.188	01:20
Breath Volume		4.097	01:20
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENTON,AMY,B
BREATH TEST OPERATOR AGENCY: Astatula PD
ARRESTING OFFICER: DEVINE,MICHAEL
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:31

SUBJECT'S LAST NAME: ENDSLEY

DATE/TIME OF TEST: 01/28/2018 17:58:13

Diagnostics Check		OK	17:53	
Air Blank		0.000	17:54	
Control Test		0.079	17:54	
Air Blank		0.000	17:54	
Subject Sample	# 1		17:56	Subject Test Refused
Breath Volume		0.000	17:56	
Air Blank		0.000	17:57	
Control Test		0.079	17:57	
Air Blank		0.000	17:58	
Diagnostics Check		OK	17:58	

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000822

INSTRUMENT REGISTERED TO: LAKE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CORMAN,GERALD,E
BREATH TEST OPERATOR AGENCY: Lake County SO
ARRESTING OFFICER: CORMAN,GERALD
ARRESTING OFFICER AGENCY: Lake County SO
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 09:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:34

SUBJECT'S LAST NAME: BEAUCHEMIN

DATE/TIME OF TEST: 01/31/2018 11:01:59

Diagnostics Check		OK	10:55
Air Blank		0.000	10:56
Control Test		0.080	10:56
Air Blank		0.000	10:56
Subject Sample	# 1	0.233	10:57
Breath Volume		2.167	10:57
Air Blank		0.000	10:57
Air Blank		0.000	10:59
Subject Sample	# 2	0.232	11:00
Breath Volume		2.433	11:00
Air Blank		0.000	11:01
Control Test		0.080	11:01
Air Blank		0.000	11:01
Diagnostics Check		OK	11:01

Data Download Date/Time: 02/01/2018 19:29

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOMMERSDORF,JOHN,G
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: PINCUS,SETH
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:39

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 01/06/2018 03:05:51

Diagnostics Check		OK	03:02	
Air Blank		0.000	03:03	
Control Test		0.081	03:03	
Air Blank		0.000	03:04	
Subject Sample	# 1		03:04	Subject Test Refused
Breath Volume		0.000	03:04	
Air Blank		0.000	03:04	
Control Test		0.081	03:05	
Air Blank		0.000	03:05	
Diagnostics Check		OK	03:05	

Data Download Date/Time: 01/22/2018 02:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KETCHUM,CHARLES,A
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: EDWARDS,KEVIN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:24

SUBJECT'S LAST NAME: DUNNE

DATE/TIME OF TEST: 01/13/2018 02:49:25

Diagnostics Check		OK	02:46	
Air Blank		0.000	02:47	
Control Test		0.080	02:47	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.080	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 01/22/2018 02:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000824

INSTRUMENT REGISTERED TO: LEESBURG P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINCUS,SETH,B
BREATH TEST OPERATOR AGENCY: Leesburg PD
ARRESTING OFFICER: DANIELS,JOHN
ARRESTING OFFICER AGENCY: Leesburg PD
LAST AGENCY INSPECTION: 12/22/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 01316080A2
DRY GAS EXPIRY DATE: 02/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:22

SUBJECT'S LAST NAME: MILLS

DATE/TIME OF TEST: 01/21/2018 00:00:24

Diagnostics Check		OK	23:46	
Air Blank		0.000	23:47	
Control Test		0.081	23:47	
Air Blank		0.000	23:47	
Subject Sample	# 1	0.096	23:50	
Breath Volume		1.390	23:50	
Air Blank		0.000	23:50	
Air Blank		0.000	23:52	
Subject Sample	# 2	0.086	23:55	Volume Not Met
Breath Volume		0.656	23:55	
Air Blank		0.000	23:56	
Air Blank		0.000	23:58	
Subject Sample	# 3	0.090	23:58	
Breath Volume		1.839	23:58	
Air Blank		0.000	23:59	
Control Test		0.078	23:59	
Air Blank		0.000	00:00	
Diagnostics Check		OK	00:00	

Data Download Date/Time: 01/22/2018 02:22

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCCULLOCH,ADAM,G
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TAYLOR,KEITH
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 21:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:00

SUBJECT'S LAST NAME: MEADORS

DATE/TIME OF TEST: 01/13/2018 22:28:19

Diagnostics Check		OK	22:25	
Air Blank		0.000	22:25	
Control Test		0.080	22:26	
Air Blank		0.000	22:26	
Subject Sample	# 1		22:26	Subject Test Refused
Breath Volume		0.000	22:26	
Air Blank		0.000	22:27	
Control Test		0.080	22:27	
Air Blank		0.000	22:28	
Diagnostics Check		OK	22:28	

Data Download Date/Time: 01/31/2018 04:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,A,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: TAYLOR,K
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: EPPERSON

DATE/TIME OF TEST: 01/14/2018 02:16:30

Diagnostics Check		OK	02:10
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.155	02:11
Breath Volume		2.082	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.162	02:14
Breath Volume		4.242	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 01/31/2018 04:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANCEL,KAREN,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BROWN,J
ARRESTING OFFICER AGENCY: Umatilla PD
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 17:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:20

SUBJECT'S LAST NAME: LINDBLOOM

DATE/TIME OF TEST: 01/15/2018 20:47:09

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:44	
Control Test		0.081	20:44	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:45	Subject Test Refused
Breath Volume		0.000	20:45	
Air Blank		0.000	20:46	
Control Test		0.080	20:46	
Air Blank		0.000	20:47	
Diagnostics Check		OK	20:47	

Data Download Date/Time: 01/31/2018 04:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANCEL,KAREN,J
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: BEECK,B
ARRESTING OFFICER AGENCY: Mount Dora PD
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:40

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 01/16/2018 03:09:20

Diagnostics Check		OK	03:03
Air Blank		0.000	03:03
Control Test		0.079	03:03
Air Blank		0.000	03:04
Subject Sample	# 1	0.113	03:04
Breath Volume		2.617	03:04
Air Blank		0.000	03:05
Air Blank		0.000	03:07
Subject Sample	# 2	0.113	03:07
Breath Volume		2.703	03:07
Air Blank		0.000	03:08
Control Test		0.080	03:08
Air Blank		0.000	03:09
Diagnostics Check		OK	03:09

Data Download Date/Time: 01/31/2018 04:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000825

INSTRUMENT REGISTERED TO: MOUNT DORA P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RICE,A,
BREATH TEST OPERATOR AGENCY: Mount Dora PD
ARRESTING OFFICER: CARTER,T
ARRESTING OFFICER AGENCY: Eustis PD
LAST AGENCY INSPECTION: 12/23/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 583457
DRY GAS EXPIRY DATE: 10/20/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:41

SUBJECT'S LAST NAME: CONOVER

DATE/TIME OF TEST: 01/28/2018 01:09:31

Diagnostics Check		OK	01:06	
Air Blank		0.000	01:06	
Control Test		0.078	01:06	
Air Blank		0.000	01:07	
Subject Sample	# 1		01:08	Subject Test Refused
Breath Volume		0.000	01:08	
Air Blank		0.000	01:08	
Control Test		0.079	01:08	
Air Blank		0.000	01:09	
Diagnostics Check		OK	01:09	

Data Download Date/Time: 01/31/2018 04:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:32

SUBJECT'S LAST NAME: REMINGTON

DATE/TIME OF TEST: 01/19/2018 00:55:34

Diagnostics Check		OK	00:52	
Air Blank		0.000	00:53	
Control Test		0.082	00:53	
Air Blank		0.000	00:53	
Subject Sample	# 1		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:54	
Control Test		0.082	00:54	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 02/21/2018 14:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LETSON,REBECCA,G
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: LETSON,REBECCA
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: GASKIN

DATE/TIME OF TEST: 01/20/2018 03:30:05

Diagnostics Check		OK	03:17	
Air Blank		0.000	03:18	
Control Test		0.081	03:18	
Air Blank		0.000	03:19	
Subject Sample	# 1	0.155	03:22	Volume Not Met
Breath Volume		0.015	03:22	
Air Blank		0.000	03:23	
Air Blank		0.000	03:24	
Subject Sample	# 2	0.149	03:25	
Breath Volume		2.820	03:25	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 3	0.168	03:28	
Breath Volume		4.218	03:28	
Air Blank		0.000	03:29	
Control Test		0.080	03:29	
Air Blank		0.000	03:30	
Diagnostics Check		OK	03:30	

Data Download Date/Time: 02/21/2018 14:20

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DIMARE

DATE/TIME OF TEST: 01/20/2018 03:45:44

Diagnostics Check		OK	03:39
Air Blank		0.000	03:40
Control Test		0.081	03:40
Air Blank		0.000	03:40
Subject Sample	# 1	0.103	03:41
Breath Volume		2.285	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.098	03:44
Breath Volume		1.410	03:44
Air Blank		0.000	03:44
Control Test		0.081	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZEZUSKI,J,J
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ZEZUSKI,J
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 16:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/20/2018 19:44:24

Diagnostics Check		OK	19:38
Air Blank		0.000	19:38
Control Test		0.080	19:38
Air Blank		0.000	19:39
Subject Sample	# 1	0.000	19:39
Breath Volume		3.558	19:39
Air Blank		0.000	19:40
Air Blank		0.000	19:42
Subject Sample	# 2	0.000	19:43
Breath Volume		2.328	19:43
Air Blank		0.000	19:43
Control Test		0.081	19:43
Air Blank		0.000	19:44
Diagnostics Check		OK	19:44

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/21/2018 02:44:41

Diagnostics Check		OK	02:38
Air Blank		0.000	02:39
Control Test		0.081	02:39
Air Blank		0.000	02:39
Subject Sample	# 1	0.153	02:40
Breath Volume		2.000	02:40
Air Blank		0.000	02:40
Air Blank		0.000	02:42
Subject Sample	# 2	0.149	02:43
Breath Volume		2.253	02:43
Air Blank		0.000	02:43
Control Test		0.081	02:44
Air Blank		0.000	02:44
Diagnostics Check		OK	02:44

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FERNANDEZ,ALAIN,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: FERNANDEZ,ALAIN
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 04:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:10

SUBJECT'S LAST NAME: TUGGLE

DATE/TIME OF TEST: 01/21/2018 05:37:47

Diagnostics Check		OK	05:30
Air Blank		0.000	05:31
Control Test		0.080	05:31
Air Blank		0.000	05:32
Subject Sample	# 1	0.200	05:33
Breath Volume		1.937	05:33
Air Blank		0.000	05:33
Air Blank		0.000	05:35
Subject Sample	# 2	0.206	05:36
Breath Volume		2.339	05:36
Air Blank		0.000	05:36
Control Test		0.080	05:37
Air Blank		0.000	05:37
Diagnostics Check		OK	05:37

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:54

SUBJECT'S LAST NAME: WELSH

DATE/TIME OF TEST: 01/22/2018 01:23:14

Diagnostics Check		OK	01:16
Air Blank		0.000	01:16
Control Test		0.079	01:17
Air Blank		0.000	01:17
Subject Sample	# 1	0.157	01:18
Breath Volume		1.148	01:18
Air Blank		0.000	01:19
Air Blank		0.000	01:20
Subject Sample	# 2	0.155	01:21
Breath Volume		1.312	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:23
Diagnostics Check		OK	01:23

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,JEFFERY,A
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: HILLIARD,JEFFERY
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 16:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: GRANT

DATE/TIME OF TEST: 01/24/2018 17:07:23

Diagnostics Check		OK	17:00
Air Blank		0.000	17:01
Control Test		0.081	17:01
Air Blank		0.000	17:02
Subject Sample	# 1	0.242	17:02
Breath Volume		3.503	17:02
Air Blank		0.000	17:03
Air Blank		0.000	17:05
Subject Sample	# 2	0.246	17:05
Breath Volume		3.566	17:05
Air Blank		0.000	17:06
Control Test		0.079	17:06
Air Blank		0.000	17:07
Diagnostics Check		OK	17:07

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: NICOLAYSEN

DATE/TIME OF TEST: 01/24/2018 23:49:05

Diagnostics Check		OK	23:43
Air Blank		0.000	23:43
Control Test		0.081	23:43
Air Blank		0.000	23:44
Subject Sample	# 1	0.217	23:44
Breath Volume		2.031	23:44
Air Blank		0.000	23:45
Air Blank		0.000	23:47
Subject Sample	# 2	0.218	23:47
Breath Volume		2.242	23:47
Air Blank		0.000	23:48
Control Test		0.081	23:48
Air Blank		0.000	23:49
Diagnostics Check		OK	23:49

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: UMANA,RAUL,J
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: UMANA,RAUL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 644476
DRY GAS EXPIRY DATE: 12/01/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 01/25/2018 00:53:32

Diagnostics Check		OK	00:42	
Air Blank		0.000	00:42	
Control Test		0.081	00:42	
Air Blank		0.000	00:43	
Subject Sample	# 1	0.068	00:46	Volume Not Met
Breath Volume		0.773	00:46	
Air Blank		0.000	00:47	
Air Blank		0.000	00:48	
Subject Sample	# 2	0.051	00:52	Volume Not Met
Breath Volume		0.285	00:52	
Air Blank		0.000	00:52	
Control Test		0.081	00:52	
Air Blank		0.000	00:53	
Diagnostics Check		OK	00:53	

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLER,GARY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: MILLER,GARY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: COLE

DATE/TIME OF TEST: 01/27/2018 01:59:11

Diagnostics Check		OK	01:52
Air Blank		0.000	01:53
Control Test		0.080	01:53
Air Blank		0.000	01:53
Subject Sample	# 1	0.111	01:54
Breath Volume		2.871	01:54
Air Blank		0.000	01:55
Air Blank		0.000	01:57
Subject Sample	# 2	0.107	01:57
Breath Volume		2.085	01:57
Air Blank		0.000	01:58
Control Test		0.079	01:58
Air Blank		0.000	01:59
Diagnostics Check		OK	01:59

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SOLLAZZO,C,
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: SOLLAZZO,C
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: STANFORD

DATE/TIME OF TEST: 01/27/2018 03:29:51

Diagnostics Check		OK	03:18
Air Blank		0.000	03:19
Control Test		0.081	03:19
Air Blank		0.000	03:20
Subject Sample	# 1	0.192	03:20
Breath Volume		1.835	03:20
Air Blank		0.000	03:21
Air Blank		0.000	03:23
Subject Sample	# 2	0.170	03:24
Breath Volume		1.148	03:24
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 3	0.173	03:28
Breath Volume		2.371	03:28
Air Blank		0.000	03:28
Control Test		0.081	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:29

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOOREHEAD,DAVID,L
BREATH TEST OPERATOR AGENCY: Ocala PD
ARRESTING OFFICER: MOOREHEAD,DAVID
ARRESTING OFFICER AGENCY: Ocala PD
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: ANDARA ALVARADO

DATE/TIME OF TEST: 01/27/2018 23:07:23

Diagnostics Check		OK	23:01
Air Blank		0.000	23:01
Control Test		0.079	23:01
Air Blank		0.000	23:02
Subject Sample	# 1	0.158	23:02
Breath Volume		2.523	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.160	23:05
Breath Volume		2.183	23:05
Air Blank		0.000	23:06
Control Test		0.078	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,MICHAEL,T
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: THOMAS,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 14:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:42

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/29/2018 15:18:53

Diagnostics Check		OK	15:10
Air Blank		0.000	15:11
Control Test		0.081	15:11
Air Blank		0.000	15:11
Subject Sample	# 1	0.159	15:12
Breath Volume		1.160	15:12
Air Blank		0.000	15:13
Air Blank		0.000	15:15
Subject Sample	# 2	0.168	15:17
Breath Volume		1.386	15:17
Air Blank		0.000	15:17
Control Test		0.079	15:18
Air Blank		0.000	15:18
Diagnostics Check		OK	15:18

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000827

INSTRUMENT REGISTERED TO: MARION COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARTMAN,RILEY,A
BREATH TEST OPERATOR AGENCY: Marion County SO
ARRESTING OFFICER: ARTMAN,RILEY
ARRESTING OFFICER AGENCY: Marion County SO
LAST AGENCY INSPECTION: 01/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 21:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:40

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 01/30/2018 23:07:56

Diagnostics Check		OK	23:00
Air Blank		0.000	23:00
Control Test		0.082	23:01
Air Blank		0.000	23:01
Subject Sample	# 1	0.000	23:02
Breath Volume		5.089	23:02
Air Blank		0.000	23:03
Air Blank		0.000	23:05
Subject Sample	# 2	0.000	23:06
Breath Volume		2.984	23:06
Air Blank		0.000	23:07
Control Test		0.079	23:07
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 02/21/2018 14:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:53

SUBJECT'S LAST NAME: FIGUEROA

DATE/TIME OF TEST: 01/01/2018 02:21:09

Diagnostics Check		OK	02:14
Air Blank		0.000	02:14
Control Test		0.080	02:14
Air Blank		0.000	02:15
Subject Sample	# 1	0.188	02:16
Breath Volume		1.449	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.190	02:19
Breath Volume		2.218	02:19
Air Blank		0.000	02:20
Control Test		0.078	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 01/03/2018 19:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: CRAGG,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: DAVILA

DATE/TIME OF TEST: 01/01/2018 03:30:00

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.146	03:25
Breath Volume		1.710	03:25
Air Blank		0.000	03:26
Air Blank		0.000	03:27
Subject Sample	# 2	0.148	03:28
Breath Volume		2.671	03:28
Air Blank		0.000	03:29
Control Test		0.079	03:29
Air Blank		0.000	03:29
Diagnostics Check		OK	03:30

Data Download Date/Time: 01/03/2018 19:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: DUNCAN,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:09

SUBJECT'S LAST NAME: JUETTNER II

DATE/TIME OF TEST: 01/01/2018 05:35:55

Diagnostics Check		OK	05:29
Air Blank		0.000	05:29
Control Test		0.079	05:30
Air Blank		0.000	05:30
Subject Sample	# 1	0.126	05:31
Breath Volume		3.976	05:31
Air Blank		0.000	05:31
Air Blank		0.000	05:33
Subject Sample	# 2	0.128	05:34
Breath Volume		3.871	05:34
Air Blank		0.000	05:35
Control Test		0.080	05:35
Air Blank		0.000	05:35
Diagnostics Check		OK	05:35

Data Download Date/Time: 01/03/2018 19:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: BATISTA

DATE/TIME OF TEST: 01/01/2018 06:46:49

Diagnostics Check		OK	06:40
Air Blank		0.000	06:41
Control Test		0.079	06:41
Air Blank		0.000	06:41
Subject Sample	# 1	0.138	06:42
Breath Volume		1.437	06:42
Air Blank		0.000	06:43
Air Blank		0.000	06:44
Subject Sample	# 2	0.133	06:45
Breath Volume		1.117	06:45
Air Blank		0.000	06:45
Control Test		0.079	06:46
Air Blank		0.000	06:46
Diagnostics Check		OK	06:46

Data Download Date/Time: 01/03/2018 19:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,MELONIE,E
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: OTHER
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:00

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 01/01/2018 09:21:17

Diagnostics Check		OK	09:15
Air Blank		0.000	09:15
Control Test		0.079	09:16
Air Blank		0.000	09:16
Subject Sample	# 1	0.072	09:16
Breath Volume		1.339	09:16
Air Blank		0.000	09:17
Air Blank		0.000	09:19
Subject Sample	# 2	0.073	09:19
Breath Volume		1.550	09:19
Air Blank		0.000	09:20
Control Test		0.079	09:20
Air Blank		0.000	09:21
Diagnostics Check		OK	09:21

Data Download Date/Time: 01/03/2018 19:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,REBECCA,J
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NA,NA
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 15:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:52

SUBJECT'S LAST NAME: AKINS JR

DATE/TIME OF TEST: 01/01/2018 17:38:08

Diagnostics Check		OK	17:32	
Air Blank		0.000	17:32	
Control Test		0.080	17:33	
Air Blank		0.000	17:33	
Subject Sample	# 1	0.186	17:34	
Breath Volume		2.203	17:34	
Air Blank		0.000	17:34	
Air Blank		0.000	17:36	
Subject Sample	# 2		17:36	Subject Test Refused
Breath Volume		0.000	17:36	
Air Blank		0.000	17:37	
Control Test		0.081	17:37	
Air Blank		0.000	17:38	
Diagnostics Check		OK	17:38	

Data Download Date/Time: 01/03/2018 19:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 12/19/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: DEJESUS

DATE/TIME OF TEST: 01/02/2018 20:42:26

Diagnostics Check		OK	20:35
Air Blank		0.000	20:36
Control Test		0.080	20:36
Air Blank		0.000	20:37
Subject Sample	# 1	0.403	20:38
Breath Volume		2.246	20:38
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.394	20:40
Breath Volume		1.488	20:40
Air Blank		0.000	20:41
Control Test		0.080	20:41
Air Blank		0.000	20:42
Diagnostics Check		OK	20:42

Data Download Date/Time: 01/03/2018 19:45

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HERNANDEZ,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: COUTURIER

DATE/TIME OF TEST: 01/24/2018 00:40:33

Diagnostics Check		OK	00:33
Air Blank		0.000	00:34
Control Test		0.080	00:34
Air Blank		0.000	00:35
Subject Sample	# 1	0.235	00:35
Breath Volume		2.503	00:35
Air Blank		0.000	00:36
Air Blank		0.000	00:38
Subject Sample	# 2	0.216	00:39
Breath Volume		3.332	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Diagnostics Check		OK	00:40

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOODEN,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 01/24/2018 04:18:33

Diagnostics Check		OK	04:03	
Air Blank		0.000	04:03	
Control Test		0.080	04:03	
Air Blank		0.000	04:04	
Subject Sample	# 1	0.265	04:07	Volume Not Met
Breath Volume		0.925	04:07	
Air Blank		0.000	04:08	
Air Blank		0.000	04:10	
Subject Sample	# 2	0.293	04:11	
Breath Volume		1.125	04:11	
Air Blank		0.000	04:12	
Air Blank		0.000	04:14	
Subject Sample	# 3	0.298	04:16	
Breath Volume		1.160	04:16	
Air Blank		0.000	04:17	
Control Test		0.079	04:17	
Air Blank		0.000	04:18	
Diagnostics Check		OK	04:18	

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:59

SUBJECT'S LAST NAME: GREAR

DATE/TIME OF TEST: 01/25/2018 23:22:52

Diagnostics Check		OK	23:19	
Air Blank		0.000	23:19	
Control Test		0.081	23:20	
Air Blank		0.000	23:20	
Subject Sample	# 1		23:21	Subject Test Refused
Breath Volume		0.000	23:21	
Air Blank		0.000	23:21	
Control Test		0.080	23:22	
Air Blank		0.000	23:22	
Diagnostics Check		OK	23:22	

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HONEYGHON,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:49

SUBJECT'S LAST NAME: BAKER

DATE/TIME OF TEST: 01/26/2018 04:16:08

Diagnostics Check		OK	04:09
Air Blank		0.000	04:10
Control Test		0.081	04:10
Air Blank		0.000	04:11
Subject Sample	# 1	0.111	04:11
Breath Volume		1.679	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:14
Subject Sample	# 2	0.111	04:14
Breath Volume		2.406	04:14
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 05:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:05

SUBJECT'S LAST NAME: PEREZ-GOMEZ

DATE/TIME OF TEST: 01/27/2018 06:32:52

Diagnostics Check		OK	06:25
Air Blank		0.000	06:26
Control Test		0.080	06:26
Air Blank		0.000	06:26
Subject Sample	# 1	0.197	06:28
Breath Volume		2.015	06:28
Air Blank		0.000	06:29
Air Blank		0.000	06:31
Subject Sample	# 2	0.197	06:31
Breath Volume		1.300	06:31
Air Blank		0.000	06:31
Control Test		0.078	06:32
Air Blank		0.000	06:32
Diagnostics Check		OK	06:32

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:17

SUBJECT'S LAST NAME: DIRR

DATE/TIME OF TEST: 01/27/2018 23:43:40

Diagnostics Check		OK	23:37
Air Blank		0.000	23:37
Control Test		0.080	23:38
Air Blank		0.000	23:38
Subject Sample	# 1	0.151	23:39
Breath Volume		1.550	23:39
Air Blank		0.000	23:39
Air Blank		0.000	23:41
Subject Sample	# 2	0.149	23:42
Breath Volume		1.218	23:42
Air Blank		0.000	23:42
Control Test		0.079	23:43
Air Blank		0.000	23:43
Diagnostics Check		OK	23:43

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PAULINO,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: HERNANDEZ-HERNANDEZ

DATE/TIME OF TEST: 01/28/2018 00:17:17

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.188	00:12
Breath Volume		2.308	00:12
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.194	00:15
Breath Volume		2.113	00:15
Air Blank		0.000	00:16
Control Test		0.079	00:16
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VAN TREESE,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: RAMSFIELD

DATE/TIME OF TEST: 01/28/2018 00:52:20

Diagnostics Check		OK	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:46
Air Blank		0.000	00:47
Subject Sample	# 1	0.135	00:47
Breath Volume		3.074	00:47
Air Blank		0.000	00:48
Air Blank		0.000	00:50
Subject Sample	# 2	0.132	00:50
Breath Volume		3.800	00:50
Air Blank		0.000	00:51
Control Test		0.079	00:51
Air Blank		0.000	00:52
Diagnostics Check		OK	00:52

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ARUNDALE,BRITTANY,T
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PECYLAK,M
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:31

SUBJECT'S LAST NAME: FLORES VICENTE

DATE/TIME OF TEST: 01/28/2018 04:58:30

Diagnostics Check		OK	04:52
Air Blank		0.000	04:52
Control Test		0.079	04:53
Air Blank		0.000	04:53
Subject Sample	# 1	0.131	04:54
Breath Volume		2.546	04:54
Air Blank		0.000	04:54
Air Blank		0.000	04:56
Subject Sample	# 2	0.132	04:57
Breath Volume		3.199	04:57
Air Blank		0.000	04:57
Control Test		0.079	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: COSTA,A
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:32

SUBJECT'S LAST NAME: MAHAMUD

DATE/TIME OF TEST: 01/28/2018 22:00:11

Diagnostics Check		OK	21:53
Air Blank		0.000	21:53
Control Test		0.079	21:53
Air Blank		0.000	21:54
Subject Sample	# 1	0.121	21:55
Breath Volume		2.082	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.126	21:58
Breath Volume		2.007	21:58
Air Blank		0.000	21:59
Control Test		0.079	21:59
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:06

SUBJECT'S LAST NAME: PLASTIC

DATE/TIME OF TEST: 01/29/2018 01:33:28

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.118	01:28
Breath Volume		2.410	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.117	01:31
Breath Volume		2.054	01:31
Air Blank		0.000	01:32
Control Test		0.078	01:32
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BARFIELD-CRAIG,C
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: MENDOZA

DATE/TIME OF TEST: 01/29/2018 23:19:37

Diagnostics Check		OK	23:12
Air Blank		0.000	23:13
Control Test		0.080	23:13
Air Blank		0.000	23:14
Subject Sample	# 1	0.160	23:15
Breath Volume		1.800	23:15
Air Blank		0.000	23:15
Air Blank		0.000	23:17
Subject Sample	# 2	0.156	23:18
Breath Volume		2.441	23:18
Air Blank		0.000	23:18
Control Test		0.080	23:19
Air Blank		0.000	23:19
Diagnostics Check		OK	23:19

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: HEIST,A 332
ARRESTING OFFICER AGENCY: Plant City PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 14:48

SUBJECT'S LAST NAME: GIBSON

DATE/TIME OF TEST: 01/30/2018 15:15:12

Diagnostics Check		OK	15:08
Air Blank		0.000	15:09
Control Test		0.080	15:09
Air Blank		0.000	15:10
Subject Sample	# 1	0.065	15:10
Breath Volume		1.914	15:10
Air Blank		0.000	15:11
Air Blank		0.000	15:13
Subject Sample	# 2	0.068	15:13
Breath Volume		1.980	15:13
Air Blank		0.000	15:14
Control Test		0.081	15:14
Air Blank		0.000	15:15
Diagnostics Check		OK	15:15

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:16
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: MEJIA

DATE/TIME OF TEST: 01/31/2018 03:02:01

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.081	02:56
Air Blank		0.000	02:57
Subject Sample	# 1	0.205	02:57
Breath Volume		1.882	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.207	03:00
Breath Volume		2.230	03:00
Air Blank		0.000	03:01
Control Test		0.080	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:02

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: LUNDY

DATE/TIME OF TEST: 01/31/2018 04:50:02

Diagnostics Check		OK	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:43
Subject Sample	# 1	0.137	04:45
Breath Volume		2.777	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:48
Subject Sample	# 2	0.140	04:48
Breath Volume		2.394	04:48
Air Blank		0.000	04:49
Control Test		0.080	04:49
Air Blank		0.000	04:49
Diagnostics Check		OK	04:50

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WASHINGTON,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:52

SUBJECT'S LAST NAME: CARTER

DATE/TIME OF TEST: 01/31/2018 21:19:14

Diagnostics Check		OK	21:12
Air Blank		0.000	21:13
Control Test		0.080	21:13
Air Blank		0.000	21:14
Subject Sample	# 1	0.100	21:14
Breath Volume		1.890	21:14
Air Blank		0.000	21:15
Air Blank		0.000	21:17
Subject Sample	# 2	0.102	21:17
Breath Volume		2.390	21:17
Air Blank		0.000	21:18
Control Test		0.081	21:18
Air Blank		0.000	21:19
Diagnostics Check		OK	21:19

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000830

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: ORNDORFF,W
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:14

SUBJECT'S LAST NAME: CHILDERS

DATE/TIME OF TEST: 01/31/2018 23:42:41

Diagnostics Check		OK	23:34
Air Blank		0.000	23:34
Control Test		0.081	23:35
Air Blank		0.000	23:35
Subject Sample	# 1	0.160	23:38
Breath Volume		2.117	23:38
Air Blank		0.000	23:38
Air Blank		0.000	23:40
Subject Sample	# 2	0.158	23:41
Breath Volume		1.515	23:41
Air Blank		0.000	23:41
Control Test		0.080	23:42
Air Blank		0.000	23:42
Diagnostics Check		OK	23:42

Data Download Date/Time: 02/14/2018 10:12

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:55

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 01/03/2018 21:23:58

Diagnostics Check		OK	21:16
Air Blank		0.000	21:17
Control Test		0.077	21:17
Air Blank		0.000	21:18
Subject Sample	# 1	0.288	21:18
Breath Volume		1.292	21:18
Air Blank		0.000	21:19
Air Blank		0.000	21:21
Subject Sample	# 2	0.289	21:22
Breath Volume		1.312	21:22
Air Blank		0.000	21:23
Control Test		0.076	21:23
Air Blank		0.000	21:23
Diagnostics Check		OK	21:23

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMAS,CHRISTOPHER,L
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: THOMAS,CHRISTOPHER
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: TAYLOR

DATE/TIME OF TEST: 01/06/2018 01:33:40

Diagnostics Check		OK	01:27
Air Blank		0.000	01:27
Control Test		0.077	01:27
Air Blank		0.000	01:28
Subject Sample	# 1	0.179	01:28
Breath Volume		2.921	01:28
Air Blank		0.000	01:29
Air Blank		0.000	01:31
Subject Sample	# 2	0.173	01:32
Breath Volume		1.648	01:32
Air Blank		0.000	01:32
Control Test		0.076	01:33
Air Blank		0.000	01:33
Diagnostics Check		OK	01:33

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CERNUTO,MATTHEW,D
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MARTIN,WILLIAM
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 08:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 08:35

SUBJECT'S LAST NAME: DIBBLE

DATE/TIME OF TEST: 01/06/2018 09:05:40

Diagnostics Check		OK	08:59
Air Blank		0.000	08:59
Control Test		0.077	09:00
Air Blank		0.000	09:00
Subject Sample	# 1	0.000	09:01
Breath Volume		1.277	09:01
Air Blank		0.000	09:01
Air Blank		0.000	09:03
Subject Sample	# 2	0.000	09:04
Breath Volume		1.671	09:04
Air Blank		0.000	09:04
Control Test		0.077	09:05
Air Blank		0.000	09:05
Diagnostics Check		OK	09:05

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PORCELLI,ANDREW,
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: PORCELLI,ANDREW
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 06:55

SUBJECT'S LAST NAME: ANGARITA

DATE/TIME OF TEST: 01/07/2018 07:25:09

Diagnostics Check		OK	07:18
Air Blank		0.000	07:19
Control Test		0.076	07:19
Air Blank		0.000	07:20
Subject Sample	# 1	0.148	07:20
Breath Volume		1.781	07:20
Air Blank		0.000	07:21
Air Blank		0.000	07:23
Subject Sample	# 2	0.140	07:23
Breath Volume		2.058	07:23
Air Blank		0.000	07:24
Control Test		0.076	07:24
Air Blank		0.000	07:25
Diagnostics Check		OK	07:25

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAY,ERICK,J
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: DAY,ERICK
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: DE VAULT

DATE/TIME OF TEST: 01/11/2018 20:27:40

Diagnostics Check		OK	20:21	
Air Blank		0.000	20:21	
Control Test		0.076	20:21	
Air Blank		0.000	20:22	
Subject Sample	# 1	0.139	20:23	
Breath Volume		2.285	20:23	
Air Blank		0.000	20:23	
Air Blank		0.000	20:25	
Subject Sample	# 2	0.138	20:26	
Breath Volume		1.496	20:26	
Air Blank		0.000	20:26	
Control Test		0.074	20:27	Control Outside Tolerance
Air Blank		0.000	20:27	

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SERRA,CRP,
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: SERRA,CRP
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: WARREN

DATE/TIME OF TEST: 01/12/2018 01:22:19

Diagnostics Check		OK	01:15
Air Blank		0.000	01:16
Control Test		0.077	01:16
Air Blank		0.000	01:17
Subject Sample	# 1	0.176	01:17
Breath Volume		1.605	01:17
Air Blank		0.000	01:18
Air Blank		0.000	01:20
Subject Sample	# 2	0.175	01:20
Breath Volume		2.136	01:20
Air Blank		0.000	01:21
Control Test		0.077	01:21
Air Blank		0.000	01:22
Diagnostics Check		OK	01:22

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELLAS,DILLON,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BELLAS,DILLON
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: JIMENEZ JR

DATE/TIME OF TEST: 01/12/2018 23:38:36

Diagnostics Check		OK	23:32
Air Blank		0.000	23:32
Control Test		0.077	23:33
Air Blank		0.000	23:33
Subject Sample	# 1	0.052	23:34
Breath Volume		3.355	23:34
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 2	0.054	23:37
Breath Volume		3.328	23:37
Air Blank		0.000	23:37
Control Test		0.078	23:37
Air Blank		0.000	23:38
Diagnostics Check		OK	23:38

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGIGHAN,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MCGIGHAN,WILLIAM
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: LOIODICE

DATE/TIME OF TEST: 01/13/2018 05:07:11

Diagnostics Check		OK	05:00	
Air Blank		0.000	05:00	
Control Test		0.076	05:01	
Air Blank		0.000	05:01	
Subject Sample	# 1	0.127	05:02	
Breath Volume		2.617	05:02	
Air Blank		0.000	05:03	
Air Blank		0.000	05:05	
Subject Sample	# 2	0.109	05:05	
Breath Volume		1.464	05:05	
Air Blank		0.000	05:06	
Control Test		0.074	05:06	Control Outside Tolerance
Air Blank		0.000	05:07	

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGIGHAN,WILLIAM,T
BREATH TEST OPERATOR AGENCY: Stuart PD
ARRESTING OFFICER: MCGIGHAN,WILLIAM
ARRESTING OFFICER AGENCY: Stuart PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:34

SUBJECT'S LAST NAME: LOIODICE

DATE/TIME OF TEST: 01/13/2018 05:16:57

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.076	05:10
Air Blank		0.000	05:11
Subject Sample	# 1	0.129	05:12
Breath Volume		1.851	05:12
Air Blank		0.000	05:13
Air Blank		0.000	05:14
Subject Sample	# 2	0.122	05:15
Breath Volume		2.289	05:15
Air Blank		0.000	05:16
Control Test		0.075	05:16
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CUTY,J,K
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: CUTY,J
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 12:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 12:32

SUBJECT'S LAST NAME: KAUFMANN

DATE/TIME OF TEST: 01/17/2018 13:11:28

Diagnostics Check		OK	13:05
Air Blank		0.000	13:05
Control Test		0.077	13:05
Air Blank		0.000	13:06
Subject Sample	# 1	0.191	13:06
Breath Volume		1.976	13:06
Air Blank		0.000	13:07
Air Blank		0.000	13:09
Subject Sample	# 2	0.195	13:09
Breath Volume		1.976	13:09
Air Blank		0.000	13:10
Control Test		0.077	13:10
Air Blank		0.000	13:11
Diagnostics Check		OK	13:11

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:59

SUBJECT'S LAST NAME: POWELL

DATE/TIME OF TEST: 01/17/2018 20:43:19

Diagnostics Check		OK	20:35
Air Blank		0.000	20:35
Control Test		0.077	20:36
Air Blank		0.000	20:36
Subject Sample	# 1	0.132	20:37
Breath Volume		2.367	20:37
Air Blank		0.000	20:38
Air Blank		0.000	20:40
Subject Sample	# 2	0.124	20:41
Breath Volume		1.933	20:41
Air Blank		0.000	20:42
Control Test		0.077	20:42
Air Blank		0.000	20:43
Diagnostics Check		OK	20:43

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORROW,ELISHA,D
BREATH TEST OPERATOR AGENCY: FFWCC (West Palm)
ARRESTING OFFICER: CARROLL,MICHAEL
ARRESTING OFFICER AGENCY: FFWCC (West Palm)
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:14

SUBJECT'S LAST NAME: BOSHELL

DATE/TIME OF TEST: 01/17/2018 22:38:08

Diagnostics Check		OK	22:35	
Air Blank		0.000	22:35	
Control Test		0.078	22:35	
Air Blank		0.000	22:36	
Subject Sample	# 1		22:36	Subject Test Refused
Breath Volume		0.000	22:36	
Air Blank		0.000	22:37	
Control Test		0.077	22:37	
Air Blank		0.000	22:38	
Diagnostics Check		OK	22:38	

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLASZYK,JACOB,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: BLASZYK,JACOB
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: MEDINA

DATE/TIME OF TEST: 01/18/2018 23:50:24

Diagnostics Check		OK	23:43
Air Blank		0.000	23:44
Control Test		0.078	23:44
Air Blank		0.000	23:44
Subject Sample	# 1	0.214	23:45
Breath Volume		3.843	23:45
Air Blank		0.000	23:46
Air Blank		0.000	23:48
Subject Sample	# 2	0.206	23:48
Breath Volume		3.109	23:48
Air Blank		0.000	23:49
Control Test		0.076	23:49
Air Blank		0.000	23:50
Diagnostics Check		OK	23:50

Data Download Date/Time: 01/23/2018 10:57

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCDONALD,T,C
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: MCDONALD,T
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ROBINSON

DATE/TIME OF TEST: 01/20/2018 01:40:44

Diagnostics Check		OK	01:33
Air Blank		0.000	01:34
Control Test		0.077	01:34
Air Blank		0.000	01:35
Subject Sample	# 1	0.165	01:36
Breath Volume		2.417	01:36
Air Blank		0.000	01:36
Air Blank		0.000	01:38
Subject Sample	# 2	0.166	01:39
Breath Volume		2.343	01:39
Air Blank		0.000	01:39
Control Test		0.076	01:40
Air Blank		0.000	01:40
Diagnostics Check		OK	01:40

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLIZZI,CHRISTINE,A
BREATH TEST OPERATOR AGENCY: Martin County SO
ARRESTING OFFICER: POLIZZI,CHRISTINE
ARRESTING OFFICER AGENCY: Martin County SO
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:58

SUBJECT'S LAST NAME: HENDERSON

DATE/TIME OF TEST: 01/21/2018 21:25:21

Diagnostics Check		OK	21:19
Air Blank		0.000	21:19
Control Test		0.077	21:19
Air Blank		0.000	21:20
Subject Sample	# 1	0.116	21:20
Breath Volume		1.414	21:20
Air Blank		0.000	21:21
Air Blank		0.000	21:23
Subject Sample	# 2	0.117	21:23
Breath Volume		1.328	21:23
Air Blank		0.000	21:24
Control Test		0.077	21:24
Air Blank		0.000	21:25
Diagnostics Check		OK	21:25

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000831

INSTRUMENT REGISTERED TO: MARTIN COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,JUAN,A
BREATH TEST OPERATOR AGENCY: Sewall's Point PD
ARRESTING OFFICER: GARCIA,JUAN
ARRESTING OFFICER AGENCY: Sewall's Point PD
LAST AGENCY INSPECTION: 12/07/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 06616080A2
DRY GAS EXPIRY DATE: 04/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PACHUCKI

DATE/TIME OF TEST: 01/22/2018 02:28:51

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.077	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.115	02:24
Breath Volume		2.335	02:24
Air Blank		0.000	02:25
Air Blank		0.000	02:26
Subject Sample	# 2	0.114	02:27
Breath Volume		2.027	02:27
Air Blank		0.000	02:27
Control Test		0.075	02:28
Air Blank		0.000	02:28
Diagnostics Check		OK	02:28

Data Download Date/Time: 01/23/2018 10:57

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: PALFY,R
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/05/2018 02:17:36

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:14	
Control Test		0.081	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1		02:16	Subject Test Refused
Breath Volume		0.000	02:16	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:17	
Diagnostics Check		OK	02:17	

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/06/2018 00:47:43

Diagnostics Check		OK	00:41
Air Blank		0.000	00:41
Control Test		0.081	00:41
Air Blank		0.000	00:42
Subject Sample	# 1	0.170	00:42
Breath Volume		2.511	00:42
Air Blank		0.000	00:43
Air Blank		0.000	00:45
Subject Sample	# 2	0.166	00:45
Breath Volume		2.621	00:45
Air Blank		0.000	00:46
Control Test		0.080	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: ROATEN JR

DATE/TIME OF TEST: 01/06/2018 04:44:04

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.081	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.099	04:39
Breath Volume		3.312	04:39
Air Blank		0.000	04:39
Air Blank		0.000	04:41
Subject Sample	# 2	0.101	04:42
Breath Volume		3.531	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:43
Diagnostics Check		OK	04:44

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BLANCHARD,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: HARYOKO

DATE/TIME OF TEST: 01/07/2018 04:16:46

Diagnostics Check		OK	04:07
Air Blank		0.000	04:08
Control Test		0.081	04:08
Air Blank		0.000	04:08
Subject Sample	# 1	0.145	04:11
Breath Volume		1.339	04:11
Air Blank		0.000	04:12
Air Blank		0.000	04:13
Subject Sample	# 2	0.152	04:15
Breath Volume		1.386	04:15
Air Blank		0.000	04:15
Control Test		0.079	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: NIEUWENDAM,R
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:11

SUBJECT'S LAST NAME: PEARSON

DATE/TIME OF TEST: 01/07/2018 05:40:09

Diagnostics Check		OK	05:33
Air Blank		0.000	05:34
Control Test		0.080	05:34
Air Blank		0.000	05:34
Subject Sample	# 1	0.155	05:35
Breath Volume		4.339	05:35
Air Blank		0.000	05:36
Air Blank		0.000	05:37
Subject Sample	# 2	0.157	05:38
Breath Volume		4.308	05:38
Air Blank		0.000	05:39
Control Test		0.079	05:39
Air Blank		0.000	05:40
Diagnostics Check		OK	05:40

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: WOLSTENHOLM,A
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:32
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: THERIAULT

DATE/TIME OF TEST: 01/12/2018 01:28:21

Diagnostics Check		OK	01:22
Air Blank		0.000	01:22
Control Test		0.079	01:22
Air Blank		0.000	01:23
Subject Sample	# 1	0.117	01:23
Breath Volume		1.148	01:23
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.127	01:26
Breath Volume		1.246	01:26
Air Blank		0.000	01:27
Control Test		0.079	01:27
Air Blank		0.000	01:28
Diagnostics Check		OK	01:28

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BURKE,B
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:34

SUBJECT'S LAST NAME: HUDSON

DATE/TIME OF TEST: 01/12/2018 04:11:40

Diagnostics Check		OK	04:02
Air Blank		0.000	04:03
Control Test		0.080	04:03
Air Blank		0.000	04:04
Subject Sample	# 1	0.042	04:06
Breath Volume		2.027	04:06
Air Blank		0.000	04:07
Air Blank		0.000	04:09
Subject Sample	# 2	0.041	04:10
Breath Volume		1.281	04:10
Air Blank		0.000	04:10
Control Test		0.078	04:11
Air Blank		0.000	04:11
Diagnostics Check		OK	04:11

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:44

SUBJECT'S LAST NAME: NUNEZ

DATE/TIME OF TEST: 01/13/2018 03:12:23

Diagnostics Check		OK	03:04
Air Blank		0.000	03:05
Control Test		0.080	03:05
Air Blank		0.000	03:06
Subject Sample	# 1	0.189	03:07
Breath Volume		1.375	03:07
Air Blank		0.000	03:07
Air Blank		0.000	03:09
Subject Sample	# 2	0.199	03:10
Breath Volume		1.277	03:10
Air Blank		0.000	03:11
Control Test		0.078	03:11
Air Blank		0.000	03:12
Diagnostics Check		OK	03:12

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SHEA,D
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:58

SUBJECT'S LAST NAME: NGUYEN

DATE/TIME OF TEST: 01/14/2018 01:24:17

Diagnostics Check		OK	01:18
Air Blank		0.000	01:18
Control Test		0.081	01:18
Air Blank		0.000	01:19
Subject Sample	# 1	0.000	01:19
Breath Volume		1.410	01:19
Air Blank		0.000	01:20
Air Blank		0.000	01:22
Subject Sample	# 2	0.000	01:22
Breath Volume		1.656	01:22
Air Blank		0.000	01:23
Control Test		0.080	01:23
Air Blank		0.000	01:24
Diagnostics Check		OK	01:24

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: SNAGGS,S
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/14/2018 22:22:21

Diagnostics Check		OK	22:15
Air Blank		0.000	22:16
Control Test		0.080	22:16
Air Blank		0.000	22:17
Subject Sample	# 1	0.180	22:17
Breath Volume		2.062	22:17
Air Blank		0.000	22:18
Air Blank		0.000	22:20
Subject Sample	# 2	0.182	22:20
Breath Volume		1.550	22:20
Air Blank		0.000	22:21
Control Test		0.080	22:21
Air Blank		0.000	22:22
Diagnostics Check		OK	22:22

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:29

SUBJECT'S LAST NAME: PRITCHETT

DATE/TIME OF TEST: 01/15/2018 00:57:45

Diagnostics Check		OK	00:51
Air Blank		0.000	00:51
Control Test		0.081	00:51
Air Blank		0.000	00:52
Subject Sample	# 1	0.168	00:53
Breath Volume		1.480	00:53
Air Blank		0.000	00:53
Air Blank		0.000	00:55
Subject Sample	# 2	0.167	00:56
Breath Volume		1.175	00:56
Air Blank		0.000	00:56
Control Test		0.080	00:57
Air Blank		0.000	00:57
Diagnostics Check		OK	00:57

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:50

SUBJECT'S LAST NAME: BRETON-GONZALEZ

DATE/TIME OF TEST: 01/18/2018 23:16:42

Diagnostics Check		OK	23:10
Air Blank		0.000	23:10
Control Test		0.081	23:11
Air Blank		0.000	23:11
Subject Sample	# 1	0.000	23:12
Breath Volume		1.609	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.000	23:15
Breath Volume		1.902	23:15
Air Blank		0.000	23:15
Control Test		0.081	23:16
Air Blank		0.000	23:16
Diagnostics Check		OK	23:16

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCMINN,KATHY,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:32

SUBJECT'S LAST NAME: RAMIREZ

DATE/TIME OF TEST: 01/19/2018 04:00:49

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.081	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.174	03:55
Breath Volume		2.132	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.184	03:59
Breath Volume		2.781	03:59
Air Blank		0.000	03:59
Control Test		0.079	04:00
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: FRANKLIN,R
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:19

SUBJECT'S LAST NAME: THIBOUTOT

DATE/TIME OF TEST: 01/20/2018 01:46:28

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:41
Subject Sample	# 1	0.116	01:41
Breath Volume		2.253	01:41
Air Blank		0.000	01:42
Air Blank		0.000	01:44
Subject Sample	# 2	0.108	01:44
Breath Volume		1.261	01:44
Air Blank		0.000	01:45
Control Test		0.080	01:45
Air Blank		0.000	01:46
Diagnostics Check		OK	01:46

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAIL,JULIE,A
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: VISSER,A
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:11

SUBJECT'S LAST NAME: BATTISE JR

DATE/TIME OF TEST: 01/20/2018 03:38:24

Diagnostics Check		OK	03:31
Air Blank		0.000	03:32
Control Test		0.080	03:32
Air Blank		0.000	03:33
Subject Sample	# 1	0.147	03:33
Breath Volume		1.921	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.146	03:36
Breath Volume		2.070	03:36
Air Blank		0.000	03:37
Control Test		0.080	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLLAND,CHARLEENE,N
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BRABAND,M
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:32

SUBJECT'S LAST NAME: URENA

DATE/TIME OF TEST: 01/21/2018 02:58:43

Diagnostics Check		OK	02:52
Air Blank		0.000	02:52
Control Test		0.080	02:53
Air Blank		0.000	02:53
Subject Sample	# 1	0.000	02:54
Breath Volume		2.214	02:54
Air Blank		0.000	02:54
Air Blank		0.000	02:56
Subject Sample	# 2	0.000	02:57
Breath Volume		2.761	02:57
Air Blank		0.000	02:57
Control Test		0.079	02:58
Air Blank		0.000	02:58
Diagnostics Check		OK	02:58

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANNIS,RONALD,R
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: BELEN,C
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: ACEVEDO ECHEVARRIA

DATE/TIME OF TEST: 01/21/2018 03:59:21

Diagnostics Check		OK	03:52
Air Blank		0.000	03:52
Control Test		0.079	03:53
Air Blank		0.000	03:53
Subject Sample	# 1	0.130	03:54
Breath Volume		2.636	03:54
Air Blank		0.000	03:55
Air Blank		0.000	03:57
Subject Sample	# 2	0.133	03:57
Breath Volume		2.671	03:57
Air Blank		0.000	03:58
Control Test		0.078	03:58
Air Blank		0.000	03:59
Diagnostics Check		OK	03:59

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TORRES,JENIFER,
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: GONZALEZ,J
ARRESTING OFFICER AGENCY: Tampa PD
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:36

SUBJECT'S LAST NAME: ROUTH

DATE/TIME OF TEST: 01/22/2018 04:05:23

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.080	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.179	04:00
Breath Volume		2.843	04:00
Air Blank		0.000	04:01
Air Blank		0.000	04:02
Subject Sample	# 2	0.162	04:03
Breath Volume		1.277	04:03
Air Blank		0.000	04:04
Control Test		0.079	04:04
Air Blank		0.000	04:05
Diagnostics Check		OK	04:05

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000833

INSTRUMENT REGISTERED TO: HILLSBOROUGH CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SKIPPER,ROGER,G
BREATH TEST OPERATOR AGENCY: Hillsborough Co SO
ARRESTING OFFICER: TEST,TEST
ARRESTING OFFICER AGENCY: Hillsborough Co SO
LAST AGENCY INSPECTION: 01/03/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 03417080A1
DRY GAS EXPIRY DATE: 03/05/2019
VIOLATION CODE: SYS CHECK
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:00

SUBJECT'S LAST NAME: TEST

DATE/TIME OF TEST: 01/23/2018 13:51:50

Diagnostics Check		OK	13:44
Air Blank		0.000	13:44
Control Test		0.080	13:45
Air Blank		0.000	13:45
Subject Sample	# 1	0.000	13:47
Breath Volume		1.269	13:47
Air Blank		0.000	13:48
Air Blank		0.000	13:50
Subject Sample	# 2	0.000	13:50
Breath Volume		1.675	13:50
Air Blank		0.000	13:50
Control Test		0.080	13:51
Air Blank		0.000	13:51
Diagnostics Check		OK	13:51

Data Download Date/Time: 02/13/2018 18:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JINKS

DATE/TIME OF TEST: 01/19/2018 23:17:17

Diagnostics Check		OK	23:10
Air Blank		0.000	23:11
Control Test		0.078	23:11
Air Blank		0.000	23:12
Subject Sample	# 1	0.257	23:12
Breath Volume		3.054	23:12
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.263	23:15
Breath Volume		4.199	23:15
Air Blank		0.000	23:16
Control Test		0.078	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/26/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: STEWART

DATE/TIME OF TEST: 01/20/2018 01:43:10

Diagnostics Check		OK	01:36
Air Blank		0.000	01:37
Control Test		0.078	01:37
Air Blank		0.000	01:37
Subject Sample	# 1	0.007	01:38
Breath Volume		2.691	01:38
Air Blank		0.000	01:39
Air Blank		0.000	01:41
Subject Sample	# 2	0.008	01:41
Breath Volume		4.125	01:41
Air Blank		0.000	01:42
Control Test		0.078	01:42
Air Blank		0.000	01:43
Diagnostics Check		OK	01:43

Data Download Date/Time: 02/26/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NUSBAUM,K,
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: NUSBAUM,K
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:52

SUBJECT'S LAST NAME: STRUBE

DATE/TIME OF TEST: 01/20/2018 23:40:02

Diagnostics Check		OK	23:37	
Air Blank		0.000	23:37	
Control Test		0.077	23:37	
Air Blank		0.000	23:38	
Subject Sample	# 1		23:38	Subject Test Refused
Breath Volume		0.000	23:38	
Air Blank		0.000	23:39	
Control Test		0.078	23:39	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:40	

Data Download Date/Time: 02/26/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MAYFIELD,MATTHEW,E
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: MAYFIELD,MATTHEW
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: KAFKA

DATE/TIME OF TEST: 01/21/2018 02:30:03

Diagnostics Check		OK	02:27	
Air Blank		0.000	02:27	
Control Test		0.077	02:28	
Air Blank		0.000	02:28	
Subject Sample	# 1		02:28	Subject Test Refused
Breath Volume		0.000	02:28	
Air Blank		0.000	02:29	
Control Test		0.077	02:29	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:30	

Data Download Date/Time: 02/26/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: HARBESON

DATE/TIME OF TEST: 01/31/2018 17:19:55

Diagnostics Check	OK	17:18	
Air Blank	0.000	17:19	
Control Test	0.077	17:19	
Air Blank		17:19	RFI Detect
Air Blank		17:19	RFI Detect

Data Download Date/Time: 02/26/2018 16:56

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000840

INSTRUMENT REGISTERED TO: LYNN HAVEN PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DUTTON,KERRY,D
BREATH TEST OPERATOR AGENCY: Lynn Haven PD
ARRESTING OFFICER: DUTTON,KERRY
ARRESTING OFFICER AGENCY: Lynn Haven PD
LAST AGENCY INSPECTION: 01/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:40

SUBJECT'S LAST NAME: HARBESON

DATE/TIME OF TEST: 01/31/2018 17:24:26

Diagnostics Check		OK	17:21	
Air Blank		0.000	17:22	
Control Test		0.078	17:22	
Air Blank		0.000	17:22	
Subject Sample	# 1		17:23	Subject Test Refused
Breath Volume		0.000	17:23	
Air Blank		0.000	17:23	
Control Test		0.077	17:23	
Air Blank		0.000	17:24	
Diagnostics Check		OK	17:24	

Data Download Date/Time: 02/26/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BLANCO CASTILLO

DATE/TIME OF TEST: 01/01/2018 02:13:18

Diagnostics Check		OK	02:06	
Air Blank		0.000	02:07	
Control Test		0.080	02:07	
Air Blank		0.000	02:08	
Subject Sample	# 1	0.172	02:08	
Breath Volume		2.246	02:08	
Air Blank		0.000	02:09	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.166	02:12	
Breath Volume		2.421	02:12	
Air Blank			02:12	RFI Detect
Air Blank		0.000	02:13	

Data Download Date/Time: 01/10/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: BLANCO CASTILLO

DATE/TIME OF TEST: 01/01/2018 02:21:16

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Subject Sample	# 1	0.163	02:16
Breath Volume		2.214	02:16
Air Blank		0.000	02:17
Air Blank		0.000	02:19
Subject Sample	# 2	0.159	02:19
Breath Volume		2.015	02:19
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Diagnostics Check		OK	02:21

Data Download Date/Time: 01/10/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: COLGAN,STEVEN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: OTHER
ARREST TIME: 02:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:17

SUBJECT'S LAST NAME: BRINLEY

DATE/TIME OF TEST: 01/01/2018 03:45:30

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.081	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.080	03:44	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 01/10/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HAMILTON-REYES,COURTNEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:26

SUBJECT'S LAST NAME: NOGUEIRA

DATE/TIME OF TEST: 01/06/2018 01:49:52

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.081	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.081	01:49	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 01/10/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALFONSO,
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RODRIGUEZ,ALFONSO
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: SANTOS

DATE/TIME OF TEST: 01/06/2018 03:13:24

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.082	03:08
Air Blank		0.000	03:08
Subject Sample	# 1	0.231	03:08
Breath Volume		2.132	03:08
Air Blank		0.000	03:09
Air Blank		0.000	03:11
Subject Sample	# 2	0.215	03:11
Breath Volume		2.296	03:11
Air Blank		0.000	03:12
Control Test		0.081	03:12
Air Blank		0.000	03:13
Diagnostics Check		OK	03:13

Data Download Date/Time: 01/10/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEMOS,MATHEW,T
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ZALENSKI,JOSEPH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:56

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 01/06/2018 17:25:53

Diagnostics Check		OK	17:17
Air Blank		0.000	17:18
Control Test		0.081	17:18
Air Blank		0.000	17:19
Subject Sample	# 1	0.120	17:19
Breath Volume		3.476	17:19
Air Blank		0.000	17:20
Air Blank		0.000	17:22
Subject Sample	# 2	0.120	17:24
Breath Volume		3.554	17:24
Air Blank		0.000	17:25
Control Test		0.081	17:25
Air Blank		0.000	17:25
Diagnostics Check		OK	17:25

Data Download Date/Time: 01/10/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000859

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 12/15/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:35

SUBJECT'S LAST NAME: CORRIE

DATE/TIME OF TEST: 01/07/2018 21:07:49

Diagnostics Check		OK	21:01
Air Blank		0.000	21:02
Control Test		0.082	21:02
Air Blank		0.000	21:02
Subject Sample	# 1	0.000	21:03
Breath Volume		2.535	21:03
Air Blank		0.000	21:04
Air Blank		0.000	21:06
Subject Sample	# 2	0.000	21:06
Breath Volume		2.875	21:06
Air Blank		0.000	21:06
Control Test		0.081	21:07
Air Blank		0.000	21:07
Diagnostics Check		OK	21:07

Data Download Date/Time: 01/10/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000861

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 12/29/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/27/2018 03:30:47

Diagnostics Check		OK	03:24
Air Blank		0.000	03:24
Control Test		0.082	03:24
Air Blank		0.000	03:25
Subject Sample	# 1	0.072	03:26
Breath Volume		2.843	03:26
Air Blank		0.000	03:26
Air Blank		0.000	03:28
Subject Sample	# 2	0.072	03:29
Breath Volume		1.535	03:29
Air Blank		0.000	03:29
Control Test		0.083	03:30
Air Blank		0.000	03:30
Diagnostics Check		OK	03:30

Data Download Date/Time: 01/31/2018 08:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUMMER,DAVID,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRUMMER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 01/01/2018 00:20:33

Diagnostics Check		OK	00:17	
Air Blank		0.000	00:17	
Control Test		0.079	00:17	
Air Blank		0.000	00:18	
Subject Sample	# 1	0.000	00:19	
Breath Volume		2.925	00:19	
Air Blank			00:19	RFI Detect
Air Blank		0.000	00:19	
Control Test		0.078	00:19	
Air Blank		0.000	00:20	
Diagnostics Check		OK	00:20	

Data Download Date/Time: 01/10/2018 13:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUMMER,ALCIDES,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRUMMER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:56

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 01/01/2018 00:24:38

Diagnostics Check	OK	00:23	
Air Blank		00:24	RFI Detect
Air Blank	0.000	00:24	

Data Download Date/Time: 01/10/2018 13:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUMMER,DAVID,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRUMMER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:36

SUBJECT'S LAST NAME: GUEVARA

DATE/TIME OF TEST: 01/01/2018 00:35:50

Diagnostics Check		OK	00:29
Air Blank		0.000	00:30
Control Test		0.080	00:30
Air Blank		0.000	00:30
Subject Sample	# 1	0.000	00:31
Breath Volume		2.480	00:31
Air Blank		0.000	00:31
Air Blank		0.000	00:33
Subject Sample	# 2	0.000	00:34
Breath Volume		1.941	00:34
Air Blank		0.000	00:34
Control Test		0.079	00:35
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 01/10/2018 13:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONATY,JAY,T
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 19:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:32

SUBJECT'S LAST NAME: DOWELL

DATE/TIME OF TEST: 01/02/2018 19:56:15

Diagnostics Check		OK	19:53	
Air Blank		0.000	19:53	
Control Test		0.078	19:54	
Air Blank		0.000	19:54	
Subject Sample	# 1		19:54	Subject Test Refused
Breath Volume		0.000	19:54	
Air Blank		0.000	19:55	
Control Test		0.079	19:55	
Air Blank		0.000	19:56	
Diagnostics Check		OK	19:56	

Data Download Date/Time: 01/10/2018 13:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONATY,JAY,T
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 12/14/2017

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:18
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:37

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 01/06/2018 23:04:05

Diagnostics Check		OK	23:00	
Air Blank		0.000	23:01	
Control Test		0.080	23:01	
Air Blank		0.000	23:02	
Subject Sample	# 1		23:02	Subject Test Refused
Breath Volume		0.000	23:02	
Air Blank		0.000	23:03	
Control Test		0.080	23:03	
Air Blank		0.000	23:04	
Diagnostics Check		OK	23:04	

Data Download Date/Time: 01/10/2018 13:21

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,ALEX,
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: O CONNELL,KEVIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 05:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:40

SUBJECT'S LAST NAME: SORENSON

DATE/TIME OF TEST: 01/12/2018 06:16:05

Diagnostics Check		OK	06:09
Air Blank		0.000	06:10
Control Test		0.079	06:10
Air Blank		0.000	06:10
Subject Sample	# 1	0.201	06:11
Breath Volume		1.699	06:11
Air Blank		0.000	06:12
Air Blank		0.000	06:13
Subject Sample	# 2	0.195	06:14
Breath Volume		2.207	06:14
Air Blank		0.000	06:15
Control Test		0.078	06:15
Air Blank		0.000	06:16
Diagnostics Check		OK	06:16

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUMMER,DAVID,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRUMMER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 17:13
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:14

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 01/12/2018 18:48:32

Diagnostics Check		OK	18:34	
Air Blank		0.000	18:34	
Control Test		0.079	18:35	
Air Blank		0.000	18:35	
Subject Sample	# 1	0.203	18:38	
Breath Volume		1.160	18:38	
Air Blank		0.000	18:38	
Air Blank		0.000	18:40	
Subject Sample	# 2	0.196	18:43	Volume Not Met
Breath Volume		1.039	18:43	
Air Blank		0.000	18:44	
Air Blank		0.000	18:46	
Subject Sample	# 3	0.211	18:47	
Breath Volume		2.285	18:47	
Air Blank		0.000	18:47	
Control Test		0.077	18:47	
Air Blank		0.000	18:48	
Diagnostics Check		OK	18:48	

Data Download Date/Time: 02/01/2018 11:15

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: BLYTH,KYLE
ARRESTING OFFICER AGENCY: FHP Troop E
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 12:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: CUPIT

DATE/TIME OF TEST: 01/13/2018 02:05:07

Diagnostics Check		OK	01:57
Air Blank		0.000	01:58
Control Test		0.078	01:58
Air Blank		0.000	01:59
Subject Sample	# 1	0.126	02:00
Breath Volume		2.828	02:00
Air Blank		0.000	02:01
Air Blank		0.000	02:03
Subject Sample	# 2	0.129	02:03
Breath Volume		1.785	02:03
Air Blank		0.000	02:04
Control Test		0.077	02:04
Air Blank		0.000	02:05
Diagnostics Check		OK	02:05

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:39

SUBJECT'S LAST NAME: HOLMACK

DATE/TIME OF TEST: 01/14/2018 02:12:38

Diagnostics Check		OK	02:06
Air Blank		0.000	02:06
Control Test		0.079	02:07
Air Blank		0.000	02:07
Subject Sample	# 1	0.138	02:08
Breath Volume		2.519	02:08
Air Blank		0.000	02:08
Air Blank		0.000	02:10
Subject Sample	# 2	0.136	02:11
Breath Volume		2.445	02:11
Air Blank		0.000	02:11
Control Test		0.078	02:12
Air Blank		0.000	02:12
Diagnostics Check		OK	02:12

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ELSMORE,JUSTIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: AMENT

DATE/TIME OF TEST: 01/17/2018 01:35:38

Diagnostics Check		OK	01:31	
Air Blank		0.000	01:32	
Control Test		0.080	01:32	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:57
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:33

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 01/19/2018 00:04:09

Diagnostics Check		OK	00:01	
Air Blank		0.000	00:01	
Control Test		0.079	00:02	
Air Blank		0.000	00:02	
Subject Sample	# 1		00:02	Subject Test Refused
Breath Volume		0.000	00:02	
Air Blank		0.000	00:03	
Control Test		0.079	00:03	
Air Blank		0.000	00:04	
Diagnostics Check		OK	00:04	

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ELSMORE,JUSTIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/21/2018 02:15:28

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.080	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:14	Subject Test Refused
Breath Volume		0.000	02:14	
Air Blank		0.000	02:14	
Control Test		0.081	02:14	
Air Blank		0.000	02:15	
Diagnostics Check		OK	02:15	

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ELSMORE,JUSTIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/21/2018 02:19:14

Diagnostics Check	OK	02:18	
Air Blank	0.000	02:18	
		02:18	RFI Detect
Air Blank	0.000	02:19	

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ELSMORE,JUSTIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: BAILEY

DATE/TIME OF TEST: 01/21/2018 02:27:44

Diagnostics Check		OK	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Subject Sample	# 1	0.149	02:23
Breath Volume		1.230	02:23
Air Blank		0.000	02:23
Air Blank		0.000	02:25
Subject Sample	# 2	0.160	02:26
Breath Volume		1.757	02:26
Air Blank		0.000	02:26
Control Test		0.079	02:27
Air Blank		0.000	02:27
Diagnostics Check		OK	02:27

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLYTH,KYLE,A
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: DERY,D
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 22:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:09

SUBJECT'S LAST NAME: CASTRO

DATE/TIME OF TEST: 01/21/2018 23:39:21

Diagnostics Check		OK	23:32	
Air Blank		0.000	23:32	
Control Test		0.079	23:33	
Air Blank		0.000	23:33	
Subject Sample	# 1	0.246	23:34	
Breath Volume		2.312	23:34	
Air Blank		0.000	23:35	
Air Blank		0.000	23:37	
Subject Sample	# 2		23:37	Subject Test Refused
Breath Volume		0.000	23:37	
Air Blank		0.000	23:38	
Control Test		0.076	23:38	
Air Blank		0.000	23:39	
Diagnostics Check		OK	23:39	

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ELSMORE,JUSTIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 21:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: SEGAR JR

DATE/TIME OF TEST: 01/24/2018 21:48:20

Diagnostics Check		OK	21:39	
Air Blank		0.000	21:40	
Control Test		0.079	21:40	
Air Blank		0.000	21:40	
Subject Sample	# 1		21:44	No Sample Provided
Breath Volume		0.000	21:44	
Air Blank		0.000	21:44	
Air Blank		0.000	21:46	
Subject Sample	# 2		21:46	Subject Test Refused
Breath Volume		0.000	21:46	
Air Blank		0.000	21:47	
Control Test		0.080	21:47	
Air Blank		0.000	21:48	
Diagnostics Check		OK	21:48	

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONATY,JAY,T
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: CONATY,JAY
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:49

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 01/26/2018 00:16:50

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.079	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.298	00:12
Breath Volume		3.480	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.287	00:15
Breath Volume		3.101	00:15
Air Blank		0.000	00:15
Control Test		0.075	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:16

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELSMORE,JUSTIN,R
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: ELSMORE,JUSTIN
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: ROBLES

DATE/TIME OF TEST: 01/26/2018 01:29:28

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.163	01:24
Breath Volume		1.132	01:24
Air Blank		0.000	01:25
Air Blank		0.000	01:26
Subject Sample	# 2	0.166	01:27
Breath Volume		1.652	01:27
Air Blank		0.000	01:28
Control Test		0.078	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRUMMER,DAVID,J
BREATH TEST OPERATOR AGENCY: Monroe County SO
ARRESTING OFFICER: BRUMMER,DAVID
ARRESTING OFFICER AGENCY: Monroe County SO
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 20:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:48

SUBJECT'S LAST NAME: LOISEAU

DATE/TIME OF TEST: 01/27/2018 22:17:14

Diagnostics Check		OK	22:09
Air Blank		0.000	22:09
Control Test		0.079	22:09
Air Blank		0.000	22:10
Subject Sample	# 1	0.126	22:12
Breath Volume		1.386	22:12
Air Blank		0.000	22:12
Air Blank		0.000	22:14
Subject Sample	# 2	0.125	22:15
Breath Volume		1.636	22:15
Air Blank		0.000	22:16
Control Test		0.078	22:16
Air Blank		0.000	22:17
Diagnostics Check		OK	22:17

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JESSE,A
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: YOUNG,JESSE
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 04:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:03

SUBJECT'S LAST NAME: BAROTOV

DATE/TIME OF TEST: 01/28/2018 05:32:31

Diagnostics Check		OK	05:26
Air Blank		0.000	05:26
Control Test		0.078	05:27
Air Blank		0.000	05:27
Subject Sample	# 1	0.151	05:28
Breath Volume		3.085	05:28
Air Blank		0.000	05:28
Air Blank		0.000	05:30
Subject Sample	# 2	0.149	05:31
Breath Volume		2.898	05:31
Air Blank		0.000	05:31
Control Test		0.077	05:31
Air Blank		0.000	05:32
Diagnostics Check		OK	05:32

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-000867

INSTRUMENT REGISTERED TO: MONROE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CONATY,JAY,T
BREATH TEST OPERATOR AGENCY: Key West PD
ARRESTING OFFICER: MOGERLEY,NEIL
ARRESTING OFFICER AGENCY: Key West PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 686331
DRY GAS EXPIRY DATE: 01/31/2020
VIOLATION CODE: DUI
ARREST TIME: 23:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: SCHNEIDER

DATE/TIME OF TEST: 01/30/2018 00:47:54

Diagnostics Check		OK	00:40
Air Blank		0.000	00:40
Control Test		0.080	00:41
Air Blank		0.000	00:41
Subject Sample	# 1	0.135	00:42
Breath Volume		2.187	00:42
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.131	00:46
Breath Volume		1.972	00:46
Air Blank		0.000	00:46
Control Test		0.078	00:47
Air Blank		0.000	00:47
Diagnostics Check		OK	00:47

Data Download Date/Time: 02/01/2018 11:15

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.