Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: CLOSIUS,RICHARD
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 01/04/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:24

SUBJECT'S LAST NAME: TORRES

DATE/TIME OF TEST: 02/01/2018 05:54:09

Diagnostics Check		OK	05:46
Air Blank		0.000	05:46
Control Test		0.080	05:47
Air Blank		0.000	05:47
Subject Sample	# 1	0.127	05:49
Breath Volume		2.343	05:49
Air Blank		0.000	05:50
Air Blank		0.000	05:51
Subject Sample	# 2	0.132	05:52
Breath Volume		2.480	05:52
Air Blank		0.000	05:53
Control Test		0.079	05:53
Air Blank		0.000	05:54
Diagnostics Check		OK	05:54

Data Download Date/Time: 02/01/2018 13:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MCGRATH,ROBERT,E
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: MCGRATH,MCGRATH
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/09/2018 17:34:57

Diagnostics Check		OK	17:28
Air Blank		0.000	17:29
Control Test		0.081	17:29
Air Blank		0.000	17:30
Subject Sample	# 1	0.000	17:30
Breath Volume		3.031	17:30
Air Blank		0.000	17:31
Air Blank		0.000	17:33
Subject Sample	# 2	0.000	17:33
Breath Volume		3.070	17:33
Air Blank		0.000	17:34
Control Test		0.080	17:34
Air Blank		0.000	17:34
Diagnostics Check		OK	17:34

Data Download Date/Time: 03/09/2018 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NADAL,JOSE,V
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: NADAL,JOSE
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:15

SUBJECT'S LAST NAME: ABBASI

DATE/TIME OF TEST: 02/10/2018 05:42:48

Diagnostics Check		OK	05:39	
Air Blank		0.000	05:39	
Control Test		0.080	05:39	
Air Blank		0.000	05:40	
Subject Sample	# 1		05:41	Subject Test Refused
Breath Volume		0.000	05:41	
Air Blank		0.000	05:41	
Control Test		0.081	05:42	
Air Blank		0.000	05:42	
Diagnostics Check		OK	05:42	

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: LOBO

DATE/TIME OF TEST: 02/16/2018 04:44:17

Diagnostics Check		OK	04:37
Air Blank		0.000	04:37
Control Test		0.080	04:38
Air Blank		0.000	04:38
Subject Sample	# 1	0.216	04:39
Breath Volume		1.519	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:41
Subject Sample	# 2	0.216	04:42
Breath Volume		2.449	04:42
Air Blank		0.000	04:43
Control Test		0.080	04:43
Air Blank		0.000	04:44
Diagnostics Check		OK	04:44

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KINSEY-SMITH,FELICIA,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: KINSEY-SMITH,FELICIA
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/17/2018 04:21:32

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.080	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.067	04:16
Breath Volume		2.531	04:16
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.070	04:19
Breath Volume		1.898	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FRANCO,GREGORY,R
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: PALACIOS

DATE/TIME OF TEST: 02/17/2018 04:33:48

Diagnostics Check		OK	04:26
Air Blank		0.000	04:26
Control Test		0.080	04:26
Air Blank		0.000	04:27
Subject Sample	# 1	0.154	04:28
Breath Volume		2.417	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:30
Subject Sample	# 2	0.145	04:32
Breath Volume		1.636	04:32
Air Blank		0.000	04:32
Control Test		0.079	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CLOSIUS,RICHARD,L
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: REYNOLDS,JAYSON
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:12

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/17/2018 05:35:38

Diagnostics Check		OK	05:32	
Air Blank		0.000	05:33	
Control Test		0.080	05:33	
Air Blank		0.000	05:34	
Subject Sample	# 1		05:34	Subject Test Refused
Breath Volume		0.000	05:34	
Air Blank		0.000	05:34	
Control Test		0.080	05:35	
Air Blank		0.000	05:35	
Diagnostics Check		OK	05:35	

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: BORREGO

DATE/TIME OF TEST: 02/17/2018 07:53:23

Diagnostics Check		OK	07:42	
Air Blank		0.000	07:42	
Control Test		0.080	07:42	
Air Blank		0.000	07:43	
Subject Sample	# 1	0.000	07:46	Volume Not Met
Breath Volume		0.019	07:46	
Air Blank		0.000	07:47	
Air Blank		0.000	07:49	
Subject Sample	# 2	0.064	07:52	Volume Not Met
Breath Volume		0.000	07:52	
Air Blank			07:52	Purge Fail
Air Blank		0.000	07:53	

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: BORREGO

DATE/TIME OF TEST: 02/17/2018 08:00:15

Diagnostics Check	OK	07:58	
Air Blank	0.000	07:59	
Control Test	0.074	07:59	Control Outside Tolerance
Air Blank	0.000	08:00	

Data Download Date/Time: 03/09/2018 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HINOJOSA,DANIEL,M
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: HINOJOSA,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 06:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 07:20

SUBJECT'S LAST NAME: BORREGO

DATE/TIME OF TEST: 02/17/2018 08:18:32

Diagnostics Check		OK	08:15	
Air Blank		0.000	08:16	
Control Test		0.077	08:16	
Air Blank		0.000	08:16	
Subject Sample	# 1		08:17	Subject Test Refused
Breath Volume		0.000	08:17	
Air Blank		0.000	08:17	
Control Test		0.079	08:17	
Air Blank		0.000	08:18	
Diagnostics Check		OK	08:18	

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GARCIA,EMILIE,M
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: GARCIA,EMILIE
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:43

SUBJECT'S LAST NAME: BORGE

DATE/TIME OF TEST: 02/17/2018 19:20:10

Diagnostics Check		OK	19:10
Air Blank		0.000	19:10
Control Test		0.080	19:10
Air Blank		0.000	19:11
Subject Sample	# 1	0.198	19:12
Breath Volume		2.093	19:12
Air Blank		0.000	19:13
Air Blank		0.000	19:14
Subject Sample	# 2	0.168	19:15
Breath Volume		2.335	19:15
Air Blank		0.000	19:16
Air Blank		0.000	19:17
Subject Sample	# 3	0.183	19:18
Breath Volume		2.066	19:18
Air Blank		0.000	19:19
Control Test		0.077	19:19
Air Blank		0.000	19:20
Diagnostics Check		OK	19:20

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVALOS,DANIEL,A
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: DAVALOS,DANIEL
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:25

SUBJECT'S LAST NAME: SILLERICO

DATE/TIME OF TEST: 02/19/2018 04:01:57

Diagnostics Check		OK	03:53
Air Blank		0.000	03:53
Control Test		0.080	03:54
Air Blank		0.000	03:54
Subject Sample	# 1	0.216	03:55
Breath Volume		1.589	03:55
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.202	04:00
Breath Volume		1.226	04:00
Air Blank		0.000	04:01
Control Test		0.077	04:01
Air Blank		0.000	04:01
Diagnostics Check		OK	04:01

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MYRTIL,REGINALD,J
BREATH TEST OPERATOR AGENCY: Miami-Dade PD
ARRESTING OFFICER: FRANCO,GREGORY
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:38

SUBJECT'S LAST NAME: MORENO

DATE/TIME OF TEST: 02/23/2018 03:07:36

Diagnostics Check		OK	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:00
Air Blank		0.000	03:01
Subject Sample	# 1	0.182	03:02
Breath Volume		1.332	03:02
Air Blank		0.000	03:03
Air Blank		0.000	03:05
Subject Sample	# 2	0.200	03:05
Breath Volume		1.683	03:05
Air Blank		0.000	03:06
Control Test		0.078	03:06
Air Blank		0.000	03:07
Diagnostics Check		OK	03:07

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LORENZO,YOSNIEL,
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: LORENZO,YOSNIEL
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:31

SUBJECT'S LAST NAME: VALDIVIA

DATE/TIME OF TEST: 02/24/2018 02:00:45

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:52	
Control Test		0.080	01:52	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:56	No Sample Provided
Breath Volume		0.000	01:56	
Air Blank		0.000	01:56	
Air Blank		0.000	01:58	
Subject Sample	# 2		01:59	Subject Test Refused
Breath Volume		0.000	01:59	
Air Blank		0.000	01:59	
Control Test		0.080	02:00	
Air Blank		0.000	02:00	
Diagnostics Check		OK	02:00	

Data Download Date/Time: 03/09/2018 11:02

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARDELAS,MAYLEE,
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: ROBERTS,TREVOR
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:58
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: OGLIASTRI-BARRIOS

DATE/TIME OF TEST: 02/26/2018 02:38:57

Diagnostics Check		OK	02:34	
Air Blank		0.000	02:35	
Control Test		0.080	02:35	
Air Blank		0.000	02:36	
Subject Sample	# 1		02:37	Subject Test Refused
Breath Volume		0.000	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:38	
Diagnostics Check		OK	02:38	

Data Download Date/Time: 03/09/2018 11:02

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHWARTZ,RODNEY,K
BREATH TEST OPERATOR AGENCY: Sweetwater PD
ARRESTING OFFICER: SCHWARTZ,RODNEY
ARRESTING OFFICER AGENCY: Sweetwater PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 16:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: CISNEROS

DATE/TIME OF TEST: 02/26/2018 18:03:22

Diagnostics Check		OK	18:00	
Air Blank		0.000	18:00	
Control Test		0.080	18:01	
Air Blank		0.000	18:01	
Subject Sample	# 1		18:01	Subject Test Refused
Breath Volume		0.000	18:01	
Air Blank		0.000	18:02	
Control Test		0.080	18:02	
Air Blank		0.000	18:03	
Diagnostics Check		OK	18:03	

Data Download Date/Time: 03/09/2018 11:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001277

INSTRUMENT REGISTERED TO: MIAMI-DADE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PEREZ,ROBERTO,J
BREATH TEST OPERATOR AGENCY: FHP Troop E
ARRESTING OFFICER: HOWARD,A
ARRESTING OFFICER AGENCY: Miami-Dade PD
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 20016080A2
DRY GAS EXPIRY DATE: 09/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: DA COSTA FREIRE

DATE/TIME OF TEST: 02/27/2018 05:23:37

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.080	05:17
Air Blank		0.000	05:18
Subject Sample	# 1	0.237	05:18
Breath Volume		2.066	05:18
Air Blank		0.000	05:19
Air Blank		0.000	05:21
Subject Sample	# 2	0.221	05:21
Breath Volume		1.839	05:21
Air Blank		0.000	05:22
Control Test		0.079	05:22
Air Blank		0.000	05:23
Diagnostics Check		OK	05:23

Data Download Date/Time: 03/09/2018 11:03

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:40

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/04/2018 19:07:08

Diagnostics Check		OK	19:00
Air Blank		0.000	19:01
Control Test		0.080	19:01
Air Blank		0.000	19:02
Subject Sample	# 1	0.043	19:02
Breath Volume		1.207	19:02
Air Blank		0.000	19:03
Air Blank		0.000	19:04
Subject Sample	# 2	0.043	19:05
Breath Volume		1.445	19:05
Air Blank		0.000	19:06
Control Test		0.080	19:06
Air Blank		0.000	19:07
Diagnostics Check		OK	19:07

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,F
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:08

SUBJECT'S LAST NAME: SOLORZANO

DATE/TIME OF TEST: 02/04/2018 20:32:04

Diagnostics Check		OK	20:29	
Air Blank		0.000	20:29	
Control Test		0.080	20:29	
Air Blank		0.000	20:30	
Subject Sample	# 1		20:30	Subject Test Refused
Breath Volume		0.000	20:30	
Air Blank		0.000	20:31	
Control Test		0.080	20:31	
Air Blank		0.000	20:31	
Diagnostics Check		OK	20:32	

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:18
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:44

SUBJECT'S LAST NAME: TANNER

DATE/TIME OF TEST: 02/04/2018 22:11:06

Diagnostics Check		OK	22:04
Air Blank		0.000	22:05
Control Test		0.080	22:05
Air Blank		0.000	22:06
Subject Sample	# 1	0.000	22:06
Breath Volume		1.679	22:06
Air Blank		0.000	22:07
Air Blank		0.000	22:09
Subject Sample	# 2	0.000	22:09
Breath Volume		1.386	22:09
Air Blank		0.000	22:10
Control Test		0.080	22:10
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOROZOV,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: ADMIN
ARREST TIME: 04:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:30

SUBJECT'S LAST NAME: ARMSTRONG

DATE/TIME OF TEST: 02/04/2018 22:57:59

Diagnostics Check		OK	22:51
Air Blank		0.000	22:51
Control Test		0.080	22:51
Air Blank		0.000	22:52
Subject Sample	# 1	0.000	22:52
Breath Volume		1.242	22:52
Air Blank		0.000	22:53
Air Blank		0.000	22:55
Subject Sample	# 2	0.000	22:56
Breath Volume		1.378	22:56
Air Blank		0.000	22:57
Control Test		0.081	22:57
Air Blank		0.000	22:57
Diagnostics Check		OK	22:57

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TAYLOR,TIMOTHY,P
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BROWN,T.A
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:47

SUBJECT'S LAST NAME: WILSON

DATE/TIME OF TEST: 02/09/2018 22:13:55

Diagnostics Check		OK	22:07
Air Blank		0.000	22:08
Control Test		0.080	22:08
Air Blank		0.000	22:08
Subject Sample	# 1	0.279	22:09
Breath Volume		1.511	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:11
Subject Sample	# 2	0.278	22:12
Breath Volume		2.433	22:12
Air Blank		0.000	22:13
Control Test		0.080	22:13
Air Blank		0.000	22:13
Diagnostics Check		OK	22:13

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,A,B
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:10

SUBJECT'S LAST NAME: THARPE

DATE/TIME OF TEST: 02/10/2018 19:37:16

Diagnostics Check		OK	19:30
Air Blank		0.000	19:31
Control Test		0.079	19:31
Air Blank		0.000	19:31
Subject Sample	# 1	0.128	19:32
Breath Volume		1.511	19:32
Air Blank		0.000	19:33
Air Blank		0.000	19:35
Subject Sample	# 2	0.131	19:35
Breath Volume		1.503	19:35
Air Blank		0.000	19:36
Control Test		0.078	19:36
Air Blank		0.000	19:37
Diagnostics Check		OK	19:37

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRIS,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: JIMENEZ,C
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: HOLDEN

DATE/TIME OF TEST: 02/13/2018 17:16:38

Diagnostics Check		OK	17:10
Air Blank		0.000	17:10
Control Test		0.081	17:11
Air Blank		0.000	17:11
Subject Sample	# 1	0.247	17:12
Breath Volume		3.757	17:12
Air Blank		0.000	17:12
Air Blank		0.000	17:14
Subject Sample	# 2	0.247	17:15
Breath Volume		2.898	17:15
Air Blank		0.000	17:15
Control Test		0.080	17:16
Air Blank		0.000	17:16
Diagnostics Check		OK	17:16

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTELL,CHRIS,D
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/01/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 17:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:31

SUBJECT'S LAST NAME: GUSENKOV

DATE/TIME OF TEST: 02/13/2018 18:58:19

Diagnostics Check		OK	18:51
Air Blank		0.000	18:52
Control Test		0.080	18:52
Air Blank		0.000	18:53
Subject Sample	# 1	0.334	18:53
Breath Volume		3.242	18:53
Air Blank		0.000	18:54
Air Blank		0.000	18:56
Subject Sample	# 2	0.338	18:56
Breath Volume		3.828	18:56
Air Blank		0.000	18:57
Control Test		0.081	18:57
Air Blank		0.000	18:58
Diagnostics Check		OK	18:58

Data Download Date/Time: 02/15/2018 13:20

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PFISTER,G
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 14:30
GENDER: Female
OBSERVATION PERIOD BEGAN: 16:15

SUBJECT'S LAST NAME: KEASLER

DATE/TIME OF TEST: 02/15/2018 16:40:30

Diagnostics Check		OK	16:37	
Air Blank		0.000	16:37	
Control Test		0.078	16:38	
Air Blank		0.000	16:38	
Subject Sample	# 1		16:39	Subject Test Refused
Breath Volume		0.000	16:39	
Air Blank		0.000	16:39	
Control Test		0.079	16:39	
Air Blank		0.000	16:40	
Diagnostics Check		OK	16:40	

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: PARKER

DATE/TIME OF TEST: 02/17/2018 02:26:50

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:21
Air Blank		0.000	02:21
Subject Sample	# 1	0.000	02:22
Breath Volume		1.449	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.000	02:25
Breath Volume		1.746	02:25
Air Blank		0.000	02:25
Control Test		0.079	02:26
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: STANLEY,T
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 02/17/2018 03:57:16

Diagnostics Check		OK	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Subject Sample	# 1	0.010	03:52
Breath Volume		2.023	03:52
Air Blank		0.000	03:53
Air Blank		0.000	03:55
Subject Sample	# 2	0.009	03:55
Breath Volume		2.007	03:55
Air Blank		0.000	03:56
Control Test		0.079	03:56
Air Blank		0.000	03:57
Diagnostics Check		OK	03:57

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SWINTON,L
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:54

SUBJECT'S LAST NAME: KOONTZ

DATE/TIME OF TEST: 02/17/2018 19:37:38

Diagnostics Check		OK	19:26	
Air Blank		0.000	19:26	
Control Test		0.080	19:27	
Air Blank		0.000	19:27	
Subject Sample	# 1		19:30	No Sample Provided
Breath Volume		0.000	19:30	
Air Blank		0.000	19:31	
Air Blank		0.000	19:33	
Subject Sample	# 2		19:36	No Sample Provided
Breath Volume		0.000	19:36	
Air Blank		0.000	19:36	
Control Test		0.080	19:37	
Air Blank		0.000	19:37	
Diagnostics Check		OK	19:37	

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MARRERO,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:14

SUBJECT'S LAST NAME: EMNAS

DATE/TIME OF TEST: 02/17/2018 21:41:41

Diagnostics Check		OK	21:34
Air Blank		0.000	21:35
Control Test		0.080	21:35
Air Blank		0.000	21:36
Subject Sample	# 1	0.039	21:37
Breath Volume		1.636	21:37
Air Blank		0.000	21:37
Air Blank		0.000	21:39
Subject Sample	# 2	0.039	21:40
Breath Volume		1.500	21:40
Air Blank		0.000	21:40
Control Test		0.080	21:41
Air Blank		0.000	21:41
Diagnostics Check		OK	21:41

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BRYANT,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 02/18/2018 03:45:35

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:42	
Control Test		0.079	03:42	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:44	Subject Test Refused
Breath Volume		0.000	03:44	
Air Blank		0.000	03:44	
Control Test		0.080	03:45	
Air Blank		0.000	03:45	
Diagnostics Check		OK	03:45	

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PEPPERS,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:47

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/18/2018 05:11:25

Diagnostics Check		OK	05:08	
Air Blank		0.000	05:08	
Control Test		0.080	05:09	
Air Blank		0.000	05:09	
Subject Sample	# 1		05:10	Subject Test Refused
Breath Volume		0.000	05:10	
Air Blank		0.000	05:10	
Control Test		0.080	05:10	
Air Blank		0.000	05:11	
Diagnostics Check		OK	05:11	

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: SENTERFITT,B
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 04:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:47

SUBJECT'S LAST NAME: PINERA

DATE/TIME OF TEST: 02/18/2018 06:15:29

Diagnostics Check		OK	06:07
Air Blank		0.000	06:07
Control Test		0.080	06:08
Air Blank		0.000	06:08
Subject Sample	# 1	0.112	06:10
Breath Volume		2.105	06:10
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.108	06:14
Breath Volume		2.144	06:14
Air Blank		0.000	06:14
Control Test		0.080	06:14
Air Blank		0.000	06:15
Diagnostics Check		OK	06:15

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BRANTON

DATE/TIME OF TEST: 02/19/2018 01:12:48

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:01	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:05	No Sample Provided
Breath Volume		0.000	01:05	
Air Blank		0.000	01:06	
Air Blank		0.000	01:08	
Subject Sample	# 2	0.166	01:11	Volume Not Met
Breath Volume		0.277	01:11	
Air Blank		0.000	01:11	
Control Test		0.081	01:12	
Air Blank		0.000	01:12	
Diagnostics Check		OK	01:12	

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: TORRES,W
ARRESTING OFFICER AGENCY: Neptune Beach PD
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:40

SUBJECT'S LAST NAME: BRANTON

DATE/TIME OF TEST: 02/19/2018 01:20:17

Diagnostics Check		OK	01:16	
Air Blank		0.000	01:17	
Control Test		0.080	01:17	
Air Blank		0.000	01:17	
Subject Sample	# 1		01:18	Subject Test Refused
Breath Volume		0.730	01:18	
Air Blank		0.000	01:19	
Control Test		0.080	01:19	
Air Blank		0.000	01:20	
Diagnostics Check		OK	01:20	

Data Download Date/Time: 02/28/2018 14:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: HAWKINS,A
ARRESTING OFFICER AGENCY: Jacksonville Beach
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:05

SUBJECT'S LAST NAME: ROSS

DATE/TIME OF TEST: 02/19/2018 02:32:45

Diagnostics Check		OK	02:26
Air Blank		0.000	02:26
Control Test		0.080	02:27
Air Blank		0.000	02:27
Subject Sample	# 1	0.228	02:28
Breath Volume		1.941	02:28
Air Blank		0.000	02:28
Air Blank		0.000	02:30
Subject Sample	# 2	0.220	02:31
Breath Volume		1.476	02:31
Air Blank		0.000	02:31
Control Test		0.080	02:32
Air Blank		0.000	02:32
Diagnostics Check		OK	02:32

Data Download Date/Time: 02/28/2018 14:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: BYRNE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:18

SUBJECT'S LAST NAME: HOWARD

DATE/TIME OF TEST: 02/19/2018 22:45:08

Diagnostics Check		OK	22:38
Air Blank		0.000	22:38
Control Test		0.080	22:39
Air Blank		0.000	22:39
Subject Sample	# 1	0.194	22:40
Breath Volume		1.187	22:40
Air Blank		0.000	22:41
Air Blank		0.000	22:42
Subject Sample	# 2	0.213	22:43
Breath Volume		2.562	22:43
Air Blank		0.000	22:44
Control Test		0.079	22:44
Air Blank		0.000	22:45
Diagnostics Check		OK	22:45

Data Download Date/Time: 02/28/2018 14:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: RUSSELL,C
ARRESTING OFFICER AGENCY: Atlantic Beach PD
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 13:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 15:49

SUBJECT'S LAST NAME: DELAUDER

DATE/TIME OF TEST: 02/23/2018 16:17:27

Diagnostics Check		OK	16:10
Air Blank		0.000	16:11
Control Test		0.080	16:11
Air Blank		0.000	16:12
Subject Sample	# 1	0.248	16:12
Breath Volume		2.121	16:12
Air Blank		0.000	16:13
Air Blank		0.000	16:15
Subject Sample	# 2	0.251	16:15
Breath Volume		1.832	16:15
Air Blank		0.000	16:16
Control Test		0.079	16:16
Air Blank		0.000	16:17
Diagnostics Check		OK	16:17

Data Download Date/Time: 02/28/2018 14:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GONZALEZ,J,
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: FISETTE,J
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:15

SUBJECT'S LAST NAME: WAHL

DATE/TIME OF TEST: 02/23/2018 22:43:05

Diagnostics Check		OK	22:39	
Air Blank		0.000	22:40	
Control Test		0.080	22:40	
Air Blank		0.000	22:41	
Subject Sample	# 1		22:41	Subject Test Refused
Breath Volume		0.000	22:41	
Air Blank		0.000	22:42	
Control Test		0.080	22:42	
Air Blank		0.000	22:43	
Diagnostics Check		OK	22:43	

Data Download Date/Time: 02/28/2018 14:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: CHRISTMAS,N
ARRESTING OFFICER AGENCY: FFWCC (Ocala)
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:35

SUBJECT'S LAST NAME: KOPPOLD

DATE/TIME OF TEST: 02/24/2018 17:58:52

Diagnostics Check		OK	17:56	
Air Blank		0.000	17:56	
Control Test		0.080	17:56	
Air Blank		0.000	17:57	
Subject Sample	# 1		17:57	Subject Test Refused
Breath Volume		0.000	17:57	
Air Blank		0.000	17:57	
Control Test		0.080	17:58	
Air Blank		0.000	17:58	
Diagnostics Check		OK	17:58	

Data Download Date/Time: 02/28/2018 14:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ADAMS,J,T
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: MOELLER,R
ARRESTING OFFICER AGENCY: Jacksonville SO
LAST AGENCY INSPECTION: 02/15/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:11

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/24/2018 22:38:05

Diagnostics Check		OK	22:31
Air Blank		0.000	22:31
Control Test		0.080	22:32
Air Blank		0.000	22:32
Subject Sample	# 1	0.149	22:33
Breath Volume		1.328	22:33
Air Blank		0.000	22:34
Air Blank		0.000	22:36
Subject Sample	# 2	0.150	22:36
Breath Volume		1.562	22:36
Air Blank		0.000	22:37
Control Test		0.080	22:37
Air Blank		0.000	22:38
Diagnostics Check		OK	22:38

Data Download Date/Time: 02/28/2018 14:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001279

INSTRUMENT REGISTERED TO: JACKSONVILLE SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MURRIELL,E,L
BREATH TEST OPERATOR AGENCY: Jacksonville SO
ARRESTING OFFICER: PATTERSON,K
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/28/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:16

SUBJECT'S LAST NAME: HOLT

DATE/TIME OF TEST: 02/28/2018 22:43:43

Diagnostics Check		OK	22:36
Air Blank		0.000	22:37
Control Test		0.079	22:37
Air Blank		0.000	22:38
Subject Sample	# 1	0.166	22:39
Breath Volume		2.292	22:39
Air Blank		0.000	22:39
Air Blank		0.000	22:41
Subject Sample	# 2	0.164	22:42
Breath Volume		2.113	22:42
Air Blank		0.000	22:42
Control Test		0.078	22:43
Air Blank		0.000	22:43
Diagnostics Check		OK	22:43

Data Download Date/Time: 03/21/2018 09:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001282

INSTRUMENT REGISTERED TO: UNIVERSITY OF FL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZARATE,JESSICA,L
BREATH TEST OPERATOR AGENCY: U of Fl PD
ARRESTING OFFICER: ACOSTA,ALEX
ARRESTING OFFICER AGENCY: Alachua County SO
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 22817080A4
DRY GAS EXPIRY DATE: 10/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:04

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/26/2018 16:32:35

Diagnostics Check		OK	16:29	
Air Blank		0.000	16:30	
Control Test		0.076	16:30	
Air Blank		0.000	16:30	
Subject Sample	# 1		16:31	Subject Test Refused
Breath Volume		0.000	16:31	
Air Blank		0.000	16:31	
Control Test		0.077	16:31	
Air Blank		0.000	16:32	
Diagnostics Check		OK	16:32	

Data Download Date/Time: 03/30/2018 10:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JEFFERSON,E
BREATH TEST OPERATOR AGENCY: Gilchrist County SO
ARRESTING OFFICER: WHITAKER,ROBERT
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: ALVARADO-VELASQUEZ

DATE/TIME OF TEST: 02/16/2018 21:09:22

Diagnostics Check	OK	21:07	
Air Blank	0.000	21:08	
Control Test	0.079	21:08	
Air Blank		21:08	RFI Detect
Air Blank	0.000	21:09	

Data Download Date/Time: 02/25/2018 23:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001285

INSTRUMENT REGISTERED TO: GILCHRIST COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,JEFFERSON,E
BREATH TEST OPERATOR AGENCY: Gilchrist County SO
ARRESTING OFFICER: WHITAKER,ROBERT
ARRESTING OFFICER AGENCY: Gilchrist County SO
LAST AGENCY INSPECTION: 01/22/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 259585
DRY GAS EXPIRY DATE: 01/12/2019
VIOLATION CODE: DUI
ARREST TIME: 20:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:44

SUBJECT'S LAST NAME: ALVARADO-VELASQUEZ

DATE/TIME OF TEST: 02/16/2018 21:17:31

Diagnostics Check		OK	21:11
Air Blank		0.000	21:12
Control Test		0.079	21:12
Air Blank		0.000	21:12
Subject Sample	# 1	0.148	21:13
Breath Volume		1.195	21:13
Air Blank		0.000	21:13
Air Blank		0.000	21:15
Subject Sample	# 2	0.155	21:16
Breath Volume		1.765	21:16
Air Blank		0.000	21:16
Control Test		0.079	21:16
Air Blank		0.000	21:17
Diagnostics Check		OK	21:17

Data Download Date/Time: 02/25/2018 23:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HILLIARD,KARI,A
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: HILLIARD,KARI
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 15:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:00

SUBJECT'S LAST NAME: PORTERFIELD

DATE/TIME OF TEST: 02/15/2018 16:27:29

Diagnostics Check		OK	16:21
Air Blank		0.000	16:21
Control Test		0.082	16:22
Air Blank		0.000	16:22
Subject Sample	# 1	0.000	16:23
Breath Volume		1.738	16:23
Air Blank		0.000	16:23
Air Blank		0.000	16:25
Subject Sample	# 2	0.000	16:26
Breath Volume		1.621	16:26
Air Blank		0.000	16:26
Control Test		0.082	16:26
Air Blank		0.000	16:27
Diagnostics Check		OK	16:27

Data Download Date/Time: 02/27/2018 21:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001287

INSTRUMENT REGISTERED TO: BAKER COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FARLEY,JOSEPH,E
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: FARLEY,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:06

SUBJECT'S LAST NAME: KNEEN

DATE/TIME OF TEST: 02/24/2018 04:33:29

Diagnostics Check		OK	04:26
Air Blank		0.000	04:27
Control Test		0.082	04:27
Air Blank		0.000	04:28
Subject Sample	# 1	0.184	04:28
Breath Volume		2.804	04:28
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.190	04:31
Breath Volume		3.597	04:31
Air Blank		0.000	04:32
Control Test		0.082	04:32
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 02/27/2018 21:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WILLIAMS,TERRELL,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: EHRNREITER,FRANKLIN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 01:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/11/2018 02:19:47

Diagnostics Check		OK	02:11	
Air Blank		0.000	02:11	
Control Test		0.079	02:11	
Air Blank		0.000	02:12	
Subject Sample	# 1		02:15	No Sample Provided
Breath Volume		0.000	02:15	
Air Blank		0.000	02:16	
Air Blank		0.000	02:18	
Subject Sample	# 2		02:18	Subject Test Refused
Breath Volume		0.000	02:18	
Air Blank		0.000	02:18	
Control Test		0.080	02:19	
Air Blank		0.000	02:19	
Diagnostics Check		OK	02:19	

Data Download Date/Time: 02/24/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRELL,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: EHRNREITER,FRANKLIN
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 23:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:12

SUBJECT'S LAST NAME: WILFRIDO

DATE/TIME OF TEST: 02/15/2018 23:47:48

Diagnostics Check		OK	23:33	
Air Blank		0.000	23:33	
Control Test		0.079	23:33	
Air Blank		0.000	23:34	
Subject Sample	# 1	0.190	23:37	Volume Not Met
Breath Volume		0.070	23:37	
Air Blank		0.000	23:38	
Air Blank		0.000	23:39	
Subject Sample	# 2	0.228	23:40	
Breath Volume		1.484	23:40	
Air Blank		0.000	23:41	
Air Blank		0.000	23:42	
Subject Sample	# 3	0.158	23:45	Volume Not Met
Breath Volume		0.316	23:45	
Air Blank		0.000	23:46	
Control Test		0.077	23:47	
Air Blank		0.000	23:47	
Diagnostics Check		OK	23:47	

Data Download Date/Time: 02/24/2018 08:55

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WITT,BRADLEY,G
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: WATSON,DAVID
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 01/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 15:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:16

SUBJECT'S LAST NAME: CONN

DATE/TIME OF TEST: 02/19/2018 16:42:30

Diagnostics Check		OK	16:36
Air Blank		0.000	16:36
Control Test		0.081	16:37
Air Blank		0.000	16:37
Subject Sample	# 1	0.273	16:37
Breath Volume		1.488	16:37
Air Blank		0.000	16:38
Air Blank		0.000	16:40
Subject Sample	# 2	0.266	16:40
Breath Volume		1.441	16:40
Air Blank		0.000	16:41
Control Test		0.079	16:41
Air Blank		0.000	16:42
Diagnostics Check		OK	16:42

Data Download Date/Time: 02/24/2018 08:55

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRANNON,AARON,K
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: WATSON,DAVID
ARRESTING OFFICER AGENCY: Bradford County SO
LAST AGENCY INSPECTION: 02/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 05:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:06

SUBJECT'S LAST NAME: DURRANCE

DATE/TIME OF TEST: 02/25/2018 06:39:01

Diagnostics Check		OK	06:32
Air Blank		0.000	06:33
Control Test		0.079	06:33
Air Blank		0.000	06:34
Subject Sample	# 1	0.159	06:34
Breath Volume		2.753	06:34
Air Blank		0.000	06:35
Air Blank		0.000	06:37
Subject Sample	# 2	0.154	06:37
Breath Volume		2.339	06:37
Air Blank		0.000	06:38
Control Test		0.078	06:38
Air Blank		0.000	06:38
Diagnostics Check		OK	06:39

Data Download Date/Time: 03/15/2018 12:44

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001290

INSTRUMENT REGISTERED TO: BRADFORD COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MORRELL,JOSEPH,B
BREATH TEST OPERATOR AGENCY: Bradford County SO
ARRESTING OFFICER: JONES,BRIAN
ARRESTING OFFICER AGENCY: Lawtey PD
LAST AGENCY INSPECTION: 02/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/25/2018 22:22:19

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:16	
Control Test		0.079	22:16	
Air Blank		0.000	22:17	
Subject Sample	# 1	0.152	22:18	
Breath Volume		3.796	22:18	
Air Blank		0.000	22:19	
Air Blank		0.000	22:20	
Subject Sample	# 2	0.154	22:21	
Breath Volume		2.671	22:21	
Air Blank			22:22	RFI Detect
Air Blank			22:22	RFI Detect

Data Download Date/Time: 03/15/2018 12:44

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: GARBOSKI,GARY
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:00

SUBJECT'S LAST NAME: CONNER

DATE/TIME OF TEST: 02/02/2018 20:49:14

Diagnostics Check		OK	20:44	
Air Blank		0.000	20:44	
Control Test		0.081	20:45	
Air Blank		0.000	20:45	
Subject Sample	# 1		20:47	Subject Test Refused
Breath Volume		1.023	20:47	
Air Blank		0.000	20:48	
Control Test		0.080	20:48	
Air Blank		0.000	20:49	
Diagnostics Check		OK	20:49	

Data Download Date/Time: 02/21/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: LONG,KATHY
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 04:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:41

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/04/2018 06:15:20

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.080	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.010	06:10
Breath Volume		2.882	06:10
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.008	06:13
Breath Volume		2.902	06:13
Air Blank		0.000	06:14
Control Test		0.079	06:14
Air Blank		0.000	06:15
Diagnostics Check		OK	06:15

Data Download Date/Time: 02/21/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LANDIS,BRYAN,S
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: PRIETO,KENNETH
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:44

SUBJECT'S LAST NAME: NELSON

DATE/TIME OF TEST: 02/04/2018 20:28:32

Diagnostics Check		OK	20:22
Air Blank		0.000	20:22
Control Test		0.080	20:23
Air Blank		0.000	20:23
Subject Sample	# 1	0.291	20:24
Breath Volume		2.980	20:24
Air Blank		0.000	20:24
Air Blank		0.000	20:26
Subject Sample	# 2	0.276	20:26
Breath Volume		2.367	20:26
Air Blank		0.000	20:27
Control Test		0.078	20:27
Air Blank		0.000	20:28
Diagnostics Check		OK	20:28

Data Download Date/Time: 02/21/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001293

INSTRUMENT REGISTERED TO: WILLISTON PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WRIGHT,MARQUIS,A
BREATH TEST OPERATOR AGENCY: Williston PD
ARRESTING OFFICER: WRIGHT,MARQUIS
ARRESTING OFFICER AGENCY: Williston PD
LAST AGENCY INSPECTION: 01/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 203498
DRY GAS EXPIRY DATE: 12/10/2018
VIOLATION CODE: DUI
ARREST TIME: 20:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:19

SUBJECT'S LAST NAME: MONROE

DATE/TIME OF TEST: 02/19/2018 22:12:49

Diagnostics Check		OK	22:04
Air Blank		0.000	22:04
Control Test		0.079	22:04
Air Blank		0.000	22:05
Subject Sample	# 1	0.159	22:08
Breath Volume		1.648	22:08
Air Blank		0.000	22:09
Air Blank		0.000	22:10
Subject Sample	# 2	0.155	22:11
Breath Volume		2.160	22:11
Air Blank		0.000	22:11
Control Test		0.078	22:12
Air Blank		0.000	22:12
Diagnostics Check		OK	22:12

Data Download Date/Time: 02/21/2018 11:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:02

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 02/01/2018 01:27:18

Diagnostics Check		OK	01:23	
Air Blank		0.000	01:23	
Control Test		0.080	01:23	
Air Blank		0.000	01:24	
Subject Sample	# 1		01:25	Subject Test Refused
Breath Volume		0.000	01:25	
Air Blank		0.000	01:26	
Control Test		0.080	01:26	
Air Blank		0.000	01:27	
Diagnostics Check		OK	01:27	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPLING,JEFF,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: APPLING,JEFF
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:54

SUBJECT'S LAST NAME: AMOS

DATE/TIME OF TEST: 02/02/2018 03:39:07

Diagnostics Check		OK	03:32
Air Blank		0.000	03:32
Control Test		0.080	03:33
Air Blank		0.000	03:33
Subject Sample	# 1	0.000	03:34
Breath Volume		2.410	03:34
Air Blank		0.000	03:35
Air Blank		0.000	03:37
Subject Sample	# 2	0.000	03:37
Breath Volume		2.503	03:37
Air Blank		0.000	03:38
Control Test		0.079	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAPMAN,MAVERICK,E
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CHAPMAN,MAVERICK
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:20

SUBJECT'S LAST NAME: GOELLER

DATE/TIME OF TEST: 02/03/2018 00:50:49

Diagnostics Check		OK	00:44
Air Blank		0.000	00:45
Control Test		0.082	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.214	00:46
Breath Volume		2.878	00:46
Air Blank		0.000	00:47
Air Blank		0.000	00:48
Subject Sample	# 2	0.198	00:49
Breath Volume		3.097	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: AVERY,BAYLEE
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: FELIX

DATE/TIME OF TEST: 02/03/2018 02:24:19

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.081	02:16	
Air Blank		0.000	02:17	
Subject Sample	# 1		02:20	No Sample Provided
Breath Volume		0.000	02:20	
Air Blank		0.000	02:20	
Air Blank		0.000	02:22	
Subject Sample	# 2		02:22	Subject Test Refused
Breath Volume		0.000	02:22	
Air Blank		0.000	02:23	
Control Test		0.081	02:23	
Air Blank		0.000	02:24	
Diagnostics Check		OK	02:24	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARTIN,SHAWN,A
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: MARTIN,SHAWN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 16:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:18

SUBJECT'S LAST NAME: COOK

DATE/TIME OF TEST: 02/03/2018 17:41:40

Diagnostics Check		OK	17:38	
Air Blank		0.000	17:38	
Control Test		0.081	17:39	
Air Blank		0.000	17:39	
Subject Sample	# 1		17:40	Subject Test Refused
Breath Volume		0.000	17:40	
Air Blank		0.000	17:40	
Control Test		0.081	17:41	
Air Blank		0.000	17:41	
Diagnostics Check		OK	17:41	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 18:21
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:05

SUBJECT'S LAST NAME: FINDLEY

DATE/TIME OF TEST: 02/03/2018 20:46:16

Diagnostics Check		OK	20:39
Air Blank		0.000	20:39
Control Test		0.080	20:39
Air Blank		0.000	20:40
Subject Sample	# 1	0.125	20:41
Breath Volume		1.222	20:41
Air Blank		0.000	20:42
Air Blank		0.000	20:44
Subject Sample	# 2	0.125	20:44
Breath Volume		1.980	20:44
Air Blank		0.000	20:45
Control Test		0.077	20:45
Air Blank		0.000	20:46
Diagnostics Check		OK	20:46

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MARANGE

DATE/TIME OF TEST: 02/04/2018 02:14:04

Diagnostics Check		OK	02:02	
Air Blank		0.000	02:03	
Control Test		0.080	02:03	
Air Blank		0.000	02:03	
Subject Sample	# 1	0.112	02:06	Volume Not Met
Breath Volume		0.023	02:06	
Air Blank		0.000	02:07	
Air Blank		0.000	02:09	
Subject Sample	# 2	0.143	02:12	Volume Not Met
Breath Volume		0.171	02:12	
Air Blank		0.000	02:13	
Control Test		0.079	02:13	
Air Blank		0.000	02:13	
Diagnostics Check		OK	02:14	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:42

SUBJECT'S LAST NAME: MARANGE

DATE/TIME OF TEST: 02/04/2018 02:25:44

Diagnostics Check		OK	02:16
Air Blank		0.000	02:17
Control Test		0.079	02:17
Air Blank		0.000	02:18
Subject Sample	# 1	0.141	02:20
Breath Volume		1.648	02:20
Air Blank		0.000	02:21
Air Blank		0.000	02:22
Subject Sample	# 2	0.147	02:24
Breath Volume		1.101	02:24
Air Blank		0.000	02:24
Control Test		0.079	02:25
Air Blank		0.000	02:25
Diagnostics Check		OK	02:25

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/04/2018 02:49:22

Diagnostics Check		OK	02:40	
Air Blank		0.000	02:41	
Control Test		0.079	02:41	
Air Blank		0.000	02:41	
Subject Sample	# 1		02:45	No Sample Provided
Breath Volume		0.000	02:45	
Air Blank		0.000	02:45	
Air Blank		0.000	02:47	
Subject Sample	# 2		02:48	Subject Test Refused
Breath Volume		0.000	02:48	
Air Blank		0.000	02:48	
Control Test		0.079	02:48	
Air Blank		0.000	02:49	
Diagnostics Check		OK	02:49	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HUBBARD,NICOLAS,D
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HUBBARD,NICOLAS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: FAIN

DATE/TIME OF TEST: 02/04/2018 04:16:40

Diagnostics Check		OK	04:06	
Air Blank		0.000	04:07	
Control Test		0.080	04:07	
Air Blank		0.000	04:07	
Subject Sample	# 1		04:11	No Sample Provided
Breath Volume		0.000	04:11	
Air Blank		0.000	04:11	
Air Blank		0.000	04:13	
Subject Sample	# 2		04:15	RFI Detect
Breath Volume		0.000	04:15	
Air Blank		0.000	04:15	
Control Test		0.080	04:16	
Air Blank		0.000	04:16	
Diagnostics Check		OK	04:16	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DANIELS,DAVID,C
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DANIELS,DAVID
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: WEATHERINGTON

DATE/TIME OF TEST: 02/05/2018 00:41:45

Diagnostics Check		OK	00:34
Air Blank		0.000	00:35
Control Test		0.080	00:35
Air Blank		0.000	00:36
Subject Sample	# 1	0.189	00:37
Breath Volume		1.378	00:37
Air Blank		0.000	00:37
Air Blank		0.000	00:39
Subject Sample	# 2	0.188	00:40
Breath Volume		1.640	00:40
Air Blank		0.000	00:40
Control Test		0.079	00:41
Air Blank		0.000	00:41
Diagnostics Check		OK	00:41

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DESOUSA,CARLOS,J
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: DESOUSA,CARLOS
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: AGRICOLA

DATE/TIME OF TEST: 02/06/2018 01:05:39

Diagnostics Check		OK	01:01	
Air Blank		0.000	01:02	
Control Test		0.080	01:02	
Air Blank		0.000	01:02	
Subject Sample	# 1		01:04	Subject Test Refused
Breath Volume		0.000	01:04	
Air Blank		0.000	01:04	
Control Test		0.080	01:05	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,CHAD,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FLOYD,CHAD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: MCCURLEY

DATE/TIME OF TEST: 02/06/2018 21:40:06

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.080	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.012	21:35
Breath Volume		1.484	21:35
Air Blank		0.000	21:36
Air Blank		0.000	21:38
Subject Sample	# 2	0.010	21:38
Breath Volume		1.695	21:38
Air Blank		0.000	21:39
Control Test		0.080	21:39
Air Blank		0.000	21:40
Diagnostics Check		OK	21:40

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WEAVER,COREY,R
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WEAVER,COREY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:02

SUBJECT'S LAST NAME: ASKEGARD

DATE/TIME OF TEST: 02/07/2018 23:29:02

Diagnostics Check		OK	23:22
Air Blank		0.000	23:23
Control Test		0.079	23:23
Air Blank		0.000	23:23
Subject Sample	# 1	0.246	23:24
Breath Volume		1.570	23:24
Air Blank		0.000	23:25
Air Blank		0.000	23:26
Subject Sample	# 2	0.233	23:27
Breath Volume		1.460	23:27
Air Blank		0.000	23:28
Control Test		0.078	23:28
Air Blank		0.000	23:28
Diagnostics Check		OK	23:29

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:00

SUBJECT'S LAST NAME: DOMINGUEZ

DATE/TIME OF TEST: 02/08/2018 01:29:26

Diagnostics Check		OK	01:21
Air Blank		0.000	01:22
Control Test		0.080	01:22
Air Blank		0.000	01:22
Subject Sample	# 1	0.152	01:24
Breath Volume		1.550	01:24
Air Blank		0.000	01:24
Air Blank		0.000	01:26
Subject Sample	# 2	0.161	01:27
Breath Volume		2.511	01:27
Air Blank		0.000	01:28
Control Test		0.081	01:28
Air Blank		0.000	01:29
Diagnostics Check		OK	01:29

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERHUNE,JEREMY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:02
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:30

SUBJECT'S LAST NAME: CAPUTO

DATE/TIME OF TEST: 02/09/2018 02:29:19

Diagnostics Check		OK	02:14	
Air Blank		0.000	02:15	
Control Test		0.080	02:15	
Air Blank		0.000	02:15	
Subject Sample	# 1	0.108	02:16	
Breath Volume		1.355	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:18	
Subject Sample	# 2	0.101	02:22	Volume Not Met
Breath Volume		0.191	02:22	
Air Blank		0.000	02:22	
Air Blank		0.000	02:24	
Subject Sample	# 3	0.095	02:27	Volume Not Met
Breath Volume		0.605	02:27	
Air Blank		0.000	02:28	
Control Test		0.078	02:28	
Air Blank		0.000	02:29	
Diagnostics Check		OK	02:29	

Data Download Date/Time: 02/28/2018 21:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: YOUNG,JOHN,F
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: YOUNG,JOHN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:57

SUBJECT'S LAST NAME: JOSEPH

DATE/TIME OF TEST: 02/12/2018 20:23:48

Diagnostics Check		OK	20:17
Air Blank		0.000	20:17
Control Test		0.081	20:18
Air Blank		0.000	20:18
Subject Sample	# 1	0.000	20:19
Breath Volume		2.281	20:19
Air Blank		0.000	20:19
Air Blank		0.000	20:21
Subject Sample	# 2	0.000	20:22
Breath Volume		2.968	20:22
Air Blank		0.000	20:22
Control Test		0.080	20:23
Air Blank		0.000	20:23
Diagnostics Check		OK	20:23

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRANTHAM,WILLIAM,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: GRANTHAM,WILLIAM
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:38

SUBJECT'S LAST NAME: RENONCE

DATE/TIME OF TEST: 02/12/2018 21:08:53

Diagnostics Check		OK	21:02
Air Blank		0.000	21:03
Control Test		0.079	21:03
Air Blank		0.000	21:04
Subject Sample	# 1	0.194	21:04
Breath Volume		2.457	21:04
Air Blank		0.000	21:05
Air Blank		0.000	21:06
Subject Sample	# 2	0.192	21:07
Breath Volume		2.218	21:07
Air Blank		0.000	21:07
Control Test		0.079	21:08
Air Blank		0.000	21:08
Diagnostics Check		OK	21:08

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CRIBB,BRIAN,S
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CRIBB,BRIAN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:41

SUBJECT'S LAST NAME: MAY

DATE/TIME OF TEST: 02/15/2018 23:11:25

Diagnostics Check		OK	23:08	
Air Blank		0.000	23:09	
Control Test		0.080	23:09	
Air Blank		0.000	23:09	
Subject Sample	# 1		23:10	Subject Test Refused
Breath Volume		0.000	23:10	
Air Blank		0.000	23:10	
Control Test		0.080	23:10	
Air Blank		0.000	23:11	
Diagnostics Check		OK	23:11	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSNAK,BRANDEN,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: RUSNAK,BRANDEN
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 06:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:08

SUBJECT'S LAST NAME: PERRY

DATE/TIME OF TEST: 02/17/2018 08:39:51

Diagnostics Check		OK	08:30
Air Blank		0.000	08:30
Control Test		0.079	08:31
Air Blank		0.000	08:31
Subject Sample	# 1	0.124	08:34
Breath Volume		3.144	08:34
Air Blank		0.000	08:35
Air Blank		0.000	08:36
Subject Sample	# 2	0.124	08:38
Breath Volume		3.203	08:38
Air Blank		0.000	08:38
Control Test		0.079	08:39
Air Blank		0.000	08:39
Diagnostics Check		OK	08:39

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: APPLING,JEFF,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: APPLING,JEFF
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:12

SUBJECT'S LAST NAME: BUCKMAN

DATE/TIME OF TEST: 02/17/2018 22:33:22

Diagnostics Check		OK	22:26
Air Blank		0.000	22:26
Control Test		0.080	22:27
Air Blank		0.000	22:27
Subject Sample	# 1	0.254	22:28
Breath Volume		1.824	22:28
Air Blank		0.000	22:29
Air Blank		0.000	22:31
Subject Sample	# 2	0.239	22:31
Breath Volume		1.140	22:31
Air Blank		0.000	22:32
Control Test		0.078	22:32
Air Blank		0.000	22:33
Diagnostics Check		OK	22:33

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:24

SUBJECT'S LAST NAME: HILBERT

DATE/TIME OF TEST: 02/22/2018 00:55:49

Diagnostics Check		OK	00:47	
Air Blank		0.000	00:47	
Control Test		0.081	00:48	
Air Blank		0.000	00:48	
Subject Sample	# 1		00:51	No Sample Provided
Breath Volume		0.000	00:51	
Air Blank		0.000	00:52	
Air Blank		0.000	00:54	
Subject Sample	# 2		00:54	Subject Test Refused
Breath Volume		0.000	00:54	
Air Blank		0.000	00:54	
Control Test		0.080	00:55	
Air Blank		0.000	00:55	
Diagnostics Check		OK	00:55	

Data Download Date/Time: 02/28/2018 21:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TERHUNE,JEREMY,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: TERHUNE,JEREMY
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 10:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 11:00

SUBJECT'S LAST NAME: PETERS

DATE/TIME OF TEST: 02/23/2018 11:50:49

Diagnostics Check		OK	11:43
Air Blank		0.000	11:44
Control Test		0.080	11:44
Air Blank		0.000	11:45
Subject Sample	# 1	0.178	11:46
Breath Volume		1.644	11:46
Air Blank		0.000	11:46
Air Blank		0.000	11:48
Subject Sample	# 2	0.182	11:49
Breath Volume		2.554	11:49
Air Blank		0.000	11:49
Control Test		0.078	11:50
Air Blank		0.000	11:50
Diagnostics Check		OK	11:50

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARVOTH,DANIEL,P
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HARVOTH,DANIEL
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:09

SUBJECT'S LAST NAME: WATSON

DATE/TIME OF TEST: 02/25/2018 03:38:11

Diagnostics Check		OK	03:29
Air Blank		0.000	03:29
Control Test		0.080	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.117	03:33
Breath Volume		2.160	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:35
Subject Sample	# 2	0.116	03:36
Breath Volume		1.308	03:36
Air Blank		0.000	03:37
Control Test		0.078	03:37
Air Blank		0.000	03:38
Diagnostics Check		OK	03:38

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPENCER,RAYMOND,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: SPENCER,RAYMOND
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 06:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:58

SUBJECT'S LAST NAME: MALPICA-BUZZULINI

DATE/TIME OF TEST: 02/25/2018 07:27:26

Diagnostics Check		OK	07:18
Air Blank		0.000	07:19
Control Test		0.080	07:19
Air Blank		0.000	07:20
Subject Sample	# 1	0.125	07:22
Breath Volume		1.339	07:22
Air Blank		0.000	07:23
Air Blank		0.000	07:24
Subject Sample	# 2	0.127	07:25
Breath Volume		1.382	07:25
Air Blank		0.000	07:26
Control Test		0.080	07:26
Air Blank		0.000	07:27
Diagnostics Check		OK	07:27

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLOYD,CHAD,M
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: FLOYD,CHAD
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 20:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: HANSEN

DATE/TIME OF TEST: 02/25/2018 21:33:32

Diagnostics Check		OK	21:30	
Air Blank		0.000	21:30	
Control Test		0.080	21:31	
Air Blank		0.000	21:31	
Subject Sample	# 1		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:32	
Control Test		0.081	21:32	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOWTON,CARSON,H
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: HOWTON,CARSON
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:55

SUBJECT'S LAST NAME: RAVEENDRAN

DATE/TIME OF TEST: 02/26/2018 02:23:00

Diagnostics Check		OK	02:15
Air Blank		0.000	02:15
Control Test		0.080	02:16
Air Blank		0.000	02:16
Subject Sample	# 1	0.284	02:17
Breath Volume		1.515	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.268	02:21
Breath Volume		1.273	02:21
Air Blank		0.000	02:22
Control Test		0.080	02:22
Air Blank		0.000	02:22
Diagnostics Check		OK	02:23

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COOK,CALEB,Z
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: WATERS,CALEB
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 19:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: LOFTIN

DATE/TIME OF TEST: 02/27/2018 20:18:54

Diagnostics Check		OK	20:16	
Air Blank		0.000	20:16	
Control Test		0.079	20:16	
Air Blank		0.000	20:17	
Subject Sample	# 1		20:17	Subject Test Refused
Breath Volume		0.000	20:17	
Air Blank		0.000	20:17	
Control Test		0.080	20:18	
Air Blank		0.000	20:18	
Diagnostics Check		OK	20:18	

Data Download Date/Time: 02/28/2018 21:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001303

INSTRUMENT REGISTERED TO: SANTA ROSA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CARLSON,JOSHUA,L
BREATH TEST OPERATOR AGENCY: Santa Rosa County SO
ARRESTING OFFICER: CARLSON,JOSHUA
ARRESTING OFFICER AGENCY: Santa Rosa County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629801
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 16:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: MCCURLEY

DATE/TIME OF TEST: 02/28/2018 17:00:15

Diagnostics Check		OK	16:47	
Air Blank		0.000	16:47	
Control Test		0.080	16:48	
Air Blank		0.000	16:48	
Subject Sample	# 1	0.000	16:51	Volume Not Met
Breath Volume		0.136	16:51	
Air Blank		0.000	16:52	
Air Blank		0.000	16:54	
Subject Sample	# 2	0.000	16:55	
Breath Volume		1.214	16:55	
Air Blank		0.000	16:56	
Air Blank		0.000	16:58	
Subject Sample	# 3	0.000	16:58	
Breath Volume		1.332	16:58	
Air Blank		0.000	16:59	
Control Test		0.080	16:59	
Air Blank		0.000	17:00	
Diagnostics Check		OK	17:00	

Data Download Date/Time: 02/28/2018 21:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: POKIPALA,JONAH
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: MEMELLI

DATE/TIME OF TEST: 02/09/2018 02:54:32

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:48
Subject Sample	# 1	0.117	02:49
Breath Volume		3.882	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.109	02:53
Breath Volume		1.761	02:53
Air Blank		0.000	02:53
Control Test		0.078	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 03/21/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CHAVIRA,BRANDON,A
BREATH TEST OPERATOR AGENCY: Panama City Beach PD
ARRESTING OFFICER: CHAVIRA,BRANDON
ARRESTING OFFICER AGENCY: Panama City Beach PD
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 00:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:45

SUBJECT'S LAST NAME: BENNEFIELD

DATE/TIME OF TEST: 02/10/2018 02:24:20

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.131	02:19
Breath Volume		1.480	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.140	02:22
Breath Volume		1.585	02:22
Air Blank		0.000	02:23
Control Test		0.079	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/21/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:16

SUBJECT'S LAST NAME: BURGE

DATE/TIME OF TEST: 02/17/2018 00:43:02

Diagnostics Check		OK	00:36
Air Blank		0.000	00:36
Control Test		0.080	00:37
Air Blank		0.000	00:37
Subject Sample	# 1	0.117	00:38
Breath Volume		1.699	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.123	00:41
Breath Volume		1.832	00:41
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:42
Diagnostics Check		OK	00:43

Data Download Date/Time: 03/21/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: CHASTAIN,KRISTA
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 23:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: HOLLAND

DATE/TIME OF TEST: 02/18/2018 00:26:36

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.080	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.150	00:22
Breath Volume		1.464	00:22
Air Blank		0.000	00:22
Air Blank		0.000	00:24
Subject Sample	# 2	0.151	00:25
Breath Volume		1.492	00:25
Air Blank		0.000	00:25
Control Test		0.079	00:26
Air Blank		0.000	00:26
Diagnostics Check		OK	00:26

Data Download Date/Time: 03/21/2018 10:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUBEL,NEIL,F
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: RUBEL,NEIL
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: Yes
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 03:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:53

SUBJECT'S LAST NAME: SKINNER

DATE/TIME OF TEST: 02/18/2018 04:20:18

Diagnostics Check		OK	04:14
Air Blank		0.000	04:14
Control Test		0.079	04:14
Air Blank		0.000	04:15
Subject Sample	# 1	0.198	04:15
Breath Volume		2.539	04:15
Air Blank		0.000	04:16
Air Blank		0.000	04:18
Subject Sample	# 2	0.197	04:18
Breath Volume		2.171	04:18
Air Blank		0.000	04:19
Control Test		0.080	04:19
Air Blank		0.000	04:20
Diagnostics Check		OK	04:20

Data Download Date/Time: 03/21/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BOTZONG,DUSTIN,R
BREATH TEST OPERATOR AGENCY: OTHER Local
ARRESTING OFFICER: BOTZONG,DUSTIN
ARRESTING OFFICER AGENCY: OTHER Local
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: BOHNEN

DATE/TIME OF TEST: 02/23/2018 22:52:43

Diagnostics Check		OK	22:45
Air Blank		0.000	22:46
Control Test		0.080	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.123	22:48
Breath Volume		1.398	22:48
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.125	22:51
Breath Volume		1.710	22:51
Air Blank		0.000	22:51
Control Test		0.078	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 03/21/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001309

INSTRUMENT REGISTERED TO: BAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POTTER,MONTREZ,R
BREATH TEST OPERATOR AGENCY: Bay County SO
ARRESTING OFFICER: POTTER,MONTREZ
ARRESTING OFFICER AGENCY: Bay County SO
LAST AGENCY INSPECTION: 02/06/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG628401
DRY GAS EXPIRY DATE: 10/10/2018
VIOLATION CODE: DUI
ARREST TIME: 21:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:33

SUBJECT'S LAST NAME: WILKERSON

DATE/TIME OF TEST: 02/25/2018 22:00:41

Diagnostics Check		OK	21:53
Air Blank		0.000	21:53
Control Test		0.080	21:54
Air Blank		0.000	21:54
Subject Sample	# 1	0.103	21:55
Breath Volume		1.589	21:55
Air Blank		0.000	21:56
Air Blank		0.000	21:58
Subject Sample	# 2	0.099	21:59
Breath Volume		1.507	21:59
Air Blank		0.000	21:59
Control Test		0.078	22:00
Air Blank		0.000	22:00
Diagnostics Check		OK	22:00

Data Download Date/Time: 03/21/2018 10:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001310

INSTRUMENT REGISTERED TO: NICEVILLE PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SULLIVAN,DANNY,
BREATH TEST OPERATOR AGENCY: Niceville PD
ARRESTING OFFICER: SULLIVAN,DANNY
ARRESTING OFFICER AGENCY: Niceville PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 01:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: WHITLEY

DATE/TIME OF TEST: 02/18/2018 02:26:33

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.078	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.177	02:22
Breath Volume		4.261	02:22
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.173	02:24
Breath Volume		2.539	02:24
Air Blank		0.000	02:25
Control Test		0.078	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 03/07/2018 18:00

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: LAIRD,ETHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 19:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:33

SUBJECT'S LAST NAME: OWENS

DATE/TIME OF TEST: 02/03/2018 20:58:12

Diagnostics Check		OK	20:55	
Air Blank		0.000	20:55	
Control Test		0.081	20:56	
Air Blank		0.000	20:56	
Subject Sample	# 1		20:56	Subject Test Refused
Breath Volume		0.000	20:56	
Air Blank		0.000	20:57	
Control Test		0.080	20:57	
Air Blank		0.000	20:58	
Diagnostics Check		OK	20:58	

Data Download Date/Time: 02/28/2018 07:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KING,JENNIFER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:37

SUBJECT'S LAST NAME: RATLIFF

DATE/TIME OF TEST: 02/06/2018 01:15:02

Diagnostics Check		OK	01:07
Air Blank		0.000	01:08
Control Test		0.080	01:08
Air Blank		0.000	01:09
Subject Sample	# 1	0.239	01:10
Breath Volume		2.750	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.227	01:13
Breath Volume		1.937	01:13
Air Blank		0.000	01:14
Control Test		0.080	01:14
Air Blank		0.000	01:14
Diagnostics Check		OK	01:15

Data Download Date/Time: 02/28/2018 07:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: AVERY,JOHN,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: COKONOUGHER,ROBERT
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 00:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: THANG

DATE/TIME OF TEST: 02/07/2018 02:21:32

Diagnostics Check		OK	02:04	
Air Blank		0.000	02:05	
Control Test		0.080	02:05	
Air Blank		0.000	02:06	
Subject Sample	# 1	0.207	02:09	Volume Not Met
Breath Volume		0.230	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:11	
Subject Sample	# 2	0.180	02:14	
Breath Volume		1.230	02:14	
Air Blank		0.000	02:14	
Air Blank		0.000	02:16	
Subject Sample	# 3	0.174	02:19	Volume Not Met
Breath Volume		0.710	02:19	
Air Blank		0.000	02:20	
Control Test		0.077	02:20	
Air Blank		0.000	02:21	
Diagnostics Check		OK	02:21	

Data Download Date/Time: 02/28/2018 07:59

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POUNCEY,KATHLEEN,R
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: POUNCEY,KATHLEEN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 12:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 12:12

SUBJECT'S LAST NAME: DUMAN

DATE/TIME OF TEST: 02/07/2018 12:43:16

Diagnostics Check		OK	12:37
Air Blank		0.000	12:37
Control Test		0.080	12:37
Air Blank		0.000	12:38
Subject Sample	# 1	0.012	12:38
Breath Volume		1.675	12:38
Air Blank		0.000	12:39
Air Blank		0.000	12:41
Subject Sample	# 2	0.013	12:41
Breath Volume		3.453	12:41
Air Blank		0.000	12:42
Control Test		0.079	12:42
Air Blank		0.000	12:43
Diagnostics Check		OK	12:43

Data Download Date/Time: 02/28/2018 07:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SANDERS,DAVID,O
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: KING,JENNIFER
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: THOMPSON

DATE/TIME OF TEST: 02/13/2018 01:09:27

Diagnostics Check		OK	01:02
Air Blank		0.000	01:03
Control Test		0.080	01:03
Air Blank		0.000	01:04
Subject Sample	# 1	0.160	01:04
Breath Volume		2.917	01:04
Air Blank		0.000	01:05
Air Blank		0.000	01:07
Subject Sample	# 2	0.145	01:07
Breath Volume		1.535	01:07
Air Blank		0.000	01:08
Control Test		0.079	01:08
Air Blank		0.000	01:09
Diagnostics Check		OK	01:09

Data Download Date/Time: 02/28/2018 07:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FUGATE,MITCHELL,W
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: BRANNON,JUSTIN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 16:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:05

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/19/2018 17:41:16

Diagnostics Check		OK	17:29	
Air Blank		0.000	17:30	
Control Test		0.079	17:30	
Air Blank		0.000	17:31	
Subject Sample	# 1		17:34	No Sample Provided
Breath Volume		0.000	17:34	
Air Blank		0.000	17:34	
Air Blank		0.000	17:36	
Subject Sample	# 2		17:39	No Sample Provided
Breath Volume		0.000	17:39	
Air Blank		0.000	17:40	
Control Test		0.080	17:40	
Air Blank		0.000	17:41	
Diagnostics Check		OK	17:41	

Data Download Date/Time: 02/28/2018 07:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001311

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GAYDOS,JONATHAN,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GAYDOS,JONATHAN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:13

SUBJECT'S LAST NAME: HILLIGOSS

DATE/TIME OF TEST: 02/26/2018 21:48:03

Diagnostics Check		OK	21:41
Air Blank		0.000	21:42
Control Test		0.081	21:42
Air Blank		0.000	21:42
Subject Sample	# 1	0.171	21:43
Breath Volume		1.570	21:43
Air Blank		0.000	21:44
Air Blank		0.000	21:46
Subject Sample	# 2	0.177	21:46
Breath Volume		1.386	21:46
Air Blank		0.000	21:47
Control Test		0.080	21:47
Air Blank		0.000	21:47
Diagnostics Check		OK	21:48

Data Download Date/Time: 02/28/2018 07:59

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAGAN,JAMIE,A
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: RAGAN,JAMIE
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 19:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:36

SUBJECT'S LAST NAME: MCCLENDON

DATE/TIME OF TEST: 02/04/2018 20:06:50

Diagnostics Check		OK	20:00
Air Blank		0.000	20:00
Control Test		0.078	20:01
Air Blank		0.000	20:01
Subject Sample	# 1	0.102	20:02
Breath Volume		2.968	20:02
Air Blank		0.000	20:02
Air Blank		0.000	20:04
Subject Sample	# 2	0.106	20:05
Breath Volume		2.972	20:05
Air Blank		0.000	20:05
Control Test		0.076	20:06
Air Blank		0.000	20:06
Diagnostics Check		OK	20:06

Data Download Date/Time: 02/28/2018 07:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001312

INSTRUMENT REGISTERED TO: WALTON COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PULLINGS,ALLEN,M
BREATH TEST OPERATOR AGENCY: Walton County SO
ARRESTING OFFICER: GLENISTER,JOHN
ARRESTING OFFICER AGENCY: Walton County SO
LAST AGENCY INSPECTION: 01/18/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 754250
DRY GAS EXPIRY DATE: 04/03/2020
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: HASTINGS

DATE/TIME OF TEST: 02/25/2018 05:11:41

Diagnostics Check		OK	05:05
Air Blank		0.000	05:05
Control Test		0.078	05:05
Air Blank		0.000	05:06
Subject Sample	# 1	0.217	05:07
Breath Volume		3.898	05:07
Air Blank		0.000	05:07
Air Blank		0.000	05:09
Subject Sample	# 2	0.213	05:10
Breath Volume		3.984	05:10
Air Blank		0.000	05:10
Control Test		0.077	05:11
Air Blank		0.000	05:11
Diagnostics Check		OK	05:11

Data Download Date/Time: 02/28/2018 07:53

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001320

INSTRUMENT REGISTERED TO: OKEECHOBEE CO SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLLOCK,STEVEN,D
BREATH TEST OPERATOR AGENCY: Okeechobee County SO
ARRESTING OFFICER: POLLOCK,STEVEN
ARRESTING OFFICER AGENCY: Okeechobee County SO
LAST AGENCY INSPECTION: 01/11/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 21:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:59

SUBJECT'S LAST NAME: HANVEY FAIRCLOTH

DATE/TIME OF TEST: 02/01/2018 22:22:55

Diagnostics Check		OK	22:19	
Air Blank		0.000	22:20	
Control Test		0.079	22:20	
Air Blank		0.000	22:20	
Subject Sample	# 1		22:21	Subject Test Refused
Breath Volume		0.000	22:21	
Air Blank		0.000	22:21	
Control Test		0.079	22:22	
Air Blank		0.000	22:22	
Diagnostics Check		OK	22:22	

Data Download Date/Time: 02/07/2018 11:58

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001321

INSTRUMENT REGISTERED TO: OKEECHOBEE COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: IGLESIAS,JORGE,L
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: IGLESIAS,JORGE
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/07/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 507129
DRY GAS EXPIRY DATE: 08/04/2019
VIOLATION CODE: DUI
ARREST TIME: 20:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: BENAVIDEZ

DATE/TIME OF TEST: 02/17/2018 21:39:49

Diagnostics Check		OK	21:33
Air Blank		0.000	21:33
Control Test		0.081	21:34
Air Blank		0.000	21:34
Subject Sample	# 1	0.016	21:35
Breath Volume		4.238	21:35
Air Blank		0.000	21:35
Air Blank		0.000	21:37
Subject Sample	# 2	0.015	21:38
Breath Volume		5.277	21:38
Air Blank		0.000	21:38
Control Test		0.080	21:39
Air Blank		0.000	21:39
Diagnostics Check		OK	21:39

Data Download Date/Time: 03/12/2018 12:41

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARGERUM,KELLEY,J
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: BURK,TODD
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:32

SUBJECT'S LAST NAME: GROODY

DATE/TIME OF TEST: 02/13/2018 23:13:58

Diagnostics Check		OK	23:01	
Air Blank		0.000	23:02	
Control Test		0.078	23:02	
Air Blank		0.000	23:03	
Subject Sample	# 1		23:06	No Sample Provided
Breath Volume		0.000	23:06	
Air Blank		0.000	23:06	
Air Blank		0.000	23:08	
Subject Sample	# 2	0.018	23:09	
Breath Volume		3.656	23:09	
Air Blank		0.000	23:10	
Air Blank		0.000	23:12	
Subject Sample	# 3	0.016	23:12	
Breath Volume		3.300	23:12	
Air Blank		0.000	23:13	
Control Test		0.077	23:13	
Air Blank		0.000	23:13	
Diagnostics Check		OK	23:13	

Data Download Date/Time: 02/16/2018 14:39

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001322

INSTRUMENT REGISTERED TO: OKEECHOBEE P.D.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MARGERUM,KELLEY,J
BREATH TEST OPERATOR AGENCY: Okeechobee PD
ARRESTING OFFICER: MARGERUM,KELLEY
ARRESTING OFFICER AGENCY: Okeechobee PD
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 04216080A3
DRY GAS EXPIRY DATE: 03/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:06
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: CLARK

DATE/TIME OF TEST: 02/18/2018 03:39:34

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:31
Subject Sample	# 1	0.156	03:33
Breath Volume		1.867	03:33
Air Blank		0.000	03:34
Air Blank		0.000	03:36
Subject Sample	# 2	0.154	03:37
Breath Volume		1.921	03:37
Air Blank		0.000	03:38
Control Test		0.077	03:38
Air Blank		0.000	03:39
Diagnostics Check		OK	03:39

Data Download Date/Time: 04/10/2018 14:09

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GENTRY,JOSHUA,J
BREATH TEST OPERATOR AGENCY: FHP Troop L
ARRESTING OFFICER: GENTRY,JOSHUA
ARRESTING OFFICER AGENCY: FHP Troop L
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: MANSFIELD

DATE/TIME OF TEST: 02/17/2018 02:07:50

Diagnostics Check		OK	02:01
Air Blank		0.000	02:01
Control Test		0.079	02:01
Air Blank		0.000	02:02
Subject Sample	# 1	0.137	02:03
Breath Volume		1.777	02:03
Air Blank		0.000	02:03
Air Blank		0.000	02:05
Subject Sample	# 2	0.135	02:06
Breath Volume		2.906	02:06
Air Blank		0.000	02:06
Control Test		0.078	02:07
Air Blank		0.000	02:07
Diagnostics Check		OK	02:07

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GETCHELL,E
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:10

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/17/2018 21:34:13

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:31	
Control Test		0.080	21:31	
Air Blank		0.000	21:32	
Subject Sample	# 1		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:33	
Control Test		0.080	21:33	
Air Blank		0.000	21:34	
Diagnostics Check		OK	21:34	

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIMAVERA,S,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BAUER,B
ARRESTING OFFICER AGENCY: Sebastian PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: HEARL

DATE/TIME OF TEST: 02/18/2018 01:49:29

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.080	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:48	Subject Test Refused
Breath Volume		0.000	01:48	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GIST,SHARON,D
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GIBSON,KATHERYN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 13:43
GENDER: Female
OBSERVATION PERIOD BEGAN: 14:55

SUBJECT'S LAST NAME: SUTHARD

DATE/TIME OF TEST: 02/23/2018 15:44:25

Diagnostics Check		OK	15:38
Air Blank		0.000	15:38
Control Test		0.078	15:38
Air Blank		0.000	15:39
Subject Sample	# 1	0.242	15:39
Breath Volume		1.472	15:39
Air Blank		0.000	15:40
Air Blank		0.000	15:42
Subject Sample	# 2	0.246	15:42
Breath Volume		1.503	15:42
Air Blank		0.000	15:43
Control Test		0.078	15:43
Air Blank		0.000	15:44
Diagnostics Check		OK	15:44

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROWLEY,SEAN,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: DELGADO,JOSE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 17:57

SUBJECT'S LAST NAME: MCMAHAN

DATE/TIME OF TEST: 02/23/2018 20:04:55

Diagnostics Check		OK	19:58
Air Blank		0.000	19:58
Control Test		0.079	19:59
Air Blank		0.000	19:59
Subject Sample	# 1	0.173	20:00
Breath Volume		2.843	20:00
Air Blank		0.000	20:00
Air Blank		0.000	20:02
Subject Sample	# 2	0.176	20:03
Breath Volume		2.914	20:03
Air Blank		0.000	20:04
Control Test		0.078	20:04
Air Blank		0.000	20:04
Diagnostics Check		OK	20:04

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KANTORSKI,MARK,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOHNSON,ANDREW
ARRESTING OFFICER AGENCY: Indian River Shores
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:40

SUBJECT'S LAST NAME: LIMORIN

DATE/TIME OF TEST: 02/24/2018 02:53:05

Diagnostics Check		OK	02:46
Air Blank		0.000	02:46
Control Test		0.079	02:47
Air Blank		0.000	02:47
Subject Sample	# 1	0.000	02:48
Breath Volume		2.027	02:48
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.000	02:51
Breath Volume		2.500	02:51
Air Blank		0.000	02:52
Control Test		0.080	02:52
Air Blank		0.000	02:53
Diagnostics Check		OK	02:53

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EGGERS,RYAN,K
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:30

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/24/2018 22:11:53

Diagnostics Check		OK	22:05
Air Blank		0.000	22:05
Control Test		0.079	22:06
Air Blank		0.000	22:06
Subject Sample	# 1	0.000	22:07
Breath Volume		2.156	22:07
Air Blank		0.000	22:08
Air Blank		0.000	22:10
Subject Sample	# 2	0.000	22:10
Breath Volume		2.746	22:10
Air Blank		0.000	22:10
Control Test		0.078	22:11
Air Blank		0.000	22:11
Diagnostics Check		OK	22:11

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 02/25/2018 22:38:11

Diagnostics Check		OK	22:33	
Air Blank		0.000	22:33	
Control Test		0.079	22:33	
Air Blank		0.000	22:34	
Subject Sample	# 1	0.199	22:34	
Breath Volume		2.851	22:34	
Air Blank		0.000	22:35	
Air Blank		0.000	22:37	
Subject Sample	# 2	0.188	22:37	
Breath Volume		2.867	22:37	
Air Blank			22:38	RFI Detect
Air Blank			22:38	RFI Detect

Data Download Date/Time: 03/23/2018 05:49

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:35

SUBJECT'S LAST NAME: KEMP

DATE/TIME OF TEST: 02/25/2018 22:49:14

Diagnostics Check		OK	22:42
Air Blank		0.000	22:43
Control Test		0.078	22:43
Air Blank		0.000	22:44
Subject Sample	# 1	0.190	22:44
Breath Volume		1.855	22:44
Air Blank		0.000	22:45
Air Blank		0.000	22:47
Subject Sample	# 2	0.192	22:47
Breath Volume		3.300	22:47
Air Blank		0.000	22:48
Control Test		0.079	22:48
Air Blank		0.000	22:49
Diagnostics Check		OK	22:49

Data Download Date/Time: 03/23/2018 05:49

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BRASHEARS,ERIC
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GUTIERREZ MERAZ

DATE/TIME OF TEST: 02/26/2018 02:23:43

Diagnostics Check		OK	02:12	
Air Blank		0.000	02:12	
Control Test		0.079	02:13	
Air Blank		0.000	02:13	
Subject Sample	# 1		02:16	No Sample Provided
Breath Volume		0.000	02:16	
Air Blank		0.000	02:17	
Air Blank		0.000	02:19	
Subject Sample	# 2		02:22	No Sample Provided
Breath Volume		0.000	02:22	
Air Blank		0.000	02:22	
Control Test		0.079	02:23	
Air Blank		0.000	02:23	
Diagnostics Check		OK	02:23	

Data Download Date/Time: 03/23/2018 05:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GREEN,BRIAN,P
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: BRASHEARS,ERIC
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:51

SUBJECT'S LAST NAME: GUTIERREZ MERAZ

DATE/TIME OF TEST: 02/26/2018 02:34:37

Diagnostics Check		OK	02:28
Air Blank		0.000	02:29
Control Test		0.079	02:29
Air Blank		0.000	02:29
Subject Sample	# 1	0.141	02:30
Breath Volume		3.324	02:30
Air Blank		0.000	02:30
Air Blank		0.000	02:32
Subject Sample	# 2	0.143	02:33
Breath Volume		3.359	02:33
Air Blank		0.000	02:33
Control Test		0.078	02:34
Air Blank		0.000	02:34
Diagnostics Check		OK	02:34

Data Download Date/Time: 03/23/2018 05:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,C,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: RUIZ,M
ARRESTING OFFICER AGENCY: Fellsmere PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 18:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:50

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/26/2018 19:21:33

Diagnostics Check		OK	19:18	
Air Blank		0.000	19:19	
Control Test		0.080	19:19	
Air Blank		0.000	19:20	
Subject Sample	# 1		19:20	Subject Test Refused
Breath Volume		0.000	19:20	
Air Blank		0.000	19:20	
Control Test		0.078	19:20	
Air Blank		0.000	19:21	
Diagnostics Check		OK	19:21	

Data Download Date/Time: 03/23/2018 05:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001328

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JONES,JASON,B
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: JONES,JASON
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: BUI
ARREST TIME: 22:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:05

SUBJECT'S LAST NAME: CONSTANTINE

DATE/TIME OF TEST: 02/27/2018 01:18:16

Diagnostics Check		OK	01:15	
Air Blank		0.000	01:15	
Control Test		0.079	01:16	
Air Blank		0.000	01:16	
Subject Sample	# 1		01:16	Subject Test Refused
Breath Volume		0.000	01:16	
Air Blank		0.000	01:17	
Control Test		0.079	01:17	
Air Blank		0.000	01:18	
Diagnostics Check		OK	01:18	

Data Download Date/Time: 03/23/2018 05:50

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CROWLEY,SEAN,W
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: FONSECA,GIOVANNI
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 21:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:40

SUBJECT'S LAST NAME: HERNANDEZ

DATE/TIME OF TEST: 02/01/2018 22:19:17

Diagnostics Check		OK	22:16	
Air Blank		0.000	22:16	
Control Test		0.081	22:17	
Air Blank		0.000	22:17	
Subject Sample	# 1		22:17	Subject Test Refused
Breath Volume		0.000	22:17	
Air Blank		0.000	22:18	
Control Test		0.081	22:18	
Air Blank		0.000	22:19	
Diagnostics Check		OK	22:19	

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,BRYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:32

SUBJECT'S LAST NAME: RODRIGUEZ

DATE/TIME OF TEST: 02/02/2018 02:01:13

Diagnostics Check		OK	01:53
Air Blank		0.000	01:53
Control Test		0.081	01:54
Air Blank		0.000	01:54
Subject Sample	# 1	0.024	01:55
Breath Volume		2.406	01:55
Air Blank		0.000	01:56
Air Blank		0.000	01:58
Subject Sample	# 2	0.022	01:59
Breath Volume		1.414	01:59
Air Blank		0.000	02:00
Control Test		0.081	02:00
Air Blank		0.000	02:01
Diagnostics Check		OK	02:01

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HATTON,NICKOLES
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:02

SUBJECT'S LAST NAME: WILLIAMS

DATE/TIME OF TEST: 02/03/2018 00:35:55

Diagnostics Check		OK	00:30	
Air Blank		0.000	00:31	
Control Test		0.081	00:31	
Air Blank		0.000	00:32	
Subject Sample	# 1		00:34	Subject Test Refused
Breath Volume		0.000	00:34	
Air Blank		0.000	00:35	
Control Test		0.082	00:35	
Air Blank		0.000	00:35	
Diagnostics Check		OK	00:35	

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PLUMB,GREGORY,T
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: PLUMB,GREGORY
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: TECKENBROCK

DATE/TIME OF TEST: 02/04/2018 02:16:29

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.080	02:10
Air Blank		0.000	02:11
Subject Sample	# 1	0.121	02:11
Breath Volume		1.800	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.128	02:14
Breath Volume		2.167	02:14
Air Blank		0.000	02:15
Control Test		0.081	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WOLSKI,JEFFREY,R
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: AVILA,L
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 19:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:35

SUBJECT'S LAST NAME: HENSON

DATE/TIME OF TEST: 02/05/2018 20:13:18

Diagnostics Check		OK	20:07
Air Blank		0.000	20:07
Control Test		0.081	20:07
Air Blank		0.000	20:08
Subject Sample	# 1	0.293	20:08
Breath Volume		1.933	20:08
Air Blank		0.000	20:09
Air Blank		0.000	20:11
Subject Sample	# 2	0.297	20:11
Breath Volume		1.847	20:11
Air Blank		0.000	20:12
Control Test		0.081	20:12
Air Blank		0.000	20:13
Diagnostics Check		OK	20:13

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIMAVERA,SEAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRICKE,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: MANGUM

DATE/TIME OF TEST: 02/07/2018 03:31:57

Diagnostics Check		OK	03:22	
Air Blank		0.000	03:22	
Control Test		0.081	03:22	
Air Blank		0.000	03:23	
Subject Sample	# 1		03:26	No Sample Provided
Breath Volume		0.000	03:26	
Air Blank		0.000	03:26	
Air Blank		0.000	03:28	
Subject Sample	# 2		03:30	Subject Test Refused
Breath Volume		0.000	03:30	
Air Blank		0.000	03:31	
Control Test		0.082	03:31	
Air Blank		0.000	03:31	
Diagnostics Check		OK	03:31	

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: EVANS,LEE,A
BREATH TEST OPERATOR AGENCY: Vero Beach PD
ARRESTING OFFICER: EVANS,LEE
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 13:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 13:36

SUBJECT'S LAST NAME: JAKUBOWSKI

DATE/TIME OF TEST: 02/07/2018 14:12:31

Diagnostics Check		OK	14:00	
Air Blank		0.000	14:01	
Control Test		0.081	14:01	
Air Blank		0.000	14:01	
Subject Sample	# 1	0.281	14:05	Volume Not Met
Breath Volume		0.121	14:05	
Air Blank		0.000	14:05	
Air Blank		0.000	14:07	
Subject Sample	# 2	0.260	14:10	Volume Not Met
Breath Volume		0.101	14:10	
Air Blank		0.000	14:11	
Control Test		0.079	14:11	
Air Blank		0.000	14:12	
Diagnostics Check		OK	14:12	

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIMAVERA,SEAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:00

SUBJECT'S LAST NAME: MEEK

DATE/TIME OF TEST: 02/08/2018 03:35:31

Diagnostics Check		OK	03:32	
Air Blank		0.000	03:33	
Control Test		0.080	03:33	
Air Blank		0.000	03:33	
Subject Sample	# 1		03:34	Subject Test Refused
Breath Volume		0.000	03:34	
Air Blank		0.000	03:34	
Control Test		0.081	03:34	
Air Blank		0.000	03:35	
Diagnostics Check		OK	03:35	

Data Download Date/Time: 02/09/2018 06:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: BRASHEARS,ERICK
ARRESTING OFFICER AGENCY: Vero Beach PD
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:55
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/10/2018 03:47:51

Diagnostics Check		OK	03:41	
Air Blank		0.000	03:41	
Control Test		0.080	03:42	
Air Blank		0.000	03:42	
Subject Sample	# 1	0.162	03:43	
Breath Volume		1.105	03:43	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 2	0.169	03:47	
Breath Volume		1.285	03:47	
Air Blank			03:47	RFI Detect
Air Blank		0.000	03:47	

Data Download Date/Time: 03/23/2018 05:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: JOERGER,SHAWN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 04:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:48

SUBJECT'S LAST NAME: SNEED

DATE/TIME OF TEST: 02/10/2018 05:24:06

Diagnostics Check		OK	05:17
Air Blank		0.000	05:17
Control Test		0.080	05:18
Air Blank		0.000	05:18
Subject Sample	# 1	0.000	05:19
Breath Volume		1.906	05:19
Air Blank		0.000	05:20
Air Blank		0.000	05:22
Subject Sample	# 2	0.000	05:22
Breath Volume		2.031	05:22
Air Blank		0.000	05:23
Control Test		0.080	05:23
Air Blank		0.000	05:24
Diagnostics Check		OK	05:24

Data Download Date/Time: 03/23/2018 05:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 03:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: CLARKSON

DATE/TIME OF TEST: 02/11/2018 04:35:37

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.080	04:29
Air Blank		0.000	04:29
Subject Sample	# 1	0.165	04:30
Breath Volume		2.167	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.160	04:34
Breath Volume		1.199	04:34
Air Blank		0.000	04:34
Control Test		0.080	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 03/23/2018 05:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HOLIDAY,LATONYA,A
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRICKE,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 02:24
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: VOTZI

DATE/TIME OF TEST: 02/12/2018 03:21:42

Diagnostics Check		OK	03:18	
Air Blank		0.000	03:19	
Control Test		0.081	03:19	
Air Blank		0.000	03:20	
Subject Sample	# 1		03:20	Subject Test Refused
Breath Volume		0.000	03:20	
Air Blank		0.000	03:20	
Control Test		0.080	03:21	
Air Blank		0.000	03:21	
Diagnostics Check		OK	03:21	

Data Download Date/Time: 03/23/2018 05:42

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WHILE,CONCETTA,J
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: HARTMANN,R
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 17:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:28

SUBJECT'S LAST NAME: SCAGLIONE

DATE/TIME OF TEST: 02/13/2018 19:01:09

Diagnostics Check		OK	18:58	
Air Blank		0.000	18:58	
Control Test		0.082	18:58	
Air Blank		0.000	18:59	
Subject Sample	# 1		18:59	Subject Test Refused
Breath Volume		0.000	18:59	
Air Blank		0.000	19:00	
Control Test		0.082	19:00	
Air Blank		0.000	19:01	
Diagnostics Check		OK	19:01	

Data Download Date/Time: 03/23/2018 05:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PRIMAVERA,SEAN,
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: FRICKE,J
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 01:04
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: BARBATO

DATE/TIME OF TEST: 02/15/2018 01:50:24

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:47	
Control Test		0.081	01:47	
Air Blank		0.000	01:48	
Subject Sample	# 1		01:49	Subject Test Refused
Breath Volume		0.000	01:49	
Air Blank		0.000	01:49	
Control Test		0.082	01:49	
Air Blank		0.000	01:50	
Diagnostics Check		OK	01:50	

Data Download Date/Time: 03/23/2018 05:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: EGGERS,RYAN
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 20:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: MCCALL

DATE/TIME OF TEST: 02/16/2018 00:25:47

Diagnostics Check		OK	00:19
Air Blank		0.000	00:19
Control Test		0.081	00:20
Air Blank		0.000	00:20
Subject Sample	# 1	0.000	00:21
Breath Volume		1.679	00:21
Air Blank		0.000	00:21
Air Blank		0.000	00:23
Subject Sample	# 2	0.000	00:24
Breath Volume		1.921	00:24
Air Blank		0.000	00:24
Control Test		0.081	00:25
Air Blank		0.000	00:25
Diagnostics Check		OK	00:25

Data Download Date/Time: 03/23/2018 05:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001329

INSTRUMENT REGISTERED TO: INDIAN RIVER CO. SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RAYMOND,FRANK,W
BREATH TEST OPERATOR AGENCY: Indian River Co SO
ARRESTING OFFICER: GETCHELL,ERIC
ARRESTING OFFICER AGENCY: Indian River Co SO
LAST AGENCY INSPECTION: 02/09/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 684131
DRY GAS EXPIRY DATE: 01/23/2020
VIOLATION CODE: DUI
ARREST TIME: 23:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: ROMER

DATE/TIME OF TEST: 02/16/2018 02:25:44

Diagnostics Check		OK	02:22	
Air Blank		0.000	02:22	
Control Test		0.082	02:23	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:24	
Control Test		0.081	02:25	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 03/23/2018 05:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASSEY,DARRELL,L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PAYNTER,JARED
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 21:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:45

SUBJECT'S LAST NAME: LOPEZ

DATE/TIME OF TEST: 02/04/2018 22:34:00

Diagnostics Check		OK	22:27
Air Blank		0.000	22:28
Control Test		0.081	22:28
Air Blank		0.000	22:28
Subject Sample	# 1	0.406	22:29
Breath Volume		2.937	22:29
Air Blank		0.000	22:30
Air Blank		0.000	22:31
Subject Sample	# 2	0.401	22:32
Breath Volume		2.523	22:32
Air Blank		0.000	22:33
Control Test		0.080	22:33
Air Blank		0.000	22:33
Diagnostics Check		OK	22:34

Data Download Date/Time: 03/06/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DITTMAN,LAVERNE,T
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: DITTMAN,LAVERNE
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 01/24/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: PROB/PAROL
ARREST TIME: 18:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:45

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 02/22/2018 19:13:22

Diagnostics Check		OK	19:07
Air Blank		0.000	19:08
Control Test		0.081	19:08
Air Blank		0.000	19:08
Subject Sample	# 1	0.000	19:09
Breath Volume		4.410	19:09
Air Blank		0.000	19:09
Air Blank		0.000	19:11
Subject Sample	# 2	0.000	19:12
Breath Volume		3.847	19:12
Air Blank		0.000	19:12
Control Test		0.080	19:12
Air Blank		0.000	19:13
Diagnostics Check		OK	19:13

Data Download Date/Time: 03/06/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001341

INSTRUMENT REGISTERED TO: DESOTO COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CAUDILL,CHERI,L
BREATH TEST OPERATOR AGENCY: DeSoto County SO
ARRESTING OFFICER: PACHECO,ERIQ
ARRESTING OFFICER AGENCY: DeSoto County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 31517080A1
DRY GAS EXPIRY DATE: 01/05/2020
VIOLATION CODE: DUI
ARREST TIME: 02:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: ZAVALA

DATE/TIME OF TEST: 02/26/2018 03:35:22

Diagnostics Check		OK	03:27
Air Blank		0.000	03:28
Control Test		0.081	03:28
Air Blank		0.000	03:28
Subject Sample	# 1	0.188	03:30
Breath Volume		2.253	03:30
Air Blank		0.000	03:31
Air Blank		0.000	03:32
Subject Sample	# 2	0.181	03:33
Breath Volume		1.707	03:33
Air Blank		0.000	03:34
Control Test		0.080	03:34
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 03/06/2018 09:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:23
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: JORDAN

DATE/TIME OF TEST: 02/23/2018 22:31:26

Diagnostics Check		OK	22:25
Air Blank		0.000	22:25
Control Test		0.078	22:26
Air Blank		0.000	22:26
Subject Sample	# 1	0.193	22:26
Breath Volume		1.621	22:26
Air Blank		0.000	22:27
Air Blank		0.000	22:29
Subject Sample	# 2	0.187	22:29
Breath Volume		1.613	22:29
Air Blank		0.000	22:30
Control Test		0.077	22:30
Air Blank		0.000	22:31
Diagnostics Check		OK	22:31

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:56

SUBJECT'S LAST NAME: LEWIS

DATE/TIME OF TEST: 02/23/2018 22:41:50

Diagnostics Check		OK	22:35
Air Blank		0.000	22:35
Control Test		0.078	22:36
Air Blank		0.000	22:36
Subject Sample	# 1	0.118	22:37
Breath Volume		2.566	22:37
Air Blank		0.000	22:38
Air Blank		0.000	22:39
Subject Sample	# 2	0.109	22:40
Breath Volume		1.937	22:40
Air Blank		0.000	22:40
Control Test		0.078	22:41
Air Blank		0.000	22:41
Diagnostics Check		OK	22:41

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BUEHLER,DEP,M
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BUEHLER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:08

SUBJECT'S LAST NAME: BROWN

DATE/TIME OF TEST: 02/23/2018 22:52:59

Diagnostics Check		OK	22:46
Air Blank		0.000	22:46
Control Test		0.077	22:46
Air Blank		0.000	22:47
Subject Sample	# 1	0.133	22:48
Breath Volume		3.609	22:48
Air Blank		0.000	22:48
Air Blank		0.000	22:50
Subject Sample	# 2	0.129	22:51
Breath Volume		3.667	22:51
Air Blank		0.000	22:52
Control Test		0.077	22:52
Air Blank		0.000	22:52
Diagnostics Check		OK	22:52

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: BALES,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:13
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:29

SUBJECT'S LAST NAME: GODFREY

DATE/TIME OF TEST: 02/24/2018 03:01:50

Diagnostics Check		OK	02:55
Air Blank		0.000	02:56
Control Test		0.080	02:56
Air Blank		0.000	02:56
Subject Sample	# 1	0.115	02:57
Breath Volume		3.230	02:57
Air Blank		0.000	02:58
Air Blank		0.000	02:59
Subject Sample	# 2	0.116	03:00
Breath Volume		3.593	03:00
Air Blank		0.000	03:00
Control Test		0.080	03:01
Air Blank		0.000	03:01
Diagnostics Check		OK	03:01

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MIKLUSCAK,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: ROYAL GORDALIZA

DATE/TIME OF TEST: 02/24/2018 23:17:40

Diagnostics Check		OK	23:11
Air Blank		0.000	23:11
Control Test		0.079	23:12
Air Blank		0.000	23:12
Subject Sample	# 1	0.138	23:13
Breath Volume		1.398	23:13
Air Blank		0.000	23:13
Air Blank		0.000	23:15
Subject Sample	# 2	0.136	23:16
Breath Volume		2.742	23:16
Air Blank		0.000	23:16
Control Test		0.080	23:17
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SILVA,OFFICER
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:03
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: MORRISON

DATE/TIME OF TEST: 02/25/2018 02:40:39

Diagnostics Check		OK	02:37	
Air Blank		0.000	02:38	
Control Test		0.080	02:38	
Air Blank		0.000	02:39	
Subject Sample	# 1		02:39	Subject Test Refused
Breath Volume		0.000	02:39	
Air Blank		0.000	02:39	
Control Test		0.080	02:40	
Air Blank		0.000	02:40	
Diagnostics Check		OK	02:40	

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: DONALD,TROOPER
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:14
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:24

SUBJECT'S LAST NAME: BOOS

DATE/TIME OF TEST: 02/25/2018 21:13:46

Diagnostics Check		OK	21:07
Air Blank		0.000	21:07
Control Test		0.079	21:07
Air Blank		0.000	21:08
Subject Sample	# 1	0.134	21:09
Breath Volume		1.500	21:09
Air Blank		0.000	21:09
Air Blank		0.000	21:11
Subject Sample	# 2	0.145	21:12
Breath Volume		1.152	21:12
Air Blank		0.000	21:12
Control Test		0.079	21:13
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MIKLUSCAK,OFFICER
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:18

SUBJECT'S LAST NAME: TRAMILL

DATE/TIME OF TEST: 02/26/2018 02:49:13

Diagnostics Check		OK	02:42
Air Blank		0.000	02:43
Control Test		0.080	02:43
Air Blank		0.000	02:44
Subject Sample	# 1	0.151	02:44
Breath Volume		3.195	02:44
Air Blank		0.000	02:45
Air Blank		0.000	02:47
Subject Sample	# 2	0.146	02:47
Breath Volume		2.906	02:47
Air Blank		0.000	02:48
Control Test		0.079	02:48
Air Blank		0.000	02:49
Diagnostics Check		OK	02:49

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIE,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: GREEN,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:21
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:11

SUBJECT'S LAST NAME: MATTOS

DATE/TIME OF TEST: 02/27/2018 22:06:25

Diagnostics Check		OK	22:00
Air Blank		0.000	22:00
Control Test		0.081	22:01
Air Blank		0.000	22:01
Subject Sample	# 1	0.162	22:01
Breath Volume		2.023	22:01
Air Blank		0.000	22:02
Air Blank		0.000	22:04
Subject Sample	# 2	0.161	22:04
Breath Volume		1.421	22:04
Air Blank		0.000	22:05
Control Test		0.079	22:05
Air Blank		0.000	22:06
Diagnostics Check		OK	22:06

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001344

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FLYNN,MICHAEL,T
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MIKLUSAK,Z
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 02/23/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:26
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:00

SUBJECT'S LAST NAME: SHANNON

DATE/TIME OF TEST: 02/28/2018 21:35:39

Diagnostics Check		OK	21:29
Air Blank		0.000	21:29
Control Test		0.079	21:30
Air Blank		0.000	21:30
Subject Sample	# 1	0.124	21:31
Breath Volume		1.656	21:31
Air Blank		0.000	21:31
Air Blank		0.000	21:33
Subject Sample	# 2	0.114	21:34
Breath Volume		1.371	21:34
Air Blank		0.000	21:34
Control Test		0.079	21:35
Air Blank		0.000	21:35
Diagnostics Check		OK	21:35

Data Download Date/Time: 03/29/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001346

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SCHUMANN,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SCHUMANN,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:46
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:47

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/21/2018 02:13:37

Diagnostics Check		OK	02:07
Air Blank		0.000	02:07
Control Test		0.079	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.141	02:09
Breath Volume		2.183	02:09
Air Blank		0.000	02:09
Air Blank		0.000	02:11
Subject Sample	# 2	0.140	02:12
Breath Volume		1.503	02:12
Air Blank		0.000	02:12
Control Test		0.079	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:13

Data Download Date/Time: 02/23/2018 07:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: PERRY,M
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:08
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:03

SUBJECT'S LAST NAME: PEREZ GARCIA

DATE/TIME OF TEST: 02/01/2018 04:00:23

Diagnostics Check		OK	03:54
Air Blank		0.000	03:54
Control Test		0.079	03:55
Air Blank		0.000	03:55
Subject Sample	# 1	0.072	03:56
Breath Volume		2.546	03:56
Air Blank		0.000	03:56
Air Blank		0.000	03:58
Subject Sample	# 2	0.072	03:58
Breath Volume		2.652	03:58
Air Blank		0.000	03:59
Control Test		0.079	03:59
Air Blank		0.000	04:00
Diagnostics Check		OK	04:00

Data Download Date/Time: 02/23/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SURRAN,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: TUCK,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 14:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:09

SUBJECT'S LAST NAME: SHUBACK

DATE/TIME OF TEST: 02/01/2018 16:36:49

Diagnostics Check		OK	16:30
Air Blank		0.000	16:31
Control Test		0.079	16:31
Air Blank		0.000	16:32
Subject Sample	# 1	0.128	16:32
Breath Volume		1.375	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:34
Subject Sample	# 2	0.125	16:35
Breath Volume		2.160	16:35
Air Blank		0.000	16:35
Control Test		0.078	16:36
Air Blank		0.000	16:36
Diagnostics Check		OK	16:36

Data Download Date/Time: 02/23/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTIN,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:15

SUBJECT'S LAST NAME: PION

DATE/TIME OF TEST: 02/01/2018 23:47:53

Diagnostics Check		OK	23:41
Air Blank		0.000	23:42
Control Test		0.078	23:42
Air Blank		0.000	23:42
Subject Sample	# 1	0.162	23:43
Breath Volume		4.250	23:43
Air Blank		0.000	23:43
Air Blank		0.000	23:45
Subject Sample	# 2	0.169	23:46
Breath Volume		3.445	23:46
Air Blank		0.000	23:47
Control Test		0.078	23:47
Air Blank		0.000	23:47
Diagnostics Check		OK	23:47

Data Download Date/Time: 02/23/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MELENDEZ-BONILLA,RONALD,
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: MELENDEZ-BONILLA,RONALD
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:25

SUBJECT'S LAST NAME: SUTTON

DATE/TIME OF TEST: 02/02/2018 02:08:49

Diagnostics Check		OK	02:02
Air Blank		0.000	02:03
Control Test		0.078	02:03
Air Blank		0.000	02:03
Subject Sample	# 1	0.088	02:04
Breath Volume		2.218	02:04
Air Blank		0.000	02:05
Air Blank		0.000	02:06
Subject Sample	# 2	0.087	02:07
Breath Volume		2.531	02:07
Air Blank		0.000	02:07
Control Test		0.078	02:08
Air Blank		0.000	02:08
Diagnostics Check		OK	02:08

Data Download Date/Time: 02/23/2018 07:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:52

SUBJECT'S LAST NAME: MILTON

DATE/TIME OF TEST: 02/03/2018 00:19:43

Diagnostics Check		OK	00:12
Air Blank		0.000	00:13
Control Test		0.077	00:13
Air Blank		0.000	00:14
Subject Sample	# 1	0.209	00:14
Breath Volume		1.515	00:14
Air Blank		0.000	00:15
Air Blank		0.000	00:16
Subject Sample	# 2	0.220	00:18
Breath Volume		2.257	00:18
Air Blank		0.000	00:18
Control Test		0.079	00:19
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,NICKKI,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VERMILLION,B
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:06

SUBJECT'S LAST NAME: PALMA

DATE/TIME OF TEST: 02/03/2018 03:48:05

Diagnostics Check		OK	03:40
Air Blank		0.000	03:41
Control Test		0.079	03:41
Air Blank		0.000	03:42
Subject Sample	# 1	0.091	03:42
Breath Volume		1.183	03:42
Air Blank		0.000	03:43
Air Blank		0.000	03:45
Subject Sample	# 2	0.091	03:46
Breath Volume		2.796	03:46
Air Blank		0.000	03:47
Control Test		0.079	03:47
Air Blank		0.000	03:47
Diagnostics Check		OK	03:48

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MORALES,R
ARRESTING OFFICER AGENCY: Venice PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:36

SUBJECT'S LAST NAME: SANTIAGO ZAMMORANO

DATE/TIME OF TEST: 02/03/2018 22:14:14

Diagnostics Check		OK	22:01
Air Blank		0.000	22:02
Control Test		0.079	22:02
Air Blank		0.000	22:03
Subject Sample	# 1	0.229	22:04
Breath Volume		1.167	22:04
Air Blank		0.000	22:05
Air Blank		0.000	22:06
Subject Sample	# 2	0.199	22:09
Breath Volume		1.417	22:09
Air Blank		0.000	22:10
Air Blank		0.000	22:12
Subject Sample	# 3	0.221	22:12
Breath Volume		2.378	22:12
Air Blank		0.000	22:13
Control Test		0.077	22:13
Air Blank		0.000	22:14
Diagnostics Check		OK	22:14

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VERMILLION,B
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:45
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:54

SUBJECT'S LAST NAME: HILL

DATE/TIME OF TEST: 02/04/2018 20:38:19

Diagnostics Check		OK	20:32	
Air Blank		0.000	20:33	
Control Test		0.078	20:33	
Air Blank		0.000	20:34	
Subject Sample	# 1		20:36	Subject Test Refused
Breath Volume		0.000	20:36	
Air Blank		0.000	20:37	
Control Test		0.079	20:37	
Air Blank		0.000	20:38	
Diagnostics Check		OK	20:38	

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:58
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: GOINS

DATE/TIME OF TEST: 02/10/2018 00:53:19

Diagnostics Check		OK	00:47
Air Blank		0.000	00:47
Control Test		0.079	00:47
Air Blank		0.000	00:48
Subject Sample	# 1	0.066	00:48
Breath Volume		2.609	00:48
Air Blank		0.000	00:49
Air Blank		0.000	00:51
Subject Sample	# 2	0.066	00:51
Breath Volume		3.472	00:51
Air Blank		0.000	00:52
Control Test		0.078	00:52
Air Blank		0.000	00:53
Diagnostics Check		OK	00:53

Data Download Date/Time: 02/23/2018 07:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MENDEZ,DEPUTY,C
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MCCORMICK,DEPUTY
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:50

SUBJECT'S LAST NAME: MICHENER

DATE/TIME OF TEST: 02/10/2018 05:29:46

Diagnostics Check		OK	05:21	
Air Blank		0.000	05:21	
Control Test		0.079	05:22	
Air Blank		0.000	05:22	
Subject Sample	# 1	0.213	05:23	
Breath Volume		1.585	05:23	
Air Blank		0.000	05:23	
Air Blank		0.000	05:25	
Subject Sample	# 2		05:28	Subject Test Refused
Breath Volume		0.000	05:28	
Air Blank		0.000	05:28	
Control Test		0.077	05:29	
Air Blank		0.000	05:29	
Diagnostics Check		OK	05:29	

Data Download Date/Time: 02/23/2018 07:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:55

SUBJECT'S LAST NAME: STREETER-FOLLANO

DATE/TIME OF TEST: 02/10/2018 20:25:54

Diagnostics Check		OK	20:18
Air Blank		0.000	20:19
Control Test		0.078	20:19
Air Blank		0.000	20:20
Subject Sample	# 1	0.227	20:20
Breath Volume		1.167	20:20
Air Blank		0.000	20:21
Air Blank		0.000	20:23
Subject Sample	# 2	0.215	20:24
Breath Volume		1.890	20:24
Air Blank		0.000	20:24
Control Test		0.078	20:25
Air Blank		0.000	20:25
Diagnostics Check		OK	20:25

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FOWLER,TYLER,J
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: FOWLER,TYLER
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:49

SUBJECT'S LAST NAME: SAMBLANET

DATE/TIME OF TEST: 02/11/2018 02:15:54

Diagnostics Check		OK	02:09
Air Blank		0.000	02:10
Control Test		0.078	02:10
Air Blank		0.000	02:10
Subject Sample	# 1	0.133	02:11
Breath Volume		2.546	02:11
Air Blank		0.000	02:12
Air Blank		0.000	02:14
Subject Sample	# 2	0.131	02:14
Breath Volume		2.753	02:14
Air Blank		0.000	02:15
Control Test		0.077	02:15
Air Blank		0.000	02:15
Diagnostics Check		OK	02:15

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NELSON,HALSTON,L
BREATH TEST OPERATOR AGENCY: North Port PD
ARRESTING OFFICER: COCHRAN,RACHEL
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:11
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:28

SUBJECT'S LAST NAME: HARVEY

DATE/TIME OF TEST: 02/11/2018 21:13:26

Diagnostics Check		OK	21:03
Air Blank		0.000	21:04
Control Test		0.079	21:04
Air Blank		0.000	21:05
Subject Sample	# 1	0.112	21:08
Breath Volume		1.789	21:08
Air Blank		0.000	21:08
Air Blank		0.000	21:10
Subject Sample	# 2	0.107	21:11
Breath Volume		1.847	21:11
Air Blank		0.000	21:12
Control Test		0.076	21:12
Air Blank		0.000	21:13
Diagnostics Check		OK	21:13

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: SUNDERLAND,OFC
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:15

SUBJECT'S LAST NAME: HODNE

DATE/TIME OF TEST: 02/16/2018 21:36:07

Diagnostics Check		OK	21:29
Air Blank		0.000	21:30
Control Test		0.080	21:30
Air Blank		0.000	21:31
Subject Sample	# 1	0.132	21:31
Breath Volume		2.484	21:31
Air Blank		0.000	21:32
Air Blank		0.000	21:34
Subject Sample	# 2	0.135	21:34
Breath Volume		3.109	21:34
Air Blank		0.000	21:35
Control Test		0.079	21:35
Air Blank		0.000	21:36
Diagnostics Check		OK	21:36

Data Download Date/Time: 02/23/2018 07:48

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,DANIEL,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VERMILLION,OFC
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:17

SUBJECT'S LAST NAME: PITTS

DATE/TIME OF TEST: 02/17/2018 02:56:44

Diagnostics Check		OK	02:50
Air Blank		0.000	02:51
Control Test		0.078	02:51
Air Blank		0.000	02:52
Subject Sample	# 1	0.000	02:52
Breath Volume		3.242	02:52
Air Blank		0.000	02:53
Air Blank		0.000	02:55
Subject Sample	# 2	0.000	02:55
Breath Volume		1.921	02:55
Air Blank		0.000	02:55
Control Test		0.078	02:56
Air Blank		0.000	02:56
Diagnostics Check		OK	02:56

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LUBRANO,DEP,
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MAYER,DEP
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:41
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: BRIGHT

DATE/TIME OF TEST: 02/18/2018 04:57:57

Diagnostics Check		OK	04:47
Air Blank		0.000	04:47
Control Test		0.078	04:48
Air Blank		0.000	04:48
Subject Sample	# 1	0.191	04:49
Breath Volume		2.863	04:49
Air Blank		0.000	04:50
Air Blank		0.000	04:51
Subject Sample	# 2	0.160	04:53
Breath Volume		1.488	04:53
Air Blank		0.000	04:54
Air Blank		0.000	04:55
Subject Sample	# 3	0.167	04:56
Breath Volume		2.972	04:56
Air Blank		0.000	04:57
Control Test		0.078	04:57
Air Blank		0.000	04:57
Diagnostics Check		OK	04:57

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: VERMILLION,B
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:00
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: RUSH

DATE/TIME OF TEST: 02/19/2018 00:50:48

Diagnostics Check		OK	00:44
Air Blank		0.000	00:44
Control Test		0.079	00:45
Air Blank		0.000	00:45
Subject Sample	# 1	0.165	00:46
Breath Volume		2.152	00:46
Air Blank		0.000	00:46
Air Blank		0.000	00:48
Subject Sample	# 2	0.163	00:49
Breath Volume		2.394	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Diagnostics Check		OK	00:50

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NAMORS,JASON,R
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: CURRY,R
ARRESTING OFFICER AGENCY: North Port PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:00

SUBJECT'S LAST NAME: LINCOLN

DATE/TIME OF TEST: 02/19/2018 23:33:59

Diagnostics Check		OK	23:27
Air Blank		0.000	23:28
Control Test		0.079	23:28
Air Blank		0.000	23:29
Subject Sample	# 1	0.241	23:29
Breath Volume		2.578	23:29
Air Blank		0.000	23:30
Air Blank		0.000	23:32
Subject Sample	# 2	0.221	23:32
Breath Volume		1.480	23:32
Air Blank		0.000	23:33
Control Test		0.078	23:33
Air Blank		0.000	23:33
Diagnostics Check		OK	23:33

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDFORD,TAYLOR,M
BREATH TEST OPERATOR AGENCY: FHP Troop F
ARRESTING OFFICER: LEDFORD,TAYLOR
ARRESTING OFFICER AGENCY: FHP Troop F
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:45

SUBJECT'S LAST NAME: PATTON

DATE/TIME OF TEST: 02/20/2018 03:15:30

Diagnostics Check		OK	03:09
Air Blank		0.000	03:09
Control Test		0.079	03:09
Air Blank		0.000	03:10
Subject Sample	# 1	0.155	03:11
Breath Volume		1.515	03:11
Air Blank		0.000	03:11
Air Blank		0.000	03:13
Subject Sample	# 2	0.160	03:14
Breath Volume		1.472	03:14
Air Blank		0.000	03:14
Control Test		0.078	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MITCHELL,N,N
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: MARTINEZ,D
ARRESTING OFFICER AGENCY: Sarasota PD
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: HARRINGTON

DATE/TIME OF TEST: 02/21/2018 23:56:33

Diagnostics Check		OK	23:49
Air Blank		0.000	23:50
Control Test		0.079	23:50
Air Blank		0.000	23:51
Subject Sample	# 1	0.000	23:52
Breath Volume		3.796	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.000	23:55
Breath Volume		3.097	23:55
Air Blank		0.000	23:55
Control Test		0.079	23:55
Air Blank		0.000	23:56
Diagnostics Check		OK	23:56

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001347

INSTRUMENT REGISTERED TO: SARASOTA COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSTYAL,MATTHEW,A
BREATH TEST OPERATOR AGENCY: Sarasota County SO
ARRESTING OFFICER: KOSTYAL,MATTHEW
ARRESTING OFFICER AGENCY: Sarasota County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09216080A2
DRY GAS EXPIRY DATE: 05/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:45

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/22/2018 00:13:18

Diagnostics Check		OK	00:05
Air Blank		0.000	00:05
Control Test		0.078	00:06
Air Blank		0.000	00:06
Subject Sample	# 1	0.145	00:08
Breath Volume		1.679	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.149	00:11
Breath Volume		1.957	00:11
Air Blank		0.000	00:12
Control Test		0.079	00:12
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 02/23/2018 07:48

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LEONEL

DATE/TIME OF TEST: 02/12/2018 02:21:39

Diagnostics Check		OK	02:18	
Air Blank		0.000	02:19	
Control Test		0.081	02:19	
Air Blank		0.000	02:20	
Subject Sample	# 1		02:20	RFI Detect
Breath Volume		0.046	02:20	
Air Blank			02:21	RFI Detect
Air Blank		0.000	02:21	

Data Download Date/Time: 02/28/2018 04:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: SCHEINER,JAMES
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 01:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LEONEL

DATE/TIME OF TEST: 02/12/2018 02:34:20

Diagnostics Check		OK	02:24	
Air Blank		0.000	02:24	
Control Test		0.080	02:24	
Air Blank		0.000	02:25	
Subject Sample	# 1	0.173	02:28	Volume Not Met
Breath Volume		0.480	02:28	
Air Blank		0.000	02:29	
Air Blank		0.000	02:31	
Subject Sample	# 2		02:32	Subject Test Refused
Breath Volume		0.070	02:32	
Air Blank		0.000	02:33	
Control Test		0.078	02:33	
Air Blank		0.000	02:34	
Diagnostics Check		OK	02:34	

Data Download Date/Time: 02/28/2018 04:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: ROBIN,GASTON
ARRESTING OFFICER AGENCY: Brevard County SO
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 02:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:11

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/12/2018 02:50:24

Diagnostics Check		OK	02:44
Air Blank		0.000	02:44
Control Test		0.080	02:45
Air Blank		0.000	02:45
Subject Sample	# 1	0.078	02:46
Breath Volume		2.753	02:46
Air Blank		0.000	02:46
Air Blank		0.000	02:48
Subject Sample	# 2	0.074	02:48
Breath Volume		3.593	02:48
Air Blank		0.000	02:49
Control Test		0.080	02:49
Air Blank		0.000	02:50
Diagnostics Check		OK	02:50

Data Download Date/Time: 02/28/2018 04:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: PAYNE,TAYOR
ARRESTING OFFICER AGENCY: Cocoa Beach PD
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 02:42
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:42

SUBJECT'S LAST NAME: LIRIANO

DATE/TIME OF TEST: 02/14/2018 03:10:23

Diagnostics Check		OK	03:04
Air Blank		0.000	03:04
Control Test		0.081	03:05
Air Blank		0.000	03:05
Subject Sample	# 1	0.101	03:06
Breath Volume		1.722	03:06
Air Blank		0.000	03:06
Air Blank		0.000	03:08
Subject Sample	# 2	0.096	03:08
Breath Volume		2.535	03:08
Air Blank		0.000	03:09
Control Test		0.081	03:09
Air Blank		0.000	03:10
Diagnostics Check		OK	03:10

Data Download Date/Time: 02/28/2018 04:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001361

INSTRUMENT REGISTERED TO: FL HIGHWAY PATROL
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HITCHCOCK,WILLIAM,D
BREATH TEST OPERATOR AGENCY: FHP Troop D
ARRESTING OFFICER: HITCHCOCK,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/29/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 539031
DRY GAS EXPIRY DATE: 08/25/2019
VIOLATION CODE: DUI
ARREST TIME: 01:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:50

SUBJECT'S LAST NAME: LISTON

DATE/TIME OF TEST: 02/17/2018 02:22:02

Diagnostics Check		OK	02:16
Air Blank		0.000	02:16
Control Test		0.082	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.040	02:17
Breath Volume		2.878	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.039	02:20
Breath Volume		2.714	02:20
Air Blank		0.000	02:21
Control Test		0.080	02:21
Air Blank		0.000	02:21
Diagnostics Check		OK	02:22

Data Download Date/Time: 02/28/2018 04:17

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LESHER,TIFFANY,A
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: HORNE,ALTON
ARRESTING OFFICER AGENCY: Levy County SO
LAST AGENCY INSPECTION: 01/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 16:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:50

SUBJECT'S LAST NAME: YUJA

DATE/TIME OF TEST: 02/20/2018 17:48:17

Diagnostics Check		OK	17:41
Air Blank		0.000	17:41
Control Test		0.081	17:42
Air Blank		0.000	17:42
Subject Sample	# 1	0.114	17:43
Breath Volume		1.726	17:43
Air Blank		0.000	17:44
Air Blank		0.000	17:45
Subject Sample	# 2	0.103	17:46
Breath Volume		1.945	17:46
Air Blank		0.000	17:47
Control Test		0.080	17:47
Air Blank		0.000	17:48
Diagnostics Check		OK	17:48

Data Download Date/Time: 02/25/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: POLK,AUSTIN,B
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: SCHULTZ,KYLE
ARRESTING OFFICER AGENCY: Chiefland PD
LAST AGENCY INSPECTION: 01/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: OTHER
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:30

SUBJECT'S LAST NAME: JUV

DATE/TIME OF TEST: 02/22/2018 01:55:53

Diagnostics Check		OK	01:49
Air Blank		0.000	01:50
Control Test		0.080	01:50
Air Blank		0.000	01:51
Subject Sample	# 1	0.000	01:51
Breath Volume		2.054	01:51
Air Blank		0.000	01:52
Air Blank		0.000	01:54
Subject Sample	# 2	0.000	01:54
Breath Volume		2.992	01:54
Air Blank		0.000	01:54
Control Test		0.080	01:55
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 02/25/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ERNST,WILLIAM,H
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: ERNST,WILLIAM
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 01/21/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 14:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:27

SUBJECT'S LAST NAME: RODEN

DATE/TIME OF TEST: 02/22/2018 16:58:21

Diagnostics Check		OK	16:49	
Air Blank		0.000	16:50	
Control Test		0.080	16:50	
Air Blank		0.000	16:51	
Subject Sample	# 1		16:54	No Sample Provided
Breath Volume		0.000	16:54	
Air Blank		0.000	16:54	
Air Blank		0.000	16:56	
Subject Sample	# 2		16:56	Subject Test Refused
Breath Volume		0.000	16:56	
Air Blank		0.000	16:57	
Control Test		0.081	16:57	
Air Blank		0.000	16:58	
Diagnostics Check		OK	16:58	

Data Download Date/Time: 02/25/2018 16:56

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PIPKIN

DATE/TIME OF TEST: 02/25/2018 19:36:58

Diagnostics Check	OK	19:35	
Air Blank	0.000	19:36	
Control Test	0.080	19:36	RFI Detect
Air Blank	0.000	19:36	

Data Download Date/Time: 03/31/2018 16:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MASH,EARL,C
BREATH TEST OPERATOR AGENCY: FHP Troop B
ARRESTING OFFICER: MASH,EARL
ARRESTING OFFICER AGENCY: FHP Troop B
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 18:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:05

SUBJECT'S LAST NAME: PIPKIN

DATE/TIME OF TEST: 02/25/2018 19:47:37

Diagnostics Check		OK	19:41
Air Blank		0.000	19:41
Control Test		0.080	19:42
Air Blank		0.000	19:42
Subject Sample	# 1	0.144	19:43
Breath Volume		5.101	19:43
Air Blank		0.000	19:43
Air Blank		0.000	19:45
Subject Sample	# 2	0.138	19:45
Breath Volume		3.339	19:45
Air Blank		0.000	19:46
Control Test		0.080	19:47
Air Blank		0.000	19:47
Diagnostics Check		OK	19:47

Data Download Date/Time: 03/31/2018 16:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001365

INSTRUMENT REGISTERED TO: LEVY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SNYDER,SCOTT,C
BREATH TEST OPERATOR AGENCY: Levy County SO
ARRESTING OFFICER: POLIMENI,NEIL
ARRESTING OFFICER AGENCY: Cedar Keys PD
LAST AGENCY INSPECTION: 02/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG619605
DRY GAS EXPIRY DATE: 07/14/2018
VIOLATION CODE: DUI
ARREST TIME: 17:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:48

SUBJECT'S LAST NAME: STABILE

DATE/TIME OF TEST: 02/28/2018 18:25:15

Diagnostics Check		OK	18:19
Air Blank		0.000	18:19
Control Test		0.081	18:20
Air Blank		0.000	18:20
Subject Sample	# 1	0.000	18:20
Breath Volume		3.898	18:20
Air Blank		0.000	18:21
Air Blank		0.000	18:23
Subject Sample	# 2	0.000	18:23
Breath Volume		3.757	18:23
Air Blank		0.000	18:24
Control Test		0.080	18:24
Air Blank		0.000	18:25
Diagnostics Check		OK	18:25

Data Download Date/Time: 03/31/2018 16:36

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,A,M
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,A
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:35

SUBJECT'S LAST NAME: CASTRO-BALDI

DATE/TIME OF TEST: 02/17/2018 01:14:32

Diagnostics Check		OK	01:08
Air Blank		0.000	01:09
Control Test		0.080	01:09
Air Blank		0.000	01:09
Subject Sample	# 1	0.000	01:10
Breath Volume		2.468	01:10
Air Blank		0.000	01:10
Air Blank		0.000	01:12
Subject Sample	# 2	0.000	01:13
Breath Volume		2.468	01:13
Air Blank		0.000	01:13
Control Test		0.079	01:13
Air Blank		0.000	01:14
Diagnostics Check		OK	01:14

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:43

SUBJECT'S LAST NAME: WHALEN

DATE/TIME OF TEST: 02/17/2018 02:06:38

Diagnostics Check		OK	02:03	
Air Blank		0.000	02:04	
Control Test		0.080	02:04	
Air Blank		0.000	02:05	
Subject Sample	# 1		02:05	Subject Test Refused
Breath Volume		0.000	02:05	
Air Blank		0.000	02:05	
Control Test		0.081	02:06	
Air Blank		0.000	02:06	
Diagnostics Check		OK	02:06	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:59
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:30

SUBJECT'S LAST NAME: ELLIS

DATE/TIME OF TEST: 02/17/2018 04:58:08

Diagnostics Check		OK	04:50
Air Blank		0.000	04:51
Control Test		0.081	04:51
Air Blank		0.000	04:52
Subject Sample	# 1	0.195	04:53
Breath Volume		2.472	04:53
Air Blank		0.000	04:53
Air Blank		0.000	04:55
Subject Sample	# 2	0.195	04:56
Breath Volume		1.527	04:56
Air Blank		0.000	04:57
Control Test		0.080	04:57
Air Blank		0.000	04:58
Diagnostics Check		OK	04:58

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HORNING,PETER,T
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FIVECOAT,CURTIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 14:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:08

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 02/17/2018 16:38:30

Diagnostics Check		OK	16:29
Air Blank		0.000	16:29
Control Test		0.081	16:30
Air Blank		0.000	16:30
Subject Sample	# 1	0.187	16:32
Breath Volume		2.066	16:32
Air Blank		0.000	16:33
Air Blank		0.000	16:35
Subject Sample	# 2	0.185	16:36
Breath Volume		1.457	16:36
Air Blank		0.000	16:37
Control Test		0.080	16:37
Air Blank		0.000	16:38
Diagnostics Check		OK	16:38

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: WEDE,BRENNAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:52

SUBJECT'S LAST NAME: SEYMOUR

DATE/TIME OF TEST: 02/17/2018 22:20:53

Diagnostics Check		OK	22:18	
Air Blank		0.000	22:18	
Control Test		0.080	22:18	
Air Blank		0.000	22:19	
Subject Sample	# 1		22:19	Subject Test Refused
Breath Volume		0.000	22:19	
Air Blank		0.000	22:19	
Control Test		0.080	22:20	
Air Blank		0.000	22:20	
Diagnostics Check		OK	22:20	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:23

SUBJECT'S LAST NAME: REEVES

DATE/TIME OF TEST: 02/18/2018 00:46:40

Diagnostics Check		OK	00:43	
Air Blank		0.000	00:44	
Control Test		0.081	00:44	
Air Blank		0.000	00:45	
Subject Sample	# 1		00:45	Subject Test Refused
Breath Volume		0.000	00:45	
Air Blank		0.000	00:45	
Control Test		0.081	00:46	
Air Blank		0.000	00:46	
Diagnostics Check		OK	00:46	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DNOHUE,LAWRENCE,E
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: DONOHUE,LAWRENCE
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:33

SUBJECT'S LAST NAME: KIDDY

DATE/TIME OF TEST: 02/18/2018 02:59:59

Diagnostics Check		OK	02:53
Air Blank		0.000	02:54
Control Test		0.081	02:54
Air Blank		0.000	02:55
Subject Sample	# 1	0.000	02:55
Breath Volume		1.906	02:55
Air Blank		0.000	02:56
Air Blank		0.000	02:58
Subject Sample	# 2	0.000	02:58
Breath Volume		1.628	02:58
Air Blank		0.000	02:59
Control Test		0.081	02:59
Air Blank		0.000	02:59
Diagnostics Check		OK	02:59

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANIGUA,MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PANAIGUA,LT MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:31

SUBJECT'S LAST NAME: ALLEN

DATE/TIME OF TEST: 02/21/2018 02:56:51

Diagnostics Check		OK	02:53	
Air Blank		0.000	02:54	
Control Test		0.080	02:54	
Air Blank		0.000	02:54	
Subject Sample	# 1		02:55	Subject Test Refused
Breath Volume		0.000	02:55	
Air Blank		0.000	02:55	
Control Test		0.080	02:56	
Air Blank		0.000	02:56	
Diagnostics Check		OK	02:56	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:16

SUBJECT'S LAST NAME: VRAGOVIC

DATE/TIME OF TEST: 02/21/2018 21:40:27

Diagnostics Check		OK	21:37	
Air Blank		0.000	21:38	
Control Test		0.080	21:38	
Air Blank		0.000	21:38	
Subject Sample	# 1		21:39	Subject Test Refused
Breath Volume		0.000	21:39	
Air Blank		0.000	21:39	
Control Test		0.080	21:39	
Air Blank		0.000	21:40	
Diagnostics Check		OK	21:40	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: HANNA,JANE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:24

SUBJECT'S LAST NAME: PORTO

DATE/TIME OF TEST: 02/22/2018 01:49:11

Diagnostics Check		OK	01:46	
Air Blank		0.000	01:46	
Control Test		0.080	01:47	
Air Blank		0.000	01:47	
Subject Sample	# 1		01:47	Subject Test Refused
Breath Volume		0.000	01:47	
Air Blank		0.000	01:48	
Control Test		0.080	01:48	
Air Blank		0.000	01:49	
Diagnostics Check		OK	01:49	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: GRIFFIN,KYLE,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: GRIFFIN,KYLE
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:26

SUBJECT'S LAST NAME: RICHARDSON

DATE/TIME OF TEST: 02/22/2018 02:54:05

Diagnostics Check		OK	02:47
Air Blank		0.000	02:47
Control Test		0.080	02:47
Air Blank		0.000	02:48
Subject Sample	# 1	0.102	02:49
Breath Volume		1.808	02:49
Air Blank		0.000	02:49
Air Blank		0.000	02:51
Subject Sample	# 2	0.101	02:52
Breath Volume		1.648	02:52
Air Blank		0.000	02:53
Control Test		0.080	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/22/2018 03:50:49

Diagnostics Check		OK	03:34	
Air Blank		0.000	03:35	
Control Test		0.080	03:35	
Air Blank		0.000	03:35	
Subject Sample	# 1	0.161	03:39	Volume Not Met
Breath Volume		0.109	03:39	
Air Blank		0.000	03:39	
Air Blank		0.000	03:41	
Subject Sample	# 2	0.159	03:43	
Breath Volume		1.640	03:43	
Air Blank		0.000	03:44	
Air Blank		0.000	03:46	
Subject Sample	# 3	0.141	03:49	Volume Not Met
Breath Volume		0.078	03:49	
Air Blank		0.000	03:49	
Control Test		0.078	03:50	
Air Blank		0.000	03:50	
Diagnostics Check		OK	03:50	

Data Download Date/Time: 03/22/2018 23:10

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:52
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: THOMAS

DATE/TIME OF TEST: 02/22/2018 04:02:28

Diagnostics Check		OK	03:52	
Air Blank		0.000	03:52	
Control Test		0.079	03:53	
Air Blank		0.000	03:53	
Subject Sample	# 1	0.154	03:56	
Breath Volume		1.414	03:56	
Air Blank		0.000	03:56	
Air Blank		0.000	03:58	
Subject Sample	# 2		04:00	Subject Test Refused
Breath Volume		0.386	04:00	
Air Blank		0.000	04:01	
Control Test		0.078	04:01	
Air Blank		0.000	04:02	
Diagnostics Check		OK	04:02	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FOSTER,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:12
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: SZMYGLEWSKI

DATE/TIME OF TEST: 02/22/2018 05:08:38

Diagnostics Check		OK	05:01
Air Blank		0.000	05:01
Control Test		0.079	05:01
Air Blank		0.000	05:02
Subject Sample	# 1	0.178	05:04
Breath Volume		2.828	05:04
Air Blank		0.000	05:04
Air Blank		0.000	05:06
Subject Sample	# 2	0.187	05:07
Breath Volume		2.914	05:07
Air Blank		0.000	05:07
Control Test		0.080	05:08
Air Blank		0.000	05:08
Diagnostics Check		OK	05:08

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: NIXON,AARON,M
BREATH TEST OPERATOR AGENCY: FHP Troop C
ARRESTING OFFICER: NIXON,AARON
ARRESTING OFFICER AGENCY: FHP Troop C
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:17

SUBJECT'S LAST NAME: HOUSER

DATE/TIME OF TEST: 02/22/2018 22:46:05

Diagnostics Check		OK	22:39
Air Blank		0.000	22:40
Control Test		0.080	22:40
Air Blank		0.000	22:41
Subject Sample	# 1	0.208	22:41
Breath Volume		1.261	22:41
Air Blank		0.000	22:42
Air Blank		0.000	22:44
Subject Sample	# 2	0.193	22:44
Breath Volume		1.921	22:44
Air Blank		0.000	22:45
Control Test		0.079	22:45
Air Blank		0.000	22:46
Diagnostics Check		OK	22:46

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELLIS,ALEXANDER,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: ELLIS,ALEXANDER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:21

SUBJECT'S LAST NAME: COLANDREA

DATE/TIME OF TEST: 02/23/2018 00:48:55

Diagnostics Check		OK	00:42
Air Blank		0.000	00:42
Control Test		0.080	00:42
Air Blank		0.000	00:43
Subject Sample	# 1	0.223	00:44
Breath Volume		1.480	00:44
Air Blank		0.000	00:44
Air Blank		0.000	00:46
Subject Sample	# 2	0.220	00:47
Breath Volume		2.359	00:47
Air Blank		0.000	00:48
Control Test		0.079	00:48
Air Blank		0.000	00:48
Diagnostics Check		OK	00:48

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:20
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: MULLINS

DATE/TIME OF TEST: 02/23/2018 02:16:16

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:08
Subject Sample	# 1	0.095	02:10
Breath Volume		1.210	02:10
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.099	02:14
Breath Volume		1.531	02:14
Air Blank		0.000	02:15
Control Test		0.080	02:15
Air Blank		0.000	02:16
Diagnostics Check		OK	02:16

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ANTHONY,MICHAEL,
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MARTIN,JEFFERY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 03:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:22

SUBJECT'S LAST NAME: MAISELSON

DATE/TIME OF TEST: 02/23/2018 04:47:50

Diagnostics Check		OK	04:44	
Air Blank		0.000	04:45	
Control Test		0.079	04:45	
Air Blank		0.000	04:46	
Subject Sample	# 1		04:46	Subject Test Refused
Breath Volume		0.000	04:46	
Air Blank		0.000	04:46	
Control Test		0.078	04:47	
Air Blank		0.000	04:47	
Diagnostics Check		OK	04:47	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:28

SUBJECT'S LAST NAME: CAMERIRE

DATE/TIME OF TEST: 02/23/2018 22:51:36

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:49	
Control Test		0.081	22:49	
Air Blank		0.000	22:50	
Subject Sample	# 1		22:50	Subject Test Refused
Breath Volume		0.000	22:50	
Air Blank		0.000	22:50	
Control Test		0.080	22:51	
Air Blank		0.000	22:51	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DORR,RANDALL,S
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: DORR,RANDLL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:01

SUBJECT'S LAST NAME: BLAIR

DATE/TIME OF TEST: 02/24/2018 00:24:04

Diagnostics Check		OK	00:21	
Air Blank		0.000	00:21	
Control Test		0.081	00:21	
Air Blank		0.000	00:22	
Subject Sample	# 1		00:22	Subject Test Refused
Breath Volume		0.000	00:22	
Air Blank		0.000	00:23	
Control Test		0.081	00:23	
Air Blank		0.000	00:23	
Diagnostics Check		OK	00:24	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: JACOBY,LARRY,K
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: JQACOBY,LARRY
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 02:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:01

SUBJECT'S LAST NAME: QUIRK

DATE/TIME OF TEST: 02/24/2018 03:28:19

Diagnostics Check		OK	03:21
Air Blank		0.000	03:22
Control Test		0.081	03:22
Air Blank		0.000	03:23
Subject Sample	# 1	0.198	03:23
Breath Volume		2.312	03:23
Air Blank		0.000	03:24
Air Blank		0.000	03:26
Subject Sample	# 2	0.200	03:26
Breath Volume		2.488	03:26
Air Blank		0.000	03:27
Control Test		0.079	03:27
Air Blank		0.000	03:28
Diagnostics Check		OK	03:28

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BLAIR,RONALD,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: BEARY,FRANCIS
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 16:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:26

SUBJECT'S LAST NAME: VARGAS JR

DATE/TIME OF TEST: 02/24/2018 17:49:00

Diagnostics Check		OK	17:46	
Air Blank		0.000	17:46	
Control Test		0.081	17:46	
Air Blank		0.000	17:47	
Subject Sample	# 1		17:47	Subject Test Refused
Breath Volume		0.000	17:47	
Air Blank		0.000	17:48	
Control Test		0.080	17:48	
Air Blank		0.000	17:48	
Diagnostics Check		OK	17:49	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KASELAK,STEPHEN,G
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: KASELAK,STEPHEN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:06

SUBJECT'S LAST NAME: TYLER

DATE/TIME OF TEST: 02/25/2018 00:31:48

Diagnostics Check		OK	00:29	
Air Blank		0.000	00:29	
Control Test		0.081	00:29	
Air Blank		0.000	00:30	
Subject Sample	# 1		00:30	Subject Test Refused
Breath Volume		0.000	00:30	
Air Blank		0.000	00:30	
Control Test		0.081	00:31	
Air Blank		0.000	00:31	
Diagnostics Check		OK	00:31	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PANIAGUA,LT MICHAEL,A
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: PANIAGUA,LT MICHAEL
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 01:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:00

SUBJECT'S LAST NAME: WALKER

DATE/TIME OF TEST: 02/25/2018 02:25:58

Diagnostics Check		OK	02:21	
Air Blank		0.000	02:22	
Control Test		0.081	02:22	
Air Blank		0.000	02:23	
Subject Sample	# 1		02:24	Subject Test Refused
Breath Volume		0.000	02:24	
Air Blank		0.000	02:25	
Control Test		0.081	02:25	
Air Blank		0.000	02:25	
Diagnostics Check		OK	02:25	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MOWATT,NATHAN
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:35
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:45

SUBJECT'S LAST NAME: WESTOVER

DATE/TIME OF TEST: 02/25/2018 05:12:49

Diagnostics Check		OK	05:05
Air Blank		0.000	05:06
Control Test		0.080	05:06
Air Blank		0.000	05:06
Subject Sample	# 1	0.132	05:07
Breath Volume		1.160	05:07
Air Blank		0.000	05:08
Air Blank		0.000	05:10
Subject Sample	# 2	0.130	05:11
Breath Volume		1.851	05:11
Air Blank		0.000	05:11
Control Test		0.080	05:12
Air Blank		0.000	05:12
Diagnostics Check		OK	05:12

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MOWATT,NATHAN,E
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: FRASER,CHRISTOPHER
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 04:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:58

SUBJECT'S LAST NAME: KRUG

DATE/TIME OF TEST: 02/25/2018 05:29:53

Diagnostics Check		OK	05:26	
Air Blank		0.000	05:27	
Control Test		0.081	05:27	
Air Blank		0.000	05:28	
Subject Sample	# 1		05:28	Subject Test Refused
Breath Volume		0.000	05:28	
Air Blank		0.000	05:28	
Control Test		0.081	05:29	
Air Blank		0.000	05:29	
Diagnostics Check		OK	05:29	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001367

INSTRUMENT REGISTERED TO: PINELLAS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ASHCOM,TIFFANY,J
BREATH TEST OPERATOR AGENCY: Pinellas County SO
ARRESTING OFFICER: MACKAIL,B
ARRESTING OFFICER AGENCY: Pinellas County SO
LAST AGENCY INSPECTION: 02/16/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: AG629802
DRY GAS EXPIRY DATE: 10/24/2018
VIOLATION CODE: DUI
ARREST TIME: 22:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:46

SUBJECT'S LAST NAME: HENSCHEN

DATE/TIME OF TEST: 02/26/2018 23:09:49

Diagnostics Check		OK	23:06	
Air Blank		0.000	23:07	
Control Test		0.081	23:07	
Air Blank		0.000	23:07	
Subject Sample	# 1		23:08	Subject Test Refused
Breath Volume		0.000	23:08	
Air Blank		0.000	23:08	
Control Test		0.081	23:09	
Air Blank		0.000	23:09	
Diagnostics Check		OK	23:09	

Data Download Date/Time: 03/22/2018 23:10

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: PERKINS,MARCUS
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/14/2018 23:21:45

Diagnostics Check	OK	23:20	
Air Blank	0.000	23:21	
		23:21	RFI Detect
Air Blank	0.000	23:21	

Data Download Date/Time: 02/27/2018 13:28

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001368

INSTRUMENT REGISTERED TO: NAS MAYPORT
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OCASIO,JOSE,A
BREATH TEST OPERATOR AGENCY: Mayport Naval Base
ARRESTING OFFICER: PERKINS,MARCUS
ARRESTING OFFICER AGENCY: Mayport Naval Base
LAST AGENCY INSPECTION: 01/30/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 17817080A3
DRY GAS EXPIRY DATE: 08/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:28
GENDER: Female
OBSERVATION PERIOD BEGAN: 22:45

SUBJECT'S LAST NAME: JOHNSON

DATE/TIME OF TEST: 02/14/2018 23:40:49

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.079	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.199	23:29
Breath Volume		1.281	23:29
Air Blank		0.000	23:29
Air Blank		0.000	23:31
Subject Sample	# 2	0.173	23:33
Breath Volume		2.246	23:33
Air Blank		0.000	23:34
Air Blank		0.000	23:36
Subject Sample	# 3	0.192	23:39
Breath Volume		2.167	23:39
Air Blank		0.000	23:39
Control Test		0.079	23:40
Air Blank		0.000	23:40
Diagnostics Check		OK	23:40

Data Download Date/Time: 02/27/2018 13:28

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: CLOTFELTER

DATE/TIME OF TEST: 02/01/2018 01:41:48

Diagnostics Check		OK	01:35
Air Blank		0.000	01:36
Control Test		0.082	01:36
Air Blank		0.000	01:37
Subject Sample	# 1	0.158	01:37
Breath Volume		2.218	01:37
Air Blank		0.000	01:38
Air Blank		0.000	01:40
Subject Sample	# 2	0.151	01:40
Breath Volume		1.140	01:40
Air Blank		0.000	01:40
Control Test		0.081	01:41
Air Blank		0.000	01:41
Diagnostics Check		OK	01:41

Data Download Date/Time: 02/01/2018 15:46

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORTES,ADAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:56
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:47

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/02/2018 04:16:50

Diagnostics Check		OK	04:10
Air Blank		0.000	04:11
Control Test		0.082	04:11
Air Blank		0.000	04:12
Subject Sample	# 1	0.135	04:12
Breath Volume		1.535	04:12
Air Blank		0.000	04:13
Air Blank		0.000	04:14
Subject Sample	# 2	0.134	04:15
Breath Volume		1.859	04:15
Air Blank		0.000	04:15
Control Test		0.081	04:16
Air Blank		0.000	04:16
Diagnostics Check		OK	04:16

Data Download Date/Time: 02/08/2018 16:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MARIANO,SABASTIAN
ARRESTING OFFICER AGENCY: Winter Garden PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:20

SUBJECT'S LAST NAME: STRAUM-LOVE

DATE/TIME OF TEST: 02/03/2018 21:53:55

Diagnostics Check		OK	21:46
Air Blank		0.000	21:47
Control Test		0.082	21:47
Air Blank		0.000	21:48
Subject Sample	# 1	0.000	21:48
Breath Volume		1.261	21:48
Air Blank		0.000	21:49
Air Blank		0.000	21:51
Subject Sample	# 2	0.000	21:52
Breath Volume		1.429	21:52
Air Blank		0.000	21:53
Control Test		0.082	21:53
Air Blank		0.000	21:53
Diagnostics Check		OK	21:53

Data Download Date/Time: 02/08/2018 16:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:29
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:14

SUBJECT'S LAST NAME: FLORESDORANTES

DATE/TIME OF TEST: 02/04/2018 03:43:55

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.082	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.160	03:39
Breath Volume		1.425	03:39
Air Blank		0.000	03:40
Air Blank		0.000	03:41
Subject Sample	# 2	0.154	03:42
Breath Volume		1.496	03:42
Air Blank		0.000	03:43
Control Test		0.081	03:43
Air Blank		0.000	03:43
Diagnostics Check		OK	03:43

Data Download Date/Time: 02/08/2018 16:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:17

SUBJECT'S LAST NAME: GONZALEZ MENDOZA

DATE/TIME OF TEST: 02/04/2018 04:50:43

Diagnostics Check		OK	04:43
Air Blank		0.000	04:44
Control Test		0.081	04:44
Air Blank		0.000	04:45
Subject Sample	# 1	0.189	04:45
Breath Volume		2.484	04:45
Air Blank		0.000	04:46
Air Blank		0.000	04:47
Subject Sample	# 2	0.198	04:49
Breath Volume		3.449	04:49
Air Blank		0.000	04:49
Control Test		0.081	04:50
Air Blank		0.000	04:50
Diagnostics Check		OK	04:50

Data Download Date/Time: 02/08/2018 16:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: MILLS,ARMAND,G
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORRIS,STEVE
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:29
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:05

SUBJECT'S LAST NAME: ARCADEAU

DATE/TIME OF TEST: 02/05/2018 04:40:43

Diagnostics Check		OK	04:32
Air Blank		0.000	04:32
Control Test		0.081	04:33
Air Blank		0.000	04:33
Subject Sample	# 1	0.185	04:36
Breath Volume		2.367	04:36
Air Blank		0.000	04:36
Air Blank		0.000	04:38
Subject Sample	# 2	0.183	04:39
Breath Volume		2.347	04:39
Air Blank		0.000	04:39
Control Test		0.080	04:40
Air Blank		0.000	04:40
Diagnostics Check		OK	04:40

Data Download Date/Time: 02/08/2018 16:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MICHALAK,TIMOTHY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:03

SUBJECT'S LAST NAME: SPRAY

DATE/TIME OF TEST: 02/08/2018 00:35:07

Diagnostics Check		OK	00:27
Air Blank		0.000	00:27
Control Test		0.081	00:28
Air Blank		0.000	00:28
Subject Sample	# 1	0.000	00:29
Breath Volume		1.179	00:29
Air Blank		0.000	00:29
Air Blank		0.000	00:31
Subject Sample	# 2	0.000	00:33
Breath Volume		1.675	00:33
Air Blank		0.000	00:34
Control Test		0.082	00:34
Air Blank		0.000	00:35
Diagnostics Check		OK	00:35

Data Download Date/Time: 02/08/2018 16:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:23
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:45

SUBJECT'S LAST NAME: VELEZ

DATE/TIME OF TEST: 02/09/2018 04:23:00

Diagnostics Check		OK	04:11	
Air Blank		0.000	04:11	
Control Test		0.082	04:12	
Air Blank		0.000	04:12	
Subject Sample	# 1	0.203	04:15	Volume Not Met
Breath Volume		0.496	04:15	
Air Blank		0.000	04:16	
Air Blank		0.000	04:18	
Subject Sample	# 2	0.186	04:21	Volume Not Met
Breath Volume		0.152	04:21	
Air Blank		0.000	04:22	
Control Test		0.081	04:22	
Air Blank		0.000	04:22	
Diagnostics Check		OK	04:23	

Data Download Date/Time: 02/13/2018 10:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BODE,JOSHUA
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:20

SUBJECT'S LAST NAME: ZAMBRANO

DATE/TIME OF TEST: 02/09/2018 22:58:29

Diagnostics Check		OK	22:47	
Air Blank		0.000	22:47	
Control Test		0.081	22:47	
Air Blank		0.000	22:48	
Subject Sample	# 1		22:51	No Sample Provided
Breath Volume		0.000	22:51	
Air Blank		0.000	22:51	
Air Blank		0.000	22:53	
Subject Sample	# 2		22:57	No Sample Provided
Breath Volume		0.000	22:57	
Air Blank		0.000	22:57	
Control Test		0.082	22:57	
Air Blank		0.000	22:58	
Diagnostics Check		OK	22:58	

Data Download Date/Time: 02/13/2018 10:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GALLUP,ELIAS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: RAMOS

DATE/TIME OF TEST: 02/10/2018 02:51:00

Diagnostics Check		OK	02:45
Air Blank		0.000	02:45
Control Test		0.082	02:45
Air Blank		0.000	02:46
Subject Sample	# 1	0.109	02:46
Breath Volume		1.542	02:46
Air Blank		0.000	02:47
Air Blank		0.000	02:49
Subject Sample	# 2	0.110	02:49
Breath Volume		2.363	02:49
Air Blank		0.000	02:50
Control Test		0.081	02:50
Air Blank		0.000	02:50
Diagnostics Check		OK	02:51

Data Download Date/Time: 02/13/2018 10:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VORHEIS,JUSTIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:43

SUBJECT'S LAST NAME: BANCHS

DATE/TIME OF TEST: 02/10/2018 03:19:41

Diagnostics Check		OK	03:08	
Air Blank		0.000	03:08	
Control Test		0.081	03:09	
Air Blank		0.000	03:09	
Subject Sample	# 1		03:12	No Sample Provided
Breath Volume		0.000	03:12	
Air Blank		0.000	03:13	
Air Blank		0.000	03:15	
Subject Sample	# 2		03:18	No Sample Provided
Breath Volume		0.000	03:18	
Air Blank		0.000	03:18	
Control Test		0.081	03:19	
Air Blank		0.000	03:19	
Diagnostics Check		OK	03:19	

Data Download Date/Time: 02/13/2018 10:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:09
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: YOUNG

DATE/TIME OF TEST: 02/10/2018 04:33:54

Diagnostics Check		OK	04:27
Air Blank		0.000	04:28
Control Test		0.081	04:28
Air Blank		0.000	04:28
Subject Sample	# 1	0.170	04:29
Breath Volume		3.679	04:29
Air Blank		0.000	04:29
Air Blank		0.000	04:31
Subject Sample	# 2	0.170	04:32
Breath Volume		3.058	04:32
Air Blank		0.000	04:33
Control Test		0.081	04:33
Air Blank		0.000	04:33
Diagnostics Check		OK	04:33

Data Download Date/Time: 02/13/2018 10:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ROBBINS,NATHAN
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 01/17/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:43

SUBJECT'S LAST NAME: TURNER

DATE/TIME OF TEST: 02/11/2018 04:13:03

Diagnostics Check		OK	04:06
Air Blank		0.000	04:07
Control Test		0.082	04:07
Air Blank		0.000	04:08
Subject Sample	# 1	0.000	04:08
Breath Volume		1.746	04:08
Air Blank		0.000	04:09
Air Blank		0.000	04:11
Subject Sample	# 2	0.000	04:11
Breath Volume		3.273	04:11
Air Blank		0.000	04:12
Control Test		0.082	04:12
Air Blank		0.000	04:12
Diagnostics Check		OK	04:13

Data Download Date/Time: 02/13/2018 10:07

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GIARRATANA,JOHN
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:47

SUBJECT'S LAST NAME: BEADLE

DATE/TIME OF TEST: 02/22/2018 00:19:17

Diagnostics Check		OK	00:13
Air Blank		0.000	00:13
Control Test		0.081	00:14
Air Blank		0.000	00:14
Subject Sample	# 1	0.084	00:15
Breath Volume		2.781	00:15
Air Blank		0.000	00:15
Air Blank		0.000	00:17
Subject Sample	# 2	0.078	00:17
Breath Volume		1.484	00:17
Air Blank		0.000	00:18
Control Test		0.081	00:18
Air Blank		0.000	00:19
Diagnostics Check		OK	00:19

Data Download Date/Time: 03/08/2018 11:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:33

SUBJECT'S LAST NAME: RIVERA ZAYAS

DATE/TIME OF TEST: 02/22/2018 04:03:15

Diagnostics Check		OK	03:57
Air Blank		0.000	03:57
Control Test		0.081	03:57
Air Blank		0.000	03:58
Subject Sample	# 1	0.218	03:58
Breath Volume		1.109	03:58
Air Blank		0.000	03:59
Air Blank		0.000	04:01
Subject Sample	# 2	0.215	04:01
Breath Volume		1.753	04:01
Air Blank		0.000	04:02
Control Test		0.081	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 03/08/2018 11:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DONAKOWSKI,JUSTIN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:48

SUBJECT'S LAST NAME: ESPINOZA

DATE/TIME OF TEST: 02/24/2018 02:18:24

Diagnostics Check		OK	02:15	
Air Blank		0.000	02:16	
Control Test		0.082	02:16	
Air Blank		0.000	02:16	
Subject Sample	# 1		02:17	Subject Test Refused
Breath Volume		0.000	02:17	
Air Blank		0.000	02:17	
Control Test		0.081	02:17	
Air Blank		0.000	02:18	
Diagnostics Check		OK	02:18	

Data Download Date/Time: 03/08/2018 11:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 02/24/2018 21:47:05

Diagnostics Check		OK	21:31	
Air Blank		0.000	21:32	
Control Test		0.082	21:32	
Air Blank		0.000	21:33	
Subject Sample	# 1	0.056	21:36	Volume Not Met
Breath Volume		0.070	21:36	
Air Blank		0.000	21:36	
Air Blank		0.000	21:38	
Subject Sample	# 2	0.068	21:39	
Breath Volume		2.171	21:39	
Air Blank		0.000	21:40	
Air Blank		0.000	21:42	
Subject Sample	# 3	0.058	21:45	Slope Not Level
Breath Volume		1.140	21:45	
Air Blank		0.000	21:46	
Control Test		0.081	21:46	
Air Blank		0.000	21:47	
Diagnostics Check		OK	21:47	

Data Download Date/Time: 03/08/2018 11:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: JORGE,JUNIOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:09

SUBJECT'S LAST NAME: COLON

DATE/TIME OF TEST: 02/24/2018 21:56:02

Diagnostics Check		OK	21:48
Air Blank		0.000	21:48
Control Test		0.081	21:49
Air Blank		0.000	21:49
Subject Sample	# 1	0.067	21:50
Breath Volume		1.324	21:50
Air Blank		0.000	21:50
Air Blank		0.000	21:52
Subject Sample	# 2	0.060	21:54
Breath Volume		1.519	21:54
Air Blank		0.000	21:55
Control Test		0.081	21:55
Air Blank		0.000	21:55
Diagnostics Check		OK	21:56

Data Download Date/Time: 03/08/2018 11:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALFRED,BRANDON
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:36
GENDER: Female
OBSERVATION PERIOD BEGAN: 20:18

SUBJECT'S LAST NAME: MARTINEZ

DATE/TIME OF TEST: 02/25/2018 20:44:24

Diagnostics Check		OK	20:41	
Air Blank		0.000	20:41	
Control Test		0.081	20:42	
Air Blank		0.000	20:42	
Subject Sample	# 1		20:43	Subject Test Refused
Breath Volume		0.000	20:43	
Air Blank		0.000	20:43	
Control Test		0.081	20:43	
Air Blank		0.000	20:44	
Diagnostics Check		OK	20:44	

Data Download Date/Time: 03/08/2018 11:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDEN,WILLIAM,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MILLER,TAYLOR
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:38

SUBJECT'S LAST NAME: SOTHERLAND

DATE/TIME OF TEST: 02/27/2018 01:05:08

Diagnostics Check		OK	01:02	
Air Blank		0.000	01:02	
Control Test		0.082	01:03	
Air Blank		0.000	01:03	
Subject Sample	# 1		01:03	Subject Test Refused
Breath Volume		0.000	01:03	
Air Blank		0.000	01:04	
Control Test		0.081	01:04	
Air Blank		0.000	01:05	
Diagnostics Check		OK	01:05	

Data Download Date/Time: 03/08/2018 11:13

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001418

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:04

SUBJECT'S LAST NAME: STONECIPHER

DATE/TIME OF TEST: 02/28/2018 22:35:31

Diagnostics Check		OK	22:29
Air Blank		0.000	22:29
Control Test		0.081	22:30
Air Blank		0.000	22:30
Subject Sample	# 1	0.161	22:31
Breath Volume		3.074	22:31
Air Blank		0.000	22:31
Air Blank		0.000	22:33
Subject Sample	# 2	0.157	22:34
Breath Volume		2.417	22:34
Air Blank		0.000	22:34
Control Test		0.081	22:34
Air Blank		0.000	22:35
Diagnostics Check		OK	22:35

Data Download Date/Time: 03/08/2018 11:14

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BAUER,TYLER
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:36

SUBJECT'S LAST NAME: KORNAKER

DATE/TIME OF TEST: 02/13/2018 23:05:28

Diagnostics Check		OK	22:58
Air Blank		0.000	22:58
Control Test		0.079	22:59
Air Blank		0.000	22:59
Subject Sample	# 1	0.180	23:00
Breath Volume		1.769	23:00
Air Blank		0.000	23:01
Air Blank		0.000	23:03
Subject Sample	# 2	0.167	23:03
Breath Volume		2.421	23:03
Air Blank		0.000	23:04
Control Test		0.077	23:04
Air Blank		0.000	23:05
Diagnostics Check		OK	23:05

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: TALLEY

DATE/TIME OF TEST: 02/14/2018 00:08:58

Diagnostics Check		OK	00:01
Air Blank		0.000	00:02
Control Test		0.078	00:02
Air Blank		0.000	00:03
Subject Sample	# 1	0.153	00:04
Breath Volume		1.902	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:07
Subject Sample	# 2	0.158	00:07
Breath Volume		1.449	00:07
Air Blank		0.000	00:08
Control Test		0.079	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/14/2018 03:41:02

Diagnostics Check		OK	03:29	
Air Blank		0.000	03:29	
Control Test		0.079	03:30	
Air Blank		0.000	03:30	
Subject Sample	# 1	0.146	03:33	Volume Not Met
Breath Volume		0.050	03:33	
Air Blank		0.000	03:34	
Air Blank		0.000	03:36	
Subject Sample	# 2	0.138	03:39	Volume Not Met
Breath Volume		0.339	03:39	
Air Blank		0.000	03:40	
Control Test		0.076	03:40	
Air Blank		0.000	03:40	
Diagnostics Check		OK	03:41	

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DAVIS,COURTNEY,L
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:05

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/14/2018 03:46:41

Diagnostics Check		OK	03:42	
Air Blank		0.000	03:43	
Control Test		0.077	03:43	
Air Blank		0.000	03:43	
Subject Sample	# 1		03:45	Subject Test Refused
Breath Volume		0.000	03:45	
Air Blank		0.000	03:45	
Control Test		0.078	03:46	
Air Blank		0.000	03:46	
Diagnostics Check		OK	03:46	

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:51
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:40

SUBJECT'S LAST NAME: TOWER

DATE/TIME OF TEST: 02/14/2018 05:10:25

Diagnostics Check		OK	05:03
Air Blank		0.000	05:03
Control Test		0.080	05:04
Air Blank		0.000	05:04
Subject Sample	# 1	0.129	05:05
Breath Volume		1.722	05:05
Air Blank		0.000	05:06
Air Blank		0.000	05:08
Subject Sample	# 2	0.134	05:08
Breath Volume		1.937	05:08
Air Blank		0.000	05:09
Control Test		0.077	05:09
Air Blank		0.000	05:10
Diagnostics Check		OK	05:10

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VORHEIS,JUSTIN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: RAMIREZ HERNANDEZ

DATE/TIME OF TEST: 02/15/2018 01:31:20

Diagnostics Check		OK	01:24
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.172	01:26
Breath Volume		3.070	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.161	01:29
Breath Volume		1.968	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KELEMEN,JANOS
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:16
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:52

SUBJECT'S LAST NAME: OTT

DATE/TIME OF TEST: 02/15/2018 02:22:12

Diagnostics Check		OK	02:15
Air Blank		0.000	02:16
Control Test		0.080	02:16
Air Blank		0.000	02:17
Subject Sample	# 1	0.193	02:17
Breath Volume		4.179	02:17
Air Blank		0.000	02:18
Air Blank		0.000	02:20
Subject Sample	# 2	0.196	02:20
Breath Volume		3.960	02:20
Air Blank		0.000	02:21
Control Test		0.079	02:21
Air Blank		0.000	02:22
Diagnostics Check		OK	02:22

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: POWELL,KRISTI
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:16

SUBJECT'S LAST NAME: ADAMCZYK

DATE/TIME OF TEST: 02/15/2018 04:46:33

Diagnostics Check		OK	04:40
Air Blank		0.000	04:40
Control Test		0.079	04:41
Air Blank		0.000	04:41
Subject Sample	# 1	0.220	04:41
Breath Volume		2.812	04:41
Air Blank		0.000	04:42
Air Blank		0.000	04:44
Subject Sample	# 2	0.229	04:44
Breath Volume		2.414	04:44
Air Blank		0.000	04:45
Control Test		0.078	04:45
Air Blank		0.000	04:46
Diagnostics Check		OK	04:46

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAUGHN,JACOB
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:12

SUBJECT'S LAST NAME: RIVERA

DATE/TIME OF TEST: 02/16/2018 03:45:44

Diagnostics Check		OK	03:37
Air Blank		0.000	03:37
Control Test		0.079	03:38
Air Blank		0.000	03:38
Subject Sample	# 1	0.151	03:41
Breath Volume		1.335	03:41
Air Blank		0.000	03:42
Air Blank		0.000	03:43
Subject Sample	# 2	0.151	03:44
Breath Volume		1.179	03:44
Air Blank		0.000	03:44
Control Test		0.078	03:45
Air Blank		0.000	03:45
Diagnostics Check		OK	03:45

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HEBEL,GRIFFIN
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:44

SUBJECT'S LAST NAME: OCONNOR

DATE/TIME OF TEST: 02/16/2018 05:16:13

Diagnostics Check		OK	05:10
Air Blank		0.000	05:10
Control Test		0.079	05:11
Air Blank		0.000	05:11
Subject Sample	# 1	0.000	05:12
Breath Volume		1.839	05:12
Air Blank		0.000	05:12
Air Blank		0.000	05:14
Subject Sample	# 2	0.000	05:14
Breath Volume		2.417	05:14
Air Blank		0.000	05:15
Control Test		0.079	05:15
Air Blank		0.000	05:16
Diagnostics Check		OK	05:16

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HARRIS,BRIAN
ARRESTING OFFICER AGENCY: Ocoee PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:10
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:31

SUBJECT'S LAST NAME: OSTRANDER

DATE/TIME OF TEST: 02/16/2018 22:02:22

Diagnostics Check		OK	21:56
Air Blank		0.000	21:56
Control Test		0.079	21:56
Air Blank		0.000	21:57
Subject Sample	# 1	0.254	21:57
Breath Volume		2.105	21:57
Air Blank		0.000	21:58
Air Blank		0.000	22:00
Subject Sample	# 2	0.261	22:00
Breath Volume		2.488	22:00
Air Blank		0.000	22:01
Control Test		0.079	22:01
Air Blank		0.000	22:02
Diagnostics Check		OK	22:02

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:43
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:15

SUBJECT'S LAST NAME: ALMONTE

DATE/TIME OF TEST: 02/17/2018 00:46:27

Diagnostics Check		OK	00:39
Air Blank		0.000	00:39
Control Test		0.079	00:39
Air Blank		0.000	00:40
Subject Sample	# 1	0.099	00:41
Breath Volume		1.339	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.099	00:44
Breath Volume		1.226	00:44
Air Blank		0.000	00:45
Control Test		0.078	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DAVILA,SAMUEL
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:54

SUBJECT'S LAST NAME: MATTHEWS

DATE/TIME OF TEST: 02/17/2018 02:24:12

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.079	02:18
Air Blank		0.000	02:19
Subject Sample	# 1	0.275	02:19
Breath Volume		3.125	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.273	02:22
Breath Volume		3.085	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:15
GENDER: Female
OBSERVATION PERIOD BEGAN: 04:04

SUBJECT'S LAST NAME: DELGADO MERAZ

DATE/TIME OF TEST: 02/17/2018 04:35:41

Diagnostics Check		OK	04:28
Air Blank		0.000	04:29
Control Test		0.079	04:29
Air Blank		0.000	04:30
Subject Sample	# 1	0.160	04:30
Breath Volume		1.242	04:30
Air Blank		0.000	04:31
Air Blank		0.000	04:33
Subject Sample	# 2	0.157	04:34
Breath Volume		1.808	04:34
Air Blank		0.000	04:34
Control Test		0.078	04:35
Air Blank		0.000	04:35
Diagnostics Check		OK	04:35

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WEAVER,WILLIAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:01

SUBJECT'S LAST NAME: LEGER

DATE/TIME OF TEST: 02/17/2018 05:38:08

Diagnostics Check		OK	05:27
Air Blank		0.000	05:28
Control Test		0.079	05:28
Air Blank		0.000	05:28
Subject Sample	# 1	0.092	05:32
Breath Volume		1.277	05:32
Air Blank		0.000	05:32
Air Blank		0.000	05:34
Subject Sample	# 2	0.088	05:36
Breath Volume		1.703	05:36
Air Blank		0.000	05:37
Control Test		0.078	05:37
Air Blank		0.000	05:38
Diagnostics Check		OK	05:38

Data Download Date/Time: 03/13/2018 10:33

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: BEHRENSHOUSER,R
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:53
GENDER: Female
OBSERVATION PERIOD BEGAN: 05:25

SUBJECT'S LAST NAME: MORALBENITEZ

DATE/TIME OF TEST: 02/17/2018 05:55:50

Diagnostics Check		OK	05:49
Air Blank		0.000	05:49
Control Test		0.078	05:50
Air Blank		0.000	05:50
Subject Sample	# 1	0.142	05:51
Breath Volume		2.261	05:51
Air Blank		0.000	05:51
Air Blank		0.000	05:53
Subject Sample	# 2	0.145	05:54
Breath Volume		1.972	05:54
Air Blank		0.000	05:54
Control Test		0.079	05:55
Air Blank		0.000	05:55
Diagnostics Check		OK	05:55

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CORTES,ADAM
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:11
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:13

SUBJECT'S LAST NAME: OVIEDO

DATE/TIME OF TEST: 02/18/2018 00:43:19

Diagnostics Check		OK	00:37
Air Blank		0.000	00:37
Control Test		0.078	00:37
Air Blank		0.000	00:38
Subject Sample	# 1	0.152	00:38
Breath Volume		1.375	00:38
Air Blank		0.000	00:39
Air Blank		0.000	00:41
Subject Sample	# 2	0.146	00:41
Breath Volume		1.761	00:41
Air Blank		0.000	00:42
Control Test		0.079	00:42
Air Blank		0.000	00:43
Diagnostics Check		OK	00:43

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AZEVEDO,FABIO
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:01

SUBJECT'S LAST NAME: VINCES

DATE/TIME OF TEST: 02/18/2018 01:31:20

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.079	01:26
Air Blank		0.000	01:26
Subject Sample	# 1	0.102	01:26
Breath Volume		1.500	01:26
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.105	01:29
Breath Volume		1.968	01:29
Air Blank		0.000	01:30
Control Test		0.079	01:30
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: SIMPSON,JOSEPH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:44
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:37

SUBJECT'S LAST NAME: WATKINS

DATE/TIME OF TEST: 02/18/2018 02:16:40

Diagnostics Check		OK	02:05	
Air Blank		0.000	02:05	
Control Test		0.078	02:06	
Air Blank		0.000	02:06	
Subject Sample	# 1		02:09	No Sample Provided
Breath Volume		0.000	02:09	
Air Blank		0.000	02:10	
Air Blank		0.000	02:12	
Subject Sample	# 2		02:15	No Sample Provided
Breath Volume		0.000	02:15	
Air Blank		0.000	02:15	
Control Test		0.079	02:16	
Air Blank		0.000	02:16	
Diagnostics Check		OK	02:16	

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:26
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:19

SUBJECT'S LAST NAME: PONCE

DATE/TIME OF TEST: 02/18/2018 03:49:10

Diagnostics Check		OK	03:42
Air Blank		0.000	03:43
Control Test		0.079	03:43
Air Blank		0.000	03:44
Subject Sample	# 1	0.140	03:44
Breath Volume		1.636	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.136	03:47
Breath Volume		1.750	03:47
Air Blank		0.000	03:48
Control Test		0.079	03:48
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KOSH,TOBI,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: CAMPBELL,CRAIG
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:15

SUBJECT'S LAST NAME: SOTO

DATE/TIME OF TEST: 02/18/2018 04:45:04

Diagnostics Check		OK	04:38
Air Blank		0.000	04:38
Control Test		0.079	04:38
Air Blank		0.000	04:39
Subject Sample	# 1	0.127	04:39
Breath Volume		4.402	04:39
Air Blank		0.000	04:40
Air Blank		0.000	04:42
Subject Sample	# 2	0.129	04:43
Breath Volume		4.355	04:43
Air Blank		0.000	04:44
Control Test		0.079	04:44
Air Blank		0.000	04:44
Diagnostics Check		OK	04:45

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 02/18/2018 06:06:44

Diagnostics Check		OK	05:51	
Air Blank		0.000	05:52	
Control Test		0.079	05:52	
Air Blank		0.000	05:53	
Subject Sample	# 1	0.113	05:56	Volume Not Met
Breath Volume		0.039	05:56	
Air Blank		0.000	05:57	
Air Blank		0.000	05:58	
Subject Sample	# 2	0.123	05:59	
Breath Volume		2.265	05:59	
Air Blank		0.000	06:00	
Air Blank		0.000	06:01	
Subject Sample	# 3	0.105	06:05	Volume Not Met
Breath Volume		0.382	06:05	
Air Blank		0.000	06:05	
Control Test		0.079	06:06	
Air Blank		0.000	06:06	
Diagnostics Check		OK	06:06	

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:27

SUBJECT'S LAST NAME: LONG

DATE/TIME OF TEST: 02/18/2018 06:15:33

Diagnostics Check		OK	06:08
Air Blank		0.000	06:09
Control Test		0.078	06:09
Air Blank		0.000	06:10
Subject Sample	# 1	0.112	06:10
Breath Volume		1.308	06:10
Air Blank		0.000	06:11
Air Blank		0.000	06:13
Subject Sample	# 2	0.114	06:13
Breath Volume		1.167	06:13
Air Blank		0.000	06:14
Control Test		0.079	06:14
Air Blank		0.000	06:15
Diagnostics Check		OK	06:15

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RODRIGUEZ,MICHAEL,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ALLEN,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:23

SUBJECT'S LAST NAME: DESANTIS

DATE/TIME OF TEST: 02/18/2018 10:51:21

Diagnostics Check		OK	10:45
Air Blank		0.000	10:45
Control Test		0.079	10:46
Air Blank		0.000	10:46
Subject Sample	# 1	0.072	10:47
Breath Volume		2.082	10:47
Air Blank		0.000	10:47
Air Blank		0.000	10:49
Subject Sample	# 2	0.073	10:49
Breath Volume		2.378	10:49
Air Blank		0.000	10:50
Control Test		0.080	10:50
Air Blank		0.000	10:51
Diagnostics Check		OK	10:51

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:19
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: SHEPHERD

DATE/TIME OF TEST: 02/19/2018 01:31:56

Diagnostics Check		OK	01:25
Air Blank		0.000	01:25
Control Test		0.080	01:25
Air Blank		0.000	01:26
Subject Sample	# 1	0.098	01:27
Breath Volume		2.207	01:27
Air Blank		0.000	01:27
Air Blank		0.000	01:29
Subject Sample	# 2	0.096	01:30
Breath Volume		1.210	01:30
Air Blank		0.000	01:30
Control Test		0.078	01:31
Air Blank		0.000	01:31
Diagnostics Check		OK	01:31

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: PEREZ,MICHAEL
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:50
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:38

SUBJECT'S LAST NAME: BRIDGES

DATE/TIME OF TEST: 02/20/2018 05:19:23

Diagnostics Check		OK	05:04	
Air Blank		0.000	05:04	
Control Test		0.079	05:05	
Air Blank		0.000	05:05	
Subject Sample	# 1	0.000	05:08	Volume Not Met
Breath Volume		0.464	05:08	
Air Blank		0.000	05:09	
Air Blank		0.000	05:11	
Subject Sample	# 2	0.000	05:12	
Breath Volume		1.761	05:12	
Air Blank		0.000	05:12	
Air Blank		0.000	05:14	
Subject Sample	# 3	0.000	05:17	Volume Not Met
Breath Volume		0.628	05:17	
Air Blank		0.000	05:18	
Control Test		0.080	05:18	
Air Blank		0.000	05:19	
Diagnostics Check		OK	05:19	

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DORAN,CONNOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:01
GENDER: Female
OBSERVATION PERIOD BEGAN: 00:31

SUBJECT'S LAST NAME: CLARK-WEBER

DATE/TIME OF TEST: 02/21/2018 01:02:21

Diagnostics Check		OK	00:54
Air Blank		0.000	00:54
Control Test		0.080	00:54
Air Blank		0.000	00:55
Subject Sample	# 1	0.106	00:57
Breath Volume		1.199	00:57
Air Blank		0.000	00:58
Air Blank		0.000	01:00
Subject Sample	# 2	0.095	01:00
Breath Volume		1.570	01:00
Air Blank		0.000	01:01
Control Test		0.077	01:01
Air Blank		0.000	01:02
Diagnostics Check		OK	01:02

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:20

SUBJECT'S LAST NAME: SIMMONS

DATE/TIME OF TEST: 02/21/2018 03:52:09

Diagnostics Check		OK	03:44
Air Blank		0.000	03:45
Control Test		0.080	03:45
Air Blank		0.000	03:46
Subject Sample	# 1	0.232	03:47
Breath Volume		1.199	03:47
Air Blank		0.000	03:47
Air Blank		0.000	03:49
Subject Sample	# 2	0.227	03:50
Breath Volume		1.597	03:50
Air Blank		0.000	03:51
Control Test		0.079	03:51
Air Blank		0.000	03:52
Diagnostics Check		OK	03:52

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GORDON,CURTIS
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 17:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:00

SUBJECT'S LAST NAME: LACERDA

DATE/TIME OF TEST: 02/21/2018 18:30:38

Diagnostics Check		OK	18:24
Air Blank		0.000	18:24
Control Test		0.080	18:24
Air Blank		0.000	18:25
Subject Sample	# 1	0.000	18:26
Breath Volume		1.183	18:26
Air Blank		0.000	18:26
Air Blank		0.000	18:28
Subject Sample	# 2	0.000	18:29
Breath Volume		1.199	18:29
Air Blank		0.000	18:29
Control Test		0.079	18:30
Air Blank		0.000	18:30
Diagnostics Check		OK	18:30

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:54
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:39

SUBJECT'S LAST NAME: SIMONSEBASTION

DATE/TIME OF TEST: 02/22/2018 00:12:29

Diagnostics Check		OK	00:06
Air Blank		0.000	00:06
Control Test		0.079	00:07
Air Blank		0.000	00:07
Subject Sample	# 1	0.128	00:08
Breath Volume		1.593	00:08
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 2	0.125	00:10
Breath Volume		2.648	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: DRACH,STEFAN
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: SANTANA

DATE/TIME OF TEST: 02/22/2018 04:04:32

Diagnostics Check		OK	03:58
Air Blank		0.000	03:58
Control Test		0.079	03:58
Air Blank		0.000	03:59
Subject Sample	# 1	0.156	03:59
Breath Volume		2.750	03:59
Air Blank		0.000	04:00
Air Blank		0.000	04:02
Subject Sample	# 2	0.154	04:02
Breath Volume		2.722	04:02
Air Blank		0.000	04:03
Control Test		0.079	04:03
Air Blank		0.000	04:04
Diagnostics Check		OK	04:04

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: OKEEFE,ALEXIS,J
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: FIGUEROA,HECTOR
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 05:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 06:20

SUBJECT'S LAST NAME: GUNESH

DATE/TIME OF TEST: 02/22/2018 06:48:37

Diagnostics Check		OK	06:42
Air Blank		0.000	06:42
Control Test		0.079	06:42
Air Blank		0.000	06:43
Subject Sample	# 1	0.091	06:44
Breath Volume		1.667	06:44
Air Blank		0.000	06:44
Air Blank		0.000	06:46
Subject Sample	# 2	0.090	06:46
Breath Volume		2.093	06:46
Air Blank		0.000	06:47
Control Test		0.079	06:47
Air Blank		0.000	06:48
Diagnostics Check		OK	06:48

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RODRIGUEZ,STEPHANIE
ARRESTING OFFICER AGENCY: Maitland PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:10
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:50

SUBJECT'S LAST NAME: COLLINS

DATE/TIME OF TEST: 02/23/2018 01:19:21

Diagnostics Check		OK	01:12
Air Blank		0.000	01:13
Control Test		0.080	01:13
Air Blank		0.000	01:14
Subject Sample	# 1	0.180	01:14
Breath Volume		2.734	01:14
Air Blank		0.000	01:15
Air Blank		0.000	01:17
Subject Sample	# 2	0.180	01:17
Breath Volume		2.742	01:17
Air Blank		0.000	01:18
Control Test		0.079	01:18
Air Blank		0.000	01:19
Diagnostics Check		OK	01:19

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TIAMSON,DAWN,P
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: GONZALEZ,GEORGE
ARRESTING OFFICER AGENCY: Windermere PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:09
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:56

SUBJECT'S LAST NAME: PIGNOLI

DATE/TIME OF TEST: 02/23/2018 02:26:14

Diagnostics Check		OK	02:20
Air Blank		0.000	02:20
Control Test		0.080	02:20
Air Blank		0.000	02:21
Subject Sample	# 1	0.206	02:21
Breath Volume		2.390	02:21
Air Blank		0.000	02:22
Air Blank		0.000	02:24
Subject Sample	# 2	0.205	02:24
Breath Volume		1.730	02:24
Air Blank		0.000	02:25
Control Test		0.079	02:25
Air Blank		0.000	02:26
Diagnostics Check		OK	02:26

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: AUSTIN,ADAM
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:05

SUBJECT'S LAST NAME: RAIS

DATE/TIME OF TEST: 02/24/2018 01:39:34

Diagnostics Check		OK	01:33
Air Blank		0.000	01:33
Control Test		0.079	01:33
Air Blank		0.000	01:34
Subject Sample	# 1	0.220	01:35
Breath Volume		2.484	01:35
Air Blank		0.000	01:35
Air Blank		0.000	01:37
Subject Sample	# 2	0.220	01:38
Breath Volume		2.320	01:38
Air Blank		0.000	01:38
Control Test		0.077	01:38
Air Blank		0.000	01:39
Diagnostics Check		OK	01:39

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: WONG SHUE,ANTHONY
ARRESTING OFFICER AGENCY: Orlando PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:00
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/24/2018 02:47:31

Diagnostics Check		OK	02:38
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Subject Sample	# 1	0.167	02:42
Breath Volume		1.941	02:42
Air Blank		0.000	02:43
Air Blank		0.000	02:44
Subject Sample	# 2	0.172	02:45
Breath Volume		1.519	02:45
Air Blank		0.000	02:46
Control Test		0.078	02:46
Air Blank		0.000	02:47
Diagnostics Check		OK	02:47

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MENDOZA,ANGEL
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:35

SUBJECT'S LAST NAME: GONZALEZ ARAMBULA

DATE/TIME OF TEST: 02/24/2018 03:05:14

Diagnostics Check		OK	02:57
Air Blank		0.000	02:58
Control Test		0.078	02:58
Air Blank		0.000	02:59
Subject Sample	# 1	0.208	03:00
Breath Volume		1.644	03:00
Air Blank		0.000	03:01
Air Blank		0.000	03:03
Subject Sample	# 2	0.203	03:03
Breath Volume		1.148	03:03
Air Blank		0.000	03:04
Control Test		0.078	03:04
Air Blank		0.000	03:05
Diagnostics Check		OK	03:05

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:59

SUBJECT'S LAST NAME: OATES

DATE/TIME OF TEST: 02/24/2018 03:31:18

Diagnostics Check		OK	03:24
Air Blank		0.000	03:25
Control Test		0.078	03:25
Air Blank		0.000	03:25
Subject Sample	# 1	0.224	03:26
Breath Volume		1.558	03:26
Air Blank		0.000	03:27
Air Blank		0.000	03:29
Subject Sample	# 2	0.235	03:29
Breath Volume		2.058	03:29
Air Blank		0.000	03:30
Control Test		0.078	03:30
Air Blank		0.000	03:31
Diagnostics Check		OK	03:31

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: MORALES,JOSE
ARRESTING OFFICER AGENCY: Central FL Univ PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:25

SUBJECT'S LAST NAME: SAN JUAN

DATE/TIME OF TEST: 02/25/2018 00:55:52

Diagnostics Check		OK	00:49
Air Blank		0.000	00:49
Control Test		0.079	00:50
Air Blank		0.000	00:50
Subject Sample	# 1	0.296	00:51
Breath Volume		3.144	00:51
Air Blank		0.000	00:52
Air Blank		0.000	00:53
Subject Sample	# 2	0.279	00:54
Breath Volume		2.125	00:54
Air Blank		0.000	00:54
Control Test		0.078	00:55
Air Blank		0.000	00:55
Diagnostics Check		OK	00:55

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: THOMPSON,CALEB,N
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: KREKEL,ALEC
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:23

SUBJECT'S LAST NAME: SCHARDT

DATE/TIME OF TEST: 02/25/2018 01:54:05

Diagnostics Check		OK	01:47
Air Blank		0.000	01:47
Control Test		0.079	01:48
Air Blank		0.000	01:48
Subject Sample	# 1	0.178	01:49
Breath Volume		1.726	01:49
Air Blank		0.000	01:50
Air Blank		0.000	01:51
Subject Sample	# 2	0.158	01:52
Breath Volume		1.269	01:52
Air Blank		0.000	01:53
Control Test		0.078	01:53
Air Blank		0.000	01:54
Diagnostics Check		OK	01:54

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ELBON,NICHOLAS,C
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAWKINS,DEBORAH
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:20
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:19

SUBJECT'S LAST NAME: VINCENTE

DATE/TIME OF TEST: 02/25/2018 05:51:29

Diagnostics Check		OK	05:44
Air Blank		0.000	05:44
Control Test		0.080	05:45
Air Blank		0.000	05:45
Subject Sample	# 1	0.105	05:47
Breath Volume		1.570	05:47
Air Blank		0.000	05:47
Air Blank		0.000	05:49
Subject Sample	# 2	0.106	05:49
Breath Volume		1.902	05:49
Air Blank		0.000	05:50
Control Test		0.078	05:50
Air Blank		0.000	05:51
Diagnostics Check		OK	05:51

Data Download Date/Time: 03/13/2018 10:34

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: ELLER,ASHLEY
ARRESTING OFFICER AGENCY: Apopka PD
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:48
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:20

SUBJECT'S LAST NAME: SCOTT

DATE/TIME OF TEST: 02/27/2018 01:47:23

Diagnostics Check		OK	01:40
Air Blank		0.000	01:41
Control Test		0.080	01:41
Air Blank		0.000	01:42
Subject Sample	# 1	0.152	01:42
Breath Volume		2.503	01:42
Air Blank		0.000	01:43
Air Blank		0.000	01:45
Subject Sample	# 2	0.165	01:45
Breath Volume		3.902	01:45
Air Blank		0.000	01:46
Control Test		0.079	01:46
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 03/13/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANER,OSVALDO,
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: HAIGHT,MONTAGUE
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 09:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 10:01

SUBJECT'S LAST NAME: MONTES BAEZ

DATE/TIME OF TEST: 02/27/2018 10:31:27

Diagnostics Check		OK	10:24
Air Blank		0.000	10:25
Control Test		0.080	10:25
Air Blank		0.000	10:26
Subject Sample	# 1	0.000	10:27
Breath Volume		2.617	10:27
Air Blank		0.000	10:27
Air Blank		0.000	10:29
Subject Sample	# 2	0.000	10:30
Breath Volume		2.035	10:30
Air Blank		0.000	10:30
Control Test		0.079	10:30
Air Blank		0.000	10:31
Diagnostics Check		OK	10:31

Data Download Date/Time: 03/13/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: RIVERA,VICTOR
ARRESTING OFFICER AGENCY: FHP Troop D
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:53
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:46

SUBJECT'S LAST NAME: VINCA

DATE/TIME OF TEST: 02/28/2018 02:14:02

Diagnostics Check		OK	02:07
Air Blank		0.000	02:08
Control Test		0.080	02:08
Air Blank		0.000	02:09
Subject Sample	# 1	0.156	02:09
Breath Volume		2.179	02:09
Air Blank		0.000	02:10
Air Blank		0.000	02:12
Subject Sample	# 2	0.158	02:12
Breath Volume		3.011	02:12
Air Blank		0.000	02:13
Control Test		0.080	02:13
Air Blank		0.000	02:13
Diagnostics Check		OK	02:14

Data Download Date/Time: 03/13/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001420

INSTRUMENT REGISTERED TO: ORANGE COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DENSON,TRAVIS,W
BREATH TEST OPERATOR AGENCY: Orange County SO
ARRESTING OFFICER: VAN EPPS,KELLY
ARRESTING OFFICER AGENCY: Orange County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A1
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 18:05
GENDER: Female
OBSERVATION PERIOD BEGAN: 18:21

SUBJECT'S LAST NAME: ADAMSON

DATE/TIME OF TEST: 02/28/2018 18:52:59

Diagnostics Check		OK	18:46
Air Blank		0.000	18:46
Control Test		0.080	18:47
Air Blank		0.000	18:47
Subject Sample	# 1	0.180	18:48
Breath Volume		1.882	18:48
Air Blank		0.000	18:49
Air Blank		0.000	18:50
Subject Sample	# 2	0.176	18:51
Breath Volume		1.871	18:51
Air Blank		0.000	18:52
Control Test		0.079	18:52
Air Blank		0.000	18:52
Diagnostics Check		OK	18:52

Data Download Date/Time: 03/13/2018 10:35

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 01:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:20

SUBJECT'S LAST NAME: TEDESCO

DATE/TIME OF TEST: 02/01/2018 02:54:13

Diagnostics Check		OK	02:47
Air Blank		0.000	02:48
Control Test		0.082	02:48
Air Blank		0.000	02:49
Subject Sample	# 1	0.222	02:49
Breath Volume		1.605	02:49
Air Blank		0.000	02:50
Air Blank		0.000	02:52
Subject Sample	# 2	0.216	02:52
Breath Volume		1.691	02:52
Air Blank		0.000	02:53
Control Test		0.081	02:53
Air Blank		0.000	02:54
Diagnostics Check		OK	02:54

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ALTSTATT,MATTHEW,B
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: LOVETT,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 22:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:49

SUBJECT'S LAST NAME: MILLHONE

DATE/TIME OF TEST: 02/03/2018 23:17:07

Diagnostics Check		OK	23:09
Air Blank		0.000	23:09
Control Test		0.082	23:10
Air Blank		0.000	23:10
Subject Sample	# 1	0.259	23:12
Breath Volume		2.238	23:12
Air Blank		0.000	23:12
Air Blank		0.000	23:14
Subject Sample	# 2	0.258	23:15
Breath Volume		2.089	23:15
Air Blank		0.000	23:16
Control Test		0.082	23:16
Air Blank		0.000	23:17
Diagnostics Check		OK	23:17

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 02/10/2018 02:02:31

Diagnostics Check		OK	01:51	
Air Blank		0.000	01:51	
Control Test		0.082	01:51	
Air Blank		0.000	01:52	
Subject Sample	# 1		01:55	No Sample Provided
Breath Volume		0.000	01:55	
Air Blank		0.000	01:55	
Air Blank		0.000	01:57	
Subject Sample	# 2	0.144	02:01	Volume Not Met
Breath Volume		0.019	02:01	
Air Blank		0.000	02:01	
Control Test		0.081	02:01	
Air Blank		0.000	02:02	
Diagnostics Check		OK	02:02	

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: WHITE,M
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:29

SUBJECT'S LAST NAME: PAUL

DATE/TIME OF TEST: 02/10/2018 02:11:00

Diagnostics Check		OK	02:03
Air Blank		0.000	02:04
Control Test		0.080	02:04
Air Blank		0.000	02:04
Subject Sample	# 1	0.164	02:06
Breath Volume		1.429	02:06
Air Blank		0.000	02:06
Air Blank		0.000	02:08
Subject Sample	# 2	0.166	02:09
Breath Volume		1.589	02:09
Air Blank		0.000	02:10
Control Test		0.081	02:10
Air Blank		0.000	02:10
Diagnostics Check		OK	02:11

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SWIM,G
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:46

SUBJECT'S LAST NAME: HORNE

DATE/TIME OF TEST: 02/10/2018 03:15:07

Diagnostics Check		OK	03:07
Air Blank		0.000	03:07
Control Test		0.081	03:07
Air Blank		0.000	03:08
Subject Sample	# 1	0.000	03:10
Breath Volume		1.109	03:10
Air Blank		0.000	03:10
Air Blank		0.000	03:12
Subject Sample	# 2	0.000	03:13
Breath Volume		2.750	03:13
Air Blank		0.000	03:14
Control Test		0.081	03:14
Air Blank		0.000	03:15
Diagnostics Check		OK	03:15

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:35

SUBJECT'S LAST NAME: STRAUGHN

DATE/TIME OF TEST: 02/10/2018 04:03:17

Diagnostics Check		OK	03:55
Air Blank		0.000	03:56
Control Test		0.080	03:56
Air Blank		0.000	03:56
Subject Sample	# 1	0.166	03:58
Breath Volume		1.675	03:58
Air Blank		0.000	03:58
Air Blank		0.000	04:00
Subject Sample	# 2	0.152	04:01
Breath Volume		1.300	04:01
Air Blank		0.000	04:02
Control Test		0.080	04:02
Air Blank		0.000	04:03
Diagnostics Check		OK	04:03

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: BAINE,D
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 21:57
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:39

SUBJECT'S LAST NAME: MILLER

DATE/TIME OF TEST: 02/10/2018 23:07:11

Diagnostics Check		OK	22:59
Air Blank		0.000	23:00
Control Test		0.080	23:00
Air Blank		0.000	23:00
Subject Sample	# 1	0.200	23:01
Breath Volume		1.371	23:01
Air Blank		0.000	23:02
Air Blank		0.000	23:04
Subject Sample	# 2	0.203	23:05
Breath Volume		1.410	23:05
Air Blank		0.000	23:06
Control Test		0.080	23:06
Air Blank		0.000	23:07
Diagnostics Check		OK	23:07

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: FORD,A
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:49
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:04

SUBJECT'S LAST NAME: DERVISHI

DATE/TIME OF TEST: 02/11/2018 02:33:37

Diagnostics Check		OK	02:27
Air Blank		0.000	02:27
Control Test		0.081	02:28
Air Blank		0.000	02:28
Subject Sample	# 1	0.094	02:29
Breath Volume		1.792	02:29
Air Blank		0.000	02:29
Air Blank		0.000	02:31
Subject Sample	# 2	0.093	02:32
Breath Volume		1.769	02:32
Air Blank		0.000	02:32
Control Test		0.080	02:33
Air Blank		0.000	02:33
Diagnostics Check		OK	02:33

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SIMMONS,K
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 01/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Male
OBSERVATION PERIOD BEGAN: 04:18

SUBJECT'S LAST NAME: KENDZIERSKI

DATE/TIME OF TEST: 02/15/2018 04:58:39

Diagnostics Check		OK	04:47	
Air Blank		0.000	04:47	
Control Test		0.081	04:48	
Air Blank		0.000	04:48	
Subject Sample	# 1		04:51	No Sample Provided
Breath Volume		0.000	04:51	
Air Blank		0.000	04:52	
Air Blank		0.000	04:54	
Subject Sample	# 2		04:57	No Sample Provided
Breath Volume		0.000	04:57	
Air Blank		0.000	04:57	
Control Test		0.081	04:58	
Air Blank		0.000	04:58	
Diagnostics Check		OK	04:58	

Data Download Date/Time: 02/19/2018 16:47

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: PINSON,T
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 18:38
GENDER: Male
OBSERVATION PERIOD BEGAN: 19:17

SUBJECT'S LAST NAME: HALL

DATE/TIME OF TEST: 02/19/2018 19:46:28

Diagnostics Check		OK	19:39
Air Blank		0.000	19:40
Control Test		0.081	19:40
Air Blank		0.000	19:40
Subject Sample	# 1	0.292	19:41
Breath Volume		1.996	19:41
Air Blank		0.000	19:42
Air Blank		0.000	19:44
Subject Sample	# 2	0.289	19:45
Breath Volume		2.472	19:45
Air Blank		0.000	19:45
Control Test		0.080	19:45
Air Blank		0.000	19:46
Diagnostics Check		OK	19:46

Data Download Date/Time: 03/24/2018 07:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: STIVERS,H
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:55

SUBJECT'S LAST NAME: MADDOX

DATE/TIME OF TEST: 02/20/2018 04:21:21

Diagnostics Check		OK	04:15
Air Blank		0.000	04:15
Control Test		0.081	04:15
Air Blank		0.000	04:16
Subject Sample	# 1	0.204	04:17
Breath Volume		3.113	04:17
Air Blank		0.000	04:17
Air Blank		0.000	04:19
Subject Sample	# 2	0.200	04:19
Breath Volume		2.000	04:19
Air Blank		0.000	04:20
Control Test		0.080	04:20
Air Blank		0.000	04:21
Diagnostics Check		OK	04:21

Data Download Date/Time: 03/24/2018 07:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: VELETANLIC,S
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 19:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:14

SUBJECT'S LAST NAME: LEE

DATE/TIME OF TEST: 02/24/2018 20:41:27

Diagnostics Check		OK	20:34
Air Blank		0.000	20:35
Control Test		0.082	20:35
Air Blank		0.000	20:36
Subject Sample	# 1	0.051	20:36
Breath Volume		1.414	20:36
Air Blank		0.000	20:37
Air Blank		0.000	20:39
Subject Sample	# 2	0.050	20:40
Breath Volume		1.308	20:40
Air Blank		0.000	20:40
Control Test		0.082	20:40
Air Blank		0.000	20:41
Diagnostics Check		OK	20:41

Data Download Date/Time: 03/24/2018 07:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BRAUNLICH,G,A
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: CRUZ,C
ARRESTING OFFICER AGENCY: Orange Park PD
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 20:27
GENDER: Male
OBSERVATION PERIOD BEGAN: 20:50

SUBJECT'S LAST NAME: PORTER

DATE/TIME OF TEST: 02/25/2018 21:22:29

Diagnostics Check		OK	21:11	
Air Blank		0.000	21:11	
Control Test		0.081	21:11	
Air Blank		0.000	21:12	
Subject Sample	# 1		21:15	No Sample Provided
Breath Volume		0.000	21:15	
Air Blank		0.000	21:15	
Air Blank		0.000	21:18	
Subject Sample	# 2		21:21	No Sample Provided
Breath Volume		0.000	21:21	
Air Blank		0.000	21:21	
Control Test		0.081	21:21	
Air Blank		0.000	21:22	
Diagnostics Check		OK	21:22	

Data Download Date/Time: 03/24/2018 07:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001443

INSTRUMENT REGISTERED TO: CLAY COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BELL,DAKOTA,W
BREATH TEST OPERATOR AGENCY: Clay County SO
ARRESTING OFFICER: SOLIS,E
ARRESTING OFFICER AGENCY: Clay County SO
LAST AGENCY INSPECTION: 02/19/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 09117080A2
DRY GAS EXPIRY DATE: 06/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:33
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:50

SUBJECT'S LAST NAME: ROBERTSON

DATE/TIME OF TEST: 02/26/2018 00:17:55

Diagnostics Check		OK	00:10
Air Blank		0.000	00:10
Control Test		0.082	00:11
Air Blank		0.000	00:11
Subject Sample	# 1	0.187	00:13
Breath Volume		1.328	00:13
Air Blank		0.000	00:13
Air Blank		0.000	00:15
Subject Sample	# 2	0.184	00:16
Breath Volume		1.265	00:16
Air Blank		0.000	00:17
Control Test		0.081	00:17
Air Blank		0.000	00:17
Diagnostics Check		OK	00:17

Data Download Date/Time: 03/24/2018 07:42

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: HARDY,JOE,E
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: HARDY,JOE
ARRESTING OFFICER AGENCY: FHP Troop A
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 17:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 17:00

SUBJECT'S LAST NAME: SOUTHWELL

DATE/TIME OF TEST: 02/07/2018 17:46:40

Diagnostics Check		OK	17:40
Air Blank		0.000	17:41
Control Test		0.080	17:41
Air Blank		0.000	17:41
Subject Sample	# 1	0.147	17:42
Breath Volume		1.910	17:42
Air Blank		0.000	17:42
Air Blank		0.000	17:44
Subject Sample	# 2	0.141	17:45
Breath Volume		1.597	17:45
Air Blank		0.000	17:45
Control Test		0.078	17:46
Air Blank		0.000	17:46
Diagnostics Check		OK	17:46

Data Download Date/Time: 02/25/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001646

INSTRUMENT REGISTERED TO: TYNDALL AFB
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARGE,SUSAN,M
BREATH TEST OPERATOR AGENCY: FHP Troop A
ARRESTING OFFICER: BIGGS,JOSHUA
ARRESTING OFFICER AGENCY: Jackson County SO
LAST AGENCY INSPECTION: 02/02/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 770784
DRY GAS EXPIRY DATE: 04/27/2020
VIOLATION CODE: DUI
ARREST TIME: 20:52
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:20

SUBJECT'S LAST NAME: MEIER

DATE/TIME OF TEST: 02/22/2018 23:57:09

Diagnostics Check		OK	23:50
Air Blank		0.000	23:50
Control Test		0.081	23:51
Air Blank		0.000	23:51
Subject Sample	# 1	0.199	23:52
Breath Volume		1.320	23:52
Air Blank		0.000	23:52
Air Blank		0.000	23:54
Subject Sample	# 2	0.212	23:55
Breath Volume		1.964	23:55
Air Blank		0.000	23:56
Control Test		0.081	23:56
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 02/25/2018 19:52

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SUSKEVICH,CURT,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: HAMILTON,C
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:38
GENDER: Female
OBSERVATION PERIOD BEGAN: 21:55

SUBJECT'S LAST NAME: FOSTER

DATE/TIME OF TEST: 02/02/2018 22:27:03

Diagnostics Check		OK	22:20
Air Blank		0.000	22:20
Control Test		0.079	22:21
Air Blank		0.000	22:21
Subject Sample	# 1	0.000	22:22
Breath Volume		2.679	22:22
Air Blank		0.000	22:23
Air Blank		0.000	22:25
Subject Sample	# 2	0.000	22:25
Breath Volume		2.832	22:25
Air Blank		0.000	22:26
Control Test		0.079	22:26
Air Blank		0.000	22:26
Diagnostics Check		OK	22:27

Data Download Date/Time: 02/14/2018 08:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOOS,SAVANNAH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:27

SUBJECT'S LAST NAME: LIMBAUGH

DATE/TIME OF TEST: 02/04/2018 04:01:24

Diagnostics Check		OK	03:49	
Air Blank		0.000	03:50	
Control Test		0.079	03:50	
Air Blank		0.000	03:51	
Subject Sample	# 1	0.217	03:54	Volume Not Met
Breath Volume		0.097	03:54	
Air Blank		0.000	03:54	
Air Blank		0.000	03:56	
Subject Sample	# 2	0.209	03:59	Volume Not Met
Breath Volume		0.535	03:59	
Air Blank		0.000	04:00	
Control Test		0.077	04:00	
Air Blank		0.000	04:01	
Diagnostics Check		OK	04:01	

Data Download Date/Time: 02/14/2018 08:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,FRANK
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:59
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: KEARNS

DATE/TIME OF TEST: 02/08/2018 01:36:17

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.081	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:35	
Control Test		0.080	01:35	
Air Blank		0.000	01:36	
Diagnostics Check		OK	01:36	

Data Download Date/Time: 02/14/2018 08:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,HAILEY,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TUCKER,HAILEY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:42
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:12

SUBJECT'S LAST NAME: WHITE

DATE/TIME OF TEST: 02/10/2018 01:35:37

Diagnostics Check		OK	01:32	
Air Blank		0.000	01:33	
Control Test		0.080	01:33	
Air Blank		0.000	01:33	
Subject Sample	# 1		01:34	Subject Test Refused
Breath Volume		0.000	01:34	
Air Blank		0.000	01:34	
Control Test		0.080	01:35	
Air Blank		0.000	01:35	
Diagnostics Check		OK	01:35	

Data Download Date/Time: 02/14/2018 08:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,HAILEY,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ESTERLE,HUGH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:04
GENDER: Female
OBSERVATION PERIOD BEGAN: 02:10

SUBJECT'S LAST NAME: CASTILLO

DATE/TIME OF TEST: 02/10/2018 02:39:12

Diagnostics Check		OK	02:32
Air Blank		0.000	02:32
Control Test		0.079	02:33
Air Blank		0.000	02:33
Subject Sample	# 1	0.125	02:34
Breath Volume		3.144	02:34
Air Blank		0.000	02:35
Air Blank		0.000	02:37
Subject Sample	# 2	0.125	02:37
Breath Volume		3.222	02:37
Air Blank		0.000	02:38
Control Test		0.079	02:38
Air Blank		0.000	02:39
Diagnostics Check		OK	02:39

Data Download Date/Time: 02/14/2018 08:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TUCKER,HAILEY,E
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: LOETHEN,KEITH
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 01/10/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:04

SUBJECT'S LAST NAME: BOX

DATE/TIME OF TEST: 02/10/2018 23:32:22

Diagnostics Check		OK	23:25
Air Blank		0.000	23:26
Control Test		0.080	23:26
Air Blank		0.000	23:27
Subject Sample	# 1	0.000	23:27
Breath Volume		2.320	23:27
Air Blank		0.000	23:28
Air Blank		0.000	23:30
Subject Sample	# 2	0.000	23:31
Breath Volume		2.644	23:31
Air Blank		0.000	23:31
Control Test		0.080	23:31
Air Blank		0.000	23:32
Diagnostics Check		OK	23:32

Data Download Date/Time: 02/14/2018 08:54

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SCARLATO,M
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:51
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:05

SUBJECT'S LAST NAME: TARDIE

DATE/TIME OF TEST: 02/17/2018 21:33:40

Diagnostics Check		OK	21:28	
Air Blank		0.000	21:29	
Control Test		0.080	21:29	
Air Blank		0.000	21:29	
Subject Sample	# 1		21:32	Subject Test Refused
Breath Volume		0.000	21:32	
Air Blank		0.000	21:32	
Control Test		0.080	21:33	
Air Blank		0.000	21:33	
Diagnostics Check		OK	21:33	

Data Download Date/Time: 03/07/2018 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: TOBLER,AMANDA
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 00:56
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:17

SUBJECT'S LAST NAME: KILIAN

DATE/TIME OF TEST: 02/18/2018 01:47:57

Diagnostics Check		OK	01:39
Air Blank		0.000	01:40
Control Test		0.080	01:40
Air Blank		0.000	01:40
Subject Sample	# 1	0.186	01:43
Breath Volume		2.468	01:43
Air Blank		0.000	01:44
Air Blank		0.000	01:46
Subject Sample	# 2	0.186	01:46
Breath Volume		2.761	01:46
Air Blank		0.000	01:47
Control Test		0.079	01:47
Air Blank		0.000	01:47
Diagnostics Check		OK	01:47

Data Download Date/Time: 03/07/2018 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: RUSH,DANIEL,C
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: RUSH,D
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:42

SUBJECT'S LAST NAME: RUIZ

DATE/TIME OF TEST: 02/22/2018 00:13:50

Diagnostics Check		OK	00:07
Air Blank		0.000	00:07
Control Test		0.080	00:08
Air Blank		0.000	00:08
Subject Sample	# 1	0.148	00:09
Breath Volume		3.046	00:09
Air Blank		0.000	00:10
Air Blank		0.000	00:11
Subject Sample	# 2	0.146	00:12
Breath Volume		2.984	00:12
Air Blank		0.000	00:12
Control Test		0.078	00:13
Air Blank		0.000	00:13
Diagnostics Check		OK	00:13

Data Download Date/Time: 03/07/2018 10:01

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: TOBLER,AMANDA,R
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SHIPPY,DYLAN
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 20:46
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:06

SUBJECT'S LAST NAME: CASTILLO PAZ

DATE/TIME OF TEST: 02/22/2018 21:37:35

Diagnostics Check		OK	21:28
Air Blank		0.000	21:28
Control Test		0.079	21:28
Air Blank		0.000	21:29
Subject Sample	# 1	0.304	21:29
Breath Volume		2.562	21:29
Air Blank		0.000	21:30
Air Blank		0.000	21:32
Subject Sample	# 2	0.271	21:33
Breath Volume		1.484	21:33
Air Blank		0.000	21:33
Air Blank		0.000	21:35
Subject Sample	# 3	0.304	21:36
Breath Volume		2.800	21:36
Air Blank		0.000	21:36
Control Test		0.079	21:37
Air Blank		0.000	21:37
Diagnostics Check		OK	21:37

Data Download Date/Time: 03/07/2018 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: CANTRELL,ROBERT,V
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: ZELKO,DAVID
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 16:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 16:35

SUBJECT'S LAST NAME: SANTORA

DATE/TIME OF TEST: 02/23/2018 16:59:41

Diagnostics Check		OK	16:56	
Air Blank		0.000	16:57	
Control Test		0.081	16:57	
Air Blank		0.000	16:57	
Subject Sample	# 1		16:58	Subject Test Refused
Breath Volume		0.000	16:58	
Air Blank		0.000	16:58	
Control Test		0.080	16:59	
Air Blank		0.000	16:59	
Diagnostics Check		OK	16:59	

Data Download Date/Time: 03/07/2018 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: FILLAK,MICHAEL,O
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: SUPER,MICHAEL
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:30
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: BENING

DATE/TIME OF TEST: 02/24/2018 00:07:22

Diagnostics Check		OK	00:00
Air Blank		0.000	00:01
Control Test		0.080	00:01
Air Blank		0.000	00:01
Subject Sample	# 1	0.000	00:02
Breath Volume		2.710	00:02
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.000	00:06
Breath Volume		3.195	00:06
Air Blank		0.000	00:06
Control Test		0.080	00:06
Air Blank		0.000	00:07
Diagnostics Check		OK	00:07

Data Download Date/Time: 03/07/2018 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: DEBERARDINIS,DREW
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 02:17
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:34

SUBJECT'S LAST NAME: SMITH

DATE/TIME OF TEST: 02/24/2018 03:01:01

Diagnostics Check		OK	02:54
Air Blank		0.000	02:55
Control Test		0.079	02:55
Air Blank		0.000	02:55
Subject Sample	# 1	0.237	02:56
Breath Volume		2.613	02:56
Air Blank		0.000	02:57
Air Blank		0.000	02:58
Subject Sample	# 2	0.249	02:59
Breath Volume		3.414	02:59
Air Blank		0.000	03:00
Control Test		0.079	03:00
Air Blank		0.000	03:00
Diagnostics Check		OK	03:01

Data Download Date/Time: 03/07/2018 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001647

INSTRUMENT REGISTERED TO: CAPE CORAL PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: BARKLEY,JEFFREY,G
BREATH TEST OPERATOR AGENCY: Cape Coral PD
ARRESTING OFFICER: BARKLEY,JEFFREY
ARRESTING OFFICER AGENCY: Cape Coral PD
LAST AGENCY INSPECTION: 02/14/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 34416080A2
DRY GAS EXPIRY DATE: 02/05/2019
VIOLATION CODE: DUI
ARREST TIME: 00:31
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:59

SUBJECT'S LAST NAME: DIAZ

DATE/TIME OF TEST: 02/25/2018 01:27:28

Diagnostics Check		OK	01:21
Air Blank		0.000	01:21
Control Test		0.080	01:21
Air Blank		0.000	01:22
Subject Sample	# 1	0.154	01:23
Breath Volume		2.625	01:23
Air Blank		0.000	01:23
Air Blank		0.000	01:25
Subject Sample	# 2	0.146	01:26
Breath Volume		2.250	01:26
Air Blank		0.000	01:26
Control Test		0.080	01:26
Air Blank		0.000	01:27
Diagnostics Check		OK	01:27

Data Download Date/Time: 03/07/2018 10:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WALDO,AARON,R
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:25
GENDER: Female
OBSERVATION PERIOD BEGAN: 23:23

SUBJECT'S LAST NAME: BARRETT

DATE/TIME OF TEST: 02/09/2018 23:48:26

Diagnostics Check		OK	23:45	
Air Blank		0.000	23:46	
Control Test		0.079	23:46	
Air Blank		0.000	23:46	
Subject Sample	# 1		23:47	Subject Test Refused
Breath Volume		0.000	23:47	
Air Blank		0.000	23:47	
Control Test		0.079	23:47	
Air Blank		0.000	23:48	
Diagnostics Check		OK	23:48	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:44
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:21

SUBJECT'S LAST NAME: LEYKAM

DATE/TIME OF TEST: 02/10/2018 01:55:21

Diagnostics Check		OK	01:49
Air Blank		0.000	01:49
Control Test		0.079	01:50
Air Blank		0.000	01:50
Subject Sample	# 1	0.268	01:51
Breath Volume		1.328	01:51
Air Blank		0.000	01:51
Air Blank		0.000	01:53
Subject Sample	# 2	0.260	01:53
Breath Volume		1.378	01:53
Air Blank		0.000	01:54
Control Test		0.078	01:54
Air Blank		0.000	01:55
Diagnostics Check		OK	01:55

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:41
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:12

SUBJECT'S LAST NAME: MULLIGAN

DATE/TIME OF TEST: 02/10/2018 23:57:57

Diagnostics Check		OK	23:51
Air Blank		0.000	23:52
Control Test		0.079	23:52
Air Blank		0.000	23:52
Subject Sample	# 1	0.124	23:53
Breath Volume		2.449	23:53
Air Blank		0.000	23:53
Air Blank		0.000	23:55
Subject Sample	# 2	0.124	23:56
Breath Volume		1.433	23:56
Air Blank		0.000	23:57
Control Test		0.078	23:57
Air Blank		0.000	23:57
Diagnostics Check		OK	23:57

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: KARAYIANES,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:17
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:03

SUBJECT'S LAST NAME: DURHAM

DATE/TIME OF TEST: 02/15/2018 01:32:47

Diagnostics Check		OK	01:26
Air Blank		0.000	01:27
Control Test		0.078	01:27
Air Blank		0.000	01:27
Subject Sample	# 1	0.089	01:28
Breath Volume		3.015	01:28
Air Blank		0.000	01:28
Air Blank		0.000	01:30
Subject Sample	# 2	0.094	01:31
Breath Volume		2.429	01:31
Air Blank		0.000	01:31
Control Test		0.078	01:32
Air Blank		0.000	01:32
Diagnostics Check		OK	01:32

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:32
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:14

SUBJECT'S LAST NAME: FREDRICK

DATE/TIME OF TEST: 02/15/2018 02:48:34

Diagnostics Check		OK	02:45	
Air Blank		0.000	02:46	
Control Test		0.079	02:46	
Air Blank		0.000	02:47	
Subject Sample	# 1		02:47	Subject Test Refused
Breath Volume		0.000	02:47	
Air Blank		0.000	02:47	
Control Test		0.078	02:47	
Air Blank		0.000	02:48	
Diagnostics Check		OK	02:48	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:15
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:55

SUBJECT'S LAST NAME: TOMA

DATE/TIME OF TEST: 02/16/2018 01:28:13

Diagnostics Check		OK	01:24	
Air Blank		0.000	01:24	
Control Test		0.079	01:24	
Air Blank		0.000	01:25	
Subject Sample	# 1		01:26	Subject Test Refused
Breath Volume		0.000	01:26	
Air Blank		0.000	01:27	
Control Test		0.080	01:27	
Air Blank		0.000	01:28	
Diagnostics Check		OK	01:28	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:27

SUBJECT'S LAST NAME: SPENCE

DATE/TIME OF TEST: 02/16/2018 02:53:25

Diagnostics Check		OK	02:50	
Air Blank		0.000	02:50	
Control Test		0.079	02:51	
Air Blank		0.000	02:51	
Subject Sample	# 1		02:52	Subject Test Refused
Breath Volume		0.000	02:52	
Air Blank		0.000	02:52	
Control Test		0.079	02:52	
Air Blank		0.000	02:53	
Diagnostics Check		OK	02:53	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: KARAYIANES,MICHAEL,F
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SMITH,EMILY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 04:02
GENDER: Male
OBSERVATION PERIOD BEGAN: 03:30

SUBJECT'S LAST NAME: PELLEGRINO

DATE/TIME OF TEST: 02/16/2018 04:19:47

Diagnostics Check		OK	04:12
Air Blank		0.000	04:13
Control Test		0.078	04:13
Air Blank		0.000	04:14
Subject Sample	# 1	0.144	04:14
Breath Volume		3.277	04:14
Air Blank		0.000	04:15
Air Blank		0.000	04:17
Subject Sample	# 2	0.144	04:18
Breath Volume		3.277	04:18
Air Blank		0.000	04:18
Control Test		0.078	04:19
Air Blank		0.000	04:19
Diagnostics Check		OK	04:19

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:12
GENDER: Female
OBSERVATION PERIOD BEGAN: 01:57

SUBJECT'S LAST NAME: TENBROCK

DATE/TIME OF TEST: 02/22/2018 02:24:13

Diagnostics Check		OK	02:17
Air Blank		0.000	02:18
Control Test		0.077	02:18
Air Blank		0.000	02:18
Subject Sample	# 1	0.174	02:19
Breath Volume		2.171	02:19
Air Blank		0.000	02:20
Air Blank		0.000	02:22
Subject Sample	# 2	0.176	02:22
Breath Volume		2.058	02:22
Air Blank		0.000	02:23
Control Test		0.078	02:23
Air Blank		0.000	02:24
Diagnostics Check		OK	02:24

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:31
GENDER: Female
OBSERVATION PERIOD BEGAN: 19:20

SUBJECT'S LAST NAME: MOORE

DATE/TIME OF TEST: 02/22/2018 19:56:46

Diagnostics Check		OK	19:53	
Air Blank		0.000	19:54	
Control Test		0.077	19:54	
Air Blank		0.000	19:55	
Subject Sample	# 1		19:55	Subject Test Refused
Breath Volume		0.000	19:55	
Air Blank		0.000	19:55	
Control Test		0.078	19:56	
Air Blank		0.000	19:56	
Diagnostics Check		OK	19:56	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 01:19
GENDER: Male
OBSERVATION PERIOD BEGAN: 01:10

SUBJECT'S LAST NAME: ANDERSON

DATE/TIME OF TEST: 02/23/2018 01:33:25

Diagnostics Check		OK	01:30	
Air Blank		0.000	01:30	
Control Test		0.079	01:31	
Air Blank		0.000	01:31	
Subject Sample	# 1		01:32	Subject Test Refused
Breath Volume		0.000	01:32	
Air Blank		0.000	01:32	
Control Test		0.078	01:32	
Air Blank		0.000	01:33	
Diagnostics Check		OK	01:33	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 02/23/2018 02:46:07

Diagnostics Check		OK	02:43	
Air Blank		0.000	02:44	
Control Test		0.078	02:44	
Air Blank		0.000	02:45	
Subject Sample	# 1		02:45	Subject Test Refused
Breath Volume		0.000	02:45	
Air Blank			02:45	RFI Detect
Air Blank		0.000	02:46	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 02:22
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:15

SUBJECT'S LAST NAME: ROSA

DATE/TIME OF TEST: 02/23/2018 02:50:39

Diagnostics Check		OK	02:47	
Air Blank		0.000	02:48	
Control Test		0.078	02:48	
Air Blank		0.000	02:49	
Subject Sample	# 1		02:49	Subject Test Refused
Breath Volume		0.000	02:49	
Air Blank		0.000	02:49	
Control Test		0.079	02:50	
Air Blank		0.000	02:50	
Diagnostics Check		OK	02:50	

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: COX,LENARD
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 21:25
GENDER: Male
OBSERVATION PERIOD BEGAN: 21:15

SUBJECT'S LAST NAME: HARTWELL

DATE/TIME OF TEST: 02/23/2018 21:43:59

Diagnostics Check		OK	21:37
Air Blank		0.000	21:38
Control Test		0.079	21:38
Air Blank		0.000	21:39
Subject Sample	# 1	0.222	21:39
Breath Volume		2.289	21:39
Air Blank		0.000	21:40
Air Blank		0.000	21:42
Subject Sample	# 2	0.221	21:42
Breath Volume		2.375	21:42
Air Blank		0.000	21:43
Control Test		0.078	21:43
Air Blank		0.000	21:43
Diagnostics Check		OK	21:43

Data Download Date/Time: 03/06/2018 09:04

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SMITH,EMILY,K
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: ZOLLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 23:55
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:37

SUBJECT'S LAST NAME: CHAPMAN

DATE/TIME OF TEST: 02/24/2018 00:19:37

Diagnostics Check		OK	00:10
Air Blank		0.000	00:11
Control Test		0.078	00:11
Air Blank		0.000	00:12
Subject Sample	# 1	0.252	00:12
Breath Volume		1.539	00:12
Air Blank		0.000	00:12
Air Blank		0.000	00:14
Subject Sample	# 2	0.250	00:15
Breath Volume		1.691	00:15
Air Blank		0.000	00:15
Control Test		0.078	00:16
Air Blank		0.000	00:16
Diagnostics Check		OK	00:19

Data Download Date/Time: 03/06/2018 09:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:03
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:13

SUBJECT'S LAST NAME: FONSECA HOYOS

DATE/TIME OF TEST: 02/24/2018 03:35:45

Diagnostics Check		OK	03:29
Air Blank		0.000	03:30
Control Test		0.079	03:30
Air Blank		0.000	03:30
Subject Sample	# 1	0.000	03:31
Breath Volume		3.929	03:31
Air Blank		0.000	03:31
Air Blank		0.000	03:33
Subject Sample	# 2	0.000	03:34
Breath Volume		2.820	03:34
Air Blank		0.000	03:34
Control Test		0.079	03:35
Air Blank		0.000	03:35
Diagnostics Check		OK	03:35

Data Download Date/Time: 03/06/2018 09:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: COX,LENARD,E
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: Z0LLOTA,VLADIMIR
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 22:47
GENDER: Male
OBSERVATION PERIOD BEGAN: 22:25

SUBJECT'S LAST NAME: SULLIVAN

DATE/TIME OF TEST: 02/24/2018 22:51:05

Diagnostics Check		OK	22:48	
Air Blank		0.000	22:48	
Control Test		0.078	22:49	
Air Blank		0.000	22:49	
Subject Sample	# 1		22:49	Subject Test Refused
Breath Volume		0.000	22:49	
Air Blank		0.000	22:50	
Control Test		0.079	22:50	
Air Blank		0.000	22:50	
Diagnostics Check		OK	22:51	

Data Download Date/Time: 03/06/2018 09:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: SAIA,MICHAEL
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:05
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:50

SUBJECT'S LAST NAME: TAVARES

DATE/TIME OF TEST: 02/25/2018 03:20:59

Diagnostics Check		OK	03:13
Air Blank		0.000	03:13
Control Test		0.078	03:14
Air Blank		0.000	03:14
Subject Sample	# 1	0.183	03:16
Breath Volume		1.523	03:16
Air Blank		0.000	03:17
Air Blank		0.000	03:18
Subject Sample	# 2	0.194	03:19
Breath Volume		1.957	03:19
Air Blank		0.000	03:20
Control Test		0.078	03:20
Air Blank		0.000	03:20
Diagnostics Check		OK	03:20

Data Download Date/Time: 03/06/2018 09:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: WALDO,AARON
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 03:07
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:52

SUBJECT'S LAST NAME: WASHINGTON

DATE/TIME OF TEST: 02/25/2018 03:49:41

Diagnostics Check		OK	03:43
Air Blank		0.000	03:43
Control Test		0.078	03:44
Air Blank		0.000	03:44
Subject Sample	# 1	0.100	03:44
Breath Volume		1.960	03:44
Air Blank		0.000	03:45
Air Blank		0.000	03:47
Subject Sample	# 2	0.090	03:48
Breath Volume		1.207	03:48
Air Blank		0.000	03:48
Control Test		0.079	03:49
Air Blank		0.000	03:49
Diagnostics Check		OK	03:49

Data Download Date/Time: 03/06/2018 09:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001653

INSTRUMENT REGISTERED TO: ST PETERSBURG PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: ZOLLOTA,VLADIMIR,P
BREATH TEST OPERATOR AGENCY: St. Petersburg PD
ARRESTING OFFICER: CRUM,ZACKARY
ARRESTING OFFICER AGENCY: St. Petersburg PD
LAST AGENCY INSPECTION: 02/05/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 26316080A2
DRY GAS EXPIRY DATE: 11/05/2018
VIOLATION CODE: DUI
ARREST TIME: 19:01
GENDER: Male
OBSERVATION PERIOD BEGAN: 18:55

SUBJECT'S LAST NAME: FISH

DATE/TIME OF TEST: 02/28/2018 19:34:19

Diagnostics Check		OK	19:29	
Air Blank		0.000	19:30	
Control Test		0.078	19:30	
Air Blank		0.000	19:31	
Subject Sample	# 1		19:32	Subject Test Refused
Breath Volume		0.000	19:32	
Air Blank		0.000	19:33	
Control Test		0.079	19:33	
Air Blank		0.000	19:34	
Diagnostics Check		OK	19:34	

Data Download Date/Time: 03/06/2018 09:05

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPROUSE,JOHN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: WEINTRAUB,ANDREW
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 05:28
GENDER: Male
OBSERVATION PERIOD BEGAN: 05:49

SUBJECT'S LAST NAME: MARTIN

DATE/TIME OF TEST: 02/03/2018 06:56:41

Diagnostics Check		OK	06:47	
Air Blank		0.000	06:47	
Control Test		0.081	06:48	
Air Blank		0.000	06:48	
Subject Sample	# 1		06:51	No Sample Provided
Breath Volume		0.000	06:51	
Air Blank		0.000	06:52	
Air Blank		0.000	06:54	
Subject Sample	# 2		06:55	Subject Test Refused
Breath Volume		0.000	06:55	
Air Blank		0.000	06:55	
Control Test		0.082	06:56	
Air Blank		0.000	06:56	
Diagnostics Check		OK	06:56	

Data Download Date/Time: 02/13/2018 15:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: DASILVA,ANDRE,F
BREATH TEST OPERATOR AGENCY: FHP Troop G
ARRESTING OFFICER: DASILVA,ANDRE
ARRESTING OFFICER AGENCY: FHP Troop G
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 22:33
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:40

SUBJECT'S LAST NAME: ROE

DATE/TIME OF TEST: 02/05/2018 00:08:46

Diagnostics Check		OK	00:01
Air Blank		0.000	00:01
Control Test		0.082	00:01
Air Blank		0.000	00:02
Subject Sample	# 1	0.232	00:03
Breath Volume		1.390	00:03
Air Blank		0.000	00:03
Air Blank		0.000	00:05
Subject Sample	# 2	0.236	00:07
Breath Volume		2.273	00:07
Air Blank		0.000	00:07
Control Test		0.082	00:08
Air Blank		0.000	00:08
Diagnostics Check		OK	00:08

Data Download Date/Time: 02/13/2018 15:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,ERIN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CARSON,WILLIE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: RENNA

DATE/TIME OF TEST: 02/05/2018 08:40:31

Diagnostics Check	OK	08:39	
Air Blank	0.000	08:39	
Control Test	0.082	08:39	
Air Blank		08:40	RFI Detect
Air Blank	0.000	08:40	

Data Download Date/Time: 02/13/2018 15:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: WINTERS,ERIN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: CARSON,WILLIE
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 01/31/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 07:14
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:18

SUBJECT'S LAST NAME: R500179731000

DATE/TIME OF TEST: 02/05/2018 08:43:30

Diagnostics Check	OK	08:42	
Air Blank	0.000	08:42	
		08:43	RFI Detect
Air Blank	0.000	08:43	

Data Download Date/Time: 02/13/2018 15:39

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: SPROUSE,JOHN,S
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: SPROUSE,JOHN
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 07:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 08:20

SUBJECT'S LAST NAME: SAUNDERS

DATE/TIME OF TEST: 02/21/2018 08:45:04

Diagnostics Check	OK	08:44	
Air Blank		08:44	RFI Detect
Air Blank		08:45	RFI Detect

Data Download Date/Time: 03/12/2018 16:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001655

INSTRUMENT REGISTERED TO: ST JOHNS COUNTY SO
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LEDOUX,LAUREN,M
BREATH TEST OPERATOR AGENCY: St. John's County SO
ARRESTING OFFICER: MONIE,JARED
ARRESTING OFFICER AGENCY: St. John's County SO
LAST AGENCY INSPECTION: 02/13/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 640976
DRY GAS EXPIRY DATE: 11/21/2019
VIOLATION CODE: DUI
ARREST TIME: 22:39
GENDER: Male
OBSERVATION PERIOD BEGAN: 23:32

SUBJECT'S LAST NAME: STODDART

DATE/TIME OF TEST: 02/26/2018 00:12:29

Diagnostics Check		OK	00:02
Air Blank		0.000	00:03
Control Test		0.082	00:03
Air Blank		0.000	00:04
Subject Sample	# 1	0.350	00:04
Breath Volume		4.824	00:04
Air Blank		0.000	00:05
Air Blank		0.000	00:06
Subject Sample	# 2	0.378	00:07
Breath Volume		4.136	00:07
Air Blank		0.000	00:08
Air Blank		0.000	00:10
Subject Sample	# 3	0.368	00:10
Breath Volume		5.140	00:10
Air Blank		0.000	00:11
Control Test		0.079	00:11
Air Blank		0.000	00:12
Diagnostics Check		OK	00:12

Data Download Date/Time: 03/12/2018 16:06

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: PINEDA,MIGUEL
ARRESTING OFFICER AGENCY: North Miami PD
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 03:37
GENDER: Female
OBSERVATION PERIOD BEGAN: 03:46

SUBJECT'S LAST NAME: INSUA-EVANS

DATE/TIME OF TEST: 02/03/2018 04:15:30

Diagnostics Check		OK	04:08
Air Blank		0.000	04:08
Control Test		0.082	04:09
Air Blank		0.000	04:09
Subject Sample	# 1	0.203	04:10
Breath Volume		2.472	04:10
Air Blank		0.000	04:11
Air Blank		0.000	04:13
Subject Sample	# 2	0.197	04:13
Breath Volume		1.949	04:13
Air Blank		0.000	04:14
Control Test		0.081	04:14
Air Blank		0.000	04:15
Diagnostics Check		OK	04:15

Data Download Date/Time: 02/13/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001656

INSTRUMENT REGISTERED TO: NORTH MIAMI PD
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: PINEDA,MIGUEL,A
BREATH TEST OPERATOR AGENCY: North Miami PD
ARRESTING OFFICER: DOMINGUEZ,ANTONIO
ARRESTING OFFICER AGENCY: OTHER State
LAST AGENCY INSPECTION: 01/25/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 32316080A3
DRY GAS EXPIRY DATE: 01/05/2019
VIOLATION CODE: DUI
ARREST TIME: 23:34
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:17

SUBJECT'S LAST NAME: LAGARDE

DATE/TIME OF TEST: 02/04/2018 00:46:15

Diagnostics Check		OK	00:39
Air Blank		0.000	00:40
Control Test		0.080	00:40
Air Blank		0.000	00:40
Subject Sample	# 1	0.201	00:41
Breath Volume		3.824	00:41
Air Blank		0.000	00:42
Air Blank		0.000	00:44
Subject Sample	# 2	0.203	00:44
Breath Volume		3.160	00:44
Air Blank		0.000	00:45
Control Test		0.081	00:45
Air Blank		0.000	00:46
Diagnostics Check		OK	00:46

Data Download Date/Time: 02/13/2018 12:01

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MALKOWSKI,JOSH
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: GONZALEZ VARGAS

DATE/TIME OF TEST: 02/22/2018 03:22:00

Diagnostics Check	OK	03:21	
Air Blank		03:21	RFI Detect
Air Blank	0.000	03:22	

Data Download Date/Time: 03/26/2018 22:08

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: MALKOWSKI,JOSH
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 02:40
GENDER: Male
OBSERVATION PERIOD BEGAN: 02:58

SUBJECT'S LAST NAME: GONZALEZ VARGAS

DATE/TIME OF TEST: 02/22/2018 03:32:52

Diagnostics Check		OK	03:23
Air Blank		0.000	03:24
Control Test		0.079	03:24
Air Blank		0.000	03:24
Subject Sample	# 1	0.195	03:28
Breath Volume		2.304	03:28
Air Blank		0.000	03:28
Air Blank		0.000	03:30
Subject Sample	# 2	0.179	03:31
Breath Volume		1.484	03:31
Air Blank		0.000	03:31
Control Test		0.076	03:32
Air Blank		0.000	03:32
Diagnostics Check		OK	03:32

Data Download Date/Time: 03/26/2018 22:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.

Florida Department of Law Enforcement

Alcohol Testing Program

Subject Test Electronic Data

CMI, Inc. Intoxilyzer 8000 Serial Number: 80-001715

INSTRUMENT REGISTERED TO: OSCEOLA COUNTY S.O.
SOFTWARE VERSION: 8100.27
BREATH TEST OPERATOR: LYONS,DANIEL,E
BREATH TEST OPERATOR AGENCY: Osceola County SO
ARRESTING OFFICER: DOUVRES,MICHAEL
ARRESTING OFFICER AGENCY: Osceola County SO
LAST AGENCY INSPECTION: 02/20/2018

Date or Time Changed: No
DRY GAS STANDARD LOT #: 163437
DRY GAS EXPIRY DATE: 10/30/2018
VIOLATION CODE: DUI
ARREST TIME: 23:36
GENDER: Male
OBSERVATION PERIOD BEGAN: 00:10

SUBJECT'S LAST NAME: CARCANO

DATE/TIME OF TEST: 02/23/2018 00:36:59

Diagnostics Check		OK	00:30
Air Blank		0.000	00:30
Control Test		0.080	00:31
Air Blank		0.000	00:31
Subject Sample	# 1	0.172	00:32
Breath Volume		1.867	00:32
Air Blank		0.000	00:32
Air Blank		0.000	00:34
Subject Sample	# 2	0.178	00:35
Breath Volume		2.242	00:35
Air Blank		0.000	00:36
Control Test		0.079	00:36
Air Blank		0.000	00:36
Diagnostics Check		OK	00:36

Data Download Date/Time: 03/26/2018 22:08

Printed: 23 May 2018

COBRA ©1997-2014 CMI, Inc.